HIGH PRESSURE CALIBRATION WITH A NEW ABSOLUTE PRESSURE GAUGE

by

Arthur L. Ruoff, R. C. Lincoln and Y. C. Chen

January 1973

Cornell University

Ithaca, New York 14850

NOTICE

This report was prepared as an account of work sponsored by the United States Government. Neither the United States nor the United States Atomic Energy Commission, nor any of their employees, nor any of their contractors, subcontractors, or their employees, makes any warranty, express or implied, or assumes any legal liability or responsibility for the accuracy, completeness or usefulness of any information, apparatus, product or process disclosed, or represents that its use would not infringe privately owned rights.

Report #1962

Issued by

The Materials Science Center

MASTER

DISTRIBUTION OF THIS DOCUMENT IS UNLIMITED

DISCLAIMER

This report was prepared as an account of work sponsored by an agency of the United States Government. Neither the United States Government nor any agency Thereof, nor any of their employees, makes any warranty, express or implied, or assumes any legal liability or responsibility for the accuracy, completeness, or usefulness of any information, apparatus, product, or process disclosed, or represents that its use would not infringe privately owned rights. Reference herein to any specific commercial product, process, or service by trade name, trademark, manufacturer, or otherwise does not necessarily constitute or imply its endorsement, recommendation, or favoring by the United States Government or any agency thereof. The views and opinions of authors expressed herein do not necessarily state or reflect those of the United States Government or any agency thereof.

DISCLAIMER

Portions of this document may be illegible in electronic image products. Images are produced from the best available original document.

HIGH PRESSURE CALIBRATION WITH A NEW ABSOLUTE PRESSURE GAUGE

by

Arthur L. Ruoff, R. C. Lincoln^{*}, and Y. C. ChenDepartment of Materials Science and Engineering
Cornell University, Ithaca, New York

Abstract

Based on the simultaneous measurement of length and transit time of a specimen under pressure, we have developed a new way of determining absolute high pressure. The mercury freezing point at 0°C has been determined with this new method at 7571.2 ± 1.6 bars. Accurate high pressure calibration of secondary pressure gauge such as manganin gauge is possible with this method. The maximum deviation from linearity of the present gauge studied is 11.6 ± 1.1 bars between atmospheric pressure and the mercury freezing point at 0°C.

* present address: Sandia Corporation, Albuquerque, New Mexico

If the linear compressibility of a material is isotropic, it is easy to show that the pressure can be expressed in the form

$$P = P_1 - 3\rho_1 \ell_1^3 \int_{\ell_1}^{\ell} \left(\frac{1}{\tau_{\ell}^2} - \frac{4}{3\tau_s^2} \right) \left(\frac{1}{1+\Delta(P)} \right) \frac{d\ell}{\ell^2} \quad (1)$$

where

$$\Delta(P) = \frac{\beta^2 B^S T}{\rho C_p} \quad (2)$$

Here ρ_1 and ℓ_1 are the density and length at atmospheric pressure P_1 , τ_{ℓ} and τ_s are transit times for [111] longitudinal wave and [100] shear wave respectively, β is the volume thermal expansion coefficient, B^S is the adiabatic bulk modulus, T is the absolute temperature and C_p is the specific heat of the material. It is obvious then from equation (1), the pressure can be determined if length ℓ and transit times τ_{ℓ} and τ_s are measured at the same temperature and pressure.

We choose silicon as our specimen for the following two reasons: (1) because $\Delta(P)$ is quite small ($\Delta(P) \leq 0.0011$ for $0 \leq P \leq P_{\text{Hg}}$, P_{Hg} is the freezing pressure of mercury at 0°C) and moreover can be accurately evaluated, (2) because excellent single crystals of silicon can be grown and hence the material can be expected to behave reversibly.

The fractional length change measurement is accurate to 4×10^{-8} and utilizes a meter long specimen; it is described elsewhere.^(1,2) Briefly, end positions of specimen are defined by special H-shaped soft magnetic cores which are located with two linear variable differential transformers (LVDT) external to the pressure vessel by a null technique. The length change is measured by a Fabry Perot

interferometer with laser beams traveling in a vacuum bellows, between mirrors aligned with centers of LVDT on carriages. The entire system along with ultrasonic pressure vessel are immersed in a constant temperature both controlled to $2 \times 10^{-3}^{\circ}\text{C}$. The initial length ℓ_1 at atmospheric pressure is measured interferometrically against a standard length. The fractional error of the original length of the specimen ℓ_1 is 1×10^{-5} .

The silicon specimens used for ultrasonic measurements were made from single crystals of the same purity as that used in the length measurement. The transit times were measured with pulse superposition technique. The reproducibility of transit time measurement was within 2×10^{-5} .

After data for length measurement and ultrasonic transit time were collected, analysis was done by nonlinear least square analysis for fitting Murnaghan equation and Simpson's rule was utilized for integration to obtain P vs ℓ data. Meanwhile, we obtained simultaneously the calibrations for manganin gauge.

The mercury freezing point at 0°C , P_{Hg} , was then determined using a pressure arrest method (zero rate change of pressure versus time at transition point). It was determined to be 7571.2 ± 1.6 bars.

The quadratic dependence of pressure on the ratio of change of manganin resistance and initial resistance $(\frac{\Delta R}{R_1})$ was fitted. The maximum deviation of the gauge from linearity about midway between P_1 and P_{Hg} was determined to be $\Delta P_M = 11.6 \pm 1$ bars.

Table 1 tabulates the value of ΔP_M observed by different workers. Direct comparison of ΔP_M may not be meaningful since it

is conceivable that ΔP_M varies with the manganin sample. It is, however, also conceivable that ΔP_M is not very strongly dependent on the sample (assuming its purity and fabrication are nearly the same) as Bridgman proposed but rather dependent on the inaccurate measurement of pressure. At this point we do not believe the answer is known. To compare the relative reliability of the non-linearity or ΔP_M measured by our pressure gauge and that measured by the free piston gauge, let us assume a random error of 2 bars at each pressure for the free piston gauge. Thus the value of ΔP_M would be known to at best ± 1 bar. The random error of an individual reading in the present system is 0.04 bars or less. Consequently, ΔP_M can easily be measured to 0.05 bars. It is important to know ΔP_M accurately in many applications. For example, if the pressure derivative of the bulk modulus of iron is measured by length change, an error of 2 bars in ΔP_M would give an error of nearly 10% in B'_0 .¹

Values of the 0°C mercury freezing pressure are given in Table 2. It should be noted that the Committee on Fixed Points Near Room Temperature at the Symposium on Accurate Characterization of the High - Pressure Environment held at the National Bureau of Standards, Gaithersburg, Md. in 1968 designated 7569 ± 2 bars as the accepted value for the 0°C mercury freezing pressure.^(3,4)

Acknowledgement

We would like to thank Robert E. Terry for his technical assistance. We acknowledge the support of this work by the United States Atomic Energy Commission.

Table 1. Values of ΔP_M

<u>ΔP_M (bars)</u>	<u>Source</u>
-10	Harwood ⁵
-11	NBS ⁶
-21	Zeto and Vanfleet ⁷
-21	Boren, Babb and Scott ⁸
-11.6 $\pm$.1	Present Work

Table 2. The 0°C Mercury Freezing Pressure

<u>Pressure (bars)</u>	<u>Source</u>
7565.4 $\pm$ 3.7	Newhall, Abbot and Dunn ⁹
7569.2 $\pm$ 1.2	Dadson and Greig ¹⁰
7571.0 $\pm$ 1.2	Yasunami ¹¹
7571.2 $\pm$ 1.6	Present

References

1. R. C. Lincoln, Ph.D Thesis, Cornell University, Ithaca, New York (1971).
2. R. C. Lincoln and A. L. Ruoff, Rev. Sci. Instr. (to be submitted for publication).
3. Nat. Bur. Std., Spec. Pub. 326 (1971) p. 313.
4. E. C. Lloyd and C. W. Beckett, Science 164, 860 (1969).
5. D. H. Newhall, Harwood Engineering, Private Communication.
6. D. P. Johnson, National Bureau of Standards, Private Communication with R. C. Lincoln.
7. R. J. Zeto and H. B. Vanfleet, J. Appl. Phys. 40, 2227 (1969).
8. M. D. Boren, S. E. Babb, Jr., and G. J. Scott, Rev. Sci. Instr. 36, 1456 (1965).
9. D. H. Newhall, L. H. Abbot and R. A. Dunn, in High Pressure Measurement, ed. by A. A. Giardini and E. C. Lloyd, Butterworths, Washington (1963) p. 339.
10. R. S. Dadson and R. G. P. Greig, Brit. J. Appl. Phys. 16, 1711 (1965).
11. K. Yasunami, Proc. Jap. Acad. 43, 310 (1967).