21-1094

UCRL-50007-74-1

HAZARDS CONTROL PROGRESS REPORT NO. 48
JANUARY THROUGH JUNE 1974

MASTER

July 23, 1974

Prepared for U.S. Atomic Energy Commission under contract No. W-7405-Eng-48

**LAWRENCE
LIVERMORE
LABORATORY**

University of California/Livermore

NOTICE

"This report was prepared as an account of work sponsored by the United States Government. Neither the United States nor the United States Atomic Energy Commission, nor any of their employees, nor any of their contractors, subcontractors, or their employees, makes any warranty, express or implied, or assumes any legal liability or responsibility for the accuracy, completeness or usefulness of any information, apparatus, product or process disclosed, or represents that its use would not infringe privately-owned rights."

Printed in the United States of America

Available from

National Technical Information Service

U. S. Department of Commerce

5285 Port Royal Road

Springfield, Virginia 22151

Price: Printed Copy \$2.35; Microfiche \$1.45

* Pages	NTIS Selling Price
1-50	\$4.00
51-150	\$5.45
151-325	\$7.60
326-500	\$10.60
501-1000	\$13.60

TID-4500, UC-41
Health and Safety

LAWRENCE LIVERMORE LABORATORY
University of California, Livermore, California, 94550

UCRL-50007-74-1

HAZARDS CONTROL PROGRESS REPORT NO. 48
JANUARY THROUGH JUNE 1974

MS date: July 23, 1974

NOTICE

This report was prepared as an account of work sponsored by the United States Government. Neither the United States nor the United States Atomic Energy Commission, nor any of their employees, nor any of their contractors, subcontractors, or their employees, makes any warranty, express or implied, or assumes any legal liability or responsibility for the accuracy, completeness or usefulness of any information, apparatus, product or process disclosed, or represents that its use would not infringe privately owned rights.

MASTER

Foreword

This report is divided into three major sections. The first section, Progress Reports, covers the status of activities undertaken or continuing during the period; additional reports or separate publications will cover the final results of these activities. The second section, Technical Notes, contains reports on interesting activities of a more limited scope on which further reporting is not anticipated. The third section lists recent publications.

Readers who are interested in more detail regarding any item may contact the authors of the reports who are listed in the Contents.

Contents

PROGRESS REPORTS

Radiation Protection

- Comparison of Si(Li) and Twinstel Detectors for
Measuring Pu Lung Burdens (*W. L. Pickles*) 1

- Separation of HT, Noble Gases and HTO Vapor
with Semi-permeable Membranes (*R. H. Howell,*
J. L. Cate and C. M. Wong) 2

- Progress in Determining Neutron Dose Equivalent at the
Hazards Control Calibration Facility (*R. Griffith, R. Kloepping,*
H. W. Patterson and T. Crites) 8

Industrial Hygiene

- Effect of Concentration on Respirator Cartridge
Efficiency (*G. O. Nelson, C. A. Harder and B. E. Bigler*) 11

Instrument Development

- Use of Ultraviolet Light Transmission for Determining
Track Densities in Dielectric Materials (*R. Griffith*
and J. Fisher) 15

- Beta Spectrometer: Second Generation (*J. L. Cate*) 16

- A System to Produce Low Concentrations of Gas
in Air (*G. O. Nelson, R. D. Taylor and H. J. Johnson*) 18

- Cartridge and Canister Test Systems for the Testing
and Certification Laboratory (*G. O. Nelson, R. D. Taylor*
and G. E. Sisti) 19

- Tritium Surface Contamination Monitor (*P. E. Barry*) 19

- Neutron Spectrometer Development (*J. Elliott and*
D. Slaughter) 20

Nuclear Safety

- A Method for the Safe Handling of $^{238}\text{UH}_3$
Waste (*A. J. Toy*) 24

Decontamination

- Plant Test of New Liquid Decontamination
Process (*P. E. Barry*) 24

Other

- A Weighted Linear Least-Squares Program with
Error Estimates (*D. Slaughter*) 25

- Tritium Analysis of Alpha-Contaminated Oil (*R. H. Howell,*
C. M. Wong and P. E. Barry) 25

TECHNICAL NOTES

Industrial Hygiene

- A Computer Program To Calculate Ear Protector Noise
Attenuation (*H. J. McDermott*) 28

- A Low-Level Fluorine Detector (*J. E. Johnson*
and W. Eneidi) 28

- A Computer Program to Size and Balance Exhaust
Ventilation Systems (*A. J. Toy and H. J. McDermott*) 29

Instrument Development

- Contribution of Alpha Particle Tracks to Fission Fragment
Spark Counting Backgrounds (*R. Griffith and J. Fisher*) 30

Other

- DOSIHD, A Program to Decode PDP-11® Disk Operating System
(DOS)® Dectapes® (*D. E. Jones and R. L. Haggin*) 31

PUBLICATIONS	32
REFERENCES	34

Edited by V. R. Mendenhall

Progress Reports

RADIATION PROTECTION

Comparison of Si(Li) and Twixel Detectors for Measuring Pu Lung Burdens

A program is underway to improve our ability to measure small (less than 16 nCi) quantities of Pu in the human lung. In particular, we are evaluating various detectors in comparison with the 200-cm² dual twixel detectors currently in use at LLL. We recently rented two 500-mm² X 5-m Si(Li) detectors from a commercial manufacturer for this purpose.

Silicon detectors have long been preferable to germanium detectors for low-energy x rays. The escape peak in silicon does not complicate the spectrum, and thin dead layers are easily achievable. The primary disadvantage of silicon detectors has been

their small area. Recently, silicon detectors with active areas of 1000 mm² have become available. This is close enough to the 20 000-mm² (200-cm²) dual twixel detectors to make an evaluation feasible.

We measured the background from the whole-body counting room and detector, the response to a ²³⁸Pu-doped lung in a phantom, the response to an individual known to have plutonium in the lungs, and the response to a ²⁴¹Am source.

The data from this experiment have not been completely evaluated, but some qualitative results are listed below.

1. Both detectors have a small but acceptable thorium contamination (Fig. 1). No other contamination was observed.

Fig. 1. Spectra obtained with one 500-mm² X 5-mm Si(Li) detector.

- A. Overnight background (55 580 sec) Det A.
- B. Overnight spectrum of 200 nCi lung in phantom using Det A.
- C. Overnight spectrum of ²³⁸Pu-doped lung (200 nCi) in phantom using Det B.
- D. Background-subtracted ²³⁸Pu spectrum (55 580 sec) i.e. (spectrum C) - (spectrum A).

2. The resolution was excellent, allowing clear separation of the L_{α} , L_{β} , L_{γ} x-ray complexes.
3. The cosmic ray background is completely eliminated from the spectra, and appear as overload pulses (Fig. 1).
4. The Compton scattering from the 60-keV line in ^{241}Am does not interfere with the L x-ray region of the spectrum (Fig. 2).
5. The background from the subject may be obtained directly from the spectra of the subject, contrasted with a twixtel system in which the background must be measured at a different time. This is possible because of the high resolution of the silicon detectors.
6. The count rate was very sensitive to the position of the silicon detectors because of their small size. This is a major limitation which can only be solved by having a detector whose area is comparable to the size of the patient's chest.
7. The minimum detectable activity for one Si(Li) detector is about 0.35 of that for one of the twixtel detectors. The gain in resolution almost offsets the loss due to the small size (Fig. 3).

The determination of minimum detectable activity is relatively subtle because of the complicated background subtraction methods used on the twixtel. When this analysis is completed, a more definitive report will be made on the usefulness of Si(Li) detectors for "Pu lung" counting.

Separation of HT, Noble Gases and HTO Vapor with Semi-permeable Membranes

The maximum permissible concentration for a 40-hr week (MPC) of tritium gas (HT) in air is 2000 nCi/l.

The MPC for tritium oxide (HTO) vapor in air is 5 nCi/l. The MPC for HTO is a factor of 400 lower than the MPC for HT. Many air monitoring techniques, such as air ionization chambers, do not differentiate among the chemical forms of the tritium measured. They respond to the total radioactivity contained in the sample. Due to the large difference in MPC's monitors and survey meters which have a higher response to HTO than to HT will reflect a more accurate assessment of the exposure.

A general class of sampling techniques which do discriminate between HTO and HT is based on collecting the HTO vapor on or in a hygroscopic medium such as a cold finger or silica gel. The measurements of HTO are later made on the collected sample. While these techniques are accurate and sensitive, they are not suitable for use as monitors because of the delayed response.

One attempt to separately measure HTO in the presence of radioactive noble gases uses a water pickup and a flow through liquid scintillation counter.¹ The HTO preferentially dissolves in the water which is counted in a scintillator mix with a short sample to analysis time. An enhancement of the ratio of HTO response to the response of radioactive noble gases has been obtained by this method.

A higher response to HTO than to HT was noted in a proportional counter covered with a thin, aluminized Formvar window, designed to allow penetration of the (18-keV max, 5.5-keV average energy) tritium decay beta.² Observation of that effect has prompted an investigation of semipermeable membranes, some of which have higher permeabilities for water vapor than for most diatomic or noble gases. These membranes are silicon compounds, and are designed to selectively pass various molecules. Two

Fig. 2. Pulse-height spectrum of a ^{241}Am source taken with one of the 500-mm² $\times$ 5-mm Si(Li) detectors. Note the excellent resolution (220 eV at 5.5 keV) and lack of Compton scattering background under the L x-ray lines.

Fig. 3. Direct comparison of one 500-mm² X 5-mm Si(Li) detector with one twixtel detector. Both spectra shown were taken for 55 530 sec on the same phantom with the 203-nCi ²¹³Pu lung in place. Both spectra are plotted with the same abscissa and ordinate scales. Apparent relative minimum detectable amounts is 0.35.

membranes which have been explored in detail are di-methyl silicone and a silicone polymer available from Purma Pure Company.* The permeation properties of Formvar films have also been investigated.

HTO Enhancement. The basic mechanism of the permeation effect is illustrated in Fig. 4. A gas at partial pressure P_1 on one side of the membrane and partial pressure P_2 on the other side of the membrane will permeate through the membrane until P_1 and P_2 become equal. The expression governing this behavior is

$$\frac{dP_1}{dt} = \frac{KA}{t} (P_1 - P_2)$$

where A is the exposed area of the membrane, t is the membrane thickness, t is time, and K is the permeation constant. Permeation constants for a number of species in dimethyl silicone membranes are

*Reference to a company or product name does not imply approval or recommendation of the product by the University of California or the U.S. Atomic Energy Commission to the exclusion of others that may be suitable.

shown in Table 1. The relative rates for H_2 (55), H_2O (3000), Ar (50) and Xe (171) indicate that H_2O permeates the membrane much faster than these other species. An enhancement in the concentration of vapor H_2O over these other species is gained in a clean dry flowing carrier gas which is separated from the original sample by the membrane.

If the sample gas flows on side 1 of the membrane with flow rate f_1 and the clean carrier gas flows on side 2 with rate f_2 , the ratio of concentration of a species s in the sample gas to that species s in the carrier gas at equilibrium is

$$\frac{C_1^s}{C_2^s} = \left(\frac{1}{f_1} + \frac{1}{R^s} \right) f_2 + 1,$$

where R^s is proportional to KA/T for species s . From this equation it is clear that the relative enhancement in concentration of HTO (R large) to HT and others (R small) will depend on the flow rates past the membrane. For low transfer rates across the membrane, the ratio of C_1/C_2 will be large and depend strongly on f_2 . For transfer rates which are high, C_1/C_2 will be smaller and have much less flow dependence.

Fig. 4. Diagram of flow in two chambers separated by a semipermeable membrane. The membrane exposed area is 10 cm² and the volume of each chamber is 10 ml.

Table 1. Gas permeabilities in dimethyl
silicone membrane

Gas	Dimethyl silicone	Gas	Dimethyl silicone
H	55	n-C ₄ H ₁₀	1670
He	30	n-C ₅ H ₁₂	785
NH ₃	500	n-C ₆ H ₁₄	715
H ₂ O	3000	n-C ₇ H ₁₆	360
CO	30	HCHO	925
N ₂	25	CH ₃ OH	1160
NO	50	COCl ₂	1250
O ₂	53	Acetone	490
H ₂ S	840	Pyridine	1595
Ar	50	Benzene	900
CO ₂	270	Phenol	1750
N ₂ O	365	Toluene	760
NO ₂	635	Xe	171
SO ₂	1250	CCl ₄	5835
CS ₂	7500	CH ₃ O	925
CH ₄	80	Freon 11	1790
C ₂ H ₆	210	Freon 12	107
C ₂ H ₄	115	Freon 22	382
C ₂ H ₂	2200	Freon 114	211
C ₃ H ₈	340	Freon 115	51
n-C ₈ H ₁₈	750		

Data of relative concentration as a function of flow rate have been taken for a 1-mil-thick film of dimethyl silicone, a 1-mil-thick film of the silicon polymer and Formvar films of varying thickness. The gases HT, ⁸⁵Kr and ¹³³Xe were diluted by air, propane or other gas by a known amount and passed through side 1 of the chamber shown in Fig. 4. Counting gas, either propane, methane or P10, flowed by side 2 in the chamber and then through a gold-plated proportional counter. The response in the proportional counter at equilibrium measured the relative concentration of each of the radioactive species studied. The data, using propane as the counting gas, are presented in Figs. 5, 6 and 7.

In each of these data sets the HT and ⁸⁵Kr permeation rates are low so that the ratio of C₁/C₂ is large. The ratio for HTO which has a high transfer rate is close to the value 1. In Fig. 8 the ratio for HTO is presented on an expanded scale.

The data for dimethyl silicone and the silicone polymer are reasonably consistent with the derived equation. The results for Formvar tended to be erratic, and were not as well described by the derived expression. The Formvar data in Fig. 7 is the best-behaved of the Formvar data obtained.

The relative enhancement of HTO over the other species may be obtained from these data by dividing

Fig. 5. Relative concentration of gases as a function of counting gas flow rate for dimethyl silicone membrane. High values and a steep slope indicate low permeabilities.

Fig. 7. Relative concentration of gases as a function of counting gas flow rate for a Formvar film. The Kr data do not follow a linear response.

Fig. 6. Relative concentration of gases as a function of counting gas flow rate for the silicone polymer membrane.

Fig. 8. Relative concentration of HTO on an expanded scale. The value of the intercept at zero propane flow should be at 1.

relative concentration of the other species by the HTO relative concentration. These values are shown in Table 2 for a flow rate on the detector side of 250 cm³/min propane and 250 cm³/min sample gas. These relative enhancements of HTO over HT are achieved with an HTO sensitivity of one-third to one-fourth the sensitivity of a direct measurement. Since this loss in sensitivity is not great, more than one membrane stage can be used to achieve even greater enhancement of HTO over other species.

In a two-stage dimethyl silicone membrane system, using the enhanced output of the first stage as an input to the second stage, we obtained final enhancements of 1881 HTO to HT and 2442 HTO to Kr. The

Table 2. Enhancement of HTO over other gases

	HT	Kr
Dimethyl silicone	51	44
Silicone polymer	60	54
Formvar	70	226

All flows were 250 cm³/min.

enhancements are approximately the product of two single-stage enhancements. The flow rates used to obtain these values are: $f_1 = 1000 \text{ cm}^3/\text{min}$ and $f_2 = 250 \text{ cm}^3/\text{min}$ in stage 1, and $f_1 = 250 \text{ cm}^3/\text{min}$ and $f_2 = 250 \text{ cm}^3/\text{min}$ in stage 2. The loss in HTO sensitivity at the final stage of this system is 1/20. Thus very high enhancements of HTO over other species are possible without a heavy loss in sensitivity to HTO.

Time Response and Sensitivity. The time response of a typical measurement of HTO passed through a single membrane is shown in Fig. 9. The detector volume is 250 cm^3 and the flow rate is $250 \text{ cm}^3/\text{min}$. The time response of the system is dominated by the time required for the gas in the detector to mix to equilibrium concentration. In general, the gas-mixing time constants will be much greater than the transfer time across the membrane. If a faster response is needed, either a smaller detector volume or higher flow can be used.

The sensitivity of our bench top monitor can be seen in Fig. 9. The concentration of HTO in the sample is 100 nCi/l which is 20 MPC's . The response in this counter with no background compensation is 20 times background. This represents a minimum detectable concentration of 0.46 MPC's at the 95% confidence level in 10 sec counting time. Higher sensitivity can be obtained by using a separate detector to measure the background which is then subtracted. Electronics packages already exist to perform this operation. In a multi-membrane system, detectors could follow each membrane so that readouts having both high sensitivity with lower enhancement and

lower sensitivity and higher enhancement are available in the same instrument.

Interferences and other Considerations. While water has a high permeability the data in Table 1 show several other molecules which have high rates. It is clear that in special environments where many different gases may contain tritium, the selectivity will not enhance HTO only, but rather many other similar species as well. This is not necessarily a difficulty as some of the other enhanced species are water soluble and also represent an increased hazard over non-soluble species.

In Fig. 6 the time required to clear the system of HTO when the source was stopped is comparable to the time required to reach equilibrium. Again, smaller volume detector and/or higher flow rates will decrease the time constant. Further, the counting level returned to the background level, indicating that there is not significant retention of HTO in the membrane or other parts of the system.

A monitor or survey meter must have minimum sensitivity to atmospheric humidity. In a series of measurements, varying the humidity from dry air (bottled gas) to saturated air, no dependence in gas transfer rates on relative humidity was observed.

Coatings, dust and films deposited on the membrane surface degrade the performance of the membrane. Both the overall transfer rate and differential rate is lower for a dirty membrane. Thus an operating monitor will require a good pre-filter on the sample in order to maintain its characteristics. This also means that a detector with thin windows made of these membranes with an aluminum coating as described by

Fig. 9. Typical time response of the membrane detector system to a HTO vapor pulse flowing by the membrane at $250 \text{ cm}^3/\text{min}$. The counting gas flow was $250 \text{ cm}^3/\text{min}$ and the detector volume 250 cm^3 . The time response is dominated by the mixing time of the counting gas in the detector. The return to original counting levels indicates that no tritium is retained in the membrane or detector.

Block *et al.*² will exhibit a much lower enhancement of HTO over HT and other species than if the membrane is uncoated. This excludes the membrane from use as a detector window since the coating is necessary for the membrane to conduct electricity. Further, with the membrane separated from the detector, shielding can be used to reduce the external background contribution as well as reducing the contribution from unwanted radioactive species within the detector.

Applications. It is clear that semipermeable membranes can be used to enhance the concentration of HTO over HT and other gases such as radioactive AR, KR and Xe. The sensitivity of measurement and the degree of enhancement can be manipulated to provide characteristics desirable in such instruments as area monitors, stack monitors, or portable survey meters. Further, most other features found in existing systems (such as background compensation, remote, multiple input samplers and automatic alarms) can be provided.

One excellent application might be as an air monitor for a water-moderated reactor, where effluent may contain appreciable quantities of HT, HTO, and ^4Ar . An instrument such as shown in Fig. 10, a two-stage, background-compensated system with four detectors, would do the following:

- Detector 1, measuring the raw effluent, would give total activity.
- Detector 2, after one stage of HTO enhancement, would give high sensitivity and quick response to large amounts of HTO.
- Detector 3, after two enhancement stages would have lower sensitivity and response times, but would measure HTO in the presence of large quantities of the other species. Finally,
- Detector 4, would measure only external radiation, and would provide input for the background compensation circuitry.

Acknowledgments. The authors wish to thank Dr. Lawrence Hrubesh for his recommendations and for

Fig. 10. Schematic of one possible detector/membrane arrangement for an air-monitoring system.

giving us the membrane material used in these experiments.

Progress in Determining Neutron Dose Equivalent at the Hazards Control Calibration Facility

The Hazards Control Calibration Facility is a 12.2 X 9.14 X 7.32-m enclosure shielded by 0.61 m of concrete on five sides. The sixth side is a concrete floor on an earth foundation. The entrance to the cell is a concrete door 1.42 X 0.83 m and 0.46 m thick. There is a false floor made of aluminum grating and supported by aluminum I-beams 2.74 m above the concrete floor. The neutron sources we used were pneumatically fired into the cell to a position 1 m above the aluminum grating, 4 m from the nearest side wall, and 6 m from the nearest end wall. The detectors were located in the same plane as the source.

The Hazards Control Calibration Facility is being upgraded to provide more intense neutron fields with more varied spectra for use in personnel neutron dosimetry development. In addition to an existing source of $^{238}\text{PuBe}$ with a yield of $7.9 \times 10^7 \text{ nsec}^{-1}$, we have recently acquired a 1.6-mg source of ^{252}Cf with a yield of $3.9 \times 10^9 \text{ nsec}^{-1}$ and a neutron generator capable of producing $2.0 \times 10^{10} \text{ nsec}^{-1}$ from the $^3\text{H}(\text{D},\text{n})^4\text{He}$ reaction or 10^8 nsec^{-1} from the $^2\text{H}(\text{D},\text{n})^3\text{He}$ reaction. To provide additional spectra, spherical moderators were made which can be placed over the ends of the pneumatic delivery tube. Thus, the $^{238}\text{PuBe}$ or ^{252}Cf neutrons can be moderated or scattered to different degrees. The spheres available are 5-, 10-, 15- and 25-cm-radius D_2O filled shells; 2-, 5-, and 10-cm-radius spheres of polyethylene, and a 20-cm-radius sphere of aluminum.

The usefulness of these sources and spheres for personnel dosimeter development and calibration depends on our ability to determine the neutron spectrum and the resulting dose equivalent at a given point with reasonable accuracy. There are at least three ways to estimate neutron dose equivalent (D.E.).

1. Calculate D.E. from spectrum and intensity measurements using fluence to D.E. conversion data, or
2. measure absorbed dose and determine quality factor from LET,³ or
3. use a moderated detector whose response is designed to follow the D.E. response curve.⁴

For this work we have chosen to compare D.E. estimates made by (1) calculation from computer-generated spectra using the ANISN Neutron Transport Code⁵; (2) calculation from spectra obtained using multisphere detector⁶ data processed with the LOUHI unfolding code,⁷ and

(3) measurements made with an Anderson-Braun rem counter. In the future, we plan to use data obtained with other spectrometry techniques.⁸ Comparison of the D.E. obtained by each technique have been made at 1 m from the center of the ^{252}Cf and $^{238}\text{PuBe}$ sources in each of the moderators. A comparison will also be made at 2 m for the ^{252}Cf source. The $^{238}\text{PuBe}$ source will not be used at 2 m because of the low source intensity.

Previous reports have dealt with our interest in multisphere spectrometry.^{9,10} For these measurements we used an active detector of ^6LiI . The crystal is 0.4 cm thick by 0.8 cm in diameter, and is connected to a photomultiplier tube by a 15-cm light pipe. The detector is used bare as well as in a 0.05-cm-thick cadmium cover and in 5.08-, 7.62-, 12.7-, 20.3-, 25.4-, 30.5-, and 45.7-cm-diameter spheres of polyethylene. We have had 0.5- and 1.0-cm shields of boron-loaded plastic made for use over the crystal to provide additional information about the intermediate energy neutrons, but the response functions for the detector in these shields are not yet available.

The multisphere measurements were made by placing the ^6LiI crystal at the prescribed distance from the source point and then taking counts on a laboratory scaler for each of the spheres. Response functions were obtained by interpolation of data from Sanna¹¹ for 0.4-g X 0.4-cm and 0.8- X 0.8-cm detectors using the average chord length of our crystal. The data were then processed using the LOUHI computer code. After LOUHI converges to a solution spectrum using the response functions and scaler response as input, the resulting spectrum is used to calculate D.E. as well as Kerma, total fluence and average energy. In addition, our version of LOUHI provides tabulation of differential spectrum, integral spectrum, integral D.E. and integral Kerma. Table 3 and Fig. 11 show the LOUHI output for one of the Calibration Facility measurements.

The ANISN code, originally developed at Oak Ridge National Laboratory, has been available for use at LLL in various versions since 1968. ANISN-L, the current version, like ANISN, solves the one-dimensional, multi-group, time-independent Boltzman transport equation by the method of discrete ordinates. ANISN-L is the creation of Thomas P. Wilcox of L Division, and can be used to solve problems involving fissionable systems, neutron-induced radioactivity, and shielding of neutron and gamma sources. Wilcox has prepared a series of internal laboratory notes which describe the code and its options, and these notes are available from the author. (For example, see Ref. 4.)

For our calculations, we only needed to specify the location of the source, its spectrum, the material surrounding it and its location (i.e. air, aluminum or

stainless steel, D₂O or polyethylene, and concrete), neutron flux-to-dose conversion factors, and the detector location. Our flux-to-dose conversion factors were obtained from ICRP Publication 21.¹² Figure 12 gives the ANISN-L neutron spectrum at 2 m from a bare ²⁵²Cf source of intensity 1 n/sec, and may be compared with Fig. 1. Agreement is seen to be good, except for the lowest energy ANISN neutron group. The width of the group obscures the true spectrum shape here. Although the code uses spherical geometry, we believe the dimensions of the Calibration Facility are such as to provide a suitably close approximation to it.

Table 4 shows a comparison of the neutron D.E. measurements made to date. For the ²³⁸PuBe source, the agreement is satisfactory except for the aluminum sphere which will be redone. The 25 cm D₂O sphere data have not been completed. In the case of the ²⁵²Cf measurements, the agreement is not as good, and we suspect that the multisphere measurements are inaccurate because of some of the high count rates we experienced ($> 1.5 \times 10^4$ counts/sec).

We now plan to make measurements on the ²⁵²Cf source at 2 m where the multisphere system should be in better shape because of reduced count rates. The data will also be expanded when the additional

Fig. 12. Bare Cf in calibration cdl, average energy-dependent fluence for Zone 10.

spectrometry becomes available. And finally, the data from the "active" multisphere system will be augmented by data from passive detectors, including gold foils and etch track fission detectors.

Table 4. Summary of dose equivalent rate estimate at 1 m from Calibration Facility neutron source-moderator

Facility neutron source-moderator			
Source	ANISN	Configurations	
		Multisphere spectrometer	Rem meter
²³⁸ PuBe (mrem · hr ⁻¹)			
Bare	90	82	83
5 cm poly	58	56	57
10 cm poly	32	32	31
5 cm D ₂ O	66	71	66
10 cm D ₂ O	45	49	46
15 cm D ₂ O	29	35	34
25 cm D ₂ O	14	—	20
20 cm al.	72	54	80
²⁵² Cf ^a			
Bare	4009	3020	3802
5 cm poly	2069	1620	350
10 cm poly	886	720	695
5 .. D ₂ O	2559	2080	2740
10 cm D ₂ O	1491	1340	1686
15 cm D ₂ O	906	890	1280
25 cm D ₂ O	388	—	413
20 cm al.	3251	2130	3717

^aAs of April 1, 1974. Source strength 3.9225×10^7 n/sec and all other measurements for ²⁵²Cf are normalized to this date.

Effect of Concentration on Respirator Cartridge Efficiency

The purpose of this present investigation is to determine how the service life varies as a function of the solvent vapor concentration. Although studies of this nature have been undertaken previously,¹³ little is known about the contributing effects of high relative humidities. This investigation will show how the breakthrough times of nine common solvent vapors and one gas vary with concentration, what empirical expressions can be used to describe their behavior, and what adsorption similarities exist between solvents of different types. Observed times are also compared to those calculated from the adsorption isotherm and Mecklenburg equation.

Test conditions include assaulting a pair of cartridges with concentrations which typically varied from 100 to 3000 ppm at 53.3 l/min and a relative humidity of 50%. Concentrations downstream from the cartridge were continuously monitored with a flame ionization detector. Additional tests demonstrated how the breakthrough time at various concentrations changed at flowrates ranging from 20.6 to 71.4 l/min and at humidities ranging from dry (< 2% RH) to 80%.

Results.

Effect of Concentration on the Breakthrough Time. Figure 13 shows a typical family of breakthrough curves for methyl chloroform (CH_3CCl_3), methyl acetate, acetone, and carbon tetrachloride at various concentrations. Additional data not shown were also generated for hexane, diethylamine, isopropyl alcohol and dichloromethane vapors. In all cases the curves exhibited similar behavior in that they were asymmetrically "s" shaped and showed longer breakthrough times at lower concentrations.

These curves can, however, be shown graphically in a somewhat different manner. If the breakthrough times at various concentrations at different breakthrough percents are plotted on logarithmic coordinates, a family of straight lines emerges, as shown in Fig. 14. Hence breakthrough times at lower concentrations at any given breakthrough percent can be extrapolated. Mathematically, these data take the form

$$t_b = ac^b$$

Fig. 13. Breakthrough curves for MSA cartridges at various concentrations of methyl chloroform.

Fig. 14. Effect of concentration on the breakthrough time of methyl chloroform at various breakthrough percents.

where t_b is the breakthrough time, C is the concentration, and a and b are constants for a given set of experimental conditions. Note for each data set the family of lines is approximately parallel, indicating

Fig. 15. The 10% breakthrough curves for acetone, hexane and methyl chloroform. (1) indicates coconut base carbon and (2) indicates petroleum base carbon.

that the slopes, but not the intercepts, are essentially constant. From these data we can conclude that once the slope of one breakthrough time-concentration relationship is known, the slope of the others can be approximated.

A composite of the 10% breakthrough times are shown in Figs. 15 and 16 for all the concentration groups run. The values of a and b , obtained from a first-degree, least-squares fit are shown in Table 5. Although the a values (the intercept at $C = 1$ ppm) appear to be unrelated, the slopes, b , appear to be similar. The average slope is -0.65 with a standard deviation of 0.19 for the nine solvent vapors tested.

Effect of Flowrate on the Breakthrough Time at Various Concentrations. Theory and experiment previously discussed¹⁴ shows the breakthrough to be inversely proportional to flowrate at 1000 ppm. The question often arises, however, if this also applies to the lower vapor concentrations normally encountered in the industrial environment.

Figures 17 and 18 show the 10% breakthrough time as a function of benzene and acetone concentrations for three flowrates. Note the following: (1) The linearity of the logarithmic breakthrough time as a function of concentration relationship; (2) the parallelism of the three flowrates, indicating approximately constant slopes; and (3) the ratios of the breakthrough times at a given concentration are inversely

Fig. 16. The 10% breakthrough curves for isopropanol, carbon tetrachloride, benzene, methyl acetate, diethylamine and dichloromethane.

Table 5. Tabulation of constants a and b from Eq. (1)

Solvent	Cartridge ^a	a	b
Acetone	AO ^c	1,040	-0.452
Acetone	AO ^p	4,460	-0.621
Hexane	MSA ^c	39,200	-0.937
Hexane	AO ^p	38,900	-0.908
Carbon tetrachloride	MSA ^c	8,040	-0.671
Benzene	MSA ^c	15,600	-0.743
Methyl Acetate	MSA ^c	1,840	-0.542
Diethylamine	MSA ^c	5,360	-0.729
Isopropanol	MSA ^c	1,470	-0.417
Methyl chloroform	MSA ^c	13,000	-0.744
Dichloromethane	MSA ^c	353	-0.395
		Av	11,700 ± 14,000
			-0.65 ± 0.19

^ac = coconut base carbon.

p = petroleum base carbon.

Fig. 17. The effect of concentration on the 10% breakthrough time of benzene at various flowrates.

Fig. 18. The effect of concentration on the 10% breakthrough time of acetone at various flowrates.

proportional to flowrate over the concentration range of 50 to 2000 ppm.

Effect of Humidity on the Breakthrough Time at Various Concentrations. Much research concerning the adsorption of vapors on activated carbon has been done either in dry atmospheres or at one specified relative humidity. Since water vapor is generally present in concentrations far exceeding the solvent vapor level, the corrections for the ambient humidity must be taken into account.

Figures 19 and 20 demonstrate how the 10% breakthrough time varies as a function of concentration at various humidities. Tests were run at concentrations ranging from 100 to 2000 ppm for benzene and methyl chloroform. In each case as the humidity is increased, the breakthrough-concentration line is depressed — sometimes subtly, sometimes dramatically — in a downward direction. In general, the biggest changes are noticed when the relative humidity is increased from 50 to 80%.

In addition, Figs. 18 and 19 show another interesting feature. As the concentration decreases, the

Fig. 19. The effect of concentration on the 10% breakthrough time of methyl chloroform at various relative humidities.

Fig. 20. The effect of concentration on the 10% breakthrough time of benzene at various relative humidities.

breakthrough-concentration lines begin to spread, indicating at lower concentrations the breakthrough times are more seriously decreased by the higher humidities.

Fig. 21. Breakthrough percent of vinyl chloride as a function of time using MSA cartridges.

Fig. 22. Breakthrough percent of vinyl chloride as a function of time using Acme cartridges.

The reason for this can be explained in part by the competition of the solvent and water vapor for available adsorption sites on the activated carbon. If the solvent vapor concentration is 100 ppm at 80% relative humidity, the ratio of water to solvent vapor is approximately 200. Even though activated carbon prefers the non-polar organic solvent vapors, the great preponderance of water vapor present does indeed force the carbon to selectively adsorb the water rather than the solvent vapor.

Cartridge Service Life as a Function of Vinyl Chloride Concentration. Additional cartridge service life studies have been conducted on vinyl chloride because of its widespread use (over five billion pounds annually) and suspected carcinogenic properties. Figures 21 and 22 show the families of curves for two types of cartridge at concentrations ranging from 125 to 1000 ppm. The 10% breakthrough time as a function of concentration is shown in Fig. 23. At this time, organic vapor respirator cartridges do not appear to be satisfactory at concentration levels above 125 ppm.

Fig. 23. The 1% and 10% vinyl chloride breakthrough curves for MSA and Acme cartridges.

INSTRUMENT DEVELOPMENT

Use of Ultraviolet Light Transmission for Determining Track Densities in Dielectric Materials

In an effort to develop a simple, rapid method of determining the number of highly ionizing particle tracks per unit area in dielectric materials, we have investigated the use of preferential transmission of light through holes in thin films of dielectric material.^{15,16} Detection of fission fragments and alpha particles are of particular interest for use in neutron dosimetry, radon dosimetry and neutron radiography. The range of these particles is of the order of a few tens of micrometers, so films used with light transmission techniques must be no more than 10 to 20 μm thick. Polycarbonate and cellulose nitrate are used quite widely for detection of fission fragments or alpha particles, in that order. A 10- μm film of either of these transmits visible light very well, so other steps must be taken to improve the contrast between holes or tracks and the background of the film.

One approach to contrast enhancement has been to add a colored dye to film so that it absorbs preferentially in one part of the visible spectrum.^{15,16} Another approach which has been suggested¹⁶ is the use of light in the infrared or ultraviolet regions where the film itself could provide the absorption needed for preferential contrast. We have decided to investigate the potential for use of ultraviolet transmission through untreated film because such film is readily available, as is equipment which can be used for preliminary investigations.

The instrument available to us is a Beckman DK-2 Ratio Recording Spectrophotometer. The ratio of optical transmission of a sample to reference material is made by rapidly switching the beam through each and detecting transmitted light with a photomultiplier tube. The readout is to a recorder which plots relative transmission as a function of wavelength. The instrument uses a diffraction grating to scan the spectrum below 350 nm.

Figure 24 shows the relative transmission of films of cellulose nitrate and polycarbonate compared with air. Both films are nominally 2 μm thick. Based on this data and subsequent comparison of blank and irradiated samples, we chose to measure relative transmission in polycarbonate at 235 nm. We have not progressed far enough with cellulose nitrate to choose a particular wavelength.

The portion reported here will deal only with determination of fission track density in polycarbonate and, as we shall see, is very preliminary. We began by exposing samples of 10- and 12.5- μm -thick polycarbonate film to fission fragments from an electroplated source of ^{252}Cf . The polycarbonate samples were supported between concentric rings of stainless steel and plastic, in the fashion of an embroidery hoop. The exposed sample area was $2.9 \times 10^{-4} \text{ m}^2$ (2.9 cm^2). The samples were exposed for 10 min and etched for 3 hr or 4 hr in 60°C, 6N NaOH. The nominal track density determined by optical counting was 2×10^7 tracks m^{-2} (2.10³ tracks cm^{-2}). The tracks after etching

form tunnels of about $10\ \mu\text{m}$ in diameter for a 4-hr etch (Fig. 25). Because the $10\text{-}\mu\text{m}$ material seemed to have a better sensitivity by a factor of 2, it was used for the next portion of the study.

Samples of material were exposed to the ^{252}Cf source for periods from 1 to 1000 min, yielding optically countable tracks at a rate of about 2×10^6 tracks $\text{m}^{-2} \text{min}^{-1}$ (200 tracks $\text{cm}^{-2} \text{min}^{-1}$). The samples were etched for 3 hr, analyzed as follows, etched an additional hour and analyzed again. The transmission ratios of two unirradiated "background" samples to the irradiated sample was measured for each irradiated sample. To do this, the background was placed in the "sample" position while the irradiated samples were placed in the "reference" position in the spectrophotometer. This was done because the background was optically more dense than the irradiated samples. We then calculated the reciprocal of the background-to-sample transmission ratios. The average value of inverse transmission ratios for the backgrounds compared against each other (ideally equal to 1) was used as a baseline and subtracted from the irradiated sample readings. The net relative transmission as a function of irradiation time (i.e. track density) is shown in Fig. 26.

Although there is a wide spread in data contributing to individual points, a definite relationship exists

Fig. 24. Relative uv transmission of films of cellulose nitrate and polycarbonate.

Fig. 25. Tracks from ^{252}Cf fission fragments in polycarbonate film ($10\text{-}\mu\text{m}$ thick) etched 4 hr in 60°C , 6N NaOH. Magnification 400x.

between track density and relative uv transmission. The sensitivity is such that one can detect track densities as low as 2×10^6 tracks m^{-2} . There are a number of possible causes for the poor reproducibility. One possibility is light leaks. Although the potential for leaks is small, it can probably be reduced by redesigning the holder so that the films are held firmly in position. Film thickness variation may also contribute to sample variation. For high track densities, that effect is expected to be negligible. There is also potential for contamination from etch products from a film that has not been properly rinsed. Also, nonuniformity in etching could contribute. These are possibilities that must be investigated before this technique could be used reliably.

In summary, transmission of uv light may represent a sensitive, rapid method of determining particle track densities in thin films. Analysis of one sample takes less than 1 min, and in fact lends itself well to potential automation. Before the technique could be used satisfactorily, severe problems in reproducibility must be resolved.

Beta Spectrometer: Second Generation

In 1967, we developed a solid-state beta spectrometer system for the Hazards Control Counting Laboratory.^{17,18} Its primary purpose was to allow identification of unknown isotopes which emitted only betas or conversion electrons by determining the end point energy of the beta spectrum, or the energy of conversion electrons. The spectrometer consisted of a cooled $80\text{-mm}^2 \times 5\text{-mm}$ Si(Li) detector, mounted in an elaborate, air-locked vacuum chamber, a cooled FET preamplifier, and standard external electronics

Fig. 26. Net relative transmission of fission fragment irradiated 10 μ m polycarbonate film at 235 nm as a function of irradiation time.

components, culminating in a conventional multi-channel pulse-height analyzer (PHA). Output in the form of spectral plots and data reduction was rather tediously done by hand and eye methods. The various technologies represented have progressed rapidly in the intervening years, and it was decided this spring to update the system.

Detector design has improved in the past few years to the point where cooling is no longer necessary to provide adequate electron resolution for our purposes. This means that a much simpler chamber, lacking both cooling and its attendant airlock, can be built. Also, larger detectors are available, thus providing enhanced efficiency for low-activity samples. Relatively easy access to the main computer facility enables one to utilize far more powerful data analysis techniques than our previous methods. Thus, a simpler, and yet more powerful spectrometry system has evolved.

The hardware is now quite simple, as shown in Fig. 27. The detector (200 mm² X 5 mm) is mounted in a light-tight evacuable chamber, with an adjustable sample stage. The preamplifier and amplifier are both external, as is the bias supply amplifier output that goes to any of several pulse-height analyzers in the counting room. Resolution of the detector is about 14 keV fwhm for ¹³⁷Cs electrons (624 keV).

Once a spectrum is acquired a rather circuitous data processing chain begins, resulting in a disk file of the data being written on one of the main computers. (This

process is undergoing changes, and so will not be detailed.) Once on disk, a code reads the file and performs a weighted least-squares quadratic fit to the higher energy portion of the spectrum. The resulting polynomial is differentiated and evaluated using operator input energy calibration values to determine E_{max} for the spectrum. Plots of the data and the fit are generated, and can be viewed on the T¹ screen, or output via RIET or DD80 printers. Once on disk, spectra can be processed at the rate of more than one per min by operators with only limited knowledge of the computer system.

Initial tests show good accuracy on spectra of reasonable quality. Several factors can compromise this quality, including:

1. Poor statistics, from weak samples and short counting times.
2. Incomplete beta stopping, with particles of >2.5-MeV energy.
3. Presence of accompanying gamma rays. These cause Compton interactions in the detector and chamber, and distort the normal spectral shape.

Fig. 27. Block diagram of spectrometer hardware.

4. Poor sample preparation, causing self-absorption and hence spectral distortion.

More experience with "real" spectra from working samples will perhaps allow operator intervention to compensate for such defects. Despite these difficulties, it appears that $E_{\max}$ can be measured to accuracies of about $\pm 5\%$, which is usually adequate for identification purposes.

A System to Produce Low Concentrations of Gas in Air

A system has been designed to produce low concentrations (0.5-2000 ppm) of gas in air. This unit represents significant improvement over a previously reported device¹⁹ in that: (1) no special glassware is required, (2) instant resetting is available, (3) it is 50% lighter, and (4) it occupies 60% less space.

The apparatus is shown in Fig. 28. The contaminant gas is introduced into the air stream by precisely controlling the syringe plunger advance rate. The plunger is driven by a 280-rpm synchronous motor and the speed (0.5-15 rpm) is controlled by a variable ratio gear box. A gear reducer with a right-angle drive is needed to obtain the proper plunger advance rates. A quick release lever provides for instant resetting of the moving slide, and a limit switch prevents syringe plunger overshoot and possible damage to the drive mechanism.

Syringe sizes from 0.05 to 20 ml capacity can be used to provide a large dynamic range of concentrations. A Teflon "T" fitted with a Teflon-faced silicon rubber septum provides the inert, gas-tight interconnection between the needle and moving air stream. A Teflon or platinum needle,

Fig. 28. The system used to produce low concentrations of gas in air.

however, must be used when working with particularly corrosive gases such as hydrogen chloride or chlorine.

This device is currently used to produce span gases for carbon monoxide, nitric oxide, and chlorine analyzers. It can, however, be utilized to calibrate any field instrumentation or detector tubes which require gas mixtures in the part-per-million range.

Cartridge and Canister Test Systems for the Testing and Certification Laboratory

The National Institute of Occupational Safety and Health has extended reimbursable contract NIOSH-1A-74-21. This contract authorizes the Special Projects Section to design, fabricate, test and deliver four canister and cartridge testing systems to the Testing and Certification Laboratory located in Morgantown, West Virginia.

The contract is scheduled to terminate July 1, 1975, but our target completion date is March 18, 1975. Approximately 1.3 man-years of effort will be expended. An allotment of \$65,000 will be split into \$25,000 for purchasable parts and \$40,000 for manpower requirements.

The specifications for each cartridge test system is shown in Table 6. Briefly stated, each system will humidify an air stream and mix it with an appropriate gas or vapor. A gas analyzer is provided for the carbon tetrachloride system. In addition to fabricating these gas-mixing systems, two gas-analyzer calibration units are being built. These will be used to establish the accuracy of the nitrogen dioxide, carbon monoxide and chlorine analyzers which were previously installed.

The two humidification systems are 80% complete. The carbon tetrachloride and miscellaneous gas systems are approximately 40% complete. One calibration system has been fabricated, and is currently being used at the Morgantown test facility.

Tritium Surface Contamination Monitor

Two PAC4G alpha survey instruments have been modified for use by the Health Physics Group as

tritium surface monitors. Modifications were made to the power supply, the grids, the windows for the probe, and the gas flow indicator vent.

The normal operating voltage of the PAC4G is 1600 to 1800 V. Experiments have shown that 2400 V is a better operating voltage for tritium detection, so the power supplies have been rebuilt to operate at 2400 V.

New grids to support the probe windows which are about one-fifth as thick as the originals have been fabricated. This minimizes the distance in air that the tritium beta must traverse before reaching the window. The windows are aluminized Formvar having a thickness of 70 to 100 $\mu\text{g}\cdot\text{cm}^{-2}$. This is thin compared to the 600 $\mu\text{g}\cdot\text{cm}^{-2}$ range of the tritium beta. The windows are supported on the grids by a fine nickel screen with approximately 90% transmission.

The normal method of checking gas flow through the probe is by holding a lighted match at the vent until a flame appears. However, as the propane counting gas flows into the probe, it diffuses out through the windows and presents a hazard for this

Fig. 29. Neutron spectrum obtained with hydrogen proportional counter. Source neutrons had nominal energy of 23 keV with 12-keV dispersion.

Table 6. System specifications for cartridge and canister testing systems

System	Flowrate (l/min)	Gas or vapor concentration (%)	Relative humidity	Analyzer sensitivity (ppm)
Humidification - I	32, 64	-	25	-
Humidification - II	32, 64	-	85	-
Carbon tetrachloride	32, 64	0.1, 0.5, 2.0	50	5
	175	0.1	50	5
Miscellaneous gas	5-70	0.05-5	25-95	-

method of checking gas flow. For this reason, a gas flow indicator vent has been plugged and rotameters have been installed on the exterior of the meter housings.

In addition to tritium, these detectors also respond well to alpha, gamma and other beta radiation so other techniques must be employed to differentiate between them.

Neutron Spectrometer Development

Introduction. A neutron spectrometer system has been under development by the LLL Hazards Control Department for some time. Its history is traced in earlier reports.^{20,21} Each of the detectors in this system has been exposed to monoenergetic neutron fluxes of known intensity* and to (α, n) and ^{252}Cf sources. Some of the important characteristics of the spectrometer system are summarized in Table 7.

Also, a computer code, NUTSPEC, is under development for the purpose of unfolding the pulse-height data. This code is intended to make suitable corrections to the raw data, and to take maximum advantage of digital filtering techniques in the unfolding of the data in the hope that a simplified version of it may be implemented on a Nuclear Data 4420 computer-analyzer system. The methods used so far seem to be quantitative, and the results are in reasonable agreement with similar measurements reported in the literature.^{22,23}

Proportional Counters. All of the proportional counters in this system have been calibrated, and their most important characteristics are included in Table 7. The hydrogen proportional counter has been calibrated at energies down to 8 keV. An example of a spectrum of 23-keV neutrons is shown in Fig. 29. It is important to note that the energy dispersion of the neutron beam in these experiments was 12-15 keV.

Processing of the pulse-height data obtained with proton recoil spectrometers requires the calculation of the slope of the proton recoil energy distribution. The neutron flux is related to the slope by the relation.^{24,25}

$$\phi(E) = -\text{const} \frac{E}{\sigma_s(E)} \frac{dp(E)}{dE}$$

where

E = energy,
 $p(E)$ = energy distribution of recoil protons,
 σ_s = elastic scattering cross section.

For the proportional counters used in these experiments, the total ionization has been found to be proportional to proton energy within the useful energy range. Consequently, the slope of the proton recoil distribution may be simply related to the slope of the pulse-height distribution after suitable corrections for gamma rays, wall- and end-effects.

A variety of digital filters has been tested in the calculation of the slope of the pulse-height distribution. The code NUTSPEC allows a choice of digital filters up to nine points in width, and includes various degrees of smoothing. Figure 30 shows the contrasting results obtained with digital filters ranging from 3 to 9 points, with and without smoothing. Note that for the filters derived from a Taylor expansion, the statistical noise in the derivative increases with the number of points in the filter. However, for filters based on a fit of a parabola to $2N + 1$ points, the noise in the spectra is reduced with increasing numbers of points in the filter. Fourier transform analysis of these filters indicates no lateral displacement of peak centroids, and the peak areas are essentially the same for all of the tested filters.

Scintillators. As indicated in Table 7, most of the proportional counters are limited by proton escape at energies above 1 MeV. Consequently, several liquid organic and organic crystal scintillators have been obtained. These scintillators were chosen because they show good pulse-shape discrimination properties which makes the suppression of gamma-ray events fairly simple. Figure 31 shows the electronics configuration used to accomplish gamma-ray suppression and pulse-height analysis.

Unfolding scintillator pulse-height spectra are complicated by the fact that the light output does not vary linearly with proton energy. However, the light output from electrons produced by Compton scattering is linear with energy. Thus the energy scale of the pulse-height distribution may be calibrated using a gamma-ray source, and then a formula which relates electron and proton light output may be used to infer proton energies. The light output response data of Verbinski²⁶ for NE213 and Schwartz and Owen²⁷ for stilbene have been fitted with polynomials using a weighted least-squares technique. Light output for protons calculated with these polynomials falls within 2-3% of the experimental data of Verbinski²⁶ and

* Monoenergetic neutron sources were produced by the $\text{D}(\text{D}, n)^3\text{He}$ and $\text{D}(\text{T}, n)^4\text{He}$ reactions on the Texas Nuclear neutron generator at the LLL calibration facility. Also, monoenergetic neutron sources were obtained from the $^7\text{Li}(p, n)^7\text{Be}$ reaction at the LLL cyclotron facility. For these experiments, thin films of LiF ($0.25\text{--}1.0 \text{ g/m}^2$) were evaporated onto tantalum backing material of 0.5- to 1.0-mm thickness. Neutrons produced by the (p, n) reaction were observed at 30° from the incident proton beam, and their energies were computed from the known proton energy and the kinematics of the reaction. In all experiments, the absolute neutron intensities were monitored using a precision long counter.

Fig. 30. Spectrum of 250-keV neutron source obtained with hydrogen recoil proportional counter.

- (a) Pulse-height data.
- (b) Neutron flux obtained using three-point, non-smoothing digital filter.
- (c) Neutron flux obtained using nine-point, non-smoothing digital filter.
- (d) Neutron flux obtained using five-point, smoothing digital filter.
- (e) Neutron flux obtained using nine-point, smoothing digital filter.

Owen.²⁷ The formulas used are given below where unit light output is that from a 1-MeV electron.

For NE 213

$$L_p(E_p) = \begin{cases} 0.001660 + 0.03556 E_p + 0.1236 E_p^2 - 0.01506 E_p^3 & \text{for } E_p \leq 3 \text{ MeV} \\ -0.3091 + 0.3519 E_p + 0.009938 E_p^2 & 3 < E_p \leq 10 \text{ MeV} \\ -1.070 + 0.5103 E_p + 0.001585 E_p^2 & E_p > 10 \text{ MeV} \end{cases}$$

For Stilbene

$$L_p(E_p) = \begin{cases} -0.0136 + 0.0868 E_p + 0.0775 E_p^2 - 0.00544 E_p^3 & \text{for } E_p \leq 4 \text{ MeV} \\ -0.307 + 0.291 E_p + 0.0253 E_p^2 - 0.000513 E_p^3 & 4 < E_p \leq 15 \text{ MeV} \\ -2.534 + 0.703 E_p & E_p > 15 \text{ MeV} \end{cases}$$

where E_p is proton energy in MeV, $L_p(E_p)$ is light output.

Table 7. Some important characteristics of spectrometer system components

Detector (mm)		Useful energy range	Effectiveness of γ -ray discrimination	Energy resolution (%)	Sensitivity (counts/nm ²)
50.8 x 152	H ₂ prop. counter	10 - 800 keV	OK	10	4×10^{-6}
50.8 x 458	CH ₄ prop. counter	50 - 800	poor	15	10^{-6}
50.8 x 458	He ³ prop. counter	200 - 2500	OK	15	-4×10^{-6}
25.4 x 152	He ³ prop. counter	60 - 800	OK	4	4×10^{-7}
50.8 x 115	NE213 liquid organic	1.5 - 20 MeV	good	7 - 10	1.5×10^{-3}
38 x 76	NE213 liquid organic	1.5 - 20	good	7	10^{-4}
12.7 x 12.7	NE213 liquid organic	1.5 - 20	good	7	5×10^{-6}
10 x 15	Stilbene organic crystal	1.5 - 20	good	7	5×10^{-6}
38 x 2	Li ⁶ I	0.5 - 20	good	10	10^{-7}

Fig. 3: Block diagram of electronics used for scintillator spectrometers.

Fig. 32. (a) Neutron flux measured with 38×76 mm NE213 liquid scintillator. Source neutrons produced by $D(D,n)^3\text{He}$ reaction.
(b) Source neutrons produced by $D(T,n)^4\text{He}$ reaction.

A proton recoil distribution is constructed from the pulse-height data and the gamma-ray calibration using the light output formulas above. This is corrected for proton leakage and secondary scattering using the relation due to Anderson and Broek²⁸ and for attenuation of the unscattered flux in the detector. Neutron spectrum measurements of the D,D and D,T reactions using these detectors are shown in Figs. 32 and 33. The asymmetry of the peak is due to elastic scattering off carbon (in the detector) which degrades the neutron energy without producing a significant light output. Resolution of these detectors is typically about 7%, as indicated in Table 7, and absolute neutron intensities are in good agreement with those obtained by a calibrated long counter.

Measurements of moderated and unmoderated PuBe and ^{252}Cf sources have also been made. An example is shown in Fig. 34. These results are in reasonable agreement with those of Anderson and Neff²² and Lorch.²³ These data are now being used to compute neutron "dose equivalent" and "Kerma" for neutron dosimetry purposes.

Conclusion. It is clear from Table 7 that a capability exists for the measurement of neutron energy spectra and dose over the energy range 10 keV – 20 MeV. Sensitivity is variable over this range, depending on the detector used. At present, at least 50 000 counts are required in the pulse-height spectrum to give reliable spectral data. Table 7 may be used to estimate the lower bound on useful neutron intensities given a constraint on the length of count. For a 24-hr counting period, the largest NE213 scintillator may be used in neutron fluxes in excess of $300 \text{ n/m}^2 \text{ sec}$ ($0.03 \text{ n/cm}^2 \text{ sec}$). This corresponds to a dose rate of only a few $\mu\text{R/hr}$. Upper limits on neutron intensities are set by pulse pile-up considerations. Good results with the scintillators are obtained up to approximately 3000 Hz while the proportional counters must be limited to about 500 Hz. Consequently, neutron fields up to $10^8 \text{ n/m}^2 \text{ sec}$ ($10^4 \text{ n/cm}^2 \text{ sec}$) or dose rates up to the order of 1 R/hr may be characterized with this system.

Development of the data-unfolding techniques is continuing to reduce the systematic errors and to include estimates of the random errors. Measurements of various moderated sources are continuing. Evaluation of the ^6LiI scintillator has not yet begun. Finally, specifications are being written for the purchase of two additional hydrogen proportional counters to extend the energy range and sensitivity of these devices.

Fig. 33. PuBe source viewed through 20 cm of Al. The detector is 38×76 mm NE213 liquid scintillator

NUCLEAR SAFETY

A Method for the Safe Handling of $^{238}\text{U}_3$ Waste

We have begun a series of experiments to discover if $^{238}\text{U}_3$ machining scrap could be handled in the same manner as metallic uranium scrap. Preliminary experiments by others²⁹ indicated that U_3 scrap would oxidize slowly in air. We are investigating the possibility that it will oxidize slowly in mixtures of water and water-soluble oil.

Three 100-g batches of U_3 were washed with "Freon TF" and centrifuged. The batches were placed in 1-l stainless steel beakers which could be capped and sealed. The first beaker also contained 250 ml of a 20:1 water-water-soluble oil mixture, the second 250 ml of a 40:1 water-water-soluble oil mixture, and the third only U_3 .

Readings were taken periodically and excess pressure was vented from the cans. No data on rates of evolution of gas are presented because x-ray diffraction studies indicate that the beaker caps leaked. Gage readings do indicate that the $\text{U}_3\text{-H}_2\text{O}$ reaction proceeded at a constant rate over the 7 weeks of the experiment.

A fourth batch of U_3 was washed with "Freon TF," filtered, and allowed to stand on the filter paper. After seven weeks, a portion of each of the four batches was heated to 85°C for 2 hr and 100°C for 8 hr. Following these treatments, samples of each batch were submitted for x-ray diffraction analysis. Table 8 gives the results.

Since these initial experiments indicate that U_3 oxidizes at reasonable rates under the conditions likely to be encountered in disposal, a larger-scale experiment has begun. In this experiment, 3 kg of U_3 has been put in a 30-gal barrel with 25 gal of water, and a pressure gage has been installed to monitor the reaction. The barrel is stored out of doors. We will

Table 8. Results of x-ray diffraction analysis:

Sample treatment	Major species
Original material	Major phase $\beta\text{-U}_3$, strong phase UO_2
20:1 cutting oil treatment	Major phase $\beta\text{-U}_3$, strong phase UO_2
20:1 cutting oil treatment + heat	Major phase $\beta\text{-U}_3$, strong phase UO_2
40:1 cutting oil treatment	Major phase $\beta\text{-U}_3$, strong phase UO_2
40:1 cutting oil treatment + heat	Major phase $\beta\text{-U}_3$, strong phase UO_2
Dry material in beaker	Strong phase $\beta\text{-U}_3$, strong phase UO_2
Dry material in beaker + heat	Major phase U_3O_8 , moderate amount U_3O_{10} , possible traces other oxides
Washed and filtered	Major phase UO_2 , some $\beta\text{-U}_3$
Washed and filtered + heat	Major phase UO_2 , some $\beta\text{-U}_3$

monitor the pressure and the chemical species at intervals to see if any new findings accrue.

At present, the conclusion of these studies is that it is reasonable to dispose of $^{238}\text{U}_3$ in vented drums under water-water soluble oil mixtures.

Acknowledgments. We would like to thank Clarence Hoenig, who did a preliminary experiment and who contributed helpful suggestions, and Bill Birka, who did the x-ray diffraction analysis.

DECONTAMINATION

Plant Test of New Liquid Decontamination Process

A British technique for the decontamination of radioactive liquids³⁰ was recently tested at LLL on a small scale. The results were sufficiently encouraging to try a larger scale test. We used 1000 gal of contaminated sea water, free of excess phosphates and detergents, for the test.

The method originally called for coprecipitating barium sulfate and copper ferrocyanide sorbent from

solution, then filtering and drying it into a cake. The cake would then be crushed into granules and packed into a suitable column. The liquid waste would be decontaminated by passing it through the column. The manipulation of the liquid in this manner seemed to be an overly tedious and inefficient procedure, so it was decided to coprecipitate the sorbent in the tank itself.

The original concentration of radioactivity was about 10^4 dis/min/l alpha and 10^5 dis/min/l beta.

The slurry was stirred for 3 hr and allowed to settle overnight. Approximately 90% of both alpha and beta contamination was removed, as in the laboratory-scale tests, reducing the activity to a level where it could be discarded into the sewer. We then dumped 800 gal, and transferred the remaining 200 gal to the dregs tank where it was processed through the rotary drum filters.

OTHER

A Weighted Least-Squares Program with Error Estimates

The linear least-squares fitting program OCTOPOLY* has been used for some time to obtain polynomial fits to various data. This program has been substantially modified to include Devington's³¹ algorithm for multiple linear regression. At present, the code offers four fitting options:

1. Unweighted linear least squares.
2. Minimum chi-squared.
3. Weighted linear least squares using statistical $(1/\sqrt{N})$ weights.
4. Weighted linear least squares with weights supplied by the user.

For fitting modes 1, 2 and 3 above, the uncertainty in each of the parameters is calculated as well as the variance, and chi-squared for the fit. The code will now accept not only polynomials as fitting functions, but any linear combination of analytic or tabulated functions.

Results of the fitting are printed, and the fitted functions are plotted together with the input data. Error bars are also shown on the plots.

Tritium Analysis of Alpha-Contaminated Oil

Oil which has been contaminated by tritium and alpha-emitting particulates presents a special tritium analysis problem. The alpha activity may be measured by wet-ashing the oil, flaming it, and counting the resulting residue. However, during the chemical preparation of the samples for alpha analysis, the tritium, as well as all the stable hydrogen, is lost. An investigation was initiated to analyze the tritium contamination of these oils using a solvent extraction and lyophilizing method.

To analyze for tritium content in these oils, the tritium must be separated from the rest of the sample and then measured. We attempted to separate the tritium from the oil matrix by chemical exchange with

It was noted that the method added a significant amount of dissolved solids to the liquid. A preferable method would be one which would reduce the concentration of dissolved solids in the effluent.

The procedure has been added to the file of methods actively used in special applications.

hydrogen in other solvents. Solvents which have a higher vapor pressure than the oil can then be either distilled or lyophilized. The tritium-tagged solvent can then be measured in a standard liquid scintillation cocktail.

Solvents which would be suitable for this application should have several characteristics. They should be soluble in the oil, nonquenching in the liquid scintillator, and higher in vapor pressure than the oil.

All of these solvents were soluble or at least immiscible in the oil, and all had higher vapor pressures than the oil. Many oils are highly quenching.

No quenching effect was observable in solvent-scintillator mixtures ranging from 0 to 2 ml of solvent dissolved in 15 ml of Hewlett Packard Insta-Gel for the following solvents:

Benzene
Ethanol, Methanol, Butanol
Iso-octane, Propanol, Octanol,
Toluene, Xylene, Dioxane.

A sample of Mobil vacuum pump oil which was tritium-contaminated at a known level and a separate alpha contaminated sample were obtained, along with clean oil of this type. Stock solutions of 4 ml of solvent and 20 ml of oil were prepared and allowed to stand. They were then lyophilized. The resulting liquid was mixed in a scintillation cocktail and counted. Table 9 presents a summary of these results.

After running several solutions in this manner, it became apparent that roughly the same total amount of tritium was obtained with each solvent. This suggested that the solvent itself was not exchanging hydrogen for tritium, but rather freeing some volatile tritiated species, possibly HTO, which was then collected in the lyophilization. Data which support this view are presented in Table 10. Tritium was still collected from oil with no solvent mixed in. The total amount of tritium collected was also independent of the amount of solvent added. This supports the supposition that some tritium-bearing fraction of the oil is independently collected by lyophilization. Moreover, this fraction is soluble in a rinse by a

* Origin unknown.

Table 9. Tritium analysis results

Solvent	Equilibration time	Collection time (hr)	% recovery	% dilution ratio
Ethanol	69.5 hr	2.0	2.47	16.43
Ethanol	17.0	5.0	2.59	17.55
Methanol	17.0	5.0	3.06	15.32
Octanol	17.0	29.0	1.31	34.82
Propanol	89.0	4.5	2.21	14.76
Butanol	20.0	5.0	2.35	14.48
Water	22.0	2.0	2.54	13.37
Dioxane	23.5	17.5	-	10.86
Benzene	30 days	89.0	-	4.24

Note: 1. All samples contained 4 ml solvent plus 20 ml of stock oil No. 2.

2. Stock oil No. 2 = 3.59×10^3 dis/min/ml

$$3. \% \text{ recovery} = \frac{\text{volume of solvent collected} \times \text{dis/min/ml}}{20 \times 3.59 \times 10^3 \text{ dis/min/ml}} \times 100$$

concentration of tritium in recovered

$$4. \% \text{ dilution ratio} = \frac{\text{solvent (dis/min/ml)}}{\text{concentration of tritium in original oil sample (dis/min/ml)}} \times 100$$

Table 10. Tritium analysis results

Sample description	Equilibration time (hr)	Collection time (hr)	% recovery	% dilution ratio
20-ml sample of oil was lyophilized for 23.5 hr. 4 ml of water was used to rinse the glassware.	-	23.5	1.16 4 ml H ₂ O rinse constituted the volume.	6.41
20-ml sample of oil was heated to 100°C while being lyophilized for 23.5 hr. then 4 ml of water was used to rinse the glassware.	-	23.5	2.06 4 ml H ₂ O rinse constituted volume	10.31
2 ml ethanol + 20 ml oil	24	2.0	2.06	22.84
4 ml ethanol + 20 ml oil	24	2.0	1.80	11.98
8 ml ethanol + 20 ml oil	24	2.0	2.33	6.13
4 ml of 0.1 M permanganate solution + 20 ml oil reflux for 1 hr; let cool for 1 hr; lyophilize.	-	2.0	6.15	49.17

Note: 1. Stock oil No. 2 = 3.59×10^3 dis/min/ml

$$2. \% \text{ recover} = \frac{\text{volume of solvent collected} \times \text{dis/min/ml}}{20 \times 3.59 \times 10^3 \text{ dis/min/ml}} \times 100$$

concentration of tritium in recovered

$$3. \% \text{ dilution ratio} = \frac{\text{solvent (dis/min/ml)}}{\text{concentration of tritium in original oil sample (dis/min/ml)}} \times 100$$

water-soluble solvent and not in a rinse by benzene or non-water miscible solvent.

Solvent was added to the alpha-contaminated oil. This oil was then lyophilized. The vapor from the oil was filtered through a wad of glass wool so that small droplets could not pass through the lyophilizing

system. The collected sample was wet-ashed, flamed and counted. All samples analyzed by this method were completely devoid of alpha contamination. Thus, the technique does separate the tritium and alpha-contaminated species in such a way that the tritium can be measured.

Samples of oil tritiated under other conditions will probably have a different tritium-bearing fraction available to lyophilizing. Since exchange with solvents does not appear to be a major effect, these variations of tritium-bearing fraction, if present, can represent an unacceptable variation in the results of the analysis. Another approach is to chemically oxidize some or all of an oil sample so that the tritium is bound in HTO. The HTO can then be collected by lyophilizing in the same manner as the previous solvents. A 20-ml sample of oil was mixed with 4 ml of a 0.1 molar solution of potassium permanganate and refluxed for 1 hr. Lyophilizing the results of this refluxed mixture resulted in releasing 6.15% of the tritium in the original oil sample. Some tritium was oxidized in this experiment.

A third approach is to oxidize the oil in a combustion bomb and then collect the resulting mixture of H_2O and HTO. Two such experiments were tried. In the first experiment, 2 ml of tritium-

contaminated Mobil vacuum pump oil was oxidized in a 1.8-l Parr bomb filled with oxygen to 300 psi. The combustion products were vacuum-distilled at methanol and dry-ice temperatures. Water was collected in this system and the CO_2 was vented. No discernible ash was left in the bomb.

About $84\% \pm 12\%$ of the tritium contaminating the oil was recovered in this analysis. The silicone-based oil, dimethyl polysiloxane, used as a cutting oil and pump oil in diffusion pumps, was also burned in the bomb. Complete combustion was obtained, leaving a residue of white silicone powder.

A second analysis, using a Packard #300 oxidation chamber, recovered about 50% of the tritium in a 0.1-ml sample of oil. In both cases, oxidation of the oil by combustion resulted in a high tritium recovery.

Acknowledgments. We wish to thank Lynn Penn for helpful conversations and Emery Willes and Paula Wood for the demonstration and use of their apparatus.

Technical Notes

INDUSTRIAL HYGIENE

Computer Program to Calculate Ear Protector Noise Attenuation

A computer program has been written to calculate the noise attenuation provided by the ear protective devices commonly used at the Laboratory, and the resulting noise exposure to the wearer.

The input is the octave band and overall noise intensities measured at the work location. The noise attenuation and resulting exposure for popular brand ear plug- and ear muff-type protectors are then calculated from equations contained in the National Institute for Occupational Safety and Health (NIOSH) "Criteria Document on Noise."

The noise attenuation data for the ear protectors, as supplied by the manufacturer, is stored in the program. However, this data can readily be replaced if other brand or type noise protective devices are considered for use. Additional information may be obtained from the author.

A Low-Level Fluorine Detector

From time to time programatic needs require the handling of fluorine gas in gloveboxes at the Lawrence Livermore Laboratory. When fluorine is handled in a glovebox containing potential airborne radioactive particulate material, it is imperative that the experimenter be able to detect any fluorine leaks within the glovebox. Any undetected fluorine leaked into a glovebox may ultimately cause a breach in the high-efficiency air (HEPA) filters which would provide a direct avenue of release of radioactive particulate matter to the environment. A survey of the literature for methods of fluorine detection resulted in no applicable fluorine sampling system which would satisfy our experimental needs. Such a sampling system must be able to detect fluorine levels in the low parts-per-million range, utilize a visual indication system, have a fast response time, and operate against 5 in. wg of negative pressure.

Since fluorine is such a strong oxidizing agent, this property has been used effectively for several different types of fluorine detectors. Several commercial instruments are available which detect oxidants using a microcoulomb sensor to monitor the oxidation-reduction reaction of a potassium iodide iodine solution. Therefore, a Mast ozone meter was purchased for evaluation at low fluorine levels. Another detection system which we felt may satisfy our experimental requirements consists of a

standard midget impinger containing a buffered potassium iodide* starch† indicating solution (Fig. 34).

To evaluate these two possible detection methods, several fluorine concentrations were prepared using a standard dilution method (Fig. 35). The large glass jar was carefully cleaned and dried under vacuum for several days. Research-grade argon was used to fill and flush the glass jar. The fluorine was removed from a low-pressure source using a standard gas syringe which was equipped with a Teflon O-ring and injected into the partially evacuated jar along with the research-grade argon. As soon as the jar reached atmospheric pressure, simultaneous sampling with both

Fig. 34. Buffered potassium iodide standard impinger system in place on glovebox exhaust.

*The buffered potassium iodide absorbing solution was prepared by dissolving the following reagent-grade chemicals in 750 ml of distilled water: 10 g potassium iodide, 13.6 g potassium dihydrogen phosphate, and 26.8 g sodium monohydrogen phosphate heptahydrate. The resulting solution volume was adjusted to 1 liter using distilled water.

†The starch solution was prepared by making a paste from 2 g of soluble starch, 0.01 g of mercuric iodide and a small amount of water. The mixture was slowly added to 1 liter of boiling water and boiled for a few minutes.

Fig. 35. Sampling train plus mixing jar.

detection systems was started. The first run was considered a passivation run. Three additional fluorine test atmospheres of approximately 22 ppm fluorine were prepared by injecting 1 ml of fluorine into the jar. The Mast ozone meter read approximately 1/140 of the calculated value, or 0.3 ppm fluorine. A dark blue iodine starch complex developed immediately in the midget impinger containing the buffered potassium iodide starch indicator solution. A 5-min time period was required for the fluorine concentration, as indicated by the Mast ozone meter, to decrease from 0.3 to 0.0 ppm. To determine how sensitive the buffered potassium iodide starch impinger system is to low-level fluorine concentrations, a sample was collected when the Mast ozone meter indicated approximately 0.05 ppm fluorine. A time of 1-1/2 min was required for the indicator solution to develop a perceivable color.

This basic set of experiments indicates that a Mast ozone meter can be used successfully to detect low levels of fluorine. Because this instrument is intended for ambient air applications, the sampling pump is not designed to overcome a negative static pressure of any magnitude. The inability of the Mast ozone meter to overcome the high static pressure of our system disqualified it from our present fluorine operation.

The buffered potassium iodide starch impinger system was found to have a quick visual response time to low levels of fluorine. Since the sampling pump is a separate part of the system, it can be sized to

overcome whatever static pressure exists in the system to be sampled. By maintaining the indicator solution volume at 10 ml, holding the air sampling rate at 2 l/min and replacing the solution with fresh indicator solution after 4 hr of operation, a sensitive, simple, inexpensive and dependable fluorine monitor will result.

A Computer Program to Size and Balance Exhaust Ventilation Systems

The American Conference of Governmental Industrial Hygienists (ACGIH) sells a computer code which designs industrial exhaust ventilation systems. Given the spatial layout of the system and the types of hood inlets, the code sizes all the ducting for air flow balance and determines exhaust fan flow and pressure drop criteria. The output of the code is in a format similar to the ACGIH Industrial Ventilation³² Manual's velocity pressure method worksheet.

The code was purchased from the ACGIH and modified to run on the LLL computers.³³ Following that, it was decided that the code's usefulness would be extended if it could be used to size fans in existing systems and to add and balance additional ducting into existing systems. The program was further modified to include these features and to improve its efficiency. It is now available for use. Full documentation is being prepared as a separate report, and will be available in the near future.

INSTRUMENT DEVELOPMENT

Contribution of Alpha Particle Tracks to Fission Fragment Spark Counting Backgrounds

For some time, LLL has made use of a set of low-level neutron area monitors which involve the use of thin foils of ^{235}U in association with thin films of polycarbonate to register fission fragments from $^{235}\text{U}(n,f)$ reactions.³⁴ The foil-film pair is used inside a cylindrical moderator designed to impose a "rem response" on the thermal neutron detector it contains. During analysis of eight of these units for the second half of 1973 we found the detectors, including background detectors (uranium foil-film pairs stored in boron carbide spheres), showed about 500 total counts which is about 10 times our normal background and is equivalent to a neutron dose of about 10 mrem. The 8- μm -thick polycarbonate films are normally etched in 60°C, 6N NaOH for 1-1/2 hr before counting and it is assumed, based on tests of the system, that ^{235}U alphas are not recorded. This time, however, we realized that 6N KOH was used inadvertently, and began to suspect that the high counts were the results of counts due to alpha tracks etched more efficiently by the KOH. Because it is nice to use spark counting of fission foils for other long-term, low-level measurements, we decided to investigate the effect of changes of some of the parameters we used on sensitivity to alpha background.

To review briefly, the spark counter is a device for detecting chemically etched, charged-particle tracks in films of dielectric material.³⁵ The film is placed between two electrodes having a potential of 500-1000 V. One of the electrodes is aluminized plastic film so that a spark passing through a particle track hole burns a spot in the aluminum layer, preventing passage of a second spark. By connecting a pulse counter to the electrodes, the particle tracks can be counted automatically. In our case, the polycarbonate film is placed between a ring of stainless steel and one of plastic, much like the use of an

embroidery hoop. The open area for irradiation and counting is 2.9 cm^2 .

To simulate the alpha background from long-term use of ^{235}U in a reasonable time span, we used a set of 5- μCi electroplated ^{239}Pu foils as a ^{235}U surrogate. Including self absorption of the thick ^{235}U foil, we estimated a 2-day irradiation with our ^{239}Pu sources to be equivalent to 6 months with ^{235}U .

We began by comparing the counts for 1-1/2 hr each times in 6N NaOH and 6N KOH at 60°C using 4-, 10- and 12.5- μm -thick registration material. Table 11 shows a comparison of the data for the ^{239}Pu alpha sources with relative responses to a ^{252}Cf fission fragment source.

The table indicates that the use of the wrong etchant for the same amount of time probably accounted for the factor-of-10 increase in our background.

When the alpha particle reaches the end of its track, the linear energy transfer (LET) increases rapidly. Since track registration improves with increasing LET,³⁶ we became concerned that the bare ^{239}Pu foils did not properly simulate the potential background contribution for degraded alphas emitted by our 5.5 $\text{mg}\cdot\text{cm}^{-2}$ ^{235}U sources. So the data for NaOH were repeated by interposing absorbers of polycarbonate film to simulate the degradation of the alphas. The use of absorbers (Table 12) suggests that the effective alphas probably come from the surface of the ^{235}U , and that degraded alphas probably do not penetrate the film far enough to be recorded. Indeed, the reduction in KOH-etched background by going from 8- to 10- μm -thick film tends to confirm this.

Finally, we used the bare ^{239}Pu irradiator to compare the change in alpha counts with etch time. The results shown in Fig. 36 include fission fragment data for comparison. It appears that the problem we encountered was not so much the preferential etching of alphas by KOH, but that we simply were too far

Table 11. Comparison of alpha background with fission fragment responses for polycarbonate film (1-1/2 hr etch in 60°C etchant)

Film thickness (μm)	6N NaOH		6N KOH	
	Alpha bkgd ^a	Fission fragment ^b count	Alpha bkgd ^a	Fission fragment ^b count
8	43 $\pm$ 4	867 $\pm$ 72	403 $\pm$ 87	936 $\pm$ 65
10	7 $\pm$ 4	626 $\pm$ 40	4 $\pm$ 4	789 $\pm$ 32
12.5	4 $\pm$ 4	443 $\pm$ 53	7 $\pm$ 4	621 $\pm$ 63

^aNormalized to 48 hour exposure from 5 μCi ^{239}Pu

^b2 minute exposure to 0.03 μCi source of ^{252}Cf

out on the KOH etch-time curve. If we had been etching in KOH at times much less than 1-1/2 hr, alpha background would not have developed.

The results of this investigation suggest that in addition to backgrounds from spontaneous fission and cosmic-ray fission of ^{235}U , there is a potential background from alpha particle damage. This background can be controlled either by using polycarbonate films with thicknesses of 10 μm or more, or by choosing an etching time which is short enough to avoid the problem, yet long enough for stable fission fragment counting. In addition, contrary to our first impression, KOH may be somewhat better for alpha suppression if the etching time is selected properly.

Table 12. Effect of polycarbonate absorbers on observed alpha background in 8- μm polycarbonate films

Absorber thickness (mg cm^{-2})	Alpha count ^a
0 mg cm^{-2}	43 $\pm$ 4
0.7	10 $\pm$ 7
1.5	1 $\pm$ 1
2.2	1 $\pm$ 1
3.0	1 $\pm$ 1

^aFoils irradiated 48 hr with 5 μCi of ^{239}Pu and etched 1-1/2 hr in 60°C, 6N NaOH.

Fig. 36. Effect of etch time on alpha sensitivity (a) 60°C, 6N NaOH. (b) 60°C, 6N KOH.

OTHER

DOS11DT, A Program to Decode PDP-11* Disk Operating System (DOS)* Decapes*

The Hazards Control Department has many laboratory instruments that output data in a variety of paper tape formats. Although magnetic tapes have higher capacity and are much faster, they are more expensive and either are not or have not been available for many instruments. Since much of the data from these instruments may be input to computer codes, there has been a need for an efficient data transfer path to the Octopus system.

The Hazards Control PDP-11 system has as peripheral devices a fast, reliable paper tape reader-punch and a Dectape transport. In addition, the

Octopus system has a Dectape transport on one of its interfacing minicomputers. Computations has prepared a program to make a disk image of a complete Dectape, and has written a routine to convert this image for PDP-10* ASCII tapes. With this background information, we decided to prepare a program to decode Decapes for data from all the Hazards Control paper tape sources, including the PDP-11 system. This program is known as DOS11DT and has been tested and used for data transmittal since the latter part of 1973. All of the known paper tape data sources in the Hazards Control Department can be automatically decoded from Dectape using the seven subroutines provided.

Publications

G. O. Nelson, L. E. Swisher, R. D. Taylor and B. E. Bigler, *Preparation of Mixtures of Humidified Air with CO, SO₂, NO₂, NH₃, NH₂CH₃, Cl₂ and HCl to Measure Respirator Cartridge and Canister Service Life*, Lawrence Livermore Laboratory Rept. UCRL-75601 (1974).

Four systems have been fabricated to measure the effective service life of respirator cartridges and canisters. The four systems employ carbon monoxide, sulfur dioxide/nitrogen dioxide, ammonia/methylamine, and chlorine/hydrogen chloride as contaminants. Each system produces test concentrations in the low percentage range (typically 0.05 to 3.0%) at flow rates of 32 and 64 l/min and at relative humidities between 20 and 95%. Suitable analyzers have been provided to measure the downstream concentration from 1 to 500 ppm as a function of time. Holders have been designed to accommodate the wide size variation of commercially available cartridges and canisters. This paper has been submitted for publication in the American Industrial Hygiene Association Journal.

R. J. Kloepping, *The Scintillation Response of Lithium Fluoride*, Lawrence Livermore Laboratory, Rept. UCRL-51528 (1974).

Lithium fluoride crystals, with TIF added to the charge, were prepared by the Bridgman-Stockbarger method. The scintillation response to photon energies between 25 and 1250 keV for these crystals and a commercially prepared LiF (Ti) crystal was studied. The response of the LiF (Ti) crystal was compared to the response of an NE109 scintillator at 1250 keV. The dose response of the LiF (Ti) crystal was evaluated over the energy range 25 to 1250 keV. The absorption and emission spectra of the LiF (Ti) crystal are discussed. A historical review precedes the discussion.

J. R. Gaskin and M. W. Magee, "The HEPA-Filter Smoke Plugging Problem," *Proc. 13th AEC AIR CLEANING CONF.*, San Francisco, Calif. (Aug. 1974).

Actual experiences indicate that during the early stages of a fire, pyrolysis and incomplete combustion of organic materials used in the furnishings or interior finishes of laboratories yield copious quantities of smoke particulates, both liquid and solid. Furthermore, the use of fire retardants in materials used for the above purpose interferes with the combustion process, so that burning of such materials in later stages of a fire will yield dense smoke. These particulates can plug up a HEPA filter or even a more porous prefilter, and thus effectively shut off the exhaust ventilation. In this

case, the fire room will pressurize and contamination may spread in an uncontrolled manner.

Both small- and large-scale tests have been conducted to evaluate the nature and degree of the problem as a function of materials involved, rate of exposure to the fire, and kinds and temperatures of smoke so generated. Some test work has also been done on scrubbing of smoke. Proposed future work is described.

C. L. Lindeken, D. G. Coles and J. W. Meadows, "Ge(Li) Spectrometric Measurements of Isotopic Uranium Ratios in Soils Treated with Apatite-Derived Fertilizers," submitted to the IAEA Symp. on Isotope Ratios as Pollutant Source and Behaviour Indicators, Vienna, 18-22 Nov., 1974, (preprint UCRL-75659).

Soil samples collected in an agricultural area in the Delta region of the San Joaquin Valley in Central California contained levels of uranium as high as 17 ppm. Isotopic uranium analysis determined by mass spectrometry showed ²³⁵U/²³⁸U ratios of 7.1×10^{-3} , typical of natural uranium. However, the ²³⁸U content inferred from gamma counting of the ²²⁶Ra daughters disclosed marked disequilibrium in the ²³⁸U decay series. Elevated uranium is attributed to its presence in ammonium phosphate fertilizer produced from apatite (Ca F₂ Ca₃ (PO₄)₂). ²³⁸U daughter disequilibrium is due to removal of ²²⁶Ra by coprecipitation with CaSO₄ (gypsum) during conversion of the raw material to phosphoric acid with sulphuric acid. A more direct method for ²³⁸U measurement using Ge(Li) gamma spectrometry is described which is based on detection of the 63-keV transition gamma of ²³⁴Th. This method, which is independent of ²²⁶Ra daughters, can detect ²³⁸U down to about 0.1 ppm in soils. A comparison with ²³⁸U measurements using mass spectrometry shows agreement within $\pm 10\%$.

P. H. Gudiksen, D. E. Jones, H. L. Beck, J. E. McLaughlin, T. P. Stuart and O. D. T. Lynch, Jr., *External Dose Estimates for Future Eniwetok Atoll Inhabitants*, (in preparation for publication).

Measurements of environmental gamma radiation levels on Eniwetok Atoll were combined with pertinent population distributions and expected future life styles to assess the external doses that could be received by the Marshallese people upon their return to the former U.S. nuclear test site in the Pacific.

C. A. Harder and G. O. Nelson, *Mobile Water Filtration System Operating Manual*, Lawrence Livermore Laboratory, Rept. M-049 (1974).

The mobile water filtration system is a skid-mounted, multi-stage filter system designed for the removal of moderate to sub-micronic particulate matter from contaminated liquid. The unit is designed as two complete filtration systems, each capable of independent operation. The two filter systems have a common precoat tank separation tank and blow-down system.

The unit is complete with all necessary electrical controls, precoat charger, diatomaceous earth separation tank, and air compressor. A separate shipping container houses spare parts, several bags of diatomaceous earth, and accessories.

The system includes:

- Two single-stage, self-priming centrifugal pumps (each capable of producing full design flow).
- Two 54-ft² diatomaceous earth (D.E.) precoat filters.
- Two 36-unit primary cartridge filters (nominal particle size removal -- 0.5 micron).
- Two 24-unit secondary cartridge filters (0.2-micron absolute removal).

R. G. Purington, "Calculating Flow from a Fire Nozzle Using SI Units," submitted to *SEPE Fire Triangle* (preprint UCRL-75786).

Two nozzle flow equations for use with SI units are derived; one is for nozzle sizes measured by diameter, the other for those measured by area. The SI units used in the equations are litres per second for flow rate (Q), kilopascals for pressure (P), and square millimetres for area (A) or centimetres for diameter (d). The equation using area is $Q = A \sqrt{2P/1000}$. Using the diameter, the equation is $Q = 0.111d^2 \sqrt{2}$. A table of SI equivalents for common nozzle sizes is provided.

References

1. R. V. Osborne, *Development of a Monitor for Tritiated Water Vapour in the Presence of Noble Gases*, Atomic Energy of Canada, Ltd., Rept. AECL-4303 (1972).
2. S. Block, D. Hodgekins and O. Barlow, *Recent Techniques in Tritium Monitoring by Proportional Counters*, Lawrence Livermore Laboratory, Rept. UCRL-51131 (1971).
3. H. H. Rossi and W. Rosenzweig, "A Device for Measurement of Dose as a Function of Specific Ionization," *Radiology* **64**, 404 (1955).
4. I. O. Anderson and J. Braun, "A Neutron Rem Counter with Uniform Sensitivity from 0.025 eV to 10 MEV," *Neutron Dosimetry*, Vol. 2, Proc. Symposium on Neutron Detection, Harwell, December 14, 1962, (IAEA, Vienna, 1963).
5. T. P. Wilcox, *ANISN-L, A CDC-7600 Code which solves the One-dimensional, Multigroup, Time-Independent Transport Equation by the Method of Discrete Ordinates*, Lawrence Livermore Laboratory, Internal Memo MOR-15 (1974). Readers outside the Laboratory who desire further information on LLL internal documents should address their inquiries to the Technical Information Department, Lawrence Livermore Laboratory, Livermore, CA 94550.
6. R. L. Bramblett, R. I. Ewing and T. W. Bonner, "A New Type of Neutron Spectrometer," *Nucl. Instr. Methods*, **2**, 1 (1960).
7. J. T. Routti, "Mathematical Considerations of Determining Neutron Spectra from Activation Measurements," *Proc. Second Intern. Symp. on Accelerator Dosimetry and Experience*, Stanford University, November 5-7, 1969, paper CONF 691101.
8. *Hazards Control Progress Report No. 47*, Lawrence Livermore Laboratory, Rept. UCRL-50007-73-3 (1973) p. 14-18.
9. *Hazards Control Progress Report No. 45*, Lawrence Livermore Laboratory, Rept. UCRL-50007-73-1 (1973) p. 18-19.
10. *Hazards Control Progress Report No. 46*, Lawrence Livermore Laboratory, Rept. UCRL-50007-73-2 (1973) p. 1-3.
11. R. S. Sanna, *Thirty-one Group Response Matrices for the Multisphere Neutron Spectrometer Over the Range Thermal to 400 MeV*, USAEC Health and Safety Laboratory, Rept. HASL-267 (1973).
12. ICRP Publication 21, *Data for Protection Against Ionizing Radiation from External Sources: Supplemental to ICRP Publication 15. The International Commission on Radiological Protection*, (Pergamon Press, 1971).
13. G. O. Nelson and C. A. Harder, "Effect of Concentration on Respirator Cartridge Efficiency," *Hazards Control Progress Report No. 44*, Lawrence Livermore Laboratory, Rept. UCRL-50007-72-3 (1972).
14. G. O. Nelson and C. A. Harder, "Respirator Cartridge Efficiency Studies: IV. Effects of Steady-State and Pulsating Flow," *Am. Ind. Hygiene Assn. J.*, **33**, 797 (1972).
15. J. Barbier, "Contrast Improvement of Images Obtained in Cellulose Nitrate Film by Trunk Film Methods," *Trans. Am. Nucl. Soc.* **13**, 530 (1970).
16. R. L. Fleishcher and R. M. Walker, *Radiation Detection Method* U.S. Patent 3 770 762, 6 Nov. 1973.
17. C. T. Prevo and J. L. Cate, "A Practical Solid-State Beta Spectrometer," *Nucl. Instr. Methods*, **ss**, 1967.
18. C. T. Prevo and J. L. Cate, "A Solid-State Beta Spectrometer," *Hazards Control Progress Report No. 25*, Lawrence Livermore Laboratory, Rept. UCRL-50007-66-1 (1966).
19. G. O. Nelson and K. S. Griggs, *A Dynamic Method for Producing Known Concentrations of Gas and Solvent Vapor in Air*, Lawrence Livermore Laboratory, Rept. UCRL-70394 (1967).
20. J. H. Elliott, W. L. Pickles and D. R. Slaughter, "Neutron Spectrometry," *Hazards Control Progress Report No. 46*, Lawrence Livermore Laboratory, Rept. UCRL-50007-73-2 (1973).
21. D. R. Slaughter, J. H. Elliott, "Neutron Spectrometer," *Hazards Control Progress Report No. 47*, Lawrence Livermore Laboratory, Rept. UCRL-50007-73-3 (1973).
22. M. E. Anderson and R. A. Neff, *Nucl. Instr. Methods* **99**, 231 (1971).
23. F. A. Lorch, *Intern. J. Appl. Radiation Isotopes*, **24**, 585 (1973).
24. P. W. Benjamin, C. D. Kemshall and A. Brickstock, *Analysis of Recoil Proton Spectra*, Atomic Weapons Research Establishment, U.K., AWRE Rept. No. 0-9/68.
25. E. F. Bennett and T. J. Yule, *Nucl. Instr. Methods* **98**, 393 (1972).
26. V. V. Verbinski et al., *Nucl. Instr. Methods* **65**, 8 (1968).
27. C. D. Schwartz and G. E. Owen, *Recoil Detection in Scintillators, Fast Neutron Physics Part I*, J. B. Marion and J. L. Fowler, Eds. (Interscience Publishers, New York, 1960).

28. H. W. Broek and C. E. Anderson, *Rev. Sci. Instr.* **31**, 1063 (1960).
29. C. Hoenig, Lawrence Livermore Laboratory, private communication.
30. *Improvements In and Relating To Decontamination Processes for Radioactive Liquids*, Belgonucleaire, S.A., Brussels, Belgium. British Patent Application No. 41723/70, filed Sept. 1, 1970.
31. P. R. Bevington, *Data Reduction and Error Analysis for the Physical Sciences* (McGraw-Hill, New York, 1969).
32. *Industrial Ventilation*, available from the Committee on Industrial Ventilation, P.O. Box 453, Lansing, Michigan 48902.
33. H. J. McDermott and A. J. Toy, *A Computer Program to Design Exhaust Ventilation Systems*, Lawrence Livermore Laboratory, Rept. UCRL-50007-73-2 (1973), p. 18.
34. *Hazards Control Progress Report No. 39*, Lawrence Livermore Laboratory, Rept. UCRL-50007-71-1 (1971).
35. W. G. Cross and L. Tommasino, *Proc. Intern. Conf. Nucl. Track Registration in Insulating Solids*, (Clermont-Turand, 1969).
36. R. L. Fleischer, P. B. Price, R. M. Walker and E. L. Hubbard, *Phys. Rev.* **133**, (5A), A1443 (1964).