**MULTIWIRE PROPORTIONAL CHAMBERS IN NUCLEAR MEDICINE:
PRESENT STATUS AND PERSPECTIVES**

V. Perez-Mendez*, L. Kaufman, M.R. Powell[†], D.C. Price

**University of California
San Francisco**

and

**L. Blumin, R. Cavalieri,
Veterans' Administration Hospital
San Francisco**

NOTICE

This report was prepared as an account of work sponsored by the United States Government. Neither the United States nor the United States Atomic Energy Commission, nor any of their employees, nor any of their contractors, subcontractors, or their employees, makes any warranty, express or implied, or assumes any legal liability or responsibility for the accuracy, completeness or usefulness of any information, apparatus, product or process disclosed, or represents that its use would not infringe privately owned rights.

MAS

ABSTRACT

The construction and operating characteristics of Xenon filled Multi-wire Proportional Chambers (MWPC) for Nuclear Medicine imaging are described. The resolution and image quality of these devices is demonstrated by various images taken with γ rays ranging in energy from 28 keV to 140 keV. From some measurements done with a 4 atmosphere chamber, it is estimated that 10 atmosphere Xenon filled chambers could provide a satisfactory combination of resolution (2.m.m.) and detector efficiency for use with γ rays in this energy range.

[†]Present address: 350 Parnassus Ave., San Francisco, Ca.

* Also at the Lawrence Berkeley Laboratory, Berkeley, California.

Multiwire proportional chambers (MWPC) were originally developed as position sensitive detectors for particle physics research. When filled with Xenon at pressures ranging from 1 - 10 atmospheres, they are excellent detectors for gamma ray distributions and can provide intrinsic spatial resolutions better than 2 mm for photon energies between 25 and 140 kev. with present data rates up to 10^5 events/sec which allows for dynamic studies.

The basic chamber (1) is shown schematically in fig.1. It consists of three grids of parallel wires enclosed in a gas-tight envelope filled with a Xenon 93% - Methane 7% mixture. The spacing between grids varies from 3-10 millimeters, and that between wires varies from 1-3 millimeters, depending on the size of the chamber and the spatial resolution required. The center grid, on which the avalanche multiplication occurs, consists of gold-plated tungsten wires typically 15-25 microns in diameter; this grid is held at a high positive potential relative to the two outside ones. The wires of the central grid are connected together and the pulse-height information obtained from it, allows for energy selection of the gamma ray interacting with the Xenon. The two outer grids which are used to obtain the spatial information consist of larger diameter wires, 50 microns or larger. In order to provide a rectilinear system of coordinates these two grids have their wire axes mutually orthogonal.

The principle of operation is the following: a α ray photon entering the active volume of the chamber through the collimator has a definite probability for interacting with the Xenon gas by the photoelectric effect. The ensuing photoelectron will deposit its energy in a small region mostly in the form of ionization. The probability of interaction and the localization of the secondary electron both improve with increase in density

(pressure) of the Xenon gas. The electric field between the wire planes will cause the secondary electrons to drift towards the anode plane where they will undergo avalanche multiplication in a small region surrounding a wire and produce an amplified pulse on it whose amplitude will be proportional to the energy of the photoelectron. Simultaneously, an induced pulse of the opposite polarity will be generated on the cathode wires by the motion of the ions produced in the avalanche.

The coordinates of an ionizing event are determined by the use of electromagnetic delay lines capacitively coupled to each of the orthogonal planes of cathode wires (1,2). The distance X or Y of an ionizing event occurring in the chamber is proportional to the time interval between the onset of the avalanche pulse and the arrival of the pulse at the end of the appropriate delay line. In order to enable the cathode plane signals to produce voltage signals of the optimum shape for electronic timing, the wires are decoupled from each other and their common bus bar through 200 k isolating resistors. The construction and characteristics of our delay lines have been described in detail elsewhere (2). Alternative readout methods using resistive delay lines have also been developed (3).

A signal that indicates the occurrence of an ionizing event is obtained from the central plane through an RC network. This signal is processed by the technique of differentiation and zero-crossing, and is used to start two time-to-amplitude converters. Similarly processed signals obtained from each delay line are used to stop the converters, one from the x-coordinate line and one from the y-coordinate line. Analog image reconstruction and display (fig. 2a) is done by using the outputs from the converters to drive the x-y deflection plates of an oscilloscope. The

resultant pattern of dots may then be photographed with a Polaroid camera.

An alternative method of recording the images (fig. 2b) is to digitize the time-delay signals from the chamber and then store the coordinates of each event on magnetic tape, disc, or random-access core of a computer.

We have built a number of chambers operating at atmospheric pressure ranging in size up to 50 x 50 cm for x-ray transmission imaging, and for gamma rays ranging from 28 kev to 80 kev, in nuclear medicine applications.

Following initial phantom studies, in vivo evaluation of several radioisotopes and radiopharmaceuticals is now under way with the single atmosphere MMPC in animals and in man. Iodine-125 (28 kev x-ray; 35 kev α) is useful for Thyroid imaging. Xenon-133 (81 kev x-ray) is well imaged in ventilatory evaluation of the lungs. Hg-197 chlormerodrin (77 kev x-ray) gives excellent images of the kidneys, although hampered somewhat by tissue depth. Dysprosium-159 (44.5 kev Terbium x-ray) complexed to HEDTA and to polyphosphate has demonstrated excellent skeletal detail in the rabbit. Cerium-141 (36 kev Praseodymium x-ray) as the hydroxycitrate forms a colloid which localizes in liver, spleen and bone marrow, resulting in excellent images obtained in the rabbit. In fig. 3 we show the image of a human thyroid taken with I-125. In fig. 4 we show the image of a rabbit taken with Dy-159. Fig. 5 shows images of the liver and bone marrow of a rabbit labelled with Ce-141 Hydroxycitrate. In all of these cases we used a Nuclear Chicago low energy high resolution collimator, and the good detail obtained is evident in the pictures.

For imaging with the 140 kev gamma from Tc-99m we built a pressurized chamber (4 atmosphere) with sensitive area of 20 x 20 cm. Figure 6 shows the detection efficiency as a function of gamma ray energy and

Figure 7 shows pictures of a bar pattern taken with 60 kev and 140 kev gamma rays. Larger chambers (30 x 30 cm.) to be used at 10 atmospheric pressure are under construction.

Figure 8 shows images of a Picker Thyroid Phantom filled with Tc-99m and of a rat skeleton and liver/spleen labeled with Tc-99m Polyphosphate and Sulfur Colloid respectively.

The spatial resolution of images obtained with the MMPC is limited by the characteristics of the multi-hole collimators currently used in Nuclear Medicine. The use of coded apertures--such as Fresnel zone plate (5) or multi-pinhole arrays (6) are especially attractive for use with MMPC since they allow for imaging with good resolution, provide images with tomographic information and make optimum use of the sensitive areas of the larger MMPC.

Imaging of gamma rays with energies above 140 kev can be accomplished by use of lead converters. Since flat lead converters have low efficiencies, their surface area--hence their detection efficiency--can be increased a factor of 10 by giving them a honeycomb shape with suitable electric fields to drift the conversion electrons into the avalanche region of the MMPC. A positron camera consisting of two MMPC, 48 x 48 cms sensitive area with honeycomb lead converters has been constructed at the Lawrence Berkeley Laboratory (7) and is being used at the Nuclear Medicine Section U.C.S.F.

The detection efficiency for the 0.51 Mev annihilation photons is 5% for each chamber with two such converters coupled to each chamber. With the two detectors spaced at 50 cms, the spatial resolution at the mid-plane is 6 mm FWHM.

Other configurations of MMPC such as linear, i.e., one-dimensional

chambers, have been built for measuring density distribution of tissue by gamma ray transmission measurements, using the 60 kev gamma rays from Am-241, in a transaxial tomography configuration. (8)

SUMMARY

Xenon-filled MWPC provide high resolution imaging of x-rays and gamma rays. The main deficiency is the low sensitivity of chambers operating at atmospheric pressure. Atmospheric pressure chambers in sizes ranging up to 50 x 50 cm. have been built and used to image radioisotopes emitting gamma rays below 60 kev. They have also been used in x-ray transmission imaging to provide digitized images which can then be processed by computer. Research is underway to evaluate the advantages and net gain in detection efficiency that may be realized with the use of coded apertures as compared to the conventional multi-hole collimators.

For imaging higher energy gamma rays such as the 140 kev line of Tc-99 pressurized chambers are necessary. For higher energy gamma rays such as the 0.51 Mev lines from positron annihilation, MWPC at atmospheric pressure with honeycomb lead converters operating as a positron camera become very useful. Their main advantage over conventional scintillation positron cameras is the ease and low cost of constructing large area devices. The main disadvantage of lead converter detection is that no energy discrimination is possible, and hence necessitates the use of positron emitters which are not accompanied by a large fraction of gamma rays in the energy region from 200 kev - 1.5 Mev. This deficiency may not be too serious in practice since accidental coincidences tend to have a fairly uniform distribution in space, and the computer logic which records the

annihilation events is programmed to subtract backgrounds of this sort. Furthermore, there are many medically useful isotopes which are pure positron emitters such as C-11, N-13, F-18, Ga-68, and others.

ACKNOWLEDGMENT

This work has been supported in part by the U. S. Atomic Energy Commission, and by a Grant from the Veterans Administration Hospital, San Francisco.

REFERENCES

- (1) Multiwire Proportional Chambers for Biomedical Application, S. N. Kaplan, L. Kaufman, V. Perez-Mendez and K. Valentine. Nuc. Inst. & Methods 106 (1973) 397-406.
- (2) Improved Delay Lines for Wire Chamber Readout, R. Grove, V. Perez-Mendez and J. Sperinde. Nuc. Inst. & Methods 106 (1973) 407-408.
- (3) Proportional Counter Photon Camera, C. J. Borkowski and M. K. Kopp. IEEE Trans. Nuc. Sci. NS-19 (1972) 161-168.
- (4) Performance of a Pressurized Xenon-Filled Multiwire Proportional Chamber, L. Kaufman, V. Perez-Mendez and G. Stoker. IEEE Trans. Nuc. Sci. NS-20 (1973) 426-428.
- (5) Gamma Ray Imaging Using a Fresnel Zone Plate Aperture, Multiwire Proportional Chamber and Computer Reconstruction, B. Macdonald, L. T. Chang, V. Perez-Mendez and L. Shiraishi. IEEE Trans. Nuc. Sci. NS-21 (1974) 678-684.
- (6) A Method of Tomographic Imaging Using a Multiple Pinhole Coded Aperture, L. T. Chang, S. Kaplan, B. Macdonald, V. Perez-Mendez and L. Shiraishi. Lawrence Berkeley Laboratory Report LBL 3001 (April 1974). To be published in J. Nucl. Med.
- (7) Multiwire Proportional Chambers for Positron Imaging, C. B. Lim, D. Chu, L. Kaufman, and V. Perez-Mendez. IEEE Trans. Nuc. Sci. NS-21 (1974) 85-88.

- (8) A High Pressure Xenon Multiwire Proportional Chamber for x-ray Laminographic Reconstruction using Fan Beam Geometry. D. Boyd, J. Conrad, J. Dehnert, D. Chu, C. Lim, B. Macdonald, V. Perez-Mendez. IEEE Trans. Nuc. Sci. NS-21 (1974) 184-187.

FIGURE CAPTIONS

- Fig. 1. Schematic diagram of Xenon-filled chamber (a) side view, (b) top view showing wire planes and delay lines.
- Fig. 2. Electronics block diagrams for (a) an analog system with scope display and (b) digital system with storage capabilities.
- Fig. 3. Image of human thyroid labelled with Iodine-125.
- Fig. 4. Image of rabbit labelled with Dy-159 HEDTA and Dy-159 polyphosphate. The vertebrae of the rabbit are seen clearly.
- Fig. 5. Image of rabbit liver and bone marrow labelled with Ce-141 Hydroxy Citrate.
- Fig. 6. Detection efficiency of 4 atmosphere Xenon-filled chamber versus gamma energy. Thickness of chamber including drift regions is 6.4 cms.
- Fig. 7. Transmission bar patterns taken with 4 atmosphere chamber: (a) 10 mm, (b) 4 mm, (c) 2 mm, (d) 1 mm -- taken at 60 kev.; (e) 10 mm, (f) 4 mm, (g) 2 mm -- taken at 140 kev..
- Fig. 8. Top image is of Picker Thyroid Phantom filled with Tc-99 solution (300 k counts). Lower images of rat skeleton and liver/spleen labeled with Tc-99m polyphosphate (300 k counts) and Tc - 99m Sulfur Colloid (100k counts).

XBL 744-771

Fig. 1.

XBL 723-639

Fig. 2.

Multiwire Proportional Chamber

Rectilinear Scanner

Functioning
Nodule

10 K, 5 min

Hot Spot = 1 K cpm
37 hole 3" focus collimator, 30 cm/min

Cold Nodule
(Follicular
adenoma,
cystic
degeneration

17.4 K, 10 min

Hot Spot = 1 K cpm
37 hole 3" focus collimator, 30 cm/min

XBB 736-3980

Fig. 3.

MULTI-WIRE PROPORTIONAL CHAMBER

Rabbit

^{159}Dy -HEDTA

24 hours post

10 min

XBB 729-4509

^{159}Dy -Polyphosphate

120 hours post

10 min

Fig. 4.

MULTI-WIRE PROPORTIONAL CHAMBER

141 Cerium-Hydroxy-Citrate

Liver (800K)

XBB 729-4508
Bone Marrow (200K)

Fig. 5.

XBL7210-4180

Fig. 6.

Fig. 7.

THYROID PHANTOM ^{99m}Tc 100 K

RAT LIVER/SPLEEN ^{99m}Tc Sulfur Colloid
100 K

XBB 748-5949
RAT ^{99m}Tc POLYPHOSPHATE 300 K

Fig. 8.