C
325
12-7-62

ORNL-3090
UC-34 - Physics
TID-4500 (17th ed., Rev.)

PULSE HEIGHT CALCULATIONS FOR A PARALLEL
PLATE IONIZATION CHAMBER CONTAINING
ELECTRON ATTACHING GASES

H. B. Eldridge

OAK RIDGE NATIONAL LABORATORY

operated by

UNION CARBIDE CORPORATION

for the

U.S. ATOMIC ENERGY COMMISSION

DISCLAIMER

This report was prepared as an account of work sponsored by an agency of the United States Government. Neither the United States Government nor any agency Thereof, nor any of their employees, makes any warranty, express or implied, or assumes any legal liability or responsibility for the accuracy, completeness, or usefulness of any information, apparatus, product, or process disclosed, or represents that its use would not infringe privately owned rights. Reference herein to any specific commercial product, process, or service by trade name, trademark, manufacturer, or otherwise does not necessarily constitute or imply its endorsement, recommendation, or favoring by the United States Government or any agency thereof. The views and opinions of authors expressed herein do not necessarily state or reflect those of the United States Government or any agency thereof.

DISCLAIMER

Portions of this document may be illegible in electronic image products. Images are produced from the best available original document.

Printed in USA. Price: \$0.50 Available from the
Office of Technical Services
U. S. Department of Commerce
Washington 25, D. C.

— LEGAL NOTICE —

This report was prepared as an account of Government sponsored work. Neither the United States, nor the Commission, nor any person acting on behalf of the Commission:

- A. Makes any warranty or representation, expressed or implied, with respect to the accuracy, completeness, or usefulness of the information contained in this report, or that the use of any information, apparatus, method, or process disclosed in this report may not infringe privately owned rights; or
- B. Assumes any liabilities with respect to the use of, or for damages resulting from the use of any information, apparatus, method, or process disclosed in this report.

As used in the above, "person acting on behalf of the Commission" includes any employee or contractor of the Commission, or employee of such contractor, to the extent that such employee or contractor of the Commission, or employee of such contractor prepares, disseminates, or provides access to, any information pursuant to his employment or contract with the Commission, or his employment with such contractor.

ORNL-3090

Contract No. W-7405-eng-26

HEALTH PHYSICS DIVISION

PULSE HEIGHT CALCULATIONS FOR A PARALLEL PLATE
IONIZATION CHAMBER CONTAINING ELECTRON ATTACHING GASES

H. B. Eldridge

Date Issued

NOV 21 1962

OAK RIDGE NATIONAL LABORATORY
Oak Ridge, Tennessee
operated by
UNION CARBIDE CORPORATION
for the
U.S. ATOMIC ENERGY COMMISSION

**THIS PAGE
WAS INTENTIONALLY
LEFT BLANK**

CONTENTS

	Page
Abstract	1
Introduction and Method	1
Results	3

PULSE HEIGHT CALCULATIONS FOR A PARALLEL PLATE IONIZATION
CHAMBER CONTAINING ELECTRON ATTACHING GASES

H. B. Eldridge

ABSTRACT

When the pulse from a parallel plate electrode ionization chamber containing an attaching gas is examined with a linear pulse amplifier having the proper integration and differentiation time constants, the pulse height will be a simple analytic function of the attachment coefficient and the electron drift velocity.¹⁻³ In this report the expressions for the pulse height as a function of the partial pressure of the attaching gas, electron drift velocity, attachment coefficient, chamber geometry and amplifier time constants are evaluated and presented in tabular form.

INTRODUCTION AND METHOD

The time variation of the electrode potential in a parallel plane electrode ionization chamber containing an attaching gas is given by:¹

$$g(t) = \frac{A}{\alpha f_1 P d} [1 - \exp(-\alpha f_1 P W t)]$$

Since α , f_1 , P , W , and d do not vary with time, one may write

$$g(t) = \frac{A}{F} [1 - \exp(-ft/\tau_0)]$$

where $f = \alpha f_1 P d$; $\tau_0 = d/W$; W = drift velocity of electrons; α = attachment coefficient; $f_1 P$ = partial pressure of attaching gas; and d = distance from electron source to collecting electrode.

If this pulse is observed with a linear pulse amplifier having a step function response given by $1(t/t_1) \exp(-t/t_1)$, where t_1 = amplifier differentiating and integrating time constant, then the output pulse will be given by

$$V(\tau) = \int_0^{\tau} \frac{dg(t)}{dt} \frac{(\tau-t)}{t_1} \exp[-(\tau-t)/t_1] dt \quad \text{for } \tau < \tau_0$$

and by

$$V(\tau) = \int_0^{\tau_0} \frac{dg(t)}{dt} \frac{(\tau-t)}{t_1} \exp[-(\tau-t)/t_1] dt \quad \text{for } \tau > \tau_0$$

Direct integration of these equations gives

$$V(\tau) = \frac{A \exp(-\tau/t_1)}{(\tau_0 - t_1 f)} \left[\frac{\exp(u\tau)}{u} - \tau - \frac{1}{u} \right] \quad \text{for } \tau < \tau_0 \quad (1)$$

and

$$V(\tau) = \frac{A \exp(-\tau/t_1)}{(\tau_0 - t_1 f)} \left[\{ \exp(u\tau_0) - 1 \} \tau - \exp(u\tau_0) \left(\tau_0 - \frac{1}{u} \right) - \frac{1}{u} \right] \quad (2)$$

for $\tau > \tau_0$

where $u = (\tau_0 - t_1 f)/t_1 \tau_0$. For convenience in what follows, let $u = \tau/t_1$, $x = \tau_0/t_1$, and $y = \tau_0/t_1 - f$. To find the pulse height, one must find the maximum of the function $V(u)$ which may be accomplished by evaluating $V(u')$ where u' is given by the equations,

$$\frac{dV(u')}{du} = 0 \quad \text{and} \quad \frac{d^2V(u')}{du^2} < 0 \quad (3)$$

However, the physical nature of the problem makes the second equation's evaluation unnecessary. The first derivative is indeterminate at $f = x$ and is evidenced in the table to follow. The maximum may be evaluated at $x = f = 0$ and is e^{-1} . When Eq. (3) is evaluated for the two regions of u , one gets for the position of the maximum of $V(u)$,

$$u' = (x/y - 1) [\exp\{(y/x)u'\} - 1] \quad \text{for } u' < x \quad (4)$$

$$u' = \frac{\exp(y) (1 + x - x/y) - 1 + x/y}{\exp(y) - 1} \quad \text{for } u' \geq x \quad (5)$$

One may note that Eq. (5) gives u' explicitly while Eq. (4) is transcendental and must be solved numerically.

A program was developed for the IBM 7090 computer which found the pulse height as a function of x and f . The principal scheme was to evaluate u' using Eq. (5) and then determine if $u' \geq x$. If $u' \geq x$, one found the pulse height by substituting the u' so found into expression (2). However, if $u' < x$, when found using expression (5), expression (4) was used to determine a u' which was substituted into Eq. (1).

RESULTS

Pulse heights were computed for the following ranges of the parameters x and f :

$$0 \leq x \leq 5, \quad \text{in increments of } 1.0$$

$$0 \leq f \leq 5, \quad \left\{ \begin{array}{ll} \text{in increments of } 0.01 & 0 \leq f \leq 1.0 \\ \text{in increments of } 0.02 & 1.0 < f \leq 5.0 \end{array} \right\}$$

The results are summarized in the following table where the elements of

the first column are different values of f , the elements of the first row are different values of x , and the elements of the second row are the normalization factors, N . For each value of x the pulse heights are normalized to the pulse height corresponding to $f = 0$, hence N equals the pulse height at $f = 0$. The remaining elements are the normalized pulse heights corresponding to the f which is the first element in the row and the x which is the first element in the column. Since the normalized pulse height is generally the quantity measured in an experiment, this quantity was tabulated.

Table I. Normalized Response of an Amplifier to Pulses from a Plane Parallel Ionization Chamber in Which Electron Capture Takes Place^{*}

X	0.0	1.0	2.0	3.0	4.0	5.0
$\frac{N}{f}$	0.367879	0.353224	0.316132	0.270666	0.227772	0.192027
0.	1.000000	1.000000	1.000000	1.000000	1.000000	1.000000
0.010	0.995016	0.994990	0.994833	0.994508	0.994075	0.993615
0.020	0.990066	0.990015	0.989701	0.989056	0.988195	0.987279
0.030	0.985149	0.985072	0.984605	0.983644	0.982357	0.980991
0.040	0.980264	0.980163	0.979545	0.978270	0.976564	0.974750
0.050	0.975411	0.975287	0.974520	0.972935	0.970813	0.968557
0.060	0.970591	0.970443	0.969530	0.967638	0.965105	0.962411
0.070	0.965802	0.965632	0.964574	0.962380	0.959439	0.956312
0.080	0.961046	0.960853	0.959653	0.957159	0.953815	0.950258
0.090	0.956320	0.956107	0.954766	0.951976	0.948232	0.944251
0.100	0.951626	0.951391	0.949912	0.946830	0.942691	0.938289
0.110	0.946962	0.946708	0.945093	0.941721	0.937191	0.932373
0.120	0.942330	0.942055	0.940306	0.936649	0.931732	0.926501
0.130	0.937727	0.937434	0.935553	0.931613	0.926313	0.920674
0.140	0.933155	0.932844	0.930832	0.926613	0.920934	0.914891
0.150	0.928613	0.928284	0.926144	0.921649	0.915594	0.909151
0.160	0.924101	0.923754	0.921489	0.916721	0.910294	0.903455
0.170	0.919619	0.919254	0.916865	0.911827	0.905033	0.897802
0.180	0.915165	0.914785	0.912273	0.906969	0.899811	0.892192
0.190	0.910741	0.910345	0.907713	0.902146	0.894628	0.886624

^{*} Elements of first column are values of f . Elements of first row are values of X . Elements of second row are normalization factors, N . Remaining elements are normalized pulse heights corresponding to particular values of X and f .

Table I. Continued

X	0.0	1.0	2.0	3.0	4.0	5.0
$\frac{N}{f}$	0.367879	0.353224	0.316132	0.270666	0.227772	0.192027
0.200	0.906346	0.905934	0.903184	0.897357	0.889482	0.881098
0.210	0.901980	0.901553	0.898687	0.892602	0.884375	0.875662
0.220	0.897642	0.897201	0.894220	0.887881	0.879305	0.870343
0.230	0.893332	0.892877	0.889784	0.883194	0.874272	0.865135
0.240	0.889051	0.888582	0.885378	0.878540	0.869276	0.860032
0.250	0.884797	0.884316	0.881003	0.873919	0.864317	0.855031
0.260	0.880571	0.880077	0.876657	0.869332	0.859395	0.850125
0.270	0.876372	0.875867	0.872341	0.864777	0.854508	0.845312
0.280	0.872201	0.871684	0.868055	0.860254	0.849658	0.840587
0.290	0.868057	0.867528	0.863798	0.855763	0.844843	0.835946
0.300	0.863939	0.863400	0.859570	0.851305	0.840063	0.831387
0.310	0.859849	0.859300	0.855371	0.846878	0.835319	0.826906
0.320	0.855784	0.855226	0.851200	0.842482	0.830609	0.822501
0.330	0.851746	0.851178	0.847058	0.838118	0.825934	0.818168
0.340	0.847734	0.847158	0.842944	0.833785	0.821293	0.813904
0.350	0.843748	0.843163	0.838858	0.829482	0.816685	0.809709
0.360	0.839788	0.839195	0.834800	0.825210	0.812112	0.805579
0.370	0.835853	0.835253	0.830769	0.820968	0.807572	0.801512
0.380	0.831944	0.831336	0.826766	0.816756	0.803065	0.797506
0.390	0.828059	0.827445	0.822790	0.812574	0.798591	0.793559

Table I. Continued

X	0.0	1.0	2.0	3.0	4.0	5.0
$\frac{N}{f}$	0.367879	0.353224	0.316132	0.270666	0.227772	0.192027
0.400	0.824200	0.823579	0.818841	0.808422	0.794150	0.789669
0.410	0.820365	0.819739	0.814919	0.804298	0.789744	0.785836
0.420	0.816555	0.815923	0.811023	0.800204	0.785397	0.782056
0.430	0.812770	0.812132	0.807153	0.796139	0.781115	0.778329
0.440	0.809008	0.808366	0.803310	0.792103	0.776897	0.774652
0.450	0.805271	0.804625	0.799492	0.788095	0.772741	0.771025
0.460	0.801557	0.800907	0.795701	0.784115	0.768645	0.767447
0.470	0.797867	0.797214	0.791935	0.780163	0.764606	0.763915
0.480	0.794201	0.793544	0.788194	0.776239	0.760625	0.760429
0.490	0.790558	0.789898	0.784478	0.772343	0.756698	0.756987
0.500	0.786939	0.786276	0.780788	0.768474	0.752825	0.753589
0.510	0.783342	0.782677	0.777122	0.764632	0.749004	0.750232
0.520	0.779768	0.779102	0.773481	0.760817	0.745234	0.746917
0.530	0.776217	0.775549	0.769864	0.757029	0.741513	0.743643
0.540	0.772688	0.772019	0.766271	0.753267	0.737841	0.740407
0.550	0.769182	0.768512	0.762703	0.749532	0.734216	0.737210
0.560	0.765698	0.765028	0.759158	0.745823	0.730637	0.734050
0.570	0.762236	0.761565	0.755638	0.742140	0.727103	0.730927
0.580	0.758796	0.758125	0.752140	0.738483	0.723612	0.727839
0.590	0.755377	0.754707	0.748666	0.734851	0.720164	0.724787

Table I. Continued

X	0.0	1.0	2.0	3.0	4.0	5.0
$\frac{N}{f}$	0.367879	0.353224	0.316132	0.270666	0.227772	0.192027
0.600	0.751981	0.751311	0.745216	0.731245	0.716759	0.721768
0.610	0.748605	0.747937	0.741788	0.727664	0.713394	0.718783
0.620	0.745251	0.744584	0.738383	0.724108	0.710069	0.715831
0.630	0.741918	0.741253	0.735001	0.720577	0.706783	0.712911
0.640	0.738606	0.737942	0.731641	0.717070	0.703535	0.710022
0.650	0.735314	0.734653	0.728304	0.713588	0.700325	0.707164
0.660	0.732043	0.731385	0.724989	0.710130	0.697152	0.704336
0.670	0.728793	0.728138	0.721695	0.706697	0.694014	0.701537
0.680	0.725563	0.724911	0.718424	0.703287	0.690911	0.698767
0.690	0.722354	0.721705	0.715174	0.699900	0.687843	0.696026
0.700	0.719164	0.718519	0.711946	0.696538	0.684808	0.693312
0.710	0.715994	0.715353	0.708739	0.693199	0.681807	0.690626
0.720	0.712844	0.712207	0.705554	0.689882	0.678837	0.687966
0.730	0.709714	0.709081	0.702389	0.686589	0.675900	0.685333
0.740	0.706603	0.705975	0.699246	0.683319	0.672993	0.682726
0.750	0.703511	0.702888	0.696123	0.680072	0.670118	0.680143
0.760	0.700439	0.699821	0.693020	0.676847	0.667272	0.677586
0.770	0.697386	0.696773	0.689938	0.673644	0.664455	0.675053
0.780	0.694351	0.693745	0.686877	0.670463	0.661667	0.672544
0.790	0.691336	0.690735	0.683835	0.667305	0.658908	0.670058

Table I. Continued

X	0.0	1.0	2.0	3.0	4.0	5.0
N	0.367879	0.353224	0.316132	0.270666	0.227772	0.192027
f	0.800	0.688339	0.687744	0.680814	0.664168	0.656176
	0.810	0.685360	0.684772	0.677812	0.661053	0.653471
	0.820	0.682400	0.681819	0.674830	0.657959	0.650794
	0.830	0.679459	0.678884	0.671867	0.654887	0.648142
	0.840	0.676535	0.675967	0.668924	0.651836	0.645517
	0.850	0.673630	0.673069	0.666000	0.648806	0.642917
	0.860	0.670742	0.670189	0.663095	0.645802	0.640342
	0.870	0.667872	0.667327	0.660209	0.642829	0.637791
	0.880	0.665019	0.664482	0.657342	0.639887	0.635265
	0.890	0.662185	0.661655	0.654494	0.636975	0.632762
	0.900	0.659367	0.658846	0.651664	0.634094	0.630282
	0.910	0.656567	0.656053	0.648853	0.631242	0.627826
	0.920	0.653784	0.653280	0.646060	0.628419	0.625392
	0.930	0.651018	0.650522	0.643285	0.625624	0.622980
	0.940	0.648268	0.647781	0.640528	0.622857	0.620590
	0.950	0.645536	0.645054	0.637789	0.620118	0.618222
	0.960	0.642820	0.642344	0.635068	0.617406	0.615875
	0.970	0.640121	0.639652	0.632364	0.614720	0.613549
	0.980	0.637438	0.636953	0.629678	0.612060	0.611243
	0.990	0.634771	0.634218	0.627009	0.609426	0.608957
						0.624727

Table I. Continued

X	0.0	1.0	2.0	3.0	4.0	5.0
$\frac{f}{N}$	0.367879	0.353224	0.316132	0.270666	0.227772	0.192027
1.000	0.632121-0.		0.624357	0.606817	0.606691	0.622655
1.020	0.626868	0.626445	0.619105	0.601674	0.602218	0.618562
1.040	0.621678	0.621292	0.613921	0.596626	0.597821	0.614532
1.060	0.616551	0.616188	0.608802	0.591672	0.593497	0.610565
1.080	0.611486	0.611148	0.603750	0.586809	0.589245	0.606659
1.100	0.606481	0.606166	0.598761	0.582032	0.585063	0.602812
1.120	0.601536	0.601246	0.593837	0.577341	0.580949	0.599023
1.140	0.596650	0.596385	0.588975	0.572733	0.576900	0.595290
1.160	0.591822	0.591582	0.584174	0.568204	0.572915	0.591612
1.180	0.587052	0.586838	0.579435	0.563753	0.568992	0.587986
1.200	0.582338	0.582150	0.574755	0.559378	0.565130	0.584413
1.220	0.577680	0.577519	0.570135	0.555076	0.561327	0.580890
1.240	0.573077	0.572943	0.565573	0.550846	0.557581	0.577417
1.260	0.568529	0.568422	0.561068	0.546685	0.553892	0.573992
1.280	0.564033	0.563955	0.556620	0.542592	0.550257	0.570614
1.300	0.559591	0.559541	0.552228	0.538564	0.546675	0.567281
1.320	0.555201	0.555179	0.547890	0.534601	0.543145	0.563994
1.340	0.550861	0.550870	0.543607	0.530700	0.539666	0.560750
1.360	0.546573	0.546611	0.539377	0.526860	0.536236	0.557549
1.380	0.542334	0.542402	0.535199	0.523079	0.532854	0.554389

Table I. Continued

X	0.0	1.0	2.0	3.0	4.0	5.0
$\frac{N}{f}$	0.367879	0.353224	0.316132	0.270666	0.227772	0.192027
1.400	0.538145	0.538243	0.531074	0.519356	0.529519	0.551271
1.420	0.534004	0.534133	0.527000	0.515689	0.526231	0.548192
1.440	0.529911	0.530070	0.522976	0.512077	0.522987	0.545153
1.460	0.525866	0.526056	0.519001	0.508519	0.519788	0.542152
1.480	0.521866	0.522088	0.515076	0.505013	0.516631	0.539188
1.500	0.517913	0.518166	0.511199	0.501558	0.513516	0.536260
1.520	0.514005	0.514290	0.507369	0.498153	0.510442	0.533369
1.540	0.510142	0.510459	0.503586	0.494797	0.507409	0.530513
1.560	0.506323	0.506672	0.499850	0.491488	0.504414	0.527691
1.580	0.502547	0.502929	0.496158	0.488226	0.501458	0.524902
1.600	0.498815	0.499229	0.492512	0.485009	0.498540	0.522147
1.620	0.495124	0.495571	0.488910	0.481837	0.495659	0.519424
1.640	0.491476	0.491956	0.485352	0.478708	0.492813	0.516732
1.660	0.487868	0.488381	0.481837	0.475621	0.490003	0.514072
1.680	0.484301	0.484848	0.478364	0.472576	0.487228	0.511442
1.700	0.480774	0.481354	0.474933	0.469572	0.484486	0.508842
1.720	0.477287	0.477901	0.471543	0.466608	0.481778	0.506271
1.740	0.473839	0.474486	0.468193	0.463682	0.479102	0.503729
1.760	0.470429	0.471110	0.464884	0.460795	0.476458	0.501216
1.780	0.467057	0.467772	0.461614	0.457945	0.473845	0.498730

Table I. Continued

X	0.0	1.0	2.0	3.0	4.0	5.0
$\frac{N}{f}$	0.367879	0.353224	0.316132	0.270666	0.227772	0.192027
1.800	0.463723	0.464471	0.458383	0.455132	0.471263	0.496271
1.820	0.460425	0.461208	0.455191	0.452354	0.468711	0.493838
1.840	0.457164	0.457981	0.452036	0.449612	0.466189	0.491432
1.860	0.453939	0.454790	0.448918	0.446904	0.463695	0.489052
1.880	0.450750	0.451634	0.445836	0.444229	0.461230	0.486697
1.900	0.447596	0.448514	0.442791	0.441588	0.458793	0.484366
1.920	0.444476	0.445428	0.439783	0.438979	0.456383	0.482061
1.940	0.441390	0.442376	0.436805	0.436402	0.454001	0.479779
1.960	0.438338	0.439358	0.433867	0.433856	0.451644	0.477520
1.980	0.435319	0.436373	0.430941	0.431340	0.449314	0.475285
2.000	0.432332	0.433421-0.		0.428855	0.447008	0.473072
2.020	0.429379	0.430501	0.425255	0.426398	0.444728	0.470882
2.040	0.426457	0.427613	0.422468	0.423971	0.442473	0.468714
2.060	0.423566	0.424756	0.419700	0.421572	0.440241	0.466567
2.080	0.420707	0.421931	0.416979	0.419201	0.438034	0.464442
2.100	0.417878	0.419136	0.414287	0.416857	0.435849	0.462338
2.120	0.415079	0.416371	0.411631	0.414540	0.433688	0.460254
2.140	0.412311	0.413636	0.409010	0.412249	0.431549	0.458190
2.160	0.409572	0.410931	0.406420	0.409984	0.429432	0.456147
2.180	0.406862	0.408254	0.403863	0.407744	0.427337	0.454123

Table I. Continued

X	0.0	1.0	2.0	3.0	4.0	5.0
N f	0.367879	0.353224	0.316132	0.270666	0.227772	0.192027
2.200	0.404180	0.405606	0.401338	0.405530	0.425263	0.452118
2.220	0.401527	0.402987	0.398844	0.403340	0.423211	0.450132
2.240	0.398903	0.400395	0.396381	0.401174	0.421179	0.448165
2.260	0.396305	0.397831	0.393948	0.399031	0.419168	0.446217
2.280	0.393735	0.395294	0.391544	0.396912	0.417176	0.444286
2.300	0.391192	0.392783	0.389169	0.394816	0.415205	0.442374
2.320	0.388675	0.390299	0.386823	0.392742	0.413252	0.440479
2.340	0.386185	0.387842	0.384504	0.390690	0.411319	0.438601
2.360	0.383720	0.385410	0.382213	0.388660	0.409405	0.436740
2.380	0.381281	0.383003	0.379949	0.386651	0.407509	0.434896
2.400	0.378868	0.380621	0.377711	0.384663	0.405632	0.433069
2.420	0.376479	0.378265	0.375499	0.382696	0.403772	0.431258
2.440	0.374114	0.375932	0.373312	0.380750	0.401930	0.429463
2.460	0.371774	0.373624	0.371151	0.378823	0.400106	0.427684
2.480	0.369458	0.371340	0.369014	0.376916	0.398298	0.425921
2.500	0.367166	0.369079	0.366901	0.375028	0.396508	0.424173
2.520	0.364897	0.366841	0.364812	0.373160	0.394734	0.422440
2.540	0.362651	0.364626	0.362746	0.371310	0.392977	0.420723
2.560	0.360428	0.362434	0.360703	0.369478	0.391236	0.419019
2.580	0.358227	0.360264	0.358683	0.367665	0.389510	0.417331

Table I. Continued

X	0.0	1.0	2.0	3.0	4.0	5.0
$\frac{N}{f}$	0.367879	0.353224	0.316132	0.270666	0.227772	0.192027
2.600	0.356049	0.358116	0.356685	0.365870	0.387801	0.415657
2.620	0.353892	0.355990	0.354709	0.364092	0.386107	0.413997
2.640	0.351757	0.353885	0.352754	0.362332	0.384428	0.412351
2.660	0.349644	0.351802	0.350820	0.360589	0.382764	0.410719
2.680	0.347551	0.349739	0.348907	0.358862	0.381115	0.409100
2.700	0.345480	0.347697	0.347015	0.357152	0.379480	0.407495
2.720	0.343428	0.345675	0.345142	0.355459	0.377860	0.405903
2.740	0.341398	0.343674	0.343289	0.353782	0.376254	0.404324
2.760	0.339387	0.341692	0.341456	0.352120	0.374662	0.402758
2.780	0.337396	0.339730	0.339642	0.350474	0.373084	0.401204
2.800	0.335425	0.337788	0.337847	0.348843	0.371520	0.399663
2.820	0.333473	0.335864	0.336070	0.347227	0.369969	0.398135
2.840	0.331540	0.333959	0.334311	0.345627	0.368431	0.396619
2.860	0.329626	0.332073	0.332571	0.344041	0.366906	0.395114
2.880	0.327731	0.330206	0.330848	0.342470	0.365394	0.393622
2.900	0.325854	0.328357	0.329142	0.340912	0.363895	0.392142
2.920	0.323995	0.326525	0.327454	0.339372	0.362409	0.390673
2.940	0.322155	0.324711	0.325782	0.337836	0.360934	0.389215
2.960	0.320332	0.322915	0.324127	0.336313	0.359473	0.387769
2.980	0.318526	0.321136	0.322489	0.334724	0.358023	0.386334

Table I. Continued

X	0.0	1.0	2.0	3.0	4.0	5.0
N	0.367879	0.353224	0.316132	0.270666	0.227772	0.192027
f	0.316738	0.319375	0.320867	0.320867	0.356585	0.384910
3.000	0.314967	0.317630	0.319260	0.331884	0.355158	0.383497
3.020	0.313212	0.315901	0.317669	0.330395	0.353744	0.382095
3.040	0.311475	0.314189	0.316094	0.328952	0.352341	0.380703
3.060	0.309754	0.312494	0.314534	0.327514	0.350949	0.379322
3.080	0.308049	0.310814	0.312989	0.326091	0.349568	0.377952
3.100	0.306360	0.309151	0.311458	0.324681	0.348199	0.376591
3.120	0.304687	0.307503	0.309942	0.323281	0.346840	0.375241
3.140	0.303030	0.305870	0.308441	0.321893	0.345492	0.373900
3.160	0.301388	0.304253	0.306953	0.320517	0.344154	0.372570
3.180	0.299762	0.302650	0.305480	0.319153	0.342827	0.371249
3.200	0.298151	0.301063	0.304020	0.317801	0.341511	0.369939
3.220	0.296554	0.299491	0.302574	0.316459	0.340205	0.368637
3.240	0.294973	0.297932	0.301141	0.315129	0.338909	0.367345
3.260	0.293406	0.296389	0.299722	0.313810	0.337622	0.366063
3.280	0.291854	0.294859	0.298315	0.312502	0.336346	0.364789
3.300	0.290316	0.293344	0.296921	0.311205	0.335079	0.363525
3.320	0.288791	0.291842	0.295540	0.309919	0.333822	0.362270
3.340	0.287281	0.290354	0.294171	0.308643	0.332575	0.361023
3.360	0.285785	0.288880	0.292815	0.307377	0.331337	0.359786
3.380						

Table I. Continued

X	0.0	1.0	2.0	3.0	4.0	5.0
N	0.367879	0.353224	0.316132	0.270666	0.227772	0.192027
f	3.400	0.284302	0.287419	0.291470	0.306122	0.330108
	3.420	0.282833	0.285971	0.290138	0.304876	0.328889
	3.440	0.281377	0.284536	0.288818	0.303641	0.327678
	3.460	0.279934	0.283114	0.287509	0.302416	0.326476
	3.480	0.278504	0.281705	0.286211	0.301200	0.325284
	3.500	0.277087	0.280309	0.284925	0.299994	0.324100
	3.520	0.275682	0.278925	0.283650	0.298798	0.322925
	3.540	0.274290	0.277553	0.282387	0.297611	0.321758
	3.560	0.272911	0.276193	0.281134	0.296433	0.320600
	3.580	0.271543	0.274845	0.279892	0.295264	0.319450
	3.600	0.270188	0.273510	0.278660	0.294105	0.318309
	3.620	0.268845	0.272186	0.277439	0.292954	0.317175
	3.640	0.267513	0.270873	0.276229	0.291813	0.316050
	3.660	0.266193	0.269572	0.275028	0.290680	0.314933
	3.680	0.264885	0.268282	0.273838	0.289556	0.313824
	3.700	0.263588	0.267004	0.272657	0.288440	0.312722
	3.720	0.262303	0.265736	0.271487	0.287333	0.311628
	3.740	0.261028	0.264480	0.270326	0.286234	0.310542
	3.760	0.259765	0.263234	0.269175	0.285144	0.309463
	3.780	0.258513	0.261999	0.268033	0.284061	0.308392

Table I. Continued

X	0.0	1.0	2.0	3.0	4.0	5.0
N f	0.367879	0.353224	0.316132	0.270666	0.227772	0.192027
3.800	0.257271	0.260774	0.266901	0.282987	0.307328	0.335673
3.820	0.256040	0.259560	0.265778	0.281920	0.306272	0.334607
3.840	0.254819	0.258356	0.264664	0.280862	0.305223	0.333548
3.860	0.253609	0.257163	0.263559	0.279811	0.304181	0.332495
3.880	0.252410	0.255979	0.262463	0.278768	0.303147	0.331450
3.900	0.251220	0.254806	0.261376	0.277733	0.302118	0.330411
3.920	0.250041	0.253642	0.260297	0.276705	0.301093	0.329379
3.940	0.248871	0.252488	0.259227	0.275685	0.300090	0.328353
3.960	0.247711	0.251344	0.258166	0.274672	0.299073	0.327334
3.980	0.246562	0.250209	0.257113	0.273667	0.298005	0.326321
4.000	0.245421	0.249084	0.256068	0.272668-0.		0.325314
4.020	0.244290	0.247968	0.255032	0.271677	0.296127	0.324314
4.040	0.243169	0.246861	0.254003	0.270693	0.295124	0.323320
4.060	0.242057	0.245763	0.252983	0.269716	0.294144	0.322332
4.080	0.240954	0.244674	0.251970	0.268745	0.293180	0.321350
4.100	0.239860	0.243594	0.250966	0.267782	0.292212	0.320375
4.120	0.238776	0.242523	0.249969	0.266825	0.291261	0.319405
4.140	0.237700	0.241461	0.248980	0.265875	0.290310	0.318441
4.160	0.236633	0.240407	0.247998	0.264932	0.289368	0.317484
4.180	0.235575	0.239362	0.247024	0.263995	0.288432	0.316532

Table I. Continued

X	0.0	1.0	2.0	3.0	4.0	5.0
N	0.367879	0.353224	0.316132	0.270666	0.227772	0.192027
f	0.234525	0.238325	0.246057	0.263065	0.287501	0.315585
4.200	0.233484	0.237296	0.245097	0.262141	0.286577	0.314645
4.220	0.232451	0.236276	0.244145	0.261223	0.285658	0.313710
4.240	0.231427	0.235264	0.243200	0.260312	0.284745	0.312781
4.260	0.230411	0.234260	0.242261	0.259407	0.283838	0.311857
4.280	0.229403	0.233264	0.241330	0.258508	0.282937	0.310939
4.300	0.228403	0.232276	0.240406	0.257615	0.282041	0.310026
4.320	0.227411	0.231295	0.239489	0.256728	0.281151	0.309119
4.340	0.226427	0.230322	0.238578	0.255847	0.280267	0.308216
4.360	0.225451	0.229357	0.237674	0.254972	0.279388	0.307320
4.380	0.224483	0.228400	0.236776	0.254103	0.278515	0.306428
4.400	0.223522	0.227450	0.235886	0.253240	0.277647	0.305542
4.420	0.222569	0.226507	0.235001	0.252382	0.276784	0.304661
4.440	0.221623	0.225571	0.234123	0.251530	0.275927	0.303785
4.460	0.220685	0.224643	0.233251	0.250684	0.275075	0.302914
4.480	0.219754	0.223722	0.232386	0.249843	0.274228	0.302048
4.500	0.218830	0.222808	0.231527	0.249008	0.273386	0.301187
4.520	0.217913	0.221901	0.230674	0.248178	0.272549	0.300331
4.540	0.217004	0.221001	0.229827	0.247354	0.271718	0.299480
4.560	0.216102	0.220108	0.228986	0.246534	0.270891	0.298633

Table I. Continued

X	0.0	1.0	2.0	3.0	4.0	5.0
N	0.367879	0.353224	0.316132	0.270666	0.227772	0.192027
f						
4.580	0.216102	0.220108	0.228986	0.246534	0.270891	0.298633
4.600	0.215206	0.219222	0.228150	0.245721	0.270070	0.297792
4.620	0.214318	0.218342	0.227321	0.244912	0.269253	0.296955
4.640	0.213436	0.217469	0.226498	0.244109	0.268441	0.296123
4.660	0.212561	0.216602	0.225680	0.243310	0.267635	0.295296
4.680	0.211693	0.215742	0.224868	0.242517	0.266832	0.294473
4.700	0.210831	0.214889	0.224062	0.241729	0.266035	0.293654
4.720	0.209976	0.214042	0.223261	0.240946	0.265243	0.292840
4.740	0.209127	0.213201	0.222466	0.240168	0.264454	0.292031
4.760	0.208285	0.212366	0.221676	0.239394	0.263671	0.291227
4.780	0.207449	0.211537	0.220891	0.238626	0.262892	0.290427
4.800	0.206619	0.210715	0.220112	0.237862	0.262118	0.289630
4.820	0.205795	0.209899	0.219338	0.237103	0.261348	0.288839
4.840	0.204978	0.209088	0.218570	0.236349	0.260583	0.288051
4.860	0.204167	0.208284	0.217807	0.235599	0.259822	0.287266
4.880	0.203361	0.207485	0.217048	0.234855	0.259066	0.286490
4.900	0.202562	0.206692	0.216295	0.234114	0.258313	0.285711
4.920	0.201769	0.205905	0.215547	0.233378	0.257566	0.284939
4.940	0.200981	0.205124	0.214804	0.232647	0.256822	0.284159
4.960	0.200199	0.204348	0.214066	0.231920	0.256083	0.283386
4.980	0.199423	0.203578	0.213333	0.231198	0.255347	0.282571
5.000	0.198653	0.202813	0.212604	0.230480	0.254616	0.

REFERENCES

1. T. E. Bortner and G. S. Hurst, "An Apparatus for Measuring Electron Attachment: Results for Oxygen in Argon," Health Phys. 1, 39-45 (1958).
2. G. S. Hurst and T. E. Bortner, "Capture of Electrons in Molecular Oxygen," ORNL-2670 (September 15, 1959).
3. G. S. Hurst and T. E. Bortner, "Attachment of Low Energy Electrons in Mixtures Containing Oxygen," Phys. Rev. 114, 116-120 (1959).

ORNL-3090
 UC-34 — Physics
 TID-4500 (17th ed., Rev.)

INTERNAL DISTRIBUTION

1. Biology Library	96. K. Z. Morgan
2-3. Central Research Library	97. J. P. Murray (K-25)
4. Laboratory Shift Supervisor	98. D. R. Nelson
5. Reactor Division Library	99. G. D. O'Kelley
6-7. ORNL — Y-12 Technical Library	100. R. H. Ritchie
Document Reference Section	101. P. S. Rudolph
8-57. Laboratory Records Department	102. F. W. Sanders
58. Laboratory Records, ORNL R.C.	103. M. J. Skinner
59. J. A. Auxier	104. J. A. Swartout
60. R. D. Birkhoff	105. A. M. Weinberg
61. T. A. Butler	106. G. M. Fair (consultant)
62. F. J. Davis	107. J. C. Frye (consultant)
63-67. H. B. Eldridge	108. W. H. Langham (consultant)
68. R. F. Hibbs (Y-12)	109. E. P. Odum (consultant)
69-93. G. S. Hurst	110. R. L. Platzman (consultant)
94. W. H. Jordan	111. L. S. Taylor (consultant)
95. C. E. Larson	

EXTERNAL DISTRIBUTION

112. Division of Research and Development, AEC, ORO
 113-746. Given distribution as shown in TID-4500 (17th ed., Rev.) under
 Physics category (75 copies — OTS)