TABLE OF CONTENTS

Abstract	v
Introduction	1
Theory	2
Clinostat Design	3
Specimen Holders	4
Operating Characteristics	5
Cesium Irradiation	6
Environment	7
Test Controls	11
Housing Structure	12
Acknowledgment	12a
Figures 1 - 13	13-25

NOTICE

This report was prepared as an account of work sponsored by the United States Government. Neither the United States nor the United States Atomic Energy Commission, nor any of their employees, nor any of their contractors, subcontractors, or their employees, makes any warranty, express or implied, or assumes any legal liability or responsibility for the accuracy, completeness or usefulness of any information, apparatus, product or process disclosed, or represents that its use would not infringe privately owned rights.

ABSTRACT

A clinostat-irradiator complex has been designed and built at the Lawrence Berkeley Laboratory. The purpose of these instruments is to study radiation effects on small organisms while they exist under various degrees of gravity compensation. Described herein is the integrated, experimental system.

The Modified and Completed 1 π Clinostat-Irradiation System

Ira Lee Silver

Donner Laboratory and Lawrence Berkeley Laboratory, University of California, Berkeley

Introduction

Two "1 π " clinostats have been designed and built in our Laboratory under a joint program of the NASA Bioscience Division (Purchase Order #W12-792 Task 6 to the AEC) and the Atomic Energy Commission.¹ One clinostat is employed in conjunction with a gamma irradiation source, and the other serves as a nonirradiation control. The purpose of these instruments is to study radiation effects on small organisms while they exist under various degrees of gravity compensation. Possible synergistic or antagonistic relations of gravity compensation to radiation can help interpret results of biological studies in space flight and serve as controls for future satellite experiments. A previous report by the author (UCRL 19369, September, 1969), delineated the initial design of the clinostat-irradiator. Described herein is the revised, final and integrated system, complete with its own housing structure.

1. Special acknowledgment is made to Dr. Solon A. Gordon of the Division of Biological and Medical Research, Argonne National Laboratory, Argonne, Illinois, who has served as a consultant on this clinostat development. The clinostats built in Argonne under Dr. Gordon's auspices served as models for the design of these present clinostats.

Theory

A 1π clinostat provides circular rotation with uniform speed about an axis normal to the force of gravity. This concept cannot provide a true weightless or free fall environment, because it precludes neither mechanical stresses nor subcellular tumbling and rotation, in biological regions off the axis of rotation. However, if the ensuing superimposed centrifugal forces are negligible, the clinostat provides the experimenter with the best feasible means of causing the vector averaged gravitational force to be zero for prolonged times. A more detailed analysis of the dynamics of clinostat motion is being prepared by the author for publication and is available upon request.

More sophisticated " 2π " devices are possible whereby variable simultaneous rotation about a vertical axis will induce any desired fractional "gravitational" force. This extension has questionable utility since the same result can be more easily obtained by tilting a 1π clinostat through the equivalent angle.

A three dimensional (4π) clinostat has been pioneered by Solon Gordon, cf: Space Biology, proceedings of the 24th Annual Biology Colloquium, Oregon University Press, 1963, pp 75-105. Here a series of gimbal mounts are computer driven, enabling a point specimen to trace over the surface of a sphere so that the time averaged position, angular velocity, and angular acceleration are zero. In this manner it is hoped

to negate the possible effects of all external tropic stimuli on biological development.

Clinostat Design

The clinostats are identical, measuring 17 inches high, 17 inches broad, and 7 inches deep (excluding attached specimen holders) and each weighs approximately 100 pounds. They consist of nine parallel, horizontal, ventilated, rotating shafts, capable of operating at speeds of 2 to 91 rpm. Eight of these shafts are evenly spaced on the circumference of a 9 inch diameter circle, at the center of which is located the ninth shaft. Simultaneous controls are provided by: a vertical shaft rotating at the same speed as the horizontal ones, and by four horizontal stationary studs.

Each clinostat is run by two, three-speed synchronous electric motors, enclosed in water jackets. The motors are placed in parallel and situated on diagonally opposite sides of the clinostat to ensure proper balance. Motion is transmitted to the shafts from the motors via capstans engaging a flywheel, which then moves an endless Mylar belt snaking over one of two sets of shaft pulleys. One set of pulleys sets the shaft speeds at 8, 16, 32, 64 or 91 rpm; the other set reduces said speeds by a factor of four. In fact, the shaft speeds can be further reduced by changing the relative diameters of the flywheel and capstans.

Figures 1 - 5 show different views of the clinostat, delineating the salient details. It is noted that several

major design changes have been made to the original clinostat, described in UCRL 19369 and shown here in Fig. 6. The readily observable revisions have been the removal of the: vibration isolation mounts, internal leveling screws, and sliding track; as well as the addition of the aforementioned stationary studs. The reason for these removals will be apparent subsequently. Bolt holes in the spacers (Figs. 1, 2, 5) will constrain the clinostats within special environmental chambers; the latter described in the section: Environment.

Specimen Holders

The ends of the horizontal shafts are accurately faced, and threaded 3/8-in. diameter by 16 threads per inch; the thread extending 3/8 inch from the end of the 3/4-inch shaft. Thus, any type of specimen holder adapted to fit onto this attachment can be employed. Because the clinostat must fit within environmental chambers, the length of the specimen holders must be less than 9 inches. The vertical shaft consists of a polycarbonate tube (1.250 in. o.d. x 1.030 in. i.d.) with an effective length of approximately 9 inches.

Interchangeable specimen holders have been made for use on the shafts. These holders are shown in Figs. 7 and 8 and shown attached to the clinostat in Figs. 1 - 4 and 6; the perforations allow for free air flow. Thus, if size permits, biological specimens can be constrained within pieces designed to fit in these holders. Some examples are shown in Figs. 7 - 9. The screw arrangement shown in Fig. 7 provides a means

of disassembling the unit.

Operating Characteristics

Reference to gravity is made by means of bubble levels (Figs. 2-4, and 6). An azimuthal level, along the axis of the horizontal shafts, is readable to 1 part in 10,000. A horizontal polar level, at right angles to this one, is readable to 1 part in 2,000. When the clinostats are tilted (see section: Environment), a circular bubble level will be used to adjust the shafts to the vertical via a mirror arrangement. Manual precision leveling screws will permit reference adjustment (see section: Environment). The shafts are parallel to one another within 1 part in 10,000, and the vertical shaft is perpendicular to the horizontal to 1 part in 1,000. The sag of a typical specimen holder on the shaft was found to be negligible.

Quartz-crystal oscillators (30, 60, and 86 Hz) are employed in the ac motor power supply. These, in conjunction with multiple-pole motor windings, provide a total of five motor speeds. Tests show that vibrations are lowest when only one drive motor is used for each clinostat. The quietest motor has therefore been selected, and for each speed the electrical LC circuit values are optimized to give the least vibration. The resulting induced vibrations, seen at the shaft ends for the higher speed range of the belt-pulley setup, are as follows for both clinostats:

Frequency (Hz)	# of poles	rpm	average fraction of IG
30	24	8	.00007
30	12	16	.00016
30	6	32	.00028
60	6	64	.00034
86	6	91	.00052

The noise level in the clinostat locale was measured with an H. H. Scott type 410C sound-level meter in combination with an H. H. Scott type 420A sound analyzer. The scale employed was a flat "C" weight, which covers the frequency spectrum from 20 Hz to 20 kHz; the zero db reference level was 2×10^{-4} dynes/cm². The noise generated by each clinostat averages about 47 decibels; varying by about 4 db in the speed range 8-91 rpm.

Cesium Irradiation

Irradiation capabilities are provided by a nominal strength 1000 curie cesium 137 source (1042 Ci on 8/28/65). As described in UCRL 19369, the source was originally confined with a seven ton lead shielded chamber. However, the source has since been installed onto a double track crane with a five ton lift capability (Figs. 10-11). The consequent irradiator system is contained within a concrete shielded room (see section: Housing Structure). The crane has horizontal and vertical movement, both at an electrically controlled speed of 15 inches/min. In addition, the source can be electrically moved through 90° from a horizontal irradiating position,

Fig. 10, to a vertical position, Fig. 11.

These motions can be regulated at the source, as well as via a control console located outside the source room. The console, of course, also controls the initiation and termination of irradiation. In addition, a manual scram button is installed within the irradiator room. A microphone-loud speaker system allows for communication between the source room and the console.

Dose calibrations show that a maximum irradiation rate for the source-clinostat geometry, is about 50 Roentgen/min. Proper positioning and shielding can reduce this flux to any desired rate. The crane-irradiator system has been designed so as to allow for the future addition of a stronger cesium source: eg. 10,000 Ci, to be employed in conjunction with the present source. Since the precise dosage seen by a biological specimen will be dependent on the particular experimental geometry, it is possible to monitor all irradiation experiments with small LiF disks which can sandwich the specimens.

Environment

Both clinostats are housed in identically designed "environmental chambers", which in turn are contained within individual rooms (see section: Housing Structure). Various views of the environmental chambers are given in Figs. 1, 10-12.

As noted in Fig. 1, the chambers are free to pivot through

90° (with a stable position at 45°) so as to provide for sequential "vertical control" irradiation. This can be necessary since the geometry of the clinostats (Figs. 1, 6) is such that the vertical shaft and the horizontal shafts will not simultaneously "see" the same dose energy deposition. Manual precision leveling screws, located on the base on which the chambers rest (Figs. 1, 10 and 11), permit reference adjustments of the shafts in relation to the force of gravity. In addition, the leveling screws can be used to tilt the shafts off of the horizontal. Thus, any fraction of the gravitational vector can be applied, because the effective component is proportional to the sine of the angle of inclination.

Conditioned air, supplied by an Aminco-Aire conditioner, will pass in parallel through both environmental chambers. The flow is through a closed system with air being bled from the main ducting via butterfly valves into each of the chambers. Fig. 10 shows the bleed-off arrangement for the irradiator clinostat: air passes through flexible tubing into a baffle system on the side of the chamber. The air is thus diffused in the chamber, prior to exhausting to the ambient. Air filters, as shown in Fig. 10, are placed in each line, immediately after the bleed-off point. In order to maintain a sterile environment, an ultraviolet germicidal lamp is installed at the air conditioner outlet.

Resistance heaters (not shown in Fig. 10), installed in each line, after the air filter, supplement the heat supplied

by the air conditioner. In addition, these heaters are used in conjunction with two Wheelco temperature regulators to maintain a constant temperature in each clinostat. These regulators work on a thermocouple sensor input; variations of less than 0.5°C can be obtained between the sensor locations at each clinostat. Thus, the procedure followed is to set the air conditioner at a temperature lower than required, and then set the regulators to the desired temperature. As a safety precaution, the resistance heaters cannot be turned on unless the air conditioner is on.

Temperatures up to 40°C are available with this system. A ring of eight copper-constantin thermocouples monitor the clinostat temperature (Figs. 1 and 12). Convective air currents can, if not mechanically regulated, cause spatial thermal variations of up to 2°C . Additional thermocouples monitor the inlet and outlet air temperature of the environmental chambers. Thermocouple readouts for each chamber can be either continuously or intermittently plotted on respective sixteen channel recorders.

A concurrent relative humidity range of 20-80% is available from the air conditioner. Drier environments can be obtained by utilizing desiccators within the chambers. The humidity is monitored by solid state probes, located within the baffle, i.e., the air inlet section of each chamber. A switch controls the continuous readout of each probe onto a

meter located adjacent to the channel recorders (see section: Housing Structure). Humidity calibrations are made with wet and dry bulb thermometry.

Fluorescent, e.g., ultraviolet, and incandescent, e.g., white, red, lighting is available within the chambers (Figs. 1 and 12). The lights are not currently water cooled, but can be so adapted if lighting requirements result in excessive heat generation. Tests with a light meter, Photovolt 200 M, indicate that leakage of visible light into the darkened chambers is less than the meter's sensitivity, i.e., less than 0.05 foot candle.

Provision has been made to allow any desired "atmosphere" to pass over the test specimens. The shafts on the clinostat are hollow. Tubes are attached to them through the clinostat face plate (Figs. 1, 2 and 11) and then passed out of the respective chambers to any desired gas supply. In fact, one tube services two horizontal shafts, so that four separate atmospheres can be employed simultaneously.

The chambers have been lined with heavy lead foam padding so as to attenuate outside noise as well as dampen internal noise (Figs. 1 and 12). In a test using a sound generator at a distance of five feet to produce pressure waves up to 6000 Hz, the attenuation was found to vary between 10 and 35 decibels. The same measuring techniques and scale were used as described in the section: Operating Characteristics. The

sound level within both environmental chambers, with the clinostats not running, was 48 and 50 decibels for the air conditioning unit being off and on respectively. With both the air conditioner and the clinostats running the noise level in each chamber averaged about 52 decibels, varying by about 1.5 db over the speed range 8-91 rpm.

Local and intermittent external "spiked" vibrations are found to be in the magnitude of 10^{-3} G for frequencies down to 2 Hz, and less than 10^{-4} G between 0.5 and 1.5 Hz. In general, therefore, the clinostat-induced vibrations are of the same magnitude as the externally generated ones. For this reason, the original vibration isolation mounts on the clinostat (Fig. 6) were removed so as to "short circuit to ground" the internally produced energy. Both environmental chambers rest on individual inertia platforms as shown in Figs. 10 and 11. For the present, these platforms are rigidly supported from underneath. If it is deemed necessary, however, isolators can be installed to further reduce external vibrations; air mounts for example can "float" the platforms.

Test Controls

As noted at the beginning of the above section, there are provided two means of running vertical irradiation controls: the vertical shaft on the clinostat allows for simultaneously running test and control (although the control dosage will be slightly different due to the system configuration), and the

rotation of the enclosure chamber provides a means for a sequential control run. Similarly, there is available the horizontal stationary studs for simultaneous irradiation controls, as well as the second clinostat for nonirradiated controls.

Housing Structure

The clinostat-irradiation complex is a self contained unit housed in Building 74-B at the Lawrence Berkeley Laboratory. A schematic of the floor plan is shown in Fig. 13. Both the control and irradiator rooms are temperature controlled and soundproofed. In addition, the latter room is concrete shielded, as indicated, and has a steel door (6.5 ft high x 4.5 ft wide x 1 ft thick) opening to the outside. All electrical wiring, water cooling, and air ducting lead from the laboratory area to both rooms via ceiling and floor trenches. An interlock system isolates the irradiator room from the rest of the complex, during irradiation.

All control operations are handled in the laboratory area. The air conditioner, temperature and humidity recorders, cooling water pump, and irradiator control console are all along the walls common to both the laboratory and the clinostat rooms. In addition, the laboratory and the adjacent Tribolium room contain the instrumentation and equipment necessary to perform exacting experiments.

ACKNOWLEDGMENT

This project could not have been successfully completed without the creativity and skill of the following: Elbert Benjamine, Pete Bringham, George Goranson, Jack Gunn, Al Kleid, Al Kleven, Jack Nelson, Jerry Nicholas, Duane Norgren, Eugene Riley, Frank Upham, George Wigle, and Al Windsor. The author most gratefully thank these gentlemen for their efforts and contributions.

CBB 716-2725

FIGURE 1: CLINOSTAT, FRONT VIEW; CONTAINED WITHIN ENVIRONMENTAL CHAMBER.

1. stationary shaft	2. rotating shaft
3. gas and water tubes	4. lights
5. ring of thermocouples	6. handle for tilting chamber
7. leveling screw	8. air outlet from chamber
9. accelerometer post and electrical outlet	

CBB 716-2480

FIGURE 2: CLINOSTAT, VIEW OF TOP AND LEFT SIDE.

1. azimuthal striders level	2. vertical bubble level and mirror
3. motor casing	4. gear for rotating vertical shaft
5. gas tubes leading to horizontal rotating shafts	
6. bolt hole for constraining clinostat	

CBB 716-2486

FIGURE 3: CLINOSTAT, VIEW OF TOP.

1. polar striders level	2. flywheel
3. stationary shaft	4. rotating shaft
5. tension adjustment for vertical rotating shaft	

CBB 716-2482

FIGURE 4: CLINOSTAT, VIEW OF TOP AND RIGHT SIDE.

1. striders levels	2. gear for rotating vertical shaft
3. motor casings, (2)	4. bolt hole for constraining clinostat
5. shaft pulley for the high speed range, with mylar belt	
6. shaft pulley for the low speed range	
7. shaft pulleys for transmitting power to the horizontal circumferential shafts	

CBB 714-2484

FIGURE 5: CLINOSTAT, REAR VIEW.

1. flywheel	2. capstan
3. motor casing	4. spacer
5. shaft for transmitting power to the center horizontal shaft	

CBB 696-3939

FIGURE 6: CLINOSTAT, PRIOR TO REVISIONS.

1. vibration isolation mounts	2. sliding track
3. mylar belt on shaft pulley	4. pulley for rotating vertical shaft
5. horizontal rotating specimen holders	
6. vertical rotating specimen holder	

XBB 699-6005

FIGURE 7: CLINDOSTAT SPECIMEN HOLDER; SPECIFIC ORGANISM ADAPTERS.

1. screw to disengage sealing plug

XBB 7011-5356

FIGURE 3: CLINOSTAT SPECIMEN HOLDER SYSTEM; TRIBOLIUM EGGS.

1. specimen holder	2. sealing plug
3. spacer	4. capillary tube, 1 mm diameter
5. capillary tube holder	

XBB 703-1490

FIGURE 9: TRIBOLIUM EGG AND PUPAE ADAPTERS TO FIT IN SPECIMEN HOLDERS.

CBB 716-2727

FIGURE 10: CLINOSTAT-IRRADIATOR SYSTEM; HORIZONTAL IRRADIATION.

1. crane	2. controls for crane and irradiator motion
3. irradiator	4. port opening for irradiation source
5. flexible tubing for air	6. environmental chamber base
7. air filter	8. location for resistance heater
9. main ducts for air	10. inertia pad

CBB 716-2729

FIGURE 11: CLINOSTAT-IRRADIATOR SYSTEM; VERTICAL IRRADIATION.

1. environmental chamber	2. opening to view striders levels
3. location for irradiation beam	4. baffle section of chamber
5. thermocouple plugs	6. water and gas inlets
7. power leads	8. manual scram button

CBB 716-2488

FIGURE 12: ENVIRONMENTAL CHAMBER; INTERIOR VIEW.

1. fluorescent lighting	2. track for clinostat spacer
3. bolt to hold clinostat	4. air inlet
5. lead - foam lining	6. power leads and gas tubes
7. snap to keep chamber door tightly shut	

XBL 7111-1602

FIGURE 13: FLOOR PLAN OF CLINOSTAT-IRRADIATOR HOUSING COMPLEX.

- A, B, C, D, G : incubator or refrigeration units
 E : sink
 F : fume hood
 H : irradiator control console
 I : ovens
 J : power console for both clinostats
 K : temperature and relative humidity recorders
 L : Aminco-Aire conditioner
 M : analytic balance