13.19.72
154

AK. 2377

LA-5077-MS

INFORMAL REPORT

MASTER

3D Magnetic Field Computations and Comparison with Measurements

This report was prepared as an account of work sponsored by the United States Government. Neither the United States nor the United States Atomic Energy Commission, nor any of their employees, nor any of their contractors, subcontractors, or their employees, makes any warranty, express or implied, or assumes any legal liability or responsibility for the accuracy, completeness or usefulness of any information, apparatus, product or process disclosed, or represents that its use would not infringe privately owned rights.

In the interest of prompt distribution, this LAMS report was not edited by the Technical Information staff.

Printed in the United States of America. Available from
National Technical Information Service
U. S. Department of Commerce
5285 Port Royal Road
Springfield, Virginia 22151
Price: Printed Copy \$3.00; Microfiche \$0.95

LA-5077-MS

Informal Report
UC-28

ISSUED: November 1972

los alamos
scientific laboratory
of the University of California
LOS ALAMOS, NEW MEXICO 87544

NOTICE

This report was prepared as an account of work sponsored by the United States Government. Neither the United States nor the United States Atomic Energy Commission, nor any of their employees, nor any of their contractors, subcontractors, or their employees, makes any warranty, express or implied, or assumes any legal liability or responsibility for the accuracy, completeness or usefulness of any information, apparatus, product or process disclosed, or represents that its use would not infringe privately owned rights.

3D Magnetic Field Computations and Comparison with Measurements

by

R. Christian*
H. F. Vogel

*Purdue University

DISTRIBUTION OF THIS DOCUMENT IS UNLIMITED

3D MAGNETIC FIELD COMPUTATIONS AND COMPARISON WITH MEASUREMENTS

by

R. Christian and H. F. Vogel

ABSTRACT

3D calculations have been made for an H frame magnet mounted in a steel box for a d-c separator. In the computations, steel is assumed to be of infinite permeability. Measurements have been made on this magnet, which are compared with the computations.

A picture frame magnet with coils on the return legs has been mounted symmetrically within a rectangular steel box with rectangular cut-out in each of its two end faces. The picture frame has inside dimensions 13-1/2-in. x 13-1/2-in. with a length of 6-1/4-in.; the frame thickness being 3/4-in. The clearance between the frame and the steel box is 3/4-in. except at the ends, in the z-direction, where it is greater, because the length of the steel box is $2L = 16-7/8$ in. The dimensions of the rectangular cut-out are 8-in. in the y-direction, which is the direction of the main flux, and 1-5/8-in. in the x-direction. The two legs of the magnet are wound with a thin coil

Hall probe measurements on this not very carefully assembled model magnet were integrated and are tabulated below together with the corresponding computations.

TABLE I

y(in.)	$\int_{-L}^L B_y(x=0;y) dz$	$\int_{-L}^L B_y(0;0) dz$
	Measurement	Computation
0	0	0
+0.5	0.05%	0.125%
-0.5	0.16%	
+1.25	0.59%	1.4%
-1.25	0.86%	
-2.25	2.66%	5.3%

The computations were made with a simple program using scalar potentials, which was written

within one day and which executed correctly the next day on the CDC 6600 computer. The edit part of the program took then a week to complete. An octant of the real model is used for reasons of symmetry. The boundary potentials are zero except at the iron core where $V = y$ ($V = \text{const} = y_{\text{coil}}$ at the yoke).

Rapid iteration by the extrapolated Liebmann method with Frankel-Young convergence of potentials is used. The edit subprogram fits a Taylor series truncated after the third order terms; central difference differentiation is being used in order to obtain the integrals $H_x dz$ and $H_y dz$. The computed results of Table I have been obtained with a problem space of 91500 points total. The potentials here are written into large core memory of the CDC 7600. Two-hundred and sixty iterations bring the residual down to less than 2×10^{-4} , corresponding to displacements of the $\int_{-L}^L B dz$ values of less than 3×10^{-3} . The cost of running this problem on the CDC 7600 was \$154.

The program has meanwhile been completed with current cuts and shims for the magnet presently under construction.

The two important uses of this program were, to develop the three-dimensional aspects for a new subroutine of the ray tracing program. A set of selected points will then be measured and used in the ray tracing computations; and to learn about accuracy requirements and shimming techniques for the magnet.