Operated by
Westinghouse
Hanford Company
for the USAEC

A subsidiary of
Westinghouse Electric
Corporation

RDT Standards Transmittal

Hanford Engineering Development Laboratory

DISTRIBUTION	DATE March 1, 1973		NO. 100
	RETURN COMMENTS TO	TYPE SUBMISSION	TYPE ACTION
	J. D. Schaffer Westinghouse Hanford Company P. O. Box 1970 Richland, WA 99325	☐ 1. PRELIMINARY ☐ 2. DRAFT ☒ 3. TENTATIVE ☐ 4. FINAL ☒ 5. Revision	☐ A. APPROVAL ☐ B. INFORMATION ☐ C. COMMENTS ☒ D. TRIAL USE ☐ E.
	NOT LATER THAN		
Distribution List 1	NUMBER	DESCRIPTION	ACTION
		RDT F 3-2T, Calibration Program Requirements, February 1973.	D
NOTICE This report was prepared as an account of work sponsored by the United States Government. Neither the United States nor the United States Atomic Energy Commission, nor any of their employees, nor any of their contractors, subcontractors, or their employees, makes any warranty, express or implied, or assumes any legal liability or responsibility for the accuracy, completeness or usefulness of any information, apparatus, product or process disclosed, or represents that its use would not infringe privately owned rights. MASTER Release for Announcement in Nuclear Science Abstracts <i>leg</i>			

SIGNATURE

Al Thiede

DISCLAIMER

This report was prepared as an account of work sponsored by an agency of the United States Government. Neither the United States Government nor any agency Thereof, nor any of their employees, makes any warranty, express or implied, or assumes any legal liability or responsibility for the accuracy, completeness, or usefulness of any information, apparatus, product, or process disclosed, or represents that its use would not infringe privately owned rights. Reference herein to any specific commercial product, process, or service by trade name, trademark, manufacturer, or otherwise does not necessarily constitute or imply its endorsement, recommendation, or favoring by the United States Government or any agency thereof. The views and opinions of authors expressed herein do not necessarily state or reflect those of the United States Government or any agency thereof.

DISCLAIMER

Portions of this document may be illegible in electronic image products. Images are produced from the best available original document.

Supersedes
RDT F 3-2T, February 7, 1969

RDT Standard

CALIBRATION PROGRAM REQUIREMENTS

FEBRUARY 1973

Any further distribution by any holder of this document or of the data therein to third parties representing foreign interests, foreign governments, foreign companies and foreign subsidiaries or foreign divisions of U.S. companies should be coordinated with the Director, Division of Reactor Development and Technology, U.S. Atomic Energy Commission.

Division of Reactor Development and Technology
United States Atomic Energy Commission

R D T

S T A N D A R D

RDT F 3-2T

UNITED STATES ATOMIC ENERGY COMMISSION
DIVISION OF REACTOR DEVELOPMENT AND TECHNOLOGY

Date February 1973

Page i of i

CALIBRATION PROGRAM REQUIREMENTS

TABLE OF CONTENTS

	<u>Page</u>
1. SCOPE	1
2. DEFINITIONS	1
3. REQUIREMENTS	2
3.1 General	2
3.2 Calibration Control	2
3.3 Use of M&TE	3
3.4 Nonconformances	4
4. QUALITY ASSURANCE AUDITS	4

UNITED STATES ATOMIC ENERGY COMMISSION
DIVISION OF REACTOR DEVELOPMENT AND TECHNOLOGY

Date February 1973Page 1 of 4

CALIBRATION PROGRAM REQUIREMENTS

---

1. SCOPE

1.1 This standard sets forth requirements for establishment and maintenance of a calibration program to measure and verify the accuracy of all measuring and test equipment (M&TE) necessary to accomplish contract or program requirements.

1.2 When invoked by written program directives, RDT standards, contracts, or purchase orders, this standard shall apply to laboratory or contractor calibration programs and shall be extended to subcontractors and suppliers who are required to perform measurements and tests.

1.3 This standard and any document executed in its implementation shall not supersede other requirements, but shall be in addition to those requirements.

2. DEFINITIONS

2.1 Accuracy. The degree of conformance of a measure to a recognized standard or true value. Accuracy is determined both by the degree of systematic error of measurement (bias), and by precision.

2.2 Calibration. Comparison of a measurement standard or item of M&TE with a standard of higher accuracy or item of M&TE of closer tolerance to detect and quantify inaccuracies, and to report or eliminate those inaccuracies by adjustment.

2.3 Measuring & Test Equipment (M&TE). Devices or systems used to calibrate, measure, gage, test, inspect, or control in order to acquire research, development, test or operational data; for process control; or to determine compliance with design, specifications, or other technical requirements. Systems used for measurement or testing include components from the sensing element through the output or recording device.

2.4 Precision. The degree of agreement between individual measurements; i.e., repeatability.

2.5 Reference Standards. Standards of the highest accuracy in a laboratory's or contractor's internal calibration program. These standards establish the basic accuracy limits for that program.

3. REQUIREMENTS

3.1 General. A program for the calibration of M&TE shall be established and maintained. The program shall be designed to measure and assure the accuracy of controlled M&TE. It shall provide for the timely detection of inaccuracies and for prompt, effective action for their correction.

3.2 Calibration Control. The calibration program shall include the following elements of control:

3.2.1 Program Description. A list of M&TE shall be prepared and maintained to identify those items included in the calibration program covered by the scope of this document. Written procedures shall be provided describing controls over M&TE.

3.2.2 Adequacy of Standards. Standards used for calibrating M&TE shall have an accuracy level, history of stability and acceptable calibration range that are adequate for the programmatic requirements for the M&TE. M&TE shall be calibrated against standards having tolerances of one-fourth to one-tenth those of the M&TE. Tolerances greater than one-fourth may be acceptable when limited by the state-of-the-art.

3.2.3 Environmental Control. M&TE shall be stored and calibrated in environments which will not adversely affect their accuracy. Environmental factors which shall be considered include, but shall not be limited to; temperature, humidity, vibration, radio frequency interference, electromagnetic interference, dust, cleanliness, fumes and voltage stability. When inaccuracy due to environmental effects cannot be avoided, compensating corrections shall be determined and applied to the calibration made.

3.2.4 Intervals of Calibration. Written methods or procedures shall be provided to recall controlled M&TE for recalibration at prescribed intervals to verify the required accuracy. Such intervals may be in calendar time or relate to usage. Interval selection should consider experience, inherent stability, purpose or use, and accuracy required. Historical records shall be maintained which contain sufficient experience data for evaluating calibration intervals.

3.2.5 Calibration Procedures. Written methods or procedures for calibrating M&TE shall be provided to assure calibration technique uniformity. Maximum use shall be made of existing methods, practices or procedures.

Calibration procedures shall include the following information as a minimum:

1. Identity of the item or system to be calibrated.
2. Calibration equipment and standards.
3. Sequence of operations.

4. Checks, tests, measurements and acceptance tolerances.
5. Method of presenting calibration data.
6. Special instructions when necessary.

3.2.6 Traceability. M&TE shall be calibrated utilizing standards whose calibration is traceable to the U. S. National Bureau of Standards, accepted values of natural physical constants, or derived by the ratio type of self-calibration techniques.

3.2.7 Records. The validity of all reference standards shall be supported by certificates, reports, or data sheets which indicate the date of calibration, calibration facility and any deviations from standard environmental conditions. These documents shall also provide evidence of traceability and shall contain data demonstrating conformance with accuracy requirements. Subordinate standards shall be supported by like documents, except that data on accuracy conformance and on environmental conditions may be omitted when not essential to achieving the necessary controls on accuracy.

Records shall be maintained to show that established schedules and procedures for the calibration of standards and M&TE have been followed. The records for M&TE shall contain a history of calibration or other means of control for each item or system of M&TE showing calibration interval, date of last calibration, and conformance or nonconformance to required tolerances before adjustments, including actual values of nonconformances.

3.2.8 Labeling. Each item or system of M&TE shall be labeled to indicate its calibration status. Systems shall show a listing of components. The label shall indicate by whom the M&TE was calibrated and when the next calibration is due. When size or functional characteristics of M&TE prevent the application of a label, an identifying code shall be applied to reflect status. When neither labeling nor coding is practical, the procedures shall provide for monitoring of records to assure control. Labels, codes, or recall records shall be used to identify and control any M&TE whose use must be limited due to imprecision or bias.

3.2.9 Precalibration Checks. The calibration of M&TE submitted for recalibration shall be checked before adjustments or repairs are made.

3.3 Use of M&TE. Controls shall be established and implemented to assure that M&TE consistently produces results of acceptable accuracy. Such measures shall include, as appropriate:

1. Environmental and handling controls.
2. Training and qualification of personnel who use and calibrate M&TE.

-
3. Checking calibration status before use.
  4. Interim checks between calibrations.
  5. Reporting and recalibrating potentially damaged M&TE.

M&TE which has not been properly maintained or calibrated shall be identified as nonconforming and removed from service.

3.4 Nonconformances. If an M&TE item or system is found to be outside of required tolerances, it shall be identified as nonconforming and the using organization shall be notified that all items inspected or tested with the nonconforming device since its last calibration are suspect and should be considered nonconforming until proven otherwise. If the items inspected or tested have been shipped, they shall be identified and reported to the receiving activity, with recommendations for any necessary reinspection.

4. QUALITY ASSURANCE AUDITS

Operations performed in compliance with this standard shall be subject to audit. Audits shall include, but not be limited to, surveillance of calibration operations, reviews of records, review of the calibration program, and checks of actual M&TE calibrations against appropriate standards.

RDT Standards Transmittal

Hanford Engineering Development Laboratory

DISTRIBUTION	DATE		NO.
	March 1, 1973		100
	RETURN COMMENTS TO	TYPE SUBMISSION	TYPE ACTION
	J. D. Schaffer Westinghouse Hanford Company P. O. Box 1970 Richland, WA 99325	☐ 1. PRELIMINARY ☐ 2. DRAFT ☒ 3. TENTATIVE ☐ 4. FINAL ☒ 5. Revision	☐ A. APPROVAL ☐ B. INFORMATION ☐ C. COMMENTS ☒ D. TRIAL USE ☐ E.
	NOT LATER THAN		
	NUMBER	DESCRIPTION	ACTION
Distribution List 1		RDT F 3-2T, Calibration Program Requirements, February 1973.	D

SIGNATURE

DeStade

RDT

F 3-2T

Supersedes
RDT F 3-2T, February 7, 1969

RDT Standard

CALIBRATION PROGRAM REQUIREMENTS

FEBRUARY 1973

Any further distribution by any holder of this document or of the data therein to third parties representing foreign interests, foreign governments, foreign companies and foreign subsidiaries or foreign divisions of U.S. companies should be coordinated with the Director, Division of Reactor Development and Technology, U.S. Atomic Energy Commission.

Division of Reactor Development and Technology
United States Atomic Energy Commission

R D T

S T A N D A R D

RDT F 3-2T

UNITED STATES ATOMIC ENERGY COMMISSION
DIVISION OF REACTOR DEVELOPMENT AND TECHNOLOGY

Date February 1973

Page i of i

CALIBRATION PROGRAM REQUIREMENTS

TABLE OF CONTENTS

	<u>Page</u>
1. SCOPE	1
2. DEFINITIONS	1
3. REQUIREMENTS	2
3.1 General	2
3.2 Calibration Control	2
3.3 Use of M&TE	3
3.4 Nonconformances	4
4. QUALITY ASSURANCE AUDITS	4

UNITED STATES ATOMIC ENERGY COMMISSION
DIVISION OF REACTOR DEVELOPMENT AND TECHNOLOGY

Date February 1973Page 1 of 4

CALIBRATION PROGRAM REQUIREMENTS

---

1. SCOPE

1.1 This standard sets forth requirements for establishment and maintenance of a calibration program to measure and verify the accuracy of all measuring and test equipment (M&TE) necessary to accomplish contract or program requirements.

1.2 When invoked by written program directives, RDT standards, contracts, or purchase orders, this standard shall apply to laboratory or contractor calibration programs and shall be extended to subcontractors and suppliers who are required to perform measurements and tests.

1.3 This standard and any document executed in its implementation shall not supersede other requirements, but shall be in addition to those requirements.

2. DEFINITIONS

2.1 Accuracy. The degree of conformance of a measure to a recognized standard or true value. Accuracy is determined both by the degree of systematic error of measurement (bias), and by precision.

2.2 Calibration. Comparison of a measurement standard or item of M&TE with a standard of higher accuracy or item of M&TE of closer tolerance to detect and quantify inaccuracies, and to report or eliminate those inaccuracies by adjustment.

2.3 Measuring & Test Equipment (M&TE). Devices or systems used to calibrate, measure, gage, test, inspect, or control in order to acquire research, development, test or operational data; for process control; or to determine compliance with design, specifications, or other technical requirements. Systems used for measurement or testing include components from the sensing element through the output or recording device.

2.4 Precision. The degree of agreement between individual measurements; i.e., repeatability.

2.5 Reference Standards. Standards of the highest accuracy in a laboratory's or contractor's internal calibration program. These standards establish the basic accuracy limits for that program.

3. REQUIREMENTS

3.1 General. A program for the calibration of M&TE shall be established and maintained. The program shall be designed to measure and assure the accuracy of controlled M&TE. It shall provide for the timely detection of inaccuracies and for prompt, effective action for their correction.

3.2 Calibration Control. The calibration program shall include the following elements of control:

3.2.1 Program Description. A list of M&TE shall be prepared and maintained to identify those items included in the calibration program covered by the scope of this document. Written procedures shall be provided describing controls over M&TE.

3.2.2 Adequacy of Standards. Standards used for calibrating M&TE shall have an accuracy level, history of stability and acceptable calibration range that are adequate for the programmatic requirements for the M&TE. M&TE shall be calibrated against standards having tolerances of one-fourth to one-tenth those of the M&TE. Tolerances greater than one-fourth may be acceptable when limited by the state-of-the-art.

3.2.3 Environmental Control. M&TE shall be stored and calibrated in environments which will not adversely affect their accuracy. Environmental factors which shall be considered include, but shall not be limited to; temperature, humidity, vibration, radio frequency interference, electromagnetic interference, dust, cleanliness, fumes and voltage stability. When inaccuracy due to environmental effects cannot be avoided, compensating corrections shall be determined and applied to the calibration made.

3.2.4 Intervals of Calibration. Written methods or procedures shall be provided to recall controlled M&TE for recalibration at prescribed intervals to verify the required accuracy. Such intervals may be in calendar time or relate to usage. Interval selection should consider experience, inherent stability, purpose or use, and accuracy required. Historical records shall be maintained which contain sufficient experience data for evaluating calibration intervals.

3.2.5 Calibration Procedures. Written methods or procedures for calibrating M&TE shall be provided to assure calibration technique uniformity. Maximum use shall be made of existing methods, practices or procedures.

Calibration procedures shall include the following information as a minimum:

1. Identity of the item or system to be calibrated.
2. Calibration equipment and standards.
3. Sequence of operations.

4. Checks, tests, measurements and acceptance tolerances.
5. Method of presenting calibration data.
6. Special instructions when necessary.

3.2.6 Traceability. M&TE shall be calibrated utilizing standards whose calibration is traceable to the U. S. National Bureau of Standards, accepted values of natural physical constants, or derived by the ratio type of self-calibration techniques.

3.2.7 Records. The validity of all reference standards shall be supported by certificates, reports, or data sheets which indicate the date of calibration, calibration facility and any deviations from standard environmental conditions. These documents shall also provide evidence of traceability and shall contain data demonstrating conformance with accuracy requirements. Subordinate standards shall be supported by like documents, except that data on accuracy conformance and on environmental conditions may be omitted when not essential to achieving the necessary controls on accuracy.

Records shall be maintained to show that established schedules and procedures for the calibration of standards and M&TE have been followed. The records for M&TE shall contain a history of calibration or other means of control for each item or system of M&TE showing calibration interval, date of last calibration, and conformance or nonconformance to required tolerances before adjustments, including actual values of nonconformances.

3.2.8 Labeling. Each item or system of M&TE shall be labeled to indicate its calibration status. Systems shall show a listing of components. The label shall indicate by whom the M&TE was calibrated and when the next calibration is due. When size or functional characteristics of M&TE prevent the application of a label, an identifying code shall be applied to reflect status. When neither labeling nor coding is practical, the procedures shall provide for monitoring of records to assure control. Labels, codes, or recall records shall be used to identify and control any M&TE whose use must be limited due to imprecision or bias.

3.2.9 Precalibration Checks. The calibration of M&TE submitted for recalibration shall be checked before adjustments or repairs are made.

3.3 Use of M&TE. Controls shall be established and implemented to assure that M&TE consistently produces results of acceptable accuracy. Such measures shall include, as appropriate:

1. Environmental and handling controls.
2. Training and qualification of personnel who use and calibrate M&TE.

-
3. Checking calibration status before use.
  4. Interim checks between calibrations.
  5. Reporting and recalibrating potentially damaged M&TE.

M&TE which has not been properly maintained or calibrated shall be identified as nonconforming and removed from service.

3.4 Nonconformances. If an M&TE item or system is found to be outside of required tolerances, it shall be identified as nonconforming and the using organization shall be notified that all items inspected or tested with the nonconforming device since its last calibration are suspect and should be considered nonconforming until proven otherwise. If the items inspected or tested have been shipped, they shall be identified and reported to the receiving activity, with recommendations for any necessary reinspection.

4. QUALITY ASSURANCE AUDITS

Operations performed in compliance with this standard shall be subject to audit. Audits shall include, but not be limited to, surveillance of calibration operations, reviews of records, review of the calibration program, and checks of actual M&TE calibrations against appropriate standards.