The DUODEHCATRON - A General Purpose
Negative and Positive Ion Source^{*†}

H.T. Richards and G.M. Klody

University of Wisconsin, Madison, Wisconsin

ABSTRACT

The DUODEHCATRON is a DUOplasmatron type Direct Extraction ion source which replaces the filament with a Hollow Cathode Arc running on argon gas. A 3 mm aperture between cathode and intermediate electrode permits use of a gas in the anode region which is different from the argon on which the hollow cathode arc runs. A 100 liter/min forepump differentially pumps the cathode region. This separation of anode and cathode gases allows use of anode gases or vapors which would poison the filament emission of an ordinary duoplasmatron. Off axis extraction of negative ions formed by dissociative attachment of the anode gas or vapor gives intense and bright beams of negative ions. We have produced $\geq 100 \mu\text{a H}^-$ and $\geq 10 \mu\text{a}$ of BH_4^- , NH_2^- , O^- , OH^- , F^- , Cl^- . On axis extraction of positive ions gives the normal bright and intense beams of a duoplasmatron. We have now installed the source for routine use on the University of Wisconsin tandem accelerator.

* Work supported in part by the U.S. Atomic Energy Commission

† This paper was presented at the 2nd International Conference on Ion Sources, Vienna, 11-16th September 1972.

NOTICE

This report was prepared as an account of work sponsored by the United States Government. Neither the United States nor the United States Atomic Energy Commission, nor any of their employees, nor any of their contractors, subcontractors, or their employees, makes any warranty, express or implied, or assumes any legal liability or responsibility for the accuracy, completeness or usefulness of any information, apparatus, product or process disclosed, or represents that its use would not infringe privately owned rights.

MASTER

DISTRIBUTION OF THIS DOCUMENT IS UNLIMITED

DISCLAIMER

This report was prepared as an account of work sponsored by an agency of the United States Government. Neither the United States Government nor any agency Thereof, nor any of their employees, makes any warranty, express or implied, or assumes any legal liability or responsibility for the accuracy, completeness, or usefulness of any information, apparatus, product, or process disclosed, or represents that its use would not infringe privately owned rights. Reference herein to any specific commercial product, process, or service by trade name, trademark, manufacturer, or otherwise does not necessarily constitute or imply its endorsement, recommendation, or favoring by the United States Government or any agency thereof. The views and opinions of authors expressed herein do not necessarily state or reflect those of the United States Government or any agency thereof.

DISCLAIMER

Portions of this document may be illegible in electronic image products. Images are produced from the best available original document.

The off-axis direct extraction duoplasmatron source as developed by Lawrence, Beauchamp, and McKibben¹ gives unexcelled performance for intense high quality beams of H^- and D^- . It also gives excellent beams of heavier negative ions when the appropriate parent gas for dissociative attachment exists.² However, for the heavy negative ions, filament lifetime is short because of sputtering. Also certain parent gases poison the emission so effectively as to preclude their use. In our experience most boron compounds that could give negative boron ions seem to fall in this latter category.

To solve earlier filament problems Dawton³ replaced the filament by a mercury pool cathode. This arrangement is still used at Harwell. However, the complexity of the source, the need of cold traps to control the mercury vapor, and perhaps some personal prejudices against the use of mercury lead us to consider alternate solutions.

In particular, the high density positive column plasmas achieved by Lidsky et al.⁴ with hollow cathode arcs (HCA) running on argon gas attracted our interest. Then Gerry and Rose⁵ showed that the positive column of the discharge would pass through small apertures and that the spectra were characteristic of the gases on each side of the aperture. We combined these approaches to give a simple general purpose negative ion source. We called the resultant source⁶ a DUODEHCATRON since it employed the basic DUOp lasmatron geometry of the Direct Extraction sources but had a Hollow Cathode Arc in place of the ordinary filament.

The simplicity and performance of the source quickly exceeded our expectations though the August 24, 1970 bombing of Sterling Hall destroyed our lab and delayed development by almost two years.

Figure 1 shows schematically a hollow cathode arc. The argon gas

flows continuously through the small tantalum tube which constitutes the hollow cathode. A 100 liter/minute forepump maintains a pressure of ~ 80-100 microns in the region exterior to the Ta tube. With 200 volts applied between the cathode and anode the arc strikes if an external r.f. field temporarily ionizes the argon external to the tube. The resultant positive ions bombard the cathode and those ions that enter the tube initiate the hollow cathode arc. As a result the end of the Ta tube glows bright red for several diameters from the tip. The mechanisms involved in the HCA are complicated and for details we refer the reader to reference 4. Lidsky et al.⁴ claim that the electron emission is primarily thermionic from the hot tip though field emission and secondary emission play a role. The superiority of argon gas with tantalum cathodes may suggest a more complicated mechanism perhaps involving metastable states of argon.

Fig. 1
Hollow Cathode Arc

Figure 2 shows an overall schematic diagram of our source, extraction system, and einzel lens arrangement. Note that the source is basically a direct extraction type duoplasmatron where the off-axis extraction results from the motion of the left hand portion of the source assembly about the zwischen pivot. The anode gas feeds in through the zwischen electrode. The aperture (D) of boron nitride separates the anode gas from the argon of

Figure 2

DUODEHCATRON

the cathode region. The argon gas enters at A and flows through the tantalum tube which in this version of the source has a heat shield surrounding it to lower the power requirements and simplify cooling requirements. The total power consumption of the source is typically between 150-500 watts and forced air cooling is simple and adequate. The cooling air enters (C.A. in) at the top of the zwischen electrode and leaves at the bottom (C.A. out). A copper disk brazed to the tip of the zwischen electrode aids in the heat transfer and also carries the groove for a viton "O" ring. The O-ring provides both the insulation for anode to zwischen voltage and a sliding vacuum seal as the assembly moves about the pivot for off axis extraction. Figure 3 shows these details more clearly. In particular a thin annular disk of boron nitride between the zwischen and anode is important to shield the viton "O" ring from the plasma and to prevent discharge and sputtering between the zwischen and anode.

The differentially pumped hollow cathode is in a Pyrex glass pipe tee with conical ends, (commercially available from Corning Glass Company). The commercial cast aluminum flanges and the teflon gaskets supplied also by Corning Glass Company are both cheap and reliable.

The source operation is similar to an ordinary direct extraction duoplasmatron except for the starting. To start the source we connect a small commercial Tesla coil (of the type used in leak chasing glass systems) to a wire around the Pyrex pipe between cathode and anode and then slowly increase the argon gas flow until the arc strikes (usually at a pressure of 200-300 μ on the thermocouple gauge). When the arc strikes we disconnect the Tesla coil and reduce the argon flow to operating range ($\sim 80 \mu$). Then we slowly increase the anode gas flow until the optimum negative ion current occurs. This gas flow typically corresponds to $\sim 20 \mu$ rise on the

thermocouple vacuum gauge in the cathode region.

Starting is easier and operating temperatures of the Ta tube tip are lower if barium carbonate or LaB_6 coatings are inside the tantalum tube.

The hollow cathode arc will run stably for currents in excess of 3-4 amperes. To run stably at low currents (and at lower gas flows) one can insert a bundle of small Mo tubes (for example 4 tubes 1.0 mm O.D. and 0.6 mm I.D.) into the tip of the Ta tube.⁷ Then at low arc currents the arc extinguishes in some tubes.

Table I shows test stand results for different negative ions

Table I

Negative ions from the DUODEHCATRON for 0.8 mm O.D. anode aperture,

$i_{\text{arc}} = 3-5 \text{ a}$, $V_{\text{arc}} = 60-100 \text{ v}$, $V_{\text{extraction}} = 10-16 \text{ kV}$,

$i_{\text{extraction}} = 1-2 \text{ ma}$, argon flow 0.5 atm ml/s

Negative Ion	Anode Gas	Typical Analyzed Current (μa)
H^-	H_2	40-100
BH_4^-	$\text{Zr}(\text{BH}_4)_4$	4-10
NH_2^-	NH_3	4-10
O^-	CO_2	10
	O_2	7
OH^-	$\text{O}_2 + \text{H}_2$	16
F^-	Contamination from exposed teflon gasket	~ 10
Cl^-	Si Cl_4	20

REFERENCES

1. G.P. Lawrence, R.K. Beauchamp and J.L. McKibben, Nuc. Inst. and Meth. 32, 357 (1965)
2. L.A. Jacobson and H.T. Richards, Bull. Am. Phys. Soc. Vol. 12, 1203, (1967)
3. R.H.V.M. Dawton, Nature, 195, 1190 (1962), Nuc. Inst. and Meth. 24, 285, (1963)
4. L.M. Lidsky, S.D. Rothleder, D.J. Rose, S. Yoshikawa, C. Michelson and R.J. Mackin Jr., J. App. Phys. 33, 2490, (1962)
5. E.T. Gerry and D.J. Rose, J. App. Phys. 37, 2725 (1966)
6. G.M. Klody and H.T. Richards, Bull. Am. Phys. Soc. 15, 642, (1970); 17, 570, (1972)
7. J.L. Deleroix, H. Minoo, and A.R. Trindade, Rev. Sci. Ins. 40, 1555 (1969)

Figure 3
Detail of the zwischen
and anode region