ARGONNE NATIONAL LABORATORY
9700 South Cass Avenue
Argonne, Illinois 60439

MAGRAM--A Computer Code for
Quantitative Electron-microprobe Analysis
of Radioactive Materials

by

R. Natesh,* E. M. Butler,
and D. R. O'Boyle

Materials Science Division

NOTICE

This report was prepared as an account of work sponsored by the United States Government. Neither the United States nor the United States Atomic Energy Commission, nor any of their employees, nor any of their contractors, subcontractors, or their employees, makes any warranty, express or implied, or assumes any legal liability or responsibility for the accuracy, completeness or usefulness of any information, apparatus, product or process disclosed, or represents that its use would not infringe privately owned rights.

October 1971

*Now with EBR-II Project

DISTRIBUTION OF THIS DOCUMENT IS UNLIMITED
fy

DISCLAIMER

This report was prepared as an account of work sponsored by an agency of the United States Government. Neither the United States Government nor any agency Thereof, nor any of their employees, makes any warranty, express or implied, or assumes any legal liability or responsibility for the accuracy, completeness, or usefulness of any information, apparatus, product, or process disclosed, or represents that its use would not infringe privately owned rights. Reference herein to any specific commercial product, process, or service by trade name, trademark, manufacturer, or otherwise does not necessarily constitute or imply its endorsement, recommendation, or favoring by the United States Government or any agency thereof. The views and opinions of authors expressed herein do not necessarily state or reflect those of the United States Government or any agency thereof.

DISCLAIMER

Portions of this document may be illegible in electronic image products. Images are produced from the best available original document.

TABLE OF CONTENTS

	<u>Page</u>
ABSTRACT	5
I. INTRODUCTION	5
II. CORRECTION PROCEDURES	6
A. Detector Dead Time, Instrument Drift, and Background Radiation	7
B. Atomic-number Correction	9
C. Absorption Correction	9
D. Fluorescence Correction	10
III. PROGRAMMING CONSIDERATIONS	11
A. Main Program	11
B. Plotting Subroutines	11
IV. MAGRAM INPUT AND OUTPUT	12
APPENDIXES	
A. Listing of the MAGRAM Code	20
B. List of Symbols	46
C. Programming Notes for Plotting Subroutines	47
1. Subroutine GRAPHS	47
2. Subroutine CALPLT	47
3. Printer Subroutine PPLOT	47
D. Listing of MAGRAM Input and Output	48
1. MAGRAM Input	48
2. MAGRAM Output	48
ACKNOWLEDGMENTS	53
REFERENCES	54

LIST OF FIGURES

<u>No.</u>	<u>Title</u>	<u>Page</u>
1.	Corrected Uranium, Plutonium, and Uranium-plus-Plutonium Concentrations as a Function of Radial Distance from the Fuel Centerline	19
2.	Corrected Weight Ratios of Plutonium-to-Uranium and Plutonium-to-Uranium-plus-Plutonium as a Function of Radial Distance from the Fuel Centerline	19

MAGRAM--A Computer Code for Quantitative Electron-microprobe Analysis of Radioactive Materials

by

R. Natesh, E. M. Butler, and D. R. O'Boyle

ABSTRACT

A FORTRAN-IV computer code, called MAGRAM, has been developed to correct and plot X-ray data collected with a shielded electron-microprobe analyzer. The measured X-ray intensities of up to three elements are corrected for system dead-time effects, beam-current drift, X-ray background, atomic number, X-ray absorption, fluorescence effects, and gamma-induced secondary radiation. The program allows a maximum of 300 separate analyses to be corrected in a single run and has printer and CALCOMP plotting options for 1 to 99 points. The application of the program to measurements of the radial redistribution of uranium and plutonium in irradiated mixed-oxide fuel is illustrated.

I. INTRODUCTION

Electron-microprobe analysis is a widely used experimental technique for determining the chemical composition of a small volume of material (approximately $1\ \mu^3$) by excitation of the microvolume with a finely focused electron beam. The beam, usually $<1\ \mu$ in diameter, generates characteristic X rays in the microvolume, which are diffracted by a crystal X-ray spectrometer and counted by standard X-ray techniques. Because of the excellent resolution of the method, microprobe analysis has been useful in diversified metallurgical studies including phase equilibria, diffusion, and corrosion phenomena. In the field of nuclear materials, the electron-microprobe analyzer has been used to study compatibility reactions between fuel and cladding materials, solid-state diffusion kinetics, and metallic and ceramic multicomponent phase systems. Thus it seemed desirable to adapt microprobe techniques to investigate the complex fuel-cladding-fission-product reactions that occur in highly irradiated fuel elements.

Early microprobe analysis of irradiated fuel specimens¹⁻³ was conducted at several laboratories using commercial microprobe analyzers without special gamma shielding. To reduce the fluorescent background due to hard gamma radiation and to facilitate handling, the mass of the specimen being analyzed was kept small so that the gamma activity was sufficiently low to allow the specimen to be handled outside a hot cell.

This technique was appropriate for some initial studies of irradiated fuel; however, an examination of an entire cross section of an irradiated fuel element would be more efficient if both the specimen chamber and X-ray detection system were shielded. In response to the need for analyzing reactor fuel specimens, a commercial microprobe⁴ was developed that has the equivalent of 4.3 in. of tungsten shielding around the specimen chamber so that a highly irradiated specimen (gamma activity, 50 R at 9 in.) could be examined safely.⁵

Although the shielding greatly reduces the gamma-induced scattered fluorescent and secondary radiation detected by the proportional counters, in practice it is necessary to measure and correct for this background contribution before applying corrections for X-ray absorption, fluorescence, and atomic-number effects. This report describes a computer program for correcting the X-ray data obtained with a shielded microprobe analyzer and for plotting the corrected weight concentrations of the elements. The X-ray intensity measurements recorded on the radioactive specimen are corrected for irradiation-induced fluorescent radiation, and the corrected concentrations of each element are plotted using a plotting subroutine adapted by Fernow.⁶ Corrections for detector dead time, beam-current drift, X-ray absorption, characteristic line fluorescence, and atomic-number effects are calculated using the MAGIC-II computer code developed by Gray.⁷ The correction program developed by Gray was selected in preference to other published programs⁸⁻¹⁴ because it performs a detailed calculation of each correction factor, has a comprehensive list of mass-absorption coefficients, and has an advanced method of calculating the background correction. For use in the MAGRAM code, Gray's table of mass-absorption coefficients was extended by Savage¹⁵ to include elements with atomic numbers up to 94. The Microprobe Analysis General Correction for RadioActive Materials (MAGRAM) code was developed to quantitatively analyze radioactive samples and, in particular, to analyze X-ray data obtained from irradiated reactor fuels and cladding materials.

II. CORRECTION PROCEDURES

The physical basis for quantitative microprobe analysis is a measurement of the ratio of the number of X-ray photons emitted from the specimen being analyzed (unknown) to the number of photons emitted from a standard of known composition under identical instrument conditions. During both measurements, the accelerating voltage, beam current, spectrometer settings, and counting-time interval are held constant. For quantitative analysis of highly radioactive samples, the measured X-ray intensities must be corrected for both experimental errors (dead time, drift, and natural and gamma-induced background) and X-ray interaction effects (atomic-number effect, mass absorption, and fluorescence effects). The ratio of the characteristic X-ray intensity of element A from the unknown

to that of the standard is defined as the "k ratio" of element A. To establish the k ratio, the measured emergent X-ray intensities are corrected for errors due to system dead time and instrumental drift and for background radiation effects. After these corrections have been made, the k ratio for each element is calculated and used as the first approximation of the mass concentration C of each element n in the unknown. Finally, to obtain the true mass concentration, the k ratio is corrected for atomic-number effects, X-ray absorption effects, and secondary fluorescence excited by other characteristic lines. The usual form of the X-ray interaction correction is

$$C_n = k_n k_z k_a k_f, \quad (1)$$

where

C_n = true mass concentration of each element in the unknown,

k_n = relative intensity of emergent radiation for each element
(after correction for dead time, drift, and background),

and

k_z, k_a, k_f = computed correction factors for atomic number, absorption, and fluorescence effects.

Since k_z, k_a , and k_f are dependent on the mass concentration, an iterative procedure to calculate C_n is used in the MAGRAM code.

A detailed treatment of the atomic number, absorption, and secondary fluorescence correction procedures that are incorporated in the MAGRAM code is given by Gray.⁷ In addition, several excellent reviews¹⁶⁻¹⁸ of the current theory of quantitative electron-microprobe analysis are available. The instrumental and X-ray corrections incorporated in the MAGRAM code are briefly discussed in the following sections.

A. Detector Dead Time, Instrument Drift, and Background Radiation

Because of coincidence losses, the microprobe X-ray detection system does not count every incident photon. The first operation performed on all measured data consists of a correction for dead-time losses using Heinrich's method¹⁹ as programmed by Gray.⁷ During an analysis, the incident electron-beam intensity may vary by as much as $\pm 2\%$; thus a correction for instrument drift is required. The MAC-450 microprobe⁴ has provisions for monitoring the incident electron-beam current on an aperture located above the objective lens. The correction for instrument drift is made by monitoring the electron-beam current and normalizing all intensity readings to the same beam current. When this correction is applied, it is assumed that the X-ray intensity is directly proportional to the beam

current. Experiments performed at ANL have shown that this is a valid assumption, provided the variation observed is $\approx 3\%$ for a properly aligned system.

The background observed during the analyses of nonradioactive samples is primarily due to the continuous X-ray spectrum. In the analyses of radioactive materials, an additional correction must be made for the background due to the scattered secondary and fluorescent radiation. Natesh *et al.*⁵ found that the radiation-induced background intensity varied with specimen location relative to the X-ray detectors and have described a procedure to correct the raw data for this background. The background due to the gamma activity of the specimen is measured by turning the electron beam off and recording the background intensity at the characteristic wavelength of the element being analyzed. While these data are being collected, the spectrometer and pulse-height-analyzer settings are the same as those used when measuring the X-ray intensity on the sample.

In the MAGRAM code, the raw X-ray data are first corrected for dead time in the detection system. Following this correction, the background due to scattered radiation from the radioactive specimen is subtracted. The program then corrects for instrument drift and background due to the X-ray continuum. The intensity measurements from the standards are not corrected for radioactive background since the standards are remote from the sample and the background contribution from the radioactive specimen is negligible.

The background due to X-ray continuum is the sum of the background contributions from all elements present in the unknown sample. Thus, if n elements are present, the weighted background contributions from all n elements must be added to calculate the total continuum X-ray background. For example, consider the case in which elements A, B, C, and D are present in an unknown. To determine the total background due to the continuum, the spectrometer is tuned to the characteristic wavelength of element A, and readings are taken on the B, C, and D standards. Next, the element A standard is placed under the electron beam, and the spectrometer is detuned by moving $5/2$ full-width-at-half-maximum off the peak on both sides of the characteristic line peak for element A, and readings are taken and averaged. The background due to the continuum for element A in the unknown is equal to the weighted background intensity obtained by detuning the spectrometer on the standard for element A plus the weighted contributions of the A peak intensities obtained on the B, C, and D standards. The advantage of this method of calculating background is that the spectrometer is never detuned while making measurements on the unknown. By subtracting the sum of the backgrounds due to the X-ray continuum and gamma activity from the observed line intensity, a corrected line intensity is obtained.

B. Atomic-number Correction

The atomic-number correction is applied to correct for the variation in efficiency of producing characteristic radiation from the elements in both the sample and the standard. The correction factor k_z depends on the relative stopping power and the backscattering efficiency of the sample and the standard and is calculated from

$$k_z = \frac{R_A(Z/A \ln 1.166E/J)}{\bar{R}\bar{Z}\bar{A}/\bar{A}_A \ln (1.166E/J_A)}, \quad (2)$$

where

R = backscatter loss factor,

Z = atomic number,

A = atomic weight,

E = mean excitation potential = $\frac{E_O + E_C}{2} A$,

E_O = electron-beam accelerating voltage,

E_C = X-ray line-excitation voltage,

and

J = ionization potential of the elements

$$= 14.0 Z [1.0 - \exp(-0.1\bar{Z})] + \frac{75.5Z}{\bar{Z}/7.5} - \frac{\bar{Z}Z}{100 - \bar{Z}}.$$

The overscored terms indicate a weighted average over the composition of the sample.

C. Absorption Correction

The absorption correction factor k_a compensates for the difference in absorption of the characteristic X radiation in the unknown and standard. In the absorption correction, both the X-ray absorbing power of the specimen and the depth below the specimen surface at which the X rays are produced are considered. The latter factor involves the depth of penetration of the electron beam, which depends on the electron-absorption process. The X-ray absorption factor given by Philibert²⁰ is

$$k_a = \frac{1 + \left(\frac{hR(O)}{R(\infty) + hR(O)} \right) \frac{\chi}{\sigma}}{\left(1 + \frac{\chi}{\sigma} \right) \left(1 + \frac{h}{1 + h} \frac{\chi}{\sigma} \right)} \times \frac{\left(1 + \frac{\bar{\chi}_A}{\sigma} \right) \left(1 + \frac{\bar{h}}{1 + \bar{h}} \frac{\bar{\chi}_A}{\sigma} \right)}{1 + \left(\frac{\bar{h}R(O)}{R(\infty) + \bar{h}R(O)} \right) \frac{\bar{\chi}_A}{\sigma}}, \quad (3)$$

where

$$\chi = \left(\frac{\mu}{\rho} \right) \operatorname{cosec} \theta,$$

$$\left(\frac{\mu}{\rho} \right)_i^A = \text{mass-absorption coefficient of the } i\text{th element for A radiation,}$$

$$\bar{\chi}_A = \sum \left(\frac{\mu}{\rho} \right)_n^A C_n, \text{ where } C_n \text{ is the weight fraction of the } n\text{th element in the unknown,}$$

$$\theta = \text{X-ray take-off angle (45° for the MAC probe),}$$

$$\sigma = \text{electron mass-absorption coefficient}$$

$$= 2.54(10^5)/(E_o^{1.5} - E_c^{1.5}),$$

$$h = 4.5A/Z^2,$$

and

$$R( ) = \text{electron backscatter factor. } R(O) \text{ is taken as 1.1, and } R(\infty) \text{ is taken as 4.0.}$$

The overscored terms indicate a weighted average over the composition of the specimen, and the subscripts refer to values for a given element.

D. Fluorescence Correction

The fluorescence correction factor k_f used in the MAGRAM code for fluorescence of element A by B radiation is given by $k_f = 1/1 + k_{fA}$, where

$$k_{fA} = 0.5PC_B \frac{r_A - 1}{r_A} \omega_B \frac{A_A}{A_B} \left(\frac{\frac{E_o}{E_{CB}} - 1}{\frac{E_o}{E_{CA}} - 1} \right)^{1.67} \frac{(\mu/\rho)_A^B}{\bar{\chi}_B} \left(\ln \frac{1 + \alpha}{\alpha} + \ln \frac{1 + \beta}{\beta} \right), \quad (4)$$

$$\alpha = \frac{(\mu/\rho)_A^A}{\bar{\chi}_B} \operatorname{cosec} \theta,$$

$$\beta = \frac{\sigma}{\bar{\chi}_B},$$

P = a constant, the value of which depends upon the X-ray series (K, L, M) to which the fluorescing and fluoresced spectral lines belong, e.g., for fluorescence of an L line by a K line $P = 2.4$,

r_A = absorption jump ratio, the ratio of the mass-absorption coefficients on either side of the relevant absorption edge,

and

ω_B = fluorescent yield for element B, i.e., the probability of ionized B atoms emitting an X-ray photon of characteristic wavelength λ_B .

III. PROGRAMMING CONSIDERATIONS

A. Main Program

The output data from the MAC microprobe are automatically punched on paper tape and are then corrected by the main MAGRAM program for the effects discussed in Sec. II. The main program has 860 statements and no subroutines, and is stored on magnetic tape. To run this program, the largest option on the IBM-FORTRAN H compiler is used. Core storage required is 250K bytes. A complete listing of the MAGRAM code, including the various plotting subroutines, is given in Appendix A. Throughout the program, comment statements have been liberally inserted to define the calculations in process and to explain the program logic. Typical running time on the IBM-360/75 required to correct data for three elements measured at 100 locations is less than 20 sec. The matrix of mass-absorption coefficients and other constants given at the end of the main program greatly reduces the number of data input cards. At ANL, the main program and plotting subroutines are stored on magnetic tape. This procedure reduces the number of input cards to a minimum.

The MAGRAM program has been designed to correct input data for one to three elements. Corrections for an additional element are calculated by difference. Up to 300 data points (three elements each) may be corrected by the main program. By suitable changes in format, raw data for eight elements may be corrected. A list of symbols used in the main program and their common designations are given in Appendix B.

B. Plotting Subroutines

To obtain concentration plots as a function of specimen position, several plotting subroutines were added to the main program. One subroutine creates a printer plot, which follows the tabular output of the main program. A second subroutine causes CALCOMP graphs of the corrected concentrations to be prepared off line. Both plotting options exist as library programs at ANL and would have to be modified for use at another installation. The present plotting subroutines allow 99 data points (three elements each) to be corrected and plotted.

By means of the GRAPHS subroutine, the user indicates whether plots are to be generated and, if so, the types desired. GRAPHS also calculates weight ratios needed for the weight-ratio plots and stores these data and the position data (called RAD for radial position). Either or both of the two plotting routines are called by GRAPHS through the use of a test variable. Additional description of the two plotting routines and the test variables is given in Appendix C.

IV. MAGRAM INPUT AND OUTPUT

The input for MAGRAM consists of two parts. One is the internally stored data matrix listed at the end of the program (see Appendix A). The other part consists of initialization cards that identify the problem and contain the data obtained with the microprobe. This section describes the latter part of the input and the entire output.

The first two cards identify the problem, identify the person submitting the problem, and describe the sample; this information appears on the printed output. The third card contains information regarding the instrument parameters used for the analysis. Included are items such as the accelerating potential, take-off angle, and detection-system dead time for each spectrometer. The fourth and fifth cards contain information regarding the elements analyzed and the types of standards used in the analysis. The remainder of the input data consists of cards listing the measurements obtained with the microprobe. If graphs are desired, an additional set of cards that lists the position of the analysis must be present. Before the plotting routines listed in Appendix A can be run at another computer installation, the computer center representative should be consulted. The following lists describe the input cards in detail.

Input Data Format

<u>Column</u>		<u>Format</u>
<u>First Card</u>		
1-5	Problem number, or blank	A1, I4 (if blank, assigned by computer)
6-7	Blank	2X
8-22	Date of analysis	(3A4, A3)
23-24	Blank	2X
25-44	Name of person submitting problem	5A4
<u>Second Card</u>		
1-71	Description of sample	71A1

<u>Column</u>		<u>Format</u>
<u>Third Card</u>		
1-4	Accelerating voltage, keV	F4.1
5-6	Blank	2X
7-8	Number of compound standards (if blank, elemental standards assumed)	I2 (number is right justified)
9-10	Blank	2X
11-15	Counting interval, sec	F5.1 (if blank, 1.0 sec is assumed)
16-17	Blank	2X
18-22	Density of sample, g/cc	F5.2 (may be left blank)
23-24	Blank	2X
25-29	Film thickness, μ	F5.2 (may be left blank)
30-31	Blank	2X
32-35	X-ray emergence angle, deg	F4.1 (if blank, 52.5° is assumed)
36-37	Blank	2X
38-40	Dead time for element 1, μ sec	F3.1 (if blank, 1.0 μ sec is assumed)
41-43	Dead time for element 2, μ sec	F3.1 (if blank, 1.0 μ sec is assumed)
44-46	Dead time for element 3, μ sec	F3.1 (if blank, 1.0 μ sec is assumed)
<u>Fourth Card</u>		
1-2	First element analyzed, chemical symbol	A2
3	Blank	1X
4-5	First element analytical line, KA, LA, MA	A2
6-7	Blank	2X
8-9	Second element analyzed, chemical symbol	A2
10	Blank	1X

<u>Column</u>	<u>Format</u>

Fourth Card (Contd.)

11-12	Second element analytical line	A2
13-14	Blank	2X
15	Repetition of above data for all elements	

If one more element is present in the sample than is analyzed (e.g., oxygen), and the concentration of that element is to be determined by difference, its chemical symbol is listed last without indicating an analytical line. Chemical symbols are right justified in their field. Analytical lines are indicated as KA, LA, MA, or MB for K alpha, L alpha, M alpha, or M beta. A typical sequence is given as: $\overline{\text{NB-LA}}\text{--}\overline{\text{Zr-LA}}\text{--}\overline{\text{N-KA}}\text{--}\overline{\text{O}}$, where the dashes indicate blanks.

<u>Column</u>	<u>Format</u>

Fifth Card

1-2	Number of the element in the list on the fourth card to which compound standard applies	I2 (right justified)
3-7	Standard name	A5 (left justified)
8-13	Weight fraction of element in binary-compound standard	F6.4
14-15	Chemical symbol of other element in compound standard	A2 (right justified)

A separate card is required for each element for which a binary-compound standard is used. The total number of cards required is equal to the "number of compound standards" listed on the second card.

Data from Standards

<u>Column</u>	<u>Format</u>

First Card

1-2	Number of observations on the standard	I2

Other Cards

1-2	Blank	2X
3-7	Beam current	F5.0
8-13	Blank	6X
14-18	Standard counts	F5.0

The above cycle is repeated for each element listed on the fourth card, in the order of listing. The total number of sets of standard data must equal the total number of elements analyzed.

Background Data from the First Standard

<u>Column</u>		<u>Format</u>
<u>First Card</u>		
1-2	Number of background observations taken on first standard	I2 (right justified)
<u>Other Cards</u>		
1-2	Blank	2X
3-7	Beam current	F5.0
8-13	Blank	6X
14-18	Background taken on first standard; spectrometer detuned	F5.0
19-24	Blank	6X
25-29	Background taken on first standard; spectrometer tuned to second analytical line	F5.0
30-35	Blank	6X
36-40	Background taken on first standard; spectrometer tuned to third analytical line	F5.0

The number of background data cards must equal the number of observations of background on the first standard.

Background Data from the Second Standard

<u>Column</u>		<u>Format</u>
<u>First Card</u>		
1-2	Number of background observations taken on second standard	I2 (right justified)
<u>Other Cards</u>		
1-2	Blank	2X
3-7	Beam current	F5.0
8-13	Blank	6X
14-18	Background taken on second standard; spectrometer tuned	F5.0

<u>Column</u>		<u>Format</u>
<u>Other Cards (Contd.)</u>		
19-24	Blank	6X
25-29	Background taken on second standard; spectrometer detuned	F5.0
30-35	Blank	6X
36-40	Background taken on second standard; spectrometer tuned	F5.0

The number of background data cards must equal the number of observations of background on the second standard. The above sequence is repeated for data taken on each standard to determine contributions to other analytical lines. The total number of sets of background data must be equal to the total number of elements analyzed.

Radioactive-decay Data Format
(Electron beam turned off)

<u>Column</u>		<u>Format</u>
1-13	Blank	13X
14-18	Decays in counting time interval from first element analyzed	F5.0
19-24	Blank	6X
25-29	Decays in counting time interval from second element analyzed	F5.0
30-35	Blank	6X
36-40	Decays in counting time interval from third element analyzed	F5.0

Sample Data Format

1-2	Blank	2X
3-7	Beam current	F5.0
8-13	Blank	6X
14-18	Counts for first element	F5.0
19-24	Blank	6X
25-29	Counts for second element	F5.0
30-35	Blank	6X
36-40	Counts for third element	F5.0

This card is repeated for each observation. A maximum of 100 observations are permitted.

Trailer Card

After the last analysis in a problem, a trailer card containing routing information is required.

<u>Column</u>		<u>Format</u>
3-7	99998 is inserted if data for a new problem follow	F5.0
	99999 is inserted if no new problem follows	

Graph-options Card

The first five columns of the card are used to designate the data to be plotted.

<u>Column</u>	<u>Plot Requested</u>
1	U versus radius, wt % (U, Pu, and (U + Pu) versus radius, wt %)*
2	Pu versus radius, wt %
3	(U + Pu) versus radius, wt %
4	Weight ratio Pu/U versus radius
5	Weight ratio Pu/(U + Pu) versus radius

The type of graph desired (printer plot or CALCOMP) is indicated by inserting code numbers 1-5 in columns 1-5.

<u>Code Number</u>	<u>Graph Option</u>
1	No graph
2	Printer plot only
3	CALCOMP graph only
4	Printer plot and CALCOMP graph
5	(Used only in column 1) A CALCOMP graph of U, Pu, and (U + Pu), wt % versus radius, and a CALCOMP and printer plot of U wt % versus radius

*A CALCOMP graph of the three functions on a single plot is produced only when 5 is used in column 1.

For example, a card with 52311 in the first five columns would request three separate CALCOMP graphs (the three weight percents versus radius, uranium versus radius, and (U + Pu) versus radius) and printer plots of uranium versus radius and plutonium versus radius.

The printer plot produces a maximum of two separate graphs. If the plots corresponding to columns 1, 2, and 3 are called, they will all appear on a single printer plot. If the plots corresponding to columns 4 and 5 are called, they will both be printed on a second graph.

The CALCOMP routine generates a maximum of six graphs. Each plot corresponding to columns 1-5 is prepared on a separate graph. A sixth CALCOMP graph containing three plots is obtained by using the number 5 in column 1 of the graph options card.

Radial-distance Card

One radial distance card (radii measured in mils) must be supplied for each observation if graphs are desired.

<u>Column</u>	<u>Format</u>
1-10 Radius	F10.0

Either right-justify the numbers to column 10, or code with a decimal point.

End Trailer Card

After the radial distance card for the last observation, a trailer card with 0.000 in the first five columns must be inserted in the deck.

A sample listing of the input and output cards for the MAGRAM code is given in Appendix D. Explanatory notes have been added to clarify possible points of confusion. The output contains a listing of the constants used in the correction process. These include the atomic numbers of the elements analyzed, absorption-edge wavelengths, mass-absorption coefficients, and backscatter factors. The standard intensities given are averages based on the input data and have been normalized to a constant incident-beam current, which is also listed. At ANL, the aperture current is monitored continuously and is assumed to be proportional to the beam current. The output listing also contains the continuum-background radiation matrix. In the example shown, the effect of interference from $UM\gamma$ on PuM_{β} is apparent. This interference, which is a function of the uranium concentration at the point being analyzed, is automatically corrected by the program. The X-ray intensity measurements, corrected for detector dead time and normalized to a constant beam current, appear in tabular form in the output. The K ratios for each element at each point are also listed. If compound standards are used, a note to that effect is

printed. The instrument data, the peak-to-background ratios for each element, and the computed minimum detectability limits are listed. The corrected concentrations, in terms of weight and atomic percents, are listed in the final output table. The depth of the region analyzed (in microns) is also included in the output. As the program is being run, a magnetic tape is generated containing all the data necessary to prepare CALCOMP plots, which are then prepared off line.

The printer plots, which are generated along with the computer output, give a good first approximation of concentration as a function of distance. Points are placed on the graph by means of a truncation method with a "resolution" of about 2 wt % on the ordinate and 1.25 mils on the abscissa. Weight percents of uranium, plutonium, and the sum of uranium and plutonium are plotted as a function of radial position. An additional printer graph contains plots of the weight ratios of plutonium-to-uranium, and plutonium-to-uranium-plus-plutonium. The CALCOMP graphs more accurately plot the same data, since the data points are positioned more accurately and the CALCOMP subroutine selects the scale range based upon the range of values plotted. Thus, if five separate CALCOMP graphs are prepared, each may have a different scale depending on the range of values plotted. Sample CALCOMP graphs of the uranium and plutonium distribution in a mixed-oxide fuel element are shown in Figs. 1 and 2.

Fig. 1

Corrected Uranium, Plutonium, and Uranium-plus-Plutonium Concentrations as a Function of Radial Distance from the Fuel Centerline. Fuel is UO_2 -20 wt % PuO_2 irradiated in EBR-II to 2.8 at. % burn-up at a linear power rating of 17 kW/ft. Neg. No. MSD-53993.

Fig. 2

Corrected Weight Ratios of Plutonium-to-Uranium and Plutonium-to-Uranium-plus-Plutonium as a Function of Radial Distance from the Fuel Centerline. Same mixed-oxide fuel element described in Fig. 1. Neg. No. MSD-53992.

DIMENSION AC(8,9), B(8,9)

```

DIMENSION ROUT(8,300),P(100,36)
DIMENSION NXI(8,300),MSI(8,300)
DIMENSION APC(9,300),DECAY(300)

```

C
C
C
C
C

INITIALIZATION AND INPUT OF BASIC PARAMETERS

```

      READ(1,81) ((P(I,J),J=1,11),I=1,100)
81  FORMAT (F7.3,A3,3F6.3,6F7.3)
      READ(1,82) ((P(I,J),J=12,22),I=1,100)
82  FORMAT (7F7.3,4F5.3)
      READ(1,7) ((P(I,J),J=23,29),I=1,100)
7   FORMAT (F5.3,4F7.3,F10.3,5X,E14.8)
      READ(1,8) ((P(I,J),J=30,34),I=1,100)
8   FORMAT (5E14.8)
      READ(1,5) (MM(I),I=1,4)
5   FORMAT (4A2)
      READ(1,530) ST2,LT2
530 FORMAT (A3,A2)
      READ(1,500) (ABET(I),I=1,26)
500 FORMAT (26A1)
      READ(1,501) PRFX,NMBR
501 FORMAT (A1,I4)
      READ(1,280) (PTITL(K),K=1,400)
280 FORMAT(80A1)
      READ(1,281) (CNAM(I),I=1,40)
281 FORMAT(20A4)

```

C
C
C
C
C

INPUT PROBLEM DATA

```

1  READ 2,TAG,NPROB,DATE,NAME
2  FORMAT(A1,I4,2X,3A4,A3,2X,5A4)
      READ 222,TITLE
222 FORMAT(71A1)
      READ 88,EQ,NB,TIME,RHO,T1,THETA,TAUM
88  FORMAT (F4.1,2X,I2,2X,F5.1,2X,2(F5.2,2X),F4.1,2X,8F3.1)
      READ 6,(EL(I),LINE(I),I=1,9)
6   FORMAT (9(A3,A2,2X))

```

C
C
C
C
C

DETERMINE NUMBER OF ELEMENTS IN SYSTEM, AND NUMBER ANALYZED

```

      DO 509 I=1,9
      IF (EL(I).EQ.ST2) GO TO 518
      IF (LINE(I).EQ.LT2) GO TO 510
509  CONTINUE
518  NEL=I-1
      NA=I-1
      GO TO 511
510  NEL=I
      NA=I-1

```

C
C
C
C
C

ASSIGN PROBLEM NUMBER.

```

511  IF (NPROB.NE.0) GO TO 502
      NMBR=NMBR+1
      IF (NMBR.LE.9999) GO TO 504
      DO 506 L=1,25
      IF (ABET(L).EQ.PRFX) GO TO 507
506  CONTINUE
507  PRFX=ABET(L+1)

```

```

      NMBR=0001
504  NPROB=NMBR
      TAG=PRFX
502  PRINT 615,TAG, NPROB, DATE

```

```

C
C
C      CHECK FOR X-RAY EMERGENCE ANGLE. IF NONE FURNISHED, ASSUME 52.5
C      DEGREES
C

```

```

      IF (THETA.EQ.0.) THETA = 52.5
      CSC=1./SIN(THETA*3.14159/180.)

```

```

C
C
C      DETERMINE OR CALCULATE, FOR EACH ELEMENT ANALYZED, ALL PARAMETERS
C      NECESSARY FOR CORRECTIONS
C

```

```

      PRINT 2000
2000  FORMAT (//4X,7HELEMENT,3X,7HAT. NO.,5X,7HAT. WT.,4X,2HPF,7X,4HED GE
      1,6X,2HEC,7X,4HWAVE,5X,1HU,8X,1HR/)
      DO 95 K=1,NA
      DO 102 I=1,100
      IF (P(I,2).EQ.EL(K)) GO TO 103
102  CONTINUE

```

```

C
C
C      DETERMINE ATOMIC NUMBER AND ATOMIC WEIGHT, AND CALCULATE
C      FACTOR FOR ATOMIC NUMBER CORRECTION
C

```

```

103  Z(K)=I
      A(K)=P(I,1)
      PF(K)=Z(K)*(14.*(1.-EXP(-.1*Z(K)))+75.5/(Z(K)**(Z(K)/7.5))-Z(K)/1
      1100.+Z(K)))
      NL=3

```

```

C
C
C      DETERMINE ANALYTICAL LINE ABSORPTION EDGE, CRITICAL EXCITATION
C      POTENTIAL AND WAVELENGTH, AND CALCULATE OVERVOLTAGE
C

```

```

      DO 4 M=1,2
      IF (LINE(K).EQ.MM(M)) GO TO 3
4      NL=NL+1
      EDGE(K)=P(I,17)
      GO TO 96
3      EDGE(K)=P(I,3*NL)
96      EC(K)=P(I,NL)
      IF (LINE(K).EQ.MM(4)) GO TO 777
      WAVE(K)=P(I,NL+3)
      GO TO 778
777      WAVE(K)=P(I,NL-2)
778      U(K)=EO/EC(K)
      IF (U(K).GT.20.) U(K)=20.
      R(K)=0.
      KK=34

```

```

C
C
C      DETERMINE OVERVOLTAGE DEPENDENT BACKSCATTER LOSS FACTOR
C

```

```

      DO 107 J=1,6
      R(K)=U(K)*R(K)+P(I,KK)
107      KK=KK-1
      PRINT 2001,EL(K),I ,A(K),PF(K),EDGE(K),EC(K),WAVE(K),U(K),R(K)
2001  FORMAT (6X,A3,7X,I2,6X,F8.3,2X,F8.3,2X,F7.3,3X,F6.3,3X,F7.3,3X,F5
      1.2,3X,F8.6)

```



```

95 CONTINUE
   IF (NEL.EQ.NA) GO TO 98
C
C
C   IF AN ELEMENT IS TO BE DETERMINED BY DIFFERENCE, DETERMINE EXTRA
C   ELEMENT ATOMIC NUMBER, ATOMIC WEIGHT, CHARACTERISTIC LINE
C   ABSORPTION EDGE, CRITICAL EXCITATION POTENTIAL AND WAVELENGTH,
C   AND CALCULATE OVERVOLTAGE AND BACKSCATTER LOSS FACTOR
C
C
DO 104 I=1,100
  IF (P(I,2).EQ.EL(NEL)) GO TO 105
104 CONTINUE
105 Z(NEL)=I
   A(NEL)=P(I,1)
   NL=3
   DO 100 J=1,2
     IF (Z(NEL).GE.81.AND.NL.EQ.3) GO TO 100
     IF (P(I,NL).LT.E0) GO TO 99
100 NL=NL+1
     EDGE(NEL)=P(I,17)
     GO TO 101
   99 EDGE(NEL)=P(I,3*NL)
101 EC(NEL)=P(I,NL)
   WAVE(NEL)=P(I,NL+3)
   LINE(NEL)=MM(NL-2)
   U(NEL)=E0/EC(NEL)
   IF (U(NEL).GT.20.) U(NEL)=20.
   R(NEL)=0.
   KK=34
   DO 108 J=1,6
     R(NEL)=U(NEL)+R(NEL)+P(I,KK)
108 KK=KK-1
   PF(NEL)=0.
   PRINT 2001,EL(NEL),I      ,A(NEL),PF(NEL),EDGE(NEL),EC(NEL),WAVE(NEL),
1   U(NEL),R(NEL)
C
C
C   CALCULATE ABSORPTION COEFFICIENT MATRIX
C
C
98 DO 200 I=1,NEL
   DO 200 J=1,NEL
     NZ=Z(J)
     F=1.0
     DO 202 M=9,18
       IF (WAVE(I).LT.P(NZ,M)) GO TO 203
202 CONTINUE
     CON=P(NZ,28)
     EX=P(NZ,23)
     GO TO 201
203 INTER=8
     GO TO (204,205,206,207,208,209,210,211,212,213),INTER
204 CON=P(NZ,24)
     EX=P(NZ,19)
     GO TO 201
205 CON=P(NZ,25)
     EX=P(NZ,20)
     GO TO 201
206 F=0.961
     GO TO 205
207 F=0.917
     GO TO 205
208 CON=P(NZ,26)
     EX=P(NZ,21)
     GO TO 201
209 F=0.984

```

```

      GO TO 208
210 F=0.972
      GO TO 208
211 F=0.946
      GO TO 208
212 F=0.894
      GO TO 208
213 CON=P(NZ,27)
      EX=P(NZ,22)
C
C
C      CALCULATE ABSORPTION COEFFICIENT OF ELEMENT J FOR I RADIATION
C
201 AC(I,J)=(CON*WAVE(I)**EX)**F
200 CONTINUE
      PRINT 52
      IF (NA.LE.3) GO TO 2060
      PRINT 648, (EL(I), LINE(I), I=1,NA)
      PRINT 646
      DO 2050 J=1,NEL
      PRINT 668, EL(J), (AC(I,J),I=1,NA)
2050 CONTINUE
      GO TO 2065
2060 PRINT 647, (EL(I), LINE(I),I=1,NA)
      PRINT 645
      DO 2065 J=1,NEL
      PRINT 667, EL(J), (AC(I,J),I=1,NA)
2065 CONTINUE
C
C
C      CALCULATE COMMON EXPRESSIONS
C
      DO 30 I=1,NEL
      30 WS(I)=Z(I)/A(I)
      PRINT 2002
2002 FORMAT ( //13X,2HWS,8X,1HE,8X,3HVAL,7X,5HSIGMA,6X,2HHS,8X,2HUS,8X,
12HFS,7X,4HFACT/)
      DO 11 I=1,NA
      ALPHA(I)=1.
      E(I)=(EC(I)+EO)/2.
      VAL(I)=7.061+ALOG(E(I))
      SIGMA(I)=2.54E+05/(EO**1.5-EC(I)**1.5)
      HS(I)=4.5*A(I)/(Z(I)*Z(I))
      US(I)=1.+CSC*AC(I,I)/SIGMA(I)
      FS(I)=((1.+HS(I))*(1.+1.1*HS(I)/(4.+1.1*HS(I))*CSC*AC(I,I)/SIGMA(I)
11))/((US(I)*(1.+HS(I)*US(I)))
      FACT(I)=FS(I)*R(I)/(WS(I)*(VAL(I)-ALOG(PF(I))))
2003 FORMAT ( 6X,A3,1X,F7.4,3X,F7.4,3X,F7.4,3X,F8.2,2X,F7.4,3X,F7.4,3X,
1F7.4,3X,F8.5)
      11 PRINT 2003,EL(I),WS(I),E(I),VAL(I),SIGMA(I),HS(I),US(I),FS(I),FACT
1(I)
      DO 900 I=1,9
      FYR(I)=0.
      AJR(I)=0.
      DO 900 J=1,8
      900 B(J,I)=0.
C
C
C      CALCULATE PHILIBERT-LENARD ELECTRON STOPPING POWER
C
      ESP=.25949E+05-EO*(.24640E+04-EO*(.10164E+03-EO*(.20072E+01-EO*
1.15447E-01)))
      PRINT 2004,ESP
2004 FORMAT ( //4X,6HESP = ,E14.8//)
C

```

```

C
C      CHECK FOR FLUORESCENCE AND CALCULATE APPROPRIATE FLUORESCENCE
C      YIELDS AND ABSORPTION JUMP RATIOS
C
C
DO 878 K=1,NEL
N=Z(K)
IF ((LINE(K).EQ.MM(3)).OR.(LINE(K).EQ.MM(4))) GO TO 878
IF (LINE(K).EQ.MM(1)) GO TO 876
AJR(K)=P(N,25)*P(N,10)**P(N,20)/(P(N,26)*P(N,12)**P(N,21))
D=-.111065+Z(K)*(.013680-Z(K)*Z(K)*.217720E-06)
GO TO 879
876 AJR(K)=P(N,24)*P(N,9)**P(N,19)/(P(N,25)*P(N,9)**P(N,20))
D=-.037948+Z(K)*(.034256-Z(K)*Z(K)*.116342E-05)
879 D=D*D*D
FYR(K)=D/(1.+D)
878 CONTINUE
DO 873 K=1,NA
IF ((LINE(K).EQ.MM(3)).OR.(LINE(K).EQ.MM(4))) GO TO 873
DO 852 I=1,NEL
IF ((EDGE(K).LE.WAVE(I)).OR.(LINE(I).EQ.MM(3))) GO TO 852
IF ((EDGE(K).LE.WAVE(I)).OR.(LINE(I).EQ.MM(4))) GO TO 852
IF (LINE(I).EQ.LINE(K)) GO TO 855
IF (LINE(I).EQ.MM(1)) GO TO 859
GO TO 860
855 FF=.5
GO TO 856
859 FF=.21
GO TO 856
860 FF=1.2
856 FY=FYR(I)
RJ=AJR(K)
C
C      CALCULATE FACTOR FOR FLUORESCENCE CORRECTION
C
C
B(K,I)=FF*FY*(RJ-1.)*A(K)*AC(I,K)/(RJ*A(I))*((U(I)-1.)/(U(K)-1.))*
1*1.67
PRINT 2005,EL(I),EL(K),B(K,I)
2005 FORMAT ( 4X,24HFLUORESCENCE FACTOR FOR ,A3,13H FLUORESCING ,A3,3H
1IS ,F9.6 )
852 CONTINUE
873 CONTINUE
DO 901 I=1,8
SAP(I)=0.
SWP(I)=0.
SAPSQ(I)=0.
SWPSQ(I)=0.
SSI(I)=0.
SR(I)=0.
901 SRSQ(I)=0.
SAP(9)=0.
SAPSQ(9)=0.
SWP(9)=0.
SWPSQ(9)=0.
C
C
C      IF COMPOUND STANDARDS ARE USED INSTEAD OF PURE ELEMENTAL
C      STANDARDS, INPUT NECESSARY DATA AND CALCULATE APPROPRIATE
C      PARAMETERS TO CORRECT STANDARD DATA USING THE SAME GENERAL
C      METHODS OUTLINED
C
C
IF (NB.EQ.0) GO TO 53
DO 603 K=1,NB
READ 13, IC,STD(K,1),STD(K,2),CS,EL2

```

```

13  FORMAT (I2,A4,A1,F6.4,A3)
    DO 106 I=1,100
    IF (P(I,2).EQ.EL2) GO TO 109
106  CONTINUE

C
C
C      DETERMINE ATOMIC NUMBER AND ATOMIC WEIGHT OF SECOND ELEMENT IN
C      COMPOUND STANDARD
C
109  Z2=I
    A2=P(I,1)
    NL=3

C

C
C      DO 110 J=1,2
    IF (I.GE.81.AND.NL.GT.3) GO TO 110
    IF (P(I,NL).LT.EO) GO TO 94
110  NL=NL+1
    94  EC2=P(I,NL)
    U2=EO/EC2
    IF (U2.GT.20.) U2=20.
    R2=0.
    KK=34
    DO 112 J=1,6
    R2=U2*R2+P(I,KK)
112  KK=KK-1

C
C
C      CALCULATE SECOND ELEMENT ABSORPTION COEFFICIENT
C
    F=1.0
    DO 302 M=9,18
    IF (WAVE(ID).LT.P(I,M)) GO TO 303
    CON= P(I,28)
302  CONTINUE
    EX= P(I,23)
    GO TO 301
303  INTER=M-8
    GO TO (304,305,306,307,308,309,310,311,312,313),INTER
304  CON= P(I,24)
    EX=P(I,19)
    GO TO 301
305  CON=P(I,25)
    EX=P(I,20)
    GO TO 301
306  F=0.961
    GO TO 305
307  F=0.917
    GO TO 305
308  CON=P(I,26)
    EX=P(I,21)
    GO TO 301
309  F=0.984
    GO TO 308
310  F=0.972
    GO TO 308
311  F=0.946
    GO TO 308
312  F=0.894
    GO TO 308
313  CON=P(I,27)
    EX=P(I,22)
301  AC2=(CON*WAVE(ID)**EX)**F
    C2=1.-CS

```

```
C
C
C
C
WAB=CS*MS(ID)+C2*Z2/A2
CALCULATE OTHER PARAMETERS FOR COMPOUND STANDARD
ZAB=CS*Z(ID)+C2*Z2
PFAB=ZAB*(14.*(1.-EXP(-.1*ZAB))+75.5/(ZAB**((ZAB/7.5))-ZAB/(100.+
1 ZAB))
RAB=CS*R(ID)+C2*R2
ACAB=CS*AC(ID,ID)+C2*AC2
UAB=1.+CSC*ACAB/SIGMA(ID)
HAB=4.5/(ZAB+WAB)
FAB=((1.+HAB)*(1.+1.1*HAB/(4.+1.1*HAB)*CSC*ACAB/SIGMA(ID)))/(UAB*(
11.+HAB*UAB))
ALPHA(ID)=CS*FAB*RAB/(FACT(ID)*WAB*(VAL(ID)-ALOG(PFAB)))
603 ELO(K)=EL(ID)
CHECK FOR TYPE OF DATA (FIXED-TIME OR FIXED-COUNT) AND
ESTABLISH DEAD-TIME CORRECTION
53 IF (TAG.EQ.ABET(26)) GO TO 331
IF (TIME.EQ.0.) TIME=1.
DO 10 J=1,NA
IF (TAUM(J).EQ.0.) TAUM(J)=1.
10 TAU(J)=1.CE-06*TAUM(J)
READ BEAM CURRENTS AND STANDARD INTENSITIES
DO 5005 J=1,NA
BCSUM(J)=0.
STDSUM(J)=0.
READ 5002,N
5002 FORMAT (I2)
NN=N
DO 5003 K=1,N
READ 5004,BC(J,K),STD(I(J,K)
5004 FORMAT(4(2X,F5.0,4X))
IF (BC(J,K).EQ.0.) BC(J,K)=1.
BCSUM(J)=BCSUM(J)+BC(J,K)
DEADTIME CORRECT STANDARD INTENSITIES AND AVERAGE DRIFT
CORRECTED INTENSITIES
STDI(J,K)=STDI(J,K)*TIME/(TIME-TAU(J))*STDI(J,K))
5003 STDSUM(J)=STDSUM(J)+STDI(J,K)
BCAVG(J)=BCSUM(J)/NN
IF (J.EQ.1) BCREP=BCAVG(1)
SI(J)=STDSUM(J)/NN*BCREP/BCAVG(J)
5005 ASI(J)=SI(J)
PRINT 2008
2008 FORMAT (/15X,53HSTANDARD INTENSITIES CORRECTED FOR DEADTIME AND C
1 RIFT)
PRINT 2009,(EL(J),J=1,NA)
2009 FORMAT (/4X,12HBEAM CURRENT,9(5X,A3,3X))
PRINT 2010,BCREF,(SI(J),J=1,NA)
2010 FORMAT (/7X,F8.1,3X,9(F9.2,2X))
DO 5008 J=1,8
DO 5008 K=1,8
5008 BKSUM(J,K)=0.
```

```

C
C      READ BEAM CURRENTS AND BACKGROUNDS
C
DO 5009 J=1,NA
BCSUM(J)=0.
READ 5002,N
NN=N
DO 5010 K=1,N
READ 5011,BC(J,K),(BK(J,K,L),L=1,NA)
5011 FORMAT(4(2X,F5.0,4X))
IF (BC(J,K).EQ.0.) BC(J,K)=1.
BCSUM(J)=BCSUM(J)+BC(J,K)
C
C      DEADTIME CORRECT BACKGROUNDS AND AVERAGE DRIFT-CORRECTED
C      BACKGROUNDS
C
DO 5010 L=1,NA
BK(J,K,L)=BK(J,K,L)*TIME/(TIME-TAU(L)+BK(J,K,L))
5010 BKSUM(J,L)=BKSUM(J,L)+BK(J,K,L)
BCAVG(J)=BCSUM(J)/NN
DO 5014 L=1,NA
5014 ABK(J,L)=BKSUM(J,L)/NN*BCREF/BCAVG(J)
BKGD(J)=ABK(J,J)
5009 NAB(J)=ABK(J,J)+.5
PRINT 5100, TIME
IF (NA.LE.3) GO TO 2080
PRINT 5201, (EL(I), LINE(I), I=1,NA)
PRINT 5202
DO 2070 J=1,NA
PRINT 5204, EL(J), (ABK(J,L), L=1,NA)
2070 CONTINUE
GO TO 331
2080 PRINT 5101, (EL(I), LINE(I), I=1,NA)
PRINT 5102
DO 2085 J=1,NA
PRINT 5104, EL(J), (ABK(J,L),L=1,NA)
2085 CONTINUE
C
C      READ BEAM CURRENTS AND RAW INTENSITIES FROM UNKNOWN, AND
C      COMPUTE CONCENTRATIONS
C
331 SWSAL=0.
READ 2086,DUMMY,(DECAY(J),J=1,NA)
2086 FORMAT(4(2X,F5.0,4X))
DO 40 I=1,300
ITER=0
READ 5015, BEAM(I),(XI(J),J=1,NA)
5015 FORMAT(4(2X,F5.0,4X))
IF (BEAM(I).LE. 99997.) GO TO 5016
KODE=BEAM(I)- 99997.
GO TO 78
5016 IF (TAG.NE.ABET(26)) GO TO 5017
DO 5018 J=1,NA
5018 SI(J)=1.
GO TO 5019
5017 IF (BEAM(I).EQ.0.) BEAM(I)=1.
5019 DO 75 J=1,NA
IF (TAG.NE.ABET(26)) GO TO 5020
IF (ITER.NE.0) GO TO 22
GO TO 333
5020 IF (ITER.NE.0) GO TO 5023
NSI(J,I)=ASI(J)

```

```

C
C
C      DEADTIME AND DRIFT CORRECT UNKNOWN INTENSITIES
C
C

```

```

XXI(J)=XI(J)*TIME/(TIME-TAU(J)*XI(J))
XXI(J)=XXI(J)-DECAY(J)
XXI(J)=XXI(J)+BCREF/BEAM(I)
NXI(J,I)=XXI(J)

```

```

C
C
C      CORRECT FOR STANDARD BACKGROUNDS
C
C

```

```

XI(J)=XXI(J)-BKGD(J)
IF (I.EQ.1) SI(J)=SI(J)-BKGD(J)
SSI(J)=SSI(J)+SI(J)
GO TO 333

```

```

5023 ABGD(J)=0.

```

```

C
C
C      CORRECT FOR CONCENTRATION DEPENDENT BACKGROUNDS
C
C

```

```

DO 5024 K=1,NA
5024 ABGD(J)=ABGD(J)+C(K)*ABK(K,J)
XI(J)=XXI(J)-ABGD(J)

```

```

C
C
C      FORM K-RATIOS
C
C

```

```

333 RINT(J)=XI(J)/SI(J)
IF (RINT(J).GT..9999) RINT(J)=.9999
IF (RINT(J).LT.0.) RINT(J)=0.
IF (ITER.EQ.2) ROUT(J,I)=RINT(J)+.0001

```

```

75 CONTINUE

```

```

IF (ITER.NE.2) GO TO 5030

```

```

DO 41 J=1,NA
SR(J)=SR(J)+RINT(J)

```

```

41 SRSQ(J)=SRSQ(J)+RINT(J)*RINT(J)

```

```

5030 DO 5040 J=1,NA
RINT(J)=ALPHA(J)*RINT(J)
IF (ITER.EQ.0) C(J)=RINT(J)

```

```

5040 CONTINUE

```

```

22 ITER=ITER+1

```

```

IF (ITER.NE.1) GO TO 5060

```

```

DO 5050 J=1,NA

```

```

CO(J)=RINT(J)

```

```

5050 CONTINUE

```

```

GO TO 5070

```

```

5060 IF (ITER.LE.2) GO TO 5070

```

```

DO 5080 J=1,NA

```

```

5080 CO(J)=C1(J)

```

```

5070 IF (NEL.EQ.NA) GO TO 25

```

```

C(NEL)=1.

```

```

DO 24 J=1,NA

```

```

24 C(NEL)=C(NEL)-C(J)

```

```

IF (C(NEL).LT.0.) C(NEL)=0.0

```

```

25 RAL=0.

```

```

ZAL=0.

```

```

WSAL=0.

```

```

DO 902 J=1,8

```

```

FIDI(J)=0.

```

```

902 ACAL(J)=0.

```

```

C
C

```

```

C      CALCULATE ABSORPTION COEFFICIENTS FOR UNKNOWN
C
C
DO 26 K=1,NEL
DO 26 J=1,NEL
ACAL(K)=ACAL(K)+C(J)*AC(K,J)
IF (ACAL(K).LE.0.) ACAL(K)=.0000001
IF (ACAL(K).GT.100000.) ACAL(K)=100000.0
26 CONTINUE

C
C
C      CALCULATE BACKSCATTER LOSS FACTOR, MEAN ATOMIC NUMBER AND MEAN
C      Z / A FOR UNKNOWN
C
C
DO 27 J=1,NEL
RAL=RAL+R(J)*C(J)
ZAL=ZAL+Z(J)*C(J)
27 WSAL=WSAL+WS(J)*C(J)
HAL=4.5/(ZAL*WSAL)
PAL=ZAL*(14.*(1.-EXP(-.1*ZAL))+75.5/(ZAL*(ZAL/7.5))-ZAL/(100.+ZAL
1))

C
C
C      CALCULATE CONCENTRATIONS
C
C
DO 28 K=1,NA
UAL(K)=1.+CSC*ACAL(K)/SIGMA(K)
YP=CSC*ACAL(K)
DO 880 J=1,NEL
Y=YP/ACAL(J)
V=ESP/ACAL(J)
880 FIDI(K)=FIDI(K)+C(J)*B(K,J)/ACAL(J)*(ALOG(1.+Y)/Y+ALOG(1.+V)/V)
FAL(K)=((1.+HAL)*(1.+FIDI(K))*(1.+1.1*HAL/(4.+1.1*HAL)*CSC*ACAL(K)
1/SIGMA(K)))/(UAL(K)*(1.+HAL*UAL(K)*
C(K)=RINT(K)*WSAL*(VAL(K)-ALOG(PAL))*FACT(K)/(FAL(K)*RAL)
IF (C(K).LT.0.) C(K)=0.0
IF (C(K).GT..9999) C(K)=.9999
IF (ITER.NE.1) GO TO 5090
C1(K)=C(K)
GO TO 28
5090 IF (ITER.GT.2) C1(K)=C3(K)
C3(K)=C(K)
CDEN=CD(K)-2.*C1(K)+C3(K)
IF (CDEN.LE.0.) CDEN=.0000001
C(K)=(CD(K)*C3(K)-C1(K)*C1(K))/CDEN
IF (C(K).GT..9999) C(K)=.9999
IF (C(K).LE.0.) C(K)=C3(K)
28 CONTINUE

C
C
C      TEST FOR CONVERGENCE AND CHECK NUMBER OF ITERATIONS
C
C
IF (ITER.LT.3) GO TO 5019
DO 85 M=1,NA
DEL=ABS(C3(M)-C(M))
IF (DEL.GE..0005) GO TO 86
85 CONTINUE
GO TO 87
86 IF (ITER.LT.20) GO TO 5019
87 LOOP(I)=ITER

C
C
C      SUM CONCENTRATIONS FOR AVERAGING
C

```



```

SWSAL=SWSAL+WSAL
DO 31 K=1,NA
  IF (SR(K).LT..0001) SR(K)=.0001
  IF (SRSQ(K).LT..00000001) SRSQ(K)=.00000001
  WP(K)=100.*C(K)
31  WPC(K,I)=WP(K)
  IF (NEL.EQ.NA) GO TO 34
  WP(NEL)=100.
  DO 33 K=1,NA
33  WP(NEL)=WP(NEL)-WP(K)
  IF (WP(NEL).LT.0.) WP(NEL)=0.
34  WPC(NEL,I)=WP(NEL)
  DO 35 K=1,NEL
35  ATOM(K)=WP(K)/A(K)
  DEN=0.
  DO 36 K=1,NEL
36  DEN=DEN+ATOM(K)
  DO 40 K=1,NEL
  AP(K)=100.*ATOM(K)/DEN
  APC(K,I)=AP(K)
  SAP(K)=SAP(K)+AP(K)
  SAPSQ(K)=SAPSQ(K)+AP(K)*AP(K)
  SWP(K)=SWP(K)+WP(K)
  SWPSQ(K)=SWPSQ(K)+WP(K)*WP(K)
40  CONTINUE
78  NS=I-1
  OBS=NS
  IF (NS.GT.1) GO TO 760
  DIV=1.
  GO TO 765
760 DIV=OBS*OBS

      CALCULATE AVERAGE K-RATIOS AND RMS DEVIATIONS

765 DO 42 K=1,NA
  AVR(K)=SR(K)/OBS+.00005
42  RDEV(K)=2.*SQRT((OBS*SRSQ(K)-SR(K)*SR(K))/DIV)+.00005

      CALCULATE PEAK-TO-BACKGROUND RATIOS AND MINIMUM DETECTABILITY
      LIMITS

  IF (TAG.EQ.ABET(26)) GO TO 334
  DO 730 K=1,NA
  PEAK(K)=SI(K)/(ALPHA(K)*BKGD(K))
  MPEAK(K)=PEAK(K)
  LIMIT(K)=329./SQRT(SI(K)*PEAK(K))
730 CONTINUE

      CALCULATE AVERAGE CONCENTRATIONS

334 DO 43 K=1,NEL
  AVWP(K)=SWP(K)/OBS+.005
  AVAP(K)=SAP(K)/OBS+.005
  WPDEV(K)=2.*SQRT((OBS*SWPSQ(K)-SWP(K)*SWP(K))/DIV)+.005
43  APDEV(K)=2.*SQRT((OBS*SAPSQ(K)-SAP(K)*SAP(K))/DIV)+.005
  IF (NEL.EQ.NA) GO TO 76
  AVWP(NEL)=100.0
  AVAP(NEL)=100.0
  DO 77 K=1,NA
  AVWP(NEL)=AVWP(NEL)-AVWP(K)
77  AVAP(NEL)=AVAP(NEL)-AVAP(K)

```

```

IF (AVWP(NEL).GE.0.) GO TO 76
AVWP(NEL)=0.
AVAP(NEL)=0.

```

```

C
C
C      START OUTPUT OF ALL DATA INCLUDING CONSTANTS CALCULATED
C

```

```

76 PRINT 615,TAG,NPROB,DATE
615 FORMAT (1H1,///40X,15HPROBLEM NUMBER ,A1,I4//68X,3A4,A3//)
PRINT 29
29 FORMAT (/17X,6HATOMIC,5X,6HATOMIC,5X,11HBACKSCATTER,5X, 10HEXCIT AT
1ION,5X,10HABSORPTION,5X,11HFLUORESCENT/5X,7HELEMENT,5X,6HNUMBER,5X
2,6HWEIGHT,7X,6HFACTOR,8X,9HPOTENTIAL,6X,10HJUMP RATIO,8X,5HYIELD //
3)
DO 39 I=1,NEL
NZ=Z(I)
AJR(I)=AJR(I)+.005
FYR(I)=FYR(I)+.0005
PRINT 38,EL(I),NZ,A(I),R(I),EC(I),AJR(I),FYR(I)
38 FORMAT (7X,A3,8X,I3,5X,F8.3,7X,F5.3,10X,F7.3,8X,F6.2,11X,F5.3)
39 CONTINUE
PRINT 52
52 FORMAT (///33X,28HMASS ABSORPTION COEFFICIENTS)
IF (NA.LE.3) GO TO 637
PRINT 648,(EL(I),LINE(I),I=1,NA)
648 FORMAT (/3X,9HRADIATION,5X,8(A3,A2,6X))
PRINT 646
646 FORMAT (/4X,8HABSORBER//)
DO 658 J=1,NEL
PRINT 668,EL(J),(AC(I,J),I=1,NA)
668 FORMAT (9X,A3,3X,8(F8.3,4X))
658 CONTINUE
PRINT 5100, TIME
PRINT 5201,(EL(I),LINE(I),I=1,NA)
5201 FORMAT (/41X,11HBACKGROUNDS//1X,14HCONTRIBUTED TO,2X,8(A3,A2,6X))
PRINT 5202
5202 FORMAT (/2X,11HBY 100 % OF//)
DO 5203 J=1,NA
5204 FORMAT (9X,A3,3X,8(F6.1,4X))
5203 PRINT 5204,EL(J),(ABK(J,L),L=1,NA)
IF (NS.LE.1) GO TO 73
PRINT 615,TAG,NPROB,DATE
IF (TIME.EQ.1.) GO TO 626
PRINT 617,(EL(I),I=1,NA)
617 FORMAT (///33X,37HINTENSITIES (CPS X COUNTING INTERVAL)/35X,33HCOR
1RECTED FOR DEAD-TIME AND DRIFT//6X,8(A3,10X)//)
GO TO 610
626 PRINT 627,(EL(I),I=1,NA)
627 FORMAT (///43X,17HINTENSITIES (CPS)/35X,33HCORRECTED FOR DEAD-TIME
1 AND DRIFT//6X,8(A3,10X)//)
610 DO 688 I=1,NS
PRINT 698,(NXI(J,I),NSI(J,I),J=1,NA)
698 FORMAT (11X,8(I7,2H /,I7,1X))
688 CONTINUE
IF (TIME.EQ.1.) GO TO 619
PRINT 620,(NAB(I),I=1,NA)
620 FORMAT (/29X,46HSTANDARD BACKGROUNDS (ICPS X COUNTING INTERVAL)//15X
1,8(I4,9X))
GO TO 621
619 PRINT 622,(NAB(I),I=1,NA)
622 FORMAT (///39X,26HSTANDARD BACKGROUNDS (CPS)//5X,8(I4,9X))
621 PRINT 623,(TAUM(I),I=1,NA)
623 FORMAT (///40X,24HDEAD-TIME (MICROSECONDS)//6X,8(F3.1,10X))
IF (TIME.EQ.1.) GO TO 625
PRINT 695,TIME
695 FORMAT (///36X,18HCOUNTING INTERVAL,,F6.1,8H SECONDS)
625 PRINT 15,TAG,NPROB,DATE

```

```

      PRINT 694,(EL(I),I=1,NA)
694 FORMAT (18X,33HINDIVIDUAL K-RATIOS CORRECTED FOR/17X, 36HDEAD-TIME
1, DRIFT AND BACKGROUND ONLY//5X,1HN,4X,8(A3,5X))
      DO 568 I=1,NS
      PRINT 578,LOOP(I),(ROUT(J,I),J=1,NA)
578 FORMAT (3X,I3,2X,8(F6.4,2X))
568 CONTINUE
      IF (NB.EQ.0) GO TO 1008
      PRINT 74
      PRINT 674,(ELO(K),STD(K,1),STD(K,2),K=1,NB)
674 FORMAT (/17X,A3,37HDETERMINED RELATIVE TO A STANDARD OF ,A4,A1)
      GO TO 1008
637 PRINT 647,(EL(I),LINE(I),I=1,NA)
647 FORMAT (/25X,9HRADIATION,5X,3(A3,A2,8X))
      PRINT 645
645 FORMAT (/26X,8HABSORBER//)
      DO 657 J=1,NEL
      PRINT 667,EL(J),(AC(I,J),I=1,NA)
667 FORMAT (31X,A3,3X,3(F8.0,6X))
657 CONTINUE
      PRINT 5100, TIME
5100 FORMAT (///36X,18HCOUNTING INTERVAL,,F6.1,10H SECOND(S))
      PRINT 5101,(EL(I),LINE(I),I=1,NA)
5101 FORMAT (/41X,11HBACKGROUNDS//19X,14 HCONTRIBUTED TO,5X,3(A3,A2,8X
1))
      PRINT 5102
5102 FORMAT (/23X,11HBY 100 % OF//)
      DO 5103 J=1,NA
5104 FORMAT (31X,A3,3X,3(F6.1,6X))
5103 PRINT 5104,EL(J),(ABK(J,L),L=1,NA)
      IF (NS.LE.1) GO TO 73
      PRINT 15,TAG,NPROB,DATE
      15 FORMAT (11F1,24X,15HPROBLEM NUMBER ,A1,I4//52X,3A4,A3//)
      IF (TIME.EQ.1.) GO TO 643
      PRINT 655,(EL(I),I=1,NA)
655 FORMAT (///16X,37HINTENSITIES (CPS X COUNTING INTERVAL)//18X,33HCOR
1RECTED FOR DEAD-TIME AND DRIFT//14X,3(A3,16X)//)
      GO TO 642
643 PRINT 654,(EL(I),I=1,NA)
654 FORMAT (///26X,17HINTENSITIES (CPS)//18X,33HCORRECTED FOR DEAD-TIME
1 AND DRIFT//14X,3(A3,16X)//)
642 DO 687 I=1,NS
      PRINT 697,(NXI(J,I),NSI(J,I),J=1,NA)
697 FORMAT (2X,3(7X,I7,2H/,I7))
687 CONTINUE
      IF (TIME.EQ.1.) GO TO 652
      PRINT 651,(NAB(I),I=1,NA)
651 FORMAT (/12X,46HSTANDARD BACKGROUNDS (CPS X COUNTING INTERVAL)//1
13X,3(I4,15X))
      GO TO 665
652 PRINT 537,(NAB(I),I=1,NA)
537 FORMAT (/22X,26HSTANDARD BACKGROUNDS (CPS)//13X,3(I4,15X))
665 PRINT 664,(TAUM(I),I=1,NA)
664 FORMAT (/23X,24HDEAD-TIME (MICROSECONDS)//14X,3(F3.1,17X))
      IF (TIME.EQ.1.) GO TO 663
      PRINT 662,TIME
662 FORMAT (/19X,18HCOUNTING INTERVAL,,F6.1,8H SECONDS)
663 PRINT 15,TAG,NPROB,DATE
      PRINT 70,(EL(J),J=1,NA)
      70 FORMAT (18X,33HINDIVIDUAL K-RATIOS CORRECTED FOR/17X, 36HDEAD-TIME
1, DRIFT AND BACKGROUND ONLY//9X,1HN,7X,3(A3,13X))
      DO 702 I=1,NS
      PRINT 71,LOOP(I),(ROUT(J,I),J=1,NA)
71 FORMAT (7X,I3,5X,3(F6.4,10X))
702 CONTINUE
      IF (NB.EQ.0) GO TO 1008
      PRINT 74

```

```

74 FORMAT (//30X,10H***NOTE***)
   PRINT 51,(ELO(K),STD(K,1),STD(K,2X,K=1,NB)
51 FORMAT (//24X,A3,19HDETERMINED RELATIVE/24X,17HTO A STANDARD OF ,
1A4,A1)
1008 IF (NS.LE.1) GO TO 73
      DO 1270 L=1,2
      IF (NA.GT.3) GO TO 1200
      PRINT 15,TAG,NPROB,DATE
      GO TO 1110
1200 PRINT 615,TAG,NPROB,DATE
1110 PRINT 89,NAME,TITLE
      PRINT 66,EO,THETA
      IF (TAG.EQ.ABET(26)) GO TO 1130
      PRINT 67
      DO 1130 K=1,NA
      PRINT 68,EL(K),MPEAK(K),LIMIT(K)
1130 CONTINUE
      IF (NA.GT.3) GO TO 1240
      PRINT 1150
1150 FORMAT (//23X,20HCHEMICAL COMPOSITION/26X,14HWEIGHT PERCENT/26X,14
1HATOMIC PERCENT/)
      IF (NEL.EQ.NA) GO TO 1152
      PRINT 1153,EL(NEL)
1153 FORMAT (19X,A3,24HDETERMINED BY DIFFERENCE/)
1152 PRINT 1154,(EL(K),K=1,NEL)
1154 FORMAT (/3X,3HOBS,9X,4(A3,13X)/)
      DO 1170 I=1,NS
      PRINT 1160,I,(WPC(K,I),K=1,NEL)
1160 FORMAT (/3X,13.5X,4(F8.3,8X))
      PRINT 1165,(APC(K,I),K=1,NEL)
1165 FORMAT (11X,4(F8.3,8X))
1170 CONTINUE
      GO TO 1270
1240 PRINT 1250
1250 FORMAT (//46X,20HCHEMICAL COMPOSITION/49X,14HWEIGHT PERCENT/49X,14
1HATOMIC PERCENT/)
      IF (NEL.EQ.NA) GO TO 1252
      PRINT 1253,EL(NEL)
1253 FORMAT (42X,A3,24HDETERMINED BY DIFFERENCE/)
1252 PRINT 1254,(EL(K),K=1,NEL)
1254 FORMAT (13X,3HOBS,7X,9(A3,8X)/)
      DO 1270 I=1,NS
      PRINT 1260,I,(WPC(K,I),K=1,NEL)
1260 FORMAT (/3X,13.4X,9(F8.3,4X))
      PRINT 1265,(APC(K,I),K=1,NEL)
1265 FORMAT (10X,9(F8.3,4X))
1270 CONTINUE
73 DO 69 L=1,2
      PRINT 15,TAG,NPROB,DATE
      PRINT 89,NAME,TITLE
89 FORMAT(5X,13HSUBMITTED BY ,5A4//5X,14HDESCRIPTION - ,71A1//)
      PRINT 49,NS
49 FORMAT (/18X,33HMEAN CHEMICAL COMPOSITION AND TWO/18X, 21HSIGMA LI
1MITS BASED ON,14,9H ANALYSES,///29X,6HWEIGHT,12X,6HATOMIC/16X,7HEL
2EMENT,6X,7HPERCENT,11X,7HPERCENT//)
      DO 45 K=1,NA
      PRINT 44,EL(K),AVWP(K),WPDEV(K),AVAP(K),APDEV(K)
44 FORMAT (18X,A3,2X,F8.3,2H -,F6.3,2X,F8.3,2H -,F6.3//)
45 CONTINUE
      IF (NEL.EQ.NA) GO TO 47
      PRINT 48,EL(NEL),AVWP(NEL),WPDEV(NEL),AVAP(NEL),APDEV(NEL)
48 FORMAT (18X,A3,1H*,1X,F8.3,2H -,F6.3,2X,F8.3,2H -,F6.3//21X, 26 H*
1 DETERMINED BY DIFFERENCE)
47 PRINT 64
64 FORMAT (//5X,65H-----
1-----//13X,42HMEAN INTENSITY RATIOS AND TWO SIGMA LIIMI
2TS//21X,7HELEMENT, 12X,1HK//)

```

```

DO 63 K=1,NA
PRINT 65,EL(K),AVR(K),RDEV(K)
65 FORMAT (23X,A3,7X,F6.4,2H -,F7.4)
63 CONTINUE
PRINT 66,E0,THETA
66 FORMAT (///13X,20HACCELERATING VOLTAGE,13X,F5.1,4H KEV//13X,21H X-
1RAY EMERGENCE ANGLE,12X,F5.1,8H DEGREES)
IF (T1.EQ.0.) GO TO 83
PRINT 84,T1
84 FORMAT (/13X,14HFILM THICKNESS,19X,F6.2,8H MICRONS)

C
C
C      CALCULATE DEPTH OF ANALYZED REGION
C
C
83 IF (RHO.EQ.0.) GO TO 92
DM=.033*OBS/(SWSAL*RHO)*SQRT(E0*E0*E0)+.005

C
C
C      CONTINUE OUTPUTTING OF DATA
C
C
PRINT 93,RHO,DM
93 FORMAT (/13X,7HDENSITY,26X,F6.2//13X,24HDEPTH OF ANALYZED REGION,9
1X,F6.2,8H MICRONS)
92 IF (TAG.EC.ABET(26)) GO TO 69
IF (NA.GT.3) PRINT 15,TAG,NPROB,DATE
747 PRINT 67
67 FORMAT (//13X,44HSTANDARD PEAK-TO-BACKGROUND RATIOS (P/B) AND/17X,
134HMINIMUM DETECTABILITY LIMITS (MDL)//16X,7HELEMENT,6X,3HP/B,13X,
23HMDL//)
DO 341 K=1,NA
PRINT 68,EL(K),MPEAK(K),LIMIT(K)
68 FORMAT (18X,A3,5X,I5,2H/1,8X,F7.4,5H MT Z//)
341 CONTINUE
69 CONTINUE

C
C
C      DETERMINE IF GRAPHS ARE DESIRED
C
C
READ 60,(MARY(I),I=1,5)
60 FORMAT(5I1)
DO 61 I=1,5
IF(MARY(I)-1) 61,61,62
62 CALL GRAPHS
GO TO 320
61 CONTINUE

C
C
C      CHECK FOR MORE DATA.
C
C
320 GO TO (1,80), KODE
80 PRINT 1002,TAG,NPROB
1002 FORMAT (1H0,//10X,27HLAST PROBLEM NUMBER USED IS,1X,A1,I4)
STOP
END
SUBROUTINE GRAPHS
C      THIS SUBROUTINE DETERMINES WHICH GRAPHING OPTIONS ARE DESIRED.
C      RADII, AND CORRECT WEIGHT PERCENTS AND WEIGHT PERCENT RATIOS ARE
C      CALCULATED AND STORED IN COMMON.
C      PROGRAMMED BY RICK FERNOW(JULY 1969)
COMMON WPC(9,300),MARY(5),RAD(102), AAA(400), Y(500)
MU5=MARY(5)
DO 35 J=1,100
READ 5,RAD(J)
5 FORMAT(F10.0)
IF(RAD(J).LT.0.001) GO TO 12
35 CONTINUE

```

```

12 NPTS=J-1
   DO 1 I=1,NPTS
     Y(I)=WPC(1,I)
     Y(NPTS+I)=WPC(2,I)
     Y(2*NPTS+I)=WPC(1,I)+WPC(2,I)
     Y(3*NPTS+I)=      WPC(2,I) / WPC(1,I)
1   Y(4*NPTS+I)=      WPC(2,I) / (WPC(1,I)+WPC(2,I))
     JIM=1
     LET=1
     DO 31 K=1,5
       IF(MARY(K).NE.5) GO TO 10
       CALL CALPLT(NPTS,7,1)
       LET=2
10  IF(K.NE.5)GO TO 2
     IF(MU5.EQ.3.OR.MU5.EQ.4)GO TO 2
     IF(LET.EQ.1)GO TO 2
     CALL CALPLT(NPTS,6,2)
     2  IF(MARY(K).EQ.0) MARY(K)=1
       IF(MARY(1).EQ.5) MARY(1)=4
       MU=MARY(K)
       GO TO(31,32,33,32),MU
32  IF(JIM.EQ.2) GO TO 34
     CALL PLOT(NPTS,JIM)
34  IF(MARY(K).NE.4) GO TO 31
33  CALL CALPLT(NPTS,K,LET)
     LET=2
     IF(K.NE.5)GO TO 31
     CALL CALPLT(NPTS,6,2)
31  CONTINUE
     RETURN
     END

```

```

C      SUBROUTINE CALPLT(NPTS,K,LET)
C      THIS SUBROUTINE PREPARES CALCOMP GRAPHS
C      PROGRAMMED BY RICK FERNDW (JULY 1969)
COMMON AAA(9,300),BBB(5),RAD(102), CCC(400),Y(500), NAME(40)
DIMENSION BUFF(1000),ORD(102),ORD1(102),ORD2(102),ORD3(102)
IF(K.EQ.6)GO TO 9
IF(LET.EQ.2) GO TO 7
CALL PLOTS(BUFF,4000,11)
7  CALL PLOT(-.75,-.5,3)
   CALL PLOT(7.75,-.5,2)
   CALL PLOT(7.75,10.5,2)
   CALL PLOT(-.75,10.5,2)
   CALL PLOT(-.75,-.5,2)
   GO TO(1,2,3,4,5, 9,15),K
1  JO=10
   KO=15
   DO 8 I=1,NPTS
     ORD(I)=Y(I)
     GO TO 6
2  JO=14
   KO=16
   DO 58 I=1,NPTS
     ORD(I)=Y(I+NPTS)
     GO TO 6
3  JO=18
   KO=18
   DO 60 I=1,NPTS
     ORD(I)=Y(2*NPTS+I)
     GO TO 6
4  JO=23
   KO=22
   DO 64 I=1,NPTS
     ORD(I)=Y(3*NPTS+I)

```

```

      GO TO 6
5   JO=29
      KO=27
      DO 73 I=1,NPTS
73  ORD(I)=Y(4*NPTS+I)
      GO TO 6
15  DO 17 I=1,NPTS
      ORD1(I)=Y(I)
      ORD2(I)=Y(NPTS+I)
17  ORD3(I)=Y(2*NPTS+I)
      ORD1(NPTS+1)=0.00
      ORD1(NPTS+2)=10.00
      ORD2(NPTS+1)=0.00
      ORD2(NPTS+2)=10.00
      ORD3(NPTS+1)=0.00
      ORD3(NPTS+2)=10.00
      CALL SCALE(RAD,7.5,NPTS,1,10.)
      CALL AXIS(0.0,0.0,NAME(10),10,10.0,90.0,0.00,10.00,10.)
      CALL AXIS(0.0,0.0,NAME(1),-33,7.5,0.0,RAD(NPTS+1),RAD(NPTS+2),10.)
      CALL LINE(RAD,ORD1,NPTS,1,-1,78)
      CALL LINE(RAD,ORD2,NPTS,1,-1,92)
      CALL LINE(RAD,ORD3,NPTS,1,-1,04)
      CALL PLOT(10.0,0.0,-3)
      RETURN
6   IF(K.NE.3) GO TO 10
      ORD(NPTS+1)= 0.00
      ORD(NPTS+2)= 10.00
      GO TO 11
10  CALL SCALE(ORD,10.0,NPTS,1,10.)
11  CALL SCALE(RAD,7.5,NPTS,1,10.)
      CALL AXIS(0.0,0.0,NAME(JO),KO,10.0,90.0,ORD(NPTS+1),ORD(NPTS+2),10.)
      CALL AXIS(0.0,0.0,NAME(1),-33,7.5,0.0,RAD(NPTS+1),RAD(NPTS+2),10.)
      CALL LINE(RAD,ORD,NPTS,1,-1,78)
      CALL PLOT(10.0,0.0,-3)
      RETURN
9   CALL PLOT(0.,0.,999)
      RETURN
END
SUBROUTINE PLOT(NPTS,JIM)
C   THIS SUBROUTINE PLOTS WEIGHT PERCENTS AND WEIGHT PERCENT RATIOS VERSUS
C   RADIAL DISTANCE IN THE SAMPLE. IT WILL GIVE A ROUGH GRAPH ON THE PRINTER.
C   PROGRAMMED BY RICK FERNOW(JULY 1969)
      COMMON AAA(9,300),MARY(5),RAD(102),TITLE(400),Y(500)
      DIMENSION ORD(100),GRAPH(55,130)
      INTEGER HORZ(100),VERT(100)
      MU1=MARY(1)
      MU2=MARY(2)
      MU3=MARY(3)
      MU4=MARY(4)
      MU5=MARY(5)
C   CALCULATE THE ARRAY OF HORIZONTAL VALUES
      DO 30 J=1,NPTS
30  HORZ(J)=IFIX(0.80*RAD(J)+14.0)
      DO 61 L=1,5
        IF(L.EQ.1) GO TO 35
        IF(L.NE.4) GO TO 17
C   CLEAR THE GRAPH ARRAY
35  DO 29 I=1,130
      DO 29 J=1,55
29  GRAPH(J,I)=TITLE(7)
C   CONSTRUCT THE AXES
      DO 1 I=1,50
1   GRAPH(I,13)=TITLE(30)
      DO 2 J=13,122
2   GRAPH(50,J)=TITLE(34)
C   INSERT MARKERS AT THE PROPER PLACES

```

```

DO 3 I=1,50,5
3 GRAPH(I,13)=TITLE(82)
DO 4 J=13,117,8
4 GRAPH(50,J)=TITLE(82)
C NUMBER THE AXES
DO 5 I=1,46
GRAPH(I,12)=TITLE(I+135)
5 GRAPH(I,11)=TITLE(I+190)
GRAPH(1,10)=TITLE(236)
GO TO(50,50,50,51,51),L
51 DO 52 I=6,46,5
52 GRAPH(I,10)=TITLE(239)
GRAPH(1,11)=TITLE(239)
50 DO 6 J=12,120
6 GRAPH(51,J)=TITLE(J+229)
C LABEL THE AXES AND DETERMINE THE CORRECT SET OF Y VALUES
DO 7 J=43,75
7 GRAPH(53,J)=TITLE(J-42)
17 IF(MARY(L).EQ.0) MARY(L)=1
MU=MARY(L)
GO TO(20,20,16,20,19),L
16 IF(MU3.EQ.2) GO TO 20
IF(MU3.EQ.4) GO TO 20
GO TO(21,18,21,18),MU1
21 GO TO(61,18,61,18),MU2
19 IF(MU5.EQ.2) GO TO 20
IF(MU5.EQ.4) GO TO 20
GO TO(61,18,61,18),MU4
20 GO TO(61,44,61,44),MU /
44 GO TO(45,46,47,48,49),L
45 DO 15 I=1,NPTS
15 ORD(I)=Y(I)
IP=49
GO TO 40
46 DO 58 I=1,NPTS
58 ORD(I)=Y(I+NPTS)
IP=38
GO TO 40
47 DO 60 I=1,NPTS
60 ORD(I)=Y(2*NPTS+I)
IP=237
GO TO 40
48 DO 64 I=1,NPTS
64 ORD(I)=Y(3*NPTS+I)*100.
IP=351
DO 54 J=82,89
54 GRAPH(55,J)=TITLE(J+269)
GO TO 9
49 DO 73 I=1,NPTS
73 ORD(I)=Y(4*NPTS+I)*100.
IP=239
DO 55 J=97,108
55 GRAPH(55,J)=TITLE(J+262)
9 DO 10 I=16,31
10 GRAPH(I,5)=TITLE(I+70)
GO TO 41
40 DO 8 I=20,29
8 GRAPH(I,5)=TITLE(I+15)
41 DO 12 J=1,NPTS
VERT(J)=IFIX(51.0-0.50*ORD(J))
IF(VERT(J).EQ.0) VERT(J)=1
IH=HORZ(J)
IV=VERT(J)
IF(GRAPH(IV,IH).NE.TITLE(7)) GO TO 53
GRAPH(IV,IH)=TITLE(IP)
GO TO 12
53 GRAPH(IV,IH)=TITLE(238)

```



```

12 CONTINUE
  IF(L.LT.3) GO TO 61
  IF(L.EQ.4) GO TO 61
C  PRINT OUT ALL THE CHARACTERS IN THE ARRAY
18 PRINT 28
28 FORMAT(1H1,'PRINTER PLOT')
  DO 26 I=1,55
26 PRINT 27,(GRAPH(I,J),J=1,130)
27 FORMAT(1H ,130A1)
61 CONTINUE
  JIM=2
  RETURN
END

```

1.008 H	.014			999.	999.	999.	<i>Constants and mass absorption coefficients stored internally</i>
4.003 HE	.025			499.	999.	999.	
6.939 LI	.055		228.	226.5	999.	999.	
9.012 BE	.111		114.	111.	999.	999.	
10.811 B	.188	.005	67.6	65.6	999.	999.	
12.011 C	.284	.006	44.7	43.68	999.	999.	
14.007 N	.400	.009	31.60	30.99	999.	999.	
15.999 O	.532	.007	23.62	23.32	524.	999.	
18.998 F	.685	.009	18.32	18.09	398.	999.	
20.183 NE	.867	.018	14.610	14.302	275.	677.	
22.990 NA	1.072	.031	11.910	11.569	247.3	398.8	
24.312 MG	1.303	.049	9.890	9.512	197.3	249.3	
26.982 AL	1.560	.073	8.339	7.948	142.5	169.49	
28.086 SI	1.840	.101	7.125	6.738	105.0	123.	
30.974 P	2.144	.132	6.157	5.784	81.0	93.7	
32.064 S	2.470	.165	5.372	5.019	64.1	75.2	
35.453 CL	2.820	.200	4.728	4.397	52.1	61.8	
39.948 AR	3.203	.245	4.192	3.871	43.2	50.2	
39.102 K	3.608	.295	3.741	3.437	36.4	41.8	
40.080 CA	4.038	.346	3.358 36.33	3.070	30.7	35.13	
44.956 SC	4.489	.402 .007	3.031 31.35	2.762	26.8	30.6	
47.900 TI	4.965	.455 .004	2.749 27.42	2.497	23.4	26.94	
50.942 V	5.464	.513 .002	2.504 24.25	2.269	19.72	23.8	
51.996 CR	5.989	.575 .002	2.290 21.64	2.070	17.84	21.24	
54.938 MN	6.538	.640 .003	2.102 19.45	1.896	16.15	19.05	
55.847 FE	7.111	.707 .004	1.936 17.59	1.743	14.65	17.202	
58.933 CO	7.710	.779 .003	1.789 15.972	1.608	13.38	15.618	
58.710 NI	8.332	.854 .004	1.658 14.561	1.488	12.3	14.242	
63.540 CU	8.980	.933 .002	1.541 13.336	1.381	11.27	13.014	
65.370 ZN	9.661	1.022 .008	1.435 12.254	1.283	10.06	11.862	
69.720 GA	10.368	1.117 .017	1.340 11.292	1.196	9.517	10.828	
72.590 GE	11.104	1.217 .029	1.254 10.436	1.117	8.773	9.924	
74.922 AS	11.865	1.323 .041	1.176 9.671	1.045	8.107	9.125	
78.960 SE	12.655	1.434 .054	1.105 8.990	.980	7.503	8.407	
79.909 BR	13.470	1.553 .069	1.040 8.375	.920	6.959	7.753	
83.800 KR	14.324	1.677 .089	.980 7.817	.866	6.47	7.168	
85.470 RB	15.202	1.807 .110	.926 7.348	.816	6.008	6.644	
87.620 SR	16.107	1.941 .133	.875 6.863	.770	5.592	6.173	
88.905 Y	17.038	2.079 .157	.829 6.449	.728	5.217	5.756	
91.220 ZR	17.999	2.223 .180	.786 6.071	.689	4.879	5.378	
92.906 NB	18.987	2.371 .205	.746 5.724	.653	4.575	5.031	
95.940 MD	20.004	2.523 .227	.709 5.407	.620	4.304	4.719	
99.000 TC	21.047	2.678 .253	.675 5.115	.589	4.058	4.436	
101.070 RU	22.119	2.838 .279	.643 4.846	.561	3.835	4.180	
102.905 RH	23.220	3.002 .307	.613 4.597	.534	3.629	3.943	
106.400 PD	24.348	3.173 .335	.585 4.368	.509	3.437	3.723	
107.870 AG	25.517	3.351 .398	.559 4.154	.486	3.256	3.516	
112.400 CD	26.716	3.538 .440	.535 3.956	.464	3.085	3.326	
114.82 IN	27.942	3.730 .443	.512 3.772	.444	2.926	3.147	
118.690 SN	29.195	3.929 .511	.491 3.600	.425	2.777	2.982	
121.750 SB	30.486	4.132 .528	.470 3.489	.407	2.639	2.830	
127.600 TE	31.811	4.342 .572	.451 3.289	.309	2.510	2.688	

126.904 I	33.167	4.559	.631	.433	3.149	.374	2.388	2.554
131.300XE	34.590	4.782	.672	.416	3.017	.358	2.274	2.429
132.905CS	35.987	5.011	.726	.400	2.892	.345	2.167	2.314
137.340BA	37.452	5.247	.780	.385	2.776	.331	2.068	2.205
138.910LA	38.934	5.484	.832	.371	2.666	14.88	.318	1.978
140.120CE	40.453	5.723	.883	.357	2.562	14.04	.306	1.893
140.907PR	42.002	5.963	.926	.344	2.463	13.343	.295	1.814
144.240ND	43.574	6.209	.973	.332	2.370	12.68	.285	1.739
147. PM	45.198	6.461	1.027	.320	2.282	12.	.274	1.667
150.35 SM	46.849	6.717	1.073	.309	2.200	11.47	.265	1.600
151.96 EU	48.519	6.981	1.126	.298	2.121	10.96	.256	1.538
157.250GD	50.233	7.243	1.185	.288	2.047	10.46	.247	1.478
158.924TB	52.002	7.515	1.241	.279	1.977	10.00	.238	1.422
162.500DY	53.793	7.790	1.295	.270	1.909	9.59	.230	1.369
164.93 HO	55.619	8.068	1.351	.261	1.845	9.20	.223	1.319
167.26 ER	57.487	8.358	1.401	.252	1.784	8.82	.216	1.271
168.934TM	59.380	8.650	1.461	.244	1.727	8.48	.209	1.225
173.04 YB	61.300	8.944	1.528	.237	1.672	8.149	.202	1.182
174.97 LU	63.310	9.249	1.589	.229	1.620	7.840	.196	1.140
178.49 HF	65.310	9.558	1.662	.222	1.570	7.539	.190	1.100
180.948TA	67.403	9.877	1.743	.215	1.522	7.252	.184	1.061
183.85 W	69.508	10.200	1.814	.209	1.476	6.983	.178	1.025
186.20 RE	71.658	10.531	1.890	.203	1.433	6.729	.173	.989
190.2 OS	73.856	10.868	1.967	.197	1.391	6.490	.168	.956
192.2 IR	76.101	11.212	2.048	.191	1.351	6.262	.163	.924
195.09 PT	78.381	11.562	2.133	.186	1.313	6.047	.158	.893
196.967AU	80.720	11.921	2.220	.180	1.276	5.840	.154	.864
200.590HG	83.109	12.286	2.313	.175	1.241	5.648	.149	.835
204.37 TL	5.249	12.660	2.406	.170	1.207	5.460	.145	.808
207.19 PB	5.076	13.041	2.502	.165	1.175	5.286	.141	.782
208.980BI	4.909	13.426	2.603	.161	1.144	5.118	.137	.757
210. PO		13.814	2.683	.156	1.114		.133	.732
210. AT		14.214	2.787	.152	1.085		.129	.709
222. RN		14.619	2.892	.148	1.057		.126	.687
223. FR		15.031	3.000	.144	1.030		.123	.665
226. RA		15.444	3.105	.140	1.005		.119	.645
227. AC		15.871	3.219	.136	.980		.116	.625
232.038TH	3.941	16.300	3.325	.133	.956	4.138	.115	.606
231. PA	3.827	16.733	3.436	.129	.933	4.022	.110	.587
238.030 U	3.716	17.165	3.545	.126	.911	3.910	.107	.570
237. NP	3.611	17.610	3.666		.889		.104	.553
239. PU	3.505	18.054	3.778		.868	3.696	.102	.537
243. AM		18.504	3.887		.848		.099	.521
247. CM		18.930	3.971				.097	.507
247. BK		19.452	4.132				.091	
251. CF		19.930	4.253				.091	.475
254. ES		20.410	4.374				.089	.461
253. FM		20.900	4.498				.087	.448

2.335
 2.786
 2.652
 2.852
 2.661
 2.9212.365
 2.8892.370
 2.8502.346
 2.8132.390
 2.8032.452
 2.7842.59
 2.8512.870
 2.7802.7701.309
 2.7802.68 1.34
 2.7732.61 1.38
 2.7482.5921.42
 2.8472.6521.46
 2.7962.5811.520

999.
 999.
 999.
 999.
 999.
 677.491999.
 405. 999.
 250.700999.
 170.490999.
 124.960999.
 94.000999.
 75.800999.
 61.990708.462 999.
 50.563490.043 999.846999.846

42.278365.572696.522696.522999.	2.8022.54 1.522
35.491283.709488.114488.114999.	2.8102.52 1.524
30.826230.447383.842383.842999.	2.7952.4951.526
27.290205.606358.326358.326999.	2.8212.4781.529
24.172186.437327.992327.992999.	2.8472.3082.453
21.581167.315291.720291.720999.	2.8472.3891.926
19.380147.772255.104255.104999.	2.8642.6601.94
17.525133.456229.594230.000999.	2.8362.6441.965
15.915123.119202.000208.371999.	2.8412.6881.98
14.525110.895182.057188.400999.	2.7342.6931.991
13.288110.600159.500166.000999.	2.7342.7491.820
12.131 91.230137.000143.900999.	2.7712.6561.85
11.100 78.419116.087120.487712.534712.534999.	2.5602.6361.88
10.187 68.878 96.936102.633431.989431.989999.	2.7692.4711.915
9.367 60.924 84.686 88.243300.924300.924999.	2.5892.6501.97
8.646 53.556 73.710 76.579218.661227.800999.	2.7822.6722.05
7.984 48.336 65.494 68.309176.863179.682399.0002.7712.6472.14	
7.392 43.07 55.672 57.989129.500139.461421.0002.6262.6332.242	
6.862 38.491 50.114 51.984110.895112.403388.0002.5932.6062.34	
6.387 34.680 44.310 46.072 91.838 93.149317.0002.6682.7912.44	
5.962 31.499 39.687 41.286 77.682 78.768275.6002.5782.5942.53	
5.579 28.475 35.565 36.972 66.137 67.185240.0002.7522.7282.583.40	
5.230 26.469 32.709 34.154 59.779 60.597214.1002.6612.7341.923.48	
4.913 24.413 30.084 31.402 53.278 54.201186.9002.7512.7121.99.70	
4.630 22.5 27.578 28.853 47.508 48.140162. 2.70 2.6752.19 1.2	
4.369 20.945 25.461 26.511 42.660 43.039145.5002.6132.6572.46 1.46	
4.130 19.454 23.342 24.492 38.561 39.286136.8002.4262.6822.55 1.56	
3.907 18.109 21.603 22.699 34.941 35.494122.6002.6742.7012.6041.59	
3.700 16.878 20.119 21.061 30.82 31.14 110.5002.4272.7002.393.385	
3.505 15.874 18.603 19.614 28.13 29.50 101.0002.4692.7032.5741.60	
3.324 14.764 17.314 18.285 26.718 27.166 91.1002.3962.7012.55 1.60	
3.156 13.867 16.050 17.200 24.28 24.90 85.8002.6562.7022.5001.593	
3.000 13.020 15.072 16.014 22.699 23.114 76.6002.4752.6182.37 1.48	
2.856 12.275 14.186 15.080 21.124 21.528 72.0002.5262.6442.16 1.26	
2.720 11.575 13.345 14.193 19.660 20.090 66.9002.4532.6481.94.92	
2.593 10.8 12.410 13.232 17.8 18.441 60.1 2.44 2.6161.880.757	
2.474 10.186 11.641 12.428 16.766 17.089 53.7182.4352.6091.92.83	
2.363 9.590 10.907 11.672 15.560 15.890 49.0042.4152.6571.96.90	
2.261 9.108 10.294 11.036 14.612 14.907 45.8512.3752.6062.04 1.24	
2.166 8.642 9.741 10.459 13.756 14.036 42.8112.3262.6152.13 1.44	
2.079 8.205 9.270 9.975 13.122 13.394 40.716 2.6172.28 1.72	
1.997 7.870 8.838 9.556 12.459 12.737 39.334 2.6112.43 1.95	
1.919 7.55 8.426 9.137 11.791 12.073 37.6 2.61 2.58 2.12	
1.846 7.196 8.047 8.732 11.288 11.552 35.864 2.6232.7222.26	
1.776 6.888 7.682 8.374 10.711 11.013 34.420 2.6742.7462.36	
1.712 6.592 7.344 8.030 10.186 10.461 32.991 2.7112.7552.4	
1.650 6.301 7.014 7.694 9.724 9.989 31.159 2.7402.6852.39	
1.592 6.057 6.732 7.399 9.304 9.574 29.782 2.7552.6642.38	
1.537 5.825 6.448 7.120 8.910 9.174 28.456 2.7602.6322.37	
1.484 5.619 6.181 6.843 8.601 8.847 27.606 2.7682.6012.36	
1.433 5.374 5.933 6.579 8.186 8.487 26.284 2.7692.6002.35	
1.386 5.170 5.706 6.359 7.865 8.115 25.448 2.7632.5682.34	
1.341 4.977 5.477 6.127 7.562 7.805 24.492 2.7642.5582.32	
1.297 4.767 5.241 5.882 7.223 7.461 23.040 2.7442.5332.31	
1.255 4.585 5.020 5.650 6.870 7.110 21.9242.5172.6132.5892.30	
1.216 4.407 4.815 5.435 6.590 6.830 20.8372.0432.6682.5412.28	
1.177 4.236 4.620 5.234 6.330 6.560 19.837 2.7252.512.25	
1.141 4.071 4.433 5.043 6.073 6.300 18.949 2.7622.4922.22	
1.106 3.915 4.260 4.861 5.830 6.090 17.966 2.7032.4682.17	
1.072 3.762 4.093 4.686 5.590 5.810 17.172 2.6532.5202.06	
1.040 3.616 3.936 4.518 5.374 5.584 16.339 2.5112.5282.068	
1.009 3.478 3.783 4.355 5.157 5.360 15.492 2.5952.4742.12	
.979 3.346 3.634 4.198 4.952 5.153 14.664 2.6992.3332.19	
.951 3.217 3.492 4.047 4.757 4.955 13.874 2.6012.4562.280	
.923 3.094 3.359 3.904 4.572 4.764 13.2152.2282.5872.5272.313	
.898 2.988 3.217 3.755 4.431 4.621 12.457 2.60 2.36 2.24	
.872 2.872 3.093 3.619 4.262 4.449 11.898 2.62 2.18 2.09	

.848	2.766	2.981	3.504	4.103	4.286	11.301	2.6532.1471.92
.825	2.665	2.865	3.385	3.953	4.133	10.753	2.6852.15 1.82
.803	2.571	2.762	3.270	3.817	3.993	10.260	2.7272.1541.76
.781	2.479	2.663	3.172	3.679	3.852	9.770	2.7152.28 1.73
.761	2.392	2.567	3.068	3.557	3.729	9.325	2.6892.6471.74
.741	2.310	2.479	2.970	3.433	3.602	8.938	2.66 1.81
.722	2.235	2.392	2.884	3.333	3.497	8.605	3.1472.6761.888
.704	2.166	2.310	2.796	3.220	3.382	8.262	2.74 1.95
.687	2.090	2.237	2.721	3.121	3.282	7.955	2.8192.03
.670	2.026	2.171	2.656	3.030	3.190	7.667	2.82 2.08
.655	1.972	2.103	2.584	2.933	3.122	7.546	2.83 2.13
.637	1.891	2.017	2.491	2.840	3.000	7.064	2.84 2.17
.622	1.836	1.950	2.427	2.757	2.915	6.892	2.85 2.22
.607	1.777	1.886	2.361	2.678	2.834	6.637	2.86 2.25
.593	1.721	1.825	2.297	2.601	2.756	6.401	2.87 2.27
000.144							.10 040335 E+01
.084							.10 059716 E+01
.250							.10 085924 E+01
.404							.10 119003 E+01
.889							.10 153530 E+01
1.350	.295						.10 194915 E+01
2.311	.471						.10 240062 E+01
3.529	.818						.10 284504 E+01
4.943	.904						.10 335616 E+01
7.322	1.098						.10 390856 E+01
9.322	.80						.10 442539 E+01
11.601	.700						.10 499690 E+01
14.87	1.110	7.338					.10 555707 E+01
18.122	1.88	8.5					.10 614998 E+01
22.020	2.6	10.0					.10 669828 E+01
27.626	3.330	11.2					.10 730918 E+01
27.864	3.548	12.5					.10 785874 E+01
35.312	4.567	13.715					.10 843181 E+01
42.038	5.8	15.					.10 905935 E+01
48.678	7.3	16.8					.10 957055 E+01
54.052	8.8	18.1					.11 017059 E+01
60.083	10.927	19.593					.11 071056 E+01
64.977	12.056	.742					.11 122901 E+01
77.899	12.661	6.095					.11 173825 E+01
83.305	10.979	6.4					.11 228172 E+01
97.144	12.734	6.8					.11 276960 E+01
108.210	14.322	7.2					.11 322652 E+01
117.098	15.520	7.839					.11 371081 E+01
123.080	15.783	18.830					.11 413858 E+01
146.550	19.124	18.2					.11 459289 E+01
142.035	20.238	17.0					.11 499011 E+01
169.094	24.425	16.246					.11 536316 E+01
156.083	24.434	14.6					.11 573735 E+01
188.918	26.852	13.5					.11 614852 E+01
207.363	29.204	12.0					.11 647602 E+01
209.178	32.008	10.150					.11 681141 E+01
217.375	35.425	9.2					.11 709942 E+01
238.313	34.296	8.6					.11 741591 E+01
261.398	42.122	7.8					.11 770344 E+01
277.847	41.921	7.356 6.					.11 796616 E+01
291.277	44.976	22.304 8.					.11 822350 E+01
325.568	49.133	20.0 14.					.11 845544 E+01
332.	53.	18.4 18.					.11 865519 E+01
338.246	57.048	15.0 28.					.11 891662 E+01
266.121	60.395	13.0 66.					.11 910259 E+01
380.209	62.906	11.593155.					.11 927648 E+01
278.270	66.473	18.1571478.19					.11 947537 E+01
325.561	70.806	14.947108.					.11 963426 E+01
291.804	75.185	16.8 44.					.11 979680 E+01
429.243	79.510	20.595 34.465					.11 993455 E+01
360.505	87.573	30.1 51.8					.12 009454 E+01
428.535	90.994	57.0 86.					.12 018723 E+01

391.501	94.983	74.0	156.		.12033243E+01
393.	102.108	83.443	403.675		.12043491E+01
396.	107.580	76.	340.		.12050691E+01
400.642	108.976	64.	259.		.12067255E+01
406.	117.488	52.	210.		.12073447E+01
415.658	128.292	42.	155.		.12081746E+01
	133.322	36.	120.		.12092015E+01
	138.475	31.8	81.		.12099530E+01
	143.5	30.8	57.		.12105282E+01
	150.262	30.254	34.		.12112746E+01
	164.053	31.414	19.5		.12120670E+01
	171.439	32.195	9.0		.12127263E+01
	182.478	35.891	5.0		.12127321E+01
	191.747	37.859	4.0		.12133597E+01
	203.031	40.695	6.0		.12139165E+01
	210.729	42.976	8.8		.12146045E+01
	222.047	45.248	9.9		.12148826E+01
	227.861	47.589	11.8		.12153547E+01
	239.732	50.412	14.5		.12156483E+01
.018	247.226	53.018	16.2	1990.0.	.12154626E+01
.018785.744	221.796	53.096	18.4	19011.799	.12159731E+01
.018344.007	244.065	55.507	21.0	18100.	.12160668E+01
.017	271.	60.5	23.0	17300.	.12162942E+01
.017	290.700	63.573	25.5	16500.	.12164937E+01
.016	291.982	67.423	27.6	15700.	.12164491E+01
.016	277.059	66.215	27.8	14935.4	.12166268E+01
.043	241.237	70.331	27.626	10444.9	.12165546E+01
.058	290.	75.655	24.6	9100.	.12168429E+01
.073	342.330	85.413	21.0	7800.	.12163713E+01
.088	311.716	81.287	20.128	6500.	.12165291E+01
.106631.182	325.828	82.053	20.591	5216.8	.12163066E+01
	348.	82.	23.0		.12162152E+01
	371.	96.	27.5		.12159503E+01
	394.339	112.607	36.0		.12157646E+01
	420.	117.	47.0		.12154985E+01
	440.969	121.092	57.0		.12152714E+01
	441.	114.	62.0		.12147656E+01
	441.914	108.024	58.0		.12144476E+01
		108.	51.0		.12141881E+01
	649.224	108.019	46.169		.12136381E+01
		122.	47.0		.12132737E+01
		144.388	48.		.12129561E+01
		155.	51.		.12121884E+01
		164.	56.		.12120564E+01
		166.	63.		.12119586E+01
		169.5	67.		.12116752E+01
		173.	74.		.12112705E+01
		175.	80.		.12111340E+01
-.51088316E-02	.11210435E-02	.11886647E-03	.58783242E-05	.10814593E-06	
-.72186773E-02	.14225237E-02	.14111194E-03	.66914334E-05	.11980721E-06	
-.10267103E-01	.19859984E-02	.19498321E-03	.91950066E-05	.16415866E-06	
-.14245915E-01	.28139927E-02	.28115056E-03	.13437614E-04	.24230877E-06	
-.18509077E-01	.36845955E-02	.36908566E-03	.17648433E-04	.31814958E-06	
-.23691260E-01	.48299260E-02	.49167519E-03	.23763519E-04	.43147057E-06	
-.29437536E-01	.61336917E-02	.63336010E-03	.30898163E-04	.56450902E-06	
-.35199645E-01	.74029719E-02	.76721501E-03	.37474758E-04	.68485599E-06	
-.41864668E-01	.89623354E-02	.93928950E-03	.46210926E-04	.84851560E-06	
-.49116674E-01	.10703541E-01	.11347223E-02	.56245418E-04	.10379369E-05	
-.56049932E-01	.12310884E-01	.13096993E-02	.65021871E-04	.12007774E-05	
-.63725327E-01	.14165786E-01	.15179592E-02	.75701781E-04	.14020896E-05	
-.71366790E-01	.15998293E-01	.17221066E-02	.86106469E-04	.15973207E-05	
-.79481274E-01	.17990243E-01	.19476971E-02	.97735780E-04	.18172649E-05	
-.87160478E-01	.19826558E-01	.21511784E-02	.10805634E-03	.20101768E-05	
-.95653455E-01	.21945888E-01	.23934259E-02	.12061385E-03	.22485051E-05	
-.10350848E+00	.23847385E-01	.26056661E-02	.13142934E-03	.24513122E-05	
-.11171634E+00	.25877884E-01	.28357954E-02	.14327977E-03	.26751226E-05	
-.12063543E+00	.28167495E-01	.31023590E-02	.15726737E-03	.29428023E-05	

-1.12824497E+00	.30004786E-01-	.33062601E-02	.16760104E-03-	.31357012E-05
-1.13696853E+00	.32247548E-01-	.35671998E-02	.18128080E-03-	.33972816E-05
-1.14503823E+00	.34273003E-01-	.37984821E-02	.19324300E-03-	.36238338E-05
-1.15291678E+00	.36249918E-01-	.40240545E-02	.20490146E-03-	.38444980E-05
-1.16072637E+00	.38216675E-01-	.42489089E-02	.21653708E-03-	.40649144E-05
-1.16898749E+00	.40360370E-01-	.44993224E-02	.22969026E-03-	.43166779E-05
-1.17663502E+00	.42308407E-01-	.47236572E-02	.24135446E-03-	.45383465E-05
-1.18395179E+00	.44157222E-01-	.49350727E-02	.25228900E-03-	.47453615E-05
-1.19162687E+00	.46154730E-01-	.51685157E-02	.26454822E-03-	.49799279E-05
-1.19865794E+00	.47943894E-01-	.53739211E-02	.27519705E-03-	.51818324E-05
-1.20602707E+00	.49873145E-01-	.56000292E-02	.28708913E-03-	.54095794E-05
-1.21274769E+00	.51595238E-01-	.57985500E-02	.29740763E-03-	.56055462E-05
-1.21919369E+00	.53240969E-01-	.59875816E-02	.30720443E-03-	.57912005E-05
-1.22566990E+00	.54918897E-01-	.61823635E-02	.31737469E-03-	.59849446E-05
-1.23259096E+00	.56772842E-01-	.64027040E-02	.32906669E-03-	.62101742E-05
-1.23852883E+00	.58300778E-01-	.65788492E-02	.33821274E-03-	.63836742E-05
-1.24456601E+00	.59884886E-01-	.67641096E-02	.34793033E-03-	.65693493E-05
-1.25004521E+00	.61292189E-01-	.69259326E-02	.35631479E-03-	.67281488E-05
-1.25585394E+00	.62832701E-01-	.71071974E-02	.36585778E-03-	.69109125E-05
-1.26131735E+00	.64271190E-01-	.72754737E-02	.37467921E-03-	.70793358E-05
-1.26647808E+00	.65619559E-01-	.74320953E-02	.38284496E-03-	.72346212E-05
-1.27157351E+00	.66965836E-01-	.75897928E-02	.39111550E-03-	.73925414E-05
-1.27634956E+00	.68213686E-01-	.77345390E-02	.39865139E-03-	.75356829E-05
-1.28073135E+00	.69337986E-01-	.78629577E-02	.40525732E-03-	.76600255E-05
-1.28582501E+00	.70728039E-01-	.80292836E-02	.41410562E-03-	.78306480E-05
-1.29001167E+00	.71814000E-01-	.81541870E-02	.42055995E-03-	.79525070E-05
-1.29403845E+00	.72858135E-01-	.82741686E-02	.42675539E-03-	.80694276E-05
-1.29833832E+00	.74010390E-01-	.84098331E-02	.43388315E-03-	.82056114E-05
-1.30214063E+00	.74997904E-01-	.85231436E-02	.43972127E-03-	.83155561E-05
-1.30597373E+00	.76011294E-01-	.86410743E-02	.44586206E-03-	.84321089E-05
-1.30948495E+00	.76919693E-01-	.87447947E-02	.45118255E-03-	.85319524E-05
-1.31323346E+00	.77921045E-01-	.88619773E-02	.45730311E-03-	.86483225E-05
-1.31616897E+00	.78646785E-01-	.89414513E-02	.46124204E-03-	.87202754E-05
-1.31969356E+00	.79586642E-01-	.90510780E-02	.46695028E-03-	.88285401E-05
-1.32269303E+00	.80351500E-01-	.91369988E-02	.47129456E-03-	.89091177E-05
-1.32530877E+00	.80987026E-01-	.92051589E-02	.47460651E-03-	.89685980E-05
-1.32900290E+00	.82008486E-01-	.93273969E-02	.48108620E-03-	.90930449E-05
-1.33145460E+00	.82601088E-01-	.93905135E-02	.48413154E-03-	.91473912E-05
-1.33412717E+00	.83274589E-01-	.94650044E-02	.48784260E-03-	.92153709E-05
-1.33701012E+00	.84027757E-01-	.95508860E-02	.49222640E-03-	.92972007E-05
-1.33954821E+00	.84664835E-01-	.96209508E-02	.49569756E-03-	.93604821E-05
-1.34185234E+00	.85223639E-01-	.96803460E-02	.49855117E-03-	.94111708E-05
-1.34432934E+00	.85842550E-01-	.97478971E-02	.50187197E-03-	.94713033E-05
-1.34684918E+00	.86483371E-01-	.98189606E-02	.50541124E-03-	.95360534E-05
-1.34919346E+00	.87066345E-01-	.98822445E-02	.50850526E-03-	.95918031E-05
-1.35074655E+00	.87372330E-01-	.99069973E-02	.50934814E-03-	.96014296E-05
-1.35301387E+00	.87932134E-01-	.99672026E-02	.51226411E-03-	.96535270E-05
-1.35518044E+00	.88459059E-01-	.10022944E-01	.51492216E-03-	.97003808E-05
-1.35747831E+00	.89032813E-01-	.10085205E-01	.51795940E-03-	.97549566E-05
-1.35927929E+00	.89432565E-01-	.10123148E-01	.51957029E-03-	.97802358E-05
-1.36129036E+00	.89909138E-01-	.10172039E-01	.52182953E-03-	.98188980E-05
-1.36307709E+00	.90307205E-01-	.10209906E-01	.52343928E-03-	.98441771E-05
-1.36428614E+00	.90505087E-01-	.10220125E-01	.52344112E-03-	.98366117E-05
-1.36629296E+00	.90982954E-01-	.10269137E-01	.52570012E-03-	.98751434E-05
-1.36780074E+00	.91288432E-01-	.10294398E-01	.52658101E-03-	.98855786E-05
-1.36944046E+00	.91638452E-01-	.10325525E-01	.52778792E-03-	.99024313E-05
-1.37103485E+00	.91975714E-01-	.10355106E-01	.52891343E-03-	.99177449E-05
-1.37232571E+00	.92206793E-01-	.10369838E-01	.52916608E-03-	.99150840E-05
-1.37385404E+00	.92521425E-01-	.10396121E-01	.53009035E-03-	.99261452E-05
-1.37506964E+00	.92727292E-01-	.10407283E-01	.53013042E-03-	.99190754E-05
-1.37668731E+00	.93074957E-01-	.10438065E-01	.53130967E-03-	.99352238E-05
-1.37738192E+00	.93098117E-01-	.10423582E-01	.52983961E-03-	.98970312E-05
-1.37879579E+00	.93375409E-01-	.10444441E-01	.53043021E-03-	.99009704E-05
-1.37972521E+00	.93481251E-01-	.10441243E-01	.52960782E-03-	.98758739E-05
-1.38078480E+00	.93635483E-01-	.10444841E-01	.52918176E-03-	.98588646E-05
-1.38159628E+00	.93700760E-01-	.10435839E-01	.52801171E-03-	.98265157E-05
-1.38245741E+00	.93781724E-01-	.10428664E-01	.52693158E-03-	.97957581E-05

```

-.38317627E+00 .93812360E-01-.10414347E-01 .52542786E-03-.97562349E-05
-.38389147E+00 .93841810E-01-.10399708E-01 .52389703E-03-.97160336E-05
-.38422579E+00 .93736706E-01-.10366139E-01 .52125106E-03-.96528487E-05
-.38470337E+00 .93676237E-01-.10338166E-01 .51890823E-03-.95955598E-05
-.38518653E+00 .93618256E-01-.10310387E-01 .51656750E-03-.95381665E-05
-.38523503E+00 .93400300E-01-.10259642E-01 .51286035E-03-.94524679E-05
-.38541494E+00 .93224969E-01-.10214321E-01 .50944720E-03-.93724564E-05
-.38554613E+00 .93025953E-01-.10165031E-01 .50577736E-03-.92868621E-05
-.38503845E+00 .92596007E-01-.10082920E-01 .50016376E-03-.91610925E-05
-.38515389E+00 .92377917E-01-.10029531E-01 .49619830E-03-.90686893E-05
-.38516918E+00 .92113527E-01-.99686511E-02 .49175586E-03-.89659814E-05
-.38481991E+00 .91711352E-01-.98877062E-02 .48610989E-03-.88382291E-05
-.38416494E+00 .91188823E-01-.97887816E-02 .47936677E-03-.86873630E-05
-.38363713E+00 .90693381E-01-.96920228E-02 .47269105E-03-.85370969E-05

```

KALAMAMB

ABCDEFGHIJKLMN OPQRSTUVWXYZ

A0001

Graph titles

RADIAL DISTANCE FROM CENTER(MILS)-WT PERCENT OF U WT PERCENT OF PU WT PERCENT OF									
U+PU	WT PERCENT	RATIO	PU/U	WT PERCENT	RATIO	PU/(U+PU)	0	0	0
0	0	0	0	0	0	0	9	8	7
0	10	20	30	40	50	60	70	80	90
00	110	120	130	* :PU/U . :PU/(U+PU)					
RADIAL DISTANCE FROM CENTER(MILS) WT PERCENT OF U WT PERCENT OF PU WT PERCENT OF									
F U+PU	WT PERCENT	RATIO	PU/U	WT PERCENT	RATIO	PU/(U+PU)			

/*

APPENDIX B
List of Symbols

<u>Common Designation</u>	<u>Program Designation</u>
Atomic number	AT.NO.
Atomic weight	AT.WT.
Analytical-line wavelength	WAVE
Analytical-line excitation	EC
Absorption edge of analytical line	EDGE
Fluorescent yield	W
Mean ionization potential, J	PF
Overvoltage ratio	U
Backscatter factor	R
Z/A	WS
$ E_O + E_C /2$	E
$\ln 1.166\bar{E}$	VAL
Electron-energy coefficient, σ	SIGMA
h in Philibert's equation	HS
$1 + \chi/\sigma$	US
$f(\chi)$	FS
$f(\chi) R/S$	FACT
Mean backscatter factor	RAL
Mean atomic number, $\bar{Z}$	ZAL
Mean $(\bar{Z}/\bar{A})$	WSAL
Mean h value, $\bar{h}$	HAL
Mean ionization potential, $\bar{J}$	PAL
Chi for an alloy, χ	YP
Mean overvoltage ratio	UAL
$f(\chi)$ for an alloy	FAL

APPENDIX C

Programming Notes for Plotting Subroutines

1. Subroutine GRAPHS

Subroutine GRAPHS calls plotting subroutines depending on the value of a variable MARY, which is read from the graph options card. This routine also calculates the plutonium-to-uranium and plutonium-to-uranium-plus-plutonium ratios, which are then plotted. The ratios to be plotted and the graph axis labels were determined when the program was written. For other ratios or axis labels, minor changes in the program would be necessary. The axis labels in use at ANL appear at the end of the listing. A test variable JIM ensures that the printer-plot subroutine PLOT is called only once. Another test variable LET prevents calling the CALPLT and PLOT subroutines more than once.

2. Subroutine CALPLT

All graphs generated by this subroutine are automatically scaled based on the range of values to be plotted, except the graphs of uranium-plus-plutonium wt % versus distance and of uranium, plutonium, and uranium-plus-plutonium versus distance. For these plots, the ordinate limits are always 0 and 100 wt %.

3. Printer Subroutine PLOT

This subroutine is more restrictive with regard to scaling of the printer graph axes. All scaling is preset, and the maximum distance that can be plotted is 134 mils. Because the routine determines the location of a point by truncation, the points on the graph are located to within 2 wt % on the y axis and 1.25 mils on the x axis.

APPENDIX D

Listing of MAGRAM Input and Output1. MAGRAM Input

AO120					<u>Problem Number and Identification</u>	
HQV15	45-82 RADIAL DISTN. OF U & PU				Microprobe analysis conditions (K.V., No. of standards, counting time, etc.)	
25.C	2	10.0	9.13	45.0		
U MA	PU MB	C	Identification of characteristic x-ray lines and compound standards			
1U02	.8810	0				
2PU02	.8835	0	X-ray intensities measured on the standards (beam current and total counts in 10 sec.)			
1	43000	22500				
1	43000	84000	Background matrix on UO ₂ $\begin{bmatrix} D & T \\ T & D \end{bmatrix}$ on PuO ₂ $\begin{bmatrix} D & T \\ T & D \end{bmatrix}$ $\begin{matrix} D = \text{Spectrometer Detuned to obtain background} \\ T = \text{Spectrometer tuned to analytical line} \end{matrix}$			
1	43000	310				
1	43000	310	1550	Radioactive background, beam turned off		
		11000	500			
42950	19500	18380	X-ray intensities measured on the sample and beam current			
43100	19640	18270				
43200	19600	18730	Card indicating end of data set			
43450	19290	19050				
43500	19470	18700	Graph Options Card			
99999						
54444					Radial position cards	
92.0						
99.0					Card indicating end of problem	
98.0						
104.0						
107.0						
0.000						

2. MAGRAM Output

PROBLEM NUMBER A 120

ELEMENT	AT. NO.	AT. WT.	PF	EDGE	EC	WAVE	U	R
U	92	238.030	1243.786	3.497	3.545	3.910	7.05	0.647279
PU	94	239.000	1270.345	3.282	3.778	3.505	6.62	0.641683
O	8	15.999	0.0	23.320	0.532	23.620	20.00	0.952346

MASS ABSORPTION COEFFICIENTS

RADIATION	U MA	PU MB
ABSORBER		
U	606.	493.
PU	764.	612.
O	172.	126.

	WS	E	VAL	SIGMA	HS	US	FS	FACT
U	0.3865	14.2725	9.7193	2146.62	0.1266	1.3992	0.6932	0.44765
PU	0.3933	14.3890	9.7275	2158.82	0.1217	1.4011	0.6928	0.43806

ESP = 0.25455039E 04

STANDARD INTENSITIES CORRECTED FOR DEADTIME AND DRIFT

BEAM CURRENT	U	PU
43000.0	23058.41	86171.44

COUNTING INTERVAL, 10.0 SECOND(S)

BACKGROUNDS

CONTRIBUTED TO	U MA	PU MB
BY 100 % OF		

U	310.0	2552.0
PU	310.0	1550.7

PROBLEM NUMBER A 120

ELEMENT	ATOMIC NUMBER	ATOMIC WEIGHT	BACKSCATTER FACTOR	EXCITATION POTENTIAL	ABSORPTION JUMP RATIO	FLUORESCENT YIELD
U	92	238.030	0.647	3.545	0.00	0.000
PU	94	239.000	0.642	3.778	0.00	0.000
O	8	15.999	0.952	0.532	21.10	0.004

MASS ABSORPTION COEFFICIENTS

RADIATION	U MA	PU MB
ABSORBER		

U	606.	493.
PU	576.4.	612.
O	472.	126.

COUNTING INTERVAL, 10.0 SECOND(S)

BACKGROUNDS

CONTRIBUTED TO	U MA	PU MB
BY 100 % OF		

U	310.0	2552.0
PU	310.0	1550.7

PROBLEM NUMBER A 120

 INTENSITIES (CPS X COUNTING INTERVAL)
 CORRECTED FOR DEAD-TIME AND DRIFT

U	PU
18636 / 23058	18002 / 86171
18712 / 23058	17829 / 86171
18629 / 23058	18250 / 86171
18211 / 23058	18466 / 86171
18370 / 23058	18095 / 86171

STANDARD BACKGROUNDS (CPS X COUNTING INTERVAL)

31C	1551

DEAD-TIME (MICROSECONDS)

31C	3.0

COUNTING INTERVAL, 10.0 SECONDS

PROBLEM NUMBER A 120

 INDIVIDUAL K-RATIOS CORRECTED FOR
 DEAD-TIME, DRIFT AND BACKGROUND ONLY

N	U	PU
3	0.8074	0.1883
3	0.8108	0.1862
3	0.8071	0.1912
3	0.7889	0.1941
3	0.7959	0.1896

NOTE

 U DETERMINED RELATIVE
 TO A STANDARD OF UO₂

 PU DETERMINED RELATIVE
 TO A STANDARD OF PUO₂

PROBLEM NUMBER A 120

SUBMITTED BY

DESCRIPTION - HOV15 45-B2 RADIAL DISTN. OF U & PU

ACCELERATING VOLTAGE 25.0 KEV
 X-RAY EMERGENCE ANGLE 45.0 DEGREES

STANDARD PEAK-TO-BACKGROUND RATIOS (P/B) AND
 MINIMUM DETECTABILITY LIMITS (MDL)

ELEMENT	P/B	MDL
U	83/1	0.2393 WT %
PU	61/1	0.1446 WT %

CHEMICAL COMPOSITION
 WEIGHT PERCENT
 ATOMIC PERCENT

O DETERMINED BY DIFFERENCE

OBS	U	PU	O
1	72.206 26.538	15.681 5.826	12.114 67.236
2	72.491 27.195	15.503 5.792	12.006 67.013
3	72.193 27.258	15.924 5.988	11.884 66.754
4	70.592 24.899	16.180 5.684	13.228 69.417
5	71.185 25.373	15.800 5.609	13.015 69.018

SUBMITTED BY

DESCRIPTION - HQV15 45-82 RADIAL DISTN. OF U & PU

MEAN CHEMICAL COMPOSITION AND TWO
SIGMA LIMITS BASED ON 5 ANALYSES

ELEMENT	WEIGHT PERCENT	ATOMIC PERCENT
U	71.738 - 1.451	26.337 - 1.993
PU	15.822 - 0.462	5.785 - 0.265
O *	12.439 - 1.121	67.878 - 2.216

* DETERMINED BY DIFFERENCE

MEAN INTENSITY RATIOS AND TWO SIGMA LIMITS

ELEMENT	I K
U	0.8020 - 0.0166
PU	0.1898 - 0.0054

ACCELERATING VOLTAGE	25.0 KEV
X-RAY EMERGENCE ANGLE	45.0 DEGREES
DENSITY	9.13
DEPTH OF ANALYZED REGION	1.13 MICRONS

STANDARD PEAK-TO-BACKGROUND RATIOS (P/B) AND
MINIMUM DETECTABILITY LIMITS (MDL)

ELEMENT	P/B	MDL
U	83/1	0.2393 WT %
PU	61/1	0.1446 WT %

ACKNOWLEDGMENTS

We wish to express our appreciation to H. Savage for expanding and modifying parts of the MAGIC-II program, to R. C. Fernow for programming the plotting routines, to J. H. Kittel for suggesting the problem, and to B. R. T. Frost for constant encouragement.

REFERENCES

1. B. T. Bradbury, J. T. Demant, P. M. Martin, and D. M. Poole, *Electron Probe Microanalysis of Irradiated UO₂*, J. Nucl. Mater. 17, 227-236 (1965).
2. B. M. Jeffery, *Microanalysis of Inclusions in Irradiated UO₂*, J. Nucl. Mater. 22, 33-40 (1967).
3. J. I. Bramman, R. M. Sharpe, D. Thom, and G. Yates, *Metallic Fission-product Inclusions in Irradiated Oxide Fuels*, J. Nucl. Mater. 25, 201-215 (1968).
4. V. G. Macres, O. Preston, N. C. Yew, and R. Buchanan, "A Shielded X-ray Microprobe for the Analysis of Radioactive Samples," *Proc. Fifth Intl. Congress on X-ray Optics and Microanalysis*, Springer-Verlag, Berlin, pp. 248-249 (1969).
5. R. Natesh, B. J. Koprowski, E. M. Butler, and D. A. Donahue, "Transfer and Analysis of Highly Radioactive Materials in a Shielded Electron Microprobe," *Proc. 16th Conf. on Remote Systems Technology*, pp. 243-252 (1969).
6. R. C. Fernow, Argonne National Laboratory, private communication (1969).
7. L. J. Gray, Ph.D. thesis, University of Illinois (1969).
8. J. D. Brown, "A Computer Program for Quantitative Electron Probe Microanalysis," in *The Electron Microprobe*, T. D. McKinley, K. F. J. Heinrich, and D. B. Wittry, eds., John Wiley and Sons, New York, pp. 189-198 (1966).
9. J. D. Brown, *A Comprehensive Computer Program for Electron Probe Microanalysis*, Anal. Chem. 38, 890-894 (1966).
10. J. Z. Frazer, R. W. Fitzgerald, and A. M. Reid, *Computer Programs EMX and EMX2 for Electron Microprobe Data Processing*, S10 Reference 66-14, University of California, San Diego, Calif. (June 1966).
11. E. Lifshin and R. E. Hanneman, *Electron Microbeam Probe Analysis*, Report 66-C-250, Part II, General Electric Co., Schenectady, N.Y. (1966).
12. D. R. Beaman, "A Computer Program for Quantitative Electron Probe Microanalysis," *Proc. Second Natl. Conf. on Electron Microprobe Analysis*, Boston, Mass., Paper 11 (1967).
13. D. R. Beaman, *A Computer Program for Use in Quantitative Electron Probe Microanalysis*, Mikrochimica Acta (Wien), pp. 117-129 (1969).
14. J. W. Colby, "Quantitative Microprobe Analysis of Thin Insulating Films," *Advances in X-ray Analysis*, Plenum Press, New York, Vol. 11, pp. 287-305 (1968).
15. H. Savage, Argonne National Laboratory, private communication (1969).
16. *Quantitative Electron Probe Microanalysis*, K. F. J. Heinrich, ed., Natl. Bur. Std. Special Publ. 298 (1968).

17. J. Philibert, "Electron Probe Microanalysis," in *Modern Analytical Techniques for Metals and Alloys*, R. F. Bunshah, ed., Interscience Publishers, New York, pp. 419-531 (1970).
18. K. F. J. Heinrich, *Present State of the Classical Theory of Quantitative Electron Probe Microanalysis*, Natl. Bur. Std. Technical Note 521 (1970).
19. K. F. J. Heinrich, D. Vieth, and H. Yakowitz, "Correction for Nonlinearity of Proportional Counter Systems in Electron Probe X-ray Microanalysis," *Advances in X-ray Analysis*, Plenum Press, New York, Vol. 9, pp. 208-220 (1966).
20. J. Philibert, *X-ray Optics and X-ray Microanalysis*, H. H. Pattee, V. E. Cosslett, and A. Engstrom, eds., Academic Press, New York, p. 379 (1963).