Marion M. Chiles¹Abstract

The primary objective of this project was to investigate the feasibility of using a $\text{CaF}_2(\text{Eu})$ scintillator for detecting low-energy beta particles from tritium. A proof-of-principle detector was designed to flow tritium-spiked nitrogen gas across the surface of a thin scintillator, which was optically coupled between two low-noise photomultiplier tubes. Electronics for operating two photomultiplier tubes in coincidence eliminated most of the tube noise pulses and allowed detection of small pulses from the low-energy tritium beta particles.

Introduction

The primary objective of this project was to investigate the feasibility of using a previously suggested $\text{CaF}_2(\text{Eu})$ scintillator for detecting tritium (^3H).¹ In order to evaluate the usefulness of $\text{CaF}_2(\text{Eu})$ for this application, it was necessary to first determine its counting sensitivity for low-energy beta radiation from tritium. The maximum energy of beta particles from tritium is 18 keV, and the average energy is about 6 keV, making it very difficult to discern small electronic pulses from these low-energy particles in the presence of the noise pulses generated in the photomultiplier tubes. This situation is improved by using a coincidence circuit to generate a gate pulse only when a pulse from both tubes occurs simultaneously, hence rejecting most of the tube noise pulses. After the counting sensitivity is determined, an adequate size detector can be designed to detect 10^{-12} Ci/mL in air (which is the goal set by the U.S. Navy RADAC Program).

To simplify the initial proof-of-principle detector and test equipment setup, a beta source was prepared by depositing a small amount of ^{63}Ni on a flat aluminum plate. After the source was dried, a thin (0.00015-in.) mylar film was placed over the plate and glued around the edges to contain the radioactive material. This source was much more convenient to handle during preliminary tests than tritium, which is in a gaseous state at room temperature. Preliminary data from the ^{63}Ni source (maximum beta energy of 66 keV and average beta energy of 17 keV) indicated that the lower-energy beta particles from tritium can be detected using this technique.

Description of the Scintillator

The optimum thickness of $\text{CaF}_2(\text{Eu})$ to absorb the maximum energy beta particles from tritium and yet not absorb enough energy from background gamma photons to generate a detectable pulse is $\sim 920 \mu\text{g}/\text{cm}^2$.² A scintillator was prepared by evaporating $\text{CaF}_2(\text{Eu})$ onto both flat surfaces of a 2-mm-thick pyrex glass plate. The layer of scintillator deposited on the glass plate had an area density of $900 \mu\text{g}/\text{cm}^2$ on one side and $1000 \mu\text{g}/\text{cm}^2$ on the other side, which is near the optimum thickness.

One conceptual design¹ for a tritium monitor consists of several glass plates (light guides) coated with $\text{CaF}_2(\text{Eu})$ arranged in parallel with ~ 0.25 -in.

spacing with two low-noise and high-gain photomultiplier (PM) tubes optically coupled to opposite edges of the glass plates as shown in Fig. 1. To investigate this concept, a glass plate coated on one side with $\text{CaF}_2(\text{Eu})$ was used to determine the fraction of light lost when the tubes are coupled to the edges of the light guides compared to the light when coupled to the flat side. Pulse amplitude measurements were made both while the glass plate was optically coupled directly on the flat face of one PM tube and while one edge of the plate was coupled to the PM tube. The ratio of the average pulse amplitudes with the scintillator flat on the face of the tube and exposed to the ^{63}Ni source compared with the average pulse amplitude with the edge of the scintillator and glass plate coupled to the tube was 2.6. This measurement indicates that a significant portion of light is lost when tubes are coupled to the edge of the light guide; but by using high-gain, low-noise PM tubes and coincidence electronics (which will be explained later), sufficient signal is still available to detect the low-energy beta particles when tubes are coupled to the edges. This arrangement allows a large area of scintillator to be coupled to fairly small PM tubes, whereas if the scintillator is coupled flat on the tube, a very large PM tube would be required.

Description of Coincidence Electronics

Inherent dark-noise pulses from a photomultiplier tube are of about the same amplitude as some of the smallest pulses generated in $\text{CaF}_2(\text{Eu})$ by low-energy beta particles. The majority of these randomly occurring, undesirable noise pulses can be eliminated by using two photomultiplier tubes to view the same scintillator and operating them in coincidence mode whereby only a simultaneous pulse from both tubes will be recorded. A block diagram of the system is shown in Fig. 2. The fraction of total events counted in the coincidence operation depends on the resolving time allowed in the coincidence unit for the two pulses to coincide and supply a gate pulse to the multichannel analyzer. A resolving time of 2.0 μs gave best efficiency in these measurements.

Fig. 1. Conceptual design for several glass plates coated with scintillator and optically coupled between two photomultiplier tubes.

¹Operated by Martin Marietta Energy Systems, Inc., for the U.S. Department of Energy under Contract No. DE-AC05-84OR21400.

²M. M. Chiles, Oak Ridge National Laboratory, Bldg. 3500, P.O. Box X, Oak Ridge, TN 37831-6006.

MASTER

Jsu

Fig. 2. Block diagram of electronics for an assembly of two photomultiplier tubes. Tubes are coupled to opposite edges of the scintillator and are operated in coincidence mode to prevent counting of pulses occurring outside the scintillator.

Tritium Tests

A different scintillator assembly was designed to allow nitrogen gas spiked with tritium (tritiated methane) to flow through the scintillator cell as beta particles interact with the $\text{CaF}_2(\text{Eu})$. Figure 3 illustrates this design. The two parameters to be considered in estimating the sensitive volume of the flow cell are range of the beta particles and area of the scintillator. The area of the scintillator used is 18.6 cm^2 , but the range varies with the energy of the beta particles (discussed in the following section). Beta activity of the ^3H -spiked nitrogen gas tested was 2.6 nCi/mL and, with optimum electronic adjustments, the count rate in coincidence operation was about 21 counts/s.

Fig. 3. Gas flow cell for testing ^3H sensitivity. The pyrex plate is coated with $\text{CaF}_2(\text{Eu})$ on both sides and in the center of the cell; the edges are coupled to two photomultiplier tubes.

Table 1. Beta Energy Spectrum of Tritium on the Scintillator

In order to calculate an approximate beta exposure to the scintillator from the tritium-spiked gas, the beta energy distribution from tritium and the range of the beta particles were investigated. A tritium spectrum reported by Curran, Angus, and Cockroft¹ (Fig. 4) was selected as representative of the beta energy distribution. This distribution was divided into three energy brackets centered on values of energies for which published range data are available. The three energy brackets selected are $5 \pm 3 \text{ keV}$, $10 \pm 2 \text{ keV}$, and $15 \pm 3 \text{ keV}$, covering an energy span of betas from 2 to 18 keV. An estimated 20% of the betas have energy $< 2 \text{ keV}$ and after light losses will not generate a sufficient pulse in each PM tube to be counted in the coincidence mode. Therefore, detectable beta activity $> 2 \text{ keV}$ energy is estimated to be 80% of the total.

A population or percentage of beta activity in each energy bracket was estimated from the spectrum curve to be as follows: ~48% of the total beta population is in the energy bracket from 2 to 8 keV, about 20% is in the bracket from 8 to 12 keV, and about 12% is in the remaining bracket from 12 to 18 keV. These values are listed in Table 1, along with the range of the median energy beta particles of each bracket.

By summing the products of the area of the scintillator (18.6 cm^2), the depth of penetration, and the beta activity in each bracket, an estimated number of betas striking the scintillator can be calculated for each bracket: 8, 11.5, and 14 respectively.

Therefore, total betas estimated to reach the scintillator = $8 + 11.5 + 14 = 33.5 \text{ betas/s}$.

The integral counts from flowing tritium-spiked nitrogen gas through the experimental detector was 21 counts/s. Hence, the detector counted ~63% of the betas that were estimated to contact the scintillator.

Conclusion

The $\text{CaF}_2(\text{Eu})$ scintillator in conjunction with the coincidence technique used in this investigation is a very stable and simple method for detecting low-energy beta radiation. In this feasibility experiment, most of the beta particles contacting the scintillator were detected; thus, it is obvious that the area of the scintillator is the only parameter that can be increased to improve the sensitivity. In order to

Fig. 4. Beta energy spectrum of tritium. Source: S. C. Curran, J. Angus, and A. L. Cockroft, "Investigation of Soft Radiations-II: The Beta Spectrum of Tritium," Department of Natural Philosophy, University of Glasgow, *Philosophical Magazine* 40, 1949.

Table 1. Beta activity in energy brackets

	Energy brackets		
	(keV)		
	2-8	8-12	12-18
Median energy (keV)	5.0	10.0	15.0
Median range (csda)* in nitrogen (mg/cm ²)	0.085 ^a	0.285 ^a	0.585 ^a
Median projected range (~1/2 median range) ^a	0.043	0.142	0.292
Penetration depth in nitrogen (cm) (median projected range/density at 23°C)	0.037	0.124	0.254
Fraction of total beta activity in bracket	0.48	0.20	0.12
Beta activity in bracket dis/s/mL (tritium conc. = 2.6 nCi/mL)	46.0	19.0	11.5
Only 50% of betas will penetrate the thickness of sample listed above β /s/mL (by definition of median range)	23.0	9.5	6.0
Geometry fraction--only 50% of betas migrate toward scintillator; therefore, specific activity in direction of scintillator = β /s/mL	11.5	5.0	3.0

*csda: Continuous slowing-down approximation.

detect low activity of 10^{-11} Ci/mL in air with 10% accuracy within 2 min., a very large area (2 m²) detector would be required.

Acknowledgments

I wish to express appreciation to the following people for their help and contributions to this project: M. L. Bauer for assigning the project to me and his continued support and technical discussions; C. E. Fowler, Jr., for deposition of the scintillator on the glass plates; V. C. Miller for designing the scintillator and source-mounting cans; R. A. Todd for his assistance in assembling the coincidence electronics; and W. L. Bryan for funding for this project under the NAVY/ORNL RADIAC Development Program.

References

- [1] G. A. Colman, "Systems Study for an Advanced Tritium Detector," thesis presented for the Master of Science Degree, The University of Tennessee, Knoxville, 1984.
- [2] "Stopping Powers for Electrons and Positrons," ICRU Report No. 37, pp. 70 and 132. October 1, 1984.
- [3] S. C. Curran, J. Angus, and A. L. Cockroft, "Investigation of Soft Radiations-II: The Beta Spectrum of Tritium," *Philosophical Magazine* 40, Department of Natural Philosophy, University of Glasgow, p. 55, 1949.

[4] "Linear Energy Transfer," *ICRU Report No. 18*, p. 30. June 15, 1970.

[5] R. D. Evans, *The Atomic Nucleus*, New York: McGraw-Hill, 1970, p. 613.

DISCLAIMER

This report was prepared as an account of work sponsored by an agency of the United States Government. Neither the United States Government nor any agency thereof, nor any of their employees, makes any warranty, express or implied, or assumes any legal liability or responsibility for the accuracy, completeness, or usefulness of any information, apparatus, product, or process disclosed, or represents that its use would not infringe privately owned rights. Reference herein to any specific commercial product, process, or service by trade name, trademark, manufacturer, or otherwise does not necessarily constitute or imply its endorsement, recommendation, or favoring by the United States Government or any agency thereof. The views and opinions of authors expressed herein do not necessarily state or reflect those of the United States Government or any agency thereof.