AIM

Association for Information and Image Management

1100 Wayne Avenue, Suite 1100
Silver Spring, Maryland 20910

301/587-8202

Centimeter

Inches

MANUFACTURED TO AIM STANDARDS
BY APPLIED IMAGE, INC.

1 of 1

Contributed paper for
10th Annual High Temperature Plasma Diagnostics Conference
Rochester, N.Y. May 8-12, 1994

**Methods for the Measuring the Surface Tritium
Inside TFTR using Beta Decay**

S.J. Zweben, D. Johnson, K. Hill, L.P. Ku, G. Lemunyan, D. Loesser,
D.K. Owens, S.S. Medley, D. Mueller, J. Timberlake

Princeton Plasma Physics Laboratory
Princeton, New Jersey 08543

and

W.R. Wampler, B.L. Doyle

Sandia National Laboratory
Albuquerque, New Mexico 85185

Abstract

Three potential methods for measuring the surface tritium content of the TFTR vacuum vessel are described, each based on a different technique for measuring the *in situ* beta emission from tritium. These methods should be able to provide both a local and a global assessment of the tritium content within the top $\approx 1\mu\text{m}$ of the inner wall surface.

DISCLAIMER

This report was prepared as an account of work sponsored by an agency of the United States Government. Neither the United States Government nor any agency thereof, nor any of their employees, makes any warranty, express or implied, or assumes any legal liability or responsibility for the accuracy, completeness, or usefulness of any information, apparatus, product, or process disclosed, or represents that its use would not infringe privately owned rights. Reference herein to any specific commercial product, process, or service by trade name, trademark, manufacturer, or otherwise does not necessarily constitute or imply its endorsement, recommendation, or favoring by the United States Government or any agency thereof. The views and opinions of authors expressed herein do not necessarily state or reflect those of the United States Government or any agency thereof.

MASTER

DISTRIBUTION OF THIS DOCUMENT IS UNLIMITED

Locating and recycling unburned tritium fuel from the inner walls of DT tokamaks like TFTR, JET and ITER is important since tritium is scarce and expensive, and since its in-vessel inventory needs to be minimized for radiological safety reasons. In this paper we describe three potential techniques for *in situ* measurements of the tritium residing on the inner surface layer of TFTR tokamak vacuum vessel wall.

Tritium produced from DD reactions has already been measured on graphite wall tiles removed from TFTR, JET and other tokamaks by counting the tritium beta emission using a PIN diode or channeltron detector[1]. Beta emission from tritium implanted in surfaces has also been measured by using a proximity-coupled phosphor screen and image intensifier[2], and by focusing the secondary electrons created by tritium betas using an electrostatic lens[3]. These methods are excellent for determining the surface tritium concentration of samples removed from the tokamak, but they can not (yet) be used in TFTR since the tritiated wall tiles can not be removed before the DT run is over.

[4, 5] Ideally, spatially resolved and global measurements of the surface tritium contamination should be made inside the TFTR vacuum vessel both during the DT experimental run and the subsequent tritium clean-up campaign. Strong spatial non-uniformities of the tritium concentration can be anticipated based on previous retention measurements of deuterium in TFTR[4] and tritium in JET[5]. For example, higher tritium concentrations are expected on the inner carbon bumper limiter (which contacts the plasma) than on the recessed metal walls. [6]

The betas emitted by tritium decay have a broad Fermi spectrum with a maximum energy of ~ 18 keV and an average energy of 5.6 keV. Since the maximum range of monoenergetic tritium betas in carbon is $\sim 3 \mu\text{m}$ [6], all tritium measurements made using beta detection are restricted to tritium within a surface layer $\sim 1 \mu\text{m}$ thick. Therefore, these measurements provide a lower limit to the tritium inventory inside the vessel. [7]

The tritium inventory inside the TFTR vessel is expected to be $\leq 2 \times 10^4$ Ci, depending on the tritium retention fraction and the number of neutral beam heated DT discharges. Assuming for the moment that all of this tritium lies within $\sim 1 \mu\text{m}$ of surface, and that the tritium is distributed uniformly over the vessel area of $\sim 10^6 \text{ cm}^2$, the averaged beta emission rate should be $\sim 10^8$ betas/ $\text{cm}^2\text{-sec}$, which corresponds to a current of $\sim 10^{-11}$ amps/ cm^2 or a power of $\sim 10^{-7}$ watts/ cm^2 . Spatial non-uniformities are likely to cause this rate to vary by at least an order of magnitude around this average value.

I. Beta Detector with Magnetic Steering

The principle of this method is illustrated in Fig. 1. A beta detector is inserted into the bottom of the vessel at the end of a movable probe drive. The horizontal and vertical magnetic fields of the normal tokamak are applied without a plasma and without any filling gas present. The betas emitted from the wall are "steered" by the magnetic field onto the probe head, allowing the area examined to be scanned poloidally and possibly toroidally (with an additional toroidal field component).

The betas emitted at the wall will spiral around the magnetic field and travel freely along the magnetic field toward the detector. The spatial resolution at the wall will be determined by the Larmor radius of the electrons in the applied magnetic field. The gyroradius ρ of an electron with a pitch angle χ with respect to a magnetic field B is $\rho(\text{cm}) = 3.4(100 \text{ Gauss}/B) \sin \chi$. Thus the viewing area at $B=100 \text{ G}$ is about $2\rho \approx 6 \text{ cm}$, which is suitable for examination of the $\approx 2 \text{ m}$ high inner bumper limiter. The available magnetic field range of 10-100 Gauss can be scanned during a standard TFTR shot cycle time of $\approx 8 \text{ sec}$ to create a map of the tritium concentration along the surface of the inner bumper limiter. a 106 u

The design of a prototype detector and probe "head" for use in TFTR is shown in Fig. 2. The aperture has a large solid angle ($\sim 2\pi \text{ str}$) to collect betas over the widest possible range of pitch angles. A biasable mesh grid over the aperture is used to repel low energy ($\leq 100 \text{ eV}$) secondary electrons, which may be made in the passage through the wall[3]. The first prototype detector will be an Amptektron channel electron multiplier with a built in high voltage power supply and pulse forming electronics[8]. This detector has no intrinsic energy resolution, so for the initial experiments the pulse signals will be converted into an analog signal with a time resolution $\approx 1 \text{ msec}$ and recorded with a transient digitizer. Later iterations may be made using a PIN diode detector, which has good energy resolution[1].

The expected beta flux steered toward this detector should depend only weakly on the magnetic field strength, since the magnetic fields will not focus or defocus the beta stream. However, the beta flux through the aperture will depend on the ratio of the electron gyroradius ρ to the aperture diameter " d ", the aperture solid angle Ω , and the angle of the field line. For the first TFTR prototype with $d \approx 0.1 \text{ cm} \ll \rho$ and $\Omega \sim 2\pi$, the collected beta flux per unit aperture area should be similar to that emitted directly at the wall area magnetically connected to the detector.

Thus the expected count rate in this detector is $\approx 10^6 \text{ counts/sec}$ for $d \approx 0.1 \text{ cm}$, which is near the limit for pulse counting. This detector has already been bench-tested using a small tritium beta source, which for calibration purposes can be mounted onto the probe head itself or near the detector in its retracted position in the shielded basement area. An additional *in situ* calibration may be obtained by introducing a low pressure

of tritium gas into the vessel while the magnetic fields are pointed into an area of low tritium beta emission; for example, a tritium gas level of ≈ 100 Ci at a pressure of $\leq 10^{-4}$ torr should be readily detectable.

This method would mainly be useful for determining the spatial distribution of the tritium on the wall, rather than the total tritium inventory in the vessel. The need to integrate these local signals over the whole wall area would most likely introduce at least a factor of two uncertainty in the total inventory measurement.

II. Beta Detection Using TFTR as an Ionization Chamber

Perhaps the method most likely to provide an accurate measurement of the *total* surface tritium inventory is based on an ionization chamber[7], which is the operating principle of most commercial tritium detectors. An ionization chamber works by collecting the electron/ion pairs made by any ionizing radiation in passing through a neutral gas; typically the ionization rate is ≈ 30 - 40 eV/ion, i.e. ≈ 200 ions/tritium beta. Tritium ionization chambers generally measure the small current (≈ 1 pA/ μ Ci) drawn between biased walls (≈ 10 - 20 volts) of a small chamber (≈ 1000 cm³) filled with ~~tritium~~ gas at near atmospheric pressure. *containing tritium*

Ideally, the TFTR vacuum vessel could be made into a large ionization chamber by biasing an electrode attached to a probe head similar to that in Fig. 1. The vacuum chamber would be filled with a standard tokamak gas such as deuterium to a pressure just large enough to stop the tritium betas, i.e. ≈ 10 Torr, in order to minimize both the gas load on the tritium recovery system and the ionization produced by background radiation. The expected ionization current would be $\approx 10^{-3}$ amps for a tritium inventory of ≈ 1000 Ci, which is easily measurable.

The bias voltage should be large enough to prevent recombination or diffusion of the ions, but small enough to avoid gas multiplication (to simplify calibration). The appropriate voltage depends on the type and pressure of the gas and on the chamber geometry[8], but is likely to be in the range of ≈ 10 - 100 volts. The current vs. voltage characteristic of the TFTR chamber will be measured directly to identify the plateau region characteristic of an ionization chamber.

This ionization current would be a simple measure of the total tritium content within the 1μ m surface of the vessel, and could be calibrated with samples of tritium gas as discussed in Sec. II. However, it is not clear that all of the tritium beta ionization would be collected by this electrode due to possible diffusion or recombination, to the small residual magnetic field inside the vessel (≈ 10 Gauss), and/or to the geometry of the vessel. It would therefore be useful to have several electrodes at different locations and with different geometries to cross-check this measurement.

III. Scintillator Detector Using a Filling Gas

A third possible method to measure tritium inside TFTR would be to fill the chamber with a scintillating gas and observe the light emission with discrete detectors or TV cameras. This technique could provide good spatial resolution if the pressure was raised high enough, e.g. the tritium beta range is ≈ 1 cm in deuterium at atmospheric pressure. Scintillation light emission from tritium beta decay in solid phosphors is used in exit signs (≈ 20 Ci) and watch dials (≈ 0.2 Ci).

The scintillation processes in gases is due to molecular excitation and subsequent photon emission. The efficiency is highest for inert gases and increases with their atomic number; for example, He and Ar make ≈ 1000 photons per 4.7 MeV alpha at atmospheric pressure[9]. Noble gases emit predominantly in the ultraviolet ≈ 300 -400 nm, but a small % of nitrogen mixed with a noble gas can shift its emission partially into the visible. This efficiency is $\approx 0.1\%$ for conversion of alpha particle energy into photons, and is presumably similar for betas.

SP. Thus the betas emitted from ≈ 1000 Ci of tritium inside the vessel would produce at least ≈ 10 mW of UV light, or $\approx 10^{-8}$ watts/cm², when stopping in a scintillating gas. This is a small but measurable level; for example, an intensified CCD camera has a threshold sensitivity of $\approx 3 \times 10^{-7}$ lux, or $\approx 5 \times 10^{-14}$ watts/cm², which should be able to image a point source of 10^{-9} watts at a distance of 1 m. For use with a UV the camera might be used with a fluorescent UV-visible converter just outside a vacuum window prior to imaging. Various existing TFTR diagnostics should also be able to measure this light level on discrete chords. If the tritium beta emission light is too low to be seen, the emission level could be increased by gas multiplication, although probably with a loss of spatial resolution[10].

how?
by applying a voltage?

IV. Radiation Backgrounds Inside the TFTR Vessel

Measurements of tritium beta decay activity inside the TFTR vessel must take into account the background due to the decay of residual radioactivity in the tokamak structure caused by the DT neutrons[11]. The primary residual decay product inside the vessel is γ -rays, which have a broad energy spectrum with a maximum ≈ 100 keV, with a peak energy ≤ 3 MeV depending on cooling times. The total γ -ray flux inside the vessel 1 day after a series of DT discharges producing $\approx 10^{19}$ neutrons is calculated to be $\approx 2 \times 10^5$ γ /cm²-sec, but is ≈ 10 times less than this after 1 month of cooling. Measurements of the dose rate outside the vessel ≈ 1 -2 months after a recent series of DT discharges producing $\approx 2 \times 10^{19}$ neutrons show a level of ≈ 35 mrem/hr, or equivalently $\approx 2 \times 10^4$ MeV/cm²-sec, which is consistent with these calculations.

These γ -rays can interact directly with a detector to produce background counts; however, the probability of such a direct interaction is low. For example, for the Amptektron detector the probability of counting an ≈ 2 keV x-ray is $\approx 2-3\%$ [6] and decreases with increased energy, so this background should be small compared to the estimated beta count rate of $\approx 10^5-10^6$ sec $^{-1}$, even at the end of the DT run ($\leq 5 \times 10^{20}$ neutrons). Similarly, the direct γ -ray background in the gas filled methods should be small due to the low ionization rate of the high energy gammas.

Activation γ -rays can also produce fast electrons on the inside surface of the vessel due to Compton scattering or the photoelectric effect, and non-tritium beta decays may also occur in the surface materials of the wall. The effective probability for creating Compton electrons is relatively small, e.g. < 0.3 electrons/ γ in carbon. The photoelectric effect is more efficient in creating electrons at low γ energy, but the probability of low energy electrons escaping from the inner surface of the vessel is smaller. Non-tritium beta decays should be negligible for the carbon inner wall where most of the tritium betas are expected, but some higher energy betas may come from decay of activation products in the metallic wall surfaces at other poloidal angles.

V. Tritium Below the Surface Layer

All the tritium beta decay measurements described above are limited to a surface layer ≈ 1 μ m thick from which tritium betas can escape. Since the tokamak wall can be eroded and co-deposited over a greater depth, depending on the number of high powered discharges, it would also be useful to make *in situ* measurements of the this sub-surface tritium. Previous surface analysis[4-5] suggests that $\approx 50-90\%$ of the tritium in the TFTR wall may ~~be~~ eventually reside in this sub-surface co-deposited layer by the end of the DT run.

Sub-surface tritium could be potentially be measured by a controlled evaporation of a selected surface layer, e.g. by the laser blow-off method. The tritium released could be measured by the beta decay methods above, or possibly by spectroscopic means[12] in the presence of an ohmic plasma.

For the particular application of evaluating the tritium inventory inside the TFTR vessel during the post-DT cleanup campaign, it may be sufficient to monitor the surface tritium levels to show that they are diminished substantially from its level at the end of the DT run, e.g. by a cleanup campaign with HeO glow discharges. After this period the extent of sub-surface tritium will also be evaluated by removing small samples of wall material for *in situ* analysis.

References:

- 1) W. R. Wampler and B.L. Doyle, submitted to Nucl. Inst. and Methods (1994).
- 2) T. Ida and Y. Ikebe, Nucl. Inst and Methods in Physics Research, **A253** (1986), 119.
- 3) M.E. Malinowski and R.A. Causey, J. Vac. Sci. Tech., **A6**(3) May/June 1988 (p. 2130-2135).
- 4) W.R. Wampler, B.L. Doyle, et al, J. Vac. Sci. Technol. **A6** (1988) 2111.
- 5) M. Ulrickson, et al, J. Vac. Sci. Technol. **A6** (1988) 2001; M. Ulrickson, Proc. IEA Large Tokamak Workshop on Tritium Experiments, 1992.
- 6) A. T. Peacock, et al, Proc. 17th Symposium on Fusion Technology, Rome, Italy, 1992, p. 329.
- 7) R.D. Evans, The Atomic Nucleus, McGraw-Hill. (New York 1955)
- 8) Amptektron Model MD-501, Amptek, Inc., Bedford, Massachusetts.
- 9) e.g. G.F. Knoll, Radiation Detection and Measurement, John Wiley and Sons, New York, 1979.
- 10) A.J.P.L. Policarpo, Space Science Instrumentation **3** (1977) 77.
- 11) L. Ku and J. Kolibal, Nuclear Technology/Fusion **4**, 586 (1983), or L. Ku and S. Liew, 15th Symposium on Fusion Engineering, Hyannis, Mass (1993).
- 12) C. Skinner, et al, Princeton Plasma Physics Laboratory Report PPPL 2993 (1994).

Acknowledgments: We thank R. Hawryluk, and K. M. Young for helpful comments and support for this work. This work was also supported by DOE contracts DE-AC02-CHO-3073 and DE-AC04-94AL85000.

Figure Captions

1) Schematic view of the TFTR vacuum vessel illustrating magnetic steering of tritium betas emitted from the inner wall onto a removable beta detector. The applied horizontal and vertical magnetic fields of ≈ 100 G. can be scanned to form a map of the surface tritium on the vessel inner wall. A similar probe mechanism could be used with a biased electrode to measure the ionization current made by the tritium betas in a neutral filling gas.

2) Detector geometry used for beta detection with magnetic steering. The Amptektron channeltron detector can count ≈ 10 keV betas with high efficiency up to a rate of $\approx 10^6$ counts/sec. The biased grid is used to suppress secondary electrons emitted from the wall, and the aperture is used to limit the beta countrate.

Figure 1

Figure 2

DATE

FILMED

6/14/94

END