RECEIVED BY OSTI SEP 24 1985

BNL 36903

CONF-850734 --2

HIGH FLUX PHOTON BEAM MONITOR

P. Mortazavi, M. Woodle, H. Rarback, D. Shu*
National Synchrotron Light Source
Brookhaven National Laboratory
Upton, New York 11973

M. Howells
Center for X-Ray Optics
Lawrence Berkeley Laboratory
Berkeley, California 94720

BNL--36903

DE85 018202

ABSTRACT

We have designed two photon beam position monitors for use on our x-ray storage ring beam lines. In both designs, a pair of tungsten blades, separated by a pre-determined gap, intercepts a small fraction of the incoming beam. Due to photoemission, an electrical signal is generated which is proportional to the amount of beam intercepted. The thermal load deposited in the blade is transferred by a heat pipe to a heat exchanger outside the vacuum chamber. A prototype monitor with gap adjustment capability was fabricated and tested at a UV beam line. The results show that the generated electrical signal is a good measurement of the photon beam position. In the following sections, design features and test results are discussed.

*Permanent address: Institute of High Energy Physics, Academia Sinica, Beijing, China.

DISCLAIMER

This report was prepared as an account of work sponsored by an agency of the United States Government. Neither the United States Government nor any agency thereof, nor any of their employees, makes any warranty, express or implied, or assumes any legal liability or responsibility for the accuracy, completeness, or usefulness of any information, apparatus, product, or process disclosed, or represents that its use would not infringe privately owned rights. Reference herein to any specific commercial product, process, or service by trade name, trademark, manufacturer, or otherwise does not necessarily constitute or imply its endorsement, recommendation, or favoring by the United States Government or any agency thereof. The views and opinions of authors expressed herein do not necessarily state or reflect those of the United States Government or any agency thereof.

MASTER

INTRODUCTION

Currently underway at the National Synchrotron Light Source (NSLS) are the design and construction of several beam lines which will use the intense radiation generated by insertion devices such as wigglers and undulators.

In each of these beam lines, accurate determination of the photon beam position is important. Since the distance from the photon source to the experimental station is relatively long (typically 15-20 meters), a small angular beam deviation at the source can produce a large movement at the experimental location. This problem is critical in the case of undulators where the usable beam dimension may be only a few millimeters. The signal from the beam position monitor can be used as a feedback signal to ring trim magnets. These in turn will alter the path of the electron beam in the insertion device to correct the photon beam error.

Because the beam intensity is so great, it must not be allowed to impinge on an uncooled surface. Hence, a reliable cooling system design is essential.

Two beam position monitors are under development: one for vertical measurement of a wiggler source and one for measurement of a soft x-ray undulator source. Both make use of thin tungsten blades which project into the fringes of the photon beam (receiving 3-4 percent of the peak intensity). Each blade is cooled by a heat pipe which transmits the heat outside the vacuum vessel. The heat pipe condensor end is then cooled by a small water heat exchanger. The blade dimensions have been selected so that in the worst case of exposure to the photon beam, the deposited thermal energy will be removed without the blade overheating or experiencing unacceptable thermal stress levels. Blade design was facilitated by a finite element analysis using ANSYS.

Because of low signal levels, it is necessary to electrically isolate the monitor subassembly from the cooling water. This is accomplished by use of a beryllia sleeve between the heat pipe and the cooling water.

DESIGN

The X1 beam line at the NSLS is primarily for soft x-ray imaging experiments and uses as its source a hybrid undulator, consisting of 37 periods of 80 mm length each[1]. The radiation produced by this device has extremely high brightness and is tunable throughout the soft x-ray range[2]. For the highest K parameter value (2.6) the beam has a peak power density of about three kilowatts per square milliradian, with a total power of 760 watts[3].

The horizontal and vertical full opening angles of the photon beam at $K = 2.6$ are about 0.9 mrad and 0.3 mrad, respectively. This produces a beam size at the position monitor, 9.2 meters from the source, of 9 mm wide by 3 mm high. Therefore, in the worst case conditions, a thin blade of 0.254 mm thickness would intercept about 9 percent of the total power or 68 watts. This high flux thermal load is transferred outside the vacuum by a heat pipe system (Fig. 1). The geometry of these blades has been selected so that the allowable stresses under thermal loading are not exceeded. The relatively small beam size and the optics for the X1 beam line require the beam be monitored in both horizontal and vertical directions. This is achieved by two separate monitors in one vacuum vessel (Fig. 2).

A transient thermal analysis using the ANSYS finite element program was performed. Our tests on the heat pipes have shown that the time to reach thermal temperature equilibrium is about 20 seconds (Fig. 3). For a heat input of 68 watts, and an elapsed time of approximately 20 sec, the transient thermal analysis (Fig. 4) indicates that the maximum blade temperature will be about 2700°F, far below melting point. A subsequent stress analysis indicated that, in case of no cooling, thermal stresses due to temperature gradient exceeded the yield strength of tungsten at an elevated temperature. However, this worst case conditions has practically been eliminated by use of heatpipe closed loop cooling system. Figure 5 shows a typical stress plot of the tungsten blade.

A commercial motorized manipulator will be used to drive the blade assemblies in two directions to "chase" the beam. Its relatively small bellows bore size requires the heatpipe/blade assembly be made in two pieces. Therefore, the thermal load must be transferred across unbonded contact surfaces in vacuum. The thermal conductance, h_c , between two adjacent materials is defined by Eq. (1). It normally takes place as a result of conduction across real contact points, conduction through entrapped fluid, and radiation through an interstitial gap (3).

$$h_c = h_s + h_f + h_r = 1/R_c = q/A \Delta T \quad (1)$$

where h_s , h_f , and h_r are conductance through solid points in contact; conductance through fluid and radiation conductance respectively. R_c is the total thermal contact resistance, q is the heat flow rate, A is the total contact area perpendicular to the direction of heat flow, and ΔT the temperature drop across the interface. In a vacuum, of course, the fluid conductance is zero.

Figure 6 shows the thermal resistance of copper to copper as a function of contact pressure in a vacuum environment[4]. At a contact pressure of about 30 lbs/sq.in., a thermal resistance across contact surfaces becomes almost identical to that of a solid piece. It should be noted that at an elevated temperature a significant portion of the deposited heat is lost through radiation and therefore reduces the thermal load on the heat pipe.

The X17 and the X25 beam lines both use wigglers whose maximum power density at the position monitor is approximately 4.3 kW/cm^2 . The beam need only be monitored in the vertical direction which in the worst case results in the deposition of approximately 41 watts on a blade. The drive mechanisms for these devices are single axis actuators and will also be commercial units. The design of these actuators allows the heat pipe/blade assemblies to be free of mechanical joints in vacuum (Fig. 7). In addition, the heat pipe may be removed from the device without breaking vacuum. A similar electrically insulated water cooled heat exchanger at the condenser end of the heat pipe removes the heat.

TEST RESULTS

We have tested the position monitor in the soft x-ray beam of beam line U15 at the NSLS. The output of the monochromator was about 1 mm high and 3 mm wide; sufficient beam was available from 10 to 50 A to produce an easily recordable signal. Although the beam size and spectral content are similar to what we expect at X1, the power loading is many orders of magnitude less. Each blade was biased negatively by a 67 volt battery and the photocurrent was measured by a Keithley 427 electrometer.

We were primarily interested in the sensitivity and reproducibility of the monitor. The noise was less than 10^{-12} A and the blade that was not in the beam produced a signal of the same magnitude. We stepped the blades vertically through the beam. Figure 8 shows the signal from one blade as it was moved 0.5 mm in 0.01 mm increments. Most of these jumps are recognizable and the signal is fairly linear with beam position.

Figure 9 better illustrates the sensitivity. It shows the ratio of the outputs over a few minutes. The jump occurs when we moved the blades only 20 microns in a single step. This sensitivity is more than adequate for our applications.

CONCLUSION

We have built and begun testing a high flux beam position monitor designed to safely take high heat loads. Preliminary tests indicate the monitor is highly sensitive. We plan to test its performance under heat load at an x-ray beam line in the near future.

ACKNOWLEDGEMENTS

We are grateful to the members of the X1 Insertion Device Team for guiding the design along. The NSLS is supported by the Department of Energy under contract DE-AC02-76CH00016.

REFERENCES

- [1] H. Rarback et al., "An Undulator Source Beam Line for Soft X-Ray Imaging", this conference.
- [2] C. Jacobsen, Undulator: A Primer for X1 Users, Preprint, Department of Physics, SUNY at Stony Brook (1985).
- [3] P. Mortazavi and D. Shu, Thermal Conductance of Metallic Interface in Vacuum, 1985 Particle Accelerator Conference, Vancouver (1985).
- [4] R.H. Norris, Ed., Heat Transfer and Fluid Flow Data Books, General Electric Co. 1977, p.18.

FIGURE CAPTIONS

- Fig. 1. X1 beam monitor.
- Fig. 2. X1 beam position monitor, general assembly.
- Fig. 3. Heat pipe's time response.
- Fig. 4. Blade temperature gradient.
- Fig. 5. Blade stress plot.
- Fig. 6. Thermal resistance vs. contact pressure.
- Fig. 7. X25 and X17 beam position monitor, general assembly.
- Fig. 8. Output current for position monitor stepping through beam.
- Fig. 9. Position monitor response to 20 μ m step.

Fig. 1

Fig. 2

Fig. 3

Fig. 4

Fig. 5

Fig. 6

Fig. 7

$I(nA)$