207
11-22-77

LA. 1607

UCRL-51873 Suppl I

CALORIMETRIC MEASUREMENTS OF LASER ENERGY AND POWER—1977 SUPPLEMENT

MASTER

Stuart R. Gunn

October 10, 1977

Prepared for U.S. Energy Research & Development
Administration under contract No. W-7405-Eng-48

LAWRENCE
LIVERMORE
LABORATORY
University of California, Livermore

DISTRIBUTION OF THIS DOCUMENT IS UNLIMITED

NOTICE

"This report was prepared as an account of work sponsored by the United States Government. Neither the United States nor the United States Department of Energy, nor any of their employees, nor any of their contractors, subcontractors, or their employees, makes any warranty, express or implied, or assumes any legal liability or responsibility for the accuracy, completeness or usefulness of any information, apparatus, product or process disclosed, or represents that its use would not infringe privately-owned rights."

NOTICE

Reference to a company or product name does not imply approval or recommendation of the product by the University of California or the U.S. Department of Energy to the exclusion of others that may be suitable.

Printed in the United States of America

Available from

National Technical Information Service

U.S. Department of Commerce

5285 Port Royal Road

Springfield, VA 22161

Price: Printed Copy \$: Microfiche \$3.00

Domestic		Domestic	
Page Range	Price	Page Range	Price
001 - 025	\$ 4.00	326 - 350	\$12.00
026 - 050	4.50	351 - 375	12.50
051 - 075	5.25	376 - 400	13.00
076 - 100	6.00	401 - 425	13.25
101 - 125	6.50	426 - 450	14.00
126 - 150	7.25	451 - 475	14.50
151 - 175	8.00	476 - 500	15.00
176 - 200	9.00	501 - 525	15.25
201 - 225	9.25	526 - 550	15.50
226 - 250	9.50	551 - 575	16.25
251 - 275	10.75	576 - 600	16.50
276 - 300	11.00	601 - up	- 1
301 - 325	11.75		

1/ Add \$2.50 for each additional 100 page increment from 601 pages up.

LAWRENCE LIVERMORE LABORATORY

University of California Livermore, California 94550

UCRL-51873 Supplement I

**CALORIMETRIC MEASUREMENTS OF LASER
ENERGY AND POWER—1977 SUPPLEMENT**

Stuart R. Gunn

MS. date: October 10, 1977

NOTICE

This report was prepared as an account of work sponsored by the United States Government. Neither the United States nor the United States Department of Energy, nor any of their employees, nor any of their contractors, subcontractors, or their employees, makes any warranty, express or implied, or assumes any legal liability or responsibility for the accuracy, completeness or usefulness of any information, apparatus, product or process disclosed, or represents that its use would not infringe privately owned rights.

CONTENTS

Abstract	1
Introduction	1
Calorimeter Designs	1
Disk Calorimeters	1
Cone Calorimeters	2
Hollow-Sphere Calorimeters	3
Volume-Absorption Calorimeters	3
Partial-Absorption Calorimeters	5
Miscellany	5
Reviews	5
Pyroelectric Meters	5
Test and Comparison Procedures	6
Absorptance Measurements	6
Black Surfaces	6
Chemical Actinometry	6
Classification	6
References	8

CALORIMETRIC MEASUREMENTS OF LASER ENERGY AND POWER—1977 SUPPLEMENT

ABSTRACT

The use of calorimeters for measuring laser output energies and powers is reviewed, primarily for the period 1975-1977 since preparation of earlier reviews.

INTRODUCTION

References 1-113 and 116-186 of this report were discussed in earlier reviews.^{114, 115, 187} The present review primarily covers papers that

have been published since; it also includes some earlier papers that were previously unavailable or overlooked.

CALORIMETER DESIGNS

DISK CALORIMETERS

Nakatsuka and Kubo¹⁸⁸ described a twin calorimeter for pulsed CO₂ lasers, somewhat similar to that of Jacob et al.¹²⁶ The 50 × 50 × 0.6 mm plates are of alumina ceramic rather than anodized aluminum; 20 thermocouple junctions are attached to the rear surface of each. The reflectivity was measured as <2% at 10.6 μm, and the response was found to be linear within 2% for 100-ns pulses up to 14 J/cm².

Kubarev et al.¹⁸⁹ described a twin calorimeter rather similar to the shallow inverted-cone-and-tube calorimeters of Geist, Schmidt, and Case¹²⁷ and of West and Case.¹²⁹ The laser beam strikes a shallow inverted cone at the end of a short tube; both are blackened. The calibrating heater is located on the rear surface of the shallow cone. A photoelectric radiometer with a spherical integrating sphere surrounding the beam at the entry to the tube is used to correct

for reflected radiation. Govor, Kubarev, and Obukhov¹⁹⁰ described another similar system in which the blackened shallow inverted cone is at the rear of the interior of a copper sphere. These designs are of course somewhat equivalent to hollow-sphere calorimeters, differing primarily in that most of the radiation is absorbed upon initially striking the inverted shallow cone.

Rubtsov¹⁹¹ calculated the temporal response of thin metal-plate calorimeters with thermocouples on the rear surface to laser pulses of specified shape.

Day, Gaddy, and Iversen¹⁹² studied thin-film thermopiles for the temporal resolution of infrared laser pulses. Film thicknesses of 0.1 to 0.2 μm were used, and response times down to 0.1 μs were obtained. Bar-Isaac and Korn¹⁹³ described a line array of thin-film thermocouples with junctions 5 μm apart for spatial resolution of focused laser beams. The technology of these devices is similar to that of other thin-film thermopiles applied to measurements of neutral beams^{194,195} and x rays.¹⁹⁶

Schimmel and Donaldson¹⁹⁷ calculated the performance of a system that absorbs short laser pulses in a thin metal film; the infrared radiation from the heated metal surface would be suppressed by placing over it a barrier, transparent at the

laser wavelength but opaque in the infrared.

Stevens¹⁹⁸ has given a general review of radiation thermopiles.

Chodzko et al.^{199,200} described a calorimeter for measuring the output power of an edge-coupled cw unstable resonator. Four water-cooled copper plates are mounted within the cavity between the mirrors; their position adjusts to define a variable rectangular aperture. The plates are coated with black Ebanol C oxide coating and grooved (thus being partially equivalent to a cone calorimeter) to absorb >98% of the DF chemical-laser radiation; the temperature rise of the flowing water is measured with thermocouples.

Durran²⁰¹ described a system for measurements of intracavity laser power by observing the power dissipation in the mirrors defining the laser cavity. The mirrors are water-cooled; the temperature rise at a known flow rate is measured with a pair of thermistors.

CONE CALORIMETERS

Edwards²⁰² described a high-accuracy cone calorimeter for pulsed lasers. The cone is of nickel-coated copper: the average temperature is sensed with a coil of nickel-alloy resistance wire wound over the whole outer surface. Less than 1% of the

incident laser radiation escapes from the cone; and this loss is monitored by a photodiode near the cone opening which looks at a diffusing reflector, in the form of a spherical segment surrounding the beam entry hole in the outer jacket, some distance in front of the cone. The instrument was compared with a National Bureau of Standards "C" series calorimeter¹³; agreement was within 1%.

Kuz'michev, Zinchenko, and Valitov²⁰³ analyzed the heat transport in a cone.

HOLLOW-SPHERE CALORIMETERS

Drong and Moeny²⁰⁴ described a system consisting of a hollow sphere of a subliming solid such as carbon dioxide; energy (or power) is to be derived from the change (or rate of change) in the weight of the sphere, as measured by a transducer and recorder.

Pierce²⁰⁵ described the use of a water-cooled hollow sphere to serve simultaneously as a calorimeter and as an integrating-sphere attenuator for another detector. The interior of the sphere is coated with a diffusely reflecting material of high reflectivity. A second hole is provided in the sphere to deliver a fraction of the energy to the second detector, which may be a fast-response type for pulse-shape measurement.

The attenuation factor between the input and output ports is readily measured.

Daehler²⁰⁶ described a liquid-helium-cooled hollow-sphere calorimeter used for infrared radiation and capable of measuring powers down to about 1 μ W.

VOLUME-ABSORPTION CALORIMETERS

Franzen and Schmidt²⁰⁷ gave a detailed description of a high-accuracy calorimeter for high-power pulses using solid volume absorbers. Neutral-density glasses are used as the absorbers at 1.06 μ m; several of these, and several crystalline dielectrics which could be used at 9 to 11 μ m, were checked for damage thresholds and bleaching effects. The calorimetric body is in the form of a shallow, slant-bottomed electroformed copper cup of square cross section. The first absorber is cemented to the bottom of the cup; the laser beam strikes it at an angle of 30° to the normal. The first-surface Fresnel reflection from this is directed to the second absorber, which is cemented to one wall of the cup; the angle of incidence on this is again 30° to the normal. The residual reflection from this is directed to black paint on the opposite wall. The temperature rise is measured with a 40-junction thermopile

whose junctions are distributed approximately uniformly on the outer surface of the cup; a calibration heater is wound on a spool on the outside of the slanted bottom, which bears the first absorber on the inside. A detailed error analysis is given.

Fisk and Gusinow²⁰⁸ described a liquid-absorption calorimeter utilizing a similar principle to reduce reflection losses. The liquid is circulated behind two fused-silica windows forming a 40° wedged cavity with an entry aperture of 18 × 20 cm. The beam strikes the first at 40° to the normal, and the first-surface reflection strikes the second at 10°. Operation is by the isoperibol, not flow, method; the absorbing liquid is recirculated by a small pump, and the temperature rise produced in the system by a laser pulse or immersed electrical calibrating heater is measured with a thermistor. Aqueous copper sulfate solution was used for 248-nm radiation and heavy water (normal water being too absorptive) at 3 μm.

Emmony and Bunn²⁰⁹ described a simple water-flow calorimeter for cw CO₂ lasers. The water flowed as a vertical film over an alumina plate perpendicular to the beam, and the temperature rise was measured with a pair of thermocouples. The reflectance is 0.8%. The absorption coefficient is very high, so the surface

of the water film is heated above the average water temperature; however, the calculated power was found to be independent of the flow rate within the precision of the measurements of flow rate and temperature rise, so heat loss by evaporation was evidently not a serious problem. Other heat losses, by conduction to the alumina plate, convection to the air, and evaporation from the average water temperature, would be largely proportional to the temperature rise at the small rises (2 to 5°C) used and hence would not be detected by varying the flow rate; the authors proposed that these could be largely eliminated by making the mean temperature of the water equal to ambient temperature.

Rice and Macomber²¹⁰ studied the properties of three types of commercially available filters for use as laser pulse attenuators. Two types of neutral density filter showed reversible bleaching (decrease in attenuation with increase of power level); a third type, consisting of colloidal carbon dispersed in gelatin, did not. The results are pertinent to volume-absorption calorimetry, which depends on the absorbing medium not showing an excessive change of absorptivity at the power densities being measured. However, other studies have shown no such saturation effects for various absorbers at 1.06 μm^{146,207} and 10.6 μm.⁹⁸

Various gases have been studied as attenuators and isolators for CO₂ laser wavelength²¹¹⁻²¹⁵; these would be applicable to gas-absorption calorimeters.

PARTIAL-ABSORPTION CALORIMETERS

Kuz'michev et al.^{216,217} described partial-absorption bolometers consisting of grids of fine nickel or platinum wire covering the whole beam

area and connected into a bridge circuit for measurement of the temperature rise. Use of two mutually perpendicular arrays of wire eliminates dependence of the response on the orientation of polarization of the beam. The response does depend on the reflectivity of the wire surface, so the system must be calibrated against an absolute calorimeter at the wavelength to be measured.

MISCELLANY

REVIEWS

Boyd²¹⁸ has briefly reviewed laser power and energy measurements, including calorimetry. Cunningham²¹⁹ has summarized commercially available calorimeters and other power meters. In some respects, the calorimetry of laser beams is related to the calorimetry of beams of x rays, gamma rays, and accelerated ions; a supplement to earlier reviews in this field has recently been published.²²⁰

PYROELECTRIC METERS

A number of papers^{180,221-233} have described the development and application of high-accuracy chopper-stabilized electrically-calibrated pyroelectric meters applicable to low-level cw and pulsed laser measure-

ments as well as general radiometry. The absorbing surface of gold black on the face of the pyroelectric materials also serves as the heater for electrical calibration in some versions; in others the heater is a separate metal film behind the absorbing layer.

Morozov, Nikolaev and Russov²³⁴ described a cylindrical pyroelectric detector surrounding a laser beam and intercepting the outer edge as the beam diverged. Baker²³⁵ described a peak-hold circuit permitting use of a moving-coil meter rather than an oscilloscope to record the peak reading of pyroelectric detectors used with pulsed lasers. Putley²³⁶ and Kremenchugskii and Noitsina²³⁷ gave general reviews of pyroelectric detectors.

TEST AND COMPARISON PROCEDURES

Thacher²³⁸ gave a detailed discussion of a procedure for calibrating pulse calorimeters with reference to cw power meters, using a timed interval of a cw beam. Falconer, Niland, and Turk²³⁹ and Neufeld *et al.*²⁴⁰ discussed the advisability of using splitters at low angles of incidence where the polarization of the beam being analyzed is unknown or variable.

ABSORPTANCE MEASUREMENTS

In this related field, calorimetric methods have been applied in a number of investigations of absorption coefficients of optical materials.²⁴¹⁻²⁵¹

BLACK SURFACES

Low-reflectivity coatings are useful for cw disc calorimeters and some

other types. Blevin and Geist²²⁶ and Peterson *et al.*²²⁷ evaluated gold-black and black paint coatings on pyroelectric detectors. Cuomo, Ziegler and Woodall²⁵² described preparation of dendritic tungsten surfaces of low reflectivity.

CHEMICAL ACTINOMETRY

Demas and coworkers have published several papers²⁵³⁻²⁵⁵ on this technique which, while in no sense calorimetric, may have certain advantages for measurements of laser energy in some situations.

CLASSIFICATION

In Table 1 the newly described calorimeters are classified according to the same scheme used earlier.¹⁸⁷

Table 1. Characteristics of laser calorimeters.

Reference	1	2	3	4	5	6	7	8
188	sur	blk	dis	tot	en	iso	tc	noh
189	sur	blk	cup	tot	pow	sst	tc	htr
190	sur	blk	cup	tot	pow	sst	tc	htr
192	sur	met	dis	tot	en	iso	tc	noh
193	sur	met	dis	tot	en	iso	tc	noh
199	sur	blk	dis	tot	pow	flo	tc	noh
201	sur	met	dis	par	pow	flo	tst	noh
202	sur	met	con	tot	en	iso	res	htr
204	sur	—	hol	tot	both	pha	—	noh
205	sur	—	hol	tot	pow	flo	—	noh
207	vol	sol	cup	tot	en	iso	tc	htr
208	vol	liq	cup	tot	en	flo	tst	htr
209	vol	liq	dis	tot	pow	flo	tc	noh
216	sur	met	—	par	pow	iso	bol	noh

REFERENCES

13. E. D. West, W. E. Case, A. L. Rasmussen, and L. B. Schmidt, *J. Res. Natl. Bur. Std.* A76, 13 (1972).
98. S. R. Gunn, *Studies of Calorimeter Absorbers for cw and Pulsed CO₂ Lasers*, Lawrence Livermore Laboratory, Rept. UCRL-51854 (1975).
114. S. R. Gunn, *A Review of Calorimetric Measurements of Laser Energy and Power*, Lawrence Livermore Laboratory, Rept. UCID-16214 (1972).
115. S. R. Gunn, *J. Phys. E: Sci. Instr.* 6, 105 (1973).
126. J. H. Jacob, E. R. Pugh, J. D. Daugherty, and D. B. Northam, *Rev. Sci. Instr.* 44, 471 (1973).
127. J. Geist, L. B. Schmidt, and W. E. Case, *Appl. Opt.* 12, (1973).
129. E. D. West and W. E. Case, *IEEE Trans. Instr. Meas.* IM-23, 422 (1974).
146. S. R. Gunn, *Rev. Sci. Instr.* 45, 936 (1974).
180. R. J. Phelan, Jr., R. L. Peterson, G. P. Klein, C. A. Hamilton, and G. W. Day, *Proc. Elect.-Opt. Syst. Des. Conf.*, 5th, New York 1973 (Industrial and Scientific Conference Management, Inc., Chicago, 1973), p. 117-123. *J. Curr. Laser Abstr.* 11, 33 (1974).
187. S. R. Gunn, *Calorimetric Measurements of Laser Energy and Power-1975 Supplement*, Lawrence Livermore Laboratory, Rept. UCRL-51873 (1975).
188. M. Nakatsuka and U. Kobo, *Japan. J. Appl. Phys.* 15, 1585 (1976).
189. A. V. Kubarev, A. S. Obukhov, I. N. Govor, and V. M. Nesterenko, *Izmeritel. Tekhn.* (8) 3 (1973). English transl. *Meas. Tech.* 16, 1113 (1973).
190. I. N. Govor, A. V. Kubarev, and A. S. Obukhov, *Izmeritel. Tekhn.* (7) 50 (1974). English transl. *Meas. Tech.* 17, 1059 (1974).
191. V. A. Rubtsov, *Telofiz. Vys. Temp.* 9, 154 (1971). English transl. *High Temp.* 9, 128 (1971).
192. G. W. Day, O. L. Gaddy, and R. J. Iversen, *Appl. Phys. Lett.* 13, 289 (1968).
193. C. Bar-Isaac and U. Korn, *Appl. Phys.* 6, 173 (1975).
194. J. E. Gardener and E. Harting, *J. Phys. E: Sci. Instr.* 5, 712 (1972).
195. L. E. Sharp, L. S. Holmes, P. E. Stott, and D. A. Aldcroft, *Rev. Sci. Instr.* 45, 378 (1974).
196. D. R. Ciarlo, H. A. Brown, T. A. Boster, and D. A. Blake, *IEEE Trans. Nucl. Sci.* NS-20, 545 (1973).

197. W. P. Schimmel, Jr., and A. B. Donaldson, *Thin Film Radiative-Pulse Calorimeter Using a Differentially Transparent Barrier*, Sandia Laboratories, Rept. SAND-74-0298 (1974). *Nucl. Sci. Abstr.* 31, 8113 (1975).
198. N. B. Stevens, *Semiconductors and Semi-Metals* 5, *Infrared Detectors*, R. K. Willardson and A. C. Beer, Eds. (Academic Press, New York, 1970), pp. 287-318.
199. R. A. Chodzko, S. B. Mason, R. R. Giedt, and D. A. Durran, *Appl. Opt.* 15, 2367 (1976).
200. R. A. Chodzko, U. S. Patent 3937079 (1976).
201. D. A. Durran, *Laser Calorimeter*, The Aerospace Corporation, Rept. SAMSO-TR-74-252 (1974). *Nucl. Sci. Abstr.* 32, 456 (1975).
202. J. G. Edwards, *J. Phys. E: Sci. Instr.* 8, 663 (1975).
203. V. M. Kuz'michev, N. I. Zinchenko, and R. A. Valitov, *Radiotekhnika* (29) 135 (1974). (In Russian). *Phys. Abstr.* 78 60176 (1975).
204. M. C. Drong and W. M. Moeny, U. S. Patent 3908457 (1975).
205. R. L. Pierce, *Laser Focus* 11, (11) 62 (1975).
206. M. Daehler, *Calorimetric Method of Sensitivity Calibration for Far-Infrared Detectors*, Naval Research Lab. Interim Rept. (1976). *J. Current Laser Abs.* 13, 287 (1976).
207. D. L. Franzen and L. B. Schmidt, *Appl. Opt.* 15, 3115 (1976).
208. G. A. Fisk and M. A. Gusinow, *Rev. Sci. Instr.* 48, 118 (1977).
209. D. C. Emmony and J. C. Bunn, *J. Phys. E: Sci. Instr.* 9, 621 (1976).
210. R. O. Rice and J. D. Macomber, *Appl. Opt.* 14, 2203 (1975).
211. F. G. Gebhardt, J. H. McCoy and D. C. Smith, *IEEE J. Quantum Electron*, QE-5, 471 (1969).
212. R. W. MacPherson, *Rev. Sci. Instr.* 45, 316 (1974).
213. T. G. Roberts, *Rev. Sci. Instr.* 47, 257 (1976).
214. B. D. Green and J. I. Steinfeld, *Appl. Opt.* 15, 1688 (1976).
215. P. Lavigne, J. L. Lachambre, and J. Gilbert, *Appl. Phys. Lett.* 28, 265 (1976).
216. V. M. Kuz'michev and M. P. Perepechai, *Kvant. Elektron.* 1 2407 (1974). English transl. *Sov. J. Quant. Electron.* 4, 1337 (1975).
217. V. M. Kuz'michev, V. M. Latynin, and I. A. Priz., *Prib. Tekh. Eksp.* 17, (2) 190 (1974). English transl. *Instr. Exp. Tech.* 17, 525 (1974).
218. H. S. Boyne, *Proc. Soc. Photo-Opt. Instrum. Engr.* 69 90 (1975).
219. R. Cunningham, *Electro-Opt. Syst. Des.* (9) 33 (1976).
220. S. R. Gunn, *Nucl. Instr. and Meth.* 135, 251 (1976).

221. R. J. Phelan, Jr., R. J. Mahler, and A. R. Cook, *Appl. Phys. Lett* 19, 337 (1971).
222. J. Geist, *Rev. Sci. Instr.* 43, 1704 (1972).
223. J. Geist, *Appl. Opt.* 12, 907 (1973).
224. R. J. Phelan, Jr., and A. R. Cook, *Appl. Opt.* 12, 2494 (1973).
225. J. Geist and W. R. Blevin, *Appl. Opt.* 12, 2532 (1973).
226. W. R. Blevin and J. Geist, *Appl. Opt.* 13, 1171 (1974).
227. R. L. Peterson, G. W. Day, P. M. Gruzensky, and R. J. Phelan, Jr., *J. Appl. Phys.* 45, 3296 (1974).
228. G. W. Day, C. A. Hamilton, R. L. Peterson, R. J. Phelan, Jr. and L. O. Mullen, *Appl. Phys. Lett.* 24, 456 (1974).
229. C. A. Hamilton and G. W. Day, *A Pyroelectric Power Meter for the Measurement of Low Level Laser Radiation*, National Bureau of Standards, Rept. NBS-TN-665 (1975).
230. C. A. Hamilton, R. J. Phelan, Jr., and G. W. Day, *Opt. Spectra* 9, (10) 37 (1975).
231. J. Geist, B. Steiner, R. Schaefer, E. Zalewski and A. Corrons, *Appl. Phys. Lett.* 26, 309 (1975).
232. Anon., *Opt. Spectra* 9, (1) 32 (1975).
233. J. Geist, H. J. Dewey, and M. A. Lind, *Appl. Phys. Lett.* 28, 171 (1976).
234. B. N. Morozov, E. P. Nikolaev, and V. M. Russov, *Opt. and Spektrosk* 38 382 (1975). English transl., *Opt. and Spectrosc.* 38, 214 (1975).
235. H. J. Baker, *J. Phys. E: Sci. Instr.* 8, 261 (1975).
236. E. H. Putley, *Semiconductors and Semi-Metals* 5, *Infrared Detectors*, R. K. Willardson and A. C. Beer, Eds. (Academic Press, New York, 1970), pp. 259-285.
237. L. S. Kremenchugskii and O. V. Roitsina, *Prib. Tekh. Eksp.*, (3) 7 (1976). English transl., *Instr. Exp. Tech.* 19, 603 (1976).
238. P. D. Thacher, *Appl. Opt.* 15, 1815 (1976).
239. I. S. Falconer, R. A. Niland, and M. I. Turk, *J. Phys. E: Sci. Instr.* 8, 216 (1975).
240. C. R. Neufeld, G. Saint-Hilaire, P. Noël, A. Robert, Z. Szili, and M. Gagné, *Rev. Sci. Instr.* 48, 697 (1977).
241. B. Bendow, A. Hordvik, H. G. Lipson, and L. Skolmick, *Optical Evaluation of Carbon Dioxide Laser Window Materials at Air Force Cambridge Research Labs*, USNTIS, AD Rept. 749864 (1972). *Chem. Abs.* 78, 90844 (1973).

242. L. H. Skolnik, *A Review of Techniques for Measuring Small Optical Losses in Infrared Transmitting Materials*, pp 405-443 in *Optical Properties of Highly Transparent Solids*, S. S. Mitra and B. Bendow, Eds. (Plenum Press, New York, 1975).
243. D. A. Pinnow and T. C. Rich, *Appl. Opt.* 12, 984 (1973).
244. D. A. Pinnow and T. C. Rich, *Appl. Opt.* 14, 1258 (1975).
245. A. Zaganiaris, *Appl. Phys. Lett.* 25, 345 (1974).
246. E. G. Bernal, *Appl. Opt.* 14, 314 (1975).
247. M. Hass, J. W. Davisson, H. B. Rosenstock, and J. Babiskin, *Appl. Opt.* 14, 1128 (1975).
248. W. E. K. Gibbs and A. W. Butterfield, *Appl. Opt.* 14, 3043 (1975).
249. T. J. Magee, N. M. Johnson, M. Lehmann, J. Peng, and J. Hannigan, *Rev. Sci. Instr.* 47, 301 (1976).
250. D. Bois, *Rev. Phys. Appl.* 11, 293 (1976).
251. G. Dube, *Appl. Opt.* 16, 1484 (1977).
252. J. J. Cuomo, J. F. Ziegler, and J. M. Woodall, *Appl. Phys. Lett.* 26, 557 (1975).
253. J. N. Demas, E. W. Harris, and R. P. McBride, *A Photooxygenation Actinometer for Laser Intensity Measurement in Lasers in Physical Chemistry and Biophysics*, J. Joussot-Dubin, Ed. (Elsevier, Amsterdam, 1975).
254. J. N. Demas, *Laser Intensity Measurements by Chemical Actinometry in Creation and Detection of the Excited State* 4, W. R. Ware, Ed. (Dekker, New York, 1976).
255. J. N. Demas, R. P. McBride and E. W. Harris, *J. Phys. Chem.* 80, 2248 (1976).