**Idaho
National
Engineering
Laboratory**

*Managed
by the U.S.
Department
of Energy*

**EGG-WTD-10080
January 1992**

APR 27 1992

**Rapid Monitoring of Soil, Smears,
and Air Dusts by Direct Large-Area
Alpha Spectrometry**

Claude W. Sill

*Work performed under
DOE Contract
No. DE-AC07-76ID01570*

This document contains new concepts or the author(s) interpretation of new calculations and/or measurements; accordingly, EG&G Idaho, Inc. is required by the United States Government to include the following disclaimer:

DISCLAIMER

This report was prepared as an account of work sponsored by an agency of the United States Government. Neither the United States Government nor any agency thereof, nor any of their employees, makes any warranty, express or implied, or assumes any legal liability or responsibility for the accuracy, completeness, or usefulness of any information, apparatus, product or process disclosed, or represents that its use would not infringe privately owned rights. References herein to any specific commercial product, process, or service by trade name, trademark, manufacturer, or otherwise, does not necessarily constitute or imply its endorsement, recommendation, or favoring by the United States Government or any agency thereof. The views and opinions of authors expressed herein do not necessarily state or reflect those of the United States Government or any agency thereof.

EGG-WTD--10080

DE92 012528

Rapid Monitoring of Soil, Smears, and Air Dusts by Direct Large-Area Alpha Spectrometry

Claude W. Sill

Published January 1992

**Idaho National Engineering Laboratory
EG&G Idaho, Inc.
Idaho Falls, Idaho 83415**

**Prepared for the
U.S. Department of Energy
Office of Environmental Restoration and Waste Management
Under DOE Idaho Field Office
Contract DE-AC07-76ID01570**

MASTER

DISTRIBUTION OF THIS DOCUMENT IS UNLIMITED

Rapid Monitoring of Soil, Smears, and Air Dusts by Direct Large-Area Alpha Spectrometry

Approved by:

Kenneth Merrill
Buried Waste Integrated Demonstration Coordinator

1/20/92
Date

Kerry L. Martin, Unit Manager
Nuclear Sciences and Radiochemistry

1-17-92
Date

Prepared by:

Claude W. Sill, Principal Scientist
Nuclear Sciences and Radiochemistry

1-14-92
Date

Reviewed by:

Bruce J. Mincher, Senior Scientist
Nuclear Sciences and Radiochemistry

1/14/92
Date

ABSTRACT

Experimental conditions to permit rapid monitoring of soils, smears, and air dusts for transuranic (TRU) radionuclides under field conditions are described. The monitoring technique involves direct measurement of alpha emitters by alpha spectrometry using a large-area detector to identify and quantify the radionuclides present. The direct alpha spectrometry employs a circular gridded ionization chamber 35 cm in diameter which accommodates either a circular sample holder 25 cm in diameter or a rectangular one 20 by 25 cm (8 by 10 in.). Soils or settled dusts are finely ground, suspended in 30% ethanol, and sprayed onto a 25-cm stainless steel dish. Air dusts are collected with a high-volume sampler onto 20- by 25-cm membrane filters. Removable contamination is collected from surfaces onto a 20- by 25-cm filter using an 18-cm (7-in.) paint roller to hold the large filter in contact with the surface during sample collection. All three types of samples are then counted directly in the alpha spectrometer and no other sample preparation is necessary. Some results obtained are described.

The lower limit of detection (LLD) for most alpha emitters for a 10-min count on a 2-gram sample of soil was about 20 pCi/g. For air samples, the LLD varied from about $1 \text{ E-}12 \mu\text{Ci/mL}$ for a 10-min sample and a 10-min count immediately after completion of sample collection without waiting for decay, to $2 \text{ E-}15 \mu\text{Ci/mL}$ for a 5-hour sample, a 4-day decay time and a 2-hour count. For smears, a 10-min count over a 32-in. by 48-in. area gave an LLD of about 5 pCi of removable activity on the filter or 0.0005 pCi/cm^2 in terms of average concentration collected.

The LLD for a 0.1-gram soil sample is 15 and 40 pCi/g for 60- and 10-min spectra, respectively. The overall counting efficiency for a 0.1-gram soil sample is about 20%, including all physical losses, attenuation, and scattering. This permits quantitative determination of the TRUs, to be made in soil and settled dusts, with an accuracy almost as good as the statistics of the measurement will permit.

EXECUTIVE SUMMARY

Some radioactive wastes containing transuranic (TRU) elements have been stored at the Idaho National Engineering Laboratory (INEL) in the past by shallow land burial. Retrieval and disposition of this waste is one of the options being considered and will require continuous monitoring of the air, dusts, and soil for alpha emitters to ensure that safety limits are not exceeded. Direct measurements by alpha spectrometry are particularly promising because of their potential speed, sensitivity, and unequivocal identification of TRU radionuclides under field conditions during waste handling and retrieval operations. The technique involves direct measurement of the sample by alpha spectrometry in a large pressurized gridded ionization chamber 35 cm in diameter. Soil samples or settled dusts are thoroughly ground, suspended in 30% ethanol, sprayed onto circular stainless steel pans 25.4 cm in diameter, and dried under an infrared lamp. Air dusts are collected with a high-volume air sampler on 20- by 25-cm membrane filters with 0.8- μ m pores. Smear samples are also collected on 20- by 25-cm membrane filters using an 18-cm paint roller to hold the filter against the surface being smeared. The samples are then counted directly in the spectrometer and no other sample preparation is required.

When the spectra are well-resolved, virtually all alpha-emitting radionuclides can be identified by the energies of their individual alpha emissions, which are generally unique, and quantified by integrating the areas under the respective alpha peaks. However, there are three major problems that are encountered in the direct spectrometry of soils and dusts. First, the range of alpha particles in soils is only about 30 μ m. If sufficient sample is taken to give even moderate sensitivity in the short counting times available, resolution will be poor and accuracy will be limited because of uncertainties in the effective sample size and counting efficiencies. Secondly, there are fourteen separate alpha emitters in the ^{238}U and ^{232}Th chains, all of which occur naturally in the earth's crust at levels of 1 to 2 pCi/g. Although these alpha emitters will not be resolvable at the small sample sizes and short counting times permitted, they contribute a continuous distribution of degraded pulses on top of which the TRUs must be detected. Thirdly, the radon and thoron daughters in the air, particularly $^{214}\text{Pb-Bi-Po}$ and $^{212}\text{Pb-Bi-Po}$, attach themselves to the sample by both direct impaction and electrostatic attraction during sample collection, and cause a severe increase in the background against which the TRUs must be detected. This background can be virtually eliminated if sufficient time is allowed for decay, but generally such time is not available during field operations. These problems are all minimized by using a large-area detector so that a relatively large sample can be spread as thin as possible to give adequate resolution and sensitivity for the purpose in the presence of the indigenous activities.

In the present work, the effect of pulse-shaping time, chamber pressure, distribution of the radioactivity in the source relative to the detector, and other experimental conditions on the resolution, shape, position, and linearity of the peaks in the spectra obtained using the large gridded detector was determined. Methods of collecting large samples of air dusts and smears quickly, and for preparing samples of soil under conditions to permit direct insertion into the spectrometer were developed. Spectra of backgrounds and samples containing ^{239}Pu and ^{241}Am of each type are presented with detection limits calculated for each. The equations for calculating lower limits of detection, concentrations of activity, and the propagation of statistical uncertainties are also given.

The lower limit of detection (LLD) for alpha emitters for a 10-min count on a 2-gram sample of soil was about 20 pCi/g. For air samples, the LLD varied from about $1 \text{ E-}12 \mu\text{Ci/mL}$ for a 10-min sample and a 10-min count immediately after completion of sample collection without

waiting for decay, to 2×10^{-15} $\mu\text{Ci/mL}$ for a 5-hour sample, a 4-day decay time, and a 2-hour count. For smears, a 10-min count over a 32-in. by 48-in. area gave an LLD of about 5 pCi of removable activity on the filter or 0.0005 pCi/cm² in terms of average concentration collected.

The LLD for a 0.1-gram soil sample is 15 and 40 pCi/g for 60- and 10-min spectra, respectively. The overall counting efficiency for a 0.1-gram soil sample is about 20%, including all physical losses, attenuation, and scattering. This permits quantitative determination of the TRUs to be made in soil and settled dusts with an accuracy almost as good as the statistics of the measurement will permit.

CONTENTS

ABSTRACT	iii
EXECUTIVE SUMMARY	v
ACRONYMS	xi
INTRODUCTION	1
BACKGROUND	2
INSTRUMENTATION AND APPARATUS	4
Instrumentation	4
Instrument Calibration	4
Preparation of Large High-Resolution Calibration Source	6
Apparatus	7
EXPERIMENTAL TECHNIQUES	9
Operation of the Spectrometer	9
Effect of Experimental Conditions	9
Sample Collection and Preparation	16
Soil and Settled Dusts	16
Air Dusts	18
Smear Samples	18
EXPERIMENTAL RESULTS	20
Backgrounds	20
Counting Efficiency	29
Calculation and Interpretation of Results	34
Propagation of Statistical Uncertainties	35
QUANTITATIVE MEASUREMENTS	38
CONCLUSIONS	40
REFERENCES	41

FIGURES

1. Gas and vacuum connections to ionization chamber.	5
2. Large three-peak standard evaporated uniformly over a 25.4-cm diameter stainless steel plate.	8
3. Small three-peak electrodeposited standard, centered under detector, and counted at atmospheric pressure and 4 μ sec shaping time.	12
4. Small three-peak electrodeposited standard at edge of detector, and counted at atmospheric pressure and 1 μ sec shaping time.	13
5. Instrument background, 60-min count	21
6. Soil background, 60-min count	22
7. Air background, 10-min count	24
8. Smear sample of dirty bench top, 10-min count	25
9. Two grams of soil containing 20 pCi/g each of ^{239}Pu and ^{234}Am , 10-min count	26
10. Same sample as in Figure 9, 60-min count	27
11. Air sample collected at 16 cfm for 10 minutes containing approximately 2 pCi each of ^{239}Pu and ^{241}Am	28
12. Same sample as in Figure 11; spectrum taken 49 minutes later after decay of about half of the ^{214}Po	30
13. Air sample collected at 24 cfm for 10 min with approximately 10 mg of 103 pCi/g ^{239}Pu soil, sample allowed to decay for 25 minutes and counted for 1 hour	31
14. Same sample as Figure 13 but only a 10-min count after further decay of two hours to decrease the ^{214}Po tail	32
15. Air sample collected at 24 cfm for 10 minutes, allowed to decay for 92 hours, and then counted for two hours	33
16. Standard soil, 100 mg of 1030 ± 14 pCi/g ^{239}Pu 60-min count.	39

TABLES

1.	Effect of shaping time, P-10 overpressure, and distribution of activity on resolution of alpha spectra (source centered under detector)	10
2.	Effect of shaping time, P-10 overpressure, and distribution of activity on resolution of alpha spectra (source at left edge of detector)	14
3.	Effect of shaping time, P-10 overpressure, and distribution of activity on resolution of alpha spectra (3-peak evaporated source)	15
4.	Effect of shaping time, P-10 overpressure, and distribution of activity on resolution of alpha spectra (evaporated ²⁴¹ Am source)	17

ACRONYMS

DAC	derived air concentration
DOE	Department of Energy
FWHM	full width at half maximum
INEL	Idaho National Engineering Laboratory
LLD	lower limit of detection
TRU	transuranic

Rapid Monitoring of Soil, Smears, and Air Dusts by Direct Large-Area Alpha Spectrometry

INTRODUCTION

Wastes of all kinds containing transuranic (TRU) elements have been stored at numerous Department of Energy (DOE) sites for many years. At the Idaho National Engineering Laboratory (INEL), some of this waste has been stored by shallow land burial which is now unacceptable by today's standards. Retrieval and disposition of this waste is one of the options being considered and will require continuous monitoring of the air, dusts, and soil for alpha emitters to ensure that operational and safety radiological limits are not exceeded. Rapid turnaround time of the analyses is of considerable importance so that the retrieval operations can proceed with minimal work stoppages due to the time required to obtain results of radiochemical analyses. Direct measurements by alpha spectrometry are particularly promising because of their potential speed, sensitivity, and unequivocal identification of TRU radionuclides under field conditions during waste handling and retrieval operations. In contrast, procedures requiring complete sample dissolution and chemical separations are generally more elaborate and time-consuming.

This report describes background, instrumentation, and experimental techniques and results for rapid monitoring of soil, smears, and air dusts for TRU radionuclides. The technique involves direct alpha spectrometry in a large pressurized gridded ionization chamber 35 cm in diameter. Soil samples or settled dusts are thoroughly ground, suspended in 30% ethanol, sprayed onto circular stainless steel pans 25 cm in diameter, and dried under an infrared lamp. The principal precaution necessary is to ensure sufficiently small particle size and sample thickness so that the alpha particles can reach the detector with as little energy degradation as possible. Air dusts are collected with a high-volume air sampler on 20- by 25-cm membrane filters with 0.8- μ m pores. Smear samples are also collected on 20- by 25-cm membrane filters using an 18-cm paint roller to hold the filter against the surface being smeared. The samples are counted directly in the spectrometer and no other sample preparation is required.

BACKGROUND

When well-resolved spectra are obtained, virtually all alpha-emitting radionuclides can be identified by the energies of their individual alpha emissions, which are generally unique, and quantified by integrating the areas under the respective alpha peaks. However, there are three major areas that must be considered for direct application of alpha spectrometry to soils and dusts. First, the range of most alpha particles in soil is only about 30 μm . Consequently, only those alphas emitted near the surface will be able to escape from the sample to the detector. For even moderate resolution with alpha spectrometry, deposits of thickness less than approximately 100 $\mu\text{g}/\text{cm}^2$ are required. Thus, detectors of large areas must be used if sufficiently large samples are to be used to obtain even moderate sensitivity without chemical dissolution and separations. For example, at a sample thickness of 100 $\mu\text{g}/\text{cm}^2$, a circular area 25.4 cm in diameter of 507 cm^2 would accommodate only 0.051 gram of sample, causing a severe limitation in the maximum sensitivity and precision that can be achieved, particularly when time is critical and counting times must be kept short. Even if the limitation on average sample thickness is met, the size of the individual particles must also be kept below a few micrometers in diameter or degradation will still result. This is no particular problem with air dusts which are already very small in size and the thickness can be controlled by the sample size. Considerable grinding or sieving will be required with soils or settled dusts to achieve good resolution. Because of consequent uncertainties in counting efficiencies obtained under a given set of conditions, the accuracy of the direct measurements will not be good when high sensitivity is desired unless the smaller samples are used.

Secondly, there are eight alpha emitters in the ^{238}U chain and another six in the ^{232}Th chain, all of which occur naturally in the earth's crust at levels of 1 to 2 pCi/g. This amounts to a total natural alpha activity of 14 to 28 pCi/g and 14 separate alpha peaks from which excess activity due to the TRU nuclides must be distinguished if the sensitivities are comparable. However, because of the limitation in sample size that can be employed in a 25-cm detector, and the counting times permitted for the present application, the sensitivity will not be great enough to require their separate identification. The main problem from indigenous activity will be from the background tail against which the TRU activity must be detected.

Thirdly, the air contains relatively high activities of ^{222}Rn and its short-lived alpha- and beta-emitting daughters, ^{218}Po and $^{214}\text{Pb-Bi-Po}$, from the natural ^{238}U chain, and ^{220}Rn and its daughters $^{212}\text{Pb-Bi-Po}$ from the natural ^{232}Th chain, that are produced by decay of ^{226}Ra and ^{224}Ra in the earth, construction materials, etc. These daughters are produced, and most of them decay, as positively-charged ions that attach themselves to the samples to be counted by both direct impaction and electrostatic attraction, particularly on non-conducting surfaces. The resultant alpha activity can be very large compared to the levels of TRUs of interest, and is variable due to short-lived growth and decay so that no realistic correction can be applied, and thus interferes seriously with sample preparation, counting, and interpretation of the results.

For much of the present work, it was assumed that TRUs would not be present in detectable activities in soils, and that the analytical procedure would be used, therefore, primarily as a screening device. Consequently, it was preferable to use as large a sample as possible to achieve maximum sensitivity in the short counting times permitted, than to try to achieve high resolution at activity levels too low for the interfering natural activities to be detected. The larger quantities of mass will also increase the uncertainties in the effective sample size and counting efficiency; however, the ability to detect the presence of TRU activity was deemed more important than the accuracy with

which that measurement was made. The detection limit can be determined unambiguously for a given set of conditions without regard to counting efficiencies and effective sample size by counting a standard soil containing the expected minimum activity under the exact conditions to be used. When accuracy becomes more important than sensitivity, smaller samples and longer counting times will have to be used.

Except for radon daughters, air dusts and smears will generally contain even smaller quantities of the other nonvolatile natural activities than soils because of the smaller quantity of soil present. However, radon and thoron daughters will be collected in air dust samples in high activities with high-volume equipment, even in short sampling times, and will produce severe interference because time cannot be allowed to permit significant decay even of the 40-min $^{214}\text{Pb-Bi-Po}$ chain, much less the 10.64-hr $^{212}\text{Pb-Bi-Po}$ chain. The conditions of sample collection and analysis will have to be chosen to minimize the collection of radon daughters, and the detection limit obtained can then be calculated from the background activity collected in the energy region of interest for the conditions used. When time becomes available, much larger samples, longer counting times and longer decay times can be used to provide much higher sensitivity.

INSTRUMENTATION AND APPARATUS

Instrumentation

A custom-built gridded ionization chamber (Ordela, Inc., Oak Ridge, TN, Model 8200A) similar to those developed by others,¹⁻³ a charge-sensitive preamplifier, a linear amplifier, and a power supply were used with a Canberra-Nuclear Data Accuspec system composed of a 16K pulse-height analyzer, an IBM PS/2 model 30-286 computer with a 30 MB hard disk drive for data storage, and a Hewlett-Packard Laserjet III printer. To permit field operation immediately adjacent to the waste burial area being remediated, and to decrease the effect of radon daughters from the air in laboratories containing concrete and cinderblock construction, the instrument was set up in a mobile laboratory free of concrete materials.

The gridded ionization chamber was housed in a steel cylinder, 35 cm in diameter by 24 cm in height. The walls and top and bottom plates were made of nickel-plated steel, 9-mm thick, and the top and bottom plates were each held by 12 3/8-in. bolts to a flange welded to the cylinder. This construction will easily accept the 2-atmosphere pressure finally used. The counter chamber was 35 cm in diameter by 15 cm in height with a volume of 14.4 liters. A sample holder was positioned reproducibly inside the chamber by fixed guides, and a circular sample pan of mirror-finish stainless steel was placed in an indent in the sample holder; both were inserted into the ionization chamber through a sealable access port. The sample pans were spun from 0.86-mm mirror-finish stainless steel sheets with inside dimensions of 25.4 cm in diameter, 3 mm in height, and an area of 507 cm². The sample holder also had a cutout to accept a 20- by 25-cm membrane filter held flat by a steel hold-down plate with 20- by 25-cm outside dimensions and an opening 18 by 23 cm wide to match the effective filtration area of the high-volume sampler.

The counting gas was P-10 gas (90% Ar-10%CH₄) used without further purification from a commercial cylinder pressurized at 2200 psi. The gas was released from the cylinder through a two-stage regulator; the second stage was set at 16 psi to permit rapid filling of the chamber while minimizing the chances for significant overpressure to occur. A small gas shut-off valve was placed immediately after the regulator to control the flow of gas out of the cylinder. The gas then passed through a 0.45- μ m Acro-50 filter disk (Gelman Sciences, Inc., Ann Arbor, MI), to filter any particulate material that might be present. The gas and vacuum connections to the ionization chamber are shown in Figure 1. The pressure-relief valve opened at about 6 psi above atmospheric pressure. The small gas shut-off valves controlled flow of P-10 gas into the chamber, and evacuation of the chamber by a Leybold-Heraeus vacuum pump. The two gauges were a 2-in. vacuum gauge reading from 1 to 30 inches of mercury, and a 2-in. pressure gauge reading from 0 to 30 psi. All connections were 1/4-in. NPT threads.

Instrument Calibration

Calibration standards containing three radionuclides having principal energies as widely separated as possible were used in the present investigation to determine the linearity of response, and to determine if any other characteristics of the detector were dependent on the energy of the alpha particles emitted. The response of the large detector to the distribution of activity over the 25.4-cm diameter sample area was determined using a small 3-peak standard electrodeposited over a 1.9-cm diameter area in the center of a 4.9-cm mirror-finish stainless steel plate. However, the apparent energy or peak position and resolution indicated by the detector, as well as its counting efficiency, were dependent on the distribution and area of radioactivity under the detector.

Figure 1. Gas and vacuum connections to ionization chamber.

efficiency, were dependent on the distribution and area of radioactivity under the detector. Furthermore, many of these effects also increased more severely with alpha particles of higher energy. Consequently, another 3-peak standard was prepared in which the radioactivity was distributed over the entire 25.4-cm area of the sample plates so that the total integrated effects of samples uniformly distributed could be determined for the actual conditions being used.

The radionuclides used in both calibration standards were ^{230}Th , whose principal alpha peak is at 4684 keV, ^{239}Pu at 5155 keV, and ^{244}Cm at 5805 keV. The activity of each radionuclide was sufficient to give excellent counting statistics in 10-minute counts. The gain and digital offset were adjusted together to bring the three peaks to their desired energy positions and linearity. The analyzer was set to cover 512 channels at 10 keV/channel beginning at 2800 keV so that the principal radon daughter ^{214}Po at 7687 keV would be on scale at the upper end of the spectrum while even degraded alpha pulses, with energies as low as 2800 keV, would still be included at the lower end. Because it was easier to remember and interpret channel numbers than energies, the peaks of the three radionuclides mentioned above occurred in channels 188, 236, and 301, respectively.

Preparation of Large High-Resolution Calibration Source

About 5 mL of concentrated nitric acid was added to one of the 25.4-cm stainless steel sample plates, and the plate was heated under an infrared lamp. The hot acid was smeared over the plate with a plastic stirring rod to remove all greasy residues so that the steel would wet more readily. The plate was rinsed thoroughly with water and dried. Using a fine felt-tipped pen, a checkerboard grid of 1-inch squares was marked on the inside bottom of the pan. Using microliter pipets, exact known activities of about 10^4 dpm each of ^{230}Th , ^{239}Pu , and ^{244}Cm in 1 mL or less of volume, were combined in a 60-mL Teflon PFA beaker and diluted with sufficient water to give 4.0 mL of total volume. Working rapidly to minimize evaporation, a 50- μL aliquot was placed in the center of each 1-inch square, keeping the aliquot as confined as possible to avoid premature spreading. The partial squares around the rim received approximate fractions of a 50- μL aliquot equivalent to the fractional square inch of each partial square. If evaporation is kept small, the eighty 50- μL aliquots available in 4.0 mL of solution comes out close enough to the 78.5 required. Any excess can be spotted on the plate as desired. Using a Teflon PFA beaker to contain the activity to be added and an Oxford-type pipet with a polypropylene tip to make the additions, the transfer could be made uniformly and quantitatively without rinsing. If deemed necessary or desirable, the beaker and pipet tip can be rinsed with a drop or two of water. To avoid premature spreading or significant spattering, the plate was allowed to dry spontaneously in air with only slight warming from an infrared lamp.

Because of the impossibility of drying the sample on the plate in an area still covered by water, virtually all evaporation occurred at the edge of the liquid droplet producing a circular atoll-like deposit containing most of the mass and activity with little inside the ring except a few tortoise-shell markings. Although very small in terms of total mass, this deposit was relatively thick in terms of mass per unit area, compared to the range of an alpha particle. The consequent sample degradation confused the already complicated selection of experimental conditions leading to maximum instrumental resolution and could be eliminated by spreading the activity physically with a stirring rod.

A single drop of concentrated nitric acid was added to the residue from evaporation of each aliquot on the plate and the plate was warmed gently and uniformly under an infrared lamp. Using

a 1-mL polypropylene pipet tip whose tip had been sealed by warming slightly and squeezing with a pair of tweezers, each drop of nitric acid was smeared over its own square-inch area repeatedly as necessary to keep it from recombining. When the solution stayed pretty well dispersed over the area, the solution was allowed to evaporate to dryness. A drop of concentrated nitric acid was added at each of two areas of the plate relatively free of sample markings, and the pipet tip used to smear the solutions was rinsed consecutively in the two wash solutions. The resolution of the "smeared" standard was improved markedly, particularly at the toe of the peaks. A spectrum of the large-area standard used in the present work under the final recommended conditions of 14 psi overpressure of P-10 gas and 4 μ sec pulse shaping time is shown in Figure 2. The markers show the channel positions in which the radionuclide peaks should fall under the calibration conditions chosen, permitting changes in gain, resolution, peak shape, peak position, etc., with changes in experimental conditions to be followed.

Apparatus

The high-volume air sampler used was the model HV-1SH manufactured by F & J Specialty Products, Inc., Miami Springs, Florida. The sampler was equipped with a flowmeter calibrated for a range of 10 to 55 cfm of air, and could be used with filter holders for both 10-cm circular filters with an effective area 9 cm in diameter and 20- by 25-cm rectangular filters with an effective area of 18 by 23 cm. The no-load sampling rate was about 97 cfm, while the rate with 20- by 25-cm membrane filters was about 38 cfm with Versapor 1200 (1.2- μ m pores, Gelman Sciences, Inc.), 16 cfm with Versapor 800 (0.8- μ m pores), and about 14 cfm with Versapor 450T (thin) (0.45- μ m pores). Particularly important, the sampler was equipped with a cooling fan in the motor housing so that the motor cooling was not dependent solely on the air being drawn through the filter. Otherwise, the restricted air flow through filters of low porosity frequently could cause the motor to burn out unless the sampling times were kept short.

For taking large samples of surface dusts or other removable contamination over large areas quickly, i.e., "smears," an adapter was machined from gray PVC plastic to connect a commercial 14-cm vacuum cleaner brush attachment to the 10-cm filter head of the high-volume sampler through a 90-cm piece of Tygon tubing (2.5 cm I.D. with a 3-mm wall). The attachment was tapered inside from 2.5 cm at one end to 9 cm at the other end over a 15-cm length with a 3-cm female opening on the narrow end to accept the Tygon tubing and an external lip on the wide end to accept the retaining rings of the sampler. A similar transition piece for the 20- by 25-cm filter holder allowed higher sensitivity and better resolution, but was rather unwieldy if made large enough to give the turbulent flow necessary for uniform deposition. However, such a sampler would undoubtedly give more uniform deposition and more complete recovery of removable contamination for quantitative purposes than the smearing technique described below.

A 10-mL all-glass air-operated hand sprayer (Kontes, Spruce Street, Vineland, N. J., No. 422550) was used to spray a suspension of the finely-ground soil uniformly over the sample plate.

Figure 2. Large three-peak standard evaporated uniformly over a 25.4-cm diameter stainless steel plate.

EXPERIMENTAL TECHNIQUES

Operation of the Spectrometer

The sample holder containing a sample pan or filter was placed into the ionization chamber, the door of the access port was closed and dogged tightly against the rubber "O"-ring, and the gas valve to the vacuum pump was opened and all of the valves closed. The Leybold-Heraus pump required about three minutes to evacuate the chamber to a vacuum of 25 inches of mercury but an extra two minutes was allowed to ensure adequate removal of oxygen. The gas shut-off valve adjacent to the regulator was opened and the P-10 gas was allowed to come out of the relief valve for a few seconds to purge the line through the filter. The shut-off valve leading to the chamber was then opened, allowing some gas to be evacuated through the vacuum pump, and the valve leading to the pump was then closed. The pressure-relief valve was manually held closed until the gas pressure in the chamber had built up to the desired operating level of 14 psi above atmospheric pressure as indicated by the pressure gauge. While still holding the relief valve closed, both gas shutoff valves were tightly closed. The second stage of the regulator was set at about 16 psi so that the gas could be admitted rapidly to minimize loss of time but not so rapidly as to blow the sample around, causing internal contamination of the counter, or to permit significant overpressure if anything should go wrong. A voltage of 2000 volts was applied to the detector and the accumulation of pulses was begun. After completion of the desired counting time, the high-voltage was reduced to zero to avoid possible damage to the grid when the chamber was next evacuated. The gas-inlet valve was opened, permitting the excess gas to be vented through the relief valve down to about 6 psi. The door to the access port was then opened, the remaining pressure released, and the sample holder removed. The entire spectrum was preserved in the memory indefinitely and printed out by a laser printer, and the alpha peaks were integrated through any regions desired. When not in use, the ion chamber was kept in an atmosphere of P-10 to prevent absorption of oxygen and nitrogen and subsequent out-gassing during the next analysis.

Effect of Experimental Conditions

As mentioned above, three different radionuclides having markedly different principal energies were used in most of the present work to identify those variables whose magnitudes depend on the energy of the alpha particle being determined. For each experimental condition being tested, the gain was always readjusted to bring the three peaks back to their correct energy positions in channels 188, 236, and 301, respectively, indicated by the markers, before drawing conclusions as to the effect of the variation on resolution, peak shape, etc. Changes in gain in most cases could be used to indicate the direction and approximate extent of the change produced. In most cases, the absolute numbers were not relevant but the relative changes for a given source were meaningful. Consideration must be given to the statistical variations in the total counts.

The effect of pulse-shaping time, the overpressure of P-10 gas above atmospheric pressure in the counting chamber, and the distribution of the radioactivity relative to the detector on the resolution and peak position are shown in Table 1. When the small 1.9-cm 3-peak electrodeposited source was placed at the center of the large 25.4-cm detector and counted at atmospheric pressure with a shaping time constant of 4 μ sec, the full width at half maximum (FWHM) of the peaks became significantly larger as the energies of the three radionuclides increased, as shown in Table 1.

Table 1. Effect of shaping time, P-10 overpressure, and distribution of activity on resolution of alpha spectra (source centered under detector).

<u>Shap. time, μsec</u>	<u>P-10 press. psi</u>	<u>Small 3-peak electrodeposited source centered under detector</u>			<u>Peak Hit., c/5 min</u>
		<u>Gain</u>	<u>Isotope</u>	<u>Resolution, FWHM, keV</u>	
1	0	0.875	^{230}Th	102	8162
			^{239}Pu	123	4076
			^{244}Cm	155	1834
4	0	0.774	^{230}Th	124	6814
			^{239}Pu	149	3478
			^{244}Cm	208	1552
1	6	0.888	^{230}Th	78	10020
			^{239}Pu	87	5384
			^{244}Cm	121	2422
4	6	0.793	^{230}Th	100	8112
			^{239}Pu	109	4360
			^{244}Cm	141	2046
1	14	0.912	^{230}Th	78	9692
			^{239}Pu	76	5832
			^{244}Cm	95	2938
4	14	0.802	^{230}Th	91	8582
			^{239}Pu	89	5150
			^{244}Cm	100	2756

Moreover, the peak shapes changed from a narrow symmetrical one with ^{230}Th to a very wide one with ^{244}Cm with a pronounced shoulder on the low-energy side, as shown in Figure 3. The shape of the ^{239}Pu was intermediate between the two. This effect could not have been caused by the branching ratios of the radionuclides, because the resolution was not nearly good enough to resolve them and because the branching ratios of ^{230}Th and ^{244}Cm are both 76:24 but the branches of ^{230}Th , which always show a smooth symmetrical peak, were separated by 66 keV, while those of ^{244}Cm , which show the misshapen peak, were separated by only 42 keV.

When the shaping time constant was decreased to 1 μsec , the three peaks all shifted to lower energies requiring a substantial increase in gain to bring them back to their original energy positions. Although the resolution as indicated by the FWHM was improved, the shape of the ^{244}Cm peak was even more degraded. The shoulder on the low-energy side was more pronounced and lower so that it did not affect the FWHM as much, but would still have resulted in poor resolution. By increasing the pressure of P-10 counting gas in the chamber above atmospheric, the resolution of all three radionuclides was greatly improved and the spread among them was reduced, particularly with a shaping time of 1 μsec and a gas pressure of 14 psi above atmospheric, although the gain required was also the highest. At 14 psi overpressure with either shaping time, all three peaks were narrow, smooth and symmetrical with no evidence of shoulders or other deformities. Taking the statistics involved into consideration, the counts in the peak channel, i.e., the peak height, correlate inversely very well with FWHM, i.e., the peak width, with different conditions for the same radionuclide.

When the small source was moved out to the edge of the detector and counted at atmospheric pressure with a shaping time of 1 μsec , all three peaks were shifted to high energies and thickened considerably. The shift in position was larger than 200 keV depending on the other conditions. This was not an "edge effect" in the usual sense because the same effect was observed in proportionately smaller degree as soon as the source was moved away from the center. The ^{239}Pu peaks were rounded on the high-energy side and the ^{244}Cm peaks had a distinct shoulder on the high-energy side as shown in Figure 4. This was true even at the higher pressures when a shaping time of 1 μsec was used. As shown in Table 2, increasing the shaping time to 4 μsec and lowering the gain to bring the peaks back to their scheduled energy positions improves the resolution, narrows the range of resolutions with energy, and restores a smooth symmetrical shape to all three peaks, particularly at 14 psi overpressure. This effect is opposite to that observed when the source is centered under the detector when the shorter shaping time gives the better resolution and the shoulder on the ^{244}Cm peaks is on the low-energy side. However, the gain required to restore the peaks to their proper energy positions is markedly different when the small source is displaced compared to when it is centered. Consequently, with a single gain setting and a large source uniformly distributed over an area the same size as the detector, as in measurement of actual samples, the shift in peak position of alpha particles originating in different parts of the sample has the effect of broadening the peak and decreasing the resolution. Thus, the calibration source used to determine the optimum conditions for actual analyses should be composed of several different energies and be uniformly distributed over an area the same size as the actual samples to be measured.

Table 3 shows the results obtained with the large 3-peak source evaporated uniformly over the entire area of the sample plate, prepared as described above. As expected, the FWHM of ^{230}Th was larger in every case than was obtained with the small concentrated source, whether centered under the detector or not, and at either shaping time or overpressure. However, as was true with other conditions, the change in resolution obtained with the other two radionuclides depended on the conditions. Generally, their FWHMs were also larger than with the small source centered under the detector at either shaping time, or at the edge of the detector with the 4 μsec shaping time.

Figure 3. Small three-peak electrodeposited standard, centered under detector, and counted at atmospheric pressure and 4 μsec shaping time.

Figure 4. Small three-peak electrodeposited source at edge of detector, and counted at atmospheric pressure and 1 μsec shaping time.

Table 2. Effect of shaping time, P-10 overpressure, and distribution of activity on resolution of alpha spectra (source at left edge of detector).

Shap. Time, (μ sec)	P-10 Press. psi	Small 3-peak electrodeposited source at left edge of detector			
		Gain	Isotope	Resolution, FWHM, keV	Peak Ht., c/5 min
1	0	0.860	^{230}Th	138	6178
			^{239}Pu	186	2800
			^{244}Cm	279	1134
4	0	0.772	^{230}Th	98	8054
			^{239}Pu	107	4398
			^{244}Cm	141	1980
1	6	0.860	^{230}Th	117	6974
			^{239}Pu	153	3296
			^{244}Cm	217	1426
4	6	0.779	^{230}Th	92	8418
			^{239}Pu	90	5116
			^{244}Cm	104	2682
1	14	0.875	^{230}Th	122	6878
			^{239}Pu	160	3252
			^{244}Cm	237	1324
4	14	0.786	^{230}Th	91	8448
			^{239}Pu	86	5190
			^{244}Cm	90	2934

Table 3. Effect of shaping time, P-10 overpressure, and distribution of activity on resolution of alpha spectra (3-peak evaporated source).

Shap. Time, (μ sec)	P-10 Press. psi	Large 3-peak evaporated source			
		Gain	Isotope	Resolution, FWHM, keV	Peak Ht., c/ $\sqrt{}$ min
1	0	0.872	^{230}Th	150	2476
			^{239}Pu	172	1944
			^{244}Cm	227	1920
4	0	0.780	^{230}Th	134	2770
			^{239}Pu	150	2278
			^{244}Cm	192	2278
1	6	0.875	^{230}Th	131	2748
			^{239}Pu	137	2338
			^{244}Cm	162	2490
4	6	0.786	^{230}Th	111	3282
			^{239}Pu	115	2852
			^{244}Cm	130	3196
1	14	0.909	^{230}Th	127	2822
			^{239}Pu	128	2544
			^{244}Cm	144	2794
4	14	0.799	^{230}Th	100	3414
			^{239}Pu	95	3246
			^{244}Cm	104	3744

However, their resolution was better than with the small source at the edge of the detector when the 1- μ sec shaping time was used. For large uniform sources, a shaping time of 4 μ sec gave better resolution and less difference in resolution with energy than 1 μ sec, and an overpressure of P-10 counting gas of 14 psi reduced the FWHM by about one-third of that obtained at atmospheric pressure. Under these recommended conditions (see Figure 2), the resolutions of all three peaks with the large diffuse evaporated source were within 20 keV of those obtained under the best conditions with the small concentrated electrodeposited source. Similarly, with an air sample containing ^{214}Po distributed uniformly over an area 18 cm by 23 cm, the resolution of its peak at 7687 keV was over 500 keV at 4 μ sec shaping time and atmospheric pressure, and decreased to 300 keV with 14 psi of P-10 gas. This lessens the interference of the ^{214}Po tail on the determination of TRUs at lower energy considerably.

Because ^{241}Am is a particularly important radionuclide in nuclear waste management and was not included in the 3-peak standard because of spectral congestion, another large evaporated source containing only this radionuclide was prepared as described above. As shown in Table 4, the gain required to bring the 5486-keV ^{241}Am peak into its scheduled channel 269, which is exactly halfway between ^{239}Pu and ^{244}Cm , is almost identical to those values shown in Table 3 for the large 3-peak source for the various conditions. However, the FWHM appears to be a little larger than expected compared to the 3-peak source, possibly because of slightly larger residual crusts left in preparing the evaporated ^{241}Am source.

Sample Collection and Preparation

Soil and Settled Dusts

Collection of soil samples was straightforward and was not expected to present any unusual problems. Samples of settled dust were collected either by brushing some into a small pile with heavier deposits, or with the vacuum cleaner attachment to the high-volume air sampler with lighter deposits. Any of the usual 10-cm filters could be used to collect the dust if present in large quantities and the spraying and grinding technique was to be used for mounting as with soils. However, samples could also be taken on 20- by 25-cm membrane filters for direct spectrometry while suspending the dust by sweeping or other activities. If necessary, the samples could be dried and screened to remove root mat, rocks, and other debris.

A 2-gram sample of soil or settled dust free of debris was ground in a diamonite mortar with 4 mL of 30% ethanol for 2-3 minutes. The mortar was tipped slightly so that the liquid containing the finely-divided suspended material could be drained away from the heavier solids and removed by suction with an Oxford-type pipet. The suspension was placed in the glass tube of the sprayer, and the grinding and decantation was repeated twice more with 2 mL of 30% ethanol each time. The intent was to recover as much of the surface material of the sample as possible and reject the larger residual particles containing relatively more of the total mass. The residue was discarded. The suspension was then shaken to mix as thoroughly as possible and sprayed with occasional additional shaking onto the 25-cm sample pan as uniformly as possible. The pan was dried under an infrared lamp. It was desirable to recover about half of the 2-gram sample giving an average thickness of about 2 mg/cm^2 . Because the sample was already nearly infinitely thick for alpha emission, this procedure was not intended to be accurate quantitatively, but to provide reliable

Table 4. Effect of shaping time, P-10 overpressure, and distribution of activity on resolution of alpha spectra (evaporated ^{241}Am source).

Shap. Time, (μsec)	P-10 Press. psi	Large evaporated ^{241}Am source			
		Gain	Isotope	Resolution, FWHM, keV	Peak Ht., c/30 min
1	0	0.875	^{241}Am	217	1446
4	0	0.780	^{241}Am	185	1758
1	6	0.880	^{241}Am	147	1936
4	6	0.785	^{241}Am	137	2208
1	14	0.900	^{241}Am	155	1852
4	14	0.792	^{241}Am	111	2580

qualitative identification at the highest sensitivity possible for the conditions. Larger samples did not increase either sensitivity or precision significantly because the further increase in absorber offset the additional activity added. When activity was detected, smaller samples were used to obtain better resolution, known counting efficiency, effective sample size, and consequently, better accuracy. However, dissolution and appropriate chemical separations followed by high-resolution alpha spectrometry will always provide the best accuracy and precision possible.

Air Dusts

Relatively large samples of air dusts were collected on 10-cm or 20- by 25-cm membrane filters using the high-volume air sampler for the shortest time possible consistent with the sensitivity desired to minimize loss of time and pickup of radon daughters. For example, a 10-min sample at 16 cfm on a Versapor 800 filter, with a 10-min count begun immediately without waiting for decay, gave adequate sensitivity for most purposes. Selection of the proper filter for collection of air dusts for subsequent direct alpha spectrometry was very important. For best resolution of the spectra, the filter must be a membrane-type filter with sufficiently small pores to retain the dusts at the surface. With the usual fibrous filters, penetration of the particles into the fiber caused unacceptable degradation of the alpha particles coming out. However, the flow rate through the membrane filters of small pore size was much lower than with the fibrous filters so that the pump motor had to be cooled separately to avoid possible burnout. Millipore AA or HA membrane filters were very smooth and gave higher flow rates and slightly better resolution than other filters with comparable pore sizes, but were rather brittle and required more care in use. Versapore 800 or 450T were very rugged and much less expensive, and are recommended.

Smear Samples

Brodsky and Gallagher⁴ pointed out that, "Contamination monitoring of surfaces using what are variously called 'smears', 'wipes' or 'swipes' is a common procedure in facilities using radioactive materials in dispersible form. The Nuclear Regulatory Commission (NRC) has provided general action levels for removable contamination above which specified control measures should be taken to protect against internal radiation exposure . . . For workers, the action levels are usually many orders of magnitude lower than those that, if the material were resuspended, would produce maximum permissible concentrations in air."

If qualitative detection of loose contamination is desired, smear samples will provide greater sensitivity per unit counting time with greater convenience, speed, and simplicity than other methods such as direct measurement with survey instruments, air sampling, etc. Certainly, smear sampling will provide much earlier warning of the presence of loose contamination at much higher sensitivity than using air sampling to detect it after it has become resuspended in air. We agree with Brodsky and Gallagher that many recorded measurements of radioactivity are woefully incomplete and lacking in adequate statistical interpretation. However, even with the best statistics possible, it is doubtful that we can be sure that, "the recorded amount of radioactivity on a smear is quantitatively correct" (see Reference 4). Smear samples were virtually impossible to interpret quantitatively with any high degree of accuracy because of the nonuniformity of the sample deposition on the smear, the consequent uncertainty in the counting efficiency, the uncertainty of the completeness with which the removable activity was recovered from the area sampled, and the presence of natural activity from the air which makes gross alpha counting inadequate. It is more desirable that the detection of loose contamination be as sensitive as possible rather than as accurate as possible. The

large-area alpha spectrometry used in the present work suffers from the same quantitative deficiencies common to other smearing techniques, i.e., uncertainties in counting efficiencies and recovery of the loose contaminant, but provided positive identification of the TRU present at much higher sensitivity than other methods. When activity was detected, the measurement technique was revised to provide more uniform samples to achieve the necessary precision and accuracy.

Collection of removable activity from surfaces onto the large 20- by 25-cm filters can be accomplished simply and more efficiently and uniformly using an 18-cm paint roller with a cover having a 1-cm Orlon pile. A Versapor-800 filter is placed on the area to be smeared with its front 20-cm dimension placed at the beginning of the area to be smeared. While holding the filter firmly against the surface by pressure from the roller, the filter is pulled out from under the roller and away from the surface slowly and smoothly as the filter is drawn across the area being traversed. The roller is allowed to slide forward with the filter while rolling to the back of the filter as the filter is drawn across the designated surface. The back end of the filter should reach the roller at the same time as the roller reaches the end of the area being traversed. The same area can be covered repeatedly to increase the efficiency of removal of activity depending on the dirt loading present, or additional areas can be included to increase the sensitivity of coverage. The filter is then placed between two sheets of aluminum or stainless steel to avoid absorption of radon daughters from the air during movement to the counter. The filter smear is placed in the rectangular cutout of the filter holder, secured by the hold-down plate, and counted in the alpha spectrometer. Millipore AA or HA membrane filters are very brittle and easily broken under the conditions used, and are not recommended.

EXPERIMENTAL RESULTS

Backgrounds

Proper blanks and instrument backgrounds are of particular importance to demonstrate that a statistically positive result will not be obtained when the substance being sought is in fact below its detection limit, i.e., all positive interferences have been eliminated and contamination from all sources has been successfully avoided. This is particularly important with respect to pickup of radon and thoron daughters from the air. The blank is also of primary importance in determining a quantitative value for the detection limit under a given set of experimental conditions. The definition of lower limit of detection (LLD) used throughout the present work is that of Currie.⁵ Currie's equation for paired observations of sample and background with a probability of 0.05 that activity will be detected when none is present and will not be detected when some is present is as follows:

$$\text{LLD} = 2.71 + 4.66 B^{1/2} .$$

In terms of concentration in pCi/g, the equation becomes:

$$\text{LLD} = (2.71 + 4.66 B^{1/2})/2.22\text{TEYg}$$

where the symbols B, T, E, Y, g, and 2.22 have the meanings given below.

Figure 5 shows a typical 60-minute instrument background. The curves were virtually identical whether an empty stainless steel pan or a Versapor 800 membrane filter was placed under the detector. The shape of the curve was very similar to that obtained by others (see References 1 and 2) but the level of activity present was several times higher, perhaps because of radioactivity in the ceramic insulators used internally. In fact, the shape of the curve was almost identical to that obtained from a soil sample (see Figure 6). The curve began at a little over 6 MeV and then increased as it went to lower energies, suggestive of the successive addition of degraded pulses from the natural series, each building on the sum of those from higher energy. The barely detectable peak of ²¹⁴Po was evidently sorbed from the air during mounting because it disappeared after a couple hours out of contact with air. The areas identified by markers in this and all subsequent spectra were those used to integrate the main part of the ²³⁹Pu and ²⁴¹Am peaks. The ²⁴¹Am region was smaller than that taken for the ²³⁹Pu region because there was less room above the high-energy edge of the ²³⁹Pu, and because its higher energy puts it in a region of slightly lower background. Consequently, its counting efficiency will be proportionately smaller. The LLDs, for the background spectra obtained with no absorbing materials present, were about 0.6 pCi for 60-min counts and 1.5 pCi for 10-min counts. A minimum of three counts in a particular energy region was used for background for any counting time.

Figure 6 shows a 60-minute background of approximately 1 gram of soil, ground, and sprayed over the 25.4-cm diameter plate. As with the instrument background, the soil background increases in counting rate as it goes from high to low energies. The background begins at the 7687-keV peak of ²¹⁴Po and increases as it goes to lower energies at about the rate that one might expect from successive addition of degraded members of the natural uranium and thorium series, each building up on the sum of all previous ones. Otherwise, no identification of individual members of either series is possible at their natural level of about 1 pCi/g each in a 60-min count on a 1-gram sample

Figure 5. Instrument background, 60-min count: ^{239}Pu region, 47 counts; ^{241}Am region, 9 counts.

Figure 6. Soil background, 60-min count: ^{239}Pu region, 103 counts; ^{241}Am region, 65.

in a badly-degraded spectrum. Assuming a 50% attenuation of the alphas present and another 50% scattered out of the energy region being quantified for a 1-gram sample, the LLDs are about 3 pCi/g for 60-min counts and about 8 pCi/g for 10-min counts. The LLD would be the same for both radionuclides because the smaller count in the ^{241}Am region would be offset by the smaller counting efficiency for the smaller energy region integrated. Although these projections are a little less than the 10 to 20 pCi/g determined empirically below for 10-min counts of soils containing known concentrations, they confirm the philosophy that a factor of 2-3 in accuracy is probably less important for screening purposes than a sensitivity of a few tens of pCi/g.

Figure 7 shows a background air sample collected at 24 cfm for 10 minutes on a 20- by 25-cm Millipore AA filter (0.8 μm pore size) and counted for 10 minutes beginning 11 minutes after completion of sampling. The major peak is the principal alpha-emitting ^{222}Rn daughter, ^{214}Po , which decays with about a 40-minute half-life because of its beta-emitting ^{214}Pb -Bi precursors. The resolution of this peak is about 360 keV FWHM at a shaping time of 4 μsec and 14 psi overpressure of P-10 counting gas. The minor peak is a residual 3-min radon daughter, ^{218}Po , not yet completely decayed. This spectrum gives an LLD of about $9 \text{ E-}13 \mu\text{Ci/mL}$ for both ^{239}Pu and ^{241}Am , or a little less than half of the derived air concentration (DAC) for occupational exposure to most of the TRUs of $2 \text{ E-}12 \mu\text{Ci/mL}$ without waiting for significant decay. The contribution from the tail of the radon daughters is larger in the ^{241}Am region than in the ^{239}Pu region which is further down the tail toward lower energies. As time passes and the ^{214}Po decays, the detection limit becomes significantly lower. For example, 2 hours later, the detection limit for this same sample was lowered to $4 \text{ E-}13 \mu\text{Ci/mL}$ because of the decreased contribution of the ^{214}Po tail to both regions.

Figure 8 shows a 10-minute of a 32-in. by 48-in. smear sample taken from a dusty bench top. The spectrum is almost identical to that from an air sample except that the ^{214}Po activity is much less because of the much smaller quantity of air collected, resulting in far less interference and greater sensitivity. This smear would have detected about 5 pCi of activity collected from the bench top, or an average concentration of about 0.0005 pCi/cm^2 .

When maximum sensitivity on soils was desired, the 2-gram sample taken for analysis made the effective sample size and counting efficiency of the residue sprayed onto the plate very difficult to determine with any accuracy. Consequently, the LLD was determined empirically by mounting a standard^{6,7} containing the expected minimum detectable activity under the exact conditions to be used to determine whether or not the activity could in fact be detected. Figure 9 shows a 10-minute spectrum of soil containing 20 pCi/g each of ^{239}Pu and ^{241}Am , about half of which was ground sufficiently fine to be sprayed onto the plate. The right and center markers clearly show the high-energy edges of the ^{241}Am and ^{239}Pu peaks, respectively. The tail shows the spectrum of the degraded alphas from both peaks superimposed on the soil background. For comparison, the background for the same quantity of soil without added radioactivity is one sixth of the 60-minute spectrum shown in Figure 6. Figure 10 shows a 60-minute spectrum of the same sample, confirming that there are two peaks present that increase proportionately to counting time, that their high-energy edges do agree with the correct positions for these radionuclides, and that "most" of the activity appears within the energy region being quantified. Thus, it is concluded that the detection limit for a 10-minute count of a 2-gram sample mounted as described is about 20 pCi/g for any alpha emitter. When lesser sensitivities are acceptable and time is not so pressing, smaller samples should be taken to obtain better resolution, and longer counting times used to restore as much sensitivity and precision as possible.

Figure 11 shows a spectrum of an air sample collected for 10 minutes at 16 cfm on Versapor 800 (0.8 μm pores, Gelman Sciences) and counted for 10 minutes beginning 15 minutes after

Figure 7. Air background, 10-min count: ^{239}Pu region, 133 counts; ^{241}Am region, 141 counts.

Figure 8. Smear sample of dirty bench top, 10-min count: ^{239}Pu region, 74; ^{241}Am region 79.

Figure 9. Two grams of soil containing 20 pCi/g each of ^{239}Pu and ^{241}Am , 10-min count: ^{239}Pu region, 108 counts; ^{241}Am region, 69 counts.

Figure 10. Same sample as in Figure 9, 60-min count: ^{239}Pu region, 657 counts; ^{241}Am region, 351 counts.

Figure 11. Air sample collected at 16 cfm for 10 minutes containing approximately 2 pCi each of ^{239}Pu and ^{241}Am .

completion of sampling. During sampling, -200 mesh soil standards (see References 6 and 7), one containing 103 pCi/g of ^{239}Pu , the other containing 200 pCi/g of ^{241}Am , were shaken vigorously; the container was opened as quickly as possible and the very finest dust was allowed to be sucked onto the filter. This process was repeated until an estimated 20 mg of the ^{239}Pu soil and 10 mg of the ^{241}Am soil had been collected, giving a total of about 2 pCi of each radionuclide. The filter showed only a barely detectable light brown color from the soil collected. The spectrum in Figure 11 shows two distinct humps on the tail of the ^{214}Po peak at the proper positions for the two radionuclides used, even when the sample was counted as quickly as possible after collection before the ^{214}Po had decayed appreciably. Figure 12 shows a spectrum of the same sample taken 49 minutes later. The display has been expanded by a factor of 2 to offset the approximate decay of the ^{214}Po . This spectrum shows that the peaks were real and that increased sensitivity can be gained if time can be allowed for even partial decay of the ^{214}Po .

When time is not so urgent and high-volume air samples can be taken for relatively long times and counted for relatively long times, exceptionally high sensitivities for detection of TRUs in air can be achieved. However, the sample size must be controlled to keep the dust loading small, and sufficient time must be allowed between the end of sampling and beginning of counting to permit most of the radon and thoron daughters to decay. The resolution and accuracy will also be improved markedly. As the $^{214}\text{Pb-Bi-Po}$ chain decays with about a 40-min half-life, the ^{212}Bi thoron daughter, supported by its 10.64-hour ^{212}Pb precursor, begins to appear at approximately the same position as was occupied by ^{218}Po much earlier, and just above the ^{241}Am position. This relationship is shown in Figure 13 in an air sample to which a small quantity of ^{239}Pu was added but no ^{241}Am . Even when 3 hours is allowed to permit nearly complete decay of ^{214}Po , most of the ^{212}Bi will remain, as shown in Figure 14, and becomes the limiting interference in the determination of ^{241}Am . Also, the ^{212}Bi must not be mistaken for an artificial contaminant. About 5 days is required for 99.9% decay of the $^{212}\text{Pb-Bi-Po}$ chain to eliminate most of the interference on large air samples. An example is given in Figure 15. The detection limit is $1.9 \text{ E-15 } \mu\text{Ci/mL}$, or approximately 1/1000 of the DAC for most TRU nuclides in air for occupational exposure. If activity is detected, its concentration can be readily calculated from the known sample size and a counting efficiency that can be determined because of the small dust loading permitted.

Counting Efficiency

The large 3-peak distributed source was counted for 10 minutes in the spectrometer under the final recommended conditions, giving the spectrum shown in Figure 2. The sum of all three components down to 2800 keV was used in the determination of the counting efficiency. By dividing the counting rate of the source by the sum of the three known activities, the ^{244}Cm component being corrected for decay since standardization, the counting efficiency for massless sources uniformly distributed over the 25.4-cm stainless steel plate was obtained. The counting efficiency obtained with the present detector was 0.473 ± 0.004 , very close to the theoretical limit for 2π geometry. The other source prepared for Am^{241} alone in exactly the same manner gave 0.494 ± 0.004 . The small difference was not investigated further because the uncertainties in the Oxford-type pipets alone could have accounted for the difference. Clearly, the counting efficiency obtained with such distributed sources did not depend significantly on either the energy of the radionuclide or its distribution within the sample area under the conditions used. When maximum sensitivity on soils is desired, the lower limit of detection should be determined empirically using soil standards of known concentration as described above. However, when activity is detected and quantitative results are desired, the analysis should be redone using a smaller sample and longer

Figure 12. Same sample as in Figure 11; spectrum taken 49 minutes later after decay of about half of the ^{214}Po : ^{239}Pu region, 493; ^{241}Am region, 410.

Figure 13. Air sample collected at 24 cfm for 10 min with approximately 10 mg of 103 pCi/g ^{239}Pu soil, sample allowed to decay for 25 minutes and counted for 1 hour: ^{239}Pu region, 3700 counts; ^{241}Am region, 2394 counts.

Figure 14. Same sample as Figure 13 but only a 10-min count after further decay of two hours to decrease the ^{214}Po tail: ^{239}Pu region, 361; ^{241}Am region, 151.

Figure 15. Air sample collected at 24 cfm for 310 minutes, allowed to decay for 92 hours, and then counted for two hours: ^{239}Pu region, 73; ^{241}Am region, 43. Detection limit is $1.9 \text{ E-}15 \mu\text{Ci/mL}$, or 1/1000 of DAC.

counting time to obtain spectra of higher resolution for which a reliable counting efficiency can be determined or estimated.

The counting efficiency to be used with air samples or smears taken on 20- by 25-cm membrane filters can be determined by counting a standard prepared similarly on a checkerboard grid drawn on a 20- by 25-cm stainless steel plate with a 1.25-cm margin around the perimeter to conform to the effective area of the filter holder and hold-down plate inside the frame. The attenuation factor should be small with light dust loadings and can be estimated with reasonable accuracy. By exercising control of sampling time relative to the dust loading in the air, thick sources can be avoided and should not be necessary to give reasonably high sensitivity.

Calculation and Interpretation of Results

In a well-resolved spectrum, identification of the radionuclides is easily made by comparing the energy of each peak with the known energy for each radionuclide, most readily and easily used and remembered in terms of the corresponding channel numbers. The area under the peak is then integrated, corrected for background and blanks, and the activity is calculated from the following equation:

$$\frac{(X - B)}{2.22 T E Y g} = \text{Act, pCi/g} \quad (1)$$

where

- X = the total number of counts in the energy interval of the alpha spectrum for the radionuclide being determined
- B = the total number of counts in the same energy interval for the same counting time for the background and other corrections
- T = the counting time in minutes of the alpha spectrum
- E = the counting efficiency of the spectrometer in cpm/dpm
- Y = the fractional recovery of the radionuclide sought
- g = the size of sample taken for analysis in grams
- 2.22 = the number of dpm/pCi.

If the spectrum is badly degraded, as will be the case when large samples are taken to obtain maximum sensitivity in short counting times, the particle sizes are too large, or the sample is not

uniformly distributed, identification, and determination will be more approximate. However, ^{239}Pu and ^{241}Am are virtually the only TRU nuclides needing to be resolved from the composite background plus degraded continuum from the natural activities present. Their alphas will build on top of the natural background to give a hump the high-energy edge of which will agree with the correct energy of the TRU, provided that there is no layer of barren soil covering the entire sample. As mentioned above, the LLD can be determined empirically from known standards to obtain maximum sensitivity, or the sample size can be decreased to give adequate resolution to permit a reliable counting efficiency to be determined. In some cases in which high sensitivity and accuracy are both required, or the sample cannot be duplicated, the existing sample can be dissolved; the TRUs of interest can be separated chemically and determined by high-resolution alpha spectrometry.⁸

Propagation of Statistical Uncertainties

Every measured value of a random process has a statistical uncertainty with which the measurement was made that must be recorded with each analytical result, preferably expressed as one standard deviation. As is frequently done, comparison of the results without consideration of the uncertainty inherent in the measurements themselves is meaningless and can only lead to incorrect interpretation. The uncertainty of each measured value must include every significant random uncertainty incurred anywhere in the entire measurement process. Each uncertainty should be propagated to the final result by the partial differential equation known as the Law of Propagation of Errors, or more correctly, the propagation of uncertainties. Error propagation should also be extended to all values derived from the actual measurements such as ratios, percentages, sums and differences, material balances, etc. An estimate of the systematic uncertainties should also be made but should be kept separate and not propagated with the random uncertainties.

The estimated standard deviation for the activity in pCi/g given in Equation (1) above is the following:

$$S_A = \{[(X + B)^{1/2} / (X - B)]^2 + \{S_T/T\}^2 + \{S_E/E\}^2 + \{S_Y/Y\}^2 + \{S_g/g\}^2\}^{1/2} \times \text{Act pCi/g} \quad (2)$$

where the S_i 's are the estimates of the individual standard deviations and the other symbols have the same meanings given above. With thick samples taken to obtain maximum sensitivity, the effective mass from which alpha particles can escape to the detector and the resultant effective counting efficiency (and their uncertainties) will have to be estimated. When activity is present, it is preferable to use a smaller sample and longer counting time so that the parameters in Equation (1) and their uncertainties in Equation (2) can be determined with better accuracy.

In the operation of constant monitors for alpha activity in air, it is currently wide-spread practice for the correction **B** for the radon and thoron daughters from the air that tail into the region of interest to be calculated from the product of a measurement **C** from some other part of the spectrum outside the region of interest and a "constant" **k**, which gives the correlation ratio between the region of interest and the other region used in the measurement. In this case, **kC** will be used in Equation (1) instead of **B**. Also, because the "constant" **k** is itself the ratio of two measurements of random processes made at some other time, the uncertainty in its value must also be included in the error propagation given in Equation (2). Combining the uncertainty in the product $S_{kC} = [(S_k/k)^2 + (S_c/C)^2]^{1/2} \times kC$ with the uncertainty in **X**, and assuming that $S_c = (C)^{1/2}$, the following relationship results for the first term in Equation (2), the other terms remaining the same:

$$S_A = [\{ (X + S_{kC}^2 C^2 + k^2 C) / (X - kC) \}^2 + \dots]^{1/2} \times Act \quad (3)$$

The following real example illustrates the problems that arise with this approach. An air sample was collected for 5 minutes at 15 cfm on a Versapor 1200 membrane filter and counted in the spectrometer for 5 minutes immediately, and again after 17 and 51 minutes, and then for 30 minutes after 60 minutes total elapsed time. The 30-minute count gave 118 total counts in the ^{239}Pu region and 2,824 in the ^{214}Po region, down to the high-energy side of the ^{218}Po . The constant **k** and its standard deviation S_k were calculated from this data: $k = 118/2824 = 0.042$, and $S_k = 0.004$. In the initial 5-minute count, 78 counts and 1,454 counts, respectively, were obtained giving a net count in the ^{239}Pu region and its standard deviation of 17 ± 11 . The second 5-minute count gave 39 and 1,211 counts, respectively, for a net and standard deviation of -12 ± 8 . The third 5-minute count gave 24 and 747 counts, respectively, for a net and standard deviation of -7 ± 6 .

Although all the net values are statistically zero at the 95% confidence level, the initial spectrum contained a detectable peak of ^{218}Po , the tail of which undoubtedly increased the counts in the ^{239}Pu region more than with the later measurements by which time the 3-minute ^{218}Po would have decayed. As time passes, tailing from the higher-energy ^{212}Bi and ^{212}Po daughters of 10.64-hour ^{212}Pb from the natural ^{232}Th chain then contributes an increasingly large part of the total activity present in both regions due to decay of the 3-minute ^{218}Po and the ^{214}Po supported by the approximately 40-minute $^{214}\text{Pb-Bi}$ chain. An example is shown in Figures 13 and 14. Interference of the thoron daughters will be particularly severe with longer sampling times. Consequently, the values of **k**, **X**, and **C** must vary non-proportionately with the conditions under which they are measured, such as time of day (inversions), length of sampling and decay, etc. Clearly, if time permits, it is far preferable to allow as much decay of the radon-thoron daughters as possible before using such inexact and variable correlations.

Air monitors operated continuously will permit equilibria to be established which will tend to reduce the variability but not necessarily eliminate errors from day to day. Experimental data on continuous monitors generally confirm good reproducibility under certain conditions and extremely

variable under others. The reliability of any given measurement is, therefore, open to question. In any event, mere reproducibility does not necessarily indicate accuracy, and theoretical standard deviations should always be calculated to determine whether or not more accuracy is being claimed from the experimental data alone than the procedure can be counted on to deliver consistently.

QUANTITATIVE MEASUREMENTS

When significant radioactivity is present, the sample size recommended above to produce maximum sensitivity is much too large to permit acceptable accuracy to be obtained. With deposits thicker than about $100 \mu\text{g}/\text{cm}^2$, the total loss of activity from attenuation and scattering out of the energy region being quantified becomes so large and variable that acceptable quantification becomes virtually impossible. A 100-mg sample of a standard soil (see References 6 and 7) containing $1030 \pm 14 \text{ pCi/g}$ of ^{239}Pu was finely ground and sprayed uniformly over a 507-cm^2 area, giving an average thickness of about $200 \mu\text{g}/\text{cm}^2$. The spectrum is shown in Figure 16. About 72% of the alpha particles reaching the detector are in the chosen energy region about 540 keV wide around the ^{239}Pu peak. The overall counting efficiency of alpha particles reaching the detector in the energy region being monitored per disintegration in the original sample is about 21%. Losses of activity in the mortar, in the spraying process, from mass attenuation, and from scattering out of the proper energy region are all included in the empirical measurement. It should be mentioned that the 47% counting efficiency obtained during calibration of the detector was obtained with essentially carrier-free activity, and included all alpha pulses counted from the high energy intercept of the peak down to the low-energy end of the spectral range at 2800 keV. Since the sample background from 100 mg of soil is also little more than the instrument background shown in Figure 5, the detection limit is still very good, i.e., about 40 and 15 pCi/g for 10- and 60-min counts, respectively. Of equal importance, this permits quantitative measurements of TRUs to be made in soils and settled dust with an accuracy almost as good as the statistics of the measurements will allow.

Two other points of interest from Figure 16 should be noted. The distinct peak of ^{241}Am visible comes from the ^{241}Am ingrown from the ^{241}Pu in the old ^{239}Pu used to prepare the standard soil, equivalent to about 2% of the ^{239}Pu activity. This also shows how well small activities of ^{241}Am can be detected in the presence of relatively larger activities of ^{239}Pu . A second point of interest is the excellent resolution of about 120 keV obtained at the FWHM, but the much poorer resolution obtained at the toe of the curve. This is probably due to the large surface of the soil standard after grinding and spraying from which alpha particles can escape without losing any energy.

Figure 16. Standard soil, 100 mg of 1030 ± 14 pCi/g ^{239}Pu 60-min count.

CONCLUSIONS

The lower limit of detection (LLD) for alpha emitters for a 10-min count on a 2-gram sample of soil was about 20 pCi/g. For air samples, the LLD varied from about $1 \text{ E-}12 \mu\text{Ci/mL}$ for a 10-min sample and a 10-min count immediately after completion of sample collection without waiting for decay, to $2 \text{ E-}15 \mu\text{Ci/mL}$ for a 5-hour sample, a 4-day decay time, and a 2-hour count. For smears, a 10-min count over a 32-in. by 48-in. area gave an LLD of about 5 pCi of removable activity on the filter or 0.0005 pCi/cm^2 in terms of average concentration collected.

The LLD for a 0.1-gram soil sample is 15 and 40 pCi/g for 60- and 10-min spectra, respectively. The overall counting efficiency for a 0.1-gram soil sample is about 20%, including all physical losses, attenuation, and scattering. This permits quantitative determination of the TRUs to be made in soil and settled dusts, with an accuracy almost as good as the statistics of the measurement will permit.

REFERENCES

1. H. Hotzl and R. Winkler, "Large Area Gridded Ionization Chamber and Electrostatic Precipitator. Application to Low-Level Alpha Spectrometry of Environmental Samples," Nucl. Inst. Methods 150, pp. 177-181, 1978.
2. H. Hotzl and R. Winkler, "Experiences With Large-Area Frisch Grid Chambers in Low-Level Alpha Spectrometry," Nucl. Inst. and Methods in Phys. Res. 223, pp. 290-294, 1984.
3. R. V. Osborne and C. R. Hill, "High Resolution Alpha Particle Spectroscopy at Very Low Specific Activities," Nucl. Inst. and Methods 29, pp. 101-106, 1964.
4. A. Brodsky and R. G. Gallagher, "Statistical Considerations in Practical Contamination Monitoring," Radiation Protection Management 8, pp. 64-78, 1991.
5. L. A. Currie, "Limits for Qualitative Detection and Quantitative Determination," Anal. Chem. 40, pp. 586-593, 1968.
6. C. W. Sill and F. D. Hindman, "Preparation and Testing of Standard Soils Containing Known Quantities of Radionuclides," Anal. Chem. 46, pp. 113-118, 1974.
7. C. W. Sill, "Long-Term Stability of Solid Standards for Radiochemical Analysis," Anal. Chem. 61, pp. 2255-2258, 1989.
8. C. W. Sill and D. S. Sill, "Determination of Actinides in Nuclear Wastes and Reference Materials for Ores and Mill Tailings," Waste Management 9, pp. 219-229, 1989.

**DATE
FILMED**

6/15/92