53
3-31-92 JSQ

PREPARED FOR THE U.S. DEPARTMENT OF ENERGY,
UNDER CONTRACT DE-AC02-76-CHO-3073

PPPL-2829
UC-426

PPPL-2829

MIDPLANE FARADAY ROTATION: A DENSITOMETER FOR BPX

BY

F.C. JOBES AND D.K. MANSFIELD

February 1992

PPPL

PRINCETON
PLASMA PHYSICS
LABORATORY

NOTICE

This report was prepared as an account of work sponsored by an agency of the United States Government. Neither the United States Government nor any agency thereof, nor any of their employees, makes any warranty, express or implied, or assumes any legal liability or responsibility for the accuracy, completeness, or usefulness of any information, apparatus, product, or process disclosed, or represents that its use would not infringe privately owned rights. Reference herein to any specific commercial produce, process, or service by trade name, trademark, manufacturer, or otherwise, does not necessarily constitute or imply its endorsement, recommendation, or favoring by the United States Government or any agency thereof. The views and opinions of authors expressed herein do not necessarily state or reflect those of the United States Government or any agency thereof.

NOTICE

This report has been reproduced directly from the best available copy.

Available to DOE and DOE contractors from the:

Office of Scientific and Technical Information
P.O. Box 62
Oak Ridge, TN 37831;
Prices available from (615) 576-8401.

Available to the public from the:

National Technical Information Service
U.S. Department of Commerce
5285 Port Royal Road
Springfield, Virginia 22161
703-487-4650

Midplane Faraday Rotation: A Densitometer for BPX

F. C. Jobes and D. K. Mansfield
Princeton Plasma Physics Laboratory
Princeton, N. J.

PPPL--2829

DE92 010006

Abstract

The density in a high field, high density tokamak such as BPX can be determined by measuring the Faraday rotation of a $10.6 \mu\text{m}$ laser directed tangent to the toroidal field. If there is a horizontal array of such beams, then $n_e(R)$ can be readily obtained with a simple Abel inversion about the center line of the tokamak. For BPX operated at full field and density, the rotation angle would be quite large -- about 75° per pass. A layout in which a single laser beam is fanned out in the horizontal midplane of the tokamak, with a set of retro-reflectors on the far side of the vacuum vessel, would provide good spatial resolution, depending only upon the number of reflectors. With this proposed layout, only one window would be needed. Because the rotation angle is never more than 1 "fringe," the data is always good, and it is also a continuous measurement in time. Faraday rotation is dependent only upon the plasma itself, and thus is not sensitive to vibration of the optical components.

Simulations of the expected results show that BPX would be well served even at low densities by a Midplane Faraday Rotation densitometer of ~ 64 channels. Both TFTR and PBX-M would be suitable test beds for the BPX system.

MASTER

EP

Introduction

The measurement of the density in a high field, high density tokamak such as BPX¹ will be very difficult with standard techniques; interferometers must be of such short wavelength as to be extremely sensitive to vibration, and, on BPX at least, restricted in access to all regions of the plasma; Thomson scattering doesn't provide continuous coverage, and generally must be normalized to an interferometer. However, the density can be determined by measuring the Faraday rotation² of a 10.6 μm laser³ directed tangent to the toroidal field. If there is a horizontal array of such beams, then $n_e(R)$ can be readily obtained with a simple Abel inversion about the center line of the tokamak. For BPX operated at full field and density, the rotation angle would be quite large -- about 75° per pass.

Faraday rotation in the midplane involves much simpler physics than the perpendicular Faraday rotation^{4,5} used to measure poloidal field. For "small" angles between the direction of propagation and the magnetic field, the modes excited in the plasma are very nearly circularly polarized. [For 10 μm , "small" means $\theta < 89^\circ$ (or $> 91^\circ$).] A linearly polarized beam excites both equally; upon exiting from the plasma, the two circularly polarized modes sum back to a linearly polarized beam. A slight difference in propagation velocity between the two modes causes a phase shift ϕ between the right and left hand modes, and this translates into a rotation of $\phi/2$ in the exiting linearly polarized beam. In the vicinity of the optical components, the plasma density (if any) is very low, and the velocity difference between the two modes is essentially zero. Thus, any motion of the optical components will not contribute to the phase shift ϕ ; therefore, the density measurement is insensitive to vibration of the optics.

One way to implement this concept would be to utilize a large number of lasers, each looking at a different tangent radius. A more elegant approach would be to fan a single laser beam out in the midplane of the tokamak, and mount cube-corner retro-reflectors on the far side of the vacuum vessel from the dispersing optics. The returning set of beams would follow the optical path of the incoming beam back to an optical separator near the laser. Figure 1 is a schematic diagram of this implementation in BPX, showing a few of the 50 $\rightarrow$ 100 retro-reflected beams that would be required to give 3.2 $\rightarrow$ 1.6 cm resolution. For such a two pass system, the rotation would be up to 150°, for the beam that just misses the inner wall.

In BPX, only radial access is available. Figure 2 shows a schematic of how the dispersing optics might work in this situation. The laser beam, which has been expanded in the horizontal plane,

emerges from some relay optics, passes through the vacuum window and on to the dispersing optics. These consist of a plane mirror at $\sim 23^\circ$ to the beam and a cylindrically concave mirror mounted on the opposite wall of the port duct. The plane mirror deflects the expanded beam $\sim 46^\circ$ onto the cylindrical mirror, which disperses the light into a $\sim 32^\circ$ wide fan. This fan is more than wide enough to cover the whole plasma in the midplane. Together, the two mirrors utilize the full width of the duct, but are not very high. The alignment of the mirrors is not very critical, since they only have to produce a fan of laser light. They could be mounted in a separate periscope which could be installed (and replaced) by a remote manipulator. (In other machines, with tangential access, or at least windows close to the plasma, the separate measurement beams could be achieved by fanning one or several laser beams, or by using an array of lasers, or some combination.)

If, somehow, there was continuous retro-reflection from the far wall, then the returning light would go back to the beamsplitter along the path it took going out, and it would then go to the detector as a wide, but not high, beam of light, in which the angle of polarization would vary across the beam. In practice, there would be a finite set of cube-corner retro-reflectors, so the returning light would consist of a set of beams, each with a different polarization. The use of retro-reflectors of $\sim 1 \rightarrow 2$ cm width, (which will be forced upon us by Fraunhofer diffraction) will permit us to displace the returning beams vertically a few centimeters, so that they are separated from the incoming beam. Returning beams, which are neighbors in the horizontal plane, could be displaced vertically by different amounts, and thus could be resolved from each other, even in the presence of a fair amount of diffraction.

The retro-reflectors would not be uniformly placed around the far wall, because of port ducts and sundry equipment mounted at the ports. In some cases, retro-reflectors could be mounted on the sides of port ducts or the sides of other equipment. The resulting non-uniformity could be of use in identifying the retro-reflector belonging to each of the returning beams.

A simple way to measure the polarization angles of all the returning beams would be to rotate the polarization of the laser beam at a frequency f_r , and to pass the returning beams through a polarizing filter before being detected. The detected signals would be modulated at $2f_r$, and would have a phase shift ϕ which is equal to the difference in phase of the two nearly circular modes excited in the plasma; the reference phase angle would be the beam not passing through the plasma. The rotation can be realized either with a simple acousto-optic

or magneto-optic modulator, or by intracavity modulation of the laser output, which would result in an extremely high f_r (~100 MHz). Figure 3 is a schematic of the laser and detector system; the out-going and returning beams are in different planes, having been displaced by the retro-reflectors. With a rotating plane of polarization, the left and right circularly polarized components have a slightly different frequency. Because of this, phase measurements are no longer completely free of the effects of vibration; the resultant variation in phase is $\delta\phi = 2\pi \cdot \delta l \cdot f_r / c$. The amplitude of the vibration δl and the required phase resolution will determine the maximum rotation frequency f_r .

It makes sense to specify a large number of channels, since channels are cheap in this scheme. If one chooses retro-reflectors that displace the returning beams into i different horizontal planes, and if there are k different lasers, each with a different rotation frequency and each illuminating a different sector of the vacuum vessel, then there could be $i \times k$ different channels, but only k lasers and i detectors. Each of the detectors would have information in k different frequencies; this could be decoded in parallel with k phase detectors, or sequentially with one detector.

The Abel inversion is straightforward. The rotation is the integral of $n_e(R) \cdot R_t \cdot dl / R^2$; the function returned by Abel inversion is $n_e(R_t) / R_t$, where R_t is the tangent radius of the beam in question. In the case of uniform density the rotation angle is proportional to the toroidal angle subtended by the (returning) beam in passing through the plasma. Indeed, for BPX at full field ($BR = 21$ T-m) and 80% full density ($n_e = 7.8 \times 10^{20}$) the rotation angle is the subtended angle. If the toroidal beta is high enough to worry about, the returned function is $(1 - \beta_T) \cdot n_e(R_t) / R_t$.

In order to determine whether the inversion can be done fast enough to provide real time density values for density (gas injection) feedback, we have done some timing of calculation times, using a VAX 6420 as an analog of a future dedicated microprocessor. A 64 point Abel inversion takes 2.4 msec; 32 points, 0.6 msec. If the polynomial fit is included, 64 channels takes 4.8 msec, and 32 takes 1.8 msec. On TFTR, the gas injection system samples density every 8 msec, so that these calculations are fast enough to provide real time values of either n_e-l , peak n_e , or $n_e(R)$ for feedback purposes. With a polynomial fit, it is likely that n_e-l by itself can be calculated in ~1 msec.

Simulated Phase Shifts and Abel Inversions.

In Figures 4-10, some examples of assumed (solid line) and returned (dashed line) density profiles are shown in a), and the

resultant Faraday rotation angles in b), respectively. Figure 4 is a standard BPX density profile -- a parabola to the 0.5 power, with $n_{e0} = 6 \times 10^{20}/\text{m}^3$ (60% of rated density). The density profile is shifted outward in a rough simulation of a Shafranov shift of 30 cm. (Note: the BPX density profiles, Figures 4a-7a, are all shown in units of $10^{20}/\text{m}^3$; the phase graphs are all in radians.) Figure 5 is a flat density profile at full density ($10.0 \times 10^{20}/\text{m}^3$) to show the largest expected phase shift -- 4.2 radians, $\sim 240^\circ$. Figure 6 is a peaked profile. Figures 4-6 all have 64 channels. Figure 7 is a test of the sensitivity of this technique for BPX at 1% full density ($0.1 \times 10^{20}/\text{m}^3$ peak, parabola to the 0.5 power). The phase signal has been "digitized" at 10 bits full scale. In Figure 7, the stepped curve is the digitized signal, re-scaled back to radians, and the smooth dashed curve is a least squares polynomial fit (from IMSL⁶) of this data. The dashed curve in the density graphs is an Abel inversion of the polynomial fit and the solid curves are the "real" values. The line integral density n_{e-l} calculated from the fit is within 1% of the "real" value. The great accuracy comes from knowing just where the steps in the staircase are; i.e., from having a lot of channels.

TFTR

Since Figure 7 shows good (simulated) low density sensitivity for BPX, at values of $n_e BR$ lower than typical TFTR conditions, TFTR would make a suitable test bed for a BPX system; the tested system would, in turn, provide TFTR with fine resolution, continuous and unambiguous density measurements. Figures 8-10 show the results of TFTR modeling. Figure 8 is a simulation of a typical TFTR "Supershot" density profile (albeit with somewhat low density), and Figures 9 and 10 are a low density target plasma case. (The TFTR density profiles are all in units of $10^{19}/\text{m}^3$.) Figures 8 and 9 have 64 channels, and Figure 10 has 32 channels. In spite of the extremely low bit count of some of the data (4 in Figs 9 and 10), accuracies of 2% in n_{e-l} are achieved for TFTR at low densities, for the 64 channel case (or 6% for 32 channels); the higher density case (Figure 8) is even better. In practice on TFTR, there would be a gain of 4 amplifier after the phase detector (since full scale rotation on TFTR is about 1/4 that of BPX), equivalent to using 12 bits digitization. This would make the density profiles just as accurate with fewer channels.

Figure 11 shows a possible layout in TFTR. Because of existing hardware (especially armor plates protecting the vacuum vessel from accidental neutral beam blasts), the layout is not so simple as that proposed for BPX. The scheme consists of two laser beams, one shining through port A and one through the pump duct at port C; each

is expanded into a horizontal fan illuminating 2 regions of the vacuum vessel -- $P \rightarrow Q$, illuminated by both, and $R \rightarrow S$, illuminated by the port A laser.

This arrangement of the three fans provides continuous coverage in R (tangency radius), except for a 24 cm gap between 2.34 and 2.58 m and a 13 cm gap between 3.15 and 3.33 m. If 15 retro-reflectors were mounted in the $P \rightarrow Q$ region (about 1 every 10 cm) there would be 5 cm resolution in R ($R = 1.67 \rightarrow 2.34$ m).

The laser beams are shown fanning across port covers S and P, which have midplane diagnostics mounted on them. In order to avoid shining the lasers directly into the lens of the plasma TV at bay P, as well as the other diagnostics, the laser beams could be $5 \rightarrow 10$ cm below the midplane; however, all standard TFTR windows are opaque to 10μ radiation, so this precaution might not be necessary. The retro-reflectors mounted on the port covers would presumably have to be located directly inward from the various shutters mounted there already.

PBX-M and CDX-U

In a test on PBX-M, a 10μ laser would have about 5 mrad (peak) rotation, and a FIR laser at 100μ about .5 rad. Although the half radian phase shift is desirable, a FIR test on any machine would be a very big project. Alternatively, doing it with a CO_2 system would, in principle, only require a gain of 100 following the phase detector, plus some very careful balancing on the electrical lengths of the measurement and reference paths. Also, if vibrations of ~ 1 mm are expected, then the rotation frequency would have to be ~ 1 MHz. In practice, however, there would probably be other systematic effects of comparable (or larger) size to confuse the data. However, if one could learn how to overcome these effects on PBX-M, the BPX system would then have extremely good dynamic range, and could probably pick out very fine density fluctuations.

A test on CDX would require using 2mm radiation (for ~ 1 radian rotation of the reflected signal), which would be a completely different type of measurement.

Conclusions

There are a great many advantages of this scheme:

- 1) The data is always good, because the rotation angle is never more than 1 "fringe"; the quality of the data at any one time does

not depend upon the history of the measurement, as it does in an interferometer.

- 2) It is a continuous measurement in time.
- 3) An Abel inversion about the machine center line is much simpler than an inversion about a wandering magnetic axis.
- 4) High spatial resolution can be achieved, limited only by the number of retro-reflectors.
- 5) The rotation angles can be detected by phase detectors, which are a simple and well known technology.
- 6) Only one window required (for proposed BPX layout).
- 7) The effect is insensitive to vibration of the components, allowing them to be mounted inside the vacuum vessel.
- 8) No massive anti-vibration frame required.
- 9) The components are readily available and relatively inexpensive.
- 10) This technique enjoys all the advantages of the "standard" Faraday rotation, with none of the disadvantages.
- 11) Real time calculations of peak density, or of $n_e(R)$ could be made available for density feedback; these signals would always be good, because there is no "fringe" jump to worry about.

BPX would be well served even at low densities by a Midplane Faraday Rotation densitometer of ~ 64 channels. TFTR would be a suitable test bed for the BPX system, especially since the questions have to do with low density operation and the ability of some of the components to stand up to the environment inside a fusion reactor. A test in PBX-M could also provide information concerning the feasibility of this proposed densitometer, as well as providing PBX-M with another density diagnostic.

Acknowledgement

This work was supported by the U.S. Department of Energy, contract no. DE-AC-02-76CHO3073.

References

1. R.J. Goldston, G.H. Nielson, *et. al.* BPX Project Report #X-910331-PPL-22 (1991).
R.J. Goldston, G. Bateman, *et. al.* in 18th European Conf. on Controlled Fusion and Plasma Physics, Berlin, Germany, Europhysics Conf. Abstracts Vol. 15C, Part 1, 1-93 (1991).
2. F. De Marco and S.E. Segre, *Plasma Phys.* **14**, 245 (1972).
3. D.K. Mansfield, H.K. Park, *et. al.*, *Appl. Opt.* **26**, 4469 (1987).
4. D.P. Hutchinson, C.H. Ma, P.A. Staats, and K.L. Vander Sluis, *Nucl. Fusion* **21**, 1535 (1981).
5. H. Soltwisch and T.F.R. Equipe, *Infrared Physics* **21**, 287 (1981).
6. *The IMSL Statistics Library User's Manual.*, International Mathematical and Statistical Libraries, Inc., Houston, Texas (1987).

Figure Captions

- Figure 1. Schematic layout of laser beams in BPX.
- Figure 2. Fan-out scheme for BPX.
- Figure 3. Possible arrangement of optical components. Note that the outgoing and returning beams are in different planes, so that the returning beam is reflected by a mirror, not a beam splitter.
- Figure 4. Assumed and returned BPX density profiles (a) and associated phase shift (= twice rotation angle) (b). $n_{e0} = 6.00 \cdot 10^{20}/\text{m}^3$; $n_{e1} = 6.96 \cdot 10^{20}/\text{m}^2$; standard BPX profile; $B_T = 8.1 \text{ T}$; 64 channels.
- Figure 5. Assumed and returned BPX density profiles (a) and associated phase shift (= twice rotation angle) (b). $n_{e0} = 10.0 \cdot 10^{20}/\text{m}^3$; $n_{e1} = 12.6 \cdot 10^{20}/\text{m}^2$; broad profile; $B_T = 8.1 \text{ T}$; 64 channels.
- Figure 6. Assumed and returned BPX density profiles (a) and associated phase shift (= twice rotation angle) (b). $n_{e0} = 6.00 \cdot 10^{20}/\text{m}^3$; $n_{e1} = 4.37 \cdot 10^{20}/\text{m}^2$; narrow profile; $B_T = 8.1 \text{ T}$; 64 channels.
- Figure 7. Assumed and returned BPX density profiles (a) and phase shift (= twice rotation angle) (b). The phase associated with the assumed density is a solid line; the stepped dashed line is the "digitized" phase data; the smooth dashed line is a polynomial fit to the data. The returned density curve is an Abel¹ inversion of the polynomial fit. $n_{e0} = 0.10 \cdot 10^{20}/\text{m}^3$; $n_{e1} = 0.12 \cdot 10^{20}/\text{m}^2$; standard BPX profile; $B_T = 8.1 \text{ T}$; 64 channels; 6th order polynomial fit.
- Figure 8. Assumed and returned TFTR density profiles (a) and phase shift (= twice rotation angle) (b). The phase associated with the assumed density is a solid line; the stepped dashed line is the "digitized" phase data; the smooth dashed line is a polynomial fit to the data. The returned density curve is an Abel inversion of the polynomial fit. $n_{e1} = 5.00 \cdot 10^{19}/\text{m}^2$; $n_{e0} = 6.54 \cdot 10^{19}/\text{m}^3$; profile is parabola squared; $B_T = 5.0 \text{ T}$; 64 channels; 6th order polynomial fit.

Figure 9. Assumed and returned TFTR density profiles (a) and phase shift (= twice rotation angle) (b). The phase associated with the assumed density is a solid line; the stepped dashed line is the "digitized" phase data; the smooth dashed line is a polynomial fit to the data. The returned density curve is an Abel inversion of the polynomial fit. $n_{e1} = 1.00 \cdot 10^{19}/\text{m}^2$; $n_{e0} = 1.31 \cdot 10^{19}/\text{m}^3$; profile is parabola squared; $B_T = 5.0 \text{ T}$; 64 channels; 5th order polynomial fit.

Figure 10. Same as Figure 9, except 32 channels.

Figure 11. A possible TFTR layout. Shaded areas show regions that could be measured by beams at ports A and C. The gap in the coverage of the port A laser beam is the result of there not being any retro-reflectors in neutral beam port Q.

Fig. 1: BPX Geometry

Fig. 2: Fan-out Optics

Fig. 3: Optics Layout

Fig. 4a. BPX Density Profiles: — Assumed profile
 - - - Returned profile

Fig. 4b. Phase Shift

Fig. 5a. BPX Density Profiles: — Assumed profile
 - - - Returned profile

Fig. 5b. Phase Shift

Fig. 6a. BPX Density Profiles: — Assumed profile
 - - - Returned profile

Fig. 6b. Phase Shift

Fig. 7a. BPX Density Profiles: — Assumed profile
 - - - Returned profile

Fig. 7b. Phase Shifts: — Actual Phase shift:
 - - - Returned Phase shifts

Fig. 8a. TFTR Density Profiles: — Assumed profile
 - - - Returned profile

Fig. 8b. Phase Shifts. — Actual Phase shift
 - - - Returned Phase shifts

Fig. 9a. TFTR Density Profiles: — Assumed profile
 - - - Returned profile

Fig. 9b. Phase Shifts. — Actual Phase shift
 - - - Returned Phase shifts

Fig. 10a. TFTR Density Profiles: — Assumed profile
 - - - Returned profile

Fig. 10b. Phase Shifts. — Actual Phase shift
 - - - Returned Phase shifts

Fig. 11 TFTR Layout

EXTERNAL DISTRIBUTION IN ADDITION TO UC-420

Dr. F. Paoloni, Univ. of Wollongong, AUSTRALIA
 Prof. M.H. Brennan, Univ. of Sydney, AUSTRALIA
 Plasma Research Lab., Australian Nat. Univ., AUSTRALIA
 Prof. I.R. Jones, Flinders Univ, AUSTRALIA
 Prof. F. Cap, Inst. for Theoretical Physics, AUSTRIA
 Prof. M. Heindler, Institut für Theoretische Physik, AUSTRIA
 Prof. M. Goossens, Astronomisch Instituut, BELGIUM
 Ecole Royale Militaire, Lab. de Phy. Plasmas, BELGIUM
 Commission-Europeen, DG. XII-Fusion Prog., BELGIUM
 Prof. R. Boudiqué, Rijksuniversiteit Gent, BELGIUM
 Dr. P.H. Sekanaka, Instituto Fisica, BRAZIL
 Instituto Nacional De Pesquisas Espaciais-INPE, BRAZIL
 Documents Office, Atomic Energy of Canada Ltd., CANADA
 Dr. M.P. Bachynski, MPB Technologies, Inc., CANADA
 Dr. H.M. Skarsgard, Univ. of Saskatchewan, CANADA
 Prof. J. Teichmann, Univ. of Montreal, CANADA
 Prof. S.R. Sreenivasan, Univ. of Calgary, CANADA
 Prof. T.W. Johnston, INRS-Energie, CANADA
 Dr. R. Bolton, Centre canadien de fusion magnétique, CANADA
 Dr. C.R. James, Univ. of Alberta, CANADA
 Dr. P. Lukáč, Komenského Univerzita, CZECHO-SLOVAKIA
 The Librarian, Culham Laboratory, ENGLAND
 Library, R61, Rutherford Appleton Laboratory, ENGLAND
 Mrs. S.A. Hutchinson, JET Library, ENGLAND
 Dr. S.C. Sharma, Univ. of South Pacific, FIJI ISLANDS
 P. Mähönen, Univ. of Helsinki, FINLAND
 Prof. M.N. Bussac, Ecole Polytechnique, FRANCE
 C. Moutet, Lab. de Physique des Milieux Ionisés, FRANCE
 J. Radet, CEN/CADARACHE - Bat 506, FRANCE
 Prof. E. Economou, Univ. of Crete, GREECE
 Ms. C. Rinri, Univ. of Ioannina, GREECE
 Dr. T. Mui, Academy Bibliographic Ser., HONG KONG
 Preprint Library, Hungarian Academy of Sci., HUNGARY
 Dr. B. DasGupta, Saha Inst. of Nuclear Physics, INDIA
 Dr. P. Kaw, Inst. for Plasma Research, INDIA
 Dr. P. Rosenau, Israel Inst. of Technology, ISRAEL
 Librarian, International Center for Theo Physics, ITALY
 Miss C. De Palo, Associazione EURATOM-ENEA, ITALY
 Dr. G. Grosso, Istituto di Fisica del Plasma, ITALY
 Prof. G. Rostangni, Istituto Gas Ionizzati Del Cnr, ITALY
 Dr. H. Yamato, Toshiba Res. & Devel. Center, JAPAN
 Prof. I. Kawakami, Hiroshima Univ., JAPAN
 Prof. K. Nishikawa, Hiroshima Univ., JAPAN
 Director, Japan Atomic Energy Research Inst., JAPAN
 Prof. S. Itoh, Kyushu Univ., JAPAN
 Research Info. Ctr., National Inst. for Fusion Science, JAPAN
 Prof. S. Tanaka, Kyoto Univ., JAPAN
 Library, Kyoto Univ., JAPAN
 Prof. N. Inoue, Univ. of Tokyo, JAPAN
 Secretary, Plasma Section, Electrotechnical Lab., JAPAN
 S. Mori, Technical Advisor, JAERI, JAPAN
 Dr. O. Mitarai, Kumamoto Inst. of Technology, JAPAN
 Hyeon-Sook, Korea Atomic Energy Research Inst., KOREA
 D.I. Choi, The Korea Adv. Inst. of Sci. & Tech., KOREA
 Prof. B.S. Liley, Univ. of Waikato, NEW ZEALAND
 Inst of Physics, Chinese Acad Sci PEOPLE'S REP. OF CHINA
 Library, Inst. of Plasma Physics, PEOPLE'S REP. OF CHINA
 Tsinghua Univ. Library, PEOPLE'S REPUBLIC OF CHINA
 Z. Li, S.W. Inst. Physics, PEOPLE'S REPUBLIC OF CHINA
 Prof. J.A.C. Cabral, Instituto Superior Tecnico, PORTUGAL
 Dr. O. Petrus, AL I GUZA Univ., ROMANIA
 Dr. J. de Villiers, Fusion Studies, AEC, S. AFRICA
 Prof. M.A. Hellberg, Univ. of Natal, S. AFRICA
 Prof. D.E. Kim, Pohang Inst. of Sci. & Tech., SO. KOREA
 Prof. C.I.E.M.A.T, Fusion Division Library, SPAIN
 Dr. L. Stenflo, Univ. of UMEA, SWEDEN
 Library, Royal Inst. of Technology, SWEDEN
 Prof. H. Wilhelmson, Chalmers Univ. of Tech., SWEDEN
 Centre Phys. Des Plasmas, Ecole Polytech, SWITZERLAND
 Bibliotheek, Inst. Voor Plasma-Fysica, THE NETHERLANDS
 Asst. Prof. Dr. S. Cakir, Middle East Tech. Univ., TURKEY
 Dr. V.A. Glukhikh, Sci. Res. Inst. Electrophys. Apparatus, USSR
 Dr. D.D. Ryutov, Siberian Branch of Academy of Sci., USSR
 Dr. G.A. Eliseev, I.V. Kurchatov Inst., USSR
 Librarian, The Ukr.SSR Academy of Sciences, USSR
 Dr. L.M. Kovrizhnykh, Inst. of General Physics, USSR
 Kernforschungsanlage GmbH, Zentralbibliothek, W. GERMANY
 Bibliothek, Inst. Für Plasmaforschung, W. GERMANY
 Prof. K. Schindler, Ruhr-Universität Bochum, W. GERMANY
 Dr. F. Wagner, (ASDEX), Max-Planck-Institut, W. GERMANY
 Librarian, Max-Planck-Institut, W. GERMANY
 Prof. R.K. Janev, Inst. of Physics, YUGOSLAVIA

END

**DATE
FILMED**

4 / 21 / 92