December 15, 1989

DI

Progress Report on Fluid Dynamics of Supercritical
Helium within Internally Cooled Cabled Superconductors

DOE/ER/52149--2

DE92 008745

DOE Contract No. DE-FG02-88ER52149
Principal Investigator: Steven W. Van Sciver

The first quarter of the research program was spent largely on preparing for the preliminary examination by the graduate student, Mark Daugherty. A literature search was carried out to review the existing work in fluid mechanics and heat transfer related to pressure drop and quenching in cable-in-conduit conductors. The scope of the doctoral thesis proposal was then developed to extend the existing knowledge and meet the requirements outlined in the research proposal submitted to DOE. The preliminary examination was passed and the research proposal was thereby approved by the Mechanical Engineering Department in April of 1989. A copy of the report written for the preliminary examination is being forwarded under a separate cover.

During May and June the internal cryostat wiring system was designed and the mechanical components required for heat sinking were assembled. The system uses modular plug-in connectors to facilitate the installation of different test sections and is designed to easily accommodate future experiments carried out in the Liquid Helium Flow Facility.

Preliminary design work was also carried out to estimate the power required for the heat pulse experiment. This experiment will use an inductive energy pulse to simulate the events occurring at the start of a quench in a cable-in-conduit conductor. Initial estimates were made to determine the penetration of the magnetic field through the conduit wall and into the cable. It appears that power requirements for the inductive pulse will be on the order of several hundred thousand watts for a duration of a few milliseconds.

We also performed an experiment with a helium II heat exchanger during June and July to develop experience with helium II experimental techniques. The experiment

MASTER

was reported in the paper: "Design and Performance of Forced-Flow He II Heat Exchangers," A. Shajii, Y. Huang, M. Daugherty, R. J. Witt, S. W. Van Sciver, 1989 CEC/ICMC, Paper-FF-07, July 24-28, Los Angeles, CA. (A copy will be forwarded separately.)

During July a room temperature pressure drop rig was assembled to aid in the characterization of conductor geometries. The room temperature pressure drop measurements can be relatively easily conducted and repeated as required to assist in the development of methods to characterize the geometry of the cable-in-conduit conductors. The resulting room temperature correlations should provide a starting point in the development of the low temperature helium pressure drop correlations.

During July and August the selection of the flow meters, sensors, couplings, flanges and connectors required for the main pressure drop experiments was finalized. All items from external vendors required for the pressure drop experiments have been received.

Much of August was spent developing programs for a new data acquisition system based on the Macintosh II computer. Programs were written to collect and average steady state data and to acquire transient data. A series of programs were developed to display data on the computer screen as it was acquired. These programs promise to prove most useful during setup and debugging of the data acquisition systems for future experiments. In addition data acquired on the new system will be directly available for use with our Macintosh data analysis programs.

September was devoted to preparing the mechanical components required for the first experiment. The test sections received from Lawrence Livermore National Laboratory were fitted with connecting flanges, pressure taps and thermometer mounts. The venturi flow meters were fitted with connecting flanges and pressure transducer couplings. Bellows sections and adaptor sections were made and fitted to the pump and heat exchanger in the Liquid Helium Flow Facility.

October was used to become familiar with the operation of the Validyne pressure

transducers. These transducers will be used cold for differential pressure measurements. Cables were made to monitor these transducers during calibration and to test their performance when connected to the cryostat internal wiring system. Estimates were made of the pressure drop to be expected from each test section and pressure transducers were selected for each test section based on the magnitude of the expected pressure drops.

We are now in the process of setting up a differential pressure transducer calibration rig. The rig, which will be used to calibrate the Validyne pressure transducers in liquid helium, was modified to incorporate a Texas Instruments Precision Pressure gauge as the reference for the differential pressure measurements. Differential pressures from zero to two atmospheres can be applied by the rig.

Our current schedule calls for an experiment in the Liquid Helium Flow Facility by late December 1989 or early January 1990. The experiment will be conducted in the Liquid Helium Flow Facility to measure the pressure drop with supercritical helium flowing in three test sections over a wide range of Reynolds numbers. In addition we will attempt to measure pressure drops with helium II as the fluid in the sections. It is not known if the present heat exchanger will be able to provide enough cooling to permit all of these measurements.

INSTRUCTIONS

Who uses this form: DOE contract administrators will designate the DOE contractors and grantees who are to use this form. Generally speaking, it will be used by educational institutions and other not-for-profit institutions. DOE national laboratories and other major contractors will generally use the longer Form DOE F 1332.15.

When to use: DOE contractors are required under their contracts to transmit specified types of documents to the DOE. Some, but not all, of these are transmitted by DOE contract administrators to the DOE Office of Scientific and Technical Information (OSTI) and may be incorporated into the DOE technical information documentation system. Types of documents that will be transmitted to OSTI are identified in instructions that the contractor receives from his contract administrator. Each document title is to be accompanied by one copy of this transmittal form recommending to OSTI appropriate announcement and distribution of the document. Documents which the contractor may be required to submit to the DOE under his contract but which are not of the type to be transmitted to OSTI, e.g., contract proposals, should not be accompanied by a copy of this transmittal form. Questions concerning this form may be referred to OSTI on FTS 626-1268 or commercial 615-576-1268.

Where to send: Send this form with the document attached to the DOE contract administrator for transmittal to OSTI unless the DOE contract administrator specified otherwise.

Item instructions:

Item 1. The DOE report number will be constructed as follows:

- (a) Use DOE and the final seven characters (two letters and five digits) from the applicable contract or grant number. A slash mark must separate the letters from the digits. Complete the number with dashes followed by a sequential number for each report generated under the contract. Example for the first report generated under contract number DE-AC03-79NE01834:
DOE/NE/01834-1

- (b) Reports that are issued in multivolume, parts, or revisions, etc., should be numbered as follows:

DOE/NE/01834-1 Vol. 1
DOE/NE/01834-1 Pt. 1
DOE/NE/01834-1 Rev. 1

- (c) **Caution:** Report numbers are to be structured exactly as specified. Modifications, if essential, must be approved by OSTI.

Item 2. Self-explanatory.

Item 3. Give title exactly as on the document itself.

Item 4. Self-explanatory.

Item 5. The unrestricted, unlimited distribution includes abstracting in *Energy Research Abstracts (ERA)* and distribution to appropriate DOE/TIC-4500, Standard Distribution of Unclassified Scientific and Technical Reports addressees, to libraries that through purchase of microfiche maintain collections of DOE reports, and to the National Technical Information Service (NTIS) for sale to the public. Check box b or c if there is need for limiting announcement and distribution procedures described above.

Item 6. If box 5b or 5c is checked, give reasons.

Item 7. Self-explanatory.

Item 8. Enter information about the person to whom inquiries concerning the recommendations on this form may be addressed.

Item 9. DOE contract administrator or patent group representative should check box a or b and forward this form and document to:

Office of Scientific and Technical Information
U. S. Department of Energy
Technical Information Center
P. O. Box 62
Oak Ridge, TN 37831

END

**DATE
FILMED**

4/10/92