DOE/ER/53214--T1

DE92 013133

PROGRESS REPORT AND
REQUEST FOR CONTINUED FUNDING

NOTICE FOR
HANDLING PROPOSALS

"This proposal is to be used only for DOE evaluation purposes and this notice shall be affixed to any reproduction or abstract thereof. All Government and non-Government personnel handling this proposal shall exercise extreme care to insure that the information contained herein is not duplicated, used, or disclosed in whole or in part for any purpose other than to evaluate the proposal, without the written permission of the offeror (except that if a contract is awarded on the basis of this proposal, the terms of the contract shall control disclosure and use). This notice does not limit the Government's right to use information contained in the proposal if it is obtainable from another source without restriction. This is a Government notice, and shall not by itself be construed to impose any liability upon the Government or Government personnel for any disclosure or use of data contained in this proposal."

TO: U.S. Department of Energy
Division of Magnetic Fusion Energy
Washington, D.C. 20545

FROM: The Johns Hopkins University
Charles and 34th Streets
Baltimore, Maryland 21218

PROJECT: Extreme Ultraviolet and Soft X-ray Diagnostics of
High-Temperature Plasmas

PROJECT DIRECTOR: Principal Investigator

NAME: Warren Moos
TITLE: Professor of Physics
SOC.SEC.NO.: 060-30-5107
DEPARTMENT: Department of Physics & Astronomy

TYPE OF APPLICATION Continuation

PROPOSED STARTING DATE: January 1, 1987

PROPOSED DURATION IN MONTHS: 12

AMOUNT REQUESTED FROM SPONSOR: \$340,000

ENDORSEMENTS

PROJECT DIRECTOR

DEPARTMENT CHAIRMAN

SIGNATURE

NAME

DIVISION

TELEPHONE

Warren Moos
Warren Moos

Dept. of Physics &
Astronomy

(301) 338-7337

Lloyd Armstrong, Jr.
Lloyd Armstrong, Jr.

Dept. of Physics
& Astronomy

(301) 338-7362

DEAN

ADMINISTRATIVE OFFICER

SIGNATURE

NAME

DIVISION

TELEPHONE

George W. Fisher
George W. Fisher

Faculty of Arts & Sciences
(301) 338-8212

John Bearden
John Bearden, Director
Research Administration
(301) 338-8138

DATE

10/2/86

DOE/ER/53214--T1

ABSTRACT

This report describes recent progress and plans for calendar year 1987 in the Johns Hopkins University program to develop and improve spectroscopic diagnostics for the high temperature plasmas used in magnetic fusion research. An EUV spectrograph which provides time resolved spectra along fifteen chords of a plasma device has been completed and evaluation on DIII-D will begin in late 1986. Other instrumentation work includes the evaluation of a sensitive detector for ion temperature/velocity distribution determinations and a feasibility study of Zeeman polarimetry for determining magnetic fields. A comprehensive data set taken on the TEXT tokamak is undergoing analysis as a means of improving the ionic parameters used in diagnostic studies and to expand the capabilities of existing instruments. Potential new advances in spectroscopic technology are being monitored to determine if they provide advantages for fusion research.

DISCLAIMER

This report was prepared as an account of work sponsored by an agency of the United States Government. Neither the United States Government nor any agency thereof, nor any of their employees, makes any warranty, express or implied, or assumes any legal liability or responsibility for the accuracy, completeness, or usefulness of any information, apparatus, product, or process disclosed, or represents that its use would not infringe privately owned rights. Reference herein to any specific commercial product, process, or service by trade name, trademark, manufacturer, or otherwise does not necessarily constitute or imply its endorsement, recommendation, or favoring by the United States Government or any agency thereof. The views and opinions of authors expressed herein do not necessarily state or reflect those of the United States Government or any agency thereof.

MASTER

DISTRIBUTION OF THIS DOCUMENT IS UNLIMITED

I. INTRODUCTION

A research program to develop new types of fusion spectroscopic diagnostics has existed at this laboratory since 1976 with an emphasis on producing rugged but compact and portable instrumentation, and implementing new advances in technology for diagnostics. A major component of the program has been the development of detectors which would simultaneously measure a number of spectral and/or spatial elements. In addition, atomic physics parameters for interpreting the diagnostic data were obtained using theoretical techniques and more recently by using tokamak and mirror data. This brief report describes progress to date for this research program and plans for the remainder of the grant period which covers the calendar years 1986 and 1987 and should be considered an update on the original proposal which covered the two year period.

The report is broken into sections covering different projects. The largest project is the Space and Time Resolving Spectrograph (STRS) which simultaneously covers a number of spectral resolution elements and viewing chords in the 50-370 Å region, one of the richest for impurity emissions from the core of a tokamak. It will be installed on the DIII-D Tokamak for evaluation shortly. (Section II). A very sensitive high spectral resolution detector for ion temperature/velocity distributions has been installed on the Alcator C tokamak for evaluation. (Section III). The circular polarization analysis of impurity emission lines broadened by the Zeeman effect may permit the determination of the magnetic fields in a plasma to good accuracy with low cost apparatus. After initial feasibility tests of this technique on the TEXT tokamak, we are exploring the feasibility of this technique further (Section IV). We have used the TEXT

tokamak and three of our instruments to obtain a large data set for injected impurities. Analysis of this data is providing density sensitive line pair ratios, new information on forbidden lines and a number of other atomic physics parameters needed for better utilization of the data produced as a matter of course by modern diagnostic instrumentation (Section V). Finally several potential advances in spectroscopic technology merit monitoring and if appropriate should be used to upgrade fusion diagnostics (Section VI). A list of recent publications is given in Section VII followed by a discussion of the budget in Section VIII.

The program described in the original proposal was sufficiently broad that completion of all projects was not expected at the requested funding level. However, during the two year period we will complete several of the projects and make significant inroads into the rest. The construction of the STRS is now complete. The field evaluation will begin in October 1986 and continue through most of 1987. Analysis of the data from the field evaluation of the ion temperature detector will be completed during early 1987. Based on preliminary studies performed during this calendar year, we expect to begin construction of an apparatus for extending the studies of the Zeeman polarization study in 1987. Determination of ionic parameters using the TEXT data, the potential use of existing instruments to diagnose TEXT edge plasmas, and the new technology assessments will continue on an available effort basis. During the grant period, T-L. Yu completed his Ph.D. dissertation with support from the TMX-U project.

The information and experience in applying new technology to fusion diagnostic instrumentation gathered by this group is also utilized by workers at other DOE laboratories through publications, the move of Hopkins Postdoctoral Fellows and students to the laboratories, and by a considerable

number of personal conversations and phone contacts. Instruments constructed or owned by this laboratory have been used by a large number of DOE laboratories and also in France and Japan. As the field evaluation on these instruments is completed, personnel from Johns Hopkins University are not usually required. However the instruments do require calibration at this laboratory and occasional maintenance.

II. SPACE AND TIME RESOLVING SPECTROGRAPH

Radiation emitted by impurities in magnetic fusion devices has long been of vital interest, not only as the source of significant power loss, but also as a comprehensive diagnostic tool. Detectors are now available which permit time resolved measurements of a number of spatial and spectral elements simultaneously. These advances in detector technology allow comprehensive investigations such as the effects of auxiliary heating, power losses, impurity poloidal evolution, distributions, and impurity transport. This group has completed construction of the Space and Time Resolving Spectrograph (STRS) based on detector and optical principles demonstrated in earlier instruments and experiments in our laboratory. The STRS has a "three dimensional" (spectral, temporal and spatial) capability that will permit observations and analysis for a single shot that have previously been either impossible or extremely time consuming and expensive to obtain on a shot by shot basis.

The STRS is capable of 2 to 50 ms time resolution and can therefore produce time histories of temperature, density, etc., which are needed for evolution and transport studies. The instrument images a 15 degree viewing angle with good spatial resolution in order to obtain the spatial

distribution of radiation in a single shot. In addition to a large decrease in the number of shots required for a given study, the STRS will be able to study transient effects such as disruptions within a single shot. The large number of detector elements associated with the STRS leads to large data sets. As a consequence, a large data handling capacity is required for both the initial data reduction for operational use, and later detailed analysis. Thus, a major portion of the remaining diagnostic/development research will consist of new software tools and data handling techniques in addition to the hardware which is now complete.

The integration of the CCD's to the microchannel plate detector and in turn this detector to the spectrograph and supporting electronics was completed by mid 1986. This was accompanied by the development of the software necessary for testing and calibration. End to end testing of the complete system, including some of the data handling programs was accomplished using laboratory EUV sources. Finally photometric calibration of the complete system was performed at the NBS SURF II facility in September 1986.

During October, the instrument will be shipped to San Diego to be mounted on the DIII-D Tokamak for spatially resolved impurity studies. We emphasize that this instrumentation represents a new type of diagnostic whose capabilities need to be explored. The instrument will generate large data sets which will require the development of specialized software tools in order to optimize the extraction of information without an inordinate amount of human labor. It is expected that the evaluation process will be an evolutionary one in which experience during the evaluation will suggest additional software tools in order to more fully exploit the instrumentation. The construction of this instrument and the scientific

data from the evaluation process will form the subjects of the Ph.D. dissertations of D. Content and M. Perry.

III. DETECTOR FOR ION TEMPERATURE/VELOCITY DISTRIBUTION DIAGNOSTICS.

This laboratory has developed a new type of detector for high spectral resolution measurements of impurity line shapes broadened by the Doppler effect, a technique now used by a number of laboratories to routinely determine ion temperatures. Two major experimental difficulties are the low brightnesses of the emission lines and the presence of underlying continua which must be accurately subtracted for precise determination of the line widths. The present detector measures the spectrum in the photon counting mode across 128 pixels simultaneously, providing a significant improvement in the signal to noise and the ability to remove continua. Also, the image intensified Reticon used for this detector is geometrically stable permitting reliable measurements of the line shape (velocity distribution) and shifts (plasma flow velocity).

The detector was placed on a 1 meter Ebert system, owned by this laboratory and presently on loan to the Alcator group, which was then installed on the Alcator C Tokamak. Most of the data necessary for evaluation of the detector has been obtained. Particularly notable was the measurement of the Fe XXI 1354.1 Å line which had been undetectable with the detector system previously in use. Detailed analysis of the data for evaluation of the detector will be completed by June 1987 and will be the subject of R. Benjamin's Ph. D dissertation.

IV. FEASIBILITY OF ZEEMAN EFFECT POLARIMETRY AS A MAGNETIC FIELD DIAGNOSTIC

This technique is potentially a new low-cost method for reliably determining the internal magnetic field in high-temperature, magnetically confined plasmas, a type of measurement in which there is growing interest. In investigations of tokamak devices, the measurement of the poloidal field distribution yields the plasma current density profile which is essential for understanding the confinement, stability and energy balance of the plasma. The measurement of the internal magnetic field also is of importance in the studies of reversed field pinch (RFP) devices where the field structure and its dynamics, essential for understanding of this configuration, are known only approximately, and in the studies of high- β currentless plasmas in magnetic mirrors, stellarators, etc.

Feldman et al. [1] have suggested that the magnetic field in a tokamak can be measured by using the Zeeman splitting and polarization of the magnetic dipole radiation of heavy ions (intrinsic or injected impurities). The proposed method is a variation of the technique which has been successfully used in the measurement of solar and stellar magnetic fields [2,3]. The profile of a line emitted by a magnetically confined, high-temperature plasma is determined by two processes: Doppler broadening due to the thermal motion of the ions and the Zeeman splitting in the magnetic field. The relative effect of the magnetic field (Zeeman effect) is larger for long-wavelength transitions and heavy ions. Thus, the magnetic dipole transitions in heavy impurity ions, many of which are observed in the visible and UV spectrum of tokamak discharges [1], can be used in the measurement. The fact that the impurity ions are localized in plasma in regions where the electron temperature is approximately equal to the ionization potential of a given ion may provide sufficient spatial resolution to avoid the use of Abel inversion in order to obtain the local

emissivities. An alternative approach is to use the emissions from light impurity ions (such as stripped oxygen) due to charge exchange with a neutral beam. This approach removes any uncertainty in the location of the emitting ions, although it does require a neutral beam. Note that in both cases the measurement uses visible and near UV technology, does not require a vacuum, and is of modest cost. In addition, the technique can be used with both edge and core species and from very low to very high electron densities.

The basic principles of this technique and the encouraging results of a brief feasibility study performed on the TEXT tokamak were described in the proposal for this grant period. A simple apparatus to provide more detailed information on the accuracy of this diagnostic method will be constructed during the remainder of the grant period. Preliminary measurements on the TEXT Tokamak during the first part of the grant period have provided a evaluation of the required signal-to-noise ratio and instrumental problems associated with the proposed polarimeter which employ a scanned Fabry-Perot etalon to scan the line profiles. Also, detailed theoretical calculations of charge-exchange recombination line profiles, the results of which will be needed in the interpretation of experimental data, are being completed.

If the method proves to be a useful diagnostic, a second instrument will be designed which will measure simultaneously the polarization of several lines, with spatial and temporal resolution adequate to provide instantaneous information about the poloidal field and plasma current profile, suitable for installation and evaluation at a major DOE facility. We will also investigate the applicability of this technique to other devices particularly high beta devices such as spheromaks.

V. EXPERIMENTAL STUDY OF ATOMIC PARAMETERS FOR SPECTROSCOPIC DIAGNOSTICS

Modern spectroscopic diagnostic instruments provide a large comprehensive data base for each shot which can be utilized as a sophisticated method of measuring a number of plasma quantities. It is often the case that many more plasma parameters can be determined from the data than was contemplated when the equipment was originally installed on a machine. However, in order to exploit this capability, reliable values of the relevant atomic parameters are necessary. To a large extent, these can be supplied by theory. But critical experimental checks in well diagnosed plasmas are needed. A recent study by the National Research Council Survey on the Physics of Plasmas and Fluids chaired by R. C. Davidson and J. W. Dawson stated under Future Research Opportunities: "Quantitative plasma spectroscopy is needed to test the validity of the atomic data base and reaction kinetics, as are more precise wavelengths."

The instrumentation designed and constructed by this laboratory primarily for magnetic fusion diagnostics can also be used for measurements of this type. The absolute photometric calibration of these instruments against synchrotron radiation makes the data from these instruments particularly valuable. Although we have performed several studies of this kind using diagnostic data sets, for more precise and comprehensive work, data sets are required which have been obtained under more controlled plasma conditions than are possible on machines dedicated to exploring confinement and auxiliary heating concepts. The best facility for this purpose is the TEXT Tokamak at the University of Texas, Austin. Starting in September 1984, several instruments from the Johns Hopkins University, which covered the spectral region from 20 to 2200 Å were used to obtain simultaneous

spectral measurements during the injections of more than twenty atomic species into well characterized plasmas with electron densities from 1.5×10^{13} to 8.0×10^{13} . This allowed a comparison between theoretical and experimental determinations of electron density sensitive line brightness ratios over a wide wavelength and atomic number range. The data is useful as a calibration of spectroscopic electron density diagnostics to supplement the standard techniques, a capability which will be particularly useful in complex magnetic geometries where the density varies radically between different parts of the plasma device. In addition the study will provide the most comprehensive evaluation of the atomic calculations and level population models to date.

A number of research topics based on the data set are under investigation. During this grant period, an experimental determination of the effect of quantum mechanical resonance structures on electron excitation rates for Si III based on TMX-U and TEXT data was completed. Two studies concerned with effects of electron density on the Be-I-like isoelectronic sequence and a study of calcium have also been completed. These studies have shown that for moderate and high Z Be I-like ions, the simpler distorted wave calculations are adequate for producing diagnostic parameters, whereas for low Z the more complex R-matrix calculations are necessary. The database which has been obtained in the TEXT experiments is quite large and analysis of the data is expected to continue through the remainder of the grant period. A small additional experimental effort is expected during the remainder of the grant period in order to fill any critical gaps in this survey.

Further work is planned on additional research topics developed during the previous experiments at TEXT. A particularly attractive avenue appears

to be a more detailed exploration of the use of spectroscopic diagnostics to study the edge plasma of the TEXT Tokamak.

VI. NEW TECHNOLOGY

We are exploring new types of technology which could bring significant advances to spectroscopic diagnostics as part of our on going program. As plasma electron temperatures increase, the wavelengths of the forbidden transitions used for ion temperature determinations shift to shorter wavelengths. Below 1200 Å, there is a change in the technology caused by the need for windowless detectors and of more importance the low (but still usable) optical reflectivities of all metals when the radiation is normally incident. The construction of a high spectral resolution normal incidence grating instrument which also has a large acceptance angle to improve the sensitivity may now be feasible. We plan to perform optical ray traces for aspheric single reflection grating systems suitable for fusion to see if feasible solutions are available.

Recent advances in multilayer coatings for the soft x-ray and EUV spectral regions permit the use of normal incidence optics which cannot be used with ordinary metallic coatings in this spectral region, leading to much higher throughput, good imaging quality and possibly lower scattered light. One application would be to upgrade the normal incidence time resolving spectrograph, which now covers from 300 to 2200 Å, with such a coating in order to extend its coverage down to much shorter wavelengths.

The MCP intensified Reticon detectors developed at this laboratory appear to have important applications in x-ray diagnostics. Their pixel size is quite small, 25 μ . As a result the instrument dimensions can also be

made small, leading to a significant decrease in the weight of the lead shielding. For present designs such as those used on TFTR this can make the difference between an essentially fixed instrument and a movable instrument with more flexible applications.

A new kind of detector with potential promise for fusion diagnostics is the back thinned CCD. NASA supported studies at Lockheed and NRL indicate that there is good reason to believe that this detector will be sensitive all the way down through the ultraviolet into the soft x-ray region with essentially a 100% efficiency for detecting photons. Such a detector could provide unparalleled sensitivity for diagnostic devices with a two dimensional imaging capability. We are monitoring these programs and if future results indicate that significant benefits are possible for fusion diagnostics, will initiate additional studies.

REFERENCES FOR SECTION I-VI

1. U. Feldman, J. F. Seely, N. R. Sheeley, Jr., S. Suckewer and A. M. Title, J. Appl. Phys. 56, 2512 (1984).
2. E. F. Borra and J. D. Landstreet, Ap. J. Suppl. 42, 421 (1980).
3. E. F. Borra, Ricerche Astronomiche, 10, 35 (1980).

RECENT PUBLICATIONS

M. Finkenthal, B.C. Stratton, H. W. Moos, W. L. Hodge, S. Suckewer, S. Cohen, P. Mandelbaum and M. Klapisch), Spectra of Highly-Ionized Zirconium and Molybdenum in the 60-150 A Range from PLT Tokamak Plasmas, Journal of Physics B 18, 4393 (1985).

M. Finkenthal, S. Lippmann, L. K. Huang, T.L. Yu, B. C. Stratton H. W. Moos, M. Klapisch, A. Krasnitz, P. Mandelbaum, W. Hodge, P. E. Phillips, T. R. Price, J. C. Porter, B. Richards and W. L. Rowan, The Spectrum of Highly Ionized Praseodymium and Dysprosium from the TEXT Tokamak Plasma in the 50-250 A Range, Journal of Applied Physics 59, 3644 (1986).

T.L. Yu, M. Finkenthal, and H. W. Moos, Experimental Evidence for the Effect of Resonance Structures on Electron Excitation Rates for Si III, Ap.J. 305, 880 (1986).

D. Content, D. Wroblewski, M. Perry and H. W. Moos, Space and Time Resolving Spectrograph for Fusion Plasma Diagnostics, Bull. A.P.S. 30, 1418 (1985).

D. Wroblewski, H. W. Moos, W. L. Rowan, Demonstration of Zeeman Effect Polarimetry as a Magnetic Field Measurement Method in a Tokamak, Bull. of A.P.S. 30, 1569 (1985).

R. D. Benjamin, J. L. Terry, H. W. Moos, Observations of Doppler Broadening of Far UV Lines Emitted from Alcator C Discharges, Bull. A.P.S. 30, 1496 (1985).

T. L. Yu, S. L. Allen, H. W. Moos, T. J. Nash, Impurity Transport in TMX-U, Bull. A.P.S. 30, 1432 (1985).

W. L. Hodge, W. L. Rowan, M. Finkenthal, S. Lippmann, L. K. Huang, T. L. Yu, B. C. Stratton, H. W. Moos, The Measurement of Neon-like Isoelectronic Transitions in a Tokamak Plasma Relevant to Soft X-ray Lasing Schemes, Bull. A.P.S. 30, 1568 (1985).

D. Wroblewski, H. W. Moos and W. L. Rowan, Zeeman Effect Polarimetry of Ti XVII 3834 A Line in Texas Experimental Tokamak, Applied Physics Letters 48, 21 (1986).

M. Finkenthal, T. L. Yu, S. Lippmann, L. K. Huang, H. W. Moos, B. C. Stratton, A. K. Bhatia, R. D. Bengtson, W. L. Hodge, P. E. Phillips, J. L. Porter, T. R. Price, T. L. Rhodes, B. Richards, C. P. Ritz, and W. L. Rowan, A Comparison of the C III, O V, F VI and Ne VII $\Delta n = 0(2-2)$ Line Emissions Emitted by a Laboratory Plasma with Theoretical Predictions and Astrophysical Observations, Ap.J. in press.

D. Content, D. Wroblewski, M. Perry and W. Moos, Space and Time Resolving Spectrograph for High Temperature Plasma Diagnostics, Review of Scientific Instruments 57, 2041 (1986).

D. Wroblewski and W. Moos, Magnetic Field Measurement Based on Zeeman Effect in Lines excited by Charge Exchange Recombination, Review of Scientific Instruments 57, 2029 (1986).

S. Lippmann, M. Finkenthal, L. K. Huang, W. Moos, B. C. Stratton, T. L. Yu, A. K. Bhatia and W. L. Hodge) Laboratory Study of the Spectrum of Highly Ionized Calcium in the 100-250 A Range Applied to Solar Flare Diagnostics, Ap.J. to be published.

L. K. Huang, S. Lippmann, T. L. Yu, B. C. Stratton, H. W. Moos, Experimental Determination of Line Intensity Ratios of Transitions within the $n=2$ Complex of Be I-like Ions (Cl XIV to Cr XXI), submitted to Physical Review.

R. D. Benjamin, J. L. Terry, and H. W. Moos, A High Spectral Resolution, Photon Counting Detector for Doppler Temperature Measurements in Magnetically Confined Plasmas, submitted to Review of Scientific Instruments.

S. L. Allen, T. L. Yu, J. H. Foote, and W. L. Pickles, The Influence of Impurity and Particle Control on Tandem Mirror Experiment Upgrade Plasma Operation, J. Vac. Sci. Technol. A 4, 1075 (1986).

Budget removed
do

END

DATE
FILMED

6/09/92