FINAL REPORT FOR US DEPARTMENT OF ENERGY

PROJECT TITLE:

**SOFT X-RAY MICRORADIOGRAPHY AND LITHOGRAPH
USING A LASER PRODUCED PLASMA SOURCE**

SPONSOR ID NO: DEFG0389SF18012

P. C. Cheng

Advanced Microscopy and Imaging Laboratory (AMIL)
Department of Electrical and Computer Engineering
Center for Electronic and Electrooptics Materials (CEEM)
State University of New York, Buffalo, NY 14260, USA

DISCLAIMER

This report was prepared as an account of work sponsored by an agency of the United States Government. Neither the United States Government nor any agency thereof, nor any of their employees, makes any warranty, express or implied, or assumes any legal liability or responsibility for the accuracy, completeness, or usefulness of any information, apparatus, product, or process disclosed, or represents that its use would not infringe privately owned rights. Reference herein to any specific commercial product, process, or service by trade name, trademark, manufacturer, or otherwise does not necessarily constitute or imply its endorsement, recommendation, or favoring by the United States Government or any agency thereof. The views and opinions of authors expressed herein do not necessarily state or reflect those of the United States Government or any agency thereof.

1. BACKGROUND

Study of 3D microstructures in opaque specimens has been a problem in biology and in material sciences. Because of the high penetration ability, X-ray microscopy and microtomography become a promising tool [Cheng et al., 1989, Cheng et al., 1990, Schmahl and Cheng, 1991, Cheng and Jan, 1987, Cheng et al., 1992, Graeff and Engelke, 1991]. Among different kinds of X-ray microscopies, shadow projection microscopy is the simplest and the most convenient technique. Furthermore, it has been demonstrated that a resolution in the $0.1 \mu m$ domain can be achieved by using a point X-ray source generated by a microfocused e-beam. Recently there are increasing interests in developing a microtomography capability on conventional shadow projection microscopes.

As X-rays from the point source, produced by an laser-induced plasma, irradiating to a sphere-like specimen form a cone, this approach is commonly called cone-beam tomography.

Existing cone-beam tomography algorithms can be categorized into two broad classes: analytic and algebraic [Kak and Slaney, 1987, Smith, 1990]. The algebraic algorithms demand more computing resources than the analytic ones. Among the analytic algorithms, Feldkamp's convolution and back-projection algorithm is the most efficient and produces satisfactory images [Feldkamp et al., 1984]. Feldkamp's cone-beam reconstruction algorithm is subject to two constraints. Firstly, the specimen must be contained in a sphere-like reconstruction region. Secondly, the X-ray source must be moved along a circle in the specimen coordinate system. In addition, the conventional algorithm suffers from larger errors in reconstructing off-midplane structures. These constraints are serious in cone-beam X-ray microtomography.

2. RESULTS

2.1 General Cone-beam Reconstruction Algorithm

Considering the hardware characteristics of the laser-induced plasma X-ray source and the limitations of the conventional cone-beam reconstruction algorithm, a general cone-beam reconstruction algorithm has been developed at our laboratory, in which the motion locus of the X-ray source is an arbitrary curve corresponding to at least a 2π continuous horizontal angular displacement in the coordinate system of the specimen [Wang et al., 1991a, Lin et al., 1991, Wang et al., 1991b, Wang et al., 1991c, Wang et al., , Wang et al., 1992a]. The preliminary simulation shows that the general cone-beam reconstruction algorithm consistently results in visually satisfactory images. the fifth section.

In general cone-beam reconstruction, there are mainly two kinds of scanning loci, planar and helix-like. A planar scanning locus (a circle) is used in the conventional cone-beam reconstruction. A helix-like scanning locus is preferred to deal with rod-like specimens. Without losing generality, a locus turn of the X-ray source can be defined in cylindrical coordinates by the following equation:

$$\rho = \rho(\beta), \quad h = h(\beta), \quad \beta \in [0, 2\pi] \quad (1)$$

In the first scanning mode, $h(\beta) = 0, \beta \in [0, 2\pi]$. In the second mode, $h(\beta) \in [z - \frac{h_p}{2}, z + \frac{h_p}{2}], \beta \in [0, 2\pi]$, where a horizontal slice with a vertical axis coordinate z of a specimen is interested, and h_p is the pitch of the helix-like scanning locus. We define the mid-plane is the $x - y$ plane in the first

scanning mode, the plane with a vertical axis coordinate z in the second mode.

The general cone-beam formula can be derived by modifying the convolution and back-projection formula for equispacial fan-beam. In a three-dimensional coordinate system (x, y, z) the cone-beam reconstruction is performed by tilting the fan out of the mid-plane, thus changing the coordinate system (t, s) , where (t, s) denotes the rotated coordinate system described by:

$$t = x \cos \beta + y \sin \beta, \quad s = -x \sin \beta + y \cos \beta \quad (2)$$

β is the rotation angle of the specimen. A new coordinate system $(\tilde{t}, \tilde{s})$ is defined to represent the location of the reconstructed point with respect to the tilted fan. Because of the geometry of the tilted fan, both the source-to-origin distance and the angular differential must be modified. After the modification, we obtain

$$g(x, y, z) = \frac{1}{2} \int_{-\pi}^{\pi} \frac{\rho(\beta)^2}{(\rho(\beta) - s)^2} \int_{-\infty}^{\infty} R_{\beta}(p, \zeta) f\left(\frac{\rho(\beta)t}{\rho(\beta) - s} - p\right) \frac{\rho(\beta)}{\sqrt{\rho(\beta)^2 + p^2 + \zeta^2}} dp d\beta \quad (3)$$

where $f(\cdot)$ is the reconstruction filter, and

$$\zeta = \frac{\rho(\beta)z'(\beta)}{\rho(\beta) - s} \quad (4)$$

the local coordinate system $(x, y, z'(\beta))$ associated with the angle β is defined by:

$$z'(\beta) = z - h(\beta) \quad (5)$$

In the first scanning mode, $z'(\beta) = z, \beta \in [0, 2\pi]$. In the second mode, $z'(\beta) \in [z - \frac{h_p}{2}, z + \frac{h_p}{2}], \beta \in [0, 2\pi]$. The specimen is assumed to be centered at the origin of the coordinate system. In both cases, the projection data collected from the turn described when β varies from 0 to 2π are used to reconstruct the image in the mid-plane. The equation (3) is referred to as the general cone-beam reconstruction formula.

As mentioned earlier, in general cone-beam reconstruction, there are two kinds of scanning loci, planar and helix-like. The polygonal cone-beam algorithm and the dashed-line cone-beam algorithm are described as follows, which require typical scanning loci for each kind of scanning loci respectively.

For tomographic imaging, a large number of two-dimensional images must be recorded at various orientations. The specimen therefore must be rotated, and a mechanical stage is installed for this purpose. Because the variation of rotation axis position of the mechanical stage is one of main error sources in microtomography, an elaborate feed-back system is being implemented to correct this. However, experiments have shown that the calibration process of the rotation axis position is both time-consuming and complicated. On the other hand, the X-ray source of our X-ray microscope can be precisely moved along a straight line by steering the e-beam electromagnetically. Therefore, to minimize the number of the recalibrations of the mechanical axis position, a polygonal scanning mode can be used. The straight line portions of the polygon are achieved by the X-ray source movement and the number of the polygon sides determines the number of the stage rotation. Consequently, the number of the stage rotation is greatly reduced. Like the conventional cone-beam algorithm, the reconstruction region of the polygonal cone-beam algorithm is sphere-like.

In the case of microtomography, most specimens are not spherical but rather cylindrical. With a helix-like scanning locus, the general cone-beam formula can be used to handle rod-shaped specimens. Typically, we can use the following definitions of ρ and h to describe a turn of scanning locus of the X-ray source.

$$\rho = \frac{\rho_0}{\cos 2\pi(t - \frac{\lfloor N_S t + 0.5 \rfloor}{N_S})}, \quad t \in [0, 1] \quad (6)$$

$$h = z + h_p \frac{\lfloor \frac{N_S(\beta - \pi)}{2\pi} \rfloor}{N_S}, \quad \beta \in [0, 2\pi] \quad (7)$$

In expressions (6) and (7), ρ_0 is a constant, N_S is the number of the scanning polygon sides, h_p is the pitch of the scanning locus, and z is the vertical coordinate of a horizontal slice to be reconstructed. Thus, a three-dimensional dashed-line cone-beam algorithm is obtained. The scanning locus required by the dashed-line cone-beam algorithm is a sequence of horizontal straight lines which approximately segments a helix, similar to “winding stairs”. To achieve this scanning mode, two kinds of movements must be combined. First, the X-ray source is steered electronically horizontally. Second, the mechanical stage is rotated and shifted mechanically with specified steps. It is evident that the mechanical errors are much greater than the electronic one. The dashed-line algorithm is valuable for reducing the number of both the angular rotation and the vertical shifting of the mechanical stage. This means a much less number of the recalibrations of the stage position, that is, a higher resolution. Note that while reconstructing each horizontal slice of the specimen, it is required that the horizontal scanning line segment closest to it correspond to $\beta = \pi$ and projection data collected from the upper and lower halves of the scanning locus turn are used to reconstruct the slice, so that the reconstruction error could be minimized.

Figure 2 shows two reconstructed horizontal slices of a specimen at different z coordinates using

the dashed-line reconstruction algorithm when ρ_0 , N_s and h_p were set to 5, 8 and 1 respectively. The detector plane is 5 by 5 having 128 by 128 pixels and was assumed to be placed on the z axis for the ease of computation. One hundred two-dimensional projections were used for the reconstruction.

2.2 Estimation of the Reconstruction Accuracy

Briefly speaking, the reconstruction error comes from two aspects, the limitation of theoretical precision and the influence of practical factors including system noise, sampling and quantization effects and inaccuracy of the imaging parameter measurement. It would be extremely useful in practice, if we could evaluate the reconstruction error quantitatively.

Some preliminary error analysis is being done [Wang et al., 1992b]. It shows that the reconstruction error depends on to a great extent the properties of the specimen image function, specifically, the variation of the specimen image function along z axis. If specimens are modeled as stochastic fields, the reconstruction error can be studied conveniently.

In stochastic representations, an image can be considered as a sample function of an array of random variables which is called a stochastic field. This characterization of a set of images is useful in studying image processing algorithms that are valid for the whole set statistically, instead of just for an individual image.

In many applications, it is often sufficient to characterize a set of images by its mean and covariance functions[Jain, 1989]. We will also use this covariance model in our reconstruction error analysis. Often one starts with a stationary random field representation where the mean is

Figure 1: The scanning locus in the dashed-line approach and a mid-plane.

Figure 2: Two reconstructed images using the three-dimensional dashed-line cone-beam reconstruction algorithm when N_s and h_p are set to 8 and 1.0 respectively.

held constant and the covariance function is represented by either a separable or a nonseparable exponential-like function. The separable covariance model is very convenient for both simulation and analysis, however, the nonseparable covariance function is a better model. In our case, the nonseparable covariance model is certainly preferred.

S. E. Reichenbach *et al.* [Reichenbach et al., 1991] describes a software simulation environment for image processing research. The simulation is based on a comprehensive model that accounts for statistical characteristics of the scene, image formation, sampling, noise, and display reconstruction. The simulation uses a stochastic process to generate digital scenes with variable spatial detail and structure. Based on their work, stochastic images could be synthesized for our error analysis. Because the point spread function of our general cone-beam reconstruction algorithm is very complicated and spatially varying, there is a lot of work to be done in this aspect.

2.3 Cone-beam Reconstruction of Plate-like Specimen

The reconstruction of a plate-like sample appears to be a limited angle problem which can be addressed using standard methods described previously in the literature. Since the methods based on a point source and shadow projection microscopy have great advantages experimentally, We propose a novel scheme using the general cone-beam algorithm to reconstruct plate-like specimens. The method can be explained as follows. Assume that the locus of the X-ray source is a two-dimensional curve (for example it may be a circle, polygon, or a rectangle). The plate-like (or thick film) sample is placed parallel to the plane of the source locus. Arrange the relative position so that the perpendicular passing through the center of the scanning locus passes through the center

of the specimen area of interest. This perpendicular is labeled the principal axis. The specific locus used is made to be similar to the shape of the area of interest, and substantially larger than that area. The detector plane is placed behind the specimen, parallel to the specimen, and the two-dimensional projection data recorded for each X-ray source position. After each frame of these projection data is mapped onto the plane facing the X-ray source and passing through the principal axis, the general cone-beam algorithm is applied to the modified projection data. In this way the problems associated with the standard limited angle approaches are avoided [Wang et al., 1991c]. Preliminary simulation results supported this scheme. However, further studies are needed to optimize the imaging parameters and reveal the reconstruction accuracy characteristics.

References

- [Cheng and Jan, 1987] Cheng, P. C. and Jan, G. J. (1987). *X-ray Microscopy – Instrumentation and Biological Applications*. Springer-Verlag.
- [Cheng et al., 1992] Cheng, P. C., Lin, T. H., and Wu, W. L. (1992). *3D Microscopies*. Springer-Verlag. to be published.
- [Cheng et al., 1989] Cheng, P. C., Newberry, S. P., Kim, H. G., and Wittman, M. D. (1989). X-ray contact microradiography and shadow projection X-ray microscopy. *Europ. J. Cell Biology*, 48(25):169–172.
- [Cheng et al., 1990] Cheng, P. C., Newberry, S. P., Kim, H. G., Wittman, M. D., and Hwang, I.-S. (1990). X-ray contact microradiography and shadow projection microscopy. In Duke, P. and Michette, A., editors, *Modern microscopy*, pages 87–117. Plenum Press.
- [Feldkamp et al., 1984] Feldkamp, L. A., Davis, L. C., and Kress, J. W. (1984). Practical cone-beam algorithm. *J. Opt. Soc. Am.*, 1 (A):612–619.
- [Graeff and Engelke, 1991] Graeff, W. and Engelke, K. (1991). Microradiography and microtomography. In E. E. Koch, T. S. and Winick, H., editors, *Handbook on Synchrotron Radiation*, volume 4, pages 361–405. North-Holland.
- [Jain, 1989] Jain, A. K. (1989). *Fundamentals of digital image processing*. Prentice-Hall, Inc., USA.

- [Kak and Slaney, 1987] Kak, A. C. and Slaney, M. (1987). *Principles of Computerized Tomographic Imaging*. IEEE Press, USA.
- [Lin et al., 1991] Lin, T. H., Wang, G., Cheng, P. C., and Shinozaki, D. M. (1991). A helical cone-beam reconstruction algorithm for X-ray microtomography. *J. Scanning Microscopy*, 13 (I):11-13.
- [Reichenbach et al., 1991] Reichenbach, S. E., Park, S. K., Alter-Gartenberg, R., and Rahman, Z.-u. (1991). Artificial scenes and simulated imaging. In *Proc of SPIE*, vol. 1569, pages 422-433.
- [Schmahl and Cheng, 1991] Schmahl, G. and Cheng, P. C. (1991). X-ray microscopy. In E. E. Koch, T. S. and Winick, H., editors, *Handbook on Synchrotron Radiation*, volume 4, pages 481-536. North-Holland.
- [Smith, 1990] Smith, B. D. (1990). Cone-beam tomography: recent advances and a tutorial review. *Opt. Engineering*, 29(5):524-534.
- [Wang et al., 1992a] Wang, G., Lin, T. H., and Cheng, P. C. (1992a). A new extended fan-beam reconstruction formula. In *Proc. of SPIE*, vol. 1660.
- [Wang et al.,] Wang, G., Lin, T. H., Cheng, P. C., and Shinozaki, D. M. A general cone-beam reconstruction algorithm. *Submitted to IEEE Trans. on Med. Imag.* In review.
- [Wang et al., 1991a] Wang, G., Lin, T. H., Cheng, P. C., and Shinozaki, D. M. (1991a). A polygonal cone-beam reconstruction algorithm for X-ray microtomography. *J. Scanning Microscopy*, 13 (I):126-128.
- [Wang et al., 1991b] Wang, G., Lin, T. H., Cheng, P. C., and Shinozaki, D. M. (1991b). Scanning cone-beam reconstruction algorithms for X-ray microtomography. In *Proc. of SPIE*, vol. 1556.
- [Wang et al., 1991c] Wang, G., Lin, T. H., Cheng, P. C., and Shinozaki, D. M. (1991c). Simulation of cone-beam image reconstruction. In *Proc. of SimTec '91*.
- [Wang et al., 1992b] Wang, G., Lin, T. H., Cheng, P. C., and Shinozaki, D. M. (1992b). Preliminary error analysis on the general cone-beam reconstruction algorithm. In *Proc. of SPIE*, vol. 1660.

preprints & reprint
removed
as

END

**DATE
FILMED**

4 / 13 / 92