Fermilab

TM-1209
1183.000

APPLICATIONS OF BONNER SPHERE DETECTORS
IN NEUTRON FIELD DOSIMETRY

Miguel Awschalom
Fermilab Neutron Therapy Facility
P. O. Box 500, Batavia, Illinois

and

Robert S. Sanna
U. S. Department of Energy
Environmental Measurements Laboratory
376 Hudson Street, New York, New York 10014

September 1983

I INTRODUCTION

Since the second world war, a number of neutron flux detectors have been developed to achieve neutron detection efficiencies that were approximately directionally isotropic and either that were independent of energy (flux detectors),¹⁻³ had energy dependence proportional to the reciprocal of the flux to dose equivalent (DE) conversion factors (rem meters)⁴⁻⁶ or had energy dependence that maximized at some neutron energy (spectrometers).^{4,7-9} One of the most famous of all such detectors is the Bonner sphere spectrometer (BSS).⁴ In its original form it consisted of a 4 mm x 4 mm high $^6\text{LiI(Eu)}$ crystal which could be positioned at the center of five polyethylene spherical moderators 2 to 12 inches in diameter.

A large bibliography exists on the BSS attesting to the usefulness of this technique. The authors are citing only a part of these references considered representative of the literature.

II THEORY.

As they traverse matter, epithermal and fast neutrons undergo elastic and/or inelastic scattering losing energy until they reach thermal equilibrium or leave the moderator. This process is known as neutron slowing down or moderation. Moderators should

have high scattering to absorption cross section ratios so that the largest number of incident neutrons will be thermalized before being absorbed. Good moderating materials are composed primarily of hydrogen, deuterium, beryllium, carbon and oxygen.

The energy dependence of a neutron detector is the result of convoluting the moderated neutron spectrum at the center of the sphere with the energy dependent detection efficiency of the neutron detector. The moderator shapes may be spherical,⁴ pseudospherical,¹⁰ cylindrical,^{1,3,5,10-12} or cubical.^{13,14} A number of such detection efficiencies have been calculated and/or measured (Tables 2 through 8, and Ref 15). Attempts have been made to enhance the low energy resolution by wrapping some of the moderators in cadmium^{16,17} and/or boron.¹⁸ Another response modification for better matching the reciprocal of the flux-to-DE energy curve was made by inserting a Cd-shell at an intermediate radius.⁶

If a neutron detector is placed at the center of a number 'r' of spherical moderators of various diameters each detector-moderator combination will have a different response to neutrons as a function of energy. The response of the BSS, as a function of energy, for one such set of moderator spheres, with a 4mm x 4mm ⁶LiI(Eu) scintillator, is shown in Figure 1.⁸

The response of each detector of an array of 'r' elements to a neutron spectrum, $N(E)$, may be written as a homogeneous Fredholm equation,

$$B_j = \int_0^{\infty} N(E) R_j(E) dE \quad (j=1, r) \quad (1)$$

where,

B_j is the response for the j-th detector in the array,

$N(E)$ is the energy spectrum of the neutron field, and

$R_j(E)$ is the absolute response for j-th detector as a function of neutron energy.

When the detector responses are known for 'n' discrete energy groups, equation (1) may be rewritten:

$$B_j = \sum_{g=1}^n N_g R_{j,g} \Delta E_g \quad (j=1, r) \quad (2)$$

where,

N_g is the differential neutron energy flux for the g-th energy group,

$R_{j,g}$ is the response of j-th detector to neutrons of group 'g', and ΔE_g is the energy width of the group.

The response distributions were calculated for a 4mm x 4mm ^6Li detector (instead of $^6\text{LiI(Eu)}$) bare and at the center of polyethylene spheres 2, 3, 5, 8, 10, and 12 inch diameter, exposed to several neutron spectra using 31 group responses⁸ (Fig. 1), for

three realistic and three "monochromatic" spectra. The realistic spectra were: the leakage spectra of the Oak Ridge unreflected Health Physics Research Reactor (HPRR),¹⁹ the atmospheric cosmic-ray neutron spectrum,²⁰ and the neutron spectrum of a bare ^{252}Cf fission source;²¹ while the "monochromatic" spectra were energy groups 31, 20, and 7. Figure 2 shows the three realistic spectra while Figures 3 and 4 display the response distributions.

III MODERATORS

Spherical polyethylene moderators are the nearly universal choice of those using the BSS method of neutron detection.^{16,22-25} The spherical symmetry, with the detector at the center of the sphere often results in an isotropic response.

Density of commercial polyethylene ranges between 0.91 and 0.96 g cm⁻³.²⁶ In fact, one user²⁷ found that the densities of a set of seven spheres were all different and varied from 0.912 to 0.961 g cm⁻³. Hence, density and proper response must be matched. The response ratio for 10 inch diameter polyethylene spheres having densities 0.90 and 0.95 g cm⁻³ is shown by curve 1,⁸ Fig. 5. Such changes in polyethylene density may have serious effects on interpretation of measurements. Most commonly, the polyethylene used is the so called "linear" high density polyethylene ($\rho = 0.96 \text{ g cm}^{-3}$).

Another source of errors is the use of polyethylene responses for a different moderating material. Curve II of Figure 5 shows the response ratio for a 0.5 inch x 0.5 inch scintillator at the center of 10 inch spheres of polyethylene (0.95 g cm^{-3}) (Table 6) and water (Table 7). Here, the effect on the results is even greater than for a mismatch in moderator densities for the same material.

There is no best BSS. In practice, there are compromises. In some situations, the classical BSS⁴ may be augmented with additional spheres²⁸ and even with activation detectors²⁹ in an attempt to increase the energy resolution of the deconvoluted neutron energy spectrum. When an acceptable solution to the set of Fredholm equations is found, integral parameters such as fluence, absorbed dose and DE may be obtained with the largest practical precision. Attempting a large area survey with such an extensive multidetector system may be a task of mammoth proportions. Therefore, for expediency and portability the system has often been reduced to a few, one to five, detectors.³⁰⁻³³ As the number of elements in a BSS decreases, the accuracy of the determination of the integral parameters may decrease. However, in many circumstances the DE determinations using a reduced set, even a single detector, may still be satisfactory for health physics purposes.⁴⁻⁶

IV INSTRUMENTATION AND MEASUREMENTS

The choice of the neutron detector is a compromise. The choice may be due to the user's experience, availability of suitable equipment, gamma ray discrimination, portability, absolute detection efficiency, information desired in real time or integrated over a period, DE rate and/or duty cycle of the neutron field.

Neutron detectors may be divided in two groups, active and passive. In either case, the gamma sensitivity of the neutron detector must be investigated and, in general, compensated by suitable means.

ACTIVE DETECTORS.

Active detectors may be either scintillators plus photomultipliers or gas filled counters operated in the proportional or Geiger-Muller mode. Their main advantage is real time data acquisition and, at times, good gamma ray discrimination. They may be used to monitor fields and set-off alarms. Data may be collected and displayed automatically.

Scintillators.

The most common scintillator used in BSS, is the ${}^6\text{LiI}(\text{Eu})$.⁴ In this detector the neutron-photon discrimination is achieved through the exoergic ${}^6\text{LiI}(\text{n},\alpha){}^3\text{H}$, $Q=4.8$ MeV, reaction which causes a large light output due to the neutron capture. The dimensions of the crystal itself are chosen to limit the light output from electrons or external charged particles crossing it.⁴ The consequence of using a suitably small crystal (4 mm x 4 mm) is that the absolute neutron detection efficiency is typically 0.1 percent.⁴

There are occasions when satisfactory discrimination against background charged particles cannot be achieved even with small crystals as in the evaluation of the weak neutron fields from cosmic rays. Various methods have been tried to overcome this problem. One uses large (0.50 inch x 0.50 inch) ${}^6\text{LiI}(\text{Eu})$ and ${}^7\text{LiI}(\text{Eu})$ crystals and the net neutron contribution is the difference of the two signals.³⁴ A second method is to use a single ${}^6\text{LiI}(\text{Eu})$ crystal and collect the data on a multichannel analyser. Values for the background may be interpolated and subtracted from the (n,α) peak.^{35,36} These two methods use small differences of two large signals. This causes large uncertainties in the net response. A third approach uses a phoswich^{37,38} to reject external particles. The phoswich consists of a small ${}^6\text{LiI}(\text{Eu})$ crystal inside a "cup" made of plastic scintillator

(Figure 6). Since the decay time constants of the two scintillators are about three orders of magnitude apart, a simple pulse shape discriminator permits excellent rejection of external particles.³⁸ Figures 7 and 8 show environmental data collected for 48.3 hours using a 4 mm x 4 mm crystal and for 70.8 hours using an 8 mm x 8 mm crystal, in both cases the moderator was a ten inch pseudosphere.³⁸ It may be seen that the phoswich dramatically reduces the background noise allowing the use of an 8mm x 8mm crystal which increased the sensitivity of the BSS by approximately a factor of four.

Just as with the moderator, care must be taken to use the appropriate response for the detector chosen. Curve III, of Figure 5, shows the ratio of the response of a 4 x 4 mm to that of a 12.7 mm x 12.7 mm ^6Li detector at the center of a 2 inch polyethylene sphere.⁸

LiI is an extremely hygroscopic substance and even well sealed crystals tend to degrade in time. Degradation is evidenced by a change in crystal color and energy resolution of the (n,α) peak. To avoid these problems ^6Li loaded glass could be used. However, no reference to their use in BSS has been found by the authors. UV lucite is recommended for light pipes between crystal and photomultipliers because it is much clearer and absorbs much less of the visible part of the spectrum than ordinary lucite. The light pipe should make a snug fit into the moderator to

minimize neutron leakage from the environment to the detector.³⁹ This last recommendation also applies to the installation of any neutron detector holder in a sphere. Use of small diameter photomultipliers is also recommended to minimize interference with the polar angle neutron detection efficiency. The ambient environment in which measurements are made may dictate taking special precautions to protect the electronics. For example, in making measurements inside the containment of a PWR reactor, it was necessary to cool and dessicate the electronics.³⁵

Gas Filled Detectors.

Proportional counters. Detectors filled with suitable gases and operated in the proportional mode may provide discrimination against photons. Two suitable gases are ^3He and $^{10}\text{BF}_3$. ^3He has been tried satisfactorily in a 3 cm diameter, spherical proportional counter at the center of a 20.8 cm sphere.⁴⁰ However, the $^3\text{He}(n,p)^3\text{H}$ has a Q of only 765 keV. Therefore, it does not furnish a strong neutron signal to simplify photon rejection. Elsewhere, a small cylindrical $^{10}\text{BF}_3$ counter was tried with some success in spheres 9 and 10 inches in diameter.⁴¹ $^{10}\text{BF}_3$ counters have good photon discrimination via the $^{10}\text{B}(n,\alpha)^7\text{Li}$, $Q = 2.8$ MeV reaction, even though most of the transitions occur to the 0.48 MeV excited state of ^7Li decreasing somewhat the photon rejection. The signals from these gas filled counters are generally processed

via charge integrating preamplifiers. However, fast current type instrumentation may be used such as the very fast signal processing 50 ohm systems. This yields not only better discrimination against background by reducing pulse pile-up, but also allows higher neutron counting rates and operations in smaller duty cycle neutron fields.

Due to the mechanisms for energy loss, incoming monochromatic neutrons collide different numbers of times with the nuclei of the moderator while slowing down, causing a spread in path lengths and slow down times. The neutron fluxes encountered in field applications have continuous energy spectra. Thus, the spread out in time for the arrival of slow neutrons at the detector location in a BSS, even from short neutron pulses should be of the order of a hundred or more microseconds for the larger spheres (six inches or more in diameter).⁴² Thus, BSS with electronic counters may work well, using appropriate correction factors, in small duty cycle fields having fairly high dose equivalent rates.

Geiger-Muller Counters. These detectors have been used successfully in very short duty cycle neutron fields. A thin walled G-M tube wrapped in silver foil allows for the activation of the silver in a short time and then counting their activities during a relatively long time, $^{109}\text{Ag}(n,\gamma) \ ^{110}\text{Ag}(t_{1/2}=24.4 \text{ sec})$ and $^{107}\text{Ag}(n,\gamma) \ ^{108}\text{Ag}(t_{1/2}=2.4 \text{ min})$.^{10,43-45} In cases where the duty cycle of the accompanying photon radiation is large, a dual system

of identical G-M's tubes may be used, one wrapped in silver and the other in tin.^{10,45} The output of the tin wrapped G-M is subtracted from the silver wrapped one and the difference is the neutron component. The use of a low power microcomputer makes this differential system easy to calibrate and operate.⁴⁵

PASSIVE DETECTORS.

Neutron interactions in the detectors may be recorded by the radioactivation or radioluminescence they produce. Passive detectors have some advantages over counters such as portability, lower unit cost and independence from neutron field duty cycle.

Radioactivation. Nuclear reactions in a number of elements may be used to record the thermal neutron flux. The actual choice of foil will depend on the length of time the neutron field is to be monitored and its intensity. In the past, such elements as indium,⁴⁶ gold (Table 8),⁴⁴ tantalum,⁴⁷ and cobalt² have been used successfully in moderated neutron flux integrators.

Thermoluminescent Dosimeters. These TLD dosimeters have been used as neutron detectors, usually ^6LiF and ^7LiF pairs, both in powder form⁴⁸ and extruded chips.^{29,36,49} The photon response of these dosimeters is essentially the same, but the (n,α) reaction in ^6Li is orders of magnitude larger than in ^7Li . Hence, the difference between the responses of the ^6LiF and the ^7LiF is a

measure of the neutron signals. Great care must be taken transporting the TLD detectors to protect them from unwanted exposure to neutrons.³⁶ Transport is that period between annealing and start of measurement plus that from the end of the measurement to readout. A successful manner to handle TLD pairs is to load them in typically polyethylene⁵⁰ inserts that go into the BSS moderators and put the inserts in Cd-cylinders. Then, upon reaching the measurement site the inserts are rapidly removed from the Cd-cylinders and inserted in the moderators. Upon completion of the measurement, the procedure is reversed.

V. SPECTRAL DECONVOLUTION

Spectral deconvolution is, simply, the solution of Eq. 1 for $N(E)$. A very good review paper on the subject was presented at the 2nd ASTM-EURATOM Symposium on Reactor Dosimetry in 1977.⁵¹ Deconvolution methods were grouped into four categories: derivative, parametric, quadrature, and Monte Carlo.

The primary application of the derivative method is with recoil particle detectors, and not for BSS.

The parametric procedure may be used if a functional representation of the spectrum exists, e.g., evaporation or fission spectra. The BSS responses are used to determine the parameters for the appropriate functional representation.^{52,53}

The quadrature method involves the solution of Eq. 2 by linear estimation, least squares, iterative or mathematical programming techniques, or a combination thereof. The three most common deconvolution codes (BON,⁵⁴⁻⁵⁶ LOUHI,^{57,58} and SAND⁵⁹) used with the BSS are found in this category. BON,^{29,36,60-62} LOUHI^{16,63-65} and SAND^{66,67} have been used to deconvolute spectra ranging from PWR containment to accelerator shield leakage spectra from BSS measurements.

The Monte Carlo category uses Monte Carlo techniques to randomly chose a neutron spectrum. This spectrum is then convoluted with the BSS response functions to yield a response distribution. The process is repeated a great number of times and the response distributions, thus obtained, are compared with the measured one. The spectrum yielding the best agreement is said to be an adequate solution of Eq. 2 for N_g . The only known code using this technique is SWIFT.^{68,69} SWIFT saves the four spectra whose response distributions best agree with the measured response distribution and averages them to yield the N_g .

A word of warning: since the systems dealt with here are usually underdetermined, a solution to Fredholm's equation is not unique. However, this has not been a serious problem in practice.

VI. CALIBRATION

In the field, the absolute neutron detection efficiencies of BSS elements are most often checked using $^{241}\text{AmBe}$, ^{241}AmB , and ^{252}Cf sources.^{42,70-73} In so doing, the effect of walls, floor, and ceiling scattering^{74,75} as well as sphere to sphere scattering must be taken into account.

To check the energy response of BSS elements, one may want to use nearly monochromatic beams^{73,76,77} if an accelerator is available, or a number of continuous energy spectra (α, n) and/or (γ, n) sources,^{42,73,78} otherwise. The use of radioactive sources for detector calibration implies that the user has (1) a calibration for the total neutron emission into 4π ; (2) knowledge of the source output as a function of polar and azimuthal angles, particularly in the case of the more massive sources, and (3) knowledge of the neutron energy spectrum for the source. Item (2) is very important if one uses a long half-life alpha emitter such as ^{239}Pu ($t_{1/2} = 24,000$ years) since the outputs at 0° and 90° (polar angles) may be significantly different. For this reason, shorter lived alpha sources such as ^{241}Am ($t_{1/2} = 455$ years) and ^{238}Pu ($t_{1/2} = 86.4$ years) have become quite common. Mixtures or combinations of either ^{238}Pu or ^{241}Am with Be, B, F and Li are readily available commercially.⁷⁹ Prices ranges (August, 1983) for 5 Ci, sources are US\$3000 ($^{238}\text{PuBe}$) to US\$4800 ($^{238}\text{PuLi}$) and US\$3600 ($^{241}\text{AmBe}$) to US\$6500 ($^{241}\text{AmLi}$).⁷⁹

The evaluation of the BSS output in terms of neutron flux, absorbed dose or dose equivalent per unit of response requires the evaluation of the constants C_j in equation 3, using a neutron source with known strength and energy spectrum.

$$B_j = \sum_{g=1}^n S C_j N'_{s,g} R'_{j,g} \Delta E_g \quad (3)$$

where,

$N'_{s,g}$ is the normalized differential energy flux of source neutrons for the g -th energy group.

S is the source strength,

$R'_{j,g}$ is the average relative energy response of the j -th detector in the g -th energy interval, and other terms have the same definitions as in equations 2 & 3.

The C_j 's are, then, given by

$$C_j = B_j / S \sum_{g=1}^n N'_{s,g} R'_{j,g} \Delta E_g \quad (4)$$

Once the C_j 's are known, the deconvolution techniques described in Section V may be used.

If dosimetry is to be done with a single detector ' j ' and information exists on the energy spectrum of the field neutrons, the procedure that follows is preferable to the use of average properties of neutron sources to calculate the detector response in terms of flux, absorbed dose or DE.

The flux, F , can be obtained from

$$F = B_j / C_j \sum_{g=1}^n N'_{f,g} R'_{j,g} \Delta E_g \quad (5)$$

and the specific dose equivalent, $\overline{DE}$, from

$$\overline{DE} = \sum_{g=1}^n DE_g N'_{f,g} \Delta E_g \quad \sum_{g=1}^n N'_{f,g} \Delta E_g, \quad (6)$$

where,

$N'_{f,g}$ is the normalized field neutron energy spectrum ($F N'_{f,g}$ is the neutron field energy spectrum).

and

DE_g = average flux-to-dose equivalent conversion factor for the g -th energy interval.

Finally, the dose equivalent is obtained from,

$$DE = F \overline{DE}$$

VII. SUMMARY

The major advantage of the BSS is that it covers an energy range from thermal to hundreds of MEV. No other spectrometer covers this entire energy range.

Among its disadvantages are its inherently low energy resolution and its weight (up to 47 kg). Another disadvantage is that if one uses a single detector alternating the moderators, one has to be concerned with time variations of the neutron field. On the other hand, if one uses multiple detectors, one for each moderator, and makes a simultaneous exposure of the entire system, the concern is about spatial variations of the neutron field and interactions between detectors.

Problems that should be studied in the future include:

- (1) benchmarking for testing of spectral deconvolution codes;
- (2) determination of what is an adequate BSS, i.e., how many and what size moderators should be used;
- (3) development of deconvolution codes for use with microcomputers so that real time data reduction becomes a reality, and
- (4) a study of moderator to moderator interactions.

VIII. ACKNOWLEDGMENTS

We want to thank J. Couch and K. O'Brien for reading a draft of this manuscript and making valuable suggestions, and M. Gleason for her cheerful and accurate typing. This investigation was supported by US PHS Grant 5P01 CA18081-09, awarded by the National Cancer Institute, DHHS, and the US Department of Energy.

REFERENCES

1. Hansen, A. O. and McKibben, J. L., A Neutron Detector Having Uniform Sensitivity from 10 keV to 3 MeV, Phys. Rev. 72, 673 (1947).
2. Smith, A. R., A Cobalt Neutron Flux Integrator, Health Phys. 7, 40 (1961).
3. DePangher, J., Nichols, L. L., A Precision Long Counter for Measuring Fast Neutron Flux Density, Battelle North West, Pacific North West Laboratory, Richland, WA, Report BNWL-260 (June, 1966).
4. Bramblett, R. I., Ewing, R. I., Bonner, T. W., A New Type of Neutron Spectrometer, Nucl. Instr. Meth. 9, 1 (1960).
5. Andersen, I. O., Braun, J., A Neutron Rem Counter, Nukleonik 6 (5) 237 (1964)
6. Hankins, D., A Modified-Sphere Neutron Detector, LA-3595 (Jan. 17, 1967).
7. McGuire, S. A., A Dose Monitoring Instrument for Neutrons from Thermal to 100 MeV, LA-3435, March 1, 1966.

8. Sanna, R. S., Thirty One Group Response Matrices for the Multisphere Neutron Spectrometer Over the Energy Range Thermal to 400 MeV, USEAC, HASL-267 (March 1973).
9. Piltingsrud, H. V., Engelke, M. J., A Passive Broad-Energy Response Neutron Spectrometer Dosimeter, Proc. Symp. Neutron Monitoring for Radiation Protection Purposes, Vienna, Dec. 11-15, 1972, IAEA, 1973, pp. 123-138.
10. Awschalom, M., Borak, T., Howe, H., A Study of Spherical, Pseudospherical and Cylindrical Moderators for a Neutron Dose Equivalent Rate Meter, Fermilab, Batavia, IL, Report TM-291 (July 12, 1971).
11. Tatsuta, H., Ryufuku, H., Shirotani, T., A New Rem-Counter for Neutrons, Health Phys. 13, 559 (1967).
12. Nakamura, T., Yamaguchi, C., Yamamoto, T., Neutron Dosimetry in the Environment Around the Accelerator, proceedings of the 1st KEK Symposium on Radiation Dosimetry, Tsukuba, Japan, July 17-18, 1978 KEK-78-18, CONF-7807109, 1978.
13. Sanna, R. S., O'Brien, K., A New Type of Moderate-and-Capture Neutron Spectrometer, to be presented at 6th International Congress of the International Radiation Protection Association, West Berlin, May 7-12, 1984.

14. Lautenbach, G., Leakage Neutron Spectrum Measurements with the PPN-Spectrometer, 7th DOE Workshop on Personnel Neutron Dosimetry, London, England, 1978 PNL2807, UC-48, 1978.
15. Zaborowski, H. L., Le Systeme Multiphere de Dosimetrie et Spectrometrie Neutron (S.M.-D.S.N.), Etudes d'Applications a la Radioprotection, presented at the Symposium International Sur l'Utilization du Californium-252, Paris 26-28 Avril, 1976.
16. Endres, G. W. R., Brackenbush, L. W., Neutron Dosimetry and Spectral Measurements in PWR Containment, 4th Symposium on Neutron Dosimetry, Munich-Nureherberg, F. R. Germany, June 1, 1981, PNL-SA-9461 (1981).
17. Griffith, R. V., Hankins, D. E., A Systematic Approach to Personnel Neutron Monitoring, 5th International Congress of the International Radiation Protection Association, 327, 1980 CONF-800304-, 1980.
18. Mehta, S. K., A New Method for the Evaluation of Neutron Emmission, Kerma, Kerma Equivalent, Dose Equivalent and the Spectrum Below 1 MeV for (alpha,n) Sources, Using Bonner Spheres, ORNL-TM-1171, 1965.

19. Poston, J. W., Knight, J. R., Whitesides, G. E., Calculation of the HPRR Neutron Spectrum for Simulated Nuclear Accident Conditions, Health Phys. 26, 217-221 (1974).
20. Hess, W. N., Patterson, H. W., Wallace, R., Cosmic-Ray Neutron Spectrum, Phys. Rev. 116(2), 445-457 (1959).
21. Green, L., Transmission Measurement of the ^{252}Cf Fission Neutron Spectrum, Nucl. Sci. & Eng. 37, 232-242 (1969).
22. Hewitt, J. E., Hughes, L., McCaslin, J. B., Smith, A. L., Stephens, L. D., Syvertson, C. A., Thomas, R. H., Tucker, A. B., Exposures to Cosmic-ray Neutrons at Commercial Jet Aircraft Altitudes, Natural Radiation Environment III, CONF-780422(Vol.2), 1980.
23. McGinley, P. H., Wood, M., Mills, M., Rodriguez, R., Dose Levels Due to Neutrons in the Vicinity of High-Energy Medical Accelerators, Med. Phys. 3, 397, 1976.
24. Miller, W. H., Sohl, J. D., Jr., Brugger, R. M., Experimental Tests of the Bonner Sphere Spectrometer Using Filtered Neutron Beams, Nucl. Instr. & Meth. 206, 219 (1983).

25. Sanna, R. S., Hajnal, F., McLaughlin, J. E., Energy-Dependent Effects on Neutron Monitor Performance in PWR Containments, Health Phys. 43, 263, 1982.
26. Extruding and Molding Grades, Book A. The International Plastics Selector, Cordura Publications, Inc., La Jolla, CA, 1978.
27. Griffith, R. V. Fisher, J. C., Measurement of the Responses of Multisphere Detectors with Monoenergetic Neutrons, 0.1 to 18.5 MeV, Hazards Control Progress Report No. 51, 29-33, UCRL-50007-75-2.
28. Davison, J. W., Hertel, N. E., Bonner Ball Detector Responses for Neutrons from Thermal Energies to 17.3 MeV, presented at the 28th Annual Meeting of the Health Physics Society, Baltimore, Maryland, June 19-24, 1983.
29. Distenfeld, C. H., Improvements and Tests of the Bonner Multisphere Spectrometer, Brookhaven National Laboratory, BNL-21293 (1975).
30. Nachtigal, D., Burger, G., Dose Equivalent Determinations in Neutron Fields by Means of Moderator Techniques, Chapt. 7 in Topics in Radiation Dosimetry, Suppl. 1, F. H. Attix, Ed., New York:Academic Press, 1972.

31. Nachtigal, D., Rohloff, F., Verfahren Zur Messung Der Flussdichten Und Dosisleistungsäquivalent Von Neutronen Im Energiebereich Zwischen Thermischer Energie Und 50 MeV, Nucl. Instr. Meth. 50, 137 (1967).
32. Hankins, D. E., The Multisphere Neutron-Monitoring Technique, LA-3700 (June 7, 1968).
33. Nachtigall, D., Rohloff, F., Sphere Techniques for Measuring Flux Densities and Dose Rates of Thermal, Intermediate and Fast Neutrons, JUL-213-ST, German, SLAC-TRANS-23, Trans., 1965.
34. Hajnal, F., McLaughlin, J., Weinstein, M., O'Brien, K., Sea Level Cosmic Ray Neutron Measurements, US AEC, HASL-241 (1971).
35. Hajnal, F., Sanna, R. S., Ryan, R. M., Donnelly, E. H., Stray Neutron Fields in the Containment of PWRs, in Occupational Radiation Exposures in Nuclear Full Cycle Facilities, IAEA-SM-242/24 (1979).
36. Sanna, R. S., Hajnal, F., McLaughlin, J. E., Gulbin, J. G., Ryan, R. M., Neutron Measurements Inside PWR Containments, U.S. DOE Report EML-379, 1980.

37. Crawford, R. L., Erteza, A., Discrimination Dependence of a Phoswich System on the Integrating Time Constant, Nucl. Instr. Meth. 30, 303 (1964).
38. Awschalom, M., Coulson, L., A New Technique in Environmental Neutron Spectroscopy, Proc. Third Int'l. Congress of IRPA, Washington, DC (1973) pp. 1464-1469.
39. Dhairyan, M. P., Nagarajan, P. S., Venkataraman, G., Response Functions of Spherically Moderated Neutron Detectors, Nucl. Instr. Meth. 169, 115 (1980).
40. Leake, J. W., An Improved Spherical Dose Equivalent Neutron Detector, Nucl. Instr. Meth. 63, 329 (1968).
41. Hankins, D. E., The Substitution of a BF_3 Probe for the LiI Crystal in Neutron rem-Meters, Health Phys. 14, 518 (1968).
42. Patterson, H. W., Thomas, R. H., Accelerator Health Physics, Academic Press: New York, 1973, Chapter 5.
43. Lanter, R. J., Bannerman, D. E., The Silver Counter A Detector for Bursts of Neutrons, LA-3498-MS (July 14, 1966).
44. Smith, A. R., Measurement of Radiation Field Around High-Energy Accelerators, UCRL-10163 (April, 1962).

45. Brown, D., Buchanan, R. J., Koelle, A. R., A Microcomputer-Based Portable Radiation survey Instrument for Measuring Pulsed Neutron Dose Rates, Health Phys. 38, 507 (1980).
46. Mandl, M. E., A Method of Measuring the Average Energy of a Neutron Spectrum in the Range 0.3 to 3.0 MeV, AERE T/R 1008, August, 1952.
47. Smith, A. R., A Tantalum Fast Neutron Integrator, University of California, Berkeley, Cal, Report UCRL-17051 (August 5, 1966).
48. Weinstein, M., Hajnal, F., McLaughlin, J., O'Brien, K., Neutron Dose Equivalents from Multisphere Accelerator Shield Leakage Spectra, US AEC, HASL-223 (March, 1970).
49. Ryan, R. M., Evaluation of Personnel Neutron Dosimetry at Operating Nuclear Power Plants, NUREG/CR-2524 (March, 1983).
50. Rohloff, F., Heinzelmann, M., Influence of Detector Type and Equipment on the Sensitivity of Bonner Spheres, in Neutron Monitoring for Radiation Protection Purposes, Vienna: IAEA, 1973, pp. 269-277.

51. Oster, C. A., Review of Unfolding Methods Used in the US and Their Standardization for Dosimetry, 2nd ASTM-Euratom Symposium on Reactor Dosimetry: Dosimetry Methods for Fuels Cladding and Structural Materials, October 3-7, 1977, Palo Alto, CA, 1977.
52. Bramblett, R. L., Bonner, T. W., Neutron Evaporation Spectra from (p,n) Reactions, Nucl. Phys. 20, 395 (1960).
53. Bricka, M., Neutrons Measurements near P.W.R. Power Reactors, Seventh DOE Workshop on Personnel Neutron Dosimetry, London, England, 1978, PNL2807, UC-48, 1978.
54. Gold, R., An Iterative Solution of the Matrix Representation of Detection Systems, TID-18304, 1963.
55. O'Brien, K., Sanna, R. S., McLaughlin, J. E., Inference of Accelerator Stray Neutron Spectra from Various Measurements, proceedings of the USEAC First Symposium on Accelerator Radiation Dosimetry and Experience, Brookhaven National Laboratory, Upton, New York, Nov. 3-5, 1965, CONF-651109, 1965.
56. Sanna, R. S., A Manual for BON: A Code for Unfolding Multisphere Spectrometer Measurements, US DOE Report EML-394, 1981.

57. Routti, J. T., Mathematical Considerations of Determining Neutron Spectra from Activation Measurements in Proceedings of 2nd International Symposium on Accelerator Dosimetry and Experience, Stanford University, November 5-7, 1969, Conf-691101, 1969.

58. Routti, J. T., High Energy Neutron Spectroscopy with Activation Detectors Incorporating New Methods for Analysis of Ge(Li) Gamma-Ray Spectra and the Solution of Fredholm Integral Equations, PhD Thesis, UCRL-18514 (April 1969).

59. McElroy, W. N., Berg, S., Crockett, T., Hawkins, R. A., Computer-Automated Iterative Method for Neutron Flux Spectra Determination by Foil Activation Data, AFWL-TR-67-41, Vol. 1 (1967).

60. Awschalom, M., Bonner Spheres and Tissue Equivalent Chambers for Extensive Radiation Area Monitoring Around a $1/2$ TeV Proton Synchrotron, IAEA/SM-167/69, probably 1973.

61. McGinley, P. H., Wood, M., Mills, M., Rodriguez, R., Dose Levels Due to Neutrons in the Vicinity of High-Energy Medical Accelerators, Med. Phys. 3, 397 (1976).

62. Holeman, G. R., Shaw, D. McM., Price, K. W., Stray Neutron Spectra and Comparison of Measurements with Discrete Ordinates Calculations, Second International Conference on Accelerator Radiation Dosimetry and Experience, held at the Stanford Linear Accelerator Center, Stanford, CA, November 5-7, 1969, CONF-691101, 552.

63. Aldrich, J. M., Haggard, D. L., Endres, G. W. R., Fix, J. J., Cummings, F. M., Thorson, M. R., Kathren, R. L., Evaluating Existing Radiation Fields, PNL-3536, 1981.

64. Hankins, D. E., Griffith, R. V., A Survey of Neutrons Inside the Containment of a Pressurized Water Reactor, UCRL-81346, ORNL/RSIC-43, ANS Winter Meeting, 1978.

65. Stephens, L. D., McCaslin, J. B., Smith, A. R., Thomas, R. H., Hewitt, J. E., Hughes, L., Ames Collaborative Study of Cosmic Ray Neutrons, II: Low- and Mid-Latitude Flights, LBL-6738, UC-34b, TID-4500-R66, 1978.

66. Huyskens, C. J., Jacobs, G. J. H., Calibration and Application of Multisphere Technique in Neutron Spectrometry and Dosimetry, 5th International Congress of the International Radiation Protection Association, 396, 1980, CONF-800304-, 1980.

67. Jacobs, G. J. H., van den Bosch, R. L. P., Calibration Measurements with the Multisphere Method and Neutron Spectrum Analysis Using the SAND-II Program, Nucl. Instr. Meth. 175, 483 (1980).
68. O'Brien, K., Sanna, R. S., Neutron Spectral Unfolding Using the Monte Carlo Technique, Nucl. Instr. Meth. 185, 277 (1981).
69. Sanna, R. S., O'Brien, K., Monte Carlo Unfolding of Neutron Spectra, Nucl. Instr. Meth. 91, 573 (1971).
70. Hanson, A. O., Radioactive Neutron Sources, Chapt. 1A in Fast Neutron Physics, Marion, J. B., Fowler, J. L., Eds. NY Interscience Publishers, 1960.
71. Kathren, R. L., Standard Sources for Health Physics Instrument Calibration, Health Phys. 29, 143 (1975).
72. Chartier, J. L., Champlong, P., Cosack, M., Wagner, S., Alberts, W. G., H., Schraube, H., Delafield, H. J. Hunt, J. B., Thompson, I. M. G., Lembo, L., Schwartz, R. B., Widell, C. O., Neutron Reference Radiations for Calibrating Neutron Measuring Devices Used for Radiation Protection Purposes and for Determining Their Response as a Function of Neutron Energy, Proc. Fourth Symposium on Neutron Dosimetry, Neuherberg, June 1-5, 1981.

73. Neutron Sources for Basic Physics and Applications, S. Cierjacks, editor, NY: Pergamon Press, 1983. Chapters II and III.

74. Jenkins, T. M., Simple Recipes for Ground Scattering in Neutron Detector Calibration, Health Phys. 39, 41 (1980).

75. Eisenhauer, C. M., Schwartz, R. B., Johnson, T., Measurement of Neutrons Reflected from the Surfaces of a Calibration Room, Health Phys. 42, 489 (1982).

76. Clough, A. S., Batty, C. J., Bonner, B. E., Tschalar, C., Williams, L. E., Neutron Spectra from the D(p,n) 2p Reactions at 30 and 50 MeV. Rutherford High Energy Lab, Didcot, England, report RPP/P 18.

77. Schwartz, R. B., Grundel, J. A., NBS Standard Reference Neutron Fields for Personnel Dosimetry, National and International Standardization of Radiation Dosimetry, Vienna: IAEA, 1978, Vol. I, p. 367.

78. Lorenz, A., A Survey of Neutron Sources and Their Applications, University of California, Livermore, CA, TID-4500,UC-34 (June 30, 1972).

79. Mr. Lester W. Jones, Monsanto Research Corporation, Dayton, Ohio, 45407, USA. Private communication.

80. Burrus, W. A., Bonner Spheres and Threshold Detectors for Neutron Spectroscopy, Neutron Physics Division Annual Progress Report for period ending Sept. 1, 1962, ORNL-3360, 296-305, 1962.

81. Maerker, R. E., Williams, L. R., Mynatt, F. R., Greene, N. M., Response Functions for Bonner Ball Neutron Detectors, ORNL-TM-3451, 1971.

82. Burgart, C. E., Emmett, M. B., Monte Carlo Calculations of the Response Functions of Bonner Ball Neutron Detectors, ORNL-TM-3739, 1972.

83. Nakamura, T., Kosako, T., Hauashi, K., Ban, S., Katoh, K., A Systematic Study on the Neutron Skyshine from Nuclear Facilities - Part 11. Experimental Approach to the Behavior of Environmental Neutrons Around an Electron Synchrotron, Nucl. Sci. & Eng. 77, 182 (1981).

DETECTOR ¹ (Dimens. in mm)	MODERATORS ²	ENERGY RANGE ³	METHOD	REF.	COMMENTS
4x4	2,3,5,8 & 12	Therm.-15	Exper.	4	Original BSS
4x4	Bare, Cd.,cvd. 2,3,5,8, & 12	Therm.-160	Extrap. Grp. av.	80	Used data of Ref. 4
4x4	2,3,5,8,10,18,16&18	Therm.-192	Theo.	7	DTK
4x4	2,3,5,8,10,12, 16 & 18	Therm.-192	Grp. av.	55	Used data of Ref.7
51 (diam.)BF ₃	2,3,4,5,6,8,10&12	Therm.-14	Theo.	81	ANISN
51 (diam.)BF ₃	Bare, 3 & 6	Therm.-15	Theo.	82	MORSE
4x4, 8x8, 12.7x12.7 & gold foil	Bare,2,3,5,8,10, 12 & 18	Therm.-100	Theo.	8	DTF-IV
12.7 x 12.7	Bare,2,3,5,8,10, 12 & 18, H ₂ O	Therm.-100	Theo.	8	DTF-IV
12.7 x 12.7	2,3,5,8,10,12&18	0.1-18 MeV	Exper.	27	Verification of Ref. 8 data
4x4	2,3,5,8,10&12	Therm.-15	Theo.	39	Monte Carlo
51 (diam.)BF ₃	3,5,8,10 & 12	Therm.-15	Theo.	39	Monte Carlo
203 (long)BF ₃	Bare,1.18 & 2.36 thick cyl.	Therm.-15	Theo.	83	ANISN-JR
4x4	Bare & 2 to 20 1/2 inch increm'ts.	Therm.-17	Theo.	28	ANISN

Grp.	ENERGY, MeV
1	2.312E+02 TO 4.000E+02
2	1.336E+02 TO 2.312E+02
3	7.725E+01 TO 1.336E+02
4	4.465E+01 TO 7.725E+01
5	2.581E+01 TO 4.465E+01
6	1.492E+01 TO 2.581E+01
7	7.408E+00 TO 1.492E+01
8	3.679E+00 TO 7.408E+00
9	1.827E+00 TO 3.679E+00
10	9.072E-01 TO 1.827E+00
11	4.508E-01 TO 9.072E-01
12	2.237E-01 TO 4.508E-01
13	1.111E-01 TO 2.237E-01
14	5.248E-02 TO 1.111E-01
15	2.479E-02 TO 5.248E-02
16	1.171E-02 TO 2.479E-02
17	5.531E-03 TO 1.171E-02
18	2.613E-03 TO 5.531E-03
19	1.234E-03 TO 2.613E-03
20	5.930E-04 TO 1.234E-03
21	2.754E-04 TO 5.930E-04
22	1.301E-04 TO 2.754E-04
23	6.144E-05 TO 1.301E-04
24	2.902E-05 TO 6.144E-05
25	1.371E-05 TO 2.902E-05
26	6.476E-06 TO 1.371E-05
27	3.059E-06 TO 6.476E-06
28	1.445E-06 TO 3.059E-06
29	6.826E-07 TO 1.445E-06
30	4.140E-07 TO 6.826E-07
31	THERMAL

Table 2

Energy structure for 31 group response functions.⁸

Table 3

Multisphere spectrometer response matrix for polyethylene

moderators ($\rho=0.95 \text{ g cm}^{-3}$), 4 mm x 4 mm ^6Li detector.⁸

GRP.	BARE	2''	3''	5''	8''	10''	12''	18''
1	4.6976E-06	1.2274E-04	1.0990E-03	8.1887E-03	2.5791E-02	3.6530E-02	4.5101E-02	6.0432E-02
2	7.9545E-06	1.4230E-04	1.2418E-03	9.0718E-03	2.7837E-02	3.8754E-02	4.7023E-02	5.9752E-02
3	1.2187E-05	1.7340E-04	1.4579E-03	1.0216E-02	2.9912E-02	4.0501E-02	4.7852E-02	5.6262E-02
4	1.8732E-05	2.2399E-04	1.8495E-03	1.2700E-02	3.5823E-02	4.7160E-02	5.4066E-02	5.7515E-02
5	3.0015E-05	3.3619E-04	2.7406E-03	1.8369E-02	4.9672E-02	6.3345E-02	7.0151E-02	6.6262E-02
6	4.5946E-05	5.6625E-04	4.6141E-03	3.0087E-02	7.7386E-02	9.5157E-02	1.0138E-01	8.4305E-02
7	5.0065E-05	1.1240E-03	9.5620E-03	6.2602E-02	1.5676E-01	1.8753E-01	1.9337E-01	1.4300E-01
8	1.0212E-04	2.8111E-03	2.1791E-02	1.2033E-01	2.4234E-01	2.5271E-01	2.2732E-01	1.1088E-01
9	1.7899E-04	6.3270E-03	4.3506E-02	1.9417E-01	2.9716E-01	2.5929E-01	1.9443E-01	5.3320E-02
10	2.7026E-04	1.3117E-02	7.8327E-02	2.7672E-01	3.1410E-01	2.2464E-01	1.3709E-01	1.9273E-02
11	4.4492E-04	2.2515E-02	1.1167E-01	2.9930E-01	2.3996E-01	1.3148E-01	6.5539E-02	4.2358E-03
12	1.8542E-03	3.5256E-02	1.4577E-01	3.0607E-01	1.8344E-01	8.7274E-02	3.5767E-02	1.4869E-03
13	1.1751E-03	4.8106E-02	1.7119E-01	2.9329E-01	1.4143E-01	6.0500E-02	2.2973E-02	8.1618E-04
14	8.6945E-04	6.0893E-02	1.8931E-01	2.7627E-01	1.1537E-01	4.6842E-02	1.7308E-02	5.8120E-04
15	1.1018E-03	7.3447E-02	2.0392E-01	2.6292E-01	9.9956E-02	3.9555E-02	1.4468E-02	4.7145E-04
16	1.4500E-03	8.5947E-02	2.1740E-01	2.5440E-01	9.0743E-02	3.5413E-02	1.2893E-02	4.1165E-04
17	1.9348E-03	9.9158E-02	2.3095E-01	2.4873E-01	8.4475E-02	3.2668E-02	1.1862E-02	3.7243E-04
18	2.5755E-03	1.1357E-01	2.4477E-01	2.4427E-01	7.9611E-02	3.0582E-02	1.1083E-02	3.4271E-04
19	3.6579E-03	1.2949E-01	2.5862E-01	2.3989E-01	7.5388E-02	2.8809E-02	1.0425E-02	3.1762E-04
20	5.2735E-03	1.4707E-01	2.7216E-01	2.3497E-01	7.1450E-02	2.7190E-02	9.8278E-03	2.9500E-04
21	7.6454E-03	1.6634E-01	2.8493E-01	2.2916E-01	6.7632E-02	2.5652E-02	9.2632E-03	2.7387E-04
22	1.0811E-02	1.8717E-01	2.9629E-01	2.2211E-01	6.3817E-02	2.4144E-02	8.7115E-03	2.5357E-04
23	1.5240E-02	2.0928E-01	3.0564E-01	2.1382E-01	6.0000E-02	2.2657E-02	8.1696E-03	2.3401E-04
24	2.1592E-02	2.3206E-01	3.1219E-01	2.0421E-01	5.6152E-02	2.1176E-02	7.6313E-03	2.1502E-04
25	3.0286E-02	2.5451E-01	3.1486E-01	1.9308E-01	5.2201E-02	1.9669E-02	7.0849E-03	1.9626E-04
26	4.1793E-02	2.7498E-01	3.1239E-01	1.8023E-01	4.8075E-02	1.8107E-02	6.5193E-03	1.7746E-04
27	5.6322E-02	2.9081E-01	3.0303E-01	1.6529E-01	4.3553E-02	1.6441E-02	5.9170E-03	1.5819E-04
28	7.4135E-02	2.9784E-01	2.8439E-01	1.4762E-01	3.8738E-02	1.4594E-02	5.2508E-03	1.377E-04
29	9.4778E-02	2.8930E-01	2.5294E-01	1.2606E-01	3.2995E-02	1.2442E-02	4.4758E-03	1.1514E-04
30	1.1327E-01	2.6279E-01	2.1168E-01	1.0261E-01	2.6903E-02	1.0161E-02	3.6564E-03	9.2207E-05
31	1.6854E-01	1.2852E-01	9.7515E-02	4.7543E-02	1.2653E-02	4.8054E-03	1.7332E-03	4.2525E-05

Table 4

Multisphere spectrometer response matrix for polyethylene

moderators ($\rho=0.95 \text{ g cm}^{-3}$), 8 mm x 8 mm ^6Li detector.⁸

GRP.	BARE	2''	3''	5''	8''	10''	12''	18''
1	3.7765E-05	5.2507E-04	4.0441E-03	2.8583E-02	8.8714E-02	1.2521E-01	1.5420E-01	2.0542E-01
2	6.3898E-05	6.1803E-04	4.5805E-03	3.1676E-02	9.5759E-02	1.3284E-01	1.6078E-01	2.0311E-01
3	9.7882E-05	7.6236E-04	5.3868E-03	3.5679E-02	1.0290E-01	1.3883E-01	1.6360E-01	1.9124E-01
4	1.5034E-04	9.9289E-04	6.8447E-03	4.4364E-02	1.2323E-01	1.6165E-01	1.8484E-01	1.9548E-01
5	2.4014E-04	1.4920E-03	1.0141E-02	6.4161E-02	1.7085E-01	2.1710E-01	2.3979E-01	2.2516E-01
6	3.6683E-04	2.4879E-03	1.7041E-02	1.0505E-01	2.6613E-01	3.2605E-01	3.4647E-01	2.8640E-01
7	4.0010E-04	4.8064E-03	3.5168E-02	2.1846E-01	5.3903E-01	6.4242E-01	6.6063E-01	4.8565E-01
8	8.1679E-04	1.1865E-02	7.9972E-02	4.1967E-01	8.3256E-01	8.6483E-01	7.7570E-01	3.7619E-01
9	1.4310E-03	2.6362E-02	1.5930E-01	6.7653E-01	1.0193E+00	8.8557E-01	6.6190E-01	1.8064E-01
10	2.1546E-03	5.4062E-02	2.8617E-01	9.6275E-01	1.0746E+00	7.6472E-01	4.6493E-01	6.5336E-03
11	3.5567E-03	9.2192E-02	4.0714E-01	1.0387E+00	8.1749E-01	4.6113E-01	2.2104E-01	1.4569E-03
12	1.4685E-02	1.4402E-01	5.3013E-01	1.0587E+00	6.2211E-01	2.9421E-01	1.2013E-01	5.3011E-03
13	9.3027E-03	1.9307E-01	6.1980E-01	1.0107E+00	4.7788E-01	2.0335E-01	7.7027E-02	3.0143E-03
14	6.9113E-03	2.4274E-01	6.8326E-01	9.4898E-01	3.8894E-01	1.5722E-01	5.8017E-02	2.1983E-03
15	8.7511E-03	2.9165E-01	7.3388E-01	9.0069E-01	3.3652E-01	1.3268E-01	4.8502E-02	1.8123E-03
16	1.1498E-02	3.4006E-01	7.8024E-01	8.6961E-01	3.0526E-01	1.1875E-01	4.3232E-02	1.6013E-03
17	1.5303E-02	3.9089E-01	8.2658E-01	8.4863E-01	2.8401E-01	1.0952E-01	3.9782E-02	1.4630E-03
18	2.0310E-02	4.4594E-01	8.7350E-01	8.3198E-01	2.6754E-01	1.0251E-01	3.7180E-02	1.3585E-03
19	2.8730E-02	5.0620E-01	9.2008E-01	8.1576E-01	2.5326E-01	9.6553E-02	3.4982E-02	1.2702E-03
20	4.1096E-02	5.7195E-01	9.6496E-01	7.9777E-01	2.3996E-01	9.1119E-02	3.2988E-02	1.1902E-03
21	5.8971E-02	6.4301E-01	1.0065E+00	7.7682E-01	2.2709E-01	8.5959E-02	3.1102E-02	1.1151E-03
22	8.2564E-02	7.1846E-01	1.0424E+00	7.5183E-01	2.1423E-01	8.0897E-02	2.9260E-02	1.0422E-03
23	1.1463E-01	7.9678E-01	1.0706E+00	7.2275E-01	2.0139E-01	7.5909E-02	2.7451E-02	9.7113E-04
24	1.5878E-01	8.7535E-01	1.0885E+00	6.8938E-01	1.8845E-01	7.0944E-02	2.5653E-02	9.0127E-04
25	2.1603E-01	9.4994E-01	1.0926E+00	6.5103E-01	1.7517E-01	6.5893E-02	2.3827E-02	8.3118E-04
26	2.8685E-01	1.0143E+00	1.0786E+00	6.0709E-01	1.6132E-01	6.0657E-02	2.1936E-02	7.5964E-04
27	3.8739E-01	1.0590E+00	1.0411E+00	5.5632E-01	1.4648E-01	5.5072E-02	1.9920E-02	6.8469E-04
28	4.5725E-01	1.0699E+00	9.7249E-01	4.9656E-01	1.2998E-01	4.8884E-02	1.7688E-02	6.0329E-04
29	5.4393E-01	1.0248E+00	8.6126E-01	4.2383E-01	1.1071E-01	4.1671E-02	1.5087E-02	5.1052E-04
30	6.0763E-01	9.1903E-01	7.1837E-01	3.4491E-01	9.0270E-02	3.4030E-02	1.2334E-02	4.1419E-04
31	7.2354E-01	4.4430E-01	3.3029E-01	1.5980E-01	4.2455E-02	1.6092E-02	5.8542E-03	1.9493E-04

Table 5

Multisphere spectrometer response matrix for polyethylene

moderators ($\rho=0.95 \text{ g cm}^{-3}$), 12.7 mm x 12.7 mm ^6Li detector.⁸

GRP.	BARE	2''	3''	5''	8''	10''	12''	18''
1	1.5104E-03	2.6322E-03	9.3415E-03	5.3314E-02	1.5879E-01	2.2232E-01	2.7255E-01	3.5990E-01
2	8.0742E-04	2.0846E-03	9.6389E-03	5.8179E-02	1.7054E-01	2.3501E-01	2.8335E-01	3.5509E-01
3	4.6054E-04	1.9979E-03	1.0811E-02	6.5046E-02	1.8279E-01	2.4516E-01	2.8790E-01	3.3396E-01
4	3.3241E-04	2.2982E-03	1.3459E-02	8.0626E-02	2.1868E-01	2.8522E-01	3.2503E-01	3.4113E-01
5	3.9104E-04	3.3305E-03	1.9862E-02	1.1650E-01	3.0308E-01	3.8293E-01	4.2152E-01	3.9275E-01
6	7.4214E-04	5.7175E-03	3.3449E-02	1.9084E-01	4.7209E-01	5.7504E-01	6.0890E-01	4.9932E-01
7	1.5989E-03	1.1802E-02	6.9598E-02	3.9724E-01	9.5633E-01	1.1329E+00	1.1608E+00	8.4621E-01
8	3.2678E-03	2.8593E-02	1.5759E-01	7.6214E-01	1.4749E+00	1.5224E+00	1.3601E+00	6.5352E-01
9	5.7206E-03	6.2287E-02	3.1238E-01	1.2262E+00	1.8007E+00	1.5534E+00	1.1558E+00	3.1182E-01
10	8.5852E-03	1.2550E-01	5.5859E-01	1.7400E+00	1.8897E+00	1.3339E+00	8.0668E-01	1.1141E-01
11	1.4210E-02	2.1165E-01	7.9128E-01	1.8685E+00	1.4269E+00	7.9741E-01	3.7993E-01	2.4143E-02
12	5.6420E-02	3.2835E-01	1.0251E+00	1.8928E+00	1.0774E+00	5.0486E-01	2.0499E-01	8.4538E-03
13	3.6809E-02	4.2943E-01	1.1886E+00	1.7949E+00	8.2257E-01	3.4731E-01	1.3099E-01	4.6589E-03
14	2.7451E-02	5.3390E-01	1.3024E+00	1.6756E+00	6.6698E-01	2.6797E-01	9.8543E-02	3.3292E-03
15	3.4721E-02	6.3689E-01	1.3912E+00	1.5829E+00	5.7586E-01	2.2594E-01	8.2341E-02	2.7071E-03
16	4.5538E-02	7.3752E-01	1.4713E+00	1.5227E+00	5.2170E-01	2.0214E-01	7.3374E-02	2.3681E-03
17	6.0466E-02	8.4181E-01	1.5506E+00	1.4813E+00	4.8497E-01	1.8639E-01	6.7503E-02	2.1457E-03
18	7.9981E-02	9.5311E-01	1.6299E+00	1.4481E+00	4.5656E-01	1.7444E-01	6.3073E-02	1.9772E-03
19	1.1246E-01	1.0728E+00	1.7071E+00	1.4161E+00	4.3196E-01	1.6429E-01	5.9332E-02	1.8350E-03
20	1.5998E-01	1.2005E+00	1.7794E+00	1.3814E+00	4.0911E-01	1.5503E-01	5.5938E-02	1.7067E-03
21	2.2750E-01	1.3347E+00	1.8439E+00	1.3418E+00	3.8702E-01	1.4625E-01	5.2729E-02	1.5867E-03
22	3.1454E-01	1.4723E+00	1.8962E+00	1.2957E+00	3.6502E-01	1.3763E-01	4.9594E-02	1.4712E-03
23	4.2927E-01	1.6093E+00	1.9331E+00	1.2429E+00	3.4306E-01	1.2915E-01	4.6514E-02	1.3598E-03
24	5.8118E-01	1.7395E+00	1.9502E+00	1.1832E+00	3.2097E-01	1.2070E-01	4.3455E-02	1.2515E-03
25	7.6835E-01	1.8542E+00	1.9421E+00	1.1156E+00	2.9833E-01	1.1211E-01	4.0350E-02	1.1442E-03
26	9.8385E-01	1.9419E+00	1.9025E+00	1.0388E+00	2.7472E-01	1.0321E-01	3.7135E-02	1.0364E-03
27	1.2096E+00	1.9869E+00	1.8229E+00	9.5091E-01	2.4943E-01	9.3708E-02	3.3709E-02	9.2544E-04
28	1.4239E+00	1.9674E+00	1.6915E+00	8.4809E-01	2.2133E-01	8.3184E-02	2.9920E-02	8.0757E-04
29	1.6029E+00	1.8491E+00	1.4899E+00	7.2351E-01	1.8852E-01	7.0913E-02	2.5509E-02	6.7628E-04
30	1.7127E+00	1.6329E+00	1.2379E+00	5.8863E-01	1.5371E-01	5.7914E-02	2.0844E-02	5.4273E-04
31	1.8617E+00	7.8070E-01	5.6810E-01	2.7270E-01	7.2292E-02	2.7389E-02	9.8840E-03	2.5113E-04

Table 6

Multisphere spectrometer response matrix for water

moderators ($\rho=1.0 \text{ g cm}^{-3}$), 12.7×12.7 ^6Li detector.⁸

GRP.	BARE	2''	3''	5''	8''	10''	12''	18''
1	1.5104E-03	2.5296E-03	8.7698E-03	6.3491E-02	2.4975E-01	3.9218E-01	5.2049E-01	7.8702E-01
2	8.0742E-04	1.9519E-03	8.8890E-03	6.8267E-02	2.6149E-01	4.0135E-01	5.2029E-01	7.3250E-01
3	4.6054E-04	1.9269E-03	1.0684E-02	8.4035E-02	3.1406E-01	4.7339E-01	6.0275E-01	8.0450E-01
4	3.3241E-04	2.2526E-03	1.3293E-02	1.0049E-01	3.5086E-01	5.0673E-01	6.1869E-01	7.2782E-01
5	3.9104E-04	2.5843E-03	1.5057E-02	1.1126E-01	3.7406E-01	5.2558E-01	6.2322E-01	6.6466E-01
6	7.4214E-04	3.6308E-03	2.0266E-02	1.4789E-01	4.8521E-01	6.6785E-01	7.7389E-01	7.6150E-01
7	1.5989E-03	6.4220E-03	3.4901E-02	2.5497E-01	8.2649E-01	1.1215E+00	1.2771E+00	1.1772E+00
8	3.2678E-03	1.4492E-02	7.6378E-02	4.9293E-01	1.3582E+00	1.6671E+00	1.7194E+00	1.1662E+00
9	5.7206E-03	3.2494E-02	1.6607E-01	9.0431E-01	1.9715E+00	2.0890E+00	1.8603E+00	7.9403E-01
10	8.5852E-03	6.7546E-02	3.1881E-01	1.3841E+00	2.2451E+00	1.9703E+00	1.4471E+00	3.2286E-01
11	1.4210E-02	1.1685E-01	4.7817E-01	1.6280E+00	1.9940E+00	1.4712E+00	9.0866E-01	1.1358E-01
12	5.6420E-02	1.9009E-01	6.5312E-01	1.7680E+00	1.6836E+00	1.0722E+00	5.7828E-01	4.7507E-02
13	3.6809E-02	2.5747E-01	8.3769E-01	1.9214E+00	1.5311E+00	8.8931E-01	4.4576E-01	2.9933E-02
14	2.7451E-02	3.3788E-01	9.8542E-01	1.9308E+00	1.3278E+00	7.2260E-01	3.4677E-01	2.0641E-02
15	3.4721E-02	4.2613E-01	1.1138E+00	1.9024E+00	1.1887E+00	6.2223E-01	2.9224E-01	1.6228E-02
16	4.5538E-02	5.1715E-01	1.2319E+00	1.9180E+00	1.1001E+00	5.6207E-01	2.6095E-01	1.3826E-02
17	6.0466E-02	6.1534E-01	1.3494E+00	1.9237E+00	1.0377E+00	5.2105E-01	2.4009E-01	1.2247E-02
18	7.9981E-02	7.2418E-01	1.4695E+00	1.9315E+00	9.8776E-01	4.8915E-01	2.2416E-01	1.1046E-02
19	1.1246E-01	8.4622E-01	1.5922E+00	1.9358E+00	9.4281E-01	4.6151E-01	2.1059E-01	1.0036E-02
20	1.5998E-01	9.8212E-01	1.7156E+00	1.9323E+00	8.9925E-01	4.3584E-01	1.9820E-01	9.1338E-03
21	2.2750E-01	1.1315E+00	1.8367E+00	1.9183E+00	8.5543E-01	4.1111E-01	1.8646E-01	8.3014E-03
22	3.1454E-01	1.2925E+00	1.9509E+00	1.8907E+00	8.1012E-01	3.8656E-01	1.7496E-01	7.5162E-03
23	4.2927E-01	1.4620E+00	2.0531E+00	1.8483E+00	7.6345E-01	3.6216E-01	1.6367E-01	6.7752E-03
24	5.8118E-01	1.6343E+00	2.1369E+00	1.7899E+00	7.1526E-01	3.3774E-01	1.5248E-01	6.0733E-03
25	7.6835E-01	1.8001E+00	2.1936E+00	1.7129E+00	6.6482E-01	3.1284E-01	1.4114E-01	5.4004E-03
26	9.8385E-01	1.9466E+00	2.2126E+00	1.6151E+00	6.1144E-01	2.8702E-01	1.2946E-01	4.7498E-03
27	1.2096E+00	2.0556E+00	2.1798E+00	1.4926E+00	5.5373E-01	2.5954E-01	1.1706E-01	4.1108E-03
28	1.4239E+00	2.1010E+00	2.0756E+00	1.3393E+00	4.8937E-01	2.2925E-01	1.0343E-01	3.4678E-03
29	1.6029E+00	2.0418E+00	1.8709E+00	1.1443E+00	4.1429E-01	1.9417E-01	8.7670E-02	2.7957E-03
30	1.7127E+00	1.8646E+00	1.5825E+00	9.2772E-01	3.3506E-01	1.5730E-01	7.1124E-02	2.1523E-03
31	1.8617E+00	9.1860E-01	7.2293E-01	4.2098E-01	1.5451E-01	7.3026E-02	3.3172E-02	9.2931E-04

Table 7

Multisphere spectrometer response matrix for polyethylene

moderators ($\rho=0.95 \text{ g cm}^{-3}$), 12.7 mm diam x 0.051 mm thick, ^{197}Au detector.⁸

GRP.	BARE	2''	3''	5''	8''	10''	12''	18''
1	1.0958E-06	3.5764E-05	2.6695E-04	1.7654E-03	5.3156E-03	7.4577E-03	9.1699E-03	1.2250E-02
2	1.4338E-06	4.0742E-05	3.0074E-04	1.9544E-03	5.7352E-03	7.9092E-03	9.5577E-03	1.2108E-02
3	1.8589E-06	4.8941E-05	3.5196E-04	2.1991E-03	6.1595E-03	8.2618E-03	9.7218E-03	1.1396E-02
4	2.6307E-06	6.0172E-05	4.4350E-04	2.7306E-03	7.3727E-03	9.6154E-03	1.0978E-02	1.1642E-02
5	3.9027E-06	8.9492E-05	6.5595E-04	3.9478E-03	1.0220E-02	1.2911E-02	1.4238E-02	1.3402E-02
6	5.8073E-06	1.5223E-04	1.1056E-03	6.4672E-03	1.5920E-02	1.9387E-02	2.0565E-02	1.7034E-02
7	9.2158E-06	3.2013E-04	2.3067E-03	1.3470E-02	3.2257E-02	3.8200E-02	3.9204E-02	2.8860E-02
8	1.5544E-05	8.0242E-04	5.2567E-03	2.5876E-02	4.9749E-02	5.1308E-02	4.5898E-02	2.2246E-02
9	2.7383E-05	1.8176E-03	1.0502E-02	4.1700E-02	6.0706E-02	5.2276E-02	3.8919E-02	1.0562E-02
0	4.1547E-05	3.7879E-03	1.8916E-02	5.9257E-02	6.3591E-02	4.4743E-02	2.7044E-02	3.7297E-03
1	5.8559E-05	6.5192E-03	2.6965E-02	6.3683E-02	4.7807E-02	2.6576E-02	1.2633E-02	7.8133E-04
2	1.0299E-04	1.0203E-02	3.5155E-02	6.4491E-02	3.5881E-02	1.6704E-02	6.7614E-03	2.5998E-04
3	1.3888E-04	1.4141E-02	4.1231E-02	6.1057E-02	2.7229E-02	1.1430E-02	4.2997E-03	1.3727E-04
4	1.4831E-04	1.7970E-02	4.5396E-02	5.6813E-02	2.1980E-02	8.7955E-03	3.2276E-03	9.5408E-05
15	2.3205E-04	2.1691E-02	4.8621E-02	5.3481E-02	1.8923E-02	7.4070E-03	2.6937E-03	7.6145E-05
16	4.3266E-04	2.5375E-02	5.1508E-02	5.1273E-02	1.7112E-02	6.6225E-03	2.3984E-03	6.5711E-05
17	7.4569E-04	2.9229E-02	5.4336E-02	4.9719E-02	1.5887E-02	6.1046E-03	2.2051E-03	5.8873E-05
18	1.2867E-03	3.3384E-02	5.7130E-02	4.8448E-02	1.4942E-02	5.7118E-03	2.0592E-03	5.3692E-05
19	2.3173E-03	3.7888E-02	5.9809E-02	4.7219E-02	1.4126E-02	5.3787E-03	1.9359E-03	4.9329E-05
20	5.1383E-03	4.2727E-02	6.2263E-02	4.5900E-02	1.3370E-02	5.0751E-03	1.8240E-03	4.5416E-05
21	6.5401E-03	4.7736E-02	6.4362E-02	4.4425E-02	1.2642E-02	4.7872E-03	1.7183E-03	4.1784E-05
22	6.5372E-03	5.3006E-02	6.5970E-02	4.2738E-02	1.1918E-02	4.5050E-03	1.6150E-03	3.8328E-05
23	4.0437E-03	5.8564E-02	6.6955E-02	4.0846E-02	1.1198E-02	4.2272E-03	1.5136E-03	3.5032E-05
24	1.4549E-02	6.4474E-02	6.7139E-02	3.8746E-02	1.0475E-02	3.9507E-03	1.4130E-03	3.1870E-05
25	3.2680E-04	6.9178E-02	6.6197E-02	3.6407E-02	9.7351E-03	3.6695E-03	1.3109E-03	2.8791E-05
26	6.1889E-03	7.4207E-02	6.4027E-02	3.3804E-02	8.9642E-03	3.3780E-03	1.2053E-03	2.5757E-05
27	2.7340E-01	7.7257E-02	5.9891E-02	3.0868E-02	8.1391E-03	3.0669E-03	1.0930E-03	2.2700E-05
28	1.4460E-02	5.3951E-02	5.2539E-02	2.7492E-02	7.2224E-03	2.7226E-03	9.6926E-04	1.9555E-05
29	1.0625E-02	5.2258E-02	4.6780E-02	2.3482E-02	6.1518E-03	2.3209E-03	8.2538E-04	1.6138E-05
30	1.1569E-02	4.7870E-02	3.9254E-02	1.9118E-02	5.0160E-03	1.8954E-03	6.7355E-04	1.2752E-05
31	2.9284E-02	2.3721E-02	1.8123E-02	8.8588E-03	2.3592E-03	8.9630E-04	3.1872E-04	5.7643E-06

DETECTOR ¹ (Dimens. in mm)	MODERATORS ²	ENERGY RANGE ³	METHOD	REF
4x4	2,3,5,8 & 12	Therm.-15	Exper.	4
4x4	Bare, Cd.,cvd. 2,3,5,8, & 12	Therm.-160	Extrap. Grp. av.	80
4x4	2,3,5,8,10,18,16&18	Therm.-192	Theo.	7
4x4	2,3,5,8,10,12, 16 & 18	Therm.-192	Grp. av.	55
51 (diam.)BF ₃	2,3,4,5,6,8,10&12	Therm.-14	Theo.	81
51 (diam.)BF ₃	Bare, 3 & 6	Therm.-15	Theo.	82
4x4, 8x8, 12.7x12.7 & gold foil	Bare,2,3,5,8,10, 12 & 18	Therm.-100	Theo.	8
12.7 x 12.7	Bare,2,3,5,8,10, 12 & 18, H ₂ O	Therm.-100	Theo.	8
12.7 x 12.7	2,3,5,8,10,12&18	0.1-18 MeV	Exper.	27
4x4	2,3,5,8,10&12	Therm.-15	Theo.	39
51 (diam.)BF ₃	3,5,8,10 & 12	Therm.-15	Theo.	39
203 (long)BF ₃	Bare,1.18 & 2.36 thick cyl.	Therm.-15	Theo.	83
4x4	Bare & 2 to 20 1/2 inch increm'ts.	Therm.-17	Theo.	28

List of Tables & Titles

Table 1. Summary of Published Response Data for BSS Detectors.

Notes:

(1) All dimensions in mm. 4x4 means 4 mm diameter, 4 mm long right cylinder. ${}^6\text{Li}$ is assumed in place of ${}^6\text{LiI(Eu)}$ unless otherwise indicated.

(2) Diameter of moderators in inches. Material is polyethylene unless otherwise indicated.

(3) Energy in MeV.

Table 2. Energy Structure for 31 Group Response Functions.⁸

Table 3. Multisphere spectrometer response matrix for polyethylene moderators ($\rho=0.95\text{ g cm}^{-3}$), 4 mm x 4mm ${}^6\text{Li}$ detector.⁸

Table 4. Multisphere spectrometer response matrix for polyethylene moderators ($\rho=0.95\text{ g cm}^{-3}$), 8 mm x 8 mm ${}^6\text{Li}$ detector.⁸

Table 5. Multisphere spectrometer response matrix for polyethylene moderators ($\rho=0.95\text{ g cm}^{-3}$), 12.7 mm x 12.7 mm ${}^6\text{Li}$ detector.⁸

Table 6. Multisphere spectrometer response matrix for water moderators ($\rho=1.0 \text{ g cm}^{-3}$), 12.7 mm x 12.7 mm ^6Li detector.⁸

Table 7. Multisphere spectrometer response matrix for polyethylene moderators ($\rho=0.95 \text{ g cm}^{-3}$), 12.7 mm diam x 0.051 mm thick, ^{197}Au detector.⁸

FIGURE CAPTIONS

1. The response of a multisphere spectrometer with a 4 mm x 4 mm ^6Li detector: I = bare, and at the center of polythene moderators, II = 2 inch, III = 3 inch, IV = 5 inch, V = 8 inch, VI = 10 inch, and VII = 12 inch diameters.⁸
2. Neutron energy spectra for: I = HPRR neutrons,¹⁹ II = Cosmic-ray neutrons,²⁰ and III = ^{252}Cf neutron source.²¹
3. Multisphere response distributions for: I = HPRR neutrons,¹⁹ II = Cosmic-ray neutrons,²⁰ and III = ^{252}Cf neutron source.²¹
4. Multisphere response distributions for three "monochromatic" neutrons. I = energy group 31, II = energy group 20, and III = Energy group 7.
5. Ratio of responses as a function of moderator density, material, and detector size. I = response ratio as a function of density ($\rho = 0.90$ to $\rho = 0.95 \text{ g cm}^{-3}$) for polyethylene spheres 10 inch in diameter, II = ratio of the responses as a function of materials (H_2O to polyethylene) for 10 inch diameter spheres, with a 12.7 mm x 12.7 mm ^6Li detector,⁸ III = ratio of the response of a 4 mm x 4 mm to that of a 12.7 mm x 12.7 mm ^6Li detector at the center of a 2 inch polyethylene sphere.

6. Phoswich assembly. ${}^6\text{LiI}(\text{Eu})$ crystal is surrounded by a plastic scintillator cup cemented to a glass disc which is cemented to an aluminum can. The inside of the can is coated with Al_2O_3 .

7. Events recorded by a 4 mm x 4 mm ${}^6\text{LiI}(\text{Eu})$ crystal by a phoswich, for 48.3 hours, in a 10 inch polyethylene pseudosphere, due to environmental radiation. The solid squares represent all events. The open circles represent neutron signatures: slow pulses not accompanied by fast pulses.

8. Events recorded by a 8 mm x 8 mm ${}^6\text{LiI}(\text{Eu})$ crystal in a phoswich, for 70.8 hours, in a 10 inch pseudosphere, due to environmental radiation. The solid squares represent all events. The open circles represent neutron signatures: slow pulses not accompanied by fast pulses.

Figure 1

Figure 2

Figure 3

Figure 4

Figure 5

Figure 6

Figure 7

Figure 8