27
9/24/79
2467515

SAND 79-0549
Unlimited Release

The Physical Properties of $\text{NH}_4\text{Cl} \cdot 3\text{NH}_3$

MASTER

D. A. Nissen

Sandia Laboratories
energy report

THE PHYSICAL PROPERTIES OF $\text{NH}_4\text{Cl} \cdot 3\text{NH}_3$

D. A. Nissen
Exploratory Chemistry Division I 8313
Sandia Laboratories, Livermore

ABSTRACT

The physical properties (viscosity, surface tension, and density) of $\text{NH}_4\text{Cl} \cdot 3\text{NH}_3$ were measured over the temperature range 14–26°C and at the equilibrium NH_3 vapor pressure ~ 5 atm, using a recently fabricated linear oscillation viscometer. Over this temperature range the viscosity varied from ~ 4 to ~ 6 cP and the density varied from 0.935 to 0.939 g/cm³; the surface tension at 24°C was 28 dyne/cm. The melting point of $\text{NH}_4\text{Cl} \cdot 3\text{NH}_3$ was found to be $5.1 \pm 0.1^\circ\text{C}$.

NOTICE
This report was prepared as an account of work sponsored by the United States Government. Neither the United States nor the United States Department of Energy, nor any of their employees, nor any of their contractors, subcontractors, or their employees, makes any warranty, express or implied, or assumes any legal liability or responsibility for the accuracy, completeness, or usefulness of any information, apparatus, product, or process disclosed, or represents that its use would not infringe privately owned rights.

CONTENTS

	<u>Page</u>
Introduction	9
Apparatus and Experimental Technique	10
A. Viscosity	10
B. Density	12
C. Surface Tension	14
Experimental Procedure	14
Results and Discussion	15
References	19

ACKNOWLEDGMENT

The author wishes to acknowledge, with gratitude, the help of Ken Geis in designing and fabricating the viscometer and Carl Schoenfelder in synthesizing the ammoniate.

THE PHYSICAL PROPERTIES OF $\text{NH}_4\text{Cl} \cdot 3\text{NH}_3$

Introduction

The compound $\text{NH}_4\text{Cl} \cdot 3\text{NH}_3$ is an attractive candidate for use in thermochemical heat pump/storage systems. This is principally because it is a liquid at temperatures above 5°C and consequently possesses a heat transfer coefficient approximately 50-fold greater than comparable solid ammoniates, e.g. $\text{MgCl}_2 \cdot 6\text{NH}_3$ or $\text{CaCl}_2 \cdot 8\text{NH}_3$.¹ Furthermore, the rate of the ammoniation reaction

appears to be an order of magnitude greater than that of the solid ammoniates.¹

While preliminary investigations of $\text{NH}_4\text{Cl} \cdot 3\text{NH}_3$ yielded encouraging results regarding its use in heat pump systems, it was not possible to make detailed system design and performance studies because nothing was known about the physical properties (viscosity, surface tension, and density). Because of the high dissociation pressure of $\text{NH}_4\text{Cl} \cdot 3\text{NH}_3$ -- the equilibrium NH_3 pressure over $\text{NH}_4\text{Cl} \cdot 3\text{NH}_3$ is ~ 5 atm at 25°C (Fig. 1) -- these measurements could not easily be made using conventional methods. Furthermore, liquid $\text{NH}_4\text{Cl} \cdot 3\text{NH}_3$ will not wet glass,¹ which means that the use of capillary methods of measuring viscosity and surface tension were ruled out. However, an instrument was recently built at Sandia Laboratories Livermore with the capability of making physical property measurements at elevated temperatures (1000°C) and pressures up to 10 atm. This apparatus was used to determine the physical properties of liquid $\text{NH}_4\text{Cl} \cdot 3\text{NH}_3$ from -14 to 26°C , and the results are presented here.

Figure 1. Pressure versus Reciprocal Temperature for $\text{NH}_4\text{Cl} \cdot 3\text{NH}_3$; taken from Troost, Ref. 2. The dashed line is a linear extrapolation of the data.

Apparatus and Experimental Technique

The physical-property measuring instrument is based on the principle of a damped, one-dimensional, harmonic oscillator, i.e., the motion of a body suspended from a spring and oscillating in a fluid. The instrument is a variation of a continuous ultrasonic viscometer developed by Roth and Rice³ and is similar in concept to the linear oscillation viscometer of White and Solomons,⁴ who present a detailed discussion of the performance of their apparatus as well as the equations which describe its motion.⁵ A detailed discussion of both the theoretical principles which form the foundation for this instrument as well as its performance will be published elsewhere.⁶ Only a very brief description of its various modes of operation will be presented here.

A. Viscosity

The heart of the instrument consists of a quartz spring oscillator, an electromagnet for remotely starting the spring oscillation, a position transducer for remote readout of the spring's extension (linear variable differential transform, LVDT), and a gold plate which is suspended in the liquid whose viscosity is being measured (Fig. 2). It is the viscous drag

Figure 2. Schematic Drawing of Experimental Apparatus

exerted on this plate by the liquid which causes damping of the oscillatory motion of the quartz spring, which in turn provides a measure of the viscosity of the liquid being studied.

The lower section of the viscometer is made of heavy-walled stainless steel tubing and contains the sample and provision for its motion. The liquid under study is contained in a crucible mounted on a pedestal fastened to the screw and yoke arrangement which is attached to the bottom flange shown in Fig. 2. Connection between the pedestal and screw is a machined stainless steel rod which passes through a compression fitting designed to maintain the integrity of the atmosphere within the viscometer. This arrangement permits the crucible and its contents to be raised or lowered, smoothly and slowly, about 4 cm. A thermocouple is fed through a fitting in the bottom flange and bent to allow its tip to be immersed in the liquid. In this way the temperature of the liquid under study could be monitored continuously. A tube for admitting NH_3 under pressure is also welded onto this flange.

An MKS Instruments Baratron capacitance manometer with a range of 0 - 10,000 torr is attached to the top flange in order to monitor the pressure continuously. Thick (1.3 cm) quartz windows are mounted on flanges located 180° apart to allow observation of the crucible and its contents.

When the crucible is properly positioned, the spring is started oscillating by applying a 6-V, 17-msec pulse to the coil of the electromagnet. By measuring the rate at which the oscillations of the spring decay (see Fig. 3 for typical data output), it is possible to determine the viscosity of the liquid under study from the equation

$$\sqrt{np} = 37.04 \delta + 0.148, \quad (1)$$

where n and p are the viscosity and density of the liquid, respectively, and δ , the logarithmic decrement, is the rate of decay of the spring oscillation. The constants 37.04 and 0.148 were determined by calibration with liquids of known viscosity and density under the appropriate pressure of argon. The calibration curve is shown in Fig. 4, and a detailed discussion of the calibration method is given in Ref. 6.

B. Density

In addition to functioning as a viscometer it is possible, with a few minor modifications, to convert the apparatus into an instrument for measuring density and surface tension.

If the gold plate (Fig. 2) is replaced by a solid body of known volume, the apparatus can function as an Archimedean densitometer. The change in buoyancy (ΔW) of a solid of known volume (V) is related to the density of the liquid in which it is immersed by

$$\rho = \frac{\Delta W}{V}.$$

Figure 3. Typical Data Trace for a Viscosity Measurement

Figure 4. Calibration Curve (at 5 atm) for Viscosity Measurements

Therefore, the density of a liquid is given simply by the difference in the LVDT reading (converted to weight) with and without liquid surrounding the bob divided by the known volume of the bob.

The bob used in these experiments was made of gold in the shape of a right cylinder. Its volume, $0.2747 + 0.00006 \text{ cm}^3$, was determined prior to these experiments by immersing it in a liquid whose density was accurately known and measuring the change in buoyancy.

C. Surface Tension

By slowly and continuously separating the liquid from the gold plate immersed in it and measuring the maximum force exerted on the plate, which occurs just prior to breaking the column of liquid supported by the plate, it is possible to determine the surface tension of the liquid from the expression

$$\gamma = \frac{F}{P \cos \theta}$$

In this expression γ is the surface tension, F is the maximum force exerted on the plate (determined as an apparent increase in the weight of the plate), P is the perimeter of the plate, and θ is the wetting angle, generally assumed to be zero if the liquid wets the plate.

The surface tension of $\text{NH}_4\text{Cl} \cdot 3\text{NH}_3$ was determined by this method at the conclusion of the viscosity experiments.

Experimental Procedure

Synthesis of the ammoniate was achieved by condensing the stoichiometric quantity of liquid ammonia onto dried NH_4Cl contained in a stainless steel reactor.⁷ Prior to use the NH_4Cl was oven dried at 80°C for a minimum of 48 hours. A sample of this material ($\sim 25 \text{ g}$) was weighed into a stainless steel reactor, 35 mm in diameter and 60 mm high. The reactor was then sealed and attached to a vacuum line. Following evacuation, the reactor was cooled with liquid nitrogen and the stoichiometric quantity of ammonia, which had previously been condensed in another part of the system, was condensed onto the NH_4Cl , forming the triammoniate. After the transfer was complete (~ 2 hours) the reactor was valved off, removed from the vacuum line, weighed, and transferred to the viscometer. To ensure that the ammoniate had in fact been synthesized an experiment was performed to measure several pressure-temperature data points and compare them with the curve established by Troost² (Fig. 1). On this basis as well as from the increase in weight of the reactor and its contents there appears to be little doubt that $\text{NH}_4\text{Cl} \cdot 3\text{NH}_3$ was synthesized.

The transfer of the reactor to the viscometer was done by first cooling the reactor and its contents in liquid nitrogen, removing the reactor cover, positioning the reactor on the pedestal, inserting the whole arrangement into the apparatus, and attaching the bottom flange to the viscometer. After being sealed the apparatus was pressurized with NH_3 gas to 80 psia, the equilibrium pressure at 25°C . When the ammoniate had melted, the reactor was raised to immerse the gold plate and thermocouple.

The oscillation curve for the ammoniate was found at several different temperatures. Each measurement was made at least five times. These data were then used to determine a value for the logarithmic decrement δ .⁶ This value was then used in Eq. (1) to determine a value for the viscosity-density product. At the conclusion of the viscosity experiments the surface tension was measured by slowly lowering the crucible and its contents away from the gold plate and measuring the maximum force exerted on the plate as the supported liquid column broke (cf. paragraph C in the previous section).

The procedure for determining the density of the ammoniate was identical to that used in the viscosity experiments except that the gold plate was replaced by a gold density bob (cf. paragraph B in the previous section).

Results and Discussion

The experimentally determined values of the viscosity, density, and surface tension for $\text{NH}_4\text{Cl} \cdot 3\text{NH}_3$ are given in Table I and plotted in Fig. 5. Based on the accuracy with which the viscosity and density of standard liquids could be determined,⁶ it is estimated that the viscosity data are accurate to 5% and the density data to 0.5%. The melting point of $\text{NH}_4\text{Cl} \cdot 3\text{NH}_3$ was determined by allowing the gold plate used in the viscosity experiments to touch the surface of the solid. As the sample gradually warmed up the surface began to liquify causing a slight but perceptible motion of the plate. This temperature was taken to be the melting point. Using this method, three separate experiments gave the melting point of $\text{NH}_4\text{Cl} \cdot 3\text{NH}_3$ as $5.1 \pm 0.1^\circ\text{C}$.

TABLE I
VISCOSITY, DENSITY, AND SURFACE TENSION OF $\text{NH}_4\text{Cl} \cdot 3\text{NH}_3$

T ($^\circ\text{C}$)	Viscosity (cP)	Density (g/cm^3)	Surface Tension (dyne/cm)
14	7.6	0.9395	--
18	5.35	0.938	--
22	4.80	0.937	--
24	--	--	28
26	3.90	0.935	--

Figure 5. Viscosity and Density of $\text{NH}_4\text{Cl}\cdot 3\text{NH}_3$ as a Function of Temperature

Viscous flow is a thermally activated process. As a consequence, we may write

$$\eta = \eta_0 e^{E/RT}$$

where E is the activation energy of flow. Figure 6 is a plot of viscosity versus the reciprocal absolute temperature. From the slope of this curve the value for E was determined to be 5.7 kcal.

With a viscosity of $\sim 5\text{cP}$ and an activation energy of flow of $\sim 6\text{ kcal}$, $\text{NH}_4\text{Cl}\cdot 3\text{NH}_3$ closely resembles a 40% sucrose-water solution.⁸ However, the value of 28 dyne/cm for the surface tension is more typical of an organic liquid than an aqueous solution. These data probably reflect a situation in which there is no extensive interaction, e.g., hydrogen bonding, between $\text{NH}_4\text{Cl}\cdot 3\text{NH}_3$ molecules. Rather, the relatively large value of the viscosity is probably due to hindered motion as a consequence of the large $\text{NH}_4\text{Cl}\cdot 3\text{NH}_3$ molecules.

Figure 6. Log n versus T^{-1} . The slope is E, the Activation Energy of Flow

REFERENCES

1. F. A. Jaeger, W. R. Haas, and J. E. Anderson, "Development of Ammoniated Salts Thermochemical Energy Storage Systems Phase II," Martin Marietta, 1979.
2. L. Troost, Compt. Rend. 88, 578, 1879.
3. W. Roth and S. R. Rich, J. Appl. Phys. 24, 940 (1953).
4. M. S. White and C. Solomons, Rev. Sci. Instr. 40, 339 (1969).
5. C. Solomons and M. S. White, Trans. Faraday Soc. 64, 305 (1968).
6. D. A. Nissen, to be published.
7. C. W. Schoenfelder, to be published.
8. J. R. Van Wazer, J. W. Lyons, K. Y. Kim, and R. E. Colwell, Viscosity and Flow Measurement (Interscience, New York, 1963), p.37.

UNLIMITED RELEASE

INITIAL DISTRIBUTION

U. S. DOE/STOR
Washington, D. C. 20545
Attn: J. H. Swisher
J. Gahmer

Solar Energy Research Institute
1536 Cole Boulevard
Golden, CO 80201
Attn: B. Butler
C. Wyman

F. A. Jaeger
Martin Marietta Corp.
Denver, CO 80118

W. E. Wentworth
Department of Chemistry
University of Houston
Houston, Texas 77004

Dr. M. Taube
Swiss Federal Institute
for Reactor Research
CH-5305 Würenlingen
Switzerland

R. S. Claassen, 5800; Attn: J. J. Saxton, 5840

T. B. Cook, 8000; Attn: A. N. Blackwell, 8200
L. Gutierrez, 8400

B. F. Murphey, 8300; Attn: T. S. Gold, 8320
G. W. Anderson, 8330
W. Bauer, 8340
D. Hartley, 8350

D. M. Schuster, 8310

D. R. Adolphson, 8312

R. W. Mar, 8313

D. A. Nissen, 8313 (5)

W. R. Hoover, 8314

L. A. West, 8315

J. C. Swearingen, 8316

R. C. Wayne, 8450; Attn: W. G. Wilson, 8451

J. D. Gilson, 8453

T. T. Bramlette, 8453

Publications and Public Information Division, 8265, for TIC (27)

F. J. Cupps, 8265/Technical Library Processes Division, 3141

Technical Library Processes Division, 3141 (2)

Library and Security Classification Division, 8266 (3)