E&TR

MASTER

UCRL--52000-86-2

DE86 009611

Energy and Technology Review

Lawrence Livermore National Laboratory

February 1986

New Laser Amplifier Improves Laser Doppler Interferometry	1
A specialized laser amplifier for use with our velocity-measuring systems makes possible more detailed measurements of explosion-driven targets extending over longer times.	
The Bunker 801 Project	8
Bunker 801, our recently completed hydrodynamics test facility at Site 300, is the most advanced of its kind in the free world.	
Neutron Spectrometry for Dosimetry	18
We have developed improved spectrometry systems and are testing a new dosimeter material to make our neutron dosimetry program for personnel protection more effective.	
Abstracts	26

UCRL-52000-86-2
Distribution Category UC-2
February 1986

DISTRIBUTION OF THIS DOCUMENT IS UNLIMITED

About the Cover

Streak cameras, which produce records like the one shown on this cover, are one of the many diagnostic techniques used to follow hydrodynamics experiments at Site 300. Hydrodynamics tests of explosive materials and exploding devices take place very quickly—the detonation often lasts only a few tens of millionths of a second. Gathering data about such short-lived events requires diagnostic instruments that can respond in a very short time frame and that can be precisely synchronized to the instant of detonation. At Site 300's Bunker 801, we have assembled the experimental and diagnostic facilities that enable us to conduct extremely sensitive and thorough hydrodynamics tests of nonnuclear explosions. For more information about this facility, see the article beginning on p. 8.

Scientific Editor

Hriar S. Cabayan

General Editors

Wallace Clements
Richard B. Crawford
Lauren de Vore
Derek P. Hendry
Eleanor M. O'Neal

Graphics Designers

John Schuster
Kimberly Smith

Compositor

Louisa Cardoza

New Laser Amplifier Improves Laser Doppler Interferometry

We have developed a specialized laser amplifier for use with our velocity-measuring systems that enables us to obtain more detailed measurements of explosion-driven targets over longer periods of time.

Many Laboratory research programs require the study of processes that occur in time scales of microseconds or less. These include studies of the equations of state of a wide variety of materials, the behavior of materials subjected to strong shock waves and other hydrodynamic phenomena, the explosion process, and the behavior of projectiles and targets upon and immediately after impact. These investigations continually demand advances in the state of the art in high-speed measurement. Therefore, we are constantly developing better and better methods of making time-resolved measurements of changes that occur in the explosion-driven targets over very short time periods.

Laser Doppler Interferometry

Laser Doppler interferometry advances our ability to measure high velocities of explosion-driven objects or "targets" (see the box on p. 2).¹ In our system, 514.5-nm light from an argon-ion laser is focused onto the

target. Return (reflected) light is frequency-analyzed by a Fabry-Perot interferometer. The interference pattern is optically transformed from a bull's-eye pattern to a row of light spots, from whose separation we can calculate target velocity. The change in the separation of the light spots is recorded by a streak camera to give a time-resolved history of the velocity of the target.

This laser Doppler interferometry system produces excellent results. With it we are able to measure velocities of 10 to 10 000 m/s over time periods of less than 4 μ s to about 10 μ s. Time resolution depends on the interferometer fill time (that is, the time it takes the light to build up to its steady-state value), which is typically less than 20 ns.

Figure 1 is a typical streak camera record of an accelerated object. Before the first shock arrives, the surface being viewed is stationary, and the record shows straight lines—that is, the bright interference spots are at constant radii. The first shock wave rapidly accelerates the surface,

For more information contact
Lloyd L. Steinmetz (415) 422-6897

producing a jump in the radii of the bright spots on the streak camera slit. Relatively straight lines again show the surface coasting at a constant velocity, until a second shock wave arrives to accelerate the object to a new higher velocity. Although further accelerations occur later, the record is

lost because there is insufficient return light. Here, the readable record extends for about $6\ \mu\text{s}$.

The typical record in Fig. 1 does provide valuable information about surface velocity, acceleration, and position as a function of time. However, after the first few

Laser Doppler Interferometry

In one form of a laser-based velocity-measuring system (illustrated below), the laser light reflected from the target under study is frequency-analyzed with a Fabry-Perot interferometer. Passage through the interferometer turns the laser beam into a bull's-eye interference pattern. So long as the target is moving at a constant velocity (including zero), the bull's-eye fringe pattern does not change. However, a changing target velocity shifts the frequency of the return light and thus changes the interference pattern. If the target accelerates toward the laser source, the frequency of the return light increases and the bull's-eye expands. Contraction of the bull's-eye may also occur if the target slows during its flight toward the laser source. It is this expansion (or contraction) that we want to measure.

Before the reflected laser light enters the interferometer, it passes through a cylindrical lens that

concentrates the light in each fringe into a pair of spots. By thus concentrating the faint return light, we are able to see some changes more clearly and to follow the target for a longer period of time than would otherwise be possible. We record the pattern of spots with a streak camera, synchronized to record when the target starts to move. The streak camera slit sees the radial expansion of the interference fringe pattern as expansion of a row of spots. The diagram below illustrates how an increase in target velocity affects the streak camera record. As the target velocity increases, the spots move apart radially and a new pair appears in the center.

With the unamplified system, we can measure target velocities from 10 to 10 000 m/s as the target moves 2 to 3 cm in as little as $4\ \mu\text{s}$. The change in frequency is typically about 0.001%, and the velocity measurements are accurate to about 1%.

microseconds, the record is faint and, in some cases, unusable; this fading is believed to be a result of deterioration of the target's reflectivity and perhaps of other physical changes in the target. After a few more microseconds, the return signal from the target is lost completely, even though the rest of the recording system is still operating perfectly.

Moreover, we would like to record velocities simultaneously from as many locations on the target as possible to verify that the target surface behaves as our calculations predict. Given the current state of the art, we would like to follow simultaneously two or three locations on the target. Clearly however, if we were to split the available light between two or three spots, the return signals would fade and be lost even sooner.

Amplifier Design Considerations

To improve and extend our measurements of target velocity, therefore, we undertook to increase the laser light intensity, the most straightforward way to increase the signal-to-noise ratio. To this end, we designed and built an amplifier for the laser light.² By building a separate amplifier instead of redesigning the original laser, we achieved greater flexibility; the system can be operated with or without the amplifier as required by each particular experiment.

We selected an organic dye amplifier for its simplicity and because suitable organic dyes are available covering the spectral range that we wanted to use. We chose a green coumarin dye (C.314) as the lasing medium for the argon-ion laser. We also chose a blue dye (C.120) as a frequency shifter to improve the coupling between the light-pump power source (flashlamps) and the lasing medium. Figure 2 shows the arrangement of the major components of the amplifier and the frequency characteristics of the flashlamps and the dyes.

Great care must be taken in the design of a laser amplifier; otherwise, an amplifier will distort the laser. The amplifier can easily change the dominant laser frequency or the spread of the frequency distribution (bandwidth), or the mode of the laser, or other important properties.

Frequency Control

The amplifier must not change the frequency of the laser light. If the amplifier changed the frequency, we could not distinguish a change in target velocity from a laser artifact. Standing waves in the amplifier medium can produce abrupt shifts in frequency (chirping). We prevent chirping by using the double-pass traveling-wave design shown in Fig. 3.

Bandwidth

The bandwidth of the laser light is controlled to 80 MHz, corresponding to a wavelength difference of about 0.01 nm. This bandwidth must be smaller than an anticipated change in frequency

Fig. 1

A typical streak camera record of the output of the Fabry-Perot interferometer in our original velocity-measuring system. The record shows the changes in frequency of the laser light reflected from a target accelerated by an explosion. Before the first shock, the surface is stationary, as shown by the straight lines. The first shock wave produces a jump in the lines as it accelerates the surface. The following relatively straight lines reveal that the surface is coasting at constant velocity until a second shock arrives to further accelerate the target. The readable record extends for about 6 μ s.

Fig. 2

Diagram of the main parts of our laser amplifier, including lasing medium, frequency-shifter dye, and relative positions of the flashlamps, diffusing surface, and interfaces. The plot shows the spectral characteristics of the flashlamp power source, the frequency-shifter dye (C.120), and the laser dye (C.314). The C.120 dye absorbs wavelengths around a peak of 350 nm and emits or fluoresces around a peak of 440 nm. The C.314 dye absorbs wavelengths around a peak of 445 nm and lases around a peak of 505 nm. Thus, energy from the flashlamp absorbed by the frequency-shifter dye is ultimately transferred to the lasing dye.

resulting from the Doppler shift. The traveling-wave amplifier design (Fig. 3) ensures bandwidth fidelity as it ensures frequency stability.

Acoustic Isolation

Any vibration in the laser medium creates density differences that defocus the beam. Thus, we must isolate the amplifier volume from the shock waves induced in the air by the sudden flashlamp heating. In our amplifier design, acoustic isolation of the laser medium from the flashlamps is provided by physical separation and by several intervening layers of acoustic impedance mismatches (air-glass interfaces).

Pointing Stability

Pointing stability of the laser beam is required to keep it from wandering on the target too much. In our experiments, the pointing stability limit is 10^{-4} rad.

Thermally Induced Artifacts

The amplifier design must also address problems associated with thermal gradients in the lasing dye solution. Thermal gradients in the amplifier medium of even a few tenths of a degree per centimetre are not tolerable because they distort the beam, producing focusing and steering effects. Since the amplifier must operate for a relatively long time (more than $10 \mu\text{s}$), it is essential that thermally induced effects be controlled.

Because the lasing dye solvent's index of refraction is a function of temperature, thermal gradients produce a lens effect. This is a time-integrated process, and the thermally induced positive lens effect becomes stronger with time. There are two principal sources of the thermal gradient. One is the nonuniformity of the pump light absorption, which permits some of the energy from the absorbed pump light to be distributed radially among the active molecules as heat. The other source of the thermal gradient is the broadband output of

the flashlamps, most of which would only heat the lasing medium without contributing to gain. The key to amplifier performance is to filter out the nonproductive wavelengths and to ensure radial uniformity of the pump wavelengths.

Solvent

We use a 50/50 water/methanol mixture as the dye solvent. The lasing dye readily dissolves in this mixture, it has high heat capacity, and its refractive index is only moderately temperature-dependent.

A further complication arises from the fact that the metastable excited triplet state in the lasing dye acts as a bottleneck to inhibit the lasing process. A "triplet quencher" is almost always required for dye lasers with pulses longer than a few microseconds. A ubiquitous triplet quencher is solvent-dissolved, ground-state oxygen; the oxygen solubility ($10^{-3} M$) in both water and methanol provides an adequate density for quenching. Without the dissolved oxygen, the usable amplified pulse is shortened to less than $2 \mu\text{s}$. (Other potential triplet quenchers, such as anthracene derivatives, have favorable energy levels, but are incompatible with the laser dye solvent.)

Dye Concentration

It is important to provide uniform excitation (and hence heating) across the active volume of the amplifier. Because the excitation light is

Fig. 3

Double-pass traveling-wave design for the prevention of frequency instabilities in the amplified laser pulse. The laser light makes only two passes through the amplifier, thereby precluding the formation of standing waves with their attendant frequency instabilities. The quarter-wave rotator turns the light polarization 90 deg after the two passes, allowing separation of input and output light by a polarization-selective prism.

absorbed transverse to the axis, uniform excitation requires a judicious choice for the active dye concentration. The concentration is adjusted to give an excitation absorption length of three diameters of the active cross section. A higher dye concentration gives more gain but earlier distortion, whereas a lower concentration gives less gain but a longer undistorted pulse. Other schemes, such as longitudinal pumping, may permit longer pulses to be amplified without distortion.

Streak camera records of the amplified beam show it being focused along the beam axis, with the focal point moving back toward the amplifier. The rate at which the focal point moves is a strong function of lasing dye concentration (and hence gain). With the modest gain employed in our amplifier, the beam focus recedes toward the amplifier at a few meters per microsecond.

A radial temperature difference of just a few tenths of a degree from axis to edge (2 mm) will produce this lens effect. A weak symmetric lens is tolerable, but a nonsymmetric lens produces steering of the beam, making it unusable at the target. Steering of more than 10^{-4} rad is unacceptable. Measurements on our prototype amplifier show that the steering is within this limit over the usable pulse width.

Optical Frequency Shifter

To increase the amount of pumping light reaching the laser dye and, simultaneously, to decrease the unwanted wavelengths, we surround the active amplifier volume with an optical frequency shifter (see Fig. 2). This is a layer of solution with a different dye (C.120) that absorbs ultraviolet light and reemits blue light that the amplifier can use. Unwanted infrared light from the flashlamps also is absorbed in the shifter solvent, where the thermally induced distortions are not important.

The frequency shifter routinely improves the amplifier gain by 50%. Temporary elimination of the shifter

fluid not only reduces the gain, but produces a premature termination of the amplified pulse. We have used a nonresonant probe laser (HeNe) to confirm the presence of thermal distortions and the effectiveness of using the frequency-shifter filter.

Reflective Coating

We coat the inner wall of the amplifier housing with a barium sulfate reflective coating that reflects flashlamp photons back toward the frequency-shifter dye. It is important to use a good coating to prevent circumferential thermal gradients that would otherwise arise from the localized heat sources, namely, the two flashlamps. The barium sulfate coating is significantly superior to several others we tried, including titanium dioxide, surface polish, surface etch, and combinations thereof. For example, amplifier gain with a barium sulfate coating is twice that with bare polished aluminum.

Shaped Gain

As the experiment progresses, the target disintegrates and target reflectance degrades. Therefore, greater light intensity is needed later in the record. A constant gain in light intensity would produce "overexposure" artifacts early in the record. We are able to shape the gain of the amplifier so that increased intensity (up to ten times) is produced later in time, when the target reflectivity is decreasing.

The gain of the amplifier is a function of the intensity of the pumping, and this can be varied by changing the current to the flashlamps. Thus, if the target's reflectivity is known to decrease late in the time record, the flashlamp current and hence the gain can be increased or "ramped up" to compensate.

Results

With this new laser amplifier, we can produce sufficient light to look at two or more spots on the target. Thus, we can now follow

the motion of one area of the target relative to another. We can also choose different recording parameters for the two spots and thereby correlate details of target motion. For example, we can record with different camera speeds on the two spots to detail different phenomena or to achieve better control of the time resolution at some critical point.

The amplifier also makes it possible to obtain longer records. As the target disintegrates, it reflects less light, and at some point the measuring system fails. However, by sending more light to the target, we extend the duration of the usable returned signal.

Figure 4 is a typical record obtained with our improved system using the laser amplifier. Comparison with the record in Fig. 1 reveals that the late periods of the record are now much clearer. This record extends to 12 μ s after the arrival of the shock wave, in contrast to the 6 μ s of Fig. 1, and the return signal is readable to the very end of the record. With the improved records, velocity can now be read to 0.5% accuracy, compared to the 1% accuracy without the amplifier.

Another feature that is revealed by Fig. 4 and similar records is the diffuseness of the streaked spot. This effect is quite visible at about 9 μ s after the shock. This is not an artifact, but represents a real velocity dispersion. That is, since particles are being ejected from the surface, the return light is coming from particles with different velocities. This phenomenon has been confirmed by other photographic evidence. Records like Fig. 4 give the first proof of velocity dispersion in submillimetre-sized target spots.

Prospects

Future experiments of high-speed shock-wave, hydrodynamic, or explosive processes will be less restricted as a result of the tenfold increase in light levels provided by our new laser amplifier.

Fig. 4

A typical streak camera record of the output of the Fabry-Perot interferometer with our amplified-laser velocity-measuring system. The record shows the changes in frequency of the laser light reflected from a target accelerated by an explosion. The target was followed for about 12 μ s, during which time it traveled about 6 cm. (An additional streak camera was used to follow the movement for a further: 4 μ s at higher resolution.) The fuzziness of the recorded spots after about 9 μ s indicates a velocity dispersion resulting from the ejection of particles from the surface of the target.

Measurements on more than two target locations are now possible. In addition, low-reflectivity targets or those with pronounced surface deterioration are now candidates for velocity measurements. The availability of increased light also opens the possibility of multipoint optical-fiber interferometry. □

Key Words: amplifier—laser; laser interferometry—Doppler; target—high-velocity; velocimetry.

Notes and References

1. Earlier work on measuring velocities by laser Doppler interferometry was discussed in the March 1979 issue of *Energy and Technology Review* (UCRL-52000-79-3), pp. 17-24.
2. Our new laser amplifier was briefly described in the July 1985 issue of *Energy and Technology Review* (UCRL-52000-85-7), pp. 32-33.

The Bunker 801 Project

The experimental and diagnostic facilities in Bunker 801 at Site 300 enable us to conduct extremely sensitive and thorough hydrodynamics tests on the high-explosive components of nuclear devices.

For further information contact
Larry L. Shaw (415) 422-4561.

Since its inception, the Laboratory has developed and studied high explosives. The necessary detonation tests of the explosives are performed at Site 300, an isolated area some 25 km east of the Laboratory's main site. Distributed among the barren hills of Site 300 are various specialized test installations, including facilities for formulating, machining, and assembling high explosives, for subjecting them to various environmental stresses, and for exploding them to observe their hydrodynamic behavior.

Bunker 801 is our newest, most elaborate, and most fully integrated facility for studying the hydrodynamics of nonnuclear explosions (see Fig. 1). Completed in

1985, Bunker 801 represents a major improvement in our ability to define and analyze the hydrodynamic characteristics of new weapon designs. It is the most advanced facility of its kind in the free world and probably in the entire world.

Hydrodynamics tests of high explosives take place very quickly; the detonation lasts only for a few tens of millionths of a second. Gathering information about such a short-lived event requires diagnostic instruments that can respond in such a short time frame and that can be precisely synchronized to the instant of detonation. To obtain a broad spectrum of information about an experimental explosion, we have developed the ability to use several diverse, synchronized, high-speed diagnostics to obtain data about a single explosive event. The equipment in Bunker 801 enables us to use advanced techniques in high-speed photography, pulsed radiography, electro-optic diagnostics, interferometry, and electrical-probe tests. In addition, we use computers in innovative ways to prepare and monitor the experiment and diagnostics, enabling us to conduct very complex experiments.

Forerunners of Bunker 801
Bunker 801 is notable among the installations at Site 300 in that it is the first hydrodynamics

Fig. 1

View of Bunker 801
as it looks today,
taken from the hill-
side rising behind
it.

facility that we were able to upgrade almost completely. We were thus able to incorporate state-of-the-art technology in all facets of the new facility, and upgrade rather than merely improve existing inadequate features. Nevertheless, Bunker 801 did have antecedents that fulfilled similar roles.

In 1955, the first hydrodynamics test facility, Bunker 301, was constructed at Site 300. This bunker, after several alterations, was eventually encapsulated in the new Bunker 801; it is now a machine shop in the new facility.

Located at the end of a winding, gravel road, the old bunker was a quonset hut, about 3 by 15 m, buried under a mound of earth and sand bags. Only the massive blast-protection door, stenciled with the bunker number, remained visible (see Fig. 2). The interior of Bunker 301 was divided into two compartments. The front compartment contained a few pieces of electronic control and diagnostic equipment, a desk, a work bench, and a few book shelves.

The rear half of the bunker was the camera room. Four camera ports in the ceiling of the subterranean room provided access to glass port holes that protruded above the earth and gravel covering the facility. Cameras inside the bunker viewed the shot table (the gravel surface, located near the bunker, on which the experiment took place) by way of mirrors mounted over the camera ports.

In 1955, the only diagnostic technique employed in the bunker was high-speed photography, and multicamera synchronization was not possible. The photography was done with high-speed mechanical cameras that incorporated fast rotary mirrors to sweep an image of the hydrodynamic event across a stationary film.^{1,2} Two types of rotary-mirror camera were used: one, the framing camera, used internal relay optics to create a series of sequential pictures of the event; in the second, the streak camera, an aperture slit viewing a thin plane of the experiment was swept along the

film. The framing camera gave us a series of snapshots of the exploding material, whereas the streak camera provided a continuous record over time of a narrow slice of the surface being photographed (see Fig. 3). From the streak image, we could measure position versus time and could infer velocity versus time and the uniformity of the moving surface.

The high-speed framing cameras then available provided up to 160 pictures at a rate of 1.5 million frames per second. The streak camera then available could record images at the rate of approximately 4 mm/ μ s. The streak camera was a synchronized camera, meaning that the camera rotor had to be in position in its rotation relative to the recording film before the experiment could commence. All our film records were made on black-and-white negative film.

By 1958, our optical diagnostic capabilities had increased to the point that we could synchronize two cameras to photograph a hydrodynamics test. In addition, 12

Fig. 2

The original Bunker 301 as it looked in 1957. The face of the bunker with its numbered door is on the right. The structures to the left of the main bunker are small storage bunkers. The firing table, where the detonations took place, is on the dirt fill above and to the left of Bunker 301. The gas bottles in front of Bunker 301 held supplies of helium and nitrogen gas used to drive the rotating mirrors in the high-speed cameras. A camera port can be seen on the top of Bunker 301.

raster oscilloscopes had been installed in Bunker 301. These oscilloscopes displayed the electrical signals from numerous metal pins, or probes, arrayed within the high-explosive charge being tested. The 12

oscilloscopes could accommodate as many as 84 electrical probes in a single test.

In 1959, Site 300 underwent an extensive construction program. Several facilities were added or

Fig. 3

Examples of high-speed photographs of hydrodynamics experiments at Site 300. In the two experiments here, a metal cylinder filled with high explosive is detonated at one end, and the dynamic effect of the explosive on the metal cylinder is monitored. From the data obtained, we can determine parameters such as detonation velocity and chemical energy released. (a) Framing-camera record of an exploding cylinder. The three

consecutive frames were taken $1.5 \mu\text{s}$ apart. (b) Streak-camera record of a similar test. The left end of the film strip shows the cylinder set up for the shot; the vertical, white hair line indicates the slice of the cylinder the camera observed. The streak on the film, moving left to right here at a rate of $3.3448 \text{ mm}/\mu\text{s}$, records the outward movement of the exploding cylinder's surface.

remodeled. Bunker 301 was closed, and construction was begun at the same spot on a reinforced concrete structure that represented, in size and capability, a major step forward in our ability to conduct hydrodynamics tests. The redesigned bunker was outfitted with a complement of 48 raster oscilloscopes and six improved high-speed cameras. The advances in high-speed camera technology and the availability of fast color film made high-speed photography a much more useful diagnostic tool. Where before we had been able to use only two cameras, now we could photograph a given high-explosive experiment with up to five cameras simultaneously.

At the same time Bunker 301 was being restructured, another facility, Bunker 351, was being readied as the main center for radiographically or pin-diagnosed hydrodynamics experiments. Bunker 351 also acquired a linear accelerator for pulsed radiography. Because the equipment in it afforded more diagnostics options, Bunker 351 gradually superseded Bunker 301 as a weapon-test facility. For some time, Bunker 301 was used exclusively by LLNL chemists as a laboratory for studying high-explosive formulations. (In 1969, the 300-series building designations at Site 300 were changed to 800-series, and Bunker 301 thus became Bunker 801.)

In the mid-1970s, we undertook yet another general modernization of the hydrodynamics test facilities at Site 300.² Bunker 801 was selected to house the new linear-induction accelerator that would serve as the powerful, pulsed, flash-x-ray machine known as the FXR. Construction of the FXR facility was begun in 1978 and completed in 1982.

The Bunker 801 Project

While work was progressing on the FXR, we began another construction effort, the Bunker 801 Project, at the same location. The facility envisioned in the Bunker 801 Project would far exceed the old quonset hut in both size and

capability. New Bunker 801 is actually a complex of protected enclosures, largely underground. Figure 4 is an artist's drawing of the new Bunker 801. With more than 1330 m² of floor space, the new bunker is much larger than any other at Site 300. More significantly, it is the only hydrodynamics test facility at Site 300 that is not dependent on some 20-year-old technology. Until the new Bunker 801 was built, we had been locked into the technology of the 1960s for much of our hydrodynamics experimentation.

The FXR is just one of the advanced diagnostics housed in the new bunker. In addition to the FXR, Bunker 801 now contains as its major diagnostics:

- An electrical-probe diagnostics system, including 48 raster oscilloscopes, that can accommodate up to 450 to 500 probe pins.
- A camera room with positions and controls for 10 high-speed cameras.
- A Fabry-Perot interferometer for measuring the velocity of explosion-driven surfaces.
- A laser-pulsed timing system for synchronizing the diagnostics with the detonation.

Fig. 4

Artist's drawing of the new Bunker 801. The large, two-story building in the background houses the FXR. To the right can be seen the firing table, which is viewed by ten camera ports. The camera room beneath the ports is visible in the cutaway. The computer room can be seen in the center cutaway. The original corrugated metal Bunker 301, now a machine shop, is in the foreground. The fill over the bunkers is gravel, which decouples the high-explosive shock wave from the bunker.

● Accommodations for electro-optic cameras and lasers to provide illumination for high-speed photography.

The Flash-X-Ray Machine

The FXR is a very powerful, pulsed radiographic device designed for diagnosing hydrodynamics tests.³ With the FXR, it is possible to radiograph the interior of imploding device components. A pulse of electrons, accelerated to an energy of 20 MeV in the linear induction accelerator, is the source of the x rays for the FXR; these electrons strike a thin tantalum target and spew out bremsstrahlung x rays, which are directed toward the test object (usually, the detonating high explosive). The x rays penetrate or are absorbed by the test object, depending upon the density and absorption cross section of its various parts. The x rays that get through strike a film behind the object, creating a radiographic image of the test object's interior (see Fig. 5). We use the radiographic

data to verify and normalize our computer calculations of device implosions.

In comparison with our older flash-x-ray machine, the FXR represents an improvement of five times the dose in one-third the pulse length. With its capability of delivering an x-ray dose of 500 rad at 1 m from the electron-beam target in a pulse of 70 ns duration, the FXR enables us to radiograph much denser objects; this, in turn, allows us to radiograph late in the implosion when the test device is densest. Also, since the FXR pulse is narrower, the radiographic exposure lasts for a shorter time, reducing motion blur and providing better resolution.

Pin Diagnostics

Pin diagnostics have been in use at Site 300 for many years and still play an essential role in our modern hydrodynamics experiments. In hydrodynamics tests of nuclear-device designs, detonation of the high explosive surrounding the device causes it to implode. We use pin diagnostics to measure the movement of the shock wave toward the center as a function of time. As the pin dome in Fig. 6 shows, the pins (electronic probes of thin wire) are arrayed in such a way that the tip of each will give a unique measurement. Each pin is connected to an electrical discharge circuit; when the surface to be measured strikes the end of a pin, the contact acts as a switch and discharges the energy stored in the pin circuit. The shape of the discharge wave form and the time it took place relative to the start of the experiment are recorded on a raster oscilloscope.

Because pin domes can contain up to several hundred pins, we must use a large number of oscilloscopes to record the pin signals (see Fig. 7). Since each oscilloscope traces the signal of several pins, we must also make each pin signal unique in order to identify it. The raster oscilloscope trace in Fig. 8 demonstrates how the pin signals are recorded. The raster trace consists of a series of consecutive

Fig. 5

An FXR radiographic image of a static test object. This image of a multilayered metal sphere illustrates the FXR's ability to define metal interfaces and material densities.

horizontal sweeps. The wave form of the signal from the pins can be varied as to polarity and discharge-time constant, making it possible for us to track many pins, positioned to report at different times, with the same oscilloscope.

High-Speed Cameras

The use of high-speed cameras in the manner developed at Site 300 is unique to LLNL. From the days of our first hydrodynamics tests, we have attempted to extract as much photographic information from each optically diagnosed test as possible. This implies using a number of rotating cameras simultaneously and synchronizing them to the experimental time history so that useful data can be obtained. In the 1950s, we used two cameras; in the 1960s and 1970s, we used up to five cameras (with three of them synchronized); more recently, we have used up to eight cameras on a single experiment. With the new Bunker 801 facility, we now can use ten cameras on a single shot.

The new camera room itself was designed to maximize our photographic capability (see Fig. 9). It is approximately 5 m high, allowing plenty of room for the complex telescopic optical systems. Optics ports in the 100-cm-thick roof of gravel-covered reinforced concrete allow the cameras to view the experiment. These ports can be as large as 50 cm in diameter, permitting the use of high-light-gathering optics or, if necessary, an array of cameras at a single port.

The camera room also was designed to accommodate future developments in optics technology. During the past year, we have installed equipment for Fabry-Perot interferometry and electro-optic photography that uses laser illumination. These technologies, developed in the last few years, greatly enhance our ability to obtain information on nuclear device hydrodynamics.

Interferometry

Fabry-Perot interferometry provides a continuous velocity record of a fast-moving surface in the imploding device.⁴ This is accomplished by reflecting a continuous laser beam from the moving surface. Light from the laser, on reflecting from a moving body, shifts its color (frequency) by an amount proportional to the object's

(a)

(b)

Fig. 6

Pin-dome diagnostic for use in hydrodynamics tests. (a) Photograph of an actual pin dome. (b) Diagram of how the pin dome is employed; this array of bristling metal probes is placed inside a hollow sphere. Each of the pins gives an electrical signal as the imploding sphere strikes it. The long and short pins are distributed so as not to interfere with each other during the test.

Fig. 7

Several of the 48 raster oscilloscopes used in Bunker 801 to trace pin-dome signals. Here, all the oscilloscopes except the one on the leftmost right are covered by mounted cameras that give us permanent images of the oscilloscope traces.

Fig. 8

Pin-dome signals registered on a raster oscilloscope. The oscilloscope makes a series of consecutive horizontal sweeps, each 2.5 millionths of a second long with timing blips every 0.5 millionth of a second. Each sweep starts at the left and traverses the screen toward the right, then takes a fraction of a microsecond to jump back to the left edge of the screen. Each pin signal is designed to be unique; the wave form of the pins can be varied as to polarity and discharge-time constant so they can be distinguished. Each signal has an echo signal of reversed polarity 0.6 μ s later to allow for any interruptions while the sweep is jumping back to the left. Here, the signals from ten pins are shown.

line-of-sight velocity. The time dependence of this color (or frequency) is then recorded with a Fabry-Perot interferometer. The output of the interferometer is a set of bright dots whose separation increases as the object's velocity increases. We measure this separation to determine the change in velocity, or acceleration, of the moving surface under scrutiny (see Fig. 10). The interference pattern is recorded by a high-speed electronic streak camera for later analysis. (See the article beginning on p. 1 for a more complete discussion of laser interferometry.)

Electro-Optic Photography

Electro-optic photography has enabled us to record images of higher resolution than is possible with conventional photography. The electronic shutter speeds of a few nanoseconds freeze fast-moving material, greatly reducing motion blur. In addition, electro-optic photography captures more detail because the large image plane (up to 75 mm in diameter) and the light-amplification characteristics of the electro-optic image tube permit magnification of the object being photographed.⁵

Computer Control System

Our ability to do complex experiments rests as much on the computer control of the diagnostic systems as it does on the sophistication of the diagnostics themselves. The control and monitoring of a hydrodynamics test is an extremely detailed procedure. Each test consists of a series of steps, each of which is briefer and more complex than the preceding one. The series culminates in the execution of a high-explosive-driven experiment and the simultaneous triggering of the diagnostics that record the event.

This sequence of steps can be divided into three accelerating phases that are reflected in the functions of the computerized fire-and-control system (see Fig. 11). The first phase is the human time scale. During this

time, a Modicon 584 slave computer gathers information on such things as blast doors being open or closed, the status of camera shutters, and the condition of personnel-safety monitors. The Modicon 584 also checks for such potential hazards as high voltage, x rays, and laser light. Even if other parts of the control

system fail or are modified, the Modicon will maintain a safe environment. Electronic and diagnostic devices are set, checked, and operated through a graphics touch panel on the computer that supervises the Modicon 584. The supervising computer displays the status of the experiment and gives the

Fig. 9

The camera room in Bunker 801. Telescopic lens systems mounted below the ten camera ports overhead allow us to focus on an experiment above ground. The cameras are mounted on moveable shafts so they can be raised or lowered for different degrees of image enlargement. The hoses on the cameras supply the helium gas that drives them; the cone-shaped shrouds above the cameras keep the optical path free of helium, which would distort the image.

Fig. 10

Diagram of the Fabry-Perot interferometer used in Bunker 801 to measure the velocity of an exploding surface. An argon-ion laser (operating at a wavelength of 514.5 nm) shines through a small hole in a mirror and through a lens that focuses the light onto the surface of a reflecting test object. The motion of the object shifts the frequency of the reflected laser light, some of which comes back through the same lens and is diverted by the mirror into the Fabry-Perot interferometer.

experimenters a convenient and centralized way to manipulate the system (see Fig. 12).

In this first phase of events, the high-explosive device is positioned, diagnostics and other equipment are set up and aligned, and the bunker is sealed for protection from the blast. After the diagnostic systems are checked for readiness, the supervising computer assists the Modicon 584 in confirming that the bunker is ready to proceed to the next phase.

The second phase of the experiment consists of setting the input parameters for electronic timing devices. These devices control the sequence of events during the experiment time immediately after detonation—less than 100 μ s. The camera controller starts the various rotating mirrors in the high-speed cameras turning in the correct sequence and sets the correct rate of acceleration to synchronize them. Valves and camera shutters are opened automatically. When the experimenter judges that all systems are ready, he pushes the fire button. This triggers the electrical signal that, in turn, actuates the test event.

The third phase, the last and fastest, is a series of short events

that lead up to and include the experimental blast. Because this portion of the experiment is too fast for the computers to react to, its sequence is controlled by preprogrammed, high-speed electronic devices. Electronic pulses flow through a network of cables, stopping for predetermined periods at electronic holding areas called pulse-delay units, before arriving precisely on schedule at their destinations. Experimental components and diagnostics are actuated like a string of dominoes.

In this last phase, the computerized fire-and-control system performs one of its most important functions, that of information gathering. Many experiments do not perform precisely as expected, and thus it is very important to determine where the experiment differed from predictions. The control-system computer collects and stores all the information it was given as well as all the commands and responses during the course of the experiment. With this data, we can make corrections as necessary in equipment and experimental procedures in subsequent tests.

Bunker 801's computer-based control system has made it possible to conduct increasingly complex

Fig. 11

Chronology of a typical hydrodynamics experiment at Site 300.

experiments with greater efficiency and success. Furthermore, the system is an open-ended framework that will permit expansion as new diagnostics are developed or as experimental needs appear.

Summary

An indispensable part of the LLNL's nuclear-design effort is the hydrodynamics tests we perform on the high explosives and nonnuclear device components that drive the nuclear detonation. We conduct these exacting and potentially hazardous tests at Site 300, a remote and uninhabited area about 25 km east of the main Laboratory site. Our hydrodynamics testing facilities at Site 300 have grown and improved greatly over the Laboratory's 30-year history. Our most recent construction project, the new hydrodynamics facility of Bunker 801, is the most advanced of its kind. With its novel computerized control system, flash-x-ray machine, high-speed cameras, and other state-of-the-art diagnostics, Bunker 801 enables us to obtain more complete and accurate information on device implosions than was ever before possible. □

Key Words: bunker—301, 351, 801; flash-x-ray machine (FXR); high explosive; interferometry—Fabry-Perot; photography—electro-optic, high-speed, streak; Site 300; testing—hydrodynamics.

Notes and References

1. For more information about our high-speed framing camera, see *Energy and Technology Review* (UCRL-52000-79-2), February 1979, p. 1.

Fig. 12

One corner of the control room in Bunker 801, showing the touch panels the bunker operator uses to conduct and monitor a hydrodynamics experiment.

2. For more information on the Laboratory's hydrodynamics diagnostics, see G. T. Carpluk, T. G. Innes, and S. R. Treast *LLNL Hydrodynamics Upgrade Program*, Lawrence Livermore National Laboratory, Rept. UCRL-52466 (1978).
3. The FXR is discussed at length in *EE Technical Review, Special Issue: Flash-X-Ray Accelerator*, Lawrence Livermore National Laboratory, Rept. UCRL-50025-83-2 (1983).
4. The Fabry-Perot velocimeter is described in *Energy and Technology Review* (UCRL-52000-79-3), March 1979, p. 17.
5. L. L. Shaw *et al.*, "A High-Speed, Eight-Frame Electro-Optic Camera with Multipulsed Ruby Laser Illuminator," presented at the 16th International Congress on High Speed Photography and Photonics, Strasbourg, France (1984); published in *SPIE* 491, 267 (1985).

Neutron Spectrometry for Dosimetry

We have developed an improved neutron spectrometry system, covering the energy range from 0.025 eV (thermal) to 20 MeV, for use in the Laboratory's radiation monitoring program, and we are testing a new material for the neutron dosimeter worn with employees' identification badges.

For further information contact
John H. Thorngate (415) 422-5216.

At LLNL, as at any laboratory working with sources of radiation, all employees wear several radiation detectors ("dosimeters") as part of their identification badges. Systems for measuring the radiation dose of x rays, gamma rays, and charged particles are well established and relatively straightforward, but measuring neutron dose accurately is much more difficult. Basically, this is because neutrons, unlike charged particles and x or gamma rays, do not affect detectors directly but only by secondary effects.

Another complication is that the response of many detectors is actually related to the number of neutrons seen, not to the neutron dose. Moreover, what must be determined is not the neutron dose itself but the "neutron dose equivalent," a measure of the effect of the neutrons on the human body.

The neutron dose equivalent is determined by multiplying the dose

(energy absorbed per gram of tissue as a result of the neutron irradiation) by a quality factor (Q) related to the biological effect of neutrons as compared with x rays. Suitable detectors for measuring the absorbed energy have existed for some time, but converting to neutron dose equivalent is complicated by the variation of Q (from 2 to 11) with neutron energy. Thus, to make an accurate conversion to dose equivalent, we must know the energy distribution (the spectrum) of the neutrons being detected.

Although there are approximate methods for estimating the neutron dose equivalent that are useful for many purposes,¹ a complete dosimetry program requires more accurate data. At LLNL, we must evaluate the spectrum for each neutron source near which employees will be working. These include the rotating target neutron source (RTNS), various fusion experiments using laser pulses or magnetic fields to compress plasma,

heavy elements that produce neutrons by spontaneous fission, and high-energy x rays that generate photoneutrons. By applying the spectrum for the particular source involved, the data recorded by an employee's dosimeter or measured by a simpler dosimeter can be converted accurately to neutron dose equivalent. We have been developing ever-more advanced neutron spectrometers for dosimetry applications for more than a decade.²

Neutron Spectrometry of Continuous Sources

Because neutrons are electrically neutral, they do not produce ionization directly and hence have no direct effect on conventional radiation detectors. Instead, they must be detected by the secondary charged particles they produce.

Low-Energy Neutrons

The multisphere system for measuring neutron energies from 0.025 eV to 5 MeV (Fig. 1) uses a small lithium iodide (LiI) scintillator enriched in lithium-6 to detect neutrons via a thermal-neutron reaction that produces an alpha particle and a triton from the lithium. This detector is used either bare, enclosed in cadmium, or at the center of a moderator sphere of polyethylene (ranging in diameter from 76 to 305 mm) to vary the sensitivity of the detector to neutrons of different energies.³

A moderator such as polyethylene decreases the energy of the neutrons passing through it by providing abundant opportunities for neutron reactions. Although neutron interactions are possible with the carbon in polyethylene, most of the moderation for neutrons below 10 MeV results from interactions with the hydrogen in successive "billiard-ball" collisions. On the average, a neutron loses half of its energy with each such collision, and some of the neutrons eventually reach thermal energy. These thermal neutrons may be captured by hydrogen to form

deuterium or they may interact with the lithium in the detector and be counted.

The number of thermal neutrons detected depends on the thickness of the moderator and on the energy of the incident neutrons. The original neutron spectrum can be inferred by calculating the spectrum that would be required to produce the thermal neutron counts measured for the different detector configurations.

Because thermal neutron yield changes slowly with neutron energy or moderator sphere size, the resolution of this technique is limited. Furthermore, the sensitivity of this type of spectrometer decreases rapidly for neutron energies above 10 MeV because neutrons at these higher energies are less likely to collide with hydrogen in the desired reactions and are more likely to interact with carbon instead.

In operation with this multisphere spectrometer, a multichannel analyzer

Fig. 1

The multisphere system for measuring the low-energy portion of the neutron spectrum (0.025 eV to 5 MeV) uses a thermal-neutron detector (LiI) inside a sphere of neutron-moderating material. The detector is shown in the center foreground of this photograph, with the polyethylene moderator spheres displayed in the background. The thinner the layer of moderator around the detector, the more high-energy neutrons fail to be thermalized. We use spheres of various sizes to vary the sensitivity of the detector to neutrons of different energies. By combining readings taken with different moderators we can reconstruct the neutron energy spectrum.

records the pulse-height distribution of the amplified signals from the LiI scintillator. The alpha particle and triton produced in the interaction between a thermal neutron and a lithium-6 atom have a combined energy of about 4.8 MeV. Hence, the light produced by these charged particles has a consistent value, which the pulse-height analyzer records as a peak rising from a background continuum. The net detector response is obtained by fitting a logarithmic function that approximates the continuum under the thermal peak and subtracting that area from the total (Fig. 2).

Data from this type of spectrometer are analyzed using a computer program such as the LOUHI code, which converges to a solution that is independent of the initial estimate (although a good estimate shortens the running time significantly).⁴ Because LOUHI requires a large computer (a Cray or CDC 7600), data cannot be reduced in the field and it takes considerable time to obtain the results. A major advantage of LOUHI is its ability to make use of whatever part of the spectrum is already known and concentrate on finding solutions to the unknown portion of the

spectrum. This feature significantly improves the efficiency of the data reduction and the usefulness of the data.

To obtain approximate spectra quickly in the field, we have been using the SPUNF code.⁵ SPUNF will run on a desk-top computer, but its results depend heavily on the accuracy of the initial estimate. We have also developed our own program, MINKY, which runs on a desk-top computer (an LSI-11), allows the use of *a priori* information (like LOUHI), and produces results more quickly.

High-Energy Neutrons

To measure neutrons of higher energies (0.5 to 20 MeV), we use a liquid scintillator to measure directly the energy of protons scattered by neutrons striking hydrogen atoms in the scintillator material. For a neutron with energy below the relativistic region, there is an equal probability that the recoil proton produced in collision with a hydrogen atom will have any energy from zero up to the initial energy of the neutron. Thus, a beam of monoenergetic neutrons (all with the same energy) directed at a radiator would produce a rectangular spectrum of recoil protons (equal numbers at all energies up to a maximum). The usual beam of neutrons, however, contains neutrons of various energies and thus makes a nonrectangular proton spectrum. A computer can "unfold" such a distribution to deduce the energy distribution of the neutrons that produced it.⁶

Most of our effort with the liquid scintillator has been in optimizing the detector and the data-reduction procedures. To optimize the detector, we have considered the shape of the detector, use of a light pipe between the scintillator and the photomultiplier tube, and the light output of the detector as a function of electron and neutron energy. Efforts to improve data reduction have concentrated on correcting for multiple scattering within the detector and on

Fig. 2

Typical pulse-height response of the multisphere system, illustrating the exponential fit used to subtract background.

computational details of the actual data-reduction program.

We use the NE-213 detector, in which the sensor material is a mixture of organic compounds consisting essentially of only hydrogen and carbon atoms (primarily paraxylene with dissolved naphthalene, scintillator material, and a material that shifts the frequency of the emitted light for more efficient detection by the photomultiplier). The naphthalene makes the light pulses produced by gamma-ray interactions different from those made by neutrons, enabling a discriminator circuit to tell them apart.⁷

The need to accommodate the thermal expansion of the scintillator liquid complicates the design of the detector cell. Each detector cell must have a small expansion chamber, mounted on the outside, partly filled with nitrogen. The cell must be oriented so that the nitrogen bubble cannot enter the sensitive volume and interfere with the transmission of light from the scintillator to the photomultiplier tube. Hence, in addition to a wide range of detector sizes (from 13 mm in diameter by 13 mm high to 114 mm in diameter by 51 mm high), we also have some cells designed to look upward, downward, or sideways. This variety of configurations allows us to use a detector optimized for the energy range, intensity, and physical arrangement of the source in question.

Including a carefully designed light pipe between the scintillator and the photomultiplier tube can improve the resolution of the detector system.⁸ For example, a given amount of energy may produce larger pulses if it is deposited in the cell near the photocathode than if it is deposited at the opposite end of the cell. We determined the magnitude of this effect, for a 51-mm detector, by measuring the pulse-height distributions produced by a collimated gamma-ray source placed at the bottom, middle, or top of the cell (Fig. 3). Because the scintillator liquid has a low atomic number, gamma

rays interact predominantly via Compton scattering, producing electron distributions in the detector with a peak at a well-defined upper edge; with no light pipe in place, the locations of the Compton edges for the three irradiation positions differed considerably, with the largest pulse-height values occurring for bottom irradiation.

We also considered the effect of the length of the pipe. When the entire volume of the cell was irradiated, the longer the light pipe the better the detector resolution. At the same time, however, the additional length decreased the total light reaching the photocathode, thereby reducing the detector sensitivity. We therefore compromised on a 25-mm-long light pipe in our new detectors, which we made of optical glass instead of plastic to reduce the amount of hydrogen present and thereby minimize the scattering of stray neutrons into the detector.

Total Neutron Spectrometry System

The final proof of our work in developing a neutron spectrometry

Fig. 3

Experimental set-up to determine how resolution is affected by the length of the light pipe between the liquid scintillator and the photomultiplier tube.

system for dosimetry applications is the quality of the data produced. We have measured several spectra using the combined multisphere spectrometer and liquid scintillator system. One of these was on SHEBA, an aqueous, homogeneous critical assembly at the Los Alamos National Laboratory, to provide dosimetry information relevant to the design of gas centrifuges for uranium enrichment. We took spectra at 5 m from the assembly, both bare and with iron or concrete shielding (Fig. 4). Another measurement was made 14 m from the bare assembly.

Another critical assembly at the Los Alamos National Laboratory is COMET, a mock-up of the Hiroshima bomb. Much of the information on which current radiation-protection standards are based comes from

studies of the survivors of Hiroshima and Nagasaki, and knowledge of the source is important for normalizing the doses assigned to each of the survivors. We made measurements with the combined spectrometer system at two points on COMET, both 2 m from the center of the assembly. One was at 90 deg to the cylindrical axis of the assembly and the other was on the axis at the top of the assembly, corresponding to the nose of the bomb (Fig. 5). Additional scintillator measurements were made at 45 deg, and the multispheres were used at other distances from the assembly. Our measurements are in excellent agreement with existing calculations.

Here at LLNL, we have used our spectrometry system to measure the neutron spectra produced by the californium and plutonium-beryllium sources at the Hazards Control Standards and Calibration Facility. This involved measurements of the spectra made with the source bare and in each of nine moderator spheres. Each spectrum was measured three times at both gains, and the data were combined (Fig. 6).⁹ When corrections were made for source decay, the agreement between consecutive measurements of the same spectrum was within the allowable uncertainties. We have also taken multisphere measurements of these spectra and are making the combined results available to users of the facility as an aid for planning exposures.

General improvements planned for the neutron spectrometry system are to make it more portable and to provide complete data reduction in the field. We are redesigning the pulse-shape discrimination unit to reduce its power requirements and thereby to improve the portability of the scintillation detector. We may also be able to decrease the size and width of the multisphere system by using nested spheres. We are also developing a more efficient data-reduction program capable of running on a battery-powered computer to

Fig. 4

Combined results from the liquid scintillator and multisphere systems, obtained 5 m from the SHEBA critical assembly at the Los Alamos National Laboratory. Measurements were made both with the assembly unshielded and with shields of iron (133 mm thick) or concrete (400 mm thick).

make possible complete data reduction in the field.

Detector for Pulsed Neutron Sources

Many neutron detectors, including those we use for spectrometry, were designed for continuous neutron sources and do not respond properly when the source is pulsed. This is particularly true if the detector signals and the source pulses have comparable lengths. The fusion experiments at LLNL are examples of pulsed neutron sources that common dose-equivalent survey meters may not measure correctly.

Several years ago, we were asked to devise an instrument to measure the neutron dose equivalents outside the shielding surrounding the Mirror Fusion Test Facility (MFTF). Because the sensitivity of the 250-mm-diameter multisphere detector varies as a function of neutron energy in much the same way as the dose equivalent, we used this concept with a fast glass scintillator to detect the thermal neutrons. By coupling the detector with fast electronics to count the detected pulses, we produced an instrument capable of making the necessary measurements and giving

prompt results. In some cases, this instrument can be used to measure the even shorter pulses produced by laser fusion experiments.

In addition, this detector is more sensitive than most dose-equivalent meters, making it suitable for fence-line measurements at LLNL and at large distances from the COMET critical-assembly facility at Los Alamos. It can measure intense neutron fields that are beyond the capabilities of most other dosimeters. This instrument has also been used to measure air-scattered neutrons from photoneutron reactions at a pulsed x-ray facility at Site 300 as well as neutrons produced by a pulsed neutron generator at the Nevada Test Site.

New Sensor Material for Personnel Dosimeters

Accuracy, wide energy range, stability, ease of evaluation, and low cost are the desirable characteristics in sensor materials used for detecting neutrons in personnel dosimeters such as those worn as part of the badges of LLNL employees. One widely used sensor material for such dosimeters is a nuclear emulsion (film) similar to the films used for cosmic-ray

Fig. 5

Combined results from the liquid scintillator and multisphere systems, obtained 2 m from the midpoint of the Los Alamos COMET assembly, a mock-up of the Hiroshima bomb. (a) Detector on axis, at the nose of the bomb assembly. (b) Detector at side of assembly, opposite the midpoint (90 deg from the assembly axis).

measurements. However, film has little response to slow neutrons (energies below 0.5 MeV), it is difficult to read, and under certain conditions the tracks left by the neutron reactions may fade.

The albedo dosimeters previously used at the Laboratory employed a thermoluminescent dosimeter material to detect neutrons that have been slowed down to thermal energies (0.025 eV) and reflected by the body. Although this dosimeter's measurements are stable and easy to evaluate, its response falls off rapidly above 100 keV. Therefore, to obtain accurate dose evaluations, we must acquire additional information about the neutron field in which the albedo dosimeters were worn.

Within the last decade, a new class of neutron detector materials has evolved in which ions recoiling from neutron impacts produce damage that

can be made visible later by selective etching. One of these materials, a commercially available plastic known as CR-39, has shown particularly good sensitivity to neutrons, and we are developing it into a successor to our albedo dosimeter material.

To investigate CR-39's adaptability to large-scale processing, we modified a commercial etch cell to accommodate about 100 detectors a day. This rate is of practical interest because we will have to evaluate thousands of these dosimeters each month if they are adopted as the standard at LLNL. We found that etching in such quantities is quite feasible.

The range of energies that can be measured with CR-39 depends on the etching technique used. We use electrochemical etching, in which the foil of exposed CR-39 plastic is immersed in potassium hydroxide and subjected to an alternating electric field. An etch for 5 hours in an electric field of 3.0 kV at 60 Hz, followed by 28 minutes of 2.7 kV at 2 kHz, and then 15 minutes with no applied potential, is near optimum, although small changes may be required from one batch of plastic to another. Completing the etch at the higher frequency produces tracks of uniform size (Fig. 7), which can then be counted semiautomatically with a commercial biological colony counter.

The neutron dose equivalent is proportional to the number of tracks per unit area. The response presently obtained with CR-39 is reasonably uniform for neutron energies from 100 keV to 4 MeV, which is better than that of other materials used for personnel neutron dosimeters. Sensitivity depends on the background count obtained from unexposed sensor material, which is affected by the manufacture of the plastic and its subsequent handling. Our newest plastic is unusually good in this respect, with a background that is low enough to permit dose-equivalent measurements as low as 25 mrem. The dosimeter has a linear response up to 400 mrem, but it

Fig. 6

Spectrum obtained with the liquid scintillator 2 m from the large $^{238}\text{PuBe}$ source at the Standards and Calibration Facility of LLNL's Hazards Control Department, with no moderator present. A calculation of the $^{241}\text{AmBe}$ spectrum (Ref. 9) is shown for comparison. The alpha particles from ^{241}Am have essentially the same energies as those from ^{238}Pu , so the spectra should be similar; no effort has been made to smooth the calculations to simulate the detector resolution. The peak near 6 MeV in the calculations may represent an artifact of the simplifications used. Excluding this point, the agreement is quite good.

remains usable though nonlinear up to about 10 rem.

Summary

We are continuously improving and modernizing our neutron dosimetry program for personnel protection at LLNL. As part of this effort, we have developed an improved neutron spectrometry system for measuring the various steady-state neutron sources at the Laboratory. We also have designed a detector for measuring pulsed neutron sources, such as the large magnetic fusion devices. In addition, we are evaluating a new plastic material that is considerably more sensitive to neutrons than previously used plastics and will soon be used for the dosimeters worn by every LLNL employee. □

Key Words: computer code—LOUHI, MINKY, SPUNF; neutron—detection, dose equivalent, thermal; neutron source—continuous, pulsed; nuclear reaction; personnel dosimetry.

Notes and References

1. Ö. Andersson and J. Braun, *A Neutron Rem Counter*, A. B. Atomenergi, Rept. AE-132 (1964).
2. D. R. Slaughter, D. W. Rueppel, and D. A. Fuess, *Neutron Spectrum Measurements for Radiation Protection Purposes*, Lawrence Livermore National Laboratory, Rept. UCRL-52415 (1978).
3. R. L. Bramblett, R. I. Ewing, and T. W. Bonner, "A New Type of Neutron Spectrometer," *Nucl. Instrum. Meth.* **9**, 1-12 (1960).
4. J. T. Routti and J. V. Sanberg, *General Purpose Unfolding Program LOUHI 78 with Linear and Nonlinear Regularizations*, Helsinki University of Technology, Rept. TKK-F-A359 (1978).
5. F. Hajnal, *Environmental Measurements Laboratory*, New York, N.Y., private communication (1982).
6. E. Baldinger, P. Huber, and H. Staub, "Ausbeute und Energiebestimmung der Neutronen aus der Kernreaktion $D + D = {}^3\text{He} + n$," *Helv. Phys. Acta* **11**, 245-268 (1938).
7. F. T. Kuchnir and F. J. Lynch, "Time Dependence of Scintillations and the Effect on Pulse-Shape Discrimination," *IEEE Trans. Nucl. Sci.* **NS-15**(3), 107-113 (1969).
8. H. Schölerman and H. Klein, "Optimizing the Energy Resolution of Scintillation Counters at High Energies," *Nucl. Instrum. Meth.* **168**, 25-31 (1980).
9. F. de Guarrini and R. Maladora, "Two Different Technique Measurements of the Neutron Spectrum of an Am-Be Source," *Nucl. Instrum. Meth.* **92**, 277-284 (1971).

Fig. 7

Photomicrograph of an etched CR-39 plastic foil (above, $9.6 \times$) and enlargement (left, $85 \times$), showing the uniformity of the etched recoil tracks. Each pit in this picture represents the point where a proton, knocked ahead by being hit with a neutron, went through the plastic breaking molecular bonds. Because these pits are so uniform, we can use semiautomatic machinery to count them and determine the neutron dose to which the film has been exposed.

New Laser Amplifier Improves Laser Doppler Interferometry

A specialized laser amplifier has been developed for use with high-speed velocity-measuring systems that enables us to obtain more detailed measurements of explosion-driven targets over longer periods of time. With the amplifier, sufficient light is produced to permit examination of two or more spots on a target. Experimenters can now follow the motion of one area of the target relative to another; they can also choose different recording parameters for the two spots and thereby correlate details of target motion. The amplifier also makes it possible to obtain longer records. As the target disintegrates, it reflects less light, and at some point the measuring system fails; however, using the amplifier to send more light to the target extends the duration of the usable returned signal.

Contact: Lloyd L. Steinmetz (415) 422-6897

The Bunker 801 Project

The nuclear-device design effort at LLNL relies heavily on accurate data about device implosions. Since the implosions are driven by chemical high explosives, it is important that we understand the hydrodynamic behavior of these explosives and their effects on all the other components of the device design (these components include nonfissile materials we substitute for fissile ones in hydrodynamics tests). We have recently completed an advanced hydrodynamics test facility at Site 300. This facility, Bunker 801, features a unique computer-control system, a pulsed flash-x-ray machine, synchronized high-speed cameras, electrical-probe diagnostics, interferometry radiography capabilities, and other diagnostic equipment that enable us to obtain hydrodynamics data of a quality and completeness far superior to that collected in the past.

Contact: Larry L. Shaw (415) 422-4561.

Neutron Spectrometry for Dosimetry

As part of our continuing effort to improve and modernize LLNL's neutron dosimetry program for personnel protection, we have developed an improved neutron spectrometry system for measuring steady-state neutron sources in the energy range from 0.025 eV (thermal) to 20 MeV as well as a detector for measuring pulsed neutron sources. In addition, a new, more sensitive material, CR-39 plastic, is being evaluated for the personnel-monitoring neutron dosimeter worn with employees' identification badges.

Contact: John H. Thorngate (415) 422-5216.

Recent Titles

Articles published in recent issues of *Energy and Technology Review* are grouped below according to their chief sponsors, the Assistant Secretaries of the U.S. Department of Energy. Research funded by other federal agencies is listed under Work for Others.

ENERGY RESEARCH

Preparing for the Cray-2 (Brief—September 1985)
AVLIS Chosen as the Uranium-Enrichment Technology of the Future (Brief—August 1985)
Monoclonal Antibodies (August 1985)

DEFENSE PROGRAMS

Office of Inertial Fusion

A New Era in Laser Technology (November 1985)
X-Ray Lasing: The Diagnostics (November 1985)
X-Ray Lasing: The Experiments (November 1985)
X-Ray Lasing: Future Directions (November 1985)
X-Ray Lasing: The Novette Laser (November 1985)
X-Ray Lasing: The Targets (November 1985)
X-Ray Lasing: Theory (November 1985)
Optical Coatings by the Sol-Gel Process (October 1985)
Rapid Growth of KDP Crystals (August 1985)

Office of Military Application

The Bunker 801 Project (February 1986)
New Laser Amplifier Improves Laser Doppler Interferometry (February 1986)
Integrating Computer-Aided Engineering at LLNL (January 1986)
Railgun Research and Development (September 1985)
Solid Modeling Research (September 1985)
Radioactive Ion Beams: Hot Stellar Reactions in the Laboratory (August 1985)

LABORATORY REVIEWS

Neutron Spectrometry for Dosimetry (February 1986)
Windpower: Assessing the Potential (September 1985)
State of the Laboratory Issue (July 1985)

SCIENCE REVIEWS

Modern Computers, Ancient Skies (October 1985)

WORK FOR OTHERS

Department of Defense

Precision Engineering at LLNL (January 1986)
Generating Radar Images of the Earth with the S-1 Computer (December 1985)
Laser Pantography (December 1985)
Wafer-Scale Integration (December 1985)
Combustion Instabilities in Liquid Propellants (October 1985)

National Institute of Environmental Health Science

Mutagens in Cooked Food (January 1986)