A STUDY OF THE EFFECTS OF PHOTON STATISTICS ON
THOMSON SCATTERING DATA

By

G.W. Hart, F.M. Levinton, and D.H. McNeill

DECEMBER 1985

PLASMA
PHYSICS
LABORATORY

PRINCETON UNIVERSITY
PRINCETON, NEW JERSEY

PREPARED FOR THE U.S. DEPARTMENT OF ENERGY,
UNDER CONTRACT DE-AC02-76-CDO-3073.

EXTERNAL DISTRIBUTION IN ADDITION TO UC-20

Plasma Res Lab, Austro Nat'l Univ, AUSTRALIA
 Dr. Frank J. Paoloni, Univ of Wollongong, AUSTRALIA
 Prof. I.R. Jones, Flinders Univ., AUSTRALIA
 Prof. M.H. Brennan, Univ Sydney, AUSTRALIA
 Prof. F. Cap, Inst Theo Phys, AUSTRIA
 Prof. Frank Verheest, Inst theoretische, BELGIUM
 Dr. D. Ralumbo, Dg XII Fusion Prog, BELGIUM
 Ecole Royale Militaire, Lab de Phys Plasmas, BELGIUM
 Dr. P.H. Sakanaka, Univ Estadual, BRAZIL
 Dr. C.R. James, Univ of Alberta, CANADA
 Prof. J. Teichmann, Univ of Montreal, CANADA
 Dr. H.M. Skarsgard, Univ of Saskatchewan, CANADA
 Prof. S.R. Greenivasan, University of Calgary, CANADA
 Prof. Tudor W. Johnston, INRS-Energie, CANADA
 Dr. Hannes Barnard, Univ British Columbia, CANADA
 Dr. M.P. Badynski, MPB Technologies, Inc., CANADA
 Chalk River, Nucl Lab, CANADA
 Zhengwu Li, SW Inst Physics, CHINA
 Library, Tsing Hua University, CHINA
 Librarian, Institute of Physics, CHINA
 Inst Plasma Phys, Academia Sinica, CHINA
 Dr. Peter Lukac, Komenského Univ, CZECHOSLOVAKIA
 The Librarian, Culham Laboratory, ENGLAND
 Prof. Schatzman, Observatoire de Nice, FRANCE
 J. Redet, CEN-BP6, FRANCE
 AM Dupas Library, AM Dupas Library, FRANCE
 Dr. Tim Mual, Academy Bibliographic, HONG KONG
 Preprint Library, Cent Res Inst Phys, HUNGARY
 Dr. R.K. Chhajlani, Vikram Univ. INDIA
 Dr. B. Dasgupta, Saha Inst, INDIA
 Dr. P. Kaw, Physical Research Lab, INDIA
 Dr. Phillip Rosenau, Israel Inst Tech, ISRAEL
 Prof. S. Capeman, Tel Aviv University, ISRAEL
 Prof. G. Postagni, Univ Di Padova, ITALY
 Librarian, Int'l Ctr Theo Phys, ITALY
 Miss Clelia De Palo, Assoc EURATOM-ENEA, ITALY
 Biblioteca, del CNR EURATOM, ITALY
 Dr. H. Yamato, Toshiba Res & Dev, JAPAN
 Direc. Dept. Lg. Tokamak Dev. JAERI, JAPAN
 Prof. Nobuyuki Inoue, University of Tokyo, JAPAN
 Research Info Center, Nagoya University, JAPAN
 Prof. Kyoji Nishikawa, Univ of Hiroshima, JAPAN
 Prof. Sigeru Mori, JAERI, JAPAN
 Prof. S. Tanaka, Kyoto University, JAPAN
 Library, Kyoto University, JAPAN
 Prof. Ichiro Kawakami, Nihon Univ, JAPAN
 Prof. Satoshi Itoh, Kyushu University, JAPAN
 Dr. D.I. Choi, Adv. Inst Sci & Tech, K. REA
 Tech Info Division, KAERI, KOREA
 Bibliothek, Fom-Inst Voor Plasma, NETHERLANDS
 Prof. B.S. Wiley, University of Waikato, NEW ZEALAND
 Prof. J.A.C. Cabral, Inst Superior Tecn, PORTUGAL
 Dr. Octavian Petrus, ALI OJZA University, ROMANIA
 Prof. M.A. Hellberg, University of Natal, SO AFRICA
 Dr. Johan de Villiers, Plasma Physics, Nuoor, SO AFRICA
 Fusion Div. Library, JEN, SPAIN
 Prof. Hans Wilhelmson, Chalmers Univ Tech, SWEDEN
 Dr. Lennart Stenflo, University of UMEA, SWEDEN
 Library, Royal Inst Tech, SWEDEN
 Centre de Recherches, Ecole Polytech Fed, SWITZERLAND
 Dr. V.T. Tolok, Kharkov Phys Tech Ins, USSR
 Dr. D.D. Ryutov, Siberian Acad Sci, USSR
 Dr. G.A. Elisaeiev, Kurchatov Institute, USSR
 Dr. V.A. Glukhikh, Inst Electro-Physical, USSR
 Institute Gen. Physics, USSR
 Prof. T.J.M. Boyd, Univ College N Wales, WALES
 Dr. K. Schindler, Ruhr Universitat, W. GERMANY
 Nuclear Res Estab, Jülich Ltd, W. GERMANY
 Librarian, Max-Planck Institut, W. GERMANY
 Bibliothek, Inst Plasmaforschung, W. GERMANY
 Prof. R.K. Janev, Inst Phys, YUGOSLAVIA

MASTER

A STUDY OF THE EFFECTS OF PHOTON STATISTICS ON
THOMSON SCATTERING DATA

G.W. Hart,^(a) F.M. Levinton,^(b) and D.H. McNeill

Princeton Plasma Physics Laboratory

P. O. Box 451, Princeton, N.J. 08544

Abstract

A computer code has been developed which simulates the counting statistics of a Thomson scattering measurement. The scattered and background signals in each of the wavelength channels are assumed to obey Poisson statistics, and the spectral data are fitted to a Gaussian curve using a nonlinear least-squares fitting algorithm. This method should be applicable to Thomson scattering measurements in which the signal-to-noise ratio is low due to low signal or high background. Thomson scattering data from the S-1 Spheromak have been compared to this simulation, and they have been found to be in good agreement. This code has proved to be useful in assessing the effects of counting statistics relative to shot-to-shot variability in producing the observed spread in the data. It was also useful for designing improvements for the S-1 Thomson scattering system, and this method would be applicable to any measurement affected by counting statistics.

1 Introduction

The effect of photon statistics on a Thomson scattering measurement is difficult to quantify. While the statistics of a single channel, i.e., Poisson statistics, are well known, the interaction of multiple channels of data with the least-squares fitting used to evaluate the temperature and density is not easily calculated. This makes it difficult to assess whether the variations in the measurement of a plasma temperature are due to statistical fluctuations or to real shot-to-shot changes. This paper discusses a method of making this assessment.

2 Experiment

The S-1 Spheromak Thomson scattering system is a single-point system that can be moved to different points in the plasma on a shot-to-shot basis. The system has six wavelength channels spread over the range from 35 to 314 angstroms from the ruby laser line. The output from the photomultipliers on each of these channels is gated into a LeCroy 2249A current integrator for approximately 80 nsec during the laser pulse. A second set of integrators is gated on for approximately 160 nsec immediately after the laser pulse to obtain a background subtraction signal. After the signals are read and digitized, the result is then computer-readable [1].

Since the integrators do not actually count photons, but rather accumulate charge, it is necessary to calibrate the correspondence between the

signal level on the integrators and the number of photoelectrons produced in the photomultipliers. We used two methods to obtain this number, both of which produced the same result. The first was to measure the height and width of single photoelectron pulses coming from the photomultipliers and use the known charge sensitivity of the integrators to calculate the number of photoelectrons per digitizer count. The second method used the fact that a repeated measurement of a photon signal will be distributed according to Poisson statistics. A constant signal from a lamp was put into the polychrometer, and the signal level from the photomultipliers was repeatedly sampled. The mean and standard deviation of those measurements was then calculated. From the properties of the Poisson distribution, the standard deviation of the measurement is proportional to the square root of the number of photoelectrons [2]. The constant of proportionality is the same as that relating the signal to the number of photoelectrons. Therefore, using the above data we can calculate that constant. If S is the signal level, σ is the standard deviation of the signal, C is the constant of proportionality between the signal and the number of photoelectrons, and P is the number of photoelectrons, then

$$S = CP \quad (1)$$

and

$$\sigma = C\sqrt{P}. \quad (2)$$

This gives the result that

$$C = \sigma^2/S. \quad (3)$$

The gain of the photomultipliers was adjusted until one count on the integrator approximately equaled one photoelectron on the photomultiplier.

We shall consider the case of S-1 operating with Figure-8 coils [3]. At the operating parameters of 40 eV, electron density of $3 \times 10^{13} \text{ cm}^{-3}$, and a laser energy of 13.5 J, the channel closest to the laser line ($\Delta\lambda = 46 \text{ \AA}$) has an average signal level of 50 photoelectrons, with a background light level of 63 photoelectrons during the laser pulse. The large background light level appears to be caused mostly by molecular hydrogen emission from the thick blanket of neutral gas surrounding the plasma [1].

Because of the low signal level and the large background light, there was a large spread in the measured temperatures from this system, and it became essential to differentiate the scatter in the measurement caused by photoelectron statistics from that caused by the shot-to-shot variability of the experiment. For that purpose we developed the code described here.

3 Computer Simulation

The code simulates the effect of the interaction of the Poisson statistics of the photoelectron counting process in each signal channel with the non-linear fitting routine used to calculate the temperature and density from the data channels. If a given channel is expected to have a given number of photoelectron counts, μ , then the Poisson statistics of the measurement say that the probability of measuring a number n photoelectrons is given

by the expression

$$P(n, \mu) = \frac{\mu^n e^{-\mu}}{n!}. \quad (4)$$

The value which goes into the fitting routine is a simulated signal plus background (as with the first gate in the real experiment), from which a background signal (simulating the second gate) will be subtracted.

A set plasma temperature is put into the code, along with the expected mean signal and background levels for the central channel. For a given plasma temperature, specifying the signal in the central channel is equivalent to specifying the plasma density and optical throughput. The code calculates the expected signal for each channel from a Gaussian distribution of the signal versus wavelength, using the given temperature to calculate the width and the central signal for the height. The value for the signal plus background in a given channel is randomly chosen from a Poisson distribution with a mean equal to the expected signal plus the background level. The background for that channel is also chosen from a Poisson distribution, using twice the input background level as the mean (because in the experiment the background gate is twice as wide as the signal gate) and dividing the resultant number by two before subtraction. For example, if the signal expected in a channel is 15 photoelectrons and the background level is 50 photoelectrons, the signal plus background would be chosen from a Poisson distribution with a mean of 65. The background would be chosen from a distribution with a mean of 100, with the result divided by two before subtraction from the first value.

The above process is carried out for each signal channel, making proper allowance for its relative spectral width, and the results are put into a least-squares fitting routine [4]. The routine fits the signals to a function of the form

$$y_i = Ae^{-Bx_i^2}, \quad (5)$$

where y_i is the simulated signal and $x_i = \Delta\lambda_i$. At the temperatures we were measuring, the relativistic corrections to the above expression were insignificant. It is important to use a nonlinear least-squares fitting routine of this sort to prevent the noise in the outer channels from skewing the measurement [5].

4 Results and Discussion

The results of this comparison are shown in Fig. 1. The jagged curve shows an experimental histogram of 121 shots taken on S-1 with a plasma current of 270 kA and a fill pressure of 3 mTorr. This set of data was chosen for our analysis because the plasma was reproducible and was collected over a short period of time to minimize changing conditions. The smooth curve is the result of this code for a temperature of 40 eV, with the same mean signal and background levels as in the experimental data (50 signal and 63 background photoelectrons). *The temperature distribution from the experimental data is in good agreement with the results of the simulation. This implies that in this case the variation in the measurements can be totally explained by the statistics of the measurement, without any contributions from shot-to-shot*

variations in the actual temperature.

There are many other uses for this code beside the one mentioned above. Using the code we were able to simulate the response of the system to different conditions such as changes in density and temperature. The most important function of the code has been to evaluate the various possible design changes, including spectral channel positions, photomultiplier quantum efficiency, and gate width and to choose the ones that would give the best improvement in performance without having to actually build the systems. This approach is not limited to Thomson scattering analysis, but could be applied to any system affected by counting statistics.

5 Acknowledgment

This work was supported by U. S. Department of Energy contract No. DE-AC02-76-CHO3073.

^(a) Present Address: Dept. of Physics and Astronomy, Brigham Young University, Provo, Utah 84602.

^(b) On assignment from JAYCOR, San Diego, Ca. 92138.

References

- [1] G. W. Hart and D. H. McNeill, in *Proceedings of the Sixth U.S. Symposium on Compact Toroid Research and the Fifth U.S.-Japan Joint Sym-*

- posium on Compact Toroid Research* (Princeton, N.J., 1984), p. 116.
- [2] *Photomultiplier Manual*, RCA Electronic Component⁺, Harrison, N.J., 1970, Technical Services Publication number PT-61.
 - [3] M. Yamada, R. Ellis, Jr., H. P. Furth, G. Hart, A. Janos, S. Jardin, F. Levinton, D. Meyerhofer, M. Mimura, C. H. Nam, S. Paul, A. Sperduti, S. von Goeler, F. Wysocki, and P. Young, in *Proceedings of the Tenth International Conference of Plasma Physics and Controlled Nuclear Fusion Research 1984* (IAEA, Vienna, 1985), Vol. II, p. 535.
 - [4] P. R. Bevington, *Data Reduction and Error Analysis for the Physical Sciences* (McGraw-Hill, New York, 1969). The method used is a combination of CURFIT and GRADLS.
 - [5] E. A. Lazarus, *Plasma Physics* **25**, 1271 (1983).

Figure Caption

Figure 1. Comparison between a histogram of 121 shots of experimental data from the S-1 Spheromak and a histogram of the results of a run of the code with one million samples at a temperature of 40 eV, a signal level of 50 photoelectrons, and a background level of 63 photoelectrons. The simulation is scaled down to match the scale of the data.

DISCLAIMER

This report was prepared as an account of work sponsored by an agency of the United States Government. Neither the United States Government nor any agency thereof, nor any of their employees, makes any warranty, express or implied, or assumes any legal liability or responsibility for the accuracy, completeness, or usefulness of any information, apparatus, product, or process disclosed, or represents that its use would not infringe privately owned rights. Reference herein to any specific commercial product, process, or service by trade name, trademark, manufacturer, or otherwise does not necessarily constitute or imply its endorsement, recommendation, or favoring by the United States Government or any agency thereof. The views and opinions of authors expressed herein do not necessarily state or reflect those of the United States Government or any agency thereof.

Fig. 1