MASTER

LA--10640-MS

DE86 007588

Three-Energy Gamma-Ray Absorptiometer (TEGA) for Nondestructive Assay of Plutonium and Uranium in Solution

Massimo Aparo*

DISCLAIMER

This report was prepared as an account of work sponsored by an agency of the United States Government. Neither the United States Government nor any agency thereof, nor any of their employees, makes any warranty, express or implied, or assumes any legal liability or responsibility for the accuracy, completeness, or usefulness of any information, apparatus, product, or process disclosed, or represents that its use would not infringe privately owned rights. Reference herein to any specific commercial product, process, or service by trade name, trademark, manufacturer, or otherwise does not necessarily constitute or imply its endorsement, recommendation, or favoring by the United States Government or any agency thereof. The views and opinions of authors expressed herein do not necessarily state or reflect those of the United States Government or any agency thereof.

*Visiting Scientist at Los Alamos. E.N.E.A.-C.R.E. Casaccia, Div. COMB-MEPIS, 0060 Rome, ITALY.

DISTRIBUTION OF THIS DOCUMENT IS UNLIMITED

Los Alamos Los Alamos National Laboratory
Los Alamos, New Mexico 87545

1-14

CONTENTS

ABSTRACT	1
I. INTRODUCTION	2
II. FORMULATION OF THE METHOD	3
III. BASIC INSTRUMENT FEATURES	6
IV. EXPERIMENTAL RESULTS	7
V. CONCLUSION	12
ACKNOWLEDGMENTS	12
REFERENCES	14

THREE-ENERGY GAMMA-RAY ABSORPTIOMETER (TEGA)
FOR NONDESTRUCTIVE ASSAY OF PLUTONIUM AND URANIUM IN SOLUTION

by

Massimo Aparo

ABSTRACT

An experimental approach for the nondestructive characterization of plutonium and uranium solutions is presented. The technique relies on the transmission of photons of three different properly chosen energies and allows an independent and simultaneous determination of plutonium and uranium by the different absorption of the two elements in the range of K-edge energies. The performances achievable have been evaluated through measurement of a set of solutions using the hardware of the compact K-edge densitometer. The plutonium and uranium concentrations ranged from 50 to 150 g/l. In this concentration range, the relative precision is below 3.0% for uranium assay and below 6% for plutonium assay. Further improvements of the performances of the technique are discussed.

I. INTRODUCTION

A nondestructive technique for independent and simultaneous determination of plutonium and uranium in solution is currently under development. The measuring technique relies on the transmission of photons at three different properly chosen energies and is based on differential absorption of plutonium and uranium in the range of K-edge energies. This technique can be considered as a further development of dual energy x-ray absorptiometry (DEXA),¹ which is based on differential absorption in the range of L-edge energies. DEXA is a non-destructive technique for the assay of mixed special nuclear material (for instance, thorium and uranium or uranium and plutonium) in solution. This measuring technique relies on the transmission of photons of two different energies and allows independent evaluation of one element (thorium or uranium), with the second one (uranium or plutonium) being determined on the basis of the total heavy elements. Field testing of the DEXA instrument for assay of thorium-uranium mixed solutions demonstrated assay precisions of better than 1% in a counting time of 4000 s for both elements in the range from 30 to 70 g/l. Also demonstrated was the potential of using the instrument for process control or as a safeguards assay tool in reprocessing facilities.

The DEXA technique appears to be unsatisfactory for assay of uranium-plutonium mixed solutions having high plutonium concentration because spontaneous emissions from ^{238}Pu overlap the transmitted peaks.

To overcome this problem and to make the method insensitive to the presence of fission products, we decided to shift the analysis to the range of K-edge energies where it becomes possible to independently assay both uranium and plutonium. For this reason, a feasibility study of the TEGA technique was carried out and the results were presented at the Sixth ESARDA conference.²

This report describes the physical principles of the technique and presents the first experimental results obtained using the same measuring head as that used with the compact K-edge densitometer (KED),³ an instrument developed at Los Alamos National Laboratory.

II. FORMULATION OF THE METHOD

Use of the TEGA technique in assaying solutions for plutonium and uranium concentration involves the measurement of the transmission through the solution

of three photons at energies E_0 , E_m , and E_h . The upper part of Fig. 1 shows the behavior of the photon absorption coefficient μ around the K-edge region of heavy elements (plutonium, uranium). (The elements are dissolved in a low-Z acid medium.) The lower and middle parts of Fig. 1 show the transmitted photon intensities for the three photopeaks at the energies E_0 , E_m , and E_h .

Fig. 1. Upper: an expanded view of the mass attenuation coefficient around the K-edge region vs gamma-ray energy for plutonium and uranium. Middle: a schematic representation of the transmitted photon intensities for "blank" (solid curves) and uranium-bearing solution (dotted curves). Lower: a schematic representation of the transmitted photon intensities for "blank" (solid curves) and plutonium-bearing (dotted curves) solution.

The transmissions of the photons at these three energies through a thickness d of sample solution are

$$-\ln T_{\ell} = [\mu_{Pu}(E_{\ell})\rho_{Pu} + \mu_U(E_{\ell})\rho_U]d$$

$$-\ln T_m = [\mu_{Pu}(E_m)\rho_{Pu} + \mu_U(E_m)\rho_U]d$$

$$-\ln T_h = [\mu_{Pu}(E_h)\rho_{Pu} + \mu_U(E_h)\rho_U]d \quad .$$

The transmissions $T_{\ell,m,h}$ are the ratios of photopeak intensities $R_{\ell,m,h}/R_{\ell,m,h}^*$, where the intensities in the denominator are the "blank" solution values and those in the numerator are solution values.

If E_{ℓ} , E_m , and E_h are chosen to satisfy the following conditions,

$$(1) \quad E(U \text{ K-edge}) < E_m < E(Pu \text{ K-edge})$$

$$(2) \quad \text{Back balancing: } \mu(U, E_{\ell}) = \mu(U, E_m)$$

$$(3) \quad \text{Forward balancing: } \mu(Pu, E_m) = \mu(Pu, E_h) \quad ,$$

then

$$\rho_{Pu} = \frac{-\ln (T_{\ell}/T_m)}{[\mu(Pu, E_{\ell}) - \mu(Pu, E_m)]d}$$

$$\rho_U = \frac{-\ln (T_m/T_h)}{[\mu(U, E_m) - \mu(U, E_h)]d} \quad .$$

Although it is possible to find radioisotopes that emit a line between the K edges of uranium and plutonium, satisfying both balancing conditions appears not to be achievable. Nevertheless we can create these conditions by correcting the measured transmissions T_ℓ^* and T_h^* to the values they would assume at energy E_ℓ and E_h . This correction is straightforward because of the linear behavior of $\ln \mu(E)$ vs $\ln(E)$, at least in a limited energy range far away from edges:

$$\ln T_h = [\mu(Pu, E_m) / (\mu(Pu, E_h^*))] \ln T_h^* = \beta \ln T_h^*$$

$$\ln T_\ell = [\mu(U, E_m) / \mu(U, E_\ell^*)] \ln T_\ell^* = \alpha \ln T_\ell^* .$$

Expressing Pu and U as a function of the measured transmissions T_ℓ^* , T_m , T_h^* , at energies E_ℓ^* , E_m , and E_h^* , respectively,

$$\rho_{Pu} = \frac{-\alpha \ln T_\ell^* + \ln T_m}{[\alpha \mu(Pu, E_\ell^*) - \mu(Pu, E_m)]d}$$

$$\rho_U = \frac{-\ln T_m + \beta \ln T_h^*}{[\mu(U, E_m) - \alpha \mu(U, E_h^*)]d} .$$

To fulfill these requirements, a ^{75}Se source has been chosen despite its short half-life of 120 days, and the measurements have been carried out using the lines at 96.733 keV (E_ℓ), 121.115 keV (E_m), and 198.596 keV (E_h).

III. BASIC INSTRUMENT FEATURES

The compact K-edge densitometer has been used to allow testing and evaluation of the method. This instrument consists of (1) a measurement head (Fig. 2) that allows measurement through a glove port of a glove box; (2) a portable planar detector (10 mm by 200 mm² with a resolution of 510 eV at 122 keV); and (3) a portable, computer-based multichannel analyzer (4096-channel Davidson). For our measurements, a cadmium shield (2 mm thick) was placed in front of the detector to achieve a lower energy cutoff that kept the deadtime of the multichannel analyzer below 20%. The solutions were contained in plastic vials with a transmission path of 2 cm. The distance between the source and the detector was about 28 cm, and the diameter of the output collimator was 0.5 cm. The ⁷⁵Se transmission source, located in a holder at one end of the measuring head, had a nominal activity of 52 mCi, but at the date of measurements it was 4 months old. An amplifier shaping time of 3 μ s was used during data acquisition, and pulse pileup rejection was employed. A "straight-through" measurement with a "blank" solution (3M HNO₃) was collected every day to obtain the

Fig. 2. Measuring head of the compact K-edge densitometer used for testing the performance of the TEGA technique.

unattenuated values $R_{\lambda,m,h}^*$. Because the method is based on ratios of transmitted peaks of a single radioisotope, it was not necessary to introduce a time correction that takes into account the half-life of ^{75}Se . A typical spectrum of the transmitted radiation is shown in Fig. 3.

IV. EXPERIMENTAL RESULTS

A set of standard (pure plutonium or uranium as well as mixed) aqueous solutions ($3M \text{ HNO}_3$) has been prepared to calibrate and test the performance of the technique. Calibration of the instrument has been carried out by using one pure plutonium and one pure uranium standard solution with a counting time of 4000 s. However, the choice of the calibration solutions does not affect the results.

Fig. 3. Typical straight-through spectrum of transmission source. The 88-keV peak from a ^{109}Cd source can be used for counting loss correction.

Figure 4 is a plot of the calibration constant ($\Delta\mu d$) vs plutonium concentration. For each plutonium sample, the correction constant was determined both by measurements (as the ratio between the 97- and 121-keV transmitted peaks for different uranium solutions) and by calculation (from literature values). Figure 5 is the same plot of calibration constants, but vs uranium concentration.

The same standard solutions have also been measured with the KED system to compare the performances of the two techniques, TEGA and KED. The results obtained are summarized in Table I and compared with the known values; related uncertainties (1σ) are also given.

It is worthwhile to point out that the difference in respective uncertainties between uranium and plutonium is principally due to the large correction factor (2.1) applied to the transmission at energy E_0 for satisfying the back balancing condition. Such a problem was not noticed in the previous feasibility study because the evaluation of the expected performances was carried out using the line at 66.11 keV, closer to the ideal energy E_1 (70 keV). The estimated correction factor of the 66-keV line is 0.8. Currently, the experimental apparatus will not allow the measurement of the 66.11-keV line.

Figures 6 and 7 show the relative assay precision expected for plutonium as well as for uranium as a function of the U/Pu ratio in a counting time of 2000 s. The blocks on the figures represent the measured relative precision.

It must be pointed out that, in principle, the technique is sensitive to the matrix material, which is taken into account through the measurement of "blank" solution. So matrix variation (for instance, fluctuation of acid molarity) may introduce systematic errors that are independent of uranium and plutonium concentration. Calculations have shown that a 0.5M variation in the acid (HNO_3) molarity introduces biases of about 0.5 g/l in the assay of plutonium and about 0.15 g/l in the assay of uranium. In both cases, by correcting the transmission data for the presence of the other heavy element, we have actually enhanced the effect of the variation of the matrix. The large difference in the two correction factors causes the introduced bias in the plutonium determination to be much greater than the bias in the uranium measurement.

Fig. 4. A plot of the calibration constant Δ_{121} in cm^3/g vs plutonium concentration. Each point represents a different correction factor α determined by the ratio between the 121- and 97-keV transmitted peaks for each uranium solution or by literature values.

Fig. 5. The same plot as in Fig. 4, but vs uranium concentration. The correction factor β was determined by the ratio between the 121- and 198-keV transmitted peaks for each plutonium solution or by literature values.

TABLE I
COMPARISON OF CONCENTRATION MEASUREMENT RESULTS

Known Concentration		TEGA ^a		KED ^b
Plutonium (g/l)	Uranium (g/l)	Measured Concentration		Measured Concentration
		Plutonium (g/l)	Uranium (g/l)	Plutonium (g/l)
44.8	-	45.49 ± 3.27	0.219 ± 1.58	44.31 ± 1.64
94.0	-	calibration		calibration
142.5	-	141.78 ± 4.07	-1.06 ± 1.76	142.6 ± 2.61
223.2	-	234.02 ± 5.24	-2.92 ± 1.96	228.8 ± 3.76
-	51.9	-0.717 ± 3.03	51.30 ± 1.46	0.865 ± 1.46
-	103.7	calibration		-4.65 ± 1.5
-	155.6	-2.46 ± 3.69	153.19 ± 1.67	-1.21 ± 1.53
45.2	103.7	48.43 ± 4.05	98.4 ± 1.76	40.10 ± 1.63
68.6	77.8	69.38 ± 4.08	77.26 ± 1.77	66.83 ± 1.82
93.1	51.9	97.1 ± 3.73	51.14 ± 1.63	88.78 ± 1.98
103.7	155.6	108.8 ± 4.97	152.98 ± 1.63	102.8 ± 2.23

^aThe TEGA technique used a 2000-s counting time.

^bThe KED technique used a 3000-s counting time.

Fig. 6. A plot of expected per cent precision (1σ) vs Pu/U ratio for a 2000-s counting time and different uranium concentrations. The blocks represent the measured precision.

Fig. 7. A plot of expected per cent precision (1σ) vs the U/Pu ratio for a 2000-s counting time and different plutonium concentrations. The blocks represent the measured precision.

V. CONCLUSION

The instrumental approach described here appears interesting because it allows simultaneous and independent determination of two heavy elements, plutonium and uranium, that often flow in the same stream in different parts of the nuclear fuel cycle (that is, in reprocessing plants, conversion plants, and others). In addition, by optimizing the design of the measuring head, the instrument might work both in an active mode (determination of plutonium and uranium concentration) and in a passive mode (isotopic analysis of plutonium), allowing a complete characterization of the assay solution.

Further improvement of the assay precision appears feasible either by (1) rearranging the measuring head to reduce the distance between the source and the detector and introducing a stronger ^{75}Se source (for instance, activity 100 mCi) in such a way as to use the 66.11-keV line, or (2) because the 66.11-keV line is the most important cause of imprecision, by replacing the ^{75}Se source with a mixed source of ^{75}Se and ^{133}Ba . According to calculations made on the basis of the experimental results, the first alternative [use of a stronger ^{75}Se source (activity 100 mCi) with a shorter distance between the source and detector] should allow measurement of plutonium with a relative precision under 2.5% in a counting time of 1000 s. The second alternative [use of a mixed source of ^{75}Se and ^{133}Ba (with activities of 25 and 10 mCi, respectively) and the same measuring device we used in our study] should allow a relative precision under 2% in a counting time of 1000 s. In this case, the 81-keV line of ^{133}Ba would replace the 66-keV line of ^{75}Se .

The relative precisions achievable by these two different arrangements for plutonium assay are plotted vs the U/Pu ratio in Figs. 8 and 9.

ACKNOWLEDGMENTS

I would like to acknowledge the support given by members of the Safeguards Assay Group without which I would not be able to carry out this work. I would especially like to thank R. Augustson for his continuous helpful suggestions as well as for his comments on this report and L. R. Cowder for handling the plutonium and uranium solutions and for preparing the measuring head.

Fig. 8. A plot of the expected relative precision in plutonium assay vs U/Pu ratio for 1000-s counting time and different plutonium concentrations. The ^{75}Se source activity was assumed to be 100 mCi; that of the distance source detector, 14 cm.

Fig. 9. The same plot as Fig. 8, but the source was assumed to be a mixed ^{75}Se and ^{133}Ba source, with activities of 25 and 10 mCi, respectively.

REFERENCES

1. M. Aparo, B. Mattia, F. V. Frazzoli, and P. Zeppa, "Dual Energy X-Rays Absorptiometer for Non Destructive Assay of Mixed Special Nuclear Material in Solution," in Proceedings of the Fifth Annual Symposium on Safeguards and Nuclear Material Management, Versailles, France (1983), pp. 271-275.
2. M. Aparo, P. Cresti, F. V. Frazzoli, R. Remetti, C. Vicini, and P. Zeppa, "Characterization of Plutonium Solutions (Densitometry and Isotopic Analysis by Gamma Spectrometry)," in Proceedings of the Sixth Annual Symposium on Safeguards and Nuclear Material Management, Venice, Italy (1984), pp. 249-253.
3. L. R. Cowder, S. F. Klosterbuer, R. H. Augustson, A. Esmailpour, R. Hawkins, and E. Kuhn, "A Compact K-Edge Densitometer," in Proceedings of the Sixth Annual Symposium on Safeguards and Nuclear Material Management, Venice, Italy (1984), pp. 261-268.