27
8/27/84

①

UCRL--50021-83

DE84 017287

D-0431-6

I-17153

NOTICE

PORTIONS OF THIS REPORT ARE ILLEGIBLE. It has been reproduced from the best available copy to permit the broadest possible availability.

DISCLAIMER

This report was prepared as an account of work sponsored by an agency of the United States Government. Neither the United States Government nor any agency thereof, nor any of the employees, makes any warranty, express or implied, or assumes any legal liability or responsibility for the accuracy, completeness, or usefulness of any information, apparatus, product, or process disclosed, or represents that its use would not infringe privately owned rights. Reference herein to any specific commercial product, process, or service by trade name, trademark, manufacturer, or otherwise does not necessarily constitute or imply its endorsement, recommendation, or favoring by the United States Government or any agency thereof. The views and opinions of authors expressed herein do not necessarily state or reflect those of the United States Government or any agency thereof.

DISTRIBUTION OF THIS DOCUMENT IS LIMITED

8/8

MASTER

UCRL-50021-83
Distribution Categories
UC-21, 22

Laser Program Annual Report **83**

Scientific Editor: Charles D. Hendricks

Publication Editors: M. Louise Rufer
Peter W. Murphy

MS date: June 1984

Lawrence Livermore National Laboratory

Available from: National Technical Information Service • U.S. Department of Commerce
5285 Port Royal Road • Springfield, VA 22161 • \$35.50 per copy • (Microfiche \$4.50)

Preface

In the 1983 *Laser Program Annual Report* we present the accomplishments and unclassified activities of the Laser Program at Lawrence Livermore National Laboratory (LLNL) for the year 1983. It should be noted that the report—of necessity—is a summary, and more detailed expositions of the research can be found in the many publications and reports authored by staff members in the Laser Program.

The purpose of this report is to present our work in a brief form, but with sufficient depth to provide an overview of the analytical and experimental aspects of the LLNL Inertial-Confinement Fusion (ICF) Program.

The format of this report is basically the same as that of previous years. Section 1 is an overview and highlights the important accomplishments and directions of the Program. Sections 2 through 7 provide the detailed information on the various major parts of the Program: Laser Systems and Operations, Target Design, Target Fabrication, Fusion Experiments, Laser Research and Development, and Energy Applications.

In the interest of economy, the use of four-color photographs has been reduced significantly, and the physical makeup of the report has been modified. The overall quality of the report and the research presented remains, as previously, at a high level.

For more detailed and in-depth information about the LLNL ICF Program and about research reported in this volume, contact the authors of the various articles.

The Inertial-Confinement Fusion research program at the Lawrence Livermore National Laboratory is supported by the U.S. Department of Energy.

C. D. Hendricks
Scientific Editor

Acknowledgments

I would like to acknowledge the efforts of the many people who made this report possible. First and foremost, of course, are the authors who did the research, development, and construction, and who then wrote the articles reporting the results of this work. Also, without dedicated secretaries who typed and retyped the articles, and helped to inspire the authors to get their articles in on time, this report would never have seen the printer. There was a special effort by many people in the Technical Information Department as well as by authors and secretaries to write, design, assemble, edit, proof, modify, draw, and print this report in the shortest time possible. That we have managed to deliver this high-quality Annual Report to the reader in the shortest time in the history of the Laser Program is a credit to the entire Annual Report staff. Several individuals deserve special acknowledgments for their work on the Report. Sam Tindell was instrumental in literally pulling the articles from authors and secretaries. She worked very hard at keeping everything coming in on schedule and keeping track of the entire process. Without the tenacity and determination of Louise Rufer who served as the publication editor, the report would not have been as well edited nor would it have been produced in such a short time. The individual section scientific editors developed the section outlines, made writing assignments, read their sections many times to assure technical accuracy, and encouraged their authors to produce excellent articles on time to meet deadlines. The TID editorial and production staffs provided excellent output even though some authors and their articles were difficult to handle. Publishing a book such as this one would not be possible without the highly skilled technical publications personnel. To all those who contributed their time and expertise, my sincere thanks and appreciation.

Although the Scientific Editor assignment is a dubious honor, I should thank Drs. Emmett and Holzrichter for giving me this singular opportunity for the second year.

I also would like to acknowledge the U.S. Department of Energy which supports the Laser Fusion Program at Lawrence Livermore National Laboratory under Contract W-7405-Eng-48.

C. D. Hendricks
Scientific Editor

Contents

Laser Program Overview

Section Editor:
C. D. Hendricks

Section

1

1-1

Laser Systems and Operations

Section Editor:
G. J. Suski

Section

2

2-1

Target Design

Section Editor:
L. J. Suter

Section

3

3-1

Target Fabrication

Section Editor:
A. M. Hawryluk

Section

4

4-1

Fusion Experiments

Section Editor:
V. W. Slivinsky

Section

5

5-1

Laser Research and Development

Section Editor:
J. R. Murray

Section

6

6-1

Energy Applications

Section Editor:
W. J. Hogan

Section

7

7-1

Subject and author indexes follow the last page of Section 7.

Target chamber of the Novette
laser which became operational
in 1963.

Laser Program Overview

1

Laser Program Overview

Contents

Section Editor: C. D. Hendricks

Introduction	1-1
Summary of Achievements	1-2
Laser Operations and Systems	1-2
Target Design, Target Fabrication, and Fusion Experiments	1-3
Laser Research and Development	1-4
Energy and Military Applications	1-5
Program Resources	1-5
Program Facilities	1-6
References	1-7

Laser Program Overview

Introduction

J. F. Holzrichter

During 1983, the Inertial Confinement Fusion (ICF) Program at the Lawrence Livermore National Laboratory (LLNL) conducted an active and successful experimental program. With the Novette laser—in near-constant use since its completion in December 1982—experimental emphasis has been to confirm theories on short-wavelength laser–target physics and to test the Nova laser design. In particular, experiments using Novette's 2ω ($0.53\ \mu\text{m}$) light showed:

- Efficient laser absorption and low suprathermal electron production using beam intensities greater than $10^{14}\ \text{W}/\text{cm}^2$ on target-plasma scalelengths 5 times longer than previously used. This confirms and extends the short wavelength experimental data base obtained using 10- to 100-J lasers (see Fig. 1-1).
- Improved coupling to fusion capsules with low target preheat in target compression experiments.
- That important physics demonstrations, such as producing a soft x-ray laser, can begin. For the x-ray laser experiments, the Novette beams were focused to $I > 10^{14}\ \text{W}/\text{cm}^2$ over a line focus $100\ \mu\text{m}$ wide by 1.2 cm long.
- Efficient harmonic conversion with large beams and large arrays producing $>80\%$ internal crystal conversion efficiency and $>60\%$ system conversion efficiency (see article entitled "Novette Laser System" in Section 2 of this book).

The results of the Novette experiments have confirmed our ICF target experimentation strategy. We will use Nova to verify our implosion calculations by using scale models of high-gain targets. With Nova, we expect to show that the laser–plasma interaction over scalelengths of interest for high-gain targets ($\sim 1\ \text{cm}$) is well controlled, that the ablation pressures and efficiencies are consistent with the requirements of high-gain targets, and that the compression requirements of 1000 times are compatible with fluid physics. Potential fluid physics issues associated with fluid instabilities, mixing, symmetry, and center spot ignition will be evaluated.

Fig. 1-1. Novette experiments show that $0.53\text{-}\mu\text{m}$ laser light is efficiently absorbed at power densities of 10^{14} to $10^{15}\ \text{W}/\text{cm}^2$ with plasma scalelengths $5\times$ those used with Argus experiments. These results support our models that predict successful compression of high-gain targets using short-wavelength ($\sim 0.5\ \mu\text{m}$) irradiation.

Summary of Achievements

Direct-drive Nova targets require a low-Z fuel container-ablator and a thick layer of ultra-smooth solid-density DT. The materials research needed to provide this target, and the material developments for special targets supporting studies in x-ray spectroscopy, x-ray lasers, equation of state, and opacity proceeded very well this year.

Our laser research and development activity is dedicated to supporting the Nova laser system and the target experimental program, and to providing a conceptual basis for building a large 5- to 10-MJ laser driver. Such a driver would demonstrate the gain vs energy relationship for ICF targets as well as provide an intense source of fusion neutrons and x rays for weapons-physics and weapons-effects studies.

Summary of Achievements

Laser Operations and Systems

Novette Experiments. We used the Novette laser for a variety of experiments during 1983. It was first configured for 100-ps pulse length and used for x-ray laser studies from January through March, 1983 (see Fig. 1-2). In April and May, the system was reconfigured for 1-ns operation and dedicated to compression and scaling

experiments. These experiments immediately showed improved laser-target coupling using up to 9 kJ of 0.53- μm light. (Previous experiments conducted on Shiva at 10 kJ with 1.06- μm light had shown unacceptable coupling.) The Novette laser worked as designed, confirming the Nova design basis. It has produced >12 TW/beam at 1.06 μm in 100 ps, and >6 TW/beam at 0.53 μm . At 1 ns, the system produced 9 kJ/beam at 1.06 μm , and 4.5 kJ/beam at 0.53 μm .

We experienced some difficulty with the harmonic conversion cells. The thin (10 μm) layer of index-matching fluid in the crystal array trapped minor contaminants,

Fig. 1-2. The Novette laser is configured to irradiate a variety of target designs. The 1.06- μm laser output is harmonically converted to 0.53- μm green radiation.

causing material damage that would propagate on subsequent shots. This was corrected by removing the liquid layer from the Novette array, by redesigning the Nova crystal cells with thicker liquid layers that do not trap contaminants, and by using an improved fluid.

A second set of procedural difficulties was encountered with the alignment of complex targets and with the use of beam diagnostics at multiple wavelengths. Because these diagnostic sensors had been based on early prototype designs for Nova, we were able to make corrections to the final Nova components, and to improve operations with the Novette system.

The performance of the Novette laser at 1ω and 2ω has given us confidence that our propagation codes accurately describe the performance of large-aperture lasers and that Nova will meet or exceed its projected energy and power operating points.

Nova Installation. The Nova project team continued to install the laser and target systems for scheduled completion in September, 1984. Nova remains within budget and will meet its technical objectives. A great deal of effort has gone into making the Nova system effective for laser-target experimentation.

Based on our experience with Novette, we predict Nova peak output at $0.35\ \mu\text{m}$ to be 70-kJ in 1 ns (see Fig. 1-3). This output is completely consistent with Nova line-item performance specifications—at $0.35\ \mu\text{m}$, 40- to 70-kJ light at 3 ns. In addition, with Novette's demonstrated performance at 1.06 and $0.53\ \mu\text{m}$, we confidently predict that Nova operating points will fall within the line-item specifications of 80- to 120-kJ at $1.06\ \mu\text{m}$ (3 ns), and 50- to 80-kJ at $0.53\ \mu\text{m}$ (3 ns).

Target Design, Target Fabrication, and Fusion Experiments

The target design activity has centered on designing Novette targets and analyzing experimental data, on preparing Nova experiments, and on developing a long-term strategy. To support target implosion studies, we have made many improvements to the LASNEX code allowing us to maximize our use of the LLNL computer facility as well as improving the accuracy of our

numerical modeling. At the wavelengths and intensities anticipated for high-gain, long-scalelength, high-Z targets we find that instability thresholds (see Fig. 1-4) are dominated by collisional effects (inverse bremsstrahlung absorption). Although some reactor designs exceed these thresholds, we have observed (numerically) that the hot electrons generated by plasma instability do not dominate the electron energy spectrum until they are driven by at least several times the instability threshold. Nova scaling experiments are being designed to confirm the findings of these analyses, and to verify the anticipated high quality of the implosion environment required for a high-gain target.

The direction of our target fabrication technology has changed to meet the needs of weapons-physics experiments and of high-density Nova experiments. Novette has been used to evaluate many of these changes. For Novette x-ray laser targets, special construction techniques were developed to make a high-aspect-ratio lasing channel ($40\ \mu\text{m} \times 4\ \mu\text{m} \times 14\text{-mm-long}$) filled with a lasing gas (neon or SF_6) at a prescribed pressure. These and similar micro-fabrication techniques are required to effectively use these laser systems for high-temperature and high-density physics experiments.

The most important activities for Nova high-density experiments have related to the development of cryogenic fuels and polymer fuel containers for direct-drive experiments. Target fabrication experiments demonstrating thin-layer ($\sim 10\ \mu\text{m}$) cryogenic fuel targets were successfully accomplished. Investigation of thick-layer

Fig. 1-3. As modeled by our propagation code MALAPROP, Nova will approach output levels (at 1 ns) of 90 kJ at 1ω , 75 kJ at 2ω , and 70 kJ at 3ω . The maximum power level is determined by small scale self-focusing, which dramatically increases the fluence on the crystal array as a critical power is reached.

Fig. 1-4. Nova reactor targets operate with plasma scalelengths 10^4 times the $0.35\text{-}\mu\text{m}$ laser wavelength and at intensities $I\lambda^2 \sim 1\text{--}3 \times 10^{14} \text{ W} \cdot \mu\text{m}^2/\text{cm}^2$. In low-Z targets, instability thresholds may be exceeded. In high-Z targets, strong collisional absorption of laser beam and plasma waves will suppress instabilities.

cryogenic DT fuel for Nova direct-drive targets has begun. One very promising technique investigated by our colleagues¹ at the University of Illinois uses a thermal gradient method to position the thick liquid-DT layer symmetrically inside a target sphere. In addition to the cryogenic research, the development of plastic fuel capsules has been successful. Work accomplished at LLNI and at KMS Fusion has led to successful control of fuel capsule roundness, smoothness, and diameter.

Laser Research and Development

The Laser R&D Program has three primary concerns. The first is to provide the laser technology support for Nova, the second is to support the ongoing target experimental program, and the third is to provide a basis for designing a multimegajoule driver. The 3ω antireflection (AR) coating for Nova optics is the final major R&D project for Nova. Small research samples using etched sol-gel coatings have worked very satisfactorily for ultraviolet coatings. The 3ω AR coatings have been made routinely on small substrates at the 1-ns damage level of 5 to $8 \text{ J}/\text{cm}^2$, a satisfactory improvement over the

damage level of 2 to $3 \text{ J}/\text{cm}^2$ measured on thin-film AR coatings and on earlier sol-gel coatings. These improved coatings are now being engineered for the large-area optics application. Much effort is under way to finalize the Nova harmonic cell design. This work includes measuring the onset of Raman sidescattering in the 30-cm KDP cell segments, measuring the damage levels in KDP (which when laser-annealed exceeds $5 \text{ J}/\text{cm}^2$), designing a variety of harmonic conversion schemes including the 4ω array for Novette, and finally obtaining accurate measurements on damage processes in the index-matching fluid layers.

We are conducting a variety of studies to provide a basis for conceptualizing a large laser system in the 5- to 10-MJ range. For any multimegajoule ICF driver to be economically acceptable, its cost should not differ greatly from that of the Nova laser. This statement holds true for either of the two available laser options—Nd:glass or KrF. To accomplish this objective:

1. The laser amplifier should store optical energy with $>3\%$ efficiency. This reduces the relative size of the pulsed power system by about a factor of 4 compared with that of Nova. In addition, the cost per joule of the pulsed power system must be reduced by about four times by improved technology.
2. The efficiency of laser-energy extraction from the medium must be increased by at least a factor of 2 compared with the Nova operating point in order to reduce the cost of a glass laser medium. Also, to reduce the relative area of our laser-beam-handling optics by a nominal factor of 4, we must operate at higher fluences, increasing from the present operating points of $\sim 5 \text{ J}/\text{cm}^2$ to $20 \text{ J}/\text{cm}^2$.
3. The parts count must be similar to that of Nova in order to have similar engineering design and installation costs. Possible approaches include using simple multipass architectures, and using large apertures—up to 3 m in diameter.

As a step toward these objectives, we have measured some coating damage levels of up to $40 \text{ J}/\text{cm}^2$ at 10 ns . We have seen laser glass that can absorb almost twice as much pump light in a laser cavity as present glasses and yet still maintain a long fluorescence time. Investigation of solid

state pumping schemes shows that it may be possible with xenon flashlamps to exceed 5% storage efficiencies. In addition, a completely new design for a Pockels cell optical switch has been invented that makes use of a low-density plasma to provide an electrical conduction path covering large surface areas of KDP. This device makes it possible to consider large-area optical switches for large-diameter multipass laser cavities. Finally, we continue to explore cost reduction techniques on laser optics, pulsed power, and accompanying structures.

Energy and Military Applications

In an economics analysis,² Nuckolls evaluated ICF's potential for competing with coal and breeder-supplied fission energy. In our present view of an evolving energy future ICF must be economically competitive with these technologies. The results of these studies show that ICF can compete effectively against coal or fission if several of the following innovations can occur:

- Higher gain targets (200 to 300 times vs 100 times) leading to a 2 times reduction in driver costs.
- Polarized DT fuel leading to >3 times reduction in driver costs.
- Efficient, multiplexed drivers lead to 2 times reduction in driver costs.
- Higher conversion efficiency of target-output energy to electrical energy using higher temperatures or direct conversion. This could lead to 2 times improvements in plant output power per unit cost (35% — >60%).

Reactor costs may be further reduced when compared to present designs by using lower cost containment buildings. The use of inert heat transfer media such as Li_2O rather than liquid lithium could make this possible.

The three new reactor designs introduced in the 1982 Laser Annual³ are aimed at meeting these economic and technical objectives. The Cascade reactor system, with a Li_2O granule wall, is fireproof, operates at a very high temperature (up to 1200 K), is compact, and controls the tritium inventory. Pulse*Star is a pool-type reactor in which the fusion chamber rests in a pool of

fireproof lithium-lead liquid that acts as the tritium breeder and heat-transfer medium. The Starburst reactor is an attempt to explore direct conversion techniques to raise the efficiency of ICF fusion systems. Direct conversion is particularly cost-effective because an improved thermal-to-electrical efficiency factor directly determines the cost per watt of the entire power plant.

Finally, the potential to increase the gain or lower the laser threshold energy at constant gain is very important to the future of ICF. The recent discussions of polarized fusion fuel⁴ for magnetic fusion systems have been extended by More⁵ to ICF systems. The benefits may be tremendous: reductions of three or more in ignition threshold may be available.

In summary, the ICF concept is being successfully tested using harmonically converted Nd:glass laser systems. This work is not only contributing important physics information, but is also providing a positive basis for a desirable fusion energy future.

Author: J. F. Holzhrichter

Program Resources

Laser Fusion Program funding for FY 83 consisted of \$34.8 million in operating funds, \$4.5 million in capital equipment funds, and \$34.5 million in construction funds. The operating budget represented a 5.5% dollar increase from FY 82, and supported a work force average of 29% full-time-equivalent employees (FTEs), reduced from 315 FTEs the previous year. Actual buying power in FY 83 was down 4% from FY 82 because of inflation. A 10-year history of funding and work force is shown in Fig. 1-5.

When compared with FY-82 costs, a significant shift in costs occurred in FY 83 from the laser development category to the fusion theory and experiment category because of the completion of Novette construction in early FY 83 and the subsequent operation and experimental use of that facility.

Direct work force costs were, as in the past, less than half of total costs, in keeping with our goal of maintaining maximum flexibility by using outside procurements.

Fig. 1-5. Ten-year history of ICF Program resources.

Program Facilities

Table 1-1. Distribution by budget category of costs and work force for FY 83.

Budget category	Operating costs (\$ thousand)	Work force (FTE)
Laser development	11 161	65
Fusion theory and experiments	43 638	231
	54 799	296

Table 1-1 shows a distribution by budget category of FY-83 costs and work force.

Capital equipment funds were available to support our need for increased and improved target fabrication capabilities and for diagnostic tools. The total represented a 5% increase over the previous year.

Nova was fully funded at \$176 million, with the final appropriation of \$34.5 million occurring in FY 83. Because of full funding, Nova will be completed by the end of FY 84 at a cost reduction of \$9 million from the total necessary if appropriations had been carried into FY 84. As of the end of FY 83, Nova is on schedule and on budget.

Author: L. P. Altbaum

Program Facilities

In 1983, facilities construction and renovation in the Laser Program featured the completion of a major new laboratory,

continued progress on a new office building, and the completion of many important smaller projects. Since the beginning of the current long-range development plan four years ago, we have experienced steady and significant growth; our building space now totals over 638 000 ft². This continued progress is bringing us within reach of our ultimate goal, that of providing permanent facilities for all our people and experiments.

Major work for the ICF Program in 1983 included completing and occupying the Nova laboratory addition (Building 391, Increment II). In the past two years the completion of several laboratory and office building projects—culminated by the new addition to Building 391—has provided the facilities vital to the Nova Project.

During the past year, the Facilities Group successfully managed the completion of \$3.6 million in construction and maintenance projects, and guided or assisted in the completion or continued progress of many smaller construction and modernization tasks for new and existing facilities. This major facilities work during 1983 included:

- Designing, planning, and/or managing projects for Nova facilities and experiments: liquid nitrogen supply station for

the Nova laser cooling system, laboratory activation, diagnostics loft construction, modifications to the Operations/Control Room, developmental assistance for and relocation of the sol-gel and neutral solution process experiments, activation of the completed Nova office building.

- Coordinating modifications to the fusion experiments and micro-fabrication laboratories in Bldg. 381.
- Improving existing facilities: upgrading seismic and OSHA crane specifications, upgrading environmental controls in Nova office building.

- Modifying existing laser and advanced laser laboratories; rebuilding laser banks.
- Planning and directing site improvements: grading corrections and related exterior work around Bldg. 391, construction of the crafts area and Laser Program corporation yard along the North Outer Loop Road.

In addition, we installed new landscaping in the areas around several of our buildings, and we restored the landscaping in the area between Bldgs. 381 and 481.

Author: R. A. McCardle

-
1. L. Mok, K. Kim, T. P. Bernat, and D. H. Darling, "Temperature Effects on the Formation of a Uniform Liquid Layer of Hydrogen Isotopes Inside a Spherical Cryogenic ICF Target," <i>J. Vac. Sci. Technol.</i> A1, 897 (1983).2. J. H. Nuckolls "The Feasibility of Inertial Confinement Fusion," <i>Physics Today</i> 35(9), 24 (1982).3. 1982 <i>Laser Program Annual Report</i>, Lawrence	Livermore National Laboratory, Livermore, Calif., UCRL-50021-82 (1983), pp. 8-3 to 8-34. 4. R. M. Kulsrud, H. P. Furth, M. Goldhaber, and E. J. Valeo, "Fusion Reactor Plasma with Polarized Nuclei" <i>Phys. Rev. Lett.</i> 49, 1248 (1982).5. R. M. More, "Nuclear Spin-Polarized Fuel in Inertial Fusion" <i>Phys. Rev. Lett.</i> 51, 396 (1983).
-

Fig. 2-16. Apodizer test pattern. Note strong diffraction pattern of diagonal wire.

Nova Frequency Conversion and Focusing System

lines separated by one crystal width (27 cm), intersected at 90° by a second pair, similar to a tic-tac-toe pattern. Spot patterns are required to apodize the regions around the array discrete-support points. The system generates straight lines by traversing in

Nova
prog
fram
squa
shop
sect
welc
and
10 0
feet
in 14

Nova spaceframe welding in progress. Building the spaceframe required 10 miles of 6-in.-square structural steel that was shop-fabricated into 50 large sections, then assembled and welded on site. The fabrication and assembly, which required 10 000 weld joints and 20 000 feet of weld, were accomplished in 14 months.

Laser Systems and Operations

2

Laser Systems and Operations

Contents

Section Editor: G. J. Suski

Introduction	2-1
Nova System Design and Performance	2-2
Nova System Activation	2-5
Nova Frequency Conversion and Focusing System	2-8
Optical Components for Nova	2-14
Front End Subsystem	2-20
Power Systems Energy Storage	2-24
Beam Alignment and Diagnostics Systems	2-30
Target Systems	2-39
Mechanical Systems	2-41
Integrated Controls	2-45
Nova Project Management	2-47
Novette Laser System	2-48
References	2-63

Laser Systems and Operations

Introduction

R. O. Godwin

Novette. During this first year of operation, the Novette laser demonstrated its flexibility as an experimental facility, verified many of Nova's performance parameters, and corrected several design problems for Nova.

The first series of Novette experiments was conducted to investigate the physics of a laser-driven x-ray laser. We used 0.53- μm light pulses at 100-ps duration. These first experiments addressed many of the issues central to achieving x-ray lasing, and demonstrated the versatility of Novette as a test facility. Following these initial experiments, the laser was prepared for inertial fusion experiments at 0.53 μm and 1 ns, and operated in this mode for the balance of the year. Novette fired 474 high power shots during the year, 217 of which irradiated targets of various types. Details of the Novette operation are contained in the article "Novette Laser System" at the end of this section.

During its operation, the Novette laser demonstrated the following performance per beam:

- 12.6 TW at 1.05 μm and 100 ps
- 6.8 TW at 0.53 μm and 100 ps
- 9.2 kJ at 1.05 μm and 1.0 ns
- 5 kJ at 0.53 μm and 1.0 ns.

The above results meet or exceed the goals for the Nova laser system.

Using Novette, we identified and solved the following problems in the Nova laser design: Poor reliability associated with the 19-in. xenon flashlamps used in the 46-cm disk amplifiers and the rod amplifiers; laser-induced damage in the crystal-array index matching fluid; bulk damage to some input spatial filter lenses of BK-7 glass; and difficulty with the Nova 0.53- μm target alignment method. Literally hundreds of minor concerns were also identified and addressed. As a result of solving these problems on Novette, the Nova laser system will be operational sooner and with fewer operational difficulties.

Nova. The construction phase of the Nova laser project accelerated during the past year. The fraction of the total contracts awarded increased from 60 to 90%. Low inflation and good engineering design resulted in very favorable prices for these contracts. We used the savings to add one 31.5-cm disk amplifier and one 46-cm disk amplifier to each of the 10 laser chains. These amplifiers will provide Nova with more energy and flexibility for complex pulse shaping.

During the year, we relocated the Nova control room from the basement to the main floor, adjacent to the laser bay. We learned, through operating experience on Novette, that the personnel interaction was primarily between the control room and the laser bay, rather than between the control room and the master oscillator room (MOR). We also changed the target illumination geometry from 10 opposing beams to 10 non-opposing beams. This prevents one laser beam from firing into an opposing laser beam, if the target is not present. Non-opposing geometry also provides the opportunity to use the opposing port to mount laser diagnostics for high-powered laser shots.

Installation and activation of Nova is proceeding well on the first eight beams. (The remaining two beams are installed and operational on Novette, and will be shut down and moved to Nova in June 1984.) The Nova MOR, laser

power conditioning bank (for eight beams), and much of the control room are complete and operational. Laser hardware for eight beams is installed through the 20-cm amplifiers, and six of these beams have been fired at 500 J per beam at 1 ns. Much of the mechanical hardware and some of the turning mirrors for the laser diagnostics are installed in the switchyard. The target chamber is installed, and vacuum system installation is starting in the target chamber room.

There are three remaining issues for the Nova laser: (1) selection of the method of retro-pulse protection; (2) demonstration of frequency conversion array performance at $0.53\text{ }\mu\text{m}$ and $0.35\text{ }\mu\text{m}$; and (3) determination of the need to replace BK-7 glass in two input spatial filter lenses with fused silica.

As a result of the demonstrated performance of Novette, combined with our computer models for lasers, we are confident that the Nova laser will exceed all of its performance goals by a comfortable margin. The Nova project is in excellent condition from both a cost and schedule viewpoint. We look forward to project completion at the end of FY 1984 and the initiation of experiments early in FY 1985.

Nova System Design and Performance

The Nova laser design remains fundamentally the same as has been described in previous Laser Annuals.¹ However, we have elected to procure additional large-disk amplifiers (one 31.5-cm and one 46-cm aperture) in each of the 10 Nova beamlines. We allowed for sufficient physical space² in the original design for this option, and favorable cost experience with Nova procurements enabled us to exercise it. The addition of the large-disk amplifiers, shown schematically in Fig. 2-1, constitutes a cost-effective method of ensuring maximum pulse shaping flexibility, which is important to future Nova target experiments. We selected 1.0-cm-thick KDP crystals to optimize performance at 2ω and 3ω in the important 1- to 3-ns temporal regime. Laser performance with these final amplifiers in place will be summarized later in this article. Our crystal thickness choice, and other considerations related to the final frequency-conversion-array design, are described in "Nova Frequency Conversion and Focusing System," later in this section.

In addition to the points mentioned above, we have simplified and solidified our target-focusing strategy and have decided on the diagnostics for on-target laser-beam power and energy at all wavelengths. Dispersion in the fused-silica lens

permits the target to be free from residual wrong-color light. A schematic representation of this baseline configuration is shown in Fig. 2-2.

Finally, we have decided to characterize the Nova beams in the focal volume by implementing an opposed-port diagnostic. This diagnostic is intended to evaluate energy distribution at the equivalent target planes during full power operation at any wavelength. A conceptual design, featuring an achromatic, Cassegrainian optical system, has been chosen, and a detailed design is under way.

In last year's Laser Annual,² we presented our best estimates of component damage thresholds that were likely to limit laser performance at 1ω , at a variety of temporal pulse durations from 0.1 to 5.0 ns. These components were: the neutral-solution-processed (NSP) spatial filter; entrance lens surfaces; the high-reflectivity (HR), dielectric-coated turning mirror surfaces, the KDP frequency-conversion crystals; and the Faraday-rotator glass and polarizer coatings.

NSP technology³ has been extensively proven on Novette. During day-to-day operations, employing full-scale, NSP-treated Nova lenses, damage thresholds are indistinguishable from those of untreated surfaces. No Novette lens has been replaced because of surface damage. Production high-reflectivity turning-mirror surfaces have been satisfactorily tested to high damage thresholds (typically 8 J/cm^2). The

Fig. 2-1. Baseline layout for a single Nova chain.

Fig. 2-2. Baseline layout for Nova focusing system.

additional amplifiers in the chain provide higher gain downstream from the Faraday isolator components. They are, therefore, less threatened by high fluence levels and are no longer at risk in nominal system operation. Bulk damage in the KDP crystals is our greatest concern in limiting laser performance. We have performed a number of system simulations using the MALAPROP code,⁴ to map out the safe operating performance regime for Nova.

In our performance estimates, shown below, we have run many detailed system simulations assuming that a pulse with

a rectangular power vs time profile is incident on the frequency-conversion array. Our computational procedure assumes that pulse shaping is applied at the chain input. Figure 2-3 illustrates the salient features of the simulations corresponding to 1-ns performance. The data points mark the focus of 1ω beam energy at the entrance to the KDP crystal array and the peak fluence at 1ω seen by the KDP crystals at that energy. At low energy and fluence, the trajectory is essentially linear; peak fluence is dominated by diffraction effects only. At higher energy, nonlinear B -integral effects

Fig. 2-3. Energy available at 1ω and 1 ns to drive the frequency-conversion array as a function of peak fluence incident on the KDP crystals.

Fig. 2-4. Intensity transfer functions for the Nova array design.

Fig. 2-5. Expected on-target energies per beam at 1 ns.

appear, ultimately resulting in catastrophic self-focusing of modulation peaks if the laser system is driven too hard. Measured KDP damage thresholds are shown along the abscissa.

The MALAPROP code has been extended this year to incorporate detailed modeling of the frequency-conversion process itself. (For additional MALAPROP data, see "Incorporation of Frequency Conversion in MALAPROP," in Sec. 6.) Single point intensity transfer curves from 1ω to 2ω and from 1ω to 3ω are incorporated in the

code for the Nova array design and are shown in Fig. 2-4. The code then computes performance estimates for 2ω and 3ω , as shown for 1 ns in Fig. 2-5. Notice that 7 to 8 kJ at either 2ω or 3ω are readily available from one Nova beam without excessive damage to the KDP array.

The MALAPROP computations have been repeated at pulse durations of 0.1, 0.3, 1, 2, 3, and 5 ns, and the results are summarized in Fig. 2-6. The safe operating ranges of the laser system are shown in the colored regions, which span peak-to-average intensity (or fluence) ratios between 1.5:1 and 2.0:1. Nominal on-target performance, at any of the three wavelengths available, is summarized in Fig. 2-7. It is anticipated that the laser itself can work comfortably at or below these power vs energy curves without incurring significant optical damage over its operating lifetime.

Development of the sol-gel process⁵ for reducing Fresnel reflection losses on fused-silica lens and window surfaces has progressed rapidly. We have demonstrated workable coatings, of excellent uniformity in transmission, with damage thresholds exceeding $6 J/cm^2$ at 3ω and 1 ns, on quarter-scale optical pieces. Referring to Fig. 2-5, we see that this damage threshold is quite satisfactory. We have demonstrated on Novette that scatter-plate technology⁶ works at scale and full power to ameliorate diffraction modulation on the beam, when such modulation arises from sharp, opaque edges (such as the 46-cm disk split and the KDP array interstices). This apolization technique completely and satisfactorily supplants earlier, more costly methods of spatially smoothing the beam profile at its edges. Again on Novette, we have made a good beginning at measuring the saturation fluence ($4.0 \pm 2 J/cm^2$ in phosphate glass)

Fig. 2-6. Summary performance curves: (a) 1ω to KDP array, (b) 2ω energy on target, and (c) 3ω energy on target.

Fig. 2-7. Expected on-target energy at 1ω , 2ω , and 3ω . Safe performance at each frequency is anticipated at or below the curves. Performance at 1ω assumes that the KDP array is detuned but in place.

in a working system. At high output energies, such as those planned for Nova, realizable performance is linearly dependent on this parameter, which previously was measured only with small rod amplifiers. The results from Novette are in substantial agreement with these earlier measurements.

Author: W. W. Simmons

Nova System Activation

Goals and Strategy

During the last year, the Nova project has progressed from the construction of the building to the installation of the laser and target systems. Our goal is to complete construction of the Nova project before October 1, 1984. We have started to install and test most of the major systems. We have already activated the laser MOR, the preamplifier, the laser chain driver sections, and the power conditioning system.

Nova System Activation

Nova activation is managed from the Nova project office, under the direction of the Nova Project Engineer, the Deputy Project Manager, and the Nova Systems Engineer. Each subsystem project engineer is responsible for the assembly, installation, and activation of the items designed or procured by his group. The Systems Engineer sets and monitors the schedule for the delivery and installation of products, and coordinates the tasks performed by the project teams. A small group of engineers and technicians, under the direction of the Systems Engineer, forms the Activation Team. They integrate the components and subsystems into a working facility. Their tasks include laser component alignment, product testing, and performance assessment with functional tests.

The power conditioning system was assembled and tested by the Power Conditioning Group up through the junction boxes, to which the amplifier flashlamps are connected. The flashlamps are connected only after agreement between the Activation Team and the Power Conditioning Group. The initial firing tests are performed jointly by the Power Conditioning Group and the Activation Team. System operation and maintenance is supervised by the Power Conditioning Group.

Table 2-1. Nova project milestones with scheduled completion dates. A (C) following a date indicates that milestone has been completed.

Office building complete	June 1982 (C)
Begin spaceframe installation	July 1982 (C)
Begin installation of control room	August 1982 (C)
Complete installation of MOR frame	September 1982 (C)
Complete installation of two laser chains in test-bed	December 1982 (C)
Order array windows, focus lenses, KDP crystals	December 1982 (C)
Complete laser spaceframe installation	March 1983 (C)
Begin alignment systems installation in MOR	March 1983 (C)
Begin laser installation on Nova spaceframe	March 1983 (C)
Begin target chamber installation	June 1983 (C)
Begin delivery array windows, focus lenses, KDP crystals	July 1983 (C)
Project transition plan submitted for approval	September 1983 (C)
Complete target chamber installation	October 1983 (C)
Begin turning mirror installation	November 1983 (C)
Complete capacitor bank installation	December 1983 (C)
Begin focus lens/crystal installation	March 1984
Complete power conditioning installation	May 1984
Complete laser installation through power section (first eight arms)	May 1984
Begin installation of two arms from Nova test-bed	June 1984
Complete turning mirror installation	June 1984
Complete delivery array windows, focus lenses, KDP crystals	July 1984
Complete installation of two arms from Nova test-bed	September 1984
Complete focus lens/crystals installation	September 1984
Complete alignment systems installation	September 1984
Nova operable/project complete	September 1984
Approval of operational phase	September 1984

The Mechanical Engineering Group assembles, certifies, and installs the laser components on the spaceframe and connects all utilities. The Activation Team does the optical alignment and on-line optical evaluation, and directs the test firing of the components.

The Alignment Group assembles and prealigns the sensor packages and installs them on the laser spaceframe. The Activation Team and the Alignment Group work together to align the sensor package to the beamline. The Alignment Group then sets up the sensors, including determination of cross-coupling matrices and implementation of automatic alignment.

The Laser Diagnostic Group installs and checks the diagnostics detectors in the sensor packages, and installs the data recording and reduction systems to process the signals from the detectors. The Laser Diagnostics Group participates with the activation team in the first on-line testing and calibration of the data recording system.

Schedule

Nova project milestones are summarized in Table 2-1. The Systems Engineer has prepared more detailed schedules for component delivery and alignment, and for the progressive functional testing and calibration of the installed and aligned assemblies. Each of the Project Engineers has developed a detailed schedule to implement his part of the Systems Engineer schedule. The schedules are reviewed, and the project plans are updated periodically.

Present Status Summary

In the MOR (Fig. 2-8), the short pulse oscillator and the cw alignment laser are installed and operating. The switchout, isolators, amplifiers, and optics are installed to complete the beamlines for insertion of either the pulsed oscillator or the cw alignment laser into the laser bay (Fig. 2-9). In the laser bay, the preamplifier and splitter array are completely assembled to provide laser pulses into eight beamlines. Components in eight beamlines are installed and aligned through both the mid-chain sensors,

and the two 31.5-cm-diam turning mirrors that turn the beam at the end of the laser spaceframe. The active components in six of these beamlines have been connected to the power conditioning system, and the components have been fired. Pulse propagation to the input sensors and calibration of their energy measuring detectors is complete. We

have begun pulse propagation to the mid-chain sensor of the first two beamlines. We have also begun to survey the first beamlines in the power amplifier section, and the alignment crosshairs on the first six beamlines have been installed.

All of the alignment sensors in the MOR, preamplifier, and splitter array are installed

Fig. 2-8. Nova master oscillator room (MOR).

Fig. 2-9. Nova laser bay.

and operating. The input and mid-chain sensors for eight beamlines are installed and operating, and we have begun installation of output sensor 1ω modules. Automatic alignment of the cw beam through the preamplifier, splitter array, and input sensor of eight beamlines is operational. We have also implemented the automatic alignment of the pulsed beam.

We have installed and aligned the target chamber to receive the laser beams. We have also begun installation of the vacuum system and the target diagnostics data recording system.

The assembly and activation of the Nova system has progressed well during 1983. Current schedules and management systems should ensure completion of the Nova project by the end of FY 1984.

Author: D. R. Speck

Nova Frequency Conversion and Focusing System

Introduction

We used our experience with the Novette 2ω crystal arrays and the doublet lens positioner to refine the Nova $2\omega/3\omega$ frequency conversion and focusing optics systems. Novette hardware was used extensively to support more than 250 system shots, ranging in pulse width from 0.1 to 1 ns and with a maximum 2ω on-target power and energy of 6 TW and 4.5 kJ per beam. The crystal-array conversion efficiency was greater than 75% after transmission losses in the beam dump, focus lens, and debris shield were taken into account.

Initial array performance measurements made during 0.1-ns shots were in close agreement with our plane-wave computer model except at very high power where, as predicted, the 1ω beam began to develop severe spatial modulation due to nonlinear propagation effects. Our first shots at 1 ns were also in good agreement with theory, but we observed a rapid degradation of the HC-56 index-matching fluid. Both Novette arrays were disassembled, cleaned, and reassembled

without the fluid, using antireflection-coated windows. This configuration has operated without further degradation.

The fluid damage issue has been studied extensively using small-aperture test cells. Our new design is ready for full-scale testing in the first Nova array, scheduled for completion early in 1984. The index-matching fluid damage resulted in several Nova array design changes: new fluid, thicker fluid layers, and elimination of the large, precision aluminum eggcrate in favor of a discrete glass support. Because of 3ω operations, we also investigated possible nonlinear losses, such as transverse stimulated Raman scattering (SRS), which may occur in KDP and index-matching fluid layers.

In addition to the Nova $2\omega/3\omega$ design work, we developed a 4ω array and focusing optics system for the Novette laser. This system uses a Nova fused-silica singlet lens to focus the UV light on target. The 4ω wavelength is generated using a bare crystal array design, i.e., one without index-matching fluid. This design was originally developed as a backup to the Nova $2\omega/3\omega$ array in the event of serious problems with the index-matching fluid.

We gained valuable experience using the Novette lens positioner, which is similar to that employed on Nova. The Nova focusing optics system, which must accommodate 2ω and 3ω light, uses a fused-silica singlet lens instead of the Novette BK-7-glass doublet lens design. The hardware is basically identical to that in the Novette design, with minor changes to accommodate installation on the Nova target chamber. Unlike Novette, however, laser diagnostics, alignment aids, and KDP array supports are mounted directly to the lens positioner and are significantly different. The dynamic stability of this large mechanical system was evaluated numerically to help identify potential operational problems.

Crystal Arrays

In January, 1983, we assembled, tested, and installed two 74-cm-aperture KDP crystal arrays on Novette. The measured 2ω internal conversion efficiency vs input intensity is plotted for both the Novette south (5×5) and north (3×3) arrays in

Fig. 2-10. The solid theoretical curve assumes a single 1.8-cm crystal with 100 μ rad of net angular detuning error. Overall conversion efficiencies of the system, at target chamber center, can be obtained by multiplying the curve values by 0.74. The data show that both Novette arrays closely approached the optimal conversion efficiency achievable with an equivalent single crystal.

We evaluated the optical figure of the Novette arrays by observing their transmitted wave front on an 80-cm Fizeau interferometer (Fig. 2-11). The calculated distortion would result in a focal spot blurring of twice the diffraction limit. The distortion actually observed in the array implies a focal spot blur circle of $\sim 4.5\times$ diffraction limit. This difference between our theoretical and experimental findings can be attributed to departures from ideal boundary conditions in the actual array. Since the laser system operates at $\sim 10\times$ diffraction limit without the array, the additional $4.5\times$ distortion from the array contributes less than a 10% RMS increase in the minimum focal-spot size at high power.

Between January and May, 1983, we ran Novette at 100-ps pulse widths for a series of x-ray laser experiments. During this time, the performance of the index fluid-filled crystal arrays was excellent; no optical damage was observed. In mid-May, we changed the Novette operating pulse width to 1 ns. This change increased the energy density on the arrays by roughly an order of magnitude. Almost immediately, the arrays began to accumulate darkened areas in the 5- to 10- μ m fluid layers between the crystals and the windows. Figure 2-12 shows

the index-fluid damage in a Novette array. Small pieces of NG-1 absorbing glass were attached to the array in several places to prevent further irradiation of severely damaged sites.

The damage formed around a contaminant (e.g., a dust particle) trapped in the fluid layer. The contaminant would heat rapidly and cause local decomposition of the fluid. We had not observed this type of damage in smaller test cells. The decomposition created a radial distribution of trapped bubbles and debris, each of which was encircled by dark fluid residue. In subsequent exposures, this residue would act as an additional absorber and the

Fig. 2-10. 2ω conversion efficiency of Novette 3×3 and 5×5 arrays for input pulses of Gaussian temporal shape.

Fig. 2-11. (a) Novette 3×3 array transmitted wave front quality. (b) Computer simulation of the optical path difference.

process would repeat itself. The net effect was to spread the damage rapidly across the aperture.

Index-fluid damage required an immediate rework of both arrays to ensure their continuing operation on the system. We disassembled, cleaned, and reassembled both arrays without the index-matching fluid in the KDP/window interface. In order to reduce Fresnel reflection losses for the dry arrays, both surfaces of the input and output windows were NSP-antireflection coated. Small vacuum leaks in some array diagnostic windows were also discovered and repaired.

The wave front figures for the two dry arrays, shown in Fig. 2-13, indicated good optical quality. We expected the improvement in optical quality because the array optical elements, windows, and KDP crystals all acted in transmission: with the

fluid-filled arrays, only the two outer surfaces affected the wave front and they are distorted by the hydrostatic load.

We have closely monitored the performance of the reworked, dry arrays, and have observed no growth of previously damaged sites. The arrays have been exposed to over 100 high-energy shots, and their performance has been satisfactory in all respects since the rework.

The performance requirements for the Nova KDP crystal arrays are identical to those for the Novette arrays, except that the Nova arrays must be capable of generating both 2ω and 3ω wavelengths.

Based on our experiences with Novette, we added two new requirements to the Nova array design:

- A larger crystal/window gap must be employed to avoid the index-fluid damage observed on Novette. Also, the index-matching fluid choice must be reevaluated for Nova operating conditions.
- Metallic components should be eliminated from within the array clear aperture. This permits beam apodization directly on the array windows, and eliminates the need for a separate optical element and absorbing masks to protect metal components from direct exposure to incident or reflected beams.

We used a full-scale array test cell to simulate the optical characteristics of the Nova design. This test cell has individual cross-shaped glass spacers at the crystal-intersection points. The spacers are optically finished to be flat, parallel, and of identical

Fig. 2-12. Index-fluid damage in the Novette 5×5 array.

Fig. 2-13. Transmitted wave front quality of Novette arrays without index-matching fluid.

thickness. They define the window spacing and determine the optical figure of the completed array, while permitting the fragile crystals to remain in a relatively strain-free condition.

We compared calculated surface distortion contours with experimental interferograms (Fig. 2-14) to perform checks on our analytical model. A dark fringe occurs each time the array surface contour changes by $\lambda/2$. We were able to use our model with confidence to design the Nova arrays.

For acceptable optical performance on Nova, the slope errors in surface contour must be kept below about 13 fringes per 27-cm segment. Our measurements on this test cell demonstrate that the Nova baseline design can meet or exceed this criterion.

Nova arrays must operate in various orientations with respect to gravity. At angles as large as 50° with respect to the vertical. Experimental measurements for nonvertical attitudes are difficult because our 80-cm-aperture interferometer is configured for

vertical optics only. We relied on our analytical model to optimize array parameters (internal vacuum level, window thickness, window edge supports, etc.) for operation at orientations other than vertical.

Stress Birefringence Evaluation

As normal forces are applied to the array windows to compensate for gravity and hydrostatic loads, the windows experience high local stress concentrations in the regions around the cross-shaped glass supports. Stress concentrations induce local changes of the index of refraction in the array windows. The resulting birefringence depolarizes the incoming light, potentially reducing harmonic conversion efficiency. An example of this birefringence effect is shown in Fig. 2-15. With no stress applied to the disk, the background is dark. As stress levels increase, the disk becomes increasingly birefringent around

Fig. 2-14. (a) Computer simulation of the optical path difference in full-scale test cell ($\lambda = 0.6328 \mu\text{m}$, contour spacing $\lambda/2$). (b) Reflected wave front interferogram.

Fig. 2-15. Stress birefringence in a 1.3-cm-thick fused-silica window.

Fig. 2-16. Apodizer test pattern. Note strong diffraction pattern of diagonal wire.

the stress, producing the bright regions in the photographs. Within these bright regions, degradation in conversion efficiency can be expected to occur. However, the results of our measurements, together with an analytical model of stress-birefringence distribution, indicate that this effect should be small for the Nova arrays. The decrease in conversion efficiency should be $<1\%$ for the typical case, and should not present any serious limitations on the array performance.

Apodization

We used phase-scattering apodizers on Novette to minimize hard-edge diffraction from the 46-cm amplifier-disk split and crystal-array gap regions. We used a temporary prototype facility to fabricate these apodizing elements.

We recently completed installation of an upgraded, phase-scattering apodizer system that uses a motor-driven x - y translation stage with a small nozzle through which abrasive material is forced to apply patterns onto optical substrates. The system is equipped with a high-velocity, high-volume vacuum recovery system for minimizing dust contamination of the optics, environment, and personnel. We designed this version of the system specifically to produce the linear stripe patterns required for apodizing the Nova arrays. Stripe widths and roll-off are controlled by varying travel rate, nozzle diameter and height, and abrasive size and feed rate. We are conducting tests to optimize the parameters for producing the Nova array window apodized pattern, which is basically a pair of straight

lines separated by one crystal width (27 cm), intersected at 90° by a second pair, similar to a tic-tac-toe pattern. Spot patterns are required to apodize the regions around the array discrete-support points. The system generates straight lines by traversing in the x -direction with y fixed or vice versa. The system can generate spot patterns up to ~ 30 mm in diameter by operating the nozzle at a fixed height and x - y position.

A test pattern with two orthogonal apodized lines and an apodized spot over the intersection point is shown in Fig. 2-16. The stripe width is approximately 6 mm and the spot diameter was roughly 25 mm. A wire on the diagonal illustrates the diffraction ripples produced by a hard edge as concentrated with the reduced ripples from the apodized pattern edges.

Index-Matching Fluids

The index-matching fluid damage observed in the Novette arrays prompted a review of the Nova array design. We performed tests at 1ω and 2ω , and found that a fluid layer of ~ 0.5 mm of HC-56 did not suffer the damage described above. Nonetheless, particles trapped in the fluid layer would still decompose, forming bubbles and debris, which would float out of the aperture. At 3ω , the HC-56 decomposed, independent of particulate contamination, and attacked the KDP surface. For this reason, we terminated further tests on HC-56.

Our initial interest in HC-56 was based on its very low stimulated Raman scattering (SRS) cross section. A new liquid, propylene carbonate, was recommended by the Laboratory for Laser Energetics at the University of Rochester for its damage resistance to 3ω light. The SRS gain of this fluid was measured and found to be approximately 1.5 times higher than KDP. Ethylene carbonate, a solid at room temperature, was dissolved in the propylene carbonate to bring the Raman gain to 1.3 times that of KDP. The precise ratio of components was determined by balancing the different Raman lines to yield a reduction in the gain of the strongest lines.

The primary threat to the Nova crystal arrays from SRS is the risk of internal damage from intense sidescattered light.

Growth of the Stokes wave intensity, I_{SRS} , in the small-signal regime is given by the following equation for a plane wave:

$$I_{\text{SRS}} = I_0 \exp(g_{\text{SRS}} I_p l) \quad (1)$$

I_0 is defined as the equivalent source intensity. I_0 is probably in the range of $0.001 \leq I_0 \leq 0.1 \text{ W/cm}^2$, depending on geometry.

In this transverse pump geometry, the gain length is defined by the pump time, τ :

$$l = c\tau/n \quad (2)$$

where c is the speed of light and n is the index of refraction. Therefore,

$$I_{\text{SRS}} = I_0 \exp(g_{\text{SRS}} F_p \tau / n) \quad (3)$$

where F_p is the mean pump fluence (J/cm^2). Solving for pump fluence, we obtain

$$F_p = \frac{n}{c g_{\text{SRS}} \tau} \ln \left(\frac{I_{\text{SRS}}}{I_0} \right) \quad (4)$$

The onset of intense SRS can be described by a fluence limit. Note the logarithmic dependence on I_0 , which renders F_p relatively insensitive to its precise value.

If we assume that the array performance is limited by damage of the internal structure exposed to sidescattered light, then we must limit the maximum value of I_{SRS} , for example, $I_{\text{SRS}} \leq 0.1 \text{ GW/cm}^2$ for all τ . Also, if we assume that $I_0 = 0.1 \text{ W/cm}^2$, the pump fluence would be limited to

$$F_p < \frac{1}{g_{\text{SRS}}} \quad (\text{J/cm}^2) \quad (5)$$

where g_{SRS} is in units of cm/GW .

Estimates of g_{SRS} at the harmonic wavelengths for the Nova index-matching fluids and for KDP are less than 0.25 cm/GW , which yields a pump fluence limit of $<4 \text{ J/cm}^2$. This value is comparable to the maximum Nova design fluence. In the event that the SRS light is more intense than expected, we can install absorbing glass between the $27 \times 27\text{-cm}$ crystal segments to reduce the maximum transverse dimension. We will carefully monitor the sidescattered light for this problem in the first Nova array, to be tested on Novette.

Novette 4ω System

The 4ω crystal array employs a 5×5 aluminum eggcrate type-II doubler crystals from Novette mounted on the input side, and new, type-I quadrupler crystals on the output side. The eggcrate was machined using electric-discharge machining technology similar to that used on the original Novette eggcrate. However, the faces of this eggcrate were conventionally machined flat to $\sim 30 \mu\text{m}$ instead of diamond turned. Crystal-to-crystal alignment is maintained by referencing these machined surfaces. The surface was also wedged about 4 mrad to prevent crystal-surface back reflections from propagating down the laser and diagnostic systems. This array will be used on Novette for target experiments, and to demonstrate the viability of the Nova backup array design.

Nova Lens Positioner

The Nova lens positioner is based on the Novette design described in the 1982 *Laser Program Annual Report*. The major differences are:

- We substituted a single, fused-silica lens for the two-element BK-7 lens. This allows operation at 1ω , 2ω , and 3ω , and simplifies the mechanical design by eliminating the secondary drive mechanism.
- The Nova lens positioner is supported from the target chamber. The Novette lens positioner is supported by a rail and linear bearing system.
- We have anticipated the need for single-shot, disposable plastic debris shields, as well as multiple-shot fused-silica shields.

The single element $f/4.3$ Nova fused-silica focus lens is designed to operate safely with surface reflectivities of $R < 0.7\%$ per surface at 1ω , 2ω , and 3ω . If the lens is uncoated, or if the antireflection coatings degrade to more than 1% per surface, ghost focus damage is expected to occur on-axis, from both the incident laser beam and target reflected light.

We have developed a baffle design to block the potentially damaging ghost reflections in case the surface reflectivity exceeds 0.7% . Figure 2-17 shows a ray trace of the

Fig. 2-17. Annular groove and bead-blasted (roughened) area protects lens from damaging ghost focused rays.

axial ghost due to the incident laser beam. After several reflections, the rays from 0.7 to full aperture are trapped by total internal reflection. After five reflections, these rays come to a focus near the front surface of the lens at a ray height of about 80 mm. An annular groove 20 mm wide, and 15 mm deep, with an outer radius of 81 mm, intercepts these rays and harmlessly absorbs, scatters, and refracts the energy away from the lens axis. The nearly flat rear surface of the lens can be apodized using the phase-scattering technique to frustrate the target

reflection ghost in an annular region from 10 to 15 mm in diameter.

The fused-silica focus lens is also a vacuum barrier. Significant stress exists in the groove region, with a maximum compressive stress of 1247 psi (8.6×10^6 Pa) calculated for the groove bottom. We plan to pressure-test each lens to 1.5 atm to verify that each is free of internal defects that might lead to failure under load.

We have analyzed the stresses and deflections due to both static and dynamic loads on the lens positioner as installed on the Nova chamber. Finite-element calculations using the GEMINI code with an 0.004-g white noise input at the target bay floor show maximum pointing errors of less than $10 \mu\text{rad}$ due to random vibrations. Measurements of the actual ambient noise spectra in the Nova target bay are under way. Static deflections are well within the lens positioning adjustments available.

A fluorohydrocarbon film about $12 \mu\text{m}$ thick will be used as a disposable debris shield on routine low-energy experiments. A remote-controlled reel mechanism will move this film across the aperture as required. In addition, a mounting scheme incorporated into the Nova lens positioner allows the installation of a 10-mm-thick, high-quality fused-silica debris shield for experiments at high energy.

Nonhorizontal beamlines in Nova require careful control of particles so that contamination of optical surfaces is minimized. The translation mechanisms in the Nova lens positioner is shielded, wherever possible, to trap any particles generated by the mechanisms; access doors and cable openings are carefully sealed to minimize contamination from external sources.

Authors: M. A. Summers, B. C. Johnson, J. D. Williams, and L. G. Seppala

Optical Components for Nova

Introduction

The highlights of 1983 in optical components and materials are:

- Production of sufficient good-quality KDP material to supply all our Nova KDP plates.
- Development of a single-wavelength, low-reflectivity sol-gel coating for fused silica with a high laser damage threshold (9 J/cm^2).
- Observation of no laser-induced damage on NSP coatings in Novette over a full year of operation.
- Receipt of about 75% of the total Nova optics ordered.

Laser Glass

Hoya Optics, Inc., Schott Glass Technologies, Inc., and Kigre, Inc., are supplying the phosphate laser-glass disks for Nova. By the end of 1983, they had completed melting, edge-cladding, and pre-finishing about 86% of the 617 Nova disks; 481 disks (about 78%) have been optically finished by Eastman Kodak, Inc., and Zygo Corporation.

Fused Silica

The fused-silica procurement has proceeded according to plan.⁷ We have received 15 lens blanks (15 ordered) and 13 window blanks (15 ordered) from Corning, and 9 window blanks (15 ordered) from Heraeus. All the material has met our requirements. Because of manufacturing difficulties with large fused-silica blanks, homogeneity specifications were somewhat looser than we would have liked. Looser specifications implied the probable need for compensating figuring. However, Corning improved homogeneity so that their blanks are now better than the specifications, making the optical finishing easier.

We chose 80-cm-diam fused-silica windows as Nova baseline debris shields. These shields can be thin, because they are free-standing, parallel windows. Because of smaller volume and reduced homogeneity requirements due to the small optical path, thin windows reduce cost. (Thinness also minimizes nonlinear optical effects.) Fused-silica material for debris shields costs less than $\$0.80/\text{cm}^3$. The homogeneity and

Inclusion cross
section (mm^2)
per 100 cm^3 of
glass

	Homogeneity	glass
Lenses and windows	2×10^{-6}	0.03
Debris shields	5×10^{-6}	8.00

inclusion specifications for the Nova fused silica are summarized in Table 2-2.

Filters

There are three large filters in the Nova design:

1. A 27-cm-sq central beam dump, just after the KDP array, that protects the target from 1ω irradiation during 2ω shots.
2. A 40-cm-diam filter in front of the opposed-port calorimeter; the filter absorbs 1ω light and allows 2ω light to reach the calorimeter.
3. An 80-cm-diam attenuator in front of the RBD mirror; the attenuator absorbs 95% of 3ω light and 75% of 2ω light in double pass, and protects the optics in the RBD package from damage and solarization. The attenuator will be made from Nd-doped BK-7.

We ordered a Cu-doped phosphate glass, DFF-1, from Hoya for the first two filters. The glass transmissions are 89% at 2ω and 4% at 1ω .

For 3ω shots, we attempted to identify a filter that absorbs 1ω and 2ω and transmits 3ω . Unfortunately, all the materials we tested have two-photon absorption or solarization at fluences above 0.5 J/cm^2 . This makes the filters unusable in places where transmission stability is important. We will separate 3ω from 1ω and 2ω by other methods.

KDP Growth and Production

Cleveland Crystals and Interactive Radiation achieved three important goals in producing semi-finished, 27-cm-square KDP blanks:

1. Cleveland completed their preproduction order of 28 full square blanks in early December 1983. Interactive Radiation should complete their preproduction contract in January 1984.

Table 2-2. Nova specifications for fused silica blanks.

Fig. 2-18. COMS data output showing (a) $\sim 6500\text{-}\mu\text{rad}$ mismatch of test and reference sample axes, and (b) mismatch reduced to $\sim 1.2\text{ }\mu\text{rad}$ by diamond turning of test sample.

Optical Components for Nova

2. Cleveland Crystals achieved a delivery rate of 12 blanks per month in December 1983.
3. Both vendors (with technical assistance from LLNL) have significantly improved their manufacturing capability.

Cleveland Crystals has sufficient tank capacity to supply all the Nova crystals. They doubled their fabrication capability by adding staff and machinery to allow parallel operations on all the manufacturing. The quality of their blanks meets or exceeds LLNL specifications in all regards. Interactive Radiation experienced difficulties in their early growth runs. They have made major changes in their growing methods, and growth runs now look very good.

KDP Final Finishing

In January 1983, we successfully negotiated contracts with Cleveland Crystals and Pneumo Precision for the final finishing of Nova KDP blanks by diamond turning. Cleveland Crystals received 60% of the Nova production quantity and Pneumo Precision Optics West Division received 40%. The common factors of the contracts are:

- The final finishing of 27-cm-square $\times$ 1.0-cm-thick blanks is to LLNL specifications.
- An LLNL-owned diamond turning machine (oil showered for temperature stability), is located at each vendor.
- LLNL-owned, 18-in.-diam interferometers are located at each vendor.
- Phase-matching equipment for final crystal orientation is located at each vendor.
- Adequate facilities for cleaning, handling, measuring, and storing finished crystals is at each vendor.

We allowed each vendor to select the basic machine configuration they deemed best, consistent with our 1982 qualification program.

We are using two crystal orientation measurement systems (COMS), designed and built by the Laser Diagnostic Group,⁸ to orient Nova $2\omega/3\omega$ crystals at the vendors. Figure 2-18(a) shows typical output data of the COMS, indicating that a crystal correction cut is required as measured by the difference between the standard crystal and the test sample peaks. In this example, a $6497\text{-}\mu\text{rad}$ correction was indicated. After diamond turning, a rescans of the crystal [Fig. 2-18(b)] shows that the error has been reduced to $1.2\text{ }\mu\text{rad}$, which is well within the $\pm 30\text{ }\mu\text{rad}$ tolerance.

Mirrors

Our two manufacturers, Zygo Corporation and Eastman Kodak, have finished a total of 88 mirrors, including all mirrors from 60.4 through 94 cm diam. The finish of all the mirrors meets or exceeds our specifications. To date, we have not lost any blanks because of breakage. Both Zygo Corporation and Eastman Kodak are working to finish the 109-cm mirrors by March 1984.

Focus Lenses

In 1983, we began to manufacture the 80-cm-diam, single-element, fused-silica focus lenses. Because of the difficulty of manufacturing aspheric lenses of this size, we split the focus lens order between the Applied Optics Division of the Perkin-Elmer

Corporation and the Tinsley Corporation. Perkin-Elmer has completed the first three lenses. Figure 2-19 shows a double-pass interferogram of a 77-cm-diam fused-silica aspheric focus lens at 633 nm. Perkin-Elmer used a null lens to correct for the spherochromatic deviation from the nominal design wavelength: the maximum deviation of the aspheric surface from the best-fit spherical surface is 135 μm .

Coatings

We devoted 1983 to the production of optical coatings on Nova optics ranging in size from less than 1- to 94-cm-diam. Optical elements in the Nova chains, switchyard, and target room that require optical coatings are listed by aperture in Table 2-3. The quantities listed do not include spare substrates that serve as backup for the laser optics.

The diagnostic (first turning mirror) design has spectral reflectance values of $96.5 \pm 1.5\%$ at 1.05 μm and $50 \pm 20\%$ at both 0.53 and 0.35 μm . A spectral reflectance curve for this mirror design is shown in Fig. 2-20.

All of the Nova diagnostic mirrors have been coated. The return-beam diagnostic (RBD) mirrors have also been coated, but measurements are not complete.

Eight large aperture target-chamber turning mirrors have been produced. The damage threshold of these mirrors has consistently exceeded 8 J/cm^2 .

Neutral Solution Process (NSP)

The NSP lenses and windows in Novette have demonstrated high resistance to laser-induced damage. We have found almost no surface damage on NSP surfaces during the year of Novette operation. However, in late 1982, we observed changes in reflectance on the vacuum side of the Novette spatial filter lenses (Fig. 2-21). The visual effect was an initial increase in reflectance, which stabilized after some days and remained at the increased level. In situ measurements at 633 nm on the 31.5-cm-aperture output lens in the Novette north arm showed a 3% reflectance on the vacuum side and 0.8% on the air side of the lens. Similar measurements on the

46-cm-aperture lens showed 1.3% and 0.9% reflectances, respectively. After NSP

Fig. 2-19. Double-pass interferogram of a 77-cm-diam aspheric focus lens at 633 nm.

Table 2-3. Large-aperture coated optics in Nova.

Component	Diameter (mm)	Aperture (mm)	Nova qty.	Qty. delivered
5-cm rods	50	42	16	8
Tab 01 rotators	100	95	10	8
Tab 02 rotators	166	156	10	8
Tab 03 rotators	220	210	10	8
Tab 04 rotators	331	321	10	14
Tab 04A mirrors	604	576	10	13
Tab 04B mirrors	604	509	10	13
1.5% beam splitters	420 × 240	218	10	11
Tab 10 polarizers	103 × 60	50	60	40
Tab 01 polarizers	196 × 115	100	40	47
Tab 02 polarizers	330 × 190	160	10	13
Tab 03 polarizers	420 × 240	218	20	20
Tab 04 polarizers	700 × 400	335	20	16
RBD mirrors	820	790	10	4
Diagnostic mirrors	940	908	10	—
Turning mirrors	940	900	24	—
Turning mirrors	1090	980	4	—
Objective lens	800	770	10	—

Fig. 2-20. Calculated spectral reflectance of a 94-cm diagnostic mirror at 22.5 deg from surface normal.

Fig. 2-21. Spectral transmission of neutral solution processed BK-7 before and after vacuum exposure.

processing, these parts had no difference in reflectance on the two surfaces.

We undertook a series of tests on two 52-cm-diam spatial filter lenses from Novette to determine the cause of this effect. We rinsed selected areas of the two lenses with methylene chloride solvent. Gas chromatography/mass spectrometry (GC/MS) and infrared spectroscopy analysis of the solvent clearly differentiated between the two sides of the lenses and showed that the surface contamination was much heavier on the lens side exposed to the atmosphere than on the vacuum side. Samples from the vacuum sides of the two lenses were similar and showed traces of siloxanes and phthalates. Samples from the air side had very similar total ion chromatograms and showed much larger quantities of phthalates.

Frustrated-total-internal-reflection spectra of the residues from the samples confirmed the difference in the two sides of the lenses. The residue in air-side samples was mostly phthalates, plus a small caprolactam absorption. The residues from vacuum-side samples produced very similar composite spectra, but only caprolactam and phthalates were clearly identified.

We then conducted four controlled vacuum tests on thirty 2-in.-diam BK-7 pieces

that were all NSP processed in one run. The intent of these tests was to determine whether an NSP surface exposed to vacuum loses transmission.

Test One. Eight 2-in.-diam witness pieces maintained at 68°C for 200 h in a clean vacuum system had no loss in peak transmission. We prebaked the vacuum chamber to 100°C wall temperature using silicone-rubber-encased heating tapes. The residual gases released during baking were largely of mass 17 (OH^-) and 18 (H_2O). However, the minimum transmission values increased by as much as 0.5% after exposure to vacuum bake.

Test Two. We maintained seven BK-7 witness pieces in a clean vacuum at 22°C for 500 h. The results were the same as in test one.

Test Three. We weighed three BK-7 witness pieces (two NSP-processed and one unprocessed) before and after exposure to vacuum for 16 h. The unprocessed BK-7 piece showed no weight change, while the two processed pieces each exhibited a weight loss of 0.5 ± 0.1 mg. The NSP pieces also showed the same increase in transmission as previously observed at the transmission minimum. Figure 2-21 shows the transmission of a typical NSP witness piece before and after vacuum exposure.

We used the equation for the reflectivity maximum of a simple linear ramp index profile⁹ to show that a lowering of the film index at the air interface will decrease the reflectivity. This is the case when water (index 1.3) is boiled out of the pores and replaced by air (index 1.0). Thus, we believe that the increase in the transmission in the valleys is due to boil-off of water.

Test Four. We placed eight NSP samples and one control sample in a vacuum chamber and pumped it down to 1×10^{-8} Torr. After 500 h, we removed two samples; neither sample had a loss in transmission. After removing two more samples from the chamber 1000 h later, we again found no transmission loss. The final two samples remained in vacuum for an additional 1000 h, for a total of 2500 h. They had no measurable loss in transmission.

These tests showed conclusively that NSP parts exposed, even for long periods, in a clean vacuum do not degrade. Presumably, contaminants existed in the newly constructed spatial filters or were

introduced during pumping. The incidence of this effect had fallen markedly by the end of 1983.

Sol-Gel

We continue to develop sol-gel coatings for fused silica by using the Westinghouse process. With Westinghouse's continued assistance, we installed the baseline processing system and coated a 25-cm-diam test optic. Most of our work this year, however, concentrated on establishing the parameters to obtain the desired spectral characteristics together with high damage levels at 350 nm. Our goal is a graded-index, pure fused-silica coating that will reflect no more than 0.7% per surface at 1ω , 2ω , and 3ω , and have a damage level of at least 5 J/cm^2 delivered in 1 ns at 350 nm.

The basic process is to prepare a solution of tetraethyl orthosilicate (TEOS), H_2O , ethanol, and a small amount of HNO_3 , to form a polymer. To produce a graded index, the polymer is uniformly deposited on the fused-silica substrate by draining or withdrawal, allowed to dry, baked at approximately 500°C to drive off the organics and reduce the pore size, and acid etched to taper the pores.

To achieve an effective broadband coating, the coating must have a minimum optical thickness of at least one-half of the largest wavelength of interest. If the porous coating has an average index of refraction of, for example, 1.22, the thickness required for 1053-nm wavelength is therefore about 430 nm. Our initial efforts to produce coatings this thick and with high damage resistance generally resulted in low damage thresholds. The problem was apparently caused by trapped organics near the substrate surface. We discovered, however, that with properly adjusted polymer, we could repeatedly obtain coatings which produced a spectral transmission characteristic of a quarter-wave coating at 350 nm, and had high damage threshold (9 J/cm^2). We estimate these coatings to be about 70 nm thick.

We developed a high-damage, broadband, three-layer coating by treating between each layer deposition with ammonium nitrate to further oxidize the organics. The spectral characteristics of this

Fig. 2-22. Spectral transmission of a single-layer, sol-gel coating about 70 nm thick. Laser damage threshold is about 9 J/cm^2 at 0.6 ns and 350 nm.

Fig. 2-23. Spectral transmission of a 450-nm-thick sol-gel coating consisting of three successive layers. Laser damage threshold is 6 J/cm^2 at 0.6 ns and 350 nm.

coating are shown in Figs. 2-22 and 2-23. This is one possible approach to producing high-damage, broadband coatings. We are pursuing others.

Nova Focus Lens Ghost Baffle

The single element $f/4.3$ Nova fused-silica focus lens is designed to operate with surface reflectivities of $R < 0.7\%$ per surface at 1ω , 2ω , and 3ω . Ghost focus damage is expected to occur on-axis, from both the incident laser beam and target-reflected light, if the lens is uncoated or if the antireflection coatings degrade in time to more than 1% per surface.

Authors: S. E. Stokowski, F. T. Marchi, L. G. Seppala, E. P. Wallerstein, and G. R. Wirtenson

Major Contributors: N. J. Brown, R. C. Duffus, G. C. Sclipa, C. D. Walmer, and W. T. Whistler

Front End Subsystem

Introduction

The Nova front end consists of the MOR preamplifier section and the splitter array. The front end provides optical pulses with selectable duration for input to the Nova amplifier chains.¹⁰ The MOR must also generate synchronized trigger signals for the rest of Nova. The MOR and the preamplifier section provide alignment beams, path length measurement, and adjustment for the Nova arms. The MOR will also be able to generate short duration (~ 20 ps) pulses for target diagnostic applications.

The devices now operating in the MOR are a short-pulse oscillator and switchout, two 10-mm amplifiers, a cw YLF laser, and associated alignment and diagnostic systems. All components are installed and operational in the preamplifier section except the components associated with the two arms from Novette. We are using the laser bay cw YLF laser and the MOR cw laser for alignment and gain measurements; the short-pulse oscillator is being used only in gain measurements.

Pulse Shaping with Silicon Switches

To generate optical pulses, we pass light from a long-pulse oscillator through a Pockels cell gate that is driven by a shaped high-voltage pulse. To create a properly shaped optical pulse, we need a properly shaped high-voltage pulse. Our pulse-shaping technique, which uses passive reflection from a transmission line with nonuniform impedance, has the inherent advantage that the pulse shape may be easily changed by changing the impedance characteristics of the transmission line. The principal obstacle to realizing this scheme is the limited high-voltage holdoff capability of present silicon switches.

A schematic of the voltage pulse-shaper is shown in Fig. 2-24. Initially, silicon switches S1, S2, and S3 are in high-resistance state. Transmission line A is charged to high voltage. Line C is a section whose impedance varies with distance according to the desired pulse shape. When the leading edge of the pulse encounters line C, a reflection will be created, traveling back toward S1. S2 is illuminated, and the reflected pulse passes unimpeded through

Fig. 2-24 Silicon switch pulse-shaping device.

switches S1 and S2 (the switches turn off very slowly). The voltage driving Pockels cell (PC1) will consist of the desired signal (the reflected pulse passed back through S1 and S2) and other signals, including the reflection from the falling edge of the first square pulse. To eliminate the light passing through PC1 due to the spurious signals, PC2 is driven by a square pulse precisely synchronized with the shaped pulse. Switch S3 is illuminated with a fraction of the same optical pulse that illuminates S1. This segment of the voltage pulse shaper is the same as the square-pulse gate normally used with the long-pulse oscillator.

We tested the experimental prototype of the shaped voltage pulse generator at half the expected operating voltage. Figure 2-25 shows an oscilloscope trace of the output pulse. The reflecting transmission line consists of four sections of different impedance. The four levels of this stepped pulse correspond to the four sections of line. The impedance of the line can be varied to produce continuously rising output pulses. We used a 100-ps Gaussian optical pulse to illuminate the two silicon switches. This test demonstrated the ability to produce shaped-voltage pulses.

We are working with the Engineering Research Division (ERD) to improve the high-voltage holdoff ability of the silicon switches. We are testing different doped and undoped silicon materials, contact types, and contact geometries. ERD fabricates the devices and we test them with and without laser illumination. ERD also provides expertise in semiconductor physics and aids in interpreting test data.

The crucial factor in high-voltage holdoff is the characteristics of the switch material when only voltage and no light is applied. To test high-voltage holdoff, we subject the switch chips to a 600-ns-long square voltage pulse of from 50 V to 3 kV. At these voltages and this time duration, nonlinear effects that could lead to breakdown are apparent. We measured switch parameters important to the operation of the pulse shaper (such as contact resistance and carrier lifetime), using pulsed illumination synchronized with applied voltage pulses. Switches made from gold-doped material were used successfully in Novette at 10 kV, and have the best voltage holdoff

Fig. 2-25. Oscilloscope trace of output pulse.

capability. Gold-doped switches also have the most linear current-voltage characteristic.

Another factor that strongly affects the operation of the shaped-voltage pulse generator is the reflectance of the reflective transmission line. Tests of transmission line structures indicate the reflection coefficient of the line can be increased to achieve desired output amplitude. One of the main advantages of this pulse-shaping scheme is that pulses with very high contrast ratios can be created if the impedance variations are very small, but precisely controlled.

Activation

With very few exceptions the design of the Nova MOR is identical to that of Novette's MOR. Nova MOR output requirements are pulses with widths < 100 ps to > 5 ns with better than 10^7 prepulse attenuation.¹¹ The switchout provider: pulse selection and prepulse suppression for the short-pulse oscillator; the slicer unit provides a carved pulse from a larger, Q-switched pulse from the long-pulse oscillator. The pulse is amplified by selected 1-cm rod amplifiers, then relayed and imaged through spatial filter SF1 (shown in Fig. 2-32 in the 1982 Laser Annual) and up through the periscope to the preamplifier section. Isolation between amplifiers is done through Pockels cell isolator units. Two cw alignment lasers are available for beam alignment through the amplifier chains: one is situated in the MOR, the other is in the laser high bay. A schematic for the high-bay cw alignment system is in Fig. 2-26.

We generate specified pulse durations with both a long- and a short-pulse oscillator. Provisions exist to construct four more

Fig. 2-26. High-bay cw alignment system.

Front End Subsystem

Fig. 2-27. Measurement device for 1-cm rod amplifier gain.

oscillators for pulse shaping and 20-ps probe activation. The short-pulse oscillator is a Q-switched, modelocked Nd:YLF laser with the rod situated in a double ellipse pump cavity. Building the Q-switch and modulator assemblies (which required the bonding of electroded LiNbO_3 transducers to the respective quartz substrates) were critical steps in constructing the short-pulse oscillator. The units were impedance-matched to $50\ \Omega$ with a capacitive-inductive network. The rest of the oscillator construction is purely a mechanical assembly from appropriate drawings. We activated the short-pulse oscillator first to complete gain

measurements and calibrations through the laser chains.

We brought two Quantronix cw alignment lasers online: one in the MOR, and the other in the Nova high-bay. The MOR cw laser was one of the first components activated so that the beam could be used to establish the alignment of the other components in the MOR beamline. Activating the MOR cw laser also enabled us to align the preamplifier section and main chains in the laser bay. We used the MOR cw to do cw YLF rod-gain characterizations on all rods before accepting them from vendors. We use the cw YLF laser in the laser bay for

beam alignment when the MOR cw beam is not available. We use the additional alignment laser to work in more than one area of the laser bay at the same time. A telescopic lens system is set up near the laser output to expand the beam to the desired diameter of 34 mm. Wave plates and polarizers then determine the direction in which the beam is propagated.

The switchout and slicer units, assembled following cw installations, consist of two 10-mm longitudinal Pockels cells separated by Brewster polarizers to give pulse selection and prepulse attenuation of 10^7 . We use the isolator units, PC1-1 and PC1-2, to isolate amplification stages. Both PC1 and PC2 (shown in Fig. 2-32 in the 1982 Laser Annual) are single-stage Pockels cell devices, and use the same 10-mm-diam Pockels cells as the switchout. The two isolators give an additional 10^7 prepulse attenuation. We used crossed polarizers to test for light extinction ratio.

Each of the rod amplifiers is a 10-mm-diam, 15-cm-long, high-gain phosphate rod in a double ellipse cavity pumped by two krypton arc lamps. All amplifiers were assembled in the Nova clean room, except the 10-mm amplifiers, which were assembled and tested in the MOR. The amplifiers are water cooled, and all water system fittings are stainless steel, plastic, or nickel to avoid copper contamination.

Nova 1-cm Rod Amplifiers

During our tests of Nova MOR components, we measured the on-axis gain and the gain profile of a 10-mm rod amplifier using the arrangement shown in Fig. 2-27. We used a corner cube to scan the beam across the rod and a 1-m focusing lens to provide a small sampling area in the rod. This is a convenient way to measure the gain profile in a small-diameter rod.

The upper curves in Fig. 2-28 show the measured on-axis gain, which reaches more than 100 with only 0.7 kJ flashlamp pump energy. The curve labeled "clear flow tube" in Fig. 2-29 shows the measured gain profile in the x direction perpendicular to the plane through the lamps and the rod. The gain drops off very rapidly from the center, and is down by almost a factor of 2 at the edges. Similar poor gain distribution was

also present in the Novette amplifiers, and resulted in a distorted beam if the beam was not well centered in the rod. The gain profiles of the Nova prototype 1 cm glass rod have similar shapes, but the gain variations are smaller.¹¹

The pump geometry in the amplifiers is a double ellipse, with the lamps and rod exactly on the foci, producing strong focusing of the flashlamp light in the rod. To improve the gain profile, the rod-doping density can be changed, or the ellipses can be changed to defocus the pump light in the rod. We frosted the flow tube around the rod (see detail at upper right in Fig. 2-27) to scatter the light going into the rod

Fig. 2-28. Optical gain vs pump energy for original pump and frosted flow tube.

Fig. 2-29. Optical gain vs position (x scan).

Power Systems Energy Storage

This resulted in some sacrifice in efficiency but greatly improved uniformity. The curves labeled "frosted flowtube" in Figs. 2-28 and 2-29 show the results. The minimum required gain of 30 can still easily be achieved over the center 60% of the rod, and the gain variation is within $\pm 10\%$. This change should result in a more uniform beam out of the MOR, and should relax the stringent alignment requirements for the 10-in.-amplifiers.

Authors: R. B. Wilcox, D. F. Browning, P. H. Chaffee, and R. Merrill

Major Contributors: D. J. Kuizenga and W. W. Simmons

Power Systems Energy Storage

Introduction

During the past year, the energy storage system for Nova was installed and activated through the first eight arms. The controls for the pulsed power were installed and made operational early in the activation of the bank, and the controls software was re-organized and simplified.

Capacitor Bank

The assembly of the 60-MJ capacitor bank for the Nova laser system is almost complete. The first eight arms of the Nova bank have been finished. The bank for the two arms from Novette and for additional amplifiers is scheduled to be added in mid-1984.

Bank Activation

The checkout and activation of a large capacitor bank such as Nova is a formidable task. We were successful in activating the Nova bank by adhering to basic quality control procedures.

We first performed acceptance testing on critical components, such as capacitors, ignitrons, and inductors. If we found a component to have a high failure rate, we

inspected all similar components. For example, each of 2645 high-density capacitors was subjected to 500 pulse discharges at 10% above rated voltage to remove infant mortality failures from the population.¹² If an item was off the shelf, lot acceptance techniques were used.

The second step was to test the capacitor modules themselves before placing them on the capacitor racks. Each module received 25 charge and discharge cycles. During this test, the capacitors in each module were discharged through the pulse-forming inductor into the dummy load on each module.

After all the modules were tested and installed on the racks, and all the cables pulled and connected, the third step was to check the wiring. A point-to-point resistance check was made of all the high voltage cables. Each circuit has two cables: a charge cable from the switch to the module, and a load cable from the module to the junction box. Also, each switch assembly has a cable from the power supply high-voltage distribution rack to the switch assembly.

In the fourth step, each cable was checked three ways: shield to ground, center conductor to ground, and center conductor to shield. Each check consists of a 1-min high voltage test to 25 kV dc. The cables were checked out switch by switch, because in a switch assembly, up to 30 cables are clustered together. During these checks, the capacitors in the modules are shorted.

The fifth step was to check out the power supplies, fan-outs, switches, and circuit modules as a system. This step was carried out one power supply at a time. Usually 4 of perhaps 20 switches connected to a power supply were charged. The circuit modules were charged and the energy was then transferred to the dump resistors. After the dump mechanisms and resistors were verified as operational, the circuit modules were charged and discharged into their respective dummy loads. Typically, these steps were repeated a few times at increasing voltages until the modules were charged and discharged at nominal energy. All circuits connected to a given power supply were similarly checked out.

The capacitor bank is not fully operational until the circuit modules are discharged into flashlamps in the amplifiers.

This test is done as the amplifiers become available on the spaceframe. The entire process is scheduled to be complete by mid-1984. To date, all of the amplifier circuits have been charged to nominal voltage and fired into dummy loads.

Pockels Cell Drivers

Precise control of a large chain laser such as Nova requires that Pockels-cell-based fast optical switches be included in the laser beamlines and in the master oscillator room (MOR). Functions of the fast optical switches, as shown in Fig. 2-30, include suppression of amplified spontaneous emission (ASE), pulse selection (switchout) for the mode-locked, short-pulse oscillator, and pulse carving of the output of the long-pulse oscillator.

During the early design phase of Nova, we discovered that EEV, Inc. in England, had developed a new family of ultrafast thyratrons suitable for use as Pockels cell drivers. We contracted with Cardon Instrument Co., near the EEV facility, to develop and build most of the Nova Pockels cell drivers. Current-handling capability is orders of magnitude greater than for a hard

tube of equivalent size. Also, jitter and drift can be kept well under a nanosecond with tetrode versions of the thyatron.

There are two specified versions of Pockels cells:

- Model A: less than 4-ns rise time to 5 kV for the MOR.
- Model B: less than 10-ns rise time to 10 kV for the laser bay.

Model A will drive two paralleled 93- Ω transmission lines; model B will drive five paralleled 50- Ω lines. Each Pockels cell driver contains internal circuitry for power mains isolation, EMI filtering, and a switching-regulator power supply to support the driver section. Equipment interlocks are provided for high voltage safety. Together, these two models will fill all Nova Pockels cell driver applications except for the long-pulse oscillator pulse carver, which requires switching speeds faster than 1 ns.

The evaluation of the model B prototype was reported in the 1982 Laser Annual.¹³ Since then, we have received and installed one production model B unit on Novette and three on Nova. All these units are on-line, and performance has been most satisfactory. We have also received two model A prototypes, and have evaluated their

Fig. 2-30. Nova Pockels cell driver usage.

performance both as a switchout and as a 10-mm preamplifier isolator in the MOR. One model A unit is currently supporting preamplifier isolation on Nova; the other is still under prototype evaluation.

We have continued to refine our planar-triode based technology. Through redesign of the two-chassis driver, both reliability and performance have been improved.

We have developed a single-board, three-tube preamplifier chassis that can drive either a single 93- Ω Pockels cell or the standard six-tube amplifier chassis. The new preamplifier has a provision for remote control of output pulse width via two fiber-optic fast-timing channels. By connecting two Pockels cells in a 93- Ω line, only the preamplifier is required to drive the cells. Nova MOR tests are being performed in this configuration.

We plan to keep the new planar-triode design active by using one or two in the Nova MOR indefinitely. The planar-triode design offers a viable, independent alternative to our thyatron units. The planar-triode design also promises improved speed with further development.

Work in other areas of LLNL with planar triodes has shown us that we have not yet reached the speed limits of available devices. Switching times as low as 400 ps have been demonstrated. To achieve times this low, the tubes are operated in the grounded-grid configuration, thereby bypassing the Miller effect. We have an ongoing technology-sharing policy with other groups in LLNL, with mutual benefit to all involved. We are examining some solid-state Pockels cell driver work that could lead to a faster and more robust front end for our planar-triode units.

Plasma Shutter. We are nearing completion of the production design of the plasma shutter. The plasma shutter operates by electrically vaporizing an aluminum strip into a dense plasma that is injected directly into the laser-beam path, near the focal point of the final spatial filter. A more extensive description of the concept and of research on the plasma shutter system is available elsewhere.¹⁴⁻¹⁸

We eliminated the requirement for two-axis remote alignment of the plasma shutter, including insertable cross hairs. This elimination resulted in major mechanical simplification. The laser could be refocused during alignment so that the horns of the

plasma shutter chip could be used as alignment guides.

The mechanical and high-voltage designs of the plasma shutter were finalized, and production prototype components obtained. Pulser final assembly was nearly complete by the end of the year. The design of the production version of the system control and diagnostic electronics was completed.

Three major design guidelines were followed for the plasma shutter electronics:

- Modular system assemblies were required, to minimize production, maintenance, and installation problems.
- System self-containment was required: electric power, Novabus, MOR trigger, and master-abort fiber-optic lines are the only external connections. The plasma shutter system includes a self-contained fast-diagnostic system to verify proper firing sequence.
- Attention was paid to electromagnetic compatibility: high energy discharge path will be extremely noisy electrically, so external wire connections are filtered and shielded, and critical lines are optically isolated.

Modular system assembly also aids in stand-alone testing, and has allowed the schedule of the plasma shutter to be less tightly bound to the Nova construction schedule. Installation of the production unit plasma shutter system on Nova is planned for September 1984.

Oscillator Controls

The function of oscillator controls is to control up to five laser oscillators and to synchronize high-speed system elements such as switchouts, ASE Pockels cells, streak cameras, and other beam diagnostic devices.¹⁹ From experience on Novette, we found that some changes in oscillator controls were required for operation on Nova. A block diagram of Nova oscillator controls is shown in Fig. 2-31.

Each of the up to 128 delay channels is enabled at the desired repetition rate by the power conditioning front end processor (FEP). Oscillator controls required that a majority of the FEPs interrupt time to change enable bits for each fast timing channel. A multiplexing scheme was implemented to reduce the number of full system synchronization commands from 128 to 4.

A local LSI-11/23 is used to control the oscillator controls hardware, generate local synchronization commands, and further unburden the FEPs. Remote synchronization commands from the power conditioning FEPs are used for synchronous system operation. Local or remote synchronization can be selected by a switch in the oscillator controls R-bus interface chassis. The Novanet fiber communications network is used to communicate between the oscillator controls LSI-11/23 and the power conditioning on-line VAX. The Novanet link allows control system Ramtek displays to command oscillator controls.

The programmable delay-controller panel was replaced by a CRT terminal with a touch-sensitive screen to allow more flexibility for future expansion.

The Novette fiber-optic drivers used avalanche transistors to pulse the laser diodes. The majority of timing errors were caused by avalanche transistor failure. This resulted

in excessive maintenance requirements. The Nova fiber-optic driver uses an SCR instead of the avalanche transistor.

The Novette fast timing channels were nonmonotonic and susceptible to false triggering due to noise, and had an output pulse width that was a function of the programmed time delay. The nonmonotonic behavior was due to a nonsymmetrical clock, and to the use of clock edges to introduce a delay of 16 ns. To solve the problem of nonmonotonic fast-timing channels on Nova, a 31-ns passive programmable delay line is used instead of clock edges and the 15-ns passive delay line.

Noise immunity on Nova was improved by sampling the synchronous fire timer command line. The channel enable line now acts only as a synchronous permissive. All edge-triggered events are now driven by the system's high-power, low-jitter 32-MHz clock signal. The channel output pulse width is now set to be 32 periods of the

Fig. 2-31. Nova oscillator controls block diagram.

Fig. 2-32. Fast timing channel.

Fig. 2-33. Nova power conditioning control system.

32-MHz clock, approximately 1 μ s. We implemented a 50- Ω output driver to facilitate flexible system interconnects (Fig. 2-32).

One set of switching power supplies, designed for the 10-mm rod amplifiers, was

installed for the Nova MOR. A manual control panel for the system was developed that allows the switching power supplies to be operated when the computer is not available.

A salient feature of the switching power supplies is that they can be operated remotely through a specially designed Novabus interface. This feature was not available on Novette and has resulted in significant space saving in the MOR. Two interfaces are required, each of which is capable of driving four systems or eight power supplies. In addition, a computer system interface (CSI) provides the trigger for the flashlamp power supply for synchronized timing and rep-rate capability.

Pulsed-Power Controls

The pulsed-power system is controlled and monitored by a computer system, as illustrated in Fig. 2-33. The Nova control system contains three VAX-11/780 central control computers. The pulsed-power system generally only runs on one of these computers, although any of the remaining computers could be used.

The operator console contains a Ramtek display with a touch panel and a DEC VT125 computer terminal. The operator controls the system by selecting various high-level functions available through a series of display menus. Several terminal menus are also available to provide low-level control and maintenance functions.

The VAX responds to operator requests by generating a series of commands to the hardware devices. These commands are sent to the LSI-11/23 FEPs through a shared memory or over a fiber-optic communication network. The FEPs interpret the commands and perform the appropriate functions to the hardware. In reverse, the FEPs constantly poll the hardware for status changes and send new information back to VAX through the multipoint or over the fiber-optic network. Information may be presented to the operator on one of the display menus.

The two FEPs connected to the shared memory are used to control the various pulsed power devices such as power supplies and capacitor bank interfaces. The third FEP is used to control the laser master oscillator fast-timing hardware. An additional LSI-11/23 processor is included to provide auxiliary control of the system.

diagnostics, and target diagnostics subsystems into a unified system by which an operator can control the entire laser. Although this layer extends beyond the pulsed power system, it must use many of the services provided by the pulsed power system. In addition, the laser control layer provides global shot scheduling with second to minute timing resolution.

Layer two, subsystem control, provides high-level control of the laser pulsed power subsystem to layer one. Based on the pulsed power, timing, and facility control, provided from layer three, layer two makes decisions that affect the entire subsystem, such as aborting shot sequences when facility related failures are detected, or selecting various laser component configurations.

Layer three, virtual control, provides control of high-level, pulsed-power components. Control of these components is internally translated into lower layer requests. An operator at this level may select various laser components, or sets of laser components for a particular operation. At present, four basic types of operations have been defined: a test sequence, a pulsed-ionization lamp check (PILC) sequence, a dry run sequence, and a shot sequence. The test sequence verifies that the system is

Fig. 2-34. Power conditioning controls.

Controls Layering

We developed a control model as an aid to system definition and software design. As illustrated in Fig. 2-34, the control system is divided into six layers of controls functions. The layers are numbered from highest to lowest, one to six. The highest layer provides service to the operator; the lowest layer interfaces directly with the pulsed power elements. Layers one through four are primarily software based; layers five and six are primarily hardware based. Each layer builds upon the layer below to provide higher-level services to the layer above. Communication takes place only between adjacent layers, and only through a strict, well-defined interface. Adherence to this model has enforced structured design and modular development.

Layer one, laser control, is used to integrate the pulsed power, alignment, laser

functioning properly by performing several diagnostic functions and hardware checks. The PILC sequence may then be used to determine flashlamp failures. The dry run sequence performs a shot sequence but without high-voltage. The shot sequence is used to verify charge and to fire the system.

Upon selecting a set of components, one of the four sequences can be initiated. The virtual control layer determines which lower-level devices are necessary, the functions that must be performed, and the time at which they must occur. As the sequence progresses, the selected hardware is automatically configured and verified. Verification failures generally result in the sequence being held, while severe failures at critical times result in an abort.

The virtual control layer also provides control of high-level system timing, and monitoring of the facility related subsystem; this includes the safety interlocks and the nitrogen supply system.

Layer four, logical control, provides direct control of the numerous pulsed power devices, including control of power supplies, ignitrons, and lamp-check diagnostic (LCD) packages. An operator at this level can monitor and command each pulsed power device individually. The bulk of control at this layer is carried out within the FEPs. As logical level requests are generated within the VAX, they are transferred over the network or through the shared memory to the FEPs. These requests are translated into lower level requests provided by services from the next lower layer. New status is returned to the VAX in a similar manner and is also available through the logical level interface.

Layer five, physical control, is based on individual hardware interfaces incorporated in each pulsed power device. The heart of this control structure is an interconnection network known as the Novabus. The Novabus connects the FEPs with the pulse power devices by extending the FEPs internal bus over a serialized fiber-optic bus throughout the laser facility. Bus messages from the computer are relayed down each fiber-optic chain to the appropriate device. To improve reliability, each device is serviced by two redundant computer buses. Voltage isolation and noise immunity were key issues in the design of all the Novabus components.

Layer six, hardware control, contains control electronics. These electronics are responsible for driving the actual pulsed power elements. Included in this layer are the ignitron electronics for firing and monitoring the ignitron switches, and the LCD electronics for detecting and sampling flashlamp circuit currents.

Nitrogen System Controls

Manual control switches are now in place for all 31 valves and for the 10 power contactors. Through the use of fiber optics, there is no electrical connector between the nitrogen pad and the laser bay components.

The manual control switches control the valves, vaporizer, refrigerator, heater, and blowers. The vaporizer, refrigerator, and heater each have their own temperature control systems. The blowers have a variable voltage inverter to allow speed control.

Summary

During the past year, the majority of the Nova Power System was installed and activated (without the additional amplifiers) through the first eight arms. Assembly and activation presented no unusual problems, and took place efficiently. The controls system was available from the point of first need and was very useful in activating the bank.

Author: K. Whitham

Major Contributors: R. E. Burch, G. R. Dreifuerst, D. G. Gritton, R. W. Holloway, D. Johnson, B. T. Merritt, A. M. Mihalka, J. A. Oicles, E. Padilla, J. A. Smart, K. Snyder, R. L. Thombley, and R. B. Wilcox

Beam Alignment and Diagnostics Systems

Introduction

The Nova alignment system is designed to perform automatically all preshot

Fig. 2-35. Status of sensor packages and stepper motor controllers at the end of 1983.

alignment tasks from the oscillator to target chamber center. The laser diagnostics system automatically collects all laser energy and streak camera data at shot time. Subsequently, the data is processed and displayed for operator use. We have focused our efforts in 1983 on the fabrication, installation, and activation of the front-end and laser bay subsystems, and have continued development of the output and target bay subsystems. Figure 2-35 shows the status of the alignment and diagnostic sensor packages and stepper motor controllers (SMCs) at the end of 1983.

Authors: E. S. Bliss and R. G. Ozarski

Alignment Sensor Packages and Related Components

The input sensor,^{20,21} mid-chain sensor,²² and output sensor^{23,24} designs are described in earlier Laser Annuals. They are included in Fig. 2-35, but they will not be discussed further here, because the previous descriptions are still accurate.

We have finished the multiplexed local illuminator and harmonic module through the detail design stage (Fig. 2-36).²⁵ The harmonic beam transport components are of modular design, which enables us

to make the paths from each of the beam-lines to the harmonic module optically equivalent for both 2ω and 3ω . Each optical relay section meets the following additional requirements:

- Achromatic for 2ω and 3ω , with unit magnification.
- Full field angle of 100 mrad, corresponding to 2 mrad in the 74-cm output beam.
- Designed without field lenses to avoid the high flux near the focus of the alignment beams.
- Diffraction limited over a flat image plane.

The achromatic lenses are air-spaced doublets made of fused silica and Schott LF5 glass. Both materials have small absorption coefficients and low susceptibility to solarization at 3ω . A short, fused-silica relay system adjacent to each output spatial filter is inserted for 3ω operation to compensate for the difference in the 2ω and 3ω focal lengths of the spatial filter output lens.

Figure 2-37 is a diagram of the wavelength-independent pointing reference required to obtain collinearity between the 1ω alignment beam (which propagates through the entire laser system) and the 2ω and 3ω alignment beams introduced at the output spatial filter.²⁶ The pointing reference uses a fixed Cassegrainian telescope (which is inherently achromatic) and an insertable mirror for sampling the center of

Beam Alignment and Diagnostics Systems

Fig. 2-36. The local illuminator, harmonic module, and associated beam transport components.

each beamline in sequence. Two alignment beams are collinear when they focus at the same spot on the solid-state charge-injection-device (CID) camera at the telescope's focal plane. We will use four telescope assemblies to align Nova's 10 beams; one telescope will be installed under each of the four northward propagating beamline groups in the switchyard.

A telescope with an effective focal length of 20 m is necessary to obtain the required 5- to 10- μ rad angular resolution with a corresponding 500- μ rad field of view. To keep the physical dimension of the telescope as short as possible, we chose a 1-m focal-length primary and a 105.3-mm secondary with a 900-mm spacing. The resulting back focal length (secondary to focal plane distance) is only 2 m. Since both the primary and secondary are $f/10$, diffraction limited operation can be achieved using spherical mirrors.

The target plane imager (Fig. 2-38) will be mounted on one pole of the target chamber, and will be used to position alignment beams in the target plane for all wavelengths. Images of beams incident on a ground glass reticle are optically relayed to a viewing module located outside the chamber. When the ground glass is replaced by a properly oriented mirror at chamber center, the sensor can collect light directly from a specified beamline. Then, images of near-field planes near the focus lens can also be viewed.

The multiple lens relay closest to the target operates at $f/2.7$ with a relay magnification of about 5 $\times$, and is made entirely of fused silica to maximize 3 ω transmission. This eight-element assembly is nearly diffraction limited over a 7-mm-diam flat viewing area and covers a 10-mm-diam viewing area with slightly reduced image quality.

A motorized stage moves the assembly axially from the 2 ω focus to view the target plane at 1 ω or 3 ω . The middle 1:1 relay and the field lens are achromatized for 1 ω , 2 ω , and 3 ω light, using fused silica and Schott LF5 glass. Achromatic relaying of the image and pupil planes is essential to minimize the diameter of the optical relay.

Fig. 2-37. Wavelength-independent pointing reference.

Fig. 2-38. Target-plane imager for near- and far-field viewing at three wavelengths.

The viewing module uses a CID camera at its focal plane. The module provides 1-, 3-, and 10-mm fields of view at chamber center (far field), and 20-, 40-, and 80-cm fields of view in the near field. The resolution is better than 2% of each field of view. At shot time, the optical relay assembly retracts about 1 m into the viewing module and is protected from shot debris by a closed valve.

Authors: C. D. Swift and L. G. Seppala

Major Contributors: R. D. Boyd, G. W. Harwood, Jr., J. P. Hockett, W. A. Jones, R. L. Saunders, and J. D. Wintemute

Nova Alignment Controls

The structure and operation of the alignment control system for Nova was described in some detail in the 1982 *Laser Annual*.²⁷ At that time, the prototype controls were in routine operation on the Novette system. We adapted the prototype control system software to accommodate the differences in layout between Novette

and Nova: the schematic graphics frames used to select and display the status of alignment components at the operator consoles are unique to each system. The adaptation in control system software included addition of operator station controls to handle distribution of the cw alignment beams to the 10 laser chains by adjusting the splitter array wave plates.

We control all alignment activities from the central operator consoles with computer-generated menu displays, such as the one in Fig. 2-39. Automatic alignment, which will extend all the way to the target chamber when the system is completed, is currently operating from the MOR through the chain input sensors. In the transition from Novette to Nova, we incorporated procedures to handle alignment of a number of chains in parallel for speed improvement and improved the ability of the closed-loop alignment software to recognize various kinds of unusable images—for example, to detect cross-hair images in which the cross-hair intersection is missing.

To keep the Nova system current with a minimum of effort, we established

Fig. 2-39. Control graphic display for distribution of the two cw alignment beams to the 10 amplifier chains.

We extended VAX software for the stepper motor controls (SMCs) to support the

integration of the automatic alignment with the operator console controls and to provide several maintenance functions. With the new software, alignment devices can be grouped together into sets for larger-scale

Fig. 2-40. The Nova stepper motor controllers (SMC) network map, part of an integrated documentation package for alignment controls.

control functions. Each set can be controlled as a single entity to perform specific applications (such as to change the configuration of a sensor). With a number of devices moving simultaneously, the waiting time for the completion of a specific task is greatly reduced. The system status can be checked with fewer program calls than were required on Novette. We also eliminated typed commands to load programs into the remote SMCs: programs are now loaded from a graphics menu at the central operator's console. Important alignment parameters, formerly stored in non-volatile memory in the SMCs, can now be saved in VAX disk files as well. If the SMC hardware fails, basic data (such as device set-points and cross-coupling matrices) can be retrieved from the VAX disk and restored in the SMC.

Hardware production for 32 LSI-11-based SMCs is on schedule and nearly finished. We have designed and installed all cable trays, and cabling for the MOR and laser bay beamlines is complete. We have also installed all Novanet fiber optics and a fiber optic isolation chassis to electrically isolate the stepper motors of the two MOR output gimbals from each other even though they are controlled by the same SMC as a single device. The isolation is necessary because the second gimbal is physically located not in the MOR but in the switchyard. The status of SMC installation and activation is summarized in Fig. 2-35.

Development of additional SMC hardware to meet the needs of the Nova target alignment system is continuing. We are integrating commercially available equipment (which will drive the high speed, high torque motors of the target lens positioner) into the SMC so that the focus lens can be controlled over the Nova control network and not just by a stand-alone local controller.

The SMC for a particular alignment device is not always near that device, so some form of portable control is needed to access devices in the system for maintenance functions. We are developing a portable local controller with the same set of controls available at an SMC control panel that can be attached to any of the VAX terminal lines in the laser bay, switchyard, and target bay.

Nova alignment data is entirely in video form.²⁸ A high-quality television distribution system is vital to the operation of the automatic alignment system, and facilitates operator-assisted alignment. In 1983, we installed the Nova television distribution system. New fiber-optic technology considerably improves performance, reliability, and cost over previous systems.

We are writing integrated documentation to describe all hardware in the alignment control system to facilitate maintenance and continuing improvements. We completed the documentation for the front end in 1983.

Figure 2-40 is the top-level network map of all the Nova stepping motor controllers. The map of the distributed system serves as a reference guide to the second level documentation for each SMC. Eight different customized versions of the SMC are used in the Nova system. Each component requiring SMC control is shown schematically by device on the second level layout drawings referenced by drawing number below each SMC in Fig. 2-40. The layout drawings are used by the technical staff to prepare cable specifications and software tables for the SMCs. Electronic engineering document numbers of subassemblies are referenced at each level in increasing detail: given the top level document, all necessary supporting drawings are cross referenced. The second-level layouts also provide the basis for graphic display designs used at the operator console.

Author: P. J. VanArsdall

Major Contributors: R. D. Boyd, R. D. Demaret, D. Douglas, R. A. Fertig, G. L. Hermes, F. W. Holloway, D. L. McGuigan, R. T. Shelton, and G. J. Zaragoza

Diagnostic Sensor Packages

The input sensor,^{20,21} mid-chain sensor,²² and output sensor module^{23,24} perform alignment tasks as alignment tasks. These sensor packages and the mid-disk diagnostic sensor²⁹ have all been described in previous Laser Annuals. Their status is shown in Fig. 2-35.

During 1983, we selected the primary ghost reflection from the target focusing

Fig. 2-41. Arrangement of components in the harmonic output diagnostics spool for 2ω (upper) and 3ω (lower) operation.

lens as the means to obtain output diagnostic signals at the harmonic wavelengths.³⁰ The components needed to implement this approach reside in a spool mounted between the KDP array and the focus lens positioner. Figure 2-41 shows these components and their positions relative to the target chamber center and the KDP array. The 8-cm diam calorimeter receives no incident light directly. It resides in the shadow of a pick-off mirror that sends a 15-cm-diam sample from the center of the beam into a streak camera interface module on the side of the diagnostic spool. Also, the calorimeter faces away from the target, so it sees no light reflected from the target.

The 15-cm-diam streak-camera sample is attenuated and focused onto a fiber optic that transports it to a shared streak camera. A multiplexing fixture allows us to interface fibers from the 10 beams plus a reference signal for relative timing measurements onto the same 4-cm photocathode. An alignment edge on the fixture registers

against the streak camera slit to assure proper positioning.

The absorbers are called ABRs, because they are designed to be absorbing beam dumps for laser wavelengths not desired on the target, and because they reflect a fraction of the ghost focus light into the calorimeter. There are a number of constraints on the ABRs and other components in the diagnostic spool:

- There must be separate ABRs for 2ω and 3ω shots, and each must be moved in or out of the beam as appropriate for the next shot.
- The beam centerline must be clear of all non-transmitting components for certain alignment tasks.
- The ghost focus beam diameter at the ABR must be greater than 15 cm to avoid high flux caustics in the aberrated ghost.
- The edges of the ABR must fall in a previously apodized shadow, such as the one for the KDP array, or additional apodization must be added.

Beam Alignment and Diagnostics Systems

- The calorimeter must be far enough ahead of the focus lens to avoid mechanical interference over the full 1ω to 3ω range of the lens.
- All parts of the mechanical structure used to hold and position optical components must be shadowed from the main beam at the time of a shot.
- The beam diameter on the calorimeter must be kept greater than about 4 cm or damage will occur from high on-axis flux.
- The calorimeter's outside diameter should be no greater than about 8 cm or it will block too much of the ghost focus light before it can reach the ABR.

Consideration of the above factors led us to the spacings and dimensions indicated in Fig. 2-41 and to the arrangement of motor driven translation stages shown in Fig. 2-42. Supports for the mirrors, absorbers, and calorimeter do not block any of the beam because they are positioned in the shadow of the split disks in the 46-cm amplifiers.

A stable reflectivity on the second surface of the chamber focus lens is required in order to obtain a stable calibration of the diagnostic spool calorimeter. The insertable silicon diode (shown in Fig. 2-42) provides a way to monitor the fraction of harmonic alignment beam light that gets to the calorimeter. If the ratio of the signal on this diode to the signal on a similar diode located

in the output sensor splitter box begins to change, a recalibration of the calorimeter may be required.

The opposed-port diagnostic sensor package³¹ is still in the detail design stage. This sensor is intended to take advantage of the unused port opposite each target chamber lens. When no target is in place, the sensor can obtain an unobstructed view of the focused beam profile in the target plane. Only one or two of these instruments will be built. They will be designed to:

- Meet length, diameter, and weight limits to allow relatively easy insertion into any of the target chamber opposing ports.
- Avoid adding aberrations due to nonlinear propagation of the beam in the instrument itself.
- Safely dump most of the laser beam energy to avoid damaging small-diameter optical components and saturating the photographic film.
- Provide a well-resolved picture of the full-power laser beam in planes near the target chamber center.

Figure 2-43 shows the current baseline design, an inverse Cassegrainian telescope with a collimating lens. The primary and secondary mirrors are uncoated BK-7 spherical mirrors, each with a reflectivity of about 4%. The tertiary lens is spherical fused silica that adjusts over a 50-mm range to obtain

Fig. 2-42. Mechanical design for moving harmonic output diagnostics components in and out of the beam.

Fig. 2-43. Inverse Cassegrainian telescope design for the opposed-port diagnostic sensor package.

Fig. 2-44. The Nova target chamber at the end of 1983.

correct focus for 1ω , 2ω , and 3ω operation. The optical assembly has acceptable alignment tolerances and operates over a 3-mm diam field of view with near diffraction-limited performance. The obscured central region corresponds to a 160-mm-diam hole in a 740-mm-diam beam. However, the central obscuration does not compromise the image: this part of the beam is already removed in the diagnostic spool (see Fig. 2-41) and at the target chamber lens (see Fig. 2-17).

The target plane image will be recorded on film with a multiple-image camera to provide simultaneous images of several planes in the target chamber at several exposure levels. The target plane will also be viewed by a video camera for setup and alignment before a shot.

All diagnostic data (except that recorded on film) are acquired, processed, and stored electronically. The diagnostic acquisition and processing components have been described in earlier *Laser Annuals*^{32,33} and are now in various stages of fabrication and installation.

Authors: R. G. Ozarski and L. G. Seppala

Major Contributors: B. V. Beeman, D. A. Bender, D. M. Benzel, R. D. Bolt, J. H. Ewins, R. E. Hugenberger, D. W. Myers, J. A. Prior, and J. L. Wilkerson

alignment viewing optics, final calorimeter supports and handling equipment for lens positioners, KDP arrays, beam diagnostics, and calorimeters associated with the target chamber.

During 1983, we installed and aligned the Nova target chamber. Most of the vacuum system was installed and designs were completed on target viewers, target positioner, and target insertion hardware. Figure 2-44 is a photograph of the chamber taken at the end of 1983.

Target Systems

Introduction

Nova target systems include the target chamber with its supports and vacuum system, and the target positioner, target

Target Chamber Installation and Alignment

The target chamber is composed of three sections: a central structural ring and two hemispherical heads. The target

Target Systems

chamber has an outside radius of 2.3 m, and aluminum alloy walls 12.7 cm thick. The structural ring provides support for the target positioner, viewing optics, vacuum pumps, and numerous target diagnostics packages. The hemispherical heads support the lens positioners, handling equipment, and some target diagnostics packages. A total of 150 ports provide target diagnostics capabilities for many positions, defined in spherical coordinates.

The support structure carries the weight of the chamber on four points that are adjustable in all directions to locate the chamber precisely in the center of the room. Four additional legs brace the chamber against seismic movements. All supports are located to avoid interference with any of the diagnostic lines of sight, as developed with the aid of our computer aided design drafting (CADD) system.

We aligned the target chamber by placing a theodolite on a tower in the center of the chamber at the precise location to be occupied by the target. We verified the location with survey marks in the floor and in four walls of the target room. We stretched fine steel wires across the ports that had been originally defined as primary references while machining the chamber. We then moved the chamber with hydraulic jacks until the cross wires coincided with the lines of sight defined by the survey marks. The locations of other ports were then verified by means of the theodolite still located in the target position. Following alignment, we welded the seismic braces in place.

Vacuum System Design

Two vacuum systems, main and auxiliary, provide independent service for the target chamber and for the diagnostic packages.

The main system is designed to provide a base pressure in the target chamber of 10^{-7} Torr and a working pressure of 10^{-6} Torr. Starting from atmospheric pressure, 10^{-5} Torr will be reached in 30 minutes. The roughing system consists of two 500-cfm mechanical pumps and two 2700-cfm Roots blowers located outside the building on a covered concrete pad with a roughing line 10-in. in diameter and 150-ft long. The high vacuum is provided by four CTI model 400 cryogenic pumps. These

pumps are mounted on two plenums attached directly to the chamber. Two 5000-cfm Balzers turbomolecular pumps are also mounted on the plenums because helium will be used for leak checking and for cooling the cryogenic targets.

The auxiliary system provides enough roughing vacuum to the diagnostic packages to pump the package through the target chamber. If higher vacuum is required for a given experiment, individual cryogenic pumps will be installed on each package. The roughing system consists of a 107-cfm Kinney pump and an 850-cfm Roots blower. The pump and the blower are connected to the chamber by a 4-in. roughing line. Two manifolds, shown in Fig. 2-44, surround the chamber and provide 68 possible connections. Pressure in the manifolds will normally be in the 10^{-3} Torr range, but will rise to as much as 200 Torr when a new experiment is initially connected. To prevent pressure surges to other experiments, roughing valves are interlocked to allow only one opening at a time. The time required to rough a 150-liter experiment to 0.1 Torr is less than two minutes.

For experiments that require individual cryogenic pumps, we have provided helium supply and return manifolds around the structural ring. Automatic quick connectors allow connection and removal without disturbing the rest of the system. There are four helium compressors in the main system and eight in the auxiliary system. The compressors are computer controlled and are brought on line as required to maintain the proper supply and return pressures in the manifolds.

Controls for the vacuum control system are computer-based and interlocked to prevent any injury to equipment or personnel through operator error or equipment failure.

Target Positioner Development

The Nova target positioner design is based on our experience with Shiva and Novette. We use crossed roller-slides to provide three orthogonal motions and one rotation about the axis of the unit. The overall arrangement is shown in Fig. 2-45. All motions are stepper motor actuated, and positions are indicated by rotary encoders attached to the

Fig. 2-45. Overall layout of the Nova target positioner.

motor shafts. The center of the positioner is a tube with a 133-mm i.d. Into this tube, we will insert a target support pylon by means of a carriage mounted on rails on the top of the target chamber. An air lock allows pylons to be changed without affecting the chamber vacuum. The pylons can be specially designed to provide support features for any of the present targets. A pylon can also be used to support an optical relay system. A duplicate target positioner will be used to support and manipulate the target plane imager pylon.

Author: F. Rienecker

Major Contributors: S. B. Rodriguez, K. A. Hurst, T. L. Schwab, K. J. Haney, R. L. Horton, and D. C. Kumpf

Mechanical Systems

Introduction

During 1983, we completed fabrication of the hardware for Nova and started

assembly and installation of component parts. We completed construction of all four spaceframes (MOR, laser, switchyard, and target), and activated two of the laser rooms (MOR and laser bay) as class-10 000 clean operating facilities.

By building Novette, we were able to put 20% of Nova's hardware through service for a full one to two years before it was needed on-line for Nova. The additional time permitted us to deal with assembly, installation, operation, and debugging of each Nova component. We documented all changes, errors, and improvements made on Novette hardware. This information led to action items for the remaining 80% of the Nova hardware.

The advantages of having dependable parts, and good planning on the basic requirements of quality, schedule, and cost, assured us of troublefree assemblies. We developed a system to inspect, store, and assemble parts kits. The parts were then withdrawn from inventory and cycled through shops to correct discrepancies found while building Novette. Corrected parts were returned to inventory for subsequent withdrawal and assembly in our

Mechanical Systems

clean rooms. The flow of assemblies through our clean rooms is phased with the schedule for installation on the spaceframe.

Spaceframe

We began construction of the steel support frames in July 1982, and completed it in May 1983. Sections of each frame were shop fabricated, painted, and shipped by truck.

We fitted the sections of each frame together, aligned them in position, and welded them. We installed a total of 50 prefabricated sections in the laser, switchyard, and target rooms. In several cases, entire structures were assembled for dimensional control, then cut apart for shipping. We specified dimensions to be within 3 mm. Our acceptance survey showed deviation from this. On the laser spaceframe (which consists of two structures 10 m high, 3 m wide, and 62 m long) 30% of the final dimensions were perfect and the remaining tolerances varied to within 6 mm. A few points were badly out of tolerance. Other frames had similar tolerances. Our system allowed us to compensate for this with shim plates that develop a vertical plane over the entire surface within 3 mm. There are four 10- × 62-m vertical planes, which are the mounting surfaces for all system components. We removed the 3-mm deviation in these surfaces using component cradle adjustments

to give us near-perfect mechanical alignment. (Even though the frame was badly out of tolerance on one of the four planes, our system of *shimming* and final adjustment permitted final alignment.)

The spaceframe must hold more than 400 components stable enough to maintain alignment. We used finite-element analyses to predict dynamic response of the frame to disturbing vibrations. These data were used to determine the structural requirements necessary to accomplish stability for each frame. We similarly analyzed support and anchor containment conditions due to seismic activity. Additional structural members were required to meet seismic design standards. A balance between minimum mass and high stiffness had to be met to assure a high fundamental frequency.

Gravity deflection of the frame is of little concern because of the large amount of steel required for a high fundamental frequency compared to that required to carry the gravity load. However, we looked for the best assembly plan to avoid tilt (caused by uneven loading during assembly), which would lead to movement of previously aligned parts.

We performed dynamic testing³⁴ following the installation of each frame to link analytical and experimental data to verify design values. Table 2-4 shows a comparison of design values. The first two modes are within several Hz and indicate good correlation between design and actual values. The first and most important mode, 8.27 Hz for the laser frame, gives us assurance that laser beam deflections from the resulting motion are acceptable.

The first step in dynamic testing was to construct a mathematical transfer function model of each structure. First, we divided the structure into nodes. We then excited the structure with a hammer blow at one convenient node, and measured the response with accelerometers at the other nodes. We determined the transfer functions between nodes using Fourier transform techniques. We then extracted modal parameters from the transfer functions using the TRANSF computer code.

We painted the frame with a very tough polyurethane paint, and the frames were thoroughly washed on all surfaces with soap and water to bring each room to class-10 000 clean conditions.

Table 2-4. Calculated and measured frequencies of the Nova spaceframe.

	Calculated frequency (Hz)	Measured frequency (Hz)
Laser bay frames		
Mode 1	9.95	8.27
Mode 2	10.08	8.45
Mode 3	11.32	8.72
Mode 4	12.11	9.57
Mode 5	13.61	11.25
Switchyard frames		
Mode 1	6.66	5.5
Mode 2	7.06	6.3
Mode 3	8.25	8.6
Mode 4	15.56	19.1
Mode 5	16.13	21.4
Target frame		
Mode 1	4.15	5.6
Mode 2	5.45	5.9
Mode 3	5.72	6.3
Mode 4	5.91	6.5
Mode 5	6.56	7.5

Amplifiers

There are 166 power amplifiers on Nova, ranging in size from 5- to 46-cm output beam aperture. Six of these are for the front-end splitter and preamplification section. The remainder (160) are on the 10 main laser beams.

Our reflective surfaces have a 90% reflectivity, which is better than any of our previous large production surfaces. Highly-polished tooling was used for electroforming of nickel parts that were subsequently silverplated. A very thin coating of super-bright silver was deposited with no subsequent polishing. All silverplated surfaces were 100% inspected using a tape-pulling method, visual inspection, and reflectivity testing. Silverplated surfaces degrade with time from exposure to oxygen, so we attempted to keep them in a nitrogen atmosphere. When the components are assembled and on-line, they are protected by the system nitrogen. While in storage, the components are also protected, but during assembly in the clean rooms, they can be tarnished. We clean the components by dipping them in a thio-urea solution that combines with silver sulfide and is removed as hydrogen sulfide. This procedure, carried out just before final placement in the assembly, removes no silver and restores the reflective surface. We have processed approximately 50% of our amplifiers. The assemblies are then checked on an interferometer, validated for accuracy and cleanliness, wrapped with protective covers, and delivered to the laser bay.

Spatial Filters

Spatial filters are assembled in our class-10 000 clean room. Approximately 50% have been assembled and installed. These are sizes up to 31.5 cm.

We recently performed tests on Novette³⁵ to study the propagation of energy through a system of spatial filters at a high pressure (1 to 10 m Torr). Three shots through a 7.4-cm spatial filter with energies between 7.5 and 8.5 kJ indicated gas breakdown at 10 m Torr and above; gas breakdown is not noticeable at 1 m Torr. On the smaller spatial filters we use small vac-ion pumps, and

operate them at lower pressures than 1 m Torr to extend their lifetimes. Cryo pumps are used on the larger spatial filters as a best choice. Diffusion pumps would require extensive trapping, valves, and backing pumps. Turbo-molecular pumps would have the risk of vibration, and vac-ion pumps would be too large for the required speed. Our experience has been that the lens holders and positioners support the lenses without distortion and are easily adjustable under vacuum load.

Mirrors

We have assembled and installed approximately 50 mirrors. Interferograms of all installed mirrors show most of them to display a surface flatness of 1/4 wave, with the worst cases at no worse than 1/3 wave.

It takes two people approximately 16 hours to assemble one complete mount (with mirror), test it in front of an interferometer, and mount it on the spaceframe. Although this can be considered quick for optics as large as 109 cm in diameter, it should be pointed out that by the time all of the Nova mirrors are installed, an estimated 2200 man-hours will have been spent assembling the largest area of precision optical mirrors in the world.

Faraday Rotators

There are 42 Faraday rotators on Nova, 2 of 5-cm clear aperture and 10 each of 10-, 15-, 20.8-, and 31.5-cm apertures. We have assembled and partially installed most of the Faraday rotators. Interferograms indicate that the mounted polarizer plates and rotator disks are better than $\lambda/4$ in transmission.

Switchyard and Target Beam Tubes

The 10 Nova beams exit the last spatial filter with a diameter of 74 cm. They enter the switchyard (Fig. 2-46), where each beam is pointed by two mirrors to the target bay, and then pointed at the target chamber by another set of mirrors. The beams are enclosed in 36-in.-diam aluminum tubes through all this distance to avoid distortion by air currents. The total

Mechanical Systems

length of beam tubes required for the 10 Nova beams is 455 m. Originally, we intended to make the entire beam-path length hermetically sealed so that nitrogen gas

could be used inside the tubes. The Novette beam tubes were designed to be a non-sealed system as a test bed for Nova with the option of modifying it to a sealed-beam

Fig. 2-46. 74-cm-diam beam tubes in the Nova switchyard.

Fig. 2-47. Nitrogen re-circulating system.

path if necessary. Novette experiments have shown that nitrogen is not required in the beam tubes; it is possible to build Nova as a nonsealed system as well.

In addition to beam tubes, the system has many panels and weldments required to cover the switchyard mirror towers and the target bay mirror housings. The system also includes an additional tower in the switchyard for supporting the tubes, and has independent supports throughout the target bay. The system is designed to go together with as little field fabrication as possible so that sections of beam tubes can be removed during maintenance periods.

Recirculating Nitrogen Gas System

The nitrogen recirculating gas system³⁶ is designed to minimize contaminants. The nitrogen system (Fig. 2-47) has three modes of operation:

- Recirculating, in which nitrogen is recirculated, filtered, and temperature adjusted.
- Open loop, in which gas is drawn from the liquid nitrogen tank, temperature adjusted, passed through the laser amplifiers in a single pass, and exhausted to the atmosphere.
- Purge mode, which allows a single laser arm (or, if necessary, the entire laser) to be purged of atmosphere after a beamline has been opened.

In the recirculating mode, liquid nitrogen is drawn from a 28 000 gallon tank, vaporized in a water-heated heat exchanger to about 50°F, and regulated to a pressure of 0.2 psig. The gas flows to the suction side of three Roots-type blowers capable of producing flow rates between 100 and 5000 cfm up to 4 psig. The Roots blowers draw either gas from the laser return pipe or make-up gas from the nitrogen tank.

A pressure of 0.2 psig in the return ensures that the amplifiers and return piping are above atmospheric pressure. A selected combination of Roots blowers pressurize the gas to about 1 psi and cause it to flow through two stages of HEPA filters, refrigerators to remove the heat of compression, and a heater that sets the final temperature to within $\pm 0.5^\circ\text{F}$. Flow rate through an amplifier or a series of amplifiers is established by membrane filters installed in the amplifier. We achieved predicted flow rates by selecting the pore size and area of a filter at

a particular pressure. Gas flows in two parallel cavities in an amplifier—the pump and flashlamp cavities. We selected filters to cause the pump cavity to be at slightly higher pressure than the flashlamp. Higher pressure in the pump cavity reduces the possibility of contaminants being swept from the lamp assembly into the more critical laser cavity. Used gas flows in the return duct to the suction side of the blowers.

In addition to multiple filtering, the gas, ducts, valves, and other in-line components were cleaned and assembled using class-10 000 clean-room technology. The aluminum pipes were caustic etched and passivated to remove the mill finish and oil from inside surfaces. We use low-outgassing polyurethane tubing as the final connecting hose. A special silicone rubber coupling was designed to provide a clean, low-outgassing joint for the pipe sections. This joint allowed pipe sections to be prefabricated and set in place without welding or use of common flanges (which are not conducive to particulate-free assembly).

Author: C. A. Hurley

Major Contributors: F. A. Frick,
A. Martos, and H. G. Patton

Integrated Controls

Introduction

We continued to install and integrate Nova controls during 1983. We installed and activated the three VAX-11/780 control computers, and started their connection to the laser control system. We continued to enhance the central controls software tools throughout the year.

Shortly after we brought the three VAX control computers on-line, we began to test shared memory between them. We have demonstrated that the memory can work as the hardware communication mechanism, and we are continuing work to tie the control system software processes together through the memory.

We moved the Nova central control room from the basement to an area adjacent to the main entrance to the laser bay on the ground floor. We now have the central control room next to the operator's room, with both rooms close to the laser bay. A

Integrated Controls

clean-room environment in both rooms will allow the operating staff to move freely between the laser bay and the control room.

By the end of 1983, important major subsystems of the control system were in operation, and we had met one of the primary goals established at the beginning of the project: the control system was in use to help install and activate the laser. With the control system, we can automatically align major portions of the laser path. Automatic alignment is controllable from either the central control room or a special single display console inside the laser spaceframe.

Software Enhancements

We enhanced the operator console graphics system in the Novette control system for use in Nova. Sequential display updates now occur up to twice as fast as before. Also, we implemented a general screen switching mechanism to allow a single screen console to do anything that a full three-screen console in the central control room can do (with slight speed restrictions). The single-screen units are used in the laser bay, switchyard, and diagnostics loft to facilitate installation, activation, and pre-shot setup.

We upgraded the Novanet software driver for LSI-11 processors to support the entire memory-address space of the LSI-11/23. This upgrade was done primarily to support the laser diagnostics code, which has been expanded in the past year.

All significant problems in the Praxis language compiler were fixed during the year. We used our extensive control system software (~300 000 lines), which is written in Praxis, to study the impact of modern block-structured language features on large control systems. We are using this data to draw conclusions about languages of this generation such as Ada.³⁷

Software Quality Assurance Plan

A Nova software quality assurance plan was written and approved during the past year.³⁸ The plan serves two purposes: First, it documents many of the techniques and procedures we have followed over the years to ensure the expandability, reliability, and maintainability of our code. Second, it

establishes criteria by which to judge both past and future software products.

The approach is to divide our control system software into products, which can be separately traced against the criteria in the plan. These products, which interface with each other, make up the entire software portion of the Nova control system. New versions of software are managed on a product-by-product basis.

Supplementary Guidelines

We have developed and formalized specific guidelines to organize the computer system software and to control software transport from the development VAX computer to the three control VAX computers. These guidelines describe how the computational load of the control system software will be shared among the three control computers. They also describe how the system hardware and software is to be organized to maximize the availability of portions of the control system in the event of an unscheduled failure or scheduled down-time of one of the three control VAX computers.

Under normal conditions we will operate the power conditioning and target diagnostics subsystems on the north VAX, the laser diagnostics and central manual alignment subsystems on the south VAX, and the auto-alignment software on the west VAX.

Auto-alignment described in "Beam Alignment and Diagnostics Systems," earlier in this section, minimizes the manual adjustments required of laser operations personnel. The laser may be fired without auto-alignment, but at a much lower shot rate. In the event of a single VAX failure, we can rotate all the machine assignments so that the entire system, together with the auto-alignment software, will operate in a degraded mode on the remaining two VAX computers. The rotation of computer assignments is aided by a computer data bus switching scheme we built that allows the peripherals of any one VAX computer to be used by another computer.

For the machine rotation scheme to work, the software on all disk packs on all three machines must be identical at all times. When necessary, we must be able to switch peripherals to the remaining two operational VAX computers within one hour. We will ensure that the control system software

on all disk packs is identical by periodic re-writes of all control computer disk packs on the Nova development VAX. Newly developed code from the development VAX will be transferred over the Novanet network to the three control VAX computers.

The procedures for developing and maintaining software include specifications for how program files are grouped on disk storage. Approximately 24 separate software products make up the total Nova control system. Each software product is composed of many files. Up to three complete sets of these files may be maintained at a time. Each set, called a release, represents the product at one of three stages of functional development: development, test, or operation.

We established a structure to store, access, and manage all of the software products at each of the three levels that make up the total system. As we add new features to a product at the development level, previous software is pushed upward to the test level, and then to the operational level as confidence in proper operation of that feature is established. The same structure

for storage is used on all Nova VAX computers. Operators and programmers can easily determine and track the relative level of the software products with which they are involved. Some software products have already evolved to the point of maturity during the past year.

Authors: F. W. Holloway, J. E. Krammen, and J. R. Severyn

Major Contributors: A. J. De Groot, J. M. Duffy, J. R. Hill, D. L. McGuigan, M. K. Rash, M. Schlicht, J. A. Smart, and J. W. Taylor

Nova Project Management

Introduction

At the beginning of 1983, the Nova project (configured as 10 arms with $2\omega/3\omega$ frequency conversion), was fully funded at \$176 million. During January and February,

Total project	03/01/83 Baseline	171.300
December 1983	Estimate at completion (EAC)	171.755
Planned % complete 94.68%	Variance over (-) under (+)	-0.455
Actual % complete 91.11%	Contingency	4.700
	Total estimate at completion (TEC)	176.000

Fig. 2-48. Nova project financial and completion schedule.

Novette Laser System

we prepared detailed cost and schedule estimates to support the project cost of \$176 million, with completion planned for the end of September, 1984. Our estimates became the March 1, 1983 project baseline that we presented to the Department of Energy at the Nova project semiannual review, held in April, 1983 at LLNL. During the remainder of 1983, we tracked the project against this baseline.

The baseline financial plan for the total project is illustrated in Fig. 2-48. At level 1 of the project work breakdown, we allocated funds as follows (amounts in thousands):

Nova laser/target system	\$135 310
Nova laboratory/office building	35 990
Contingency	4 700
	<hr/> \$176 000

The laboratory/office building figure represents construction that was completed in 1982. The contingency amount of \$4700 thousand was 10.1% of the project uncommitted balance on March 1, 1983.

Project Management Controls and Tracking

The Nova project management controls provided effective cost and schedule tracking during 1983. As the project entered its final phase, we put additional procedures in place to ensure project completion on schedule and within budget.

A key to controlling schedule and budget during the final phase of assembly was to reflect in the project cost plan all new expected costs not included in the baseline. We accomplished this by entering negative variances in anticipation of the expected costs. The inclusion of expected future variances, as well as variances to date, allowed us to accurately report total estimated project cost at completion. Also during the final phase, all level-of-effort work packages were classed as cost variances instead of schedule variances, and we changed the cost tracking system to reflect all labor and supplies/expense variances as cost variances. Finally, to ensure that the project commitment curves accurately reflect remaining procurement issues, all unplaced procurements are now reviewed and

evaluated monthly for their schedule impact. Project engineers are given summaries to distinguish behind-schedule procurements from those that are schedule insensitive.

Cost and Schedule Systems

The key control documents of the Nova cost system are the commitment performance measurement graphs. At the end of December, 1983, the remaining contingency for the Nova project was \$4245 thousand, which is 18.9% of the uncommitted project balance for that date.

The Nova activation schedule is the key document in the Nova schedule system. We established detailed milestones for Nova assemblies so they are available at the required time for activation. Through 1983, we updated milestone status of assemblies every two months, and revised assembly milestones schedules after each update. These periodic reviews highlighted potential problems to allow timely corrective action.

Cost/Schedule Summary

The March 1, 1983 cost and schedule baseline established a projected total cost of \$176 million, with a planned completion at the end of September, 1984. At the end of 1983, with all known future costs and assembly schedules considered, we expect the Nova project to be completed within budget and on schedule.

Author: F. J. Holcomb

Novette Laser System

Introduction

As recently as three years ago, the relative effectiveness of different lasers as fusion drivers was still controversial. Today, the majority of evidence indicates that a $2\omega/3\omega$ laser will be effective as a fusion driver. We assembled Novette with components from Shiva and Nova as a short wavelength source for multikilojoule target experiments.

We assembled Novette on an accelerated schedule in an existing building adjacent

to the new Nova laboratory. Construction began in January of 1982, and system activation was completed in 13 months. Novette includes a Nova-style control system, the refurbished Shiva target chamber, and a full suite of target diagnostics. Novette delivers 18-kJ, 1-ns, 1 ω pulses, which are frequency doubled and focused onto targets. During the past year a large body of data has been gathered concerning the performance of Novette as a system. In the following article, some of our operating experience will be described, new subsystems that we implemented this year will be discussed, and finally, measurements on Novette's beams, taken at several locations within the laser chains, will be compared with simulated calculations.

Each of Novette's two beamlines is optically similar to a Nova arm: the emerging beams have a diameter of 74 cm. Harmonic conversion takes place in two unique mosaic arrays of type II potassium dihydrogen phosphate (KDP) crystals. The two ~ 4.5 TW 2ω pulses are concentrated onto targets by two 74-cm-aperture $f/4$ doublet lenses. Because the two pulses arrive at the target within 5 ps of each other, a typical target may be irradiated simultaneously from two sides by ~ 9 kJ in 1 ns. We devoted about half of Novette's experimental time to plasma physics studies, whose aim was to develop a better understanding of short-wavelength laser-plasma interaction phenomena relevant to inertial confinement fusion (ICF). The balance of the experimental time was divided between high-density implosion research and non-LTE plasma experiments. In short, Novette has provided a flexible, high-energy-density experimental facility to bridge the gap between Shiva and Nova, while allowing us to probe the details of the Nova design. Detailed description of the Novette laser system is available in previous Laser Annuals.^{39,40}

Novette Design

We based the design of the Novette test bed on the requirements of the Nova laser. Because the output of a repeatable, high-quality master oscillator is typically $\sim 10^{-4}$ J, Nova's power amplifier must provide a gain of 10^5 . In Novette, the weak pulse produced by the master oscillator first

passes through a preamplifier, which consists of crystalline and glass rod amplifiers, beam-expanding spatial filters, and beam-splitting elements, to produce two parallel, spatially uniform beams with energies of up to 1 J each. At the master oscillator, the laser beam is spatially Gaussian with a diameter of 1.5 mm at 10% of its maximum intensity. As the beam proceeds through the preamplifier, it is expanded and passed through an aperture so that only the smoothest central region exits the preamplifier and is presented to each of the arms. In each arm, after the preamplifier, seven beam-expanding spatial filters separate stages consisting of rod and disk amplifiers with increasing apertures. The peak output flux from each stage is chosen to be near its operating limit. Propagation is controlled by these spatial filters to minimize local intensity fluctuations and keep the maximum laser flux below optical-damage limits.

The principal design constraint on the maximum energy delivered by each stage is the damage threshold of its optical components, which depends on the purity of the optical medium, the optical surface, the pulse duration, and the wavelength of the laser light. To raise the damage threshold in Novette, components such as spatial filter lenses, which in former laser systems were antireflection coated with dielectric films, were replaced with optics treated by a neutral solution process.⁴¹ Such optics not only display excellent transmission (typically $>99\%$), but they also have surface damage thresholds of 12 to 16 J/cm² for 1-ns pulses, approaching those of the polished glass substrate.

A major design goal for both Novette and Nova is to maximize the output energy without damaging optics. We would like to propagate an ideal beam that exactly fills the power amplifier with an output flux just below the damage limit. However, such a beam does not propagate for any useful distance without diffraction of the beam edges, which leads to constructive interference and regions of excessive optical flux. The first spatial filter in each of Novette's arms has a 2-mm-diam pinhole in its focal plane. The beam profile that emerges is nearly uniform, except for six faint rings. Succeeding spatial filters relay this beam onto the entrance lenses of the following spatial filters and eventually onto the

Novette Laser System

target-chamber focusing lens. The net effect of this strategy has been to achieve a measured fill-factor of 85% in Novette. This is substantially greater than the ~70% or less of earlier lasers.

The MALAPROP code has proven invaluable for estimating peak fluxes impinging on optical components.⁴² We used MALAPROP to calculate the electric field distribution experienced by each component during a hypothetical shot, and to predict the ultimate performance of Novette. We introduced realistic dirt and damage effects into each calculation to model typical operating conditions.

The conversion of laser light into its harmonics in nonlinear birefringent media is a well-known process that was first demonstrated more than 20 years ago. Harmonic conversion of Novette's output beams requires doubler optics with diameters greater than 74 cm. Array architectures for Novette's initial operation were dictated by the size of available KDP crystals. We built one 5- by 5-element array of 15- by 15-cm

crystals and another 3- by 3-element array of 27- by 27-cm crystals. In the Novette assemblies, the crystals and their eggcrate support structures are sandwiched between NS-processed windows to ensure that the crystals are all oriented within a 100- μ rad full angle error cone.

Novette Operation

During the summer of 1982, we began to integrate the central control system software and operation of the smaller stages of the laser. The large laser components, with optical apertures of 31.5 and 46 cm, began to arrive in August. By early September, Novette's 31.5-cm amplifier sections delivered approximately 4 TW each in separate 100-ps shots. Next, the final 46-cm amplifiers and the two frequency-conversion arrays were installed. These two 120-m optical trains had to be folded twice to fit the 70- by 15-m space previously occupied by Argus.

Fig. 2-49. Novette operational history summary.

In January of 1983, we began a series of full-power, two-beam shots. On January 24, 1983, the highest powers yet recorded for a two-beam laser were achieved: 25 TW at 1ω converted to 13 TW at 2ω for 100 ps. By this time Novette was already engaged in 100-ps x-ray laser studies. With the completion of x-ray laser studies at the end of March, Novette was converted to 1-ns operation. By mid-April it was apparent that both of our harmonic conversion arrays were being damaged. Consequently, initial 1-ns target studies had to be performed with reduced energy during May and June until corrective action eliminated the problem in July and deliverable energy rose to nominal values.

The cooling of Novette's output amplifiers and subsequent laser system realignment would limit Novette to about eight target shots during a 16-h working day. In practice, however, target and target diagnostic preparation and alignment have generally cut potential system usage by more than a factor of 2. Dramatic reductions in productivity, measured by number of target shots per month, come at the end of an experimental sequence when an entirely new target type is first introduced. Figure 2-49 suggests that shot rates in excess of 50 target shots per month could be achieved if a standard target design were used.

Integration of Novette/Nova Controls

Novette's control system is a subset of the system we are implementing for Nova; hardware elements are identical and software elements are similar. We continue to improve Nova control system computer programs, which now contain enhancements not available on Novette. Detailed descriptions of both Nova and Novette laser systems are reported in previous *Laser Annals*.^{39,40}

The control system for the laser and target facilities was proven to be flexible and reliable during 1983. Flexibility facilitated the incorporation of changes; high reliability enhanced the overall effectiveness of laser-target operations. During operation of the Novette facility, numerous control system changes were identified to provide greater visibility into laser-target performance and to make the system more reliable and easier

to operate. Each change, by itself, yields only a small improvement, but the aggregate effect is substantial. In the following discussions we highlight the hardware, software, and operational progress made on the Novette system that benefit both Novette and Nova.

Operation of the Control System. Both Nova and Novette use similar control consoles and operator displays. Differences between the two control systems relate to specific differences between the two laser systems: Nova has three VAX computers where Novette has one, Nova has four control consoles where Novette has two, and Nova has 10 laser arms where Novette has two.

We integrated the subsystem control console software so that any subsystem software could coexist with any other subsystem utilizing three control consoles. Previously, a console could service only one subsystem at a time. An operator was required to stop the subsystem software running on the desired control console and then start the selected new subsystem software. Now the operator can change from one subsystem control menu to another by touching the button displayed on any control frame to return to the root control menu. The operator then follows the control menu path to the desired subsystem control menu.

Hardware, software, and procedural changes to large operating systems such as Novette pose many problems. Seemingly innocuous changes can have unexpected effects on system operation. To ameliorate these problems, we refined our quality control provisions for hardware and software to stress configuration control. We now perform compatibility analysis on all scheduled changes before incorporating them. Software changes are implemented using three qualitative tiers: the lowest ("development") tier, for programs in development; the middle ("library") tier, for programs being tested; and the highest ("system") tier, for tested, reliable programs routinely used to control the laser/target system.

Long-Pulse Oscillator and Slicer Development. To generate pulses with durations from 1 to 20 ns, a pulse of the desired length is gated out of a much longer pulse from a Q-switched, single-axial-mode laser.^{43,44} The configuration is similar to that

Fig. 2-50. Optical layout of the long-pulse oscillator.

of a short-pulse oscillator, except the laser is not mode-locked. Instead, interactivity etalons limit oscillation bandwidth to a single axial mode of the cavity.

Two uncoupled etalons, shown in Fig. 2-50, are used to limit oscillation to a single axial mode. Etalon 1 is a 3-cm-thick, 1-in.-diam quartz optic, coated for 40% reflectivity on both faces. Etalon 2 is uncoated, 3-mm quartz. Etalon 1 is designed so that modes adjacent to the mode at the transmission peak are attenuated by no more than 1 to 2% in a single pass. Etalon 2 suppresses the other modes that are above the lasing threshold.

The cavity length must be maintained constant to $<0.5 \mu\text{m}$ to prevent mode hopping or simultaneous oscillation of two modes at the same gain level. To accomplish this, the rear mirror is moved by a piezoelectric positioner driven by an analog feedback circuit that has the prelude signal as its input. By applying a small dither signal to the piezoelectric positioner during the prelude, the position of the oscillating mode under the transmission curve of the etalons can be determined by detecting the change in prelude level. The amplitude of the dither signal is $\sim 10\%$ of the spacing between cavity modes. If the cavity mode is not tuned to the etalon mode, a monotonic increase or decrease in the prelude will be detected by the feedback circuit. An error signal is then generated and amplified to drive the rear mirror positioner and tune the cavity.

Oscillator pulses 400 ns long pass through a Pockels cell gate (slicer) to produce pulses of the desired duration. The slicer is driven by a silicon switch^{45,46} illuminated by a single 100-ps to 1-ns pulse from the short-pulse oscillator. The switch is a photoconductor that responds to high-energy pulsed illumination by dropping resistance from $\sim 50 \Omega$ to $<5 \Omega$ during the

illumination time. The switch discharges a charge line at 9 kV into a 50- Ω cable and transmits a 4-kV pulse to the Pockels cells. We can select the duration by varying the charge-line length using 50- Ω cables that are carefully connected to avoid reflection.

The switch is a high-sensitivity silicon chip with metallic surface contacts at opposite ends of the chip. If high-voltage dc is applied to the switch, a thermal runaway condition will quickly develop, resulting in breakdown. Therefore, high voltage is applied in a 600-ns pulse that is synchronized with the incident optical pulse. A fiber-optic trigger starts the high-voltage generator, and the voltage on the charge line reaches its peak when the light is applied to the switch. The charge-line voltage then drops to zero. A thyatron shorts the output of the high-voltage pulse generator and discharges the charge line (in the absence of a light pulse) after a fixed delay.

Even with the charge-line voltage applied for less than a microsecond, the switch can develop thermal runaway and break down if its initial resistance is low enough. The dark resistance is a function of silicon material, contact spacing, contact material and processing, and geometrical configuration. Due to the limited energy available from the short-pulse oscillator (100 to 200 μJ), the spacing must be small enough to allow the device to reach low on-resistance with available illumination.

Contact spacing is 2.5 mm, which works well given certain types of material and/or contacts. We have found that devices made of high-purity silicon with aluminum contacts exhibit supralinear current vs voltage characteristics, resulting in excessive dark current at high voltage. Devices of gold-doped material with similar contacts exhibit linear behavior and better performance at high voltage. Changing contact material and

processing greatly affects the device characteristics. Presently, we use gold-doped devices with aluminum contacts, which have $\sim 50 \Omega$ dark resistance at 9 kV.

The chips are silver epoxied to a copper microstrip transmission line on a fiberglass printed-circuit board with type HN connectors on either end. The switch is then cast in clear epoxy potting material. The resulting module is inserted into a holder on the optical table in the MOR. The short-pulse beam is directed onto the switch-active area, and the timing of the pulse-charging unit is adjusted to coincide with the arrival of the 100-ps pulse.

The electrical pulse from the switch typically has a rise time of < 250 ps from 10 to 90% of full amplitude. At present, the optical rise time is limited by the Pockels cells, but is approximately 500 ps. By simply reducing charge-line length, subnanosecond optical pulses can be created, overlapping pulse width range of the short-pulse oscillator.

Laser Diagnostics. The laser diagnostics data acquisition system on Novette has been operating satisfactorily, and accurately measures the fraction of laser light that reaches the detectors. Successful measurement requires the fraction of laser light to be constant from shot to shot. This has not been true on Novette where, in two cases, critical mirror coatings have been shown to vary with humidity. Because detectors are typically designed to sample

the small ($\sim 3\%$) fraction of light transmitted through turning mirrors, a small change in reflectance can result in a large change in signal. This effect has been observed at the mid-chain location when the nitrogen flow is interrupted; we can avoid this problem with a constant-humidity environment during calibration shots as well as on-target shots.

At the turning mirror that directs the diagnostic beam into the output sensors, humidity presents a similar problem because the mirror reflectance is only 7% at the 2ω wavelength and therefore leads to large humidity-induced effects. On Novette, we solved this problem by adding an energy diagnostic behind the highly transmissive mirror. On Nova, no such mirror now exists; we simplified the output packages to obtain more direct measurements—a more reliable design.

Target System Controls. Target system controls for both Nova and Novette are essentially the same—nearly all of the Novette target diagnostic hardware is reused on Nova. Target system computer programs for Nova have been implemented and validated on Novette. Successful experiments were performed to support the x-ray laser, scaling, implosion, and cryogenic target series.

We designed and implemented the control system to support cryogenic targets on the Novette laser system (Fig. 2-51). The target support equipment consists of two

Fig. 2-51. Novette cryogenic control block diagram.

cryogenic support systems, one for the target positioner and one for the shroud retractor, a heating laser, and a switched capacitor bank to power the retractor. We control the cryogenic support systems with two LSI-11/23 microprocessors, both of which communicate with the central VAX computer via Novanet. One LSI-11/23 is a standard Nova-style stepping motor controller (SMC) that controls the target positioner and shroud retractor movement. The second LSI-11/23 is a CAMAC-based system that can be run in two different configurations: first, with the retractor and positioner sealed together, or second, with the shroud not sealed against the target positioner. In either configuration, we monitor six temperature-sensing diodes, two thermocouples, one thermistor as a reference temperature for the thermocouples, five pressure sensors, three flow sensors, and three user-selectable set points. Based on these transducers, we control flow rates, pressures, and temperatures throughout the cryogenic system.

System control can be either manual or automatic. The system can also warn the operator of illegal or bad conditions, such as low cryogen levels, out-of-range flow rates or pressures, etc. We have furnished an automatic setup mode that pumps the system down and purges various cryogen lines before operation. We also support an automatic control mode in which the system is capable of changing various set points and pressures to achieve a predefined target temperature.

When the system is running, all major functions can be controlled from the VAX, which stores and archives all cryogenic control data, parameters, and valve positions.

In 1983, we corrected several hardware and software problems. We integrated several devices into Novette: a fiducial system designed for Nova, a target-plane imager, a fiber-optic bundle, and a multiple-image camera.

Power System Controls. Operation of the Novette pulsed power system has proven to be routine; operation is successful in 95% of the laser shots.

As 1983 progressed, the occurrences of high-voltage, high-ampere fuses blown in the energy storage circuits increased, primarily with the 5000-A fuses used for the

31.5-cm amplifiers. We found that these fuses were being operated too close to their current ratings and that they were deteriorating with time and use. We replaced them with 7000-A fuses. We also encountered failure of several pulse-forming inductors—the inductor casing would crack after an unspecified period of use. A change in the mold design should correct this fault. Other changes made to Novette that will be incorporated in Nova include new flashlamp design, modified flashlamp current diagnostics, and the rewiring of high-voltage fanout racks to prevent internal arcing.

Hardware and Software Change Management. The Novette control system matured to the point where the primary emphasis of development was to support the day-to-day operation of the laser and target facilities. As daily operation of Novette relied more heavily on the control system, changes to the system had greater effects on operations. Also, as the pace of Nova controls development increased, so did the numbers of suggestions for changes in the design based on experience with Novette.

To help both controls developers and users to identify and clarify issues, we introduced a formal communication mechanism. The procedure is based on the use of a simple form, called a controls change request (CCR). The CCR is used by both Nova and Novette controls users and developers to document needed changes to software and hardware. The form and procedures for its use are included in the Nova software quality assurance plan.

Since we introduced the CCR procedure, an average of five CCRs have been submitted and processed each month. Only a few requests were specifically addressed to Nova controls. However, we are applying 26% of the requests to change Novette directly to Nova designs; nearly all Novette requests are being applied to the Nova system at some level. Thirty-seven percent of the Novette CCRs involved hardware only, 42% requested software changes, and 21% required both hardware and software changes.

We have resolved 58% of all Novette requests; the remaining 42% are either in progress or are waiting for manpower to

become available. We have resolved 70% of requests that involved only software, and have completed 56% of hardware requests. Only 10% of the requests that involve both hardware and software have been resolved.

We continue to improve the CCR mechanism. However, the changes do not replace the interchange between users and developers. The quantity and quality of this communication is increasing.

Performance of Novette Compared to Theoretical Models

The laser pulse that eventually reaches the target emerges from Novette's actively mode-locked, Q-switched neodymium:yttrium-lithium-fluoride (Nd:YLF) oscillator as one of a train of about 20 pulses at 8-ns intervals. Each pulse exits the master oscillator with subnanosecond accuracy ($\pm 2\%$ in energy and $\pm 1\%$ in pulse duration) relative to the Q-switch time and is consistently selected by means of a Pockels cell gate. Future Novette experiments will require longer shaped pulses produced by the system described above.

Triggered by Novette's computer control system, the master oscillator and preamplifier system provides all of the subnanosecond

timing and trigger pulses required by the laser and target systems. In addition, the computer control system automatically measures and aligns the beam and shapes the pulses in time and space. In 1983, the devices installed in Novette's master oscillator system have provided temporally Gaussian pulses ranging from 100 ps to 1 ns.

At the preamplifier output, the single optical pulse is split into two, whose relative timing is adjusted by means of variable-path-length sections and whose amplitudes are set using waveplate/polarizer pairs. The wings of the spatially Gaussian beam are next truncated so that only the central 26-mm-diam portion is allowed to pass into the amplifier chains. A typical beam profile, measured just after the first spatial filter, is in Fig. 2-52.

The first 28-cm-long by 5-cm free-aperture rod amplifier is presented with a circular beam having an average intensity of about 64 MW/cm^2 . Following the rod amplifiers, the beam passes through relay-ing spatial filter and glass-disk amplifier stages of increasing size. The first two amplifier stages, which have 9.4- and 14.5-cm clear apertures, respectively, were removed from Shiva and refitted with phosphate glass disks. The 20.8-cm-diam stage uses the more efficient box amplifier design.

Fig. 2-52. (a) Beam profile at Novette's input sensor diagnostic package, (b) beam's measured power distribution function.

The large Nova box amplifiers store approximately 2% of the electrical energy in the capacitor bank as inversion in the disks at pulse-propagation time. Following the 20.8-cm stage, the beam is expanded to 31.5 cm, passes through the final Faraday rotator isolator in the chain, and is turned by two mirrors down the final leg of the spaceframe, which holds four 31.5-cm amplifiers and three 46-cm output amplifiers. Each of these amplifiers contains two disks.

Gain depletion by amplified spontaneous emission (ASE) or nonresonant parasitic oscillations in each disk impose a size limitation on glass disks. With well index matched absorbing cladding on the disk edges, ASE depumping limits the gain-length product, al , along the major axis of the disk to between 3 and 4. To preserve high stored energy, the disks in the largest amplifiers, those with 46-cm apertures, are split by an absorbing stripe along their minor axes. The photographs in Fig. 2-53 were taken at Novette's output for shots with approximately 10, 11, and 13 TW per arm. The prominent dark shadow bisecting each beam is caused by this absorbing region in the 46-cm amplifier disks. So long as these and other necessary absorbing regions produce no intense diffraction fringes, their effect on the beam is only to reduce its area and thus its total energy. Three uncontrolled damage sites, about two thirds of the way through the laser, produced the spots on the right hand side of the photos. Diffraction fringes arising from these sites grow dramatically as the output power is increased. Very intense filaments can destroy the optical quality of component

surfaces, as well as damage the bulk material. The level of beam break-up that threatens optics is also sufficient to impair the focusability of the beam and can render the laser useless for target studies at these powers.

Even as it has proven to be impractical to model every target shot with two-dimensional fluid dynamics codes such as LASNEX, so it has been unreasonable to calculate all aspects of each Novette firing ahead of time with MALAPROP.⁴¹ Given the groundwork laid by MALAPROP, simpler simulators should be capable of steering us away from the more obvious damage risks; such routines can now be run before each Novette laser shot.

Infrared Output Laser Beams. Each power amplifier chain in Nova or Novette starts with a 26-mm-diam aperture and ends with a focusing system that concentrates the light from that chain onto a target. Energy and beam profile data are routinely obtained by the optical diagnostic packages in each chain. Given the (easily measured) input-beam statistics, we can estimate the output-beam statistics and compare them with observations.

Most of the nonlinear optical effects associated with small scale self-focusing will arise in the thicker output components of the amplifier chain. The first simulation challenge is to correctly model the large-signal gain of this power amplifier.

We now have an extensive body of data on the performance of the Novette laser system when propagating 100-ps and 1-ns pulses. The importance of these data lies in our ability to verify some assumptions

Fig. 2-53. Deterioration of Novette's 74-cm-diam beams with increased output power.

Fig. 2-54. Novette's measured performance at 100 ps, and 1 ns.

about the effect of saturation on the ultimate performance of Novette and of Nova. The data demonstrate reasonable agreement with curves generated using published values for the saturation fluence.

Our calculations predicted that 100-ps pulses at intensities consistent with a chain output power of 11 TW would contain too little energy to display gain saturation effects. In early experiments this year, Novette's performance confirmed that 100-ps gains for each portion of the laser system are essentially constant. Short-pulse laser output power was limited, instead, by B growth effects, seen in Fig. 2-54(a) as a slight roll-off in the output energy above 1.2 kJ.

In April 1983, we began target experiments requiring 1-ns pulses. We predicted a significant decrease in gain at the higher energy fluence possible with longer pulses. This was confirmed by the data plotted in Fig. 2-54(b). As expected, there is a decrease of nearly a factor of 2 in gain from the lowest to the highest energy. The theoretical curve shown is calculated using a fit to LHG-8 phosphate saturation data,⁴⁷ which indicate effective saturation fluxes of 4.02 to 4.05 J/cm² over the modes' range of fluxes so far encountered by Novette.

We plotted the ratio of the mid-chain and mid-disk sensor energy readings against the mid-disk energy to show the performance of the 20.8-cm disk amplifiers (Fig. 2-55). A curve fitted through these

Fig. 2-55. Novette south arm mid-chain gain (1-ns pulse width).

data, corrected for passive losses, determined the loss-free small signal gain of the 20.8-cm disk amplifiers. Using the familiar Franz-Nodvik expression⁴⁸

$$F = S \ln[1 + G \{\exp(E/S) - 1\}] \quad (6)$$

where E is the input fluence, F the output fluence, G the loss-free small-signal gain, and S the saturation fluence, we calculated the saturation fluence as a function of output fluence. The calculation included a correction for the internal amplifier losses. In Fig. 2-56, the saturation values are

compared with data published in Refs. 48 and 49.

At longer pulse lengths, gain saturation appears as a modification of the pulse shape. In Fig. 2-57, the response of the gain to a rectangular input pulse nearly 18-ns long is the result of gain saturation in the 20.8-cm disk amplifiers for an output energy of 1270 J. For the simple saturating amplifier, gain is given by

$$\frac{P_E(t)}{P_E} = \frac{e^{P_E t / S}}{e^{P_E t / S} - 1 + \frac{1}{G}} \quad (7)$$

Fig. 2-56. Saturation fluence data from Eq. (6), and published data for LG-750 and LHG-8 glasses.

Fig. 2-57. Ratio of mid-chain power to input power for an 18-ns, approximately rectangular input laser pulse.

Fig. 2-58. Output beam profile on an 11-TW pulse: (a) photo; (b) scan.

Assuming that the saturation fluence is the linearly increasing function of output fluence of Fig. 2-56, determined from the 1-ns data, we may calculate the gain variation shown by the solid curve in Fig. 2-57 using the approximate expression

$$G(t) \approx \frac{G_0}{1 + G_0 \frac{P_E t}{S}} \quad (8)$$

Although the data we have are not sufficient to determine the actual value of the saturation fluence to better than $\pm 10\%$, in this case the agreement is particularly good.

Laser flux on each component has been normalized to measurements taken from typical shots. Such calculations have often been used to alert the laser system operator: when a particular choice of gains and pulse durations is likely to induce optical damage somewhere in the system.^{50,51}

Since November of 1982, we have accumulated photographs of the output infrared beams presented to Novette's 2ω crystal arrays at the rate of about two per day. Virtually all of these images, like the 10-TW sample in Fig. 2-53, are quite uniform in appearance, with slight evidence of nonlinear breakup. We have densitometered and analyzed several of the images, most notably the higher power beam profiles such as the 11-TW shot in Fig. 2-58. We constructed film sensitivity curves from witness samples of film from the same batch and processed in the same way as the output

beam photos (such as that shown in Fig. 2-58). Figure 2-59 is a typical intensity-weighted distribution function derived from an 11-TW beam. Small scale self-focusing shows up in this figure as a long, high-intensity tail, which carries less than 20% of the output infrared power. The specific film calibration chosen changes this distribution somewhat. Generally, the 20 and 50% points remain almost unchanged while the data reduction of the highest-power, most-saturated regions of the photographic plate can lead to as much as a 20% uncertainty in locating the 80% point.

Statistical optics estimates were used to predict the statistics of Novette's beams.⁵² Figure 2-52 is a typical intensity profile measured by one of Novette's input sensor packages. The beam at this station might reasonably be approximated by Gaussian statistics with an intensity standard-deviation-to-mean ratio of between 0.1 and 0.2. This low-signal portion of Novette experiences neither high flux nor high intensity, so that only passive phase errors and diffraction influence the beam. Although the beam varies with each alignment of the master oscillator and preamplifier, its statistics are reliably stable. Given the measured small-signal gains,⁵⁰ and a beam input like that in Fig. 2-52, we calculated the two output infrared probability distribution functions shown in Fig. 2-60.

The measured data points are taken from two laser power distribution functions. Because the near field experienced by optical

components has been treated as a complex Gaussian random variable, the laser intensity must be "Rician"; i.e., the probability density function of the laser intensity takes the form

$$P(I) = \frac{1}{\sigma_A^2} \exp\left(-\frac{I + I_s}{\sigma_A^2}\right) I_0 \frac{2\sqrt{I I_0}}{\sigma_A^2} \quad (9a)$$

for $I > 0$, and

$$P(I) = 0 \quad \text{for } I < 0 \quad (9b)$$

Fig. 2-59. Laser output power distribution function.

Fig. 2-60. Calculated (broad curves) and measured (data points) laser power distribution functions.

where σ_A^2 is the noise intensity, I_s is the signal intensity, I is the random variable intensity, and I_0 is the first modified Bessel function.⁵³ The distribution of phase error across the beam remains essentially Gaussian; consequently, far-field focal distributions have been modeled as the sum of two spatial Gaussians.

In some applications, the ratio of the intensity standard deviation to its mean is of interest. The variance of the total intensity is

$$\sigma_I^2 = \langle I^2 \rangle - \langle I \rangle^2 = \sigma_A^4 + 2I_s\sigma_A^2 \quad (10)$$

so the contrast, $\sigma_I/\langle I \rangle$, may be written

$$\frac{\sigma_I}{\langle I \rangle} = \frac{\sqrt{\sigma_A^4 + 2I_s\sigma_A^2}}{I_s + \sigma_A^2} \sim \sqrt{\frac{2}{I_s}} \sigma_A \quad (11)$$

The preceding leads one to expect that the rms ripple gain through a lossless slab of dielectric material could be approximated by the ratio of the output contrast to the input contrast; i.e., rms ripple gain $\approx e^{B/4}$. This contrast gain shows a growth rate four times slower than the worst possible ripples. The fastest growing ripples experience a gain of e^B (Ref. 54). Although these fastest growing modes propagate at angles of one to several milliradians from the main beam, and are thus well outside the bandpass of Novette's spatial filters, they can be excited

by damage sites such as that in Fig. 2-53 and deplete the beam locally. Monitoring only rms ripple amplitude can therefore give the operator or designer a false sense of security.

The meaning of the laser power distribution must be clarified because the statistical optics model adopted presumes that the beams are infinite in extent. Although passive diffraction prevents the propagation of perfectly uniform or flat-topped laser beams in any laser, well designed laser chains mitigate diffractive effects through optical relaying. Novette's amplifier chains have seven relay spatial filters that permit approximately 85% of the amplifier's aperture area to be used efficiently. The product of this fill factor and the physical amplifier aperture area is the effective beam area used by the simulation to calculate output energy. In effect, an arbitrarily chosen region of the (x, y) plane, having the effective beam area, simulates the beam. The calculated power distribution function is given by

$$F(I) = (I_s + \sigma_A^2)^{-1} \int_0^I I' p(I') dI' \quad (12)$$

When reducing film data, the measured film density is first converted to optical intensity point by point with a film calibration file. A histogram is then created by counting the pixels within given intensity intervals. To avoid the problem posed by large areas of blank film leading to large pixel counts at zero intensity, the number of pixels in each bin is multiplied by the mean intensity for that bin. Power, as opposed to intensity, distribution functions like those in Figs. 2-58 and 2-59 are found by integrating the histograms over intensity.

Figure 2-60 shows that the "Rician" has the feature that, as the signal to noise ratio is reduced, the density function is increasingly dominated by a high intensity tail. This behavior is evident in Fig. 2-60 and is explicit in Fig. 2-61. Reduced statistics on Novette's 0.1-ns infrared output pulses are compared in Fig. 2-61 to contours of $F = 0.8, 0.5$, and 0.2 . Nonlinear optical effects increase the parametric optical spatial noise at the expense of the signal so that, for example, the intensity below

Fig. 2-61. Intensities at the 20, 50, and 80% points of the Novette 1ω power distribution function (at SHG arrays, for 100-ps pulses).

which 80% of the laser power is delivered ($F = 0.8$) grows more rapidly than total output power.

When we converted Novette to 1-ns operation, we added another 46-cm-aperture amplifier. The additional glass in the highest flux region of the chain hastens the onset of small-scale beam breakup by increasing B . Figure 2-61 displays measured and calculated distributions for the Novette laser in this new configuration.

As expected, beam breakup becomes significant at slightly lower output power, and risk of damage limits safe operation to <9 TW in 1 ns rather than <12 TW in the 100-ps case (Fig. 2-62). MALAPROP can also produce power distribution functions by imitating the procedure used in reducing the photographic data. Figure 2-63 compares actual near-field 9-TW output beam statistics with corresponding MALAPROP output and statistical optics approximations.

Frequency Conversion. Before Novette's laser pulses reach the target, they are frequency doubled in type II KDP crystals.⁵⁵ Frequency conversion efficiency for an ideal plane wave can be described by the relation

$$\eta = \tanh^2 \left\{ \frac{1}{2} \tanh^{-1} \left[\operatorname{sn} \left(2\eta_0^{1/2}, \frac{1 + \delta^2}{4\eta_0} \right) \right] \right\} \quad (13)$$

where δ is the phase retardation due to offset of the crystal axis orientation, η_0 is the small-signal efficiency for $\delta = 0$ on exact phase matching, and sn is one of the Jacobi elliptic functions.⁵⁶ If the beam impinging on Novette's frequency conversion apparatus is assumed to be made up of a large number of beamlets, each propagating in a slightly different direction and each having a different intensity, a first-order estimate of frequency conversion may be made by averaging over intensity and angle. The spread of intensities in Novette's beams dominates the conversion process. Figure 2-64 compares the expected value of the harmonic intensity to observed conversion efficiencies. The calculated curve was obtained by numerical integration of the relation

$$\frac{I_{2\omega}}{I_{1\omega}} = \int_0^{+\infty} p(I') dI' \quad (14)$$

Fig. 2-62. Same as Fig. 2-61, but for 1-ns pulses.

Fig. 2-63. MALAPROP calculations, statistical output approximation, and measured 9-TW near-field power distribution.

Fig. 2-64. Predicted and measured efficiency of $1\omega \rightarrow 2\omega$ conversion.

Fig. 2-65. Total 2ω energy on-target vs laser pulse duration.

derated by known window and KDP crystal attenuations. Any additional angular detuning of the crystals (typically aligned to within $\pm 25 \mu\text{rad}$), fabrication error when aligning phase match direction in different crystals in each segment of Novette's mosaic arrays (typically $< 50 \mu\text{rad}$), or polarization misalignment ($\pm 0.5 \text{ deg}$), should reduce conversion, so this estimate is an upper bound.

Damage Limit Assessment. We can estimate risk to the system with the specification of a mean and variance for damaging flux, together with an intensity distribution for the laser field at each component. The risk to a given component may be quantified as the probability of observing damage following exposure to the beam; i.e.,

$$\text{risk} = \int_0^{+\infty} p_D(I) \int_0^I p(I') dI' dI \quad (15)$$

where $p_D(I)$ is the probability density function for that component's optical damage. Because the optical damage threshold typically increases as the square root of the irradiating laser's pulse duration, there is always less chance of damage in well

designed systems when they are required to propagate short pulses. Figure 2-65 quantifies this situation for Novette by indicating as safe the energy on target below which the risk of damage to any component is less than about 10%. The shaded interval in Fig. 2-65 ends at a risk level of about 15%. Operation in this interval presents nonnegligible, but still acceptable, risk to the laser. The frequency conversion arrays, the output turning mirrors that must route the pulse to the target, and the input lenses to the output spatial filters seem to be the components most at risk. As Fig. 2-65 shows, Novette is capable of energy fluxes above this shaded acceptable risk band at pulse lengths longer than about 0.6 ns.

Summary

Novette is the latest embodiment of the rapid evolution of powerful ICF laser systems. Each of its two relatively compact arms has exceeded the total output of all of Shiva's 20 arms. And Novette was assembled in well under a third of the time required to build Shiva and has achieved a shot rate on target exceeding Shiva's. An evolutionary step of this magnitude in so short a time approaches a revolution in high-power laser technology.

As a test bed for the Nova laser, Novette has provided the first operational test of split disk amplifiers and of harmonic generation in large-aperture, multi-element KDP crystal arrays.

Authors: K. R. Manes, D. J. Christie, A. J. DeGroot, D. G. Gritton, J. S. Hildum, J. E. Krammen, J. B. Richards, W. W. Simmons, and R. B. Wilcox

21. 1981 *Laser Program Annual Report*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-50021-81 (1982), pp. 2-68 to 2-69.
22. 1981 *Laser Program Annual Report*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-50021-81 (1982), pp. 2-69 to 2-70.
23. 1980 *Laser Program Annual Report*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-50021-80 (1981), pp. 2-128 to 2-131.
24. 1982 *Laser Program Annual Report*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-50021-82 (1983), p. 2-24.
25. 1982 *Laser Program Annual Report*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-50021-82 (1983), p. 2-17 to 2-19.
26. 1982 *Laser Program Annual Report*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-50021-82 (1983), p. 2-20.
27. 1982 *Laser Program Annual Report*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-50021-82 (1983), pp. 2-31 to 2-32.
28. 1977 *Laser Program Annual Report*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-50021-77 (1978), pp. 2-95 to 2-96.
29. 1980 *Laser Program Annual Report*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-50021-80 (1981), p. 2-122.
30. 1982 *Laser Program Annual Report*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-50021-82 (1983), pp. 2-22 to 2-23.
31. 1982 *Laser Program Annual Report*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-50021-82 (1983), p. 2-23.
32. 1979 *Laser Program Annual Report*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-50021-79 (1980), pp. 2-114 to 2-118.
33. 1980 *Laser Program Annual Report*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-50021-80 (1981), pp. 2-125 to 2-127.
34. J. Weaver, J. Pasternak, and L. Fields, *Dynamic Testing of Nova Laser Structural System*, Lawrence Livermore National Laboratory, Livermore, Calif., 2nd International Modal Analytical Conference, July 1, 1983.
35. K. Manes et al., *Novette Monthly Progress Report*, Lawrence Livermore National Laboratory, Livermore, Calif., OPS 83-227, July 18, 1983.
36. 1981 *Laser Program Annual Report*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-50021-81 (1982), p. 2-5.
37. F. W. Holloway, G. J. Suski, and J. M. Duffy, *Experience with a High Order Programming Language (Similar to Ada) on the Development of the Nova Distributed Control System*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-53466 (1983).
38. F. W. Holloway, *Nova Object Subordinate Quality Assurance Program Plan*, Control System Software, Lawrence Livermore National Laboratory, Livermore, Calif., M-78-060A-3357H, October 12, 1983.
39. 1982 *Laser Program Annual Report*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-50021-82 (1983), pp. 2-53 to 2-81.
40. 1981 *Laser Program Annual Report*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-50021-81 (1982), pp. 2-2 to 2-27.

References

41. G. R. Willenison, N. J. Brown, and L. M. Cook, "Scaling Up the Neutral Solution Antireflection Process," *Optical Engineering*, **22**(4) (1983).
42. W. W. Simmons, J. T. Hunt, and W. E. Warren, "Light Propagation Through Large Laser Systems," *JQE QE-17*(9), 1727 (1981).
43. 1980 Laser Program Annual Report, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-50021-80 (1981), p. 2-224.
44. D. J. Kuizenga, "Short-Pulse Oscillator Development for the Nd:Glass Laser-Fusion Systems," *IEEE Quant. Electron.*, **QE-17**(9), (1981).
45. R. B. Wilcox, "Silicon Switch Development for Optical Pulse Generation in Fusion Lasers at Lawrence Livermore National Laboratory," *Proceedings of SPIE Conference on Picosecond Electronics*, vol. 439, p. 112, San Diego, Calif., 1983.
46. G. Morou and W. Knox, "High Power Switching With Picosecond Precision," *App. Phys. Lett.*, **35**, 492 (1979).
47. J. L. Emmett, W. F. Krupke, and J. B. Trenholme, *The Future Development of High-Power Solid State Laser Systems*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-53344 (1983).
48. W. E. Martin and E. Milam, "Gain Saturation in Nd: Doped Laser Materials," *IEEE J. Quantum Electronics QE-18*(7), 1155 (1982).
49. S. M. Yarema and D. Milam, "Gain Saturation in Phosphate Laser Glasses," *IEEE J. Quantum Electronics QE-18*(11), 1941 (1982).
50. 1982 Laser Program Annual Report, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-50021-82 (1983), pp. 2-56 to 2-57.
51. 1982 Laser Program Annual Report, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-50021-82 (1983), p. 2-61.
52. K. R. Manes and W. W. Simmons, *Statistical Optics Applied to High Power Glass Lasers*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-90158 (1984).
53. J. W. Goodman, "Statistical Properties of Laser Speckle Patterns," *Laser Speckle and Related Phenomena; Topics in Applied Physics*, vol. 9, J. C. Dainty, Ed. (Springer-Verlag, New York, N.Y., 1975).
54. D. C. Brown, "High Peak Power Nd:Glass Laser Systems," *Optical Sciences*, **25**, 188 (1981).
55. J. F. Holzzrichter, D. Eimerl, E. V. George, J. B. Trenholme, W. W. Simmons, and J. T. Hunt, *High Power Glass Lasers*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-52868, Rev. 1 (September 1982).
56. M. Abramowitz and I. A. Stegun, *Handbook of Mathematical Functions* (Dover, New York, N.Y., 1964) p. 567.

Laser rays propagating in vacuum, all with the same nonzero angular momentum about the z -axis, lie on a hyperboloid of revolution. In three-dimensional space, each ray moves in a straight line, but in (r, z) coordinates each ray traces out a hyperbola.

Target Design 3

Target Design

Contents

Section Editor: L. J. Suter

Introduction	3-1
Laser Targets	3-2
Evolution of Rayleigh-Taylor Instabilities in Multilayered Fluids	3-2
Simple Models of High-Gain Targets—Comparisons and Generalizations	3-5
Model for Gain vs Laser Energy in Direct-Drive Targets	3-9
Charged-Particle Fusion	3-13
Introduction	3-13
Studies for the High-Temperature Experiment in Heavy-Ion Fusion	3-13
Testing Hydrodynamic Models for Simulation of Pinched Beams	3-21
Plasma Physics	3-25
Introduction	3-25
Stimulated Raman Scattering in an Inhomogeneous Plasma with an Obliquely Incident Pump	3-26
Thresholds of Convective Raman Backscattering: Effects of ∇n , Noise, Collisions, and Landau Damping	3-28
The $2\omega_{pe}$ Instability in Current Experiments	3-32
Stabilization of Filamentation by Angular Spread	3-35
2-D LASNEX Ray-Trace Calculations of Thermal Whole-Beam Self Focusing	3-36
Beat-Wave Heating in Inhomogeneous Plasma Irradiated with 0.53- and 1.06- μm Light	3-40
Reheating of Hot Electrons by Laser-Excited Plasma Waves	3-42
Direct-Implicit Plasma Simulation	3-46
Theory of Photodiodes	3-48
LASNEX and Atomic Physics	3-50
Overview	3-50
Three-Dimensional Laser Ray Tracing on an r - z Lagrangian Mesh	3-51
Thomas-Fermi-Based Equation of State Package	3-55
Thermal Occupation and Pressure of Resonance States	3-59
Convergence Acceleration for Multigroup Radiation-Matter Coupling Iteration	3-62
Energy Spectrum of Thermonuclear Neutrons	3-64
References	3-67

Target Design

Introduction

J. D. Lindl

The Target Design Program combines the efforts of the Plasma, Code Development, and Design Groups in three basic areas:

- Developing theories of beam-plasma interaction, implosions, and thermonuclear microexplosions.
- Building plasma and implosion-burn computer codes.
- Using these theoretical and computational tools to design targets and simulate experiments.

Here are some highlights of our work in 1983.

A major focus of our work in target design was on implosion, hydrodynamic scaling, and plasma-scaling targets for ongoing experiments on Novette. Hohlraum plasma-scaling experiments extended the Argus data base at 2ω and confirmed our predictions of the utility of short-wavelength light for decreasing the production of hot (suprathermal) electrons in hohlraums. This reduction made it possible to design Novette targets that achieved compressions 2 to 3 times greater than were obtained with similar targets on Shiva.

We began detailed design work for Nova capsules whose implosions will be hydrodynamically equivalent to those of high-gain capsules. An important feature of this analysis is the trade-off between hydrodynamic instability and plasma instability for targets that are consistent with high gain. We have developed an analytic model that indicates that direct-drive targets of this type will either be hydrodynamically inefficient and subject to plasma instabilities or will break up because of fluid instabilities. The high intensity required to drive a thick (hydrodynamically stable) shell leads to a significant distance between the critical-density and ablation surfaces, which lowers the hydrodynamic efficiency. To achieve hydrodynamic stability at lower intensity, we would have to reduce the growth rate of fluid instabilities by a greater amount than currently appears possible on the basis of theory or numerical simulation. But radiation-driven implosions, which form the basis for our high-gain capsules, are consistent with current estimates of stability, symmetry, and implosion efficiency, and remain our favored approach.

We significantly improved the physics in LASNEX. We have corrected a long-standing instability in the artificial viscosity. This greatly improves our ability to carry out two-dimensional calculations of high-convergence shells. In addition, we have nearly completed a scheme for three-dimensional laser ray tracing, and a Thomas-Fermi-based equation-of-state package.

We increased our effort on development of an advanced plasma code. We are constructing an implicit two-dimensional electromagnetic particle code that will extend our ability to simulate low-frequency kinetic phenomena.

The following articles report details of our progress on laser targets, ion-beam targets, plasma physics, atomic physics, and code development. Our work on the design and interpretation of experiments is reported in Section 5, "Fusion Experiments."

Laser Targets

Evolution of Rayleigh-Taylor Instabilities in Multilayered Fluids

Introduction. Fluid instabilities are a critical issue in the inertial-confinement fusion (ICF) program. As the shell accelerates inwards and compresses the DT fuel, small imperfections on the shell's outer surface can grow large enough to cause breakup of the shell before peak compression is achieved. Unstable growth of imperfections on the inside of the shell causes mixing of high-Z shell material into the DT fuel, degrading the yield. These problems arise even in single-shell targets, such as that shown in Fig. 3-1.

Rayleigh-Taylor instabilities are among those of concern in ICF. Last year¹ we reported how they can be suppressed by the use of graded-density anti-mix layers. Now we have developed a formalism to calculate the evolution of these instabilities in multilayered fluids of arbitrary density profiles. Our equations permit us to determine the extent to which perturbations feed through from one surface to the other as the shell accelerates inward. This will permit us to study shell breakup and to estimate the extent of mixing in ICF targets.

Our formalism assumes linearity, incompressibility, no heat transfer, no surface tension, no viscosity, and constant acceleration. This article briefly describes the theory²⁻⁵

Fig. 3-1. Perturbations on either surface of the tamper of a single-shell ICF capsule may grow during implosion, causing mixing or shell breakup.

and gives some examples of its application to cases of interest in ICF.

General Equation. Figure 3-2 shows the general system to which our equations apply. The geometry assumed is planar, not spherical. The system consists of N fluid layers with densities $\rho_1, \rho_2, \dots, \rho_N$ and thicknesses $t_1, t_2, \dots, t_N$, between which lie $N - 1$ boundaries numbered $1, 2, \dots, N - 1$. The perturbation at the interface between layers i and $i + 1$ is assumed to have amplitude $\eta_i(\tau)$; for simplicity, we treat here the relatively simple case in which all perturbations have the same wavelength λ . We consistently use τ to indicate time.

Initial (i.e., $\tau = 0$) conditions are specified through the initial amplitudes $\eta_i(0)$ and their first time derivatives $\dot{\eta}_i(0)$ at each interface.

The amplitudes $\eta_i(\tau)$ are obtained by an expansion in terms of the eigenvalues and eigenfunctions of the system.²⁻⁵ The eigenvalue equation reads:

$$\mathbf{M}\mathbf{W} = \frac{gk}{\gamma^2} \mathbf{W} \quad (1)$$

where $\mathbf{M}$ is the tri-diagonal matrix given in Refs. 1, 2, and 3; $\mathbf{W}$ is the eigenfunction, a vector whose components are the velocity amplitudes at the $N - 1$ interfaces; g is the (constant) acceleration; $k = 2\pi/\lambda$; and γ is the growth rate. Equation (1) is a shorthand for $N - 1$ linear equations, in which each interface is coupled to its adjacent interfaces. Two distant interfaces communicate with each other via signals transmitted through the intermediate layers.

As discussed in Refs. 1, 2, and 3, we find $N - 1$ eigenvalues γ_l and associated eigenfunctions $\mathbf{W}_l(y)$. The eigenfunctions are functions of y (see Fig. 3-2); we let $W(i, l)$ denote the value of the l th eigenfunction at interface i . The normal modes, labeled by the index l , grow exponentially in time. Depending on the density gradient, we also find modes that oscillate in time, given by imaginary values of the γ_l .

References 4 and 5 report our work in 1983 in detail. There we derive the equations for the evolution of the perturbations $\eta_i(\tau)$ at each interface i in terms of the initial conditions at the interfaces. These

equations are

$$\eta_i(\tau) = \sum_{l=1}^{N-1} \sum_{j=1}^{N-1} W(i, l) W^{-1}(l, j) \times \left[\eta_j(0) \cosh(\gamma_l \tau) + \frac{1}{\gamma_l} \dot{\eta}_j(0) \sinh(\gamma_l \tau) \right] \quad (2)$$

The double sum in Eq. (2) implies that the evolution of the perturbation $\eta_i(\tau)$ at each interface i is influenced by all the modes l and by the initial conditions $\eta_j(0)$ and $\dot{\eta}_j(0)$ at all the interfaces j . The amplitudes of the perturbations at the outer and inner surfaces, for example, are given by $\eta_1(\tau)$ and $\eta_{N-1}(\tau)$.

Equation (2) indicates that a perturbation at a single interface will eventually induce perturbations in the rest of the system. Assume, for example, that all $\eta_j(0)$ and $\dot{\eta}_j(0)$ are zero except for some $\eta_k(0)$. At time τ there will be perturbations at each interface i , given by

$$\eta_i(\tau) = \eta_k(0) \sum_{l=1}^{N-1} W(i, l) W^{-1}(l, k) \times \cosh(\gamma_l \tau) \quad (3)$$

How large these perturbations grow depends on the wavelength, the density profile, and the initial conditions. The weighting factor $W(i, l) W^{-1}(l, k)$, in which W^{-1} is the inverse of the matrix whose elements are the $W(i, j)$, automatically accounts for the fact that interfaces close to the k th interface feel its effect more strongly.

We have written a code that calculates $\eta_i(\tau)$ for a given density profile, wavelength, and initial conditions. We have checked the code, which can handle an arbitrary number of fluids, for the cases $N = 2$ and $N = 3$, for which explicit expressions for the growth rates can be obtained.

$N = 2$. For $N = 2$, Eq. (3) reduces to the well-known expression for the classical case.^{6,7} The system consists of two semi-infinite fluids of densities ρ_1 and ρ_2 ;

perturbations at their common interface evolve according to the equation

$$\eta(\tau) = \eta(0) \cosh(\gamma \tau) + \frac{1}{\gamma} \dot{\eta}(0) \sinh(\gamma \tau) \quad (4)$$

where γ is the classical growth rate

$$\gamma_{\text{class}} = \left(gk \frac{\rho_2 - \rho_1}{\rho_2 + \rho_1} \right)^{1/2} \quad (5)$$

$N = 3$. For $N = 3$ the system, shown in Fig. 3-3(a), consists of two semi-infinite boundary fluids of densities ρ_1 and ρ_3 and an intermediate fluid of density ρ_2 and thickness t_2 . Taylor⁷ treated the case $\rho_1 = \rho_3 = 0$. For arbitrary ρ_1 , ρ_2 , and ρ_3 , we find that the two eigenvalues are given by

$$\left(\frac{\gamma^2}{gk} \right)_{\pm} = \frac{-b \pm (b^2 - 4ac)^{1/2}}{2a} \quad (6)$$

Fig. 3-2. We follow the growth of perturbations of initial amplitude $\eta_i(0)$ and wavelength λ at the $N - 1$ interfaces i between N fluid layers of arbitrary densities and thicknesses, under a constant acceleration $\vec{g} = g\hat{y}$.

Laser Targets

Fig. 3-3. (a) The case $N = 3$. (b) For fixed ρ_1 and ρ_3 , and for all λ , the growth rates $(\gamma^2/gk)_\pm$ have extrema at $\rho_2 = (\rho_1\rho_3)^{1/2}$.

Fig. 3-4. The case $N = 5$, with densities $(0, 1, 3, 1, 0)$, thicknesses $(\infty, t, t, t, \infty)$, and arbitrary perturbation amplitudes $\eta_1, \eta_2, \eta_3, \eta_4$.

where

$$\begin{aligned} a &= (1 + ST)(\rho_3 + \rho_1) + S(\rho_2 + \rho_1\rho_3/\rho_2) \\ b &= -(1 + S + ST)(\rho_3 - \rho_1) \\ c &= \frac{S}{\rho_2}(\rho_3 - \rho_2)(\rho_2 - \rho_1) \end{aligned} \quad (7)$$

with $S = \sinh(kt_2)$ and $T = \tanh(kt_2/2)$. For $\rho_1 = \rho_3 = 0$, we recover Taylor's result $(\gamma^2/gk)_\pm = \pm 1$.

Holding ρ_1 and ρ_3 fixed and non-zero, we have studied the two eigenvalues given by Eq. (6) as functions of the intermediate density ρ_2 . As shown in Fig. 3-3(b), the larger growth rate $(\gamma^2/gk)_+$ has a minimum at $\rho_2 = (\rho_1\rho_3)^{1/2}$, and the smaller growth rate $(\gamma^2/gk)_-$ has a maximum at the same place. In fact, we find (algebraically) that this is so for all perturbation wavelengths.

This implies that the middle fluid layer would perform best as an anti-mix layer if its density were the geometric mean of the two adjacent densities. More elaborate applications of our technique with multiple layers, and with continuous density profiles, show that gentle density gradients can help suppress the growth of Rayleigh-Taylor instabilities (see Ref. 8).

Our expressions for the evolution of perturbations at the interfaces 1 and 2 in

Fig. 3-3(a) show, as implied by Eq. (2), that these interfaces are coupled, and that the evolution of the perturbation at one interface depends very much on the perturbation at the other interface, particularly at long wavelength λ or small thickness t_2 .

A Numerical Example. For more than three layers, we must use our computer code to evolve the perturbations according to Eq. (2). The input consists of the density profile (the ρ_i and t_i); the perturbation wavelength λ ; and the initial values $\eta_i(0)$ and $\dot{\eta}_i(0)$. All these quantities are arbitrary. The output consists of the values $\eta_i(\tau)$ of the perturbation amplitudes at chosen values of τ .

For example, Fig. 3-4 shows the density profile $\rho = (0, 1, 3, 1, 0)$, with thicknesses $(\infty, t, t, t, \infty)$. This is a special case of $N = 5$ and density profile $(0, \rho, R\rho, \rho, 0)$. There are $N - 1 = 4$ eigenvalues: two are given by

$$\frac{\gamma^2}{gk} = \pm 1 \quad (8)$$

and the other two are given by

$$\frac{\gamma^2}{gk} = \pm \frac{R - 1}{[(R + 1)^2 + R(R + 2)/S^2]^{1/2}} \quad (9)$$

For $R = 3$, as in Fig. 3-4, Eq. (9) reduces to

$$\frac{\gamma^2}{gk} = \pm \frac{1}{2} \left(1 + \frac{15}{16S^2} \right)^{1/2} \quad (10)$$

One reason we chose the density profile shown in Fig. 3-4 is that it permitted us to derive these analytic expressions for

Fig. 3-5. Evolution of perturbations in the system of Fig. 3-4, with $\lambda = 3t$, $t = 1$, and $g = 1$. (a) All $\eta_i = \eta(0)$. (b) $\eta_1 = \eta_4 = \eta(0)$, $\eta_2 = \eta_3 = 0$.

the growth rates. If we choose $\lambda = 3t$, so that $kt = 2\pi/3$, Eqs. (9) and (10) yield $\gamma^2 = \pm 2\pi/3$, $\pm 0.3240\pi$; for this same case, the code gives the values $\gamma^2 = \pm 1.018$, ± 2.094 , satisfactorily close to the analytic values. (Our units of length and time are determined by the choices $t = 1$ and $g = 1$.)

Figure 3-5 shows the evolution of the perturbations from two sets of initial conditions. In Fig. 3-5(a), the initial amplitudes all have the common value $\eta_1(0) = \eta_2(0) = \eta_3(0) = \eta_4(0) = \eta(0)$. Note that even perturbations at the nominally stable interfaces 3 and 4 grow under the influence of the perturbations at the unstable interfaces 1 and 2. Classically, one might expect η_3 and η_4 to oscillate at amplitudes that never exceed their initial values. Figure 3-5(a) shows, however, that after oscillating for a short time, η_3 and η_4 begin to grow. Note that η_3 , being closer to the unstable interfaces 1 and 2, grows more than η_4 .

In Fig. 3-5(b), we have assumed $\eta_2(0) = \eta_3(0) = 0$, and $\eta_1(0) = \eta_4(0) = \eta(0)$, so that the initial perturbations are only on the external surfaces. Internal perturbations develop rather quickly and the outer-surface amplitude η_1 increases by a factor of 100 by time $\tau \approx 3.6$.

Summary. We have developed a formalism and written a code to calculate how perturbations evolve by the Rayleigh-Taylor mechanism in a system subjected to a constant acceleration. The system consists of an arbitrary number of fluid layers with

arbitrary thicknesses and densities. We find that perturbations feed through from one interface to another, and that even surfaces with no initial perturbations can develop rather large perturbation amplitudes. We plan to use the formalism to study shell breakup and to estimate mix in ICF targets.

Author: K. O. Mikaelian

Simple Models of High-Gain Targets—Comparisons and Generalizations

Introduction. Several simple models of high-gain ICF targets have been published. Here we examine those of Kidder⁹ (K) and Meyer-Ter-Vehn¹⁰ (MTV); by generalizing the MTV model, we show that the two models can be smoothly linked. We solve the MTV model analytically, rather than numerically. We further generalize the MTV model by allowing ignition criteria to vary with target scale size, and we use it to predict the improvement in ICF target performance that would result from the use of polarized DT fuel.

The starting point for these models is an assembled DT fuel configuration, i.e., one that is about to burn. A hot-DT ignition region (hot spot) of radius R_{H1} is surrounded by nearly Fermi-degenerate (cold) DT fuel of radius R_1 . The assembly is assumed to have internal energy ηE , where E is the original driver energy and η represents the

Fig. 3-6. Configurations of hot spot (H) and cold fuel (c) according to (a) the isobaric MTV model; (b) the isochoric K model.

many efficiency factors involved in getting to this final, pre-burn state. We wish to find G^* , the highest possible gain from such a configuration at this energy ηE .

Figure 3-6 shows variation of ρ , P , and T with R in the K and MTV models. The isobaric MTV model enforces a constant pressure throughout the assembled fuel. Kidder's isochoric model assumes constant density, which leads to a much higher pressure in the hot spot than in the cold fuel and to a G^* five times higher than MTV. We will shortly find out why.

Generalizations and Comparisons.

We generalize the MTV model by allowing the pressure to drop by a factor ϵ from the hot spot to the cold fuel, so that $P_c = \epsilon P_H$. For $\epsilon = 1$ we recover the MTV model; for ϵ quite small (as shown in Table 3-1) we recover the K model. (We have further generalized the model to include an outside tamper material; for simplicity we ignore this feature here.) We generalize the MTV ignition criteria with the parameters β and γ : $T_H = 5\gamma$ keV and $(\rho R)_H = 0.4\beta$ g/cm². Anticipating interest in the use of polarized DT fuel, we define the parameter δ : $\langle \sigma v \rangle = \delta \langle \sigma v \rangle_{DT}$. Finally, we follow MTV by defining α , the factor by which the cold fuel is non-degenerate:

$$P_c = 2.3 \times 10^{12} \alpha \rho_c^{5/3} \text{ dyne/cm}^2, \quad (11)$$

$$E_c = 0.35 \alpha \rho_c^{2/3} M_c \text{ MJ}, \quad (12)$$

where the cold-fuel density ρ_c is in g/cm³ and M_c is in grams. The hot-spot energy and pressure are easily found to be

$$E_H = 10^3 \beta \gamma R_H^2 \text{ MJ}, \quad (13)$$

$$P_H = 1.6 \times 10^{15} \beta \gamma / R_H \text{ dyne/cm}^2. \quad (14)$$

By setting $P_c = \epsilon P_H$, we immediately find

$$\rho_c = 51 \left(\frac{\beta \gamma \epsilon}{\alpha R_H} \right)^{0.6}; \quad (15)$$

Eqs. (12) and (15) then yield the cold-fuel mass

$$M_c = 0.21 \left(\frac{R_H}{\beta \gamma \epsilon} \right)^{0.4} \alpha^{-0.6} (\eta E - E_H). \quad (16)$$

We find the outer fuel radius R_c from the identity

$$M_c = \frac{4\pi \rho_c}{3} (R_c^3 - R_H^3). \quad (17)$$

The burn efficiency ϕ is given by

$$\phi = \frac{\rho R}{7/6 + \rho R}, \quad (18)$$

where

$$\rho R = 0.4 + \rho_c (R_c - R_H). \quad (19)$$

Finally, the gain G is given by

$$G = \frac{\phi q_{DT} M_c}{E}, \quad (20)$$

where $q_{DT} = 3.3 \times 10^3$ MJ/g.

We can now see why, for a given ηE , there is an $R_H \approx R^*$ that leads to the optimal gain G^* . If R_H is too large, E_H becomes too large and $E_c = \eta E - E_H$ too small, leading to a small M_c and low gain. If R_H is too

small, the pressure must then be too high (if the ignition criteria are to be met) as will ρ_c : high ρ_c (at fixed ηE) again leads to small M_c and low gain. Thus there is an optimal R^* and G^* .

To derive G^* analytically we anticipate that $M_c \gg M_H$, $E_c \gg E_H$, $R_c^3 \gg R_H^3$ and $\rho_c(R_c - R_H) \gg 0.4$. Under these assumptions, we obtain

$$R_c = 0.1 R_H^{1/3} (\eta E / \beta \gamma \epsilon)^{1/3} \quad (21)$$

Another key trick is to fit the function ϕ to a power law:

$$\phi \approx \frac{1}{2} \left(\frac{\rho R}{7/\delta} \right)^{1/2} \quad (22)$$

Despite its simplicity, Eq. (22) is quite good over the entire range $2 \leq \rho R \leq 20$. Thus

$$\phi = \frac{1}{2} \sqrt{\frac{51}{7\delta}} \left(\frac{\beta \gamma \epsilon}{\alpha R_H} \right)^{0.3} \sqrt{R_c - R_H} \quad (23)$$

with R_c from Eq. (21), and we use Eq. (23) in

$$G = \phi \frac{q_{DT}}{E} \times \left[0.21 \left(\frac{R_H}{\beta \gamma \epsilon} \right)^{0.4} \frac{1}{\alpha^{0.6}} (\eta E) \left(1 - \frac{10^3 \beta \gamma R_H^2}{\eta E} \right) \right] \quad (24)$$

Defining

$$W \equiv 10 R_H^{2/3} (\beta \gamma / \eta E)^{1/3} \quad (25)$$

we obtain

$$G = 1.2 \times 10^4 \eta \left(\frac{\eta E}{\alpha^3} \right)^{0.3} \frac{\delta^{0.5}}{(\beta \gamma)^{0.4} \epsilon^{0.26}} \times W^{2/5} (1 - W^3) \sqrt{1 - \epsilon^{1/3} W} \quad (26)$$

For $0.01 \leq \epsilon \leq 1$, we find that $\partial G / \partial W = 0$ at $W^* = 0.34 \epsilon^{-0.07}$, or

$$= 63 \epsilon^{-0.1} \eta E / \beta \gamma \quad \mu\text{m} \quad (27)$$

and

$$G^* = 6 \times 10^3 \eta \left(\frac{\eta E}{\alpha^3} \right)^{0.3} \frac{\delta^{0.5}}{(\beta \gamma)^{0.4} \epsilon^{0.3}} \quad (28)$$

Isobaric (Meyer-Ter-Vehn)	Isochoric (Kidder)
$G^* = 6 \times 10^3 \eta \left(\frac{\eta E}{\alpha^3} \right)^{0.3} \epsilon^{-0.3}$	$G^* = 30 \times 10^3 \eta \left(\frac{\eta E}{\alpha^3} \right)^{0.4}$
$(\rho R)^* = 13 \left(\frac{\eta E}{\alpha^3} \right)^{0.2} \epsilon^{0.1} \text{ g/cm}^2$	$\rho R = 4 \left(\frac{\eta E}{\alpha^3} \right)^{0.11} \text{ g/cm}^2$
$\rho_H = 63 (\eta E)^{-0.5} \text{ g/cm}^3$	$\rho = 33 (\eta E)^{-0.5} \text{ g/cm}^3$
$\rho_c = 1050 (\eta E)^{-0.4} \alpha^{-0.6} \epsilon^{0.1} \text{ g/cm}^3$	
$M_c = 29 (\eta E)^{1.2} \alpha^{-0.6} \epsilon^{-0.44} \text{ mg}$	$M_c = 230 (\eta E)^{1.1} \alpha^{-0.4} \text{ mg}$
$E_H = 0.04 E \epsilon^{-0.2}$	$E_H = 0.16 E$
$R_H = \frac{63}{\epsilon^{0.4}} (\eta E)^{1.2} = \frac{R_c}{3} \mu\text{m}$	$R_H = 120 (\eta E)^{1.2} = \frac{R_c}{10} \mu\text{m}$
$\epsilon \equiv \frac{P_c}{P_H}$	$\epsilon \equiv \frac{P_c}{P_H} = 6 - 0.1 \frac{\eta E}{\alpha^3}$

Table 3-1. Expressions for optimum gain G^* and corresponding quantities from the J and MTV models.

Fig. 3-7. Capsule gain G vs R_H and ρR in the MTV model; G depends only weakly on R_H near the optimum R^* .

Figure 3-7 shows that $G(R_H)$ is rather flat near R^* ; in fact, if R_H differs from R^* by a factor of 2, $G \approx 0.8G^*$.

The same analytic approach can be applied directly to Kidder's model. Table 3-1 gives the various optimal quantities according to the two models. To simplify, we set $\beta = \gamma = \delta = 1$.

Table 3-1 shows why the K model gives larger gains. Since the model is isochoric, its ρ_c is quite low, leading to an M_c 10 times higher than that of MTV. But the K model's low ρ_c leads to a ρR that is 1/4 that of the MTV model. Since $\phi \sim (\rho R)^{1/2}$, the K model has half the burn efficiency; it burns 10 times more fuel, so its gain is greater by a factor of 5.

For $\epsilon = 1$, we recover the numerical results of the MTV model. Calculating ϵ in the K model (from the bottom line of Table 3-1), and inserting it into the ϵ -dependent formulas of the generalized isobaric model, we recover the Kidder results. A check of this analytic work with a

numerical model similar to MTV confirms the ϵ scaling. In general, the true assembled-fuel configuration is somewhere between the two models, so that this ϵ formalism can be of practical use.

Polarized-Fuel Scaling. If the DT nuclei can be spin-polarized, the reaction rate should be 50% higher, corresponding to $\delta = 1.5$. (We ignore here the possibility that anisotropic effects may lead to even more favorable results.) Equation (28) indicates that $G^* \sim \delta^{0.5}$ gives the final answer [recall that the δ scaling of Eq. (28) comes from the burn-efficiency term, ϕ]. But we expect further improvements because of the relaxed ignition criteria, namely $\beta\gamma = \beta\gamma(\delta)$. A simple model for ignition will suffice here to give us an idea of how $\beta\gamma$ should scale with δ .

For ignition, we require that the self-heating time t_H be shorter than the hydrodynamic disassembly time t_D . We define the self-heating time by requiring that $\rho\langle\sigma v\rangle t_H \sim T$, i.e., that the burn rate times t_H yields enough energy to double the temperature. The disassembly time is the usual $t_D \sim R_H/C_s \sim R_H T^{-1/2}$, where C_s is the sound speed. Since $\langle\sigma v\rangle$ contains the dominant T dependence, $t_H \leq t_D$ leads (roughly) to a criterion of the form $\rho R\langle\sigma v\rangle \sim \text{constant}$, or (at fixed γ), $\beta \sim \delta^{-1}$. Equation (28) now yields

$$G^* \sim (\eta E)^{0.3} \frac{\delta^{0.5}}{(\delta-1)^{0.4}} \sim \delta^{0.9} (\eta E)^{0.3} \quad (29)$$

Thus if we are satisfied with a particular gain G^* , the required energy E scales as δ^{-3} . Polarized fuel (with $\delta = 1.5$) would thus reduce the required driver energy by a factor of 3.

Further Generalizations. We have assumed that the ignition criteria are independent of the scale size of the assembled fuel, in particular of the ρR of the cold fuel. This is not quite correct, because the ρR of the cold fuel acts to tamp the hot-spot disassembly. Since tamping usually enters as a term of the form $[1 + (5/3)\mu]^{1/2}$, where $\mu = \Lambda_{\text{tamp}}^2/M_{\text{fuel}}$, it would not be surprising if (at fixed γ) we had $\beta \sim (\rho R)^{-1/2\beta'}$. Thus β is no longer a constant, and the analytic optimization procedure, although similar, yields different scalings. Assuming that $\beta' \sim 1/\delta$, as above, we find (at fixed G) $E \sim \delta^{-2.5}$, a

somewhat weaker dependence than the δ^{-3} scaling of the previous section.

A final result stemming from our generalized isobaric model arises from considering the minimum energy (of a frozen DT shell, of mass M_s moving at a given implosion velocity V_0) required to achieve the isobaric assembled fuel configuration. We require that

$$\frac{1}{2} M_s V_0^2 = E_c + E_H = \eta E, \quad (30)$$

or, using Eqs. (12), (13), and (15), that

$$M_s \left[\frac{1}{2} V_0^2 - 4.8 \left(\frac{\beta\gamma\epsilon}{R_H} \right)^{2/5} \alpha^{3/5} \right] = 10^3 \beta\gamma R_H^2. \quad (31)$$

Minimizing M_s is equivalent to minimizing the input energy $M_s V_0^2/2$, since V_0 is given. Thus minimizing M_s with respect to R_H leads to

$$V_0 = 2.6 \times 10^7 (\eta E)^{-0.1} (\beta\gamma\alpha)^{0.3} \epsilon^{0.2} \text{ cm/s} \quad (32)$$

and

$$R_{\min} = 120 \sqrt{\eta E/\beta\gamma} \text{ } \mu\text{m}. \quad (33)$$

Note that this minimum-energy configuration has $R_H = 2R^*$. The optimization here does not concern itself with burn efficiency (and hence with ρR), so it gives slightly different results than those of Eqs. (27) and (28), which sought to optimize gain. As noted above, at $R_H = 2R^*$ we have $G = 0.8G^*$, so this minimum-energy configuration is quite acceptable. Since $R_H = 2R^*$ relaxes convergence-ratio (i.e., symmetry) constraints by a factor of 2, it would seem to be the preferred state from that point of view as well. Note, too, that in this state the mass M_c is within 10% of M^* (as given in Table 3-1). It is therefore no surprise that the required ignition velocity of Eq. (32) is approximately the same as what one would obtain by setting $M^* V_{ig}^2/2 = \eta E$.

Summary. We have generalized the isobaric model of high-gain targets in several important ways. We have carried out the optimization process analytically and have obtained scaling laws that indicate the

possible improvements in capsule performance with polarized fuel. In the next article we consider the accessibility of the optimal assembled fuel state, and find η from first principles, in order to make this simple model even more useful in ICF target design.

Authors: M. D. Rosen, J. D. Lindl, and A. R. Thiessen

Model for Gain vs Laser Energy in Direct-Drive Targets

Introduction. In this article we derive, from first principles, an optimal gain vs laser energy for viable ICF targets. We wish analytically to answer the question: What is the highest gain achievable with a canonical 2-to 4-MJ driver? Is it 2, 20, or 200?

We attack this problem in two steps. First we appeal to results obtained (in the preceding article) from a simple model for the optimal gain G^* of an assembled DT fuel (i.e., one about to ignite and burn), with a fixed internal energy ηE . Then we cull numerous relevant quantities, in particular the hydrodynamic efficiency η , from analytic models of steady-state spherical laser ablation. We then find η as a function of ηE , subject to numerous physics constraints that make the design viable.

The first step yields $G = G(\eta E)$; the second yields $\eta = \eta(\eta E)$, from which we obtain our desired result $G = G(E)$.

The Optimal Assembled Fuel Configuration. Meyer-Ter-Vehn¹¹ has recently considered an assembled DT fuel (he treats an ignition hot spot surrounded by a cold DT fuel at the same pressure as the hot spot). If the assembly has total energy ηE , and the cold fuel is on an adiabat that is α times higher than Fermi-degenerate, the optimal gain G^* is

$$G^* = 6 \times 10^3 \eta (\eta E / \alpha^3)^{0.3} , \quad (34)$$

(where ηE is in MJ); the mass of the system is

$$M^* = 29(\eta E)^{1.2} \alpha^{-0.6} \text{ mg} , \quad (35)$$

and the implosion velocity required to bring the fuel into this state is at least

$$V_{\text{imp}}^2 = \eta E / (M^* / 2), \text{ or}$$

$$V_{\text{imp}} = 2.6 \times 10^7 (\eta E / \alpha^3)^{-0.1} \text{ cm/s} . \quad (36)$$

These results hold quite well even if the assembled fuel is not exactly in the optimum configuration. For example, if the hot-spot radius were twice the optimum value [$R = 2R^* = 2 \times 63(\eta E)^{1/2}$], then we would have $G = 0.8G^*$, $M \approx M^*$ and $V \approx V_{\text{imp}}$.

These parameters serve as both goals and constraints as we try to achieve this optimal fuel assembly in practice. We use all of them in the course of this article.

Steady-State Ablation Results. Here we appeal to the results of Max et al.^{12,13} for steady-state ablation in spherical geometry. We consider plastic (a typical ablator) with a flux limiter of about 0.035 (or $\phi = 0.3$, in their notation). We then obtain for the ratio of critical radius r_c to ablation radius R_A the expression

$$r_c / R_A = 1.3 I^{0.11} \lambda^{-0.39} R^{-0.08} , \quad (37)$$

where (as throughout this article) I is in units of 10^{14} W/cm^2 , λ is in microns, and R is in μm ; we have dropped the subscript A from the ablation radius R_A . (The ratio r_c / R_A is a measure of the standoff distance $r_c - R_A$, itself an important parameter in connection with symmetry union.) Defining $S = r_c / 1.3 R_A$ as a normalized standoff parameter, we obtain

$$R = S^{-12.5} I^{1.4} \lambda^{4.8} ; \quad (38)$$

we find for the pressure (in Mbar)

$$P = 8.4 I^{0.57} \lambda^{-0.99} R^{0.07} \\ = 8.4 I^{0.66} \lambda^{-0.66} S^{0.9} , \quad (39)$$

and for the mass ablation rate (in g/s)

$$\dot{M} = 2.3 \times 10^4 I^{0.48} \lambda^{-0.82} R^{1.89} , \quad (40)$$

or for the areal ablation mass (in $\text{g/cm}^2 \cdot \text{s}$)

$$\dot{M}_A = 1.8 \times 10^5 S^{1.4} I^{0.33} \lambda^{-1.33} . \quad (41)$$

For a rocket-payload velocity, we then find

$$V = (P / \dot{M}) \ln \lambda \\ = 4.6 \times 10^7 I^{0.33} \lambda^{0.7} S^{-2.3} \ln X , \quad (42)$$

Laser Targets

where $X \equiv m/m_0$ is the fraction of the initial rocket mass remaining. We require that this payload reach the implosion velocity given by Eq. (36). This then yields for X the result

$$\ln \bar{X} = -0.57\alpha^{0.3}(\eta E)^{-0.1}I^{-0.33}\lambda^{-0.7}S^{2.3} \quad (43)$$

Calculation of In-Flight Aspect Ratio.

The laser pulse that drives an ICF target must be shaped in time to keep the fuel compressed on a low adiabat [α much greater than a few in Eq. (34) leads to low gain]. Typically the shell is compressed, becoming thinner and denser as it stays in contact with the driving pressure. Later it speeds up, and it is usually at terminal, ignition velocity by the time it has moved in to about half its initial radius.

We now put these descriptions into quantitative terms. The density of the shell can be found by assuming that it stays in pressure contact with the drive, so that the P of Eq. (39) becomes $P = 2\alpha\rho^{5/3}$, yielding

$$\rho = 2.4I^{0.4}S^{-0.55}\lambda^{-0.4}\alpha^{-0.6} \quad (44)$$

Getting the shell up to terminal velocity by the time it moves halfway in means that

$$\frac{1}{2}R = \int V dt \quad (45)$$

or, using Eq. (42),

$$R = (9.2 \times 10^7 I^{0.33} \lambda^{0.7} S^{2.3}) \frac{m_0}{\dot{m}} \times [1 - \bar{X}(1 - \ln \bar{X})] \quad (46)$$

where $\bar{X}$ is derived in Eq. (43). To find the thickness of the compressed shell, we simply find

$$\Delta X_0 = \frac{\dot{m}}{\rho} t_{BT} = \frac{\dot{m}}{\rho} \frac{m_0}{\dot{m}} \quad (47)$$

Using Eqs. (41) and (44) we obtain

$$\Delta X = 7.5 \times 10^4 S^2 I^{-0.07} \lambda^{-0.9} \alpha^{0.6} m_0 / \dot{m} \quad (48)$$

Thus the in-flight aspect ratio (IFAR) of the shell is given by

$$\text{IFAR} \equiv R/\Delta X_0 = 1.2 \times 10^3 I^{0.4} \lambda^{1.6} \times S^{-0.4} \alpha^{-0.6} [1 - \bar{X}(1 - \ln \bar{X})] \quad (49)$$

This is a transcendental equation. To make further progress we fit it to a power law applicable over the range of interest of all parameters:

$$\text{IFAR} \approx 140\lambda^{0.5}I^{-0.1}S^{-0.5}(\eta E)^{-0.15}\alpha^{-0.1} \quad (50)$$

Closing the Set of Equations. We can now start eliminating variables (such as I). On the one hand we can represent the in-flight mass as

$$M = 4\pi \frac{R^3 \rho}{\text{IFAR}} = 0.2 R^3 I^{0.5} \lambda^{-0.9} \alpha^{-0.5} (\eta E)^{0.15} \quad (51)$$

On the other hand, after $1 - \bar{X}$ of the mass is ablated away, we must be left with the optimal mass given by Eq. (35). Therefore

$$M = 29 \times 10^{-3} (\eta E)^{1.2} \alpha^{-0.6} \bar{X}^{-1} \quad (52)$$

from which we find

$$R = 5.2(\eta E)^{0.35}I^{-0.17}\lambda^{0.3}\bar{X}^{-0.33} \quad (53)$$

As an aside, we note that there is still another way to represent the mass in its initial configuration. This will give us an expression for the initial aspect ratio $(AR)_0$, which is quite different from the IFAR. Thus, equating $M = 4\pi R^3 \rho_0 / (AR)_0$ and the M of Eq. (52), and defining d by $\rho_0 \equiv 0.2d$, we find

$$(AR)_0 = 12(\eta E)^{-0.15}I^{-0.5}\lambda^{0.9}\alpha^{0.6}d \quad (54)$$

Comparing Eq. (53) with Eq. (38) we obtain, for consistency,

$$I = 2.9(\eta E)^{0.23}S^8\lambda^{-2.8}\bar{X}^{-0.22} \quad (55)$$

Thus Eq. (43) becomes

$$\bar{X} = \exp[-0.39\alpha^{0.3}(\eta E)^{-0.18}S^{-0.4} \times \lambda^{0.23}\bar{X}^{0.07}] \quad (56)$$

Eq. (53) becomes

$$R = 4.3(\eta E/\bar{X})^{0.3}\lambda^{0.77}S^{-1.4} \text{ mm} \quad (57)$$

Eq. (50) becomes

$$\text{IFAR} = 125(\eta E)^{-0.17}\lambda^{0.8}S^{-1.3}\alpha^{-0.1} \quad (58)$$

and Eq. (39) becomes

$$P = 17(\eta E / \bar{X})^{0.15} S^{4.4} \lambda^{-2.5} \text{ Mbar} \quad (59)$$

Finally, we are most interested in calculating the hydrodynamic (rocket) efficiency

$$\eta_{H1} = \frac{mV^2}{2It} = \frac{mV^2}{2I} \left(\frac{\bar{X}}{1 - \bar{X}} \right) \quad (60)$$

Using Eqs. (41), (36), and (55) we obtain

$$\eta_{H1} (\%) = 3 (\eta E)^{-0.35} \lambda^{0.54} S^{-4} \bar{X}^{0.15} \frac{\bar{X}}{1 - \bar{X}} \quad (61)$$

For completeness we substitute Eq. (55) into the expression for $(AR)_0$ and find

$$(AR)_0 = 7(\eta E)^{-0.26} S^{-4} \lambda^{2.3} \alpha^{0.6} \bar{X}^{0.1} d \quad (62)$$

Another quantity of interest as a possible constraint is the convergence ratio CR, defined as the ratio of the original radius to the hot-spot radius. We recall that the optimum hot-spot radius is given by $R^* = 63(\eta E)^{1/2}$; thus, with Eq. (57), we obtain

$$CR = 68(\eta E)^{-0.2} \bar{X}^{-0.3} \lambda^{0.77} S^{-1.4} \quad (63)$$

The characteristic pulse length t can be found by setting $t = m_0(1 - \bar{X})/m$, with $m_0 = \rho_0 R / (AR)_0$. Thus Eqs. (41), (55), (57), and (62) yield

$$t = 48(1 - \bar{X}) \eta E^{0.5} \times \lambda^{0.8} S^{-0.14} \alpha^{0.6} \bar{X}^{-0.3} \text{ ns} \quad (64)$$

We can make an interesting self-consistency test: we compare the I of Eq. (55) with that obtained by setting $I = E/4\pi R^2 t = (\eta E)/(4\pi \bar{X} R^2 t)$. Using Eqs. (61), (57), and (64) we indeed find that these expressions for I agree.

Results. Equations (56) and (61) are the core of the solution, because they yield $\bar{X}$ and then η as functions of ηE , for given λ and S . Figures 3-8 and 3-9 show $\eta = \eta(\lambda, S)$ at $\eta E = 1$ and 0.1 MJ respectively. These figures are reminiscent of Fig. 7 of Ref. 13, except that our analysis is for arbitrary energy (in Ref. 13 the calculations were carried out at a single, fixed energy), and that we explicitly put in the IFAR isocontours.

▲ Fig. 3-8. Rocket efficiency vs λ_L for assembled DT with 1 MJ internal energy. Preferred regimes ($IFAR \leq 50$ and $\lambda^2 \leq 3$) overlap, giving an acceptable operating point.

Fig. 3-9. Rocket efficiency vs laser wavelength for 0.1 MJ of assembled DT. No operating point.

There are two regimes in the figures. In the lower left is the low-IFAR regime, which corresponds to high I (this is required if a low-IFAR shell is to attain ignition velocity); this regime is hydrodynamically more stable, so one would prefer to work there. In the upper right is the low- I , high-IFAR regime, which is less likely to lead to plasma instabilities and hot-electron generation. One would like to work in this regime, which can also be characterized as having $\lambda^2 \leq 3$ (as shown in the figures), to avoid deleterious plasma processes and to ensure good absorption. In Fig. 3-8, for $\eta E = 1$, these preferred regions overlap, and we see a suitable operating

Fig. 3-10. Rocket efficiency and required laser intensity vs ηE and $\eta E/\eta = E_{\text{abs}}$, with $\lambda = 1/4 \mu\text{m}$ and IFAR = 50.

Fig. 3-11. Gain vs laser energy for IFAR = 50 and 70, with $\lambda = 1/4 \mu\text{m}$ and $f = 0.03$.

Fig. 3-12. Gain vs laser energy for $f = 0.1$ and 0.03 , with $\lambda = 1/4 \mu\text{m}$ and IFAR = 50.

point at $\lambda = 1/4 \mu\text{m}$ corresponding to a rocket efficiency $\eta_H = 6.5\%$. Figure 3-9, however, shows that for $\eta E \leq 0.1$, and for IFAR < 50, the best we can do is 3% efficiency at an I that is uncomfortably high from a plasma-physics point of view.

Figures 3-8 and 3-9 do not show it, but the CR constraint (say $\text{CR} \leq 50$) is less stringent than the $\text{IFAR} \leq 50$ constraint. We chose $\text{IFAR} \leq 50$ as follows. If, as usual, the growth rate of a typical manufacturing defect is given by $\gamma = f(kg)^{1/2}$, we find that $\gamma t = f(2\pi \cdot \text{IFAR})^{1/2}$. To achieve $\gamma t \leq 3$ (which corresponds to defect growth by a factor of 3000), we require that $\text{IFAR} < 10/f^2$. Simulations show that Rayleigh-Taylor growth is about half the classical rate, so that $f = 1/2$; thus $\text{IFAR} \leq 40$ or 50.

If we optimistically assert that symmetry will not be a problem (so that 5 is not a constraint), Figs. 3-8 and 3-9 imply that we should maximize η at the shortest possible λ , which we take to be $1/4 \mu\text{m}$. Then with $\text{IFAR} = 50$, we can solve for $\bar{X}$ and η_H as functions of ηE ; Fig. 3-10 shows the results of this analysis. Once $\bar{X}$ and η_H are known, we can determine I and $\eta E/\eta$; these quantities are also plotted in Fig. 3-10.

Another efficiency to account for is absorption. Many high- f /number lenses are needed for a direct-drive target so as not to take up too much reactor-chamber solid angle. This means a considerable amount of tangential incidence on the pellet, and as the target implodes some of the rays will miss it completely. Therefore we take $f_{\text{abs, max}}$ to be $2/3$. In addition, if very high I is required (for $\text{IFAR} \leq 50$) the absorption may not be too efficient even with $1/4\text{-}\mu\text{m}$ light. We therefore take $f_{\text{abs, min}}$ to be $1/3$.

We can now plot $G(E_L)$, because Eq. (34) gives $G(\eta E)$ and Fig. 3-10 gives η_H and governs our choice of f_{abs} . Figure 3-11 shows how sensitive $G(E_L)$ is to IFAR constraints, and shows low gain (~ 30) at 3 MJ.

If we redo the entire calculation for a flux limit of $f = 0.1$, we get the result shown in Fig. 3-12. For these conditions we obtain $G(3 \text{ MJ}) = 60$, which is an improvement. Still, these gains do not compare favorably with the published Livermore gain curves for hohlraum targets,¹⁴ which are over 100 at 3 MJ.

Conclusions. We have shown how targets perform on the basis of simple models. Keeping a reasonably low IFAR constrains

the intensity to be high, leading to low hydrodynamic efficiency because the rocket exhaust velocity is not well matched to the required payload velocity. It seems clear that the IFAR is a major lever, and that learning how to reduce the Rayleigh–Taylor growth can lead to higher allowable IFARs and thus to much higher gains.

Authors: M. D. Rosen and J. D. Lindl

Charged-Particle Fusion

Introduction

In 1983 we continued a series of studies that highlight the usefulness of high-quality ion beams for ICF applications. The studies involve the interaction of the physics of ICF targets and the physics of ion beams. Several results of these studies are summarized in the 1980, 1981, and 1982 Laser Annuals.

For ICF applications, many target- and beam-physics considerations favor the design of ion accelerators to produce quality beams (that is, beams of low emittance). Among the target-physics advantages of such beams are the following:

- Significant beam-target-interaction experiments with plasmas at $\leq 10^6$ K (86 eV) could be performed with only a few kilojoules of ions of range ≤ 0.01 g/cm² in matter, provided that they can be focused to ~ 1 mm spot radius (an example is the planned high-temperature experiment for heavy-ion fusion, described in the next article).
- Estimates of reactor-target gain^{15,16} reveal the advantages of focusing beams to smaller focal spots (diameters of a few millimeters).
- We have recently found that peak power requirements for driving single-shell ICF reactor targets could be reduced below those quoted in Ref. 15, provided that the beams have smaller focal spots. For example, preliminary results show that target gain ~ 100 is possible with less than 200 TW of peak power.

The beam-physics advantages of low emittances can be seen, for example, in the final-propagation properties of

heavy-ion beams across fusion chambers; these advantages, described in more detail in Ref. 17 and in the next two articles, include the following:

- Ballistic propagation in soft vacuum (10^{-3} to 10^{-4} Torr) allows ≤ 1 -kA beams.
- Neutralized-beam ballistic transport allows ≤ 4 -kA beams.
- Self-pinched beams are expected to give the smallest beam-entrance ports (≤ 1 cm) and highest beam current per channel; the possibility of propagating such beams across a reactor chamber is enhanced if the beam emittance is low. For heavy-ion fusion, this additional non-ballistic propagation mode is more a long-term possibility, but it offers significant advantages. We are continuing aspects of this work in cooperation with the LLNL electron-beam work (see "Testing Hydrodynamic Models for Simulation of Pinched Beams," later in this section; see also Ref. 18).
- Scaled experiments on most reactor-beam propagation modes could already be performed with a few-kilojoule machine (see the next article).
- Good beam quality also allows more leeway for final-focus techniques that have advantages over standard quadrupoles in the high-temperature experiment (see the next two articles).

Author: J. W-K. Mark

Studies for the High-Temperature Experiment in Heavy-Ion Fusion

The heavy-ion high-temperature experiment¹⁹⁻²² (HTE) and its associated accelerator test facility will provide a data base for the application of high-current, heavy-ion accelerator technology to inertial confinement fusion. The HTE is intended to achieve physical-parameter values that are within an order of magnitude of reactor-scale parameters, with a few exceptions (such as total beam energy and power) dictated by the budget for this first accelerator test facility. Plasma temperatures of up to 10^6 K (86 eV) are expected with disk targets.

The HTE accelerator and physics experiments are designed to study physical phenomena in three general categories:

1. The handling of high ion currents during the acceleration, bunching, and final focusing of the ion beam.

2. Neutralization, stability, spot size, and multibeam interaction during final transport to the target.
3. Beam-target interaction: energy-deposition anomalies, hot-electron generation, plasma-plume stability, and target charge-up.

The HTE does not address pellet-implosion and reactor-chamber design, which are decoupled from driver type and for which one needs higher beam energies than the several kilojoules expected from the HTE accelerator.

Accelerator Beam Physics in the HTE.

The HTE is designed to address all of the major accelerator issues for a heavy-ion fusion reactor driver: namely, the space-charge limits at low values of $\beta = v/c$, the final beam-compression process, and the design of final focusing lenses.

For most of the length of the heavy-ion reactor driver, nothing more than a rather classical, low-current accelerator is needed; only at the low- β and final-beam-compression ends must the accelerator be operated close to space-charge limits. The HTE accelerator, however, is basically *all* low- β section plus final compression, so that it tests all the difficult aspects of the reactor driver.

From this point of view, the HTE is a definitive and yet relatively low-priced experiment for the accelerator-beam physics in heavy-ion fusion. An indication of the relative scales of the HTE and a reactor driver is given in Table 3-2.

Tests of Fusion-Chamber Beam Physics in the HTE. The HTE also addresses major issues for final transport of heavy-ion beams across fusion chambers. Although we are also studying other modes of transport such as self-pinching (see the next article), the most relevant modes are in the ballistic-transport regimes with chamber vapor pressures $\lesssim 10^{-3}$ Torr in a lithium HYLIFE chamber. Theory predicts that for

these ballistic regimes the ion beams of heavy-ion fusion should not encounter any problems for reactor applications. Nevertheless, experiments are being planned (as part of the HTE) to confirm these theoretical predictions.

Heavy ions may be ballistically focused in vacuum, or may require a cloud of electrons for neutralization. A dimensionless parameter that clearly distinguishes ballistic-transport issues for heavy ions is²³

$$\nu = \frac{Z^2 e I_0}{\beta^2 M c^3} \quad (65)$$

where I_0 is the particle current per beamlet, Z is the charge state at injection, $\beta = v/c$, and M is the ionic mass. The quantity ν is a measure of electrical stiffness: a sufficiently low value of ν allows ballistic transport. For example, to avoid space-charge blowup in vacuum we need $\nu/\theta^2 \ll 1$, where θ is the angle of convergence at the wall. Other problems, such as emittance growth, pellet charge-up, and two-stream growth rates, are all directly related to ν multiplied by geometric factors that are common to all the various ICF schemes. (The stripped state of the ion, Z^* , must also be considered if the reactor-chamber lithium pressure is greater than $\sim 10^{-3}$ Torr, or if the dynamics in the plume are studied.) The geometrical factors are much less variable than ν . A possible defining condition for the HTE is, therefore,

$$\nu(\text{HTE}) \approx \nu(\text{reactor}) \quad (66)$$

Fulfillment of this condition would allow a test of several of the final-transport issues (described above) appropriate to a heavy-ion driver. For the reactor and a sodium-ion HTE, we have the following values (assuming $Z = 1$ in both cases):

	I_0 (A)	β	M (M_p)	ν
Reactor	10^3	0.3	200	4.9×10^{-6}
Na HTE	16	0.1	23	2.2×10^{-5}

The proposed HTE test facility should also permit useful experiments on beam neutralization (W. Fawley, quoted in Ref. 21). A 2.4-keV electron beam has the necessary longitudinal velocity to travel in parallel with a 100-MeV sodium beam. This kinetic energy per electron is quite low in

Table 3-2. Ratios of various quantities in the HTE and the megajoule accelerator driver.

Quantity	Ratio, HTE/reactor
Number of electric quadrupoles	>0.1
Number of magnetic quadrupoles	>0.1
Number of pulvers	>0.1
Number of betatron oscillations	>0.3
Space-charge limit, e/s	1.0
Ratio of space charge to beam plasma frequency, L	>0.3
Beam plasma frequency, ω_p	>0.3

comparison with the electrostatic potential drop of more than 180 kV associated with an unneutralized 3-kA sodium beam. Extrapolating from results given by a 2-D particle-code simulation (using the HIPPO code of W. Fawley and S. S. Yu), we expect that the coinjection of electron beams with transverse temperatures below 1 keV will produce very good neutralization, even if the ion beam is subdivided into as many as 50 beamlets.

Role of HTE in Testing Beam-Target Interaction Physics. Heavy-ion fusion must demonstrate that the energy in multi-kiloampere ion-beam currents will be deposited in reactor targets with ion ranges similar to those used in target calculations. It must also demonstrate experimentally that effects such as high incident-beam current densities (especially after stripping of incident ions in the target) would not have deleterious effects on beam deposition.

According to a recent JASON review, theory and experiments have not indicated that any problems should arise. For example, one might be concerned about the high current densities (megamperes per square centimeter) after stripping of incident ions. As indicated in Fig. 3 of Ref. 24, 10-GeV thallium ions (for example) would be stripped to $Z^* \approx 70$ in dense target material, raising the incident beam current from about $16 \text{ kA}/4\pi$ (0.25 cm^2) $\approx 20 \text{ kA}/\text{cm}^2$ to $\sim 1 \text{ MA}/\text{cm}^2$. Experiments at the Naval Research Laboratory (NRL) and at Sandia National Laboratories (Albuquerque), discussed below, indicate that high ion-beam current densities did not cause anomalous deposition rates. (Incident-beam current densities in light-ion fusion reactor targets would be up to an order of magnitude higher than those for heavy-ion fusion.)

Other questions have been raised regarding beam-target interactions for heavy ions:

1. Are there streaming instabilities (with or without resistivity) or beam-generated magnetic fields that could affect deposition rates?
2. Does range shortening with increasing plasma temperature create an unstable situation (i.e., do shorter ranges result in hotter plasmas with even shorter ranges)?
3. Are nonisotropic or non-Maxwellian plasmas generated that give rise to further instabilities (such as the Weibel

instability) or to fast electrons (with consequent preheat)?

4. Are there important effects (such as the generation of magnetic fields) due to inhomogeneities?

Further theoretical and experimental work should be helpful in answering such questions. Swegle's recent theoretical work²⁵ indicates that plasma effects do not interfere with energy deposition by either heavy or light ions. Swegle considers electrostatic and electromagnetic instabilities, including some resistivity effects, and considers Weibel and range-shortening instabilities. Extensions of Swegle's analysis to lower, blowoff-plasma densities might be helpful. Greater stability is generally expected for heavy ions because they are harder to deflect ("stiffer") and because the associated return-current velocities are more strongly subthermal (and thus more stable). The higher quality of heavy-ion beams means less stabilizing longitudinal energy spread, however, so that further detailed theoretical work might still be relevant.

Carefully designed, well-diagnosed experiments are clearly even more desirable. Data from NRL²⁶ on the deposition of $250\text{-kA}/\text{cm}^2$ light-ion beams in Mylar and aluminum targets agree with computer codes.²⁷ Agreement with experimental deposition rates is found even though the only plasma effect in the codes is the generation of electron plasma waves by the beam. It is important to note that these experiments are already at reasonably high plasma temperatures (~ 10 to 15 eV). Preliminary results of light-ion experiments at Sandia²⁸ at $\leq 1 \text{ MA}/\text{cm}^2$ also agree with theory.

Additional experiments with heavy ions (or some scaled equivalent) are important for the heavy-ion fusion program. Of course, it would be ideal if funding for the HTE were to permit direct verification of deposition physics for (say) thallium ions at several kiloamperes and at kinetic energies of $\sim 10 \text{ GeV}$.

Before the HTE, experiments in the following categories have already been proposed as part of various ICF-related programs:

1. Deposition in cold matter of low-current, heavy-ion beams of 20 to $50 \text{ MeV}/\text{amu}$ (such work is proposed at Lawrence Berkeley Laboratory and at the

Gesellschaft für Schwerionenforschung in West Germany).

- Deposition of similar ions in high-temperature plasmas heated by auxiliary pulsed-power or laser sources (e.g., at the Service de Physique Nucléaire et Nucléaire, in France).
- Scaled experiments using different ions and energies but retaining realistic ratios of such heavy-ion beam-target interaction parameters as v_i/v_e and $n\lambda_D^3$. (Better diagnostics and better isolation of plasma effects might be possible.²⁹)

A mosaic of experiments in these categories could cover essentially all individual effects in heavy-ion beam-target interaction physics, and could reduce uncertainties in the minds of critics of heavy ions.

Still, it will be important for heavy-ion fusion to perform experiments that combine all effects under near-reactor conditions. In this connection, a suitable goal for the HTE might be to attain a ~ 1 kA of an intermediate

ion, such as sodium, at ~ 125 MeV, focused to a spot radius of 0.1 cm. The unneutralized current density might reach 250 kA/cm^2 under these conditions. Similar considerations influenced a number of other HTE parameters.

Table 3-3 summarizes many beam and target parameters. The values given are, of course, just typical numbers (for example, the beam energy needed to achieve target gain of ~ 100 was estimated in the 1982 Laser Annual³⁰).

Beam-Target-Interaction Conditions in the HTE—A Point Model. We have yet to show that kiloamperes of ~ 125 -MeV sodium ions focused to ~ 0.1 cm will indeed generate the realistic hot-plasma target claimed in Table 3-3 and at the same time avoid target disintegration before such temperatures are reached. A convenient way to do this might be to show that such an incident beam will heat a disk target to appropriate temperatures for relevant time scales.

The target might be arranged as shown in Fig. 3-13 (see Refs. 21 and 22). The target material should not be solid in the usual bulk sense—felted aluminum of ~ 0.1 normal density should yield an adequate disk-disassembly time scale.²¹ (Some low-Z foams could be used at even lower densities.) Longitudinal expansion is limited during heating (and ease of observation of the internal temperature is correspondingly improved) if the target is made deep enough by reducing the density of deposition material. A lead tamper is placed around the felted-metal target to reduce hydrodynamic temperature losses by radial expansion.

Here we describe a convenient way to calculate the temperatures of spot-heated materials in these disks. It is based on a simple point model that includes particle range, beam irradiance, energy loss, and the specific energy of the target material. An extension of this model includes a scale time for hydrodynamic disassembly. Detailed numerical simulations (summarized in the next part of this article) indicate that the point model should be quite satisfactory for comparative studies; such studies are likely to be needed, because the final HTE parameters might differ somewhat from those of Table 3-3 because of considerations such as accelerator cost.

Table 3-3. Typical HTE and reactor parameters in heavy-ion fusion.

	HTE	Reactor
Ions	125-MeV Na	10-GeV U
Beams from accelerator		
Energy E (kJ)	3.7	10^3
Power P (TW)	0.12	10^2
Total beam current I (kA)	0.97	10
Number of beamlets	16	16
Beam on target		
Focal spot radius r (cm)	0.1	0.25
n_i (cm^{-3})	6.0×10^{13}	3.5×10^{13}
Zn_i (cm^{-3})	6.6×10^{14}	3.2×10^{15}
Ion range (g/cm ²)	0.017	0.1
$E/\pi r^2 R$ (MJ/g)	6.9	51
Unneutralized current density of ion beam (MA/cm ²)	$ZI/\pi r^2 = 0.34$	$ZI/4\pi r^2 = 1.2$
Heated deposition material		
Peak temperature (eV)	75	270
Heating time (ns)	31	10
Beam stiffness, v	2×10^{-5}	6×10^{-6}

Fig. 3-13. Spot-heating disk target.

As a simplified point model of the relevant physics, we evaluate the temperature rise by solving the rate equation

$$\frac{d\epsilon}{dt} = \frac{S - \sigma T^4}{R}, \quad (67)$$

where

$\epsilon(T)$ is the specific energy of target (J/g),
 $S(t)$ is the irradiance by the beam (W/cm²),
 R is the ion range (g/cm²),
 σ is the Stefan-Boltzmann constant (1.028×10^5 W/cm²·eV⁴),
 t is the time (s),
 T is the temperature (eV).

The slight mathematical difficulties encountered in solving Eq. (67) arise from the dependence of S on time, which destroys separability, and from the complex dependence of ϵ on temperature in realistic situations. A numerical treatment is needed, in general, but below we introduce approximations that are adequate for most situations of interest and that yield a simple, tabular solution.

We consider, an ideal pulse having zero rise length and constant beam energy flux $S = S_f$ at the focal spot, so that the formulas below for T_f , τ , and τ_d have a simple interpretation. A large collection of interconnected parameters is available for survey studies. The procedure adopted here covers only a small set of possible variations, and we give only a summary of our results.

Specifically, we have solved Eq. (67) under the following assumptions:

1. Particle type, particle energy (E), total particle current (I_0), beam-edge radius (a), and pulse length (τ_p) are specified.
2. Beam power (P), pulse energy (W), and beam irradiance (S_f) are determined from the specified quantities.
3. The target is melted aluminum, with specific energy given by the power-law fit

$$\epsilon = (7750 \text{ J/g}) T^{1.5} \rho^{-0.12} \equiv C T^{1.5}, \quad (68)$$

with T in eV and ρ in g/cm³. (There is no systematic error in this fit for $10 < T < 150$ and $10^{-2} < \rho < 1$; random errors are $\sim 15\%$.)

4. For particle range (R) we fitted a power law in beam kinetic energy E to numerical results used at LLNL and Los Alamos (see Fig. 2 of Ref. 22).
5. The depth of penetration ($l = R/\rho$) is set equal to twice the beam radius to get the longest disassembly time consistent with an ~ 30 -deg angular cone of beam irradiation.
6. The target density is obtained from ion range and penetration depth.
7. The final temperature (T_f) is obtained from Eq. (67) in the form

$$T_f = (S_f/\sigma)^{1/4}. \quad (69)$$

8. The energy-loss scale time (τ) is defined through the characteristic cooling rate realized at temperature T_f if S is suddenly turned off:

$$\tau \equiv -\frac{T}{\dot{T}_f} = \frac{3}{2} \frac{RC}{\sigma T_f^{5/2}}; \quad (70)$$

Eq. (68) gives the required value of C .

When S is suddenly turned on, the temperature reaches $0.96T_f$ in time τ .

9. The disassembly time (τ_d) is defined to be

$$\tau_d = \frac{l}{c_s} \approx \frac{R/\rho}{\sqrt{\frac{5}{3} \frac{(Z^* + 1)kT}{M_0}}}. \quad (71a)$$

Here Z^* is the mean ionization state of the target, T is the target temperature when heating is 50% complete, M_0 is the target particle mass, and ρ is the target mass density. The quantity Z^* is a continuous value for multiple ionization, corresponding to a continuous ionization potential $\phi(Z^* + 1/2)$ obtained by linear interpolation of discrete values; Z^* is determined from the continuous approximation to the Saha equation,

$$\phi\left(Z^* + \frac{1}{2}\right) = kT \ln \left[\frac{10^{-2} (kT)^{3/2} A_T}{\rho Z^*} \right], \quad (71b)$$

where A_T is atomic number. Because an additional averaging of values of Z^* and T during heating should be used, an approximate τ_d is found by using half of the

Fig. 3-14. Heating of a disk target by an ion beam, according to Eq. (3).

product $(Z^* + 1)T_i$ for the factor $(Z^* + 1)T$ in Eq. (71a).

Figure 3-14 shows the scaled temperature T/T_i vs the scaled time t/τ for these conditions. (Note that T reaches $0.96T_i$ in time τ , as claimed above.) At early times, $T/T_i \approx [(3/2)(t/\tau)]^{2/3}$.

If T has reached T_i and S is suddenly turned off at time t_0 , the subsequent decay of T is given by

$$\frac{T}{T_i} = \left(1 + \frac{5}{2} \frac{t - t_0}{\tau}\right)^{-2/5}. \quad (72)$$

Before presenting detailed numerical results, we collect the formulas used:

$$P = 10^{-6} I l \quad (73)$$

$$W = P \tau_p \quad (74)$$

$$S = \frac{P}{\pi a^2} \quad (75)$$

$$R = \begin{cases} 0.01(E/119)^{1/42} & (\text{Na}^+) \\ 0.01(E/51.5)^{1/55} & (\text{C}^+) \end{cases} \quad (76)$$

$$l = 2a \quad (77)$$

$$\rho = \frac{R}{l} \quad (78)$$

$$T_i = 55.8 S_i^{0.25} \quad (79)$$

$$\tau = 14.9 \left(\frac{R}{0.01}\right) \left(\frac{\rho}{0.1}\right)^{0.12} \left(\frac{T_i}{100}\right)^{2.5} \quad (80)$$

$$\tau_d = \sqrt{2} \sqrt{27} (788.0) \frac{I}{\sqrt{T_i(Z^* + 1)}} \quad (81a)$$

$$\phi \left(Z_i^* + \frac{1}{2} \right) = T_i \ln \left(\frac{10^{-2} T_i^{3/2} A_T}{\rho Z_i^*} \right) \quad (81b)$$

In these formulas, we have used the following quantities and units:

Quantity	Unit
P	TW
E	MeV
l	A
W	kJ
τ_p	ns
S	TW/cm ²
a	cm
R	g/cm ²
l	cm
ρ	g/cm ³
T_i	eV
S_i	TW/cm ²
τ	ns
τ_d	ns
Z_i^*	none
ϕ	eV
A_i	none

The numerical factor of $\sqrt{2}$ in Eq. (81a) results from using T_i and Z_i^* instead of mean values; the factor of $\sqrt{27}$ reflects the mass number of aluminum (27).

Table 3-4 gives results for Na^+ and C^+ beams for a variety of plausible HTE parameter sets. T_i is consistently higher for $a = 0.05$ cm, but the smaller radius will be difficult to achieve because of the smaller required momentum spread. C^+ would be preferred on the basis of the lower kinetic energies required to achieve a given T_i , but the higher currents required for high irradiance may be very unattractive for the economics of accelerator beam transport.

In the discussion above, and particularly with reference to Table 3-4, the pulse length was actually selected to slightly exceed the scale time for energy loss, so that the value T_i could be approached. If the disassembly time turns out to be shorter than the radiation scale time, then we have made a poor choice of parameters. A reduction of radius and proportionate reduction of pulse length is one possible direction of improvement. This may strain focal requirements, but it reduces total pulse energy.

Ion	Mass	E (MeV)	a (cm)	T_f (eV)	Z_i	R (g/cm ²)	ρ (g/cm ³)	τ (ns)	τ_d (ns)	I (A)	τ_p (ns)	P (TW)	S (TW/cm ²)	W (kJ)
Na ⁺	23	100	0.10	70.5	9.5	0.00781	0.0391	31.2	42.6	800	32.0	0.080	2.55	2.56
		100	0.05	99.7	10.5	0.00781	0.0782	12.1	17.1	800	13.0	0.080	10.2	1.04
		150	0.10	78.0	10.0	0.0139	0.0695	40.3	39.5	800	42.0	0.120	3.82	5.04
		150	0.05	110.3	10.2	0.0139	0.139	15.6	16.5	800	17.0	0.120	15.3	2.04
		200	0.10	83.8	9.4	0.0209	0.105	48.3	39.2	800	50.0	0.160	5.09	8.00
		200	0.05	118.5	10.4	0.0209	0.209	18.6	15.8	800	20.0	0.160	20.4	3.20
C ⁺	12	50	0.10	70.5	9.6	0.00955	0.0478	37.3	42.4	1600	40.0	0.080	2.55	3.20
		50	0.05	99.7	10.3	0.00955	0.0955	14.4	17.3	1600	15.0	0.080	10.2	1.20
		75	0.10	78.0	9.3	0.0179	0.0895	50.3	40.9	1600	52.0	0.120	3.82	6.24
		75	0.05	110.3	10.0	0.0179	0.179	19.5	16.6	1600	21.0	0.120	15.3	2.52
		100	0.10	83.8	9.1	0.0280	0.140	62.3	39.8	1600	65.0	0.160	5.09	10.40
		100	0.05	118.5	10.3	0.0280	0.280	24.1	15.8	1600	25.0	0.160	20.4	4.00

LASNEX Simulations Confirm Point-Model Disk Temperatures. The simulation described in Refs. 21 and 22 demonstrates the temperature rise and disassembly described here (even though parameters in that study differ from current HTE parameters). Among other examples, Refs. 21 and 22 give details for 100-MeV Na⁺ ions incident on an aluminum target in a lead tamper, with the following parameters (case A):

$$\begin{aligned}
 a &= 0.017 \text{ cm}, \\
 \rho &= 0.15 \text{ g/cm}^3, \\
 P_i &= 0.3 \text{ TW}, \\
 S_i &= P_i/\pi a^2 = 330 \text{ TW/cm}^2, \\
 \tau_p &= 10 \text{ ns}, \\
 l &= 0.1 \text{ cm}.
 \end{aligned}$$

The ion pulse has a 1-ns rise and fall time and an 8-ns flat top, so that $W = 2.7$ kJ.

Case A differs from values usually assumed for the HTE, primarily in the very small radius ($a = 0.017$ cm), which was selected to drive the temperature to over 200 eV. While our preliminary work^{21,22} concentrated on an accelerator test facility that could reach $\sim 10^6$ K (86 eV) in deposition-physics experiments, we used a conceptual accelerator design with 400 beamlets (see Ref. 31). Such a facility was considered capable of exceeding 10^6 K in physics experiments, so that a number of examples given in Ref. 21 were for higher temperatures, even though there were other examples close to the typical goal of 10^6 -K plasmas.

The computed $T(t)$ for case A is shown in Fig. 3-15, reproduced from Ref. 21. The peak temperature of 225 eV, reached at 2.3 ns, agrees well with the radiation-limited value

$$T_i = (55.8 \text{ eV})(330)^{1/4} = 238 \text{ eV} \quad (82)$$

The 238-eV value is approached very closely (235 eV) at a depth of 0.025 cm. More generally, we have reported²¹ that surface temperature is $\sim 20\%$ below the peak interior value.

We attribute the steep falloff of surface temperature after 2.3 ns to disintegration of the surface by hydrodynamic expansion.²¹ This behavior can be understood by observing that the disassembly time for this target is about 5.6 ns. (We use $R = 7.8 \text{ mg/cm}^3$ and $T_i = 238 \text{ eV}$.) After about 2 ns, as shown in a LASNEX zone diagram in Ref. 21, an 0.01-cm thickness of aluminum will be exiting the target, but the interior should be relatively undisturbed. The exiting layer will expand sideways and receive a much-reduced energy deposition from the ion beam. Crudely, we expect the temperature of this layer to drop on the energy-loss time scale given by

Table 3-4. HTE parameters for Na⁺ and C⁺ ions of various energies with $a = 0.1$ and 0.05 cm. (Depth $l = 2a$.)

Fig. 3-15. Surface temperature vs time for case A (from LASNEX).

Charged-Particle Fusion

Eq. (70), namely $\tau \approx 9$ ns, in agreement with Fig. 3-15.

Another check of the model assumptions is provided by the temperature rise during the first 1 ns, during which we can neglect expansion and energy loss. In $t = 1$ ns the deposited energy is

$$\frac{t}{2} \frac{S_i}{R} = 21\,000 \text{ kJ/g} \quad (83)$$

Fig. 3-16. Rates of energy deposition and loss at front of disk for case B (from LASNEX).

Fig. 3-17. Total energy deposited and lost, and surface temperature, for case B (from LASNEX).

Fig. 3-18. Internal temperature in aluminum deposition material at 31 ns for case B (from LASNEX).

which corresponds to $T \approx 200$ eV, in good agreement with the LASNEX value of ~ 180 eV (Fig. 3-15).

The 72- and 92-eV results of Fig. 7 in Ref. 21 or Table 1 in Ref. 22 are typical of current numbers for the HTE. Disassembly of deposition material is more pronounced for the long pulses now being considered, so that a new simulation is in order. Accordingly, we consider 125-MeV Na^+ ions incident on an aluminum target in a lead tamper, with the following parameters (case B):

$$\begin{aligned} a &= 0.1 \text{ cm}, \\ \rho &= 0.09 \text{ g/cm}^3, \\ P_i &= 0.122 \text{ TW}, \\ S_i &= P_i / \pi a^2 = 3.87 \text{ TW/cm}^2, \\ \tau_i &= 30.9 \text{ ns}, \\ l &= 0.2 \text{ cm}. \end{aligned}$$

Figure 3-16 shows the assumed ion-pulse shape: a 0.5-ns rise and fall time, and a 29.9-ns flat top, corresponding to $W = 3.7$ kJ. Figure 3-16 also shows the rate of energy loss from the front of the disk.

Figure 3-17 gives total energy deposited, total energy lost, and surface temperature vs time. Figure 3-18 gives contour plots of internal temperature in aluminum deposition material at 31 ns. The simulation includes only a small part of the high-Z disk that supports and tamps the low-Z deposition material.

Inclusion of Hydrodynamic Disassembly. We can obtain a slightly more complex point model than that of Eq. (67) by including some effects of hydrodynamic disassembly, as follows:

$$\begin{aligned} \frac{d}{dt} \left[\epsilon(t) + \frac{1}{2} \eta c_s^2(t) \right] &= \frac{1}{R} \\ &\times \left\{ S(t) - \sigma T^4(t) - \frac{\sigma T^4(t)}{\pi r_0^2} 2\pi r [l(t) - l_0(t)] \right\} \quad (84) \end{aligned}$$

where η is some typical order-1 parameter to relate kinetic energy per unit mass to the instantaneous sound speed of the medium, which depends on temperature. Equation (84) becomes one for $T(t)$ if we make simple additional assumptions, such as

$$l(t) \approx l_0 + \int_0^t c_s(t') dt' \quad (85)$$

and use

$$\rho(t)/\rho_0 \approx \rho_0 I_0 \quad (86)$$

Further realism might be obtained by including heat losses to the high- Z material²¹; but in case B this loss is only $\sim 20\%$ (at 31 ns) of the already small energy loss indicated in Fig. 3-17. Another improvement might be to include the temperature dependence of the factor Z^* in Eq. (71). With these improvements, however, the point model becomes too complex for evaluation with a hand calculator, and we might as well use LASNEX calculations to obtain accuracy at these levels.

Authors: E. P. Lee (LBL) and
J. W.-K. Mark

Major Contributors: R. O. Bengter and
R. Kopp (Los Alamos)

Testing Hydrodynamic Models for Simulation of Pinched Beams

Hydrodynamic models have been developed by Mark and Yu³² and by others to describe energetic pinched beams, such as those used in ion-beam fusion. The closure of the Mark-Yu model is obtained with adiabatic assumptions mathematically analogous to those of Chew, Goldberger, and Low³³ for magnetohydrodynamics (MHD). The other models treated here use an ideal-gas closure and a closure by Newcomb based on an expansion in V_{th}/V_z . Predictions of these hydrodynamic beam models are compared with those obtained from a kinetic treatment. The Mark-Yu model can reproduce the global axisymmetric phase-mix damping typical of real beams. For non-axisymmetric disturbances, it also exhibits resonance structures like those of the kinetic theory.

The development of self-pinched beams will be advantageous for long-term fusion reactor applications. Among the expected advantages are higher beam currents, fewer and smaller beam ports, as well as beam propagation at somewhat higher fusion-chamber pressures (~ 1 to 10 Torr, as against $\lesssim 10^{-3}$ Torr for ballistic-propagation modes), which will permit higher liquid-metal temperatures for reactor efficiency and for additional wall protection. If we

develop supplementary final-focus techniques that increase the focal-spot beam energy flux beyond the fluxes envisioned in the previous article, the high-temperature experiment for heavy-ion fusion could reach beam and target plasma parameters even closer to those of reactor requirements.

Clearly, for a physics experiment, we could use supplementary final-focus schemes that (because of such matters as efficiency or repetition rate) are not directly related to the major purpose of testing heavy ion fusion (HIF) accelerator hardware capable of meeting reactor requirements for long-term use at high repetition rates. For example, in the external-current pinch proposed in Refs. 34 and 35, the beam energy flux is increased even at the possible sacrifice of up to 2/3 of the particles.

G. Krafft³⁶ of LBL is performing experiments related to a final-focusing scheme due to S. Robertson.³⁷ The Mark-Yu hydrodynamic beam model^{32,38} was proposed in part for the purpose of simulating such phenomena.

Fluid-beam models (such as the one described in Refs. 32 and 38) promise to be more economical than particle simulations for the three-dimensional, nonlinear beam-dynamics problems now under study. The fundamental question posed by these fluid models is whether they can adequately describe a self-pinched particle beam in which there are kinetic effects (such as phase-mix damping of radial beam oscillations and wave-particle resonances for non-axisymmetric instabilities) transverse to the direction of beam propagation. In this paper we compare the results of several fluid-beam models with the descriptions of these two phenomena given by the correct kinetic theory. For generality, we include the natural extensions to relativistic beam velocities.

All of the fluid beam models we have considered are obtained by closing a set of moments of the Vlasov equation by making some appropriate assumption about the beam pressure. We make several additional assumptions here to reduce the number of equations in each set. For a beam propagating in the positive z direction, the parallel velocity $V_z \equiv \beta c$ is assumed to be nearly uniform, so that the distance from the beam head $\xi = \beta ct - z$ is constant for each beam slice, and all z components of the beam pressure tensor vanish. This assumption is quite well satisfied when the beam is

axially cold, as in heavy-ion fusion or in a highly relativistic beam, when the transverse velocity of particles is small compared with V_z . When these conditions are satisfied, the transverse force on the beam can be written (to order γ^{-2}) as the gradient of a scalar pinch potential $a \approx A_z - \phi$. We neglect all energy losses, so that γ is constant.

The MHD theory of Chew, Goldberger, and Low³³ for neutral plasmas was generalized by Berman and Mark^{39,40} for application to a non-neutral gravitational analog of a Vlasov plasma. Within the Wentzel-Kramers-Brillouin (WKB) regime of short radial wavelengths, Berman and Mark also compared this model numerically with kinetic theory and with an ad hoc isothermal gas model. Mark and Yu^{32,38} modified this model for application to self-pinch beams that could be nonaxisymmetric, nonlinear, and partially neutralized. Mark⁴¹ developed a version of this model for some accelerator-beam problems. Recently, Buchanan and Lee⁴² proposed an ad hoc axisymmetric model of pinched beams, and Newcomb⁴³ developed his hydrodynamic model, that overcame some of the deficiencies in the model of Siambis.⁴⁴

Mark-Yu Hydrodynamic Pinched-Beam Model. The principal assumption of the Mark-Yu model^{32,38} is that the bulk of beam particles have rosette orbits with at most about a 2:1 ratio of maximum to minimum radial excursions. These orbits are decomposed into a drifting guiding-center motion and a gyration. In steady, axisymmetric fields, the guiding center does not drift radially, so that its orbit is circular; the gyration about the guiding center is an ellipse. Even in nonaxisymmetric, time-varying fields, the angular frequency of the guiding center is the generalized betatron frequency, given by

$$\Omega_d^2(r, \theta, z, t) \approx -\frac{q\beta}{\gamma m r} \frac{\partial n}{\partial r}, \quad (87)$$

and the radial gyration frequency ν is approximately the value for small radial oscillations about a betatron orbit,

$$\nu^2(r, \theta, z, t) \approx \frac{1}{r^3} \frac{\partial}{\partial r} (r^4 \Omega_d^2). \quad (88)$$

The frequency ν provides the analogy to the cyclotron frequency of the MHD theory of Chew, Goldberger, and Low.³³

A drift-kinetic equation for the distribution function of guiding-center orbits is obtained by averaging over the radial gyration period. For a slowly evolving beam with small longitudinal velocity spread, this equation yields adiabatic relations between ν and the components P_{rr} , $P_{\theta\theta}$, and P_{zz} of the stress tensor $\mathbf{P}$ that closes the set of moment equations. In cylindrical coordinates, the equations are the exact continuity and momentum equations,

$$\frac{DN}{Dt} = -N \left[\frac{1}{r} \frac{\partial}{\partial r} (r V_z) + \frac{1}{r^2} \frac{\partial j}{\partial \theta} \right], \quad (89a)$$

$$\begin{aligned} \frac{DV_z}{Dt} &= \frac{r^2}{r^3} \\ &\quad - \frac{1}{\gamma m N} \left[\frac{\partial P_{rr}}{\partial r} + \frac{1}{r} (P_{\theta\theta} - P_{zz}) \right] \\ &\quad + \frac{q\beta}{\gamma m} \frac{\partial a}{\partial r}, \end{aligned} \quad (89b)$$

$$\frac{Dj}{Dt} = -\frac{1}{\gamma m N} \frac{\partial P_{\theta\theta}}{\partial \theta} + \frac{q\beta}{\gamma m} \frac{\partial a}{\partial \theta}, \quad (89c)$$

and the adiabatic pressure relations,

$$\frac{D}{Dt} \left(\frac{r P_{rr}}{|j|^{1/2} N^{3/2}} \right) = 0, \quad (89d)$$

$$\frac{D}{Dt} \left(\frac{|j|^{1/2} P_{\theta\theta}}{r N^{5/2}} \right) = 0, \quad (89e)$$

$$P_{zz} \approx 0. \quad (89f)$$

For initial values, the supplementary relation

$$P_{\theta\theta} = \frac{r^4 \nu^2}{4 j^2} P_{rr} \quad (89g)$$

might be useful when coupled with Eq. (88).

Because there are two directions of rotation for rosette orbits in a nonrotating beam, the theory in this case naturally calls for two counter-rotating fluid components, each evolving according to Eq. (89).

The guiding-center orbit theory and the closure equations (89d) to (89g) were derived under the adiabatic assumptions

$$\frac{\delta r}{r} \sim \frac{v_{\perp}}{v_{\theta}} \sim \frac{\partial}{\partial \theta} \sim \epsilon, \quad \text{with } \epsilon < 1. \quad (90)$$

Here δr is the deviation from average radius in the particle orbit, p_r , p_{θ} and p_z are the particle momenta, and it is assumed that there is some reference frame moving at speed $v_{z,p}$ in the z direction, where

$$\left| \frac{\partial}{\partial t} + (v_z - v_{z,p}) \frac{\partial}{\partial z} \right| = O(\epsilon). \quad (91)$$

This frame is useful only for showing the validity of the particular mode of derivation. We note that nonlinear waves are allowed under these assumptions. Also, Mark and Yu did not require as restrictive assumptions on axially cold beams as those adopted in the fifth paragraph of this article. They obtained an adiabatic relation for the stress component P_{zz} , whose effects we have ignored for convenience of testing transverse dynamics.

Exactly analogous hydrodynamic models have been developed by Berman and Mark^{39,40} for so-called gravitational plasmas, studied in connection with spiral-structure phenomena in disk galaxies. They found ample evidence⁴⁰ that this type of hydrodynamic model remains applicable in the regime of short radial wavelengths, i.e., the WKB regime, in which the adiabatic assumptions used in the derivation are supposedly no longer valid. In this work the most accurate numerical results were obtained when all terms of the order of $P_{\theta\theta}$ were included.

Numerical Comparisons of Axisymmetric Phase-Mix Damping. We have made several numerical comparisons between fluid models and particle simulations. These comparisons were made in cylindrical geometry, assuming axial symmetry and no variation in z along the direction of beam propagation. This assumption eliminates the z component of the electric field from the model and reduces the problem to one dimension. With this simplification, quantities have only a radial dependence, and it is

advantageous to use a Lagrangian fluid formulation. For both the fluid model and the particle simulation, we assume that the radial motion is decoupled from the longitudinal motion.

The fluid equations in Lagrangian form identically satisfy the continuity equation. Given a closure relation for pressure, one need then only solve for position according to $(dr/dt) = V_r$, and for velocity V_r , according to Eq. (89b).

The pinch force can be expressed in terms of enclosed mass M by integrating Ampere's equation, with the result

$$\frac{\partial a}{\partial r} = -\frac{2qM}{\gamma m r}, \quad (92)$$

where

$$M = 2\pi \int \rho r dr. \quad (93)$$

Using Eq. (92) in Eq. (89b) and expressing the pressure derivative in terms of M , one obtains the momentum equation solved by the fluid code,

$$\begin{aligned} \frac{DV_r}{Dt} &= \frac{j^2}{r^3} - \frac{2M}{r} \left(\frac{q}{\gamma m} \right)^2 \\ &\quad - 2\pi r \frac{\partial P_{zz}}{\partial M} - \frac{1}{r\rho} (P_{zz} - P_{\theta\theta}). \end{aligned} \quad (94)$$

In the case of the cold beam, the kinetic treatment reduces to the fluid model, and equilibrium is established as a balance between the angular momentum and the pinch force. With j^2 expressed in terms of M , Eq. (94) becomes

$$\frac{DV_r}{Dt} = 2 \left(\frac{q}{\gamma m} \right)^2 M \left(\frac{r_0^2}{r^3} - \frac{1}{r} \right) \quad (95)$$

at position r_0 . A perturbation $r = r_0(1 + \epsilon)$ then yields the first-order solution

$$r = r_0 - \epsilon r_0 \cos \omega t, \quad (96a)$$

$$V_r = \omega \epsilon r_0 \sin \omega t, \quad (96b)$$

$$\omega = \frac{2}{r_0} \frac{q}{\gamma m} M^{1/2}. \quad (96c)$$

For a Gaussian radial current profile $I_z = I_B (\pi b^2)^{-1} \exp(-r^2/b^2)$, Eq. (96) gives the rms radius R , to first order, as

$$R = \left[b^2 - 2\epsilon b^2 \int_0^\infty y e^{-y} \cos \lambda dy \right]^{1/2}, \quad (97)$$

$$\lambda = \frac{2Z}{b} \left(\frac{I_B}{I_A} \right)^{1/2} \left(\frac{1 - e^{-y}}{y} \right)^{1/2}, \quad (98)$$

where I_B is total beam current and I_A is the Alfvén current.

Figure 3-19 shows the behavior of a cold, Gaussian-profile beam with an initial rms radius of 1 cm, subjected to a self-similar compression; the beam oscillates, exhibiting phase-mix damping. The motion is obtained from the analytic Eq. (97) and from computer runs with the particle and fluid codes. The fluid-code result agrees well with the solution of Eq. (97). The amplitude of the particle-simulation curve is slightly greater, partly because of the difference between the profile cutoffs in the two codes. The oscillation frequencies of the particle and fluid results agree with that given by the analytic solution.

For systems with nonzero pressure, Mark, Buchanan and Yu⁴⁵ have compared the Mark-Yu fluid model with axisymmetric phase-mix damping of radial oscillations in particle codes. They used 3000 particles in a Gaussian-profile electron beam with an initial radius of 1 cm, assuming full space-charge neutralization but no current neutralization. Phase-mix damping resulted in an oscillation frequency and a damping rate similar to those in the particle simulation. Further investigations are being

pursued to determine the effect of different initial conditions.

Beam Response to Nonaxisymmetric Disturbances. We have used the various fluid models to determine beam response to small-amplitude, nonaxisymmetric field disturbances. As appropriate equilibrium configurations, we look for solutions of the fluid equations that are axisymmetric and time-independent in the particle frame of reference. Because $\xi = \beta ct - z$ is a constant of the particle motion, and β is assumed to be essentially the same for all particles, time-independence is ensured by changing variables from (z, t) to (ξ, z) , and then requiring that $\partial/\partial z = 0$. Because the equilibrium quantities depend only on r and ξ , an arbitrary perturbation may be Fourier-analyzed into linearly independent modes of the form

$$\psi_1(r, \theta, \xi, z) = \psi(r, \xi) \exp(i\ell\theta - i\Omega z/\beta c), \quad (99)$$

where ψ represents any of the beam quantities and Ω is the wave frequency seen by a fluid element. To obtain a set of linear differential equations for the perturbation amplitude ψ , we substitute $\psi = \psi_0(r, \xi) + \psi_1(r, \theta, \xi, z)$ for each quantity in the fluid equations. The resulting equations have no explicit ξ dependence because a constant- ξ beam slice couples to the rest of the beam only through the field equations. The first-order perturbation equations are then reduced to a single inhomogeneous differential equation for the perturbed density amplitude $\hat{n}$. The complexity of the $\hat{n}$ equation makes it expedient to solve for $\hat{n}$ analytically near the singular points of the

Fig. 3-19. Phase-mix damping.

equation, and to compare the analytic form with the results of a Vlasov treatment.

For the Mark-Yu model of Eq. (89), a differential equation is derived for $\hat{n} = \hat{N}/N_0$ of the form

$$\frac{d^2 \hat{n}}{dr^2} + \lambda_1 \frac{d\hat{n}}{dr} + \lambda_0 \hat{n} = S \quad (100)$$

Inspection of the coefficients in Eq. (100) shows three regular singular points. The strongest singularity occurs at a radius r_1 given by

$$\Omega = i \frac{j_0}{r_1^2} = i \Omega_p(r_1) \quad (101)$$

where Ω_p is given by Eq. (87) and j_0 is the equilibrium fluid angular momentum. A power-series solution of Eq. (100) for small $x = r - r_1$ gives two solutions of the form

$$\hat{n} = x^p (1 + C_1 x + C_2 x^2 + \dots) \quad (102)$$

where $p = 0$ or -1 in the limit of zero pressure, and $p > -1$ for nonzero pressure in a beam with monotonic N_0 and A_0 profiles. A formal calculation of $\hat{N}$ from the Vlasov equation⁴⁶ shows that, for a beam with $P_{rr} = \gamma m N_{00}^2 / r^2 + P_{\theta\theta}$, $\hat{n}$ has a logarithmic singularity at the radius given by

$$\Omega = \Omega_p(r) \quad (103)$$

where Ω_0 is the time average of $\dot{\theta}$ of a particle with perpendicular energy $H_\perp = qB A_0(r)$ and zero angular momentum. Because it can be shown that $\Omega_p = \Omega_g$ for $r = 0$ and that $\Omega_p/\Omega_0 \approx 2/3$ for large r , the fluid-resonance condition Eq. (101) is quite similar to Eq. (103). The weakening of the $\hat{n}$ singularity with nonzero pressure in the fluid treatment is also qualitatively correct. To determine whether the fluid model gives the correct functional dependence on pressure, however, one must numerically solve the dispersion equation obtained by combining Eq. (100) with Ampere's equation, and compare the results with the Vlasov result. The other two singular points of Eq. (100) yield much weaker singularities than that given by Eq. (101) and are thus physically inconsequential.

Equations for $\hat{n}$ have also been derived from a set of beam fluid equations with an ideal-gas closure (similar to the isothermal

gas model of Berman and Mark⁴⁰ and to the Buchanan model⁴²), and for the Newcomb mod. I.⁴³ In the ideal-gas model, two counter-rotating fluid components are used, and the transverse pressure is assumed to be isotropic. The dominant singular point is again given by

$$\Omega = \frac{j_0^2}{r_1^2} \quad (104)$$

and near this point $\hat{n}$ has the analytic form of Eq. (102) with a qualitatively similar pressure dependence. An important difference between Eq. (104) and the previous resonance condition, Eq. (101), is that j_0 is arbitrary here, whereas the relation $j_0 \approx r \Omega_p$, needed for approximate agreement with Eq. (103), is fundamental to the Mark-Yu formulation. In the Newcomb model, only a single fluid component is used instead of two counter-rotating components, so the fluid stress $\gamma m N_i^2 / r^2$ is absorbed into $P_{\theta\theta}$, and the off-diagonal pressure component $P_{r\theta}$ must be retained. These features lead to more complicated closure relations and to a fourth-order differential equation for $\hat{n}$. All the singularities in $\hat{n}$ found from the Newcomb model are much weaker than those found in a Vlasov treatment, with the most singular having a form $\hat{n} \sim x^4 \ln x$ near $x = r - r_1 = 0$. Evidently the kinetic-like resonances seen in the Mark-Yu and ideal-gas models result from coupling between a wave and the equilibrium azimuthal fluid motion and are absent from a single-component model like that of Newcomb.

Authors: J. K. Boyd, J. V. K. Mark, W. M. Sharp, and S. S. Yu (SLAC)

Plasma Physics

Introduction

The coupling of laser light with plasmas is a key issue in laser fusion and in other laser applications. Our plasma physics program consists of a three-pronged effort: the theory and simulation of laser-plasma coupling and electron-energy transport, the design and interpretation of coupling experiments, and the development of

advanced plasma codes. Highlights of our progress in 1983 are briefly reviewed here.

We have continued to focus our effort in theory and simulations on coupling mechanisms in long-scalelength plasmas. These mechanisms include Raman and Brillouin scattering, the $2\omega_{pe}$ instability, and filamentation and self-focusing due to ponderomotive and thermal mechanisms. We have determined the theoretical intensity threshold for Raman sidescattering in an inhomogeneous plasma with obliquely incident pumping light, and have carried out simulations that show the reduction of Raman scattering by electron-ion collisions. A theory of the stabilization of filamentation by diffuse angular spread has been developed. Thermal self-focusing, including the stabilizing influence of beam divergence, has been calculated for reactor-size plasmas. We also report work on electron heating by beats between 0.53- μm and 1.06- μm laser light, and on electron reheating by localized plasma waves.

A sizable effort was again devoted to the design and interpretation of coupling experiments in collaboration with the Fusion Experiments Group. We completed an extensive analysis of the Argus experiments on the wavelength scaling of electron transport.⁴⁷ Initial coupling experiments with the Novette laser have extended our short-wavelength data base to include larger regions of underdense plasma than were accessible with the Argus laser. Initial Novette experiments with exploding foils have allowed us to access underdense plasmas with longer density scalelengths than before. This experimental work is reported in Section 5.

We have increased our attention to the development of advanced plasma codes. We are constructing an implicit two-dimensional electromagnetic particle code that will extend our capability to simulate low-frequency kinetic phenomena. This code uses the so-called direct method, which we developed earlier.

Author: W. L. Kruer

Stimulated Raman Scattering in an Inhomogeneous Plasma with an Obliquely Incident Pump

The theoretical threshold for stimulated Raman scattering (SRS) is lower at oblique

than at normal incidence. Since some degree of oblique incidence is unavoidable in laser fusion applications if uniform illumination is to be achieved, this implies that SRS may be a greater source of hot electrons than has been assumed. Experiments are in order to determine if oblique incidence in fact leads to an increased production of hot electrons.

In SRS, incident laser light is scattered by stimulated electron plasma waves (epw) in the underdense ($n < n_c/4$) target corona. These epws, when strongly driven, accelerate background electrons to suprathermal velocities, giving rise to a hot-electron component that can preheat the DT fuel and cause the target to implode inefficiently.

Recent experiments on the Novette laser (reported in "Novette Large-Plasma Interaction Experiments" in Sec. 5) show that hard x rays from hot electrons correlate strongly with the Raman-scattered light from the plasma, indicating a close connection between SRS and the production of hot electrons.

As a resonant three-wave interaction, SRS must occur at plasma densities at which the so-called frequency-matching condition is satisfied:

$$\omega_0(\mathbf{k}_0) = \omega(\mathbf{k}) + \omega_e(\mathbf{k}_0 - \mathbf{k}) \quad (105)$$

Here $(\omega_0, \mathbf{k}_0)$, $(\omega, \mathbf{k})$, and $(\omega_e, \mathbf{k}_0 - \mathbf{k})$ are the frequencies and wave numbers of the pumping laser light, the scattered electromagnetic (em) wave, and the epw, respectively; they satisfy the dispersion relations

$$\omega_0^2 = \omega_p^2 + c^2 k_0^2,$$

$$\omega^2 = \omega_p^2 + c^2 k^2, \quad (106)$$

$$\omega_e^2 = \omega_p^2 + 3 v_T^2 (\mathbf{k}_0 - \mathbf{k})^2,$$

where v_T is the electron thermal velocity.

In inhomogeneous plasmas, the electron-density gradient in the corona localizes the region of SRS instability; the requirement that the scattered em wave grow (by some stated but arbitrary amount) before it propagates out of this region gives rise to a threshold laser intensity for the instability. This threshold is much lower for sidescattering, in which the em wave propagates at right angles to the density gradient and thus leaves the instability region only slowly (by refraction).

In previous work^{48,49} the laser light was considered to be normally incident on a cold plasma. We have extended this work to include the effects of oblique incidence and of warm (but fluid) plasma.

Nonzero plasma temperature gives rise to frequency shifts of the scattered light of order (v_T^2/c^2) , and alters the spatial structure of the unstable plasma wave, but does not substantially affect the thresholds or the growth rate.

The effects of oblique incidence, however, are significant, at least for s-polarized light. For p-polarization, sidescattering occurs preferentially out of the plane of incidence defined by $\mathbf{k}$ and $\mathbf{E}_0$; as might be expected from symmetry considerations, oblique incidence of p-polarized light has little effect on the instability. However, for s-polarization, sidescattering occurs preferentially in the plane of incidence and towards the pump. This is understandable, because the corresponding epw wave vector (to which the homogeneous growth rate is proportional) is larger in this case.

We used the technique⁵⁰ of transforming the coupled wave equations into k -space, so that linear density gradients gave rise to k -derivatives and to a second-order eigenvalue problem for the growth rate in inhomogeneous plasmas. This technique has also been applied to the $2\omega_p$ instability.^{50,51}

The threshold condition for SRS incident at a vacuum angle of incidence θ_0 turns out

to be

$$\begin{aligned} \left(\frac{v_0}{c}\right)^2 \left(\frac{\omega_0 L}{c}\right)^{4/3} &\geq f(\omega_p, \theta_0) \\ &= \left(\frac{\omega_p}{\omega_0}\right)^{2/3} \left[\left(2 - \frac{2\omega_p}{\omega_0} - \frac{\omega_p^2}{\omega_0^2} \right) \pm 2 \right. \\ &\quad \times \left. \left(1 - \frac{2\omega_p}{\omega_0} \right)^{1/2} \sin \theta_0 \right]^{-1}, \quad (107) \end{aligned}$$

where

$$\left(\frac{v_0}{c}\right)^2 = \frac{I \lambda^2}{1.37 \times 10^4}, \quad (108)$$

here I is the peak (local) laser intensity, in units of 10^{14} W/cm², and λ_0 is in microns. For normal incidence this agrees with the cold-plasma numerical results of Liu, Rosenbluth, and White. For p-polarization there is no $\sin \theta_0$ term, so there is no obliquity effect.

Figure 3-20 is a plot of $f(\omega_p, \theta_0)$ from Eq. (107). Note that the threshold drops monotonically with density (through ω_p) from the value often quoted for quarter-critical density. However, at sufficiently low density ($k\lambda_0 \sim 0.3$) kinetic effects (neglected here) should be included in the analysis, and the threshold can be expected to rise because of Landau damping of the epw.

Fig. 3-20. Threshold for Raman side-scattering in an inhomogeneous plasma, for an s-polarized pump, at various angles of incidence. From Eq. (107).

(Landau-damping effects on the SRS threshold are discussed in the next article.)

As an aside, we note that Eq. (107) can be obtained by a suitable interpretation of the well-known Rosenbluth result⁵²

$$\gamma_0^2 > \kappa' v_1 v_2, \quad (109)$$

in which γ_0 is the homogeneous-plasma growth rate, κ' is the spatial gradient of the wave-number mismatch, and v_1 and v_2 are the components (in the direction of the density gradient) of the group velocities of the scattered em wave and epw respectively. For sidescattering v_1 vanishes, but an effective group velocity can be obtained by taking the wave number in the direction parallel to ∇n to be the "Airy wave number" $(\omega_p^2/c^2 L)^{1/3}$. Then (within undetermined numerical factors of order 1) one has

$$\begin{aligned} \gamma \sim & \left(\frac{\omega_p}{\omega_0 - \omega_p} \right)^{1/2} \left(\frac{v_0}{c} \right) \omega_0 \\ & \times \left[\left(2 - \frac{2\omega_p}{\omega_0} - \frac{\omega_p^2}{\omega_0^2} \right) \right. \\ & \left. \pm 2 \left(1 - \frac{2\omega_p}{\omega_0} \right)^{1/2} \sin \theta_0 \right]^{1/2}, \\ \kappa' v_2 \sim & \omega_p / L, \end{aligned} \quad (110)$$

$$v_1 \sim \frac{c^2}{\omega_0 - \omega_p} \left(\frac{\omega_p^2}{c^2 L} \right)^{1/3}.$$

Substituting Eqs. (110) into the Rosenbluth result Eq. (109), we recover our Eq. (107).

The lowering of the threshold for SRS at oblique incidence has implications for the efficiency of ICF target heating. We are proposing oblique-disk experiments to test the implications of our results.

Authors: E. A. Williams and B. Afeyan (LLE, Rochester)

Thresholds of Convective Raman Backscattering: Effects of ∇n , Noise, Collisions, and Landau Damping

We present kinetic simulations and theory of the threshold of Raman backscattering for a variety of electron-density profiles,

electron-ion collision frequencies, and background temperatures (which affect Landau damping). The simulations show Raman-backscattering fractions of $\sim 4 \times 10^{-4}$ at intensities ~ 30 times below the ∇n threshold; we suggest that this is due to light scattering from noise electron plasma waves. We find that in many experiments, collisions significantly affect Raman scattering.

The Raman instability⁵³⁻⁶³ is an important consideration for laser fusion reactors, because their targets will be characterized by large regions of underdense plasma. (Reference 56 gives a more complete list of references.) Raman scattering can generate high-energy electrons, which can cause significant preheat. The Raman instability is sensitive to the density gradient,⁵³ ∇n , which acts to detune the resonance between the laser light, the scattered light, and the electron plasma wave.

Last year, Tanaka et al.⁵⁴ at Rochester, Phillion et al.⁵⁵ at Livermore, and Shepard et al.⁵¹ at KMS Fusion, Inc. (KMS) reported experimental observations of convective Raman scattering well below the traditional density-gradient threshold. Here we report theory and electromagnetic, relativistic, one-dimensional kinetic (particle) simulations^{56,58} that suggest interpretations for such low-level observations, and we expand the threshold calculations to include collisional and Landau damping.

Figure 3-21 shows the Raman-backscattered spectrum in a numerical simulation in which the intensity $I_0^2 = 3 \times 10^{13} \text{ W} \cdot \mu\text{m}^2/\text{cm}^2$, is a factor of 30 below the ∇n threshold; the signal amplitude is roughly an order of magnitude above the noise. The electron temperature was 1 keV; n/n_c ran from 0.125 to 0.175 in a distance $128\lambda_0$. We interpret this spectrum as scattering from electron plasma wave noise that is inherent in the particle simulation and that, in fact, emulates the noise in a real plasma. The scattered frequency ω_s corresponds to the solution of $k_0 = k_s + k_{\text{epw}}$ and $\omega_0 = \omega_s + \omega_{\text{epw}}$ as in Raman scattering (the subscripts 0, s, and epw refer to the pumping, scattered, and electron plasma wave, respectively.) The reflected spectrum shows a similar peak at $\omega_0 + \omega_{\text{epw}}$; the transmitted light shows a similar peak (although at lower intensity) at $\omega_0 \pm \omega_{\text{epw}}$. A control simulation without the pump does not show such a peak.

Figueroa et al.⁶⁰ quantify the competition between bremsstrahlung and Thomson scattering for an experimental situation and find that Thomson scattering dominates as a noise source.

Figure 3-22 shows the fraction of light absorbed by Raman backscattering in simulations in which the density-gradient scalelength is long enough to be compared to reactor targets. The ions were fixed to isolate the Raman process. The absorption measurements were taken from the maximum slope of the absorbed-energy curve over at least 60 laser cycles. As described in Ref. 56, the Raman absorption varies considerably in time. For high-gain Raman, the variation can easily be more than an order of magnitude, because the amplification is inversely proportional to the damping, which is itself mostly due to the rapidly convecting hot-electron tail. The long-time-scale average absorption is roughly a factor of 2 less than shown for high-gain Raman scattering, but about the same for low-gain Raman. In low-level cases, absorption is inferred from the scattered-light spectrum using the Manley-Rowe relation, fraction absorbed = $\omega_s/\omega_0 = 1 - \omega_s/\omega_0$ (Ref. 64).

The cold, collisionless threshold in the convective regime is obtained from the relation

$$\frac{\gamma_0^2}{\kappa' v_{\text{gepw}} v_{\text{gs}}} > 1, \quad (111)$$

using

$$\gamma_0 = \frac{k_{\text{ep}}}{4} v_{\text{osc}} \left(\frac{\omega_p}{\omega_s} \frac{\omega_p}{\omega_{\text{epw}}} \right)^{1/2} \times \left(1 - \frac{n}{n_c} \right)^{-1/4}, \quad (112)$$

where the wave number and frequency-matching conditions are used to find k_{epw} :

$$k_0 = k_s + k_{\text{epw}}, \quad (113)$$

$$\omega_0 = \omega_s + \omega_{\text{epw}}, \quad (114)$$

and

$$\kappa' = \frac{d}{dx} (k_0 - k_s - k_{\text{epw}}). \quad (115)$$

Fig. 3-21. Fourier transform in time of Raman-backscattered light for which λ_0^2 is a factor of 30 below the classical threshold.

Fig. 3-22. The effect of doubling the number of electrons from 80 000 to 160 000 is shown in the lower open squares: the absorption is lower, mostly because of lower noise.

Here

$$v_{\text{gepw}}, v_{\text{gs}} = c \left[1 - \left(\frac{\omega_p}{\omega_s} \right)^2 \right]^{1/2} \quad (116)$$

are the group velocities of the electron plasma wave and scatter light wave respectively, and $\kappa' v_{\text{gepw}} = \omega_p/(2L)$, where $L = (\Delta x/\Delta n)n$ and ω_p is the plasma frequency at the density n of interest; n changes by an amount Δn in distance Δx . Using

$$v_{\text{osc}} = \frac{eE_0}{m_e \omega_0} = 8.55 \times 10^{-10} \lambda_0 (\mu\text{m}) c [I (\text{W}/\text{cm}^2)]^{1/2}, \quad (117)$$

the threshold condition of Eq. (111) becomes

$$I\lambda_0^2 > 1.5 \times 10^{18} \frac{k_s c}{\omega_0} \frac{\lambda_0}{\Delta x} \frac{\Delta n}{n} \left(\frac{k_{\text{epw}} c}{\omega_0} \right)^{-2} \quad (118)$$

for $n/n_c \lesssim 0.15$, where the laser light has intensity I and peak vacuum electric field E_0 . For a cold electron temperature ($k_{\text{epw}} \lambda_D \ll 1$), we have

$$\frac{k_{\text{epw}} c}{\omega_0} = \left(1 - \frac{n}{n_c} \right)^{1/2} + \left[1 - 2 \left(\frac{n}{n_c} \right)^{1/2} \right]^{1/2} \quad (119)$$

Equation (5) of Ref. 53 gives the backscattering amplification:

$$I = I_0 \exp \left(\frac{2\pi \gamma_0^2}{\kappa' v_{\text{gepw}} v_{\text{gs}}} \right), \quad (120)$$

where I and I_0 are, respectively, the intensities of the scattered and noise light at frequency ω_s . While Eq. (111) gives the usually quoted instability threshold, Eq. (120) shows that the amplification at this threshold is $\exp(2\pi) = 535$. Part of the ambiguity, of course, arises from the definition of the threshold; the traditional threshold does indicate where the instability becomes large.

Of course, other plasma effects can reduce the Raman threshold. Raman sidescattering has a lower threshold, and is commonly observed. Filamentation can increase the light intensity,⁵⁹ and a parabolic density profile has a lower threshold than a linear profile.⁶⁰ Flat spots in the density profile,⁶⁵⁻⁶⁷ and Raman scattering enhanced by the Brillouin scattering wave number $2k_0$ (in quite underdense plasmas) are other sources. The density of Raman scattering is usually inferred to be quite low, where

$$k\lambda_D = \frac{k_{\text{epw}} c}{\omega_0} \frac{v_e}{c} \left(\frac{n_c}{n} \right)^{1/2} > 0.3, \quad (121)$$

and consequently the Landau damping $\omega_i \approx \omega_p$ is large. An alternative explanation for Raman scattering may be stimulated Compton scattering,⁶² which occurs when the damping is great enough to allow

enough width in ω -matching for scattering to occur for $n > n_c/4$. The growth rate, ω_s , and $k\lambda_D$ for $T_e = 10$ keV in this regime are plotted in Fig. 1b of Ref. 56. We have seen Compton scattering in a simulation with $T_e = 2$ keV and $n/n_c = 0.35$ for $64\omega_0$, $\nabla n = 0$. Of course, any electron plasma wave in a density gradient can radiate in an inverse resonant absorption process.⁶⁸ A localized electron plasma wave can, in principle, cause a local spectrum.

Nishikawa and Liu⁶³ have derived a threshold including Landau damping and the effects of ν_{ei} on the scattered wave but not the incident wave:

$$I\lambda_0^2 > 3 \times 10^{18} \frac{k_s c}{\omega_0} \frac{\Delta n}{n} \frac{\omega_{\text{epw}}}{\omega_p} \times \frac{1 + \frac{\nu_{ei}}{\omega_0} \frac{\Delta x}{\lambda_0} \frac{\omega_p}{\omega_s} \frac{\omega_p}{\omega_0} \frac{\omega_0}{k_s c}}{\left(\frac{k_{\text{epw}} c}{\omega_0} \right)^2 \frac{\Delta x}{\lambda_0} \frac{n}{n_c} \arctan \alpha}, \quad (122)$$

where

$$\alpha = \frac{\Delta n/n_c}{4 \left(\frac{\omega_i}{\omega_0} + \frac{\nu_{ei}}{2\omega_0} \right) \frac{\omega_p}{\omega_0}} \quad (123)$$

More precision can be obtained by integrating $n(x)$ and other terms over x . One could approximate the collisional absorption of the incident wave by calculating the inverse bremsstrahlung absorption on the pumping wave between the laser and the position of interest. The far right column of the legend in Fig. 3-22 shows the threshold prediction of Eq. (119) for significant Landau damping; the prediction is in reasonable agreement with the data.

As experiments tend toward higher frequency lasers, the effect of electron-ion collisions can become significant. For example, for $T_e = 3$ keV, $\bar{z} = 50$, $I = 5 \times 10^{14}$, and $\lambda_0 = 0.53 \mu\text{m}$, the ratio of collision frequency to growth rate

$$\frac{\nu_{ei} [1 + 2(\omega_i/\nu_{ei})]^{1/2}}{\gamma_0} \quad (124)$$

is ~ 0.8 at $n/n_c = 0.15$. The collision frequency reduces Raman scattering in three ways:

1. It directly damps the electron plasma waves.
2. It absorbs the scattered light wave.
3. It absorbs the incident laser light directly.

In the heavily damped limit ($\omega_i \gg \kappa' \Delta x v_{\text{epw}}/2$), the Raman gain is^{58,63}

$$\frac{k_{\text{epw}}^2 \Delta x}{P'} \left(\frac{v_{\text{osc}}}{c} \right)^2 \frac{\omega_p}{\omega_i + \nu_{ei}/2} \frac{\omega_p}{\omega_{\text{epw}}} \left(1 - \frac{n}{n_c} \right)^{-1/2}$$

$$\frac{\nu_{ei}}{\omega_0} \left(\frac{\omega_p}{\omega_s} \right)^2 \frac{\Delta x}{\lambda_0} \left(1 - \frac{\omega_p}{\omega_s} \right)^{-1/2}; \quad (125)$$

this includes the effect of ν_{ei} on the scattered wave but not the pump. The scattering can be calculated from the gain by means of Fig. 13 of Ref. 56. A fluid analysis shows the reduced growth rate to be

$$\gamma = - \left[\frac{\omega_i}{2} + \frac{\nu_e}{4} \left(1 + \frac{\omega_p^2}{\omega_s^2} \right) \right]$$

$$+ \left\{ \left[\frac{\omega_i}{2} + \frac{\nu_e}{4} \left(1 - \frac{\omega_p^2}{\omega_s^2} \right) \right]^2 + \gamma_0^2 \right\}^{1/2}. \quad (126)$$

This simulation models electron-ion collisions by randomly rotating the velocity vector (as in Ref. 69); the mapping of three velocity components to two was done by J. Denavit). Figures 3-23 and 3-24 show simulation results for $\nabla n = 0$ and $\neq 0$ respectively. Note that Raman scattering is strongly reduced in Figs. 3-23 and 3-24 for γ_0 a factor of 2 above the threshold

$$\gamma_0^2 > \frac{\nu_{ei}}{2} \left(\frac{\omega_s}{\omega_p} \right)^2 \left(\frac{\nu_{ei}}{2} + \omega_i \right). \quad (127)$$

This threshold was derived for Brillouin scattering by Bodner and Eddleman.⁵⁷ The fact that simulations show a more dramatic effect with ν_{ei} suggests that nonlinear Landau damping may be playing a major role. In vertically paired data points in Fig. 3-23, the lower point corresponds to late-time absorption. The noise level of Fig. 3-23, measured by Fourier analyzing the electromagnetic

Fig. 3-23. Fraction of light absorbed by Raman backscattering in a constant-density plasma for a variety of collision frequencies where ν_{ei} is defined at the n/n_c indicated.

Fig. 3-24. Fraction absorbed by Raman backscattering as a function of laser intensity and wavelength. Equation (118) predicts the threshold for the circles to be $\sim 3 \times 10^{15}$, ν_{ei} uses $n = n_c/4$ for the squares and $n_c/10$ for the circles

noise, was found to be $\sim 5 \times 10^{-5}$ independent of ν_{ei} ; this is the numerical noise inherent in a particle simulation. Preliminary evidence indicates that ν_{ei} also reduces the T_{hot} of Raman backscattering in a density gradient. For $n = n_c/4$, T_{hot} is ~ 120 , 62, and 48 for $\nu_{ei}/\omega_0 = 0$, 0.0025, and 0.005, respectively; for $n = n_c/10$, T_{hot} is ~ 16 , 10, and 9 for $\nu_{ei}/\omega_0 = 0$, 0.001, and 0.002, respectively (data are from the runs of Fig. 3-24).

One can roughly estimate ν_{ei}/γ_0 for the parameter range of interest by doing least-squares fits to LASNEX data for 450- μm , 1-ns disk irradiations with the T_e and $\bar{z}$ of interest at $n_c/4$ at the peak of the laser

Fig. 3-25. Rough scaling of collision frequency/Raman growth rate ν_{ei}/γ_0 vs laser wavelength for a variety of laser intensities for $n = n_c/4$.

pulse: T_e is $\sim 10^{-6} \rho^{0.45} (\omega_0/1.06 \mu\text{m})^{0.85}$, and $\bar{z}$ is $\sim 4.8 I^{-0.67} (\omega_0/1.06 \mu\text{m})^{-0.06}$. Note that T_e and $\bar{z}$ increase with the spot size and pulse length and are sensitive to other geometric factors. Putting these fits into the formulas for the ν_{ei} gives

$$\nu_{ei} \sim 3 \times 10^{12} \left(\frac{I}{10^{15}} \right)^{-1.1} \left(\frac{\lambda_0}{1.06} \right)^{-3.3} \quad (128)$$

Figure 3-25 shows a plot of this simplified scaling.

In conclusion, we have presented kinetic simulations and theory of the threshold of Raman backscattering for a variety of density profiles, background temperatures, and electron-ion collision frequencies. The simulations show Raman backscattering fractions of $\sim 4 \times 10^{-4}$ at intensities ~ 30 times below the ∇n threshold; we suggest that this is due to light scattering from noise electron plasma waves. We find that in many experiments, collisions need to be included in Raman modeling, as shown in Figs. 3-23 and 3-24.

Authors: K. G. Estabrook and W. L. Kruer

Major Contributors: W. Seka, C. J. Randall, B. F. Lasinski, D. W. Phillion, R. P. Drake, R. E. Turner, and E. M. Campbell

The $2\omega_{pe}$ Instability in Current Experiments

The $2\omega_{pe}$ instability, through its nonlinear saturation mechanism, is thought to generate some of the hot and/or superhot

electrons in the underdense corona of laser fusion targets. This instability has been studied theoretically^{70,71} and in computer simulations.⁷²⁻⁷⁶ However, the $2\omega_{pe}$ instability does not have a direct experimental signature, and its presence is usually inferred from $3\omega_0/2$ or $\omega_0/2$ emission.

As part of our series of articles⁷³⁻⁷⁶ on this instability, we review here the evidence for the $2\omega_{pe}$ instability in current experiments at incident wavelengths of $1 \mu\text{m}$ or less. While we find such evidence, it is difficult to obtain reliable values characterizing the plasma condition; we therefore cannot tell whether the threshold for the instability was exceeded in many of these experiments, or by how much.

The $2\omega_{pe}$ instability is the decay of an incident electromagnetic wave into two electron plasma waves. The frequency-matching conditions limit this instability to the quarter-critical density region, where critical density (n_c) is the plasma density at which the plasma frequency (ω_{pe}) equals the frequency (ω_0) of the incident electromagnetic wave, whose wavelength is λ_0 . The indirect observations ($3\omega_0/2$, $\omega_0/2$, and hot electrons) that usually establish the presence of this instability result from the subsequent interaction of the decay plasma waves; these observations do not necessarily yield information about the original decay-wave spectrum. Further, it is often difficult to disentangle the hot-electron component due to the $2\omega_{pe}$ instability from the complete hot-electron spectrum.

Typical $2\omega_{pe}$ threshold estimates suggest that this instability may have occurred in many recent experiments. For example, Fig. 1 of Ref. 77 shows Shiva and Argus data with respect to the parametric instability thresholds. Much of the Shiva and Argus parameter regime is apparently above the $2\omega_{pe}$ instability threshold.

This threshold is set by density gradient stabilization for the current experimental parameter regime and is given by

$$k_0 L \frac{v_0^2}{v_e^2} > 16 \frac{\sqrt{3}}{2}$$

or

$$I > \frac{5 \times 10^{15} T_e}{\Lambda n} \quad (129)$$

Here k_0 is the wave vector of the incident light, L is the local gradient scalelength, defined by $L^{-1} = (dn/dx)/(4/n_c)$, v_0 is the electron quiver velocity in the electric field of the laser, and v_e is the electron thermal velocity. Then

$$\frac{v_0^2}{v_e^2} = \frac{0.04I\lambda^2}{T_e}, \quad (130)$$

where I is the intensity of the incident laser light in units of 10^{14} W/cm², L and λ are in microns, and T_e is the electron thermal temperature in keV.

The $2\omega_{pe}$ instability gradient threshold given in Eq. (129) comes from Ref. 73, and is an adaptation of the results of Ref. 70. This expression agrees to better than a factor of 2 with the later results of Ref. 72. Thus there is no theoretical disagreement on the $2\omega_{pe}$ instability gradient threshold.

The lack of information on the plasma conditions at quarter-critical density is the main uncertainty in applying the theoretical results to experiments. Neither the gradient scalelength nor the electron thermal temperature are routinely determined experimentally. Few experiments monitor the half-harmonic optical emissions and determine the underdense plasma conditions.

In practice, phenomenological estimates are often used to evaluate threshold expressions. The background plasma temperature is estimated from flux-balance arguments in which the absorbed laser light is carried away by heat conduction. Thus T_e is determined from $f_a I = f_e n_e m_e v_e^3$, where f_a is the absorption fraction and f_e is the flux limit. Then an isothermal expansion is assumed, and the density-gradient scalelength is estimated from $L = c_s t/2$, where c_s is the sound speed and t is the laser pulse length. This type of phenomenological estimate was used to develop Fig. 1 of Ref. 77.

Hydrocode calculations may in principle be used to evaluate plasma conditions in the neighborhood of quarter-critical density. Typical density and temperature profiles for gold disk irradiations are shown in Fig. 2 of Ref. 78. Although the temperature profile is flat throughout the corona, the density profile is not very well resolved. It is uncertain whether quarter-critical density lies in the rather steep region below the critical density or in the relatively flat shelf at $\sim 0.1n_c$.

The usual phenomenological estimates agree with the hydrocode results for the region at $0.1n_c$. But the ambiguity about the quarter-critical-density scalelength may result in threshold uncertainties by as much as a factor of 10. Thus, many Argus and Shiva experiments that appear to be above the $2\omega_{pe}$ density-gradient threshold may, in fact, not satisfy the threshold conditions. These concerns are consistent with $3\omega_p/2$ emission at the 10^{-5} level for gold-disk irradiations at $\sim 10^{15}$ W/cm². (By "level" we mean the intensity of the emission as a fraction of that of the incident wave.)

Another factor-of-2 uncertainty in the threshold evaluation may come from determining the corona temperature: comparison with experimental results (such as frequency shifts of half-harmonic emissions) suggests that the hydrocode result may be high by about this factor. As we move to shorter incident wavelengths, the uncertainty in the corona temperature becomes more important than the scalelength determination, because collisional damping is then setting the intensity threshold. For example, for an $0.35\text{-}\mu\text{m}$, 3×10^{14} W/cm² irradiation of a gold disk, the estimated plasma temperature is 2 keV and the $2\omega_{pe}$ instability is collisionally damped.

If the background plasma temperature becomes too high, the $2\omega_{pe}$ instability becomes kinetic. This change in character for $T_e \geq 7$ keV was seen in our ZOHAR simulations, reported in Ref. 76. The kinetic $2\omega_{pe}$ instability has since then also been studied theoretically.⁷⁹ These considerations are expected to be relevant to experiments with wavelengths of $1\text{ }\mu\text{m}$ or longer.

The early Novette data from gold-disk irradiations confirm the Argus results. At incident intensities $\leq 10^{15}$ W/cm², the $3\omega_p/2$ emission levels are less than 10^{-4} . Here, too, simple estimates suggest that the incident intensity is several times above the $2\omega_{pe}$ instability density-gradient threshold. But again, LASNEX calculations indicate that these simple estimates give too low a threshold, since quarter-critical may be in a steeper part of the density profile. The measured T_{hot} is ~ 30 keV, which is more indicative of phase velocities of plasma decay waves at $0.1n_c$. Hot electrons generated by plasma waves at quarter-critical density are expected to be ≥ 40 keV, on the basis of the

($I\lambda^{2.0-3.1}$) scaling of T_{hot} observed in the simulations of the $2\omega_{\text{pe}}$ instability in this parameter regime.⁷³ Thus a consistent assumption is that there is at most a low level of the $2\omega_{\text{pe}}$ instability occurring in the Novette gold-disk irradiations.

There have, however, been experiments in which the $2\omega_{\text{pe}}$ instability appears to have occurred at significant levels. In these, the experimental parameter regime is usually a factor of a few above a typical threshold estimate. Then the $3\omega_{\text{q}}/2$ emission level increases to 10^{-4} .

One such example is the beryllium-disk irradiations reported in Ref. 80. The incident $1.06\text{-}\mu\text{m}$ intensity was $3 \times 10^{14}\text{ W/cm}^2$. Hydrocode calculations suggest a scalelength of $130\text{ }\mu\text{m}$ and a temperature of 2 keV near quarter-critical. For these conditions, the incident intensity is a factor of 4 above the gradient-stabilization threshold given by Eq. (129). The experimental results include: $3\omega_{\text{q}}/2$ level of $\sim 10^{-4}$, $f_{\text{hot}} \sim 1\%$, and $T_{\text{hot}} \approx 10\text{ keV}$. No Raman emission was detected from the $0.1n_c$ region, and the calculated conditions are below the Raman convective threshold. There was a weak signal at $\omega_{\text{q}}/2$; such signals are commonly ascribed to the $2\omega_{\text{pe}}$ instability. The estimated T_{hot} for the $2\omega_{\text{pe}}$ instability is $\sim 20\text{ keV}$, a factor of 2 higher than the measured value. Nonetheless, threshold estimates and the detected optical signatures suggest that the $2\omega_{\text{pe}}$ instability is generating the hot electrons. Similar results on the $2\omega_{\text{pe}}$ instability for CH-foil irradiations were reported in Ref. 81.

A correlation of $3\omega_{\text{q}}/2$ emission and hot-electron production was seen in exploding-foil irradiations on Shiva.⁸² The calculated conditions ($T_e = 3\text{ keV}$, scalelength near quarter-critical of $200\text{ }\mu\text{m}$) indicate that the laser intensity of $2 \times 10^{15}\text{ W/cm}^2$ was at least a factor of 10 above the $2\omega_{\text{pe}}$ gradient-stabilization threshold. Experimentally, the $3\omega_{\text{q}}/2$ level was $\sim 4 \times 10^{-4}$. But early Novette results on thin gold foils suggest a correlation between Raman scattering and hot-electron production.

Conditions assumed in the ZOHAR simulations of the quarter-critical density region correspond to those in the exploding-foil experiments, since the critical surface is absent in both. The simulation conditions

($3 \times 10^{15}\text{ W/cm}^2$ incident intensity, $1.06\text{-}\mu\text{m}$ light, $T_e = 3\text{ keV}$, and density profile from 0.195 to $0.235n_c$, corresponding to a scalelength of $150\text{ }\mu\text{m}$) are similar to those of the Shiva exploding-foil experiment. The $2\omega_{\text{pe}}$ instability was more robust in the simulations: the instability absorption was 35% , T_{hot} was 90 keV , and $3\omega_{\text{q}}/2$ emission in the forward and backward directions was $\sim 10^{-4}$ and $\sim 10^{-4}$ respectively. Note, however, that in the experiment the quarter-critical surface was present for only a fraction of the incident laser pulse, so that the agreement between experiment and simulation may in fact be better than the numbers indicate.

The $3\omega_{\text{q}}/2$ and $\omega_{\text{q}}/2$ frequency spectra have been measured in many experiments. Discrepancies between these spectra and those implied by the linear theory of the $2\omega_{\text{pe}}$ instability require that we invoke non-linear mechanisms. As detailed in Ref. 83, for $0.53\text{-}\mu\text{m}$, $5 \times 10^{15}\text{ W/cm}^2$ irradiations of gold disks, the $\omega_{\text{q}}/2$ emission spectrum corresponds to plasma waves with wavelengths shorter than the incident wavelength. Linear theory predicts decay waves of longer wavelength (shorter k).

The fact that some fraction of the $3\omega_{\text{q}}/2$ emission is produced out of the plane of polarization also illustrates the complexity of the half-harmonic production mechanisms. Since the $2\omega_{\text{pe}}$ decay waves necessarily lie in the plane of polarization, a possible $3\omega_{\text{q}}/2$ production mechanism may involve combining a $2\omega_{\text{pe}}$ decay wave with an out-of-plane Brillouin-sidescattered wave. Alternative explanations that involve rescattering of the constituent waves are also possible.

From the definitive CO_2 laser experiments on the $2\omega_{\text{pe}}$ instability,⁸⁴ it is clear that for the conditions of that experiment, the $2\omega_{\text{pe}}$ decay wave spectrum does not correspond to the spectrum associated with the half-harmonic spectrum and the detected hot electrons. In these CO_2 laser experiments, the $2\omega_{\text{pe}}$ decay waves have been observed directly by Thomson scattering. The results correspond to linear theory. But the decay wavelengths are too short to permit k -matching for a $3\omega_{\text{q}}/2$ electromagnetic wave made from a decay plasma wave at quarter-critical and an electromagnetic wave

at ω_0 . Also, the observed T_{hot} is associated with waves whose phase velocity is greater than that of the $2\omega_{pe}$ decay waves.

In summary, threshold estimates for the $2\omega_{pe}$ instability have many uncertainties because of the lack of experimental information on plasma conditions at quarter-critical density. For experiments with incident wavelength $\leq 1 \mu\text{m}$, there is reasonably good evidence for the presence of the $2\omega_{pe}$ instability for some low-Z targets. Only for CO_2 laser experiments, where the decay plasma waves have been observed directly, is it possible to relate the linear theory to the usual (indirect) signatures of the $2\omega_{pe}$ instability.

Author: B. F. Lasinski

Major Contributors: A. B. Langdon, W. J. Krue, W. C. Mead, R. E. Turner, E. M. Campbell, R. P. Drake, and D. W. Phillion

Stabilization of Filamentation by Angular Spread

Motivated by recently proposed illumination schemes, we consider in this article filamentation by the ponderomotive force mechanism for the situation in which the light consists not of a single plane wave but of many nonparallel, overlapping beamlets. Such illumination is a feature of experiments at Rochester,⁸⁵ the Osaka random phase shifter (RPS),⁸⁶ and especially of the Naval Research Laboratory's "induced spatial incoherence" (ISI) proposal.⁸⁷ The analysis shows that filamentation is stabilized (that is, filamentation is eliminated) by illumination whose mean-square angular spread $\bar{\theta}^2$ is large enough:

$$\bar{\theta}^2 \geq \frac{n}{8n_c} \frac{\bar{v}_0^2}{v_e^2}, \quad (131)$$

where $n/n_c = \omega_p^2/\omega_0^2$ is the ratio of the local electron density to the critical density, $v_e^2 = T_e/m_e$, and where $\bar{v}_0^2$ represents the electron's oscillatory velocity corresponding to the sum of intensities, i.e., $\bar{v}_0^2/v_e^2 = 4 \times 10^{-16} \Sigma I \lambda^2/T_e$ (W/cm^2 , μm , keV).

Why do oblique rays tend to stabilize filamentation? Consider a beam of light

whose propagation vector $\mathbf{k}_0$ makes a sufficient angle to the filament. The magnitude of the electric field in the low-density plasma inside the filament is less than outside (just as the electric field increases as a wave approaches critical density). This is the opposite of the situation for a ray of light with $\mathbf{k}_0$ almost parallel to the filament, which is intensified by focusing. It is clear that the oblique rays of the diffuse light not only do not contribute to the self-focusing mechanism, but that in fact they oppose it. Hence they can provide stabilization, not just a reduced growth rate. The same argument holds for filamentation by the thermal mechanism.

The analysis here begins by extending the dispersion relation^{88,91} that describes filamentation (and Brillouin scattering). The overlapping beamlets, assumed to be mutually incoherent as in the ISI proposal, are characterized by their propagation vectors $\mathbf{k}_0$ and vector amplitudes $\mathbf{v}_0 = e\mathbf{E}_0/\omega m_e$. The dispersion relation becomes

$$\frac{\omega^2 - k^2 c_s^2}{\omega_{pi}^2} = \frac{1}{4} \sum \left[\frac{(\hat{\mathbf{k}} \times \mathbf{v}_0)^2}{\omega^2 - \omega_{pe}^2 - (ck_i)^2} + \frac{(\hat{\mathbf{k}}_i \times \mathbf{v}_0)^2}{\omega^2 - \omega_{pe}^2 - (ck_i)^2} \right], \quad (132)$$

where the sum is over beamlets, $(\omega, \mathbf{k})$ are the frequency and wave vector of the density perturbation, $\omega_{\pm} \equiv \omega \pm \omega_0$, $\mathbf{k}_{\pm} \equiv \mathbf{k} \pm \mathbf{k}_0$, $\hat{\mathbf{k}}_i \equiv \mathbf{k}_i/k_i$, and $c_s \equiv \omega_{pi}/k_{D0}$ is the isothermal sound speed. With $\omega^2 \ll \omega_0^2$ and $k^2 \ll k_{D0}^2$, Eq. (132) becomes

$$-\frac{\omega^2 - k^2 c_s^2}{\omega_{pi}^2} = \frac{1}{2} \sum \frac{v_0^2}{c^2} \times \frac{1}{1 - 4(\mathbf{k} \cdot \mathbf{k}_0 - \omega\omega_0/c^2)^2/k^4}. \quad (133)$$

Let $\mathbf{k}_0 \approx k_0(\hat{\mathbf{z}} + \theta\hat{\mathbf{x}} + \theta_y\hat{\mathbf{y}})$ and $\mathbf{k} = k\hat{\mathbf{y}}$, so that $\mathbf{k} \cdot \mathbf{k}_0 = \theta_y k_0 k$. Light beamlets, whose $k_{0y} = \theta_y k_0$ exceeds half the k -value associated with filamentation, act to stabilize

filamentation, because such beamlets contribute negatively to the right-hand side of Eq. (133). This observation is sufficient to allow us to make the heuristic estimate below of the new threshold.

The filamentation threshold may be estimated as follows. For a spread $\bar{\theta}^2$ of directions of $\mathbf{k}_0$, filamentation requires $(k/k_0)^2 \geq 2\bar{\theta}^2$ to get enough flux contributing. For this k , filamentation requires $(k\lambda_{De})^2 \leq \bar{v}_0^2/4c^2$. Eliminating k then leads to Eq. (131).

Quantitative results can be obtained for the special case of a Gaussian spread of angles. For an angular spread given by

$$\frac{v_0^2}{c^2} (\theta_x, \theta_y) = \frac{\bar{v}_0^2}{c^2} \frac{1}{2\pi\bar{\theta}^2} \exp\left[-\frac{\theta_x^2 + \theta_y^2}{2\bar{\theta}^2}\right], \quad (134)$$

the dispersion relation becomes

$$\begin{aligned} -\frac{\omega^2 - k^2 c_s^2}{\omega_{pi}^2} &= \frac{1}{4} \frac{\bar{v}_0^2}{c^2} \frac{k}{2(2\bar{\theta}^2)^{1/2} k_0} \\ &\times \left[-Z\left(\frac{2\omega\omega_0/c^2 k + k}{2(2\bar{\theta}^2)^{1/2} k_0}\right) \right. \\ &\quad \left. + Z\left(\frac{2\omega\omega_0/c^2 k - k}{2(2\bar{\theta}^2)^{1/2} k_0}\right) \right], \quad (135) \end{aligned}$$

where

$$Z(\xi) \equiv \frac{1}{\sqrt{\pi}} \int_{-\infty}^{\infty} \frac{e^{-x^2}}{x - \xi} dx \quad (136)$$

is the Fried-Conte plasma dispersion function. For large $\bar{\theta}^2$,

$$-\frac{\omega^2 - k^2 c_s^2}{\omega_{pi}^2} = \frac{1}{2} \frac{\bar{v}_0^2}{c^2} \frac{k^2}{4k_0^2 \bar{\theta}^2}; \quad (137)$$

that is, the effective intensity is reduced by the factor $k^2/4k_0^2 \bar{\theta}^2$.

The threshold condition for purely growing instability is obtained by setting $\omega = 0$ in Eq. (135):

$$\begin{aligned} (k\lambda_{De})^2 &< \frac{1}{2} \frac{\bar{v}_0^2}{c^2} \\ &\times \operatorname{Re} \left[-\frac{k}{2(2\bar{\theta}^2)^{1/2} k_0} Z\left(\frac{k}{2(2\bar{\theta}^2)^{1/2} k_0}\right) \right]. \quad (138) \end{aligned}$$

For small $\bar{\theta}^2 \operatorname{Re}[\dots] \rightarrow 1$, and we recover the usual threshold, $(k\lambda_{De})^2 < \bar{v}_0^2/2c^2$, applicable to well-collimated light. For general $\bar{\theta}^2$, using $\operatorname{Re}[\dots] \leq k^2/4\bar{\theta}^2 k_0^2$, we obtain Eq. (131).

The case of finite angular spacing of a small number of beamlets, as at Rochester and Osaka, is more complicated. Preliminary results indicate a reduction in growth rate, but not complete stabilization. The maximum growth rate decreases as the spacing Δk_0 of the $\mathbf{k}_0$ is decreased.

Discrete beamlets produce multiple resonances in the dispersion relation. These resonances support Brillouin *forward* scattering if the Doppler-shifted beam frequencies are spaced closer than the sound frequency.

Angular spread may also partially suppress Brillouin backscattering. It has been asserted, without demonstration,⁹¹ that the gain length is increased when $\bar{\theta}^2 \geq v_0/v_e$. (Here $\bar{\theta}$ refers to the angle of convergence of light from a lens of small f/λ number.) This condition is more stringent than that for filamentation, but both conditions can be satisfied when $v_0/v_e < 1$.

Experimental realization may be a side effect of NRL's ISI technique⁸⁷ for improving the uniformity of illumination of direct-drive targets. In their proposal, spatial incoherence induced by an echelon smooths the focal-spot irradiance of a broadband laser. In addition, the mutually incoherent beamlets have an angular spread $\bar{\theta}^2 \approx (2f)^{-2/2}$.

Author: A. B. Langdon

2-D LASNEX Ray-Trace Calculations of Thermal Whole-Beam Self Focusing

Self focusing⁹²⁻¹⁰⁴ increases the intensity of laser light. It can thereby increase absorption by plasma waves, producing hot electrons and causing preheat; it might also be a source of perturbations that could grow by the Rayleigh-Taylor instability. In this article, we explore the parameter space of laser wavelength, spot size, and intensity, plasma temperature, density, and species (hydrogen and gold) for diverging and parallel rays, with some comparison to theory.

Thermal self focusing occurs when laser light heats a region of plasma, causing it to expand and decreasing its density. The index of refraction of light in a plasma is given by $[1 - (\omega_p/\omega_0)^2]^{1/2} \approx [1 - (n/n_c)]^{1/2}$, so the light bends towards regions of decreasing density; this further heats the plasma, further decreasing the density, and so on. [Here ω_p is the plasma frequency, $\omega_p = (4\pi ne^2/m_e)^{1/2}$, ω_0 is the laser light frequency, n is the plasma density, and n_c is the critical plasma density, at which $\omega_0 = \omega_p$.]

If the initial temperature is high enough, however, thermal transport will diffuse the heat away fast enough to dissipate the process. Because thermal diffusion varies as $T_e^{3/2}/Z$, and inverse bremsstrahlung absorption varies as $Zn(n/n_c)/T_e^{3/2}$, self focusing is sensitive to temperature, Z , and n . This process is quantified in several papers.⁹⁴⁻⁹⁹

Other, related forms of self focusing are ponderomotive and relativistic. Ponderomotive self focusing¹⁰⁰⁻¹⁰² arises because the ponderomotive pressure $E_0^2/8\pi$ of the light can accelerate the plasma away from the light. Relativistic self focusing¹⁰³ can occur without ion motion in a homogeneous plasma. It occurs if the peak oscillatory velocity of the electron in the transverse laser electric field, given by $v_{osc} = eE_0/m_e\omega_0 = c(\lambda_0^2/1.4 \times 10^{16})^{1/2}$, approaches the speed of light, c ; ω_p then decreases (because of the relativistic increase in the electron mass), and the local index of refraction accordingly decreases. (Here E_0 is the electric field amplitude of the laser light, I is the intensity of the light in W/cm^2 , and λ_0 is its wavelength in μm .)

For low v_{osc} , thermal self focusing is generally the most important process. All these self-focusing mechanisms are exacerbated by so-called hot spots in the laser source¹⁰⁴ and by inhomogeneities in the plasma. An independent mechanism for concentrating the laser light can arise in special target and focusing geometries that cause the light to be refracted to a caustic focus. Such a situation arises, for example, in the funnel formed by a burn-through foil.

Inverse bremsstrahlung is the dominant absorption mechanism. Collective absorption processes such as Raman scattering and $2\omega_{pe}$ may be significant, and they can be enhanced by the higher intensities

in a filament. Brillouin backscattering and sidescattering may be important competing processes.¹⁰²

We used the LASNEX code to model the dynamics of thermal self focusing of laser light in hydrogen and gold plasmas initially at constant density and temperature. LASNEX, a 2-D cylindrical hydrodynamics simulation modeled the laser light as a number of 2-D rays, refracted according to geometrical optics and absorbed by inverse bremsstrahlung.¹⁰⁵

We idealize the geometry and modeling so as to isolate thermal self focusing: The plasma has constant initial density and temperatures, and fixed boundary conditions on the left, right and top; $r = 0$ is the center line. The simulation includes a plasma 1 cm in radius so that ion waves will not be reflected from the top. Internal plasma motion and electron heat conduction were allowed; radiation transport was only allowed in one case.

In most runs, we used $I = 3 \times 10^{15} \text{ W}/\text{cm}^2$ and $\lambda_0 = 0.266 \mu\text{m}$. In some cases, initially parallel light self focused to $\sim 10\times$ the incident intensity in 0.5 to 2 ns (the exact value of the multiplication factor is uncertain because of the resolution of the simulation). A diverging lens (such as would be used in currently planned experiments) increased the self-focusing length or prevented focusing altogether.

We define the "whole-beam" self-focusing length as the distance from the right wall ($z = 1 \text{ cm}$) to the first point at which most of the rays converge. Similarly, the "filamentation" length is the distance for a few rays to converge locally. Often there is no intensity multiplication, because most of the light is absorbed before it can come to a focus. In current reactor-target designs, the laser light will have to pass through a maximum effective plasma length of $\sim 1 \text{ cm}$. Thus, if the self-focusing length is greater than 1 cm (or if absorption is strong enough that no intensity multiplication occurs), these targets will be unaffected by self focusing. As the plasma heats up, the distance to focus for both whole-beam and filamentation self focusing increases in time. Essentially the laser drills a hole in the plasma, both by moving plasma aside and by bleaching (increasing T_e and thereby decreasing inverse bremsstrahlung absorption).

Figure 3-26 shows the paths of initially parallel rays (corresponding to infinite f/λ number), in a snapshot at $t = 3$ ns. Initial values were $T_e = 1$ keV, $n/n_c = 0.09$, $z = 68$, $I = 3 \times 10^{15}$ W/cm², $\lambda_0 = 0.266$ μ m. The laser spot diameter was 0.5 cm; radiation transport was not included. (The horizontal scale is 10 times the vertical scale.) There is no point focus, but rather a line focus (at $z \sim 0.35$ cm) whose position can oscillate in z . There are small filaments at $r = 0.002$ and 0.005 , $z \sim 0.25$.

Figure 3-27 shows the self-focusing length for parallel light in a gold plasma as

a function of n/n_c and I , for $\lambda_0 = 0.266$ μ m and a spot diameter of 0.5 cm. The transport multiplier was $f = 0.03$. For simplicity, we took $\bar{z}$ as the value in Saha equilibrium. In most runs with the equilibrium temperature at these intensities (~ 5 keV), there was no whole-beam or filamentation self focusing. In one case, for $n/n_c = 0.045$, $T_e = 5$, and $I = 10^{16}$, the self-focusing length was < 1 cm. A rerun with diverging $f/4$ optics showed no self focusing. A run with $f = 0.6$ (the default value) showed less pronounced filamentation, but the whole-beam self-focusing length was only slightly longer.

Figures 3-28 and 3-29 show the self-focusing length for parallel light in hydrogen plasmas as a function of T_e , λ_0 , spot diameter, and n/n_c . All points in Fig. 3-28(a) had $T_e = T_i$; in Fig. 3-28(b), all points had $T_e/T_i = 30$.

In actual experiments, both the plasma and the laser are much noisier than the simulation. We simulated laser noise as a hot spot of intensity $3 \times$ the background intensity; the focusing length was about the same as for the hot spot alone. We have also found that a ring beam (one with intensity peaks away from $r = 0$) will self focus.

We calculate the quasi-equilibrium temperature, T_e , inside the filament and use that temperature in Lehmberg's theory^{94,95} to find the filamentation length for comparison with the results of our simulations. In order to determine T_e , we balance the inverse bremsstrahlung

Fig. 3-26. Ray trace of thermal self-focusing. The temperature is unrealistically low: for the expected $T_e \sim 3$ to 5 keV, the light would be absorbed before it came to a focus.

Fig. 3-27. Whole beam self-focusing length vs (a) n/n_c and (b) laser intensity I . No focus was found for intensities $\lesssim 3 \times 10^{14}$.

absorption inside the laser spot, whose FWHM diameter is D , and use the heat transport equation to take the heat radially away from T_e to T_{cold} (given at large radius), with the result

$$T_e \text{ (keV)} = \left[\frac{T_e}{T_e - T_{\text{cold}}} \frac{1}{f} \frac{1}{\delta(z)} \frac{1/3.5 \times 10^{14}}{(1 - n/n_c)^{1/2}} \right]^{1/5} \times \left[\frac{D^2 \ln \Lambda (n/n_c)}{\lambda_0} \right]^{2/5}, \quad (139)$$

where f is the transport multiplier, δ and Λ are from Spitzer,¹⁰⁶ and D is the full width at half maximum (FWHM) diameter of the laser spot.

Lehmberg's equations⁹⁴ are

$$\alpha = \left[\frac{\alpha_{\text{pm}} k_{\perp}^2}{k_0} + \alpha_{\text{tm}}^2 - \left(\frac{k_{\perp}^2}{2k_0} \right)^2 \right]^{1/2}, \quad (140)$$

where α_{tm} is the inverse thermal filamentation length,

Fig. 3-28. Whole beam self-focusing length vs (a) T_e and (b) laser wavelength.

Fig. 3-29. Self-focusing length in hydrogen vs (a) laser spot diameter and (b) n/n_c .

Fig. 3-30. Self-focusing length vs laser intensity for diverging $f/30$ optics (as in a projected reactor).

$$\alpha_{tm} = 4.2 \times 10^{-7} \frac{\bar{z} \ln \Lambda}{\lambda_0} \left[\frac{(n/n_c)^{3/2}}{\bar{z}_f \epsilon_0^{3/2} T_e^5} \right]^{1/2}, \quad (141)$$

α_{pm} is the inverse ponderomotive filamentation length,

$$\alpha_{pm} = \frac{2.9 \times 10^{-12} (n/n_c) \lambda_0}{\bar{z}_f \epsilon_0 T_e}, \quad (142)$$

$$\epsilon_0 = 1 - (n/n_c), \quad k_0 = 2\pi \epsilon_0^{1/2} / \lambda_{0cm}, \quad k_{\perp} = 2\pi/D, \quad \bar{z}_f = 1 + (1/\bar{z}), \quad \text{and}$$

$$\phi = \frac{\bar{z} + 0.24}{1 + 0.24\bar{z}}. \quad (143)$$

The theoretical filamentation lengths, α^{-1} , are shown in Figs. 3-27 to 3-29. The theory assumes that the light rays are initially parallel (lens f/∞), as was assumed for the simulation data shown in these figures. The theory is Cartesian; this is a limitation, because the simulation and the experiment are cylindrical. The simulation lengths are for whole-beam self focusing; these are comparable to the theoretical small-scale filamentation lengths α^{-1} , but they correspond to a somewhat different phenomenon.

Figure 3-30 shows self-focusing lengths in a gold plasma for 0.266- μm light for a reactor target with diverging $f/30$ beams, with radiation transport included. The electron temperature was $T_e = 3 \text{ keV}$; spot diameter = 0.05 cm. Present reactor-target designs have intensities below 3×10^{15} , which do not show intensity multiplication in Fig. 3-30. Higher intensities could lead to self-focusing problems.

Conclusion. We have studied the parameter space of thermal self focusing with respect to laser wavelength, intensity, and

spot diameter, plasma density, temperature and species, and parallel and divergent optics with some comparison to theory. Work is continuing on enlarging the parameter space because these effects may not be negligible.

Authors: K. G. Estabrook, W. L. Kruer, and D. S. Bailey

Beat-Wave Heating in Inhomogeneous Plasma Irradiated with 0.53- and 1.06- μm Light

The output of frequency-doubling crystals includes light at both ω_0 and $2\omega_0$, where ω_0 is the original frequency. If both 1.06- μm and 0.53- μm light impinge on a target, the light waves can beat to resonantly excite a plasma wave at points in the plasma where the electron density is one-fourth the critical density for the 0.53- μm light. Here we present computer simulations and a simple model for this beat-wave heating in an inhomogeneous, expanding plasma. The model allows for a local steepening of the density profile and provides estimates for the absorption. We find that steepening by the plasma wave greatly reduces the absorption; it is hard to get even 10% absorption.

Plasma heating and electron acceleration by the beat between two electromagnetic waves have been considered by many workers.¹⁰⁷⁻¹¹² An application of the phenomenon that is receiving considerable attention is the plasma accelerator. Very short laser pulses are envisioned in the accelerator, so that changes in plasma density do not have time to develop and spoil the resonant acceleration of electrons.

Our work is complementary to that being done on the plasma accelerator in that we explore beat-wave heating in a self-consistently steepened density profile.

A Simple Model of Beat-Wave Heating. We consider two electromagnetic waves, with electric-field amplitudes E_1 and E_2 , frequencies ω_1 and ω_2 , and wave numbers k_1 and k_2 , propagating in a planar plasma with density $n_0(x)$. If we treat the electrons as a fluid, the force equation gives for the longitudinal component of the electron fluid

velocity, u_L , the differential equation

$$\frac{d\mathbf{u}_L}{dt} = -\frac{e\mathbf{E}}{m} - \nabla(\mathbf{v}_1 \cdot \mathbf{v}_2) - \nu_e \mathbf{u}_L \quad (144)$$

Here $\mathbf{E}$ is the electrostatic field, $\mathbf{v}_{1,2} = e\mathbf{E}_{1,2}/m\omega$, and ν_e is the damping rate. The term $\nabla(\mathbf{v}_1 \cdot \mathbf{v}_2)$ expresses the ponderomotive force due to the beat between the two electromagnetic waves, which can resonantly drive an electron plasma wave. Using Eq. (1) in the continuity and Poisson equations, we obtain

$$\left(\frac{\partial^2}{\partial t^2} + \nu_e \frac{\partial}{\partial t} + \omega_{pe}^2 \right) \mathbf{E} = -4\pi en_0 \nabla(\mathbf{v}_1 \cdot \mathbf{v}_2) \quad (145)$$

Equation (145) shows that an electron plasma wave will be resonantly driven if $|\omega_1 - \omega_2| \approx \omega_{pe}$.

To compute the energy transfer, we approximate the driving field by its value at the resonance point. Noting that at the resonance point we have $\omega_{pe} = \omega_1 = \omega_2 - \omega_1$, we then have

$$\frac{\partial}{\partial x}(\mathbf{v}_1 \cdot \mathbf{v}_2) \approx \frac{k_2 v_1 v_2}{2} \sin(\omega_1 t - k_2 x), \quad (146)$$

where v_1 is treated as a standing wave at its cutoff. The steady-state solution for E is

$$|E|^2 = \beta^2 \left[\left(1 - \frac{\omega_{pe}^2}{\omega_1^2} \right)^2 + \left(\frac{\nu_e}{\omega_1} \right)^2 \right]^{-1}, \quad (147)$$

where $\beta = 4\pi en_0 k_2 v_1 v_2 / 2\omega_1^2$. The absorbed energy flux I_{abs} is

$$I_{\text{abs}} = \int_0^\infty \nu_e \frac{|E|^2}{8\pi} dx = \frac{\omega_1 L}{8} \beta^2 \quad (148)$$

Here the density has been modeled as a linear variation through the resonance point with local density scalelength L . Substituting for β then gives

$$I_{\text{abs}} \approx \frac{\pi}{4} \frac{\omega_1 L}{c} \left(\frac{k_2 c}{2\omega_1} \right)^2 \left(\frac{v_1}{c} \right)^2 \frac{c E_0^2}{8\pi} \quad (149)$$

Note that I_{abs} is proportional to E_0^2 , Σ_2^2 , and L .

Beat-wave heating can be strongly affected by local steepening of the density profile, either by the resonantly driven electrostatic wave or by the reflecting light wave (when $\omega_2 = 2\omega_1$). To illustrate this effect crudely, we estimate the scalelength by

$$\frac{1}{L} = \frac{1}{L_{\text{em}}} + \frac{1}{L_{\text{es}}} \quad (150)$$

where L_{em} is the steepened scalelength due to the reflecting light wave with frequency ω_1 , and L_{es} is the steepened scalelength due to the resonantly generated plasma wave. Each of these scalelengths can be estimated from earlier work. In particular, from Ref. 113, we have $\omega_1 L_{\text{em}}/c \sim 2v_e/v_{\text{IFs}}$ for $0.03 < v_{\text{IFs}}/v_e < 1$, where v_{IFs} is v_1 evaluated using the free-space value of E_1 , and where v_e is the electron thermal velocity. Likewise,¹¹⁴

$$\frac{\omega_1 L_{\text{es}}}{v_e} \sim 5 \left(\frac{v_e}{v_d} \right)^{0.75} \quad \text{for } 0.03 < \frac{v_d}{v_e} < 1 \quad (151)$$

where $v_d \approx k_2 v_1 v_2 / 2\omega_1$.

Computer Simulations. To explore beat-wave heating more quantitatively, we carried out computer simulations using the 1.5-D relativistic electromagnetic particle code OREMP. In these simulations, both 1.06- and 0.53- μm light were simultaneously propagated into an inhomogeneous slab of plasma. Reflected and/or transmitted light waves were allowed to propagate freely out of the system. Particles expanded into a vacuum on the low-density side and were re-emitted, with their initial temperature, on the high-density side.

As an illustrative example, we consider a case in which the intensity of the 0.53- μm light is $4 \times 10^{16} \text{ W/cm}^2$, the intensity of the 1.06- μm light is $1.33 \times 10^{16} \text{ W/cm}^2$; the plasma, with an initial electron temperature of 4 keV, becomes overdense to both light waves. Figure 3-31 shows the results at a time $\omega_0 t / 2\pi = 1500$.

The plot of electron phase space in Fig. 3-31(a) shows (at the right) localized

Fig. 3-31. Results of a computer simulation of beat-wave heating, at time $\omega_0 t/2\pi = 1500$: (a) Electron phase space; (b) ion-density profile; and (c) electron distribution function.

heating due to the beat-wave-driven plasma wave. Figure 3-31(b) shows the expected energetic electron tail, which has a quasi-Maxwellian velocity distribution with a temperature of about 26 keV. Figure 3-31(c) shows the local steepening of the ion-density profile at the critical densities of the 0.53- and 1.06- μm light. This steepening reduces absorption: in this example, only about 8% of the total energy flux was absorbed via the beat-wave heating. These general features of the beat-wave heating compare favorably with the model estimates. For example, Eqs. (148) and (149) predict a net absorption of about 6%.

In another simulation, the intensities of both incident beams were reduced by a factor of 5. This time the plasma, with an initial electron temperature of 4 keV, remained underdense to the 0.53- μm light. The beat-wave heating in the locally steepened density profile was rather weak. The net absorption was about 1.5% (the model predicts 2.5%).

Summary. We have presented computer simulations and a simple model for beat-wave heating in an inhomogeneous, expanding plasma irradiated with 1.06- and 0.53- μm laser light. The heating is proportional to the local density scalelength and is strongly reduced by profile steepening due to both the reflecting light wave and the resonantly driven plasma wave.

Authors: W. L. Kruer and K. G. Estabrook

Reheating of Hot Electrons by Laser-Excited Plasma Waves

The total number and the energy spectrum of hot electrons in laser fusion targets are of

interest because of the considerable effects such particles can have on target implosions. In particular, preheating of the DT fuel by hot-electron penetration can significantly degrade performance.

Theories of the production and transport of these long-mean-free-path electrons have been confirmed qualitatively in the laboratory. Hot electrons are not observed directly: rather, one deduces their number and energy spectrum from the hard x rays and fast ions they generate. The relationship between the production of hot electrons and the observable quantities may be subtle, because the number and energy spectrum of hot electrons may be altered by transport and by reheating by the plasma waves that produced them initially. Indeed, x-ray data^{115,116} show a weak but significant dependence of hot-electron temperature on target material or Z, while the theory of their production seems to possess no such feature.

We have previously shown¹¹⁷ that transport via Coulomb collisions introduces a Z dependence to the hot-electron number that establishes equilibrium between the source and target interior. We also showed that the energy spectrum is altered (independent of Z). Here we report initial results of an analysis of the effect of reheating on the energy spectrum. Hot electrons can have long lifetimes in the system, because of weak thermalization, and can therefore traverse the plasma waves that produced them many times. Thus, even though the single-pass wave-particle interaction turns out to be weak, the cumulative effect of many such interactions on the hot-electron distribution can be significant. This improved modeling may help to reconcile

theory and experiment quantitatively. Rosen and Estabrook^{118,119} have also considered the reheating problem, employing a different approach.

Here we consider the interaction of test electrons with a model plasma-wave electric field, given by the product of a static envelope (which provides spatial localization) and a high-frequency traveling wave. Previous efforts¹²⁰⁻¹²⁴ have been primarily concerned with standing waves that arise in homogeneous plasma. The new feature of our work is the introduction of the phase velocity of the high-frequency field, which is localized according to wave-matching conditions in an inhomogeneous plasma.

The essential physics of reheating in the weak-interaction limit can be stated simply: electrons moving in the direction of the traveling wave gain energy in passing through it if they enter with a velocity less than its phase velocity, but lose energy if they enter with velocity greater than its phase velocity. This is the familiar physics of Landau damping of infinite plane plasma waves. Electrons transiting against the phase velocity of a localized traveling wave gain energy.

The energy gains and losses are second-order in the field amplitude and induce a diffusion of the particles in energy space. We show that the test-electron momentum change, averaged over the phase of entrance into the wave, is sufficient to describe the transformation of a distribution function via a Fokker-Planck operator.

Quasi-steady localized structures of intense high-frequency electron plasma waves and self-consistent ion-density depressions or shelves have been of continuing interest to workers on theories of anomalous interaction of external pumps with plasmas and anomalous plasma transport in unstable plasmas. Such structures can be responsible through wave-particle interaction for the conversion of wave energy to thermal energy. These structures also contribute to electron-transport phenomena by the same mechanism.

Investigations of self-consistent models of these systems have been illuminated by test-particle studies in which the structure acts on transiting electrons that do not themselves contribute to the dynamics of the structure. To this end, we solve the

motion of a particle through the field structure:

$$\dot{X} = V, \quad (152)$$

$$\dot{V} = \frac{-e}{m} E(X, t).$$

The electric field is regarded as known, and the change in momentum and energy of the test particles are investigated as a function of initial conditions. For the sake of definiteness, we write the field as the product of an envelope function, $a(x)$, and a high-frequency traveling wave:

$$E(x, t) = E_0 a(x) \cos \left[\omega \left(t - \frac{x}{v_p} \right) \right]. \quad (153)$$

E_0 is the strength of the field for $a(0) = 1$. The envelope function becomes negligibly small for $|x/L| > 1$, so that the length of the wave structure is $2L$.

Figure 3-32 shows a particular example of this model field. The field makes one complete cycle in the envelope $a(x)$; this is characteristic of resonant absorption. Note that a localized standing wave is obtained

Fig. 3-32. A localized traveling-wave electric field excited by a laser beam in an inhomogeneous plasma of electron density n_e . The subscript m indicates the density and position of laser/plasma-wave matching.

in the limit $v_0 \rightarrow \infty$; $v_0/\omega L$ and $(eE_0/m\omega)/\omega L$ or $eE_0 L/m(\omega L)^2$ are dimensionless parameters that characterize the wave structure.

We assume that the energy of electrons of interest is much greater than that of an electron oscillating in the wave [that is, the effective trapping energy $m(eE_0/m\omega)^2$]. This means that, after they are produced by the wave, the hot electrons interact only weakly with it. We calculate analytically the effects of the model field for such weakly interacting electrons. Because in plasmas of practical interest many electrons of nearly equal velocity will transit the field during one temporal period of the wave, we deal throughout with the phase average of the test-electron interaction. Thus we solve Eqs. (152) and (153) with ωt replaced by $\omega t + \phi$, and average the results over the phase at entry, ϕ .

To understand the effect of the structure on the plasma, we consider the transformation of a distribution of electrons as they transit the localized field. It should be clear that when the changes in test-electron momenta are small, a Fokker-Planck description of the transformation of the distribution function will emerge. To show this, we observe that an exact solution of the Vlasov equation can be written as

$$vf(L, v) = \int_0^{2\pi} \frac{d\phi}{2\pi} \int_0^\infty dv' v' f(-L, v') \delta[v - V(v', \phi)] \quad (154)$$

Equation (154) results from integrating over all contributions of the entering flux in v' and ϕ to the emerging flux in v . On the right-hand side of Eq. (154) we must evaluate the exit velocity, V , as a function of the entry velocity v' and phase ϕ . We assume that ϕ is distributed uniformly in the range $(0, 2\pi)$. Equation (154) is normalized properly, as can be verified by integration over v ; conservation of particles is assured because the total exiting flux equals the total entering flux. For given values of v' and ϕ , the δ -function in Eq. (154) has one zero, given by the solution of Eqs. (152) and (153) with $X(0) = -L$ and $V(0) = v'$.

It is convenient to write the solution of the equation of motion as

$$v = V(v', \phi) = v' + \Delta V(v', \phi) \quad (155)$$

When the change in velocity, ΔV , is small compared with the velocity scale of the distribution function, we can expand the integrand to obtain the change in the distribution to second order:

$$\Delta_{\text{wave}}(vf) = -\frac{\partial}{\partial v} [\langle \Delta V \rangle vf] + \frac{1}{2} \frac{\partial^2}{\partial v^2} [\langle (\Delta V)^2 \rangle vf] \quad (156)$$

In Eq. (156), $\Delta_{\text{wave}}(vf) = vf(L, v) - vf(-L, v)$, and $\langle \Delta V \rangle$ and $\langle (\Delta V)^2 \rangle$ are the phase-averaged changes in velocity of a test particle:

$$\begin{aligned} \langle \Delta V \rangle &= \int_0^{2\pi} \frac{d\phi}{2\pi} \Delta V(v, \phi) \\ \langle (\Delta V)^2 \rangle &= \int_0^{2\pi} \frac{d\phi}{2\pi} [\Delta V(v, \phi)]^2 \end{aligned} \quad (157)$$

Thus we have recovered the familiar Fokker-Planck form for the change in the distribution due to small changes in velocity of its particles.

To model the reheating on the longer time scale of interest in laboratory experiments, we introduce into the electron-transport equation the effective wave collision term

$$\frac{df}{dt_{\text{wave}}} = N_{\text{wave}} \Delta_{\text{wave}}(vf) \quad (158)$$

where N_{wave} is the number density of wave structures in the system.

The phase-averaged energy change of a test particle, $\langle \Delta \epsilon \rangle = \langle mv^2/2 - mv'^2/2 \rangle$, is

$$\langle \Delta \epsilon \rangle = \frac{m}{2} [2v \langle \Delta V \rangle + \langle (\Delta V)^2 \rangle] \quad (159)$$

For test electrons that interact weakly with the wave, Eqs. (156) to (158), combined with Eqs. (152), (153), and (155), suffice to describe the system. Considerable analytic reduction of the problem can be accomplished.

We solve Eqs. (152) and (153) by expanding about the free-streaming orbit:

$$X = -L + \Delta X^{(0)} + \Delta X^{(1)} + \dots \quad (160)$$

and

$$V = v' + \Delta V^{(0)} + \Delta V^{(1)} + \dots \quad (161)$$

To lowest order,

$$\Delta X^{(0)} = v't \quad (162)$$

$$\Delta V_{\phi}^{(0)} = -\frac{eE}{m} [X^{(0)}, t] \quad (163)$$

To first order,

$$\Delta \dot{X}^{(1)} = \Delta V^{(0)} \quad (163)$$

$$\Delta \dot{V}^{(1)} = -\frac{e}{m} \Delta X^{(1)} \frac{\partial E}{\partial X} [X^{(0)}, t] \quad (163)$$

Inspection of the model electric field in Eq. (153) gives

$$\langle \Delta V^{(0)} \rangle = 0 \quad (164)$$

For a symmetrical spatial envelope, $a(x) = a(-x)$, the second moment of $\Delta V^{(0)}$ gives the dominant term in $\langle (\Delta V)^2 \rangle$, namely

$$\langle (\Delta V)^2 \rangle = \left(\frac{eE_0 L}{m} \right)^2 \frac{1}{2v^2} \times \int_{-1}^1 d\xi \int_{-1}^1 d\xi' a(\xi) a(\xi') \cos[\alpha(\xi - \xi')] \quad (165)$$

Here $\alpha = \omega L(1/v - 1/v_p)$. Equation (165) is the result of substituting Eq. (153) into Eq. (157) and introducing the variable $\xi = [-L + \Delta X^{(0)}]/L$ into the term that survives phase averaging. Note that $\langle (\Delta V)^2 \rangle$ is of second order in the field strength, and that both traveling and standing waves are included in the expression.

The dominant term in $\langle \Delta V \rangle$ comes from $\Delta V^{(1)}$; it may be written as

$$\langle \Delta V \rangle = \frac{1}{2v} \frac{d}{dv} [v \langle (\Delta V)^2 \rangle] \quad (166)$$

From Eqs. (159) and (166), we see that the phase-averaged energy change of test electrons is

$$\langle \Delta \epsilon \rangle = \frac{m}{2v} \frac{d}{dv} [v^2 \langle (\Delta V)^2 \rangle] \quad (167)$$

Using Eq. (166) in Eq. (156), we find that the reheating operator of Eq. (158) becomes

$$\frac{df}{dt_{\text{wave}}} = \frac{N_{\text{wave}}}{2} \frac{\partial}{\partial v} v \langle (\Delta V)^2 \rangle \frac{df}{dv} \quad (168)$$

The phenomenology of the test-particle interaction physics can be obtained from Eqs. (165) to (167). In particular, Eq. (167) means that for velocities greater than the phase velocity of a traveling wave, we have

$$\langle \Delta \epsilon \rangle_{\text{TW}} \sim \frac{1}{4} \frac{eE_0 L}{m v^2} \frac{eE_0 L}{v} \text{sign}(\alpha) \quad (169)$$

Thus $\langle \Delta \epsilon \rangle_{\text{TW}}$ changes sign with the direction of entry ($v < 0$ or $v > 0$) of test particles into the field structure. We find that particles moving with the wave, but at speeds greater than the phase velocity of the wave, lose energy: $\langle \Delta \epsilon \rangle_{\text{TW}} < 0$. Particles moving with the wave at speeds less than the phase velocity, and particles moving against the wave, gain energy: $\langle \Delta \epsilon \rangle_{\text{TW}} > 0$. In fact, Eq. (167) shows that the sign of $\langle \Delta \epsilon \rangle$ is given by the sign of α , so that $\langle \Delta \epsilon \rangle_{\text{TW}} < 0$ only if $v > v_p$.

The discussion above does not apply to a standing wave (SW). In this case the high-velocity limit of Eq. (167) is

$$\langle \Delta \epsilon \rangle_{\text{SW}} \sim \frac{1}{4} \frac{eE_0 L}{m v^2} \frac{eE_0 L}{v^2} \quad (170)$$

Fig. 3-33. Test-particle interaction $\langle \Delta V \rangle$, $\langle (\Delta V)^2 \rangle$ and $\langle \Delta \epsilon \rangle$ of Eqs. (165) to (167) for $a(x) = \exp(-x^2/L^2)$. For the traveling wave, $v_p/\omega L = 1/\pi$ and $eE_0 L/m(\omega L)^2 = 1/25\pi$. For the standing wave, $v_p/\omega L \rightarrow \infty$.

Because the phase velocity of a standing wave is effectively infinite, all particles have velocities less than the phase velocity and thus gain energy from the interaction. Note that there is no asymmetry in the direction of entry of particles to the standing wave as there is for the traveling wave.

There is some structure in the phase-averaged quantities of Eqs. (165) to (167) at relatively low velocities. We will also see that the localized field structure exhibits an effective phase velocity, sometimes not exactly v_p , at which the interaction with test electrons is strongest. The effective phase velocity can be defined, for example, by maximizing $\langle \Delta v \rangle$ with respect to v .

Figure 3-33 shows $\langle \Delta v \rangle$, $\langle (\Delta v)^2 \rangle$, and $\langle \Delta v \rangle$ for a localized traveling wave and for the associated standing wave. All the quantities plotted agree well with the results of direct numerical integration of Eqs. (152), (153), (157), and (159).

Figure 3-34 shows the single-pass transformation of a Maxwellian distribution obtained by employing the results of Fig. 3-33 in Eq. (156). The direct numerical transformation by implementation of Eq. (154) compares favorably with these analytic results.

Finally, Figure 3-35 shows the numerical solution of $\partial f / \partial t = df / dt_{\text{wave}}$ from Eq. (158). The initial distribution is reheated on a time scale

$$\tau = (cE_0 L / m v_{\text{hot}}^2)^{-2} (N_{\text{wave}} v_{\text{hot}})^{-1} \quad (171)$$

and particles of higher and higher energy appear as the result of diffusive upscattering on the wave.

Authors: J. R. Albritton and
E. A. Williams

Direct-Implicit Plasma Simulation

We have experimented with several variants of the direct-implicit plasma simulation algorithms¹²⁵ with the aim of isolating those forms most suitable for extension to a two-dimensional, electromagnetic code. Previous code experimentation¹²⁶ emphasized best performance in an electrostatic implementation. In the work reported here, we selected a number of the simpler, though less rigorously justified, algorithms. Most of the variants were eliminated by test problems with parameters chosen to stress the linearization on which the field predictor is based. The survivors are as simple as any algorithm that has been proposed, yet are (so far) more robust than we expected.

The tests used values of $\omega_{pe} \Delta t$ as large as 120, much larger than have been reported previously^{126,127} (ω_{pe} is the electron plasma frequency and Δt is the time step). Some care was required to avoid nonlinear numerical instability when $(\omega_{pe} \Delta t)^2$ exceeded N_c , where N_c is the number of particles per cell (in one dimension, $N_c = n \Delta x$). Many applications of two-dimensional explicit codes require only that $N_c \gtrsim 10$. If the nonlinear instability were not circumvented, we would need more than $(\omega_{pe} \Delta t)^2$ particles per cell, which would be a severe limitation.

In this article we outline the method that works best and explain why other variants

Fig. 3-34. The single-pass transformation of a Maxwellian distribution according to Eq. (156) for the same field as in Fig. 3-33. Only positive v are shown; a one-dimensional, isotropic system is modeled.

Fig. 3-35. Reheating distribution functions from Eq. (17) for $t = 0$ and $t = 10\tau$, with the same field as in Figs. 3-33 and 3-34.

are less robust. The principal test problem, a variation of that of Ref. 126, features initially sharp gradients and a large range of density.

Several suitable schemes for time-differencing the equations of particle motion have been analyzed and applied.¹²⁸ Here we use the so-called D_1 scheme, written as

$$\frac{x_{n+1} - x_n}{\Delta t} = v_{n+1/2} \quad (172a)$$

$$\frac{v_{n+1/2} - v_{n-1/2}}{\Delta t} = \bar{a}_n + \frac{v_{n+1/2} + v_{n-1/2}}{2} \frac{qB_n}{mc} \quad (172b)$$

where

$$\bar{a}_n = \frac{1}{2} (\bar{a}_{n-1} + a_{n+1}) \quad (173)$$

a is the acceleration due to the electric field only, and n is the time level. The variables x , v , and a are quantities associated with each particle.

The direct-implicit method is used here in the following one-dimensional, unmagnetized, electrostatic algorithm. The position x_{n+1} of a particle at time level $n+1$, as given by an implicit time-integration scheme, can be written as

$$x_{n+1} = \beta \Delta t^2 a_{n+1} + \bar{x}_{n+1} \quad (174)$$

where $\beta > 0$ is a parameter that controls implicitness ($\beta = 1/2$ for the D_1 scheme), and $\bar{x}_{n+1}$ is the position obtained from the equation of motion with the acceleration a_{n+1} omitted:

$$\bar{x}_{n+1} = x_n + v_{n+1/2} \Delta t + \frac{1}{2} \bar{a}_{n-1} \Delta t^2 \quad (175)$$

$\bar{x}_{n+1}$ is known in terms of positions, velocities, and accelerations at times t_n and earlier. In its simplest form, which we adopt for this example, the direct-implicit algorithm is derived by linearization of the particle positions relative to $\bar{x}_{n+1}$.

At the grid point located at X_j , the charge density $\bar{\rho}_{j,n+1}$ is formed by adding the

contributions of the simulation particles at positions $\{\bar{x}_{i,n+1}\}$:

$$\bar{\rho}_{j,n+1} = \sum_i q_i S(X_j - \bar{x}_{i,n+1}) \quad (176)$$

where $X_j \equiv j\Delta x$, Δx is the grid spacing, i is the particle index, q_i is the charge, and S is the particle-grid interpolation spline. With this extrapolated charge density and the linearized implicit contribution $-\partial(\chi E)/\partial x$, the field equation is

$$\bar{\rho}_{j,n+1} = \frac{1}{\Delta x} [(1 + \chi_{j+1/2}) E_{j+1/2,n+1} - (1 + \chi_{j-1/2}) E_{j-1/2,n+1}] \quad (177)$$

where $\chi(x) = \Sigma \beta \bar{\rho}(x)/(q/m)\Delta t^2$ summed over species; i.e., $\chi = \beta(\omega_p \Delta t)^2$. We call χ the implicit susceptibility, because of the similarity of Eq. (177) to the field equation in nonuniform dielectric media. Note that $\chi \gg 1$ where $\omega_p \Delta t$ is large. Two representations used here are¹²⁵

$$\chi_{j+1/2} = \frac{1}{2} (\chi_j + \chi_{j+1}) \quad (178a)$$

and

$$\chi_{j+1/2} = \max(\chi_j, \chi_{j+1}) \quad (178b)$$

where

$$\chi_j = \Delta t^2 \sum_s [\beta \bar{\rho}_{j,n+1} q/m]_s \quad (179)$$

is a sum over species index s .

In terms of the field

$$E_{j,n+1} = \frac{1}{2} (E_{j-1/2,n+1} + E_{j+1/2,n+1}) \quad (180)$$

formed from values of E at half-integer positions, the particle acceleration is evaluated at $\bar{x}_{n+1}$ by an interpolation of the form

$$m_i a_{i,n+1} = q_i \Delta x \sum_j E_{j,n+1} S(X_j - \bar{x}_{i,n+1}) \quad (181)$$

where the sum is over grid points j and where S is the spline function used in Eq. (176). We have chosen not to include

spatial smoothing, because the resulting loss of resolution may be too costly in two dimensions and because smoothing makes the field equation more expensive to solve.

Two parameters measure the stress on the algorithm. The more important parameter is $\chi_1 = \beta q^2 \Delta t^2 / m |\Delta x|$, where q and m are the particle charge and mass and $|\Delta x|$ is the zone volume. This is a worst-case measure of the validity of linearization in the field prediction, as stressed by short-wavelength sampling fluctuations in the charge density. Written as $\beta(\omega_{pe} \Delta t)^2 / N_c$, this parameter is seen to be greater than unity when $\beta(\omega_{pe} \Delta t)^2 > N_c$; its significance is mentioned above.

The second parameter is $v_{th} \Delta t / \Delta x$, the ratio of the thermal electron transit distance per cycle to the zone size. Although a direct-implicit code is stable with $v_{th} \Delta t / \Delta x > 1$, energy conservation is degraded without spatial smoothing.

Depending on the differencing of the field equation, the ratio $(x_{n+1} - \bar{x}_{n+1}) / \Delta x$, a measure of the validity of linearization, can be as great as $\beta \Delta t^2 (q^2 / m) / |\Delta x| = \chi_1$ due to the field of a single particle, in regions of large density gradient or low density. It is essential that a particle contribute enough to the χ s multiplying all the $\{E_{j+1/2}\}$ that accelerate the particle and its neighbors in the same zone, so that the accelerations do not become too large. Equations (178a) and (178b) work well in the test problems considered. However, in the momentum-conserving algorithm,¹²⁶ the formation of χ must be more "local" than it is in Eq. (178) (that is, a particle must contribute to fewer χ s). With algebraic momentum conservation in any form attractive for extension to two dimensions and electromagnetic fields [i.e., a form that does not use spatial filtering, iteration, or "strict" formulations of the field predictor (in the sense given in Refs. 125 and 126)], we have been unable to run with $\chi_1 \approx 1$. With the formulation outlined here, we have obtained reasonable results with χ_1 well over 100, which considerably reduces the required number of particles per cell when $(\omega_{pe} \Delta t)^2$ is large. We are proceeding along these lines in implementing the two-dimensional electromagnetic algorithm.¹²⁹

Authors: A. B. Langdon, D. W. Hewett, and A. Friedman

Theory of Photodiodes

Photodiodes consist of aluminum electrodes, separated by a vacuum gap as shown in Fig. 3-36(a), with the cathode coated with a high-Z metal, such as gold. X rays with energies from ~ 10 to 500 keV are converted by the photoelectric effect into photoelectrons, which are emitted preferentially from the gold. Below the so-called Child-Langmuir limit (defined below), the resulting photocurrent is an approximately linear function of x-ray intensity; the "post-saturation" current above this limit is an approximately logarithmic function of the intensity.

Although photodiodes are commonly used as x-ray detectors in the linear regime,¹³⁰⁻¹³² their use in the post-saturation regime would lead to a new type of detector operable with thermal sources whose intensities covered several orders of magnitude. This current can be measured with a B-dot detector. The analysis presented here yields an expression for the current forced through the diode in the high-energy, relativistic case, which has not been treated before. The diode potential was obtained as long ago as 1921 for the nonrelativistic case.¹³³⁻¹³⁵

Photoelectrons emerge from the cathode with broad energy and angular distributions. Their space charge creates a potential minimum ϕ_1 at some position x_1 between the electrodes, as shown in Fig. 3-36(b). This potential minimum reflects low-energy electrons, allowing only those with energies greater than $-\epsilon\phi_1$ to contribute to the anode current, I_1 . In one dimension, these effects are conveniently represented on the phase-plane diagram of Fig. 3-36(c). The figure indicates that I_1 is only a fraction of the current I_0 emitted from the cathode. At energies above 500 keV, conversion occurs partially by the Compton effect. In this case, electrons are emitted preferentially in the direction of the incoming radiation and the angular distribution is narrower.

Under steady-state conditions, the distribution function of the electrons in the gap must be a function of the constants of the motion; in the one-dimensional case considered here, these reduce to the electron energy $W = \gamma mc^2 - \epsilon\phi$, where ϕ is the

potential, γ is the Lorentz factor, c is the speed of light, $-e$ is the electron charge, and m is the electron mass. If the incident x-ray (or γ -ray) spectrum is approximately thermal, it is reasonable to assume that photoelectrons are emitted from the cathode with a Maxwellian distribution. Since $\phi = 0$ at the cathode, the electron distribution in the gap is then

$$f(x, p_x) = A e^{-(W - mc^2)/kT}, \quad (182)$$

where k is the Boltzmann constant and T is the photoelectron temperature. At low energies, where relativistic effects can be neglected, such a distribution follows, in one dimension, from an isotropic distribution after integrating with respect to the transverse momenta p_y and p_z . In the relativistic case, these perpendicular momenta cannot be integrated out, and Eq. (182) implies emission normal to the cathode. (The effect of angular spread for relativistic electrons is considered later.)

The normalization constant A is related to the emitted current by

$$J_0 = \int_0^\infty e v_x f dp_x = e k T A, \quad (183)$$

and the current that reaches the anode is

$$J_1 = J_0 e^{e\phi_1/kT}. \quad (184)$$

Poisson's equation is now used to evaluate the potential $\phi(x)$. For $x < x_1$, we have

$$\frac{d^2\phi}{dx^2} = 4\pi e \int_{-\infty}^{+\infty} f(x, p_x) dp_x - 4\pi e \int_{p_T}^{\infty} f(x, p_x) dp_x, \quad (185)$$

where $p_T(x)$ is the momentum at the separatrix, shown in Fig. 3-36(c). The first integral in Eq. (185) represents the density if all electrons returned to the cathode, and the second integral gives the correction for those electrons that do not return. For $x > x_1$, Poisson's equation gives

$$\frac{d^2\phi}{dx^2} = 4\pi e \int_{p_T}^{\infty} f(x, p_x) dp_x. \quad (186)$$

▲ Fig. 3-36. (a) Physical configuration of a photodiode. (b) Corresponding space-charge potential. (c) Corresponding phase-space diagram showing returning, transiting, and void regions.

Fig. 3-37. Normalized potential minimum $e\phi_1/kT$ vs normalized emitted current J_0/J_{CL} for several values of the relativistic parameter $\tau = kT/mc^2$.

Substituting $f(x, p_x)$ from Eq. (182) into Eqs. (185) and (186), and integrating numerically, one obtains¹³⁶ the potential minimum, ϕ_1 . The potential $\phi(x)$ is plotted in Fig. 3-37 as a function of the anode current J_0 for

Fig. 3-38. Normalized anode current I_1/I_{CL} vs normalized emitted current I_0/I_{CL} for several values of $\tau = kT/mc^2$.

several values of the relativistic parameter $\tau = kT/mc^2$. In Fig. 3-37, I_0 is normalized to the Child-Langmuir current

$$I_{CL} = \frac{1}{9\pi} \left(\frac{2e}{m} \right)^{1/2} \frac{(kT'e)^{3/2}}{d^2}, \quad (187)$$

where d is the gap distance.

In Fig. 3-38 we plot the anode current I_1 , as evaluated from the potential by using Eq. (184). The figure shows that it is possible to drive currents in photodiodes well above the Child-Langmuir current limit, which corresponds to a voltage kT/e . Further, we see that the anode current I_1 is an approximately logarithmic function of the emitted current I_0 in this "post-saturation" regime.

In the nonrelativistic limit $\tau \rightarrow 0$, the result of Fig. 3-38 fits the simple scaling law¹³¹

$$\frac{I_1}{I_0} \approx \frac{2}{\left[1 + 22.4 \left(\frac{d I_0^{1/2}}{\tau^{3/4}} \right)^2 \right]}, \quad (188)$$

where I_0 is in kA/cm^2 , T is in keV, and d is in cm. This scaling law is valid for $I_0 \gtrsim 10 I_{CL}$; it does not extend to the limit $I_0 \rightarrow 0$, for which we should have $I_1 \rightarrow 0$. Figures 3-37 and 3-38 clearly show the relativistic effects on the potential minimum ϕ_1 and on the current I_1 . For $\tau < 0.16$ (corresponding to $T < 82$ keV), relativistic corrections to the current are less than 8%, and corrections to the potential are not visible in Fig. 3-37. For large values of τ , however, relativistic effects on I_1 are important, and Eq. (188) overestimates the anode current.

As already mentioned, the analysis above does not account for the relativistic effects of the perpendicular momenta. To estimate

these effects we consider an electron distribution defined by

$$f(x, p_x, p_{\perp 0}) = \frac{A}{2\pi p_{\perp 0}} e^{-\frac{W - mc^2}{kT}} \delta(p_{\perp} - p_{\perp 0}). \quad (189)$$

where $W = \gamma mc^2 - e\phi$, $\gamma = [1 + (p_x^2 + p_{\perp}^2)/m^2 c^2]^{1/2}$, and $p_{\perp}$ is the perpendicular component of the momentum, which is a constant of the motion. Because of the δ -function dependence of the distribution on $p_{\perp}$, all particles have the same perpendicular momentum, $p_{\perp 0}$. For a given energy, they are emitted at some polar angle θ from the x axis, given by

$$\tan \theta = \frac{p_x}{p_{\perp}} = \left[\frac{m^2 c^2 (\gamma^2 - 1) - 1}{p_{\perp 0}^2} \right]^{-1/2}. \quad (190)$$

This angle of emission is a decreasing function of the energy, in qualitative agreement with photoelectric and Compton emission laws.

Because the Lorentz factor may be written as $\gamma = \gamma_x \gamma_{\perp}$, where $\gamma_{\perp} = [1 + (p_{\perp 0}/mc)^2]^{1/2}$ is a constant and $\gamma_x = [1 + (p_x/\gamma_{\perp} mc)^2]^{1/2}$, it follows that the results of Figs. 3-37 and 3-38 remain valid in the present case if the electron rest mass m is replaced by an effective rest mass $\gamma_{\perp} m$ in the definitions of τ and I_{CL} .

Authors: J. Denavit and G. D. Craig

LASNEX and Atomic Physics

Overview

Several years ago, we converted LASNEX to run on the Cray I computer. At the same time, we rewrote the entire code to take advantage of modern structured-design techniques, which promised to significantly improve code-development efficiency. This investment is now paying off: we were able to spend a substantial fraction of 1983 solving long-standing problems, several of

which are detailed in this section. During 1984 we will move LASNEX to at least one new computer, and will convert it to run under a new operating system, so further work of this kind may have to wait for one or two more years.

We have nearly completed rewriting the LASNEX routines that geometrically track laser rays (see the following article). A major enhancement resulting from this rewrite is the ability to handle skew rays (those with a propagation component around the problem's symmetry axis). This will allow us to determine the propagation of real laser beams more accurately, and to simulate three-dimensional backscattering and sidescattering of light. Our previous model forced light rays to travel in the (r, z) plane of the LASNEX calculation, and often created anomalously high intensities near the symmetry axis.

We have developed a Thomas-Fermi-based equation-of-state package (see the article by that title later in this section) that will soon be incorporated into LASNEX. The package emphasizes fundamental calculations, rather than extensive tabulation, and can therefore be used for any material. Table-interpolation errors have disappeared. Because it is based on a free-energy calculation, the new package returns entropy as well as pressure and energy, and also guarantees that all values are thermodynamically consistent. In addition, our new routines calculate the equations of state separately for electrons and ions; this is necessary if the two fluids have different temperatures. The routines also return the degree of ionization, for use by other LASNEX physics packages.

We have improved our ability to calculate high-density implosions. We finally isolated and fixed a long-standing two-dimensional numerical instability, involving artificial viscosities, which had seriously restricted the design of high-convergence multi-shell targets. We also significantly improved the mesh-positioning and advection algorithms in our semi-Eulerian hydrodynamics package. We have begun to model material mixing caused by hydrodynamic instabilities. Finally, we improved our theoretical understanding of the thermal occupation and pressure of resonance states (see "Thermal

Occupation and Pressure of Resonance States," later in this section); these states can be important to the atomic physics of high-density materials.

We made the iterative method for self-consistently calculating multigroup radiation-matter coupling more efficient by accelerating its convergence rate (see "Convergence Acceleration for Multigroup Radiation-Matter Coupling Iteration," later in this section). We incorporated improvements in the theoretical energy spectrum of thermonuclear neutrons into codes that simulate experimental time-of-flight measurements and pinex pictures (see "Energy Spectrum of Thermonuclear Neutrons," later in this section).

Author: G. B. Zimmerman

Three-Dimensional Laser Ray Tracing on an r - z Lagrangian Mesh

A ray-tracing package is used in LASNEX to determine the spatial distribution of deposited laser energy. In the package used for the past decade,¹³⁷ all rays are assumed to be primary—that is, to have trajectories that everywhere lie in the meridional plane. In some cases this leads to an anomalous brightness on-axis.

We have developed a new package that computes the trajectories of an arbitrary set of laser rays through the mesh. We have also introduced a facility for modeling the effects of scattering processes on ray trajectories. This package will allow us to calculate real laser beams more accurately, and to simulate three-dimensional scattering of laser light.

The LASNEX plasma is axisymmetric, and is specified on an r - z Lagrangian mesh of quadrilaterals. In order to specify a unique value of the electron density n at each point in space, the mesh is triangulated by drawing one of the two possible diagonals in each zone. Ray trajectories are computed using the gradient-index geometrical-optics equations. On the basis of computed ray path lengths, energy is deposited in each zone by inverse bremsstrahlung; the ray intensity is correspondingly

Fig. 3-39. Typical LASNEX mesh, illustrating the difficulty of efficiently obtaining a zone index given (r, z) of a point in space.

attenuated. Resonance absorption at turning point[†] is modeled. Instability absorption is modeled in various ways.

The ray-trajectory calculation must detect the crossing of each triangle boundary, because no efficient algorithm exists for determining zone indices given (r, z) on a distorted mesh (see Fig. 3-39). The trajectories are recomputed each LASNEX cycle, because the mesh moves and densities evolve.

The conserved angular momentum about the z axis, $h = rv_\phi$, is zero for a primary ray and nonzero for a skew ray. In vacuum (or in matter with $n = \text{constant}$) a ray is a straight line in (x, y, z) coordinates and a hyperbola in (r, z) coordinates. Coordinates (x, y, z) are evolved for each ray. This directly provides three-dimensional information on the deflection of each ray by the plasma. Each ray, insofar as it affects the plasma, represents a set of rays distributed over all azimuthal angles ϕ . In vacuum, this set forms a hyperboloid of revolution.

The new algorithm was designed to satisfy the following criteria:

1. It must reduce to the old "exact" algorithm for $h = 0$.
 2. It must be exact for all rays in vacuum or in constant-density plasma.
 3. It must lead to small trajectory errors when h and ∇n are both nonzero.
- In addition, efficiency was an important concern.

The equation of motion for a ray is

$$\ddot{x} = \nabla(\eta^2/2) \quad (191)$$

where η is the plasma index of refraction; a dot denotes differentiation with respect to a pathlike parameter along the path. In this

equation the quantity $-\eta^2/2$ plays the role of a potential energy field: the ray "energy" is $v^2/2 - \eta^2/2 = 0$, and the ray speed is just $|\dot{x}| = |\dot{x}| = \eta$. In these units, $c = 1$. In a vacuum, $\eta = 1$.

Because the refractive index is given by

$$\eta = \sqrt{1 - n/n_c} \quad (192)$$

where $n_c = m\omega^2/4\pi e^2$ is the critical density, we can write the ray equation of motion as

$$\ddot{x} = -\frac{1}{2n_c} \nabla n(x) \quad (193)$$

In the new package, ray positions are advanced using a sequence of time steps. The mesh has been triangulated, and zone-centered n -values have been interpolated onto vertices. Thus $|\nabla n|$ is constant within each triangle, and ∇n lies in the r - z plane. The direction of ∇n varies as a function of azimuthal position ϕ ; it rotates as a skew ray (one for which $h \neq 0$) moves. We take a time step using the approximation $\nabla n = \text{constant}$ during the step (this will be valid if the time step is small enough). Thus, during each step the path is a parabola in (x, y, z) space, although not necessarily in r - z space as it was in the old package.

For notational simplicity, we define the vector K , whose components are

$$\begin{aligned} K_r &= \frac{(\nabla n)_r}{2n_c} \\ K_z &= \frac{(\nabla n)_z}{2n_c} \\ K_\phi &= 0 \end{aligned} \quad (194)$$

In terms of K , the ray equation of motion becomes

$$\ddot{x} = -K(x) \quad (195)$$

In finite-difference form, the ray position is advanced using

$$x(\Delta t) = x_0 + \dot{x}_0 \Delta t - \frac{1}{2} K_0 \Delta t^2 \quad (196)$$

where

$$\mathbf{K}_0 = \mathbf{K}(\mathbf{x}_0) = \frac{K_r}{r_0} (x_0 \hat{\mathbf{x}} + y_0 \hat{\mathbf{y}}) + K_z \hat{\mathbf{z}} \quad (197)$$

and

$$r = \sqrt{x^2 + y^2} \quad (198)$$

We do not explicitly employ conservation of h , and work in (r, z) space, because doing so would provide an exact treatment of the limits $h = 0$ and $K_r = 0$.

The velocity advance is *not* implemented as a simple derivative of the position:

$$\dot{\mathbf{x}}(\Delta t) \neq \dot{\mathbf{x}}_0 - \Delta t \mathbf{K}_0 \quad (199)$$

rather, for greater accuracy, it is obtained from

$$\begin{aligned} \dot{\mathbf{x}}(\Delta t) &= \dot{\mathbf{x}}_0 - \Delta t \bar{\mathbf{K}} \\ &\approx \dot{\mathbf{x}}_0 - \Delta t \mathbf{K}[\mathbf{x}(\Delta t/2)] \quad (200) \end{aligned}$$

The approximation used for the time-centered force in Eq. (200) is

$$\begin{aligned} \bar{\mathbf{K}} &= \left[\frac{x_0 + x(\Delta t)}{r_0 + r(\Delta t)} \right] K_r \hat{\mathbf{x}} \\ &+ \left[\frac{y_0 + y(\Delta t)}{r_0 + r(\Delta t)} \right] K_r \hat{\mathbf{y}} + K_z \hat{\mathbf{z}} \quad (201) \end{aligned}$$

This form does not diverge for r_0 or $r(\Delta t)$ small, and generally produces a better trajectory than other forms we have tried.

On exit from a triangle, the velocity is reset to conserve energy and angular momentum. First,

$$v_\phi(\Delta t) = \frac{h}{r(\Delta t)} \quad (202a)$$

then,

$$|\mathbf{v}_{rz}|^2 = \eta^2 - v_\phi^2 \quad (202b)$$

finally,

$$\begin{aligned} v_r(\Delta t) &= \text{sign}[v_r(\text{after step})] \\ &\times \sqrt{\max(|\mathbf{v}_{rz}|^2 - v_\phi^2, 0)} \quad (202c) \end{aligned}$$

because the z advance is "exact." Resetting v after every step generally produces a poorer trajectory.

The crossing of every triangle edge must be computed, to give knowledge of a ray's position in the logical mesh of zone indices. The equation of edge i is

$$g_i(r, z) = U_i r + V_i z + W_i = 0 \quad (203)$$

The sign of g_i changes as one crosses an edge. Along the trajectory we take a sequence of N_s substeps of size $\Delta t_l = (l/N_s)\Delta t_{\text{accuracy limit}}$ for $l = 1, 2, \dots, N_s$, and compute (for each edge) the product $g_{\text{before step}} \times g_{\text{after step}}$. If this product is negative for any edge, a crossing has occurred and we exit the substep loop. We compute crossing times of all edges for which g has changed sign, using a new, high-order variant of *regula falsi* iteration to solve for $t(g = 0)$. The ray path is well defined, and the triangle generates a closed volume when rotated about the z axis, so the exit time is the minimum crossing time. The time step size Δt will thus be either the minimum of the crossing times, if a crossing was found by substepping, or the accuracy-limit time associated with the $\mathbf{K} = \mathbf{K}_0$ approximation (and other limits, discussed later), if no crossing was found.

Additional logic is used to ensure that the crossing computed is, in fact, a true exit of the triangle.

For greater efficiency, variable time-step sizes are employed. Since the algorithm is "exact" for either $h = 0$ or $K_r = 0$, lowest-order trajectory errors vary as hK_r . We thus wish to control the dimensionless quantity $hK_r \Delta t^3/r^3$. We set

$$\begin{aligned} \Delta t \text{ (accuracy limit due to fixed } K) &= \frac{f_{\text{acc}}}{|hK_r|^{1/3}} (r_{\text{entry}} - r_{\text{min}}) \left(\frac{r - r_{\text{min}}}{r_{\text{entry}} - r_{\text{min}}} \right)^{1/3} \quad (204) \end{aligned}$$

where $r_{\min}$ and f_{acc} are input parameters, r_{entry} is the radius of the ray upon entry to the triangle, and $\alpha = 0, 1/2$, or 1. (So far, $\alpha = 1/2$ seems to work best.)

We disallow four kinds of time step:

1. Steps that would rotate the ray through more than 90 deg; we do this because the approximation $K = \text{constant}$ has completely broken down by then. This requirement prevents primary rays from stepping past $r = 0$.
2. Steps past a turning point, because stopping at turning points both decouples the resonance absorption attenuation from the inverse bremsstrahlung and facilitates the implementation of a parameterized scattering at turning points.

Fig. 3-40. A bowtied mesh, and a ray being tracked through the negative area part of a bowtied zone. The dotted portion of the path represents tracking backward in time.

Fig. 3-41. The mesh near the center of a target, showing zero-area triangles.

Fig. 3-42. Ray trajectories through a zone for (a) $f_{\text{acc}} = 0.01$; (b) $f_{\text{acc}} = 0.05$; (c) $f_{\text{acc}} = 0.2$.

3. Steps larger than $64 \times \Delta t_{\text{substep}}$.
 4. Steps larger than $4 \times \Delta t_{\text{acc, previous}}$.
- The limit Δt_{acc} is modified to incorporate these further limits.

For robustness, we also employ a *posteriori* consistency and accuracy controls. The next time step's accuracy-limited Δt is computed before the present step is implemented: if $\Delta t_{\text{acc, next}}$ is $< \Delta t/4$, then Δt is halved and the present step is retried. This reduces inaccuracies arising from taking a large jump to small radius.

The new package includes many other elements, such as:

- Logic to loop over ray states for each ray in turn.
- An entry search scheme that sweeps over boundary zones to find the entry point at smallest time for each ray. The package also checks for reentry of escaped rays.
- Logic to advance triangle and zone indices.
- The turning-point scattering package.
- A ray-plotting facility.
- A facility to track through a bowtied zone, tracking backwards in time if the ray has entered the negative-area triangle of the bowtie (see Fig. 3-40).
- A facility to track through a zero-area triangle (see Fig. 3-41).

Figure 3-42 shows the trajectory of a typical skew ray ($h = 0.1$) refracting through a turning point. The critical density is 2.0 (arbitrary units); densities at the vertices are indicated. The dots, which represent the ray's position after each time step, are connected for clarity. Initial and final tangent

directions are indicated by straight lines at the borders of the zone.

In Fig. 3-42(a), the accuracy criterion f_{acc} was 0.01, and a very accurate trajectory was obtained using 43 steps; in Figs. 3-42(b) and 3-42(c), f_{acc} was 0.05 and 0.2, respectively. Figure 3-42 and Table 3-5 show that the trajectory is quite good even with a minimal number of steps; we generally find this to be so, except for highly skew rays.

In Fig. 3-43, six rays of increasing skewness are traced through a LASNEX mesh. Here $z = 0$ is a symmetry plane, and the density decreases from well above critical to well below with increasing distance from $(0, 0)$. Ray A is a primary ray, while the others are skew; their hyperbolic trajectories in vacuum are evident to the left of their starting point at $r = 0.2, z = 0.9$. One ray (F) has missed the mesh altogether.

In general, the new package uses about three times as much cpu time to track a primary ray as the old package; cpu times for the deposition should be of the same order of magnitude as the tracking time, and should be similar for both old and new packages, leading to an overall increase in cpu time of about a factor of 2. (In a real system, of course, most rays are at least somewhat skew, so such comparisons may not be meaningful.)

We plan to use the new ray-tracing package to model more realistic distributions of laser rays as they come to a volume focus (as opposed to a point or line focus); to model scattering by nonaxisymmetric density ripples near critical density; and to model diagnostic laser beams.

Author: A. Friedman

Thomas-Fermi-Based Equation of State Package

We have developed a new equation of state (EOS) package intended as a utility code. While based on previous work,¹³⁸⁻¹⁴¹ the package will extend our capabilities in several ways. It has the following characteristics:

- The package is self-contained, with no linkage to external data bases. The user must supply only a small number of inputs that describe the material composition and properties of the cold solid.

Run	f_{acc}	No. of steps	z_{exit}	ϕ_{exit} (deg)	t_{exit}	cpu time (s)
a	0.01	43	0.50341	4.734	4.07573	1.2×10^{-3}
b	0.05	10	0.50342	4.727	4.07573	4.1×10^{-4}
c	0.2	4	0.50356	4.662	4.07577	2.8×10^{-4}

▲ Table 3-5. Data corresponding to the runs of Fig. 3-42.

Fig. 3-43. Six rays traced through a typical LASNEX mesh. Ray A is primary; the rest are skew.

Fig. 3-44. Pressure $p(\rho)$ for room-temperature aluminum. The new package is very close to typical EOS tables.

- The package supplies thermodynamic data (including electron and ion entropy) that will allow designers to measure the importance of heat flow, shocks, and other nonadiabatic processes. This will provide a tool for the optimization of implosions or other hydrodynamic flows. Figures 3-44 to 3-47 illustrate typical results obtained from the package.
- The electron and ion equations of state can be obtained separately; this

Fig. 3-45. Aluminum adiabats (lines of constant entropy).

Fig. 3-46. Aluminum ionization contours as predicted by the TF method.

Fig. 3-47. Aluminum isobars (pressure contours).

is a necessity if the two species have different temperatures.

- The EOS model is derived from the Los Alamos SESAME equation of state, and has been extensively tested for many cases. Although the method is not suitable for all applications, we can identify materials and conditions for which the greatest errors or uncertainties arise.
- The package generates EOS data from an accurately calculated free energy. While conventional library data bases represent the thermodynamic functions in a way that is inherently inconsistent, we employ a new interpolation method that greatly reduces interpolation errors and assures exact thermodynamic consistency.
- The call-up time is about 16 μ s per call (on the Cray 1), about 10 times greater

than that of the usual EOS library schemes. This disadvantage is offset by improved numerical quality of the data and improved thermodynamic consistency, and by the fact that a single call to the package returns all EOS quantities, their derivatives, and the plasma ionization state $Z^*(\rho, T)$, which otherwise would have required additional calculations.

- The package permits time-dependent alteration of the material composition, as produced by thermonuclear reactions or by atom mixing of materials because of the Rayleigh-Taylor instability.

EOS Structure. The EOS model is based on the assumption that electron and ion quantities are additive to sufficient accuracy for our applications. A more rigorous description^{141,142} would be significantly more complicated.

Within the additive assumption, we take the free energy per gram to be

$$F(T_e, T_i, \rho) = F_e(T_e, \rho) + F_b(T_e, \rho) + F_i(T_i, \rho) \quad (205)$$

In this representation, F_e is the electronic free energy obtained from the Thomas-Fermi statistical model, scaled to refer to the material in question. F_b is a semi-empirical correction for chemical bonding effects in the solid state. F_i is a model form for the ion equation of state, which reduces to ideal-gas and solid Grüneisen models in the appropriate limits. These models are discussed below.

The pressure p , entropy S , and energy E are obtained from the free energy according to the following equations:

$$p_e = \rho^2 \frac{\partial F_e}{\partial \rho}, \quad p_b = \rho^2 \frac{\partial F_b}{\partial \rho}, \quad p_i = \rho^2 \frac{\partial F_i}{\partial \rho};$$

$$S_e = -\frac{\partial F_e}{\partial T_e}, \quad S_b = -\frac{\partial F_b}{\partial T_e}, \quad S_i = -\frac{\partial F_i}{\partial T_i};$$

$$E_e = F_e + T_e S_e, \quad E_b = F_b + T_e S_b, \quad E_i = F_i + T_i S_i \quad (206)$$

The total pressure, entropy, and energy are sums of the additive contributions shown. The complete EOS is thermodynamically consistent because each part is consistent;

for example,

$$p_e - T_e \frac{\partial p_e}{\partial T_e} = \rho^2 \frac{\partial E_e}{\partial \rho} \quad (207)$$

The tabular interpolation scheme we employ for the electronic free energy is designed to satisfy Eq. (207) exactly. The bonding correction terms are independent of temperature, so $S_b = 0$ and $E_b = F_b$. Both bonding and ion free energies are obtained from analytic formulas that obey the consistency conditions exactly.

The EOS model is intended principally for application to calculations on laser fusion targets, in which the main emphasis is on hot plasma conditions. In general terms, the model should be adequate for temperatures above 1 eV. We have made no effort to describe the two-phase liquid-vapor region that occurs below the critical point. Because ICF target calculations usually start from cold, condensed matter, we describe the initial solid state accurately enough to assure the following:

1. Zero pressure at the correct normal density ($T = 0$ case).
2. Approximately correct density jumps for strong shocks beginning with the cold solid.
3. Smooth connection to the high-pressure plasma state, assuming there are no important phase transitions.

Below we describe how these goals are achieved. Because the entire method is based on a Thomas-Fermi description, we anticipate that there may be reduced accuracy for materials of very low or very high Z .

Electron Equation of State. The electronic free energy $F_e(T_e, \rho)$ is taken from the Thomas-Fermi (TF) model. Among the physical effects included by this model are partial ionization produced by temperature or compression and the core-repulsion pressure of high-density atoms. In previous work we have compared the TF theory with other approaches.¹⁴¹⁻¹⁴³ The greatest differences are produced by shell-closure effects and by the need for correction near the cold solid state to reproduce effects of chemical bonding. The significant differences between the TF model and more accurate theoretical models are greatly

reduced by the semiempirical bonding correction described below.

The TF data are calculated by a self-consistent-field solution of the TF equations for a spherically compressed atom. This is a numerical calculation, but the results scale with Z according to the formulas

$$F(\rho, T) = \frac{Z^{7/3}}{A} F_0(\rho_0, T_0) \quad ,$$

$$\rho = Z^{1/3} \rho_0 \quad , \quad (208)$$

$$T = Z^{1/3} T_0 \quad .$$

Together with these equations, a single table of $F_0(\rho_0, T_0)$ generates the electron free energy for any other element.

Although the range of interest extends over 6 to 8 decades in density, centered at 1 g/cm³, and over 3 to 5 decades in temperature, the TF table for F_0 covers a much larger range so as to include the desired conditions for all elements. We constructed a large and accurate TF free-energy table for this purpose. The method used to fit the table is described below.

Electron Bonding Corrections. The TF theory describes the pressure resulting from free electrons and from the ion-core repulsion of bound electrons. It does not describe the situation near the cold solid state, in which electron chemical bonds support a finite matter density without a large applied pressure. To obtain an approximately correct answer in this region, we add a semiempirical bonding correction to the energy. The bonding energy is independent of temperature and therefore applies also to the free energy:

$$E_b = E_0 \{ 1 - \exp b [1 - (\rho_s/\rho)^{1/3}] \} \quad , \quad (209)$$

where ρ_s = initial solid density and E_0 and b are parameters that characterize the solid bonding. The energy E_b approaches E_0 at very low density, is zero at solid density, and approaches a negligible constant at high density.

The bonding pressure is then

$$p_b = \rho^2 \frac{\partial E_b}{\partial \rho} = - \left(\frac{E_0 \rho_s b}{3} \right) \left(\frac{\rho}{\rho_s} \right)^{4/3} \times \exp b [1 - (\rho_s/\rho)^{1/3}] \quad . \quad (210)$$

We determine E_0 and b by requiring that the total pressure and bulk modulus be correct for the initial cold solid. The experimental bulk modulus is required to be

$$B = \rho \left(\frac{\partial p}{\partial \rho} \right)_{\rho_s} = \left[\rho \frac{\partial p_e}{\partial \rho} + \rho \frac{\partial p_i}{\partial \rho} \right] - (2 + b) \left(\frac{E_0 \rho_s b}{9} \right), \quad (211)$$

and this equation, together with the requirement that $p_e + p_b + p_i = 0$, establishes E_0 and b . With a correct bulk modulus, the model predicts reasonably accurate shock-wave Hugoniot.

Ion Equation of State. The ions are described by a unified phenomenological model that combines several well-known theories. These include a high-temperature ideal-gas law, a scaling-law EOS for the dense vapor, a melting temperature determined by the Lindemann melting law, and a combination of quantum (Debye) and classical specific heats for the solid at low and high temperatures. The ion pressure is determined by a Grüneisen equation of state with a density-dependent Grüneisen parameter γ . These quantities are calculated by simple formulas developed originally by R. D. Cowan. Although no simple model could accurately predict all properties of the enormous variety of elements and compounds known to chemistry, the Cowan model works surprisingly well for most materials.

The equations of the Cowan model are described elsewhere,¹⁴⁰ and will not be summarized here.

Thomas-Fermi Table. Although the TF theory is based on relatively simple equations, it would not be practical to evaluate the thermodynamic variables with a direct in-line calculation. Therefore it is necessary to tabulate the TF free energy. Fortunately, scaling laws allow us to obtain the EOS for any element by scaling from a single TF table. The specific free energy of any composite material is expressed as

$$F = \frac{Z}{A} \xi^{4/3} F_0 \left(\frac{T}{\xi^{4/3}}, \frac{Z\rho}{A\xi^2} \right), \quad (212)$$

where F_0 is the specific electronic free energy for hydrogen, Z and A are the average nuclear charge and mass of the mixture,

and ξ is a mixture-dependent quantity that must approach Z as the mixture approaches a single element. The uncertainty in the exact formulas for ξ arises because the TF scaling laws do not directly apply to mixtures of more than one element. We use

$$\xi = \exp \left[\frac{Z \ln Z}{Z} \right], \quad (213)$$

which is correct near full ionization. We will experiment with various forms for ξ .

Our basic TF table consists of values of energy, entropy, pressure, and the derivative dP/dT on a rectangular grid logarithmically spaced in temperature and density. The table extends to low density and temperature to handle the scaled values generated by arbitrary materials. We use a coarse table, with two points per decade in density (from 10^{-9} to 10^4 g/cm³) and three points per decade in temperature (from 10^{-6} to 10 keV). (Originally we constructed a table with much finer resolution, which we used to generate the values of dP/dT . The finer table enabled us to verify the accuracy of interpolation from the coarse table: agreement to one part in 10^3 is obtained.)

Interpolation in the TF table has been designed so that it reduces exactly to correct forms in both degenerate and nondegenerate limits. The functions provide for continuity of energy, entropy, pressure, and dP/dT across the table box boundaries. This is accomplished by defining the new quantities

$$u = \rho^{-2/3}, \quad (214)$$

$$H(u, T) = uF = u(E - TS),$$

for which it follows that

$$\begin{aligned} \frac{\partial H}{\partial T} &= -uS, \\ \frac{\partial H}{\partial u} &= F - \frac{3}{2} \frac{p}{\rho}, \\ \frac{\partial^2 H}{\partial u \partial T} &= -S - \frac{3}{2\rho} \left(\frac{\partial p}{\partial T} \right). \end{aligned} \quad (215)$$

H and these three derivatives are thus defined at each point in the table. Within any table box we use these 4 quantities, evaluated at each of the 4 corners of the box, to

determine the 16 unknowns in the following interpolation function for H :

$$H(u, T) = A(u) + B(u)T + C(u)T^2 + D(u)T \ln T, \quad (216)$$

where

$$\begin{aligned} A(u) &= A_a + A_b u + A_c u^2 + A_d u \ln u, \\ B(u) &= B_a + B_b u + B_c u^2 + B_d u \ln u, \\ C(u) &= C_a + C_b u + C_c u^2 + C_d u \ln u, \\ D(u) &= D_a + D_b u + D_c u^2 + D_d u \ln u. \end{aligned} \quad (217)$$

This form reproduces both degenerate and nondegenerate limits exactly.

The fitting can be done very efficiently by breaking it into several one-dimensional fits. The interpolating function is, in effect, a cubic spline with the cubic term replaced by $x \ln x$. Along each table box edge, H and its derivatives are determined entirely by values specified at the two end-points of that edge. This ensures continuity of energy, pressure, and entropy. This is necessary because derivatives of these quantities must be calculated by the EOS package.

To handle the rare cases for which T and/or u lie outside the table, we have established extrapolation boxes at the table perimeter that assure the following limiting forms:

$$\begin{aligned} H &\rightarrow H_0 + uB(T) + u^2C(T) & u \rightarrow 0, \\ H &\rightarrow uB(T) + D(T)U \ln u & u \rightarrow \infty, \\ H &\rightarrow A(u) + C(u)T^2 & T \rightarrow 0, \\ H &\rightarrow A(u) + B(u)T - \frac{3}{2}T^2 \ln T & T \rightarrow \infty, \end{aligned} \quad (218)$$

where the constant $H_0 = 3E_F u/5$, and where $E_F \propto \rho^{2/3}$ is the Fermi energy of fully ionized electrons. Finally, the degree of ionization is calculated from the TF pressure by using fits to the Fermi functions.

Summary. Because of the many constraints satisfied by the interpolation scheme outlined above, the package functions very smoothly and generates very satisfactory EOS data. Because of this, and because of the interest in entropy edits for

use by designers, we anticipate that this package will be used for many practical applications. Because many of our EOS tables were generated by essentially the same equations, there should be no significant difference in the physics, and a significant improvement in the numerical representation of EOS data.

Authors: R. M. More and
G. B. Zimmerman

Thermal Occupation and Pressure of Resonance States

Atomic properties of hot, dense plasmas (such as those that arise in laser fusion) are often calculated by application of the self-consistent-field approximation to a single atom surrounded by a continuous positive charge background.¹⁴⁴⁻¹⁴⁶ The electron eigenstates in the spherically symmetric potential $V(r)$ are populated according to the Fermi-Dirac distribution law

$$f(\epsilon) = [1 + \exp(\epsilon - \mu)/kT]^{-1}. \quad (219)$$

The positive charge distribution is determined by the plasma density through an extended ion-sphere model.¹⁴⁴ The total charge density is required to be consistent with the original potential $V(r)$.

The one-electron energy levels can be divided into two categories: localized *bound states*, with discrete negative energies, and *continuum states*, which take the form of an incident plane wave plus outgoing scattered waves. At certain densities there are shape resonances or resonant scattering states in the continuum. In practice, the resonant states appear at high densities for strongly screened potentials, e.g., with ion-sphere boundary conditions [$V(r) = 0$ for $r > R_0$]. The resonances are directly associated with pressure ionization of bound electrons.

Although resonances occur in the continuum, they closely resemble the discrete bound states. Indeed, as density increases, any particular bound state, of negative energy $E_{n,l} < 0$, changes into a resonance as its energy increases through zero. After this, it may be assigned a complex energy eigenvalue $\tilde{E}_{n,l} = E_{n,l} + i\Gamma_{n,l}$, whose imaginary part $\Gamma_{n,l}$ is proportional to the width of the resonance.

When the resonance state has significant thermal occupation, the compression that raises it to higher positive energies is resisted by a pressure. The thermal average occupation and pressure of resonance states can be rigorously determined by resolving the continuum formulas into additive contributions.

We begin with the density of states for a free-electron gas interacting with a single spherically symmetric potential¹⁴⁷:

$$g(\epsilon) = g_0(\epsilon) + \frac{2}{\pi} \sum_{l=0}^{\infty} (2l+1) \frac{d\delta_l}{d\epsilon} \quad (220)$$

The free-electron density of states is

$$g_0(E) = \frac{1}{2\pi^2} \left(\frac{2m}{\hbar^2} \right)^{3/2} V \sqrt{\epsilon} \quad (221)$$

The scattering phase shifts are conveniently calculated from the Jost function for potential scattering. (Our analysis applies to any potential having finite range R_0 .) The Jost function $f_l(k)$ is defined as the limit

$$f_l(k) = \lim_{r \rightarrow 0} \frac{(kr)^l}{(2l-1)!!} f_l(k;r) \quad (222)$$

where the Jost solution $f_l(k;r)$ of the radial Schrödinger equation is uniquely defined by the boundary condition

$$f_l(k;r) \rightarrow i^l e^{-ikr} \quad (223)$$

The energy is $\epsilon = (\hbar^2/2m)k^2$. The definitions and notation are consistent with those of Ref. 148.

In terms of the Jost function, the resonance energies are determined from complex wave vectors $Q_{n,l}$ for which the boundary condition at small radii is also obeyed:

$$f_l(Q_{n,l}) = 0 \quad (224)$$

The resonance energy is $\tilde{E}_{n,l} = (\hbar^2/2m)Q_{n,l}^2$. On the assumption that the l th partial wave has no bound states, but only the family of resonances at wave vectors $Q_{n,l}$, the Jost function can be expressed as an infinite product^{148,149}:

$$f_l(k) = [f_l(k)] e^{i\delta_l(k)} \\ = f_l(0) e^{-ikR_0} \prod_n \left(1 - \frac{k}{Q_{n,l}} \right) \quad (225)$$

Taking the imaginary part of the logarithm of this expression, we obtain a simple series expansion of the phase shift. Using the symmetry property $Q_{-n,l} = Q_{n,l}^*$, it is straightforward to show that

$$\frac{d\delta_l}{d\epsilon} = \frac{a}{v\epsilon} + \text{Im} \sum_{n>0} \left(\frac{\tilde{E}_{n,l}}{\epsilon} \right)^{1/2} \frac{1}{\epsilon - \tilde{E}_{n,l}} \quad (226)$$

The density derivative $d\delta_l/d\rho$ has a similar form. The various resonances in Eq. (226) give simple additive contributions that are formally independent.

From Eq. (222) for the density of states, we express the number and pressure of electrons as

$$N - N_B - N_0 = \frac{2}{\pi} \sum_{l=0}^{\infty} (2l+1) \\ \times \int_0^{\infty} f(\epsilon) \frac{d\delta_l}{d\epsilon} d\epsilon \quad (227a)$$

$$p - p_B - p_0 = -\frac{2}{\pi} \frac{\rho^2}{AM_p} \sum_{l=0}^{\infty} (2l+1) \\ \times \int_0^{\infty} f(\epsilon) \frac{d\delta_l}{d\rho} d\epsilon \quad (227b)$$

In Eqs. (227), $f(\epsilon)$ denotes the Fermi-Dirac distribution; N_B and p_B are possible bound-state contributions, and N_0 and p_0 are the ideal-gas terms associated with $g_0(\epsilon)$. Equation (227b) is derived (by an integration by parts) from the more general form

$$p = \frac{\rho^2 k T}{AM_p} \int_0^{\infty} \frac{\partial g(\epsilon)}{\partial \rho} \log |1 - f(\epsilon)| d\epsilon \quad (228)$$

appropriate to an arbitrary density-dependent one-electron density of states. AM_p is the atomic mass.

The conditions under which Eq. (228) gives the correct pressure, without an additional correction for the double-counting of the electron-electron interaction energy, are discussed in another context by V. Heine.¹⁵⁰ Briefly, that discussion indicates that omitting the double-counting correction compels us to reinterpret the density derivatives [of $g(\epsilon)$, $\delta(\epsilon)$, and E_n] as partial derivatives with respect to changes of the assumed positive charge distribution, neglecting the contribution due to changes in the self-consistent electron density.

Inserting the Regge series for the phase shift into these equations, one readily obtains

$$\delta N_{n,l} = 2(2l+1) \operatorname{Re} [F(\tilde{E}_{n,l})] , \quad (229a)$$

$$\delta p_{n,l} = 2(2l+1) \frac{\rho^2}{AM_p} \operatorname{Re} \left[F^{(1)}(\tilde{E}_{n,l}) \frac{d\tilde{E}_{n,l}}{d\rho} \right] \quad (229b)$$

for the population $\delta N_{n,l}$ and pressure $\delta p_{n,l}$ associated with the n th resonance of angular momentum l . In these equations, the generalized Fermi functions are defined by

$$F(\tilde{E}_{n,l}) = \frac{1}{i\pi} \int_0^\infty \left(\frac{\tilde{E}_{n,l}}{\epsilon} \right)^{1/2} \frac{f(\epsilon)d\epsilon}{\epsilon - \tilde{E}_{n,l}} , \quad (230)$$

$$\begin{aligned} F^{(1)}(\tilde{E}_{n,l}) &= \frac{1}{i\pi} \int_0^\infty \left(\frac{\epsilon}{\tilde{E}_{n,l}} \right)^{1/2} \frac{f(\epsilon)d\epsilon}{\epsilon - \tilde{E}_{n,l}} , \\ &= F(\tilde{E}_{n,l}) - \frac{i}{\pi} \sqrt{\frac{kT}{\tilde{E}_{n,l}}} F_{-1/2} \left(-\frac{\mu}{kT} \right) , \end{aligned} \quad (231)$$

where F_i is the usual Fermi-Dirac integral. Again, $f(\epsilon)$ is the Fermi-Dirac distribution of Eq. (219).

If the width of the resonance is small, so that $\tilde{E}_{n,l}$ is approximately real (that is, $\tilde{E}_{n,l} \approx E_{n,l}^0$) one finds

$$\operatorname{Re} [F(\tilde{E}_{n,l})] \approx \operatorname{Re} [F^{(1)}(\tilde{E}_{n,l})] \approx f(E_{n,l}^0) . \quad (232)$$

The Cauchy principal-value integrals that give $\operatorname{Im} [F]$ and $\operatorname{Im} [F^{(1)}]$ are not necessarily small in this case.

Equations (229) show how to resolve the total continuum pressure into additive resonance-state contributions to be added to a smooth background contribution. The resolution is unique and rigorous, and applies to any potential of ion-sphere type [one for which $V(r) = 0$ for $r > R_0$]. However, the derivation is limited to the one-electron approximation defined by Eq. (226).

For real-energy eigenstates, the eigenvalue $E_s = E_s(\rho)$, of degeneracy D_s , gives a delta-function contribution to the density of states:

$$g(\epsilon) \sum_s D_s \delta[\epsilon - E_s(\rho)] . \quad (233)$$

Inserting this form into Eq. (228), we find the bound-state pressures to be

$$p = \frac{\rho^2}{AM_p} \sum_s D_s f(E_s) \frac{dE_s}{d\rho} . \quad (234)$$

In the limit in which the resonance width approaches zero, the resonance pressure formula Eq. (229b) becomes identical with Eq. (234). This comparison proves the continuity of the pressure through the pressure-ionization transition^{151,152}; this transition occurs at the density determined by $E_{n,l}(\rho) = 0$, at least in the central-cell model.

The resonance pressure formula can be taken one step further by employing the resonance-state perturbation theory developed by Humblet,¹⁵³ Zel'dovich,¹⁵⁴ and More.^{155,156} According to this theory,^{157,158} the first-order change in the complex resonance eigenvalue produced by a perturbation potential $\delta V(r)$ is

$$\delta \tilde{E}_{n,l}^{(1)} = \langle \phi_{n,l} | \delta V | \phi_{n,l} \rangle = \int_0^\infty \phi_{n,l}^2 \delta V(r) dr , \quad (235)$$

where $\phi_{n,l} = N_{n,l} f(Q_{n,l}; r)$ is the resonance-state wave function. The coefficient $N_{n,l}$ is

$$N_{n,l} = \sqrt{\frac{f_l - Q_{n,l}}{if_l(Q_{n,l})}} , \quad (236)$$

and the integral that gives $\delta \tilde{E}_{n,l}^{(1)}$ converges as long as $\delta V(r)$ vanishes beyond $r = R_0$ (this is true despite the fact that the wave functions $\phi_{n,l}$ grow asymptotically with radius). With this expression for the change in eigenvalue, one obtains

$$\begin{aligned} \delta p_{n,l} &= 2(2l+1) \frac{\rho^2}{AM_p} \\ &\times \operatorname{Re} \left[F^{(1)}(\tilde{E}_{n,l}) \left\langle \tilde{\phi}_{n,l} \left| \frac{\delta V}{\delta \rho} \right| \phi_{n,l} \right\rangle \right] , \end{aligned} \quad (237)$$

which gives the resonance-state pressure as an expectation value of the change in potential associated with compression.

An explicit pressure-ionization model is required for the application of the screened-hydrogenic ionization model to dense plasmas capable of pressure ionization. Equations (229) provide a sound, fundamental basis for this application,

and appear to differ from the approximation usually employed in laser-fusion target calculations.¹⁵⁹

Author: R. M. More

Convergence Acceleration for Multigroup Radiation-Matter Coupling Iteration

An iterative scheme is used in LASNEX to solve the coupled multigroup radiation-diffusion and electron-temperature equations.¹⁶⁰ In the iteration, two different regular matrix splittings are used alternately. Generally this scheme converges rapidly, requiring fewer than five iterations to reach acceptable accuracy. In certain circumstances, however, particularly when the time step is large and the rates associated with transport and radiation-matter coupling are similar, the convergence can be quite slow. We have developed a convergence-acceleration scheme for use in these unusually stiff cases.

Consider a linearly convergent iteration

$$T^{n+1} = L(T^n), \quad (238)$$

where the T^n are successive iterates toward the solution of some equation. In most textbook treatments of convergence acceleration, L is a nonlinear scalar operator. For our problem L becomes a linear or nonlinear matrix operator, and T becomes a vector.

We first describe the scalar case. For example, to solve

$$\sin x - 2x + 0.1 = 0, \quad (239)$$

one might use $L(x) = (\sin x + 0.1)/2$. (In this case Newton iteration, with

$$L_N(x) = x - \frac{\sin x - 2x + 0.1}{\cos x - 2}, \quad (240)$$

is superior, because it is quadratically convergent.) Defining Y to be the change in T due to a single application of the iteration,

$$Y^n \equiv Y(T^n) \equiv L(T^n) - T^n, \quad (241)$$

we see that the exact solution T_e satisfies

$$Y(T_e) = 0. \quad (242)$$

The term "fixed-point iteration" is thus applied to iterations of this type,¹⁶¹ because T_e is a stable fixed point of the mapping defined by Eq. (238).

If we define an error $E^n \equiv T^n - T_e$, linear convergence implies that

$$E^{n+1} \approx L'(T_e)E^n; \quad (243)$$

that is, the error changes by a constant factor L' with each iteration. (The error decreases in each iteration if $L' < 1$; it changes sign if $L' < 0$.) If we subtract Eq. (243), with $n \rightarrow n-1$, from itself, we see also that $Y^{n+1} \approx L'Y^n$. Figure 3-48 shows a typical $L(T)$ and the corresponding $Y(T)$.

Classical acceleration schemes are closely related to the solution of Eq. (242) by Newton iteration. The iteration is initialized using $T^1 = L(T^0)$; for the next iteration, however, the secant formula (the usual two-point approximation to Newton's method) is employed. We wish to achieve $Y^2 = 0$, so [assuming that $Y(T)$ is a straight line] we solve the equation

$$0 = Y^2 \text{ (straight line)} \\ = \frac{Y^1 - Y^0}{T^1 - T^0} (T^2 - T^1) + Y^1 \quad (244)$$

for T^2 , with the result

$$T^2 = T^1 - Y^1 \frac{T^1 - T^0}{Y^1 - Y^0}. \quad (245)$$

In general, at iteration level n , this becomes

$$T^{n+1} = T^n + \alpha^n [L(T^n) - T^n], \quad (246)$$

where the acceleration (or "over-relaxation") parameter α^n is given by

$$\alpha^n = \frac{T^n - T^{n-1}}{[L(T^{n-1}) - T^{n-1}] - [L(T^n) - T^n]}. \quad (247)$$

The acceleration parameter α^n is thus the same finite-difference approximation to $-1/Y'$ that would appear in a secant

Fig. 3-48. (a) Dependence of $L(T^n)$ on T^n for a typical scalar fixed-point iteration; note the slow convergence. (b) Dependence of Y on T ; T^2 generated using secant iteration.

iteration. We may also write

$$\alpha^n = \frac{\alpha^{n-1}}{1 - Y^n/Y^{n-1}} \quad (248)$$

When this procedure is applied to a precomputed sequence of iterates of Eq. (238), in order to obtain a faster-converging sequence, it is called "Aitken's Δ^2 process"; when the procedure is applied concurrently with the calculation of the iterates, but every other step (i.e., α^n is set to unity when n is even), it is called "Steffenson acceleration." In a typical implementation,¹⁶¹ α is initially unity and is only set to the value given by Eq. (247) or (248) after the ratio of errors E^{n+1}/E^n at successive iterates has become constant (within about 10%) at a value < 1 .

For the present application, the scalars T , Y , and E become the vectors $\mathbf{T}$, $\mathbf{Y}$, and $\mathbf{E}$, and $L(T)$ becomes the matrix operation $\mathbf{L}\mathbf{T}$ ($\mathbf{L}$ may depend weakly on $\mathbf{T}$ in some cases). The solution T_e satisfies the relation

$$\|\mathbf{Y}(\mathbf{T}_e)\| = 0 \quad (249)$$

where the double vertical bars denote the L_2 norm (length) of the vector inside. This motivates the generalization of Eq. (248) used here:

$$\alpha^n = \frac{\alpha^{n-1}}{1 \pm \frac{\|\mathbf{Y}^n\|}{\|\mathbf{Y}^{n-1}\|}} \quad (250)$$

The sign indeterminacy in Eq. (250) arises because the lengths $\|\mathbf{Y}^n\|$ convey no information about the direction of $\mathbf{Y}^n$, much as if we had replaced the Y^n of Eq. (248) with their absolute values $|Y^n|$. To determine the correct sign, we compute the inner product

$$S \equiv \mathbf{Y}^n \cdot \mathbf{Y}^{n-1} \quad (251)$$

When $S > 0$, the negative sign is chosen in Eq. (250), corresponding to the case $Y^n/Y^{n-1} > 0$ in Eq. (248). When $S < 0$, the positive sign is chosen, corresponding to $Y^n/Y^{n-1} < 0$; in this case, we reduce α and (if $\alpha^{n-1} = 1$) "under-relax," because the iteration is oscillating.

Other generalizations of Eq. (248) are possible: one attractive one might use the information in the magnitude of S :

$$\alpha^n = \frac{\alpha^{n-1}}{1 - \frac{2\mathbf{Y}^n \cdot \mathbf{Y}^{n-1}}{\|\mathbf{Y}^n\|^2 + \|\mathbf{Y}^{n-1}\|^2} \frac{\|\mathbf{Y}^n\|}{\|\mathbf{Y}^{n-1}\|}} \quad (252)$$

We may try this in the future.

We set $\alpha = 1$, corresponding to no acceleration, under any of the following conditions:

1. When $n = 0$ (the first iteration).
2. When convergence (based on unaccelerated quantities) has been

Fig. 3-49. Relative L_2 change in T iterates for Milne problem. The new acceleration scheme reduces the required number of iterations by a factor of .0.

declared—convergence with $\alpha > 1$ is likely to be nonmonotonic, and we do not want to risk making the answer worse.

3. When the final pass to calculate fluxes, and other diagnostic quantities is being performed—for the same reason.
4. When n is even—this Steffenson variation works better in practice.

We also check the size of the T_i changes in each zone j ; if any zone has experienced a change (up or down) by more than a factor of 10, we allow only under-relaxation. This ensures that all the T_i will be positive and serves to make the iteration more robust. Finally, the maximum allowable value of α is an input parameter; $\alpha_{\max} = 100$ seems to be a reasonable setting.

We do not wait for the ratio of errors to reach a constant value before accelerating. Such a delay might be expected to be beneficial, because we expect linear convergence in the limit of a large number of nonaccelerated iterations. In this limit, error eigenvectors associated with all but the dominant eigenvalue of L' have decayed, and L' acts like a scalar $L' < 1$. However, we have found that acceleration works well without such a delay.

A test problem that illustrates the convergence-acceleration scheme is the calculation of the steady state of a very simple

model stellar atmosphere, the Milne problem of a plane photosphere, in the diffusion approximation.^{162,163} A doubly infinite slab of matter with a gray opacity is open to vacuum on one end (optical depth $\tau = 0$) and is held at a fixed temperature on the other end ($\tau = 25$). The LASNEX time step is set very large, and the temperature at $\tau = 25$ is initialized to the fixed value.

Results are shown in Fig. 3-49, in which the abscissa corresponds to iteration number and the ordinate to relative L_2 change in temperature during the iteration, before acceleration is applied. The uppermost (no-acceleration) curve is smooth, indicating monotonic convergence of the fixed-point iteration. Adequate convergence requires about 2500 iterations. The middle curve corresponds to acceleration every iteration, with changes in α limited to a factor of 4 (larger limits led to instability); the curve is noisy, because the convergence is nonmonotonic.

The lowest curve in Fig. 3-49 corresponds to the generalized Steffenson iteration described above. The convergence is again nonmonotonic, but the oscillations are much smaller. The peaks appear to occur when the iteration is correcting a set of newly introduced error eigenvectors arising from the previous acceleration; presumably these have small eigenvalues and are quickly damped. The generalized Steffenson scheme reduces the number of iterations required for adequate convergence by a factor of 10. Of course, smaller reductions are to be expected in time-dependent problems with smaller time steps.

We plan to incorporate an improved convergence criterion based on energy conservation and/or changes in interzonal energy fluxes, and to experiment with alternative forms such as Eq. (252).

Author: A. Friedman

Energy Spectrum of Thermonuclear Neutrons

We have improved the accuracy of our theoretical model of the energy spectrum of neutrons produced by the DT thermonuclear reaction. Our previous model, which

assumed a Gaussian shape in neutron energy centered at 14 MeV, was adequate for estimating fuel-ion temperatures below 10 keV, but was not accurate enough to measure high ion temperatures or to simulate the neutron spectrum far from its peak. Several authors^{164,165} have worked on this problem, but they made unnecessary approximations that give poor results for neutron energies above 16 MeV.

We begin by deriving the energy spectrum of neutrons in the reference frame moving with the thermonuclear fuel. We then use this spectrum to produce the spectrum as seen by the detector. Although the discussion centers on neutrons from the DT reaction, the formulas are generally valid for all particles generated in two-particle-breakup thermonuclear reactions.

From Eq. 48 of Ref. 164, the neutron creation spectrum (normalized to unity) is

$$n(E) = \frac{2M}{\pi\sqrt{2}m_D m_T m_n \lambda} T^2 \langle \sigma v \rangle \times \int_0^\infty \frac{dF_r E_r \sigma(E_r) e^{-E_r/T}}{\sqrt{Q + E_r}} \times \left[e^{-\frac{M}{m_n T} (\sqrt{\lambda(Q + E_r)})^2} - e^{-\frac{M}{m_n T} (\sqrt{\lambda(Q + E_r)})^2} \right], \quad (253)$$

where

$$M = m_D + m_T$$

$$\lambda = \frac{m_n}{M}$$

$$E_r = \frac{m_D m_T}{2M} v_r^2 = \text{relative DT energy in cm system}$$

$$\langle \sigma v \rangle = \text{Maxwell-averaged reactivity at ion temperature } T.$$

Q and E are much larger than T and E_r , so we ignore the second exponential term and move the first exponential term outside the integral to obtain a spectrum of the form

$$n(E) = \frac{1}{\sqrt{4\pi} E_0 \alpha (1 + \beta)} e^{-\frac{1}{\alpha^2} [\sqrt{E/E_0} - (1 + \beta)]^2}, \quad (254)$$

where $E_0 = \lambda Q \approx 14$ MeV, and where we have again normalized to unity.

We now determine α and β by forcing $n(E)$ to reproduce the correct average values $\langle E \rangle$ and $\langle E^2 \rangle$. Multiplying Eq. (253) by powers of E and integrating, we obtain

$$\frac{\langle E \rangle}{E_0} = \left(1 + \frac{\langle K \rangle}{Q} \right) + \frac{3}{2} \left(\frac{m_n}{m_\alpha} \frac{T}{Q} \right), \quad (255a)$$

$$\frac{\langle E^2 \rangle}{E_0} = \left(1 + 2 \frac{\langle K \rangle}{Q} + \frac{\langle K^2 \rangle}{Q^2} \right) + 5 \left(\frac{m_n}{m_\alpha} \frac{T}{Q} \right) \times \left(1 + \frac{\langle K \rangle}{Q} \right) + \frac{15}{4} \left(\frac{m_n}{m_\alpha} \frac{T}{Q} \right)^2, \quad (255b)$$

where

$$\langle K^m \rangle = \frac{\int_0^\infty K^{m+1} \sigma(K) e^{-K/T} dk}{\int_0^\infty K \sigma(K) e^{-K/T} dk}. \quad (256)$$

These kinetic energy moments can be conveniently expressed in terms of derivatives of $\langle \sigma v \rangle$:

$$\langle K \rangle = T \frac{\partial \ln(T^{3/2} \langle \sigma v \rangle)}{\partial \ln T}, \quad (257a)$$

$$\langle K^2 \rangle = \langle K \rangle^2 + T \frac{\partial \langle K \rangle}{\partial \ln T}. \quad (257b)$$

Forming the same moments by integrating Eq. (254), we obtain

$$\frac{\langle E \rangle}{E_0} = (1 + \beta)^2 + \frac{3}{2} \alpha^2, \quad (258a)$$

$$\frac{\langle E^2 \rangle}{E_0^2} = (1 + \beta)^4 + 5(1 + \beta)^2 \alpha^2 + \frac{15}{4} \alpha^4. \quad (258b)$$

Equations (255) and (258) can now be solved for α^2 and β :

$$\alpha^2 = \frac{m_n}{m_\alpha} \frac{T}{Q} + \Delta^2 + O\left(\frac{T^3}{Q^4}\right), \quad (259a)$$

$$\beta = \delta - \frac{1}{2} \left(\frac{3}{2} \Delta^2 + \delta^2 \right) + O\left(\frac{T^3}{Q^3}\right), \quad (259b)$$

Fig. 3-50. Simulated neutron time-of-flight streak picture of a spherically symmetric thermonuclear explosion.

where

$$\delta = \frac{\langle K \rangle}{2Q} \quad (260a)$$

$$\Delta = \frac{\langle K^2 \rangle - \langle K \rangle^2}{2 \left(1 + \frac{\langle K \rangle}{Q} \right) Q^2} \quad (260b)$$

Our final solution for the neutron energy spectrum in the frame moving with the fuel is contained in Eqs. (254), (257), and (259).

Actual neutron detectors measure the time of arrival τ of the neutron. This time is affected by the neutron energy E , the

component u of the matter velocity toward the detector, the time t of neutron creation, and the source-to-detector distance d , and by neutron scattering in the fuel and the surrounding material. We have not yet attempted to simulate neutron scattering, but we include all the other effects in the time-of-flight spectrum

$$f(\tau) = n(E) \left| \frac{dE}{d\tau} \right| \quad (261)$$

where

$$E = \frac{1}{2} m_n \left(\frac{d}{\tau - t} - u \right)^2 \quad (262)$$

The detector response is obtained by integrating Eq. (261) over all space and time, weighting $f(\tau)$ by the neutron creation rate at each point. These formulas have been incorporated into LASNEX and into the TDG post-processing code, which produces two-dimensional pinex simulations. Figure 3-50 is a simulated neutron time-of-flight streak picture of a spherically symmetric thermonuclear explosion. The detector distance of 500 cm is great enough that the early neutron distribution (at the bottom of the picture) is determined primarily by their energy spectrum and thus forms a useful ion-temperature measurement. The late-arriving neutrons, at the top of the picture, were actually created later in the explosion process. The slight increase in target radius with time is due to hydrodynamic expansion during the burn.

Author: G. B. Zimmerman

1. 1982 *Laser Program Annual Report*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-50021-82 (1983), pp. 3-12 to 3-16.
2. K. O. Mikaelian, *Phys. Rev. Letters* **48**, 1365 (1982).
3. K. O. Mikaelian, *Phys. Rev. A* **26**, 2140 (1982).
4. K. O. Mikaelian, *Phys. Letters* **99A**, 46 (1983).
5. K. O. Mikaelian, *Phys. Rev. A* **28**, 1637 (1983).
6. Lord Rayleigh, *Scientific Papers* (Dover, New York, N.Y., 1900) vol. 2.
7. G. I. Taylor, *Proc. Roy. Soc. London, Ser. A* **201**, 192 (1950).
8. K. O. Mikaelian and J. D. Lindl, *Phys. Rev. A* **29**, 290 (1984).
9. R. Kidder, *Nucl. Fusion* **16**, 405 (1976).
10. J. Meyer-Ter-Vehn, *Nucl. Fusion* **22**, 561 (1982).
11. Ibid.
12. C. E. Max, C. F. McKee, and W. C. Mead, *Phys. Fluids* **23**, 1620 (1980).
13. C. E. Max, J. D. Lindl, and W. C. Mead, *Nucl. Fusion* **23**, 131 (1983).
14. J. H. Nuckolls, "The Feasibility of Inertial-Confinement Fusion," *Physics Today* **35** (9), 24 (1982).
15. 1982 *Laser Program Annual Report*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-50021-82 (1983), pp. 3-17 to 3-20.
16. 1982 *Laser Program Annual Report*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-50021-82 (1983), pp. 3-20 to 3-28.
17. 1980 *Laser Program Annual Report*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-50021-80 (1981), pp. 3-19 to 3-24.
18. 1982 *Laser Program Annual Report*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-50021-82 (1983), pp. 3-28 to 3-32.
19. R. O. Bangerter, *The U.S. Program in Heavy-Ion Fusion*, Los Alamos National Laboratory, Los Alamos, N. Mex., LA-UR-82-1192 (1982).
20. *A Proposed Program to Develop Linear Induction Accelerator Technology in Support of the National Accelerator Inertial Fusion Plan*, Lawrence Berkeley Laboratory, Berkeley, Calif., LBL PUB-5065 (1982).
21. J. W.-K. Mark et al., "Heavy-Ion Inertial Fusion: Suggested Experiments on Disk Heating and Beam Transport Using Accelerator Test Facilities," in *Proc. High-Power Beams 81*, H. J. Doucet and J. M. Buzzi, Eds. (Ecole Polytechnique, Palaiseau, France, 1981), p. 649.
22. J. W.-K. Mark, "Recent U.S. Target Physics Related Research in Heavy Ion Fusion," in *Proc. Symp. on Accelerator Aspects of Heavy Ion Fusion*, D. Boehne, Ed., Gesellschaft für Schwerionenforschung, GSI-82-8 (1982).
23. 1980 *Laser Program Annual Report*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-50021-80 (1981), pp. 3-19 to 3-24.
24. D. S. Bailey, Y. T. Lee, and R. M. More, "Ab Initio Calculations of the Charge State of Fast Ions Stopping in a Finite Temperature Target," in *Proc. Symp. on Accelerator Aspects of Heavy Ion Fusion*, D. Boehne, Ed., Gesellschaft für Schwerionenforschung, GSI-82-8 (1982).
25. J. A. Swegle, *Analytic Scaling Laws Related to Beam and Return Current Stability in Particle Beam Driven ICF Target Regions*, Sandia National Laboratory, Albuquerque, N. Mex., SAND-83-0168 (1983).
26. F. C. Young, D. Mosher, S. J. Stephanakis, S. A. Goldstein, and T. A. Mehlhorn, *Phys. Rev. Lett.* **49**, 549 (1982).
27. T. A. Mehlhorn, J. M. Peek, E. J. McGuire, J. N. Olsen, and F. C. Young, *Current Status of Calculations and Measurements of Ion Stopping Power in ICF Plasmas*, Sandia National Laboratory, Albuquerque, N. Mex., SAND-83-1519 (1983).
28. J. N. Olsen, T. A. Mehlhorn, D. J. Johnson, and P. L. Dreike, "Ion Stopping in Heated Targets," paper 1P21 presented at the Fifth International Conference on High-Power Particle Beams, San Francisco, Calif. (September, 1983).
29. J. S. de Groot, University of California, Davis, Calif., personal communication (1983).
30. 1982 *Laser Program Annual Report*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-50021-82 (1983), pp. 3-17 to 3-20.
31. W. B. Hermannsfeldt, *High Intensity Multi-Megajoule Heavy-Ion Accelerator: A Plan for Research, Design, Development and Demonstration*, Lawrence Berkeley Laboratory, Berkeley, Calif., Report HIFAN-158 (1980).
32. J. W.-K. Mark and S. S. Yu, *Bull. Am. Phys. Soc.* **26**, 916 (1981), and *Phys. Lett.* **92A**, 179 (1982).
33. G. F. Chew, M. L. Goldberger, and F. E. Low, *Los Alamos Lecture Notes on Physics of Ionized Gases*, Los Alamos National Laboratory, Los Alamos, N. Mex., LA-2055 (1955).
34. J. W.-K. Mark, R. O. Bangerter, W. M. Fawley, S. S. Yu, and D. L. Judd, *Some Proposed Disk Heating, Focusing and Beam Transport Experiments for Heavy-Ion Inertial Fusion Test Facilities*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-86189, Rev. 1 (1982).
35. J. W.-K. Mark, "Recent U.S. Target Physics Related Research in Heavy Ion Fusion," in *Proc. of the Symposium on Accelerator Aspects of Heavy Ion Fusion, Darmstadt, Germany (March 29-April 2, 1982)*, Gesellschaft für Schwerionenforschung, Darmstadt, Report 82-8 (1982), p. 454.
36. G. Krafft, *Usefulness of Collective Focusing for Final Focusing*, Lawrence Berkeley Laboratory, Berkeley, Calif., LBL-14800 (1982).
37. S. Robertson, *Phys. Rev. Lett.* **48**, 149 (1982).
38. 1982 *Laser Program Annual Report*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-50021-82 (1983), pp. 3-28 to 3-32.
39. R. H. Berman and J. W.-K. Mark, *Bull. Am. Astron. Soc.* **7**, 377 (1975), and *Astrophys. J.* **216**, 257 (1977).
40. R. H. Berman and J. W.-K. Mark, *Astrophys. J.* **231**, 388 (1979).
41. J. W.-K. Mark, *Soc. Indus. Appl. Math. (SIAM) J. Appl. Math.* **42**, 914 (1982).
42. H. L. Buchanan, Ph.D. thesis, University of California, Davis, Calif. (1982) (to be published).
43. W. A. Newcomb, *Phys. Fluids* **25**, 846 (1982).
44. J. G. Siambis, *Phys. Fluids* **22**, 1372 (1979).
45. J. W.-K. Mark, H. L. Buchanan, and S. S. Yu, *IEEE Trans. Nucl. Sci.* **30**, 2552 (1983).
46. W. M. Sharp, M. Lampe, and H. S. Uhm, *Phys. Fluids* **25**, 1456 (1982).
47. W. C. Mead et al., *Characteristics of Lateral and Axial Transport in Laser Irradiations of Layered-Disk Targets at 1.06- and 0.35- μ m Wavelengths*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-89335 (1983).

48. C. S. Liu, M. N. Rosenbluth, and R. B. White, *Phys. Fluids* **17**, 1211 (1974).
49. M. Mostrom, *Raman Scattnr Instability in a Nonuniform Plasma*, Ph.D. Thesis, University of California at Berkeley, Berkeley, Calif. (1977).
50. C. S. Liu and M. N. Rosenbluth, *Phys. Fluids* **19**, 967 (1976).
51. A. W. Simon, R. W. Short, E. A. Williams, and T. Dewandre, *Phys. Fluids* **26**, 3107 (1983).
52. M. N. Rosenbluth, *Phys. Rev. Lett.* **29**, 565 (1972).
53. Ibid.
54. K. Tanaka, L. M. Goldman, W. Seka, M. C. Richardson, J. M. Soutar, and E. A. Williams, *Phys. Rev. Lett.* **48**, 1179 (1982).
55. D. Phillion et al., *Phys. Fluids* **25**, 1434 (1982) and *Phys. Rev. Lett.* **49**, 1405 (1982).
56. K. Estabrook and W. L. Kruer, *Phys. Fluids* **2**, 1892 (1963).
57. S. E. Bodner and J. L. Eddleman, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-73378 (1971).
58. K. Estabrook et al., *Phys. Rev. Lett.* **45**, 1399 (1980).
59. C. Joshi, C. E. Clayton, and F. F. Chen, *Phys. Rev. Lett.* **48**, 874 (1982).
60. H. Figueroa et al., *Phys. Fluids* (to be published).
61. C. C. Shepard, G. Busch, R. R. Johnson, R. J. Schroeder, and J. A. Tarvin, *Bull. Am. Phys. Soc.* **28**, 1058 (1983).
62. A. T. Lin and J. M. Dawson, *Phys. Fluids* **18**, 201 (1975).
63. K. Nishikawa and C. S. Liu, Eq. (III-58) in *Advances in Plasma Physics* (Interscience, New York, N.Y., 1976), A. Simon and W. Thomson, Eds.
64. J. M. Manley and H. E. Rowe, *Proc. I.R.E.* **44**, 904 (1956).
65. L. V. Powers, G. R. Monroy, and R. L. Berger, *Nucl. Fusion* **19**, 659 (1979).
66. O. Willi, R. G. Evans, and A. Raven, *Phys. Fluids* **23**, 2061 (1980).
67. C. R. Gwinn et al., *Phys. Fluids* **26**, 275 (1983).
68. V. L. Ginzburg and V. V. Zheleznyakov, *Sov. Astron. A. J.* **2**, 653 (1958).
69. R. Shanny, J. M. Dawson, and J. M. Greene, *Phys. Fluids* **10**, 1281 (1967).
70. C. S. Liu and M. N. Rosenbluth, "Parametric Decay of Electromagnetic Waves into Two Plasmons and Its Consequences," *Phys. Fluids* **19**, 967 (1976).
71. A. Simon, R. W. Short, E. A. Williams, and T. Dewandre, "On the Inhomogeneous Two-Plasmon Instability," *Phys. Fluids* **26**, 3107 (1983).
72. A. B. Langdon, B. F. Lasinski, and W. L. Kruer, "Nonlinear Saturation and Recurrence of the Two-Plasmon Decay Instability," *Phys. Rev. Lett.* **43**, 133 (1979).
73. *Laser Program Annual Report 1977*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-50021-77 (1978), p. 4-49.
74. *Laser Program Annual Report 1979*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-50021-79 (1980), p. 3-25.
75. *1980 Laser Program Annual Report*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-50021-80 (1981), p. 3-30.
76. *1981 Laser Program Annual Report*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-50021-81 (1982), p. 3-30.
77. C. E. Max, E. M. Campbell, W. C. Mead, W. L. Kruer, D. W. Phillion, R. E. Turner, B. F. Lasinski, and K. G. Estabrook, *Scaling of Laser Plasma Interactions with Laser Wavelength and Plasma Size*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-88056 (1983).
78. W. C. Mead, E. M. Campbell, K. G. Estabrook, R. E. Turner, W. L. Kruer, P. H. Y. Lee, B. Pruett, V. C. Rupert, K. G. Tirsell, G. L. Stradling, F. Ze, C. E. Max, M. D. Rosen, and B. F. Lasinski, "Laser Irradiation of Disk Targets at 0.53 μ m Wavelength," *Phys. Fluids* **26**, 2316 (1983).
79. L. V. Powers and R. L. Berger, "Kinetic Theory of Two-Plasmon Decay," *Phys. Fluids* **27**, 242 (1984).
80. W. C. Mead, E. M. Campbell, W. L. Kruer, R. E. Turner, C. W. Hatcher, D. S. Bailey, P. H. Y. Lee, J. Foster, K. G. Tirsell, B. Pruett, N. C. Holmes, J. T. Trainor, G. L. Stradling, B. F. Lasinski, C. E. Max, and F. Ze, *Characteristics of Lateral and Axial Transport in Laser Irradiations of Layered-Disk Targets at 1.06- and 0.35- μ m Wavelengths*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-89335 (1983).
81. E. M. Campbell, M. D. Rosen, L. W. Phillion, R. H. Price, K. Estabrook, B. F. Lasinski, S. P. Obenschain, E. A. McLean, R. R. Whitlock, and B. H. Ripin, "Laser Plasma Coupling in Long-Pulse, Long Scale Length Plasmas," *Appl. Phys. Lett.* **43**, 54 (1983).
82. D. W. Phillion, E. M. Campbell, K. G. Estabrook, G. E. Phillips, and F. Ze, "High-Energy Electron Production by the Raman and $2\omega_p$ Instabilities in a 1.064- μ m Laser-Produced Underdense Plasma," *Phys. Rev. Lett.* **49**, 1405 (1982).
83. R. E. Turner, D. W. Phillion, B. F. Lasinski, and E. M. Campbell, "Half- and Three-Halves Harmonic Measurements from Laser-Produced Plasmas," *Phys. Fluids* **27**, 511 (1984).
84. H. Baldis and C. Walsh, "Growth and Saturation of the Two-Plasmon Decay Instability," *Phys. Fluids* **26**, 1364 (1983).
85. *Laboratory for Laser Energetics Review* **17** (Oct.-Dec., 1983) (to be published).
86. Y. Kato and K. Mima, "Random Phase Shifting of Laser Beam for Absorption Profile Smoothing and Instability Suppression in Laser Produced Plasmas," *Applied Physics B* **29**, 186 (1982).
87. R. H. Lehmberg and S. P. Obenschain, "Use of Induced Spatial Incoherence for Uniform Illumination on Laser Fusion Targets," *Optics Commun.* **46**, 27 (1983).
88. L. M. Gorbunov, *Zh. Eksp. Teor. Fiz.* **55**, 2298 (1968) [*Sov. Phys. JETP* **28**, 1220 (1969)].
89. J. Drake, P. K. Kaw, Y. C. Lee, G. Schmidt, C. S. Liu and M. N. Rosenbluth, *Phys. Fluids* **17**, 778 (1974).
90. W. M. Manheimer and E. Ott, *Phys. Fluids* **17**, 1413 (1974).
91. R. Z. Sagdeev, D. W. Forslund, and A. Galeev, "Report of the Working Group on Stimulated Backscatter," *Comments Plasma Phys.* **2**, 97-108 (1976).
92. K. G. Estabrook and W. L. Kruer, "2-D LASNEX Ray Trace Studies of Thermal Self-Focusing in Cylindrical Geometry," *Proceedings of 13th Annual Anomalous Absorption Conference* (Laser-Plasma Research Laboratory,

- University of Alberta, Edmonton, Alberta, Canada, 1983), A. A. Offenberger, Ed.
93. R. S. Craxton and R. L. McCrory, "The Hydrodynamics of Thermal Self-Focusing in Laser Plasmas," *Proceedings of 13th Annual Anomalous Absorption Conference (Laser-Plasma Research Laboratory, University of Alberta, Edmonton, Alberta, Canada, 1983)*, A. A. Offenberger, Ed.; and Laboratory for Laser Energetics, University of Rochester, Rochester, N.Y., URLLE Report No. 144 (1983).
 94. M. J. Herbst, J. A. Stamper, R. H. Lehmberg, R. R. Whitlock, F. C. Young, J. Grun, and B. H. Ripin, U.S. Naval Research Laboratory, Washington, D.C., NRL Memorandum Report 4983 (1981); and *Proceedings of the 1981 Topical Conference on Symmetry Aspects of Inertial Fusion Implosions* (U.S. Naval Research Laboratory, Washington, D.C., unpublished), S. Bodner, Ed.
 95. A. J. Palmer, *Phys. Fluids* **14**, 2714 (1971).
 96. F. W. Perkins and E. J. Valeo, *Phys. Rev. Lett.* **32**, 1234 (1974).
 97. M. S. Sodha, A. K. Ghatak, and V. K. Tripathi, *Prog. Opt.* **13**, 169 (1976).
 98. M. S. Sodha, R. K. Khanna, and V. K. Tripathi, *Opto-Electronics* **5**, 533 (1977).
 99. V. K. Tripathi and L. A. Pitale, *J. Appl. Phys.* **48**, 3288 (1977).
 100. A. B. Langdon and B. F. Lasinski, *Phys. Rev. Lett.* **34**, 934 (1975).
 101. C. E. Max, *Phys. Fluids* **19**, 74 (1976).
 102. Kent Estabrook, *Phys. Fluids* **19**, 1733 (1976).
 103. C. E. Max, J. Arons, and A. B. Langdon, *Phys. Rev. Lett.* **33**, 209 (1974).
 104. R. A. Haas, M. J. Boyle, K. R. Manes, and J. E. Swain, *J. Appl. Phys.* **47**, 1318 (1976).
 105. D. S. Bailey, *Ray Trace Laser Sources in LASNEX*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-75881 (1974).
 106. Lyman Spitzer, *Physics of Fully Ionized Gases* (Interscience, New York, N.Y., 1962), chap. 5.
 107. B. I. Cohen, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-89871 (1983).
 108. N. M. Kroll, A. Ron, and N. Rostoker, *Phys. Rev. Lett.* **13**, 85 (1964).
 109. C. R. James and W. B. Thompson, *Can. J. Phys.* **45**, 1771 (1967).
 110. B. I. Cohen, A. N. Kaufman, and K. M. Watson, *Phys. Rev. Lett.* **29**, 581 (1972); M. N. Rosenbluth and C. S. Liu, *Phys. Rev. Lett.* **29**, 701 (1972).
 111. J. McGrath and R. L. Berger, in *1979 Annual Technical Report, KMS Fusion, Inc.*, Ann Arbor, Mich., KMSF-41046 (1979), p. 3-8.
 112. T. Tajima and J. M. Dawson, *Phys. Rev. Lett.* **43**, 267 (1979).
 113. Kent Estabrook and W. L. Krueger, *Phys. Fluids* **26**, 1888 (1983).
 114. J. R. Aibritton and A. B. Langdon, *Phys. Rev. Lett.* **22**, 1794 (1980).
 115. R. A. Haas, H. D. Shay, W. L. Krueger, M. J. Boyle, D. W. Phillips, F. Rainer, V. C. Rupert, and H. N. Kornblum, "Interaction of 1.0 μ m Laser Radiation with Planar Target," *Phys. Rev. Lett.* **39**, 1533 (1977).
 116. K. R. Manes, H. G. Ahlstrom, R. A. Haas, and J. F. Holzrichter, "Laser-Plasma Interaction Studies with High Power Glass Laser," *J. Opt. Soc. Am.* **67**, 717 (1977).
 117. *1981 Laser Program Annual Report*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-50021-81 (1982), p. 3-44.
 118. M. D. Rosen and K. G. Estabrook, *Analysis of the Z Dependence of Laser Generated Suprathermal Electron Temperature*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-82280 (1979).
 119. *Laser Program Annual Report 1979*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-50021-79 (1980) p. 3-25.
 120. E. J. Valeo and W. L. Krueger, *Phys. Rev. Lett.* **33**, 750 (1974).
 121. G. J. Morales and Y. C. Lee, *Phys. Rev. Lett.* **33**, 1534 (1974).
 122. B. Bezzerides and D. F. Du Bois, *Phys. Rev. Lett.* **34**, 1381 (1975).
 123. C. P. DeNeef and J. S. DeGroot, *Phys. Fluids* **20**, 1074 (1977).
 124. Peter DeNeef, *Plasma Phys.* **19**, 737 (1977).
 125. A. B. Langdon, B. I. Cohen, and A. Friedman, *J. Comput. Phys.* **51**, 107 (1983).
 126. *1982 Laser Program Annual Report*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-50021-82 (1983), pp. 3-56 to 3-59.
 127. J. Denavit, *J. Comp. Phys.* **42**, 337 (1981).
 128. B. I. Cohen, A. B. Langdon, and A. Friedman, *J. Comput. Phys.* **46**, 15 (1982).
 129. *1982 Laser Program Annual Report*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-50021-82 (1983), p. 3-53 to 3-56.
 130. R. Kalibjian and G. G. Peterson, "Analysis of Space Charge in Photonic Tubes," *J. App. Phys.* **54**, 4295 (1983).
 131. R. B. Spielman and J. P. Anthes, "A Study of Bare X-Ray Diode Saturation due to a High Fluence X-Ray Pulse," *AIP Conf. Proceedings No. 75* (American Institute of Physics, New York, N.Y., 1981), D. Attwood and B. Henke, Eds., p. 278.
 132. P. J. Ebert and A. F. Lauzon, "Vacuum Diode Detector for Measuring High Intensity Gamma-Ray Flux," *Rev. Sci. Instr.* **38**, 1747 (1967).
 133. T. C. Fry, *Phys. Rev.* **17**, 441 (1921).
 134. I. Langmuir, *Phys. Rev.* **21**, 419 (1923).
 135. C. B. Wheeler, *J. Phys. A* **5**, 1337 (1972).
 136. G. Craig and J. Denavit, "The Photoelectron Diode," Lawrence Livermore National Laboratory, Livermore, Calif., Internal Memoranda (parts I and II, November 1982; part III, January 1983).
 137. D. S. Bailey, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-75887 (1974).
 138. R. More, J. Barnes, and R. Cowan, *Bull. Am. Phys. Soc. II* **21**, 1153 (1976).
 139. J. Barnes, *Phys. Rev.* **153**, 269 (1967).
 140. C. Cranfill and R. More, Lawrence Livermore National Laboratory, Livermore, Calif., Internal Report LA-7313-MS (October, 1978).
 141. R. More, *Atomic Physics in Inertial Confinement Fusion*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-84991 (1981).
 142. R. More, in *Atomic and Molecular Physics of Controlled Thermonuclear Fusion* (Plenum, New York, N.Y., 1983), C. Joachain and D. Post, Eds., p. 399.

References

143. R. More, *Material: at Extreme Conditions*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-84115 (1980).
144. D. A. Liberman, *J. Quant. Spectrosc. Radiat. Transfer* **27**, 535 (1982).
145. C. Dharma-Wardana and F. Perrot, *Phys. Rev. A* **26**, 2096 (1982).
146. F. Perrot, *Phys. Rev. A* **26**, 1035 (1982).
147. C. Kittel, *Quantum Theory of Solids* (McGraw-Hill, 1962).
148. R. Newton, *J. Math. Phys.* **1**, 319 (1960).
149. T. Regge, *Nuovo Cimento* **8**, 671 (1958).
150. V. Heine, *Solid State Physics* (Academic Press, New York, N.Y., 1980), vol. 35, H. Ehrenreich, F. Seitz, and D. Turnbull, Eds., p. 114.
151. F. Rogers, H. Graboske, and H. DeWitt, *Phys. Lett.* **34A**, 127 (1971).
152. A. Petschek, *Phys. Lett.* **34A**, 411 (1971).
153. J. Humblet, *Mém. Soc. R. Sci. Lille* **12**, 9 (1952).
154. B. Ya. Zel'dovich, *Sov. Phys. JETP* **12**, 542 (1961).
155. R. More, *Phys. Rev. A* **4**, 1782 (1971).
156. R. More and E. Gerjuoy, *Phys. Rev. A* **7**, 1288 (1973).
157. T. Berggren, *Nucl. Phys. A* **389**, 261 (1982).
158. J. Bang et al., *Nucl. Phys. A* **339**, 89 (1980).
159. G. Zimmerman and R. More, *J. Quant. Spectrosc. Radiat. Transfer* **23**, 517 (1980).
160. 1982 Laser Program Annual Report, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-50021-82 (1983), pp. 3-68 to 3-70.
161. S. D. Conte and C. de Boor, *Elementary Numerical Analysis: An Algorithmic Approach* (McGraw-Hill, New York, N.Y., 1972), pp. 50-57.
162. Ya. B. Zel'dovich and Yu. P. Raizer, *Physics of Shock Waves and High-Temperature Hydrodynamic Phenomena* (Academic, New York, N.Y., 1966), pp. 157-164.
163. D. Mihalas, *Stellar Atmospheres* (Freeman, San Francisco, Calif., 1970), pp. 34-64.
164. M. M. R. Williams, "A Generalized Energy Exchange Kernel for Inelastic Neutron Scattering and Thermonuclear Reactions," *J. Nucl. Energy* **25**, 489 (1971).
165. H. Brysk, "Fusion Neutron Energies and Spectra," *Plasma Phys.* **15**, 611 (1973).

Weathering crystals growing
on the surface of glass laser
spheres.

Target Fabrication 4

Target Fabrication

Contents

Section Editor: A. M. Hawryluk

Introduction	4-1
Target Assembly and Microfabrication	4-2
Production	4-2
Precision Engineering Research Lathe	4-2
Cryogenics	4-4
Introduction	4-4
Novette Cryogenic Target Support System	4-5
Thermophysical Properties of Cryogenic DT	4-7
Developments in Fuel-Layering Techniques	4-8
Materials and Coatings	4-10
Introduction	4-10
Seeded Microspot X-Ray Diagnostic Targets	4-11
Rayleigh-Taylor Targets	4-13
Target and Material Measurements	4-15
Introduction	4-15
Measurement Development	4-15
Ion-Beam Drill	4-16
Ion-Beam Drilling Experiments on Metal Foils	4-18
References	4-22

Target Fabrication

Introduction

I. F. Stowers

The task of the Fusion Target Fabrication (FTF) Program is to make targets for current experiments and develop techniques for making targets for future experiments. In 1983, FTF did both these things, developing and producing targets that were shot on Novette and making significant gains in developing fabrication techniques that will be used to make targets for Nova.

During 1983, we supplied several hundred disk targets to measure plasma phenomena and laser-matter interactions, and about 30 targets meant to induce soft x-ray lasing. The soft x-ray laser targets were an entirely new type of target for FTF and the Inertial Confinement Fusion (ICF) Program. Much was learned in this series of experiments, which we anticipate will be resumed in the spring of 1984. Details of the design and construction of these targets were given in the 1982 Laser Annual.¹

Although Nova will not become operational until FY 1985, work is under way to develop fabrication techniques for implosion targets. The goal of many Nova experiments will be to demonstrate a target that is hydrodynamically equivalent to and scalable to a reactor target. The following are major unclassified technical accomplishments in the development of fabrication techniques.

Machining technology was advanced with the development of a small, precision, single-point diamond-turning lathe. This piece of equipment should help us turn out significantly more accurate and repeatable target parts.

Several cryogenic fuel-layering techniques have been evaluated, and the most promising will be selected in 1984 and intensely investigated through 1985. Efforts to improve our capability to model cryogenic processes resulted in our developing analytic expressions for the thermophysical properties of cryogenic deuterium-tritium (DT).

Hydrocarbon polymer (CH_x) is currently applied by plasma polymerization to glass microspheres as an ablator in Novette experiments. The ability to seed CH polymer with higher-Z materials may be essential in certain Nova targets. This year the seeded microspot diagnostic target was developed. These targets represent a unique fabrication capability that combines micromachining, plasma etching, and plasma polymerization.

Measuring capability goes hand in hand with the ability to fabricate precision targets. This year we extended our analysis capabilities in several key areas that include the measurement of the gas fill and surface regularity, and interferometry of transparent targets. The measurement group, with the ion-beam drill work, has advanced its capability for micromachining. This work shows that the ion beam can be a useful tool for machining a variety of materials and shapes.

These subjects are treated in more detail in the articles that follow.

Target Assembly and Microfabrication

Production

The production of unclassified targets for inertial-confinement fusion and weapons physics experiments was one of several major tasks of the Fusion Target Fabrication Group during 1983. Targets were produced for developing a soft x-ray laser, for studying the generation of large-scale plasmas, for evaluating the interaction of light with matter, and for evaluating fusion fuels and target materials at cryogenic temperatures.

The period from January to April was devoted to producing soft x-ray laser targets. Sealing the minute lasing channel to hold the lasing gas at 50 Torr continued to be the major hurdle in providing these targets (see 1982 Laser Annual¹). In addition to those targets constructed for experiments on Novette, we provided targets for similar experiments at the Laboratory for Laser Energetics at the University of Rochester. Of these targets, four were of a simpler one-sided design and were the only ones used for actual experiments at Rochester. In late October, we resumed the development of targets for x-ray laser experiments scheduled for Novette in April 1984. We spent November and December redesigning the target to eliminate deficiencies found in the original designs.

For plasma phenomena and laser-matter interaction experiments, we supplied several hundred disk targets of various materials. These targets were mainly produced by micromachining and microassembly techniques developed during the past few years (see 1979 and 1980 Laser Annuals^{2,3}) and include gold, beryllium, aluminum, and polymer disks. Many of these targets were supplied by the Fusion Target Assembly section of the Future Systems Group at Rockwell International's Rocky Flats Plant. Ball-in-film targets were fabricated for experiments meant to determine the structural properties of target assemblies and evaluate the properties of fusion fuel at cryogenic temperatures. The ball-in-film targets were constructed using procedures previously described.⁴

In an effort to improve our capabilities for micromachining, we undertook the development of a small, precision, single-point diamond-turning lathe (see "Precision Engineering Research Lathe" in this section). This new lathe, based on state-of-the-art hydraulic bearings and computer controls, will provide target parts on a production basis with previously unobtainable precision and accuracy.

A new development in disk targets is reported in the article "Seeded Microspot X-Ray Diagnostic Targets" also in this section. These targets were developed to provide more precise data on the expansion of plasmas. They represent a unique fabrication capability that makes use of micromachining, plasma etching, and plasma polymerization techniques.

Table 4-1 summarizes the target types and quantities that were produced during 1983.

Author: C. W. Hatcher

Precision Engineering Research Lathe

We have recently completed the design and fabrication of a small, two-axis, ultra-high precision, numerically controlled lathe for the turning of small contoured parts. Final compensation of the servos and testing of the software is now under way. Preliminary single-axis cutting tests, using a radiused diamond tool on a flat copper part, resulted in a surface finish of 0.1 to 0.2 μin . R_a (0.5 to 1 μin . peak-to-valley). The projected size contour accuracy of the machine is 10 μin . peak-to-valley. Called the Precision Engineering Research Lathe (PERL), this machine will both increase our production capacity and improve our accuracy capability.

The approach taken on the design of the PERL was fundamentally the same as that used on its predecessor, the Baby Optics Diamond Turning Machine (BODTM); in both cases we attempted to maximize machine stiffness and resonant modes and to minimize sensitivity to temperature fluctuations by (1) minimizing the size of the machine and (2) minimizing the structural loop of the machine by selecting a T-base configuration.

Target Assembly and Microfabrication

While the configurations of the two machines are similar, we attempted to improve the accuracy of the PERL by utilizing more accurate slides and higher performance servos. Some of the more significant design features of the PERL are shown in Fig. 4-1 and described below.

Slides and Slide Drives. To obtain the required accuracy and friction characteristics, the two linear slides are supported by 4-in. travel hydrostatic bearings developed at LLNL. Oil was used to pressurize the bearings to provide viscous damping. The deviation from straightness of travel for both slides has been demonstrated to be less than 8 μ in. in full travel.

Capstan drives were used to provide slide motions to minimize the Coulomb friction in the slide drives. A steel roller, preloaded against a steel drive bar by a hydrostatic (air) pad, is driven by a dc torque motor. The drive bar is attached to the slide by a pair of flexures in order to minimize the effects of misalignment on slide straightness. Velocity feedback is provided by a tachometer mounted integrally with the drive motor.

Position Feedback. Position feedback for both slides is provided by interferometers operating at 1 μ in. resolution. In the case of the x (tool) slide, the consequences of slide angular motion are minimized by locating the laser in line with the tool. On the z-slide, where the effects of slide angular motion on positioning accuracy cannot be measured directly, two interferometers, one on either side of the spindle, are used to provide angular motion data to the computerized numerical control system (CNC), which will correct for the consequent positioning error in real time.

Spindle and Spindle Drive. Both the radial and axial motion error values of the workpiece spindle are less than 1 μ in. To obtain the best surface finish possible, we minimized asynchronous (nonrepeatable) spindle motion by driving the spindle directly with a brushless dc torque motor. The motor rotor is cantilevered directly from the spindle rotor, and the motor stator is mounted from the spindle stator. This configuration also served to eliminate separate motor bearings, a potential source of vibration. The brushless dc torque motor was selected because with this configuration

Target type	Number of targets
X-ray laser	32
Ball-in-film	13
Ball-on-stalk	32
Disks	291

Table 4-1. Summary of target types and quantities produced in 1983.

most of the heat is generated in the stator, where it can be most easily removed. By enclosing the stator in a finned housing through which we pump temperature-controlled water, the temperature of the spindle and motor are maintained at $68 \pm 0.05^\circ\text{F}$.

Temperature Control. As mentioned previously, the size of the PERL was minimized to reduce sensitivity to variations in temperature. This, combined with direct cooling of the spindle, the only major heat source on the machine, allowed us to utilize an air-shower temperature control on the machine. This approach was a departure from previous designs for larger machines where liquid shower systems were utilized.

We have built an enclosure around the machine through which we recirculate 600 cfm of air. This system maintains the machine environment at $68 \pm 0.1^\circ\text{F}$. Thermal

Fig. 4-1. Precision Engineering Research Lathe (PERL) for making small contoured parts.

Fig. 4-3. Novette cryogenic system shown mounted on the target chamber.

target shots simulated. We describe the entire cryogenic support system, its operations and our full system tests in the following article ("Novette Cryogenic Target Support System").

We also continued our efforts to model the formation and behavior of condensed fuel layers in cryogenic targets. An important ingredient in such modeling is a set of thermodynamic equations of state for cryogenic DT. We developed such a set, as we report in the article "Thermophysical Properties of Cryogenic DT." These equations will be useful in designing systems to support Nova-class cryogenic laser fusion targets.

We have been exploring the possibility of supporting liquid layers on the interiors of spheres by applying a thermal gradient to them. A model of this type of target has been developed, and we describe it briefly in the article "Developments in Fuel-Layering Techniques." Finally, we have modified the limits of the fast-refreeze process of forming solid cryogenic layers of DT fuel in laser fusion targets. Most Nova-class targets will exceed these limits, which are

determined largely by gravitational flow of hydrogen condensates. We have therefore continued our investigation of formation of solid layers in free-fall, where the effects of gravity are absent. The status of the studies is also reported in the article "Developments in Fuel-Layering Techniques."

Author: T. P. Bernat

Novette Cryogenic Target Support System

The system we began developing several years ago to support experiments with cryogenic laser fusion targets has now been completed and implemented on Novette. The system includes the cryogenic apparatus and the Integrated Test System (ITS).

The Cryogenic Apparatus. The important parts of the cryogenic apparatus are a cryogenic target mount (pylon), a rapidly removable cryogenic protective shroud, control systems, a heating laser for applying heat pulses to targets, a viewing system, and an interface point to the Novette timing and sequencing controls (see Fig. 4-2).

Fig. 4-4. The heating laser system provides automated focusing, steering, and safety interlocking for the heating and illuminating beams.

Figure 4-3 shows the system assembled into the laser target chamber.

The cryogenic hardware itself and its operation have been described in earlier *Laser Annuals*.⁵⁻⁷ The heating laser system for heat-pulsing cryogenic targets has not been so described. It is schematically shown in Fig. 4-4. This system allows remote pointing and focusing of the heating laser beam over a distance of greater than 1 m, the radius of the Novette target chamber. It also automatically protects the system against debris or optical damage at shot time. The viewing system was a target viewer salvaged from the old Shiva facility, but modified to have a higher numerical aperture for greater resolution. Backlighting was provided by a small He-Ne laser.

The electronic control system provides the means of controlling the temperatures of the pylon tip and the protective shroud by controlling cryogen flow rates. It also controls the flow of helium "exchange gas" into the shroud that provides thermal coupling to the target. The temperatures and exchange-gas flow-rate set points can be changed by commands given at a computer keyboard. The electronic system also monitors, displays, and logs all system data during an entire cryogenic run and provides

means of positioning the target and the shroud through stepping-motor-driven, four-axis positioners.

The retraction of the cryogenic shroud, the pulsing of the heating laser, and the closing of the protective shields were critically timed operations that had to be included in the Novette firing sequence. This was accomplished through an interface unit that was driven by timing signals provided by Novabus. The interface unit allowed us to locally set times with millisecond resolution for the cryogenic system operation without changing the timing signals from Novabus.

Integrated Test System. The complexity of our entire cryogenic support system, and the interface with the Novette firing sequence, made it necessary to test our equipment and simulate shots. The ITS was built to perform these functions (see Fig. 4-5). It was designed to be stable enough that targets could be characterized in place interferometrically. This required that the entire spaceframe be decoupled from floor vibrations by being lifted on air mounts.

We used the ITS to proof test and debug the cryogenic units and their controls. We measured the thermal histories of the cryogenic pylon and the shroud as they were

cooled to their lowest operating temperatures. We determined temperature stabilities, ability to change temperature set points, ability to control exchange-gas flow rates, interaction between exchange gas flow and system temperatures, liquid helium consumption, and many other details of the system performance.

We gained valuable experience in simulating shots of cryogenic targets. With a target mounted on the cryogenic pylon in the ITS vacuum chamber, we tested positioning ability of the stepping-motor controlled stages, the alignment of the shroud and pylon, our ability to make a seal between the shroud and pylon, firing the shroud removal system, and the heat pulsing sequence to be used on our Novette shots. These total system tests allowed us to move the entire cryogenic support system to the Novette facility and bring it to operational status in less than one week.

Author: T. P. Bernat

Major Contributors: D. H. Darling, J. I. Miller, B. Myers, P. S. Rasmussen, R. K. Reed, D. H. Roberts, J. J. Sanchez, and J. R. Severyn

Thermophysical Properties of Cryogenic DT

By computer modeling cryogenic processes involving DT, we can mitigate the need for costly experiments. To facilitate modeling we have developed analytic expressions (equations of state) to match recently compiled thermophysical properties of 25%D₂-50%DT-25%T₂, a common laboratory mixture. We are currently using these expressions in the ICA2CON code.

Accurate equations of state depend on reliable thermophysical data. Although DT has the greatest potential as a laser fusion fuel, until this year there had not been a comprehensive compilation of DT thermophysical data. This year P. C. Souers of LLNL compiled and compared DT data from research done in the United States, England, France, Japan, and the Soviet Union.⁸ He also calculated some of the thermophysical properties of DT from fundamental physics. Souers has developed a specific set of data for the 25/50/25 mixture.

Analytic expressions for thermal conductivity, specific heat, density, enthalpy, compressibility, and liquid viscosity are now available for the solid, liquid, and gas phases⁸⁻¹⁰ (as indicated in Fig. 4-6). These expressions for the 25/50/25 mixture are applicable in the temperature range from 4.2 to 100 K, at subcritical pressures.

Using this data, we generated algebraic equations of state. Particular attention was directed to equations to model the saturation curve. The saturation conditions can be determined from more than one of the independent variables. The equations of state are specifically intended for use with numerical algorithms that deal with condensation and solidification, where it is necessary to be able to identify a phase boundary. Saturation conditions can be approached from the liquid or vapor phases, but the equations of state are different in those phases. A similar statement can be made for the solid/liquid boundary. The probability

Fig. 4-5. The integrated test system allows complex experiments with cryogenic target support hardware.

Fig. 4-6. Analytic expressions of the P , v , T surface have been developed for DT at low temperature (4.2 to 190 K) and subcritical pressure.

of numerical instabilities due to use of the DT equations of state has been reduced by assuring that the same phase boundaries are identified by the equations in solid, liquid, or vapor phase.

Author: L. V. Griffith

Developments in Fuel-Layering Techniques

Solid layers of hydrogen fuels can be formed on the interiors of glass spherical shells by quickly refreezing the fuel after it has been warmed by a pulse of heat. We have described this technique extensively in previous *Laser Annuals*.^{11,12} As the amount of fuel contained in a target increases, however, the rapid isothermal refreezing technique breaks down because the characteristic fuel-freezing time is no longer negligible compared with the gravity-induced fuel-sagging time. This breakdown results in a fuel layer that is thicker at the bottom than at the top. Cryogenic targets with large fuel masses can only be formed by overcoming the effects of gravity. We are trying two approaches: applying a thermal gradient to a target and forming fuel layers in free-fall.

Thermal Gradient Approach. Workers at KMS Fusion, Inc. (KMS) first pointed out that the distribution of liquid hydrogen inside glass spheres could be affected by an anisotropic thermal environment.^{13,14} We have reported our own earlier qualitative investigations of this effect and the behavior of liquid layers on the interiors of spheres.¹⁵⁻¹⁷ However, none of the investigations, thus far, has provided clear-cut

guidelines on how to choose a set of predetermined control parameters that will work for a specific target. During the past year, we have extensively studied the behavior of liquid hydrogen isotopes (i.e., the fusion fuel liquid) inside a spherical-shell ICF target subject to a thermal gradient. Our purpose was both to assess the feasibility of, and formulate a recipe for, fabricating a uniform fuel layer by applying a thermal gradient across the target. Much of this work was carried out by K. Kim and his co-workers at the University of Illinois.

Unlike most other liquids, liquid hydrogen isotopes and their mixtures (henceforth, simply referred to as hydrogen) exhibit a vanishingly small contact angle on all substrate surfaces because of strong *van der Waals attractions*.^{18,19} Consequently, when a spherical-shell target containing hydrogen gas is cooled below the liquefaction point of hydrogen, the resulting liquid hydrogen inside the target spreads out and wets the entire target interior. Kim and Rieger, by setting up a nonuniform temperature field, were previously able to form a uniform liquid hydrogen layer inside a spherical target.²⁰ In published work,²¹ we described mechanisms by which nonuniform temperature effects compensate for gravity, thus resulting in a uniform liquid hydrogen layer inside an ICF target.

The nonuniform temperature effects included in our calculations are the effects of the surface tension gradient and the mass transfer across the liquid-gas interface, i.e., the condensation and evaporation. The net result is an expression for the temperature gradient at the target exterior that will sustain a uniform liquid layer of hydrogen isotopes inside a spherical ICF target. Missing in this theory, however, are the isotopic effects such as the differences in the thermodynamic properties and transport coefficients of the isotopic species. For these reasons, although it currently accounts for the behavior of single-component liquids, the theory is not capable of predicting the behavior of a mixture of hydrogen isotopes inside an ICF target.

In multicomponent liquids, differences in the vapor pressures of the various isotopic components result in a nonuniform distribution of isotopes throughout the entire liquid layer. This in turn gives rise to a gradient in the total interfacial tension. (Note

Fig. 4-7. Three possible liquid layer configurations (a), (b), (c) of a binary mixture of hydrogen isotopes inside a spherical-shell target.

that the values of the interfacial tension of the individual isotopic components are different even at the same temperature.)

Depending on the magnitude of the temperature gradient, the interfacial tension gradient induced by the component separation can easily be larger in magnitude than (but always opposite in direction to) the one that would have prevailed in the case of a single-component liquid. In a single-component liquid, the interfacial tension is highest at the coldest part of a target, since surface tension is a decreasing function of temperature.

These effects can be better understood by scrutinizing the behavior of a binary liquid mixture of hydrogen isotopes, for example, a DT mixture. There are three possible liquid layer configurations: the liquid layer can be thick at the bottom, uniform throughout, or thick at the top (see Fig. 4-7). In Fig. 4-7(a), since the bottom of the target is warmer than the top, there will, in general, be net evaporation at the bottom and net condensation at the top. More importantly, since deuterium has higher vapor pressure than tritium at a given temperature, more deuterium than tritium will evaporate at the bottom of the target, resulting in buildup of tritium in the bottom liquid layer. Similarly, there will be buildup of deuterium at the target top. This component separation, in turn, gives rise to a surface tension gradient that is directed downward. In this case, the surface tension gradient and gravity act on the same direction and pull the liquid to the bottom of the shell. If the temperature field is reversed, however, the surface tension gradient opposes gravity. For a temperature field chosen such that gravity and the surface tension gradient exactly cancel, a uniform layer will result, as in Fig. 4-7(b). Further increase of the temperature difference across the target can actually

Fig. 4-8. External temperature gradient required to maintain a uniformly thick liquid layer of 50% D₂-50% T₂ mixture vs liquid layer thickness, for a glass target of 100 μm outer radius and 5 μm wall thickness. T_{avg} is the temperature along the target equator.

overwhelm gravity, and cause a layer thicker at the top, as in Fig. 4-7(c).

To understand the behavior of a multicomponent liquid hydrogen inside a target subject to a temperature gradient, we have performed an extensive theoretical analysis. In particular, the degree of the component separation has been determined by solving the diffusion equations with the appropriate boundary conditions. One of the most important results of this analysis is the prediction of the value of the temperature gradient that is required to sustain a uniformly thick layer of multicomponent liquid hydrogen. The predicted temperature gradients obtained for a spherical-shell glass target containing a binary liquid mixture of 50% D₂ and 50% T₂ are plotted in Fig. 4-8 for different values of liquid layer thickness. The outer radius and wall thickness of the target are, respectively, 100 μm and 5 μm, and the target is maintained at a specified temperature (denoted as T_{avg}) along the equator. We have also determined the dependences of the thermal gradient on the target size for the same type of target containing a uniform liquid layer 6 μm thick (results are shown in Fig. 4-9). We are

Fig. 4-9. External temperature gradient required to maintain a uniform liquid layer of 50% D₂-50% T₂ mixture vs target outer radius, for a spherical glass shell 5 μ m thick containing a liquid layer 6 μ m thick.

Fig. 4-10. Schematic of the prototype free-fall system.

presently performing quantitative experiments on the thermal gradient effect at the University of Illinois.

Free-Fall Fabrication of Cryogenic Targets. Our second approach to offsetting the effects of gravity on the fuel layers is to form targets in free-fall, where gravity is in effect absent. We have constructed a system for testing the feasibility of this idea.

A schematic of this free-fall target-freezing system is shown in Fig. 4-10. In this system, a DT-filled target is mounted in a target freezing cell, which is first filled with helium exchange gas to a desired

pressure and then sealed. In a typical target-freezing experiment, the following sequence occurs: First, the target-freezing cell is brought into good thermal contact with a cryostat at the top of a vertical drop tower. Since the cryostat is maintained at a temperature well below the freezing point of DT, the DT gas inside the target freezes, more than likely, into a nonuniform layer.

Second, a heating laser pulse is introduced into the cell to vaporize the DT condensate but not to substantially heat the target environment.

Third, a locking device that keeps the freezing cell in good thermal contact with the cryostat is quickly deactivated to cause the cell to free-fall through the drop tower. During this free-fall, because the thermal mass of the cell is much larger than that of the target and because the exchange-gas pressure has been chosen to rapidly conduct heat, the evaporated DT mixture inside the target refreezes into a solid layer. The layer should now be uniform because the refreezing is done under zero-gravity condition. The drop tower is designed with sufficient height to allow for the DT mixture to completely refreeze during the target free-fall.

Fourth, at the bottom of the drop tower is another cryostat (similar to the one at the top) that "catches" the free-falling target-freezing cell and retains the integrity of the frozen DT layer inside the target.

By inspecting the target at the top cryostat where the DT condensate is supposed to be nonuniform and at the bottom cryostat where the DT condensate has been formed free from gravity effects, we can evaluate this target-freezing scheme.

Author: T. P. Bernat

Major Contributors: J. Buzzard, K. Kim, and L. Mok (University of Illinois); and D. Darling (LLNL)

Materials and Coatings

Introduction

Because target surfaces are converted to plasmas by high-energy laser pulses, it is crucial to know their temperature and density as the plasma develops. A first step is

to characterize thick, free-standing CH films that contain a spot seeded with sulfur, silicon, or chlorine. To make microspot targets, we developed new film-forming, seeding, release, stress-control, and micropatterning methods. Characterization of flat, 2- μm -thick CH films seeded with 2 to 3 at.% of sulfur, silicon, or chlorine was done in close conjunction with the Target and Material Measurements Group.

Theoretical attention has also been directed toward possible Rayleigh-Taylor (R-T) instabilities during compression of targets (100 to $1000\times$ initial density). To check and calibrate computer predictions of R-T response in target materials, we designed slab-type targets of CH plastic and glass with one well-characterized sinusoidal ripple surface. Sinusoidal grooves were diamond turned in copper to make the molds for the slab targets. Release coatings, etching, and non-intrusive characterization had to be developed. Slabs with ripple periods of 10 through 50 μm , 1 to 5 μm deep, were made for Novette and Nova experiments.

Author: D. E. Miller

Seeded Microspot X-Ray Diagnostic Targets

A new target to measure the temperature and density-time history of a laser-induced plasma was proposed by the Fusion Experiments Group. The specifications for the target called for a freestanding hydrocarbon (CH) film with a sharply defined spot (microspot) in the center seeded with either silicon, sulfur, or chlorine. The target film was fabricated using a three-step procedure consisting of depositing the hydrocarbon film, defining the microspot, and depositing a seeded spot through a mask. In the final assembly step, the film containing the microspot was mounted over a 1.5-mm-diam hole in a support. The support was either a plastic ring or a copper foil electroplated with 3 μm of gold. The remainder of this article will discuss the technologies developed to fabricate the target.

Hydrocarbon Film Deposition. In the completed microspot target, the seeded microspot is supported in a hydrocarbon film of equal thickness. Three distinct technologies for depositing a hydrocarbon film

were investigated: plasma polymerization, Parylene coating, and solution casting. In each of the deposition technologies, the CH film is deposited onto a flat substrate. Since a 2- μm -thick, freestanding film is needed, it is necessary to provide a release layer. Our first attempts to form freestanding films used water-soluble polyacrylic acid (PAA) chips as coating substrates. After the PAA chip was coated by plasma polymerization, it was floated on water to dissolve the substrate. Though successful, this technique was slow because of the slow rate of dissolution of the PAA. In addition, it was difficult to make flat PAA chips. We found that a layer of potassium chloride (KCl) approximately 0.05 μm thick, evaporated onto a glass slide, provided a flat substrate that released easily as the film was floated onto water. Salt-coated glass slides were used as releasable substrates for each of the CH film deposition technologies investigated.

The first CH film-forming technology we investigated was our established plasma polymerization process, which is used to coat microshell targets. We found that film intrinsic stress caused the CH film to release as a wrinkled layer on the KCl-coated glass slides. Without the KCl layer, we could not strip the layer from the glass. We tried to reduce the intrinsic stress by changing the process conditions of the helical resonator coater. Since we do not have facilities to measure film intrinsic stress, we were unable to systematically find conditions that reduced stress. However, we were able to compare results from the four helical resonator coaters we have in the Target Fabrication Group. We found that two of the coaters produced a satisfactory low-stress coating. These systems were then used to deposit all the plasma polymerized CH films used for the microspot target.

The second CH deposition technique we tried was Parylene deposition on KCl-coated glass. Parylene was deposited from the gas phase as a uniform, stress-free film. However, Parylene coatings were not a practical alternative because of delays and insufficient accuracy in the coating thickness.

In the third technique, solution casting of polymer films, the film is cast from a solution of polystyrene in dichloromethane. The solution is pumped into a conical reservoir containing a KCl-coated glass slide. The

Fig. 4-11. Microspot targets were fabricated by two alternative technologies. The technologies differed in the CH hole defining step. On the left the hole is defined by masking. On the right the hole is etched.

solution is drained out of the reservoir at a controlled rate leaving behind a smooth polymer film deposited on the glass slide. The polymer coating thickness is controlled by two variables, the solution concentration and the draining rate of the solution. Increasing the solution concentration increased the thickness. Decreasing the draining rate decreased the film thickness because of greater refluxing action of the solution on the forming film. The polystyrene films were stress-free and varied slightly in thickness along the length of the glass slide because of the changing liquid level descent rate in the conical vessel. The film thickness was measured and the back of the slide was marked at the desired $2\text{ }\mu\text{m}$ thickness.

Microspot Definition. To keep the film thickness constant, the seeded spot must be deposited in a hole in the CH film. Creating this hole was the single most difficult step in fabricating the microspot target. A large number of techniques for removing the CH were investigated, including mechanical drilling, laser drilling, ion milling, UV decomposition, reactive-ion etching (RIE) in a commercial plasma stripper, reactive-ion etching in a helical resonator discharge, punching, and masking. All of the techniques showed some promise; however, only the last three techniques produced uniform, burr-free, sharply defined holes.

To punch a hole, the $2\text{-}\mu\text{m}$ -thick film was stripped from the substrate by floating on water and then mounted on a steel washer. The restanding film was supported on a smooth piece of aluminum foil. The

supported film was then put in a press and the cutting tool pressed into the film. Cutting a hole in the $2\text{-}\mu\text{m}$ -thick films required that an experienced operator apply exactly the correct pressure to the cutting tool. Too little pressure resulted in the central spot remaining in the film. Excessive pressure indented the film and caused the thin film to crack in a concentric circular pattern. Punching holes required an extra film transfer and skill to achieve a burr-free hole. Because of the additional time-consuming handling and the skill required to apply the proper pressure to cut the film, punching was abandoned in favor of the more direct etching and masking techniques of hole definition.

Reactive-ion etching using a commercial apparatus appeared successful initially, and much time was spent developing masking techniques and studying the effect of discharge power on etching rate. This form of RIE undercut the mask, creating a thin region around the central spot. The prevalence of undercutting was not recognized until the microspot targets were carefully characterized. We found that when the mask used to define the hole did not make uniform, intimate contact with the film, undercutting occurred. Undercutting could extend inward up to $100\text{ }\mu\text{m}$ under the edge of the mask. Undercutting during RIE became our major target loss mechanism. However, RIE also offered us a simple, self-aligning technology to fabricate high quality microspot targets, and consequently we devoted great attention to solving its problems. An eventual solution was found in using our own plasma polymerization equipment as an etcher.

Etching a hole in a CH film through a gold washer mask produced well-defined, sharp-edged holes when a helical resonator discharge was used. By feeding $2\text{ cm}^3/\text{s}$ of oxygen at 75 mTorr pressure and 20 W power into our plasma polymerization coating apparatus we were able to remove a $2\text{-}\mu\text{m}$ -thick polymer film in 15 minutes. The helical resonator discharge has shown less tendency to undercut the mask than the commercial reactive-ion etching apparatus. Using etching technology, we fabricated microspot targets using the procedure shown in Fig. 4-11.

The other successful fabrication technique followed the procedure shown under "masking" in Fig. 4-11. A gold disk was

used as a mask to keep the CH film from depositing where the hole was to be. After the CH layer was deposited, the disk mask was removed and a washer-shaped mask with an inside diameter matching the removed disk was aligned with the defined hole. The film with the aligned mask was at this point ready to have the seeded spot deposited in the defined hole.

Microspot Development. Deposition of the microspot required development of both new materials and masking technology. In the past, we have been able to seed polymers with chlorine, bromine, and iodine. Our goal then was to achieve a high seeding composition for high density materials. For the microspot target, new seed elements (silicon, sulfur, and chlorine) were required at a relatively low (2 at.%) concentration. Because of the low concentration, even seeding with chlorine, which we did last year, required new feedstock gases.

The feedstocks were selected on the basis of volatility, atomic composition of the seed element, and purity. If the feedstock is to be flowed into the plasma polymerization process, the material must have a vapor pressure of approximately 1 Torr. To limit the seed atom composition to 2 at.%, it is desirable that the seed atom be a minor constituent of the feedstock. The feedstocks we selected are tetramethyl silane, $(\text{CH}_3)_4\text{Si}$, for silicon seeding, dimethyl sulfide, $(\text{CH}_3)_2\text{S}$, for sulfur seeding, and 2-chloro-2-methyl propane, $(\text{CH}_3)_3\text{CCl}$, for chlorine seeding.

We successfully prepared silicon- and sulfur-seeded polymers up to 3 at.% using tetramethyl silane and dimethyl sulfide, respectively, as feed gases. Seeding with chlorine was demonstrated although compositions were not quantitatively measured. Figure 4-12 shows the variation of silicon and sulfur content with gas feed rate into the plasma reaction vessel. Silicon and sulfur composition were measured by x-ray fluorescence.

The seeded microspot was deposited after aligning a gold mask to the hole in the hydrocarbon polymer film. The assembly was transferred to a plasma coating system where the seeded polymer was deposited through the hole in the mask. Adhesion between the seeded microspot and the polymer support film was excellent. Uniformity of the microspot is shown by the interferogram in Fig. 4-13. In the 20- μm annular region near the edges of the mask the

Fig. 4-12. The atomic concentration of seed atom was measured by x-ray fluorescence as a function of gas feed rate. For sulfur seeding the gas used was dimethyl sulfide. For silicon seeding the gas used was tetramethyl silane.

2.0- μm -thick CH film; 2.0- μm -thick, 300- μm -diam microspot

Fig. 4-13. The thickness uniformity of microspot targets is shown interferometrically. The same order of interference passed through both the CH film and the seeded spot.

microspot is thinner by a maximum of 0.5 μm , which was determined to be satisfactory. The film was next floated on water and mounted on a gold or polyethylene supporting ring.

In summary, microspot targets were successfully fabricated for plasma physics experiments. The microspot, seeded with silicon or sulfur from 100 to 300 μm in diameter and from 1 to 3 μm thick, was deposited by plasma polymerization into a hole in a hydrocarbon polymer film of equal thickness.

Author: S. A. Letts

Major Contributors: R. A. Corley, T. M. Tillotson, L. A. Witt, J. F. Kimmons, D. J. Lenz, V. K. Chen, C. M. Mazuch, C. S. Cordial, P. L. McCarthy, G. J. Greiner, C. M. Walkup, and R. W. Ryon

Rayleigh-Taylor Targets

Because of Rayleigh-Taylor (R-T) instabilities all surfaces of targets for implosion experiments must be exceptionally smooth

Fig. 4-14. Definitions of λ and η for machined-in target defects.

Fig. 4-15. Targets for tests of Rayleigh-Taylor instability were made by solution casting films on precision machined molds.

(see Section 3, "Evolution of Rayleigh-Taylor 1: stabilities in Multilayered Fluids"). The dreaded "100 Å" surface smoothness requirement is, in part, expensive insurance against R-T instabilities. Can we survive a "200 Å" surface smoothness? Are there approaches that can minimize the R-T instability effects? The present R-T experiments are designed to answer some of those questions.

Our approach to the problem has been to fabricate targets with defects so that their performance can be compared with theoretical predictions. Since it is more difficult to make theoretical calculations for a point defect, our R-T experiments are designed to test continuous, single-frequency sinusoidal defects of various wavelengths and peak-to-valley heights.

These targets are flat disks and can be made with sinusoidal defects of wavelengths (λ) of 10, 20, 30, and 50 μm , and peak-to-valley heights (η) of 0.2, 0.5, 1, 2, 3, 4 and 5 μm (see Fig. 4-14). Various combinations of λ and η are needed, and in all cases a surface smoothness of 500 Å (0.05 μm) is desired. The average target thickness depends on the target material, ranging from 20 μm for glass to 40 μm for hydrocarbon (CH) polymers. The range of target materials is confined to materials similar to those used in ICF targets. Glass, CH polymer, and polyvinyl chloride (PVC) were used in our initial experiments.

Fabrication Technique. Based on the physical dimensions and the materials specified for the R-T targets, we narrowed the fabrication options to three approaches that we investigated: (1) direct precision machining of a film, (2) ion milling through a mask, and (3) replication from a precision-machined mold. All three approaches are feasible, but we chose the replication technique because it gave us the most flexibility. Different molds can be produced ahead of time, and any R-T target can be replicated quickly from the existing molds. Fast response time is important because the surfaces of some of the target materials degrade with time. Also, it is more economical to fabricate one high quality mold than to fabricate many high quality targets.

The molds, 2-cm-diam by 0.6-cm-thick oxygen-free copper disks with either gold or copper electroplated surfaces, were all turned on the No. 2 precision diamond turning machine. Patterns cut on the

electroplated materials gave the smoothest surfaces. Gold, being inert, is the ideal mold material. The copper mold, which can be dissolved easily without affecting the target material, is useful if there are serious problems in releasing the R-T target from the mold.

The target surface can be replicated by various methods. Thermal pressing would have been ideal except that the process may damage the mold, negating the advantage of replication. Fortunately, the R-T target materials under consideration can all be dissolved in solvents. Thus the targets can be replicated by spreading a solution of the target material onto the mold and removing the film from the mold when the solvent has evaporated. To spread a uniformly thick liquid layer on the 2-cm-diam surface, we use the same spinning technique used by industry to apply photoresist to silicon wafers. By adjusting the spinning speed and the viscosity of the liquid, we can obtain a uniform layer in the thickness range of interest. This liquid molding technique is illustrated in Fig. 4-15.

Eight micromachining runs have been made so far, with six molds machined per run. The 8 mold types contain a total of 58 target patterns, and more than 10 R-T targets can be made from each pattern on a single mold. A typical mold is shown in Fig. 4-16. The surface smoothness varies from 0.05 to 0.1 μm depending on the peak-to-valley height of the R-T target pattern. The diamond-turning machine can produce a 0.025- μm surface under the best of conditions. The present collection of molds meets the target designer's specifications and should satisfy his requirements for the present and foreseeable future.

We have been successful using the liquid molding technique with liquids that wet the mold surface, producing polystyrene and PVC targets that meet our thickness and surface smoothness specifications. We were less successful with glass material (NaSiO_3). The liquid glass is water based, and water does not wet a metal surface naturally. A few wetting agents were tried without much success.

Author: E. J. Hsieh

Major Contributors: G. T. Jameson, C. W. Hatcher, D. R. Ciarlo, R. W. Clouser, J. B. Bryan, and T. C. Trevino

Fig. 4-16. A typical precision machined mold for fabricating Rayleigh-Taylor targets.

Target and Material Measurements

Introduction

Measurement technologies are a vital component of ICF target development and production. Material and process developments require precise measurement of physical properties, structure, and chemical composition. Target production depends heavily on quality assurance studies and documentation of target dimensions, surface finish, and symmetry. During the past year, we have continued our efforts to provide the measurement capabilities needed to keep pace with changes in target design, materials, production processes, and quality requirements.

The following articles describe developments in measurement techniques and improved facilities. The completion of our high intensity ion-beam drill is also described. These developments improve our existing capacity to measure, analyze, and document the condition of advanced ICF targets.

Author: W. G. Halsey

Measurement Development

In our continuing effort to provide measurements needed for target development and production, we have extended our analysis

capabilities in several key areas. Several new facilities were added this year. One important measurement we must make is the DT fuel content in glass microspheres. A new, low-background, gas-flow proportional counter has been installed to measure x rays emitted by the target when beta particles from tritium decay strike the glass shell. When properly calibrated, this measurement indicates how much fuel is in the target.

Our surface analysis system, which includes a scanning Auger microprobe and secondary-ion mass spectrometer, has been upgraded with the addition of a high vacuum transfer stage. This stage allows us to switch materials that are sensitive to oxygen or water vapor between the dry box and the surface analysis system without exposure to air.

New techniques are developed for existing facilities to respond to changing target analysis needs. Three examples illustrate this evolutionary process. The existing automated x-ray image analysis system (AXIAS), originally designed to examine x-ray micro-radiographs, was modified to examine interferograms from the cryogenic research projects. Our split-image measuring microscope can now measure vertical heights with a new stage motion encoder and precise focus system. To support the beryllium research activities, our metallographic and microsectioning area has been modified to handle toxic materials.

Author: W. G. Halsey

Ion-Beam Drill

A three-year project to design and construct a system that uses a highly focused beam of gold ions to drill submicron holes through different target materials was completed. The heart of the drill system is a high intensity, scanning ion probe that was constructed and installed at Hughes Research Laboratory (HRL). In this article we discuss the general specifications and the design philosophy used to meet them.

Requirements. The general design specifications for the ion-beam drill were quite severe. First the system had to be capable

of drilling holes of 1 μm diam or less in conducting and nonconducting materials up to 20 μm thick. Also, because the milling rates of most materials in the ion beam are so large, the beam cannot be used to locate or scan the spheres to be drilled. Such a scan would produce unacceptable surface damage. Therefore, the spheres must be automatically positioned under the ion drill without visual reference to an image produced by scanning. The tolerances for positioning the spheres are also quite stringent; each hole must be within $\pm 1 \mu\text{m}$ of the sphere top dead center. These general specifications determined the overall design of the drill system.

Operating Procedure. In order to meet these system requirements, the spheres are placed on a carrier, much like a tray, which is then placed in a scanning electron microscope (SEM). The distances between two fiducial marks on the carrier and the centers of the spheres are measured with the interferometer system and the SEM, and this coordinate information is stored in a computer (see Fig. 4-17). The sphere carrier is then manually transferred to the ion-beam probe where the fiducial marks are located by scanning with the ion beam. After the fiducial marks are located, the ion beam is blanked. The carrier automatically moves to center the spheres under the ion beam using the position information stored in the computer.

Major Components. In brief there are five major components of the ion-beam drill system: an SEM, an ion-beam probe, an interferometer measuring system, a secondary-ion mass spectrometer (SIMS), and a computer control system. Figure 4-18 is a functional diagram of the ion-beam drill showing the relationship of the different components. Each part of the system is in reality a complex subsystem.

A standard aluminum SEM stub was modified for use as a carrier to demonstrate the performance of the ion-beam drill. The stub was polished and then coated with a thin film of gold (about 1000 Å). Fiducial marks were machined on the carrier by use of the ion-beam drill. The gold film was milled away to reveal the aluminum, which provided a high contrast for the fiducial marks.

Fig. 4-17. Spheres are positioned under the ion-beam drill by measuring x , y , Δx , and Δy with the interferometer and SEM, locating the fiducial marks with the ion beam, and using coordinate data to position the blanked ion beam.

Fig. 4-18. Functional diagram of the ion-beam drill.

A Bausch and Lomb Nanolab 2100LE SEM was selected for use in the ion drill. The only required modification to the SEM was the addition of the interferometer measuring system. Two plane interferometers are used to measure the x and y coordinates of the stage to an accuracy of $0.08 \mu\text{m}$.

The ion probe is similar to an SEM in that it has both power and viewing electronics, a vacuum system, and a vibration isolation system. The ion-beam column and all of its associated power and control electronics were designed and fabricated by Hughes Research Lab (HRL). The column produces a focused, high intensity beam of 60-keV gold ions using a gold/silicon "liquid metal" ion source. The beam spot size is less than $2000 \text{ \AA}$; beam current is 0.1 nA . The column design has been described elsewhere.^{22,23}

The interferometer was fabricated from commercially available components and measures to within $0.08 \mu\text{m}$. The components are essentially the same as those used on our LLNL diamond-turning machines. The x and y coordinate positions of both the SEM and probe stages are monitored.

The SIMS serves as part of the ion-beam control system. By monitoring secondary-ion production during milling of the target, the control system will allow us to determine when the beam has completed a specific machining operation. In particular, for the case of drilling holes through the wall

of a glass microballoon, it is necessary to determine when the beam first breaks through the inner wall. Because the depth of field of the beam is larger than a shell, the ion beam must be shut off to prevent damage to the opposite wall.

Stage movements and measurements are all under computer control. We modified our Tracor-Northern TN-2000/1310 system so that it could perform all of the tasks required by the ion drill. The modifications include the addition of the interferometer and SIMS interfaces. Also the software was modified to make use of the increased position accuracy information provided by the

Fig. 4-19. Completed ion-beam system during final testing.

interferometer system. The completed ion-beam system is shown in Fig. 4-19.

Author: R. F. Schenz

Major Contributors: G. J. Greiner,
A. Hernandez, and G. F. Jacobson

Ion-Beam Drilling Experiments on Metal Foils

As part of a continuing development program, we have been conducting experiments to evaluate the potential of the Hughes Research Laboratory (HRL) ion column for micro-machining different materials. We find that besides being useful in drilling glass microspheres, ion-beam drilling has great potential for machining other materials. In earlier work²⁴ we reported drill times for gold and glass and developed empirical expressions relating drill time to material thickness. More recently we also examined the drilling and machining of

polyvinyl alcohol films.²⁵ In this article we discuss methods for estimating the drill rates based on equations of sputtering efficiency. We also present some recent results on ion drilling of beryllium and make some predictions of drill rates for other metals based on the data for beryllium and gold.

Estimates of Sputtering Efficiency. The drill time for any given material depends on the sputtering efficiency S . The sputtering efficiency for a given ion/target combination can be approximated using the analytic expression developed by Thompson²⁶; this expression, which is valid for bombardment with high energy, heavy ions, is

$$S = \frac{(\pi a_0 n^{1/3})^2}{\delta} \left(\frac{E_R}{E_B} \right) \left(\frac{m_i}{m_i + m_t} \right) (z_i z_t)^{5/6} \text{ sec } \phi \quad (1)$$

where a_0 is the Bohr radius, n is the number density, E_R is the Rydberg energy, E_B is the binding (i.e., sublimation) energy, m_i , m_t , z_i ,

z_i are the masses and atomic numbers for the incident and target atoms, respectively, and ϕ is the angle of incidence of the beam relative to a normal drawn to the target surface. The sputtering efficiency is expressed as the number of atoms of target material removed per incident ion.

The Thompson expression is generally found to overestimate²⁶ the efficiency of sputtering by a factor of 2 to 4. For our work, calculation of the absolute sputtering efficiency is not as critical as being able to estimate relative values for different target materials. Thus, from our measured sputtering efficiency for Au^+ on Au, we would like to estimate sputtering rates (and hence machining rates) for Au^+ on other materials. From Eq. (1), the ratio of sputtering efficiencies for two different target materials (x and y) is

$$\frac{S_x}{S_y} = \left(\frac{n_x}{n_y}\right)^{2/3} \left(\frac{E_{B_y}}{E_{B_x}}\right) \left(\frac{m_i + m_y}{m_i + m_x}\right) \left(\frac{z_x}{z_y}\right)^{5/6} \quad (2)$$

It is assumed here that the ion source is the same. Note, however, that a more general expression can be derived from Eq. (1) for ratios of any ion/target combination.

The sputtering efficiencies for the Au^+/Au and Au^+/glass ion/target combinations were estimated from the measured drill times. Using gold as an example, the measured drill-through time t in seconds as a function of thickness x in microns is

$$t = 0.78 \exp(0.44x) \quad (3)$$

The drill rate v_s in microns/second is

$$v_s = \frac{dx}{dt} = 2.91 \exp(-0.44x) \quad (4)$$

The drill rate can be expressed in terms of the volume V of material removed by knowing the hole geometry vs hole depth. Our measurements have shown that the holes are roughly conical with a base of constant radius r . Therefore if x is the hole depth,

$$V = \frac{\pi r^2 x}{3} \quad (5)$$

For this geometry, the change in volume with increasing hole depth is

$$\frac{dV}{dx} = \frac{\pi r^2}{3} \quad (6)$$

Thus combining Eqs. (4) and (6), the volumetric rate of material removal, in cubic microns/second, is

$$\frac{dV}{dt} = 3.06r^2 \exp(-0.44x) \quad (7)$$

or expressed as the sputtered rate of gold atoms

$$\left(\frac{dN}{dt}\right)_{\text{Au}} = 1.82 \times 10^{11} \times r^2 \exp(-0.44x) \text{ atom/s} \quad (8)$$

Going through a similar argument for the Au^+/glass combination, the sputtering rate can be shown to be

$$\left(\frac{dN}{dt}\right)_{\text{Gl}} = 7.2 \times 10^{10} \times r^2 \exp(-0.47x) \text{ atom/s} \quad (9)$$

The sputtering efficiency values for gold and glass shown in Fig. 4-20 were determined by dividing Eqs. (8) and (9) by the incident flux ($\sim 10^9$ ions/s). The beryllium curve was calculated using Eq. (2).

Fig. 4-20. Sputtering efficiency vs thickness for gold, glass, and beryllium. The gold and glass curves are based on experimental data. The beryllium curve was calculated from the Thompson equation and the gold sputtering ratios.

Target and Material Measurements

Extrapolation of the drill times to zero depth eliminates the interference effect of the hole and gives sputtering efficiencies that can be compared with published values.

We were unable to locate published data for Au^+ on Au; however, sputtering efficiencies for Kr^+ , Xe^+ , and Hg^+ on Au and for Au^+ on Ag and Cu have been compiled by Carter and Colligon.²⁷ Their data are based on ion energies of 10 and 45 keV, compared to the 60-keV Au^+ source we use. It is known²⁷ that the sputtering efficiency reaches a "plateau" between 10 and 100 keV, so their values are probably only slightly lower than would be expected with a 60-keV source. We have used Eq. (2) to estimate sputtering efficiencies for Au^+ on Au based on these results for other ions. These sputtering efficiencies, which range from 14.8 atoms/ion for Au^+ /Cu to 80 atoms/ion for Kr^+ /Au, compare reasonably well with our measured value of 30 atoms/ion for Au^+ on Au.

The reported²⁷ sputtering efficiencies for quartz and glass bombarded by 25-keV Kr^+ ions are about 120 and 135 amu per incident ion respectively. (Atomic mass units are used because of the multi-element composition of the material.) For comparison our sputtering efficiency is 235 amu/ion for 60-keV Au^+ on an alkali silicate glass. (The mean formula weight for our glass is 20.4 g/g-mole. Thus a sputtering ratio of 11.5 atoms of glass per ion is equivalent to 235 amu/ion.) This is probably fair agreement since both the ion's energy and its mass are greater for our experiment. Unfortunately, little data exists on glass sputtering

with high energy ions, so extensive comparisons are not possible.

Experimental Measurements. A beryllium foil 2.5 μm thick was mounted over a 10-mil hole drilled in a silicon wafer. The foil thickness was measured directly from SEM photographs. This foil was used to establish the drill rate for beryllium. The measured drill time for this foil was 15.9 s. This is an average value based on several successive drilling runs. For comparison, the time required to drill 2.5- μm gold foil is 2.4 s. Thus, our measured sputtering rate for gold is about 6.6 times that for beryllium (assuming equivalent hole geometries for both materials).

Equation (2) can be used to predict the ratio of sputtering efficiencies for Au/Be to compare with our observed values. Values for the various terms in Eq. (2) are listed in Table 4-2; based on these data, the calculated sputtering ratio (Au/Be) is 5.6, which is in fair agreement with the 6.6 experimental value. The low sputtering efficiency for beryllium is known and, as can be seen from Eq. (1), results primarily from its low mass. Carter and Colligon²⁷ have compiled sputtering data for many different materials, ion sources, and incident energies. Drawing from this data, reported values for Au/Be sputtering ratios for low energy (500 eV) Kr^+ and Xe^+ are 6.4 and 8.6, respectively. Unfortunately, because the data are for low energy ions, comparison of absolute sputtering rates is not possible. Nevertheless, the ratio of sputtering rates we observed agrees rather well with these data.

The utility of Eq. (2) for predicting sputtering ratios is quite clear. It provides a simple method for estimating milling rates for other materials, provided sputtering rate data are available for at least one reference material (in our case, gold). To illustrate this, we used Eq. (2) to predict the ratio of drill times for five different metals vs gold. The results are given in Table 4-3. Values for the various parameters used in the calculation are summarized in Table 4-2. In general, these predictions can only be expected to be accurate to within a factor of 2. Reported sputtering ratios for lower energy ion sources are also given in Table 4-3; these results are from Carter and Colligon.²⁷ Note that the Thompson equation often

Table 4-2. Values of various parameters in the Thompson equation (Eq. 2) for several different metals.

Quantity	Metal							
	Au	Be	Ta	Ni	Ag	Cu	Al	W
E_B (kcal/mole)	88.0	78.25	186.9	102.7	68.1	80.5	378.0	201.8
z	79	4	73	28	47	29	13	74
m (g/g-atom)	196.97	9.01	180.95	58.71	107.87	63.55	26.98	183.85
n (in 10^{22} atoms/cm ³)	5.9	12.4	5.5	9.1	5.9	8.5	6.3	6.3

Table 4-3. Anticipated ratio of drill times for various metals vs gold of the same thickness. Calculated results are based on Eq. (2), whereas the experimental data are from Ref. 27.

	Ta/Au	Ni/Au	Ag/Au	Cu/Au	Al/Au
Calculated	2.3	1.4	0.9	1.1	2.2
10 keV Ar^+	—	—	0.82	1.19	—
10 keV Kr^+	—	—	1.0	1.81	—
500 eV Xe^+	3.45	2.4	0.91	1.47	3.7
500 eV Kr^+	3.57	2.4	0.93	1.30	3.2
500 eV Hg^+	4.17	2.7	0.95	1.43	4.2

Fig. 4-21. SEM photographs of an approximately $5 \times 5 \mu\text{m}$ square hole milled in the metal coating on a microsphere.

predicts faster drill rates than would be expected based on the low energy data.

Micromachining Demonstration. One of the potential uses of the ion mill is to carry out complex micromachining tasks on various materials. Experiments were recently carried out at HRL to demonstrate the machining capabilities of the focused ion-beam column. This work was carried out on a machine similar to the one purchased by Fusion Target Fabrication. The experiments consisted of cutting an approximately $5\text{-}\mu\text{m}$ -square hole in the wall of a metal-coated microsphere. The sphere was $180 \mu\text{m}$ in diameter with an approximately $1.2\text{-}\mu\text{m}$ -thick metal wall. Figure 4-21 shows a representative hole drilled by this method.

In conclusion, the ion-beam mill has potential in machining materials in various

Material	Rate of material removal ($\mu\text{m}^3/\text{s}$)	Milling time (s)
Au	0.146	196.2
Al	0.050	479.4
Au	0.132	217.8
Be	0.0235	1221.6
Cu	0.119	239.4
Glass	0.062	457.2
Ni	0.094	304.8
Ta	0.036	501
W	0.058	487.8

shapes. Table 4-4 shows estimated drilling times for a comparable hole in various materials. Estimated volumetric rate of material removal is also given.

Authors: J. H. Campbell and B. H. Ives

Major Contributors: R. F. Schenz, G. J. Greiner, J. Ward, and G. Stevens (Hughes Research Lab)

Table 4-4. Estimated milling times for square cuts in various materials (using same conditions and sample dimensions as for the metal tested).

References

1. 1982 *Laser Program Annual Report*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-50021-82 (1983).
2. *Laser Program Annual Report - 1979*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-50021-79 (1980).
3. 1980 *Laser Program Annual Report*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-50021-80 (1981).
4. C. W. Hatcher, L. E. Lorenzen, and B. W. Weinstein, *J. Vac. Sci. Technol.* **18**(3), 1187 (1981).
5. *Laser Program Annual Report - 1979*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-50021-79 (1980), pp. 4-59 to 4-61.
6. 1980 *Laser Program Annual Report*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-50021-80 (1981), pp. 4-58 to 4-60.
7. 1981 *Laser Program Annual Report*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-50021-81 (1982), pp. 4-52 to 4-53.
8. P. C. Souers, Lawrence Livermore National Laboratory, Livermore, Calif., private communication (March 15, 1983).
9. C. Souers, *Cryogenic Hydrogen Data Pertinent to Magnetic Fusion Energy*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-52628 (1979).
10. P. C. Souers, *Hydrogen Properties for Fusion Energy* (University of California Press, Berkeley, to be published).
11. *Laser Program Annual Report - 1977*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-50021-77 (1978), pp. 5-38 to 5-42.
12. 1980 *Laser Program Annual Report*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-50021-80 (1981), pp. 4-54 to 4-57.
13. T. M. Henderson, R. B. Jacobs, D. L. Musinski, R. J. Simms, and G. H. Wuttke, "Point-Contact Conduction-Cooling Technique and Apparatus for Cryogenic Laser Fusion Pellets," *Advances in Cryogenic Engineering* **23**, 690 (1978).
14. T. M. Henderson, R. B. Jacobs, D. E. Solomon, and G. H. Wuttke, "Techniques for Obtaining Cryogenic Laser Fusion Targets by Condensing Fuel Gases in Microshell Pellets," *Advances in Cryogenic Engineering* **21**, 455 (1975).
15. *Laser Program Annual Report - 1977*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-50021-77 (1978), pp. 4-58 to 4-59.
16. 1980 *Laser Program Annual Report*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-50021-80 (1981), pp. 4-51 to 4-53.
17. K. Kim, B. J. Smoot, R. L. Woerner, and C. D. Hendricks, "A New Technique for Fabricating Cryogenic Laser-Fusion Targets Using Cold-Gas Jets," *Appl. Phys. Lett.* **34**, 282-283 (1979).
18. R. J. Good and G. V. Ferry, "The Wetting of Solids by Liquid Hydrogen," *Advances in Cryogenic Engineering* **8**, 425 (1980).
19. K. Kim, L. Mok, and T. P. Bernat, "Behavior of Liquid Hydrogen Inside an ICF Target," *Proceedings of the Second International Colloquium on Drops and Bubbles*, Dennis H. LeCroissette, Ed. (Jet Propulsion Lab, Pasadena, Calif., 1981), pp. 96-106.
20. K. Kim and H. Rieger, "Cryogenic Inertial Confinement Fusion Target Fabrication System Directly Operable Inside a Room-Temperature Target Chamber," *Appl. Phys. Lett.* **37**, 425 (1980).
21. L. Mok, K. Kim, T. P. Bernat, and D. H. Darling, "Temperature Effects on the Formation of a Uniform Liquid Layer of Hydrogen Isotopes Inside a Spherical Cryogenic ICF Target," *J. of Vac. Sci. and Technol.* **A1**, 897 (1983).
22. R. L. Seliger, J. W. Ward, V. Wang, and R. L. Ruben, *Appl. Phys. Lett.* **34**, 310 (1979).
23. 1980 *Laser Program Annual Report*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-50021-80 (1981), pp. 4-18 to 4-19.
24. J. Campbell, D. Schenz, G. Greiner, and B. Ives, *Results from Recent Ion Drilling Experiment at Hughes Research Lab*, Lawrence Livermore National Laboratory, Livermore, Calif., Laser Program R&D Memo TF82-838 (Nov. 1982).
25. J. Campbell, R. F. Schenz, G. J. Greiner and B. H. Ives, *Results from Further Ion Drill Work at Hughes Research Lab*, Lawrence Livermore National Laboratory, Livermore, Calif., Laser Program R&D Memo TF83-080 (Jan. 1983).
26. M. W. Thompson, *Phil. Mag.* **18**, 377 (1968).
27. G. Carter and J. S. Colligon, *Ion Bombardment of Solids* (American Elsevier, New York, 1968) Chap. 7.

Basic x-ray spectroscopy measurements on neon-like selenium yield the ion charge states. These experiments support the x-ray laser program.

Fusion Experiments 5

Fusion Experiments

Contents

Section Editor: V. W. Slivinsky

Introduction	5-1
Diagnostics	5-1
Introduction	5-1
CR-39 Neutron Spectrometer for Fuel ρr Measurements	5-2
Novette Dante X-Ray Spectrometers	5-4
Design of Nova Dante Spectrometer	5-6
Photoemissive Materials for 0.35- μm Wavelength	5-7
Gamma-Ray Streak Camera	5-8
Streaked Recording of Simultaneous Optical and X-Ray Signals	5-10
Design of Ultrafast Gated Intensifier for Laser Fusion X-Ray	
Framing Applications	5-12
Novette Near-Axis Imaging System	5-18
X-Ray Phase Lens	5-19
Time-Resolved Crystal Spectrography	5-20
Imaging 20- to 100-keV X-Ray Emissions from Laser Fusion Targets	5-22
Co-Propagating Beam on Novette for X-Ray Backlighting	5-24
50-ps Neutron Diode	5-28
Time-Resolved Spectral Measurements Above 80 Å	5-28
Data Management and Analysis	5-31
Introduction	5-31
Applications of the ORACLE Data Base Management System	5-32
Processing of Laser Data	5-32
Processing of Data from the Streaked Crystal Spectrograph	5-34
UFO Code on the FEAF VAX Computer	5-36
Electronic Shotbook	5-37
Experiments	5-39
Introduction	5-39
Novette Laser-Plasma Interaction Experiments	5-41
Novette Soft X-Ray Laser Experiments	5-46
X-Ray Laser "Fail-Safe" Diagnostics: Performance and Results	5-53
Line Diagnostics of Laser-Produced Plasmas	5-55
Measurements of the Two-Plasmon Instability and Electron Heating	5-57
Electron Heating from the Ion-Acoustic-Decay Instability in a Finite-Length System	5-58
References	5-60

Fusion Experiments Program

Introduction

E. M. Campbell

The major objectives and responsibilities of the Fusion Experiments Program continue to be the development of advanced diagnostics and fielding for inertial confinement fusion (ICF) experiments, and, in conjunction with X Division, the design and implementation of the ICF experimental program. During the past year, we have also expanded our program to include the study of laboratory soft x-ray lasers. The x-ray laser program is a collaborative effort with R Program at LLNL.

In this section, we present results of the highly successful laser-plasma interaction studies conducted with the Novette laser operating at 0.53 μm . These Novette experiments have advanced our understanding of the physics of laser-plasma coupling in moderate and long scalelength plasmas. In addition, the experiments continue to show the improvement in the coupling physics that occurs with submicron light.

Our advanced diagnostic efforts continue to focus on improving our measuring capability for the complex physics of ICF. Progress has been made in the development of our picosecond x-ray framing cameras, and in identifying techniques and instruments for characterizing the fuel areal density.

Diagnostics have also been developed in support of the x-ray laser experiments. We have developed and fielded sensitive, time-resolved spectrometers for observing amplified spontaneous emission in the energy range of 40 to 160 eV.

Diagnostics

Introduction

In this part of the Fusion Experiments section we review progress in the development of new diagnostic capabilities in support of the Fusion Experiments Program. Of the 14 articles featured, 4 discuss novel refinements of "classical" instrumentation while 10 discuss the incorporation of new technologies or the synergism of multiple technologies to produce new instruments for meeting the ever-changing demands of our rapidly moving program.

Newly developed diagnostic instruments are a good indicator of the future directions of the Fusion Experiments Program. In 1983 we developed the instruments for use in 1984 and beyond. The articles here reflect the major themes of the next few years in the Laser Program.

1. In future ICF experiments on the Nova facility, we anticipate much higher target areal densities than we have ever before achieved—we hope to achieve ρr values well in excess of 100 mg/cm². Under such conditions we must rely more heavily on thermonuclear neutrons to characterize the compressed fuel. This concern motivates the work on fast-neutron diodes and on particle track detection methods for fuel ρr measurement using scattered neutrons.
2. On the Nova facility, we will seriously confront the issues of detailed measurement of the hydrodynamic stability of ICF targets. We will require detailed temporal and spatial information on dynamic processes in materials that may be relatively dense and cool (i.e., opaque to self-emission) and moving at speeds on the order of 10^7 cm/s. The investigation of such processes will require time-resolved imaging capabilities and auxiliary laser-driven x-ray sources well

beyond the current state of the art in short-pulse x-radiography. The investigation of rapidly moving hydrodynamic processes further motivates the need for absolute timing references on all time-resolved instrumentation.

3. As we proceed to shorter-wavelength laser radiation for driving ICF targets, the expectation of reduced suprathermal electron production does not diminish our interest in quantitative spatial and temporal characterization of suprathermal electron production and transport. This interest motivates time-resolved and space-resolved suprathermal x-ray instrumentation with significantly increased sensitivity.
4. Another important development for the Laser Program is the growing appreciation of the usefulness of large laser facilities for applications other than ICF. A prime example of this is the work directed at producing the world's first laboratory x-ray laser. New instrumentation is being developed for time-resolved spectral characterization of the pump radiation and the laser output. The development of x-ray optics for the efficient transport and focus of x-ray laser radiation is also under way.

Author: N. M. Ceglio

CR-39 Neutron Spectrometer for Fuel ρr Measurements

The fuel ρr may be measured by counting the number of recoil deuterons and tritons elastically scattered by fusion neutrons and recorded on CR-39 plastic film.¹ Because our high-density targets have high pusher areal densities ($\rho\Delta r > 0.1 \text{ g/cm}^2$) that cannot be penetrated by these recoil fuel ions, we are developing a detector capable of responding to scattered neutrons. These particles can easily escape the target. For highly compressed targets, the dominant neutron interaction is elastic scattering with the DT fuel, where the neutrons lose some of their initial energy. The fuel ρr can be estimated from $\rho r = 5.4 / \epsilon \text{ (g/cm}^2\text{)}$, where the factor 5.4 contains the cross-section and atomic weights and ϵ is the fraction of 14.1-MeV neutrons that have been elastically scattered. The fraction of scattered neutrons that can be observed by the detector, whose efficiency is ϵ , depends on the detector response and on the neutron spectral shape. Typical values for ϵ are expected to be 1 to 10% for Nova targets. The background level can be determined by employing high-yield, low-density targets that produce no signal due to scattering in the

Fig. 5-1. Schematic of neutron detector. The neutron energy can be found from the recoil proton energy E_p and various parameters that describe the geometry of the detector and the etch pit.

fuel. Contributions to the background in the energy range of interest (3 to 14 MeV) include TT neutrons, scattered neutrons from the target chamber walls, and effects from target-produced x rays. We describe here the first efforts in developing such a detector, and the results of a preliminary test.

The detector, placed a few centimeters from the target, as shown schematically in Fig. 5-1, has a 1-mm-thick tantalum shield to protect the track detector from x rays and debris generated by the implosion. A polyethylene radiator 50 to 200 μm thick is used to produce recoil protons that in turn can produce damage tracks in the CR-39. Additional data are obtained by using two layers of 1-mm-thick CR-39. The first layer serves to slow down the highest-energy protons, thus yielding larger and more readable tracks on the surface of the second layer.

After etching, the neutron energies E_n can be derived (Fig. 5-1) from several parameters that describe the geometry of the tracks and the detector, using

$$E_n = E_p$$

$$\times \frac{l^2 + k^2 - tk \tan \gamma \cos \phi - tl + \left(\frac{l}{2 \cos \gamma}\right)^2}{\left(k \sin \gamma \cos \phi + l \cos \gamma - \frac{l}{2 \cos \gamma}\right)^2} \quad (1)$$

From the major and minor axes and the off-center distance of the track apex, and with the aid of proton calibration curves, we determine the recoil proton energy E_p and the dip angle $(\pi/2 - \gamma)$. Additional parameters required are the angle ϕ between the major axis of the ellipse and the line passing through the center of the ellipse and the detector centerline; the distance k between the centers of the ellipse and the detector; the target-to-detector distance l ; and the radiator thickness t . Errors will be introduced by making the radiator thickness large because all elastic scattering events are assumed to occur a distance $t/2$ into the radiator. Derivations of a similar equation used in connection with a neutron dosimeter are given by Frank, Benton, and Albergotti.²

Because it is necessary to accurately measure features of the tracks, the requirements

Monomer	Initiator	Additives	Manufacturer
Allymer	2% IPF	0.01% antioxidant	American Acrylics
Allymer	2% IPP	1% DOP + 0.01% antioxidant	American Acrylics
Allymer	7% CHPC	None	Pershore
Allymer	7% CHPC	1% DOP	Pershore
PPG	2% IPP	None	American Acrylics

Table 5-1. CR-39 samples tested.

Fig. 5-2. An unfold of a 14.7-MeV neutron spectrum obtained at the Livermore RTNS.

on the CR-39 are quite demanding. To maximize registration sensitivity and track readability we evaluated five different types of CR-39 (described in Table 5-1) by exposing them to monoenergetic protons having energies of 3, 5, 8 and 11 MeV at dip angles ranging from 60 to 90 deg. The CR-39 samples were etched at 70°C for 24 h in an unagitated bath of 6.25-N NaOH with 0.5% Dowfax surfactant. From qualitative criteria we judged the CR-39 made from the Allymer monomer containing 0.01% antioxidant to be the most acceptable and convenient. We obtained this plastic with the aid of Dr. P. B. Price at the University of California, Berkeley.

As an initial test we exposed the Allymer monomer CR-39 to 14.7-MeV neutrons generated by the Livermore rotating target neutron source (RTNS). Using the calibration curves, we were able to unfold a simple neutron spectrum (Fig. 5-2). The neutron fluence was 4×10^7 n/cm², which produced 4×10^3 tracks/cm², thus yielding an efficiency of 10^{-4} tracks/neutron. For a parallel source of neutrons Eq. (1) simplifies to

$$E_n = E_p / (\cos \gamma)^2 \quad (2)$$

The measured FWHM of the 14.7-MeV peak was 1.4 MeV, compared to actual FWHM of 400 keV.

Author: S. M. Lane

Novette Dante X-Ray Spectrometers

Instruments to measure soft x-rays are mounted on the Novette chamber on axes roughly orthogonal to the axes of the two laser beamlines. In this article we describe the Novette x-ray diode (XRD) arrays, locally called Dante spectrometers, that are designed to measure absolute, time-dependent, x-ray spectra from 0.15 to 1.5 keV, using 10 K- or L-edge filtered channels with relatively broad (200 eV) energy resolution.³

The two 10-channel Dante XRD arrays (designated H and A) are horizontally positioned on opposite sides of the chamber. To facilitate the correlation of results, the spatially discriminating streak camera^{4,5} (SDSS) has been mounted directly above and as close as possible to the Dante-H instrument (18 deg away), and efforts have been made to match the energy responses of the two instruments. We intend to rely on the SDSS for superior time resolution and spatial discrimination and on the Dante spectrometers for absolute x-ray spectral measurements. The data obtained with these instruments will be compared in detail.

The Novette Dante spectrometers are similar to Shiva facility instruments previously described.³ Several technical improvements have been incorporated. The nearly identical Novette Dante systems each consist of two sets of five filtered XRDs located on a flight tube that is evacuated to 10^{-6} Torr by a vertically mounted cryostat

pump. The flight tube is supported on a new front pivot assembly so that its axis can be easily and accurately aligned with the aid of a rear-mounted, centerline laser. Use of a more rigid stainless-steel flight tube has resulted in a lower base vacuum pressure and better internal alignment for each instrument.

The energy response of each Dante channel is determined by calibrating its components separately.⁶ Nearly all of our energy-response measurements have been made with the help of the LLNL Calibration and Standards Laboratory. Several new filter sets, obtained for use on Novette, have been calibrated at the Ionac facility. A very good set of 1- μ m boron filters has been produced at LLNL and is being used for the lowest-energy channel. A good backup boron set was developed by the Lebow Co. The Shiva Formvar filter (carbon K-edge) was replaced by thick Parylene, thus decreasing the contribution to the channel response immediately below the oxygen K-edge. (Instead of Formvar, we now use Parylene-N for the Dante carbon K-edge filter and carbon for the SDSS carbon K-edge filter. We continue to calibrate and use Formvar foils made at LLNL for proportional counter windows.) We obtained magnesium filters of a more suitable thickness.

Our standard x-ray diodes were retrofitted with a new cathode and anode design for Novette. New sets of cathodes were diamond-turned and calibrated.^{7,8}

Three grazing-incidence mirrors are being used on the back array of each Novette Dante spectrometer. These self-contained

Fig. 5-3. Reflectivities of (a) two grazing-incidence carbon mirrors, (b) two beryllium x-ray mirrors, and (c) carbon mirror at 2.4 deg.

Fig. 5-4. Absolute energy response of a typical carbon K-edge channel obtained using a set of calibrated Parylene-N filters, an aluminum XRD, and the fit to the beryllium mirror reflectivity data shown in Fig. 5-3.

Fig. 5-5. Calibrated energy sensitivities of Dante-II channels (a) with filter edges below 800 eV, and (b) with filter edges above 700 eV.

mirror cassettes, designed by Lockheed, have been recently recalibrated at the Ionac.⁶ Figure 5-3(a) shows the Ionac x-ray reflectivity data for two carbon mirrors at 5 deg used for the boron-filtered channels. The best-fit theoretical curve for both mirrors was obtained by scaling by a factor of 0.86 the reflectivity values given by the REFLECT2 computer code written by H. F. Finn of our Laboratory. The measured reflectivity values for the two beryllium mirrors at 3.4 deg, shown in Fig. 5-3(b), are reasonably well fitted with the REFLECT2 code by assuming a weighted composition of beryllium and beryllium oxide. The experimental data in the important x-ray energy region below the vanadium L-edge filter for the carbon mirror at 2.4 deg, shown in Fig. 5-3(c), are fitted reasonably well using REFLECT2 and a 0.82 scale factor. Experimental values are comparatively high near the carbon K-edge, as expected from modifications of the reflectivity near edges calculated by A. Toor of our Laboratory; the experimental values are also high at the 704-eV Ionac iron-L source line, for reasons yet to be explained. Work on the mirror calibration data analysis is still in progress.

Figure 5-4 shows the absolute energy response of a typical carbon K-edge channel and the energy calibration data for each of the components from which the total response was derived. The absolutely calibrated channel energy responses of the 10 Novette Dante-H channels are summarized in Figs. 5-4 and 5-5. Comparison with the data produced by Shiva instruments³ shows that thicker filters, grazing-incidence mirrors, and the addition of the boron channel have resulted in significant improvements in coverage and in energy resolution below 600 eV.

Authors: K. G. Tirsell and D. G. Nilson

Design of Nova Dante Spectrometer

Dante spectrometers, consisting of arrays of filtered x-ray diode channels, are designed to measure absolute x-ray spectral emission below approximately 1.5 keV. The energy response of each channel is defined by a

K- or L-edge filter. Energy FWHMs are typically 200 eV. For the lowest energy channels, grazing-incidence reflection from a flat mirror removes contributions to the response from the energy region above the filter edge. The main advantage of a Dante spectrometer is its ability to measure spectral emission absolutely. Each of the Nova channel components will therefore be calibrated at the LLNL X-ray Calibration and Standards Laboratory.

The Nova target irradiation experiments will use two 10-channel Dante spectrometers. They will view the target from opposite directions from ports orthogonal to the axis of the laser beam cones. Complementary soft x-ray diagnostic instruments will be located in adjacent ports.

Our Nova Dante spectrometer design, shown schematically in Fig. 5-6, incorporates most of the mechanical improvements in the Novette version. A permanent-magnet deflection structure is located close to the front entrance to sweep electrons laterally out of each x-ray beamline. The forward collimation is designed to restrict each

Fig. 5-6. Schematic of the Nova Dante x-ray spectrometer.

detector's view to a 1-cm-diam circular region at the target center. To simplify alignment, the flight tube is supported by a pivot point on the chamber flange, the centerline is configured for a laser alignment system, and the back of the system is adjustable for pointing. A major innovation is a filter housing that can be conveniently removed from the flight tube. The filter assembly has been designed so that the conductance between target chamber and detector stations can be greatly decreased by using the Parylene prefilter as a vacuum barrier. Thus the low pressures required for high diode bias voltage can be maintained when target chamber pressures exceed 2×10^{-5} Torr.

For greater flexibility in experimental coverage, the spectrometer has provision for 15 x-ray lines of sight. Figure 5-6 shows one possible arrangement. In this case the basic array consists of 10 standard x-ray diodes, four with mirror cutoffs. Each of these detectors is coupled to a Tektronix 7912 transient digitizer by means of a 50- Ω 1/2-in. air heliax signal cable. Three of the lines of sight have been allocated to 50-ps x-ray diodes that are capable of better than 120-ps FWHM time response when used with Thompson TS-660 oscilloscopes and short signal cables. For best results, we will use the shielded recording system enclosure designed for Novette. One of the detector stations is reserved for an optically sensitive detector that is being designed to record light emission from Nova targets. The bottom position on the outer ring is reserved for an x-ray instrument to complement the Dante spectrometer, such as the soft x-ray streak camera (SDSS) or any other x-ray diagnostic instrument that could be used at 4 or 5 meters. This arrangement will permit comparison of results between instruments with the same view of the target.

Authors: K. G. Tirsell, D. G. Nilson, and E. A. Ries

Photoemissive Materials for 0.35- μ m Wavelength

A proposal is under way to identify light-sensitive photocathode materials and evaluate various methods for implementing a hybrid (optical and x-ray) streak camera. We are examining the response of

photocathodes to 0.35- μ m light, because this is the output wavelength of the frequency-tripled Nd:glass Nova laser, and because multiple fiducials can be obtained from the output laser sensor packages. Such a hybrid camera will be based on our soft x-ray streak camera, so that the photocathode will be demountable and also transparent. Furthermore, the photocathode should be able to withstand eventual exposure to air without degradation of performance. Consequently, we devoted particular attention to metallic photocathodes with low work functions. (The photon energy at 0.35 μ m is 3.54 eV.)

The suitability of various photocathodes is related to the quantitative response of the streak camera. Threshold corresponds to the minimum detectable camera output, as recorded on film. The maximum usable signal is the one that produces blooming on the photographic film, with deterioration of both temporal and spatial resolution. (We are principally concerned with obtaining an accurate timing fiducial rather than an accurate representation of the laser pulse.) These limits are illustrated in Fig. 5-7, which shows the performance of several materials of various thickness on different substrates, including our standard x-ray photocathode material, gold and cesium iodide. In absolute terms, the

Fig. 5-7. Response of photocathode materials to radiation of wavelength 0.35 μ m. Lower and upper bars mark threshold and blooming irradiances, respectively. See Table 5-2.

Table 5-2. Photoemissive materials, substrates, and thicknesses tested.

Material	Material thickness (Å)	Substrate	Substrate thickness
Niobium	100	SiO ₂	1 mm
	300	SiO ₂	1 mm
	500	SiO ₂	1 mm
Barium	200	SiO ₂	1 mm
Aluminum (1)	200	CH	1000 Å
Aluminum (2)	300	SiO ₂	1 mm
Gold	300	CH	1000 Å
Inconel	200	SiO ₂	1 mm
CsI	1000	CH	1000 Å

threshold corresponds to a cathode emission of about 10^5 electrons/cm², and blooming corresponds to a cathode emission of about 10^8 electrons/cm².

Another consideration is the amount of damage sustained by the material. Substantial damage to the niobium cathode is revealed by electron microscopy.

Results obtained so far indicate that aluminum is significantly more sensitive and exhibits a wider dynamic range than the other materials. These data were obtained with a well-characterized 0.35- μ m-wavelength laser pulse of 40 ps duration obtained by frequency tripling the output of a Nd:YAG laser. Calibrated photodiodes measured laser intensity, an optical streak camera gave temporal pulse shape, and a conjugate-plane film diagnostic provided time-integrated beam shape.

Three options were considered to implement photocathodes.

1. If the photocathode is suitable for both light and x rays, a uniform, transparent photocathode can be built with the candidate materials. One portion will be illuminated with the laser pulse and the rest with the x-ray flux.
2. If light and x rays require different photocathodes (different materials or thicknesses on substrates), a double photocathode can be implemented.

3. The 0.35- μ m photocathode can be implemented on the tip of an optical fiber. This simplifies the transmission and distribution of the laser pulse signal.

We have fabricated and tested aluminum photocathodes deposited on optical fiber tips, with good results. Irradiances of about 8×10^6 W/cm² provide a good laser pulse profile with negligible temporal dispersion because of transmission through the fiber. We plan to use fiber optics for the implementation of the fiducial network at Nova. In a parallel effort, we are extending the photocathode research to 0.27- μ m wavelength, to be used at the Novette laser facility. An appealing feature of the optical fiducial is that streak cameras can be tested and timed without the need to generate soft x rays, alleviating the need for target shots in the target chamber.

Authors: C. P. Hale, H. Medeck, and P. H. Y. Lee

Gamma-Ray Streak Camera

We have designed and tested a streak camera that is far more sensitive to photons (UV to γ -ray) than the conventional x-ray streak camera (XSC), with resolution better than 150 ps.

Fig. 5-8. A prototype γ -ray streak camera and the experimental setup.

Our prototype streak camera is essentially the same as the conventional XSC except that the photocathode is replaced by a microchannel plate (MCP) electron multiplier (Fig. 5-8). The reason that this arrangement can provide high sensitivity is evident by comparing Figs. 5-9 and 5-10. Figure 5-9 is the back-surface secondary electron quantum yield for a 230-Å gold transmission photocathode.⁹ (The extrapolation uses unpublished front-surface data of Slivinsky and Leipelt normalized to the data of Henke et al.⁹) Figure 5-10 is a compilation of the MCP detection efficiencies for 0.1- to 1000-keV photons. The MCP retains its >1% efficiency for γ -like photon energies. This new streak camera, originally conceived for soft x-ray application, offers an outstanding opportunity for γ -ray detection with excellent temporal resolution.

Let us compare this instrument with a conventional streak camera without an MCP. The detection efficiency of the MCP version is ~1% for MeV photons (Fig. 5-10). The quantum efficiency of a gold photocathode for MeV photons is ~0.05% (Fig. 5-9), and nominally, ≥ 10 photoelectrons are required for a detectable streak trace on the film, so the detection efficiency of the conventional XSC is 0.005%. Thus, the increase in the streak camera sensitivity for γ rays due to the MCP cathode is ≥ 200 .

We have obtained preliminary data using UV and soft x rays at our Monojoule laser facility. The streak camera is so sensitive that we turned off the laser amplifiers and used only the oscillator, obtaining 100-ps laser pulses of about 50 μ J. For our preliminary

tests, an etalon in the laser beam provided multiple pulses at 800-ps intervals.

The UV and x rays generated in a tantalum disk target were detected with the streak camera. Figures 5-11(a) and (b) show typical streak traces and their exposures, respectively. The FWHM of the weakest trace

Fig. 5-9. Back-surface secondary electron quantum yield for a 230-Å gold transmission photocathode.

Fig. 5-10. Compilation of microchannel plate efficiencies from various authors: Parkes (Ref. 10), Dolan (Ref. 11), and Adams (Ref. 12). (After Ref. 11.)

Fig. 5-11. (a) Typical streak traces and (b) their exposures, obtained for UV and x rays generated at our 1.06- μ m, 170-ps Monojoule laser facility operated at 50 μ J.

is 150 ps. The exact duration of the UV and x-ray pulse is not known; the resolution of our streak camera is estimated to be less than 150 ps. The major part of this time resolution is attributed to the large energy spread of the secondary electrons from the MCP cathode. We intend to improve the resolution by using an electrostatic analyzer to select a narrow energy band of electrons. Meanwhile, the present resolution is adequate for time-resolved suprathermal x-ray measurements from Novette experiments. Accordingly, we are installing this streak camera on the Novette chamber. It should provide temporal information for the 20-, 45-, 70-, and 100-keV x-ray images that we have obtained with the MCP pinhole camera described in the article "Imaging 20- to 100-keV X-Ray Emissions from Laser Fusion Targets," later in this section.

We believe that this highly sensitive streak camera will be of value for time-resolved UV, x-ray, and γ -ray measurements.

Authors: C. L. Wang, H. Medeck, C. P. Hale, G. R. Leipelt, S. W. Thomas, and J. D. Wiedwald

Major Contributors: J. F. Bushman and R. L. Griffith

Streaked Recording of Simultaneous Optical and X-Ray Signals

The OX-2 streak camera was designed to record information that allows one to infer the time history of suprathermal electrons and, simultaneously, the time history of various parametric instabilities in the plasma. Since these instabilities are a possible source of suprathermal electrons, observing the time correlations should enable us to identify which instabilities, if any, contribute significantly to the suprathermal population.

The electrons are detected indirectly by observing the high-energy (30 keV) bremsstrahlung emitted when they interact with the dense target material. Stimulated Raman scattering produces light at wavelengths between ω_0 and $2\omega_0$, while two-plasmon decay is inferred from the presence of scattered light at $3/2 \omega_0$ (where ω_0 is the frequency of the laser light). Other possible suprathermal generating

mechanisms, such as resonance absorption or filamentation, are not detected by this instrument, although, crudely speaking, the conditions for filamentation are similar to those required for stimulated Brillouin scattering (SBS). Brillouin scattered light (very near the laser wavelength) is also measured by the OX-2.

The OX-2 camera (a modified version of the OX-1 streak camera used on Shiva) can record eight simultaneous channels of information. There are three optical channels, two x-ray channels, one background channel in which both the optical and x-ray signals are blocked, and two channels for fiducials. The three optical channels are for the Raman, $3/2 \omega_0$, and Brillouin light. The two fiducials are taken from the driving pulse. All signals are recorded with a single optical streak camera with an S-1 photocathode. As demonstrated with the OX-1 camera, the S-1 photocathode has adequate sensitivity, when used with a wide slit, to x rays above 25 keV. The OX-2 camera has a 2.6-mm-wide, 6.4-mm-thick lead slit for the x-ray channels. A 100- μ m-wide, 0.1-mm-thick stainless steel slit, which is taped to the cathode and covers all channels, acts as the slit for the optical channels. The wider slit for the x-ray channels is necessary to increase the sensitivity of the camera to x rays. Studies have shown that slits can be widened to 1.5 mm without loss of temporal resolution.¹

The configuration of the OX-2 is shown in Fig. 5-12. The entire filter section of the camera is under vacuum and exposed to the target chamber during shots. With the streak tube angled slightly downward, the lead shield stops high-energy x rays from interacting with the streak tube, output phosphor, image intensifier, and photographic film.

The filter section can be divided into optical and x-ray channels. The optical channels consist of a plasma shield, a set of fixed and movable filters that are used to balance the relative intensity of the different channels, and 10 mm of lead glass to block x rays. These filters, which cover all the optical channels, are followed by a discrete filter section for individual channel definition. A 1.06- μ m blocking filter is included in the fixed-filter section to prevent any $1/2 \omega_0$ light from interfering with the Raman

Fig. 5-12. OX-2 configuration.

Fig. 5-13. Typical OX-2 data.

channel. To achieve proper exposure of the film, the movable filters are used to offset any differences in intensity between channels anticipated for a given shot. The movable filters can be selected individually or in various combinations without breaking vacuum. The discrete filter section is made up of two colored glass filters per channel. The channels are 2 mm wide with 1 mm separation. Fiducials are delivered via graded-index optical fiber from a 1.06- μm laser beam diagnostic sensor package. The fiducials, which provide a measure of the incident pulse shape, are recorded along both edges of the data. By lining up the two fiducials, we increase the accuracy of temporal measurements. The entrance holes for the optical fibers are shielded from x rays with a 3.2-mm-thick lead sheet.

The x-ray channel consists of two 25- μm -thick aluminum foil light shields. One channel was used to experiment with a piece of NE-111 scintillator material for increased x-ray response. The scintillator material was placed in contact with the 100- μm slit.

Figure 5-13 represents typical data from one of our experiments. The channels are

labeled in the figure and time proceeds vertically as indicated by the arrow. Note that in this particular experiment, no $3/2 \omega_0$ light was generated along with the high-energy x rays. The fine structure in the optical

channels is an interesting feature of the scattering process and demonstrates the good temporal resolution achieved with this instrument. Also, note that the two x-ray channels have equal density profiles, indicating a lack of signal contribution from the scintillator. Although this simple experiment is far from conclusive, we have identified other scintillator materials that have improved performance, enhancing the x-ray response of this camera.

We are making several improvements in the OX-2 camera for future experiments. We have begun installation of an x-ray-transmitting LiF vacuum window before the filter section to allow this section to be at atmospheric pressure during shots. This will simplify the mounting of the streak tube to reduce risk of damage and to allow convenient exchange of tubes to suit different experiments. To increase x-ray sensitivity, we will also investigate a BaF_2 scintillator material in combination with an S-1 streak tube having a sapphire window. Because Nova will have UV light that can produce a fiducial signal on our standard soft x-ray streak cameras, we have abandoned our mica window effort. The mica window was to be used as the cathode substrate of an optical streak tube to correlate low-energy x-rays and scattered visible light.

Authors: R. J. Ellis and R. E. Turner

Major Contributors: R. P. Drake,
H. Medeck, G. E. Phillips, and
J. B. Richards

Fig. 5-14. Target instabilities can result in a nonuniform implosion similar to this gas bubble. (Courtesy of Jane Lin, California Institute of Technology.)

Design of Ultrafast Gated Intensifier for Laser Fusion X-Ray Framing Applications

A major challenge for laser fusion is the study of the symmetry and hydrodynamic stability of imploding fuel capsules. Streak x-radiography, in one space and one time dimension, does not provide sufficient information about these phenomena. We need frames of at least 200 pixels in each of two spatial dimensions, with frame times of 10 to 100 ps, good image quality, minimal geometrical distortion (approximately 1%), and dynamic range greater than 1000:1. A gated transmission line imager (TLI) can meet these requirements with frame times between 30 and 100 ps. (This instrument is also called a gated intensifier, and is sometimes called a framing camera.) An instrument of this type is being developed. In this article we describe our progress on this instrument, including theory of operation, ultrafast pulse generation and propagation, component integration, and development of a high-resolution phosphor screen.

Diagnostic Requirements. The ablation of the surface of a fuel capsule creates pressures $> 10^8$ atm, far exceeding the strength of materials. The shell is accelerated inward at more than 10^{12} gravities and reaches velocities $> 10^7$ cm/s. Under these extreme conditions it is clear that the forces on the capsule must be exerted very symmetrically and uniformly or the capsule will not converge to a small high-density point.

The problem of imploding the fuel capsule is complicated by the high accelerations and by the fact that the low-density ablating material accelerates the high-density shell. Under these conditions the shell is potentially unstable against small perturbations that arise from Rayleigh-Taylor and Richtmyer-Meshkov instabilities. Figure 5-14 shows a gas bubble undergoing Rayleigh-Taylor instability. This figure gives an idea of how damaging the instability is to symmetry and suggests what an unstable imploding fuel capsule might look like.

These instabilities can be controlled by several methods, including adjusting the shell's thickness and grading its density.

These methods involve increases in capsule complexity and reductions in implosion efficiency. Since it is desirable to implode the capsule with the minimum amount of energy, it is important to establish the parameter-space boundary between capsule stability and instability.

To summarize, laser fusion requires the production of a hydrodynamically stable and symmetric implosion, which presents laser fusion diagnostics with a major challenge, namely, to develop a method to measure the symmetry and stability of the imploding shell and the symmetry of the compressed core.

The method that holds the most promise for meeting this challenge is x-ray back-lighting. Figure 5-15 shows how an x-ray backlit photo of the imploding target might be taken. The sample shown is the fuel capsule to be radiographed. Under conditions of interest the capsule would be irradiated by a number of laser beams, not shown here for simplicity. The x rays for radiographing the fuel capsule are produced by a laser-driven x-ray source that consists of a small (1×1 mm) foil of material having the desired characteristic emission lines. The foil is irradiated by one of the laser beams not used to irradiate the target. The x rays pass around and through the fuel capsule and are collected by an x-ray focusing optic (a grazing-incidence x-ray mirror, here shown schematically as a lens). The x-ray optic focuses the transmitted x-ray image onto an imaging detector, which records it on a recording medium, such as film or a CCD array. In this case the imaging detector is a high-speed framing camera, the subject of this article.

Framing Camera Requirements. The framing camera requirements, discussed below, are summarized in Table 5-3.

The fuel capsule implodes from its initial radius to its final radius in roughly the length of the laser drive pulse, ~ 1 to 10 ns. To resolve the image of the capsule without serious blur, it is therefore necessary to have an exposure time much shorter than a nanosecond.

With an implosion velocity of $\sim 3 \times 10^7$ cm/s and a desired spatial resolution at the target of a few microns, it is necessary to have a shutter time of 10 to 100 ps. The

Fig. 5-15. Implosion instabilities and symmetry can be measured by radiographing the target with a gated imaging detector.

Table 5-3. Framing camera requirements.

Time resolution	10-100 ps
Spatial resolution	20-60 μm
Image distortion	<1%
Image size	>200 $\times$ 200 pixels
X-ray energy range	1-8 keV
Dynamic range	>1000:1
Shutter ratio	>10,000:1
Edge response	$\sim 1000:1$ in 50 μm
Quantum efficiency	>5%

requirement for spatial resolution is determined by the fact that symmetry errors of less than 1% of the initial capsule radius can cause degradation of burn efficiency. This spatial resolution is also consistent with the fact that the wavelength of maximum growth rate for the Rayleigh-Taylor instability is on the order of the capsule wall thickness, again typically a few microns. This requirement has several implications for the framing camera design:

- The image must have a distortion of less than 1%, or be easily and convincingly correctable to that level.
- With magnifications of 10 to 20 for the x-ray optic, the camera resolution must be in the range of 20 to 60 microns.
- The camera must have $\sim 200 \times 200$ pixels to measure symmetry errors of 1%.

Taken together, these specifications imply an active photocathode 4 to 12 mm on a side. Generally speaking, it is easier to gate a small image quickly than a large one. Many framing-camera designs use deflection systems to displace the image from place to place on the screen between frames. We have not chosen this method, in the belief that an electrostatic deflection system offers too much opportunity for distortion of the image and will compromise the credibility of symmetry measurements at the 1% level. This problem is avoided by using the proximity-focused geometry, which inherently has small distortion. The residual distortion is fixed and can be easily and convincingly corrected. In a

proximity-focused gated intensifier, however, the gate time is limited to the transit time of the gating pulse across the photocathode at the speed of light. Thus the resolution of the camera and the number of pixels in the frame limit the speed of a proximity-focused gated imager. With a photocathode dimension of 4 to 12 mm the limiting gate time is 13 to 40 ps.

Other important considerations for viewing backlit targets and other laser fusion phenomena are dynamic range and edge response. Typically the target is dense and quite attenuating to the backlighting x rays, but the backlighter extends beyond the edge of the fuel capsule, so one must view a small, dark object against a large, bright background. The background from the backlighter may be more than 100 times brighter than the x rays transmitted through the target. Any scatter in the x-ray optical system or the intensifier will tend to flood the shadow image of the target with background and destroy the dynamic range of the image. To record this type of scene one needs a dynamic range $>1000:1$ and an edge response that drops nearly the full dynamic range within a few pixels.

A similar problem arises in the time domain, in that both the backlighter and the target emit intense background for a long time compared to the gate time. If the gated intensifier is not opaque enough when it is gated "off," leakage through the device will fog and blur the recorded image. For this reason the shutter ratio must be greater than 10 000:1.

Although the laser-driven backlighter source that we plan to use is one of the brightest available, the high resolution of the instrument in both space and time divides up the emitted energy so finely that

it is not easy to obtain sufficiently intense signals. Thus we require that the photocathode quantum efficiency and the electron transport through the camera be as high as possible.

Status of Gated Imaging. Fast gating of electrooptic imaging tubes has been the objective of researchers for many years. Kotecki and Lear¹⁴ summarize the technology of microchannel plate (MCP) intensifier tubes. They conclude that the gating speed of conventional MCP image tubes is limited to slightly under a nanosecond. Laviron and Delmare¹⁵ of the French Laboratory for Physical Electronics (LEP) described a tube driven through two parallel 50- Ω inputs, designed to minimize lead inductance and photocathode resistivity. This tube achieved 300-ps FWHM gating for visible-light inputs. Only recently has an effort been made to construct an imaging tube with the propagation of a fast gating pulse as a prime design criterion. Workers at LEP have built a sealed tube with 50- Ω characteristic impedance between the front surface of an active MCP and a phosphor screen.¹⁶ The 7-kV gating pulse is capacitively divided between the MCP (~ 1 kV) and the MCP-screen vacuum gap (~ 6 kV). The MCP is used as the x-ray converter/detector. This structure was among the first to include the gated gap as part of a transmission line, and it achieved a gating time of a few hundred picoseconds.

Transmission-Line Imager. From this background a device called a Transmission-Line Imager (TLI) has evolved to serve the needs of LLNL's Laser Fusion Program. The TLI is shown schematically in Fig. 5-16. Its principal features are an integral, geometrically matched pulse generator, a parallel-plate transmission line operating above its

Fig. 5-16. Schematic of the gated transmission line imager.

normal cutoff frequency, a photocathode and phosphor screen anode nondisruptively integrated into the transmission line, and an integral thick-film termination resistor to minimize electrical reflections. The implementation of the complete TLI design is a combination of innovative electrooptical design, a blending of electrical and mechanical technologies, and careful attention to selection of materials compatible with each other and with the required fabrication process.

A side view of the TLI design is shown in Fig. 5-17. We will describe it from left to right. The basis for the very short electrical gating pulse is the photoconductive (Auston) switch.¹⁷⁻²¹ This switch is a piece of near-intrinsic, high-resistivity ($>10^7 \Omega \cdot \text{cm}$) GaAs cut to the same width as the transmission line. It has two Au-Ge contacts²²⁻²⁴ plated and annealed on one surface and attached with Au-Sn or Ag-Sn solder to the transmission line. The gap between the contacts is approximately 2 mm wide, adequate to hold off 20 kV for several tens of nanoseconds. Surface breakdown is prevented by encapsulating the area above the gap in transparent nonconductive epoxy. The photoconductive switch behaves as an insulator until carriers are available to conduct between the contacts. It is possible to create a great number of carriers very quickly by illuminating the gap with a high-energy laser pulse ($\sim 1 \text{ mJ}$), turning

the GaAs into a quasi-metal for a period corresponding to the carrier recombination time, about 1 ns. One by-product of the Laser Fusion Program is the availability of laser pulses coincident with the experiment and of sufficient energy to cause the GaAs to conduct. The speed with which the device switches is the same as the width of the illuminating pulse. In our case, a 20-ps, 1.06- μm laser pulse or a 5-ps dye laser pulse will be available.

Near-uniform illumination of the switch gap ($2 \times 25 \text{ mm}$) is achieved by a fiber-optic geometry-changer, round at one end and rectangular at the other. Experiments at LLNL have shown the damage threshold of step-index fibers to be approximately $5 \times 10^{10} \text{ W/cm}^2$, well above the power required to drive the switch.

As the GaAs suddenly becomes conductive, the charge line and the transmission line are connected electrically. The charge line has been pulse-charged to a potential of $\sim 20 \text{ kV}$, and the transmission line has been biased slightly positive to deter photoelectrons from crossing the gap. The result is a pair of 10-kV waves moving away from the switch at the speed of light, one down the transmission line and one down the charge line. The duration of the pulse is limited to the round-trip propagation time of the charge line—for our work this time ranges from about 30 ps to about 300 ps,

Fig. 5-17. Cross section of the transmission line imager.

Fig. 5-18. Experimental 35-mm-wide transmission-line pulse as measured by a 100-ps rise-time oscilloscope.

depending on the phenomenon being diagnosed. The plane wave generated by switching along the full width of the switch gap is carried along a 25-mm-wide transmission line. The cutoff frequency of this line is 6 GHz, but pulses with rise times corresponding to much higher frequencies can be successfully transmitted, so long as the wave remains plane. As an example, Fig. 5-18 shows the leading edge of a 2.4-kV pulse transmitted 10 cm down a 35-mm-wide (4-GHz cutoff) transmission line. The laser pulse width was 50 ps. The deconvolved oscilloscope response has a rise time of approximately 60 ps (or 100 ps after an additional 20 cm of travel and reflection from the open end of the transmission line). Our TLI design requires pulse fidelity to be maintained for only 5 to 10 cm, so a rise time of 30 to 40 ps is anticipated (with the 20-ps laser excitation pulse).

After generation and transmission, the pulse reaches the active region of the device, the x-ray photocathode. The photocathode will be either gold or CsI on a 25- μ m beryllium substrate. The cathode emits photoelectrons with a quantum efficiency in the range of 0.1 to 1 (Ref. 9) and varying kinetic energy, typically 1 to 10 eV for gold and 0.5 to 4 eV for CsI. The lower energy spread of the CsI electrons, due to their smaller transverse velocities, is much preferred, but we will use gold at first because it is much easier to handle.

The importance of small transverse velocities is a consequence of our way of improving the dynamic range of the imager. The photocathode is essentially transparent

to the x rays being measured, allowing most of them to continue on to the phosphor screen. X rays stopped by the phosphor will create noise in the image, because there is no gating mechanism for the x rays. Our solution to this problem is to install an electron collimator in front of the screen. This collimator is essentially a thin, 0-deg-bias-angle MCP without gain. Electrons moving normal to its surface will pass through it to the phosphor, if there is not too much transverse electron energy. By projecting the x-ray image on the photocathode at a slight angle away from the normal (about 5 deg) x rays will be absorbed in the high-lead-content MCP glass. MCP attenuation is calculated to be >500 in the operating energy range.

After passing the photocathode, the gating pulse must be absorbed in some manner that keeps reflections to an acceptable level. The wide line and narrow spacing precludes discrete resistors. Working with F. Uribe of Sandia National Laboratory in Livermore, we have made large-area thick-film resistors that appear to have both the required high-frequency performance and the required short-pulse, high-current capability. The terminating resistor will be approximately 25 mm square and will have a resistance of about 18 Ω , the characteristic impedance of the TLI.

The terminating resistor is long compared to the wavelength of significant frequencies in the pulse. In order to minimize reflections, the local characteristic impedance will be controlled at a level nearly equal to the resistance remaining. This will be accomplished by placing an inclined ground plane under the resistor. The ground plane will stop short of actually touching the "ground" end of the resistor. Instead, there will be a small gap with relatively large surface area. This gap, in conjunction with discrete capacitors, will form an AC ground and still allow for a DC reverse bias to be applied across the gap.

The resolution and dynamic range requirements of the TLI are based, in part, on the fact that we are studying a dark image on a bright background. This is generally quite difficult because electrons and photons scatter from the bright areas into the dark. Elastic scattering of accelerated

photoelectrons²⁵ from a standard phosphor screen is depicted in Fig. 5-19. A significant fraction ($\sim 30\%$) of the electrons are scattered and eventually find their way back to the screen with a large fraction of their energy, resulting in image noise. Eliminating these scattered electrons is the second function of the electron collimator. Electrons scattered from the top of the collimator will not be able, because of their transverse velocities, to pass through other channels. Electrons scattered from the screen will be absorbed by the collimator. The overall result is a great increase in resolution, with an apparent loss in efficiency of about 50%. However, many of these lost electrons would have resulted in noise rather than signal, so there is a net increase in the information capacity of the system.

Finally, special care is being taken to ensure that the phosphor screen minimizes loss of information by photon scattering. The image will be coupled out of the TLJ via a fiber-optic faceplate, either directly onto film or into an ungated image intensifier. It has been determined that the high numerical aperture (NA) of most fiber-optic faceplates results in a negligible depth of field when directly coupled to a subsequent stage. As a result, the finite phosphor thickness degrades the resolution well below that found by observing the screen with low-NA viewing optics. In addition, light scattered among the phosphor grains is eventually collected by fibers, possibly far from its point of creation. This process can drastically impair the edge response of the imager. A new phosphor screen design, being pursued to minimize these effects, is shown in Fig. 5-20 (Ref. 26). The two important features of this screen are optical isolation of phosphor islands corresponding to fiber cores, and the inclusion of a monolayer of small phosphor grains.

The optical isolation is achieved by standard lithographic techniques. A method that uses the fiber-optic faceplate as its own mask allows the webbing between the fibers of the faceplate to be coated with about $1\ \mu\text{m}$ of metal, while the ends of the fibers are left coated with photo-resist. The faceplate is then heated to about 150°C to 175°C and the resist becomes viscous. The brushing of fine-grain phosphor (0.25 to

▲ Fig. 5-19. Electrons scattered from the phosphor screen degrade imager edge response.

Fig. 5-20. Electron collimator and segmented phosphor screen greatly enhance resolution and contrast ratio.

Fig. 5-21. Low internal cross talk of AO 2790A fiber optic results in sharp edge response.

$1\ \mu\text{m}$) onto this sticky surface lay down a monolayer of phosphor over the individual fibers. The temperature of the fiber-optic screen is then raised to 350°C , evaporating the resist and bonding the phosphor grains to the faceplate glass. The result is the creation of islands of phosphor with little possibility of cross talk.

Fiber-optic faceplates have been evaluated for internal cross talk, and American Optical 2790A has been found to have very low internal fiber cross talk (see Fig. 5-21). We anticipate that screens built using this technique will demonstrate point spread

Diagnostics

functions with widths on the order of the fiber pitch (6 μm), although overall screen resolution will be limited to the 15- μm collimator pitch.

If this design is successful in all respects it will make possible unprecedented quality of short frame-time quantitative images of ICF phenomena.

Authors: R. H. Price, J. D. Wiedwald, J. R. Kalibjian, S. W. Thomas, and W. M. Cook

Novette Near-Axis Imaging System

We built the Near-Axis Imaging System to look for filamentation by imaging the Raman and $3/2 \omega_0$ light generated along these filaments in the near-backwards direction. Refraction should not be an issue for Raman scatter at electron densities near or below $0.1 n_c$, or for $3/2 \omega_0$ light, which is parametrically generated near $n_c/4$.

Because we wish to image light ranging spectrally from the UV to the infrared, we required an imaging system that was almost completely achromatic. Because the

filaments may extend hundreds of microns longitudinally, our imaging system had to be able to look close to the laser beam axis and to have a large depth of field. To get close to the laser beam axis, part of the optical system had to be inside the target chamber. After considering several possible mirror systems, we decided to have a turning mirror and a lens collect the light and focus it outside the chamber.

It is not possible to construct a lens that is achromatic over more than a narrow spectral range. However, there is a way of constructing a completely achromatic system with a lens as the first optic. This article describes the optical design of this instrument.

Figure 5-22 depicts the optical system. It is an $f/10$, diffraction-limited system with neither lateral nor longitudinal color aberration at any wavelength in Gaussian optics. Its magnification is 10. The re-entrant tube, which holds the filter, deflection mirror, and Suprasil I lens doublet, can be rotated and moved in and out while the chamber is at vacuum, so that the field of view can be centered on the target. The filter can also be

Fig. 5-22. Novette Near-Axis Imaging System.

moved into and out of position while the chamber is at vacuum. This is a necessary feature, because in many of the planned experiments the filter will be opaque to the visible light used to do the alignment. One possibly unfamiliar optical element is the Mangin mirror, which is a lens whose back surface is aluminized.

The principle behind the exact color correction is explained by Fig. 5-23. The positive and negative thin lenses are made of the same optical material, so that their focal lengths are in constant ratio independent of wavelength. The idealized achromatic relay (a concave mirror) images the back surface of the first lens onto the front surface of the second lens. The color correction is exact only in Gaussian optics, so ray matrices are used in the explanation that follows.

The ray matrices relate the output ray height y and slope dy/dx to the input height and slope. Thus

$$\begin{bmatrix} y \\ dy/dx \end{bmatrix}_2 = \begin{bmatrix} A & B \\ C & D \end{bmatrix} \begin{bmatrix} y \\ dy/dx \end{bmatrix}_1 \quad (3)$$

The matrix element B for the idealized relay is zero, since all rays emerging from any given point on the input plane 1 are focused to another point on the output plane 2, regardless of the ray slope. The ray matrix $\mathbf{Q} = \mathbf{WVU}$ for the combined optical system is

$$\mathbf{Q} = \begin{bmatrix} A & 0 \\ C - \left(\frac{A}{f'} + \frac{D}{f}\right) & D \end{bmatrix} \quad (4)$$

Because the relay is achromatic, the matrix elements for the relay ray matrix do not depend on wavelength. The matrix $\mathbf{Q}$ will also be independent of wavelength if

$$\frac{A}{f'} + \frac{D}{f} = 0 \quad \text{or} \quad \frac{f}{f'} = -\frac{D}{A} \quad (5)$$

If Eq. (5) is true at any wavelength, it is true at all wavelengths, given that the two lenses are thin and of the same material.

The cancellation of the lateral and longitudinal color aberrations does not depend critically on the alignment, but there is a prism effect whose elimination requires

Fig. 5-23. Principle behind the color-correction scheme. The matrix elements of $\mathbf{V}$ are independent of wavelength. The combined matrix $\mathbf{WVU}$ is also wavelength-independent if the quantity in parentheses equals zero.

very precise alignment. The prism effect causes images at different wavelengths to be displaced laterally from each other, even though they may come to focus in the same plane and with the same magnification.

Having specified the desired collecting lens and the magnification, the spherical aberration can be eliminated by choosing the relay magnification. (This is the previously discussed relay that images the collecting lens onto the Mangin mirror.) The collecting lens contributes a positive spherical aberration, while the Mangin mirror contributes a negative one. For a smaller relay magnification, the Mangin mirror must be more powerful and will contribute a negative spherical aberration of greater magnitude. Final iteration of this design was obtained with the ACCOS V ray-tracing code.

Author: D. W. Phillion

Major Contributor: L. Seppala

X-Ray Phase Lens

We report an important early step toward the development of efficient, high-resolution x-ray lensing elements. We have designed and fabricated the first Fresnel phase plate for use at x-ray wavelengths.

Workers in a variety of research disciplines, from plasma diagnostics and x-ray microscopy to materials microanalysis and x-ray lithography, have a strong interest in the development of efficient, high-resolution x-ray lensing elements. Fresnel zone plates (FZP) have already been used as high-resolution imaging components.^{27,28} However, zone plates, in common with all

Fig. 5-24. SEM micrographs of the silver phase lens. The 5500-Å-thick structure has 100 zones, with a minimum zone width of 3200 Å.

diffractive optical elements, suffer from efficiency limitations. The inefficiency of zone plates arises from their multi-ordered output. Specifically, an ideal zone plate has a first-order diffraction efficiency of only 10%, while 25% of the radiation incident on the FZP is undiffracted, and appears as a troublesome background illumination in the image plane. The diffraction efficiency of Fresnel structures can be significantly improved by the judicious use of phase effects.^{29,30} Replacement of the opaque zones of a Fresnel zone plate with transparent, phase-shifting zones can, under ideal conditions (π phase shift and negligible absorption) increase the first-order diffraction efficiency to 40% while virtually eliminating the zero-order background in the image plane.

The structure we have fabricated is made of silver approximately 5500 Å thick, and has 100 zones, with a minimum zone width of 3200 Å. It is designed to operate at the aluminum $K\alpha$ line ($E \approx 1.5$ keV). The phase lens is shown in Fig. 5-24. Its calculated operating characteristics at 1.5 keV are as follows:

Attenuation parameter (η)	0.16
Energy fraction in first-order focus (e_1)	0.26
Energy fraction in zero order (e_0)	0.04
Energy fraction absorbed (e_{ab})	0.32
Energy fraction in other orders (Σe_i)	0.38
Spatial resolution (Δ)	0.4 μm
f/number	387

Fabrication of the silver phase lens involves three major process steps:

1. Generation of a zone plate x-ray lithography mask, using electron-beam lithography to write the zone plate pattern.³¹

2. Replication of the zone plate mask into a thin PMMA pattern on a plating base, using x-ray lithography.

3. Silver electroplating the zone plate pattern to a controlled thickness in the PMMA mold.

Details of phase lens design and fabrication are reported elsewhere.³² We are planning resolution and efficiency tests.

Authors: N. M. Ceglio, A. M. Hawryluk, and M. Schattenburg

Time-Resolved Crystal Spectrography

We have built an x-ray streak crystal spectrograph for making time-resolved x-ray spectral measurements. The instrument uses an elliptically curved crystal to concentrate and focus the x rays onto the entrance slit of an x-ray streak camera. With this spectrograph we have made spectral measurements with resolving power greater than 200 and temporal response of 20 ps. Such an instrument is valuable in diagnosing laser-produced plasmas.

X-ray spectroscopy is an important plasma diagnostic because the relative intensities of x-ray lines can be used to infer plasma temperatures and densities.³³ Moreover, if there is sufficient energy resolution, Stark-broadened individual line profiles are a plasma density diagnostic.³⁴ The streak spectrograph can also assist in developing a short-pulsed x-ray line source. These x-ray line sources are needed for making time-resolved short-pulsed absorption

measurements, such as those proposed for measuring the core size of an ICF pellet.³⁵

The instrument described here was developed for diagnostic work on our Novette laser system. It uses an elliptically curved Bragg diffraction crystal to focus and concentrate the x-ray spectrum onto the streak camera slit.³⁶ With such a design we have obtained moderately high resolving powers of $E/\Delta E > 200$ at x-ray energies around 1 keV. The instrument is versatile, being able to cover spectral regions from 200 eV to 10 keV. This range is achieved by using various geometries that change the Bragg angle in the range from 11 to 38 deg and by using various diffracting crystals. For the first application, we used a KAP crystal ($2d = 26.632 \text{ \AA}$), permitting us to measure spectra in the 1-keV range.

The spectrograph design takes advantage of the point-to-point focusing properties of an ellipse. X rays emitted from a laser-produced plasma at one focus of the ellipse are reflected from the elliptical surface and converge on the other focus (Fig. 5-25). A Bragg diffraction crystal is elastically bent to conform to the elliptical surface. Each ray is reflected from the surface at its unique Bragg angle, providing wavelength selection. The streak camera in the far field behind the focal spot measures the dispersed spectrum. The dispersion of the system depends on the geometry of the ellipse and the distance of the streak camera slit from the focus.

We use a modified version of the standard LLNL x-ray streak camera as a detector. The modifications are to the front end and not to the electrooptics. The flange-to-cathode distance has been shortened by eliminating the gate valve from the original design. The shorter front end allows a larger range of spectral coverage by moving the slit closer to the crystal. Also, the cathode slit length has been extended to 2.5 cm for covering a larger spectral range. This extension results in a curvature of the slit image. The curvature is a dynamic effect and is not observed in the static mode of operation. The curvature, much greater than would be expected from the difference in transit times of the electrons in the tube, is a result of nonuniform fringing fields in the deflection plates.

The focusing properties of the spectrograph have been tested with a standard

x-ray source. A copper L x-ray source is placed at approximately the correct target-to-crystal distance, as in the Novette target chamber. The diffracted x rays are recorded on film, as shown in Fig. 5-26. The three parts of Fig. 5-26 correspond to crystal-to-film distances of 5, 10, and 20 cm, which, in turn, correspond to positions of the near-field, at-focus, and far-field, respectively. As the film is moved away from the source, the positions of the $L\alpha$ and $L\beta$ lines reverse, indicating the focusing properties predicted by the elliptical geometry.

Densitometer scans through the spectrum in the far field are shown in Fig. 5-27. The energy of the copper $L\alpha$ and $L\beta$ lines is 930 eV and 950 eV, respectively, indicating a spectral dispersion of 7.14 eV/mm. The copper $L\alpha$ line exposure slightly saturates the film, but the $L\beta$ line can be used to determine resolution. The measured $L\beta$ line width is 5 eV, which implies a resolving power of 190. This estimate includes effects due to the source size, which for these tests was 500 μm . Also, the test geometry does not exactly duplicate the real geometry, and there may be broadening due to aberrations of the elliptical crystal.

The first tests of the streak spectrograph have been to measure L-shell spectra around 1 keV from Cu, Fe, and Ni which are potential pumping sources for an x-ray laser. Typical intensities on target are

Fig. 5-25. Schematic of the spectrograph design, showing elliptical focusing of x rays from target onto the streak camera slit.

Fig. 5-26. X rays recorded at three crystal-to-slit distances, showing reversal of the positions of the copper $L\alpha$ and $L\beta$ lines.

Fig. 5-27. Density measurement of the far-field data of Fig. 5-25, showing the spectrograph resolving power.

Fig. 5-28. Time-resolved data from the streak spectrograph. (a) Raw data; (b) digitized spectrum.

10^{14} W/cm^2 . An example of the chromium data is shown in Fig. 5-28. On the left is streak camera output. The image curvature discussed above is quite noticeable. The total curvature is on the order of 100 ps. The data are reduced by measuring the density of the two-dimensional image and converting film density to intensity with a calibration wedge placed on the data. The time curvature is corrected digitally with an accuracy of about 20 ps. The data are then corrected for the dispersion and efficiency of the spectrometer and for the detection efficiency of the streak camera. As an example of the final determination of the reduced x-ray spectrum as a function of time, Fig. 5-28(b) shows the chromium spectrum around the peak of the pulse from the data in Fig. 5-28(a). We can also obtain time histories of the different lines to compare trends in emission profiles.

The feature at 980 eV in Fig. 5-28(b) is a doublet separation of 3 eV is just barely resolvable, which implies a resolving power of about 300. This resolving power is at least as good as that measured on the x-ray source. This improvement may be due to the smaller source size from the laser-produced plasma.

Author: R. L. Kauffman

Imaging 20- to 100-keV X-Ray Emissions from Laser Fusion Targets

Images of high-energy x-ray emission from laser fusion targets reveal features of transport of hot/superhot electrons. By

Fig. 5-29. A schematic of a high-sensitivity microchannel plate pinhole camera for x-ray imaging.

pre-heating the target, these electrons may be detrimental to efficient target implosion. We report here successful imaging of 20-, 45-, 70-, and 100-keV x-ray emissions from fusion targets at the Novette laser facility.

Our imaging device is a four-pinhole camera with a microchannel plate (MCP) for x-ray detection and signal amplification (Fig. 5-29). The pinholes are 127, 254, 508 and 508 μm in diameter on a 1.5-mm-thick gold disk. The axes of the pinholes converge at the target.

Since the sensitivity of the MCP is nearly flat for 20- to 100-keV x rays,¹¹ the energy coverage of each channel is determined by the K-edge filters (50- μm silver, 381- μm dysprosium, 254- μm gold, and 381- μm uranium) and the suprathermal x-ray spectrum, which is typically an exponentially decaying function of the x-ray energy. For example, Fig. 5-30 shows the photon transmission for the 45-keV channel using a 381- μm dysprosium filter and a suprathermal x-ray spectrum of 24-keV temperature.

Figure 5-31 is an example of the images obtained for a gold disk target irradiated

Fig. 5-30. The photon distribution for the 47-keV channel using a 381- μm dysprosium filter and a suprathermal x-ray spectrum of 24 keV temperature.

with the 0.53- μm Novette laser at 4.4 kJ and $3.5 \times 10^{15} \text{ W/cm}^2$. At 70 and 100 keV, the x-ray flux is low, and relatively large pinholes are required even though the MCP is sensitive enough for detection of a single photoelectron. For planar targets, the suprathermal x rays appear to originate at the center of the region on which the laser is focused.

Our imaging system is more sensitive than conventional pinhole cameras. In

Diagnostics

addition, because we can gate the MCP we can develop our camera into a very

Fig. 5-31. Images of a gold disk target irradiated with $0.53 \mu\text{m}$ laser light at 4.4 kJ and $3.5 \times 10^{15} \text{ W/cm}^2$.

Fig. 5-32. Relay from hard aperture to $f/4$ lens.

sensitive time-resolved framing camera for high-energy x rays.

Authors: C. L. Wang, G. R. Leipelt, and D. G. Nilson

Co-Propagating Beam on Novette for X-Ray Backlighting

By making use of its relay architecture, we have provided Novette with an independently focusable x-ray backlighting beam. As constructed, this beam fulfills all of the backlighter requirements without needing a separate amplifier chain.

As an alternative to the expensive method of constructing a separate laser chain for the x-ray backlighter beam, we have chosen to use a portion of one of Novette's two beams. We retain the flexibility of having independent pulse length, timing, and focusing for the backlighter beam, but we pay the cost of having less energy on target and of having a hole in one of the two main beams. The backlighter pulse is frequency-doubled in the same crystal array

as the co-propagating main beam. The backlighter beam takes up one entire side segment of the north beam 3×3 KDP crystal array, which covers about 15% of the total beam area. The basic scheme is shown in Fig. 5-32.

Table 5-4 summarizes the criteria that went into this choice. This article describes the backlighter beam design, but not the planned experiments that will use the backlighter.

Novette's 74-cm output optics permit the propagation of laser beams with areas of about 3500 cm^2 that can deliver to the target about 5 TW of $0.53\text{-}\mu\text{m}$ light per beam. The maximum power and energy for the backlighter beam are of course about 15% of the corresponding numbers for a single Novette beam.

In the Master Oscillator Room the backlighter pulse is injected into the common beamline, lagging the main pulse by 15 to 20 ns. The two pulses co-propagate up to a Pockels cell (PC5-2). The input and output polarizers are parallel for PC5-2 (or, rather, the crossed polarizers are made to appear parallel to the beam by the use of a 90-deg quartz rotator). Up to PC5-2, the two beams have the same spatial characteristics but are temporally multiplexed. The leading edge of the high-voltage gate to PC5-2 is midway between the main pulse and the lagging backlighter pulse. The backlighter pulse is switched out and reinjected at the north 26-mm hard aperture, which is farther down the beamline. The main beam is transmitted through the output polarizer for PC5-2 and then goes through a splitter, which is the starting point for the individual north and south beamlines. Each beam then goes through a trombone delay line followed by a 26-mm-diameter hard aperture. Figure 5-33 shows the extraction and reinjection of the backlighter pulse. The backlighter pulse is steered a short distance down the north side of the center space frame, while the beam that will become the north main beam has to propagate to the end of the trombone line and back again. Thus the backlighter pulse, when reinjected at the 26-mm hard aperture, is synchronous with the main pulse. At this point, the backlighter pulse and main beam are spatially multiplexed.

Item	Reason
Independent timing and pulse width	Needed for flash radiography and desirable for streaked backlighting, since the peak target compression generally occurs well after the peak of the driving laser pulse.
Frequency doubling	X-ray line conversion efficiencies are much higher at $0.53 \mu\text{m}$ than at $1.05 \mu\text{m}$, particularly when $<100 \text{ ps}$ x-ray flashes are needed.
Independent focusing	Different experiments will use different x-ray lines and targets of different size. Spot diameters from 1.0 to $600 \mu\text{m}$ are needed and the backlighter focus must be capable of lateral displacement of several millimeters with respect to the drive beam focus.

▲ Table 5-4. Backlighter beam requirements.

Fig. 5-33. Backlighter beam extraction and reinjection. The mirror for the reinjection is 10 mm square in its projection along the beam direction. The relaying of the Novette laser causes this 10-mm-square projection to be imaged onto one 27-cm-square side segment of the north 3×3 KDP crystal array.

Novette uses spatial-filter apertures to expand the beam between stages of differing aperture. These spatial filters are evacuated Keplerian telescopes whose focal planes are fitted with aperture stops. The stops remove high spatial frequencies, which constitute a beam instability and otherwise would grow because the index of refraction of transmitting components depends on intensity. The 26-mm hard aperture is just before the input lens of the $3F5$ spatial filter. This spatial filter removes the diffraction fringes produced by the sharp edges of the reinjection mirror. The plane in which the 26-mm hard aperture lies is reimaged at the input lens of every spatial filter as well as at the target chamber. The relayed image of the injection mirror precisely fills one side segment of the north 3×3 KDP crystal array.

Independent focusing of the backlighter beam is achieved by a weak lens doublet between the KDP crystal array and the $74\text{-cm-diam } f/4$ lens doublet (Fig. 5-34). The two lenses form a Galilean telescope that shrinks the beam by about 7% when the spacing between them is such as to make

Fig. 5-34. (a) Location of the Novette backlighter focusing lens. (b) Focusing lens assembly slide mechanism.

the telescope afocal. Both lenses are square. The edges of the first, positive lens hide in the shadow cast by the bead-blasted glass apodizer. This lens is stationary. The second, negative lens can be smaller, because the beam is smaller, and yet still have room to move around while keeping its edges in shadow. The independent focusing of the backlighter beam is achieved by stepping-motor-controlled xyz motion of this second, negative lens.

The lens doublet can be moved in and out of position on a ball slide. The entire lens doublet assembly is mounted on a carriage that slides freely along two parallel rails that are supported on linear, recirculating ball bushings.

Laser fields in an amplifier having circular apertures, such as Novette, may readily be represented by an expansion in terms of Laguerre-Gaussian modes. Adopting the notation of Ref. 37, the scalar field, u , may be written:

$$u(r, \theta, z) = \sum_{p,l} \tilde{a}_{pl} u_{pl}(r, \theta, z), \quad (6)$$

where

$$u_{pl}(r, \theta, z) = \frac{2}{\sqrt{1 + \delta_{0l}}} \frac{p!}{\pi(l + p)!} \frac{1}{w(z)} \times \left(\frac{\sqrt{2}r}{w(z)} \right)^l L_p^l \left(\frac{2r^2}{w^2(z)} \right) \begin{pmatrix} \cos l \theta \\ \sin l \theta \end{pmatrix} \times e^{-i(k/2)r^2/\tilde{q}(z)} e^{-jkz + i(2p+l+1)\psi(z)}, \quad (7)$$

where p and l are integers, L_p^l is the associated Laguerre polynomial, and δ_{0l} is 1 for $l = 0$ and 0 for $l > 0$. Given a set of modal amplitudes, $\tilde{a}_{pl}$, free propagation between obstructing apertures may be calculated by keeping track of changes in the beam spot size parameter, $w(z)$, and in the modal phase, $\psi(z)$, with distance. Within the Fresnel degree of approximation, these quantities are determined by the Rayleigh range, z_R , and the distance, z , to the spherical wave origin or "beam waist":

$$\psi(z) = \arctan(z/z_R), \quad (8)$$

$$w(z) = w(0) \sqrt{1 + \left(\frac{z}{z_R} \right)^2} \quad (9)$$

One convenient way to find z and z_R anywhere in an optical system like Novette is to employ ray matrices. Such a system of simple lenses and distances can always be described by an $ABCD$ ray matrix that has unit determinant. A general property of any unit-determinant matrix is that it can be decomposed into a product of three matrices, as follows:

$$\begin{bmatrix} A & B \\ C & D \end{bmatrix} = \begin{bmatrix} 1 & B/D \\ 0 & 1 \end{bmatrix} \begin{bmatrix} 1 & 0 \\ CD & 1 \end{bmatrix} \begin{bmatrix} 1/D & 0 \\ 0 & D \end{bmatrix} \quad (10)$$

(distance) (lens) (relay)

Of course, other decompositions exist that order the simple distance, lens, and relay ray matrices differently. The choice above

avoids ambiguity, because a relay having a magnification of $1/D$ leaves the modal phase of the input field unchanged. The lens of Eq. (10), of focal length $-1/CD$, shifts z and z_R without altering the beam profile. Finally, the beam propagates a distance B/D to the output plane.

Formally, the equation

$$\tilde{q}_{out} = \frac{A\tilde{q}_{in} + B}{C\tilde{q}_{in} + D} \quad (11)$$

is to be applied for each of the three matrices in the decomposition. The middle ray matrix, representing the lens, requires special treatment. The phases of the complex coefficients $\tilde{a}_{pl}$ in Eq. (6) must be changed

so as to keep the same relative phases of the terms $\tilde{a}_{pl} u_{pl}$ before and after the lens. This is necessary because the lens changes z and z_R in such a way as to make the relative phases of the new normal modes u_{pl} different from the old. The relay does not change the ratio z/z_R and so does not change $\psi(z)$. Hence, the relay preserves the relative phases of the u_{pl} as defined by Eq. (7).

The ray matrix that maps rays from the 26-mm-diam hard aperture to the target focusing lens is

$$M = \begin{bmatrix} -27.4 & -7562.2 \\ 0 & -0.0365 \end{bmatrix} \quad (12)$$

Fig. 5-35. Beam profiles at several points along the north beamline. The diffraction code used to obtain these results demands spherical symmetry, so we took the backlighter beam as the center 10-mm-diam circle at the 26-mm hard aperture.

which shows that the Novette power amplifier has a magnification of 27.4 and that the target focusing lens is located only 276 mm inside the ideal relay plane. Indeed, the component that sets the profile of both pulses is the input spatial filter, with its 2-mm pinhole. Figure 5-35 shows calculated beam profiles at key locations within the laser. After the first spatial filter relay (SF5) in the power amplifier chain, the beam profiles are set and are changed little theoretically when they reach the conversion crystals.

Authors: K. R. Manes and D. W. Phillion

Major Contributors: F. A. Frick,
D. J. Kuizenga, D. J. Uroff; and
J. Victoria (EG&G)

50-ps Neutron Diode

A neutron diode is similar to an x-ray diode, but is sensitive to neutrons. Such a device is of interest because it is easy to fabricate and the system (detector-cable-recorder) resolution should be very good—about 115 ps when the recorder is a Thomson 4-GHz oscilloscope. When we can record with a neutron diode and a fast optical or x-ray diode, we will be able to accurately measure interval times and average implosion velocities. In this article we report our initial evaluation and test of a neutron diode. We selected the 50-ps diode previously designed and fabricated for x-ray applications³⁸ and converted it to a neutron diode by coating its aluminum cathode with ²³⁸U.

One of the principal concerns is the noise that may be encountered by electronic devices near the target. To address this issue, we have carried out experiments with a neutron diode inside the Novette chamber. We placed the neutron diode 40 cm from the target, at a re-entry tube inside the Novette chamber. The tube is 40 cm long and 10 cm in diameter, with a 3-mm-thick iron wall.

The test results show that in the Novette chamber environment (8 kJ, 0.1 to 1 ns, 0.53- μ m laser), with a 1.2-mm iron shield in front of the diode, the noise level is negligible (<5 mV). We detected no EMP background. X-ray signals were observed with a 12.7- μ m aluminum shield, assuring us that the diode was operating properly.

The mechanism of signal generation with the uranium cathode is described in some detail in Ref. 39. In brief, fission fragments of neutron-uranium interaction drag out a large number of secondary electrons (200 to 400 electrons per fission fragment for 1- to 10- μ m-thick uranium). The signal transmitted to an oscilloscope through a cable of characteristic impedance Z is given by

$$V = QZ/t, \quad (13)$$

where Q is the charge emitted by the cathode during time t , the duration of incident neutrons. Q is given (in coulombs) by

$$Q = 200 \times 1.6 \times 10^{-19} \sigma N_n N_U x, \quad (14)$$

where σ is the fission cross section, N_n is the number of neutrons incident on the cathode, N_U is the number of uranium atoms/cm², and x is the thickness of the uranium coating. We choose $x = 10 \mu\text{m}$, approximately the range of fission fragments. N_U is given by

$$N_U = N_A \rho / M_U, \quad (15)$$

where N_A is Avogadro's number, M_U is the mass number, and ρ is the density of the uranium coating, estimated at 8.9 g/cm³ from the work of Anno.⁴⁰

For $Z = 50 \Omega$, $t = 100$ ps, $\sigma = 1.15$ b, a cathode of 0.79 cm² area placed at 10 cm from a target that emits 10^{12} 14-MeV neutrons isotropically, V is calculated to be 0.26 V, an easily readable signal. The overall time resolution is $(50^2 + t_c^2 + t_s^2)^{1/2} = 114$ ps, where $t_c = 20$ ps and $t_s = 100$ ps are the impulse response times for the cable and the Thomson oscilloscope, respectively.

In conclusion, the combination of 50-ps neutron diode and fast oscilloscope may provide a detector system of 114-ps resolution and a detection threshold of about 10^{12} neutrons from the fusion target.

Authors: C. L. Wang, D. E. Campbell,
G. R. Leipelt, D. G. Nilson, and V. W.
Slivinsky

Time-Resolved Spectral Measurements Above 80 Å

We have built a time-resolved XUV spectrograph as the principal diagnostic for measuring x-ray lasing in the 80 to 300 Å range

for the Novette x-ray laser experiments (see "Novette Soft X-Ray Laser Experiments," later in this section.) The principal components of the spectrograph are a cylindrically curved x-ray mirror for light collection, a transmission grating for wavelength dispersion, and a soft x-ray streak camera for x-ray detection. These components have been combined to produce an instrument with high spectral resolution ($E/\Delta E > 200$), continuous wavelength coverage from 80 to 300 Å, good time resolution (~ 20 fs), and high sensitivity ($\sim 10^{20}$ photons/s-sr). In addition, because the instrument measures x-ray lasing output, we can align it to the target with an accuracy of better than 1 mrad.

Figure 5-36 is a schematic of the instrument design. X rays from the target are collected and focused at the detection plane by a cylindrically curved grazing-incidence x-ray mirror. The mirror acts like one element of a Kirkpatrick-Baez x-ray microscope, producing a line focus perpendicular to the plane of dispersion.⁴¹ The mirror has an 8-m radius of curvature, operates at an angle of incidence of 4 deg, and is 57 cm from the target, nearly midway between the target and the detection circle. This arrangement produces almost unit magnification along the focal direction. Spherical aberrations for these conditions, estimated to be less than 10 μm , should not significantly affect the resolving power of the instrument.

The dispersing element of the system is a free-standing transmission grating placed directly behind the x-ray mirror.⁴² The grating disperses the x rays from the zeroth order established by the reflected direction of the mirror. The grating used for the experiment has a periodicity of 2000 Å and a line-to-space ratio of 1.9. The grating is 62 cm from the detection plane, which gives a focal-plane dispersion of 3.2 Å/mm. The active area of the grating is about 3 mm in diameter; the dimensions of the mirror overfill the grating area with x rays.

An LLNL-designed soft x-ray streak camera is used as a detector.⁴ The camera is designed to pivot around the grating from the zeroth order to 9 deg to allow continuous coverage through the desired wavelength region. The camera has a temporal response of 20 ps, providing good time resolution for the experiment. The camera's entrance slit is aligned with the dispersion plane of the

▲ Fig. 5-36. Schematic of the transmission-grating streak spectrograph for measuring spectra above 80 Å.

Fig. 5-37. Photocathode response of 900 Å CsI.

grating. The 1.2-cm slit allows a 38-Å portion of the spectrum to be measured at one time. To cover other regions of the spectrum, the streak camera must be moved along the detection plane. The 100- μm resolution of the streak camera along the cathode slit translates to a wavelength resolution of 0.32 Å. Neglecting source size effects, the resolving power, $E/\Delta E$, is 250 for 80 Å and increases for longer wavelengths.

For these experiments, the streak-camera photocathode is 900-Å CsI film deposited on a 1000-Å Parylene substrate. A 50-Å layer of aluminum is placed over the Parylene to increase electron conduction. Cesium iodide has been shown to have high sensitivity.⁹ We have estimated the quantum efficiency of the cathode (Fig. 5-37). For these estimates we use the measured front-surface yields of CsI (Ref. 43) and the expression of Henke et al.⁴ to infer rear-surface yields, which are needed for x-ray streak-camera efficiencies. The calculated back-to-front ratio is also shown in Fig. 5-37. It is interesting to note that the large increase in the CsI yield around 100 Å observed in the front-surface yield is moderated in the rear-surface yield. This increase in yield is due to increased absorption in the CsI, which decreases the number of photons reaching the rear surface of the cathode.

Diagnostics

The instrument has been calibrated in order to make absolute intensity measurements. The streak-camera signal, S (intensity), is related to the target flux, F (in $\text{W}/\text{sr} \cdot \text{keV}$) at a given energy, E , by

$$S = d\Omega \xi \eta R(E) \alpha (dE/E) F \quad (16)$$

The term $d\Omega$ is the solid angle subtended by a resolution element of the streak camera, including the focusing properties of the cylindrical mirror. The mirror reflectivity ξ has been measured at energies above 200 eV and is 30% of the theoretical value. We have used the theoretical curve to extrapolate to lower energy, correcting for the 30% decrease. Reflectivities in this range are greater than 50%, so that small errors in reflectivity do not greatly change the instrument sensitivity. The grating-diffraction efficiency, η is calculated to be nearly constant at 0.025 over the energy range of interest. Calibrations of the grating are in progress. $R(E)\alpha$ is the streak-camera detection efficiency, where $R(E)$ is the photocathode detection efficiency in electrons/photon, and α is the streak-camera response, which has been measured using the calibration technique previously established.⁵ The calibrations are done using broad-band channels around 200 eV. We have extrapolated the cathode response $R(E)$ using the curves in Fig. 5-37. The

minimum sensitivity of the instrument ranges from $10^8 \text{ W}/\text{sr} \cdot \text{keV}$ at 250 eV to $8 \times 10^6 \text{ W}/\text{sr} \cdot \text{keV}$ at 80 Å. Errors on the intensity measurements are estimated to be about a factor of 2.

Figure 5-38(a) demonstrates performance of the instrument and shows an example of the data. The spectrograph is viewing the edge of a Formvar foil irradiated on both sides by 100-ps pulses of 450 J of 0.525- μm light focused in a cylindrical spot 2 by 12 mm. (See the article "Electron Heating from the Ion-Acoustic-Decay Instability in a Finite-Length System," at the end of this section.) The intense line feature in the middle of the cathode is the 3p-2s transition from O VIII at 102.4 Å. The result shows that the emission turns on early during the irradiation pulse and persists for about 2 ns. Spectra at two different times are shown in Fig. 5-38(b). The upper trace represents emissions early in the laser pulse. The line is relatively sharp and is associated with a high continuum level. Later, as shown in the lower trace, the continuum level is reduced, but the line is equally intense, although slightly wider. It appears that the emission continues after the irradiation as the foil blows up. As the foil expands, the line broadens because of the larger source.

In summary, we have built a streak spectrograph for time-resolving laser-plasma

Fig. 5-38. (a) Example of the streak-camera data for the O VIII 3-2 line from a Formvar foil, and (b) the time-resolved spectra.

spectra above 80 Å. It uses a cylindrically curved mirror for light collection, a transmission grating for spectral dispersion, and a soft x-ray streak camera for time-resolved detection. We are calibrating the system, but we estimate that we can measure spectral intensities of about 10^7 to 10^8 W/sr-keV. We plan to increase the sensitivity of the instrument by increasing the collection efficiency of the mirror optic.

Authors: R. L. Kauffman, N. M. Ceglio, and H. Medeck

Data Management and Analysis

Introduction

During 1983 we passed two significant milestones in the management and analysis of data from laser fusion experiments. The most important milestone was the initiation of classified data operations with our fusion experiments analysis facility (FEAF) VAX computer. This event allowed us for the first time to process all experiments on a single computer with the same processing software. Our physics and target-parameter data bases have been expanded to include classified data. We also established facilities to process classified film data. Users can now process many types of film data much faster than on the Octopus system. For example, color-enhanced images of targets are available within a few minutes of obtaining digitized film data. An important aspect of our enhanced data operations is the successful operation of the secure one-way fiber-optic link for sending data files from the Novette computer to FEAF. Hundreds of data and parameter files have been transmitted to FEAF in a reliable, automated, and rapid manner. The link has experienced no downtime. We have not had to carry magnetic tape or disk packs from Novette to FEAF.

The second milestone is the initiation of development of an "electronic shotbook" that will allow physicists to examine shot data on the screens of computer terminals,

as well as in the traditional hard copy shotbook. The electronic shotbook is limited to final processed physics and shot summary parameters. Using several FEAF VAX utility programs, the user will access raw data in digital form. In addition, tables and/or plots of raw data and partially processed data will be kept in the hard copy volumes. The electronic shotbook development is described in detail in the article "Electronic Shotbook"; the main features are summarized here.

The ORACLE data base management system described in the 1981 and 1982 Laser Annuals is the working center of the shotbook. (We describe the use of ORACLE in or present activities in the next article.) Physics and shot summary data are contained in ORACLE data base tables and files of two-dimensioned (x, y) and three-dimensioned (x, y, z data.) The user interface to the shotbook is menu-driven; all options and instructions are presented on a terminal to the physicist. Selection of an option is accomplished by positioning a cursor on the screen next to the option.

We have done a great deal of work to improve data processing. Software has been written to process data from new diagnostics such as the streak crystal spectrograph (see "Processing of Data from the Streaked Crystal Spectrograph") and the transmission-grating streak spectrograph (see Sec. 5 of the *1981 Laser Program Annual Report*). This year we implemented software to process laser diagnostic data and store the results in a form accessible by the shotbook. Raw data from laser-energy monitors and optical streak CCD cameras are sent from the Novette laser-control system via the one-way link to FEAF, where they are converted to laser power histories and energy-balance parameters. The processing is described in detail in the article "Processing of Laser Data."

Progress is also being made to implement our Eikonix film-digitizer system into FEAF for rapid turnaround in digitizing negatives and Polaroid positives. We have written software that controls the digitizer and converts the raw digitizer data (transmission through the film) to an array of film-density data. These data are stored in a file in the PDP 11/34 computer disk (the PDP 11/34

controls the Eikonix digitizer). The file, on magnetic tape, is transported to the FEAR VAX for analysis. We are preparing modifications that will permit us to transfer the file from the PDP 11/34 to the VAX via DECNET, the Digital Equipment Corporation inter-computer network. This installation will result in a fast and efficient system for film digitization and processing. During 1983 we installed an array processor in our system in preparation for an increased image-processing capability. Implementation will be completed in 1984.

**Authors: J. M. Auerbach, R. A. Lerche,
and R. W. Carey**

Applications of the ORACLE Data Base Management System

The ORACLE relational data base management system (DBMS) has become a foundation of our data processing and analysis effort at FEAR. We are now using it at every level of data management. We use it, for example, for keeping track of equipment configurations and for display and manipulation of physics data. The ORACLE DBMS is the main building block of our electronic shotbook, as described in the article "Electronic Shotbook" later in this section.

In last year's Laser Annual we described how we used ORACLE as a mapping between diagnostic detectors and the associated electronic modules that record data from the detectors. This mapping is now maintained on-line in an ORACLE data base for all shots. An experimenter who wants to verify or reprocess data can use this data base to quickly find the setup for any shot. In addition, we maintain an ORACLE data base of component and calibration data for all x-ray detectors for all shots. This allows us to rapidly reprocess x-ray data or reconstruct a past x-ray detector configuration. With ORACLE we have for the first time established a computerized archiving system for experimental setups. In fact, this is the first time a complete and up-to-date configuration archive has been achieved for our facilities.

We are also using ORACLE as a data management tool for archiving of raw data

files. A typical shot data file package consists of two CAMAC data files, two laser diagnostic data files, and several files from diagnostics that record data on film. A typical file set consumes several megabytes of disk space. With our present on-line disk storage, we cannot maintain more than a month's raw data on-line. We have to store most data off-line, so we must have a good archiving system that lets us retrieve files on tape easily if further analysis is necessary.

We are using ORACLE to keep track of data files stored on magnetic tape. An ORACLE data base table has been set to relate a data file for a specified shot number and diagnostic to the name of the magnetic tape volume on which it is stored. The table has the following columns:

SHOTNO	shot number
FILENAME	name of data file
FILE TYPE	name of diagnostic or system corresponding to data
TAPE ID	volume name of the magnetic tape that contains the data file

Using the search capabilities of ORACLE, one can quickly locate the tape on which a particular file (FILENAME) is stored or find the storage location and names of all files for a specified shot (SHOTNO). ORACLE does all searches quickly, without custom software. This system gives experimenters a fast and easy way to locate specific items of raw data. All queries are done on-line with an electronic terminal.

Authors: J. M. Auerbach and R. W. Carey

Major Contributor: B. J. DeMartini

Processing of Laser Data

During 1983 we developed and implemented software on the FEAR VAX to process in a highly automated fashion the large amounts of laser diagnostic data generated with Novette. Laser-diagnostic data consist of:

- Disk files of incident- and reflected-beam energy data for various wavelengths.
- Disk files of streak-camera CCD data arrays. The CCD data contain the time

histories of the incident and reflected laser pulses.

- Digitized film data from near- and far-field focal-spot cameras.

The first two sets of files are created on the Novette laser computer system and transmitted to the FEAF VAX via the one-way link. Film from the near- and far-field cameras is digitized on a photometric densitometer system. The results are transported to FEAF by magnetic tape. In the near future, a film digitizer system will be on-line at FEAF, allowing us to digitize and process film data in one rapid and continuous process.

The processing software described below converts the large amount of raw data from each experiment into a concise form. Outputs include the temporal and spatial distribution of laser energy on target and various average quantities, such as total energy and pulse FWHM. Average or summary parameters are stored in tables of the ORACLE data base, described in the previous article.

Laser Energy Data Files. The control system of the Novette laser is set up so that groups of laser diagnostic sensors are associated with a number of microprocessors. After shot time the laser energy data are collected from the microprocessors by the main VAX control computer and stored in ASCII disk files. Each data item is labeled by the microprocessor by which it was collected. A mapping between the microprocessor label and physical quantity is provided in a mapping section at the beginning of the file. FEAF software processes this file as follows. First the mapping data are read into memory and the location in the file of the incident and reflected energies on target are selected. These pointers are then used by the program to extract the energy values (incident and reflected energies for 1ω , 2ω , and 3ω) out of the data section of the file. The data are then inserted into the ORACLE data base table called LASER SHOT SUMMARY. The table is accessed by several processing codes, such as those for calculating energy-balance parameters and for calculating laser-power histories. After processing is complete, the raw data file is archived on magnetic tape.

Streak-Camera CCD Data Files. The Novette optical streak cameras record the

temporal history of the incident and reflected light pulses at the irradiating wavelength. A small fraction of each pulse is spatially averaged by passing through ground glass and focused onto an optical fiber that is terminated at the entrance slit of the streak camera. One streak camera handles four fibers, which contain the pulse histories for the incident and reflected portions of both beams. The four streak patterns from the pulses are recorded on a CCD mounted on the back of the camera. The CCD data are written into a binary file that is sent to FEAF via the one-way link.

Processing of the streak camera data requires specification of the CCD data file, the laser parameters from the ORACLE data base, and the camera sweep speed. The user must examine the CCD image for quality, and must specify locations for lineouts for time profiles. The user selects lineout profiles with the help of a color-enhanced image of the CCD data, produced on one of the FEAF color display units (Fig. 5-39). The four ellipses correspond to the streak patterns from the reflected beam (outer patterns) and from the incident beam (inner patterns) for a Novette two-beam shot. The user of the processing program uses a trackball to select the optimum position for taking a lineout profile in each spot. The peak of each incident pulse profile is referenced as zero time. The CCD amplitude-vs-position profile is converted into a power history (with the use of data for laser energy and camera sweep speed). Figure 5-40 shows a sample power history plot produced by the software. In addition, the software calculates energy on target as a function of time (the running integral of

Fig. 5-39. Color-enhanced display of optical streak-camera CCD data.

Data Management and Analysis

laser power). Figure 5-41 shows a sample plot of this quantity. Laser power histories are also stored in disk files for use in later

Fig. 5-40. Laser power history produced by FFAF optical streak CCD processing software.

Fig. 5-41. Laser-power running integral produced by FFAF optical streak CCD data processing software.

Fig. 5-42. Energy-fraction running integral produced by FFAF focal-spot data processing software.

analysis or for display in the electronic shotbook facility (see the article "Electronic Shotbook," later in this section).

Focal Spot Digitized Film Data. Analysis of focal-spot digitized film data was described in the 1980 Laser Annual.⁴⁴ The processing was done on the Octopus computer system. Since that time we have developed software to process all focal-spot data on the FFAF VAX. For a beam focal-spot pattern, the FFAF VAX software can compute the following:

- A D -log E curve to convert the distribution of density to a distribution of relative exposure and then, using laser parameters (beam energy and pulse width), to a distribution of intensity.
- Vertical and horizontal lineouts of the intensity distribution.
- Radial average of the intensity distribution.
- Histogram of the intensity distribution (relative occurrence vs intensity).
- A running integral of the histogram, giving the fraction of beam energy at intensities up to the intensity I (upper limit of the integral). Figure 5-42 shows a sample plot of this last quantity.

Authors: J. M. Auerbach and B. J. DeMartini

Processing of Data from the Streaked Crystal Spectrograph

Early in 1983, during the Novette x-ray laser series of experiments, data were obtained for the first time by the streak crystal spectrograph. This instrument is described in detail in another article earlier in this section, "Time-Resolved Crystal Spectrography." Here we describe the processing of the data from the spectrograph.

The data-recording medium of the streak crystal spectrograph is film. The film contains both energy-related and time-related information on x-ray emission from the target. The film position coordinate parallel to the streak direction corresponds to time, and the coordinate perpendicular to the streak direction corresponds to photon energy. The film data are digitized and stored in disk files, in which form they become the input to be processed. Each pixel value

in the data file corresponds to a value of film density. Film density is converted by the processing code to relative x-ray emission, taking into account the response of the film and the streak camera sensitivity.

To convert position to photon energy and film density to relative fluence, the processing performs the functions listed below, allowing for the geometry and material parameters of the spectrograph and streak camera and the film response (density-exposure relationship). A code named XCSSPAR1 performs as follows:

1. It calculates the dispersion relationship of the spectrometer (the photon energy corresponding to a horizontal coordinate on the streak camera photocathode). Taken into account are Bragg reflection, spectrometer geometry, and the connection geometry between spectrometer and streak camera.
2. It calculates the response function of the combination of spectrometer and streak-camera photocathode (relative response vs photon energy). The response function takes into account the following factors as a function of energy:
 - (a) the geometrical sensitivity of the spectrograph crystal.
 - (b) the reflectivity coefficient of the crystal material.
 - (c) the transmission of a thin-foil blast shield between the spectrometer and streak-camera photocathode. The transmission is calculated as a function of energy for x-ray cross section data contained in a FEAF VAX x-ray physics data library named CSLOW.
 - (d) the quantum efficiency of the streak-camera photocathode as a function of photon energy. The quantum efficiency is calculated from atomic cross sections from a second FEAF VAX physics data library for elements with atomic numbers 1 to 94 and for x-ray photon energies from 0.1 to 10 keV.

The factors (a) through (d) are multiplied together to give the total response of the streak spectrograph. The dispersion and response-function data are stored in a disk file. The file can be used to unfold any spectrogram corresponding to the same configuration data. XCSSPAR1 also produces plots of the dispersion, the total response, and the contributions to the

response of items (a) through (d) above as functions of energy.

The code for processing streak-spectrograph data files is named XCSSPROC1. It allows the user to take horizontal and/or vertical lineouts of a streak record and convert them to spectra at fixed times and/or time profiles at constant energy. These are the final forms of data needed by physicists for experimental analysis.

The conversion of film density-position data to spectra and time profiles is accomplished as follows:

First, a known spectral feature on the spectrograph is identified in order to align the dispersion relationship with the position data of the spectrograph. The user takes a horizontal lineout of the data and specifies the energy of a known spectral feature. The program then adds a correction for position so that the dispersion relationship gives the energy of the feature. The vertical spectrograph coordinates are converted to time by multiplying those values by the camera sweep speed (units of time/length).

After a user selects a lineout, the following calculations take place:

1. Horizontal lineout (spectrum for fixed time).
 - (a) Each position on the lineout is converted to energy—the conversion uses the dispersion relationship. The vertical lineout coordinate is converted to time—the conversion uses camera sweep speed.
 - (b) Film density is converted to relative exposure—the conversion uses $D\text{-log } E$ data from a file that contains wedge calibration data.
 - (c) Relative exposure is converted to relative x-ray emission—the conversion uses the spectrometer response function. Figure 5-43 shows a sample spectrum.
2. Vertical lineout (emission at fixed energy).
 - (a) Each position or lineout is converted to time—the conversion uses the camera sweep speed.
 - (b) Horizontal coordinate of lineout is converted to photon energy.
 - (c) Film density for each point of the lineout is converted to relative exposure—the conversion uses the $D\text{-log } E$ curve.
 - (d) Relative exposure is converted to relative intensity—the conversion uses the value of the spectrometer response

Fig. 5-43. Sample plot of a spectrum from the streak spectrograph data processing. Spectrum is for a single time.

Fig. 5-44. Sample plot of the time history of x-ray emission at a single photon energy.

function at the photon energy corresponding to the lineout location.

Figure 5-44 shows a sample time profile.

The processing software is written so that the user can process as many time profiles or spectra as he wants in one run of the program. He can also store each spectrum or time profile in a separate disk file for use with other analysis software or as a data source for the electronic shotbook (see the article "Electronic Shotbook" later in this section).

Author: J. M. Auerbach

UFO Code on the FEAR VAX Computer

In 1983 we implemented on the FEAR VAX computer a versatile program called the UFO (unfold operator) code that permits interactive data-manipulation and solution of integral equations. UFO was developed by Lynn Kissel⁴⁵ and Frank Biggs of Sandia National Laboratories, Albuquerque. To install the code we extensively modified the FEAR VAX graphic software and made minor modifications of the user interface.

UFO offers features desired by fusion experiments physicists and diagnosticians. Much of the experimental data consists of x, y data pairs, which must be manipulated and displayed in a variety of ways. The basic data entity of UFO is a set of x, y data pairs. UFO data sources are ASCII files containing these data pairs. In this respect, UFO is similar to the Octopus code SOCKITTOME,⁴⁶ which is used by almost all physicists in fusion experiments. The UFO code eliminates the need to convert SOCKITTOME for a VAX computer and implement graphics for many devices not used by SOCKITTOME. UFO on the FEAR VAX can produce plots on Retro-Graphics terminals, Ramtek color displays, and

the VERSATEC hard copy plotter. UFO is interfaced to the DISSPLA graphics package, a commercial graphics routine software library installed on the FEAFF VAX in 1981. In addition, the sources for UFO are written in structured form with Fortran 77, making them more maintainable than the SOCKITTOME sources.

The user interface for UFO makes use of a free-field interpreter. The user types in character strings corresponding to command key-words and parameters. The interpreter then parses the fields between blanks and processes them appropriately. If the user gives a command in incorrect form, UFO identifies the error, displays an appropriate message, and waits for another command.

Table 5-5 summarizes the capabilities of the code for manipulating files (x, y data sets). The code also allows for editing of the files without leaving the program.

UFO solves systems of integral equations of the form

$$\int_{x_1}^{x_2} R_i(x)S(x)dx = Q_i \quad (17)$$

This equation is relevant to typical x-ray diagnostics problems, where one wants to determine a spectrum $S(E)$ (E = photon energy) given a set of detector signals Q_i and a set of detector response functions $R_i(E)$.

UFO calculates $S(E)$ in terms of basis functions. The solution is assumed to be of the form

$$S(E) = \sum_{x_1}^{x_2} C_j D_j(E) \quad (18)$$

The basis functions $D_j(E)$ are specified as x, y data pairs in files. The C_j are calculated so as to give a least-squares solution.

Work is in progress to evaluate results from the UFO integral equation solver in comparison with other unfold codes now in use. One such code is the UNSPEC code on Octopus, which is used for unfolding spectral data recorded by the Dante x-ray spectrometer.

UFO can also run calculations from a string of commands stored in a file. This capability allows quick repetitive processing with different data files. Processing was

1. Interactively plot (specify graph configuration) one or more x, y data pair sets.
2. Generate cubic-spline interpolators.
3. Generate x, y data pairs from analytic expressions which can be interpolated with a preselected accuracy.
4. Integrate over a specified interval in x , or generate a cumulative (running) integral.
5. Add, subtract, multiply, or divide the entries in tables of x, y data pairs.
6. Report statistics on x, y data pairs and interpolate arbitrary intermediate values.
7. "Thin" a table of x, y data pairs by removing pairs not needed to ensure accurate interpolation.
8. Perform an arbitrary linear transformation on the x or y values of x, y data pairs on input from a disk file.
9. Generate least-squares polynomial fits.

significantly slower when the same commands had to be entered repeatedly.

Author: J. M. Auerbach

Electronic Shotbook

In 1983 we completed activation of the FEAFF computer system (fusion experiments analysis facility).⁴⁷ Its hardware and software are selected to reduce effort on manual tasks related to data collection and shotbook preparation while improving shot data turnaround time, accuracy, readability, and accessibility. In this article we report developments towards providing laser scientists with an "electronic shotbook," a system for on-line computer storage and retrieval of shot data.

Data collected from target shots at our laser facilities form the basis of our experimental program. By studying these data we are able to develop an understanding of the interactions between high power lasers and targets. The ability to have later access to these data and to analyze and compare them is essential for the experimental program.

The data from a target shot are assembled in a binder we call an experiment shotbook.⁴⁸ The data are separated by categories, such as shot summary, shot setup, target, energy balance, scattered-light spectroscopy, x rays, charged particles, and neutrons. In this form, there is only one copy for each target experiment, so

Table 5-5. Data file manipulations available in UFO.

access to shot data is cumbersome and at times difficult.

The main design objective for the electronic shotbook project is to provide rapid, easy access to experimental shot results. Our strategy is to use fully the capabilities of the sophisticated computer hardware and software available to us. Most of our scientists have FEAF graphics terminals in their offices, so there is no problem with physical access to the FEAF system. We are developing a computer program called SHOTBOOK to make access to shot results independent of a detailed knowledge of our computer and its software and data base.

For the development of SHOTBOOK we elected to follow Yourdon's structured design techniques.^{49,50} From an analysis of the requirements of users, we developed a simple user/SHOTBOOK interface, which provides the user with a combination of menus, forms, and single-keystroke commands. From a detailed survey we defined the data base tables and directories needed to hold shot information. Then with data flow analysis and structure charts we developed a modular program design.

The overall program strategy is based on the concept of an active shot list. Through a systematic sequence of menu interactions, the scientist creates a list of shots that meet a set of selection criteria. To display specific shot results, two more menu selections are necessary. Selection of a shot from the shot list causes the shot Table of Contents to be displayed on the terminal screen. Selection of the desired section from the Table of Contents results retrieves the data for that section from the data base and displays the first page of that section on the user's terminal.

When a shotbook page is displayed, the user can browse through shot results with four single-keystroke commands, called NEXT PAGE, PREVIOUS PAGE, NEXT SHOT, and PREVIOUS SHOT.

NEXT PAGE causes the next page of the current section to be displayed. If the last page of a section is currently displayed, then NEXT PAGE causes the first page of the next section listed in the Table of Contents to be displayed. PREVIOUS PAGE works in a similar manner. With these two commands the user can easily thumb through results related to a single shot.

Often it is desirable to look at data taken with one diagnostic instrument over a series of shots. The NEXT SHOT command causes the first page of the currently displayed section from the next shot in the active shot list to be displayed on the terminal. PREVIOUS SHOT functions in a similar manner.

A HARDCOPY command may be used to generate a printed copy of the current terminal displays.

Section displays are created with the DISPLA software library,^{47,51} which gives us great flexibility with display format. Individual section displays are designed to the requirements of the physicists responsible for analyzing shot results and the diagnostician responsible for analyzing data from a specific instrument. Section displays may consist of multiple pages with text and graphics. Typical text includes headings, data verification, classification, and comments. Graphics may include two- and three-dimensional plots, contour plots, and three-dimensional images. A sample page display is shown in Fig. 5-45.

Besides data retrieval and display, the electronic shotbook will provide users with facilities for data analysis. With a set of single-keystroke commands the user can change the presentation form of graphs. A CROSSHAIR command activates this set of commands. In the crosshair mode, a crosshair appears on the terminal screen at the center of the first-defined graphics area. The ADVANCE command rapidly advances the crosshair from one graph to the center of the next. If a displayed graph is too small on the standard page display, it may be expanded to the full screen with the ZOOM command. Commands are available for altering graph scale limits, scale type (linear or logarithmic), and for overlaying a lineout on a contour plot.

We use the ORACLE relational data base management system⁵² (DBMS) computer program to manage our shot data. Data to be accessed by SHOTBOOK are initially placed into the data base by our various processing codes. Over 40 tables are defined for shot-related data. The SHOTBOOK program accesses these data for use with its output displays. SHOTBOOK also provides users with access to forms for updating information stored in the data base. With the

Fig. 5-45. A sample high-energy x-ray spectrum, as produced by the SHOTBOOK program and displayed on the user's terminal.

Experiments

Introduction

During the past year most unclassified experiments at the Novette facility have been in two areas: laser-plasma coupling and soft x-ray laser research. The multiterawatt ($P_{2\omega} \lesssim 10$ TW) frequency-doubled Nd:glass Novette laser allows wavelength scaling to be extended to significantly larger plasmas than has previously been possible (with coronal temperatures >1 keV). Novette's higher laser output power also permits us to carry out x-ray laser experiments. With the use of an appropriate cylindrical lens, Novette can produce moderate intensities ($\lesssim 5 \times 10^{15}$ W/cm²) over dimensions (and potential gain lengths) exceeding 1 cm (transverse dimensions $\sim 10^{-2}$ cm).

The large-plasma experiments conducted to date on Novette are summarized in the next article by E. P. Drake and R. E. Turner. A wide variety of x-ray and optical diagnostics have enabled us to study many aspects of the physics of laser-plasma coupling in coronas having radial and calculated axial

DBMS the privilege to update data is granted only to the personnel responsible for that particular piece of information.

On-line help is available throughout the SHOTBOOK program with the HELP command. Our help messages, menus, and forms are generated with DEC's forms management system (FMS) software.⁵³

The motivation for developing an electronic shotbook comes from the potential benefits, such as improved turnaround time and better accuracy, readability, and accessibility of data. The electronic shotbook is an attractive alternative to the present loose-leaf binder form of the experiment shotbook. In 1983 we completed the system analysis and design for the SHOTBOOK program. The program modules controlling the user interface, program commands, and data base access are operational. Only the individual section displays remain to be programmed. We expect SHOTBOOK to be available for use by the end of the first quarter of 1984.

Authors: R. A. Lerche, J. M. Auerbach, and R. W. Carey

Major Contributors: E. J. DeMartini, M. M. Wakakuwa, and F. Ze

Experiments

dimensions exceeding 300 μm . Previous wavelength scaling experiments conducted at LLNL and elsewhere were limited to plasmas having characteristic dimensions [$L \approx (\nabla \ln n_e)^{-1} \approx \text{spot radius}$] less than 100 μm , where n_e is one quarter the critical density. Since many of the instabilities that can degrade the coupling (absorption and suprathermal-electron production) are velocity- and density-gradient stabilized and have small gain lengths in these small-volume coronas, the importance of extending the experiments to larger plasmas to simulate larger targets for larger lasers is obvious.

As a consequence of the larger interaction regions, we have seen several features of coupling physics previously observed only at 1.06 μm . For example, while the absorption by gold and beryllium disk targets on Novette agrees well with Argus data, stimulated Brillouin sidescattering is clearly observed at Novette. This observation is a direct consequence of the large radial and axial dimensions of the Novette plasmas. Thus while the total absorption remains unchanged, the mix of inverse bremsstrahlung and Brillouin backscattering and sidescattering is different.

Experimental data (discussed in the following article) show that the stimulated Raman instability (SRS) can reach efficiencies of approximately 1%. This represents an increase of a factor of 100 over that seen previously for sub-micron light. The experiments also show an excellent correlation between the detected optical energy from SRS and the inferred energy in suprathermal electrons as measured from the x-ray bremsstrahlung spectrum.

Attempts at creating even larger underdense plasmas by using exploding foils are also discussed in the following article. Many experiments (including the use of tracer dots for x-ray spectroscopy) have been performed, but analysis of the data has only just begun. Nonetheless several interesting features have already been observed. SRS yields greater than 5% have been observed from 1200- $\text{\AA}$ gold foils irradiated at peak intensities of $2 \times 10^{15} \text{ W/cm}^2$. These targets should produce extremely flat density profiles at electron densities of $\sim 0.2 n_c$; this is consistent with

the high SRS yields. In addition, the time histories of the SRS light and the 50-keV x-ray emission, measured on a specially configured optical streak camera (see "Streaked Recording of Simultaneous Optical and X-Ray Signals," earlier in this section) are well correlated for the exploding-gold-foil experiments.

A lack of detailed diagnosis of the plasma temperature and density profiles (spatial and temporal) limits our understanding of laser-plasma coupling and energy transport. In principle, x-ray spectroscopy can yield such information, and we are developing our capability to both measure and analyze relevant spectra. In "Line Diagnostics of Laser-Produced Plasmas," we present results of analysis of time-resolved, space-integrated emission spectroscopy. The results clearly show the need to space-resolve or space-localize the spectra so that averaging over the severe gradients near n_c and the ablation front are minimized.

The articles "Time-Resolved Spectral Measurements Above 80 $\text{\AA}$ " and "X-Ray Laser 'Fail Safe' Diagnostics: Performance and Results," in this section, present the results of our first attempts at producing single-pass extreme ultraviolet (XUV) and soft x-ray lasers. The experiments concentrated on the resonant photoexcitation scheme using either fluorine or neon gas as the laser medium. A wide variety of flash-lamp materials was employed, with particular emphasis on L-shell transitions in nickel and chromium and K-shell emission from sodium. No output signals above the threshold of the laser diagnostics were detected in these experiments (see "Time-Resolved Spectral Measurements Above 80 $\text{\AA}$," earlier in this section). Instrument calibrations have determined that this threshold corresponds to less than about 5 to 7 gain lengths of amplification for the various schemes that were examined.

In addition to the photopumped laser scheme, several experiments were performed to examine possible electron-collisional and three-body recombination lasers. The former scheme uses electron monopole excitation to create inversions in neon-like selenium in the corona of a laser-produced plasma. No evidence for lasing was

detected in these experiments; one possible explanation is refraction of the x rays out of the gain medium due to the steep axial gradient of the laser-produced plasma.

Finally, experimental evidence for a population inversion between the $n = 3$ and $n = 2$ levels in hydrogen-like oxygen was obtained on several shots that examined 3-body recombination and cascade as the principal pumping mechanism. The inversion resulted from the rapid non-equilibrium expansion of a Formvar foil heated by Novette. These experiments are also discussed in the article "Novette Soft X-Ray Laser Experiments."

Author: E. M. Campbell

Novette Laser-Plasma Interaction Experiments

This article summarizes the unclassified laser-plasma interaction experiments performed on Novette during 1983 and presents some preliminary results. These experiments will be completed during early 1984, and final analysis will be reported in the next Laser Annual.

The experiments were intended to explore laser-plasma interaction physics in large, Nova-size plasmas. To do this, we used the Novette laser to irradiate solid (disk) targets and to explode thin gold or plastic foils. The disk targets produced axial plasma scalelengths hundreds of laser wavelengths long, and the foil targets produced plasma scalelengths of several thousand laser wavelengths, as illustrated in Fig. 1-4. A plasma of this size is large enough to approach or exceed the thresholds for many laser-plasma instabilities. Figure 1-4 shows the thresholds for several of them; the threshold is exceeded for points above or to the right of the curves on the figure. These thresholds are calculated on the assumption of a plasma with a charge (Z) of 4 and with specific model-dependent parameters. For example, if the drift velocity of the plasma is close to the sound speed, then the Brillouin threshold is much lower than shown. Radial scalelengths are also important, because they may allow growth of sidescatter

instabilities. The Novette plasmas studied are hundreds to thousands of laser wavelengths in radial extent, permitting large growth lengths.

Our disk-target experiments showed that solid targets irradiated with large spots can produce significant scattered light and hot-electron fractions. For average intensities of $\sim 10^{15}$ W/cm², more than 20% of the incident green light was scattered from the target, with much of the scattering detected at angles greater than 30 deg from the direct backscattered direction. As noted below, we believe that most of this scattering is Brillouin scattering. At similar intensities, as much as 1% of the incident laser energy appeared as broadband Raman scattering (at wavelengths below 1 μ m), while at most 0.01% was scattered at $3/2 \omega_0$. The scattered $3/2 \omega_0$ light is indicative of plasma wave-producing processes occurring near $n_c/4$, such as the two-plasmon decay. The hot-electron fraction was also as large as 1% and was strongly correlated with the Raman scattering. These and other data indicate that Raman scattering from densities below quarter-critical is the dominant cause of hot-electron production in these plasmas.

The exploding-foil experiments have produced data that can be compared in some detail to calculations by the LASNEX computer code. The plasma density can be estimated from the spectrum of Raman backscattering, the absolute magnitude of the bremsstrahlung emission as a function of axial distance and time, and the timing of the $3/2 \omega_0$ emission and transmitted light. During early 1984, we hope to obtain density information from x-ray line ratios as well. The temperature is indicated by line ratios obtained from x-ray spectroscopy, by the spectrum of Raman backscattering, by the expansion of the bremsstrahlung emission with time, and possibly by other features of the optical spectra. Using both gold and plastic (CH) foils, we have observed more than 5% of the incident green light to be scattered as oblique Raman scattering from below quarter-critical density. This is comparable to the value observed in exploding-foil experiments on Shiva using 1.06- μ m light,⁵⁴ although more Raman scattering was observed in the forward direction in the Shiva experiments than in our

Experiments

Novette experiments. These experiments show that although short-wavelength irradiation may reduce the magnitude and impact of laser-plasma instabilities and the resultant suprathermal electrons, it does not eliminate them, and that sidescattering instabilities may become an increasing concern as plasmas become larger.

Targets and Diagnostics. Here we describe the targets used in these experiments, summarize the target shots obtained, and describe the important diagnostics. Most of the disk targets were gold disks (we also shot three beryllium disks and one PVC disk during 1983), and most of these disk shots were obtained in the course of testing the laser and the target diagnostics. More than 40 such targets were irradiated. Most of the targets were 1200 μm in diameter, with larger targets being used for experiments with focal spots larger than 900 μm . The gold targets were usually 5 μm thick. The focal spot size on each target was measured using x-ray microscope photographs of x-ray emission at energies of 1 to 3 keV. When the target is irradiated with a converging beam, the edge of the x-ray emission is sharp for most shots, and the diameter of the x-ray emitting region is

defined to within about $\pm 50 \mu\text{m}$. The average intensity values defined by the total incident laser energy and spot size have an absolute uncertainty of a factor of 2; their relative uncertainty is $\pm 25\%$. Figure 5-46 shows the laser energies and spot sizes for the gold disk experiments. Based on the measured beam structure, the peak-to-average intensity variations are less than 10 to 1.

The exploding-foil experiments used foils whose thicknesses were specified by analyzing 1-D LASNEX calculations, with the aim of producing a large plasma at approximately quarter-critical density at the peak of the laser pulse. We irradiated nine plastic foils from 1 to 3 μm thick and six gold foils from 0.09 to 0.12 μm thick. Seven of the plastic foils were doped with 2 at.% sulfur to allow x-ray spectroscopy measurements. In two of these targets, the doped region was confined to a 300- μm spot in the center of the target. The small spot reduced the radial gradients in density and temperature of the region emitting x rays. (See "Time-Resolved Crystal Spectroscopy," earlier in this section.)

Several diagnostics measured the bremsstrahlung x rays produced when the hot electrons are stopped in the target. The FFLEX diagnostic measured the time-integrated spectra of energetic x rays. We infer the number of hot electrons and the temperature of the hot-electron distribution from these data. The pinhole microchannel plate two-dimensionally imaged the energetic x rays in the 20- to 100-keV range. (See "Imaging 20- to 100-keV Emissions from Laser Fusion Targets," earlier in this section.) Finally, the optical/x-ray (OX) streak camera measured the time dependence of the 25- to 60-keV x rays, when the hot-electron fraction was above about 1%. In the experiments reported here, the fraction of the laser energy converted into hot electrons, f_{hot} , was not well correlated with the laser intensity. For some shots of equal intensity, we observe factor-of-10 changes in f_{hot} . However, the hot-electron production was very strongly correlated with the amount of Raman scattering.

Several scattered-light measurements were made on Novette, involving time-resolved optical spectroscopy, time-integrated

Fig. 5-46. The normal-incidence gold-disk target shots at 1 ns on Novette during 1983 included significant variations of both laser energy and spot size.

multi-channel spectroscopy, and scattered-light energy measurements. The time-resolved spectroscopy was performed using two 0.5- μm grating spectrometers coupled to S-1 photocathode streak cameras. We typically operated one of these spectrometers with a low-dispersion grating to measure Raman-scattered light in the wavelengths from 600 to 1200 nm. Colored-glass filters were employed to partially compensate for the strong spectral dependence of the S-1 photocathode sensitivity. The other instrument typically measured either 1/2 or 3/2 harmonic light. The 600-lines-per-mm grating in this spectrometer introduced a 100-ps temporal spread in the data. In addition to the half-meter instruments, a 1-m spectrometer resolved the Brillouin-scattered light (near the laser wavelength), and a 0.25-m spectrometer was used as an additional $\omega_0/2$ diagnostic. Both of these instruments used a time-integrating optical multi-channel analyzer (OMA) as the detector. Light is brought to these instruments, which are remotely located in the Novette basement, by either fiber-optic links from light collectors located 15 deg from the laser axis, one at each end of the chamber, or by means of a light pipe. The latter used a Cassegrainian collector (at $\theta = 45$, $\phi = 18$, 63 deg from the plane of the electric vector of the green laser beam) plus mirrors to bring the scattered light to the basement laboratory without the temporal broadening which occurs in multi-mode fiber optics. The light pipe is also the only means of observing wavelengths at $3\omega_0/2$ or shorter, because the glass fibers do not transmit in this ultraviolet region. The temporal dispersion of the fiber optics was not directly measured; however, in some 100-ps experiments, the $\omega_0/2$ emission is clearly less than 200 ps in duration, indicating relatively little pulse broadening.

Other time-integrated measurements use photodiodes with bandpass filters for the multi-channel spectroscopy, and broadband filtering for the energy measurements. A total of 94 photodiodes measured the energy scattered from the target as green light, Raman light, or $3/2 \omega_0$ light. In addition, the green light reflected into and transmitted through the chamber optics was measured by the output sensor package. For each type

of scattering, we integrated the scattered fluence over all solid angles to obtain the total scattered energy.

Disk-Target Results. The results summarized here are from our initial analysis of the data, and are subject to revision. We compared the absorption of gold disk targets as a function of average irradiation intensity for Novette and previous Argus experiments (see Fig. 1-1). The absorption is calculated from the scattered-light measurement described above. The Novette and Argus data overlap, showing the absorption to be roughly the same. However, some of the Novette experiments produced significantly more Brillouin (and Raman) sidescattering than others. When sidescattering is strong, the absorption at a given intensity is about 10% smaller than otherwise. The sidescattered green light is strongly out-of-plane, as is expected for Brillouin sidescattering and as Fig. 5-47 shows. The spectrum of this radiation is only a few angstroms wide and is

Fig. 5-47. The sidescattered green light is strongly out-of-plane, as expected for stimulated Brillouin sidescattering.

Experiments

slightly red-shifted from the incident light. [Note that the red stimulated Brillouin scattering (SBS) shift and the blue Doppler shift are expected to approximately cancel.]

The time-resolved Raman spectroscopy has produced data that are remarkably similar to those observed during the Argus green-light experiments.⁵⁵ In particular, during the high intensity irradiation of disk targets, we observe the SRS light to begin abruptly, and over a wide range of wavelengths—typically 750 to 900 nm. Pulsations in the emissions are frequently observed, especially at the more intense irradiations.

The SRS energy measurements have produced some new results. The data points in Fig. 5-48 show the hot-electron fraction and Raman yield for the gold disk target shots. These two quantities are very well correlated. The line in the figure shows the energy partition between Raman photons and plasma waves for Raman scattering at 850 nm. Assuming that most of the plasma wave energy is converted into hot electrons, that most of the hot electrons strike the target, and that no other process produces hot electrons, this line should reproduce the data. As the figure shows, all the data (except one point) are consistent with this line, within the experimental uncertainties.

From Fig. 5-48 and other data, we draw the following conclusions: First, the observed Raman yield is sufficient to account for all the observed hot electrons, although the absolute uncertainties would allow another process of comparable magnitude and scaling to contribute also; second, the

excellent correlation of hot-electron fraction and Raman yield with one other, but not with intensity or spot size, strengthens the claim that the hot electrons result dominantly from Raman scattering. Third, the comparatively small variation of $3/2 \omega_0$ yield throughout these experiments (and its poor correlation with other observed quantities) suggests that instabilities at quarter-critical density are not significant sources of hot electrons in these experiments.

Moreover, the observed hot-electron temperatures are consistent with modeling of Raman scattering. When the plasma waves are damped, they should produce hot electrons with velocities comparable to the phase velocity of the plasma waves. This has been borne out by simulations,⁵⁶ and is consistent with previous experiments using 1.06 μm irradiation.⁵⁷ For 850-nm Raman-scattered light that emerges from a planar plasma at 135 deg to the incident laser beam, the phase velocity of the electron plasma wave corresponds to a hot-electron energy of 24 keV. The electron energies corresponding to the range of scattering angles and wavelengths observed on Novette are 14 to 40 keV, which is consistent with the observed range of 20 to 33 keV for these experiments. The Bohm-Gross shift in the plasma wave frequency is small for these experiments, so that the phase velocity of the plasma waves is not strongly sensitive to the thermal electron temperature. In contrast to experiments with shorter pulses or smaller spots, resonance absorption is probably not an important source of hot electrons here. The long plasma scale-lengths reduce the amount of light reaching critical density, and the hot electrons produced there should have a temperature below 10 keV.

Exploding-Foil Results. The exploding-foil experiments have produced a great deal of interesting data, as described above. We have only begun to analyze them, and will report on them in detail later. Figure 5-49 and Table 5-6 show some of these data. The OX streak camera measures the time dependence of the incident 1ω laser light, the scattered green, Raman, and $3/2 \omega_0$ light, and the 25- to 60-keV x rays. The camera is located perpendicular to the plane of polarization of the laser and at 45 deg

Fig. 5-48. The hot-electron fraction correlates very well with the observed Raman scattering, and less well with several other experimental variables. The data indicate that Raman scattering is the principal source of hot electrons in these plasmas.

from the backscattered direction. In addition, the output streak cameras measure the incident green light as well as the green light reflected into and transmitted through the chamber optics. Figure 5-49 shows this data for an exploding gold foil. We see that the reflected light comes early in the pulse and the transmitted light comes late in the pulse. The timing of the transmitted light indicates when the plasma density drops below the critical density of $0.53\text{-}\mu\text{m}$ light. The obliquely scattered green light is strongest during the first half of the pulse. The top three traces show the Raman-scattered light (produced by the Raman instability), the $3/2\ \omega_0$ light (a consequence of instabilities at quarter-critical density), and the x rays. The Raman-scattered light and the x rays are both centered at the peak of the 1-ns laser pulse, with a full width at half maximum of about 500 ps. In contrast, the $3/2\ \omega_0$ light has a 1-ns width and most of it comes during the first half of the pulse. These data support the conclusion drawn above that the source of the hot electrons in these gold plasmas is Raman scattering. The time-resolved spectrum of Raman backscattering for this shot shows that the density stays above 0.15 critical until very late in the pulse, which is consistent with the observed duration of the Raman and $3/2\ \omega_0$ light on the OX. Some of the CH foils burned through earlier in the pulse. In these cases the Raman spectrum was quite broad early in the pulse, but later the maximum wavelength of SRS decreased steadily, indicating the decay of the plasma to lower densities.

Table 5-6 presents scattered light yields from the exploding-foil shots. The photodiodes, reflected energy measurements, and

transmitted energy diagnostics determine the reflected, transmitted, and absorbed energies. These results, which will be compared with predictions from LASNEX,

Fig. 5-49. These absolutely timed traces show several features of the laser-plasma interactions present in an exploding-foil experiment using a $0.12\text{-}\mu\text{m}$ -thick gold foil. The text explains their origin and significance.

Table 5-6. Some results from exploding-foil experiments.

Shot No.	Target (μm)	Energy (kJ)	Intensity (10^{14} W/cm^2)	2ω reflected (%)	2ω transmitted (%)	Raman (%)	$3/2\ \omega$ (%)	Absorbed (%)
93091409	2.0 CH	3.2	8	19 ± 9	—	3.6	0.014	—
93100408	2.1 CH	3.8	6	31 ± 11	7 ± 3	>2.9	0.013	59 ± 16
93101003	2.5 CH	4.1	11	20 ± 7	21 ± 8	>3.6	0.017	55 ± 17
93102705	2.1 CH	3.8	9	32 ± 13	14 ± 3	>4.3	0.015	50 ± 21
93110216	3.0 CH	3.5	22	24 ± 7	10 ± 2	9.8	0.021	56 ± 19
93111711	1.0 CH	3.6	13	11 ± 5	38 ± 9	1.9	0.006	49 ± 15
93112105	2.1 CH	3.7	13	27 ± 12	21 ± 4	8.0	0.015	44 ± 20
93121310	1.8 CH	3.7	13	14 ± 3	26 ± 5	5.6	0.012	55 ± 11
93102003	0.10 gold	4.0	6	11 ± 4	28 ± 5	0.25	0.005	61 ± 9
93103111	1.12 gold	3.5	22	16 ± 5	13 ± 3	>5.2	0.010	65 ± 11
93110113	1.12 gold	3.3	20	16 ± 5	7 ± 2	7.3	0.010	70 ± 11

Fig. 5-50. $3\omega/2$ spectra from a thin CH target. (a) Observation angle of 45 deg from laser axis, backward direction. (b) Observation angle of 135 deg from laser axis, forward direction.

Experiments

should be sensitive to the competition between inverse bremsstrahlung and stimulated Brillouin scattering. Raman yields were often high, exceeding 5% on some shots. Under appropriate conditions, the Raman instability produced significant scattering of the incident green irradiation. Finally, the yield of $3/2 \omega_0$ light is roughly 0.01% of the incident laser energy for these shots (and for disk shots at similar intensity). This shows that the magnitude of the plasma waves that can produce $3/2 \omega_0$ light does not increase substantially as the size of the plasma increases.

The $3/2 \omega_0$ data from thick targets typically shows two peaks, one on either side of the exact $3/2$ -harmonic wavelength. A simple theory⁵⁸ predicts that the blue-shifted spectrum is forward scattered while the red shifted peak is the result of a back-scattering process. We can test these ideas with thin CH foil targets. When a foil's density is between $9/4$ and $1/4$ of critical, we should expect only blue-shifted $3/2$ light in the forward direction, and only red-shifted light in the backwards direction. This is not in complete accordance with our observations, several of which are shown in Fig. 5-50. In particular, while we have observed only the red-shifted peak in the backwards direction, we have seen both peaks in the forward direction, at 45 deg from the laser and target axes. This agrees with other measurements,⁵⁹ and is consistent with a theory which attributes $3/2$ harmonic production to the presence of filamentation.⁶⁰ A puzzling aspect of these data is the spectral asymmetry obvious in Fig. 5-50(b). The width of the blue-shifted peak is similar to that observed from disk targets, while the red-shifted peak is much broader. A simplistic theory, using the very

broad red-shifted peak, results in a rather large estimate for the electron temperature. We do not currently understand the $3/2$ generation mechanism well enough to explain the details of this spectrum.

Authors: R. P. Drake and R. E. Turner

Major Contributors: B. F. Lasinski, K. C. Estabrook, S. A. Letts, D. W. Phillion, E. M. Campbell, E. A. Williams, R. L. Kauffman, C. W. Hatcher, W. L. Kruer, and C. L. Wang

Novette Soft X-Ray Laser Experiments

In December of 1982 we began a set of measurements to demonstrate a laser-pumped amplifier with high gain at x-ray wavelengths. This was the first experimental series carried out at the Novette laser facility. The experiments were done in collaboration with LLNL's R- and Nuclear Test Programs, which have the responsibility for studying such lasing schemes pumped by a variety of sources.

We studied, with decreasing levels of effort, three different x-ray laser schemes: resonant photoexcitation,^{61,62} collisional excitation,^{63,64} and collisional recombination.⁶⁵ Each scheme can produce lasing transitions in the wavelength interval from 230 to 80 Å. These and other methods for producing population inversions leading to transitions in XUV region are reviewed by Waynant and Elton.⁶⁶

The resonant-photoexcitation scheme received the primary emphasis in these experiments. Gains predicted for the scheme are extremely high, but the physics is the most complicated, so the likelihood of a

demonstration is probably the lowest. The concept was first suggested by Vinogradov et al.,⁶¹ with refinements by others.^{62,67,68}

Many of the ideas for our experiments were expressed by Hagelstein.⁶⁹ The basic method for obtaining the population inversion is shown in Fig. 5-51. A rectangular-cross-section gas cell is bounded on two sides by x-ray flashlamp foils that are irradiated by the high-power 0.53- μm Novette laser. The plasma produced by the laser interaction (laser intensity $\sim 10^{14} \text{ W/cm}^2$) produces copious amounts of continuous and discrete x rays, which are used to ionize the lasant gas between the foils and to resonantly pump excited states of the ions. Electron-ion collisions then transfer the population to higher-angular-momentum states that cannot subsequently decay to the ground state from which the ion was photoexcited. A specific case for this scheme is shown in Fig. 5-52, where we illustrate a level diagram for the helium-like neon ion. In one of the schemes attempted, we chose a flashlamp material that emitted the 11,000- $\text{\AA}$ x rays necessary to resonantly photoexcite helium-like neon to the 1s4p level. We then depend on collisions in the plasma to populate the 1s4f level. This can

lead to a population inversion over the 1s3d, thus leading to a laser transition at 231 $\text{\AA}$. Obviously this same idea can be used to pump the 1s3p level with collisions, again leading to a population inversion of the 1s3d over the 1s2p. This concept was also used to pump helium-like fluorine to produce a possible laser transition at 104.2 $\text{\AA}$. Finally, we also applied the technique to hydrogen-like fluorine in order to pump the 3p level by collisional transfer, thus leading to an inversion of the 3d level over the 2p. Table 5-7 summarizes all the schemes attempted, as well as the flashlamp material that was used to provide the resonant transition for pumping the lasant gas. This is a reduced version of Hagelstein's list⁶⁹ of potential pump candidates, selected as a result of some high-precision spectroscopy performed by

Fig. 5-51. A cross-sectional view of the resonant photoexcitation x-ray laser target. The original concept is from Ref. 61.

Fig. 5-52. A simplified level scheme for the helium-like neon laser. Lines that should not be present in the pump are shown dashed with an X.

For example:

P. Burkhalter et al.⁷⁰ in support of these experiments.

of photoexcitation). Prior to these experiments, only the degree of resonance of the Na-Ne scheme was known with sufficient accuracy. In addition, we were able to certify the brightness of only one of the schemes, namely, the chromium-pumped fluorine line at 14.458 Å. The plasma conditions inside the possible lasing channel were not determined.

Another serious concern is the complication of target construction for this scheme. These targets are described in the article "Production," in Section 4.

The collisional-excitation laser was initiated much later in the planning phase of this particular set of experiments. Considerable work has been done on the scheme,^{71,72} particularly for the krypton atom, but we elected to avoid the difficulties of working with a gas target and chose to use a nearby solid element in the periodic table, selenium.

Figure 5-53 illustrates the principal physics mechanisms responsible for producing an inversion with this scheme. Laser light is

Pump	Lasant
Na (1s 2p $^1P_1 \rightarrow 1s^2 \text{ } ^1S_0$) $\lambda = 11.003 \text{ \AA}$	Ne (1s ² $^1S_0 \rightarrow 1s4p \text{ } ^1P_1$) 4d $\rightarrow$ 3p (53.6 eV) $\lambda_0 = 11.000 \text{ \AA}$
Ni (N-like, 3d $\rightarrow$ 2p) $\lambda = 11.000 \text{ \AA}$	Nc (1 s ² $^1S_0 \rightarrow 1s4p \text{ } ^1P_1$) 4d $\rightarrow$ 3p (53.6 eV)
Cr (B-like, 3d $\rightarrow$ 2p) $\lambda = 14.456 \text{ \AA}$	F (1 s ² $^1S_0 \rightarrow 1s3p \text{ } ^1P_1$) 3d $\rightarrow$ 2p (119.0 eV) $\lambda_0 = 14.458 \text{ \AA}$
Mn (Be-like, 3d $\rightarrow$ 2p) $\lambda = 12.641 \text{ \AA}$	F (1s $2S_{1/2} \rightarrow 3p \text{ } ^2P_{1/2, 3/2}$) 3d $\rightarrow$ 2p (153.0 eV) $\lambda_0 = 12.643 \text{ \AA}$

Table 5-7. Resonant photoexcitation x-ray laser schemes.

Fig. 5-53. An illustration of the neon-like selenium collisional excitation laser scheme. Lasing can possibly occur between the 3p and 3s levels because of preferential radiative decay out of the 3s.

Experiments

used to produce a freely expanding coronal plasma having a specific density and temperature profile in the dimension parallel to the incident laser beam. Irradiation conditions are chosen to maximize the abundance of the neon-like ion species in the density regime around $10^{21}/\text{cm}^3$. Collisional population of the 3s, 3d, and 3p levels leads ultimately to an inversion between the 3p and the 3s levels because the upper (3p) level cannot dipole-decay back to the neon-like ground-state configuration, whereas the 3s has a strong radiative decay to the 2p level. The strongest candidates for lasing are those states having a $j = 0$ initial and $j = 1$ final state. A full level diagram showing some of the transitions emitted by the neon-like selenium M-shell is shown in Fig. 5-54.

Unfortunately, the collisional population of the $n = 3$ levels is non-trivial, whereas producing neon-like ground-state configurations is trivial (it requires only the removal of weakly bound M-shell electrons). Exciting one of the L-shell electrons to the 3p state requires approximately 30% higher plasma temperature (800 instead of 600 eV). Any laser design must therefore incorporate plasma conditions that involve significant plasma temperature and density. Table 5-8 illustrates the sensitivity of the calculated gain to the plasma electron temperature for a krypton plasma (from the calculations of Feldman, Bhatia, and Suckewer⁷²). Figure 5-55 shows the gain for our scheme as calculated by Hagelstein.⁷³ Note that at $\sim 25 \mu\text{m}$ from the surface, the gain is ~ 10 ; this corresponds to a density of 10^{21} cm^{-3} , which would be easily observable by our diagnostics.

The principal physics issues that can affect the collisional laser scheme are (a) the refraction of x rays due to the steep density gradient along the dimension parallel to the incident laser beam, (b) the optimum plasma conditions necessary for producing the maximum population of $j = 0$ neon-like states, and (c) significance of absorption from the copious transitions from other ionic species.

The final scheme we studied, with only a few experiments, involves preferential recombination of electrons into high- n states in hydrogen-like oxygen ions. The original

N_e (cm^{-3})	Gain coefficients (cm^{-1}) for electron temperatures of	
	$1 \times 10^7 \text{ K}$	$3 \times 10^7 \text{ K}$
10^{19}	0.045	0.067
10^{20}	0.92	1.7
10^{21}	8.8	16.5
10^{22}	35	66

Table 5-8. Calculated gain coefficients for the neon-like krypton $2p^5 3p(^1S_0) \rightarrow 2p^5 3s(^3P_1)$ transition.

Fig. 5-54. A simplified level scheme illustrating the wavelengths, radiative rates and gf values for selenium.

Fig. 5-55. The small-signal gain for the neon-like selenium scheme as a function of distance off the surface of the selenium-coated foil.

idea for this scheme comes from Gudzenko and Shelepin,⁶⁵ with numerous later extrapolations to laser-pumped plasmas (Irons and Peacock⁷⁴ or Per⁷⁵). The basic recipe for producing the inversion is as follows: (a) ionize a plasma rapidly and completely with a short-pulse, high-intensity optical laser; (b) cool the plasma rapidly by adiabatic expansion, electron conduction, radiative power loss, or, preferably, some combination of the three.

Some of the physics that affects the performance of this laser scheme concerns the following questions: (a) Is population inversion

Experiments

maintained for a plasma of this density and temperature, i.e., can collisional recombination into the $n = 2$ level dominate the collisional excitation into the $n = 2$ level, and can the plasma be limited in size so that the trapping of the ground state for the laser (i.e., the $n = 2$ to $n = 1$ Lyman- α line) remains negligible, at least during the times of peak upper-state population? (b) Can we truly model the atomic physics of this problem, i.e., the ionization kinetics, level populations, line widths, and radiation transport?

This scheme was not optimized before the experiments, and our post-experiment modeling has revealed some flaws in our design. We overstripped the plasma, i.e., the electron temperature was too high for optimum production of hydrogen-like ion species during the lifetime of the plasma. Moreover, we did not attempt to cool the plasma by any exotic means other than natural adiabatic expansion (a cooling rate that scales as the expansion velocity of the plasma density distribution).

Early in the planning for these experiments, we adopted a strategy to perform integrated tests of the targets. Our plan was (a) to ensure that the new Novette laser could produce 5 TW of green light per beam, with a pulse length of 100 ps FWHM, (b) to verify that the new cylindrical lenses were functional, which means that they produced a fairly uniform 1.2-by-0.01-cm focal spot, (c) to check flashlamp line brightness at green wavelengths (our previous studies had been at Shiva at

1.06 μm), (d) to confirm proper operation and spectral line coverage for the principal x-ray laser diagnostics, (e) to perform approximately 20 integral (complete) x-ray laser target tests, and (f) to perform null tests or other verification tests to certify success in the event that we obtained positive results in any experiments. A complete table of criteria for verification of the achievement of lasing is shown in Table 5-9.

The diagnostics we used for these experiments are shown schematically in Fig. 5-56. The transmission-grating streak camera, described in the article "Time-Resolved Spectral Measurements Above 80 $\text{\AA}$," measures the time history of the potential x-ray laser line with 20-ps resolution. The fail-safe diagnostic, discussed in the next article, measures the time-integrated x-ray spectra with a high tolerance for misalignment of the x-ray laser target. Finally, the streak x-ray spectrograph, discussed in "Time-Resolved Crystal Spectrography," earlier in this section, measures the history of the x-ray line emission from the flashlamp. These three main diagnostics were complemented by the following diagnostic array: x-ray microscopes and a pinhole camera to measure the beam size from x-ray images; time-integrated x-ray crystal spectrographs to determine the absolute intensity of the flashlamp lines; optical diodes to determine laser absorption; and several other of our normal diagnostics for studying laser-plasma conditions. The optical axes of the principal diagnostics (fail-safe and transmission-grating

Table 5-9. Criteria for the verification of lasing.

Element and shot No.	Pulse widths (ps, FWHM)		Peak line intensity (10^{16} eV/keV-sphere)		Line width (eV)	Emission area (10^{-3} cm 2)	Brightness (10^{-4} photons/mode)	
	Laser	Line	Front	Back			Front	Back
NaCl								
93012706	96	165	7	3.3	~2.5	7.85	8	3.4
93012804	98	150	8.3	4.3	~2.5	7.85	10.4	4.9
93020107	230	190	9.5	—	~2.5	7.85	9.43	—
93020204	102	60	14	9.0	~2.5	7.85	44	26
93020206	100	170	4.7	—	~2.5	7.85	5.2	—
Ni								
93013110	116	190	2.7	—	2.5	7.85	2.7	—
Cr								
93011906	94	100	40	—	2.0	7.85	1100	—
93012506	90	110	42	—	—	2.0	20.1	41
93012509	100	100	23	—	—	2.0	7.85	63
93012512	100	120	11	—	—	2.0	20.1	9.8
Mn								
93013119	105	140	26	—	—	2.3	7.85	50
93020104	101	160	23	—	—	2.3	7.85	49

streak camera), together with the target-chamber center, defined the line of sight along which we pointed the longitudinal axis of the laser focus, as well as the axis to which we aligned our x-ray laser target to high precision (± 0.5 mrad). The alignment scheme for the whole system, shown in Fig. 5-57, was designed by H. Medeck.

Our experiments were divided into two parts: (a) flashlamp shots to determine brightness, and (b) x-ray laser target shots. The results of a limited number of flashlamp shots are shown in Table 5-9 for chromium (14.458-Å helium-like fluorine scheme), nickel and sodium (helium-like neon scheme at 11.000 Å) and manganese (hydrogen-like fluorine at 12.643 Å). The tables include the time-integrated intensity and x-ray pulse length data (when available), in addition to an estimate of the brightness in photons/mode. In most cases, the brightness is less than optimum for achieving gain. For all cases except chromium, the data are too limited to permit a conclusive statement that sufficient brightness cannot be achieved. More experiments

to determine the optimum laser irradiance and target-thickness parameters for these flashlamps are planned for the near future.

The results of the x-ray laser target experiments can best be discussed by the type of scheme that was tested. For the resonant photoexcitation scheme three potential laser scenarios were tested and are tabulated in Table 5-7. Our basic result is that we observed no lasing transitions in some 22 target tests. The sensitivity limits of our

Fig. 5-56. A schematic illustration of the placement of the principal x-ray laser diagnostics on the Novette target chamber.

Fig. 5-57. Alignment scheme for the x-ray laser targets and the x-ray laser diagnostics: the transmission-grating streak camera and the fail-safe.

Experiments

diagnostics required achievement of gain lengths greater than 5 to 6. Although there were many ways to fail in the execution of the experiments, we believe that a proper test of the schemes took place. Some possible reasons for the lack of positive results are: (1) we did not properly certify a resonance between lasant and pump lines (a $\pm 0.0015\text{-}\text{\AA}$ overlap is required), (2) our pump lines may not have been bright enough to provide enough gain for lasing to be observed with instruments of the sensitivity of those we used, (3) the calculated laser kinetics may be slightly inaccurate, or perhaps the lasant cavity ion temperature was too high (neither was measured in the experiments), or (4) the laser cavity may have collapsed before the inversion was initiated.

The collisional-excitation laser scheme with simple, selenium-coated Formvar targets was also tested. No lasing lines were observed for this scheme either. Prior to finishing the experiments we already knew where a fault resided in our design. Figure 5-58 illustrates the effects of a steep density gradient (which is produced with this type of target) on the trajectory of our potential x-ray laser beam. In short, our scheme would have been compromised by refraction of the x-ray laser beam out of the maximum gain region, as well as out of the viewing area of our x-ray laser diagnostics. We have planned some major improvements to this scheme for our next series, to be held in the spring of 1984.

The final scheme we tested was the recombination laser. Here we irradiated a simple Formvar foil target (15% oxygen) with two of the Novette beams operated at relatively low intensity. Figure 5-59 shows our experimental setup and some of the results.

We measured the time-dependent emission from the Balmer- α ($n = 3$ to $n = 2$) transition of hydrogen-like oxygen for both long and short foils. Its value shows whether the signal is exponentiating in intensity with the length of the plasma, as

Fig. 5-58. An illustration of the effects of a steep density profile [$n_e = n_e(r)$] on the trajectory of a beam of x rays.

Fig. 5-59. The experimental setup for the recombination laser experiment. Data obtained for plasmas of two lengths are also shown.

would be the case if it were lasing. An analysis of our data and comparison with theory indicate that a small amount of gain was achieved ($\alpha \approx 0.5 \text{ cm}^{-1}$). Since this scheme was not optimized prior to the experiments, we are encouraged by these results and hope to further test this scheme in the future planned experiment. Further details of these results are presented in a report.⁷⁶

The results of all these experiments led us to conclude that another, more rigorous series of measurements is warranted. Our latest effort incorporates improved physics understanding, target design, and diagnostics instrumentation.

Experiments are under way to study underlying physical processes that could have compromised the demonstration of these lasers or the measurement of their performance. The work is being done with a relatively small laser at KMS Fusion, Inc., Ann Arbor, Michigan. In these experiments we are concentrating on the measurement of most of the major plasma physics issues that could have compromised the schemes. As an example, in the resonant photoexcitation scheme we must verify that a pump line overlaps the energy of an absorbing laser state to within 1 part in 10^4 . This is currently being accomplished with some extremely accurate spectroscopic measurements.

Our strategy to achieve success includes more than a certification of the relevant plasma conditions. We are also constructing more sensitive diagnostics to permit detection of the laser lines, even if they are much weaker than predicted. In addition, we are improving our x-ray laser target construction techniques in order to make the critical components of a target more closely resemble the complex geometries that have been generated during redesign of the laser scheme. Improved target design also enables better execution of the experiment, since we must often certify alignment of our diagnostics to extremely close tolerances.

In the next Novette experiments we will concentrate on only one or two schemes, iterating target design and optical-laser pump parameters around the values predicted by our simulation codes, in order to reduce the possibility that a small

uncertainty in the calculation caused us to overlook a successful demonstration.

We envision our program expanding in the near future, due to an increased understanding of the relevant physics issues, as well as the emergence in October 1984 of an even bigger optical pump laser, Nova. This will enable us to study even more exotic laser design and to extend our measurements to even shorter x-ray laser wavelengths.

**Authors: D. L. Matthews and
R. L. Kauffman**

X-Ray Laser "Fail-Safe" Diagnostics: Performance and Results

In the beginning of 1983, we completed and put into operation a newly developed high-sensitivity medium-resolution spectrograph of large angular acceptance and dynamic range.⁷⁷ This diagnostic system was named "Fail-Safe" because of its ability to cope with a large directional uncertainty in the x-ray laser. We review here some highlights of its performance and experimental results.

The diagnostic system performed quite well, even though this was the first time an open microchannel plate (MCP) electron multiplier was used on the target chamber, and we were concerned that the radiation background might be a problem. In fact, the MCP proved to be so useful that we have since developed two more new diagnostics that use MCPs (see the earlier articles "Gamma-Ray Streak Camera" and "Imaging 20- to 100-keV X-Ray Emissions from Laser Fusion Targets"). For the setup of Fail-Safe, the reader is referred to Ref. 77. Here we simply mention that x-rays are dispersed by a reflection grating and detected by a MCP.

Figure 5-60 shows a spectrum obtained in a calibration shot for the 155-eV manganese-pumped-fluorine laser series. The abrupt cutoff of the spectrum at the high-energy side is due to the K absorption edge of a boron filter, which was used to provide an unambiguous energy reference. This single-laser-beam calibration shot shows that a relatively well defined, small emission source was seen by the spectrograph.

Experiments

Fig. 5-60. Spectrum obtained from a calibration shot for the 153-eV x-ray laser experiments.

Fig. 5-61. A typical spectrum observed from the x-ray laser target shots (see text for details).

Fig. 5-62. Another type of spectrum observed from the x-ray laser target shots (see text for details).

Fig. 5-63. Candidate recombination lasers at 121 and 164.2 eV (the latter possibly stronger). Spectra obtained with Formvar foils of effective lengths (a) 1.27 cm and (b) 0.27 cm.

The x-ray laser target shots with both laser beams, on the other hand, indicate that:

1. The emission region was often so large that the spectrum was blurred. (See Fig. 5-61; note the difference between Fig. 5-60 and Fig. 6-61.) Or
2. There sometimes appeared to be two distinct emission regions, presumably due to the two laser beams hitting the target flashlamp materials at different heights. (See Fig. 5-62; note the overlapping spectra.)

The Fail-Safe was sensitive enough to obtain data from virtually all target shots. We have not noticed, however, indications of x-ray lasing expected from either the resonant photoexcitation or selenium collisional schemes at 53.6, 70, 119, and 153 eV. In view of the negative results and the great time and effort required, we chose not to pursue the absolute calibration of our diagnostic system. However, since our apparatus performed to the design specification, we may use our estimated system sensitivity⁷⁷ to draw the preliminary conclusion that no resonantly pumped x-ray lasers were observed at the gain-length level of ~ 7 .

However, during the calibration shots using Formvar foils, we did observe an indication of O VIII H α recombination laser at 121 eV. Figures 5-63 (a) and (b) show the spectra obtained with 2000-Å-thick Formvar foils of 1.27 and 0.27 cm respectively, irradiated with two opposite, cylindrically focused (14×3 mm) $0.53\text{-}\mu\text{m}$ Novette beams at $1.4 \times 10^{13} \text{ W/cm}^2$ and 100 ps. The total power in each beam is 5 TW. The measured intensity ratio was 5.6, which is greater than the ratio of 3.8 expected from a medium exhibiting no gain.⁷⁶ This is in agreement with the stronger evidence obtained with the time-dependent intensity measurement.⁷⁶

A much larger intensity ratio (~ 50) is observed for the O VIII H β line at 163.3 eV. Unfortunately, this line is outside the energy range covered by the time-resolved diagnostics, the transmission-grating streak camera. It appears that the large intensity ratio may indicate a stronger recombination laser at 163.3 eV. A further experiment is needed to confirm this observation.

Authors: C. L. Wang, E. M. Campbell, C. W. Hatcher, G. R. Leipelt, D. L. Matthews, and V. W. Slivinsky

Line Diagnostics of Laser-Produced Plasmas

We have continued experiments to investigate the validity of x-ray line diagnostics for laser-produced plasmas.

Densities of various x-ray lines depend on the local temperature and density of the plasma. Many lines have been used to diagnose a wide variety of laboratory and astrophysical plasmas.⁷⁸ Because of their transient nature, special care must be used in applying x-ray spectroscopy to laser-produced plasmas.⁷⁹⁻⁸¹ In the 1982 Laser Annual, we reported on measurements to test the sensitivity of the diagnostic to local plasma conditions. Here, we report more data that demonstrate the consistency of line ratios in measuring local plasma conditions at different times.

In this experiment, we investigated x-ray line emission from various dopants in a CH plasma. We measured line ratios from Si, Cl, and Ca, examining consistency among various hydrogen-like and helium-like x-ray line emissions. The dopant concentrations were less than 5% to ensure that they did not significantly modify the laser-plasma physics and hydrodynamics. Low concentrations of dopants also minimize opacity effects.

The experiments were performed at Rutherford Laboratory⁸² in collaboration with Imperial College. The experimental setup was similar to that reported in the 1982 Laser Annual.⁸³ Planar targets were irradiated at near normal incidence with a single beam of $2\omega_0$ light. The beam had an energy of about 20 J in 1 ns, and all experiments were performed with a peak irradiance of about 5×10^{14} W/cm². The principal diagnostic is a time-resolving crystal spectrograph⁸⁴ to measure the time history of the relative intensity of the various spectral lines. Figure 6-64 is an example of the data, showing chlorine lines from a chlorine-doped CH disk at the peak of the laser pulse. The various spectral lines used for diagnostics are identified. The ratio of the Cl XVII $L\alpha$ to the Cl XVI 1P_1 lines measures the ionization balance between the two charge states. This ratio is used as a temperature diagnostic. The $2^1P/3^2P$ helium-like emission ratio is used as an

electron-density diagnostic. This ratio is density-sensitive above about 10^{21} cm⁻³. The relative intensity of the lithium-like Cl XV is also a temperature-sensitive diagnostic, but the theoretical modeling for this line is not as certain. Similar spectra have been measured for silicon and calcium ions, so their spectra can also be analyzed to form a consistent picture of the CH plasma.

In order to relate relative line intensities to plasma density and temperature, we have developed a computer code to calculate relative intensities of x-ray spectra.⁸⁵ The calculations solve the rate equations for the level populations of the various ionization states of the ion of interest. For the present results, ground-state and all excited-state levels are included for the lithium-, helium-, and hydrogen-like ionization states. Only the ground-state levels are included for all of the lower ionization states. For lithium-like and helium-like ions, autoionizing states that form satellites of the resonance transition of the next ionization state are also included. The model solves for the level populations, assuming that the ion is in steady state with a plasma at a given temperature and density. Using these populations and the radiative transition probability, we can predict the relative intensity of x rays emitted from the different excited states. Examples of the results from the calculations are shown in Fig. 5-65. In Fig. 5-65(a) the $L\alpha/1P_1$ ratio for chlorine is

Fig. 5-64. Example of the time-resolved chlorine spectrum from a laser-produced plasma. The identified transitions are used as plasma diagnostics.

Fig. 5-65. Line intensity ratios for chlorine predicted by a steady-state level population model: (a) ratio vs T_e for various n_e ; (b) ratio vs n_e for various T_e .

Fig. 5-66. For the x-ray-emitting region laser-irradiated CH foil disk, (a) shows temperature curve and points from measured chlorine line ratios at corresponding times; (b) shows density curve from measured chlorine line ratios at corresponding densities.

plotted for several densities. The ratio does show some variations with density, but the overall ratio can provide a unique temperature determination. In Fig. 5-65(b) the $3P_1/1P_1$ ratio for chlorine is plotted for several temperatures. Above electron densities of 10^{21} cm $^{-3}$ the ratio is a good measure of the density. Below 10^{21} cm $^{-3}$ the ratio varies with temperature and is insensitive to density. Similar sets of curves have been generated for silicon and calcium in order to interpret their spectra.

The time-resolved data have been analyzed by measuring the time history of the relative intensities of the various lines from the data shown in Fig. 5-64. The relative intensities are corrected for spectrograph dispersion and detector response. Time-dependent ratios can then be found from the data.

The hydrogen-like $L\alpha$ to helium-like $1p$ ratio is plotted in Fig. 5-66(a) for chlorine.

On the right-hand scale, the electron temperature corresponding to the measured ratio is indicated for an electron density of 10^{21} cm $^{-3}$. The data indicate that the plasma temperature is peaking around the peak of the pulse at about 900 eV. The temperature falls to about 600 eV at the half widths. This agrees with the trends found for the silicon-doped plasma under the same irradiation conditions. At these temperatures hydrogen-like calcium is not produced, so we could not use it for comparison.

The ratio of the 2^1P to 3^1P lines for Cl XVI is shown in Fig. 5-66(b). This ratio decreases monotonically over time during the laser pulse. The predicted densities are plotted on the right-hand scale for 1-keV plasma temperature. The ratio indicates that the lines are being emitted from around critical density (4×10^{21} cm $^{-3}$) early in the pulse. The average density in the emitting region falls

to about 10^{21} cm^{-3} at the peak of the pulse. Below a value of 1.8, the ratio is not a density-sensitive diagnostic, but begins to vary with temperature. The observed decrease in the ratio after the peak of the pulse is in the temperature-sensitive regime and is not indicative of a decreasing density. We also were able to measure this line ratio for calcium. It showed qualitatively the same trends as for chlorine, although the lines are sensitive to densities higher than those produced in these disks. Unfortunately, we could not probe the lower-density region during this experiment because our spectrographs could not resolve the silicon lines. Lower-Z ions must be used to probe these lower densities.

In summary, we have completed preliminary experiments to identify useful line diagnostics for laser-produced plasmas. The technique appears promising as an independent measure of the local plasma parameters. We hope to extend these techniques to our present Novette experiments and to correlate these measurements with other plasma diagnostics to better understand the laser heating of matter.

Author: R. L. Kauffman

Major Contributors: D. L. Matthews, R. W. Lee, and J. D. Kilkenny

Measurements of the Two-Plasmon Instability and Electron Heating

At the University of California, Davis, we have continued our measurements of microwave-plasma interactions. Our experiments have been performed on Prometheus III, which is a larger version of Prometheus I,⁸⁶ the major difference being that the underdense plasma can be longer ($L/\lambda_{De} <$

5×10^3). The electron mean free path is about 1 m, which is greater than the device dimensions, so the plasma is essentially collisionless, with electron temperature about 1 to 3 eV and critical density $n_c = 1.8 \times 10^{10} \text{ cm}^{-3}$. Microwaves in the TM_{02} mode interact with plasma near the quarter-critical surface.

When the microwave power is above a clearly defined threshold, the two-plasmon decay (TPD) instability is observed near the quarter-critical surface [Fig. 5-67(a)] with a frequency of one-half of the microwave frequency, i.e., $\omega_0/2$ ($\omega_0/2\pi = 1.2 \text{ GHz}$). The measured threshold, v_0/v_w , is about 0.1, close to the approximate theoretical value⁸⁷ of $0.7/(k_0 L)^{1/2}$, which is about 0.14 for our experimental geometry. Microwaves are reflected at the cutoff surface, so that standing waves are formed in the underdense region.

Figure 5-67(b) shows a typical spectrum of the high-frequency electric field that we measure locally with a probe antenna. The decay spectrum has a single peak with a FWHM of $\Delta\omega/\omega_0 \approx 3 \times 10^{-3}$, because the electron temperature is low ($v_e \ll c$) in our experiments. We also have observed the emission of microwaves with frequencies near $3/2 \omega_0$ [Fig. 5-67(c)]. The amplitude of the decay electron plasma waves grows exponentially with time, up to saturation. The measured growth rates are proportional to the square root of the microwave power, in approximate agreement with theory.

For high-power microwaves (5 kW, $I\lambda_p^2/T_{e0} \sim 10^{15}$), a series of density depressions is created in the underdense region by the ponderomotive force of the standing microwave. The decay plasma waves are confined in the density depressions.

The two-plasmon instability creates significant heating of suprathermal electrons. The hot-electron temperature T_H increases.

Fig. 5-67. (a) Density-profile modification near the quarter-critical surface. $P_0 = 0.5 \text{ kW}$. (b) Spectrum of two-plasmon instabilities and incident microwave. (c) $3/2 \omega_0$ emission.

Experiments

with density and peaks at $n_c/4$ in the finite instability region. The heated-electron distribution is nearly bi-Maxwellian. The supra-thermal electron temperature is plotted as a function of P_0 in Fig. 5-68. The hot-electron temperature increases with microwave power ($T_H \propto P_0^{0.5}$). Thermal electrons are not significantly heated. The estimated anomalous collision frequency due to the two-plasmon instability, ν^* , is typically $\nu^*/\omega_0 \approx 0.01$.

We also have observed strong high-frequency waves in the underdense region, with frequencies near $\omega_0/2.5$. The threshold of this instability is 10 times higher than that of the TPD instability. At high microwave power, the amplitude of the waves near $\omega_0/2.5$ can be larger than the waves near $\omega_0/2$.

Authors: K. Mizuno, J. S. DeGroot, J. Rogers, and W. Woo (all University of California, Davis)

Fig. 5-68. T_H vs incident microwave power (for $I\lambda^2/T_{e0}$).

Fig. 5-69. Saturated amplitude of ion-density perturbation ($k\gamma_{Di} = 0.32$) vs electron-ion temperature ratio in cooled simulations, $E_0/(4\pi n_c T_{e0})^{1/2} = 0.6$, $n = 0.7 n_c$, $M/m = 100$.

Electron Heating from the Ion-Acoustic-Decay Instability in a Finite-Length System

The ion-acoustic-decay (IAD) and oscillating-two-stream (OTS) instabilities can be a source of significant absorption and production of hot electrons. Previous authors⁸⁸⁻⁹¹ have concentrated on plasma densities near critical density; we consider lower plasma densities ($n \sim 0.7 n_c$), and show the importance of the finite size of the interaction region.⁹²

For a system of finite size, heated electrons leave the interaction region, thus altering the distribution function, which in turn affects the damping rate of the electron plasma waves. This sequence is most important for low densities, where the unstable waves are of lower phase velocity and Landau damping is appreciable.

To illustrate this effect, we ran one-dimensional particle simulations with the following initial conditions: $n = 0.7 n_c$, $M/m = 100$, $E_0/(4\pi n_c T_{e0})^{1/2} = 0.6$. Particles were loaded with Maxwellian and uniform-density distributions. With periodic boundary conditions, appropriate for an infinitely long system, the saturated values of the ion-density fluctuation and electric-field energy were very low; electron heating was small and ion heating negligible. The saturation mechanism was found to be increased Landau damping due to the hot-electron tail. We modeled the loss of hot electrons in a finite system by artificially cooling the electrons. Increasing the cooling rate from zero (decreasing the system length from infinite to finite) allows the instability to grow for a longer period of time, until the increased heating overtakes the cooling and the electron tail again causes saturation.

Since the electron heating is proportional⁹³⁻⁹⁵ to $(\delta n/n)^2$, an important question is the maximum level that ion waves can be expected to reach. It has been suggested⁹⁶ that ion-wave saturation is due to trapping. We found that the ion waves are indeed limited by trapping (Fig. 5-69) when the electron cooling is large enough that saturation is not caused by the hot-electron tail formation. Figure 5-69 shows that the simulation results are in good agreement with the predictions of the simple ion-trapping theory. (For $T_{e0}/T_{i0} = 30$, the electron

heating becomes so large that the hot-electron tail again becomes a factor.)

This investigation was prompted by microwave experiments^{97,98} that revealed large-amplitude ion fluctuations due to parametric instability on low-density shelves ($n/n_c \geq 0.7$, $T_e/T_i \approx 10$). Only when the finite interaction region is modeled by cooling the heated electrons do the simulation results compare favorably with experimental results. The experimentally observed

ion fluctuations are 0.20 to 0.30, consistent with the observed trapping limit of 0.28 in the cooled simulation. In addition, strong electron heating ($\nu^*/\omega_0 = 0.08$) and ion heating are observed in the cooled simulations, in rough agreement with experiment.

Authors: P. W. Rambo, W. Woo, J. S. DeGroot, and K. Mizuno (all University of California, Davis)

References

1. S. Kacenjar, S. Skupsky, A. Entenberg, L. Goldman, and M. Richardson, *Phys. Rev. Lett.* **49**, 463 (1982).
2. A. L. Frank, E. V. Benton, and J. C. Alberghetti, *Nucl. Instr. Meth.* **122**, 433 (1974).
3. K. G. Tirsell, *New Filtered-XRD Low Energy X-ray Spectrometer*, Lawrence Livermore National Laboratory, UCID-19394 (May 1982).
4. G. L. Stradling, D. T. Attwood, and R. L. Kauffman, "A Soft X-Ray Streak Camera," *IEEE J. Quantum Electron.* **QE-19**, 604 (1983).
5. R. L. Kauffman, G. L. Stradling, D. T. Attwood, and H. Medeck, "Quantitative Intensity Measurements Using a Soft X-Ray Streak Camera," *IEEE J. Quantum Electron.* **QE-19**, 616 (April 1983).
6. K. G. Tirsell, D. G. Nilson, J. M. Auerbach, T. G. Woehrl, R. M. Penparaze (LLNL), and L. N. Koppel, C. E. Isaacson (Aracor Co.), *Absolute Calibration of Filtered X-Ray Diode Systems With Grazing Incidence Mirror Cutoffs*, Lawrence Livermore National Laboratory, UCRL-89101 (to be published).
7. *1981 Laser Program Annual Report*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-50021-81 (1982), p. 5-8.
8. J. L. Gaines, "Low Energy X-Ray Calibration Sources at the Lawrence Livermore National Laboratory," in *Proceedings of Low Energy X-Ray Diagnostics-1981*, Monterey, D. T. Attwood and B. L. Henke, Eds., p. 246 (AIP, New York, N.Y., 1981).
9. B. L. Henke, J. P. Knauer, and K. Premaratne, "The Characterization of X-ray Photocathodes in the 0.1-10 keV Photon Energy Region," *J. Appl. Phys.* **52**, 1509 (1981).
10. W. Parkes, R. Gott, and K. A. Pounds, *IEEE Trans. Nucl. Sci.* **NS-17**, 360 (1970).
11. K. W. Dolan and J. Chang, "Microchannel Plate Response to Hard X-rays," *Proc. SPIE* **106**, 178 (1977).
12. J. Adams and I. C. P. Millar, *Electronica* **14**, 237 (1971).
13. *1980 Laser Program Annual Report*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-50021-80 (1981), pp. 5-19 to 5-21.
14. D. E. Kotecki and R. Lear, "Optical Shutters using Microchannel Plate (MCP) Intensifier Tubes," *High Speed Photography, Videography and Photonics*, San Diego, Calif., SPIE Conference No. 427, August, 1983.
15. E. Laviron and C. Delmare, "Realization of an Image Converter with a 300 ps Exposure Time," *9th International Congress on High Speed Photography*, Denver, Colo., 1970, pp. 128-201.
16. N. Fleurot, J.-P. Gex, M. Rostaing and R. Sauneuf, "High Speed (<=250 ps) High Gain X-ray Shutter Camera," *15th International Congress on High Speed Photography and Photonics*, San Diego, Calif., SPIE Volume 341, 1982, pp. 772-776.
17. R. A. Lawton and A. Scavennec, "A Photoconductive Switch Gated by a Modelocked Dye Laser," presented at the *International Conference on Submillimeter Waves and their Applications*, Atlanta, Ga., June 5-7, 1974.
18. R. A. Lawton and J. Andrews, *Waveform Sampler*, U.S. Patent 4,030,840 (June 21, 1977).
19. D. H. Auston, "Picosecond Electronics with Optical Pulses," *High Speed Optical Techniques*, SPIE Vol. 94, pp. 78-82 (1976).
20. G. Mourou and W. Knox, "High Power Switching with Picosecond Precision," *Appl. Phys. Rev.* **35**(7) (October 1, 1979).
21. *Proceedings, "Picosecond Optoelectronics"*, San Diego, Calif., SPIE Conference No. 439, August 21-26, 1983.
22. T. G. Finstad, "The Annealing Behavior of Ge-Au-Ni, Ge-Au-Pt and Ge-Au-Pd Trilayered Films," *Thin Solid Films* **47**, 279-290 (1977).
23. M. Ogawa, "Alloying Behavior of Ni/Au-Ge Films on GaAs," *J. Appl. Phys.* **51**(1), 406 (1980).
24. A. Piotrowska, A. Guivarch, and G. Pelous, "Ohmic Contacts to III-IV Compound Semiconductors: A Review of Fabrication Techniques," *Solid State Electronics* **26**(3), 179-197 (1983).
25. M. Fouassier, J.-C. Rosier, and J. Dietz, "Resistance à l'Éblouissement des Tubes Intensificateurs d'Images à Galette de Microcanaux et à Double Focalisation de Proximité," *Acta Electronica* **20**(4), 369-378 (1977).
26. R. Price, *Improved Micro-Channel Screen*, Invention Case IL-7264, Lawrence Livermore National Laboratory, Livermore, Calif. (September 20, 1983).
27. G. Schmahl, D. Rudolph, and B. Niemann, *Ann. N.Y. Acad. Sci.* **342**, 368 (1980); *Proc. Conf. on High Resolution Soft X-Ray Optics* **316**, E. Spiller, Ed. (SPIE, Bellingham, Wash., 1982).
28. N. M. Ceglio, *Ann. N.Y. Acad. Sci.* **342**, 65 (1980).
29. J. Kirz, *J. Opt. Soc. Amer.* **64**, 301 (1974).
30. N. M. Ceglio and H. I. Smith, *VIII int'l. Cong. on X-Ray Optics and Microanalysis*, D. R. Beaman, R. E. Ogilvie, and D. B. Wittry, Eds. (Pendell Publishing Co., Midland, Mich., 1980).
31. D. C. Shaver et al., *J. Vac. Sci. Tech.* **16**, 1626 (1979).
32. N. M. Ceglio, A. M. Hawryluk, and M. Schattensburg, *V. Vac. Sci. Tech.* **20** (Nov.-Dec., 1983).
33. C. De Michele and M. Mattioli, "Soft X-ray Spectroscopic Diagnostics of Laboratory Plasmas," *Nucl. Fusion* **21**, 677 (1981).
34. R. W. Lee, J. D. Kilkenny, R. L. Kauffman, and D. L. Matthews, "Electron Density from Time-Resolved Stark Profiles in Ablation Plasmas," *J. Phys. B* (to be published).
35. D. T. Attwood, N. M. Ceglio, E. M. Campbell, J. T. Larsen, D. L. Matthews, and S. L. Lane, "Compression Measurements in Laser-Driven Implosion Experiments," *Proceedings of the Fifth Workshop on Laser Interaction with Matter* (University of Rochester, Rochester, N.Y., 1979).
36. B. L. Henke, H. T. Yamada, and T. J. Tanaka, "Pulsed Plasma Source Spectrometry in the 80-8000 eV X-ray Region," *Rev. Sci. Instr.* **54**, 1311-30 (1983).
37. A. E. Siegman, *An Introduction to Lasers and Masers* (McGraw Hill, New York, N.Y., 1971).
38. D. Campbell, G. Tirsell, and W. Laird, *Improved 50 ps Response X-Ray Detector With Optical Applications*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-86340 (1981).
39. C. L. Wang, R. Kalibjian, and M. S. Singh, "Design of Neutron Streak Camera for Fusion Diagnostics," *Proc. SPIE* **348**, 276 (1982).

40. J. N. Anno, "Secondary Electron Production From Fission Fragments Emerging from Thin Layers of Uranium Dioxide," *J. Appl. Phys.* **33**, 1678 (1962).
41. P. Kirkpatrick and A. V. Baez, "Formation of Optical Images by X-rays," *J. Opt. Soc. Am.* **38**, 766-744 (1948).
42. N. M. Ceglio, R. L. Kauffman, A. M. Hawryluk, and H. Medeck, "Time-Resolved X-ray Transmission Grating Spectrometer for Studying Laser-Produced Plasmas," *Appl. Opt.* **22**, 381-327 (1983).
43. E. B. Saloman, J. S. Pearlman, and B. L. Henke, "Evaluation of High Efficiency Csl and Cul Photocathodes for Soft X-ray Diagnostics," *Appl. Opt.* **19**, 749-753 (1980).
44. 1980 *Laser Program Annual Report*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-50021-80 (1981), pp. 6-7 to 6-11.
45. L. Kissel and F. Biggs, *UFO (Unfold Operator) Computer Program Abstract*, Sandia National Laboratories, Albuquerque, N.M., SAND 82-2250, November 9, 1982.
46. W. S. Derby, M. E. Hummel, E. C. Lee, R. D. Neifert, *SOCKITTOME: An Interactive Data Processing Code*, Lawrence Livermore National Laboratory, Livermore, Calif., UCID-17733 (1977).
47. 1982 *Laser Program Annual Report*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-50021-82 (1983), pp. 6-66 to 6-99.
48. 1980 *Laser Program Annual Report*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-50021-80 (1981), pp. 6-32 to 6-33.
49. T. DeMarco, *Structured Analysis and System Specification* (Prentice-Hall, Englewood Cliffs, N.J., 1978).
50. E. Yourdon and L. Constantine, *Structured Design* (Yourdon Press, New York, N.Y., 1978).
51. *DISPLA User's Manual* (Integrated Software Systems Corporation, San Diego, Calif., Ver. 9, 1981).
52. *ORACLE User's Manual* (Oracle Corporation, Menlo Park, Calif., Ver. 3.0, 1982).
53. *VAX-11 FMS Form Drivers Reference Manual* (Digital Equipment Corp., Maynard, Mass., Ver. 2.0, 1983).
54. D. W. Phillion, E. M. Campbell, K. G. Estabrook, G. E. Phillips, and F. Ze, *Phys. Rev. Lett.* **49**, 1405 (1982).
55. R. Turner et al., *Phys. Fluids* **26**, 579 (1983).
56. K. G. Estabrook, W. L. Kruer, and B. F. Lasinski, *Phys. Rev. Lett.* **45**, 1399 (1980).
57. D. W. Phillion, D. L. Banner, E. M. Campbell, R. E. Turner, and K. G. Estabrook, *Phys. Fluids* **25**, 1434 (1982).
58. R. E. Turner, D. W. Phillion, B. F. Lasinski, and E. M. Campbell, *Phys. Fluids* **27**, 511 (1984).
59. W. Seka, L. M. Goldman, M. C. Richardson, K. Tanaka, R. Short, and E. A. Williar s, paper A10, 13th Annual Anomalous Absorption Conference (1983), Banff, Alberta, Canada; *LLE Quarterly Review* **15**, 19 (Rochester, N.Y., 1983).
60. R. Short and E. A. Williams, paper D2, 13th Annual Anomalous Absorption Conference (1983), Banff, Alberta, Canada; *LLE Quarterly Review* **15**, 12 (Rochester, N.Y., 1983); and to be published.
61. A. V. Vinogradov, I. I. Sobelman, and E. A. Yukov, *Sov. J. Quant. Elect.* **5**, 59 (1975).
62. B. A. Norton and N. J. Peacock, *J. Phys.* **B8**, 989 (1975).
63. R. C. Elton, *Appl. Opt.* **14**, 2243 (1975).
64. A. V. Vinogradov, I. I. Sobelman, and E. A. Yukov, *Sov. J. Quant. Elect.* **7**, 32 (1977).
65. L. I. Gudzenko and L. A. Shelepin, *Sov. Phys. Dok.* **10**, 147 (1965).
66. R. W. Waynant and R. C. Elton, *Proc. IEEE* **64**, 1059 (1976).
67. V. A. Bhagavatula, *J. Appl. Phys.* **47**, 4535 (1976).
68. J. P. Apruseze, J. Davis, and K. Whimery, *J. Phys.* **B11**, 643 (1978).
69. P. Hagelstein, Ph.D. Dissertation, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-53100 (1981).
70. P. Burkhalter, G. Charatis, and P. Rockett, Naval Research Laboratory Report #5022 (1983) and to be published.
71. R. Stewart (Lawrence Livermore National Laboratory), private communication.
72. M. Feldman, A. K. Bhatia, and S. Suckewer, *J. Appl. Phys.* (to be published).
73. P. L. Hagelstein, *Plasma Physics* **25**, 1345 (1983).
74. F. Irons and N. Peacock, *J. Phys.* **B7**, 1109 (1974).
75. G. Pert, *J. Phys.* **B9**, 330 (1976).
76. D. L. Matthews, E. M. Campbell, K. Estabrook, W. Hatcher, R. L. Kauffman, R. W. Lee, and C. Wang, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-90279 (1982).
77. 1982 *Laser Program Annual Report*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-50021-82 (1983), pp. 5-10 to 5-12.
78. C. DeMichelis and M. Mattioli, *Nucl. Fusion* **21**, 677 (1981).
79. M. Galanti and N. J. Peacock, *J. Phys.* **B8**, 2427 (1975).
80. V. A. Boiko, S. A. Pikuz, and A. Ua. Faenov, *J. Phys.* **B12**, 1889 (1979).
81. P. G. Burkhalter, M. J. Herbst, D. Duston, J. Gardner, M. Emery, R. R. Whitlock, J. Grun, J. P. Apruzese, and J. Davis, *Phys. Fluids* **26**, 3650-3659 (1983).
82. *Central Laser Facility Annual Report*, 1981, Rutherford and Appleton Laboratories, Chilton, Didcot, England, RL-81-040.
83. 1982 *Laser Program Annual Report*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-50021-82 (1983), pp. 5-54 to 5-57.
84. C. L. S. Lewis, M. J. Lamb, J. D. Kilkenny, S. Keats, and M. H. Key, *Central Laser Facility Annual Report*, 1980, Rutherford Laboratory, Oxford, England, RL-80-026, p. 61.
85. R. W. Lee, B. L. Whitten, and R. E. Strout III (to be published).
86. K. Mizuno, J. S. DeGroot, and F. Kehl, *Phys. Rev. Lett.* **49**, 1004 (1982).
87. W. Woo et al., *Proc. 12th Anomalous Absorption Conf.*, Santa Fe (1982), GP-12 (Los Alamos National Laboratory, Los Alamos, N.M., 1982).
88. W. L. Kruer and J. M. Dawson, *Phys. Fluids* **15**, 446 (1972).
89. J. J. Thompson, R. J. Faehl, W. L. Kruer, and S. Bodner, *Phys. Fluids* **17**, 973 (1974).
90. J. S. De Groot and J. I. Katz, *Phys. Fluids* **15**, 401 (1973).

References

91. W. Woo, P. Rambo, and J. S. DeGroot, *Bull. Am. Phys. Soc.* **26**, 861 (1981). To be published in *Phys. Fluids*.
 92. P. Rambo, W. Woo, J. S. DeGroot, and K. Mizuno, UCD Report PRG-R-93, submitted to *Physics of Fluids* (1983).
 93. J. M. Dawson and C. Oberman, *Phys. Fluids* **5**, 517 (1962).
 94. J. M. Dawson and C. Oberman, *Phys. Fluids* **6**, 394 (1963).
 95. R. J. Jaehl and W. L. Kruer, *Phys. Fluids* **20**, 55 (1977).
 96. J. M. Dawson, W. L. Kruer, and B. Rosen, in *Dynamics of Ionized Gases*, M. J. Lighthill, I. Imai, and H. Sato, Eds., p. 47 (University of Tokyo, Tokyo, 1973).
 97. K. Mizuno and J. S. DeGroot, *Phys. Rev. Lett.* **35**, 219 (1975).
 98. K. Mizuno and J. S. DeGroot, *Phys. Fluids* **26**, 609 (1983).
-

Electron micrograph of low-density silica particles produced by a sol-gel process. These particles, applied to a glass substrate, produce high-density threshold antireflection coatings. The individual particles are about 20 nm in diameter.

Laser Research and Development

6

Laser Research and Development

Contents

Section Editor: J. R. Murray

Introduction	6-1
Solid-State Amplifiers	6-2
Gain Performance of Nova Amplifiers	6-2
Single-Segment Amplifier	6-7
Silicone Edge Claddings	6-8
Parasitic Oscillations in Rectangular-Slab Amplifiers	6-11
Compact Amplifiers	6-13
Flashlamps	6-18
Large-Bore Flashlamps	6-18
Transient Response of Flashlamps	6-23
Blackbody Pumping and Depumping of Nd:Glass	6-29
Optical Materials, Surfaces, and Coatings	6-31
Introduction	6-31
Sol-Gel Coatings	6-32
Etched Nuclear Tracks	6-34
Laser-Damage Studies	6-35
Thermal-Lensing Studies: Two-Photon Absorption and Solarization	6-39
Laser Glass	6-40
Concentration Quenching	6-40
Thermomechanical Properties	6-44
Ultraposphates	6-46
Broad-Band Energy Extraction	6-48
Comparison of Laser Glasses	6-56
Frequency-Conversion Systems	6-60
Damage in Conversion Cells	6-60
Spontaneous and Stimulated Raman Scattering in KDP and Index-Matching Fluids	6-61
KDP Growth	6-69
High-Dynamic-Range Frequency Conversion	6-69
Noncollinear Phase-Matching Schemes	6-74
Incorporation of Frequency Conversion in MALAPROP	6-77
Large-Aperture Optical Switches	6-79
Plasma-Electrode Pockels Cell	6-79
Plasma-Electrode Harmonic-Conversion Switch	6-84
Subharmonic Injection into Multipass Cavities	6-88
Advanced Solid-State Systems	6-91
Theory of the Electronic Structure of Impurities in Crystals	6-91
High-Average-Power, Solid-State Crystalline Lasers	6-94
Other Laser Systems	6-96
Rare-Gas-Halide Lasers: Experiments on the Rapier-B Amplifier	6-96
Free-Electron Lasers: Experiments at the Electron-Laser Facility	6-100

Atomic and Molecular Theory	6-102
Collisional Redistribution of Radiation	6-102
Photodissociation of H ₂ S	6-103
TDHF Theory of Molecular Collisions	6-105
Resonances in Molecular Collisions	6-106
Resonance-Resonance Coupling in Dissociative Electron Attachment	6-107
Hole Localization in Inner-Shell Excitation	6-109
Scattering Amplitudes Using Complex Basis Functions	6-111
Atomic Physics of Laboratory Soft X-Ray Lasers	6-113
References	6-118

Laser Research and Development

Introduction

W. H. Lowdermilk

The Laser Research and Development Program supports Nova and advanced laser systems for inertial-confinement fusion (ICF) experiments, for weapons-physics experiments, and for eventual ICF commercial-power generation through development of required materials, components, and systems. During 1983, we extensively reviewed the design of high-energy Nd:glass laser systems to reduce system costs. We considered a variety of system architectures, amplifier geometries, and requirements for improved laser glass and harmonic-conversion materials. One outcome of our study was that the efficiency and cost of the largest amplifiers are dominant factors in the cost of the system. Consequently, we have been developing new amplifiers that are designed to approach theoretical limits of efficiency. We have designed and built a test-bed amplifier with adequate flexibility to test the effects of variables such as flashlamp-packing fraction, reflector shape and coating, and basic amplifier geometry.

We have extended our earlier measurements on the overall radiative efficiency and neodymium-pumping efficiency of xenon flashlamps with various diameters. In addition, we have improved our understanding of the transient response of flashlamps.

Our work this year on laser damage and coatings focused on damage-resistant, graded-index, antireflective (AR) coatings for fused-silica optics on Nova. Two different coating processes were developed independently by Westinghouse R&D Center and Corning Glass Works. We also attempted to produce highly transmissive surfaces on fused silica using etched nuclear tracks. Although this process appears to be applicable to other materials, it leads to unacceptable losses from surface scattering when applied to fused silica.

Because of damage to the frequency-conversion cells on Novette, we reopened our search for index-matching fluids. We found improvement with a fluid closely related to one used by the Laboratory for Laser Energetics at the University of Rochester.

It is important to know the Raman-scattering cross sections of materials used in the frequency-conversion cells so that the cells can be designed to prevent loss due to stimulated Raman scattering. Consequently, we measured the Raman cross sections of potassium dihydrogen phosphate (KDP) and of some index-matching fluids of interest.

To supply frequency-conversion crystals for large laser systems of the future, we began research on increasing the growth rate of KDP. We also initiated a systematic search for new nonlinear materials having performance advantages over KDP.

We are working with Schott Glass Technologies, Inc. and Hoya Optics, Inc. (Fremont, California and Tokyo, Japan) on the development of new glass compositions with low concentration quenching. Such glasses could find application in future lasers designed to produce high average power.

The advantages of large regenerative-amplifier systems over current master-oscillator power-amplifier (MOPA) laser designs are well known. However, such systems require a large-aperture optical switch; to date, no practical

concepts for such switches have emerged. We have recently demonstrated a longitudinal Pockels-cell switch in which low-density plasma forms a transparent, conductive layer. This concept is readily scalable to large diameters.

We are searching for candidate laser systems for eventual ICF commercial-power generation. Such systems will need to produce high average power at a short wavelength (<500 nm) and with high efficiency ($>10\%$). In principle, both crystalline-host lasers and gas lasers have appropriate characteristics to satisfy these requirements; however, no known system does so at present.

We completed our measurements on the KrF Rapier-B amplifier, and the system performed as predicted. We also continued to support the free-electron-laser experiments being conducted in M Division.

Solid-State Amplifiers

Gain Performance of Nova Amplifiers

Measurement Techniques. During 1983, we improved the versatility and data-taking accuracy of our amplifier measurement facility. We physically separated the amplifier measurement facility from the flash-lamp measurement laboratory (see "Large-Bore Flashlamps," later in this section) and established independent, flexible capacitor banks in both. Each bank now consists of 20 separate circuits (connected to a common ignitron switch); each circuit has a diverse set of capacitors and inductors, where the elements of each set can be connected electrically either in series or in parallel. Capacitor selection consists of three $52.5\text{-}\mu\text{F}$ and three $14.5\text{-}\mu\text{F}$ capacitors; inductor selection consists of 200-, 300-, 450-, and $600\text{-}\mu\text{H}$ inductors. With this choice of components, we can obtain a broad range of

input energies and pulse lengths while maintaining reasonable values for the circuit-damping factor.

We made extensive measurements¹ of circuit resistance, energy-transfer efficiency, and balance among the 20 circuits for our new bank. Using a 1-kHz impedance meter, we found that circuit resistances varied from 0.37 to $0.43\ \Omega$. The values were in good agreement with those found by firing into a short circuit and measuring the time decay of ringing. The energy-transfer efficiencies into a $2.1\text{-}\Omega$ load (similar to the effective lamp impedance) were $85 \pm 2\%$ and generally agreed with the values expected from resistance measurements. We obtained excellent balance among circuits.

Our previous range of useful charge voltages for the banks was from 12 to 24 kV. The upper limit was set by component stress, while the lower limit was previously set by our ability to trigger the lamps. To aid with lamp triggering at low voltage, we modified the circuit that was previously used solely for the pulse-ionization lamp check (PILC). The PILC circuit can now provide a trigger pulse $100\ \mu\text{s}$ before the main bank is switched. The pulse, typically 12 kV from a $14.5\text{-}\mu\text{F}$ capacitor is injected on all circuits in parallel and adequate to break down the lamps. The pulse allows reliable lamp triggering down to bank voltages as low as 4 kV and greatly increases the dynamic range of our gain measurements.

We made gain measurements with an etalon-tuned, continuous-wave (cw), neodymium:yttrium-aluminum-garnet (Nd:YAG) laser. For measurements of phosphate laser glasses, we used the $1.052\text{-}\mu\text{m}$ YAG line, which lies close to the peak of

Fig. 6-1. Schematic diagram of setup for gain measurements.

the gain for phosphate glass. A schematic of our measurement setup is shown in Fig. 6-1. Both signal and reference beams were monitored with silicon photodiodes. A grating was used on the reference beam to check for operation on the desired YAG line. The transmitted beam passed through a spatial filter and a bandpass filter to eliminate unwanted background light. We used a television camera to monitor the position of the focused beam relative to the pinhole.

To improve the accuracy of data taking, we used digital oscilloscopes to record the transmitted and reference beams. The oscilloscopes were connected to an ISI-11/23 computer for signal processing. Thus, we corrected for fluctuations in laser power by dividing the transmitted signal values by the reference signal values, and smoothed the resulting gain signals to find an accurate value of the peak gain.

Figure 6-2 shows an example of our data treatment. The raw, transmitted signal, shown in Fig. 6-2(a), is the level of the transmitted signal plus a positive excursion reflecting the amplifier gain vs time. Dividing out the reference signal and normalizing gives the gain signal shown in Fig. 6-2(b). Notice that we have obtained a flat baseline and that gross fluctuations have been removed. To obtain the peak gain, we digitally smoothed around the peak via the Savitsky-Golay method.² The smoothing, shown in Fig. 6-2(c), derived from a weighted average from each of 49 neighboring points. The weighting coefficients are fixed numbers that can be calculated by fitting 49 equally-spaced points to a cubic polynomial to determine the value at the central point. Such smoothing substantially increased the accuracy of our peak

determination. The measured gain coefficients were generally reproducible to $\pm 0.5\%$ for a given amplifier.

For gain measurements, the amplifiers were purged with dry nitrogen, and the ends were terminated with bead-blasted aluminum plates. Although the effect of end-plate termination is not known, our findings were in good agreement with the available Novette measurements³ without end plates.

Gain Results. Figure 6-3 is a summary of our measurements of the peak gain on the centerline for the five Nova disk amplifiers

operating at $1.053\ \mu\text{m}$. We obtained these values by increasing the gain coefficients measured at $1.052\ \mu\text{m}$ by 3% (Ref. 4). The gains are plotted as a function of explosion fraction, the ratio of the energy stored in the bank per lamp to the explosion energy per lamp. The explosion fraction is a convenient variable for comparing the spectral output of lamps (see "Large-Bore Flashlamps") and is also a measure of how severely the flashlamps are stressed. The explosion energy of a flashlamp in joules is obtained from the empirical formula⁵

$$E_{\text{exp}} = 20\,000\,ld\,(lC)^{1/4} \quad (1)$$

where l is the arc length (cm), d is the bore diameter (cm), L is the circuit inductance (H), and C is the capacitance (F).

The highest explosion fractions in Fig. 6-3 were taken at a charge of 24 kV. Voltages decreased (by 2-kV intervals) to 4 kV for the two smallest-aperture amplifiers and to 6 kV for the others. The major difference in gain among these amplifiers is the varying numbers of disks in each head. The 9.2-cm amplifier has six disks, the 15-cm has four,

Fig. 6-2. Gain waveforms (a) as received from the signal photodiode, (b) after dividing by the reference signal to remove laser fluctuations and normalizing, and (c) after smoothing near the peak.

Fig. 6-3. Gain performance of Nova disk amplifiers vs input energy.

Fig. 6-4. Gain profiles of a 31.5-cm amplifier measured in the horizontal (major) and vertical (minor) axis directions.

the 20.8-cm has three, and the 31.5- and 46-cm amplifiers both have two disks. Mechanical and electrical details regarding these amplifiers can be found in Ref. 6.

Measurements of horizontal and vertical uniformity in the peak gains are shown in Fig. 6-4. These data were taken at a charge of 20 kV for the 31.5-cm amplifier. The variation of gain with horizontal position (along the major axis of the disk) shown in Fig. 6-4 is typical of amplifiers with an even number of disks. The observed asymmetry arises from the less intense pumping of laser glass near the ends of the amplifier. Since the disks are mounted in a zig-zag pattern, as seen from above, with an even

number of disks the more weakly-pumped disk ends are on the same side of the horizontal centerline. Higher gain occurs on the side (+ direction) where the two disks meet. Of the five Nova amplifiers, only the 20.8-cm amplifier had an odd number of disks. Its major-axis profile was both symmetric and flat to within 1% across 90% of its clear aperture. The gain profiles shown for the 31.5-cm amplifier in Fig. 6-4 are typical of box amplifiers in which pumping occurs on only two sides. For cylindrical amplifiers, in which flashlamps are positioned all around the disks (our 9.4- and 15-cm amplifiers), the profiles are more rounded.

A comparison of the Nova amplifier efficiencies as a function of input energy is shown in Fig. 6-5. Efficiency is defined as the ratio of the stored laser energy at peak gain to the energy initially stored in the capacitor bank. Efficiency is again plotted vs explosion fraction for comparison with the results in "Large-Bore Flashlamps." The stored laser energy was calculated from the centerline peak-gain coefficients adjusted to $1.053 \mu\text{m}$. At that wavelength, we assumed that the gain cross section was $4.0 \times 10^{-20} \text{ cm}^2$ for laser glass⁴ from both Schott Glass Technologies, Inc. (LG-750) and Hoya Optics, Inc. (LHC-8). The volume of laser glass was assumed to be uniformly excited and included only material that was optically accessible to the beamline.

Figure 6-5 shows an overall ordering of amplifier efficiency with respect to size; the larger amplifiers are generally more efficient. A contributing factor in the ranking is that the two smallest amplifiers are cylindrical rather than box designs. Previous measurements⁷ have shown that the box design is $\sim 30\%$ more efficient at comparable pumping conditions. In addition, there is some increase in efficiency for the larger amplifiers because of their larger disk thicknesses and, hence, greater doping-thickness products, giving larger absorption of pump light. The final factor favoring larger size is the ratio of surface area of laser glass to surface area of reflectors. The ratio relates to the relative absorption of pump light by laser glass as opposed to absorption by reflectors. This area ratio monotonically increases with size from ~ 0.15 to 0.3 for these five amplifiers.

The variation of amplifier efficiency with explosion fraction resembles the relative

pumping efficiency of Nd:glass by an open flashlamp (see Fig. 6-17 in "Large-Bore Flashlamps"). In both cases, the peak efficiency occurs at low input. However, the maximum efficiency point of these amplifiers is near an explosion fraction of 0.04, while that of an open lamp is near 0.15. In addition, the decline in efficiency with input for amplifiers is much more dramatic than for open-lamp pumping of Nd:glass. At an explosion fraction of 0.25, the decline in amplifier efficiency from its peak is ~50%, while the decline of the open-lamp pumping efficiency from its peak is <10%. Part of this difference can be explained by loss due to amplified spontaneous emission (ASE).

Optical feedback in a cavity can also effectively load the flashlamps at a higher value than expected from the electrical input alone. However, we find that these effects are not large enough to account for the observed rolloff in pumping efficiency.

We must attribute a major portion of the efficiency decline in Fig. 6-5 to the competition between absorption of pump light by the laser disks and by the flashlamps themselves. At low lamp loadings the flashlamps are relatively transparent, favoring absorption by the laser disks. At high lamp loadings the lamps become relatively "black," thus decreasing the transfer efficiency to the laser glass. Based on results in "Large-Bore Flashlamps," when pump radiation that lies within the Nd-pumping region (from 0.4 to 1 μm) is absorbed by a flashlamp, it is recycled to the same region with an efficiency of <50%. Thus, in the design of amplifiers for optimum efficiency, it is important that the lamps not view each other directly and that pump light first pass through the laser glass before striking another lamp.

Nd-Doping Variations. Since the output energy of the Nova laser is strongly affected by the performance of the largest amplifier, we were particularly interested in increasing the gain of the 46-cm amplifier. Hence, we investigated its performance using laser disks with 3% Nd doping, as opposed to the standard 2% doping. Greater Nd doping may improve performance because the larger doping-thickness product produces greater absorption of pump light. However, this improvement trades off against the reduction in excited-state lifetime that results from higher doping

▲ Fig. 6-5. Peak-storage efficiency vs input energy ranks amplifier efficiency by size.

Fig. 6-6. Improved gain coefficient of the 46-cm amplifier using 3 vs 2%-doped disks.

densities. Calculations with VODEP indicate that going from 2 to 3% doping should increase instantaneous pumping by 13% in the 4.3-cm-thick, 46-cm-amplifier disks. Although this is a relatively small effect, gain improvements of a few percent were deemed significant.

We obtained two 3%-doped, 46-cm half disks (46-cm-amplifier disks are split in half to reduce ASE) from Schott Glass Technologies, Inc. The disks were mounted on one side of an amplifier; standard 2% disks were mounted on the other side. The peak gain coefficients measured for the 3% disks are compared in Fig. 6-6 to gain coefficients with 2% disks mounted in both positions in the amplifier. Although the improvement declines with input, the advantage in using higher-doped disks is clear. At an operating point of 20 kV, the measured gain increased from 1.69 to 1.80. Slightly smaller improvements might be expected for an amplifier completely loaded with 3%-doped disks.

The measured improvement factor is also consistent with the relatively small decrease in lifetime of the upper laser level by increased doping. For small samples of the glass used in the amplifier experiments, Schott found that the lifetime decreased from 373 to 357 μs . Our measurements of the lifetimes of the large disks (including radiation trapping, which increases the lifetime) decreased from 402 to 393 μs (Ref. 8). Such a relatively small change in lifetime with Nd concentration is consistent with the high purity of laser glass with respect to other contaminant ions. Based on these measurements, we purchased 3%-Nd-doped disks for both the 46- and 31.5-cm amplifiers. The disks will be added to the baseline Nova design.

Measurements of ASE. We also measured the influence of ASE in the three largest amplifiers. Even with perfect edge claddings, which absorb all ASE incident on the disk edges, a large fraction of the excited-state population can be lost to single-pass ASE by the time peak inversion is reached. To measure the gain reduction produced by ASE, we sampled flashlamp light from the amplifier and monitored the fluorescence it produced in a piece of Nd:glass. The greatly reduced pump intensity and small volume of pumped glass involved in this peak-fluorescence detector (PFD) virtually eliminated ASE effects. If the PFD sees the same flashlamp spectrum as the laser disks, its signal should be proportional to what the stored energy in the amplifier would be without ASE. The falloff of amplifier gain coefficient with increased

pumping, compared to the PFD signal, would then show the influence of ASE.

The difficulty, in practice, was to ensure that the PFD sampled the same excitation spectrum as the laser disks. Of particular concern was the change in emission spectrum of the flashlamps with angle relative to the lamp axis. The transversely oriented lamps of the 46-cm amplifier allowed the PFD to view the lamps at right angles from the end of the amplifier. The longitudinal lamps of the other amplifiers, however, prevented viewing lamp emission normal to their axes and, thus, required a different light-collecting scheme. For these amplifiers, we used a probe on the PFD that was inserted into the amplifier. Quartz shards, filling a small quartz bulb on the end of the probe, scattered light into a quartz fiber optic, which coupled it to the PFD.

Figure 6-7 shows data from the 20.8-, 31.5-, and 46-cm amplifiers compared to a GAINPK calculation for the 20.8- and 46-cm amplifiers. The GAINPK prediction for the 31.5-cm amplifier drops somewhat faster than the curve shown in Fig. 6-7 because of the longer flashlamp pulse used with that amplifier. However, the difference was not significant compared to the data spread. The figure shows the ratio of gain coefficient, α , to PFD signal as a function of αD , where D is the major axis of the elliptical laser disks. We normalized this ratio by fitting each data set to a second-order polynomial, in a least-squares sense, and scaling the ratio to make the fit unity at $\alpha D = 0$. We then allowed a fine-tuning of each normalization by as much as the error in the zero intercept to obtain the best agreement with the GAINPK calculation. The intercepts found from least-squares fitting for the three data sets in Fig. 6-7 were fine-tuned by +2, 0, and +3% for the 20.8-, 31.5-, and 46-cm amplifiers, respectively.

Data for the 46-cm amplifier could not be independently compared to data from the other amplifiers or to GAINPK calculations. Its disks, which are split along their minor axes, have a much different shape, and the influence of ASE on stored energy depends on shape as well as the dimension⁹ of the largest disk. Instead, we used the PFD data to determine an effective l for the split disk, which corresponded to the major axis of an equivalent elliptical disk that would produce the same ASE effect. The fit of 46-cm data shown in Fig. 6-7 required an

Fig. 6-7. Experimental ASE data for 20.8-, 31.5-, and 46-cm amplifiers compared to GAINPK calculation.

effective D of 75 ± 4 cm, or 85% of the major axis of the unsplit disk.

The effective D obtained for the 46-cm split disks appears to be too large as judged by Monte Carlo calculations of ASE for a split-disk geometry.¹⁰ We suspect that looking directly at the lamps, as the PFD did for the 46-cm data, gives a different spectrum than the disks see, thus causing an overestimate of the ASE effect. The disks see an angle-averaged spectrum, including all emission angles from the lamps, as well as spectral components degraded by reflection off the reflectors. However, the fact that the 20.8- and 31.5-cm data fit a common curve that closely matches the GAINPK calculation indicates a reasonably accurate model in GAINPK for predicting the influence of ASE. Both the data and calculation for the 31.5-cm amplifier indicate that stored energy in the amplifier would be 50% higher at maximum bank energy without ASE.

Authors: J. E. Murray and H. T. Powell

Major Contributors: D. H. Ford, R. W. Holloway, B. T. Merritt, and B. W. Woods

Single-Segment Amplifier

We have constructed an extremely flexible Nd:glass amplifier, called the single-segment amplifier (SSA), which we will use to measure the dependence of amplifier efficiency on a wide variety of parameters. The SSA could represent a single segment of a very large segmented-array amplifier;

however, in a more general sense, it will be a testbed to provide the fundamental data necessary to design a new generation of glass-laser amplifiers. The baseline design of the SSA was chosen to lie in a parameter space that should provide better pumping efficiency¹¹ than past designs. By contrast, Nova amplifiers were largely designed to optimize the gain coefficient, which is of paramount importance for glass-laser systems designed for ≤ 1 -ns pulse lengths.

A schematic diagram of the SSA is shown in Fig. 6-8. The disks are rectangular to maximize the area available for pumping, and the nominal aperture is 32-cm square. The laser disks presently are 4% Nd-doped (4.3-cm-thick) laser glass of the LG-660 and LG-750 types manufactured by Schott Glass Technologies, Inc. LG-660 is a relatively low-gain-cross-section (2×10^{-20} cm²) silicate glass; LG-750 is the high-cross-section (4×10^{-20} cm²) phosphate glass used in Nova. The disks were purchased without an absorbing cladding to prevent internal, parasitic oscillation. We plan to use the silicone-rubber-based edge cladding, discussed in the next article. We analytically investigated possible parasitic oscillation using silicone claddings, as discussed in "Parasitic Oscillations in Rectangular-Slab Amplifiers," later in this section, and found that such oscillation is not a problem.

The SSA is four disks long to minimize end losses. To check the effect of the ends, both absorbing and reflective end pieces are being fabricated. Since our long amplifier requires unusually long flashlamps,

Fig. 6-8. Schematic diagram of the single-segment amplifier.

we purchased 1.5-cm-bore $\times$ 230-cm arc-length lamps. They have been found to perform reliably.

The flashlamp reflector shown in Fig. 6-8 is of the Winston-Rabl type.¹² These shapes reflect light emitted by a lamp without passing any light through the lamp plasma. We will compare this reflector geometry to simpler shapes and to other Rabl-shaped reflectors that are of various depths, which provide either more or less collimation of the pump radiation. Thus, in general, we will test a variety of reflectors to determine the sensitivity of performance to reflector shape and lamp-packing fraction.

The top and bottom reflectors of the SSA are designed so that the disk holders are outside the pump cavity to the extent possible. This design maximizes the ratio of glass-to-reflector area, a positive trend, as discussed in the previous article. We have obtained vacuum-deposited silver reflectors with reflectivity $>95\%$ at 500 nm compared to the 90% value for the electroplated silver reflectors used in Nova. By using reflectors of two silver types and polished stainless steel (70% reflectivity), we will determine the sensitivity of amplifier performance to cavity reflectivity.

We plan to make numerous other variations, such as changing the pump pulse length and AR processing of the laser disks and blast shields. We hope to obtain a substantial increase in efficiency over the Nova amplifiers and to obtain a sizeable fraction of the potential performance of Nd:glass amplifiers, as discussed in "Large-Bore Flashlamps," later in this section.

Author: H. T. Powell

Major Contributors: F. A. Frick, G. C. Lee, J. B. Trenholme, and M. D. Vergino

Silicone Edge Claddings

Disk amplifiers require edge claddings on each laser disk to prevent feedback of ASE in the disks. Measurements on a 31.5-cm-aperture disk amplifier indicated that single-pass ASE reduces stored energy by one-third at peak gain. Feedback from the disk edge allows additional passes for ASE, rapidly increasing the loss of stored energy and presenting the possibility of parasitic lasing in the disks. Glass edge claddings used on Novette and Nova perform

extremely well but are expensive and must be reformulated for each type of laser glass. Because we plan to test several different laser glasses in the SSA, we developed a silicone rubber cladding to avoid the time and expense associated with developing several new glass claddings.

The rubber cladding consists of room-temperature-vulcanized (RTV) rubber poured around the edge of a laser disk with plates of filter glass imbedded in the rubber to absorb ASE. The clear, nonabsorbing rubber acts only to couple ASE from the disk into the absorbing filter glass. The elastic and RTV properties of the rubber accommodate a wide variety of thermal and mechanical properties in the laser and filter glasses; the elastic properties provide an excellent stress-distributing and shock-absorbing mount for the disk.

The rubber must be index-matched to the laser and filter glass to eliminate Fresnel reflection, the principal source of feedback from the edges. The rubber must also be free of bubbles and other scattering centers, and it must survive both flashlamp and ASE fluences of several J/cm^2 without being damaged.

Commercial RTV silicone rubbers meet most edge-cladding requirements. Small edge-cladding samples of commercial products showed damage thresholds several times the expected ASE fluence, and the rubbers cured without scattering centers and with good elastic properties. However, their indexes of refraction fell below 1.42 at 1000 nm, compared to the 1.52 typical of laser glasses of interest. An index match of 1.42/1.52 is adequate to keep bulk parasitic modes below threshold (see the following article), but is not necessarily adequate to suppress surface modes.

Figure 6-9(a) shows bulk and surface parasitic modes. Bulk modes reflect from the disk surfaces at angles greater than the critical angle for total internal reflection and experience loss only at the disk edges. The maximum-gain bulk mode reflects at the critical angle that maximizes its path length between edges at n times the disk dimension for a disk with refractive index n . Surface modes run parallel to and just under the disk surfaces, taking advantage of higher gains near the surfaces.

Figure 6-9(b) shows that surface gain can be much higher than the gain averaged through the disk thickness.

Fig. 6-9. (a) Cross section of laser disk showing surface- and bulk-parasitic modes and (b) ratio of gain coefficient to its average over the thickness of the disk vs distance into the disk for three values of doping-thickness product.

Doping-thickness products (wt% doping of Nd ions times disk thickness in centimeters) greater than about 3 give a higher gain for surface modes than for bulk modes (despite the longer path length of the bulk mode), and surface modes reach threshold first. Since we want to investigate amplifier performance for doping-thickness products as great as 20, we require a much better index match than is available from commercial RTV rubber.

A damage threat to the cladding from above-threshold parasitic modes also necessitates a better index match. Without lasing action, ASE fluence on the cladding is fairly uniform, because most of the ASE paths bounce between surfaces and average the gain. The average ASE fluence on the cladding of a 31.5-cm amplifier, for example, is $\sim 5 \text{ J/cm}^2$ and presents no threat to the rubber. Above-threshold parasitic modes, either surface or bulk, collect energy from large volumes of the disk and preferentially deposit it in higher-loss regions of the cladding. Therefore, ASE energy can be concentrated into very small absorbing regions, giving much higher than average fluences in those regions. Thus, ensuring no damage also requires a much better index match to suppress surface modes than is available with commercial RTV rubber.

We have tailored a silicone rubber to increase its refractive index without losing other desirable properties, such as a sufficiently low viscosity to allow pouring,

Fig. 6-10. Refractive index of experimental silicone polymers increases with wt% of phenyl-containing monomer.

room-temperature curing, and product clarity. The silicone polymer, which can be cross-linked to form a rubber, is synthesized from two siloxane monomers: one contains methyl substituents; and the other, phenyl. A catalyst breaks the ring-like structure of these monomers to form short, linear segments that attach end-to-end to form chains. The ratio of phenyl to methyl substituent determines the refractive index. Figure 6-10 shows the results of varying this ratio. We found that a polymer made with 63 wt% of the phenyl-containing monomer gives the desired refractive index.

The average length of the chain, which determines the viscosity of the uncured rubber, can be controlled by introducing a third reactant, called an end blocker. This material has a linear structure and contains vinyl

groups. The catalyst that cleaves the rings of the two monomers breaks the end blocker into two segments. Each attaches to an end of the siloxane chain, stopping its growth. The amount of end blocker in the original mixture determines the average length of the polymer molecule and the viscosity of the mixture. We found that a viscosity of $<10^4$ cp was necessary for reasonable pourability of the uncured rubber. The vinyl in the end blocker also provides the cross-linking sites by which the liquid polymer is converted into a solid rubber. Polymer formation is completed by heating the reaction mixture to destroy the catalyst, two filtrations (0.5- μ m Teflon) to remove the residual monomers and catalyst end products, and an intermediate stripping process to drive off volatiles (150°C under argon).

Rubber formation requires a second cross-linking polymer containing silane to interconnect the molecules of the parent polymer. In the presence of a second catalyst, each of a large number of silane subgroups in the cross-linking molecule can attach to either vinyl group at the ends of the parent polymer. The many attaching sites of each cross-linking molecule connect many parent polymer molecules, giving a strongly interconnected, spaghetti-like mass of molecules in the final cured rubber. We found that ~5 wt% cross-linker gives the desired hardness and friability of the final rubber.

We require clarity in the final rubber as well as the proper refractive index to avoid ASE feedback from the edge cladding. The most common potassium-based catalyst for the polymer-forming step uses a silica filler to prevent decomposition of the cured

rubber with heat; however, the silica filler makes the cured rubber cloudy. We use an ammonia-based catalyst that does not require the silica filler for stability. We also found that a high phenyl cross-linker was needed for clarity.

The catalyst used in the cross-linking step determines both the temperature and rate of curing. Our edge-cladding application requires a room-temperature cure because of the large thermal-expansion coefficient of the rubber. If cured at an elevated temperature, the rubber contracts on returning to room temperature and tends to separate from the disk or absorbing plates. A platinum catalyst at a concentration of 5 ppm gives curing at room temperature and an adequate working time of ~2 h. The platinum catalyst can be mixed with the parent polymer and has no effect until the cross-linking agent is added. The parent polymer with platinum catalyst is mixed with the cross-linking agent to initiate curing, and the mixture can be de-aired with a mechanical pump to eliminate bubbles.

We demonstrated the performance of the new rubber edge cladding with full-scale tests in the largest available unsegmented amplifiers. We first clad a 31.5-cm phosphate-glass disk and compared its performance in a Nova amplifier to a similar disk with the Nova glass cladding. Figure 6-11 shows the results for two passes through the amplifier. Performance of the rubber-clad disk was identical to that of the glass-clad disk up to the nominal operating point of the amplifier at 312 kJ. At higher inputs, performance of the rubber-clad disk fell off slightly, but its gain was down by <1% at the maximum input of 449 kJ. The means by which we sealed the disk into its holder to pour the cladding may have caused this falloff. We formed a seal between disk and holder, prior to pouring the cladding, with a small quantity of the commercial RTV rubber with an index of 1.42. The seal formed an irregular 2- to 4-mm-wide band around one face, with an index mismatch of 1.42 to 1.52. This mismatch and the disk's doping-thickness product of 8.6 meant that a surface mode reached threshold at an amplifier gain of 1.48. Since the rubber-clad disk reached gains as high as 1.87, surface modes would have been well above

Fig. 6-11. Performance comparison of one glass-clad to one silicone-clad disk in a 31.5-cm amplifier (2-pass gain).

threshold and may have reduced the gain near that surface enough to explain the observed falloff.

Our first test demonstrated the desired ASE suppression capability of the cladding, but the rubber was extensively damaged. Most of the damage resulted from direct exposure to flashlamp light. Tests showed that the rubber survived direct exposure within 1 cm of an open flashlamp, but the environment inside the amplifier proved to be considerably more severe. The directly exposed rubber in the gap between disk and holder survived only when recessed well back from the amplifier interior. Rubber exposed through a 0.5 mm gap survived if recessed by 5 mm.

Our first test also resulted in some bulk damage. All such damage seemed to originate at either the disk edge or a surface of the absorbing plates and probably resulted from surface contamination. To minimize such problems for subsequent claddings, all components were cleaned in a class-100 clean room, and the rubber was prepared and poured in the same facility.

To prove survivability of the rubber cladding, we switched to the SSA hardware. The 31.5-cm holder had been designed for Nova glass-clad disks and had inherent sealing and shielding problems; the SSA disk holder had been specifically designed with adequate shielding and provision for a liquid seal between disk and holder. The SSA test used rectangular disks of silicate glass, Schott LG-660, with a cross section of $2 \times 10^{-20} \text{ cm}^2$, one-half that of the 31.5-cm phosphate disk previously tested. Consequently, its rubber cladding was not subjected to a parasitic gain as large as in the previous tests. However, the damage threat in the SSA test was comparable because the stored energy in the SSA disk was comparable. Both tests generated ASE fluences, estimated at 5 J/cm^2 , on the cladding. Inspection of the cladding after testing showed no change and confirmed the durability of the cladding.

In summary, we have developed and tested in large-aperture amplifiers a new RTV, silicone-rubber-based edge cladding that is potentially much less expensive than our currently used glass claddings. The refractive index of the rubber has been

tailored specifically to 1.52, but the index can be adjusted from 1.42 to 1.52 to allow index matching to most silicate, phosphate, and fluorophosphate glasses. The elasticity of the rubber cladding provides an excellent stress-distributing and shock-absorbing mount for the disk. The low thermal conductivity of the rubber could also be used to reduce transverse thermal gradients in high-average-power applications.

Author: J. E. Murray

Major Contributors: M. O. Riley, R. J. Poli, H. T. Powell, and B. W. Woods

Parasitic Oscillations in Rectangular-Slab Amplifiers

The basic limitations placed on circular- and elliptical-disk amplifiers by parasitic oscillations were derived by Trnholme¹¹ over a decade ago. In 1982, we used a similar approach to calculate the effect of parasitics in rectangular glass slabs proposed for use in the SSA. We have identified the lowest-threshold modes and have determined the requirements for their suppression.

To model parasitic oscillations, we used the ray-tracing approximation developed previously.¹¹ The analysis treats only the growth of radiation amplitude and ignores the phase of the waves and any diffraction effects. We also restricted our treatment to specular (Fresnel) reflections because diffuse reflection cannot be dealt with simply in a ray-tracing approach. Under these limitations, both lossless (total internal reflection) and lossy (partial internal reflection from at least one interface) parasitic modes may exist, depending on the construction of the amplifier slab.

We assumed a rectangular plate with dimensions l , w , and h , with $l \geq w > h$. Both n_2 , the index of refraction of the glass at the oscillation wavelength, and α , the gain coefficient, were assumed to be independent of position. We assumed that the slab was surrounded (around its edges) with a material that had an index of refraction equal to n_3 and that absorbed all of the entering radiation. Finally, the edge-clad slab was assumed to be completely enclosed by a medium with an index of refraction n_1 , with

the requirement that n_1 be less than both n_2 and n_3 .

Since we were interested in the lowest-threshold oscillation, we required that the reflection from the faces of the slab take place at the critical angle $\theta_c = \sin^{-1}(n_1/n_2)$. Each parasitic mode may then be identified by the integer pair (M, N) , where M is the number of bounces off each long edge (length l), and N is the number of bounces off each short edge (length w) required for one, round trip of the ray. It then follows that the angle of incidence on the short edge, viewed in the face plane, is

$$\Psi = \tan^{-1} \left(\frac{Mw}{Nl} \right) \quad (2)$$

By choosing M and N appropriately, we can come arbitrarily close to any possible value for the angle Ψ .

In a lossless parasitic mode, the rays must intersect both the long and short edges at an angle of incidence greater than the critical angle for total internal reflection, $\phi_c = \sin^{-1}(n_1/n_2)$. This can only occur if the indexes of refraction satisfy the relation

$$\left(\frac{n_3}{n_1} \right)^2 \leq \left(\frac{n_2}{n_1} \right)^2 - \frac{1}{2} \quad (3)$$

For values of n_3 that are less than n_2 , but do not satisfy Eq. (3), no lossless parasitic modes exist in rectangular amplifier slabs. The onset of lossless parasitic oscillation occurs suddenly and fills the entire slab as soon as n_3 is small enough to satisfy Eq. (3) (e.g., for $n_2/n_1 = 1.52$, when $n_3/n_1 = 1.35$). The behavior is in contrast with that in cylindrical-disk amplifiers,¹³ where lossless, parasitic rim modes exist along all values of

$n_3 < n_2$ and fill an increasing portion of the disk as the index of the edge cladding decreases, occupying the entire volume only when $(n_3/n_1)^2 \leq (n_2/n_1)^2 - 1$.

For lossy modes, where the angle of incidence is less than the critical angle along at least one of the edges, the threshold required for parasitic oscillation was calculated by setting the round-trip gain equal to the total loss incurred by incomplete reflection at the edges. Lower bounds on the gain-coefficient thresholds were calculated numerically according to the above prescription for modes with N and M ranging from 1 to 20, along with several modes that required an infinite number of bounces off one or the other pair of edges. For all cases considered, the mode with the lowest threshold was found to be dependent only on whether n_3 was greater or less than n_2 .

When the index of refraction of the cladding is greater than that of the glass, there is no possibility of total internal reflection at any of the edges. The mode with the lowest threshold was found to be the one that oscillated directly between the two short edges. With $n_3 < n_2$, the mode with the lowest loss occurred when Ψ was adjusted so that the ray made use of the maximum possible path length in the glass at which the internal reflection from the long edge was total. In both cases, any attempt to lower the thresholds by increasing the path length in the glass was thwarted by increased loss at the edges.

Figure 6-12 shows the results of our analysis applied to a rectangular slab of LG-750 phosphate glass, where $n_1 = 1$ and $n_2 = 1.52$. LG-750 is the highest-gain material proposed for use in the SSA. The lower limit on the threshold for parasitic oscillation is plotted as a function of the index mismatch between the amplifier glass and edge cladding. Also shown is the value of $\alpha l = 3.3$, which has been estimated as the largest value for the disks in the SSA.

Figure 6-12 implies that damping of the parasitic oscillation in the SSA will require that the value of $n_3 - n_2$ lie in the range of -0.07 to $+0.12$. A silicone based edge cladding, with an index differing from that of LG-750 by <0.01 , has been demonstrated. Thus, it appears that parasitic oscillation will not pose a problem for the operation of the SSA.

Author: K. S. Jancaitis

Fig. 6-12. Lower bound on parasitic oscillation threshold vs index mismatch of edge cladding and laser glass.

Compact Amplifiers

The next generation of fusion lasers will require much larger and more efficient amplifiers at lower cost than current single-disk or segment designs. An ideal amplifier consists primarily of laser glass and flashlamps and avoids, to the extent possible, exposure of metal surfaces and self-absorption of flashlamp light. To reduce cost, individual components must be standardized.¹³

The design of a compact amplifier consisting of arrays of identical, rectangular segments interleaved with rows of flashlamps, is shown in Fig. 6-13. Such a compact amplifier can provide very large beam apertures with improved efficiency at reduced cost. Although the design was developed for multipass options of Nova,¹⁴ the considerations of multimegajoule lasers have renewed our interest in this concept.

The main design features of the compact amplifier are:

- Reduced length of both the amplifier and the laser system because of beam-reimaging requirements.
- Low vignetting losses for off-axis, multipass-system applications.
- Reduced lamp-reflector areas, improved pump efficiencies, and reduced cost because of smaller volume, higher efficiency, and the use of standardized components.

Concerns with respect to access for assembly and maintenance have been addressed and can be successfully resolved. The compact-amplifier design in combination with long-lifetime, low-cost silicate laser glass can reduce the projected cost of energy storage to $< \$4/\text{J}$ for hardware, energy storage, and optics.

Vignetting Losses. For off-axis multipass systems, the effective clear aperture of an amplifier is reduced by vignetting because the laser beams traverse the amplifier at small angles relative to their optical axis. The vignetting loss is directly proportional to the length of the amplifier, which, in turn, is inversely proportional to the number of transverse segments.

For square segments nominally 50 cm on each side, about 10 transverse segments are anticipated for a megajoule laser, leading to an order-of-magnitude reduction of the vignetting losses for the compact design. The vignetting losses for standard amplifier designs would be so large that off-axis multipass systems would not be feasible. However, with compact amplifiers, vignetting losses can be reduced to $< 10\%$ (Ref. 15); thus, the compact design can provide a viable option for off-axis multipass systems.

A harmonically switched multipass system still appears to be the best option for a megajoule laser because the harmonic

Fig. 6-13. A compact amplifier consists of identical, rectangular segments interleaved with rows of flashlamps to provide large beam apertures and high pump efficiency.

Solid-State Amplifiers

converter can also be used as a switch at almost no extra cost. It is unlikely that a regenerative system can be made competitive by requiring a separate or high-voltage (>100-kV) second-harmonic-generation (SHG) switch, an input beam injector, an exploding foil to hold off the gain, and a prepulse discriminator to eliminate a 7% vignetting loss. The additional cost and potentially higher losses and nonlinear effects will probably exceed this maximum potential improvement of ~7%, especially if the cost of developing the new components is taken into account.

Pump Efficiency. A rough estimate of the cavity losses for the standard and compact designs can be obtained by considering the relative surface areas exposed to pump light and their average absorptivities. The standard amplifier has about 2 to 3× as much reflector surface exposed to pump light as the compact design, with correspondingly higher cavity losses. The surface area of a full-involute reflector is

$$A_I = \frac{\pi^2}{2} dL \quad (4)$$

where d is the lamp diameter and L is the arc length. In contrast, the surface area for two half-involute reflectors is

$$A_H = \frac{\pi^2}{4} dL = \frac{A_I}{2} \quad (5)$$

or one-half that of the full-involute reflector. The number of outside lamps for full-involute reflectors can be approximated by

$$N_{IC} \approx 2 \frac{nD}{\pi d} N_r \quad (6)$$

and

$$N_{IS} \approx 2 \frac{nD}{\pi d} N_r N_T = N_{IC} N_T \quad (7)$$

for compact and standard amplifiers, respectively, where n is the index of refraction, D the clear aperture of the segments in the plane of polarization, N_r the number of disks intercepted along the beam axis, and N_T the number of transverse segments. The

number of inside lamps for half-involute reflectors is approximately

$$N_{HC} \approx 2 \frac{nD}{\pi d} N_r (N_T - 1) \quad (8)$$

The surface areas of the top and bottom reflectors are

$$A_{sC} \approx 2nD^2 N_r N_T \quad (9)$$

and

$$A_{sS} \approx 2nD^2 N_r N_T^2 = A_{sC} N_T \quad (10)$$

which add up to total reflector surface areas of

$$\begin{aligned} A_C &= A_I N_{IC} + A_H N_{HC} + A_{sC} \\ &\approx \left[\frac{\pi L}{2} \left(1 + \frac{1}{N_T} \right) + 2D \right] nD N_r N_T \end{aligned} \quad (11)$$

and

$$\begin{aligned} A_S &= A_I N_{IS} + A_{sS} \\ &\approx \pi L + 2D N_T + nD N_r N_T \end{aligned} \quad (12)$$

for the compact and standard amplifier, respectively. The ratio of these areas for square-output beams with $D N_T \approx L$ is

$$\frac{A_S}{A_C} \approx \frac{2N_T + \pi + 2}{\pi(1 + N_T) + 4} \approx 2.67 \quad (13)$$

for $N_T = 10$ transverse segments.

These reflector surfaces compete for pump light with the segments, which have a total surface area of

$$A \approx 2\sqrt{1 + n^2} D L N_r N_T \quad (14)$$

The relative surface areas for nominally square output beams are then

$$\frac{A_C}{A} \approx \frac{(1 + N_T) \pi / 4 + 1}{N_T \sqrt{1 + 1/n^2}} \approx 0.80 \quad (15)$$

and

$$\frac{A_S}{A} \approx \frac{\pi/2 + 1}{\sqrt{1 + 1/n^2}} \approx 2.14 \quad (16)$$

for $N_1 = 10$.

A first-order scaling of the cavity efficiency can be obtained from the probability of absorbing pump photons proportional to the relative surface areas and their effective absorptivities in the form

$$\eta \approx \frac{\eta_A A}{\eta_A A + (1 - R)A_R + (1 - \eta_A)A_1 + \eta_1 A_1 + \dots} \quad (17)$$

where $\eta_A \approx 0.4$ is the absorption efficiency of the laser glass, $R \approx 0.9$ is the effective reflectivity of the metal surfaces, $\eta_1 \approx 0.8$ is the flashlight-conversion efficiency in the lamp, and

$$A_1 \approx 2\pi D L N_1 N_2 \approx \frac{A}{\sqrt{1 + 1/n^2}} \quad (18)$$

is the flashlamp surface area for involute reflectors. This leads to relative improvements in pump efficiency in the range of

$$\frac{\eta_C}{\eta_S} \approx \frac{1.25 + 2.14/4}{1.25 + 0.8/4} = \frac{0.69}{0.56} = 1.23 \quad (19)$$

in favor of the compact amplifier. The potential improvement obviously depends on the effective reflectivity of all metal surfaces and on the detailed design of the amplifier.

Energy Storage. Involute reflectors limit the number of flashlamps in both designs and, hence, the obtainable stored energy. Somewhat higher gains could be obtained by increasing the number of lamps, but direct-flashlamp absorption would significantly reduce the cavity efficiency. Because of the relatively high cost of pulsed power, adding more lamps would probably decrease the cost-effectiveness.

Access. We examined several options with respect to providing access to individual segments and flashlamps without sacrificing efficiency. Access could be provided by mounting the segments and flashlamp modules in movable sections, as illustrated in Fig. 6-13. Here, the weight of the movable sections is suspended from a main

support beam. Rollers or linear bearings with additional guidance or partial lower supports can provide accurate motion. Such a design reduces the required stiffness and, cross-sectional area for the individual segment and flashlamp supports.

A sketch of a dividable segment support is shown in Fig. 6-14. This support could be separated by releasing a wedged lock slide or some other locking mechanism. The flashlamp modules would only be mounted on one side of the movable sections; access to any desired row of lamps and segments would be obtained by releasing the locks in the appropriate row of segment supports and moving the amplifier sections apart. The flashlamp modules would then be directly accessible, and the segments could be inspected.

Segment Supports. The segments are stacked on top of each other to minimize the surface area of support structures exposed to pump light. Stacking the segments does not cause any problems so long as the segments are supported reasonably uniformly. Silicone edge claddings would be ideal, while solid edge claddings would require a thin silicone or other elastic sheet to eliminate point loadings. For a 2.5-m-high amplifier, the average loading for the lowest segments is ~ 10 psi, which leads to a negligible stress optic distortion of 0.003 waves at $1.06 \mu\text{m}$.

Retainer strips would hold the segments in place and would have to be removed to exchange a segment. To remove a segment from a particular stack, one would attach a handling fixture to H-shaped edge-cladding shields and lift the segment from the stack.

The cross-sectional area of each segment support is $\sim 20 \text{ cm}^2$, which would limit the elongation (due to the total weight of segments and structure) to $< 10 \mu\text{m}$ for stainless steel. The moment of inertia in the narrow direction of the stainless-steel supports is about 10 cm^4 , which would lead to deflections of 0.2 mm/kg of transverse force. This deflection can be reduced by mounting tension bands between segments or by interlocking the H-shaped edge-cladding shields with the supports to form a more rigid structure. While cross-webbing of the individual support beams is possible in two directions for the compact amplifier, it is only available in one direction for the

Fig. 6-14. Interlocked segment supports allow separation of sections while minimizing exposure to flashlamp light.

standard segmented amplifier. Providing equal rigidity for both designs will require larger support beams for the standard amplifier, increasing the exposed-metal surface area of the internal structure relative to the segment area from ~5 to ~20% for the compact and standard amplifiers, respectively.

Beam Obstructions. Because of the internal flashlamps, alternating 3- and 7-cm-wide gaps will occur in the overall beam. Within the amplifier, these gaps are irrelevant because they do not affect the pump efficiency. However, the output optics, such as turning mirrors and focusing optics, would have to be larger to accommodate the full beam aperture. These components

would probably be segmented as well, and it may be a blessing instead of a liability to have some extra space between components for easier mounting.

Cost. To avoid beam aberrations and eliminate the observed 7% loss in pump efficiency due to the shield glass, the amplifiers would have to be evacuated. In this case, the volume of the amplifier becomes quite important. The cost of an evacuated amplifier can be approximated by

$$C_{ME} = \$1k N_s + \$0.3k N_L + \frac{\$70k}{m^{1.8}} V^{0.6}, \quad (20)$$

where N_s is the number of individual segments, N_l the number of lamps, and V the active volume of the amplifier determined by the beam cross-sectional area and the active length of the amplifier.

With such cost scaling, some rough comparisons of the cost for compact and standard amplifier designs can be made. To provide a gain of 7 for a near-optimal 5-pass system with a long-lifetime glass requires about 16 segments in series. With 10 transverse segments stacked 5 high, one amplifier would contain about 800 segments. For square segments nominally 50 cm on each side, about 1600 flashlamps with a bore diameter of 2 cm and an arc length of 2.5 m would be required; these lamps would be pumped at ~ 50 kJ each. A 5- by 5-array amplifier would be about 7 m long for the compact design, while the standard design is about $10\times$ longer, or 70 m, with corresponding active volumes of 45 and 450 m^3 , respectively. The mechanical cost, in \$ million, for these two designs is then $C_{\text{ME,C}} \sim 0.8 + 0.5 + 0.7 = \2 million for the compact design and $C_{\text{ME,S}} \sim 0.8 + 0.5 + 2.7 = \4 million for the standard design, without counting the cost for laboratory space. Obviously, there are costs that will reduce this doubling in cost for the standard amplifier, but a vacuum system that is $10\times$ larger is simply more expensive. Even without evacuation, the compact amplifier is probably less costly than the standard design. The extra cost of about \$0.6 million for the vacuum system of the compact amplifier is compensated for by eliminating beam aberrations and losses in pump efficiency due to the shield glass.

With a glass cost of about $\$0.5/\text{cm}^3$, which includes solid edge claddings for silicates, $\$0.15/\text{cm}^2$ for polishing, an electrical cost of $\$0.08/\text{J}$ and $\$500/\text{lamp}$, the total cost, in \$ millions, of the compact amplifier with 800 segments and 1600 lamps amounts to about

- \$ 2.0 for mechanical hardware
- 3.0 for 6 m^3 of laser glass (3-cm-thick segments)
- 0.6 for 400 m^3 of polishing
- 6.4 for 80 MJ of energy storage
- 0.8 for 1600 flashlamps
- \$12.8 million total.

With this design and long-lifetime laser glass, it should be possible to obtain an energy-storage efficiency of $\sim 4\%$ or 3.2 MJ.

This translates into an energy-storage cost of $\$4/\text{J}$, which is not yet an optimized case because the cost of bank energy is disproportionately high at 50%, while a fully optimized system balances the cost of ME, EE, and optics more equally.

It is clear from our estimates that we do not want more lamps in the amplifier than the involute reflectors provide. The optimization trend will definitely be toward more segments pumped at lower energies but at higher efficiency. Such a reoptimization may even bring the energy storage cost down to $\$3/\text{J}$, which is about an order of magnitude lower than for the relatively small Nova amplifiers. The improvement breaks down into a:

- Twofold increase in energy storage for long-lifetime glasses.
- Cost reduction by $2.5\times$, primarily due to silicate glass costs that are $6\times$ lower.
- Twofold improvement in efficiency and lower cost of large but compact amplifiers with rectangular segments.

System Considerations. Cooling an evacuated amplifier is not a trivial consideration because large pumps are required to remove the cooling gas. The compact amplifier provides better flow channels than the open-box structure, and it may be possible to restrict the volume of gas leakage to the rest of the system by low-pressure cooling flows. In addition, a large fraction of the heat comes from the lamps, and this heat could be carried away with water-cooled lamps. Such lamps were found to be remarkably reliable and relatively inexpensive. Liquids are not only very effective coolants but they also reduce the reflection losses between lamp and sleeve, which is even more important. A fluorescence converter added to the coolant or sleeve can provide some extra pump light from the unused UV emission of the lamps. The outside surface of the sleeve may have to be rippled to allow the fluorescence from the coolant or sleeve to escape.

Another consideration is the size and shape of the segments. There is no reason to use nominally 2:1 rectangular segments to obtain individual, square beamlets. From considerations about ASE, a square segment would be much superior. However, the fill factor for rectangular beams is somewhat lower than for square beams, and it is more difficult to match the shape of the

second-harmonic crystal to the individual beams. From these considerations, an aspect ratio of 1.3:1 appears to be more reasonable. For maximum efficiency, the sides of the segments should be cut at the Brewster angle to reduce the dead volume in the amplifier, as indicated in Fig. 6-14.

The overall height of the compact amplifier is determined by the arc length of the lamps; length is determined by the segment size and the number of segments along the beam required for a desired gain; width is determined only by mechanical constraints of supporting and servicing the amplifier. In principle, it would be possible to build only two amplifiers for a megajoule system and to use individual, wedged lenses to focus the beam on the target, thereby eliminating all turning mirrors and support frames. Fine alignment could be accomplished with the individual input-beam mirrors. A single 3-m-high by 10-m-wide amplifier could deliver 7 MJ at a fluence of $30 \text{ J}/\text{cm}^2$ and a fill factor of 0.8, which is sufficient to obtain 10 MJ in the green from two beams. For an evacuated system, it is better to build a few, large vessels instead of many small ones.

Such a system would require about 50 m^3 of laser glass in the form of 10 000 identical segments with a total surface area of about 4000 m^2 , pumped by 12 000 flashlamps, and delivering 700 MJ of bank energy to store 30 MJ at a cost of about \$110 million. An energy-storage cost of $< \$4/\text{J}$ for the components of large but compact amplifier designs appears to be feasible.

Author: W. F. Hagen

Flashlamps

Large-Bore Flashlamps

Introduction. We shall first briefly review our understanding of how flashlamps operate. The flashlamp plasma is in local thermodynamic equilibrium at $\sim 1 \text{ eV}$ under normal operating conditions. The photon field is not in equilibrium with the particle field, and, consequently, the flashlamp radiates as a gray body. The emissivity varies with wavelength from small values in the UV to near unity at $1 \mu\text{m}$. As energy into the flashlamp increases, the emissivity rises

at short wavelengths, shifting the spectrum toward the blue. Another consequence of the gray-body nature of the plasma is that the spectrum shifts toward the blue as the viewing angle increases from a direction normal to the lamp axis to directions more collinear with the lamp.

Last year¹⁶ we determined the radiant efficiency and relative Nd-pumping efficiency for clear-fused-quartz, xenon-filled flashlamps as a function of bore diameter (1, 1.5, and 2.7 cm), explosion fraction, and pulse width. We monitored the input energy electrically and the radiant-output energy calorimetrically. We used both a total-enclosing calorimeter and a flashlamp-envelope temperature monitor (a measure of energy loss) to establish lower and upper limits for the radiant output. All other optical diagnostics viewed the flashlamp via a barium sulfate (packed powder) scattering plate, which averaged the tubular flashlamp light over viewing angle. Among our diagnostics were a pyroelectric detector and a Nd-fluorescence detector. The pyroelectric detector had a spectrally flat response and was able to track the output power of the lamps. The Nd-fluorescence detector consisted of a small slab of ED-2 glass and a photodiode carefully configured so that the photodiode viewed only the Nd fluorescence near $1.06 \mu\text{m}$.

We determined experimentally that both the radiant efficiency and the Nd-pumping efficiency, for a fixed pulse width and pulse shape, are independent of the bore diameter when these data are plotted as a function of explosion fraction (input energy/explosion energy). The radiant efficiency increases with increasing explosion fraction up to an explosion fraction of 0.1 and remains constant, at a great value of $75 \pm 5\%$, for higher explosion fractions. The Nd-pumping efficiency also increases with increasing explosion fraction, peaking at an explosion fraction of 0.15, and decreasing at higher input energies. We attribute the fall-off in the pumping efficiency at very low inputs to the decline in the radiant efficiency and the falloff at high input to the blue shift of the spectrum.

Experiments. During 1983, we tested both sealed and refillable lamps obtained from commercial sources and our own lamps consisting of quartz tubing, sintered tungsten electrodes, and O-ring seals. The

End view

Side view

Fig. 6-15. Setup to measure the emission characteristics of flashlamps of various bore diameters.

Fig. 6-16. Radiant efficiency as a function of explosion fraction for 1.5-, and 4.2-cm-diam flashlamps (600- μ s pulse length).

1.5-cm-bore data in Fig. 6-16. These data were determined from the wall-temperature rise and from the total-enclosing calorimeter. The error bars represent statistical errors, which we determined from averaging many shots on various days. As found previously, the overall efficiency curves are quite similar when plotted as a function of explosion fraction.

Figure 6-17 shows our results for Nd-pumping efficiency (peak fluorescence/input energy) vs explosion fraction for the 4.2-cm-bore tube together with the results from the other flashlamp sizes previously tested. Although the scale of the vertical axis is arbitrary, the relative normalization of the data points is absolute. The curves are remarkably similar, suggesting that there is little reason, on performance grounds alone, to choose one lamp diameter over another for use in Nd:glass amplifiers.

performance of the three types of lamps was indistinguishable.

Specifically, we tested lamps with bores of 1, 1.5, 2.7, and 4.2 cm filled with 450, 300, 100, and 75 Torr of xenon, respectively.^{1b} As previous, either one or two series lamps were connected to several parallel circuits via an ignitron. In addition, a series trigger pulse to break down the gas was injected through the normal LC circuit from a 0.66- μ F capacitor 200 μ s before the main ignitron closed. This technique reduced the scatter of data at low explosion fractions and permitted us to obtain data at even lower input energies. Each lamp was housed in a rugged brass structure with symmetric ground returns, as shown in Fig. 6-15. A fine helical wire, which was wound around the lamp and connected to the current return, helped initiation of the discharge. The lamps and bank were interconnected with RG-17 cable, which has ample current capability. The symmetrical current-return geometry minimized magnetic forces, which can be significant at high currents.

The pyroelectric Nd-fluorescence detectors and the spectrometer were all apertured to view only light scattered from the barium sulfate (packed powder) scattering plate. All data were fed into an LSI-11/23 computer. A 1.5-cm-bore reference flashlamp was periodically inserted into the experimental geometry to ensure that all data were absolutely comparable. As before, the baseline circuit for the 1.5-cm-bore lamp contained 87 μ F, 450 μ H, and had a 0.3- Ω resistive loss.

The calorimetrically measured radiant efficiency for the 4.2-cm-bore lamp with 75 Torr xenon is compared with the

Fig. 6-17. Nd-pumping efficiency vs explosion fraction for flashlamps of four different bore diameters. A representative error bar is shown.

Fig. 6-18. The presumed opacity factor, j^2d , at maximum current vs explosion fraction for 1.0-, 1.5-, 2.7- and 4.2-cm-diam lamps.

Although our Nd-fluorescence detector effectively integrates the spectral/temporal output of the flashlamps, the similarity of the curves in Fig. 6-17 suggests that their detailed optical performance may be similar at a given explosion fraction. This behavior is not unexpected. Previous experiments¹⁷ found that the absorption at a given wavelength scales as current density squared (j^2). Hence, the emissivity should be a function of j^2d , where d is the bore diameter. If we plot j^2d (at peak current) as a function of explosion fraction, as shown in Fig. 6-18, we again find that the curves from the different flashlamps are very close together. The agreement helps to explain the similarity in optical performance.

We also studied the performance of 1.5- and 4.2-cm-bore flashlamps as a function of xenon fill pressure. The input was fixed at

an explosion fraction of 0.19. Holzrichter¹⁸ previously found a rather broad maximum in the Nd-pumping efficiency with pressure as well as a greatly reduced lamp lifetime at higher fill pressures. We obtained information on radiant efficiency from the time integral of the pyroelectric-detector signal as normalized by the calorimetric data. Figure 6-19(a) compares the radiant efficiency as a function of pressure for the 1.5- and 4.2-cm lamps, while Fig. 6-19(b) compares the Nd-pumping efficiencies. Notice that the performance of the smaller lamp declines much more dramatically at low pressure than that of the larger-bore lamp. It is likely that the decline is caused by increased wall losses. For pressures near our standard values (75 and 300 Torr for the 4.2- and 1.5-cm lamps, respectively) the performance is quite insensitive to fill pressure.

Spectrally Resolved Measurements. We chose to use a Tektronix J-20 to obtain time-resolved flashlamp spectra based primarily on the breadth of its spectral response—nominally 0.3 to 1.1 μm , and on the convenience of interfacing it to our LSI-11/23 computer system via a digital processing oscilloscope. The J-20 is, in essence, a 512-element detector, which is attached to the output plane of an 0.25-m grating spectrograph. The breadth of its spectral response, compared to that of conventional photocathode, results from the direct use of a silicon Vidicon for light detection. Unfortunately, however, the J-20 has no fast-gating capability to obtain time-solved spectra. In our case, the detector integrates light for 50 ms. This interval is divided into a 30-ms quiescent period and a 20-ms period during which the acquired change is individually read out from each element. The cycle occurs continuously. Triggering consists simply of choosing to store the output of a given cycle.

To obtain adequate time resolution for our flashlamp studies, we interfaced the J-20 to a fast, mechanical shutter provided by a chopper wheel rotating at 8000 rpm. A 1-mm-wide slot in the wheel produced a gating interval with a full width at half maximum (FWHM) of 25 μs and a 10%-width of approximately 60 μs . This opening was complemented by an identical slot on the opposite side of the wheel,

Fig. 6-19. (a) Radiant efficiency and (b) Nd-pumping efficiency vs. xenon fill pressure for 1.5- and 4.2-cm-bore flashlamps at an explosion fraction of 0.19.

which provided a timing signal by shuttering a He-Ne laser beam. The timing of the discharge to the wheel rotation was complicated by the fact that the rotating wheel and the J-20 operated independently of one another. Timing was made possible because several wheel openings, spaced by 8 ms, occurred within the 30-ms interval over which the J-20 was quiescent. The normal fire signal was first routed to a timing circuit, which provided an output at the beginning of the next J-20 read-scan. After a specified time delay (~ 35 ms), the first available wheel opening was then sensed to allow firing of the bank at any desired position with respect to the next wheel opening. The net result of the timing circuit was that the shutter opening could be placed anywhere within the current pulse with a precision better than 10 μ s.

A major problem with a silicon Vidicon detector for quantitative spectral measurements is that its output is not linear with input when used with pulsed-light sources. The problem results from the phenomenon known as lag.¹⁹ However, we have found it possible to correct for this nonlinearity and to obtain a linear response over a dynamic range of ~ 100 . For calibration, we used a cw tungsten light together with a moderately fast (1- to 2-ms) electrical shutter. Our understanding of the lag phenomenon, as supported by tests at several pulse lengths, suggested that the response function should be the same at all pulse lengths provided the light pulse is short compared to the 50-ms integration time.

We constructed signal-vs-intensity curves using neutral-density filters having accurately known transmissions at specific wavelengths. We fitted the results to the formula

$$\ln(I) = A[\ln(S)] + B[\ln(S)]^2 + C \quad (21)$$

where I and S are the relative intensity and measured signal levels, respectively, and A , B , and C are fitting parameters. We obtained good fits to Eq. (21) with relatively good agreement for the fitting parameters A and B for data taken at several points within the spectral range of from 400 to 1000 nm. Outside this range, the sensitivity of the J-20 was quite low and its response did not fit this function. This limited our use of the J-20 to the 400- to 1000-nm range.

After applying the correction formula to convert from signals to relative intensities, we generated a response curve for the J-20 as a function of wavelength using the shuttered tungsten source as our standard. We obtained calibration spectra on three separate shots to isolate the blue (300- to 700-nm grating range), yellow (500- to 900-nm), and red (700- to 1100-nm) spectral regions using long-pass color filters to prevent order overlap. Within each range, there were 512 spectral points. Spectral correction factors were then computed by dividing the expected intensity by the measured intensity at each element.

We obtained actual flashlamp spectra in a similar manner using three separate shots and the same bandpass filters. The spectra

Fig. 6-20. Output spectral distribution at peak current for (a) 1.5-cm-bore and (b) 4.2-cm-bore flashlamps.

were then compacted to a single 1000-point array with splice points at 550 and 800 nm. Generally, the splice produced a smooth union; however, in some cases there was a noticeable jog in the spectra presumably caused by shot-to-shot variations. Thus, the J-20 provided detailed spectra over the range of from 400 to 1000 nm with a resolution of ~ 1 nm. This is narrower than the width of the spectral lines emitted by the flashlamp plasma.

To use our calorimetric data on lamp efficiency to calibrate the spectra absolutely, we needed to determine the fraction of the radiated power that was outside the spectral bandwidth of the J-20 i.e., at wavelengths shorter than 400 nm and longer than 1000 nm. This information was readily obtained by measuring signals in three ways from a pyroelectric detector (which has a spectrally flat response): first unfiltered, then with a 400-nm long-pass filter, and finally with a 1000-nm long-pass filter. By knowing the average transmissions of the filters, we were then able to determine the fraction of instantaneous radiated power that was in each of the spectral bands.

Examples of spectra obtained in this manner for a 1.5-cm-bore flashlamp are shown in Fig. 6-20(a). These spectra were taken at the peak of the current pulse for two different explosion fractions, f_x . Notice that the units of the y axis are fraction/ μm . The pyroelectric-detector measurements were used to calibrate these data. To determine the absolute power radiated in a given spectral region, we must find the area under the curve for that bandwidth and multiply by the instantaneous radiation efficiency and the electrical input power. By normalizing the spectral distribution in this manner, the shift of the radiation towards the blue

with increasing input becomes apparent, as shown in Fig. 6-20(b). These spectra also show, as expected, that there is substantial near-infrared line structure (regions of high emissivity corresponding to well-known neutral xenon lines) at low input. The structure washes out at higher input as the output more closely approaches that of a blackbody. In some cases, the xenon lines become self-reversed, suggesting the presence of a colder absorbing region surrounding the hotter central plasma.

Spectra taken for the 4.2-cm-bore lamp are shown in Fig. 6-20(b). Notice that the pairs of spectra in Figures 6-20(a) and (b), which correspond to constant explosion fractions, are quite similar. The similarity is expected, as previously explained. However, the larger-bore flashlamps consistently show more line structure at a given explosion fraction. This effect may be related to the lower fill pressure used in the larger-bore lamps, which may decrease the lamps' emissivity.

The correspondence between the spectra of these two different lamps vs explosion fraction is further illustrated in Fig. 6-21. The fractions of the radiant power at peak current in the blue, yellow, and red spectral regions, as determined by the pyroelectric detector, are plotted vs explosion fraction. The agreement between the two lamps is extremely good.

Using the absolutely calibrated spectra, we calculated the instantaneous pumping efficiency for Nd:glass. The pumping efficiency is defined as the ratio of extractable laser power at $1 \mu\text{m}$ to the electrical power deposited in the lamp. We obtained efficiency values by convolving the measured flashlamp spectra with the absorption spectrum for ED-2 (as calculated from the cross

Fig. 6-21. Radiant-output fraction in three spectral regions vs explosion fraction.

Fig. 6-22. Calculated Nd:glass pumping efficiency vs explosion fraction based on quantitative experimental data. The assumed doping-thickness product was 1.1×10^{21} Nd ions/cm².

sections in Ref. 20). We presumed that the pump photons passed through the glass once and that there was a unit quantum yield for producing the ${}^4F_{5/2}$ upper laser state for absorption at any wavelength.

Computed results for Nd-pumping efficiency at the peak of the current pulse are shown in Fig. 6-22 for both the 1.5- and 4.2-cm-bore lamps. Note that the values plotted are the instantaneous pumping efficiency even though the λ axis is explosion fraction, a time-integrated quantity. The general shape of the plot is similar to Fig. 6-17. However, note that the pumping efficiency in Fig. 6-22 is both absolute and instantaneous while that in Fig. 6-17 is relative and time-integrated by the response of the Nd fluorescence. The 1.5- and 4.2-cm-bore results are again quite similar when plotted as a function of explosion fraction despite the slight differences in their detailed spectra for these conditions.

Finally, it is important to recognize that the absolute values for the pumping efficiency, as shown in Fig. 6-22, are quite large but are in good agreement with previous estimates.²¹ After taking into account the loss of excited states during pumping, we conclude that a 5% peak-storage efficiency should be possible in a Nd:glass amplifier. Our planned experiments with the single-segment

amplifier, described in "Blackbody Pumping and Depumping of Nd:Glass," are directed at testing that prediction.

Authors: L. P. Bradley and H. T. Powell

Major Contributors: K. S. Jancaitis, D. A. Masquelier, and C. E. Thompson

Transient Response of Flashlamps

The Trenholme-Emmett model²² uses a steady-state approximation to describe the emission properties of long-pulse xenon flashlamps. The model assumes that the radiative and electrical characteristics of a flashlamp are simply a function of the instantaneous input power or, equivalently, of

Flashlamps

the instantaneous current density. Subsequent work,²³ however, has shown differences in the emission spectrum at equal current points on the rise and fall. Thus, questions have arisen regarding the validity of the steady-state approximation.

We have used two methods to study the radiative response time of xenon flashlamps when excited at conditions suitable for pumping Nd:glass lasers. In the first method, we introduced a short-circuiting (crowbar) switch to interrupt the arc and then measured the instantaneous decay time of radiation, τ_r , using a detector with a flat spectral response. We then compared the results for τ_r to two different models. We later observed that the radiative response time was evident as a time lag between the input and radiated power for lamps that were not short-circuited; we were thereby provided a second method for determining the response time. Finally, we found that differences in the electrical characteristics on the rise and fall of the current can also be described by a simple time lag. This seemed reasonable since both the electrical and radiative characteristics should be a function of the internal energy or, equivalently, of the temperature of the plasma. For pulses that are long compared to the response time, the internal energy simply tracks the input power with a time lag.

Crowbar Measurements. Our experimental apparatus for measuring radiation

decay with a crowbar-switched lamp is shown in Fig. 6-23. We used a clear-fused-quartz flashlamp with a 1.9-cm bore, a 49-cm arc length, and discharge-circuit parameters of $C = 750 \mu\text{F}$ and $L = 56 \mu\text{H}$. A 20-kV pulse from a 20-nF capacitor (not shown) triggered the lamp coincident with closure of switch S_1 . After a specified time delay, a second switch, S_2 , in parallel with the flashlamp, was closed. Both switches were size A ignitrons. Thus, the short-circuited the arc by interrupting the current through and, hence, power deposition in the flashlamp. To minimize circuit inductance associated with diverting the current through S_2 , the diverted current path was run parallel to the lamp a short distance away, and S_2 was mounted directly at one end of the lamp.

To measure the change in the total radiated power, we used a pyroelectric detector with a flat spectral response and a rise time of $1 \mu\text{s}$. Radiation from the lamp was also angle-averaged using a BaSO_4 scatter plate.²⁴ We determined the instantaneous decay constant of the radiation at the time of crowbarring, τ_r , from oscillograms. The value of τ_r (from 20 to $50 \mu\text{s}$) was always long compared with the instantaneous current decay time of the crowbar discharge ($\sim 2 \mu\text{s}$). Thus, we assumed that the energy deposition was instantaneously interrupted. Nonetheless, there was approximately a $\pm 10\%$ systematic uncertainty in the determination of the decay time associated with the finite fall time of the current.

We measured radiative-decay times for a large range of gas pressures, charging voltages, and time delays before closure of the crowbar switch. A typical set of results is shown in Fig. 6-24. Here, τ_r is plotted as a function of $V \cdot I$, the instantaneous input power (per unit discharge length) at the crowbar time for an initial xenon fill pressure of 350 Torr and three different energy loadings. These data were obtained by crowbarbing the discharge at successively longer times after breakdown. The arrow shows the direction of increasing time, with the data points for $V \cdot I$ values separated by $\sim 50 \mu\text{s}$.

The decay constant is obviously not a unique function of the instantaneous $V \cdot I$ throughout the discharge pulse; however, the curves of τ_r vs $V \cdot I$ appear eventually to

Fig. 6-23. Apparatus used to investigate the radiation decay of a crowbar-switched xenon flashlamp.

join a common band of decay times that is, more or less, a unique function of $V \cdot I$. We refer to this band as the main sequence. The time after breakdown required to reach this band decreased as the input energy was increased. The approximate value of this time delay is noted in Fig. 6-24. The time at which the τ_r vs $V \cdot I$ curve joins the main sequence may be a measure of the time at which the arc-expansion phase is complete and the arc has reached its steady state conditions. Such a conjecture is consistent with observations and calculations that the bore-filling time decreases with increasing input energy to the flashlamp.^{25, 26}

Values of τ_r vs $V \cdot I$ for a charging voltage of 4 kV and a range of initial xenon fill pressures (50 to 600 Torr) are plotted in Fig. 6-25. Only those data points that we have qualitatively assigned to the main sequence are shown. Note that the radiation decay times in the main-sequence region are quite insensitive to the xenon fill pressure. The values of τ_r may be slightly lower for lower initial fill pressures, but the functional dependence is weak and somewhat obscured by the scatter in the data.

We also briefly investigated the radiation decay times for argon and krypton gas in

the same lamp. Data for these two gases were taken only at a 300-Torr fill pressure and at a charge voltage of 4 kV. Again, only decay times judged to be on the main sequence are plotted in Fig. 6-25. The

Fig. 6-24. Radiation decay time of a crowbar-switched flashlamp (350 Torr xenon) vs instantaneous power deposition. Approximate time at which τ_r joins the main sequence is shown.

Fig. 6-25. Radiation decay times vs input power for xenon at various pressures and for argon and krypton.

observed decay times for argon and krypton were similar to those for xenon; however, in general, there appears to be a trend toward somewhat shorter decay times for the heavier rare gases.

We have used a simple theoretical treatment to model these experiments. The experiments measure the instantaneous decay rate, τ_r , defined by

$$\frac{dP_r}{dt} = -\frac{1}{\tau_r} P_r \quad (22)$$

where P_r is the total power radiated by the lamp, and τ_r is the time constant for such radiation. Equation (22) can be rewritten as

$$\left(\frac{dP_r}{dE}\right)\left(\frac{dE}{dt}\right) = -\frac{1}{\tau_r} P_r \quad (23)$$

where E is the internal energy of the lamp. If the system has reached steady state, then energy balance requires that

$$-\frac{dE}{dt} = P_{in} \quad (24)$$

where P_{in} is the power into the lamp just before crowbarring. Substituting P_{in} into Eq. (23) and solving for τ_r yields

$$\tau_r = \frac{P_r}{P_{in}(dP_r/dE)} \quad (25)$$

The power into the lamp per unit length is a well-known empirical function²⁷ of current density, j (A/cm²), tube diameter, d (cm), and gas pressure, P (Torr), given by

$$P_{in} = 1.3 \left(\frac{P}{450} \right)^{0.2} \left(\frac{\pi}{4} d^2 \right)^{3/2} j^{3/2} \text{ (W/cm)} \quad (26)$$

Given a model for $T(j, d)$, which is the temperature as a function of current density and diameter, and assuming thermodynamic equilibrium at a constant temperature throughout the lamp, then the internal energy, E , can be calculated by the Saha equation.²⁸ Therefore, to determine τ_r , we require a model for both $T(j, d)$ and $P_r(j, d)$.

We investigated two such models to fit our results with xenon. The first was the Trenholme-Emmett model,²² which presumes

$$T(j, d) = [(9450 D^{0.03} j^{0.01})^6 + (93 D^{0.27} j^{0.34})^6]^{1/6} \quad (27)$$

where j (A/m²) is the current density and D is the effective plasma diameter given by

$$D = \frac{d}{1 + 0.0008/d} \quad (28)$$

The real tube diameter, d , is in cm. This temperature relationship was empirically derived to fit experimental measurements of flashlamp radiant efficiency. The empirical fitting used the emissivity of the plasma, which was obtained from separate measurements.

We applied a second model derived from Rovinskii.²⁹ He obtained an implicit relationship for $T(j)$ given by

$$j \approx 3.5 p^{1/2} T^{5/4} U_i^{-1/2} \exp\left(-5000 \frac{U_i}{T}\right) \quad (29)$$

where p is the pressure in atm and U_i is the ionization potential in eV. The formula was derived by setting the input power, determined from the Spitzer conductivity formula³⁰ (scaled by a constant to fit experimental data), equal to the radiated power, determined from the Biberman and Norman theory³¹ (assuming an optically thin plasma). Electron density, n_e (cm⁻³), was related to current density by

$$j \approx 2.4 \times 10^{-15} n_e \quad (30)$$

through the use of the electron drift velocity.

The similarity in predicted radiated power as a function of current density for the two different models is not surprising since, in both cases, temperature and emissivity functions were related to give overall radiant efficiencies near unity. However, the predicted temperatures as a function of current density are different in shape although similar in magnitude. Moreover, the

dependences of the predicted temperature on fill pressure and bore diameter are different for the two models.

Using Eq. (25), we calculated τ_r vs P_{in} for the two models, as shown in Fig. 6-26. The Rovinskii model gives values for τ_r that agree with experiment within 30% and that vary with P_{in} in the direction observed in the experiment. Agreement with the Trenholme-Emmett model is not as good. Predicted values for τ_r vary in a direction opposite to observed input power. We also show the Rovinskii results for argon and krypton in Fig. 6-26. Differences arise from the different ionization potentials, U_i , in Eq. (29). The model predicts the same general trend as the experiment and shows relatively small differences among rare gases, as observed.

The variation of τ_r with P_{in} in the two models depends rather subtly on the $T(j)$ function. Since the term P_r/P_{in} in Eq. (25) is roughly constant, the major variation of τ_r with P_{in} is given by

$$\frac{dP_r}{dE} = \left(\frac{dP_r}{dj} \right) \left(\frac{dj}{dT} \right) \left(\frac{dT}{dE} \right) \quad (31)$$

The factor dT/dE is independent of the two models and depends only on the Saha equation. As previously discussed, dP_r/dj is essentially identical for both models. Thus, only dj/dT distinguishes one model from the other at a given pressure and filling fraction. For the two models, dj/dT varies in opposite directions with j , which accounts for the difference in predictions.

Time-Lag Measurements. As expected from our crowbar measurements output radiation from a xenon flashlamp generally lags behind input power. Figure 6-27(a) is a comparison of the time dependence of radiant power measured by a pyroelectric detector and input power determined by $V \cdot I$. These data were obtained for a 1.5-cm-bore $\times$ 112 cm-arc-length xenon lamp excited at an explosion fraction of 0.19 and using the apparatus described in "Optical Materials, Surfaces, and Coatings," later in this section. Both curves were arbitrarily normalized to their peak values. Notice that output radiation appears to lag behind input power with a simple time shift.

We obtained an optimal value for the time lag by plotting P_r vs P_{in} and then finding the best value for the time shift to minimize the area between the curves. Replotting P_{in} and P_r vs time with the optimal

Fig. 6-26. Radiation decay time vs input power for a 1.9-cm-bore flashlamp using two different models.

Fig. 6-27. Input electrical power and output radiant power vs time (a) without a time offset and (b) with a time offset of 37 μ s.

Fig. 6-28. Optimized time lags to minimize hysteresis.

time shift yields the curves shown in Fig. 6-27(b). Notice that with this time offset, the output radiation closely follows the input power and that there is no significant difference in the radiative efficiency between the rise and fall of the current. The instantaneous power efficiency for radiation must then equal the time-integrated radiation efficiency.

We determined optimal time lags from the pyroelectric detector measurements for various input energies to the flashlamp. The results are plotted in Fig. 6-28 both as a function of peak discharge power and of explosion fraction for a 1.5-cm-bore xenon flashlamp. Although exact agreement is not expected, the results can be compared to Fig. 6-25. In one case, we measured the instantaneous decay rate and, in the other case, we determined the optimum time lag averaged over the full pulse. Nonetheless, agreement between the two sets of data is quite good in both magnitude and trend.

Since the radiative output of the lamp lags behind the input, it seems reasonable that the electrical properties should lag behind the input as well. The appropriate quantity to consider in this regard is the conductivity of the lamp. The conductivity should simply be a function of the thermodynamic state of the plasma. However, to obtain the conductivity, the current density must be determined. Since we know that

the arc expands in time and typically fills the lamp bore nonuniformly, the actual current density is somewhat difficult to determine. However, one can naively define an average current density by simply dividing the current by the bore area and neglecting other complications. After defining current density in this manner, we determined the conductivity of a 1.5-cm-bore xenon flashlamp vs input power. The results are shown in Fig. 6-29 for three different input energies. Notice that, as observed for plots of voltage vs current, the curves exhibit hysteresis and do not follow a common path for the three input energies. The oscillations on the power rise are associated with acoustical waves,³² which are observed as a modulation on the voltage pulse.

Once again, we found an optimal time shift at each input energy to obtain the closest common characteristic for the conductivity vs P_{in} . The hysteresis is nearly eliminated in Fig. 6-29(b), and the curves at the three inputs essentially collapse to a common characteristic. The common characteristic for conductivity is approximately given by $(V \cdot I)^{1/3}$, as expected from the well-known empirical relation $V \approx I^{1/2}$ (Ref. 33).

The optimal time lags determined by this procedure are also shown in Fig. 6-28. Once again, the time lags are in close agreement in both magnitude and trend with those

Fig. 6-29. Conductivity vs input power (a) with no time shifts and (b) with optimized time shifts to minimize hysteresis.

found from the radiative-response measurements. Exact agreement is not expected since radiated power and conductivity depend on the state of the plasma in functionally different ways.

Conclusions. We have shown that both the radiative and electrical characteristics of xenon-flashlamp plasma exhibit obvious transient effects. For long input pulses, the response time of the plasma can be treated as a roughly constant time lag of $\sim 30 \mu\text{s}$ between the radiative and electrical properties of the plasma and the input power. Arc expansion does not have a strong effect on such a description.

Authors: A. E. Orel and H. T. Powell

Major Contributors: L. P. Bradley, K. S. Jancaitis, and M. J. Kushner

Blackbody Pumping and Depumping of Nd:Glass

We have calculated the pumping and depumping rates in Nd:glass driven by a blackbody distribution of radiation that is located outside a glass disk. Our calculation is particularly relevant for Nd pumping by chemical (pyrotechnic) flashlamps that can exhibit a blackbody spectrum. Such lamps are being studied under contract support by GTE Sylvania. The calculation is also pertinent to flashlamp pumping of laser amplifiers. Flashlamps are assumed to be gray-body radiators, which are at a well-defined thermodynamic temperature. The radiation

field that they can produce, even in a highly reflective optical cavity with many lamps, cannot exceed an intensity corresponding to their blackbody temperature. Thus, the calculation places upper limits on flashlamp pumping and depumping rates.

For the depumping rate (stimulated emission caused by flashlamp radiation), the upper limit provided by a blackbody calculation should be close to the actual value achieved in amplifiers. For three out of the four flashlamp-depumping bands (at 1.06, 1.3, and $1.8 \mu\text{m}$), the flashlamps themselves are the major light sink since, at these wavelengths, the reflectivity of the cavity is quite good, and the Nd:glass is transparent. Then the emitter and the absorber (Flashlamp) are at the same temperature, which is precisely the condition necessary to achieve a blackbody radiation field.

The geometry for which we calculated the photon-induced reaction rates was a thin, cylindrical glass slab completely surrounded by a blackbody radiator at temperature T . We presumed that radiation could enter only through the faces. Though our analysis required that the transverse dimension of the disk be much greater than its width, the exact shape of the disk was irrelevant. For the cases studied, second-order effects, such as depletion of the ground-state density of absorbers or additions to the radiation field by emission from the glass, were negligible ($<10\%$). We also found it sufficient to calculate only the rates generated by single-sided illumination; the corresponding results for dual-sided

Fig. 6-30. Calculated excitation efficiency for blackbody pumping of LG-750 Nd:glass disks vs doping-thickness product and blackbody temperature.

pumping could then be obtained from consideration of symmetry.

In the geometry described above, the radiation field inside the glass disk differs from the isotropic, homogeneous blackbody distribution presumed to be outside the disk as the result of two effects. First, refraction at the glass surface compresses the angular distribution inside the glass to angles less than the critical angle $\phi_c = \sin^{-1}(1/n)$, where n is the ratio of the index of refraction of the glass to that of the surrounding medium. Second, absorption in the disk results in a spatially-dependent spectral distribution due to the dependence of the absorption coefficient on the photon wavelength.

Taking both processes into account, the local reaction rate at a distance z from the glass face caused by single-sided illumination is given by the expression

$$R(z) = 4\pi N_R \int d\nu \frac{\sigma_R(\nu) (\nu/c)^2}{e^{h\nu/kT} - 1} \times \int_{\cos \theta = 1/n}^1 d(\cos \theta) \exp \left[-N\sigma_A(\nu) \frac{z}{\cos \theta} \right], \quad (32)$$

where N_R is the density of the reactant ion in the glass, $\sigma_R(\nu)$ is the reaction cross section, and $\sigma_A(\nu)$ is the overall absorption

cross section of the glass, which may include processes in addition to the excitation of the ion of interest. In Eq. (32), the two integrals represent the convolution of the blackbody emission with the reaction cross section and the depletion of the radiation by absorption in the glass, respectively.

With the reaction rate given by Eq. (32) and with $\sigma_R(\nu) = \sigma_A(\nu)$, the energy-transfer efficiency for blackbody pumping of a Nd:glass disk of total width w is

$$\eta(w) = h\nu_l \int_0^w \frac{R(z) dz}{\sigma_B T^4}, \quad (33)$$

where $h\nu_l$ is the energy of the laser photon, and σ_B is the Stefan-Boltzmann constant.

Figure 6-30 shows the calculated results for the excitation efficiency of LG-750 phosphate glass as a function of doping-thickness product and blackbody temperature, using the absorption cross section for Nd given in Ref. 34. The peak energy transfer occurs near 5500 K. At this temperature, the efficiency for the doping-thickness product of 10^{21} ions/cm² was ~18%.

Calculation of the excitation level that can be achieved in finite-sized disks, in general, must take into account such complications as radiation trapping and ASE. For glass samples that are not too large, however, such effects should be dominated by the first-order pumping and decay terms. With this point in mind, we have calculated the steady-state inversion ratio, N^*/N_R , generated by blackbody pumping of finite-signal disks under the assumption of a constant lifetime for the excited upper level. By a derivation similar to that for Eq. (33),

$$\frac{N^*}{N_R} = \frac{2\tau_D}{wN_R} \int_0^w R(z) dz, \quad (34)$$

where τ_D is the upper-state lifetime and N_R is the Nd³⁺ ion density. Notice that Eq. (34) gives the excitation density averaged over the full disk width.

The results of Eq. (34) for LG-750 phosphate glass disks with $\tau_D = 350 \mu s$ are shown in Fig. 6-31, where we have plotted

the inversion ratio as a function of the blackbody temperature and the doping-thickness product, $N_R w$. As expected, the average inversion ratio is a decreasing function of doping-thickness product.

Figure 6-31 shows that a 7000-K blackbody could produce Nd^* densities of interest ($N^*/N \approx 0.01$) for laser fusion but that a 5500-K blackbody probably could not. The lower temperature is more typical of very hot chemical reactions. By burning a 1-mm zirconium shroud in an UF_6 atmosphere, GTE Sylvania has been able to produce Nd^* concentrations roughly 20% of those produced by a xenon flashlamp of comparable dimensions. This is generally consistent with Fig. 6-31.

We also used Eq. (32) to estimate the magnitude of the stimulated radiative depumping of Nd ions by blackbody radiation, in the limit of a small disk size, where there was no significant variation of optical intensity with position in the glass. The upper laser level of Nd may decay radiatively into emission bands centered at either 0.88, 1.06, 1.35, or 1.80 μm . For each of these transitions, identified by the subscript i , the ratio of stimulated depumping rate for the transition i to the total spontaneous decay rate for dual-sided illumination of a small disk is given by

$$\frac{\tau_{i,stim}}{\tau_{i,spont}} = \left(1 - \sqrt{1 - \frac{1}{n^2}}\right) \int \frac{\beta_i g_i(\nu)}{e^{h\nu/kT} - 1} d\nu \quad (35)$$

where $\tau_{i,spont}$ is the spontaneous radiative lifetime of the upper state, β_i is the fluorescence branching factor ($\sum \beta_i = 1$), and $g(\nu)$ is the normalized fluorescence-line shape [$\int g(\nu) d\nu = 1$] for the line of interest.

As a specific example, we used Eq. (35) to calculate the effects of blackbody radiative depumping in LG-750 phosphate glass, using experimental optical data.^{34,35} The results of this calculation are presented in Fig. 6-32. In the temperature range of typical pyrotechnic sources (from 4000 to 5000 K), the stimulated depumping rate is $\leq 1\%$ of that from spontaneous decay. At elevated temperatures, the radiative depumping rate rises so that at 11 000 K, the increase in the effective upper-state decay

Fig. 6-31. Predicted average steady-state Nd^{3+} inversion ratio produced by blackbody pumping of LG-750 glass disks vs doping-thickness product and blackbody temperature.

Fig. 6-32. Prediction of the small-disk model for the ratio of stimulated decay rates for various Nd transitions to total Nd^* spontaneous decay rate vs blackbody temperature.

rate due to stimulated emission on the 1.06- μm transition alone is $\sim 5\%$. Thus, stimulated depumping is a small but not negligible effect when xenon flashlamps are used to drive the laser inversions. At high energy loading, such lamps display blackbody spectral radiance at temperatures of from $\sim 10\,000$ to $12\,000$ K in the relevant energy range of from ~ 1.1 to 1.2 eV (Ref. 36).

Authors: K. S. Jancaitis and H. T. Powell

Optical Materials, Surfaces, and Coatings

Introduction

Our primary task during 1983 was development of a 351-nm AR coating for the silica lenses and windows in Nova. The goal was to provide coatings that withstand irradiation

by 1-ns, 351-nm pulses at fluences of 5 J/cm^2 . We evaluated three types of AR surfaces: porous-silica, sol-gel films; porous surfaces made by etching damage trails caused by passage of nuclear particles through silica surfaces; and coatings made by conventional physical-vapor deposition. Our results indicate the feasibility of making porous-silica coatings that meet Nova specifications.

We also measured thresholds for bare, polished, fused-silica surfaces made by potential vendors of Nova components. Most of the bare surfaces had thresholds well above the goal of 5 J/cm^2 for coated surfaces. However, some bare surfaces were damaged at fluences as low as 4 J/cm^2 .

We continued evaluation of reflective and AR coatings made by physical-vapor deposition. We studied the influence of deposition parameters and coating designs on damage thresholds and the scaling of 351-nm thresholds with laser-pulse duration. Results of our experiments provide an adequate data base for 351-nm thresholds measured with 1-ns pulses, although additional data will be needed at longer pulse durations.

Author: D. Milam

Sol-Gel Coatings

Porous-silica coatings are now viewed as the most promising solution to the problem of fabricating damage-resistant, 351-nm AR coatings. Research on these coatings was conducted at Westinghouse Research and Development Center (Pittsburgh, Pa.), Corning Glass Works, Inc. (Corning, N.Y.), Hoya Optics, Inc. (Japan), and on site at Livermore. All efforts, except those at Corning, focused on coatings made from organometallic solutions. Corning investigated an alternate (proprietary) process using water-based chemistry.

Our goal was to provide 351-nm AR coatings with thresholds of 5 J/cm^2 for 1-ns pulses. Coatings that met this goal were made using processes developed by both Westinghouse and Corning. Development of both processes is continuing; currently, the two processes yield rather similar

coatings. Because some aspects of the Corning process are proprietary, most of the available data are for coatings made by the Westinghouse procedure. Thus, we will concentrate on experiments with the latter process.

At Westinghouse, B. E. Yoldas demonstrated fabrication of sol-gel AR coatings made from liquids containing polymer silica molecules.³⁷ Early studies (Westinghouse work contracted by LLNL) provided coated test substrates with two-surface transmission exceeding 99% for all frequencies between 351 and 1053 nm. The optical thickness of these coatings was from 500 to 600 nm. Damage thresholds measured for many coatings (82 samples) made at Westinghouse are given in Fig. 6-33(a). The highest threshold was 5 J/cm^2 ; the median was 2.1 J/cm^2 . The two coatings with thresholds $>9 \text{ J/cm}^2$ were much thinner than those with thresholds $\leq 5 \text{ J/cm}^2$. The coating apparatus at Westinghouse was not housed in a clean room, and the morphology of damage in Westinghouse coatings suggested that damage was caused by laser heating of dirt particles in the coating layer.

Because contaminants were apparently limiting thresholds, a research apparatus for coating 5-cm-diam substrates was built in a class-100 clean room at LLNL. Here, we first studied filtration of the precursor liquid. The coating solution is made by adding ethyl silicate to a mixture of ethyl alcohol, water, and nitric acid. A detailed recipe of the coating process evolved in studies at LLNL is given at the end of this article. We found that both ethyl silicate and the mixture of other liquids could be filtered to $0.02 \mu\text{m}$. The completed solution containing polymer molecules of silica could readily be filtered to $0.20 \mu\text{m}$ and, with some difficulty, could be filtered to $0.02 \mu\text{m}$. However, coatings made at LLNL from carefully filtered solutions usually had damage thresholds $<2 \text{ J/cm}^2$ regardless of whether they were made by LLNL or Westinghouse personnel.

To investigate the possibility that dirt in the coatings was contamination left on or in substrate surfaces, we then evaluated several techniques for cleaning substrates. In a typical cleaning cycle, the substrate was soaked in an acidic or basic solution,

scrubbed with a rotating brush under soap and water, rinsed in deionized water or ethyl alcohol, and spun dry. In some cases, the polished surface was completely removed by HF acid.

In evaluating a procedure, we coated substrates that had been subjected to all steps in the procedure, substrates cleared by various permutations of the steps, and some control samples cleaned only by

drag wiping the substrate with tissue wetted first with acetone and then ethyl alcohol. Thresholds of coatings were found to be rather insensitive to the substrate cleanliness.

Figure 6-33(b) is a histogram of damage thresholds of 36 coatings made at LLNL. All coatings had thicknesses ranging from 500 to 600 nm. The median, 1.2 J/cm^2 , was less than the median threshold of the

Fig. 6-33. Laser-damage thresholds (0.6 ns, 355 nm) of porous-silica AR coatings deposited on silica substrates at (a) Westinghouse, (b) LLNL, and (c) for 90- to 120-nm coatings deposited at LLNL.

coatings made at Westinghouse without the aid of a clean-room environment. Therefore, we concluded that damage thresholds were limited by an intrinsic property of the coating. The most likely problem was trapping of carbon residues.

Carbon is present in two forms in the precursor liquid. The polymer molecules have ethyl side groups, and the carrier liquid is primarily ethyl alcohol. We believed that the carrier liquid would readily evaporate. The coatings were usually baked at from 400° to 500°C. Baking should have eliminated ethyl side groups, which decompose at ~175°C. However, carbon from either the solvent or the ethyl side groups could be trapped if pores were sintered shut before the carbon was oxidized to carbon dioxide.

To increase the probability that carbon was eliminated, we thinned the coatings from the 500-nm thickness (needed to provide broadband antireflectivity) to ~90 nm. Thinning produced a large increase in thresholds, as shown in Fig. 6-33(c). Some thin coatings had thresholds as high as 9 J/cm², which is the median for bare, polished fused silica. Windows bearing these thin coatings transmitted >99% at 351 nm, but the narrow bandwidth of the transmission spectrum indicated that the refractive index of the thinned layers was homogeneous, not graded. We believe, however, that observation of large thresholds for thin coatings confirms that carbon was being trapped in the thicker layers that we initially made.

Several attempts were made to produce thicker carbon-free layers. The ethyl side groups can be eliminated by hydrolysis, but films dried in water vapor were usually dense. Use of multiple layers, each 90 nm thick, produced thick films with broadband antireflectance, but these stacked-layer coatings usually had low thresholds. The problem of fabricating thick, damage-resistant coatings remains unsolved.

Fabrication of Coating Solution. Our recipe for making thin sol-gel coatings has undergone constant evolution at LLNL. Although our process may not yet be optimized, the recipe in current use is given here in detail.

Mix 127.5 g ethyl alcohol, 38.7 g water, and include 10 drops nitric acid. Add 208.0 g ethyl silicate to the solvent. Hold

this mixture for 4 h at the reflux temperature of ethyl alcohol, mixing continuously. Cool the solution to room temperature, and add 382 g ethyl alcohol. Stir thoroughly. Store the solution in a sealed container to prevent contact with atmospheric water. The shelf life of the polymerized solution is not known.

Chemical reactions believed to occur during formation of the polymer molecules are:

where -X represents -OCH₂CH₃. In the second and fourth steps, polymerization may also proceed through reaction of hydrolyzed ethyl silicate molecules:

Application of Coating. Withdraw a substrate from the polymer solution at the rate of ~4 cm/min into a dry nitrogen environment. Coating thickness depends primarily on the silica content, the viscosity of the solution, the withdrawal rate, and the atmosphere in which drying occurs. Adjustment of withdrawal rate can be used to tune the optical thickness of the coating, which should be ~90 nm.

Curing the Coating. Dry the wet layers thoroughly in dry nitrogen, then bake them at 480°C for from 0.5 to 1 h. Cool the coated substrate, and etch the coatings for from 4 to 5 min in 2×10^{-5} M HF. Wash the substrate in deionized water and then ethyl alcohol.

In summary, we developed a narrow-bandwidth, 351-nm AR coating with a damage threshold in excess of our goal, 5 J/cm². Development of a coating with broadband antireflectivity is continuing.

Author: D. Milam

Major Contributors: I. M. Thomas, J. G. Wilder, F. Rainer, and M. C. Staggs

Etched Nuclear Tracks

Our previous attempt³⁸ to prepare satisfactory AR surfaces on fused silica using etched nuclear tracks was unsuccessful. Passage of fission fragments through a silica

surface creates trails of damaged material that etch at a rate greater than undamaged material. Etching a single trail produces an approximately conical pit. A surface with a large spatial density of conical pits has a graded refractive index.

Our theoretical analysis of the depth profile of the average refractive index was done under the assumptions that the fission fragments strike the surface at random incidence and that the pits are perfect cones. The calculated profiles were smooth curves; the refractive index decreased quadratically near the solid glass and decayed exponentially far from it. Both the discontinuity at the free surface and the maximum index gradient could be made arbitrarily small by increasing the depth of etching. Broadband antireflectivity would be expected for surfaces with such index profiles.

In 1983, we studied the effects of varying both the dosage of fission fragments and the etch solution. We obtained the best results with P-etch,³⁰ which is $3\text{HF}-2\text{HNO}_3-60\text{H}_2\text{O}$ (by volume). Examination of surfaces by scanning electron microscopy (SEM) indicated that P-etch created cones with apex angles of $\sim 25^\circ$. The apexes and ridges between pits were slightly rounded, and the rounding affected a larger proportion of smaller pits. Our attempts to reduce scattering losses by producing cones with smaller apex angles were unsuccessful.

The antireflectivity that we achieved was broadband but modest. In the best case, for silica substrates treated on both surfaces, transmission at normal incidence was 95.5% at both 0.35 and 0.53 μm ; the corresponding spectral reflectances were 1.3 and 2.1%. For untreated fused silica, the transmission at both wavelengths was $\sim 93\%$. SEM analysis of the etching of isolated tracks indicated that the etching depth was adequate. We attribute the remaining surface reflectance to apex and ridge rounding, which causes large gradients in the index profile.

The remainder of the transmission losses are due to diffuse surface scatter, since absorptive losses are negligible. The scattering losses increased with etching depth. The spectral reflectance of the glass surfaces could be reduced to very low values by simply increasing the scattering, but the total transmission was also reduced (in one case, the transmission was 90.0%, and the reflectance was 0.4% at 0.35 μm). The

scattering losses were found to be proportional to λ^{-2} . For heavily etched surfaces (large pits), $n = 4.0$ as expected from standard theories. The value of n was as high as 1^4 for more lightly etched surfaces. High values of n are not understood theoretically.

Stevens and Cody⁴⁰ prepared AR surfaces on fused silica with the same method, but they used ions from an accelerator (which gave a normal, rather than random, incidence angle) to produce the damage tracks. Their pits had a rather large cone angle, resulting in greater scattering losses at short wavelengths than in our experiments. Nevertheless, they were able to achieve a maximum transmission of 96.8% at 600 nm. Thus, normal incidence of the tracking particles appears to provide some advantage.

Author: R. W. Hopper

Laser-Damage Studies

During 1983, we conducted several experimental studies of laser-damage thresholds on AR coatings and reflectors. Unless otherwise specified, our reported damage thresholds were measured with 0.6-ns, 0.35- μm pulses.

Thresholds of bare-fused-silica windows polished by Zygo, Inc. ranged from 8 to 13 J/cm^2 , as shown in Fig. 6-34. Thresholds of

Fig. 6-34. Laser-damage thresholds on bare-fused-silica windows polished by Zygo, Inc.

Fig. 6-35. Laser-damage thresholds on multilayer dielectric AR coatings made of several combinations of materials.

Fig. 6-36. Laser-damage thresholds on AR coatings made only of low- and moderate-index materials.

samples from major suppliers of optical fabrication for Nova were comparable. However, samples polished by some vendors had thresholds below 3 J/cm^2 . Etching of low-threshold surfaces with dilute HF revealed extensive subsurface polishing damage, which was not observed when high-threshold substrates were etched. Therefore, we believe that the low thresholds resulted either from rapid polishing or from high load pressure that caused grinding marks to fill with polishing residues.

We tested several $0.35\text{-}\mu\text{m}$ AR coatings made by the research group at Optical

Coating Laboratory, Inc. (OCLI). The coatings were deposited by electron beam evaporation onto fused silica that was bowl feed polished. Each of the coatings consisted of four layers, of which two were low-index material and two were high-index material. Thresholds for these coatings ranged from 2.4 to 3.6 J/cm^2 , as shown in Fig. 6-35. Coatings made of the combination $\text{Sc}_2\text{O}_3/\text{MgF}_2$ had the highest thresholds. A second set of coatings contained only materials with either low or moderate refractive indexes. Thresholds for these coatings, shown in Fig. 6-36, were

Fig. 6-37. Laser-damage thresholds on four-layer AR coatings as a function of the thickness of layers used in each coating.

Fig. 6-38. Laser-damage thresholds on all-quarter-wave reflectors made of four combinations of materials.

slightly lower than those of conventional coatings containing both high- and low-index materials. Reasonably high thresholds were observed for the combinations SmF_3/SiO_2 , SiO_2/MgF_2 , and SmF_3/MgF_2 . We expected higher thresholds for coatings containing only low-index materials and do not yet understand the results.

In a study of coating designs, OCLI made four-layer $0.35\text{-}\mu\text{m}$ AR coatings of varied thickness. Thickness of the individual layers ranged from $\sim\lambda/4$ in the first design to $\sim\lambda/8$ in the last design. The range provided an interesting test of a proposed damage mechanism that suggests that thresholds should decrease as coating thickness

increases.⁴¹ The mechanism was proposed to explain an observed relationship between thresholds and thicknesses of single-layer coatings.⁴² That observation was complicated by the fact that internal standing-wave intensities in single-layer coatings vary with layer thickness. However, in the four-layer AR coatings that we studied, standing-wave intensities are almost independent of layer thickness. The results of our tests are given in Fig. 6-37. Thresholds varied only slightly with layer thickness and were greater for thicker coatings. Therefore, we were unable to demonstrate a fundamental relationship between layer thickness and damage thresholds.

We also tested a large number of 0.35- μm , multilayer, dielectric reflectors fabricated by OCLI. Thresholds for standard 21-layer and 15-layer reflectors made of 4 combinations of materials are shown in Fig. 6-38. The highly reflective (HR) coatings with the largest thresholds, 3.5 J/cm², were composed of $\text{Sc}_2\text{O}_3/\text{MgF}_2/\text{SiO}_2$. The outer 12 layers of the coatings, which experienced the greatest intensities, were made of $\text{Sc}_2\text{O}_3/\text{MgF}_2$, which is highly stressed and not easily used for fabrication of the entire reflector. The addition of half-wave-thick overcoats of either MgF_2 or SiO_2 reduced thresholds. This outcome is counter to results of similar tests made at 0.25 and 1.06 μm .

We also tested reflectors with either two or four outer pairs of either $\text{SmF}_3/\text{MgF}_2$ or

$\text{SmF}_3/\text{SiO}_2$. Thresholds for these coatings were from 1.9 to 2.5 J/cm², which is below the thresholds of 3 to 4 J/cm² observed for some all-oxide coatings.

For some reflectors, the thicknesses of the outer layers were adjusted to minimize standing-wave intensity in the high-index material. Such reflectors had the highest thresholds, ~ 5 J/cm², of any 0.35- μm design tested. Data for multilayer reflectors with modified outer layers are shown in Fig. 6-39. In a typical all-quarter-wave design, the standing-wave intensities are equal in both the high- and low-index layers. Several authors have shown that thickening the outer low-index layer reduces the intensity in the adjacent high-index layers.⁴³⁻⁴⁵ Since the high-index layer is thought to be more easily damaged than the low-index layer, reduction of intensity in the high-index layer should increase damage resistance. We have frequently tested this design in reflectors for 0.25 and 1.06 μm and have found no correlation between thresholds and coating alterations. Our success in current experiments is encouraging, and the experiment will be repeated.

We performed two scaling experiments with 0.35- μm pulses and 0.6-, 1.0-, 5-, and 9-ns durations to measure thresholds for a selection of HR and AR coatings. The data shown in Fig. 6-40 cannot be described by a single scaling relationship of the form $T = T_0 t^{\delta}$. Values of the coefficient δ ranged from 0.21 to 0.50. Therefore, provision of

Fig. 6-39. Laser-damage thresholds on multilayer reflectors with outer layers modified to reduce standing-wave intensities in the high-index material.

Fig. 6-40. Laser-damage thresholds of randomly selected (a) thin-film reflectors and (b) thin-film AR coatings measured with laser pulses of varied durations.

adequate design data will require generation of a histogram of thresholds for each coating type at each pulse duration of interest.

We also used pulses of ~ 1 ns in duration to measure thresholds at 0.35, 0.53, and $1.06\text{ }\mu\text{m}$ for a selection of $\text{ZrO}_2/\text{SiO}_2$ reflectors. The reflectors were made in a single deposition at OCLI. Masking of the vapor stream in the coating chamber allowed simultaneous fabrication of reflectors tuned to the three different wavelengths. The median thresholds at 1.06 and $0.35\text{ }\mu\text{m}$ were 8.1 and 2.5 J/cm^2 , respectively, as shown in Fig. 6-41. Thresholds for $0.53\text{ }\mu\text{m}$ were uniformly spread between 3.7 and 10 J/cm^2 .

In summary, we have been unable to fabricate AR coatings by physical-vapor deposition to meet the Nova requirement of 5 J/cm^2 . Further, since many material combinations and coating designs have been tested, it is unlikely that significant improvements will be made in the immediate future. The best reflectors, with $0.35\text{-}\mu\text{m}$ thresholds of $\sim 5\text{ J/cm}^2$, were made by a technique that tests our basic understanding of damage in coatings. Therefore, the experiment illustrated in Fig. 6-39 deserves repetition.

Author: D. Milam

Major Contributors: M. C. Staggs, F. Rainer, and S. E. Peluso (LLNL); C. K. Carniglia and T. Tuttle-Hart (Optical Coating Laboratory, Inc.)

Thermal-Lensing Studies: Two-Photon Absorption and Solarization

In 1983, we expanded the measurements that we reported in 1982 to include solarization kinetics in BK-7 and two-photon-absorption (2PA) measurements in BK-3, BK-7, and BK-10 borosilicate glasses. We also began a systematic study of 2PA near the band edge using photothermal-lensing spectroscopy. The need for better understanding of 2PA in fusion-laser materials, especially in glass, is now well established. The general technique and approximate layout of our experiment have been discussed in previous reports.⁴⁶

The results of our measurements in BK-7 indicate that a two-photon process initiates

Fig. 6-41. Thresholds of $\text{ZrO}_2/\text{SiO}_2$ reflectors measured at (a) $0.35\text{ }\mu\text{m}$, (b) $0.53\text{ }\mu\text{m}$, and (c) $1.06\text{ }\mu\text{m}$.

solarization, whereas one-photon bleaching limits it. Other work on solarization in borosilicate glasses indicates that the increased absorption is attributable to color-center formation, that is, to charge-trapping mechanisms.⁴⁷ Figure 6-42 shows that the absorption first rises and then approaches a limiting, steady-state value with repeated exposure to optical pulses of a fixed fluence. At high intensity optical damage occurs before a limiting solarization is reached.

Figure 6-43 shows the estimated long-exposure limit, as a function of fluence, of each of the curves in Fig. 6-42. In steady-state solarization, the rate of formation equals the rate of color-center annihilation. Thus, the linear dependence in Fig. 6-43 indicates that, if a two-photon process dominates the creation of color centers, then a one-photon process dominates the annihilation of color centers. (We assume that the density of possible charge-trapping sites remains fixed and is much greater than the number of actual color centers.) Separate tests confirmed that one-photon absorption, while initially absent, was much greater at low fluences in the solarized glass than was 2PA. Further, by exposing various solarized volumes of BK-7 glass to low-intensity, 10-Hz pulses, the absorption coefficients fell

Fig. 6-42. Two-photon-induced solarization in BK-7 glass.

Fig. 6-43. Steady-state solarization vs peak fluence.

Table 6-1. Relative 2PA coefficients, β , at $0.53 \mu\text{m}$ in borosilicate glasses.

Glass	β
BK-3 and similar glasses	0.25 ± 0.15
BK-7	1.0
BK-10 and similar glasses	0.2 ± 0.1

Fig. 6-44. Thermal-lensing signal vs fluence squared.

smoothly from 25 to 35% to new steady-state values. However, absorption did not bleach entirely away. Apparently, more than one type of color center was formed in the solarization process, but not all types of color centers formed could photobleach at $0.53 \mu\text{m}$.

We also measured the relative 2PA coefficients in various borosilicate glasses at $0.53 \mu\text{m}$. Table 6-1 shows that at this wavelength BK-7 has a higher 2PA coefficient than either BK-3 or BK-10. This outcome was expected because of the slightly narrower band gap of BK-7 (3.9 eV) relative to either BK-3 (4.0 eV) or BK-10 (4.1 eV). Figure 6-44 shows the peak single shot thermal-lensing signals, normalized by sample thickness, in three samples. For 2PA, the signal should be a straight line when plotted against the square of the incident peak fluence. Only the BK-7 data in Fig. 6-44 appear to lie along a straight line (within experimental uncertainty). At high fluences, the BK-3 signal appears to saturate, and the BK-10 signal appears to exhibit a greater-than-quadratic dependence on the applied fluence. The reasons for such disparate behavior among these three borosilicate glasses are not yet understood.

We have begun measuring relative 2PA spectra using photothermal lensing. As a variable wavelength source, we use a Moletron DL-18 dye laser pumped by a Q-switched Moletron MY-34 Nd:YAG laser. Energy per pulse is monitored by a spectrally flat pyroelectric detector. Apart from the different pulsed-laser source and the different energy detector, the experimental details are essentially the same as reported for our previous thermal-lensing measurements at $0.53 \mu\text{m}$ (Ref. 46).

Authors: W. T. White and M. J. Weber

Major Contributor: T. W. Hindley

Laser Glass

Concentration Quenching

Concentration quenching in laser glass is the shortening of Nd-fluorescence lifetime as Nd concentration increases. This property of laser glass is important in the design of laser amplifiers because amplifier efficiency is a function of the fluorescence lifetimes, τ , of the excited Nd ions in the laser medium. The function depends on cavity geometry, flashlamp-packing density, and reflector geometry. Trenholme⁴⁸ found that

amplifier efficiency is proportional to $\tau^{0.67}$ for a spare-lamp, rectangular-disk-amplifier cavity and to $\tau^{0.5}$ for a round-head cavity. Ideally, we want the longest possible lifetime at a Nd concentration that is sufficient to absorb most of the flashlamp light.

Nd-fluorescence lifetime in a glass is determined by several processes, but the most important are:

- Radiative decay of the excited state. Radiative decay sets the maximum possible lifetime when all other decay modes are absent.
- Nonradiative transitions between Nd-ion pairs. Nonradiative transitions are the primary cause of concentration quenching.
- Energy transfer to molecular-impurity vibrations. The fundamental stretching mode of OH⁻ (from 2800 to 3500 cm⁻¹) is the most effective.
- Energy transfer to electronic states of impurity ions. The transition metals (such as Fe, Cu, V, and Ni) are the strongest quenchers; the rare earths (such as Sm, Dy, Pr, Er, and Yb) are less strong.
- Nonradiative energy transfer to phonon vibrations in the host material. This transfer is more important in materials with high-frequency vibrations, such as borate glasses.

In 1983, we studied nonradiative transitions between Nd-ion pairs in a variety of phosphate and silicate glasses. Our goal was to elucidate the sources of observed variations in concentration quenching from glass to glass. Schott Glass Technologies, Inc. and Hoya Optics, Inc. (Fremont, Calif. and Tokyo, Japan) melted samples for us that were dry (low OH⁻ content) and relatively free of transition-metal and rare-earth impurities. We found that concentration

quenching takes a variety of forms in the different glasses and that no single theory fits all the experimental data.

To accurately measure fluorescence lifetime, we must eliminate radiation-trapping effects, which derive from the resonance re-absorption of 880-nm Nd fluorescence. Radiation trapping is commonly neglected and can lengthen the lifetime by as much as 30%. We found⁴⁹ that standard spectroscopic samples (volumes were ~2 cm³) may have up to a 15% increase in fluorescence lifetime.

To minimize radiation trapping, we designed and tested the sample illustrated in Fig. 6-45. The sample is an array of 1-mm-diam × 1-mm-thick cylinders surrounded by an index-matched, absorbing optical cement that is placed over an absorbing filter glass. The 1-mm cylinders are small enough that <2% of the fluorescence is reabsorbed in the cylinder. The optical cement absorbs 800- to 1000-nm light and prevents fluorescence from one cylinder from reaching a

Fig. 6-45. Specially prepared samples eliminate radiation trapping in measurements of Nd-fluorescence-decay times.

Fig. 6-46. Nd-fluorescence-decay time decreases with increasing Nd concentration in ultraphosphate UP-16 (70% P₂O₅, 20% K₂O, 10% Ln₂O₃) and metaphosphate LG-750 (Nova glass).

Fig. 6-47. Nd-fluorescence-decay time vs Nd concentration in three silicate glasses. LG-650 has low radiative-transition probabilities; LG-660 has moderate probabilities; LG-670 is a Shiva silicate.

Fig. 6-48. Concentration quenching in the new ultraphosphate (UP-16), a Nova phosphate (LG-750), and a Shiva silicate laser glass (LG-670).

neighboring one. The cement is also index matched to keep fluorescence from being totally internally reflected within a cylinder. By using several cylinders, we increased the available signal.

Figures 6-46 and 6-47 show our results on concentration quenching in two phosphate and three silicate glasses respectively. LG-750 is one of the Nova laser glasses; UP-16 is a new ultraphosphate glass; LG-670 (ED-2) is one of the silicate glasses

used in Shiva; LG-660 is a tank-meltable silicate glass; and LG-650 is a low-cross-section silicate. Figure 6-48 shows a comparison of the two phosphate glasses and LG-670. Note that the phosphate glasses have less concentration quenching than the silicate and that the Nd lifetimes of the two phosphates differ considerably for Nd concentrations $> 5 \times 10^{20} \text{ cm}^{-3}$. Figure 6-49 shows the second e-folding time for three silicate glasses. (The first e-folding time refers to the time interval for fluorescence intensity to drop from its initial value to $1/e$, the second e-folding time is the interval from $1/e$ to $1/e^2$, the third e-folding time is the interval from $1/e^2$ to $1/e^3$.) The Nd lifetimes differ considerably, and the concentration at which the onset of quenching occurs varies substantially.

We must consider the theory of concentration quenching to explain our observations. The most successful theory of concentration and impurity quenching in laser materials is the dipole-dipole interaction theory, first discussed by Förster. In summary, the energy-transfer rate, W_{DA} , between two ions, one that donates energy (the donor, D) and the other that accepts

Fig. 6-49. Concentration quenching in three silicates with differing radiative-transition probabilities expressed by the peak $^4F_{3/2} \rightarrow ^4I_{11/2}$ cross section, σ .

energy (the acceptor, A), depends on three basic parameters expressed below in Eq. (36):

- The ion-ion distance to the inverse sixth power.
 - The transition strengths of an emission band of the donor and an overlapping absorption band of the acceptor.
 - The spectral overlap of the two bands.
- Then,

$$W_{DA} = K \left(\frac{1}{R_{DA}^6} \right) \left(\frac{\beta}{\tau_D} \right) (S_A) \times \int f_D(\nu) g_A(\nu) \nu^{-4} d\nu, \quad (36)$$

where K is a constant, R_{DA} is donor-acceptor distance, β is the branching ratio of the fluorescence band, τ_D is radiative lifetime of the donor, S_A is absorption-band strength, f and g are the emission- and absorption-band spectral line shapes, respectively, and ν is frequency.

In a system where an impurity ion is both acceptor and donor and is randomly distributed throughout the material, the R_{DA}^6 term will make the quenching decay rate depend quadratically on the ion concentration. We find that the quadratic dependence does not always appear, as illustrated in Fig. 6-50. Here, we plot the decay rate

minus the rate at low Nd concentrations as a function of concentration. (Because the Nd decay in glass is nonexponential, we calculate the rates at the three e -folding times.)

The quenching rates of LG-670, LG-660, and LG-650 increase approximately quadratically with concentration. However, UP-16 has a nearly linear dependence, and LG-650 has a 3.3-power dependence. UP-16 has little concentration quenching primarily because its ultraphosphate structure keeps the Nd-Nd distance large. We estimate that the Nd-Nd distance is $>6 \text{ \AA}$ based on the known structure of the crystalline analogue, $\text{KNd}(\text{PO}_3)_4$. We suspect that transition-metal impurities may be the primary cause for quenching in UP-16 because the dipole-dipole theory predicts a linear dependence in this case. We are investigating such a possibility with careful measurements of impurity concentrations.

LG-650 has an unusual quenching rate. We speculate that the source of the dependence may be structural because large Nd concentrations can change the glass structure. The LG-650 samples with Nd concentrations $< 5 \times 10^{20} \text{ ions/cm}^3$ have a compensating amount of La_2O_3 added so that the total lanthanide content is 5 wt%. However, the compositions of samples with higher Nd content were diluted by the Nd additions, which may have introduced structural changes.

Fig. 6-50. Nonradiative decay rates increase with Nd concentration of (a) UP-16, (b) LG-750, (c) LG-650, (d) LG-660, and (e) LG-670. Power-law curves are shown for comparison.

The second parameter, the transition strength of overlapping transitions in donor and acceptor ions, appears to have a significant effect on the concentration quenching of silicates. In Fig. 6-49, we can see a correlation between transition strength and onset of concentration quenching. On the other hand, we find that the third parameter, spectral overlap, varies by $<5\%$ in the same glasses. Thus, in silicate glasses, long lifetime implies low transition strengths and emission cross sections. Although low-cross-section glasses may be desirable for long lifetime, they unfortunately have low gain as well, which makes them susceptible to impurity absorptions at the laser wavelength. They also require high-fluence quenching for efficient energy extraction.

In contrast to silicates, phosphate glasses do not have a correlation between quenching rates and transition strengths. However, they still have a variation in concentration quenching, as shown by the two glasses in Fig. 6-48. We believe that concentration quenching in phosphates is primarily determined by the Nd-Nd distance, which can vary by a factor of 2 or more because phosphate glass structures can change significantly with composition. Phosphate glasses

are polymeric in structure because the PO_4 tetrahedra join at only two or three vertices. (Silicate glasses have a three-dimensional structure because the SiO_4 connect at three or four vertices.) The Nd-Nd distance in phosphates is sensitive to the chain length and to cross-linking of the PO_4 groups.

Author: S. E. Stokowski

Major Contributors: J. E. Lynch and M. J. Weber

Thermomechanical Properties

Thermomechanical properties of laser glasses are critically important in high-average-power lasers. In such lasers, heat-removal rates and thermal stresses determine the maximum power that the laser medium can sustain. In 1983, we initiated a program to collect data on these properties in both glasses and crystals. The thermomechanical figure of merit, W_{FM} , for a laser material designed to run at its fracture limit is

$$W_{\text{FM}} = \frac{\kappa(1 - \mu)}{\alpha l} S_1 \text{ (W/m)} \quad (37)$$

where κ is thermal conductivity, μ is Poisson's ratio, S_T is fracture strength, α is the coefficient of thermal expansion, and E is Young's modulus. Data on α , μ , and E are readily available. However, accurate measurements of κ are hard to find, and consistent measurements of S_T are almost nonexistent.

We measured κ by a thermal-comparator technique, which is only accurate to $\sim 50\%$. We have started designing new equipment to measure this quantity more accurately. The new technique is based on measuring the time decay of a thermal wave that is generated by crossed laser beams.

Fracture strength, S_T , is a function of two parameters: critical-fracture toughness, K_{Ic} , and maximum flaw size, c , in the surface. This function is

$$S_T = Y K_{Ic} c^{-1/2}, \quad (38)$$

where Y is a dimensionless parameter that depends on the location and geometry of the flaw. Because most materials are finished and polished with the same process, the distribution in flaw size will be approximately the same. Therefore, K_{Ic} is a relative measure of strength.

S. Freiman at the National Bureau of Standards has been measuring the critical-fracture toughness, K_{Ic} , of silicate glasses. In 1983, we contracted with him to measure K_{Ic} in phosphate glasses. Table 6-2 lists his first results on four phosphate glasses. The glass with the highest K_{Ic} contains Al_2O_3 , whereas K_{Ic} is little affected by MgO .

We have also looked for correlations among thermomechanical properties and glass compositions. Figure 6-51 shows that thermal conductivity and thermal expansion are strongly correlated. Unlabeled data points are multicomponent glasses. The inverse dependence between the two properties arises from the dependence of both on the anharmonics of glass structural vibrations. Large anharmonics result in low thermal conductivities and high thermal expansions.

Figure 6-52 shows the approximate correlation of Young's modulus and thermal

Table 6-2. Critical-fracture toughness, Young's modulus, and hardness in four phosphate glasses.

Composition (mole %)	Fracture toughness, K_{Ic} (MPa $\cdot$ m ^{1/2})	Young's modulus, E (GPa)	Hardness (GPa)
70P ₂ O ₅ , 20K ₂ O, 10La ₂ O ₃	0.33 ± 0.03	44.2	3.39
75P ₂ O ₅ , 15K ₂ O, 10La ₂ O ₃	0.33 ± 0.06	46.7	3.39
70P ₂ O ₅ , 10K ₂ O, 10MgO, 10Ln ₂ O ₃	0.35 ± 0.08	53.9	3.47
68P ₂ O ₅ , 15Al ₂ O ₃ , 15MgO, 2Ln ₂ O ₃	0.55 ± 0.04	67.2	4.47

Fig. 6-51. Thermal conductivity and thermal expansion are inversely correlated in phosphate glasses. Lines of constant κ/α are shown.

Fig. 6-52. Young's modulus and thermal expansion are inversely correlated in phosphate glasses. A constant αE curve and a straight-line fit to the data, excluding Kigre glasses, are shown.

expansion. Once again, unlabeled data points are multicomponent glasses. The Kigre glasses are in a separate region, perhaps because of their high ZnO content. The product α/E of thermal expansion coefficient and Young's modulus is approximately constant in these limited data. In summary, thermal conductivity and fracture toughness are the two most important thermomechanical properties for a glass in high-average-power lasers.

Author: S. E. Stokowski

Ultraphosphates

In collaboration with Schott Glass Technologies, Inc. and Hoya Optics Inc., we are studying new phosphate laser glasses in the ultraphosphate region. The ultraphosphate region is defined by $r < 1$, where $r = (a + b + 3c)/d$ in the glass composition $a(R_2O) + b(RO) + c(R_2O_3) + d(P_2O_5)$. Metaphosphates are glasses with $r = 1$, and polyphosphates have $r > 1$, as shown in Fig. 6-53. Our goal is to find new laser glasses with long fluorescence lifetimes, low concentration quenching, and good thermomechanical characteristics. We could use such glasses for future fusion-laser systems or high-average-power lasers.

In our past work on phosphate laser glass, we have concentrated on metaphosphate and polyphosphate compositions. For example, the Nova laser glasses, LHG-8 and LG-750, are metaphosphates. Russian researchers, on the other hand, have done extensive work in ultraphosphates, primarily because they observed reduced concentration quenching in some Li-La-Nd ultraphosphates.

Schott and Hoya have supplied us with a total of 200 new compositions, some of which are shown in Fig. 6-54 and in Table 6-3. For each composition, the companies melt at least two samples containing 0.5×10^{20} and 10×10^{20} ions/cm³, respectively. They use La_2O_3 in the glasses to maintain a constant (~ 10 mole %) total lanthanide ($La_2O_3 + Nd_2O_3$) content, which keeps the glass structure relatively independent of Nd content. The companies melt the glasses under conditions that minimize water content (OH⁻), because water quenches the Nd-fluorescence lifetime.

As a result of our studies, two compositional regions stand out. The first group, alkali-lanthanide-phosphates, such as $K_2O-Ln_2O_3-P_2O_5$ or the lithium-sodium equivalent, have reduced concentration quenching. For instance, the fluorescence lifetime is 200 μs at a Nd concentration of 10×10^{20} ions/cm³ (see UP-16 in Fig. 6-46). The second group, Al_2O_3 -containing ultraphosphates, also have reduced concentration quenching, as shown in Fig. 6-55. Although the quenching is not as large as in the first group, the thermomechanical characteristics are better.

In the aluminum-free, alkali-ultraphosphate region, there is a range of compositions in which concentration quenching is minimal, as shown in Fig. 6-56. This compositional range seems to be correlated with the appearance of a stable glass structure, because, at high K_2O/P_2O_5 ratios, the glass-melt crystallizes; at low ratios, the P_2O_5 volatilizes excessively. UP-16 is in the center of this region and makes a good glass. Adding >5 mole % of alkaline earths (Mg, Ca, or Ba) or alumina to UP-16 results in increased concentration quenching. Unfortunately, the thermomechanical characteristics of this glass, as shown in Table 6-4, are not very good because of its high thermal expansion and low thermal conductivity.

To find glasses with better thermomechanical properties, we investigated

ultraphosphates with alumina-alkali ratios of from 0.5 to 2, because earlier Russian work had indicated low concentration quenching in this region. The best composition that we found is UP-91. Figure 6-56 shows fluorescence lifetime vs concentration for both UP-91 and UP-16. The thermomechanical properties of UP-91 are much better than those of UP-16, as shown in Table 6-4.

We have found that adding up to 10 mole % of B_2O_3 can increase the fluorescence lifetime by 10% at high Nd concentrations. Normally, B_2O_3 acts as a strong quencher of the Nd fluorescence because of the high-frequency vibrational modes of the BO_3 group. However, boric oxide in concentrations of <10 mole % in ultraphosphates incorporates itself into the glass structure as BO_4 side groups on the phosphate chains so that BO_3 chain terminators are absent.

Author: S. E. Stokowski

Major Contributor: J. E. Lynch

Alkali substitutions:

- A_1 Li — Na — K
 A_2 Li — Na — K
 A_3 Li — Na — K — Rb — Cs

Alkaline-earth substitutions:

- E_1 Mg — Zn — Ba
 E_2 Mg — Ca — Zn — Ba
 E_3 Mg — Ca — Ba
 E_4 Mg — Ca — Zn — Ba

Table 6-3. Alkali and alkaline-earth substitutions.

Fig. 6-53. Compositional diagram of phosphate glasses.

Fig. 6-54. Compositional diagram of quaternary phosphate glasses melted in the P_2O_5 - K_2O - Al_2O_3 - BaO system. Alkali substitutions are designated by A; alkaline-earth substitutions by E (see Table 6-3); single-glass compositions by data points.

Fig. 6-55. Concentration quenching of an aluminophosphate glass, UP-91, compared to an alkali-phosphate, UP-16.

Fig. 6-56. Nd-fluorescence lifetimes at 13×10^{20} Nd ions/cm³ in K₂O-In₂O₃-P₂O₅ glasses.

	UP-16 (aluminum-free)	UP-91 (14-mole% Al ₂ O ₃)
Thermal expansion, α_{20-300} (10 ⁻⁷ /°C)	127	90
Thermal conductivity, λ (W/m·K)	0.49	0.65
Young's modulus, E (GPa)	45	61
Poisson's ratio, ν	0.270	0.247

Table 6-4. Thermomechanical characteristics of two ultraphosphate glasses.

Broad-Band Energy Extraction

Introduction. Narrow-band energy extraction for long-lifetime silicate laser glasses is ~20% lower than for phosphate glasses. The reduced energy extraction would partially compromise the improved energy-storage capabilities of long-lifetime, low-cost silicate glasses. The lower extraction efficiency is attributed to spectral hole-burning of the inhomogeneously broadened emission line.⁵⁰ Broad-band extraction could reduce the loss in

extraction efficiency to <10% of that of an ideal, homogeneous amplifier.

Saturation data on long-lifetime glasses are, however, quite unreliable. Our uncertainties about the calculated emission cross sections and differences in the time-resolved saturation data make it difficult to assess whether broad-band extraction would be necessary. Fluorescence-line-narrowing (FLN) data⁵¹ show no large differences in the ratio of homogeneous to inhomogeneous line widths, $\Delta \equiv \Delta\nu_{\text{H}}/\Delta\nu_{\text{IH}}$, for phosphate and silicate laser glasses. Other contributing factors, such as off-line extraction, must be considered to explain the observed differences (a factor of ~2) in effective saturation fluences.

The actual differences in extraction efficiency are much smaller than implied by the departure of effective saturation fluence from saturation fluence for a homogeneous medium. For laser-beam fluences that are in the low range, the effective saturation fluence becomes meaningless and can be misleading because its determination from experimental data involves large uncertainties. Relative extraction efficiencies, in comparison to those for a homogeneous medium, provide a more realistic and direct correlation with system performance. Our analysis of amplifier elements represents the performance of amplifiers quite well by applying effective, average beam fluences.

Experimental Observations. Measurements of calorimetric saturation show that silicate laser glasses exhibit lower energy extraction than phosphates. Extraction from the low-cross-section silicate glass, LG-56, appears to be especially poor.⁵⁰ Differences in energy extraction are assumed to be caused by differences in the spectral homogeneity of the glasses. The fluorescence-line shapes support such a qualitative explanation, but spectral homogeneities for LG-56 and LG-750 (a phosphate glass) would have to differ by almost an order of magnitude to give quantitative agreement between experimental data and simulations of energy extraction from a single transition at line center.⁵¹

Our saturation measurements on phosphate laser glasses were primarily made at 1.053 μm , which is near the fluorescence peak; LG-56 data were taken at 1.064 μm , or ~70 Å from peak wavelength. The difference in wavelength can also explain some

of the differences in measured saturation fluences (10 J/cm^2). The saturation fluence for a homogeneous medium is $F_s = h\nu/\sigma \approx 20 \text{ J/cm}^2$ for LG-56, where $h\nu$ is photon energy and σ is the stimulated-emission cross section at the extraction wavelength. Limited data for LG-5 at $1.053 \mu\text{m}$ show a saturation fluence of 12 J/cm^2 at output fluences of 8 J/cm^2 . The corresponding saturation-fluence ratios, F_s/F_a , are 0.58 at $1.053 \mu\text{m}$ and 0.50 at $1.064 \mu\text{m}$.

For Q-88 phosphate glass, a reversed wavelength scaling was observed at an output fluence of 2 J/cm^2 , with $F_s \approx 0.67 I_a$ at $1.064 \mu\text{m}$. The corresponding initial $1.053 \mu\text{m}$ measurements lead to an effective saturation fluence of $0.55 I_a$. Improved energy extraction at the long-wavelength side of the emission line⁵² agrees qualitatively with initial Q-88 data but conflicts with LG-56 data.

Further complications arise from uncertainties in stimulated-emission cross sections, which may vary by up to 20% from the Judd-Ofelt value.⁵³ For energy extraction off line center, additional corrections must be applied. Because LG-56 glass is not identical to LG-650 glass, the use of LG-650 spectroscopic data for LG-56 saturation measurements is questionable.

Other difficulties in LG-56 data arise from differences in the beam profiles. The beam for LG-56 measurements was down-collimated and quite hot in the center; the Q-88 beam was more uniform. Even though saturation is relatively insensitive to spatial beam modulations, it is important to equalize low- and high-fluence beam areas to obtain an equivalent, average fluence. An effective beam aperture of $\sim 4 \text{ mm}$ appears to be more realistic than the 5-mm aperture applied in the evaluation. The change would raise the effective saturation fluences from ~ 10 to 15 J/cm^2 , corresponding to 75% of F_s at $1.064 \mu\text{m}$. Similar discrepancies were observed⁵⁴ after reduction of photometric data. Depolarization losses at the Brewster output face could also reduce energy extraction.

Other discrepancies are found in the small-signal gain of the extraction pulse and the cw probe beam. For LG-56, the pulsed gain at a beam-fluence ratio of $R = F/F_a < 0.02$ is 8% lower than the cw gain. For ED-2, a 6%-lower pulsed gain was observed.⁵⁰ The discrepancy is similar to the

Fig. 6-57. Input and output pulse shapes for LG-56 glass. Pulse sharpening corresponds to an effective saturation fluence of 20 J/cm^2 .

observed 5% reduction in gain coefficient at $R \approx 0.01$ (Ref. 55). This low-level saturation may actually be an induced loss mechanism correlated with the induced losses observed in unpumped glasses.⁵⁰

Uncertainties about saturation in the time-resolved data also remain unexplained. There are actually three different effective saturation fluences obtainable from every saturation measurement. The time-resolved gain reduction of the amplifier leads to a lower effective saturation fluence than the calorimetric data, while data on time-resolved pulse sharpening lead to higher effective saturation fluences. While such observations are qualitatively compatible with spectral and polarization hole burning and with gain or loss recovery, no consistent model that can simulate all the data has yet emerged. For example, the effective saturation fluence of LG-56 was measured at $\sim 10 \text{ J/cm}^2$ from the calorimetric data, while the pulse sharpening data shown in Fig. 6-57 lead to a value of $\sim 20 \text{ J/cm}^2$, which is equal to $h\nu/\sigma$ at $1.064 \mu\text{m}$. Pulse-distortion data for ED-2 glass lead to $\sim 30\%$ larger effective saturation fluence than the calorimetric data.⁵⁰

Simulation of Fluorescence-Line

Shapes. Energy extraction from Nd-laser glasses with 12 transitions depends on the spectral position, strength, and correlation of the individual transitions. The complexity of simulating fluorescence-line shapes can be reduced by cooling the sample to 4 K and eliminating one-half of the transitions. However, full freedom for selecting position, strength, and line width of the individual transitions (31 free variables)^{56,57} was allowed for the simulation of the fluorescence line at room temperature for 3669 A silicate laser glass from American Optical Corp. and for ED-2. With so many degrees of freedom, no unique set of variables can be distinguished from others in reproducing the observed line shapes.

Fig. 6-58. Fluorescence-line shapes of (a) LG-650 and (b) LG-750 simulated with the same Stark splitting ($\Delta\nu_s = 75 \text{ cm}^{-1}$) and two inhomogeneous widths ($\Delta\nu_{IH} = 110$ and 80 cm^{-1} , respectively).

To decrease the number of degrees of freedom and uncertainty, we reduced the number of variables from 31 to 9 and 16 for the 4- and 300-K fluorescence simulations, respectively. We selected an average Stark splitting for the lower laser level and an average line width for the individual transitions. Our approach confined the variables to a relatively narrow range, but was sufficiently flexible to accurately reproduce the measured fluorescence-line shapes, as shown in Fig. 6-58.

The corresponding Stark splittings, line widths, and relative line strengths are summarized in Table 6-5. The Stark splitting of the upper $^4F_{3/2}$ level was obtained from low-temperature $^4I_{9/2} \rightarrow ^4F_{3/2}$ absorption spectra.⁵⁸ The simulations include a Lorentzian homogeneous line width of $\Delta\nu_H = 20 \text{ cm}^{-1}$, while the inhomogeneous broadening was assumed to be Gaussian.

Even though there is some flexibility in selecting a line width, the values are bound by the steepest slopes and the degree of smoothness or modulation of the fluorescence spectra. Wide transitions will not match steep slopes; narrow transitions exhibit too much modulation. The inhomogeneous line width for the three silicates is essentially the same at $\Delta\nu_{IH} \approx 110 \text{ cm}^{-1}$, while the phosphate glass exhibits somewhat narrower lines with $\Delta\nu_{IH} \approx 80 \text{ cm}^{-1}$. The line widths obtained from our simulations are compatible with those reported in Refs. 56, 57, and 58.

Spectral Homogeneity of Laser Glasses.

It is interesting to note that the average Stark splitting of the lower laser level ($\Delta\nu_s \approx 75 \text{ cm}^{-1}$) is essentially identical for all four glasses. The differences in line shapes are primarily due to variations in the line strengths of the individual transitions. The phosphate LG-750 appears to be more homogeneous than the silicate LG-650, as shown in Fig. 6-58. In reality, the homogeneous line widths of the $^4I_{9/2} \rightarrow ^4F_{3/2}$ transitions for the two glasses were measured as 25.3 and 59.5 cm^{-1} , respectively,⁵⁸ which would make the silicate more homogeneous than the phosphate. More recent FLN measurements indicate homogeneous line widths for the $^4I_{11/2} \rightarrow ^4F_{3/2}$ transitions of from 22 to 53 cm^{-1} for LG-750 and 30 to 24 cm^{-1} for LG-650 at 1.05 and 1.06 μm , respectively.⁵⁹ The measurements are in general agreement with hole-burning experiments^{60,61} and exhibit the same wavelength scaling as the saturation data.

Narrow-Band Extraction. Simulation of energy extraction for multitransition, inhomogeneously broadened lines is difficult and uncertain because the positions, widths, strengths, and correlations of the individual transitions are not accurately known. Our simulations were primarily carried out for single transitions. We found

that simulations of an amplifier element were sufficient to describe amplifier performance when we applied effective, average beam fluences.

Energy extraction for a single transition depends on the spectral position of the extracting laser pulse relative to the fluorescence peak, as shown in Fig. 6-59. Curves showing extraction off-line include a spectral offset, $\Delta\nu_{\text{off}}$, of the extracting laser pulse by 30% of the inhomogeneous line width. In our example, a reduction of the effective saturation fluence to $F_s = 0.45 F_{\sigma}$ at a fluence ratio of $R = F/F_{\sigma} = 0.25$ could be explained by either extraction at line center, with $\Delta = 0.2$, or by off-center extraction at $\Delta\nu_{\text{off}}/\nu_{\text{IH}} = 0.3$, with $\Delta = 0.8$. The primary difference in the two explanations is that off-center extraction could lead to a much faster increase in effective saturation fluence with beam fluence, as shown in Fig. 6-59(a).

The low effective saturation fluence of $0.45 F_{\sigma}$ in Fig. 6-59(a) corresponds to an actual loss in extraction efficiency of only 13%, as shown in Fig. 6-59(b). Therefore, a comparison of laser glasses on the basis of effective saturation fluence could be misleading.

The emission line of actual laser glasses consists of 12 overlapping lines, and the narrow-band extraction pulse interacts with several transitions at their wings. Off-center extraction can also reduce the effective saturation fluence at low beam fluences. Knowledge of the line-width ratio alone is insufficient to completely characterize energy extraction. It is also necessary to know the spectral position at which the extracting

Table 6-5. ${}^4F_{3/2} - {}^4I_{11/2}$ transitions of Nd^{3+} laser glasses.

Transition	ED-2	3669 A	LG-650	LG-750
Stark splitting and line widths (cm^{-1})				
$\Delta\nu_s {}^4F_{3/2}$	160	143	139	114
$\Delta\nu_s {}^4I_{11/2}$	75	75	75	75
$\Delta\nu_{\text{IH}} (\text{cm}^{-1})$	110	110	110	80
Relative line strengths ($a-a = 100$)				
a-a	100	100	100	100
a-b	60	42	50	56
a-c	30	12	15	34
a-d	28	5	10	18
a-e	23	27	17	10
a-f	10	21	31	4
b-a	35	50	40	36
b-b	45	100	60	55
b-c	15	0	10	20
b-d	0	5	15	28
b-e	40	55	35	14
b-f	0	21	31	10

Fig. 6-59. (a) Calculated effective saturation fluence for off-line extraction is lower than for extraction at line center. (b) Corresponding deviations in extraction efficiencies are less severe, especially at low beam fluences.

laser pulse intercepts the various transitions. Since it is difficult to determine the location of individual transitions, it may not be possible to characterize energy extraction with spectroscopic data alone. Such data will help to improve our understanding of saturation but will probably lack the precision required for simulation of megajoule lasers.

Broad-Band Extraction. Broad-band extraction is only effective for relatively inhomogeneous media with $\Delta < 0.5$. Figure 6-60(a) shows the relative extraction efficiencies, η/η_H , vs line-width ratio, $\Delta\nu_L/\Delta\nu_{IH}$, of a spectrally flat extraction pulse (width = $\Delta\nu_L$) for various spectral

homogeneities of the laser medium and for various normalized beam fluences. For a homogeneous system, the relative extraction efficiencies decrease monotonically because portions of the extracting pulse are spread out over the wings of the emission line, where extraction is less effective.

For relatively inhomogeneous media, the relative extraction efficiencies initially increase with spectral broadening of the extraction pulse, reach a maximum, and then decrease for very broad extraction pulses. The optimal extraction line width depends on the beam fluence and on the homogeneity of the medium. From such an

Fig. 6-60. Broad-band extraction improves the extraction efficiency of inhomogeneous laser media for (a) extraction on line center and (b) extraction off line center.

Fig. 6-61. Broad-band extraction yields more pronounced changes in (a) ion-site distribution than in (b) observable gain distribution. (c) Off-line extraction accentuates the inhomogeneous character of a glass.

analysis, it appears to be possible to maintain the relative extraction efficiency for an inhomogeneous medium with $\Delta \approx 0.1$ to well within 20% of the relative extraction efficiency of a homogeneous medium by proper selection of the line width of the extracting laser pulse.

Figure 6-61 shows examples of broad-band extraction at line center for various homogeneities of the medium and for various beam fluences. Figure 6-61(a) shows the ion-site distribution, which was assumed to be initially Gaussian. Figure 6-61(b) shows the corresponding spectral-gain distribution after the pulse was transmitted through an amplifier element. Spectral hole burning in the gain profile becomes less obvious for homogeneities with $\Delta > 0.5$. The change in logarithmic gain from an initial Voigt profile gives some indication of the spectral pulse distortion for a spectrally uniform input pulse.

Examples of broad-band extraction with an offset from line center by $\Delta\nu_0 = 0.25 \Delta\nu_{\text{H}}$ are shown in Fig. 6-61(c) for the same conditions. Here, it is easier to observe spectral hole burning even for

relatively homogeneous materials with $\Delta \approx 1$ because the gain profiles are asymmetric.

The corresponding relative extraction efficiencies for various line-width ratios, $\Delta\nu_l/\Delta\nu_{\text{H}}$, are shown in Fig. 6-60(b). Again, the optimal line-width ratio depends on the beam fluence and homogeneity of the medium. Obviously, broader extraction pulses are required for maximum extraction efficiency off line center.

Optimal Extraction for a Single Transition. Potential improvements in extraction efficiency, η , for multiline and broad-band sources relative to narrow-band extraction, η_N , for a single transition are shown in Fig. 6-62. Figure 6-62(a) shows the results for symmetric, dual-line extraction, while Fig. 6-62(b) shows three symmetrically spaced and equally strong lines. Figures 6-62(c) and (d) correspond to symmetric and off-center broad-band extraction, respectively, with a spectrally uniform intensity distribution. It is obvious that broad-band extraction is superior to multiline extraction. Dual-line extraction is not very effective for a single transition but may be quite sensible for laser glasses, if the two

Fig. 6-62. Multiline extraction is less effective than broad-band extraction for a single transition.

Fig. 6-63. (a) Optimal extraction-line width depends on beam fluence and inhomogeneity of the laser medium. (b) Extraction efficiency can be maintained close to that of a homogeneous system by broad-band extraction.

extraction lines are matched to the centers of two transitions.

For extraction off line center, even larger improvements are available by broadening the laser pulse, as shown in Fig. 6-62(d). Figure 6-63 shows optimal extraction-line-width ratios, Δ_L^* , and maximum extraction efficiencies, η^* for extraction on line center. (Here, $\eta_H = 1 - e^{-R}$ is the extraction efficiency of a homogeneous medium.) The relative extraction efficiency, η^*/η_H , exhibits a minimum at a relative beam fluence of $R \approx 1.1$. For example, the extraction efficiency of a relatively inhomogeneous medium with $\Delta = 0.2$ can be maintained within 15% of that of a homogeneous medium by extraction-line broadening. The optimal extraction-line width for the worst condition at $R = 1.1$ would be $\Delta_L^* \approx 0.9 \Delta_{H1/2}$ in this case. For an actual amplifier, the deviation from η_H would be somewhat greater because the optimal extraction-line width cannot be maintained simultaneously for both input and output fluence.

In general, the optimal extraction-line width, Δ_L^* , for a single transition is about equal to the inhomogeneous line width, $\Delta_{H1/2}$, for average fluence ratios of $R \approx 1$, which corresponds to typical operating conditions of cost-optimal fusion lasers.⁶²

Energy Extraction for Two Overlapping Transitions. To assess energy extraction for several transitions, we simulated the extraction of two equally strong transitions for various extraction-line widths, Stark splittings, ratios of homogeneous to inhomogeneous line width, and beam fluences. To restrict the number of variables, we centered the extraction pulse between the two transitions.

The correlation of individual transitions at various ion sites was an important variable. Three conditions are possible. Narrow-band extraction at the wing of the first transition may produce a spectral hole in the second transition at the same wing (correlated condition), at the opposite wing (anticorrelated condition), or may produce no hole at all (uncorrelated condition). Although phosphate glasses exhibit a lack of correlation between Stark components, silicate glasses do not show a similarly strong extraction pedestal, but are instead dominated by narrow-band hole burning.⁶¹

Examples of broad-band extraction for correlated Stark components are shown in Fig. 6-64. The spacing of the two equally strong transitions, $\Delta_s = \Delta_{\nu_s}/\Delta_{H1/2}$, increases from 0.5 to 0.75 to 1.0 in Figs. 6-64(a), (b), and (c), respectively. The laser line width

Fig. 6-64. Correlated transitions exhibit extraction side bands, leading to more effective overall extraction.

was held constant at $\Delta_1 = 0.5$. For correlated transitions, secondary spectral holes that develop in the site distributions are symmetrically spaced relative to the extraction pulse. Extraction with correlated transitions is more effective than for uncorrelated or antio. related transitions. For correlated transitions, most ion sites can be accessed with the laser line width equal to the Stark splitting, even for extremely inhomogeneous laser media.

The improvements in extraction efficiency with increasing extraction-line width are shown in Fig. 6-65. The spacings of the two transitions are the same as in Fig. 6-64. In most cases, the optimal extraction-line width is equal to the Stark splitting, $\Delta\nu_s$, of the two transitions. Broader extraction lines become less effective because most ion sites were already accessed at their peak cross section. Our examples of correlated conditions indicate that optimal broad-band lasers could maintain extraction within 10% of that of a homogeneous medium even for

a relatively inhomogeneous medium, with $\Delta = 0.2$.

Our difficulty is that we have no firm data on the degree of correlation or the relative strength and spectral position of the individual transitions. These uncertainties make it difficult to pinpoint the optimal extraction-line width; however, it should be in the range of the inhomogeneous line width or the Stark splitting, whichever is less. A line width of ~ 50 Å should be sufficient to obtain most of the potential improvements. Such a line width can be obtained with Q-switched glass oscillators.

LG-56 Saturation. The measured extraction efficiencies for LG-56 glass relative to a homogeneous medium are shown in Fig. 6-66. The data indicate about a 25% reduction in extraction efficiency at an output-fluence ratio of 0.58. However, the effective beam aperture for these experiments was probably closer to 4 mm than to the 5 mm actually used. Such a correction would move the data toward higher

Fig. 6-65. Broad-band extraction from two correlated transitions is more effective than for a single transition (compare Fig. 6-60).

Fig. 6-66. Potentially poor narrow-band extraction for LG-56 glass can be reduced to <10% by broad-band extraction.

effective fluences and closer to those from a homogeneous medium, as shown by the arrow in Fig. 6-66 (the arrow also serves as an error bar). Estimates for narrow- and broad-band extraction are also included for an amplifier element operating at an average fluence $F = (F_{in} + F_{out})/2$. Calculations for

a full amplifier lead to extraction that is lower by a few percent. Such estimates suggest that it should be possible to maintain extraction efficiency to within 10% of that of a homogeneous medium using a broad-band source.

Author: W. F. Hagen

Comparison of Laser Glasses

Introduction. The relative performance of different laser glasses depends on laser-pulse duration, pulse shape, system architecture, and amplifier geometry. Thus, it is difficult to establish a simple figure of merit to develop new laser glasses. Instead, elaborate system optimizations are required to determine the projected gains and costs of amplifiers with different sizes. The projections are then used for different system architectures to determine the laser performance for various pulse durations and pulse shapes.

We can introduce simplifications if the laser is designed for either high-power or high-energy applications. The next generation of fusion lasers will primarily consist of high-energy designs, for which nonlinear effects become less important although they cannot be ignored. In a long-pulse (10-ns) regime, the amplifier can be optimized for maximum energy storage per unit cost, while system performance is dominated by extraction efficiency. Multipass systems are, in general, more effective in extracting stored energy because, for the same output-fluence limitations, the total effective fluence in a multipass system is much higher than for a single-pass system.

The optimization procedure can be further simplified by evaluating the most costly output stage of a system and appropriately scaling the driver stages. Systematic evaluation of high-cross-section phosphate glasses (LG-750 or LHG-8) with $\sigma = 4 \text{ pm}^2$ and a low-cross-section silicate glass (LG-650) with $\sigma = 1.1 \text{ pm}^2$ shows that multipass systems could be about twice as cost-effective as single-pass systems. The long-lifetime silicate is $\sim 50\%$ more cost-effective at 10 ns than the phosphate glasses for a fluence limit of 30 J/cm^2 . Other laser glasses lie between these extreme cases.

Amplifier Simulations. We used the 46-cm Nova amplifier for our comparisons of laser glasses. Gain data for LG-750 can be applied for code normalization and simulation of LG-650. The primary differences between the two glasses are in their stimulated-emission cross sections and associated fluorescence lifetimes. The fluorescence lifetime for 3% Nd-doped LG-750 is $364 \mu\text{s}$, and the inversion efficiency for 44.6-mm-thick disks is $\sim 8.7\%$. In contrast, cost-optimal LG-650 disks would be $\sim 35 \text{ mm}$ thick, Nd-doped at 6% with a lifetime of $\sim 690 \mu\text{s}$, and would have an inversion efficiency of 9.6%. Because of the longer fluorescence lifetime, the optimal pump-pulse duration for LG-650 is $\sim 1.7\times$ longer than for LG-750.

Cost scaling for the 46-cm Nova amplifier can be approximated by

$$C_{46} \approx 90\text{k\$} + 21 \frac{\text{k\$}}{\text{cm}} t + 0.15 \frac{\text{\$}}{\text{J}} E_B \\ \approx 195\text{k\$} + 0.15 \frac{\text{\$}}{\text{J}} E_B, \quad (39)$$

where t is disk thickness and E_B is bank energy. With this scaling, the highest cost effectiveness for LG-750 glass is obtained at a gain coefficient of $\sim 5.8 \text{ m}^{-1}$ and amounts to an energy storage of 8.3 kJ at a cost of $\$270\text{k}$ or $70 \text{ J/k\$}$. In contrast, the highest cost effectiveness for LG-650 glass is obtained at $\sim 4.5 \text{ m}^{-1}$, which amounts to an energy storage of 17.8 kJ at a cost of $\$255\text{k}$ or $70 \text{ J/k\$}$, which is more than twice the LG-750 value assuming the same glass cost ($\sim \$3/\text{cm}^3$). If we include the lower glass cost of $\$0.5/\text{cm}^3$ for LG-650, then the cost effectiveness of LG-650 would increase to $93 \text{ J/k\$}$, which is about three times higher than for LG-750. This is a high-leverage item even if the extraction efficiency is sacrificed to some extent. Refinements of our simulations with respect to ASE, radiation trapping, and the effective fluorescence lifetime due to nonlinear decay will not significantly alter our conclusions.

Energy Storage. Previous simulations of various laser glasses show that the cost effectiveness of storing energy scales quite well with the inverse square root of the stimulated-emission cross section, σ , and can be approximated⁶³ by

$$Z_s = \frac{E_s}{C_A} \approx 55 \frac{\text{J} \cdot \text{pm}}{\text{k\$}} \sigma^{-0.5}, \quad (40)$$

where E_s is stored energy and C_A is cost of the amplifier. However, energy-storage density does not scale in this manner. The optimal gain coefficients are almost independent of the cross section due to ASE losses and parasitic limitations; therefore, we require relatively thick disks for high-cross-section materials. A $\sigma^{0.15}$ scaling of the gain coefficient and a $\sigma^{0.35}$ scaling of the disk thickness is quite typical for segment sizes of about 30 cm , leading to a logarithmic gain per isocost disk of

$$g_o \approx 0.15 \text{ pm}^{-1} \sqrt{\sigma}. \quad (41)$$

In other words, large disks made from low-cross-section materials are, in general, more highly doped and are pumped at higher energies with longer pump-pulse durations than the high-cross-section disks.

The simple scaling of Eq. (40) is more conservative than the 46-cm amplifier simulations indicate, since the relative

Fig. 6-67. Measured and calculated extraction efficiency for ED-2 laser glass relative to a perfect laser medium.

improvement in cost effectiveness is only $\sqrt{4/1.1} \approx 1.9$ rather than 3 for LG-650 over LG-750. In addition, low-cross-section materials reduce pulse distortions, ASE, and isolation requirements. Such requirements have not yet been included in our analysis.

Energy Extraction. Saturation measurements indicate better extraction characteristics for phosphates than for silicates. The use of a dual-ion saturation model or a spectral-hole-burning simulation is required for accurate system evaluations. An example of saturation data normalized to output fluences is shown in Fig. 6-67. A loss of $\sim 10\%$ in extraction efficiency relative to a homogeneous medium is shown. Simulation of an inhomogeneous medium matches the experimental data quite well. For LG-650, the departure from ideal extraction may be as large as 25%, but the data are not reliable, as discussed in the previous article, "Broad-Band Energy Extraction." A more likely departure from perfect extraction for LG-650 is in the range of 15 to 20% for narrow-band lasers and $<10\%$ for broad-band or multiline extraction. For simplicity, energy extraction was calculated as a fixed fraction (75%) of ideal extraction for the worst-case simulation of LG-650.

System Evaluations. The simulation of laser systems can be accomplished quite well by evaluating the cost effectiveness⁶³

$$Z = \frac{fF_o}{c_A + c_f} \left(1 - \frac{k}{G_s} \right) \approx f\eta_s \frac{1 - 0.25/G_{SF}^{0.3}}{1 + \Delta} \quad (42)$$

where f is the fill factor, F_o the output fluence, c_A the area-normalized amplifier

cost, c_f the fixed cost, k the drive-cost factor, G_{SF} the saturated-fluence gain, η the extraction efficiency, and $\Delta = c_f/c_A$ the fixed-cost fraction. The area-normalized amplifier cost can be approximated by

$$c_A \approx \frac{nh\nu \ln G_o}{\sigma Z_s} \approx 5 \times 10^{-10} \frac{\$}{\text{cm}} \ln \frac{G_o}{\nu \sigma} \quad (43)$$

where G_o is the small-signal gain of the amplifier.

Nonlinear effects for fusion pulses were approximated⁶⁴ by

$$\Delta B \approx \frac{2\pi n_2 I_o (o - i)}{\lambda n^2 \alpha (1 - e^{-\alpha})} \quad (44)$$

where $I_o = F_o/\tau$ is the peak output-beam intensity at the end of the pulse of duration τ , o and i are the lossless output- and input-fluence ratios, respectively, and α is the gain coefficient before the pulse enters the amplifier. For multipass systems, the average gain coefficient along the amplifier is reduced to $\alpha = \alpha_0 \exp(i_0 - i)$ due to previous passes, where α_0 is the initial gain coefficient and $i_0 = F_i/nF_s$ is the input-fluence ratio for the first pass.

The equations above are sufficient to evaluate single-pass and multipass systems in general terms for various glasses and constraints. We evaluated the cost effectiveness of 1- and 5-pass systems for LG-750 and LG-650 under the following conditions: a fill factor of $f = 75\%$, a loss coefficient of 0.2 m^{-1} , a cavity transmission of 90%, a fixed cost of $c_f = \$1 \text{ million/m}^2$, a ΔB limit of 3, and a fluence limit of 30 J/cm^2 . LG-750 phosphate glass was evaluated for perfect extraction; LG-650 silicate glass was evaluated with an effective extraction efficiency of 0.75 and $0.9 \eta_H$ for narrow- and

broad-band extraction, respectively. For our examples, more conservative gain coefficients of 4.7 and 3.7 m^{-1} were applied for the phosphate and silicate glasses, respectively.

Figure 6-68(a) shows the results for narrow-band extraction from various gain stages with a ΔB limit of 3 but with no fluence limits. These examples correspond to isofluence system designs, which are valid for nominal Gaussian pulse shapes. However, isofluence designs would lead to very high intensities in the driver stages for fusion pulses, where peak intensities are required at the end of the pulse. Hence, isofluence designs are not a sensible option.

Maintaining the intensity in the driver stages at or below that in the output stages leads to isointensity designs with

appropriately larger and more costly driver stages than for isofluence designs. The results for such isointensity designs, shown in Fig. 6-68(b), lead to a significant reduction in the cost effectiveness of 1-pass systems from ~ 10 to $6 \text{ J/\$}$ at 10 ns for phosphate glass. For 5-pass systems, the difference in cost effectiveness between isofluence and isointensity designs is very small because the cost of the driver becomes insignificant (this is not the case for 1-pass systems). Hence, the cost effectiveness of multipass systems is potentially about twice as high as for a MOPA chain.

With 75% of perfect extraction, which can be considered a worst case, and assuming equal glass costs, the silicate would outperform the phosphate for pulse durations

Fig. 6-68. Cost effectiveness of 1- and 5-pass systems. (a) Nominal Gaussian pulse shapes leading to isofluence designs. (b) Isointensity designs required for fusion pulse shapes. (c) Broad-band extraction. (d) Output-fluence limit = 30 J/cm^2 .

above 2 ns. In reality, the glass costs are projected to be much lower for the silicates, and this would accentuate the difference in cost effectiveness even more.

In contrast, the silicate has potential for further improvements by broad-band or multiline extraction, as shown in Fig. 6-68(d). In this case, the silicate would be almost twice as cost-effective as the phosphate for a 10-ns laser. Again, multipass systems are about twice as cost-effective as MOPA chains.

Figure 6-68(d) shows the results for a fluence limit of 30 J/cm^2 at 10 ns scaled with the square root of the pulse duration. In this case, the silicate glass is primarily fluence-limited, while the phosphate glass is still limited by ΔB . The potential improvements of multipass systems over MOPA chains by a factor of 2 or more will not be completely realized because of higher fixed costs and higher losses and vignetting losses for off-axis systems. However, at least a 50% improvement can be realized.

Author: W. F. Hagen

Frequency-Conversion Systems

Damage in Conversion Cells

Because our initial strategy for assembling large-area, damage-resistant conversion cells was unsuccessful, we completed further studies during 1983. Our original design incorporated 10- μm -thick layers of a halocarbon index-matching liquid, HC-56. The liquid was chosen because it had small Raman-scattering coefficients. Use of 10- μm -thick liquid gaps allowed for a simple mechanical design for large crystal arrays.

Studies of damage in thin layers of HC-56 showed that bubbles and black residues were produced by laser heating of contaminants that were either in the liquid or on the surfaces of the windows and crystals.⁶⁵ In small, carefully cleaned cells, the liquid layers were capable of withstanding 5 J/cm^2 delivered in 1-ns, 1.06- μm (1ω) pulses. Because of this demonstration, and because use of both HC-56 and the thin-gap design were highly desirable, we

attempted to sufficiently clean the components for second-harmonic (2ω) cells on Novette to prevent damage. At 1ω input fluences below 2 J/cm^2 , black residues began accumulating around a few, small absorptive areas in the liquid gaps and grew with successive Novette shots until obscuration in the cells was unacceptable.

To find a better design, we retested the 5-cm-aperture 2ω cells. Even for carefully cleaned cells irradiated by a collimated, 20-mm-diam, 1ω beam, we found that black areas in the thin liquid gaps could be induced on at least some test sites by a series of 5 to 25 irradiations at 4 to 7 J/cm^2 . Once a small but visible black site was induced, it grew during 5 to 10 additional shots until it covered the entire beam.

We did not find a correlation between the 2ω light itself and damage. In phase-matched cells, damage sometimes occurred in the liquid at the input surface of the KDP crystal, which experienced only 1ω light, and sometimes at the output face, which experienced both 1ω and 2ω light. In fact, the only test surface that remained undamaged during multishot testing was an exit surface. We have drawn four conclusions from our studies:

- Damage to the HC-56 layers in 2ω cells was caused by laser heating of contaminants, as initially believed.
- Since all common liquids have boiling temperatures that are low relative to temperatures easily reached by laser heating, the problem cannot be solved by use of an alternate liquid.
- It is unrealistic to assume that large arrays can be cleaned sufficiently to prevent damage; therefore, the cells must be redesigned to be tolerant of damage.
- Tests supporting redesign should involve cells with thicker liquid gaps that are deliberately salted with contaminants that would probably be present in large cells.

We then conducted a second series of tests using 2ω cells with liquid gaps having thicknesses of either 0.5 or 1 mm. The liquids were contaminated with microscopic fragments of the metals and elastomers from which cells are assembled. Provision was made to flow liquid through the cell and to filter it, should damage residues begin to accumulate. The cells were tested in phase-matched orientation using repeated irradiations by 1-ns, 1ω pulses at fluences of

from 4 to 7 J/cm². As in previous tests with 10- μ m liquid gaps, irradiation of contaminants in cells with thicker gaps produced bubbles and microscopic black particles. However, both bubbles and particles streamed upward. Transverse spread of opaque black areas, like those that occurred in the Novette 2 ω cells, was never observed in a test of a cell with gap thicknesses of 0.5 or 1 mm.

To assure ourselves that the damage problem was not peculiar to HC-56, we also tested 2 ω cells filled with propylene carbonate or with a saturated solution of water and KDP. Some damage occurred in these liquids, although the formation of visible black residues was not as apparent as it had been in HC-56. We concluded that some liquid decomposition will occur in 2 ω cells used at Nova fluences (up to 10 J/cm² for 1 ω input), but that cells with thick liquid layers should recover to a useful state between shots.

In a third series of tests, we evaluated Nova prototype 0.35 μ m (3 ω) monolithic cells with 10-mm-thick doubling and tripling crystals. For this series, we used 0.5- or 1-mm gaps of either propylene carbonate or HC-56. In addition to liquid breakdown caused by contaminants, for cells filled with HC-56 we observed formation of a spatially uniform fogging on the output 3 ω interface. With 1 ω inputs of 5 to 7 J/cm², which produced 3 ω fluences in excess of 2 J/cm², the fogging was apparent after 10 to 15 shots. The fogging was not visible after up to 40 shots at 1 ω inputs of 1 to 2 J/cm² and did not occur at inputs of from 4 to 7 J/cm² in cells not phase-matched for 3 ω . Therefore, we believe we were observing the photochemical decomposition previously reported by the University of Rochester.⁶⁶ Researchers there found that use of a hydrocarbon, Koolase,⁶⁷ prevented the fogging. Our results support their findings: fogging has not been observed in cells filled with either distilled Koolase or propylene carbonate. However, Koolase was rejected in the initial design studies for Nova because its Raman-scattering cross section was relatively large. Resolution of this problem is discussed in the next article.

Finally, some KDP crystals suffer internal damage at Nova design fluences of up to 10 J/cm² at 1 ω and up to 5 J/cm² at 3 ω . We have performed both 1-on-1 and *n*-on-1

Fig. 6-69. Histograms of 1-on-1 and *n*-on-1 bulk laser-damage thresholds of KDP crystals tested with 0.6-ns, 0.35- μ m pulses.

damage tests at 3 ω on KDP samples from Nova production boules. The results are shown in Fig. 6-69. The median 1-on-1 bulk-damage threshold was 2.5 J/cm². Laser hardening was observed in *n*-on-1 3 ω tests; the median *n*-on-1 threshold was 4.5 J/cm². Therefore, significant damage to the KDP crystals in Nova should occur only when the laser is operated at maximum output.

Authors: D. Milam, C. L. Vercimak, and R. A. McGee

Major Contributors: S. E. Peluso, B. C. Johnson, M. A. Summers, and J. D. Williams

Spontaneous and Stimulated Raman Scattering in KDP and Index-Matching Fluids

Introduction. Transverse stimulated Raman scattering (SRS) within KDP crystal segments and index-matching fluids (IMF) is a potential limitation on the harmonic-array design and performance of Nova.⁶⁸ Physical damage to the array supporting structure from intense sidescattering and energy loss through the SRS process are the main issues.

In early 1983, we developed an overall picture of anisotropic Raman-scattering behavior of KDP from a large set of spontaneous- and stimulated-Raman-scattering measurements.⁶⁹ We made absolute SRS-gain measurements in KDP harmonic-array segments and arrived at

values for the transverse-gain coefficients. We also made absolute SRS-gain measurements in benzene liquid. Measurements of spontaneous scattering in an IMF relative to benzene allowed us to accurately estimate SRS-gain coefficients for various fluids.

No large laser system ($>100\text{ J}$) was available for our work. The Cyclops laser generates insufficient energy at 526 nm (1 ns) for an SRS-gain measurement in the transversely pumped configuration of the crystal array shown in Fig. 6-70. The Cyclops laser, however, is quite adequate for making quality SRS-gain measurements with collinear pump and probe beams in the longitudinal configuration. Thus, our strategy was to:

- Determine the relative strengths and depolarizations of Raman scattering in a simple Z-cut crystal using spontaneous scattering.
- Make predictions from the step above, and then measure the spontaneous Raman behavior in a Type II ($\theta = 59\text{ deg}$) crystal.
- Make absolute stimulated-gain measurements on Cyclops in several crystal orientations and compare them with the spontaneous-scattering model.
- Extract the transverse-gain coefficients relevant to Nova and Novette from the above sets of data.

Physics of Raman Scattering in KDP

We determined that the Raman line of greatest potential SRS gain in KDP is the A_1 vibrational line at 915 cm^{-1} . This line is due

predominantly to a symmetric "breathing" vibration of the PO_4 subgroups, with some coupling to the $\text{K}^+(\text{H}_2\text{PO}_4)^-$ lattice through hydrogen bonding.⁷⁰ Theoretically, Raman scattering in a uniaxial crystal from an A_1 vibration is described by a diagonal scattering tensor:

$$\mathbf{S}_{ij} = \begin{pmatrix} a & 0 & 0 \\ 0 & a & 0 \\ 0 & 0 & b \end{pmatrix} \quad (45)$$

Consequently, an A_1 vibration cannot alter the polarization state of Raman-scattered light when it is linearly polarized along the optic (crystal z) axis or polarized in the $x-y$ crystal plane. That is to say, there is no anisotropic or depolarized scattering. The polarization state of the scattered light can be altered or depolarized, however, for scattering directions or polarizations outside the $x-y$ plane.

Unfortunately, we found the spontaneous Raman activity of the 915-cm^{-1} vibration of KDP to be more complex than would be calculated from the diagonal scattering tensor. We required a tensor of the form

$$\mathbf{S}_{ij} = \begin{pmatrix} a & c & d \\ c & a & e \\ d & e & b \end{pmatrix} \quad (46)$$

Fig. 6-70. Geometry for Raman scattering.

with nonzero off-diagonal elements to describe our measurements. Each element appeared to depend on the directions of the incident and scattered light to the crystal axes.

We proposed two physical explanations for the inadequacy of a diagonal-tensor description: first, the PO_4 tetrahedra are not oriented squarely in the KDP unit cell; second, the PO_4 cage is distorted from perfect tetrahedral symmetry by hydrogen bonding. The effect is a mixing of the nonpolar A_1 vibration with the polar B_2 and E vibrational modes of the $\text{K}^+(\text{H}_2\text{PO}_4)$ lattice, thereby adding nonzero off-diagonal elements to the scattering tensor. Since the polar vibrations are known to give angularly dependent scattering, the 915-cm^{-1} vibration acquires some of this character and shows an anomalous scattering behavior.

Experiments on Spontaneous Raman Scattering. To characterize the 915-cm^{-1} vibration, given the above complexity, we measured the spontaneous Raman scattering and depolarization ratios for a set of 25 nonequivalent orientations in a Z-cut KDP crystal. Our work was done on a conventional double-monochromator Raman spectrometer using a cw argon-ion laser at 488 nm as the pump source. The collection optics were arranged at 90° to the pump axis so that only sidescattered Raman light was measured.

The crystal was entirely submersed in a glass tank containing IMF to eliminate surface scattering and stray pump light at the sample position. To eliminate polarization bias in the Raman spectrometer, a high-quality calcite polarizer followed by a crystal quartz polarization scrambler was used at the monochromator entrance slit.

For the simplest case of spontaneous scattering in a Z-cut crystal, when the polarization and propagation directions of pump and Stokes light are along the crystal axes, the scattered Stokes intensity is proportional to the square of the relevant tensor element in Eq. (46). We concluded from the Z-cut crystal measurements that the diagonal tensor elements a^2 and b^2 are the largest, making the scattering strength for polarized scattering stronger than that for depolarized scattering. This is as expected for an A_1 -symmetry vibration. The diagonal measurements were all equal to within $\sim 25\%$.

It was assumed at the outset of our work that c^2 , d^2 , and e^2 would be zero. However,

from scattering in the x - y and y - z planes, we found the element c^2 to be very close to $1/3$ as strong as a^2 . We found d^2 and e^2 to be small, approximately $1/16$ as large as a^2 or b^2 . The surprise in our spontaneous data was that we found c^2 to be angle dependent, vanishing for scattering in the x - y plane.

We have incorporated the most important of the above features into a simple model for the first-order Raman activity of KDP. We introduced the $\mathbf{q}$ vector of the Raman vibration, defined as the difference between the $\mathbf{k}$ vectors of the pump and scattered light waves, $\mathbf{k}_p - \mathbf{k}_s$. For scattering in the x - y plane, $\mathbf{q}$ is perpendicular to the optic axis of the crystal. By defining Ψ as the polar angle between $\mathbf{q}$ and the z axis of the crystal, we arrive at an approximate form for the Raman tensor,

$$\mathbf{S}_{ij} = \begin{pmatrix} a & C(\Psi) & 0 \\ C(\Psi) & a & 0 \\ 0 & 0 & a \end{pmatrix}, \quad (47)$$

where

$$C(\Psi) = a(2/3)^{1/2} |\cos \Psi|. \quad (48)$$

With the above information, we had a first-order model of spontaneous Raman scattering for our Z-cut KDP crystal. We then used the model to predict what we should see for a Type II KDP crystal.

In Fig. 6-71 we show the labels for axes of a Y-cut, Type II KDP crystal. The crystal-line axes are x , y , and z . Unit vectors along the sample edges (laboratory frame) are α , β , and γ . For spontaneous scattering, the pump propagates along γ , and $\theta = 59^\circ$ is the phase-matching angle to the optic (z) crystal axis. For keeping track of the

Fig. 6-71. Labels for axes in a Y-cut, Type II KDP crystal and for the laboratory frame.

Table 6-6. Spontaneous Raman scattering strengths for Type II KDP ($\theta = 59^\circ$ deg).

Frequency-Conversion Systems

Scattering configuration	Tensor coupling	Approximate tensor coupling	Ψ (deg)	Spontaneous scattering strength ^a	
				Predicted	Measured ^b
$\gamma(\alpha\alpha)\beta$	a^2	a^2	102.6	80.0	80.0
$\gamma(\alpha\gamma)\beta$	$[c(\Psi) \sin \theta + d \cos \theta]^2$	$[c(\Psi) \sin \theta]^2$	102.6	1.9	12.9
$\gamma(\beta\alpha)\beta$	$[c(\Psi) \cos \theta - d \sin \theta]^2$	$[c(\Psi) \cos \theta]^2$	102.6	0.7	2.6 ^c
$\gamma(\beta\gamma)\beta$	$[1/2 (a - b) \sin 2\theta + c \cos 2\theta]^2$	0^d	102.6	0^d	3.4 ^c
$\gamma(\beta\beta)\alpha$	$[a \cos^2 \theta - b \sin^2 \theta - c \sin 2\theta]^2$	a^2	68.1	80.0	6.6
$\gamma(\beta\gamma)\alpha$	$[1/2 (a - b) \sin 2\theta + c \cos 2\theta]^2$	0^d	68.1	0^d	36.1
$\gamma(\alpha\beta)\alpha$	$[c(\Psi) \cos \theta - d \sin \theta]^2$	$[c(\Psi) \cos \theta]^2$	68.1	2.0	4.4 ^c
$\gamma(\alpha\gamma)\alpha$	$[c(\Psi) \sin \theta + d \cos \theta]^2$	$[c(\Psi) \sin \theta]^2$	68.1	5.5	4.6 ^c

^aArbitrary units.

^bPrecision, $\pm 5\%$.

^cPrecision, $\pm 30\%$.

^dValue is proportional to $a - b$, which is approximately zero.

numerous scattering configurations; it is convenient to use Porto notation, A(BC)D. Here, A, B, C, and D stand for unit vectors parallel to the propagation direction A of the input beam, electric vector B of the input beam, electric vector C of the Raman scattered light, and propagation direction D of the Raman scattered light. A bar over a symbol denotes a reverse direction.

Comparisons of the predicted and measured spontaneous-scattering strengths for a Y-cut, Type II KDP crystal are shown in Table 6-6. For each scattering configuration, we list the form of the tensor coupling, which is derived by transforming the applied electric-field components into the crystal-axis frame, performing the Raman tensor multiplication, and then transforming back to the laboratory frame. The approximate tensor coupling, based on Eq. (47) and the $\mathbf{q}$ -vector angle, Ψ , are also listed.

Agreement between the predicted and measured values for spontaneous scattering in the γ - β plane is reasonable in trend and magnitude. For scattering in the γ - α plane, however, serious discrepancies arise for configurations $\gamma(\beta\beta)\alpha$ and $\gamma(\beta\gamma)\alpha$. The totals, shown next to braces in Table 6-6, are, however, in approximate agreement. Thus, the total scattering strength for the pump field polarized along both β and α agrees with prediction. The large discrepancy in depolarization for a pump polarized along β indicates an angular variation of tensor elements (probably a and b) that has not yet been taken into account. For practical considerations, we may expect to see similar averaging in the measured gain of the polarized and depolarized gain coefficients.

Experiments on Stimulated Raman

Gain. Our absolute measurements of the stimulated gain coefficient were made using

two lasers: the Cyclops laser system to provide the intense pump pulse, and a tunable, cw dye laser to probe the gain at the Stokes frequency. The experimental layout is shown in Fig. 6-72. The 1.05- μ m Cyclops beam of up to 5 J at 1.1 ns was frequency doubled to 0.53- μ m in Type II KDP for up to 3 J of energy. The dye-laser probe beam (2-mm diam) was combined with the 25-mm-diam pump beam at a crossing angle of $\sim 1/3$ deg in the sample. Over the central 3 mm in the sample, the pump beam was flat to an excellent approximation. Hence, our fluence measurements could be made simply by using a 3-mm aperture (A_1) in front of the absorbing-glass calorimeter, and dividing the indicated energy by the area of the aperture.

The cw probe beam was generated by a CR-599 dye laser, pumped by an SP-171 cw argon laser. All the wavelengths needed for this work were obtained with a single dye, Rhodamine 110. The dye laser was equipped with frequency-stabilizing (FS), single-mode electronics giving a spectral width of $< 4 \times 10^{-4}$ Å. The probe wavelength was monitored with a wavemeter and, as backup, a scanned-grating spectrograph. The precision of the wavemeter was better than 0.005 Å, more than adequate for our work. It gave us a continual readout of wavelength without the need to incessantly scan, as with the grating instrument.

After passing through the sample, the probe beam was spatially separated from the pump and detected with a vacuum photodiode (D_1). The probe beam gave a cw signal of ~ 10 mV on the TEK 7104 oscilloscope. We spent considerable effort to maximize this signal and to reduce to < 1 mV all sources of spurious electrical noise on the scope traces. We achieved a cw-probe

Fig. 6-72. Experimental apparatus for measurements of stimulated Raman gain.

Fig. 6-73. Oscilloscope traces for measurement of stimulated Raman gain in benzene liquid and KDP samples.

signal that was acceptably quiet during laser firing, and we could clearly see the Raman-gain-induced bump on the trace.

At first, we measured the absolute gain in liquid benzene to test the operation of our apparatus. Figure 6-73(a) shows the oscilloscope traces from our benzene (1-cm cell length) measurements. The lower trace shows the zero line; the upper trace shows

a gain of $G = 4.3$. Figure 6-73(b) is from shot through the 15-cm KDP sample, for which the gain is $G = 3.2$. The temporal waveform of the SRS signal follows that of the pump pulse.

For both benzene and KDP, the frequency of the dye probe beam was scanned, and shots were taken to plot SRS gain vs frequency. Typical experimental

Fig. 6-74. Spontaneous and stimulated Raman-scattering intensity vs wavelength in benzene liquid and KDP crystal samples.

Frequency-Conversion Systems

Table 6-7. Measured absolute SRS gain coefficients for the 915- cm^{-1} vibrational mode of KDP.

Scattering configuration	Length (cm)	Tensor coupling	Approximate tensor coupling	Ψ (deg)	SRS-gain coefficient (cm/GW)	
					Predicted ^a	Measured
$\beta(\alpha\gamma)\beta$	5	a^2	a^2	31	0.21	0.23
$\beta(\gamma\gamma)\beta$	5	$[a \sin^2 \theta + b \cos^2 \theta + c \sin 2\theta]^2$	$[a \sin^2 \theta + b \cos^2 \theta]^2$	31	0.19	0.19
$\alpha(\gamma\gamma)\alpha$	5	$[a \sin^2 \theta + b \cos^2 \theta + c \sin 2\theta]^2$	$[a \sin^2 \theta + b \cos^2 \theta]^2$	90	0.19	0.23
$\beta(\alpha\alpha)\beta$	15	a^2	a^2	31	0.21	0.19
$\beta(\alpha\gamma)\beta$	15	$[c(\Psi) \sin \theta + d \cos \theta]^2$	$[c(\Psi) \sin \theta]^2$	31	0.076	0.065
$\alpha(\gamma\gamma)\alpha$	15	a^2	a^2	90	0.21	0.19
$\alpha(zz)\alpha$	15	b^2	b^2	90	0.13	0.13
$\alpha(z\gamma)\alpha$	15	c^2	0	90	0.0	0.065

^aNormalized to 0.21 cm/GW for a^2 tensor coupling.

data runs are shown in Fig. 6-74. The curves represent measured, spontaneous line shape scaled to the stimulated-gain data points. Thus, we covered the exact probe wavelength for which the SRS gain was maximal. The procedure was more important for benzene than for KDP, because the benzene line was only 11% as wide (2.3 cm^{-1}).

Raman Gain in KDP-Array Segments at 2 ω . The results for the measured absolute gain in KDP are listed in Table 6-7. We tested both 5- and 15-cm-long KDP crystals. Both crystals were spare KDP-array segments. We propagated our test beams through the edge faces of these crystals so that the Raman probe beam propagated just as the Raman wave would in the actual array. The pump beam also traveled in the same direction and was near-collinear with the probe beam, as previously described. The direction was transverse to the pump direction encountered in the array geometry.

Our approach to predicting the Raman gain coefficients was to arrive at best values

for the diagonal Raman tensor elements, a^2 and b^2 , from direct, collinear stimulated-gain measurements. We then computed the coefficients appropriate for the transverse pump geometry from tensor couplings listed in Table 6-6. A best value of 0.13 cm/GW for a^2 was determined as an average of the results from the 5- and 15-cm-long crystals for configuration $\beta(\alpha\alpha)\beta$. The 5-cm crystal gave a value of 0.13 cm/GW for b^2 from $\alpha(zz)\alpha$. The predicted collinear SRS-gain coefficients in Table 6-7 were calculated with the best a^2 and b^2 , as above, and are in good agreement with experiment. In particular, the predicted gain coefficient of 0.076 cm/GW for the depolarized scattering, $\beta(\alpha\gamma)\beta$, is in excellent agreement with the experimental value of 0.065 cm/GW and substantiates our model for the $C(\Psi)$ tensor element, Eq. (48).

In the array geometry, an extraordinary wave polarized along β will principally scatter into direction α (or $-\alpha$). From Table 6-6, the appropriate tensor coupling for this transverse configuration, $\gamma(\beta\beta)\alpha$, is

$(a \cos^2 \theta + b \sin^2 \theta)^2$. With best values for a^2 and b^2 , the gain coefficient is 0.15 cm/GW. Thus, the major results are that the KDP SRS-gain coefficient for the 526-nm pump is 0.21 cm/GW when the scattered wave is an o (ordinary) wave in the crystal, and is 0.15 when it is an e (extraordinary) wave.

Srivastava and Grow^{71,72} reported a calculation of the gain coefficient for KDP relative to benzene, which they derived from measurements of spontaneous Raman scattering. Their calculated result (converted to a 526-nm pump) is 0.36 cm/GW and is in fair agreement with our values. Unfortunately, Srivastava and Grow did not specify the orientation of the KDP crystal in their paper and did not pursue quantitatively the impact of their other angle-dependent investigations on the stimulated gain.

Raman Gain of Index-Matching Fluids. After reviewing the literature,^{71,74} we believe we have made the first direct, absolute measurements of SRS gain in benzene. The value of 4.3 cm/GW for a 526-nm pump is our result. The absolute uncertainty in this number is $\pm 20\%$. For surveys of SRS strength in index-matching fluids, it is valid to run the spontaneous spectrum and scale the peak gain with our absolute benzene number.

The results for a select group of index-matching fluids are listed in Table 6-8. For each fluid, a spontaneous Raman spectrum was obtained over a wide frequency range to determine the vibrational line with the greatest scattering strength. A high-resolution scan was then performed to determine the peak scattering strength relative

to benzene. The peak stimulated gain at 488 nm is then the product of the relative peak scattering strength with the third power of the ratio of Stokes frequencies of the benzene to the IMF. The absolute Raman gain for a 526-nm pump can then be determined from the absolute 488-nm value by making the linear frequency correction in the Stokes frequency. Since the line-center, peak scattering strength is measured rather than the line-integrated strength, the full width at half maximum (FWHM) of the spontaneous line is not used in this method to calculate the gain. Halocarbon IMFs have the lowest predicted Raman gains and should not present a stimulated-Raman threat at the intensities and apertures on Nova and Novette.

Impact of KDP Raman Gain Coefficients on Nova and Novette. We used the gain coefficients for KDP, scaled to the appropriate wavelength and the current Nova maximum-intensity data, to arrive at the single-pass gain, g , for the various waves involved in doubling or second-harmonic generation (SHG), tripling or third-harmonic generation (THG), and quadrupling or fourth-harmonic generation (4HG). Values for g are listed in Table 6-9 for 27-cm-array segments. All values are too low to cause problems on Nova or Novette on a single pass.

We then computed the loss for long pulses, which allow multiple traversals of the KDP segments. Durations beyond 1.35 ns in 27-cm-array segments enable SRS originating at one edge of a crystal segment to propagate 27 cm to the opposing edge,

Table 6-8. SRS-gain predictions for index-matching fluids based on measurements of relative spontaneous Raman cross section.

Fluid	Raman frequency, ω_R (cm^{-1})	$g/g(\text{benzene})^a$ ($\times 1000$)	Estimated SRS gain ^b (cm/GW)	
			488 nm	526 nm
Benzene	992	1000	4.66	4.30
Cargille	998	588.0	2.74	2.53
	3053	97.9	0.46	0.42
	2858	148.8	0.69	0.63
Decalin	894	70.6	0.33	0.30
Ethylene carbonate	754	16.4	0.076	0.071
Halocarbon 0.8	198	16.8	0.078	0.073
Halocarbon 1.8	437	18.4	0.086	0.079
Halocarbon 56	438	17.5	0.082	0.076
Koolase	848	39.1	0.18	0.17
	2939	68.6	0.32	0.29
Krytox	805	11.7	0.055	0.050
Perfluorodecalin	757	71.6	0.33	0.31

^aMeasurements relative to benzene (pump wavelength = 488 nm).

^bPrecision, $\pm 20\%$.

Frequency-Conversion Systems

Table 6-9. Single-traversal, SRS-gain predictions for Nova KDP segments.

Pump wave	Pump-wave orientation ^a	I_m at 1.35 ns inside KDP (GW/cm ²)	SRS gain, g^b (cm/GW)	Exponential gain coefficient, $G = g^b L_{eff}^c$	Maximum edge reflectivity, ^c (R_{eff}/I_m)
ω in Type I SHG	o	2.91	0.10	7.9	3.7×10^{-4}
2ω in Type I SHG	e	2.46	0.15	10.0	4.5×10^{-5}
$1/2 \omega$ in Type II SHG	e	1.45	0.071	2.8	6.1×10^{-2}
$1/2 \omega$ in Type II SHG	o	1.45	0.10	3.9	2.0×10^{-2}
2ω in Type II SHG	e	2.46	0.15	10.0	4.5×10^{-5}
ω in Type II THG	e	0.97	0.071	1.9	1.5×10^{-1}
2ω in Type II THG	o	1.94	0.21	11.0	1.7×10^{-5}
3ω in Type II THG	e	2.35	0.23	14.6	4.6×10^{-7}
2ω in Type I 4HG	o	2.0	0.21	11.3	1.2×10^{-5}
4ω in Type I 4HG ^d	e	1.0	0.31	8.4	2.2×10^{-4}

^ao = $\gamma(\delta\beta)/\alpha$; o = $\gamma(\alpha\alpha)/\beta$.

^bPrecision, $\pm 20\%$.

^c $L_{eff} = 27$ cm.

^dExtrapolation to 4ω ignores preresonant enhancement.

Table 6-10. Predicted risk to Nova KDP arrays from SRS.

Crystal size (cm)	Gain, G, from Table 6-9						G $\times$ 1.2				G $\times$ 0.8			
	ME	N	AGP	ME	N	AGP	ME	N	AGP	ME	N	AGP	ME	N
SHG { 15	S	S	S	U*	U	S	S	S	S	S	S	S	S	S
27	S	S	S	U*	U	S	S	S	S	S	S	S	S	S
THG { 15	U	U	S	U	U	S	U*	U	S	U*	U	S	U	S
27	U	U	S	U	U	S	U*	U	S	U*	U	S	U	S

ME = metal eggcrate
N = no eggcrate

AGP = absorbing-glass partitions
S = safe

U* = unsafe for limited subset
U = unsafe

reflect, and experience further gain in additional passes across the KDP. To account for multiple traversals, the sidescattered Stokes intensity is given by

$$I_{SRS} = (I_{in} e^{g I_m L}) [R_{eff}] e^{g I_m L} [R_{eff}] e^{g I_m L} \dots \quad (49)$$

where the first term represents the first traversal with maximum intensity I_m and crystal length L , the second term is the second traversal, with edge reflectivity R_{eff} and so forth, and

$$\sum_i I_i = \frac{c\tau}{n} \quad (50)$$

From Eq. (49), we see that a way to keep the long-pulse SRS from being more deleterious than the single-traversal situation is to maintain the crystal-edge reflectivity at

$$R_{eff} \leq e^{-g I_m L} \quad (51)$$

Then the SRS gain cannot grow beyond the single-traversal levels listed in Table 6-9. We also list the maximum allowable edge

reflectivity from Eq. (51). The most critical case is the 3ω wave in Type II THG, which requires an edge treatment with reflectivity $< 4.6 \times 10^{-7}$. If the long-pulse intensity is significantly lower than the short-pulse intensity, then the necessity of satisfying Eq. (51) is relaxed.

Using a computer code to model Eq. (49), we considered the SRS loss for metal-eggcrate arrays, arrays with no eggcrate, and arrays with absorbing-glass partitions for both 15- and 27-cm crystal segments. The criterion for unacceptable loss in all cases was a sidescatter of 1% of the incident intensity. Results are listed in Table 6-10. The left three columns in this table are calculated using gain factors from Table 6-9. The remaining columns indicate the effect of the $\pm 20\%$ error on the gain coefficient. The symbol, U*, denotes unsafe, but only over a small subset of the parameter space involving pulse duration, edge reflectivity, and spontaneous-noise-source intensity.

In our computations, the only differences between metal eggcrate (ME), no eggcrate (N), and absorbing-glass partitions (AGP) are the numbers used for the edge reflectivity and the crystal size. For the metal eggcrate, we used reflectivities of 90, 50,

and 10%; for no eggcrate, we used 100%; for the glass-partition array, we used 1, 0.5, and 0.1%. For transverse size in the no-eggcrate case, we used a 70-cm clear aperture.

SHG is safe for all three types of array. Considering the upper end of the error bar ($G \times 1.2$) for SHG, we see that the no-eggcrate array is unsafe. The absorbing-glass partitioned array is safe. The metal-eggcrate array is almost safe; that is, for a limited range of pulse duration for pessimistic assumptions of edge reflectivity and noise-intensity source, the SRS sidescatters $>1\%$ of the incident intensity. Since we have no actual SRS measurements on an array exact values for these parameters are not yet established.

The array can properly transmit the desired 3ω intensity only if absorbing-glass partitions are employed. Even at the lower error limit of G , the no-eggcrate design is a problem, and the metal-eggcrate design is a restricted problem at 3ω . For the upper error limit, even the absorbing-glass-partition array may have a restricted problem with 27-cm crystals.

We include in Table 6-9 the numbers relevant for Type I 4HG in KDP. The gain coefficient at 4ω was scaled from the measured values. The pump intensity was set at 2 GW/cm^2 maximum, and 1 GW/cm^2 maximum output at 4ω was assumed. With these intensity values, it appears that SRS at 4ω , or by the 2ω pump, will not be a problem. Finally, we have no way to estimate electronic-resonance effects, which could make the 4ω gain coefficient considerably larger.

Authors: W. L. Smith, M. A. Henesian, and F. P. Milanovich

Major Contributors: C. Vercimak, R. Lopert, and S. E. Peluso

KDP Growth

In 1983, G. Loiacono of North American Philips, under contract to LLNL, continued his studies on rapid-growth methods of KDP. He grew KDP samples in the three-chamber system described in previous reports.⁷⁵ His findings were as follows:

- KDP growth rates in the three-chamber system are limited to $\sim 6 \text{ mm/d}$ by defects

and stresses introduced when the KDP boule grows in the $x-y$ plane.

- Solution pH will not control $x-y$ growth at high supersaturation, in contrast to reported results showing control at low supersaturations.
- Growth rates are sensitive to the hydrodynamics in the crystallizer.

Author: S. E. Stokowski

High-Dynamic-Range Frequency Conversion

Our recent study of frequency conversion and the performance of future systems showed that an important parameter of the frequency-conversion subsystem is its ability to handle temporally shaped pulses. Recent experience with Argus and Novette indicates that spatial variation in intensity over the face of the beam can significantly reduce conversion efficiency. Consequently, our design goal for new systems requires that the frequency-conversion scheme perform efficiently over a large range of intensities.

The temporally shaped pulses needed for target-irradiation experiments consist of a low-power foot, a transition region, and a high-power main pulse. Depending on the target, the ratio of main-pulse power to foot power ranges from ~ 15 to 50; the ratio of the energies ranges from ~ 3 to 10. To avoid significant waste of 1ω energy, both the foot and the main pulse must be converted efficiently. Moreover, both the foot and the main pulse have spatial variations, and we expect a peak-to-average ratio of ~ 2 .

Any 1ω energy that meets the crystal with an intensity for which the conversion efficiency is small will be wasted. The crystals must, therefore, convert efficiently over the full range of beam parameters present in the beam. The relevant parameters are local intensity, local beam direction, and local polarization. For processes that require two separate inputs, such as frequency mixing (3ω), the energy balance between them is an important parameter. Finally, for some short-wavelength processes (4ω), the wavelength spread of the beam is important.

We have analyzed the performance of single-crystal schemes for mixing 1ω with

$P\omega$ to generate $(1 + P)\omega$. We analyzed the relationship between the range of intensities over which efficient conversion occurs, on the one hand, and the angular detuning and energy balance, on the other. The dynamic range of the scheme is defined as the ratio between the highest and lowest intensities for which efficient conversion takes place. We studied the mathematical relationship between the dynamic range and the angular detuning and energy imbalance. For future laser systems, we are looking for a dynamic range somewhat larger than 100.

We find that the dynamic range of a single crystal is limited either by the angular detuning or by the energy imbalance between the two input waves at $I\omega$ and $P\omega$. A dynamic range of 10 seems achievable with current technology and KDP or potassium dideuterium phosphate (KD*P). Perhaps new materials will extend the dynamic range to 30. A higher dynamic range will not be achieved with a single crystal; the conditions on the angular tolerances and energy balance are several orders of magnitude beyond present experience. For example, using a single crystal of KDP, a dynamic range of 30 requires a 1- μ rad beam divergence and energy balance to $\sim 0.1\%$. It is possible that a new material would be more forgiving of beam divergence and,

thus, develop a dynamic range of 30 with a 50- μ rad beam. However, for a dynamic range of 100, the required beam divergence is 5 nrad, and energy balance must be good to within 10^{-6} . Such values are impossible in any laser system known today.

We conclude that single-crystal schemes can be expected to have dynamic ranges of from 3 to 30. They will not meet the goals for future systems. Fortunately, there are other multicrystal schemes, such as quadrature- 2ω and polarization-switched schemes that have dynamic ranges of from 100 to 1000 under conditions in which single crystals would have a dynamic range of from 10 to 30. Such schemes are described below. However, we first describe single-crystal schemes to explain why they do not perform well.

Dynamic Range of Single Crystals.

Figure 6-75 shows the typical behavior of conversion efficiency of a single crystal as a function of drive intensity. At low power, the efficiency rises from zero, peaks at some value, and then drops away. Under perfect conditions where the angular detuning is zero and the inputs are properly balanced, conversion efficiency rises to unity as the intensity increases to infinity. To define dynamic range, an efficiency, $\bar{\eta}$, is chosen, and the operating range is defined as the range

Fig. 6-75. This typical conversion-efficiency curve vs drive intensity illustrates the operating and dynamic ranges.

of intensities for which $\eta \geq \bar{\eta}$. The limits of this range are I_2 and I_1 , and the dynamic range, Ω , is defined as

$$\Omega = \frac{I_2}{I_1} \quad (52)$$

Clearly, Ω depends on both $\bar{\eta}$ and the shape of the curve. Thus, Ω is a function of desired efficiency, $\bar{\eta}$, angular detuning, and energy balance.

Consider the mixing process $I\omega + P\omega \rightarrow (1 + P)\omega$. If F and G are the input intensities at $I\omega$ and $P\omega$, respectively, and H is the harmonic output, then

$$H = \eta(F + G) \quad (53)$$

The energy balance, m , is defined as

$$m = \frac{PF}{G} \quad (54)$$

and the energy is balanced between the two inputs for $m = 1$. The energy imbalance, σ , is defined as the deviation from energy balance:

$$\sigma = 1 - m \quad (55)$$

If σ is small and derives from experimental or system-related uncertainties in the drives, polarizations, or spatial ripples on the beams, then, locally,

$$\sigma = -\frac{\delta F}{F} + \frac{\delta G}{G} \quad (56)$$

Angular detuning, δ , is

$$\delta = \frac{\Delta k l}{2} = \frac{\beta \Delta \theta l}{2} \quad (57)$$

where $\Delta \theta$ is the angular misorientation between the local beam direction and the phase-matching direction, l is the crystal thickness, and β is the angular sensitivity. The typical value of β for KDP is $\sim 2.5 \text{ cm}^{-1}/\text{mrad}$.

We are now in a position to study Ω as a function of $\bar{\eta}$, δ , and σ . In the absence of absorption, there is an analytic solution for the output that we have used to derive analytic formulas for Ω and to study the conditions under which Ω can be made large.

In the high-drive regime of Fig. 6-75, efficiency declines from its peak value. The decline depends on angular detuning and energy imbalance; moreover, it becomes more sensitive to both as drive is increased. The case for angular sensitivity is shown in Fig. 6-76. For perfect balance, $\sigma = 0$, we plot the contours of conversion efficiency as a function of detuning, δ , and drive intensity, $I = F + G$. The horizontal axis is drive intensity, η_0 , where $\eta_0 \equiv C^2 l^2 I$. Notice that as the drive intensity increases, the central maximum becomes narrower. The width of the central maximum, Δ , is

$$\Delta = 8\eta_0^{1/2} \exp(-\eta_0^{1/2}) \quad (58)$$

and drops exponentially with the square root of drive. This implies a sharp cutoff in achievable dynamic range. The dynamic range for this case ($\sigma = 0$) is calculated as follows. To achieve a dynamic range of Ω , the angular detuning δ must be less than $\bar{\delta}$, where

$$\bar{\delta} = \alpha(\eta)\Omega^{1/2} \exp[-\lambda(\bar{\eta})\Omega^{1/2}] \quad (59)$$

The functions α and λ are plotted in Figs. 6-77(a) and (b). In Fig. 6-78(a), the limiting value, $\bar{\delta}$, is plotted as a function of $\bar{\eta}$

Fig. 6-76. Contours of conversion efficiency as a function of angular detuning, δ , and drive, η_0 . The increase in angular sensitivity at high drive limits the dynamic range.

Frequency-Conversion Systems

Fig. 6-77. Detuning limits on dynamic range for (a) $\alpha(\bar{\eta})$ and (b) $\lambda(\bar{\eta})$.

Fig. 6-78. (a) Maximum angular detuning and (b) maximum energy balance allowed as a function of η and dynamic range, Ω .

for several values of the dynamic range Ω . High dynamic range rapidly leads to unrealistic requirements on δ . For Nova, δ is typically less than ~ 0.25 . To achieve a dynamic range of 30 requires improvement by a factor of 10 in δ at $\eta = 0.7$, and by a factor of 50 at $\eta = 0.8$. For a dynamic range of 100, δ must be reduced by 3 to 4 orders of magnitude from Nova values.

A similar result is found for the energy balance; σ is constrained to be less than $\bar{\sigma}$, where

$$\bar{\sigma} = \frac{2\alpha}{\lambda} \exp(-\lambda\Omega^{1/2}) \quad (60)$$

as shown in Fig. 6-78(b). A dynamic range of 30 requires σ to be $< 5 \times 10^{-3}$ at $\eta = 0.7$ and 10^{-3} at $\eta = 0.8$. A dynamic range of 100 requires energy balance to 10^{-4} .

The case of finite absorption can be treated with an extension of these results. To achieve high dynamic range, the falloff at high drive must be forestalled. In other words, high dynamic range requires the conversion process to proceed very close to saturation. True saturation occurs only for

$\sigma = \delta = 0$. Saturation requires that the sign of the interaction between F , G , and H remain constant. Thus, the angular dephasing ($\delta \neq 0$) must be small, and both F and G should go to zero simultaneously; that is, the energy imbalance ($\sigma \neq 0$) must also be small. The only way absorption affects the FGH phase is by destroying the energy imbalance with unequal absorption for F and G . Even in this case, it is possible to find an input, σ , for which the process comes arbitrarily close to energy balance at saturation. Thus, we conclude that absorption does not preclude saturation. Our conclusion here is in sharp contrast to our conclusions for angular detuning and energy imbalance, which do preclude saturation. The main effect of absorption is energy loss and heating of the crystal material; it poses no fundamental limit on dynamic range.

The main result in Figs. 6-78(a) and (b) is that there is an exponential cliff in the dynamic range as a function of the system parameters δ and σ . The dynamic range is limited by either parameter. We draw several conclusions:

- Dynamic range is limited by angular detuning and by energy imbalance and polarization errors.
- Present experience gives $\Omega \approx 10$.
- New materials may achieve $\Omega = 30$, but higher values are not conceivable.
- In searching for new materials we need not improve δ beyond ~ 30 . Further improvement is wasted because, for a small enough δ , the dynamic range is limited by energy imbalance in Type II schemes.

Dynamic Range of Quadrature

Schemes. The dynamic range of single-crystal schemes is limited by angular detuning and energy-balance constraints. At present, by using KDP it is only possible to achieve a dynamic range of ~ 10 . However, we can envision materials that would permit a dynamic range of up to 30. Higher dynamic range with single crystals is not possible because it requires wildly unrealistic tolerances on the angular detuning and on energy balance or polarization of the input beam or beams. Here, we show that the dynamic range may be made very large (from 50 to 500) by using the two-crystal quadrature scheme.

In a quadrature SHG scheme^{76,77} (see Fig. 6-79), the optical axes of the two KD*P crystals are orthogonal, so that the green output from the first crystal passes unchanged through the second. The second crystal converts only what is left over from the first. Thus, the overall conversion efficiency is

$$\eta = \eta_1 + \eta_2(1 - \eta_1) \quad (61)$$

and the drive to the second crystal is

$$I_2 = I(1 - \eta_1) \quad (62)$$

Consider that the first crystal is set to give a dynamic range of Ω_1 and operates at an efficiency above η between the limits

$$I_{\min} < I < \Omega_1 I_{\min} \quad (63)$$

At the upper limit of this range, the drive to the second crystal is

$$I_2 = \Omega_1 I_{\min}(1 - \eta) \quad (64)$$

If the dynamic range of the second crystal is Ω_2 , then we can set the second crystal

to operate at an efficiency $> \eta$ between the limits

$$I_2 < I < \Omega_2 I_2 \quad (65)$$

Thus, the combination operates above η in the range

$$I_{\min} < I < \Omega_1 \Omega_2 I_{\min} \quad (66)$$

and the dynamic range of the combination is, therefore,

$$\Omega_q = (1 - \eta) \Omega_1 \Omega_2 \quad (67)$$

Single crystals have dynamic ranges of from 10 to 30 for $\eta = 80\%$. A quadrature scheme using the same materials then has a dynamic range approaching the square of the single-crystal range. Namely, for single crystals with $10 < \Omega < 30$, we find quadrature schemes for which $20 < \Omega_q < 200$, as illustrated in Fig. 6-80.

Actually, the design just given is quite conservative because, at the point where the first crystal efficiency is dropping through $\bar{\eta}$, the second has already reached $\bar{\eta}$. Thus, at this crossover point, the efficiency is

$$\eta = \eta + \eta(1 - \eta) = 1 - (1 - \eta)^2 \quad (68)$$

and is still considerably greater than $\bar{\eta}$. By raising I_2 , or the drive required to reach 80% in the second crystal, we can also extend the upper limit. $\Omega_2 I_2$, without letting η drop below $\bar{\eta}$ in the crossover region.

It is not possible to be more quantitative without resorting to computer calculations or analytic mathematics. It appears, though, that a less conservative approach to the

Fig. 6-79. The quadrature scheme allows high dynamic range.

Fig. 6-80. The design of thin-thick quadrature-conversion schemes. A conservative choice for I_2 would be $I_2 = (1 - \eta)\Omega_1 I_{\min}$; $I_2 = 2(1 - \eta)\Omega_1 I_{\min}$ is closer to optimum.

Fig. 6-81. A thin-thick quadrature scheme with KD*P (1% surface loss per surface) yields higher dynamic range than an equal-thickness (Nova) design.

with virtually an arbitrarily high ratio of intensities in the main pulse and foot. In this sense, their dynamic range can be made extremely large.

In contrast to the Nova quadrature design, which uses equal-thickness crystals, we call our design for high dynamic range the thin-thick design. A thin-thick design using KD*P has been determined. The optimal design depends on the details of 1ω performance—in particular, on the laser system's angular detuning, and on the absolute limits ($I_{\min}$ and $I_{\max}$) required by the targets. For the choices:

- $\Delta\theta = 100 \mu\text{rad}$,
 - $I_{\min} = 0.1 \text{ GW/cm}^2$, and
 - $I_{\max} = 10 \text{ GW/cm}^2$,
- the optimal lengths are
- $I_1 = 1.4 \text{ cm}$ and
 - $I_2 = 5.0 \text{ cm}$.

A comparison of the thin-thick quadrature scheme with the Nova design is shown in Fig. 6-81. Our design achieves a dynamic range considerably greater than 100 at an internal conversion efficiency of 70%.

Author: D. Eimerl

crossover region can increase the dynamic range by ~ 2 . This makes the achievable dynamic range somewhat higher, at $40 < \Omega_0 < 400$.

If the system does not require high efficiency in the crossover region, the extremes of this range can be extended further. Pulses with feet where nearly all the energy comes within a factor of 2 of the foot intensity or the main-pulse intensity, are useful for target experiments. Quadrature schemes can provide high efficiency for such pulses

Noncollinear Phase-Matching Schemes

Noncollinear harmonic conversion involves irradiating a nonlinear material with two overlapping but noncollinearly directed laser beams. The nonlinear material

responds by developing an oscillating dipole-moment density whose frequency ω' is the sum of the input-wave frequencies. The dipole-moment density is also spatially periodic, with a wave vector $\mathbf{k}'$ equal to the sum of the input wave vectors. Thus,

$$\omega' = \omega_1 + \omega_2 \quad (69)$$

and

$$\mathbf{k}' = \mathbf{k}_1 + \mathbf{k}_2 \quad (70)$$

The dipole moment radiates light at ω' in a manner similar to an extended radio antenna. The radiation will be weak unless the wave vector of the radiation $\mathbf{k}_3$ exactly matches that of the radiating dipole-moment density. Thus, the condition for high-power output is

$$\mathbf{k}_3 = \mathbf{k}' \quad (71)$$

In most materials, $\mathbf{k}_3$ is too long to match $\mathbf{k}'$ because the refractive index is larger at shorter wavelengths. In birefringent materials we can tune the indexes and thereby shorten $\mathbf{k}_3$ so that it matches $\mathbf{k}'$. The indexes are tuned by adjusting the orientation of the material relative to the polarization and wave vectors of the three waves involved. Thus, we speak of angle-tuned harmonic generation.

Figure 6-82 shows the coordinate system for noncollinear SHG in Type I KDP. The

green (second-harmonic) ray lies in the horizontal plane, and the optical axis of the crystal rises from this plane so that the plane containing both the optical axis and the green ray is vertical. The angle between the green ray and the optical axis is θ . The two fundamental rays (red) make angles θ_1 and θ_2 with the green ray, and all three rays are coplanar. The ray plane forms angle ϕ with the vertical axis.

First, we must visualize the collinear case, for which all three rays lie in the horizontal plane and are parallel. The phase-matching (or velocity) condition on the green ray requires that the optical axis form a particular angle with the rays, called the phase-matching angle, θ_{pm} , where $\theta = \theta_{pm}$ for the collinear case.

In the noncollinear case, the two red rays are separated and are no longer parallel. Now, the dipole-moment-density wave vector necessarily decreases with opening angle $\theta_1 + \theta_2$. To maintain phase matching, we must speed up the green ray by increasing the k or phase-matching angle θ . Eventually the opening angle is large enough that θ reaches 90° . Phase matching is not possible for larger opening angles.

The special case for which $\theta = 90^\circ$, has some interesting attributes that relate to the sensitivity of the process to beam directions and crystal cut and to the coupling between the three waves. Consider, first, the directional sensitivities. In general, there is a

Fig. 6-82. (a) Wave-vector addition and (b) laboratory angles for noncollinear harmonic generation in Type I KDP.

Table 6-11. Relative sensitivity of phase-matching schemes to engineering tolerances and beam divergence.

Frequency-Conversion Systems

Scheme	Array tolerances	Beam divergence
Collinear	0.5	0.5
Noncollinear ($\theta = 90$ deg)	0.0	1.0

slight mismatch between k_3 and k' due to imperfect beam quality and construction tolerances. The phase mismatch depends on five angles:

- θ , defined above.
- θ_1 and ϕ_1 , which relate to the direction of one red input.
- θ_2 and ϕ_2 , which relate to the direction of the other red input.

Of these angles, θ is determined by engineering, fabrication, and assembly of the crystal arrays. The other four angles are determined by the quality of the input beams.

To understand the difference between collinear and noncollinear phase matching with regard to performance, we need two definitions:

- The crystal sensitivity of the arrays is the sensitivity to the phase-matching angle, θ . The acceptance angle is the tolerance on θ .
- The angular sensitivity of the scheme is the sensitivity to the red-beam angles θ_1 and θ_2 .

For our present purposes, we need not consider tolerances on the other beam angles, ϕ_1 and ϕ_2 . In collinear schemes, the two red beams are always locally parallel, and the crystal sensitivity is identical to the angular sensitivity. To put it another way, $\theta_1 = \theta_2 = 0$ for collinear schemes, and there is only one angle in the problem. The green beam follows the local red direction everywhere, so that variations in the red direction are the same as variations in θ .

In noncollinear schemes, the crystal sensitivity is separate from the angular sensitivity. For $\theta = 90$ deg, the acceptance angle is several degrees, perhaps as much as 100 mrad. However, in all the noncollinear schemes we have studied, the angular sensitivity is, in fact, greater than in the collinear case. There are theoretical reasons for believing that this is always the case. Maximally noncollinear schemes are more angularly sensitive than collinear schemes; however, their acceptance angles are much greater. Clearly then, the trade-off between noncollinear and collinear schemes is one of sensitivity to engineering tolerances vs

sensitivity to beam quality, as illustrated in Table 6-11.

In collinear schemes, both sensitivities are about the same, whereas noncollinear schemes are sensitive only to beam divergence. Our experience with Nova and Novette is that the beam divergence is considerably smaller than the array-fabrication tolerances. Thus, going to noncollinear schemes offers some improvement in performance. If the beam divergence is greater than the array misalignments, there would be some performance penalty in going to a noncollinear scheme.

The second issue relating to 90-deg phase matching is the coupling between the three waves. At this angle, the coupling vanishes for Type II SHG, in which the two input waves are orthogonally polarized. For the Type I SHG, in which the two input waves have the same polarization, the coupling is a little larger than for collinear schemes. Thus, only Type I noncollinear SHG schemes are possible for KDP at $\theta = 90$ deg.

There are two problems with Type I schemes. The first is that the range of intensities over which they convert efficiently (their dynamic range) is small. In fact, any single-crystal scheme has too small a dynamic range, as discussed in the preceding article. To meet the requirements on dynamic range, we need a quadrature scheme that involves two KD*P crystals, each in a Type II scheme arranged so that their k - z planes are orthogonal. This is the only scheme known that has the potential to achieve the required dynamic range. Since only the Type I scheme works noncollinearly, we can apparently rule out KDP as a material for noncollinear schemes.

The second problem with Type I schemes is that both the inputs have the same polarization. Two noncollinear waves of the same polarization generate interference fringes where they overlap, giving a peak-to-average-intensity ratio of 4. For Type II schemes, the two inputs have orthogonal polarizations, and there are no interference fringes. Thus, the damage fluence required of the KDP is twice as high for Type I as for Type II schemes. Again, we appear to have good reason for rejecting Type I KDP.

The two problems described above occur with any negatively birefringent material,

including all KDP isomorphs and many other potential candidates for nonlinear material. What we need is a Type II scheme in which coupling does not vanish at 90 deg. The relevant materials are positively birefringent. We need a new material whose extraordinary index is larger than its ordinary index. Urea is such a material; however, urea is somewhat less than satisfactory because of its high Raman gain and its comparatively high angular sensitivity (20 cm⁻¹/mrad). Our conclusion is that we still need a good, positively birefringent nonlinear material for the noncollinear switch.²⁸

The single major issue concerning noncollinear schemes, other than those associated with $\theta = 90$ deg, is the energy balance between the two beams. If the beams do not have equal intensity point-by-point in the array, then back-conversion limits both the conversion efficiency and the intensity range over which the arrays give that efficiency. Calculations show that energy balance must be maintained locally to within a few percent if no reduction in performance is to be incurred. Energy balance is a more severe constraint on the laser system than the directional sensitivity, discussed above. It may, indeed, be so severe as to rule out noncollinear frequency-conversion schemes altogether.

In summary, the sensitivity of noncollinear harmonic-conversion schemes to beam directions is about the same as that of collinear schemes. Noncollinear schemes are insensitive to crystal-array fabrication misalignments, whereas collinear schemes are equally sensitive to beam divergence and array tolerances. The energy balance between the two inputs is a strong system constraint for noncollinear schemes. Finally, we need a good, positively birefringent nonlinear material for the noncollinear switch.

Author: D. Eimerl

Incorporation of Frequency Conversion in MALAPROP

We have incorporated in MALAPROP the capability of modeling the conversion

process and the propagation of the color-converted beam through the final chain elements. The capability will be useful in designing future systems. We used a fast subroutine based on new analytic results for sum-frequency generation.

Our subroutine is ~ 150 lines of Fortran code and takes $\sim 400 \mu\text{s}$ per call. Thus, a 512×512 array of the beam can be converted in a few minutes of computer time. The subroutine gives both the intensity and the phase of the color-converted beam and performs both green and blue conversions. Modifications for any conversion scheme, such as quadrature (green) or 4ω schemes, are trivial in that only about 10 lines of code are involved. At present, we have included the Novette green arrays and the Nova blue configurations.

The fast analytic routine agree with the older, numerical code to ~ 1 part in 10^4 . The agreement implies that the numerical code, which was written a few years ago, was debugged correctly without the aid (at that time) of analytic results for blue. The subroutine is also suitable for transcription to BASIC and is probably fast enough to be useful in desktop computers, such as the HP-9845.

Figure 6-83 is a contour plot of the Nova 3ω output of the MALAPROP code. The input beam was given a weak convergence by passing it through a lens with a long focal length (2 km) before passing it through the Nova blue configuration using 1-cm-thick crystals. The intensity of the red beam is very low ($\sim 12 \text{ MW/cm}^2$). The incident beam has a large range of angular detunings in both the x and y directions. We used an aperture of 8 cm; thus, the maximum angular detuning of the beam was $4000 \mu\text{rad}$.

The crystal axes are shown in Fig. 6-84. The red polarization is perpendicular to the slit in the disks (the E -field is horizontal, and the slits are vertical). The criss-cross pattern in Fig. 6-83 reflects the orientation of the crystal; the doubler z axis is at 35 deg to the vertical, and the mixer axis is at 35 deg to the horizontal or incident-red-beam polarization. The solid lines represent peaks in output; the dotted lines represent valleys. The asymptotic limit of the dotted lines is a rectangular grid that defines the

Fig. 6-83. Contour plot of 3ω light out of the Nova arrays at low drive.

Fig. 6-84. Relationship between disk slit and beam polarizations for Nova MALAPROP simulations.

zeros of the output. Zeros are due to the angular sensitivity of the crystals.

A Nova chain was simulated using these algorithms. We chose the baseline chain plus one additional 46-cm amplifier and one additional 31.5-cm amplifier. The pulse duration was 1 ns, and the input-pulse shape was assumed to be such that the laser output was of constant power. We modeled the last time slice (which

corresponds to the maximum B integral and maximum beam noisiness). Under the above conditions, the entrance lens of the final spatial filter limits the power out of the laser to 12.8 TW. This is also the power into the conversion array, which is located 60 m from the output spatial filter.

Table 6-12 lists parameters associated with the $1.05\text{-}\mu\text{m}$ beam. The focal-spot intensity is calculated for a perfect,

Large-Aperture Optical Switches

linear lens with a focal length of 3.195 m (Nova singlet). The Strehl ratio is the ratio of computed focal-spot intensity to that of a diffraction-limited beam of the same power, wavelength, and near-field diameter (74 cm).

For our simulation, both KDP crystals were 1 cm thick and the detuning angles were 100 μ rad. Passive-transmission losses (mostly at 1 μ m) and Fresnel losses were simulated using a surface loss of 0.985 per surface. Beam polarizations were assumed to be perfect. Internal beam-conversion efficiency was $\sim$ 75% (no apodization losses are included; the KDP crystals are effectively monolithic). Table 6-13 lists the 3ω beam parameters that are equivalent to those in Table 6-12 for the beam exiting the crystal array.

After frequency conversion, the reconstructed 3ω beam propagates through 4 cm of fused-silica array-output window, 1 m of free space, and 8 cm of singlet focusing optics (focal length = 3.011 m). The results are listed in Table 6-14.

Three conclusions can be drawn from our computations. First, frequency conversion significantly reduces the peak-to-average intensity ratio in the near field, presumably through back conversion of the peak intensities on the beam. Second, the peaks do not grow back in the $\sim$ 1-m propagation path at 3ω expected for Nova. Third, the frequency-conversion process resembles large-scale spot broadening (similar to that from aberrations) more than it resembles small-scale (ΔB) flare. To illustrate the latter phenomenon, we note the 40% included power in Tables 6-12, 6-13, and 6-14.

Authors: D. Eimerl and W. W. Simmons

Large-Aperture Optical Switches

Plasma-Electrode Pockels Cell

Introduction. Large ICF lasers of the future will require active optical elements with apertures on the order of 1 m (Ref. 79). Large-aperture optical switches are needed for

Power	12.8 TW
Pulse duration	1.0 ns
Mean intensity	3.2 GW/cm ²
Peak intensity	9.8 GW/cm ²
B integral	9.3 rad
90% half angle	50 μ rad
40% half angle	3 μ rad
Focal-spot intensity	1.5×10^{19} W/cm ²
Strehl ratio	0.31

Power	9.5 TW
Mean intensity	2.4 TW
Peak intensity	3.3 TW
B integral	9.3 rad
90% half angle	50 μ rad
40% half angle	9 μ rad
Focal-spot intensity	1.1×10^{18} W/cm ²
Strehl ratio	0.01

Power	9.1 TW
Mean intensity	2.3 TW
Peak intensity	3.3 TW
B integral	10.8 rad
90% half angle	50 μ rad
40% half angle	11 μ rad
Focal-spot intensity	7.5×10^{17} W/cm ²
Strehl ratio	0.007

isolation of stages, switchoff from regenerative amplifier cavities, and protection from target retroreflections. No conventional optical-switch technology can be scaled to such large apertures.

Pockels cells with ring electrodes^{80,81} require crystal thickness comparable to the aperture. Designs for thin-crystal Pockels cells using transparent, conducting metallic films⁸² or liquid layers^{83,84} have been proposed. However, such designs have unacceptable resistivity at large apertures or are not appropriate for use with high-intensity laser beams. Other approaches include fast Faraday rotators⁸⁵ or transient volumetric thermal gratings,⁸⁶ but major technical developments are required before their feasibility can be evaluated.

We are pursuing a promising new approach to large-aperture electrooptical switching with transparent plasma electrodes.⁸⁷ A thin electrooptical material is exposed to a glow-discharge plasma on each face. The plasma acts as a transparent conductive coating that allows uniform charging of each surface to the required potential. Thus, high-aspect-ratio electrooptic cells can be constructed. The cells might include Pockels cells using the linear electrooptic effects in crystals, electrooptically tuned SHG cells, and quasitransverse Kerr cells with quadratic electrooptic liquids and glasses.

Table 6-12. Nova laser beam parameters entering the frequency-conversion array.

Table 6-13. 3ω beam parameters exiting the frequency-conversion array.

Table 6-14. 3ω beam parameters at the target.

Large-Aperture Optical Switches

We have demonstrated the plasma-electrode concept with a $5\text{-} \times \text{ } 5\text{-} \times \text{ } 1\text{-cm}$ KD*P Pockels cell and are studying plasma characteristics at this aperture and in a 20-by-20-cm discharge chamber. We have also demonstrated an electrooptically tuned second-harmonic switch with Type I KDP in the longitudinal Pockels configuration, as discussed in the next article. Here, we present our recent results for a $5\text{-} \times \text{ } 5\text{-cm}$ Pockels cell with plasma electrodes. We also discuss our considerations for scaling to larger apertures.

Four-Electrode Geometry for Plasma-Electrode Cells. The most promising geometry for Pockels cells with plasma electrodes is shown in Fig. 6-85. A pair of electrodes on each side of the crystal establishes the initial sheets of glow discharge in low-pressure gas covering the entire optical aperture. Since the current flows parallel to the crystal, no electric field is applied along the optical axis, and the preionization current does not affect the polarization of the transmitted optical beam. One pair of electrodes can then be used, after the desired plasma conditions are established, to charge the crystal to the required voltage. A four-electrode design allows us to establish initial plasma conditions that are independent of voltage-pulse requirements for electrooptic switching.

Unlike metallic-film or liquid conductors, plasma is a nonlinear medium with regimes of operation that are highly undesirable. For example, with insufficient preionization, a large jitter in switching time is observed,

and copious amounts of radio-frequency radiation are generated by the cell. At moderate preionization levels, plasma resistance is a function of switching voltage, and very optical rise times are observed for low voltages. However, a proper choice of operating parameters with the four-electrode geometry allows us to obtain rise times in the nanosecond regime and with no observable jitter.

Experiments. We constructed and tested a Pockels cell of the type shown in Fig. 6-85. A 1-cm-thick KD*P crystal with 5-by-5-cm aperture was potted into a Plexiglas holder with a two-part silicone elastomer. The chamber body was constructed from Lucite. The two sections were evacuated and filled with helium gas through separate, long, coiled polyethylene fill lines that served to electrically isolate the two sides. The typical fill pressure was 10 Torr.

We chose helium gas for two reasons. The ultimate lifetime of the cell is determined by the destruction of optical surfaces by sputtering due to ion bombardment. Thus, a lighter ion leads to significantly longer lifetimes. In addition, a low breakdown voltage and low plasma resistance are desirable. We saw no damage to the crystal after more than 10^4 shots at 1 Hz. Gas contamination in our cell had a significant effect on the discharge characteristics, and the plasma resistance we observed was higher than expected for pure helium.

The electrical circuit used to pulse the cell is shown in Fig. 6-86. Two separate thyatron-switched pulsers were used. One provided the preionization voltage; the other provided the electrooptic switching pulse. The switching pulse was derived from a pulse-charged cable and a self-triggered, low-inductance spark gap. A spark-gap spacing of ~ 0.5 mm between rounded tungsten electrodes and with ~ 40 psi of compressed air gave good pulse-steepening performance. Voltage pulses of ~ 1 -ns rise time and 25-ns duration at ~ 20 kV were delivered to the Pockels cell via a 100-ns delay line. The cell was operated as a high-impedance load, so that a voltage doubling of the charged cable voltage occurred at the cell. Somewhat shorter turn-on times could have been obtained with a high-voltage pulser of twice the voltage and by shunting the cell with a $50\text{-}\Omega$ load. The preionization current was adjusted by

Fig. 6-85. Four-electrode geometry for plasma-electrode Pockels cells.

Fig. 6-86. Electrical circuit used to drive the 5- by 5-cm-aperture plasma-electrode Pockels cell.

varying the preionization voltage from 2.5 to 17 kV and by changing the ballast resistors.

To use the plasma-electrode Pockels cell as an optical switch, it must be placed between a polarizer and an analyzer. The two elements are aligned so that they are orthogonal to one another and oriented along the x and y crystal axes. The light propagates along the z (optical) axis, normal to the crystal face, which is also the direction of the applied electric field. In this longitudinal electrooptic configuration, with zero potential difference between the crystal faces, no light is transmitted through the switch. For an instantaneous voltage, $V(t)$, across the crystal, the optical transmission through the analyzer is given by⁸⁸

$$T(t) = \frac{S(t)}{S_0} = \sin^2 \left[\frac{\pi}{2} \frac{V(t)}{V_\pi} \right], \quad (72)$$

where V_π is the voltage for a half-wave phase retardation between waves polarized along the fast and slow electrooptic axes. For uniaxial crystals of tetragonal point-group symmetry, including crystals of the KDP-isomorph family,

$$V_\pi = \frac{\lambda}{2n_0^3 r_{63}}, \quad (73)$$

where r_{63} is the relevant coefficient of the electrooptic tensor. At a wavelength of 1.06 μm, the half-wave voltage for KDP is

Fig. 6-87. Electrooptic-crystal-glass-plate sandwich construction and polarizer orientation for high-voltage Pockels cells.

$V_\pi = 19.3$ kV; for 90% KD*P, $V_\pi = 7.4$ kV for the clamped or high-frequency (> 1 -MHz) electrooptic effect.⁸⁹

To minimize Fresnel-reflection losses at the crystal faces [not accounted for in Eq. (72)] and to increase the optical rise-time performance (discussed later), Pockels cells with AR-coated glass plates and an IMF at each crystal-to-glass interface can be constructed, as shown in Fig. 6-87. For such a sandwich construction, the actual voltage, V_k , across the crystal, in terms of external voltage, V_0 , is reduced as in a capacitive-voltage divider:

$$V_k = \frac{C_{\text{total}}}{C_{\text{crystal}}} \cdot V_0. \quad (74)$$

When the IMF layers are thin, the capacitive impedance of the fluid can be neglected, and Eq. (74) becomes

$$V_k = V_0 \left(1 + \frac{2l}{d} \frac{\epsilon_k}{\epsilon_w} \right)^{-1}, \quad (75)$$

where ϵ_k is the high-frequency dielectric constant of the crystal, d is the crystal thickness, ϵ_w is the dielectric constant of the windows, and l is (equal) window thickness. When $V(t)$ in Eq. (72) is taken as the external applied voltage, the half-wave voltage for the cell, $V_{0\pi}$, must be calculated from Eq. (75), with $V_k = V$, taken from Eq. (73). The high frequency dielectric constant along the optical axis for KDP is $\epsilon_z = 20$; for 90% KD*P, $\epsilon_z = 45$. For fused-silica windows, ϵ_w can range from 3.75 to 4.0. For several large-aperture, electrooptic-switch geometries that we shall consider later, the half-wave voltage for the KDP- or KD*P-glass-plate sandwich construction will be 4 to 16 times the bare-crystal value.

In our initial experiments, we used a bare KD*P crystal. The electrooptic effect was observed by probing the cell with a 1-mW beam from a He-Ne laser. Crossed polarizers were placed before and after the cell, and a fast photomultiplier was used as a detector. The response time of the detection system was estimated to be ~ 20 ns. The observed half-wave voltage at 633 nm, based on peak transmission through the polarizers, was ~ 5 kV. This value was in reasonable agreement with a predicted 4.4 kV for a clamped (i.e., high-frequency) electrooptic effect in 90% KD*P.

In our preliminary experiments, the circuit used to pulse the cell was similar to that shown in Fig. 6-86, but we did not use the charged cable and self-triggered spark gap. The preionization voltage was typically set at 10 kV with 5-k Ω ballast resistors. It is unfortunate that the combination of high plasma resistance (~ 1.5 k Ω) at the low electrooptic voltage together with high capacitance of the bare crystal (~ 100 pF) resulted in long optical turn-on times. The combination of the plasma electrode and the KD*P crystal appeared to be acting as a resistive transmission line. By probing

different points of the crystal with the laser, we could see a voltage propagation time of ~ 50 ns across the 5-cm aperture and optical rise times of ~ 100 ns.

Performance of the Pockels cell was significantly improved when a 2.5-mm-thick fused-silica plate was placed in series with the KD*P crystal and the preionization level increased above that used with the bare crystal. The plate was glued to one crystal face by means of transparent silicone elastomer to provide some index-matching and mechanical stability. The low dielectric constant of the fused silica ($\epsilon_w \approx 3.75$) relative to that of the bare KD*P ($\epsilon_z \approx 45$) resulted in most of the voltage drop occurring across the fused-silica plate. This raised the effective half-wave voltage of the cell by a factor of 4 over the bare-crystal value. The capacitance of the KD*P-fused-silica sandwich dropped by a factor of 4 to ~ 25 pF.

In our later experiments, the electrooptic effect was probed at high temporal resolution by a single-frequency (514-nm) beam from an argon-ion laser with a nominal output of 1 W. We used an S-20 biplanar vacuum photodiode as a detector. The response time of the detection system was estimated to be better than 1 ns. Figures 6-88(a) and (b) show typical results obtained with the argon-ion laser and the spark-gap-steepened pulser circuit.

The oscilloscope trace shown in Fig. 6-88(a) is a representative voltage pulse (nominally 20 kV) sent to the Pockels cell. Figure 6-88(b) is an oscilloscope trace of a typical switched-out optical pulse, indicating better than 10-ns rise time for our 5- $\times$ 5-cm-aperture Pockels cell. The performance was limited primarily by plasma resistance, feed-through inductance, and the stray capacitance (~ 25 pF) of the Plexiglas crystal holder. An advanced cell design that minimizes stray capacitance and has

Fig. 6-88. (a) Voltage pulse applied and (b) optical pulse switched out by the 5- $\times$ 5-cm-aperture Pockels cell.

low-inductance feed-throughs at the electrodes would be expected to have an optical rise time of from 3 to 4 ns when pulsed by a 50-Ω driver.

A variation of the preionization current was observed to change the initial plasma resistance and thereby affect the voltage and optical rise times. Figure 6-89 is a plot of the observed optical rise time (from 10 to 90%) for the 5 × 5-cm Pockels cell vs preionization current. The curve shows improved rise-time performance with increased preionization.

Modeling Plasma-Electrode Parameters and Optical Performance. The plasma-electrode Pockels cell can be modeled as a resistive transmission line with a nonlinear, time-dependent plasma resistance. In our design, the voltage is applied from one side only. The propagation of a voltage pulse across the cell is thus governed by the diffusion equation

$$\frac{\partial V}{\partial t} = \frac{1}{RC} \frac{\partial^2 V}{\partial x^2}, \quad (76)$$

with appropriate boundary conditions, where R and C are the resistance and capacitance per unit length along the pulse-propagation direction x .

For short pulses, the plasma resistance per plasma sheet can be approximately expressed as

$$\frac{1}{R} = \frac{1}{R_0} + \int_0^l \frac{I(t') dt'}{A_e}, \quad (77)$$

where R_0 is the initial plasma resistance determined by the level of preionization, $I(t)$ is the instantaneous current density, and A_e is a coefficient accounting for electron avalanche. In principle, A_e can be calculated from the electron mobility and Townsend's first ionization coefficient. However, because of gas impurities and the complex electrode geometry, it is best to determine the value of A_e empirically.

The initial plasma resistance was estimated from the voltage-current characteristics of the preionization circuit. Since the preionization pulse was sufficiently long for a steady-state discharge to occur, R_0 was

Fig. 6-89. Comparison of experimental data with model calculations of optical rise time from 10 to 90% transmission vs measured peak preionization current in the 5 × 5-cm-aperture Pockels cell.

only a function of the instantaneous preionization current. A resistance along the discharge-propagation direction of

$$R_0 = \frac{300}{l} \Omega \cdot \text{cm}^{-1} \quad (78)$$

at 10 Torr (with l in amperes) was found to best fit the preionization data. Using the coefficient $A_e = 3 \times 10^4 \Omega/\text{A} \cdot \text{s}$ in Eq. (77), we obtain reasonable agreement between experimental data and the optical rise time calculated from Eq. (72) and numerical modeling of Eqs. (76) and (77). The curve in Fig. 6-89 is the model calculation assuming a 50-Ω driving line and a cell capacitance of 50 pF.

Projected Performance of Large-Aperture Pockels Cells. Significant improvements in performance arise from a large-aperture plasma electrode. We now predict the voltage requirements and optical rise-time performance for several realistic large-aperture cells. In Table 6-15 we consider KDP Pockels cells with square apertures of 5, 15, and 30 cm. Since optical performance with bare-crystal surfaces is disappointing, we assume that the crystal-glass-plate construction shown in Fig. 6-87 will be used. A nominal window aspect ratio of 15:1 is assumed for all cases and should provide sufficient window thickness for mechanical strength and good optical figure.

Table 6-15 lists values for crystal dimension d , window thickness l , the ratio of external to internal half-wave voltage

Large-Aperture Optical Switches

Crystal dimensions, d (cm)	Window thickness, l (cm)	$V_{0\pi}/V_{k\pi}$	$V_{0\pi}$ (kV)	R_{driver} (Ω)	C_{total} (pF)	1.94 RC (ns)	$t_{0.95}$ (ns)
$5 \times 5 \times 1$	0.3	4	77.4	50	11.1	1.1	3.3
$15 \times 15 \times 2$	1.0	6	116.0	50	33.2	3.2	8.0
$30 \times 30 \times 3$	2.0	7.67	148.3	25	69.3	3.4	12.2

Table 6-15. KDP Pockels-cell design and projected performance.

$V_{0\pi}/V_{k\pi}$ [computed from Eq. (75)], and the resulting external half-wave voltage $V_{0\pi}$. To realistically assess optical performance, we calculate the time, $t_{0.95}$, for the optical transmission to exceed 95% over the entire crystal face.

A preionization current of 4 A per centimeter of electrode length was assumed in calculating the initial plasma resistance per sheet from Eq. (78). The high-voltage driver for the 5- and 15-cm-aperture cells was assumed to have 50- Ω impedance; the impedance for the 30-cm-aperture cell was assumed to be 25 Ω . The cell was taken as an unterminated, high-impedance load.

Conclusions. A sandwich of KDP or KD*P with fused-silica plates is a baseline large-aperture design that can yield good optical performance. However, a high-intensity-laser system requires a fixed gap between the crystal and glass-plate surfaces. This gap must then be filled with an appropriate IMF. The use of an IMF significantly complicates the mechanical design of a large-aperture, plasma-electrode Pockels cell.

Thus, we require a "hard" material with three properties: an electrooptic coefficient and a dielectric constant lower than those for KDP (the two are usually related), and good material properties for high-intensity-laser operation. Such a material would require neither an IMF nor a sandwich geometry. Our search for a thin-film, AR-coated material is in progress.

Authors: J. Goldhar and M. A. Henesian

Major Contributors: J. A. Lutz and R. A. Pasha

Plasma-Electrode Harmonic-Conversion Switch

Introduction. A new type of electrooptic switch with potential application in high-power-laser systems is based on electrooptically tuned SHG.⁹⁰ In our concept for such a switch, transparent plasma electrodes^{91,92} apply a uniform electric field to a harmonic-generation crystal in the direction of

propagation of the fundamental and harmonic light. The field alters the ordinary- and extraordinary-ray phase velocities and thereby changes the phase matching of the harmonic field to the driving polarization. Thus, the voltage applied to the transparent electrodes can tune the efficiency of conversion from the fundamental to the harmonic wavelength. Because this is a longitudinal electrooptic effect (linear in the applied electric field strength), the half-wave voltage required to switch the harmonic generation is independent of crystal length. Therefore, the length of the harmonic crystal can be optimized for the expected range of fundamental intensities and is not constrained by voltage requirements for switching.

An important application of a large-aperture electrooptic harmonic switch would be to provide switchout from a high-power regenerative or multipass amplifier.⁹⁰ The layout with a multipass amplifier is shown schematically in Fig. 6-90. The switchout voltage is independent of the harmonic-crystal aperture because the electric field is applied in the axial (propagation) direction. With plasma electrodes,⁹¹ the technique can be scaled to the large amplifier apertures⁹² required in many high-power-laser applications, such as ICF.

To investigate longitudinal electrooptic control of harmonic generation, we have demonstrated an electrooptic harmonic switch using Type I KDP with plasma electrodes. Type II harmonic generation in KDP-isomorph crystals will not show a linear electrooptic effect with axial electric fields. A modified form of Type II harmonic generation can, however, be used for an electrooptic harmonic switch. The modified form of Type II generation allows wider angular acceptance and dynamic range for the harmonic-generation process.

Theory. The SHG efficiency, $\eta(l) = I_{2\omega}(l)/I_{\omega}(0)$ for plane-wave interactions in transparent, noncentrosymmetric crystals is given by⁹³

$$\eta = \frac{1}{4} \delta^2 I_{\omega}(0) \frac{\sin^2\left(\frac{\Delta k l}{2}\right)}{\left(\frac{\Delta k l}{2}\right)^2} \quad (79)$$

in the low conversion regime, where $I_{2\omega}(l)$ is the second-harmonic intensity at position l

Fig. 6-90. Layout of a high-power multipass amplifier using a large-aperture electrooptically tuned harmonic switch.

inside the crystal and $I_{\omega}(0)$ is the fundamental intensity at the entrance face. Here,

$$\delta = \left(\frac{2}{\epsilon_0 c n_{2\omega}} \right)^{1/2} \frac{4\pi}{n_{\omega} \lambda} d_{\text{eff}} \quad (80)$$

is the second-harmonic-coupling coefficient (mks units), and $\Delta k = k_{2\omega} - 2k_{\omega}$ is the wave-vector mismatch.

In KDP-isomorph crystals (negative-uniaxial, 42m-point group), the effective nonlinear coefficient, d_{eff} , is $d_{36} \sin 2\phi \sin \theta$ for Type I SHG and is $-d_{36} \cos 2\phi \sin 2\theta$ for Type II SHG (Ref. 93). Here, θ is the phase-matching angle to the optic axis, and ϕ is the azimuthal angle to the x crystal axis. Type I harmonic generation is, therefore, optimized for propagation in a plane at 45 deg ($\phi = \pm \pi/4$) to the x - z plane, and Type II is optimized for propagation in the x - z ($\phi = 0$) or y - z ($\phi = \pi/2$) planes.

In Type I harmonic generation, the fundamental wave is polarized as an ordinary ray, and the second-harmonic wave propagates as an extraordinary ray. In Type II generation, however, the fundamental wave is polarized both as an ordinary and an extraordinary ray, and the second-harmonic wave is generated as an extraordinary ray. When an electric field is applied along the direction of propagation of the beams, both the ordinary and extraordinary refractive indexes are altered by the linear electrooptic effect. The new refractive indexes may be calculated from the Fresnel equations for propagation in anisotropic media, with the appropriate perturbations to the dielectric tensor, or by considering the refractive-index perturbations to the index ellipsoid.⁹⁴

For an applied electric field along the propagation direction, with components $E_x = E \sin \theta \cos \phi$, $E_y = E \sin \theta \sin \phi$, and $E_z = E \cos \theta$, the ordinary index at the fundamental frequency becomes

$$n_{\omega}^o = n_{\omega}^o + \frac{(n_{\omega}^o)^3 r_{63} E}{2} \cos \theta \sin 2\phi, \quad (81)$$

and the extraordinary index at the second-harmonic frequency becomes

$$n_{2\omega}^e(\theta) = n_{2\omega}^e(\theta) - \frac{[n_{2\omega}^e(\theta)]^3}{2} \times (r_{63} \cos^2 \theta - 2r_{41} \sin^2 \theta) E \cos \theta \sin 2\phi. \quad (82)$$

Note that the perturbation vanishes for propagation at 90 deg to the optic axis ($\theta = \pi/2$) or for propagation in the x - z ($\phi = 0$) or y - z ($\phi = \pi/2$) planes. Thus, the longitudinal electrooptic effect is inoperative for noncritical 90 deg phase matching and optimized Type II harmonic generation. A modified form of Type II phase matching that uses an intermediate angle, ϕ , between 0 and $\pi/4$ will show a useful longitudinal electrooptic effect and still maintain the advantages of Type II harmonic generation.

Equation (79) can be written in terms of an applied voltage $V = El$ across the crystal entrance and exit faces and a half-wave voltage V_{π} for electrooptic harmonic switching. Then,

$$\eta = \frac{1}{4} \delta^2 I_{\omega}(0) \frac{\sin^2 \left(\frac{\pi V}{V_{\pi}} \right)}{\left(\frac{\pi V}{V_{\pi}} \right)^2}, \quad (83)$$

assuming that the crystal is angle phase matched for SHG in the absence of the field. A voltage of V_{π} then fully detunes the crystal to the first zero of the phase-matching curve. The half-wave voltage for Type I harmonic generation then becomes

$$V_{\pi} = \lambda \left[n^3 \cos \theta \sin 2\phi \times [r_{63}(1 + \cos^2 \theta) - 2r_{41} \sin^2 \theta] \right]^{-1}, \quad (84)$$

where $n = n_{\omega}^o = n_{2\omega}^e(\theta)$ at the phase-matching angle. The half-wave voltage for Type II harmonic conversion is

Large-Aperture Optical Switches

Table 6-16. Electrooptic coefficients and half-wave voltages for a longitudinal Pockels cell, Type I SHG, and modified Type II SHG.

Crystal	r_{63} (10^{-12} m/V)	r_{41}/r_{63}	$-V_{\pi}$ Pockels (kV)	$-V_{\pi}$ Type I SHG (kV)	$-V_{\pi}$ Type II SHG $\phi = 22.5$ deg (kV)
KDP	-10.5	+0.82	15 (19) ^a	47 (79)	>500
90% KD*P	-22.7	+0.48	6.9 (7.4)	16 (18)	137 (160)
ADP	-8.5	+2.45	19 (29)	-89 (-44)	-87 (-73)
~80% AD*P	-10.3	+3.88	16 (21)	-23 (-19)	-38 (-35)

^aValues in parentheses are high-frequency (clamped) values.

approximately twice the value given by Eq. (84) when the appropriate azimuthal and phase-matching angles for the Type II process are used.

Table 6-16 lists the computed half-wave voltages for a longitudinal Pockels cell, Type I SHG, and modified Type II SHG (at $\phi = 22.5$ deg). The values were computed for KDP, 90% KD*P, ammonium dihydrogen phosphate (ADP), and ~80% ammonium dideuterium phosphate (AD*P). The electrooptic coefficients listed in the table are unclamped (low-frequency) values from the literature,⁹⁵ and the sign of the r_{41} coefficient is assumed to be the same as the r_{63} coefficient, in accordance with present experimental results in Type I KDP. Notice that, because of the wide variation in the ratio r_{41}/r_{63} among the KDP isomorphs, each material is optimal for a different application. For a Type I second-harmonic switch, KD*P has the lowest half-wave voltage of 16 kV. For a Type II switch, AD*P with 38 kV is considerably lower than either KD*P or ADP. For a Type II switch, KDP has a predicted half-wave voltage in excess of 500 kV.

For practical implementation in a multipass amplifier, pulsed electric fields must be applied, and optimum crystal lengths are required for efficient conversion and switchout. In this case, as shown in parentheses in Table 6-16, the half-wave voltages will change because the clamped (high frequency) r_{63} coefficients for the KDP-isomorph crystals are lower than the unclamped values. The r_{41} coefficient, on the other hand, depends little on frequency because the piezoelectric-elasto-optic coupling for this tensor element is very small. In the high-conversion regime, the harmonic efficiency falls off with wave-vector mismatch more rapidly than indicated by the low-conversion solution. Therefore, the high-conversion equations must be used to determine optimal wave-vector mismatch

and voltage applied for switchout. For instance, at high conversion (90%), the wave vector mismatch and the switchout voltage are reduced as much as 30%. We are continuing our investigations in this area.

Experiments. Using a Type I KDP crystal mounted in a plasma-electrode cell, we have verified the concept of electrooptically tuned harmonic generation. The experimental layout is shown in Fig. 6-91. A 2.5×1 -cm Type I KDP crystal (phase-matching angle $\theta = 41.21$ deg) was potted into a Plexiglas-holder disk using a two-part silicone electronic-potting compound. The disk was placed between the two ring electrodes of a cylindrical-discharge cell.⁹¹ The entrance and exit faces of the crystal were exposed directly to the glow-discharge plasma without any noticeable degradation of the crystal surfaces. Four needle points were attached to each ring electrode to facilitate gas breakdown. Voltages up to 30 kV dc could be applied for a limited time (~5 min) before the crystal suffered electrical breakdown and physical damage. With dc excitation of the plasma, a weak current flows through the plasma due to the finite conductivity of the KDP ($\sim 2 \times 10^{10}$ $\Omega \cdot \text{cm}$). Under these conditions, a finite voltage drop was observed across the plasma. The voltage drop was a weak function of gas pressure and composition and was measured at ~2.5 kV for most operating conditions. For high-voltage insulation, the discharge cell was mounted inside a larger Plexiglas tube, which was then mounted to a rugged stepper-motor-driven mirror mount.

A 10-Hz Nd:YAG laser, operating at $1.06 \mu\text{m}$, was used to generate the second harmonic in the KDP crystal. The efficiency of energy conversion to the second harmonic was approximately ~0.5% with an unfocused fundamental beam. A small pinhole was placed in front of the detector to select the central, collimated portion of

the second-harmonic beam. The KDP crystal was oriented in the discharge cell with its optical axis in the horizontal plane. The phase-matching angle for peak harmonic conversion was established by scanning the second-harmonic (i.e., Maker) fringes generated in the crystal across the detector pinhole and interpolating between the first zeros. The angular rotation of the crystal about the vertical axis was absolutely calibrated in internal angle by scanning interference fringes of an external second-harmonic beam across the detector pinhole and then comparing with theoretical calculations. The fringes were generated by placing crossed polarizers (at $\pm 45^\circ$ deg from the vertical) around the cell and illuminating the crystal with a divergent beam of second-harmonic light generated at the Nd:YAG laser. The alignment of the optical axis in the horizontal plane was verified by orienting the crystal to obtain vertical fringes.

Results and Discussion. With the Type I KDP crystal oriented at the phase-matching angle (to better than ± 0.1 mrad), the second-harmonic signal was recorded as a function of voltage applied to the discharge cell. The experimental data are shown in Fig. 6-92 after subtracting the nominal 2.5-kV drop across the plasma. The theoretical curve, plotted over the data points, indicates a half-wave voltage of ± 52 kV to detune the second harmonic to the first zero of the phase-matching curve. The high half-wave voltage indicates that the two KDP electrooptic coefficients, r_{63} and r_{41} , must have the same sign. The sign of the r_{63} coefficient is known to be negative.⁹⁵ The set of unclamped (low-frequency) electrooptic coefficients that fits our data is $r_{63} = -9.9 \times 10^{-12}$ m/V and $r_{41} = -8.6 \times 10^{-12}$ m/V (Ref. 95). These values differ somewhat from the commonly accepted values for KDP shown in Table 6-16.

The second-harmonic signal was also recorded as a function of angular detuning from the phase-matching angle at fixed voltages of ± 22.5 kV applied across the crystal. Experimental plots of these data are shown in Fig. 6-93. The corresponding theoretical curves were computed with the same r_{63} and r_{41} coefficients found to fit the data in Fig. 6-92. Figure 6-93 shows that an applied voltage of ± 22.5 kV shifts the peak of

▲ Fig. 6-91. Experimental layout demonstrating electrooptic control of harmonic generation.

Fig. 6-92. SHG signal vs voltage applied to a KDP crystal for Type I harmonic generation.

the phase-matching curve approximately halfway to the first zero of the 0-kV curve, thus confirming that a voltage of ~ 50 kV is required to switch a $1.06\text{-}\mu\text{m}$ beam by

Fig. 6-93. Normalized SHG vs angular detuning in Type I KDP for three applied voltages.

Large-Aperture Optical Switches

harmonic conversion to $0.53\text{ }\mu\text{m}$. The excellent agreement of theoretical and experimental curves indicates that thermal drift due to resistive heating was not a significant factor. We estimate a shift in phase-matching angle of 0.0238 mrad/kV for a 1-cm-thick Type I KDP crystal from our data.

Conclusion. We have demonstrated electrooptic control of harmonic generation using Type I KDP inside a plasma-electrode discharge cell. Using plasma electrodes, strong electric fields can be applied to large-aperture electrooptic harmonic switches required by high-power, multipass-amplifier architectures.

Authors: M. A. Henesian, J. Goldhar, and R. A. Haas

Major Contributors: J. A. Lutz and R. A. Pasha

Subharmonic Injection into Multipass Cavities

In multipass or regenerative cavities, a method is needed to inject the seed pulse into the cavity. We have considered leaky mirrors, polarization switches, and reflections from foils in the Q switch to accomplish injection. We have shown that we can use noncollinear subharmonic generation for seed injection and that the method has four advantages. Such injection is

- Passive, because there are no electrical components.
- Inexpensive, because it requires no new components inside the cavity.
- Characterized by isolation because the cavity is decoupled from the driver.
- Accomplished by a small driver (10 MJ) can be driven by a 1- to 10-kJ driver).

The basic layout for noncollinear subharmonic injection is shown in Fig. 6-94. The design includes a frequency-conversion array and a dichroic mirror. On the last pass, the red beam is converted to the second harmonic and passes out through the dichroic mirror. Our present design for the array is a quadrature thin-thick scheme in which two KD*P crystals, nominally 1 cm and 4 cm thick, are arranged in quadrature. (See "High-Dynamic-Range Frequency Conversion," earlier in this section.) The design will achieve high conversion efficiency ($>80\%$) over a large dynamic range (0.03 to 3 GW/cm^2) if the beam quality is sufficiently high ($\pm 50\text{ }\mu\text{rad}$).

The red beams impinging on the crystals are converted collinearly to 2ω . The phase-matching angle is 59.18 deg , and the beam polarization is midway (45 deg) between the ordinary (o) and extraordinary (e) configuration. The injection scheme involves irradiating a quadrature array with two low-intensity beams, one green and one red, both noncollinear to the cavity axis. The beams mix in the thick quadrature crystal

Fig. 6-94. Scheme for noncollinear subharmonic injection.

and generate a weak red beam directed along the cavity axis. The beams do not interact in the thin crystal.

Two injection schemes are possible, depending on the polarization of the weak red beam injected into the cavity. In both schemes, the incoming green beam is polarized so that it is extraordinary in the thick crystal. In the e-injection scheme, the injected red beam is extraordinary, and the incoming red beam is ordinary in the thick crystal. In the o-injection scheme, the injected red beam is ordinary, and the incoming red beam is extraordinary in the thick crystal. The schemes are labeled according to the polarization of the red beam, namely the red beam propagating collinearly with the cavity axis.

At first, it was not clear that such a noncollinear configuration would phase match because the crystal axes and the direction of the injected beam are dictated by the switchout scheme. To achieve phase matching, we must use the azimuthal angle, ψ . We define the azimuthal angle as the angle between two planes. The first plane is that formed by the cavity axis and the z axis of the thick crystal. The second plane is that formed by the two incoming noncollinear beams. When the two planes coincide, ψ is either 0 deg or 180 deg. We define ψ as zero when the green beam makes an angle with the z axis of the thick crystal that is less than the phase-matching angle.

To visualize this, suppose that the cavity axis lies in the plane of a desk top, and that the z axis of the thick crystal extends from the desk at ~ 60 deg so that the plane it makes with the cavity axis is vertical. Both incoming green and red beams pass down through the desk surface at the point where the z -axis vector touches the desk top. Both incoming beams propagate in the same direction as the injected red beam, which is along the cavity axis.

In the configuration just described, and with all beams coplanar with the z axis, the azimuthal angle is zero. The angles that the incoming beams make with the cavity axis are listed in Table 6-17. The angles required for the e-injection schemes are thoroughly compatible with all configurations and represent the most noncollinear cases.

For any other azimuthal angle, the incoming red and green beams must be closer

to the cavity axis. We can use approximate formulas for the noncollinear phase-matching angles such that

$$\theta_{red} \approx 2\theta_{green} \quad (85)$$

and

$$\theta_{green} \approx \theta_{green\ max} \cos \psi \quad (86)$$

Equations (85) and (86) apply to both schemes and can be derived analytically from the phase-matching condition under the assumption that both θ_{red} and θ_{green} are small. The results of a numerical evaluation of ψ and θ_{green} as functions of θ_{red} are shown in Figs. 6-95 and 6-96, respectively.

Scheme	θ_{red} (deg)	θ_{green} (deg)
e injection	5.50	2.77
o injection	1.01	0.51

Table 6-17. Maximally noncollinear injection.

Fig. 6-95. Azimuthal angles for the nonaxial 1ω beam in subharmonic injection.

Fig. 6-96. Injection angle for the 2ω nonaxial beam.

Large-Aperture Optical Switches

Three technical issues must still be addressed in the design of a noncollinear injection scheme. The issues are:

- Angular sensitivity of the scheme.
- Polarization of the injected beam.
- Amount of required input energy.

The angular sensitivity of the noncollinear scheme is about the same as that for Type II SHG. The device operates at very low power; thus, the angular tolerance is several hundred microradians. The beam quality of the injected signal is largely dependent on the quality of the red and green input beams. The angular sensitivity of even the thick quadrature crystal is not important in system design.

In either the e- or o-injection schemes, the polarization of the injected signal is, by definition, either pure e or pure o. On the other hand, the beam in the cavity must be midway between these polarizations for the quadrature array to work at switchout. However, we need not actively rotate the polarization of the injected signal.

The crystal array is aligned so that the polarization that experiences the highest transmission through the disks (i.e., experiences the zero at Brewster's angle) is appropriate for high performance of the quadrature array. That is, this polarization is at 45 deg to the plane containing the optic axis (at either crystal) and the cavity axis. The other polarization experiences considerable loss on going through the disks. The disks, therefore, form a polarizing amplifier. The injected signal from either noncollinear scheme is polarized at 45 deg to the cavity polarization, which experiences least loss in the disks. The power, P_{inj} injected into the

cavity polarization is, therefore, equal to that injected into the other polarization. We can ignore the half that goes in with the other polarization and simply include a factor of 1/2 as the injected signal. Thus, $P_{inj} = P_r/2$, where P_r is the output from the crystals. The remaining half is simply lost by reflection long before it can deplete any stored energy from the disks.

In the thick crystal, the red output intensity is $I = C^2 I_{red} I_{green}$, where $C \approx 1 \text{ GW}^{-1/2}$ and $l \approx 4 \text{ cm}$. If the gain I_{red}/I of the multipass cavity is $\sim 10^3$, then we need $I \approx 10 \text{ kW/cm}^2$, which implies that I_{red} and I_{green} can both be $\sim 1 \text{ MW/cm}^2$. Thus, we could drive a 10-MJ laser with an ~ 1 - to 10-kJ input driver.

The scheme also permits pulse shaping of the injected pulse either by shaping the red incoming beam and holding the green constant or vice versa. All the pulse-shaping could be accomplished on the red beam, leaving the green beam constant. Since the green beam must be the second harmonic of the driver, we can simplify pulse-shaping the driver by separating pulse-shaping from frequency conversion. Alternately the dynamic range (max/min) of the injected pulse can be achieved with two identically shaped red and green beams, in which the dynamic range is the square root of that required of the injected signal. Pulse shaping the green beam may then not be a problem. Clearly, there are several options for pulse shaping. Perhaps identical shaping of both beams leads to the simplest system.

A system layout for our injection scheme is shown in Fig. 6-97. Note that the driver is thoroughly isolated from the main cavity.

Fig. 6-97. Layout of scheme for noncollinear subharmonic injection in a 10-MJ system.

(Telescopes to enlarge input beams to size are not shown.)

In summary, noncollinear subharmonic generation can be used to inject multipass cavities. The method is passive, simple, and inexpensive, and is accomplished by isolation of a small driver.

Author: D. Eimerl

Advanced Solid-State Systems

Theory of the Electronic Structure of Impurities in Crystals

Characterization of the electronic structure of impurities in solids affects a large class of phenomena and devices. The electronic states associated with impurities are localized on a very few atoms. Because their energy levels are usually well separated from those of the host crystal, electronic spectroscopy of transition metal and rare-earth ion impurities offers a wealth of information. The theoretical methods used to describe the spectra and electronic structure of the d- and f-series atoms in solids do not differ in principle from those used to treat molecules. In nearly all previous theoretical studies, we have used crystal-field theory, ligand-field theory, or molecular-orbital theory, which differ in the detail of their respective treatment of the atoms surrounding the impurity.

Crystal-field theory has been widely applied and is of enormous importance to the interpretation of electronic spectra. The crystal-field theory is actually a model in which the parameters must be determined by *ab initio* calculations or experiments. During 1983, we concentrated on extending and improving our ability to carry out *ab initio* calculations on metal impurities in crystals. We have also made extensive crystal-field analyses of transition metal ions in various crystal hosts.

Our ability to predict and analyze the optical spectra of impurity metals in crystals with *ab initio* methods is far behind our ability to make similar predictions for molecules. Although *ab initio* molecular orbital methods were applied to the problem of

optical spectra of impurity metals in crystals over 20 years ago,^{1a} present-day calculations continue to give large errors in the fundamental parameters derived from the predicted energy-level splittings. The size and speed of available computers and codes limited the size of the basis sets used, which, in turn, limited the accuracy of the calculations. Although computer size and speed are no longer serious restrictions, the lack of quantitative agreement with experimental data remains. In addition to the quality of the basis set used for the orbital expansions, other factors that need to be considered are: the effects of the crystal lattice beyond the nearest neighbors, shifts due to relativistic corrections and spin-orbit coupling, and the contributions to the energy-level splittings from configuration interaction. We are including these factors in our calculations on transition metals in various crystal environments.

We are carrying out open-shell Hartree-Fock calculations, with large basis sets, on atom clusters to simulate the important effects of the crystal field and covalent bonding on the energy levels of the metal. We are also developing a new configuration interaction code that will include the effects of both electron correlation and spin-orbit coupling. We will introduce other important relativistic contributions by using an effective core potential derived from Dirac-Fock calculations on the free ion to replace the core electrons of the metal.

To understand the radiative properties of vanadium-doped magnesium fluoride, $V^{+2}:MgF_2$ and $V^{+3}:MgF_2$, we carried out Hartree-Fock molecular orbital calculations on the VF_6 cluster for both positive ions: for the vanadium ion we used Wachters' Gaussian basis sets^{2b}; and for the fluorines, we used the basis set proposed by Kahn *et al.*^{3a} along with an effective core potential to replace the $1s^2$ inner-shell electrons. All electrons on the metal were explicitly treated in the calculation.

The ground configuration of the V^{+2} ion is t_{2g}^3 , which gives rise in the cluster to the ${}^4A_{2g}$, 2E_g , ${}^2T_{1g}$, and ${}^2T_{2g}$ states. We also calculated the average of the excited 4T_2 and ${}^4T_{1g}$ states that come from the $t_{2g}^2 e_g$ configuration. The ground configuration for the V^{+3} ion, which produces the ${}^1T_{1g}$, 1E_g , ${}^1T_{2g}$, and ${}^1A_{1g}$ states, is t_{2g}^2 . We also calculated the average of the ${}^3T_{1g}$ and ${}^3T_{2g}$ states produced

Fig. 6-98. The Hartree-Fock energy level splittings for the t_{2g}^3 and $t_{2g}^2 e_g^1$ configurations of V^{+2} and V^{+3} in the octahedral field of $6F^-$.

by the $t_{2g} e_g$ excited configuration. The energy-level diagrams for the two ground-state configurations are shown in Fig. 6-98.

From the energy-level splittings, we derive the Racah parameters B and C in the combination $3B + C$ and compare them with the values from crystal-field fits to the $V^{+2}:\text{MgF}_2$ spectra.⁹⁹ The crystal-field calculations predict $B = 681 \text{ cm}^{-1}$ and $C = 2630 \text{ cm}^{-1}$, which give $3B + C = 4673 \text{ cm}^{-1}$. Depending on which splitting is used to find $3B + C$, the ab initio calculations give 6058 and 6204 cm^{-1} for V^{+2} . Fits of comparable accuracy are not available for the V^{+3} ion; however, we expect the trend to be the same.

The lack of agreement between theory and experiment illustrates the need for improved theoretical methods. The larger value for the splitting, found from the Hartree-Fock calculations, is consistent with the errors expected when comparing a high-spin state with states of lower multiplicity. A calculation that includes electron correlation would lower the energy of the excited states more than the ground state and lead to smaller values of the Racah parameters. The crystal-field parameter Dq can be calculated from the splitting between the ground state and excited state with the e_g orbital occupied. For the V^{+2} ion, the crystal-field fits give $Dq = 1045 \text{ cm}^{-1}$; the Hartree-Fock calculations predict 914 cm^{-1} . Considering that we are calculating an average of the ${}^4T_{1g}$ and ${}^4T_{2g}$ states, the agreement is good. The correlation energy differences contribute less in this case because the states have the same spin multiplicity. The Hartree-Fock value of Dq for V^{+3} is 1346 cm^{-1} . Fits to the V^{+3} spectra in phosphate glass¹⁰⁰ give $Dq = 1467 \text{ cm}^{-1}$, and McClure¹⁰¹ has found $Dq = 1750 \text{ cm}^{-1}$

in corundum, which would be expected to give a somewhat larger value than for fluoride crystals.

We carried out Hartree-Fock calculations for each of the above states at several V-F distances. The resulting potential-energy curves for V^{+2} are shown in Fig. 6-99, and those for V^{+3} are shown in Fig. 6-100. Each curve corresponds to the symmetric stretch vibration of the cluster. The calculated equilibrium internuclear distance for V^{+2} is 2.41 Å, which is somewhat longer than the 1.99-Å average spacing in the crystal. For V^{+3} the equilibrium distance is 2.02 Å, and the potentials are more strongly dependent on the internuclear separation. We compared the crystal-field splittings to the ab initio calculations given above for V^{+2} and shifted the ${}^4T_{2g}$ and the ${}^2T_{1g}, {}^2E_g$ states, as indicated by the red curves. We used a similar procedure for the ${}^3T_{1g}$ and ${}^1T_{2g}, {}^1E_g$ states of V^{+3} . The upper level for the fluorescence from each ion is given by the dashed curves. The fluorescence spectrum and lifetime of $V^{+2}:\text{MgF}_2$ have been investigated as a function of temperature, and it has been shown that the amount of emission rapidly decreases as the temperature increases.¹⁰¹⁻¹⁰⁴ The fluorescence lifetime shows only a weak dependence on temperature below 200 K and decreases rapidly at higher temperatures. Figure 6-99 shows that the upper level for the fluorescence lies 3569 cm^{-1} below the ${}^2T_{1g}, {}^2E_g$ states. This is close to the 3180- cm^{-1} activation energy estimated for the nonradiative decay of the ${}^4T_{2g}$ state,¹⁰⁵ which suggests that at lower temperatures the lifetime of the ${}^4T_{2g}$ state is determined by radiation from the low-lying vibrational modes to the ${}^4A_{2g}$ ground state. At upper levels, the higher temperatures of the ${}^4T_{2g}$ state are populated and may

intersystem cross to the ${}^2T_{1g}$ and 2E_g states or undergo radiationless decay to the high vibrational levels of the ground ${}^4A_{2g}$ state. The shifted curves for the V^{+3} ion show that the ${}^3T_{1g}$ state is above the ${}^1T_{2g}$ and 1E_g states, which means there is no activation energy for the intersystem crossing to the ${}^1T_{2g}$ and 1E_g states. We need to conduct a thorough theoretical study of the fluorescence from these crystals to put our suggestions on a firm basis. We must also include the effects of spin-orbit coupling and electron correlation, and must calculate the potential energy surfaces for the proper modes of the crystal.

The comparison between ab initio theory and experiment can make effective use of crystal-field analysis. The comparison is meaningful only if the crystal-field fit accurately reflects the experimental data. We generated Tanabe-Sugano diagrams for all 3d, 4d, and 5d transition-metal ions in hosts that appear to have potential laser uses. In previous studies, the free-ion ratio of the Slater parameters F^4/F^2 , or the equivalent Racah parameters C/B , are frequently assumed for the 3d transition-metal ions. For the 4d and 5d series there are no free-ion data, so we have used various extrapolations to obtain free-ion values. Attempts to fit the parameters to the crystal data have frequently led to spurious values.

Recent work has shown that U^{4+} behaves in many ways like a d transition-metal ion in that the crystal field is very strong, and the values of the Slater parameters are very difficult to determine uniquely.¹⁰⁵⁻¹⁰⁹ In these cases the Slater parameters are about 80% of their free-ion values, and the ratio F^4/F^2 is much larger than that in the free-ion. Furthermore, the addition of the effective parameters to account for configuration interaction is necessary to obtain a reasonable fit to the data. The latter addition has rarely been included in the analysis of the spectra of d transition-metal ions in crystals, although its use in the analyses of free-ion spectra is routine. From these comparisons, we conclude that the use for crystals of free-ion values of F^4/F^2 is certainly incorrect and that we need to explore the role of the configuration interaction parameters in the d series.

Because there are no free-ion results to provide guidelines as to the parameter

Fig. 6-99. The symmetric-stretch potential-energy curves for the states of VF_6^{4-} .

Fig. 6-100. The symmetric-stretch potential-energy curves for states of VF_6^{3-} .

values for the 4d and 5d series, a bootstrap process is necessary to avoid spurious results. Preliminary calculations on Re^{4+} and Os^{6+} indicate that, with the inclusion of the configuration interaction parameters, we can obtain least-squares fits to data with rms deviations of $\sim 100 \text{ cm}^{-1}$ for these 5d transition-metal ions. These results can be confirmed only by work on other ions or in other hosts. Since the data are so limited, only consistency of parameters in a number of systems can lead to reasonable assurance that spurious results have been avoided.

Authors: N. W. Winter and K. Rajnak

Major Contributor: R. M. Pitzer

High-Average-Power, Solid-State Crystalline Lasers

During recent years, we have been evaluating transparent fluoride crystals doped with divalent vanadium as gain media for use in high-average-power, solid-state lasers. Our initial studies,^{110,111} suggested by analysis of spectroscopic data in the literature, were carried out on vanadium-doped magnesium fluoride crystals (V:MgF₂). Although laser action was obtained, we discovered excited-state absorption at the 1150-nm laser wavelength, with a transition cross section large enough to preclude efficient operation of the laser.

We developed a simple picture of the crystal field for the electronic structure of the divalent vanadium ion in an octahedral lattice site. Our picture suggested that the offending excited-state-absorption transition would be shifted spectrally away from the laser transition upon placing the ion in a lattice site with increased crystal field strength. We identified the cubic crystal, potassium magnesium perovskite (KMgF₃), as an appropriate crystal with which to test this idea. The perovskite crystal possesses a smaller lattice constant (larger crystal field strength) than magnesium fluoride. As described in the discussion below, we have found that the magnitude of this effect is smaller in KMgF₃ than in MgF₂, but it is still too large to permit operation of the laser at adequately high efficiency.

Spectroscopic measurements on the perovskite crystal were reported last year.¹¹² We found that this crystal exhibited considerable improvement over V:MgF₂ with respect to several laser parameters: high fluorescence lifetime and quantum yield at room temperature and increased stimulated-emission cross section. Our results prompted us to grow laser-quality V:KMgF₃ crystals and to conduct laser-amplifier experiments aimed at detecting excited-state-absorption processes present in this material.

Laser-Amplifier Experiment. A 1.25-cm-long crystal of ~1%-doped V:KMgF₃, conduction-cooled on a copper heat sink, was optically pumped by the focused beam of a TEM₀₀-mode, 0.51- μ m argon-ion laser.

A probe beam from a cw, TEM₀₀-mode Nd:YAG laser, aligned by a set of apertures to be exactly collinear with the pump beam, also passed through the crystal. The probe beam was detected by a silicon photodiode whose area was large enough to minimize any effects of beam steering or lensing caused by thermal gradients in the crystal. A variable-frequency optical chopper was used to modulate the pump beam, and the resultant signal induced in the probe beam was detected by a lock-in amplifier.

Figure 6-101 shows the gain (expressed as a percentage of the 1.06- μ m probe signal) vs pump power incident on the crystal for a chopper frequency of 55 Hz. The power of the probe beam was ~100 mW. The gain reduction at high powers may be related to heating effects. Observation of the phase of the fluorescence signal from the crystal (detected by another silicon photodiode), compared to the phase of the gain signal, showed that the gain was, indeed, positive. With the probe beam blocked, no gain signal was detected, indicating that the fluorescence background was negligible.

To check that the measured signal was related entirely to stimulated emission and not to other processes, the magnitude of the gain signal at constant pump power (in the linear region of Fig. 6-101) was observed as a function of chopper frequency. Because of the relatively long lifetime of the upper laser level (~1 ms), the pump-induced modulation of the inversion density should be reduced at higher frequencies. If the pump beam was modulated with a sine wave (instead of the actual square wave) we would expect a 3-dB rolloff in amplitude at a frequency of ~1000 rad/s or ~167 Hz. To check that the gain signal had the proper frequency response, it was compared with the response of the detected fluorescence signal to varying chopper frequency. Figure 6-102 shows the relative response of both gain and fluorescence signals; the essentially equivalent response curves indicate that the gain signal was directly related to the modulation of the upper-level population.

We used a set of calibrated apertures to determine the spot size of both probe and pump beams at the crystal location. The transmission, T , of a Gaussian mode

through an aperture of diameter d can be used to find the mode-characteristic radius, w :

$$w = \frac{d}{2} \left[\frac{-2}{\ln(1 - T)} \right]^{1/2} \quad (87)$$

For the probe beam, apertures with diameters of 32, 108, and 140 μm yielded values for w of 122, 117, and 119 μm , respectively. For the pump beam, apertures with diameters of 21, 32, and 51 μm yielded values for w of 34, 36, and 31 μm , respectively. The relatively small variation in calculated spot size for apertures of different sizes indicated that the measurement technique was reasonably precise.

The measured pump-beam transmission through the crystal was 0.46. After Fresnel losses are taken into account, the value implies an absorption coefficient of 0.56 cm^{-1} . The probe-beam transmission was 0.85. With Fresnel losses accounted for, the loss coefficient at 1.06 μm is 0.065 cm^{-1} , possibly due to scattering in the crystal. Some of the measured loss at the pump wavelength is also likely to be due to scattering rather than absorption, slightly increasing the uncertainty in the following calculations.

Assuming Beer's law of attenuation of the pump beam and a Gaussian intensity distribution, the gain coefficient function in the crystal, $g(r, z)$, is given by

$$g(r, z) = \frac{2P\sigma\tau\alpha_p}{\pi h\nu_p w_p^2(z)} \times \exp \left[-2r^2 w_p^2(z) - \alpha_p z \right], \quad (88)$$

where P is the pump power incident on the crystal (corrected for the Fresnel loss), σ is the gain cross section, τ is the upper-state lifetime, α_p is the pump absorption coefficient, ν_p is the pump frequency, r is the transverse radial coordinate, and z is the distance along the length of the crystal measured from the entrance face. The pump spot size function, $w_p^2(z)$ is given by

$$w_p^2(z) = w_{p0}^2 \left(1 + \frac{\lambda_p z}{\pi w_{p0}^2} \right), \quad (89)$$

Fig. 6-101. Percentage gain in V:KMgF_3 vs pump power for a chopper frequency of 55 Hz. The curve is a best-fit straight line to data points (except at the upper right).

Fig. 6-102. Gain and fluorescence-signal levels vs chopper frequency for constant pump power.

where n is the refractive index, and where it is assumed that the beam focus is at $z = 0$. Assuming that the gain is small, the measured fractional gain, G , is given by

$$G = 4 \int_0^l dz \int_0^\infty r dr \frac{g(r, z)}{w_s^2(z)} \times \exp \left[\frac{-zr^2}{w_p^2(z) + w_s^2(z)} \right], \quad (90)$$

where l is the crystal length and $w_s^2(z)$ is the spot-size function for the probe beam.

An approximation to G can be easily obtained by neglecting the z -dependence of w_p and w_s and taking the limit of the large w_s/w_p . Then,

$$G = \frac{\epsilon P \sigma \tau}{\pi w_s^2 h \nu_p} [1 - \exp(-\alpha_p l)] \quad (91)$$

or

$$\sigma = \frac{\pi w_s^2 h \nu_p G}{2 P \tau} [1 - \exp(-\alpha_p l)]^{-1} \quad (92)$$

Using previously discussed values for the parameters above, and a value for G of 5.4×10^{-3} at a P of 0.52 W, we find a value for σ of $1.6 \times 10^{-21} \text{ cm}^2 = \sigma_{\text{meas}}$.

To test the validity of the approximations used for Eq. (92), we carried out a direct numerical calculation of Eq. (90) on a computer. Our only approximation was the assumption of a constant w_s , which is valid considering that the confocal parameter for a 120- μm spot size in the crystal is $\sim 12 \text{ cm}$. The result is a value of $1.78 \times 10^{-21} \text{ cm}^2$ for σ .

The Einstein relation can be used to relate σ at the peak of the gain curve to the fluorescence lifetime, under the assumption of unit fluorescence quantum efficiency. The equation (only approximate for a system with a broad fluorescence line, such as V:KMgF₃) is

$$\sigma = \frac{\lambda_m^4}{8\pi \Delta\lambda n^2 c \tau_{\text{rad}}} \quad (93)$$

where λ_m and $\Delta\lambda$ are the wavelengths at the peak and FWHM of the gain spectrum, respectively, and c is the speed of light. Using the available fluorescence data, we found $\sigma \approx 4.4 \times 10^{-21} \text{ cm}^2$.

The observed difference between the calculated stimulated-emission cross section and the net cross section (the difference is the observed gain coefficient) is probably due to an excited-state absorption transition. We plan an additional experiment to assess this supposition by measuring the spectral distribution of the gain-reducing transition. Although the magnitude of this effect seems to be smaller in

KMgF₃ than in MgF₂, it is still too large to permit operation of the laser at adequately high efficiency. As a result of our observations, we plan to diversify our exploratory spectroscopic and laser measurements to ions other than vanadium and to crystal hosts other than simple fluorides. Our goal is to identify new solid-state gain media with useful laser properties.¹¹³

Authors: P. F. Moulton and R. E. Fahey (Lincoln Laboratory); W. F. Krupke

Other Laser Systems

Rare-Gas-Halide Lasers: Experiments on the Rapier-B Amplifier

We described assembly and testing of the mechanical and electrical components of the Rapier-B amplifier in the 1982 *Laser Program Annual Report*.¹¹⁴ During 1983, we completed these tests and conducted a series of KrF-laser experiments using the Rapier-B amplifier (subsequently referred to as the B amplifier).

Figure 6-103 shows a front view of the B amplifier before the laser optics were installed. The laser-output aperture is a 30- $\times$ 40-cm, AR-coated, fused-silica window, and the cell depth is 120 cm. Electron beams enter the laser cell through thin, rib-supported, metal-foil windows on two facing sides of the cell. The electron beams are generated by six 450 kV, 150-ns, 90-kA water Blumlein pulse-forming lines, three on each side of the cell. The Marx banks used to charge the pulse-forming lines are in oil tanks under the pulse-forming lines. The six lines generate 36 kJ of 450-kV electrons; 14 kJ are deposited in the useful laser volume when the cell is filled with 2 atm of argon gas.

Figure 6-104 shows an isointensity plot of the ArO fluorescence across the laser aperture from a gas fill containing 2 atm of argon plus a few Torr of oxygen. The figure also shows the energy-deposition profile in the cell as calculated by the SANDYL code during our design studies for the B amplifier. The measured energy-deposition profile agrees with calculations to within the

accuracy of this measurement technique, which assumes that ArO fluorescence is exactly linear in electron-beam deposition. This assumption should be reasonably valid over the range of depositions in the B amplifier but is not valid over wide deposition ranges, as discussed in the 1975 Annual Report.¹¹⁵ Figure 6-104 confirms that electron sidescattering overcomes magnetic pinch and electron-beam nonuniformities to give a smooth deposition profile, as predicted.

The optical layout for diagnostic experiments on the B amplifier is shown in Fig. 6-105. The oscillator-amplifier system uses a KrF discharge laser, similar to those systems discussed in past Laser Annals,^{116,117} to generate a 300-nJ, 20-ns laser pulse. An etalon in the KrF oscillator limits the line width of this pulse to 5 cm^{-1} , which is about an order of magnitude less than the line width of a free-running KrF oscillator. A lens matches the curvature of

Fig. 6-103. Rapier-B amplifier before laser optics were installed.

Fig. 6-104. Energy-deposition profile across output aperture of the B amplifier. (a) ArO fluorescence, (b) digitized and linearized isointensity contours at 10% intervals, and (c) isointensity contours at 10% intervals calculated during system design.

Fig. 6-105. Experimental layout for diagnostic experiments on the B amplifier.

the wave front of the discharge-laser pulse to the unstable resonator on the B amplifier, and the pulse is then injected into that resonator. The timing is set so that the 20-ns pulse fills the unstable resonator at the time when the gain rises in the B amplifier. The input seed pulse is injected at a small angle to the axis of the unstable resonator, which allows us to use a total reflector for the convex unstable-resonator mirror, M_2 . The angle between the path of the injected pulse and the unstable-resonator axis decreases by a factor equal to the unstable-resonator magnification on each transit through the cavity. Thus, the direction of the output beam is on the unstable-resonator axis within very few round trips through the cavity. For this particular case, our measurements show that only the first two trips through the resonator are outside the normal far-field spot of the laser.

Mirrors M_1 and M_2 form the unstable resonator on the B-amplifier gain volume. M_1 is a concave mirror with a 5-m radius, a 60-cm diameter, and a coating for maximum reflectivity at $0.25\text{ }\mu\text{m}$. M_2 is a 6×8 -cm rectangular, convex mirror with a 1-m radius chosen to give an unstable-resonator magnification of $5\times$. The spacing between the mirrors is slightly greater than the

confocal distance of 2 m, so that the output beam converges to a focus at a distance of 45 m from the amplifier. This allows us to diagnose the spectrum and divergence in the far field and to use smaller optical elements for some of the beam splitters.

We measure the output-beam energy using calibrated samples of the beam: reflected from uncoated, fused-silica beam splitters. The primary calorimeter is a 30×40 -cm full-beam calorimeter with a black anodized-aluminum absorbing surface. This is checked against a commercial 15-cm calorimeter, which views a beam sample after the beam has contracted to that diameter. We photograph the near-field intensity distribution of the laser output 15 m from the laser on instant-image UV film. We also photograph the focal spot in the far field when the beam comes to a focus after propagating 45 m from the laser.

We photograph the spectrum of the laser output beam (all radiation within ~ 1 mrad of the beam axis) in the fourth order of a 1-m grating spectrograph. The part of the output beam that grows from the injected seed signal has a bandwidth of 5 cm^{-1} (similar to that of the signal), while ASE has the normal free-running oscillator bandwidth, which is about an order of

Fig. 6-106. Time dependence of the early ASE flash and unstable-resonator output beam from the B amplifier.

magnitude larger. This allows us to determine the relative energy in each of these components by studying the spectrum of the output. We record the time dependence of the early flash of highly divergent ASE using a photodiode that is outside the unstable-resonator output beam but within the ASE output beam determined by the aspect ratio of the gain volume. Another photodiode records the time dependence of the laser pulse in the unstable-resonator output beam.

Figure 6-106 shows the time dependence of the ASE flash and the laser output beam from 2 atm of a laser mixture containing argon, 5% krypton, and 0.3% fluorine. For this shot, the discharge laser injected about 100 mJ into the unstable resonator, and the ASE flash radiated ~ 20 to 50 J into a large solid angle before the injected signal grew to full saturation. The laser output beam contained 850 J, of which at least 90% had the narrow bandwidth of the injected signal rather than the wide bandwidth of the ASE signal.

Figure 6-107 is a photograph of the near-field intensity distribution 15 m from the laser on a similar shot at somewhat lower energy. The output-intensity profile is consistent with the electron-beam energy-deposition profile of Fig. 6-104.

Fig. 6-107. Near-field pattern of the B-amplifier output beam 15 m from the amplifier.

Laser-extraction kinetics of KrF are dominated by the competition between gain and absorption in the medium, as discussed in past Laser Annuals.¹¹⁸⁻¹²² Figure 6-108 shows the results of a calculation of local extraction efficiency vs local laser intensity using a recent version of a kinetic model of the KrF laser.¹²³ The average excitation rate in the B-amplifier cell during the laser pulse is $\sim 700 \text{ kW/cm}^3$, and the peak rate is perhaps 1.3 MW/cm^3 . Thus, the 1-MW/cm^3 rate used in the calculation is representative of the operating conditions in this device. The extraction efficiency first rises with

Fig. 6-108. Laser extraction efficiency vs local laser intensity for a KrF-laser mixture containing argon, 5% krypton, and 0.3% fluorine.

Other Laser Systems

laser input intensity and then falls as absorption begins to dominate the kinetics.

There are two important classes of absorbers in KrF-laser mixtures: KrF precursors, such as highly excited rare-gas atoms or the F^- and Ar_2^+ ions, and KrF deactivation products, such as the excimer Kr_2F formed from KrF. If the laser field is strong, KrF is destroyed before Kr_2F is formed so that Kr_2F absorption is reduced significantly. If the formation and loss rates of the KrF precursors are slow, photoionization of these species by the laser can also be significant.

The model prediction of maximum intensity at zero extraction efficiency is a sensitive test of how well the kinetic model handles absorption and how the absorption changes under highly saturated conditions. If there are no other cavity losses, the maximum intensity is given by $I_{max} = (\alpha/\gamma_s - 1)I_s$, where α is the unsaturated small-signal gain, γ_s is the saturated loss at I_{max} , and I_s is the laser saturation intensity. We can determine I_{max} by operating the laser in a low-loss cavity at zero extraction efficiency and measuring the power circulating in the cavity. This measurement has significant uncertainties in small devices because the total nonsaturable loss also includes cavity losses, such as window reflection and scattering, and these losses may be comparable to the absorption in the medium with short gain lengths. This is not a problem with the 120-cm gain length of the B amplifier.

We measured the maximum intensity in the B amplifier by installing a low-loss cavity consisting of two 5-cm-diam reflectors positioned in the aperture at the point of maximum electron-beam deposition ($\sim 1.3 \text{ MW/cm}^2$). At this point, we measured a maximum intensity of 46 MW/cm^2 . The kinetic model predicts 38 MW/cm^2 at 1 MW/cm^2 deposition (see

Fig. 6-108) and 44 MW/cm^2 at 1.5 MW/cm^2 . The model is in reasonable agreement with our results. Some older kinetics models that do not handle photoionization properly predict that the value for I_{max} would be $\sim 60\%$ of that measured, which is an important difference at high intensities. The differences are much smaller near the peak of the extraction curve.

In summary, the B amplifier has demonstrated laser performance in agreement with projections and has verified our ability to project KrF-laser performance to larger devices. The B amplifier has demonstrated a narrow-band, injection-locked output of 850 J at a stage gain of ~ 8000 .

Authors: J. Goldfarb, J. R. Murray, and K. S. Jancaitis (LLNL); L. G. Schlitt (Pulse Sciences, Inc.)

Major Contributors: D. A. Masquelier, J. A. Lutz, and W. F. Gee

Free-Electron Lasers: Experiments at the Electron-Laser Facility

Although they require considerable development, free-electron lasers (FELs) have the potential for very high efficiency and may be useful for an ICF reactor. We have begun to investigate their properties in an experiment at the electron-laser facility (ELF) designed to produce coherent radiation in the millimeter-wave regime. The ELF and its design goals were described in a previous Laser Annual.¹²⁴ Here, we summarize our initial experiments during 1983 at the ELF. Our first experiments concentrated on beam conditioning and beam transport through the interaction region of the FEL.

Beam current and emittance are reduced by passing the beam through a long ($\sim 2\text{-m}$) tube immersed in an axial magnetic field. The emittance of the beam emerging from such an emittance selector can be determined from the beam-envelope equation,¹²⁵ and is given by

$$\epsilon_1 = \frac{\Omega}{2c} R_{\text{ex}}^2 \quad (94)$$

where R_{ex} is the emittance-selector pipe radius, Ω is the relativistic electron cyclotron frequency in the applied magnetic field, and c is the speed of light. In addition, for a uniformly filled beam phase space, the

brightness of the beam is conserved as the beam goes through the emittance selector. The transmitted current is then given by

$$I_t = \left(\frac{\epsilon_1}{\epsilon_0} \right)^2 I_0 \quad (95)$$

where ϵ_0 is the emittance and I_0 is the current of the beam going into the emittance selector. Measurements to characterize this beam conditioning show that the transmitted current scales with B_{cw}^2 as would be predicted by Eqs. (94) and (95).

Assuming our model to be correct, we determined the initial (normalized) emittance ($\epsilon_{n0} = \beta\gamma\epsilon_0$) of the beam from the experimental test accelerator (ETA) to be $0.903 \text{ rad}\cdot\text{cm}$. The outcome is in good agreement with other emittance measurements of this beam.¹²⁶ To improve the beam quality for the FEL, we typically reduce the current from the ETA by a factor of 10 to 20 while reducing the emittance by a factor of 3 to 5. These are acceptable beam parameters for good operation of the FEL.

While trying to transport the beam through the wiggler,¹²⁴ we encountered an instability in beam motion that was catastrophic at a parametric regime (beam energy and wiggler magnetic field) where we were trying to match the FEL resonance at 8.6-mm output radiation. As indicated previously,¹²⁴ the interaction region in the ELF has a planar wiggler magnetic field. Vertical stabilization of the electron beam is provided by the off-axis axial component of this field. Horizontal stabilization is provided by horizontal-focusing quadrupoles placed along the interaction region. The quadrupoles are capable of a maximum field gradient of 60 G/cm.

The horizontal wiggler motion and vertical oscillation due to vertical stabilization are coupled through the quadrupole field for an electron vertically displaced from the wiggler midplane.¹²⁷ This effect is characterized in Fig. 6-109, which shows the horizontal displacement of the electron as a function of beam energy for various magnetic fields. Each graph shows the horizontal displacement for several values of vertical displacement from the wiggler plane.

Fig. 6-109. Horizontal displacements of electron beam vs energy for various peak wiggler magnetic fields and various vertical displacements of the beam.

Atomic and Molecular Theory

The curves show that a beam that enters the wiggler off-axis or with too large a beam diameter will be lost on the waveguide walls.

Figure 6-110 shows the fraction of current transmitted through 2 m of wiggler as a function of magnetic field. As the magnetic field increases, the transmitted current decreases. The result is in qualitative agreement with the description of the beam motion presented above. For the data presented in Fig. 6-110, a magnetic field of 4.6 kG is required for an output frequency of 34.6 GHz at which we attempt to operate the FEL. At this magnetic field, no measurable current was transmitted through the wiggler. (The minimum current detectable with the beam-current monitors is 50 A.)

Although no current was detected at the end of the wiggler magnet, microwave power was detected from the interaction region; when the FEL was run in an amplifier mode, we measured a gain of 2.3 in the microwave signal. In a superradiant mode, we

measured 2 to 4 W of output power. Assuming 50 A in the wiggler, our calculations predicted single-particle emission to be on the order of a few milliwatts. Figure 6-111 shows the spectrometer signal for the input (magnetron) signal alone and for the FEL output when operated as an amplifier. The width of the signal is the resolution of the spectrometer and not the line width of the radiation.

Improved diagnostics before and in the wiggler should aid us in better matching and centering the beam into the wiggler. We should achieve better beam-current transmission through the wiggler and subsequent increases in gain and radiated power.

Authors: T. J. Orzechowski and D. Prosnitz

Atomic and Molecular Theory

Collisional Redistribution of Radiation

Collisional redistribution of radiation has been the subject of numerous theoretical and experimental investigations because it affects the propagation of light through any gaseous medium. Recently Havey et al.¹²⁸ reported ratios of the fluorescence-line intensities of sodium atoms in an argon buffer gas pumped by a laser tuned to the far wings of the resonance transitions. They studied the behavior of the relative intensities of the D_1 ($^2P_{1/2} \rightarrow ^2S_{1/2}$) and D_2 ($^2P_{3/2} \rightarrow ^2S_{1/2}$) lines as functions both of buffer-gas pressure and laser detuning. At low pressures the branching ratio reflects the state distribution produced, after absorption, in one (or more) of the excited molecular states of NaAr. For large detunings, absorption occurs at small internuclear separations and is followed by mixing between the excited states during dissociation. Havey et al. found that as the pressure increases, the ratio becomes statistical ($D_1/D_2 \rightarrow 0.5$) independent of the detuning, because the fine-structure states become completely mixed by additional collisions occurring after collision-induced absorption. This mixing must take place on a time scale that is short

Fig. 6-110. Fraction of current transmitted through the wiggler vs peak wiggler magnetic field for a beam with $\gamma = 7.4$.

Fig. 6-111. Power measured by spectrometer for magnetron only and for the FEL when operated in an amplifier mode.

Fig. 6-112. Zero-pressure intercepts of D_1/D_2 for 'far red- and blue-wing laser excitation. (a) Blue detunings are measured from the $^2P_{3/2}$ limit and (b) red from the $^2P_{1/2}$.

compared to the fluorescence lifetime. Havey et al. have measured the zero-pressure intercepts of D_1/D_2 for a range of red- and blue-wing detunings.

We have calculated these branching ratios for the same range of photon energies by solving the quantum mechanical, coupled-channel equations for this process. The Hamiltonian for the system is given by

$$H = -\frac{1}{2\mu} \nabla_R^2 + H_{el} + H_{rad} + H_{int} = H_M + H_{int} \quad (96)$$

where μ is the reduced mass of the two nuclei, H_{el} is the electronic Hamiltonian, which includes the spin-orbit interaction terms, H_{rad} is the single-mode laser field Hamiltonian, and H_{int} is the interaction Hamiltonian for the molecule and the radiation field. We construct eigenfunctions of the uncoupled molecule-field Hamiltonian, H_M , that have good quantum numbers J , M , and n . Here J is the total electronic plus nuclear orbital angular momentum and M is its space-fixed projection. The number density of the quantized photon field is given by n . The interaction between the molecule and the field is given, in the dipole approximation, by

$$H_{int} = -\frac{h}{i} \frac{e}{m} \left(\frac{2\pi h}{\omega V} \right)^{1/2} (a_j' + a_j) \nabla_j \quad (97)$$

where V is the unit volume that defines the photon number density, a and a' are the

field creation and annihilation operators, respectively, and $-(h/i)\nabla_j$ is the component of the electronic momentum in the direction of the polarized field. We expanded our total wave function in terms of the basis functions described above and then generated a set of coupled equations by taking moments of the Schrödinger equation in the usual manner. We solved the coupled equations, using the R-matrix propagation technique and determined the transition probabilities in the dressed-state basis¹²⁹ using a modified version of the code described in Ref. 130.

From our calculations, we obtained the zero-pressure branching ratios, which we compare to the measured values of Havey et al. in Fig. 6-112. We have reproduced the reported experimental values along with our calculated intensity ratios for the same range of red- and blue-wing detunings. The agreement between the results is very good. Only in the mid-range of the red detuning do our ratios seem to fall slightly outside the experimental error bars. Further details of these calculations and additional results are given elsewhere.^{131,132}

Authors: K. C. Kulander (LLNL) and F. Rebertus (Max-Planck Institute)

Photodissociation of H_2S

Molecular photodissociation is an important process that is commonly used to initiate chemical reactions, to create laser-producing population inversions, and to separate

isotopes. The effectiveness of the photodissociation leading to these results often depends sensitively on the distributions of both internal and translational energy among the fragments. Detailed dynamical calculations involving excited molecular potential energy surfaces are required to predict these energy distributions. We have recently developed the capability to carry out such dynamical calculations for cases in which more than one excited electronic state is involved in the dissociation process.

We have studied the photodissociation of H_2S for photon wavelengths between 1850 and 2100 Å. In this energy range, the photodissociation process is indirect because the electronically excited molecular state created by the absorption of a photon is not itself dissociative, but is coupled strongly to a second, repulsive surface on which the molecule can fragment. The ground electronic state of H_2S in its C_{2v} equilibrium geometry is a $^1\text{A}_1$ state. The photon absorption in this energy range leads to an excited $^1\text{B}_1$ state that subsequently dissociates due to vibronic (nonadiabatic) coupling to the $^1\text{A}_2$ state. We studied this system because there have recently been some very detailed experimental measurements of the vibrational and rotational energy distributions for a number of photon wavelengths.^{133,134}

We performed calculations for the photodissociation of H_2S with the approximation that the H-S-H angle remains fixed. This greatly simplifies the dynamics and is quite accurate for this system because of the

relative heaviness of the central atom. The photodissociation cross section is given by

$$\sigma(\omega) = \frac{4\pi^2\omega}{c} |\langle \psi_f, \mu \cdot \epsilon \psi_i^{(-)} \rangle|^2, \quad (98)$$

where ψ_i is the bound, initial electronic and vibrational state, μ is the transition dipole coupling the initial $^1\text{A}_1$ electronic state to the excited $^1\text{B}_1$ state, $\psi_f^{(-)}$ is the final scattering state normalized to have unit outgoing flux in channel f , and ϵ is the electric field.

In the case of direct dissociation, $\psi_f^{(-)}$ is simply a scattering state for the excited electronic surface. For indirect photodissociation, $\psi_f^{(-)}$ is the solution for scattering on two coupled Born-Oppenheimer surfaces. We use the known vibrational frequencies, bond lengths, and bond angles for the ground state of H_2S to construct the initial vibrational wave functions as products of harmonic-oscillator functions in terms of the symmetry coordinates. Because the bond angle is assumed to be fixed, only the symmetric and asymmetric stretch vibrations are considered. The two excited-state surfaces are generated by using rotated Morse potentials whose parameters are chosen to model the surfaces proposed by van Veen et al.¹³³ The $^1\text{B}_1$ state has a minimum displaced from the Franck-Condon (FC) region along the symmetric stretch coordinate and has only bound vibrational levels for the photon frequencies considered. The $^1\text{A}_2$ state has a saddle point near the FC region and is completely dissociative. We integrate the coupled equations to obtain the transition probabilities using the R-matrix propagation technique.¹³⁵⁻¹³⁸ We then perform calculations for a range of photon energies and several different initial states.

Because the intermediate state is highly excited vibrationally, the coupling to the dissociative continuum, and therefore the photodissociation cross section, varies rapidly with photon energy. Thus, we see a very strong dependence of the magnitude of the absorption cross section, and of the vibrational state distribution, on the initial vibrational state. Figure 6-113 shows population distributions for the ground vibrational state and for an initial state with one quantum of excitation in the asymmetric stretch mode. Both distributions are highly structured for a particular

Fig. 6-113. HS vibrational-state distributions from the photodissociation of H_2S for two different initial vibrational states.

fixed total energy and differ significantly. In general, our results¹³⁹ agree reasonably well with the observed distributions from the ground vibrational state. More accurate evaluations of the excited-state, potential-energy surfaces and the couplings between them are needed to make a more rigorous comparison.

Author: K. C. Kulander

TDHF Theory of Molecular Collisions

The time-dependent Hartree-Fock (TDHF) method requires the explicit evaluation of the time dependent wave function of a quantum system as it evolves during a collision. The TDHF method has been helpful in developing insight into the dynamics of nuclear and atomic collisions. The TDHF method is a more efficient calculational tool than the standard state-expansion methods when applications require a large number of quantum states in the expansion set or when the most effective expansion set varies with time during the collision. The second situation arises in many collisions of composite particles where the binding and scattering energies are comparable. The TDHF method has, therefore, been useful in examining both fusion in nuclear physics¹⁴⁰⁻¹⁴¹ and charge-exchange in atomic physics.¹⁴²⁻¹⁴⁵ In view of the dynamic similarity between nuclear and molecular scattering,¹⁴⁶ it is also reasonable to expect the TDHF method to yield useful results in the study of molecular collisions.

We have applied the TDHF method to the collinear scattering of a proton, H^+ , from a hydrogen molecule, $H_2(v)$ in vibrational state v , at a collision energy of a few electron volts. The target H_2 molecule is allowed to vibrate, but due to the collinear restriction, rotations are not permitted. This collinear system allows us to study reaction, charge exchange, and vibrational excitations. The various outcomes of the collision are:

The nonreactive collisions, Eqs. (99) and (100), describe events involving charge exchange and vibrational plus translational energy exchange, while Eqs. (101) and (102) describe reactive events.

In a study of this simple, two-electron polyatomic molecule, we used the TDHF method to solve for the time-dependent electronic state and classical mechanics to solve for the nuclear motion. Because we wished to study reactive collisions over an energy range of only a few electron volts, the nuclear motion could not be approximated by either a straight line or a Coulomb trajectory, both of which had been used in past TDHF calculations. We therefore coupled the classical nuclear motion to the electronic wave function through the correspondence principle and time-developed the coupled system in a self-consistent manner. The set of coupled equations is:

$$i\hbar \frac{d}{dt} \Phi(\mathbf{r}, t) = h \Phi(\mathbf{r}, t) \\ = \left\{ -\frac{\hbar^2}{2m} \nabla^2 + \frac{e^2}{2} \int \frac{d\mathbf{r}' \rho(\mathbf{r}', t)}{|\mathbf{r} - \mathbf{r}'|} - \sum_{\lambda=1}^3 \frac{e^2}{|\mathbf{r} - \mathbf{x}_\lambda|} \right\} \Phi(\mathbf{r}, t) \quad (103)$$

$$\Psi(\mathbf{r}_1, \mathbf{r}_2, t) \\ = \frac{1}{\sqrt{2}} \Phi_1(\mathbf{r}_1, t) \Phi_1(\mathbf{r}_2, t) (\alpha_1 \beta_2 - \beta_1 \alpha_2) \quad (104)$$

$$\frac{d\mathbf{x}}{dt} = \mathbf{v} \quad (105)$$

$$\frac{d}{dt} M\mathbf{v} = -\nabla_{\mathbf{x}} \langle \psi | H | \psi \rangle \quad (106)$$

$$H = \sum_{i=1}^2 h_0(\mathbf{r}_i) + \sum_{i=1}^3 \sum_{j=1}^3 \frac{e^2}{|\mathbf{x}_i - \mathbf{x}_j|} \quad (107)$$

$$h_0(\mathbf{r}_i) = h(\mathbf{r}_i) - \frac{e^2}{2} \int \frac{d\mathbf{r}' \rho(\mathbf{r}', t)}{|\mathbf{r}_i - \mathbf{r}'|} \quad (108)$$

where a single Slater determinant is used to represent the 1Σ electronic state, h is the usual Fock-Hamiltonian, M is the proton mass, H is the total Hamiltonian, $\mathbf{x}$ and $\mathbf{v}$ are the nuclear coordinate and velocity vectors, and the other parameters have their

usual meanings. Under the assumption of cylindrical symmetry, $\Phi(\mathbf{r}, t)$ is given by

$$\Phi(\mathbf{r}, t) = R^{-1/2} \chi(R, z, t) e^{im\phi} \quad (109)$$

with the quantum number $m = 0$ for the relevant molecular state. Here R is the radial position component in cylindrical coordinates.

The numerical solution was obtained as before¹⁴²⁻¹⁴⁵ except that the nuclear coordinates were propagated at each step in time, using Eqs. (105) and (106) with an explicit time-centered integrator. The set of Eqs. (103) to (106) was thus solved self-consistently.

Figure 6-114 shows the calculated electron-density contours in arbitrary units before, during, and after a collision. Here, with solid circles marking the positions of the protons, at time $t = 34$ a.u., we see the initial state of an H_2 molecule and an H^+ approaching from the right. The kinetic energy of the scatterer is 3 eV, and the target H_2 molecule is in the $v = 2$ vibrational state. At $t = 814$ a.u., the system has nearly achieved the highest proton density of the collision. The electron density exhibits a few cycles of collective vibration along the z axis until the system emerges in its final state, shown at $t = 1669$ a.u. The final state represents a completed reaction with a fractional charge transfer.

Fig. 6-114. Density contours at (a) beginning, (b) middle, and (c) end of a collision. Contours are separated by factors of 10. The solid circles denote the positions of the protons. Axes are in atomic units.

The calculations illustrate the feasibility of our method. A full scattering calculation, yielding reaction- and charge-exchange probabilities, will involve the calculation of numerous trajectories. In that calculation, an average^{147,148} over different initial phases of the H_2 oscillator will have to be taken as is usually done for classical and quasiclassical calculations.

Authors: D. H. Tiszauer and K. C. Kulander

Resonances in Molecular Collisions

Resonances in molecular systems manifest themselves in enhanced photoabsorption cross sections, narrow peaks and dips in reaction probabilities, nonstatistical dissociation rates, and altered fragment distributions. The origin of these resonances, and of the Feshbach resonances found in electron scattering, lies in the existence of doubly or multiply excited states of some separable, zero-order Hamiltonian, H_0 . The many-body character of the Born-Oppenheimer interaction potential governs the motion of the nuclei in the molecule, and the coordinates for which the Hamiltonian becomes almost separable are generally not the interbond distances but, rather, combinations of these coordinates. Thus, finding a coordinate system for which the Hamiltonian becomes almost separable (that is, for which $H - H_0$ is small) is tantamount to identifying the resonances of the system.

Note that, for a single system, the appropriate coordinates may change for different energy ranges. The intermode coupling varies with energy because different regions of phase space become accessible at higher energies. In addition, it is not necessary that H_0 be separable in all coordinates; partial separability also can lead to resonances.

The types of Hamiltonians and resonances that are expected to be found in molecular systems have received considerable attention in recent years.¹⁴⁹ Resonances in model, but realistic, potentials have been studied by a variety of techniques, ranging from completely classical to fully quantum-mechanical. We have studied two models that represent collinear reactive systems

Fig. 6-115. Reaction probabilities, P , vs total energy, E , above the dissociation limit for the HR and MHR models.

with a heavy central atom (X) situated between two light atoms (H). The (H) atoms have equal masses. The Hamiltonian for these two models is given by

$$H = \frac{p_1^2}{2\mu} + \frac{p_2^2}{2\mu} - \frac{p_1 p_2}{m_X} + D [1 - \exp\{-(r_1 - r_{10})\}]^2 + D [1 - \exp\{-(r_2 - r_{20})\}]^2 \quad (110)$$

where r_1 and r_2 are the two bond distances and p_1 and p_2 are their conjugate momenta. The reduced mass, μ , is given by $m_1 m_X / (m_1 + m_X)$, and $D = 0.76$ eV.

The resonances in this system have been studied by Hedges and Reinhardt (HR) for $m_X/m_H = 30$ using both bound-state quantum mechanical and classical methods. We performed scattering calculations on this and on a modified Hedges and Reinhardt (MHR) model, for which $m_X/m_H = 6$, to study the reaction probability as a function of the strength of the coupling term in the Hamiltonian.¹⁵⁰ We chose m_X and m_H so that $\mu_{HR} = \mu_{MHR}$ and the asymptotic boundary conditions are the same in both models. This Hamiltonian, in the limit $m_X \rightarrow \infty$, describes two uncoupled Morse oscillators.

The momentum cross term in the Hamiltonian makes the reaction possible. In the HR model, the coupling is so weak that the resonances are very narrow, and the reaction probabilities at nonresonant energies are less than 10^{-5} . The increased coupling strength in the MHR model leads to

substantial shifts in the positions of the resonances and to an increase in their widths by orders of magnitude. These effects are shown in Fig. 6-115. Here the reactive probabilities in the two models are shown for a range of collision energies.

We have found that very large changes in the dynamics can result from simply changing the mass of the central atom because the exchange of energy between the outer two atoms, through the central atom, becomes much more efficient. The momentum transfer between the atoms becomes more effective as the ratio of the masses approaches unity. By studying the effects of coupling strength on these resonances, we can better understand the existence of structure in reactive scattering probabilities and in the infrared absorption spectra of highly excited vibrational states of polyatomic molecules.¹⁵¹

Author: K. C. Kulander

Resonance-Resonance Coupling in Dissociative Electron Attachment

The effects of coupling between two adjacent electronic resonance states of molecular negative ions have been observed in several experiments involving dissociative attachment (DA) of electrons to diatomic molecules. When the potential energy curve of one resonance is attractive (possesses vibrational levels) and that of the other is repulsive, the coupling between the resonances may lead to the appearance of vibrational structure superimposed on a broad peak in

the DA cross section. Such structure has been seen experimentally in H_2 (Refs. 152 and 153), HCl (Refs. 154 and 155), HBr

Fig. 6-116. Potential energy curves of HBr^- resonances (from Ref. 156). The low-energy $2\Sigma^+$ resonance, dissociating to $\text{H} + \text{Br}^-$, is not shown.

Fig. 6-117. Ion-yield spectra for H^- formation by electron attachment to HBr (from Ref. 156).

(Refs. 154 and 156), and some other molecules. When the potential energy curves of both resonances are repulsive, the effects of the coupling are less obvious. Nevertheless, such effects have been observed recently in experiments involving the formation of H^- by electron attachment to HBr .¹⁵⁶

In the anions of hydrogen halides (HX^- , $\text{X} = \text{F}, \text{Cl}, \text{Br}, \text{and I}$), the dissociation limit $\text{H}^-(^1\text{S}) + \text{X}(^2\text{P})$ gives rise to two resonances with $2\Sigma^+$ and 2Π symmetries in LS coupling. However, spin-orbit coupling actually gives rise to three resonances: $2\Sigma^+$, $2\Pi_{3/2}$, and $2\Pi_{1/2}$. In all cases, except for HF^- , the $2\Sigma^+$ resonance is lower in energy, so the $2\Sigma^+$ and $2\Pi_{3/2}$ states dissociate adiabatically to the limit $\text{H}^- + \text{X}(^2\text{P}_{3/2})$, whereas the $2\Pi_{1/2}$ state dissociates to $\text{H}^- + \text{X}(^2\text{P}_{1/2})$. The potential energy curves appropriate for the case of HBr^- are shown in Fig. 6-116. The vertical lines in the figure indicate the region of the ground state of HBr .

In the heavier halogen atoms, the spin-orbit splitting of the $2\Pi_{1/2}$ and $2\Pi_{3/2}$ fine-structure levels is sufficiently large that it is possible to distinguish experimentally between the processes leading to the formation of $\text{X}(^2\text{P}_{3/2})$ and $\text{X}(^2\text{P}_{1/2})$. Based on the potential energy curves and the FC region shown in Fig. 6-116, we would expect two peaks at ~ 6.1 eV and at ~ 8.3 eV, in the ion yield curve for H^- produced from HBr^- ; only the 8.3-eV peak, due to the $2\Pi_{1/2}$ state, should appear in the $\text{H}^- + \text{Br}(^2\text{P}_{1/2})$ channel. Figure 6-117 shows the experimental results of Le Coat, Azria, and Tronc.¹⁵⁶ The most striking feature of Fig. 6-117 is the appearance of an unexpected peak in the region between 5.5 and 7.5 eV in the ion-yield curve corresponding to the $\text{H}^- + \text{Br}(^2\text{P}_{1/2})$ channel.

We have formulated an explanation¹⁵⁷ of the experimental observations in terms of indirect electronic coupling between the $2\Sigma^+$ and $2\Pi_{1/2}$ resonances. Qualitatively, this coupling can occur because states that can autodetach are not stationary states of the full electronic Hamiltonian, H_{el} . In molecules, they are usually chosen to be adiabatic eigenstates of $QH_{\text{el}}Q$, where Q is a suitable projection operator. Consequently, the resonance states can interact with each other in second order, via the electron-scattering continuum into which the resonances decay. The formal theory,¹⁵⁷ shows that the resonances are coupled by the

same complex energy-dependent optical potential that is responsible for the finite lifetimes of isolated resonances. The coupling matrix element, $W_{12}(R)$, which is a function of the internuclear distance, R , has the form

$$W_{12}(R) = \langle \phi_1 | H_{el} P G_{pp}^+(E) P H_{el} | \phi_2 \rangle, \quad (111)$$

where ϕ_1 and ϕ_2 are the electronic wave functions of the two resonances, P is $1 - Q$, and the Green operator $G_{pp}^+(E)$ describes nonresonant scattering at a total energy E . With some simplifying assumptions for the case of HX^+ , Eq. (111) reduces to

$$W_{12}(R) = -i/2c(\Gamma_{II} - \Gamma_{\Sigma}), \quad (112)$$

where Γ_{Σ} and Γ_{II} are the autodetaching widths of the two resonances and c is the coefficient describing the spin-orbit coupling between the $^2\Sigma'_{1/2}$ and $^2\Pi_{1/2}$ states. Thus, in the hydrogen halide anions, H formation via DA probes the spin-orbit interactions between the resonances. We expect that the magnitude of the proposed electronic coupling is greater than that of nonadiabatic effects, which are evidently not large enough to explain the observed branching ratio¹³⁶ for $Br(^2P_{1/2})$ and $Br(^2P_{3/2})$.

Quantitative verification of the indirect electronic-coupling mechanism must await detailed calculations of the resonance parameters of the $^2\Sigma'$ and $^2\Pi$ states and of the spin-orbit mixing between the $\Omega = 1/2$ components.

Author: A. U. Hazi

Hole Localization in Inner-Shell Excitation

Ab initio calculations of valence-shell ionization potentials have shown that orbital relaxation and correlation differences usually make contributions of comparable magnitude. In marked contrast to this observation is the situation for deep-core ionization, where correlation differences (~ 1 eV) play a relatively minor role compared to orbital relaxation (~ 20 eV). Theoretical calculations have shown that this relaxation is most easily described if the $1s$ vacancy created by a K-shell excitation is allowed to localize on

one of the atomic centers.¹⁵⁸ For molecules possessing a center of inversion, this means that the molecular orbitals that best describe the final state do not transform as an irreducible representation of the molecular point group.

This point has been thoroughly investigated for the case of $1s \rightarrow \pi^*$ excitation in molecular nitrogen. In a previous study,¹⁵⁹ we found that the $1s \rightarrow \pi^*$ excited states could be adequately described at the level of single-excitation configuration interaction (CI) using broken-symmetry orbitals. The dominant effect was the relaxation of the outer valence orbitals in the presence of a localized $1s$ hole. The calculations gave two quasi-degenerate molecular states of $^1\Pi_g$ and $^1\Pi_u$ symmetry; we also obtained the generalized oscillator strengths (GOS) for these transitions as a function of momentum transfer.

Recent experimental work by Shaw et al.¹⁶⁰ and by Stefani et al.^{161,162} has prompted us to carry out further calculations on N_2 , as well as analogous investigations of $1s_v \rightarrow \pi^*$ excitation in NO and N_2O . The generalized oscillator strengths display a striking similarity and point to the essential correctness of the localized hole picture for N_2 .

In our study, we used wave functions based on limited CI expansions with molecular orbitals obtained from self-consistent field (SCF) calculations on the ground state of the molecule. The N_2 calculations have been described in detail previously.¹⁵⁹ The calculations on NO and N_2O were similarly executed.

For the N_2 calculations, localized 1σ orbitals are constructed by taking $1\sigma_{1,R} = 1\sigma_R \pm 1\sigma_u$. The calculations are then carried out in a reduced symmetry (C_{2v}). Because the 1σ orbitals have a very small overlap with the remaining molecular orbitals, the latter can still be classified by g or u symmetry, as can the total molecular wave functions, when a balanced set of reference configurations is used. The orbitals labeled $2\pi_{g,u}$ which are unoccupied in the N_2 ground state, were chosen by inspecting the virtual orbitals and selecting those having the largest spatial overlap with the atomic nitrogen $2p$ orbital.

In the case of NO and N_2O , the distinguishability of the nuclear centers of course obviates the need for explicit construction

of localized core orbitals. The N_2O ground state is linear (with internuclear separations $R_{NN} = 2.128$ a.u. and $R_{NO} = 2.243$ a.u.) and has the configuration $(1\sigma^2 2\sigma^2 3\sigma^2 4\sigma^2 5\sigma^2 6\sigma^2 1\pi^4 2\pi^4), {}^1\Sigma^+$. The $1s_N \rightarrow \pi^*$ singlet states were obtained by including all single excitations relative to the configurations $(1\sigma^2 2\sigma^2 3\sigma^2 4\sigma^2 5\sigma^2 6\sigma^2 1\pi^4 2\pi^4 3\pi), {}^1\Pi$ and $(1\sigma^2 2\sigma^2 3\sigma^2 4\sigma^2 5\sigma^2 6\sigma^2 1\pi^4 2\pi^4 3\pi), {}^1\Pi$. The NO

calculations were carried out at the equilibrium internuclear separation of 2.173 a.u. The ground state configuration is $(1\sigma^2 2\sigma^2 3\sigma^2 4\sigma^2 5\sigma^2 1\pi^2 2\pi), {}^2\Pi$. The reference configurations for the spin-allowed nitrogen core-excited states were obtained by moving an electron from the 2σ orbital into the partially occupied 2π orbital. The CI calculations included all single excitations relative to these reference configurations and were carried out for the nitrogen core excited states of ${}^2\Sigma^+$, ${}^2\Sigma^-$, and ${}^2\Delta$ symmetry.

Our limited CI calculations were aimed at treating correlation effects in the initial and final states on an equal basis. The appropriate choice of reference configurations for the excited states and the restriction to single-excitation CI incorporate the dominant relaxation effects and give transition energies of essentially Δ SCF quality that, for these systems, should be accurate to roughly 1 eV. The use of CI wave functions that employ a common set of orthonormal molecular orbitals also simplifies the calculation of such properties as the generalized oscillator strengths, which were obtained for all the spin-allowed transitions.

Table 6-18 summarizes the results of our theoretical calculations on the nitrogen $1s \rightarrow \pi^*$ excited states of N_2 , N_2O , and NO. For N_2 , the calculated vertical singlet-triplet energy splittings are 0.94 and 0.89 eV for the ${}^1, {}^3\Pi_u$ and ${}^1, {}^3\Pi_g$ states, respectively. These results can be compared with the 0.82-eV difference in energy between the $v = 0$ levels, measured by Shaw et al.¹⁶⁰ These investigators,^{160,163} however, see no evidence of separate Π_g and Π_u states. The $g-u$ splittings that we obtain are ~ 0.1 eV, which is less than the natural (Auger) decay width of the states and is, therefore, presumably unresolvable.

In N_2O , we find two ${}^1\Pi$ states separated by 4 eV. This splitting arises from the difference in the binding energies of the $1s$ electrons residing on two nonequivalent nitrogen centers.¹⁶⁴ The calculated transition energies for these states are in excellent agreement with the experimental values obtained by Bianconi et al.¹⁶⁵ using monochromatized synchrotron radiation. There is also good agreement between the calculated optical oscillator strengths and the experimental value of 0.11 obtained by Stefani et al.^{161,162}

Species	Transition energy	Oscillator strength
$N_2 (1s^{-1}1\pi_g), {}^1\Pi_u$	399.06 (400.86 ^b)	0.236 (0.195 ^b)
$(1s^{-1}1\pi_g), {}^1\Pi_g$	399.12	—
$(1s^{-1}1\pi_g), {}^3\Pi_u$	398.12 (400.02 ^c)	—
$(1s^{-1}1\pi_g), {}^3\Pi_g$	398.23	—
$N_2O (1s_N^{-1}3\pi), {}^1\Pi$	400.6 (401.16 ^c)	0.090 (0.11 ^c)
$(1s_N^{-1}3\pi), {}^1\Pi$	404.7 (404.92 ^d)	0.119
$NO (1s_N^{-1}2\pi^2), {}^2\Sigma^-$	397.7	0.061
$(1s_N^{-1}2\pi^2), {}^2\Delta$	397.8	0.038
$(1s_N^{-1}2\pi^2), {}^2\Sigma^+$	399.0	0.019

^a ${}^1\Pi_{g,u}$ unresolved, value from Ref. 162.

^bValue from Ref. 167.

^c ${}^3\Pi_{g,u}$ unresolved, value from Ref. 160.

^dValue from Ref. 164.

^eValue from Ref. 161.

Table 6-18. Transition energies and oscillator strengths for nitrogen ($1s \rightarrow \pi^*$) excited states.

Fig. 6-118. GOS for $1s \rightarrow \pi^*$ excitation in N_2 . Experimental data from Ref. 161. Theoretical data arbitrarily normalized to experiment at $k = 1.0$ a.u.

Fig. 6-119. Comparison of calculated total and experimental GOS for $1s \rightarrow \pi^*$ excitation in N_2 , N_2O , and NO. Experimental data from Ref. 162.

The calculated results for NO provide an interesting example of the equivalent nucleus model of K-shell excitation. According to this model, a K-shell vacancy essentially increases the positive charge of the nuclear core by one.¹⁶⁶ There should then be a correspondence between the core-excited states of NO and the valence states of O₂. The core-excited $^2\Sigma$, $^2\Delta$, and $^2\Sigma^+$ states of NO that we have calculated have as their analogues the $X^1\Sigma_g$, $a^1\Delta_g$, and $b^1\Sigma_g^+$ valence states of O₂. Although the ordering of the calculated NO states is the same as the ordering of corresponding O₂ states, the energy differences do not agree exactly.

Stefani et al. have measured the GOS for the nitrogen $1s \rightarrow \pi^*$ transitions in N₂, NO, and N₂O. Figure 6-118 shows our calculated results and the experimental data for the case of N₂. Note that the data are in excellent agreement with the summed $^{11}I_{10}$ and $^{11}I_{12}$ GOS, pointing to the enhancement of the dipole-forbidden transition at values of momentum transfer $k \rightarrow 1$ a.u. Figure 6-119 compares the summed GOS results for N₂, NO, and N₂O with the experimental values. The shapes of the GOS are indistinguishable in the three molecules.

The nitrogen $1s \rightarrow \pi^*$ excited states in N₂, NO, and N₂O are quite similar; they are well described by limited CI wave functions that include the effect of orbital relaxation following creation of a $1s$ vacancy. For N₂, localized- $1s$ orbitals must be used to describe this relaxation properly. (Analogous calculations¹⁵⁹ using $g-u$ symmetrized orbitals give transition energies in error by 12 to 13 eV.) The N₂ molecule with a $1s$ hole behaves like a heteropolar molecule, although the overall symmetry of the wave functions that we calculate corresponds to $D_{\infty h}$. Similar considerations are expected to apply to $1s$ -vacancy production in any molecule containing atoms in equivalent locations.

Authors: T. N. Rescigno and A. E. Orei

Scattering Amplitudes Using Complex Basis Functions

Substantial progress has been made in recent years on the application of discrete-basis-set techniques to the computation of cross sections for electron scattering

by atoms and molecules. The method of complex basis functions¹⁶⁸ is especially useful for characterizing atomic and molecular resonance (autoionizing) states. Although there has been some work on applying complex-coordinate techniques to nonresonant scattering, this has been limited largely to the computation of total cross sections for photoabsorption, that is, the total probability for excitation and ionization into all energetically open channels.¹⁶⁹⁻¹⁷¹ The evaluation of important dynamical quantities such as partial cross sections, angular distributions, and scattering amplitudes using complex coordinate ideas has been largely unexplored.

For notational simplicity, we confine our discussion here to the case of two-body scattering by a central potential. The extension to scattering by a many-body target is straightforward. The scattering amplitude can be formally expressed as a matrix element of the resolvent,

$$\langle \Psi_0 | V | \Psi \rangle = \langle \Psi_0 | V | \Psi_0 \rangle + \lim_{\epsilon \rightarrow 0} \left\langle \Psi_0 \left| V \frac{1}{E + i\epsilon - H} V \right| \Psi_0 \right\rangle, \quad (113)$$

where Ψ_0 is the unperturbed (plane-wave) scattering wave function and V is the interaction potential. Here ϵ is a small imaginary part of the energy required for convergence, E is energy, and H is the Hamiltonian. The photoabsorption amplitude is also expressible as a matrix element of the resolvent

$$\langle \Phi_0 | r | \Psi \rangle = \langle \Phi_0 | r | \Psi_0 \rangle + \lim_{\epsilon \rightarrow 0} \left\langle \Phi_0 \left| r \frac{1}{E - i\epsilon - H} V \right| \Psi_0 \right\rangle, \quad (114)$$

where Φ_0 is the initial (ground) state of the target. The complex-scaling method consists of carrying out the transformation $r \rightarrow re^{i\theta}$ on the coordinates of the Hamiltonian and then using a finite set of square-integrable functions to form a matrix representation of the resolvent. Here r denotes position, and θ is the rotational angle. Equivalently, one can work with the unscaled Hamiltonian and use complex basis functions scaled as $r \rightarrow re^{-i\theta}$. The attraction of the method is that the complex-scaling transformation

Fig. 6-120. Complex eigenvalues of Hamiltonian obtained using a basis of 25 Slater-type orbitals and one pair of complex functions.

Fig. 6-121. Comparison of exact phase shift (black) with complex basis results (colored curve) for the potential $V = (1 - e^{-r})^3/r^4$.

rotates the continuous part of the spectrum off the real axis and hence renders innocuous the singularity of the resolvent operator appearing in Eqs. (113) and (114).

It is often stated¹⁷¹⁻¹⁷³ that complex-coordinate techniques cannot be used in a straightforward way to calculate the scattering amplitude for a potential that is not exponentially bounded, because the matrix elements in Eq. (113) formally diverge under coordinate rotation unless $V(r)$ decreases exponentially as $r \rightarrow \infty$. However, for the case of photoabsorption [Eq. (114)], the initial state, Φ_{iv} , provides the necessary exponential damping, and complex-scaling techniques can be employed.¹⁶⁹⁻¹⁷¹

We have investigated a more general class of transformations for analytically continuing the Hamiltonian to complex energies. An ideal transformation for numerical computations is one that locally distorts the continuous spectrum off the real axis in the neighborhood of the collision energy E_c but then returns to the real axis at higher

energy. The fact that the spectrum becomes real for $E \rightarrow \infty$ removes the restriction that the potential be exponentially bounded. Moreover, the local distortion of the spectrum around E_c removes the singularity from the resolvent operator. We have devised a procedure for choosing a mixed set of real and complex basis functions that effectively carries out such a transformation.

We begin by choosing a pair of complex functions of the form

$$\phi_{\pm} = re^{\pm ikr} \exp(\pm ikr^{\alpha}r) \quad (113)$$

Here, α is the orbital exponent, and r and θ are defined as before. These functions are augmented with a conventional set of real Gaussian or Slater-type orbitals. We found this combination of functions to be flexible enough to describe the (damped) oscillatory nature of continuum functions in a neighborhood of the real energy corresponding to $E = k^2/2$. Here k is the momentum in atomic units. Figure 6-120 shows the eigenvalue spectrum obtained by diagonalizing a single-particle Hamiltonian over a basis set consisting of one pair of complex functions with $k = 1.0$, and 25 real Slater orbitals. In this figure E_R and E_I are the real and imaginary parts of the energy. In Fig. 6-121 we plot the phase shift obtained using this method to solve Eq. (113) for the case of an infinite range potential $V = (1 - e^{-r})^3/r^4$. We find that in the vicinity of $k = 1$, this method gives phase shifts that are accurate to $\sim 0.01\%$. Moreover, by varying the value of k (or by including several pairs of complex functions with different values of k), we can tailor the basis to provide accurate results over a desired range of energies.

Although the mathematical foundations of this method have yet to be rigorously established, we have ample evidence, for various short- and infinite-range potentials, that the method is convergent. Its application to electron scattering by neutral targets and to photodetachment of negative ions is straightforward. However, the case of the Coulomb potential, which is required to treat electron-ion scattering, needs special care. The usual solution is to replace the asymptotic plane-wave states with distorted-wave Coulomb functions. While this

presents no formal difficulties, it complicates the numerics substantially, especially in molecular problems, to the point where many of the advantages of an analytic basis set approach would be lost.

We have developed another approach to the electron-ion problem based on Dollard's time-dependent treatment of Coulomb scattering.¹⁷⁴ In this formalism, one defines renormalized wave operators that properly convert free-particle plane-wave states into scattering states. One can develop a time-independent version of this formalism and derive renormalized versions of the integral equations (Lippmann-Schwinger equations) for the scattering wave functions and *T*-matrices when there are Coulomb forces between the colliding partners. For example, the analog of Eq. (114) for potentials that include a long-range Coulomb component with charge *Z* is

$$(\Phi_0 | r | \Psi) = e^{-\pi\gamma/2} (4f_0)^{-1\gamma/k} \Gamma(1 - i\gamma) \\ \times \lim_{\epsilon \rightarrow 0} \left(\Phi_0 | r | (E - H - i\epsilon)^\gamma \right) \\ + \frac{1}{(E - H - i\epsilon)^{1-\gamma}} V | \Psi_0 \rangle, \quad (116)$$

where $\gamma = Z/k$. The combination of complex-basis-function techniques with this formalism, then, can also provide a viable computational approach to electron-ion scattering problems.

Author: T. N. Rescigno

Major Contributor: C. W. McCurdy (The Ohio State University)

Atomic Physics of Laboratory Soft X-Ray Lasers

The x-ray emission spectra of high density plasmas containing ions of selenium and krypton have been generated recently with the Novette laser (See "Novette Soft X-Ray Laser Experiments" in Section 5) and with the Z-pinch atomic physics project (ZAPP)¹⁷⁵ and Pithon¹⁷⁶ generators. These experiments were motivated, in part, by earlier studies¹⁷⁷⁻¹⁷⁹ that had identified the

neon-like ions of elements with *Z* = 20 to 42 as potentially suitable candidates for producing ASE in the soft x-ray region of the spectrum. A comprehensive and fairly accurate atomic data base is required to understand the microscopic physics of these systems and to model the laser kinetics in detail.

In direct support of the experimental activities, we have continued our theoretical studies of the atomic physics of neon-like and sodium-like ions. In 1983, our efforts focused on two main topics: (1) the theoretical identification of the 3 ℓ -3 ℓ' emission lines of neon-like ions, including those due to the proposed 3p-3s laser transitions, and (2) a study of the so-called dielectronic satellite spectrum of sodium-like ions that is associated with the *n* = 3 to *n* = 2 x-ray spectrum originating from the neon-like ions.

In spite of the current interest in the neon-like ions of elements in the third row of the periodic table, their emission spectra are not well understood. Although the spectra due to the *n* = 3 to *n* = 2 transitions in the 1.4 to 2.0 keV region have been investigated and assigned,¹⁸⁰⁻¹⁸³ the *n* = 3 to *n* = 3 transitions remain largely unidentified in the spectra of elements heavier than argon. The primary aim of our study was to provide calculated transition energies and oscillator strengths for the 2p⁵3p → 2p⁵3s laser transitions and thereby aid the interpretation of the experimental results obtained on the Novette and Z-pinch facilities.

Our calculations were designed to take into account the effects of both configuration mixing and intermediate coupling of the levels arising from the 2s2p⁴3 ℓ and 2s²2p⁵3 ℓ (with ℓ = s, p, or d) configurations of neon-like ions. The calculations employed nonrelativistic radial orbitals.

Table 6-19. Calculated transition energies and oscillator strengths for (2p⁵3p) *j* = 0 → (2p⁵3s) *j* = 1 in Se²⁴⁺.

Transition	Oscillator strength	Transition energy Δ <i>E</i> (eV)			
		Present		Multiconfiguration-Dirac-Fock	
		FC-CI	RC-CI	Scofield ^a	Lakdawala ^b
L → L	0.108	73.6	74.2	75.3	—
L → U	0.0092	34.9	31.8	32.9	—
U → L	0.045	110.3	111.0	112.3	108.8
U → U	0.111	71.6	68.6	71.1	67.4

^aTaken from Ref. 185.

^bTaken from Ref. 186.

Upper level	Theory			Experiment		
	Present	Shorer ^a	Cowan ^b	Boiko et al. ^c	Burkhalter et al. ^d	Boiko et al. ^e
2p ⁵ 3s ¹ P ₁	1439	1441	1439	1439	1438	1439
3p ₁	1481	1483	1479	1480	1479	1481
2p ⁵ 3d ³ P ₁	1557	1553	1557	1555	1555	1556
³ D ₁	1576	1577	1575	1574	1573	1575
¹ P ₁	1615	1617	1613	1613	1611	1613

^aResults of relativistic random-phase approximation, taken from Ref. 187.

^bCalculations reported in Ref. 180.

^cPredictions taken from Ref. 182.

^dTaken from Ref. 180.

^eTaken from Ref. 182.

Table 6-20. Comparison of $n = 2$ to $n = 3$ transition energies (in eV) for Se²⁴⁺.

One-electron relativistic effects were treated approximately by evaluating the mass-velocity and Darwin corrections to the orbital energies. In constructing the many-electron Hamiltonian (energy) matrix, both the spin-orbit coupling and the Breit interactions were included along with the contributions from the electrostatic Hamiltonian. Relaxation of the orbitals was taken into account by using not only those configurations which describe spectroscopic states, but also additional configurations representing pseudostates. In this way, the L- and M-shell orbitals could be optimized approximately for each level with a given j -value and parity.

The calculations employed an orbital basis set in which 1s, 2s, and 2p were the Hartree-Fock orbitals¹⁸⁴ of the ground state (2s²2p⁶)¹S. The 3s, 3p, and 3d orbitals were optimized in simple frozen-core calculations of the 2p⁵3s, 2s3p, and 2p⁵3d excited states. In addition, the basis contained two pseudo-orbitals (denoted nl) for each l -value.

The calculations were performed at two levels of configuration interaction. The first case included the 2s²2p⁶, 2s2p⁶3l, and 2s²2p⁵3l spectroscopic configurations and the pseudostates arising from 2s2p⁶ nl and 2s²2p⁵ nl . This was essentially a frozen-core calculation, and the results are labeled FC-CI. In the second case, we also included those additional configurations that could be generated from the spectroscopic configurations through the excitations 2s → 3s and 2p → 3p, np . These excitations relax the 2s and 2p core orbitals, and lead to configurations with two holes in the L shell and two electrons in the M shell. Results obtained with this extended configuration list are labeled RC-CI.

Table 6-19 gives the transition energies that were obtained in the various

calculations for the four possible (2p⁵3p) $j = 0$ to (2p⁵3s) $j = 1$ transitions in Se²⁴⁺. The oscillator strengths (gf values) calculated at the FC-CI level are also shown. A comparison of the FC-CI and RC-CI results shows that relaxation of the core orbitals leads to an ~3-eV decrease in ΔE for the transitions terminating on the upper (U) (2p⁵3s) ³P₁ level, but relaxation shifts the other two transitions only slightly. For comparison, Table 6-19 also shows some results obtained with the multiconfiguration-Dirac-Fock method.^{185,186} As far as the two strong transitions are concerned, there is excellent agreement for the L-L transition, but there is an unexplained 1- to 2-eV discrepancy for U-U.

We can estimate the uncertainties in the calculated 2p⁵3p-2p⁵3s transition energies by comparing our values for the $n = 2$ to $n = 3$ x-ray transitions to those obtained in previous experimental and theoretical studies. As Table 6-20 shows, the energies of the 2p⁵3s levels in the RC-CI calculation are quite comparable to those computed by others using different methods,^{180,182,187} and they differ from the experimental values^{180,182} by less than 1 eV.

The x-ray spectra of high-density krypton or selenium plasmas in the 6- to 9-Å region are dominated by the $n = 3$ to $n = 2$ resonance transitions of the neon-like ions, accompanied by sodium-like satellite lines and some fluorine-like lines. The intensities of the sodium-like satellite lines, which originate from doubly excited states, can be used, in principle, to determine the electron temperature and the relative abundance of charge states in the plasma.

Figure 6-122 illustrates the various collisional and radiative processes that need to be considered. There are three possible processes for populating the doubly excited states of sodium-like ions:

1. dielectronic capture of an electron by neon-like ions, e.g.,

2. inner-shell excitation of sodium-like ions, e.g.,

3. inner-shell ionization of magnesium-like ions, e.g.,

The doubly excited states can decay via radiative transitions, e.g.,

or by autoionization (Auger transition), which is the reverse of the first process. To model the experimentally observed line intensities, we have calculated in detail the rates for all of these processes in krypton.

In the sodium-like ions, there are 12 doubly excited configurations involving a 2s or 2p hole and two electrons in the M shell ($n = 3$). These lead to 61 levels in L-S coupling and 201 levels in intermediate coupling. The doubly excited states were described by wave functions that included the effects of configuration interaction and spin-orbit coupling. The nonrelativistic radial orbitals were optimized as described above for the transition energies. Relativistic corrections were made with the Breit-Pauli approximation. The mass-velocity and Darwin corrections to the orbital energies were also included. In the calculations of the Auger rates, we used a single-configuration wave function for the neon-like ground state and expanded the continuum orbitals of the ejected electron in terms of discrete basis functions. We used Stieltjes moment-theory techniques¹⁸⁸ to compute the autoionization widths and their dependence on the energy of the ejected electron. The Hamiltonian (energy) matrices were computed with the SUPERSTRUCTURE code.¹⁸⁹

To calculate the cross sections for inner-shell excitation out of the $2s^2 2p^6 3l$ states of Kr^{25+} by electron impact, we used a distorted-wave approximation¹⁹⁰ to describe the collision dynamics. We computed the scattering matrix with L-S-coupled wave functions and then transformed it to an intermediate coupling scheme with the aid of *ab initio* spin-orbit coupling coefficients. Thus, we obtained the collision strengths for the doubly excited fine-structure levels. We computed the rate coefficients by

State	Without CI	With CI
$2p^5 3s 3d$		
$2p^0$	1.5	134
$2p^0$	9.9	40
$2s 3s 3p$		
$2p^0$	64	190
$2p^0$	117	160
$2p^5 3d^2$		
$2f^0$	65	79
$2f^0$	90	110
$2f^0$	760	760

Table 6-21. Effects of configuration interaction on the autoionization rates (in 10^{12} sec^{-1}) of some L-S-coupled states of Kr^{25+} .

averaging the cross sections in the region from 1.7 to 5.0 keV over a Maxwellian distribution of electrons.

We found that configuration mixing has a strong effect on the calculated radiative, Auger, and collisional excitation rates of some of the states. Table 6-21 illustrates the magnitude of the effect on the autoionization rates in the case of Kr^{25+} . However, for some states, the effect of configuration mixing is small, as shown by the $2p^5 3d^2$ configuration.

The effective two-body rate coefficients for dielectronic capture of an electron by specific doubly excited states were calculated from the formula¹⁹¹

$$k_c = \left(\frac{h^2}{2\pi m k T} \right)^{3/2} \frac{g_d}{2g_i} A_a e^{-E_c/kT}, \quad (117)$$

where g_i and g_d are the statistical weights of the initial ion and the final doubly excited states, respectively, A_a is the Auger (autoionization) rate, E_c is the kinetic energy of the captured electron, T is the electron temperature, and the other parameters have their usual meanings. Equation (117) was evaluated for temperatures between 0.5 and 2.0 keV, a range that covers the typical experimental conditions (see "Novette Soft X-Ray Laser Experiments" in Section 5, and Refs. 175 and 176).

Fig. 6-122. Collisional and radiative processes responsible for sodium-like satellite spectra. Dielectronic recombination and excitation-autoionization are two-step processes involving sodium-like doubly excited states as intermediate states.

Fig. 6-123. Calculated rate coefficients for the processes that populate doubly excited states of sodium-like krypton: dielectronic capture (A), inner-shell excitation (E), and inner-shell ionization (I). For the definition of the three processes, see Fig. 6-122.

Fig. 6-124. Comparison of experimental and calculated sodium-like satellite spectrum from krypton plasma. Experiment: typical ZAPP spectrum from Ref. 175. Calculation: coronal model neglecting inner-shell excitation or ionization.

The rate coefficients for inner-shell ionization of magnesium-like krypton by electron impact were evaluated from the scaled cross-section formulas of Moores, Golden, and Sampson.¹⁹² These authors had calculated the cross sections for ionization from the various sublevels of hydrogenic ions using the Coulomb-Born exchange approximation in the limit $Z \rightarrow \infty$. Using perturbation theory, one can show that for complex ions with finite Z , the ionization cross section Q for subshell nl is given by¹⁹²

$$Q(nl, u) = \pi a_0^2 \frac{N_{nl} n^2}{I_{nl} Z_{eff}(nl)^2} Q_{H1}(nl, u) \quad (118)$$

where I_{nl} and N_{nl} are the ionization potential and the number of electrons in subshell nl , respectively, a_0 is the Bohr radius, and Q_{H1} is the hydrogenic ionization cross section at the scaled energy $u = E/I_{nl}$. The effective nuclear charge $Z_{eff}(nl)$ is chosen to take into account the screening by the other target electrons. To evaluate Eq. (118) for the inner-shell ionization of Kr^{24+} , we used calculated values for I_{2s} and I_{2p} and took $Z_{eff}(2s) = 29.55$ and $Z_{eff}(2p) = 28.85$, based on the screening parameters recommended by Moores et al.¹⁹²

Figure 6-123 illustrates the relative magnitudes of the rate coefficients for the three processes populating some doubly excited

states of sodium-like krypton. For the $(2s^2 2p^3 3s^2) \ ^2P^o$ state, production by inner-shell ionization of Kr^{24+} has the largest rate coefficient for $T \geq 800$ eV. The dielectronic capture rate is quite small. On the other hand, the rate coefficients for both capture and inner-shell excitation are rather large for the $(2s^2 2p^3 3s 3p) \ ^2S^o$ case, with excitation dominating for $T \geq 800$ eV. There are other cases, e.g., the $(2s 2p^6 3s 3d) \ ^2D^o$ state (not shown), where dielectronic capture dominates for $T \geq 1.0$ keV.

Using the radiative (λ_r) and autoionization (λ_a) rates for the 201 levels arising from the $(2s^2 2p^3 3/3l')$ and $(2s 2p^6 3/3l')$ configurations of Kr^{25+} , we calculated a consistent dielectronic recombination rate for each satellite line from the equation¹⁹¹:

$$\alpha_{DR} = \left(\frac{h^2}{2\pi m k T} \right)^{3/2} \times \frac{1}{2} g_d e^{-I_r/kT} \frac{A_d A_r}{\sum A_r + A_d} \quad (119)$$

In the absence of any collisions (e.g., inner-shell excitation or ionization, or collisions among the doubly excited states), α_{DR} will be proportional to the intensity of the satellite line.

Figure 6-124 shows a high-resolution spectrum¹⁷⁵ for those sodium-like satellite lines that lie above the neon-like 2p-3s resonance line having the longest wavelength (7.51 Å). This region is the least cluttered; because there are no fluorine-like lines and few satellites, individual lines can be resolved. Figure 6-124 also shows a theoretical spectrum obtained with the assumption of negligible inner-shell excitation/ionization and with the assumption of a coronal plasma. To account for overlapping lines with finite widths, we summed the dielectronic recombination rates for the lines falling into bins 10 mÅ wide. As Fig. 6-124 shows, the general pattern of strong lines is clearly reproduced by the calculation. However, the relative intensities are different for the two spectra. This difference means that formation of doubly excited states by inner-shell excitation/ionization is important and must be taken into account if more accurate theoretical satellite spectra for plasma diagnostic purposes are to be obtained. More extensive modeling of the dielectronic satellite spectrum of Kr^{25+} is in progress.

Author: A. U. Hazi

Major Contributor: B. L. Whitten

1. B. T. Merritt and J. E. Murray, *Resistance and Energy Transfer Measurements of B174 Amplifier Test Bank and Novette*, Lawrence Livermore National Laboratory, Livermore, Calif., Internal Nova Memorandum PC 83-262 (December 2, 1983).
2. A. Savitzky and M. J. E. Golay, "Smoothing and Differentiating of Data by Simplified Least Square Procedures," *Anal. Chem.* **36**, 1627 (1964).
3. J. S. Hildum, *Status of Gain Measurements on Novette*, Lawrence Livermore National Laboratory, Livermore, Calif., Internal Memorandum OPS 82-01 (November 19, 1982).
4. S. E. Stokowski, R. A. Saroyan, and M. J. Weber, *Nd-Doped Laser Glass Spectroscopic and Physical Properties*, Lawrence Livermore National Laboratory, Livermore, Calif., M-095 (1978).
5. *Laser Program Annual Report—1974*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-50021-74 (1975), pp. 107 to 115.
6. W. W. Simmons, *Nova Performance, Configuration and Component Data*, Lawrence Livermore National Laboratory, Livermore, Calif., Internal Memorandum Nova 81-1323 (June 26, 1981).
7. *Laser Program Annual Report—1978*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-50021-78 (1979), pp. 2-69 to 2-71.
8. M. J. Kushner and R. J. Poli, Lawrence Livermore National Laboratory, Livermore, Calif., Internal Memorandum RCD 83-161 (1983).
9. J. B. Trenholme, *Fluorescence Amplification and Parasitic Amplification Limits in Disk Lasers*, Naval Research Laboratory, Washington, D.C., Memorandum Report 2480 (1972).
10. *1982 Laser Program Annual Report*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-50021-82 (1983), pp. 6-6 to 6-7.
11. J. L. Emmett, W. F. Krupke, and J. B. Trenholme, *Physics of Laser Fusion, Vol. IV, The Future Development of High-Power Solid State Laser Systems*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-53344 (1982).
12. W. T. Welford and R. Winston, *The Optics of Nonimaging Concentrators* (Academic Press, New York, N.Y., 1978).
13. J. B. Trenholme, *Fluorescence Amplification and Parasitic Oscillation Limitations in Disk Lasers*, Naval Research Laboratory, Wash. D.C., Memorandum Report 2480 (1972).
14. *Laser Program Annual Report—1977*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-50021-77 (1978), pp. 2-228 to 2-231.
15. *1981 Laser Program Annual Report*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-50021-81 (1982), pp. 7-56 to 7-64.
16. *1982 Laser Program Annual Report*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-50021-82 (1983), pp. 7-4 to 7-15.
17. J. L. Emmett, A. L. Schawlow, and E. H. Wulenberg, *J. Appl. Phys.* **35**, 2601 (1964).
18. *Laser Program Annual Report—1974*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-50021-74 (1975), pp. 107 to 112.
19. See, for example, "Multichannel Image Detectors," *Symposium Sciences*, Y. Talmi, Ed. (American Chemical Society, Washington, D.C., 1979), p. 102.
20. S. E. Stokowski, R. A. Saroyan, and M. J. Weber, *Nd-Doped Laser Glass Spectroscopic and Physical Properties*, Lawrence Livermore National Laboratory, Livermore, Calif., M-95, Rev. 2 (1978).
21. J. L. Emmett, W. F. Krupke, and J. B. Trenholme, *Physics of Laser Fusion, Vol. IV*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-53344 (1982).
22. J. B. Trenholme and J. L. Emmett, "Xenon Flashlamp Model for Performance Prediction," *Proc. Intern. Conf. High Speed Photogr.* (SMPLE, New York, N.Y., 1970), p. 299.
23. J. H. Kelly, D. C. Brown, and K. Teegarden, *Appl. Opt.* **19**, 3817 (1980).
24. L. P. Bradley, H. T. Powell, R. D. Behymer, and R. J. Poli, *Characteristics of 1.5 and 2.0 cm Bore Standard Flashlamps*, Lawrence Livermore National Laboratory, Livermore, Calif., Memorandum RCD 82-701 (1982).
25. M. J. Kushner, *Xenon Flashlamp Modeling: Arc Expansion and Bore Filling*, Lawrence Livermore National Laboratory, Livermore, Calif., Memorandum RCD 83-126 (1983).
26. D. K. Priest, *Study of Arc Growth and Lifetime Characteristics of 27 mm Bore Flashlamps*, International Lighting Corp., Sunnysvale, Calif., Technology Report R-ILC-74-5 (1974).
27. J. P. Markiewicz and J. L. Emmett, *IEEE J. Quantum Electron.* **QE-2**, 707 (1966).
28. A. E. Orel, *Degree of Ionization in Xenon Flashlamps*, Lawrence Livermore National Laboratory, Livermore, Calif., Memorandum TAMP-91-0111 (1981).
29. R. E. Rovinskii, *Tep. Vys. Temp.* **8**, 1121 (1970).
30. I. Spitzer, Jr., *Physics of Fully Ionized Gases* (Wiley, New York, N.Y., 1962).
31. L. M. Biberman and G. E. Norman, *J. Quant. Spect. Rad. Transfer* **3**, 221 (1963).
32. W. Lama and T. Hammond, *Appl. Opt.* **20**, 765 (1981).
33. J. Goncz, *J. Appl. Phys.* **36**, 742 (1965).
34. S. E. Stokowski, R. A. Saroyan, and M. J. Weber, *Nd-Doped Laser Glass Spectroscopic and Physical Properties*, Rev. 2, Lawrence Livermore National Laboratory, Livermore, Calif., Report M-95, Rev. 2 (1981).
35. S. E. Stokowski, Lawrence Livermore National Laboratory, Livermore, Calif., Internal Memorandum SS 78-327 (1978).
36. C. H. Church, R. G. Schlect, and I. Lieberman, *J. Quant. Spectrosc. Radiat. Transfer* **8**, 403 (1968).
37. B. E. Yoldas, D. P. Parlow, and H. O. Smith, "Polymer Oxide Antireflective Coating for Fused Silica," *Proc. Conf. Laser-Induced Damage Opt. Components*, Boulder, 1983 (Nat. Bur. Stand., Boulder, Colo., to be published).
38. *1982 Laser Program Annual Report*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-50021-82 (1983), pp. 7-24 to 7-26.
39. W. A. Pliskin and H. S. Lehman, *J. Electrochem. Soc.* **112**, 1013 (1965).
40. R. B. Stephens and G. D. Cody, *Solar Energy Mater.* **1**, 397-410 (1979).
41. T. W. Walker, A. Vaidyanathan, A. H. Guenther, and P. Nielson, "Impurity Breakdown Model in Thin Films," *Proc. Conf. Laser-Induced Damage*

- Opt. Mater. Boulder*, 1979 (Nat. Bur. Stand., Boulder, Colo., 1980), NBS Spec. Rept. 568, pp. 479-496.
42. T. W. Walker, A. H. Guenther, C. G. Fry, and P. Nielson, "Pulsed Damage Thresholds of Fluoride and Oxide Thin Films from 0.26 μm to 1.06 μm ," *Proc. Conf. Laser-Induced Damage in Opt. Mater., Boulder*, 1979 (Nat. Bur. Stand., Boulder, Colo., 1980), NBS Spec. Pub. 568, pp. 405-416.
 43. B. E. Newnam and D. H. Gill, *J. Opt. Soc. Am.* **66**, 166 (1976).
 44. J. H. Apfel, *Appl. Opt.* **16**, 1880 (1977).
 45. B. E. Newnam, S. R. Foltyn, L. J. Jolin, and C. K. Carniglia, "Multiple-Shot Ultraviolet Laser Damage Resistance of Nonquarterwave Reflector Designs for 248 nm," *Proc. Conf. Laser-Induced Damage in Opt. Mater., Boulder*, 1981 (Nat. Bur. Stand., Boulder, Colo., 1983), NBS Spec. Rept. 638, pp. 363-379.
 46. 1981 *Laser Program Annual Report*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-50021-81 (1982), pp. 7-51 to 7-54.
 47. D. L. Griscom and E. J. Friebele, "Effects of Ionizing Radiation on Amorphous Insulators," *Radiation Effects* **65**, 63 (1982).
 48. J. B. Trenholme, *Effect of Pump Pulse Shape and Duration in Various Cavities*, Lawrence Livermore National Laboratory, Livermore, Calif., Internal Memorandum LRD 82-20 (1982).
 49. S. E. Stokowski, *Water and Concentration Quenching in LHG-8 and LG-750*, Lawrence Livermore National Laboratory, Livermore, Calif., Nova Memorandum 82-1065 (1982).
 50. W. E. Martin and D. Milam, *Gain Saturation in Nd Doped Laser Materials*, Lawrence Livermore National Laboratory, Livermore, Calif., UCID-18868 (1980).
 51. D. Hall, Lawrence Livermore National Laboratory, Livermore, Calif., Internal Memorandum ERL 82-2 (Jan. 25, 1983).
 52. B. S. Guba, A. A. Mak, S. L. Fotapov, B. M. Sedov, and V. V. Shashkin, *Sov. J. Quantum Electron.* **12**(6), 772-775 (1982).
 53. W. E. Hagen, Lawrence Livermore National Laboratory, Livermore, Calif., Nova Memorandum 80-045 (Feb. 15, 1980).
 54. G. Linford and D. Milam, Lawrence Livermore National Laboratory, Livermore, Calif., Internal Memorandum AD-78-09 (April 7, 1978).
 55. G. Linford, Lawrence Livermore National Laboratory, Livermore, Calif., Internal Memorandum SSD 76-61 (April 22, 1976).
 56. M. M. Mann and L. G. DeShazer, *J. Appl. Phys.* **41**(7), 2951-2957 (1970).
 57. L. G. DeShazer, U. Ranon, and E. D. Reed, *Spectroscopy of Neodymium in Laser Glasses*, Lawrence Livermore National Laboratory, Livermore, Calif., Report 1434705 (1974).
 58. J. M. Pellegrino, W. M. Yen, and M. J. Weber, *J. Appl. Phys.* **51**(12), 6332-6335 (1980).
 59. D. W. Hall, M. J. Weber, and R. T. Brundage, *Fluorescence Line Narrowing in Nd Laser Glasses*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-89836 (1983).
 60. E. Snitzer and C. G. Young, *Lasers*, Vol. 2 (Dekker, New York, N.Y., 1968).
 61. V. I. Nikitin, M. S. Soskin, and A. I. Khizhuyak, *Sov. J. Quant. Electron.* **9**(10), 1314-1315 (1979).
 62. 1981 *Laser Program Annual Report*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-50021-81 (1982), pp. 7-55 to 7-56.
 63. 1982 *Laser Program Annual Report*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-50021-82 (1983), p. 7-66.
 64. 1982 *Laser Program Annual Report*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-50021-82 (1983), p. 7-58.
 65. 1982 *Laser Program Annual Report*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-50021-82 (1983), pp. 7-30 to 7-33.
 66. S. D. Jacobs and J. A. Abate, *Tech. Dig. CLEO, D.C.*, 1981 (Opt. Soc. Am., Washington, D.C., 1981), p. 86.
 67. Available from Kimball Associates, 846-B Rummell Road, St. Cloud, Fla. 32769.
 68. W. L. Smith and F. P. Milanovich, *Stimulated Raman Scattering Inside KDP Segments*, Lawrence Livermore National Laboratory, Livermore, Calif., Memorandum UVM 82-19 (August 25, 1982).
 69. W. L. Smith, F. P. Milanovich, and M. A. Henesian, *Stimulated Raman Scattering Inside KDP Segments-II*, Lawrence Livermore National Laboratory, Livermore, Calif., Memorandum UVM 83-03 (February 25, 1983).
 70. D. K. Agrawal and C. M. Perry, *Proceedings of the Second International Conference on Light Scattering in Solids* (Flammarion Sciences, Paris, 1971), p. 429.
 71. M. K. Srivastava and R. W. Grow, *Opt. Commun.* **8**, 82 (1973).
 72. M. K. Srivastava and G. W. Grow, *Studies on the Tunability of the Stimulated Raman Effect due to Polaritons in GaP and KH_2PO_4 Crystals*, Microwave Device and Physical Electronics Laboratory, Univ. of Utah, Salt Lake City, Utah, UTEC MD 73-057 (1973).
 73. F. Aussenegg and U. Deserno, *Opt. Commun.* **2**, 295 (1970).
 74. M. J. Colles, *Opt. Commun.* **1**, 169 (1969).
 75. G. M. Loiacono, J. Zola, and G. Kosteczky, *Growth of KH_2PO_4 Crystals at Constant Temperature and Supersaturation*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-15341 (1981).
 76. D. Eimerl, "Laser Fusion with Green and Blue Light," *Energy Technol. Rev.*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-52000-82-8 (August, 1982).
 77. M. Summers, D. Eimerl, and R. Boyd, *Optical Harmonic Generator*, U.S. Patent S-56,380 RL-8477 (1981).
 78. D. Eimerl, *Noncollinear Phase Matching Schemes*, Lawrence Livermore National Laboratory, Livermore, Calif., Internal Memorandum 2LP 83-029 (1983).
 79. G. W. Gibbs, "Special Report: The Status of Short-Wavelength Laser Fusion," *Laser Focus Electron-Opt.* **19**(10), 95-98 (1983).
 80. L. L. Steinmetz, T. W. Pouliot, and B. C. Johnson, *Appl. Opt.* **12**, 1460-1471 (1973).

References

81. M. A. Summers and B. C. Johnson, *Design and Performance of Large Aperture KD*P Pockels Cells*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-79952 (1977).
82. W. T. Pawlewicz, I. B. Mann, W. H. Lowdermilk, and D. Milam, *Appl. Phys. Lett.* **34**, 196-198 (1979).
83. J. M. Thorne, *Final Report, Transparent Liquid Electrodes*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-13798 (1977).
84. *Laser Program Annual Report 1979*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-50021-79 (1980), pp. 2-210 to 2-213.
85. *Laser Program Annual Report 1977*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-50021-77 (1978), pp. 2-156 to 2-161.
86. J. Goldhar, "Beam Deflecting Optical Switch Using Volumetric Thermal Gratings," *Appl. Opt.* **23**, 113-117 (1984).
87. J. Goldhar and M. A. Henesian, *Electro-optical Switches with Plasma Electrodes*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-89929 (1983); *Opt. Lett.* **9**, 73-75 (1984).
88. A. Yariv, *Introduction to Optical Electronics* (Holt Rinehart and Winston, New York, N.Y., 1976), 2nd ed., pp. 253-257.
89. I. P. Kamniov, "Linear Electrooptic Materials," *Handbook of Laser Science and Technology*, Vol. III: *Optical Materials Part A*, M. J. Weber, Ed. (CRC Press, Boca Raton, Fla., 1984).
90. *1982 Laser Program Annual Report*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-50021-82 (1983), pp. 6-11 to 6-14.
91. J. Goldhar and M. A. Henesian, *Electro-optical Switches with Plasma Electrodes*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-89929 (1983); *Opt. Lett.* **9**, 77-75 (March 1984).
92. J. Goldhar and M. A. Henesian, *Large Aperture Optical Switching Devices*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-90107 (1983); *Solid State Opt. Control Devices*, **464** (1984).
93. F. Zernike and J. E. Midwinter, *Applied Nonlinear Optics* (Wiley and Sons, New York, N.Y., 1973), pp. 41-48.
94. A. Yariv, *Introduction to Optical Electronics* (Holt, Rinehart, and Winston, New York, N.Y., 1976), 2nd ed., pp. 25-253.
95. W. R. Cook, R. E. S. Heannon, H. Laffe, and D. E. Nelson, *Piezoelectric and Electrooptic Constants*, Vol. II, K. H. Hellwege and A. M. Hellwege, Eds. (Springer-Verlag, New York, N.Y., 1979), pp. 495-670.
96. S. Sugano and R. G. Shulman, *Phys. Rev.* **130**, 517 (1963).
97. A. J. H. Wachters, *J. Chem. Phys.* **52**, 1033 (1970).
98. L. R. Kahn, P. Baybutt, and D. G. Truhlar, *J. Chem. Phys.* **65**, 3826 (1976).
99. *1982 Laser Program Annual Report*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-50021-82 (1983), p. 7-80.
100. A. G. Khudoleev and N. M. Bokun, *J. Appl. Spect.* **14**, 75 (1971).
101. U. S. McClure, *J. Chem. Phys.* **36**, 2757 (1962).
102. *Laser Program Annual Report 1979*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-50021-79 (1980), p. 7-53.
103. *1980 Laser Program Annual Report*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-50021-80 (1981), p. 8-60.
104. *1981 Laser Program Annual Report*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-50021-81 (1982), p. 7-94.
105. K. Rajnak, E. Camp, R. Banks, K. Shinomoto, and N. Edelstein, "Optical and Magnetic Properties of Uranium and Neptunium Borohydrides and Tetrakis-methylborohydrides," *Int. Conf. Chem. Technol. Lanthanides and Actinides 1st Conference, Italy, September, 1983*. To appear in *Inorganica Chimica Acta* with meeting proceedings.
106. K. Rajnak, E. Camp, R. Shinomoto, and N. Edelstein, "Optical and Magnetic Properties of Uranium Borohydride and Tetrakis-methyl borohydride." To be published in *J. Chem. Phys.*
107. K. Rajnak, R. H. Banks, E. Camp, and N. Edelstein, "Analysis of the Optical Spectrum of $\text{Np}(\text{BD})_4$ diluted in $\text{Zr}(\text{BD})_4$ and the Magnetic Properties of $\text{Np}(\text{BH})_4$ and $\text{Np}(\text{BH})_4\text{CH}_3$, to be published in *J. Chem. Phys.*
108. C. H. H. Van Duzen, K. Rajnak, and J. G. Conway, "Uranium five (U V) the 3S_0 level and a Parametric Analysis of the $5f^4$ Configuration," to be published in *J. Opt. Soc. Am.*
109. P. Delamoye, K. Rajnak, M. Genet, and N. Edelstein, "Parametric Analysis of the Energy Levels of U^{3+} in D_{2h} and Limiting D₂ Sites in Incommensurate ThBr_4 ," *Phys. Rev. B*, **28**, 4923-4930 (1983).
110. *1980 Laser Program Annual Report*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-50021-80 (1981), pp. 8-60 to 8-67.
111. *1981 Laser Program Annual Report*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-50021-81 (1982), pp. 7-94 to 7-98.
112. *1982 Laser Program Annual Report*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-50021-82 (1983), pp. 7-78 to 7-81.
113. I. L. Emmett, W. E. Krupke, and J. B. Trenholme, *Soc. J. Quant. Electron.* **10**(1), 1-23 (1983).
114. *1982 Laser Program Annual Report*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-50021-82 (1983), pp. 7-84 to 7-87.
115. *Laser Program Annual Report 1975*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-50021-75 (1976), pp. 505 to 511.
116. *Laser Program Annual Report 1978*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-50021-78 (1979), pp. 8-42 to 8-48.
117. *Laser Program Annual Report 1977*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-50021-77 (1978), pp. 7-94 to 7-98.
118. *1982 Laser Program Annual Report*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-50021-82 (1983), pp. 7-87 to 7-90.
119. *1981 Laser Program Annual Report*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-50021-81 (1982), pp. 7-64 to 7-82.
120. *1980 Laser Program Annual Report*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-50021-80 (1981), pp. 8-32 to 8-52.

121. *Laser Program Annual Report* 1979, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-50021-79 (1980), pp. 7-2 to 7-13 and 7-67 to 7-76.
122. *Laser Program Annual Report* 1978, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-50021-78 (1979), pp. 8-5 to 8-7.
123. K. S. Jancaits, *Kinetic Modelling of Krypton Fluoride Laser Systems*, Ph.D. Thesis, Univ. of Calif., Berkeley, Calif. (1983).
124. *1982 Laser Program Annual Report*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-50021-82 (1983), pp. 7-90 to 7-94.
125. J. P. Lee, and R. K. Cooper, "General Envelope Equation for Cylindrically Symmetric Charged Particle Beams," *Particle Accel.* **7**, 83 (1976).
126. T. J. Rossenden, "The Astron On-Line Beam Energy Analyzer," *Rev. Sci. Instr.* **43**, 1090 (1972).
127. T. J. Orzechowski et al., *The Status of the Lawrence Berkeley Laboratory and Lawrence Livermore National Laboratory Free Electron Laser (FEL)*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-89214 (1987).
128. M. D. Havey, G. E. Copeland, and W. J. Wang, *Phys. Rev. Lett.* **50**, 1767 (1983).
129. I. C. Light and R. B. Walker, *J. Chem. Phys.* **65**, 4272 (1976).
130. T. G. Schmalz, "The R-Matrix Propagation Method with Continuous Perturbation Corrections," *Report of the NRCC Workshop on Algorithms and Computer Codes for Atomic and Molecular Quantum Scattering Theory*, vol. II, L. D. Thomas, Ed. (NTIS, U. S. Dept. of Commerce, Springfield, Va., 1980), pp. 89-92.
131. K. C. Kulander and E. Rebertus, *Phys. Rev. Lett.* **51**, 1262 (1983).
132. K. C. Kulander and E. Rebertus, *J. Chem. Phys.*, to be published.
133. G. N. A. van Veen, K. A. Mohamed, T. Baller, and A. E. De Vries, *Chem. Phys.* **74**, 261 (1983).
134. W. G. Hawkins and P. L. Houston, *J. Chem. Phys.* **76**, 729 (1982).
135. I. C. Light and R. B. Walker, *J. Chem. Phys.* **65**, 4272 (1976).
136. K. C. Kulander and J. C. Light, *J. Chem. Phys.* **73**, 4337 (1980).
137. I. C. Light and A. Altenberger-Siczek, *J. Chem. Phys.* **70**, 4108 (1979).
138. K. C. Kulander and A. E. Orel, *J. Chem. Phys.* **75**, 675 (1981).
139. K. C. Kulander, *Chem. Phys. Letters* **103**, 373 (1984).
140. K. T. Davies, V. Maruhn-Rezwani, S. E. Koonin, and J. Negele, *Phys. Rev. Lett.* **41**, 632 (1978).
141. K. R. Sandhya Devi, A. K. Dhar, and M. R. Straver, *Phys. Rev. C* **23**, 2062 (1981).
142. V. Maruhn-Rezwani, N. Grün, and W. Scheid, *Phys. Rev. Lett.* **43**, 512 (1979).
143. K. R. Sandhya Devi and S. E. Koonin, *Phys. Rev. Lett.* **47**, 27 (1981).
144. K. C. Kulander, K. R. Sandhya Devi, and S. E. Koonin, *Phys. Rev. A* **25**, 2968 (1982).
145. C. Botcher, *Phys. Rev. Lett.* **48**, 85 (1982).
146. M. L. Goldberger and K. M. Watson, *Collision Theory* (J. Wiley, New York, N. Y., 1964), p. 59.
147. M. Faubel and J. P. Toennies, *Adv. Atom. Molec. Phys.* **13**, 229 (1977).
148. D. G. Truhlar and J. W. Duff, *Chem. Phys. Letters* **36**, 551 (1975).
149. R. M. Hegdus and W. F. Reinhardt, *Chem. Phys. Lett.* **91**, 241 (1982).
150. K. C. Kulander, *J. Chem. Phys.* **79**, 1279 (1983).
151. M. J. Coggiola, P. C. Cosby, and J. R. Peterson, *J. Chem. Phys.* **72**, 6507 (1980).
152. J. T. Dowell and J. T. E. Sharp, *Phys. Rev.* **167**, 124 (1968).
153. M. Tronc, F. Fiquet-Fayard, C. Schermann, and R. Hall, *J. Phys. B* **10**, 305 (1977).
154. J. P. Ziesel, I. Neuner, and G. J. Schulz, *J. Chem. Phys.* **63**, 2943 (1975).
155. R. Azria, Y. Le Coat, D. Simon, and M. Tronc, *J. Phys. B* **13**, 909 (1980).
156. Y. Le Coat, R. Azria, and M. Tronc, *J. Phys. B* **15**, 1569 (1982).
157. A. U. Hazi, *J. Phys. B* **16**, L29 (1983).
158. I. S. Cederbaum and W. Domcke, *J. Chem. Phys.* **66**, 5048 (1977).
159. T. N. Rescigno and A. E. Orel, *J. Chem. Phys.* **70**, 3390 (1979).
160. D. Shaw, G. King, F. Read, and D. Cvejanovic, *J. Phys. B* **15**, 1785 (1982).
161. G. Stefani, *Laboratorio di Metodologie Avanzate Inorganiche del Cnr, Rome Italy, private communication* (1983).
162. R. Camilloni, E. Iannelli, G. Petrocelli and G. Stefani, "Generalized Oscillator Strengths for the $1s \rightarrow n\pi$ Transitions in N_2 , NO, and N_2O ," *Int. Symp. Wavefunctions and Mechanisms from Electron Scattering Processes, Villa Montecarlo, Rome, Italy* (July 23-24, 1983).
163. G. C. King, F. H. Read, and M. Tronc, *Chem. Phys. Lett.* **52**, 50 (1977).
164. W. H. E. Schwartz, *Angew. Chem. Int. Ed. Eng.* **13**, 454 (1974).
165. A. Bianconi, H. Petersen, E. C. Brown, and R. Z. Bachrach, *Phys. Rev. A* **17**, 1907 (1978).
166. M. Nakamura, M. Sasanuma, S. Sato, M. Watanabe, H. Yamashita, Y. Iguchi, A. Ejiri, S. Nakai, S. Yamaguchi, T. Tagawa, Y. Nakai, and T. Oshio, *Phys. Rev.* **178**, 80 (1969).
167. R. B. Kay, Ph. E. VanderLew, and M. J. VanderLew, *J. Phys. B* **10**, 2513 (1977).
168. W. P. Reinhardt, *Ann. Rev. Phys. Chem.* **33**, 223 (1982).
169. T. N. Rescigno, C. W. McCurdy, and V. McKoy, *J. Chem. Phys.* **64**, 447 (1976).
170. C. W. McCurdy and T. N. Rescigno, *Phys. Rev. A* **21**, 1499 (1980).
171. B. R. Johnson and W. P. Reinhardt, *Phys. Rev. A* **28**, 1930 (1983).
172. J. Nuttall and H. I. Cohen, *Phys. Rev.* **188**, 1542 (1969).
173. R. T. Baumele, M. C. Crocker, and J. Nuttall, *Phys. Rev. A* **12**, 486 (1975).
174. J. D. Dollard, *J. Math. Phys.* **5**, 729 (1964).
175. R. Stewart, R. Fortner, and D. Dietrich, Lawrence Livermore National Laboratory, Livermore, Calif., private communication (1983).
176. R. Dukart, G. Dahlbacka, D. Dietrich, R. Stewart, and R. Fortner, "Imploding Plasma X-ray Lasers," *Physics International Report* (1982).

References

177. A. V. Vinogradov, I. I. Sobelman, and E. A. Yukov, *Sov. J. Quantum Electron.* **7**, 32 (1977).
178. A. V. Vinogradov and V. N. Shlyaptsev, *Sov. J. Quantum Electron.* **10**, 754 (1980).
179. P. Hagelstein, Lawrence Livermore National Laboratory, Livermore, Calif., private communication (1981).
180. P. G. Burkhalter, D. J. Nagel, and R. D. Cowan, *Phys. Rev. A* **11**, 782 (1975).
181. P. G. Burkhalter, J. Snihlo, A. Fisher, and R. D. Cowan, *J. Appl. Phys.* **50**, 4532 (1979).
182. V. A. Boiko, A. Y. Faenov, and S. A. Pikuz, *J. Quant. Spec. Rad. Transfer* **19**, 11 (1978).
183. H. Gordon, M. G. Hobby, N. J. Peacock, and R. D. Cowan, *J. Phys. B* **12**, 881 (1979).
184. E. Clementi and C. Roetti, *At. Data Nucl. Data Tables* **14**, 177 (1974).
185. J. Scofield, Lawrence Livermore National Laboratory, Livermore, Calif., private communication (1983).
186. S. D. Lakdawala, *Proc. Conf. Atomic Processes in High Temp. Plasmas*, 4th, Princeton University, April 1983.
187. P. Shore, *Phys. Rev. A* **20**, 642 (1979).
188. A. U. Hazi, *J. Phys. B* **11**, L259 (1978).
189. W. Eissner, M. Jones, and H. Nussbaumer, *Comput. Phys. Commun.* **8**, 270 (1974).
190. W. Eissner and M. J. Seaton, *J. Phys. B* **5**, 2187 (1972).
191. H. S. Massey, E. H. S. Burhop, and H. B. Gilbody, *Electronic and Ionic Impact Phenomena*, vol. 4 (Oxford Univ. Press, New York, NY 1974), p. 2118.
192. D. L. Moores, L. B. Golden, and D. H. Sampson, *J. Phys. B* **13**, 385 (1980).

The Cascade reactor is an inherently safe and efficient concept for converting inertial-confinement fusion energy into electrical power.

Energy Applications

7

Energy Applications

Contents

Section Editor: W. J. Hogan

Introduction	7-1
Cascade Reactor	7-3
Chamber Shape Experiments	7-4
Granular Flow Analysis	7-8
Chamber Design	7-12
Balance-of-Plant Design	7-16
Solid-Granule Transport System	7-18
Granule Thermal-Stress Analysis	7-20
Tritium Inventory and Recovery	7-22
Neutronics and Activation Analyses	7-25
Pulse*Star Reactor	7-29
Neutronics and Activation Analyses	7-30
Tritium Containment	7-33
Heat-Transfer Loop and Plant Layout	7-36
Neutronics Studies of a Suppressed-Fission ICF Hybrid	7-41
General Studies and Reactor Development Technologies	7-46
Cost Goals for ICF Reactors	7-46
Development Facilities for ICF Reactor Commercialization	7-50
Disassembly of Isochorically Heated Liquids	7-54
Blanket Optimization Techniques	7-60
Dependence of Displacement Damage on Blanket Design	7-65
Susceptibility of 2-1/4 Cr-1 Mo Steel to Liquid-Metal Embrittlement	7-68
Waste Disposal	7-73
Steel Type and Impurity Effects on Activation	7-76
References	7-81

Energy Applications

Introduction

W. J. Hogan — J. A. Blink

Inertial confinement fusion (ICF) offers the prospect of producing abundant energy from a virtually limitless fuel source—water. Furthermore, while ICF must solve its own safety and environmental impact problems, they appear to be more tractable than those for other energy sources (e.g., acid rain, the greenhouse effect, or fission-product disposal). Nevertheless, to be accepted in our society, ICF reactors must satisfy another basic criterion. They must produce energy at a price that is competitive with other processes.

In past years, studies of the HYLIFE reactor concept established the basic feasibility of developing a practical ICF reactor using current technology once a suitable driver and target are found. In last year's Laser Annual, three new reactor concepts were introduced to improve reactor safety and reduce the unit cost of the electricity produced. This year we completed several detailed studies on two of those concepts: Cascade and Pulse*Star.

In Cascade, rotation of a football-shaped reaction chamber keeps a 1-m-thick blanket of lithium-bearing ceramic granules against the wall as they absorb the shock of the fusion burst, breed tritium, and act as the heat-transfer medium. Because of the high temperatures that can be accommodated, thermal conversion efficiencies above 50% are expected. The Cascade electrical output could thus be 25 to 40% higher than that of HYLIFE with the same fusion power.

Work on Cascade is reported in eight articles. Small laboratory experiments with sand and a theoretical analysis of granular flow have demonstrated the feasibility of establishing and maintaining a uniform layer of granules against the wall as they flow through the reaction chamber. Analysis of the thermal and mechanical stresses in the chamber wall using ASME standards has led to the design of a smaller reaction chamber with an estimated 30-year lifetime. Minimizing chamber size minimizes reactor costs. First looks at the design of heat exchangers and particle transport systems outside the reaction chamber have revealed no fundamental problems. Analysis of granule response to the initial pulse of energy from x rays and debris has led us to conclude that some cracking, melting, and vaporization will occur. Because continuous refabrication of the ceramic granules does not appear economically feasible, we are now looking at other solutions. We have found that the tritium inventory in a Cascade reactor can be limited to 4 kg if the granules are smaller than ~0.7 mm in diameter. Most of this tritium is in the ceramic granules and would be released very slowly in an accident. Finally, a neutronics study has revealed constraints on wall cooling that must now be addressed and has shown that neutron activation would be low enough to permit hands-on maintenance for a SiC wall design.

Pulse*Star is a pool-type ICF reactor that places all primary loop components (reactor, pumps, and steam generators) inside an 8.5 m-radius pool of $\text{Li}_{17}\text{Pb}_{83}$. The liquid is allowed to jet through nozzles in a bell-jar-shaped reaction chamber and impinge on a metal screen. This wetted screen acts as a first wall, while the half-density, 1.5-m-thick, droplet-filled region between the screen and the bell jar absorbs heat and breeds tritium. The high power density within the pool [29 times higher than that within a light-water reactor (LWR) containment building] should lead to lower costs.

Three articles describe our recent work on the neutronics, tritium containment, and design of steam generators and LiPb pumps for Pulse*Star. Our neutronics studies verified the high tritium-breeding ratio and energy multiplication rate predicted earlier. However, activation of the lead coolant poses as much of a problem as activation of the structure, and its consequences are still to be determined. We have found that the tritium released in a spill of LiPb can be kept acceptably low by reducing the chamber vacuum to 5×10^{-3} Torr; this is below the pressure required for laser beam propagation and yet within the capability of reasonable-cost vacuum systems. Satisfactory LiPb pumps and steam generators based on current technology have been designed.

Another application of ICF technology is to produce fuel for fission reactors. In a parametric study of a suppressed-fission, ICF hybrid reactor, we have found that the specific production rate (kg/MW_e) of ^{233}U increases as the thorium and ^6Li content in the breeding blanket are simultaneously increased. Operating with a fairly high thorium content (30%) in such a design should reduce the costs for ^{233}U production.

The last eight articles in this section cover studies general enough to apply to several ICF reactor designs. Our parametric study of the capital costs of an ICF reactor concludes that electric power costs competitive with those of a pressurized-water reactor (PWR) are achievable if the capital costs per unit of thermal power of the ICF reactor (not including the driver or target factory) are between 0.4 and 1.4 times that of a PWR. The greatest influence on reactor cost is the number of reaction chambers built with each driver. Next in importance is pulse repetition frequency (prf). A prf of 2 to 5 Hz in each chamber is sufficient to realize most of the cost benefits. A study of the development facilities required to support the commercialization of inertial fusion shows that it is feasible to build a basic R&D facility flexible enough to be the starting point for several possible development paths. We also studied the physical processes that occur in the disassembly of suddenly heated liquids to determine their effect on prf. Finally, we have developed a simplified mathematical process using variational techniques for doing parametric design studies with only a few, complex neutron-transport calculations.

Damage to the structural wall from a variety of sources can limit reactor lifetime. Two sources of damage important to ICF reactors are radiation and chemical embrittlement. We have explored the dependence of displacement damage in the reactor wall on the location and design of the breeding blanket and reflector. We have also found that the embrittlement of structural steels from interactions with liquid metals is a serious problem in the absence of proper post-weld heat treatment. We have examined several heat treatments to reduce this problem to a manageable level.

In spite of the fact that ICF reactors produce no high-level fission-product waste, they do produce radioactive waste from neutron activation. Studies of the Nuclear Regulatory Commission's (NRC) criteria for radioactive waste disposal reveal that shallow land burial should be allowed even for the first wall of an ICF reactor, if the material is diluted with the blanket, shield, outer structure, and perhaps some grout. A study of four ferritic steels proposed to reduce the neutron activation problem shows that all four reduce activation significantly below the levels in the stainless steels currently used in reactor designs. It should be noted, however, that steel activation will be of less concern in an ICF reactor than in a magnetic-fusion-energy (MFE) reactor because of our less energetic neutron spectrum and our ability to place breeding and heat-transfer media between the hot plasma and the reactor vessel walls.

In sum, we have continued our efforts to develop safe, clean, and inexpensive ICF reactor concepts. Emphasis this year has been on two designs for converting fusion energy to electric power with greater safety and efficiency and on one design for producing fuel for fission reactors. In addition, much of our work is intended to enhance fusion reactor technology without being tied to a specific design. In the coming months, we expect to continue our work on Cascade and Pulse*Star to the point of establishing their basic feasibility, understanding their advantages and disadvantages, and identifying their development paths. We also expect to place greater emphasis on new reactor concepts that convert fusion energy into usable energy forms more directly, and on military applications of ICF.

Cascade Reactor

As we now conceive it, the Cascade fusion reactor consists of a rotating chamber in which a cascading blanket of solid-lithium-ceramic granules breeds tritium, acts as a heat-transfer medium, and protects the chamber wall from fusion neutrons, x rays, and pellet debris (Fig. 7-1). The sand-grain-sized granules are injected into the smaller ends of the chamber at 800 K and

are held against the wall in a loosely packed blanket by the centrifugal force of the rotating chamber. As the granules cascade along the wall, they absorb heat from the fusion reactions. When they reach the chamber's center, they flow out through exit apertures into a stationary granule catcher, pass through heat exchangers for energy extraction, and are then reinjected into the chamber. The open ends of the chamber act as vacuum and beam ports. They also provide space for injecting 600-MJ fusion pellets at 5 Hz; as each pellet

Fig. 7-1. Cascade fusion reactor: a blanket of lithium-ceramic granules absorbs the shock of the fusion burst, breeds tritium, and acts as a heat-transfer medium.

reaches the chamber center, it is illuminated by laser or ion beams.¹⁻³

Cascade combines design simplicity with flexibility and eliminates the safety problems associated with liquid lithium. The reactor can accommodate a range of temperatures, flow rates, and structural and blanket materials. Its solid-lithium-ceramic granules (e.g., Li_2O , Li_8ZrO_6 , LiAlO_2 , or Li_5AlO_4) can withstand high blanket-outlet temperatures (up to 1200 or 1400 K), which allows for the possibility of highly efficient, low-cost power plants.

In the following eight articles we discuss progress on this promising new reactor design.

Author: J. H. Plesch

Chamber Shape Experiments

When we first conceived of the Cascade reactor,⁴ we recognized its potential advantages in safety and efficiency, but we questioned the feasibility of building and operating an approximately 5-m-radius chamber containing 200 tons of granular solids that rotated at 30 rpm. Some of our concerns were allayed when we realized that large ore-crushing machines of nearly that size and rotating speed are in operation. One such machine, for crushing molybdenum, has a 4.3-m radius and rotates at 10 rpm. However, questions remained concerning the chamber shape and the feasibility of providing both a uniformly thick granular blanket and a clean central cavity inside the chamber.

Initially, we chose a football-shaped steel chamber with a 5-m minor radius because it could be made with today's materials and manufacturing techniques and because its lifetime was at least 30 years. The shape was chosen to maintain a nearly uniform Li_2O thickness of 1 m next to the chamber wall. At the ends of the football, where the radius of curvature and the flow area are small, the velocity must be greater than at the center, where the radius and flow area are large. The velocity of the Li_2O granules increases with the slope of the wall, but determining the best chamber shape proved difficult. Analysis⁵ of known granular flow behavior (see next article) provided some

insight but could not provide design criteria that would ensure a uniform Li_2O thickness. We conducted three sets of simple, small-scale laboratory experiments to answer these design questions.

Cylinder Experiments. In the first experiment we put various amounts of sand inside a 102-mm-long cylinder and rotated it. Our purpose was to check an analytical prediction of the behavior of sand tumbling in a horizontally rotated cylinder and to see if granular material could be held against the cylinder wall to form a void in the center. Results are shown in Fig. 7-2. The photographs of the sand profiles reveal close agreement with the theoretical predictions. The cylinder in Fig. 7-2(a) was rotated at 105 rpm; centrifugal acceleration was 4.4 m/s^2 . The cylinder in Fig. 7-2(b) was rotated at 172 rpm; centrifugal acceleration was 11.7 m/s^2 .

The predictions give sand velocity and position as a function of time using accelerations due to gravity and the centrifugal and Coriolis forces. They do not include mixing or shearing of the sand and hence do not show the lip-like buildup of sand, prominent at the bottom of the left photograph. Note that the rotational speed on a small-scale experiment must be higher than for the full-scale Cascade reactor. Centrifugal acceleration must equal or exceed that due to gravity regardless of scale, yet it is proportional to radius. Hence for small-scale experiments, centrifugal acceleration will only approach gravitational acceleration at higher rotational speeds.

Another important finding from this experiment is that the energy required for rotation is much greater if the sand leaves the wall and falls downward than if it stays against the wall. When the sand leaves the wall, potential energy gained as it was lifted to the top is lost and converted into thermal energy. When the sand stays against the wall, that potential energy is returned in the downward portion of the rotational cycle. The sand becomes part of the wall, and there is no conversion of potential to thermal energy. From an efficiency standpoint, the rotational speed of the Cascade reactor should be high enough that the granular blanket is held against the wall at all times. Higher speeds also simplify the maintenance of a clean central cavity.

Fig. 7-2. Cylinder tests at (a) 105 rpm and (b) 172 rpm. Sand profiles in photographs agree well with theoretical predictions shown below them.

Next, we placed the cylinder on a vertically tipping tripod and attached the drill motor directly to the cylinder. We established the sand pattern as before with the cylinder's axis horizontal and then tipped the apparatus so that the axis was tilted up to about 65 deg. When the rotational speed was high enough to hold the sand against the wall, tilting the axis had no effect. In this case there was no shearing action between sand grains, and no sand motion was observed.

Cone Experiments with an Unrestricted Exit. The second set of experiments involved rotating conical test sections of clear Lucite around their horizontal axis with the apparatus shown in Fig. 7-3. Sand was injected through a metal tube into an inlet hopper and flowed out the bottom of the hopper onto an auger driven by a drill motor. We could independently vary the rotational

speed of both the test section and the auger; we could also vary the size of the opening in the bottom of the inlet hopper.

We tested four single cones with half-angles of 10, 20, 30, and 40 deg and two double cones with half-angles of 20 to 40 deg and 20 to 30 deg. All cones had a maximum outside diameter of 108 mm and a minimum inside diameter of 50 mm. Initial tests were conducted with the single cones. When the 10 deg cone was rotated, we found that sand built up on it. The inner surface of the sand was also conical, but the half-angle was ~ 35 deg. Sand cascaded along this 35 deg conical surface but was stationary below. Changes in rotational speed had no effect. When the 20-, 30-, and 40 deg cones were rotated, no sand built up; sand grains entering the small end of the cone immediately cascaded down and out the large end.

Fig. 7-3. View of one conical test section with the feed auger being installed.

Sand placed on the bottom portion of the 20-deg cone remained when the cone was stationary. We had anticipated⁵ that sand would also build up when the 20-deg cone was rotated, because the measured angle that a Lucite plate must be lifted for sand to slide off is ~ 24 deg (the wall-friction angle:). Further, the measured angle of repose for sand is ~ 37 deg. Sand grains probably travel in helical paths as they cascade down a rotating cone, and we surmised that the helical path probably had a slope exceeding the wall-friction angle during part of its rotational cycle. In other words, sand accelerated circumferentially but probably never reached the circumferential velocity of the cone. The actual slope of the path of the sand grains, when both axial and circumferential motion were considered, probably exceeded the 24-deg slide angle.

Next, we added two-sided masking tape to the cones so that their inner surfaces were sticky. When we rotated the cones, a stationary sand region built up not only on the 10-deg cone but also on the 20- and 30-deg cones. In each case, the inside surface of the stationary sand region was also conical but with a half-angle of ~ 35 deg—just as it had been with the 10-deg cone without masking tape. With the 40-deg cone, no sand built up despite the tape.

We concluded that with a rough sand surface rigidly attached to the cone, sand grains cascading down the cone adhered to the sand below until the 37-deg angle of

repose was reached. No sand layer built up on the 40 deg cone because its angle exceeded the angle of repose. Results were again independent of rotational speed.

Cone Experiments with Restricted Exit. In the last sets of experiments we added a closure plate, separated slightly from the cone, and formed 15 equally spaced 2×3 -mm exit apertures by closing off most of the periphery with masking tape. A curved steel pointer penetrating through a central hole in the closure plate measured the radius of the sand surface at various axial locations. We again tested our four single and two double cones and obtained the results shown in Fig. 7-4. Sand built up in all cases. Note that for the 20- and 30-deg cones, the sand profile is within 5 deg of the cone profile, which results in a nearly uniform sand thickness.

Although no measurements were made, it appeared that the sand moved down the cone in slug flow and that the center portion of the cone was free of sand except for occasional grains kicked out by the auger. The 10-deg cone, whose sand-surface angle is the same as that of the 20-deg cone, may be an exception. There may have been a layer of stationary or slowly moving sand next to the cone's surface. However, we were unable to verify the existence of such a layer with our test equipment.

A strobe-light photograph of the downstream end of one cone (Fig. 7-5) shows the shape of the sand streams leaving the cone

as well as the scalloped sand profile against the closure plate. As expected, the sand profile dips adjacent to each exit aperture. These dips would be negligibly small in the Cascade chamber, because a large number of closely spaced apertures would be used.

We found that with the closure plate on, the sand depth on a cone's surface varied with rotational speed. As rotational speed (and centrifugal acceleration) increased, the sand depth decreased. This agrees with the results of Savage and Sayed,⁶ who found that flow rate through an orifice is proportional to $g^{1/2}$, where g is the acceleration due to gravity. We held our inlet feed rate constant, but the exit flow rate through the apertures depended on the centrifugal force, and hence sand depth changed when rotational speed changed. The 40-deg single cone and the 20- to 40-deg double cone were particularly sensitive to rotational speed, perhaps because 40 deg is above the sand's angle of repose. The flow rate's dependence on the exit aperture area and rotational speed would allow us to easily control both the flow rate and thickness of the solid-lithium-ceramic granules in the Cascade blanket.

We repeated this last series of tests with two-sided masking tape again applied to the cones' inner surfaces. Results were identical to those in tests without tape. In all tests, the most uniform sand thickness occurred when the half-angle was either 30 or 40 deg. A slightly smaller thickness existed at the large end of the 30-deg cone and at the small end of the 40-deg cone. We thus chose 35 deg as the best half-angle with which to obtain a uniform thickness. The Cascade chamber could be formed of two such cone frustums, or it could be formed of several frustums of cones with angles between 20 and 40 deg. A 10-deg frustum was not considered because of the possibility that a stationary layer of sand would form on the cone. Such a layer is unacceptable, because neutron heating would raise the temperature of the stationary granules past their melting point.

Conclusions. Simple, small-scale tests revealed that a nearly uniform layer of sand can be maintained next to a cone if its half-angle is between 20 and 40 deg. The most uniform thickness is expected when the half-angle is about equal to the 37-deg

Fig. 7-4. Sand experiments show that a nearly uniform layer of flowing sand can be expected on cones with half-angles between 20 and 40 deg.

Fig. 7-5. Strobe-light photograph showing the sand profile and shape of exit sand streams when a closure plate was added to the cone experiments.

angle of repose. If the half-angle is less than 20 deg, a stationary layer of sand may form on the cone.

Because the angle of repose and the wall-friction angle are independent of the scale of an experiment, our results should apply to the larger Cascade chamber. We conclude that a satisfactory Cascade chamber could be made from frustums of cones if the half-angles are between 20 and 40 deg. A clear central cavity without sand particles traversing it appeared to exist in our rotating-cone experiments.

Only minimum power is needed to rotate a chamber if the blanket granules are held against the wall at all times. Our test results

show that the flow rate depends on the exit aperture area and rotational speed. This allows easy control of the blanket thickness in the Cascade design.

Author: J. H. Pitts

Major Contributors: J. L. Cormier, R. C. Duffus, W. F. Gee, C. Michel, and O. R. Walton (LINL); and R. L. Yancey (Kaiser Engineers)

Granular Flow Analysis

In the Cascade reactor concept,⁷ Li₂O granules are fed into the ends of a rotating chamber; centrifugal forces keep the granules on the surface and cause them to flow toward the larger diameter central region, where they exit into a stationary catcher. To help evaluate the feasibility and economics of this concept, we examined some of the granular flow phenomena involved. Although the flow of granular solids is not well understood, we used what is known about flowing solids to put qualitative limits on the expected behavior.

Nature of the Flow. At least two significantly different types of flow are possible with granular materials: supercritical flow, in which the velocity of the flowing material is higher than the wave propagation velocity through the bulk granular material, and subcritical or quasi-static flow, in which disturbances can be propagated upstream. (See Ref. 8 for a discussion of transitions from one type of flow to the other.) In supercritical flow, granular materials behave like fluids, and much of our understanding of fluid flow can be applied directly to such granular flows. For example, for flows over inclined surfaces, the flow velocity would increase as the slope of the incline increases and would decrease as the slope decreases. Quasi-static flows, on the other hand, can be quite different from normal fluid-flow; specifically, the flow rate can be controlled from a downstream position, making it possible to stop the flow while the material remains on an incline. This ability to separate control of the flow depth from control of the flow rate makes quasi-static flows particularly interesting for Cascade applications. That we can consider the flows in Cascade to be quasi-static can be seen from the anticipated 2-m³/s granule flow rate.

This flow would result in an axial velocity of only 30 mm/s near the exit of the reactor, where the total flow area would be nearly 60 m². This is far less than wave propagation velocities, which are typically on the order of meters per second in loose, porous granular materials.

Inclined-Plane Flow Limits. For cohesionless granular materials, the total mass flow rate can be controlled at the bottom of an inclined plane over which the material is flowing if the following conditions are satisfied:

$$\phi_w < \alpha < \phi_t \quad (1)$$

where α is the angle of incline, ϕ_w is the wall-friction angle (above which sliding will occur), and ϕ_t is the drained angle of repose for the material. The drained angle of repose is the material's natural slope near an opening after it flows over an edge or through an orifice. Thus, if the conditions of Eq. (1) are satisfied, we can separately control the flow depth (at the top of the incline) and the flow rate (at the bottom of the incline) without any active feedback mechanisms.

Laboratory tests have demonstrated that a uniformly thick flowing layer can be maintained on an inclined plane when the conditions of Eq. (1) are satisfied.⁹ Our next step was to examine what restrictions Eq. (1) would place on the geometries of the Cascade chamber. The wall-friction angle depends on the properties of both the granular and wall-surface materials. For ceramic granules on a machined-metal surface, this angle is likely to lie in the range 24 deg < ϕ_w < 32 deg. The drained angle of repose for the granules depends on their size and shape distribution as well as on their surface properties, but it is likely to lie in the range 30 deg < ϕ_t < 40 deg. Thus we will probably find α limited to 32 ± 8 deg.

In a centrifugal-action chamber, whether or not the granules will slide depends on the incline's effective angle, α_{eff} , which in turn depends on the total body force acting on the granules, including gravity and the centrifugal and Coriolis forces, and on the chamber's shape and orientation.

To determine if granules will slide on the inner surface, we need to know the total normal and tangential forces acting on

them. If the ratio of the total tangential force to the total normal force exceeds the tangent of the wall-friction angle, then the granules will slide in the direction of the total tangential force.

Horizontal Rotating-Cylinder Flow

Limits. Consider a cylinder rotating about a horizontal axis, as shown in Fig. 7-6. For a granule stationary relative to the surface of the rotating cylinder, the effective body acceleration, $\mathbf{g}_{\text{eff}}$, is just the vector sum of the acceleration due to gravity, $\mathbf{g}$, and the centrifugal acceleration, $\mathbf{a}_c$:

$$\mathbf{g}_{\text{eff}} = \mathbf{g} + \mathbf{a}_c \quad (2)$$

In the laboratory frame of reference, $\mathbf{g}$ is a constant vector pointing downward, and $\mathbf{a}_c$ has a constant magnitude but rotates about the axis, always pointing radially outward. When viewed from the point of view of a granule on the rotating cylinder, however, the centrifugal acceleration is constant, always pointing radially outward, and the gravitational acceleration appears to rotate about the axis. The magnitudes of the total effective body acceleration in the radial (i.e., normal) and circumferential directions, g_r and g_θ , are given by

$$g_r = a_c + g \cos \theta = \omega^2 r + g \cos \omega t \quad (3)$$

and

$$g_\theta = -g \sin \theta = -g \sin \omega t \quad (4)$$

where the angle θ is the current orientation of the point on the rotating cylinder at which the granule is located, with $\theta = 0$ corresponding to the bottom.

Depending on the rotation speed, ω , and the wall-friction angle, ϕ_w , granules on the inside of the rotating cylinder may periodically see forces exceeding the wall-friction limit. We can expect to see circumferential sliding whenever

$$\tan \phi_w < \left| \frac{g_\theta}{g_r} \right| \quad (5)$$

or

$$\tan \phi_w < \left| \frac{g \sin \theta}{\omega^2 r + g \cos \theta} \right| \quad (6)$$

Fig. 7-6. Effective body force, $\mathbf{g}_{\text{eff}}$, acting on granule on inside of horizontal rotating cylinder. At 1, $\theta = 0$, $g_r = g + a_c$, $g_\theta = 0$; at 2, $\theta = 90^\circ$ deg, $g_r = a_c$, $g_\theta = g$.

At $\theta = 0$ and $\theta = \pi$, g_θ is zero and no circumferential sliding occurs; but at $\theta = \pi/2$ and $\theta = 3\pi/2$, this ratio reaches a maximum value given by

$$\left| \frac{g_\theta}{g_r} \right|_{\text{max}} = \pm \frac{g}{\omega^2 r} \quad (7)$$

Thus whenever $\tan^{-1}(g/\omega^2 r) > \phi_w$, we can expect to see periodic circumferential sliding, with granules traveling slightly slower than the cylinder wall on their upward motion and slightly faster on their downward motion. If the granules' wall-friction angle is near 30° deg, we would expect to see some oscillatory circumferential sliding for any rotational speeds for which $\omega^2 r \lesssim 1.73 g$.

Horizontal Rotating Cone. Now consider a cone-shaped chamber as shown in Fig. 7-7. The direction of the surface normal has changed, and the expression for the normal component of the total body acceleration, g_n , is now

$$g_n = g_r \cos \alpha = (\omega^2 r + g \cos \omega t) \cos \alpha \quad (8)$$

The total tangential body acceleration, g_t , can be considered to have two components: a circumferential component, g_θ , still given by Eq. (4), and an additional component, g_a , acting axially along the cone surface. This component is just the projection of the radial term, g_r , in the axial surface direction:

$$g_a = g_r \sin \alpha = (\omega^2 r + g \cos \omega t) \sin \alpha \quad (9)$$

At first glance we notice that g_a and g_n have the same time dependence, so we might expect the conditions for axial flow to be the same as those for inclined-plane flow given in Eq. (1), as long as $\omega^2 r > g$. Such is not the case, however, because the criterion for

Fig. 7-7. Effective body force, $\vec{g}_{\text{eff}}$, acting on granule on inside of horizontal rotating cone; 1 and 2 represent $\omega t = 0$ and $\omega t = 90$ deg in Eqs. (8) and (9).

slippage is based on the total tangential body acceleration, g_t given by

$$g_t = \sqrt{g_\theta^2 + g_a^2} \quad (10)$$

and not just on the axial component, g_a . Thus we will see some sliding (at least intermittently, as the chamber rotates) whenever

$$\tan \phi_w < \frac{g_t}{g_n} \quad (11)$$

If sliding does occur, then for any finite angle, α , we can expect to see at least some axial motion, because there will be a component of the tangential force acting in the axial direction. In the sand tests reported in the preceding article, axial flow of sand was obtained on cones with half-angles of only 20 deg, while sand has a 24-deg wall-friction angle. This could only occur if factors other than just the axial component of the body force, such as the circumferential component, were influencing the flow.

It should be noted that for a fixed rotational speed, the value of $\omega^2 r$ will increase as one moves down the incline, so that the constant-amplitude, sinusoidally varying circumferential force due to gravity will be a larger fraction of $\omega^2 r$ at the smaller diameters. The net effect of the periodic circumferential flows on a full-scale Cascade design will probably be a slightly thinner

granule layer on the top of the cone, especially near the inlet. This is because any circumferential sliding that occurs tends to move material toward the lower half of the cone.

Coriolis Effects. As granules flow down the cone, they move to regions of larger circumference; to keep up with the constant rotational speed of the chamber, the granules must be accelerated. The force causing this acceleration, which must be transmitted to the granules through frictional coupling with the wall, is called the Coriolis force. This force enters our slip-condition relations as an additional term in the circumferential acceleration, g_θ , changing Eq. (4) to

$$g_\theta = g \sin \theta - 2\omega v_r \\ = g \sin \omega t - 2\omega v_a \sin \alpha \quad (12)$$

where v_r is the radial component of the mass flow velocity and v_a is the velocity of the material moving axially down the cone surface.

In a full-scale Cascade chamber, v_a will be small compared to the circumferential velocity of rotation, so that the Coriolis force will have little effect on the overall flow. However, the Coriolis force probably contributed to obtaining flows at inclination angles less than the wall-friction angle in the sand flow tests mentioned earlier. This force will also contribute to overall shearing deformations in the flowing layer and thus may be an important factor affecting the percolation of fine granules, as discussed below. It should also be noted that accelerating the flowing mass by the Coriolis force as it moves to larger radii will require about 100 kW in the drive motors for a 5-m-diam chamber; this is in addition to any power required to compensate for frictional losses.

Mixing, Size Segregation, and Divergent Flows. In addition to our concern over whether the Li_2O granules will flow or not, we are also concerned about mixing and particle-size segregation in the flowing layer. As the flowing granules move through the chamber, they will be hit with high-intensity pulses of radiation and debris up to five times each second. The neutrons in these pulses will penetrate well into the flowing layer and deposit energy. However, a significant fraction of the energy is in other forms of radiation and in the kinetic

energy of the debris, so that total energy deposition will result in considerably higher heating on or near the layer's inner surface. We expect to see some granule breakage, melting, and vaporization, so that the inner layer may include broken granules at much higher temperatures than the remainder of the bed (see the discussion in "Granule Thermal-Stress Analysis").

For optimum operation, all granules would be heated equally. One way of approaching that optimum is to provide mixing of the flowing granular bed as it moves through the chamber. The most important factor here is the need to constantly replace the top surface layer that is absorbing the bulk of the radiation and debris energy.

Particulate solids can mix in two ways: turbulence or macro-mixing, in which large groups of particles move collectively, and diffusion or micro-mixing, in which individual particles migrate randomly through the surrounding particles.^{10,11} Unlike fluids, in which diffusion is spontaneous (arising from thermal agitation), granular solids diffuse only when energy is supplied, such as in shearing flow. In shearing granular flows involving materials with a wide size distribution, diffusion results in the downward percolation of smaller particles through the flowing layer.¹²

Controlled, quasi-static flow on an inclined plane does not necessarily involve any significant shearing; however, the chamber designs being considered for Cascade all have divergent flow geometries that will force some shearing to occur. We do not know enough about diffusion in flowing granular solids to say whether the shearing caused by the chamber's changing circumference will provide enough diffusion for percolation. However, alternating divergent and convergent flows are used in the mineral processing industry to separate heavy and light components in a slurry (see the discussion of Reichert cones in Ref. 13). If the divergence does not provide enough shearing to achieve rapid percolation of the hot fines through the blanket, other measures could enhance such effects.

We know from work in powdered metallurgy that flow around tight corners causes size segregation.¹⁴ This phenomenon could provide significant shearing in the bottom layers of a Cascade blanket if the chamber

Fig. 7-8. One-half of a modified Cascade chamber using steps to increase shearing in the flowing granule layer and thus increase percolation of hot fines.

had a series of steps in its surface. Such steps could also provide enough shearing through the thickness to cause the fines on the top surface to percolate into the flowing bed. For short distances, these steps could exceed the angle-of-repose restriction of Eq. (1), and, as long as the average slope did not exceed that restriction, the flow should still remain controllable at the outlet. Such a Cascade chamber might have a shape like that shown in Fig. 7-8.

Figure 7-9 shows details of possible granule inlet and outlet geometries. Figure 7-9(a) shows a stationary end wall with openings for granule, beam, and target injection. The granules are introduced through an almost tangential chute located on the downward-moving side of the chamber. Such a chute could be designed so that the tangential velocity of the entering granules is near the circumferential velocity of the chamber wall, thus assuring a granule-free central cavity for target and beam injection. The radial dimension of this inlet chute would also determine the thickness of the flowing granule layer. (The top of the stationary wall would, of course, be located at this height.)

One simple method of controlling the total flow rate at the outlet is shown in Figs. 7-9(b) and (c). In this configuration, granules would flow down through a radial slot at the largest circumference of the chamber and stop on a horizontal shelf rotating with the chamber. The granules could then be scraped off this shelf, making room for more to flow on, by any number of stationary ploughs located around the circumference. Judicious placement of such ploughs (e.g., only on the top or

Fig. 7-9. Chamber inlet and outlet details: (a) tangential inlet chute sets material depth; (b) stationary plough controls total flow rate; (c) ploughs could be positioned anywhere on the circumference.

upward-moving portion of the chamber) and utilization of the kinetic energy of the exiting granules could considerably reduce the complexity of solids recycling in a full-scale Cascade reactor.

Summary. A Cascade chamber rotating around a horizontal axis can achieve controlled, quasi-static, granular flows over a limited range of angles (corresponding to cone half-angles near 30 deg). In such flows, the flow rate can be controlled at the exit point, and the flow depth can be controlled at the entry point. The actual limiting angles are somewhat shallower than predicted for inclined-plane flow but will approach the inclined-plane values near the large-radius exit. The divergent nature of flows from small to large radii will cause shearing of the granules, which may induce size segregation and carry any hot fines on the inner surface down into or through the flowing granular blanket. Additional shearing can probably be induced in the flow to increase percolation, if necessary. Research into the nature of divergent quasi-static flow (and especially the diffusive percolation of fines in such flow) is needed before we can confidently predict that such effects can be used in a Cascade reactor.

Author: O. R. Walton

Major Contributor: J. H. Pitts

Chamber Design

During 1983, we revised the Cascade reactor design¹⁵ to obtain a minimum-sized reaction chamber that could last the 30-year

lifetime of a power plant and that could be built with today's materials and manufacturing techniques. A smaller chamber reduces cost, tritium inventory, and heat loss; it also mitigates vacuum pumping and cooling requirements.

Our revisions reduce the chamber volume to 50% and the blanket volume to 60% of their original estimates (see Table 7-1). Further, chamber half length has been reduced to 5.9 m, which allows driver beam cone angles up to ~18 deg for illumination of the fusion pellets. We chose 2-1/4 Cr-1 Mo ferritic steel for the chamber wall because of its wide use in the nuclear industry, its resistance to hydrogen embrittlement, its adequate creep-fatigue properties and yield stress at temperatures up to 800 K, and its low neutron-activation characteristics. A ferritic steel wall, however, will require external cooling because of its exposure to the high-temperature Li₂O blanket. One cooling method, liquid-metal spray cooling, would maintain a suitable maximum wall temperature of 800 K. Other wall materials, such as silicon carbide, are also being evaluated.

We chose Li₂O as the blanket material because of its high tritium-breeding ratio (~1.3). However, recent literature^{16,17} indicates material compatibility difficulties with Li₂O at high temperatures, particularly with the formation of LiOT, which is corrosive to most steels. The choice of the best blanket material thus remains to be determined.

In this article we describe characteristics of the revised Cascade chamber, our design method (which includes both thermal and mechanical analyses), and design variations that occur with changes in

chamber radius, wall thickness, and blanket outlet temperature.

Design Method. We combined thermal and mechanical analyses to calculate the total time-dependent stress in the chamber wall and compared this stress to allowable values determined from the material properties for 2-1/4 Cr-1 Mo steel¹⁸ and the methodology for pressure vessel design from the ASME Code.¹⁹ This design method is similar to that used for the HYLIFE chamber.²⁰

In the thermal analysis, we selected 800 K as the inside wall temperature. The steel has adequate strength up to this temperature, and a lower temperature only increases the necessary external cooling. We varied the outlet temperature of the Li₂O blanket and calculated the heat transfer from the granules to the wall using Schlünder's correlation.²¹ This heat transfer is essentially all by thermal radiation. We then calculated the maximum radial temperature difference across the wall, ΔT_w , including the effects due to neutron heating, and finally the secondary thermal stress, σ_{th} , using

$$\Delta T_w = \frac{\sigma(T_b^4 - T_w^4)t}{(1/\epsilon_b + 1/\epsilon_w - 1)k} + \frac{\dot{q}''' t^2}{2k} \quad (13)$$

and

$$\sigma_{th} = \frac{\alpha E \Delta T_w}{2(1 - \nu)} \quad (14)$$

Here σ is the Stefan-Boltzmann constant, T is absolute temperature, ϵ is the emissivity, $\dot{q}'''$ is the volumetric neutron heat-generation rate, t is the wall thickness, k is thermal conductivity, α is the linear coefficient of thermal expansion, E is Young's modulus, and ν is Poisson's ratio. The subscripts b and w refer to the Li₂O blanket and the inside wall surface. The calculation is iterative, because material properties are determined at the average radial wall temperature, which is initially unknown. For our revised design in Table 7-1, the secondary thermal stress is 120 MPa.

Mechanical stresses due to centrifugal force, vacuum inside the chamber, and gravity as well as those due to x-ray,

Item	Original chamber	Revised chamber
Chamber		
Shape	Prolate spheroid	Frustums of cones
Volume (m ³)	840	400
Blanket volume (m ³)	370	220
Half-length (m)	8	5.9
Minor radius (m)	5	4.4
Blanket thickness (m)	1.0	~0.9 to ~1.1
Wall thickness (mm)	25	27.5
Wall material	2-1/4 Cr-1 Mo ferritic steel	
Maximum wall temperature (K)	800	
Blanket		
Granule diameter (mm)	1	
Inlet temperature (K)	800	
Outlet temperature (K)	1200	
Flow rate (m ³ /s)	2.5	
Rotational speed (rpm)	30	
Fusion power (MW)	3000	
Fusion pellet yield (MJ)	600	
Repetition rate (Hz)	5	

Unchanged

neutron, and pellet-debris energy deposition in the blanket are calculated as functions of time and added to the secondary thermal stress. We calculate the steady equivalent pressure at the maximum chamber radius due to the centrifugal force of the blanket and chamber wall, subtract one atmosphere to account for the vacuum inside the chamber, and determine the resulting stress as

$$\sigma_c = \frac{P_c R}{t \cos \theta} \quad (15)$$

where σ_c is hoop stress, P_c is the equivalent pressure due to centrifugal force plus vacuum, R is the maximum chamber wall radius, and θ is the half-angle of the cone-shaped wall at its maximum radius. When calculating P_c we considered that the centrifugal force is radially outward, whereas the force due to external pressure acts normal to the chamber wall. Stress due to gravity, σ_g , is calculated in a similar fashion except that it varies sinusoidally with time, because a given location moves from bottom to top as the chamber rotates.

X-ray, pellet-debris, and neutron energy deposition in the blanket varies spatially as a decaying exponential—that from x rays and pellet debris decays rapidly and that from neutrons decays more slowly. The x-ray and pellet-debris deposition is large enough to ablate a few microns of material from the inside surface of the blanket.

Ginn²² showed that the resulting hoop

Table 7-1. Characteristics of the original and revised chambers.

Fig. 7-10. Temporal variation of the primary mechanical stresses at the maximum-radius section for our revised Cascade reactor design.

stress for a conservative short-cylinder geometry approximation limit is

$$\sigma_{xd} = \frac{M}{At} \sqrt{E/\rho} \quad (16)$$

where σ_{xd} is the hoop stress from x rays and pellet debris, M is the momentum in the ablated material, A is the wall area over which the momentum is applied, ρ is the density of the wall material, and E and t are, as before, Young's modulus and the wall thickness.

The ablated material also exerts a pressure on the wall until recondensation is complete. This pressure results in an additional hoop stress in the chamber wall, which can be calculated using

$$\sigma_p = \frac{2(\gamma - 1)eR}{Vt \cos \theta} \quad (17)$$

where γ is the ratio of specific heats, e is the specific energy in the ablated material, and V is the volume of void inside the chamber blanket's inner surface. The factor of 2 in the numerator accounts for σ_p being

applied dynamically. Because the ablated material must travel to the center of the chamber and rebound before the pressure builds up, there is a time delay (phase lag) from the momentum-generated stresses calculated in Eq. (16).

The magnitudes of σ_{xd} and σ_p decay with time, but those of σ_c and σ_R do not. Figure 7-10 shows the interaction of the primary mechanical stresses as a function of time for our revised chamber; the system has 10% critical damping. Stresses due to pressure from ablated material and to x rays and debris vary rapidly with time. By comparison, stress due to gravity varies slowly. Therefore, over the time span of interest, stress due to gravity is essentially constant and can be added to stress due to vacuum and centrifugal force. Note that the maximum primary stress and its peak-to-peak variation never exceed 82 and 87 MPa, respectively; the peak-to-peak variation is larger because of the net negative stress due to centrifugal force, gravity, and vacuum.

Stress resulting from neutron energy deposition is mitigated by the porous nature of the granular blanket. The granules are

continually cascading, and void space surrounds each one (like loose sand). The neutron deposition causes isochoric heating of the granules, and a rarefaction wave propagates through each granule. The granules expand slightly, but no net momentum is generated; according to Eq. (16), no hoop stress would be produced. Even if the granules are comminuted (and this is not expected), any kinetic energy produced is converted back to thermal energy as the granules interact. Net momentum would still be essentially zero. Hence, hoop stress due to neutron heating in the blanket is negligibly small.

In comparing the calculated stresses to those allowed if the intent and methodology of the ASME Code are followed, there are two criteria for our application that can be limiting. First, the sum of the primary mechanical and secondary thermal stresses at any instant must be less than the yield stress. For our conditions, the yield stress is about 210 MPa; the sum of our calculated stresses is 202 MPa. Second, the number of cycles needed to cause a creep-fatigue failure under the fluctuating stress conditions present in the Cascade chamber must be greater than the number of cycles in the chamber's 30-year lifetime. The analysis of creep fatigue is somewhat complex,^{18,20} but for our conditions we satisfy this requirement if the maximum peak-to-peak variation is less than 92 MPa; our calculated variation is 87 MPa. Other criteria involving the allowable primary stress that can be present under operating conditions and the maximum allowable steady primary stress are less restrictive.

Parameter Study. To find the minimum chamber size for Cascade, we conducted a parameter study by varying the chamber radius, wall thickness, and blanket outlet temperature. Figure 7-11 shows the chamber's minimum required minor radius and resulting wall thickness for various blanket outlet temperatures. We kept the maximum wall temperature on the inside surface constant at 800 K—a value at which 2-1/4 Cr-1 Mo steel has adequate creep-fatigue and yield-stress properties. Although decreasing the wall temperature below 800 K may be found desirable later, we felt that using the upper temperature limit would minimize any difficulties with wall cooling. Further,

we held the inlet blanket temperature constant at 800 K because results from our studies on a compatible balance of plant are not yet complete.

The chamber's minimum minor radius increases with blanket outlet temperature because of increasing thermal stress. A higher blanket outlet temperature results in more heat transfer from the blanket to the wall, a greater radial temperature difference across the wall for a given thickness, and hence a greater thermal stress. The chamber's wall thickness decreases with blanket outlet temperature, mainly because the minimum chamber radius changes at the same time.

In fact, total stress is surprisingly insensitive to wall-thickness variation if all other parameters are held constant. Figure 7-12 shows how stress (at the chamber's minor radius) varies with wall thickness when other parameters are at values shown in Table 7-1. We chose 1200 K as the blanket outlet temperature here to maximize thermal efficiency and yet still use the Li_2O granules in a state-of-the-art power plant that included high-pressure helium gas turbines.²³ Higher temperatures could not be used effectively in present gas turbines, and a lower temperature decreases the gross plant efficiency to below 55%.

In Fig. 7-12, primary mechanical stress—from momentum, pressure due to ablated blanket material, centrifugal force, gravity, and chamber vacuum—decreases as the wall thickness increases. However, the thermal resistance of the wall increases with

Fig. 7-11. Minimum chamber minor radius increases and the associated wall thickness decreases for higher blanket outlet temperatures.

Fig. 7-12. Total stress has a broad minimum that is less than the allowable stress over a range of wall thicknesses, if other parameters are held constant.

wall thickness, and this results in an increasing radial wall-temperature difference and secondary thermal stress. Total stress has a broad minimum and varies only slightly with changes in wall thickness, because the effects of the various component stresses nearly cancel each other. Allowable total stress increases slightly with wall thickness, because the average wall temperature decreases slightly. Actual stress is below allowable stress at the minor-radius section for wall thicknesses between ~17 and 32 mm. When the total chamber (not just the minor-radius section) is considered, the best wall thickness is 27.5 mm.

Conclusions. We revised Cascade to obtain a minimum-sized chamber in which the sum of all stresses is less than the 210-MPa yield stress of 2-1/4 Cr-1 Mo steel, and the peak-to-peak stress variation is less than 92 MPa. Exceeding the latter would cause creep-fatigue failure before the end of the chamber's 30-year lifetime. This method satisfies the intent and methodology of the ASME Code, and it permits the revised chamber design to be built with today's materials and manufacturing technology. The revised chamber volume is only 50% and the blanket volume is only 60% of their original estimates.

We found total stress to be surprisingly insensitive to wall thickness, other parameters being held constant. Secondary thermal stress increases but primary mechanical stresses decrease with wall thickness. These opposing effects result in a broad minimum to the total stress as a function of wall thickness.

Author: J. H. Pitts

Major Contributors: J. A. Blink and W. R. Meier (LLNL); M. J. Monsler (KMS Fusion); K. A. Murray (Energy Technology Engineering Center); R. L. Creedon, I. Maya, and K. R. Schultz (GA Technologies)

Balance-of-Plant Design

During 1983, we also investigated the balance of plant for Cascade. We define *balance of plant* as the equipment external to the reactor chamber used for converting fusion energy to electricity.

Because the Cascade chamber²⁴ must be kept at a vacuum of 0.1 Torr to allow laser beam propagation, we examined maintaining the entire primary heat-transport loop at

that pressure. This avoids the problem of keeping vacuum conditions in one portion of the loop and pressurized conditions in another. It does, however, restrict our ability to extract thermal energy from the Li_2O granules. GA Technologies is studying the feasibility of a primary heat-transport loop at higher pressures.

The thermal energy extracted from the Li_2O granules is used to generate steam in the once-through steam cycle of a conventional liquid-metal fast-breeder reactor (LMFBR). The parameters for this cycle are a steam-generator feedwater temperature of 500 K and a steam exit temperature of 730 K at 15.8 MPa. Although this cycle does not take advantage of the granules' high temperature to produce electricity more efficiently with a topping or bottoming cycle, we decided that it would allow us to focus on design issues for the primary heat-transport system rather than on the electrical conversion system. A higher efficiency cycle can be incorporated into the design later.

Cascade's gross thermal power output of 3300 MW, and primary-loop temperature change of 400 K result in a required Li_2O throughput of 10^7 kg/h. At an assumed packing fraction of 0.5, the volumetric throughput is $2.8 \text{ m}^3/\text{s}$. We assumed that at 1000 K, Li_2O granules would have a heat capacity of $290 \text{ J/kg} \cdot \text{K}$, a thermal conductivity of $3.7 \text{ W/m} \cdot \text{K}$, and a bulk density of 1800 kg/m^3 . The latter two properties are for particles having 10% internal voids.

Steam Generator Design. We examined a vertical shell-and-tube steam generator incorporating double-wall tubes like those used in the Westinghouse breeder reactor. The Li_2O granules flow vertically downward past the tubes in a vacuum on the shell side. The double-wall design limits the migration of tritium from the granules into the water. The tubes have an outer diameter of 20.6 mm and an inner diameter of 11.1 mm. The heat-transfer coefficients through the tube wall and on the tube side have been measured by Westinghouse²⁵ as $5.7 \times 10^3 \text{ W/m}^2 \cdot \text{K}$ and $1.4 \times 10^4 \text{ W/m}^2 \cdot \text{K}$, respectively. Because of the high temperatures on the outside of the tube wall, 2-1/4 Cr-1 Mo steel cannot be used for the tubes. However, Hastelloy C can withstand the temperatures and stresses for the plant lifetime.

Granule diameter (mm)	Residence time (s)	Interval between crossbars (m)
0.1	1.75	26.6
0.2	3.5	6.7
1.0	17.3	0.27
2.0	34.2	0.07
10.0	156.3	0.003

Table 7-2. Residence time vs interval between crossbars for varying granule diameters.

We calculated a heat-transfer coefficient between the granules and tube walls from a correlation by Schlünder.²⁶ Heat transfer is predominantly by thermal radiation, and its coefficient equals $120 \text{ W/m}^2 \cdot \text{K}$. This coefficient is for surface heat transfer only; if it is used for the entire bed, perfect mixing is implied. The other extreme, where there is no mixing and only granules adjacent to the tube walls transfer heat, is highly conservative and will result in the maximum heat-transfer surface area required.

Narrow horizontal crossbars are placed inside the steam generator on the shell side, so that granules traveling around the bars will experience turbulence and mixing. For this study, we conservatively assumed that a heat-transfer coefficient equal to half that for perfect mixing (or $60 \text{ W/m}^2 \cdot \text{K}$) could be achieved. The final design then has 10 steam generators, each 20 m tall and 6.5 m in diameter, and each with 11 700 tubes for a total heat-transfer surface area of $15 000 \text{ m}^2$.

How long the Li_2O granules must stay in the steam generator to release enough heat varies with their size, as shown in Table 7-2. The granule flow rate can be selected to achieve the desired residence time by installing the appropriate number of crossbar rows. We calculated the crossbar interval for each granule size, assuming that the granules fall freely between the crossbars and have zero velocity when they encounter a crossbar. A residence time of zero was assumed at each crossbar; i.e., when a particle hits a crossbar, it immediately begins the next free-fall. Crossbar intervals are also given in Table 7-2.

The 0.1-mm-diam granule is not feasible, because the needed crossbar spacing is greater than the steam generator's height. This is because the velocity to obtain the desired residence time is greater than the average velocity attained in free-fall from 20 m.

The 10-mm-diam granule is also not feasible, because it requires crossbar intervals

of 3 mm, and the bars cannot be packed that closely. As a result, the desired granule size is between 0.2 and 2.0 mm. Because mixing can be obtained more easily with larger granules, the 2-mm size was chosen. It results in the smallest heat-transfer area of those granules with acceptable residence times.

Vacuum Pumping Requirements. For a Cascade design having a burn fraction of 0.33 (Refs. 24 and 27), DT fuel is injected at a rate of 28 mg/s. A tritium-breeding ratio of 1.35 then results in a tritium pumping requirement of 3.1×10^{-3} g·mole/s. The pumping requirement for unburned deuterium is 1.9×10^{-3} g·mole/s. Helium is formed at 4.4×10^{-3} g·mole/s. A total of 9.4×10^{-3} g·mole/s needs to be pumped to maintain a steady vacuum.

The reaction chamber pressure is to be kept to 0.1 Torr. Assuming that the maximum temperature of the gas is 1200 K (the temperature of the granules) and using the ideal gas law, we calculated a total volumetric flow rate of 6800 liters/s. Because the gas must be discharged at atmospheric pressure, a primary pump system followed by a secondary and possibly a tertiary system is required.²⁸

Turbopumps were chosen as primary pumps. They are currently used at the Tokamak Fusion Test Reactor and eventually will be used to pump tritium. Turbopumps are frictionless—the only energy losses are due to adiabatic compression. In comparison, Roots blowers have a high friction loss and use more power. Turbopumps are available with magnetic bearings that require no lubrication. They have maximum efficiency at 0.01 Torr and discharge at 1 Torr.

A good secondary-pump candidate is the scroll pump, a moving-spiral positive-displacement pump. Scroll pumps are being used at the Tritium System Test Assembly; they discharge at 100 Torr. For tertiary pumps to compress the air from 100 Torr to atmospheric pressure, metal bellows pumps can be used.

The total power to operate the vacuum system is about 0.1 MW.

Plant Power Balance. The power requirements for the primary heat-transfer system in Cascade are lower than in most previous ICF reactor designs because of the

low primary flow rate. The system multiplication factor for Li_2O is 1.11 (Ref. 24). The electrical conversion efficiency has not yet been optimized for Cascade, but 44% has been obtained in other plants without topping and bottoming cycles. Cascade's high operating temperature should permit efficiencies substantially above this.

The in-plant recirculated power requirement for Cascade is 52 MW_e. This value is based on an 1100-MW_t LMFBR design and on the ratios to the plant gross-thermal output used in the HYLIFE design, except for the vacuum system, Li_2O circulation, lead coolant system, and reactor rotation. This results in a 40% net efficiency. Inclusion of topping and bottoming cycles could increase this net efficiency to 51%.

Authors: K. A. McFaray (Energy Technology Engineering Center) and J. H. Pitts (LLNL)

Major Contributors: J. A. Blink (LLNL) and M. W. McDowell (Energy Technology Engineering Center)

Solid-Granule Transport System

After the Li_2O granules are ejected from the reaction chamber, they must be transported to the steam generator inlets, located 20 m above the chamber outlet. We investigated a number of carrying and noncarrying transport methods to determine the one with the lowest power consumption.

Noncarrying conveyors push material along by blades or scrapers; this usually leads to a low efficiency and high power consumption. Carrying conveyors use belts or buckets to transport the material, and efficiencies tend to be higher. Table 7-3 gives the number and size of conveyors needed to transport 10^7 kg/h of 50%-dense Li_2O between the reaction chamber and the steam generators. For each case, the maximum possible inclination was used to minimize building size.

To avoid elevation of either the reactor or the steam generators, we need two sets of lifts: one to carry the granules up 20 m from the reaction chamber outlet to the steam generator inlets, and another to carry the granules up 10 m from the steam generator outlets to the chamber inlets. The

Cascade Reactor

Transport method	Type	Maximum incline (deg)	Conveyors		Power consumption (ME _g)
			Number	Size	
Flight conveyor	Noncarrying	40	100	1 m wide × 0.3 m high	3.6
Continuous flow elevator	Noncarrying	90	50	1 m wide × 0.3 m high	8.7
Screw conveyor	Noncarrying	35	330	1-m diam	5.5
Archimedes lift pump	Noncarrying	60	100	1-m diam	2.0
Gravity discharge elevator	Carrying	90	50	1 × 0.4 × 0.6-m buckets	1.7
Belt conveyor	Carrying	20	10	1 m wide	1.1

Table 7-3. Power consumption and size of granule transport methods investigated.

Fig. 7-13. Solid-granule transport system for the steam generators in Cascade. A single belt is shown instead of ten 1-m-wide belts.

power consumptions in Table 7-3 are based on the total 30-m elevation change for the two sets of lifts. With the exception of the Archimedes pump, the power consumptions were calculated from information given in *Marks Standard Handbook for Mechanical Engineers*.²⁹

The belt conveyor has the lowest power requirement of the methods investigated and is the most widely used in industry. It

is limited, though, to a maximum slope of 20 deg with loose, free-flowing, granular material. This is a disadvantage in that more space is needed than for systems with a vertical carrying capability. However, a belt conveyor has a high volume capacity; its maximum speed for light, free-flowing material is 4.3 m/s. It also has a low granule attrition. As a result, the belt conveyor seems to be the best transport method. A

schematic of such a conveyor system is shown in Fig. 7-13.

The transport system between the reaction chamber and the steam generator must operate at the Li_2O hot-leg temperature of 1200 K, which is out of the range of conventional materials used in transporting solids. The construction material and operating method of a belt conveyor at 1200 K thus need to be determined. The belt conveyors must also be housed in vacuum ducting that is thermally insulated on the outside.

Of the noncarrying conveyors, the Archimedes lift pump has the lowest power consumption. This pump is a screw conveyor in which the outer case rotates with the screw. The granules advance along the screw by means of displacement by the granules below. The material is held in the screw by gravity, so there is a limit on the incline to keep the granules from sliding backwards. In the Solase design,³⁰ this inclination is 60 deg. Although this is a steeper incline than that of a belt conveyor, 100 Archimedes pumps would be needed, compared to only 10 conveyor belts; the belts thus do not take up significantly more floor space.

At this time, therefore, the belt conveyor seems the best granule transport method because of its low power consumption, high volume capacity, and low granule attrition. However, other conveyor methods, such as the gravity discharge elevator, have the advantage of a larger maximum incline and may ultimately be chosen to minimize space requirements.

Authors: K. A. Murray (Energy Technology Engineering Center) and J. H. Pitts (LLNL)

Major Contributors: J. A. Blink (LLNL) and M. W. McDowell (Energy Technology Engineering Center)

Granule Thermal-Stress Analysis

The pulsed energy released from each DT pellet will subject the blanket granules facing the center of the Cascade reaction chamber to large successive loads from reflected laser light, x rays, alpha particles, and pellet debris.³¹ We completed a thermal

and structural analysis on these first-layer granules to determine:

- The feasibility of avoiding granule melting and vaporization by increasing the size of the Cascade reaction chamber.
- The amount of melting and vaporization that will occur if they cannot be avoided.
- The severity of thermal stress-induced cracking and the sensitivity of thermal stress to granule size.

We can estimate the size of the reaction chamber that is required to avoid melting and vaporization by using a one-dimensional, unsteady-state conduction equation and approximating the first-layer granules as a semi-infinite slab. Assuming that the heat flux, $\dot{q}''$, is constant during the pulse and that the granule is initially at constant temperature, T_0 , the granule's surface temperature, T_s , can be approximated by

$$T_s = T_0 + 1.13 \frac{\dot{q}''}{k} \sqrt{\alpha t_p} \quad (18)$$

where k (3.9 W/m K at 1200 K, 2.8 W/m K at 1700 K) is the granule's thermal conductivity; α ($4.5 \times 10^{-7} \text{ m}^2/\text{s}$ at 1700 K, $2.5 \times 10^{-7} \text{ m}^2/\text{s}$ at 2900 K) is the thermal diffusivity; and t_p is the pulse duration.¹² [Numerical values in parentheses, both here and following Eq. (22), are for Li_2O granules having a density of 2000 kg/m³.] The heat flux can be approximated by

$$\dot{q}'' = \frac{fQ}{t_p A_s} \quad (19)$$

where f is the fraction of the total fusion energy deposited on the granule's surface, equal to 0.32 (Ref. 33); Q is the total fusion energy, equal to 600 MJ (Ref. 31); and A_s is the granular surface area of the first layer. For a spherical reaction chamber, where t_p is considered equal to the blanket's inner radius, r_c , divided by the velocity, v , of the slowest fusion-reaction product [$2(10^6) \text{ m/s}$ for the pellet debris], we find that

$$r_c = 0.09 \left[\frac{fQ \sqrt{v\alpha}}{k(T_s - T_0)} \right]^{2/5} \quad (20)$$

With a Li_2O granule blanket, the distances r_c that are required between the chamber center and the first-layer granules to avoid melting and vaporization are 42 and 28 m, respectively. Because such large reaction chambers are required to avoid melting or vaporization, this approach is not

economically or mechanically feasible. Therefore, some melting and vaporization of first-layer granules will occur.

To estimate the upper bounds of melting and vaporization, we chose a blanket inner radius of 4 m; we assumed that the total energy intercepted by the frontal area of the granules is used to melt or vaporize material, and that nonneutron energy is all absorbed at the granule's surface. The maximum amounts of material that will melt, m_m , or vaporize, m_v , are

$$m_m = \frac{fQ}{H_f + c_p(T_m - T_0)} \quad (21)$$

and

$$m_v = \frac{fQ}{H_f + H_v + c_p(T_v - T_0)} \quad (22)$$

where H_f (2×10^6 J/kg) is the latent heat of fusion, H_v (2.7×10^7 J/kg) is the latent heat of vaporization, c_p (3000 J/kg K) is the specific heat capacity of Li_2O , and T_m (1700 K) and T_v (2900 K) are the granule's melting and vaporization temperatures. For conditions in the Cascade chamber, the most Li_2O that would melt, assuming no vaporization, is 55 kg. Alternatively, the most Li_2O that would be vaporized, assuming no melting, is 5.6 kg. In these limiting cases, the thickness of the maximum melt layer is 74 μm (10% of the first-layer granules), and that of the maximum vaporization layer is 7 μm . These results are not operationally restrictive, and if vaporized material is redeposited locally, 2-D thermal analysis results show that the granules would solidify between pulses.

In reality, the physics of the granules' change of state as the fusion energy is conducted through them, and the dynamics of the ablated material blowoff and potential granule shock formation, are extremely complex. However, by estimating some fundamental time constants, we can gain a rough idea of the granule dynamics.

The first-layer granules could fracture from either mechanical or thermal stresses. The mechanical stresses are caused by the compressive shock wave from the pulsed energy deposition. This shock wave travels through the granule, then reflects back in tension. Because ceramics such as Li_2O are much stronger in compression than tension

and the granules are assumed to be unconstrained at the surface, any fracture from mechanical stresses would occur in the back of the granule. The thermal stresses are caused by the sudden heating of the granule's surface, which causes a thin layer on the front face to expand. This expansion will pull on the rest of the granule, putting it in tension. Any fracture from thermal stresses would occur along the granule's front face, where the thermal gradient is highest. In this work we analyzed only thermal stresses.

We calculated the transient temperature distribution of a first-layer granule in Cascade with TACO2D, a two-dimensional, finite-element heat-transfer code.³⁴ To determine the stress in the granule, we then coupled the temperature distribution at the end of the pulse into NIKE2D, an implicit, finite-deformation, finite-element stress code.³⁵ The transient analysis was also used to determine if melted or vaporized material would recondense between pulses.

The end-of-pulse time was examined because we expect that this is when the largest thermal gradient, and thus the largest thermal stresses, will occur. The granule's center of mass was constrained from moving in any direction, but surface motion was unconstrained.

Figure 7-14 shows the thermal-stress contours of a 1-mm-diam, first-layer Li_2O granule. The granule's maximum thermal stress of 1340 MPa is almost 50 times greater than its fracture strength. However, this stress occurs only within a small surface region at the equator and will probably result only

Fig. 7-14. Thermal stress contours of a 1-mm-diam, first-layer granule. The granule would fragment with additional chipping on either side.

in local chipping. A more important stress figure is the maximum tensile-stress contour that dissects the granule; when this contour exceeds the fracture strength of Li_2O , the granule will fragment. The granule's 310-MPa maximum tensile-stress contour exceeds the Li_2O fracture strength by about a factor of 10. Therefore, we anticipate that first-layer granules will crack into large fragments with some chipping along the sides.

The remanufacturing implications of this cracking remain to be examined. However, two alternatives that avoid refabrication are available. One is not to refabricate cracked granules but to operate with a powdered blanket. Analysis of the flow dynamics of a powdered blanket would require a significant, parallel experimental program. The second alternative is to incorporate a special surface layer of granules that have a better response to the thermal and stress effects. Silicon carbide is one candidate. It has a factor-of-8 lower thermal-expansion coefficient than Li_2O , a factor-of-9 greater irradiated fracture strength, and a factor-of-2 greater thermal conductivity.

The thermal stresses decreased with increasing granule diameter because larger granules have more material to react to the expansion of their front surfaces. However, stress dependence on granule diameter was not strong; the tensile stresses decreased by only 3% from 1- to 2-mm-diam granules.

Figure 7-15 presents the temperature profile within a 2-mm-diam granule immediately before the next pulse. Granule temperature is below the 1700-K melt temperature of Li_2O everywhere except in a very thin region facing the original energy

flux direction. Because the initial temperature for this granule was a uniform 1200 K (i.e., just before it exits from the reaction chamber and ignoring temperature gradients from previous pulses), we concluded that first-layer granules will solidify between pulses. In addition, a granule's orientation would probably be altered between pulses, enhancing heat transfer to granules within the blanket.

Based on our thermal and structural analysis of the first-layer Li_2O granules, we conclude:

1. It is not feasible to avoid granule melting and vaporization by increasing the size of the reaction chamber.
2. A maximum of 10% of each first-layer granule will melt as a result of absorbing energy from a single fusion reaction. However, the granule's bulk temperature rise is not significant, and the granules will resolidify between pulses.
3. The maximum vaporization layer that will blow off into the reaction chamber is 7 μm thick, containing a total of 5.6 kg of Li_2O . We have assumed that this material is deposited locally. This assumption needs to be verified, however, because the otherwise implied mass and energy transport could affect the chamber's vacuum pumping requirements and system power balance.
4. The thermal stress contours that dissect a granule are about 10 times greater than the granule's fracture strength. Therefore, we anticipate that first-layer granules will crack into large fragments, with some additional chipping along the sides where local stresses are much higher.
5. The granule's tensile stresses decrease with increasing diameter, but only slightly. Therefore, granule size should be determined by ease of manufacturing, heat transfer, tritium retention, and cost.

Authors: H. E. Levine (MIT, formerly at GA Technologies); I. Maya and K. R. Schultz (GA Technologies); and J. H. Pitts (LLNL)

Tritium Inventory and Recovery

In the Cascade design,³⁶ a layer of high-temperature lithium-ceramic granules is centrifugally held against the walls of a rotating reaction chamber. These granules absorb the neutrons released from the

Fig. 7-15. Temperature profile of a 2-mm-diam, first-layer granule just before the next energy pulse.

Table 7-4. Cascade reference design parameters.

Parameter	Value
Thermal power (MW)	3000
Granule radius (mm)	0.5
Average blanket temperature (K)	1000
Blanket porosity (%)	50
Vacuum-chamber total pressure (Pa)	13
Vacuum-chamber T ₂ O pressure (Pa)	1.3
Tritium-breeding ratio	1.3
Total Li ₂ O mass, including Li ₂ O circulating outside the reaction chamber (kg)	10 ⁶

laser-induced fusion reactions, producing thermal energy and tritium. The granules are circulated to transport their energy to the power conversion system. Tritium is recovered by the vacuum system that sustains the chamber vacuum. This article describes how tritium inventory and recovery affect the Cascade design when Li₂O is the granule material.

Tritium, bred by neutron capture in the Li₂O, undergoes a three-step process before its recovery by the vacuum system. The first step is bulk diffusion within the granule. The chemical form in which the tritium will be released from the Li₂O is uncertain at this time. Some studies^{37,38} of neutron-irradiated Li₂O indicate that the tritium is released in its oxide form (T₂O); other preliminary studies,^{39,40} in which the tritium recovery is in situ, indicate that the tritium is released in its elemental form. Experiments on tritium recovery from Li₂O under the conditions expected in Cascade have not been performed. We assume that tritium is released in the T₂O form, based on the lesser uncertainty of the former experiments. The second step is desorption of the T₂O at the granule surface. This is assumed to be instantaneous. The third step is percolation of the T₂O through the interstices in the granule blanket.

The reference design parameters of Cascade are presented in Table 7-4. The objective of our study was to investigate the sensitivity of tritium inventory to the following design parameters:

- Granule radius from 0.01 to 10 mm.
- Blanket porosity from 0.4 to 0.6.
- Blanket temperature distribution from 800 to 1700 K.
- Vacuum-chamber tritium partial pressure from 0.13 to 13 Pa.

The results of the parametric analyses then formed the basis for establishing the tritium recovery requirements and the recommended recovery method.

Tritium Analysis Code. We used GA Technologies' TRIT4 code, a one-dimensional extension of the steady-state TRIT3 code,⁴¹ to determine the tritium inventory in an Li₂O blanket. The TRIT4 code models tritium inventory in the blanket in three principal modes:

- Bulk diffusion inventory within the Li₂O granule.
- Porous diffusion inventory in the interstices among the granules.

- Soluble inventory within the solid Li₂O.

The bulk diffusion inventory is calculated from the tritium generation rate and temperature profiles in the blanket, the granule size, and Fick's law, using the diffusion coefficient data of Nasu et al.⁴² An expression for the average tritium concentration in a granule is obtained by integrating the steady-state tritium concentration profile for a sphere over the radius and is expressed as⁴³:

$$\text{bulk diffusion inventory} = \frac{S r_g^2}{15D} \times V_B \quad (23)$$

where

S = tritium generation rate⁴⁴ (kg/m³),

r_g = granule radius (m),

D = bulk diffusion coefficient (m²/s),

V_B = solid Li₂O volume (m³).

A plot of the diffusion coefficient in Li₂O vs temperature for various gases is shown in Fig. 7-16. We used the data of Nasu et al., reproduced from Ref. 45, in constructing this plot.

The gaseous T₂O pressure profile within the blanket is calculated using a percolation model of porous diffusion.⁴⁶ Tritium released at the surface of a granule must travel through the interstices among granules before reaching the chamber vacuum system. This travel will create a concentration profile within the blanket. The profile, plus the T₂O partial pressure within the vacuum chamber, constitutes the porous diffusion inventory. Porous diffusion directly influences the soluble inventory within the granule by affecting the tritium concentration at the granule surface.

The soluble inventory of tritium in the form of LiOT in the solid Li₂O is determined by the thermodynamic equilibrium of the reaction

Fig. 7-16. Temperature dependence of bulk diffusion coefficients for T, O, Li, H, and D in Li_2O .

The equilibrium constant, K_p , for this reaction is given in Ref. 47 as a function of temperature:

$$\log K_p = -\frac{7073}{T} + 2.102 \log T - 0.00254T + 2.767 \quad (25)$$

where T is the temperature in kelvins and K_p is expressed in atmospheres. The chemical activity of LiOT can then be obtained using

$$a_{\text{LiOT}} = \sqrt{\frac{p_{\text{T}_2\text{O}}}{K_p}} \quad (26)$$

for the T_2O partial pressure of 1.3 Pa in Cascade. The mole fraction of LiOT , x_{LiOT} is then given by

$$x_{\text{LiOT}} = \frac{a_{\text{LiOT}}}{\gamma_{\text{LiOT}}} \quad (27)$$

where γ_{LiOT} is the activity coefficient. The activity coefficient was recently evaluated

by Norman⁴⁸ and found to be nearly linear on a semilog plot of reciprocal temperature from 800 to 1300 K. The ratio of tritium to lithium atoms is determined from the mole fraction using

$$\frac{\text{T atoms}}{\text{Li atoms}} = \frac{x_{\text{LiOT}}}{2 - x_{\text{LiOT}}} \quad (28)$$

The tritium inventory in kilograms follows directly from Eq. (28) and the 10^6 -kg Li_2O inventory in Cascade.

Tritium Analysis Results. The total steady-state tritium inventory of Cascade's granule transport loop, heat exchangers, and power conversion system is 6 kg, assuming an average Li_2O temperature of 1000 K. The steady-state radial profile of the tritium inventory in the Li_2O blanket is shown in Fig. 7-17. The inventory is expressed in kilograms of tritium per centimeter of radial blanket thickness. The figure shows that the major contributors are the bulk diffusion and soluble inventories. The bulk diffusion inventory is highest at the inner surface of the blanket, following the radial dependence of the tritium generation rate. The porous diffusion inventory is highest near the outer radius, because the tritium must flow toward the center of the chamber. Its contribution to the total inventory is negligible, but its magnitude is important through the square-root dependence of the soluble inventory on the T_2O pressure profile. The soluble inventory is highest toward the outer radius of the blanket, because it follows the porous diffusion inventory.

The values of the diffusion coefficient shown in Fig. 7-16 indicate that the time scale for bulk diffusion will be long relative to the approximately 1-min residence time of a granule within the chamber. In addition, upon exit from the reaction chamber, the Li_2O granules will be agitated during transport to the power conversion system. This will probably eliminate the percolation pathway and reduce the porous diffusion profile to that due to the T_2O partial pressure within the vacuum chamber, thus decreasing the soluble and porous diffusion inventories. This effect is qualitatively significant with granule radii of less than 0.1 mm or with blanket porosities of less than 30%.

The design parameter that most affects tritium inventory is the granule size. A plot of tritium inventory vs granule radius is shown in Fig. 7-18. A minimum of 4 kg of tritium is obtained for 0.1-mm-radius granules. For granules below 0.7-mm radius, the soluble contribution dominates the total tritium inventory; bulk diffusion dominates for larger granules. The porous diffusion inventory decreases with increasing granule size because of the larger pores and decreased path length for escape from the blanket.

Between 800 and 1700 K, the total tritium inventory is fairly constant. We attribute this to the opposing effects of temperature on the individual soluble and bulk diffusion contributions. As the temperature increases, the diffusion coefficient for tritium in Li_2O increases, decreasing the bulk inventory. However, the activity coefficient for LiOT in Li_2O decreases, increasing the soluble inventory. The opposing dependences effectively cancel one another.

Finally, we found that raising the chamber T_2O partial pressure tenfold to 13 Pa would increase the tritium inventory to 16 kg. This increase is due to the change in the soluble inventory brought about by the change in the porous inventory; it indicates a relative insensitivity of the tritium inventory to the chamber T_2O partial pressure.

Conclusions. The reference Cascade design would have a total tritium inventory of 6 kg (2g/MW). Although selecting the optimum granule size requires considerations of fabrication cost and thermal stress analysis, granules of up to 1.5-mm radius would satisfy the current 10-g/MW_t safety guidance on maximum tritium inventory.⁴⁵ Most of the tritium inventory in the total circulating Cascade blanket (i.e., all except the porous diffusion contribution) is considered quite secure, having release times on the order of days or longer in the parameter range of interest. Thus, instantaneous release is not possible. These features of acceptable tritium inventory, low leakage, and simplicity of tritium recovery preserve the safety, environmental, and economical appeal of the Cascade design.

Authors: I. Maya and K. R. Schultz (GA Technologies); J. J. Buksa (U. of New Mexico); D. F. Roelant (U. of Michigan); and J. H. Pitts (LLNL)

Fig. 7-17. Sensitivity of tritium inventory per centimeter of radial blanket thickness to blanket radius for the Cascade reference design.

Fig. 7-18. Tritium inventory vs granule radius for a blanket with a fully developed porous diffusion profile, other parameters as given in Table 7-4.

Neutronics and Activation Analyses

We investigated three wall configurations for Cascade during 1983: a 2.75-cm-thick 2-1/4 Cr-1 Mo steel wall, a 2.75-cm-thick silicon carbide wall that was prestressed in compression with a network of aluminum cables (SiC/Al), and the same SiC wall prestressed with cables made of 2-1/4 Cr-1 Mo steel (SiC/steel).⁴⁹ The basic steel wall was our initial choice.⁵⁰ However, heating from the hot Li_2O and from neutrons will require spray cooling of the wall's outer surface. The SiC wall was proposed because it can run hot, allowing adequate heat removal by radiation to the reflector. The prestressing cables would be insulated from the SiC so that only their internal neutron and gamma

Cascade Reactor

heating would be removed by radiation. Since aluminum cables must operate at a lower temperature than steel, the design problem is more difficult for them. The neutronics and activation calculations described here served as input to the thermal and mechanical analyses of the chamber wall reported earlier in this section in "Chamber Design."

Neutronics Model. We modeled the Cascade chamber as shown in Fig. 7-19. The equatorial plane was taken as a reflecting boundary. The compressed-DT fuel and pusher regions were included. The Li_2O thickness at the midplane was 96 cm, with an average density of 1.0065 g/cm^3 (50% of

Fig. 7-19. Cascade neutronics model. The Li_2O blanket and chamber wall rotate; the reflector is stationary.

Table 7-5. Neutronics results for the three Cascade wall configurations.

	Steel	SiC/Al	SiC/steel
Energy deposition (MeV)			
Li_2O^a	19.41	19.35	19.41
Wall	0.083	0.037	0.035
Cables		0.013	0.039
Reflector	0.126	0.130	0.117
Total	19.62	19.53	19.60
Multiplication			
$E_{\text{Li}_2\text{O}}/E_{\text{fusion}}$	1.104	1.101	1.104
$E_{\text{thermal}}/E_{\text{fusion}}$	1.116	1.111	1.115
Energy leakage (MeV)			
Port ^b	0.115	0.113	0.107
Corner	0.032	0.073	0.067
Reflector	0.062	0.015	0.016
Total	0.208	0.200	0.190
Neutron leakage (%)			
Port	2.12	2.04	1.76
Corner	2.96	2.99	2.74
Radial	0.42	0.29	0.31
Total	5.50	5.32	4.81
Tritium-breeding ratio	1.34	1.36	1.35

^aIncludes α and target-neutron interaction energies.

^bPorts represent 0.78% of reactor area.

the maximum possible value). The wall was modeled as a series of cones. Prestressing cables were modeled as a continuous shell including the same amount of material as discrete cables. Neutron leakage through the beam port, the reflector, and the corner between them was monitored. Because an earlier neutronics calculation without reflection showed considerable neutron leakage, we included a steel-canned graphite reflector in these calculations. The material compositions used in the calculation are listed in Ref. 49. Care was taken to include impurities because their activation sometimes dominates.

The neutronics calculation was performed using the ALICE⁵¹ three-dimensional, continuous-energy, Monte Carlo, neutron-photon transport code. ALICE is a modification of TART⁵² that treats the resonance region in more detail. The calculations used 18 000 neutron histories and produced reasonable standard deviations (a fraction to a few percent in the Li_2O regions; a few to a few tens of percent in the wall regions).

Neutronics Results. The calculated results for the three wall configurations are shown in Table 7-5. The fusion multiplication factor of about 1.11 is low compared to other ICF reactors (1.26 in Pulse-Star, 1.185 in HYLIFE). For the same recirculating power, this low multiplier must be compensated for by increased driver efficiency, thermal efficiency, or target gain. Fortunately, the high Li_2O temperature in Cascade offers a chance for high thermal efficiency that more than compensates for the low fusion multiplier.

If the wall and reflector energy is recovered at a low temperature, the effective multiplication factor falls further. The ratio of 1.10 for Li_2O energy to fusion energy approximates this lower value. However, the actual effective multiplier depends on the high- and low-temperature thermal efficiencies and on the amount of Li_2O energy that leaks through the wall.

The equatorial Li_2O region was zoned radially (a 1-cm-thick inner zone followed by three equally thick outer zones). The neutron-plus-gamma energy-deposition profile peaks at 96 kJ/kg at the inner surface, falls to 13 kJ/kg by the middle, and then falls to 4 kJ/kg at the outer edge. The

neutron energy leakage is about 1% of the thermal energy, with most of the loss in the port area (including the corner). About 5% of the neutrons leak, i.e., about six times the port surface-area fraction. The leakage through the ports themselves is ~2%, or about three times the port area. The tritium-breeding ratio is around a healthy 1.35 for all three configurations.

The wall/cable heating is 14.2, 8.5, and 12.6 MW_t for the three configurations, assuming 3000 MW of fusion power. These levels preclude radiation cooling of the basic steel wall, even if heat transport from the Li₂O could be prevented. For the SiC wall (~6 MW_t), radiation cooling is feasible because the allowable temperature is higher. However, neutron heating of the low-temperature aluminum cables is high enough (2.2 MW_t) to preclude radiation cooling, even with perfect insulation between the cables and the SiC. For the mid-temperature steel cables, neutron heating (6.7 MW_t) is borderline. With superb insulation from the SiC and careful attention to emissivity maximization, the steel cable design might be radiation cooled. However, it is prudent to continue to base the design on spray cooling, since only one wall configuration is even marginal for radiation cooling.

Further analysis of the SiC/Al heating indicates that most of the aluminum heating is due to gamma energy originating in the Li₂O from the ⁷Li(n,γ) reaction. Four

potential methods to reduce this heating are:

1. Mixing 5 to 10% lead with the lithium in the Li₂O.
2. Adding 5 to 10 cm of a high-temperature shield like tungsten or tantalum just inside the wall.
3. Increasing the ⁶Li enrichment of the lithium.
4. Increasing the Li₂O region thickness.

We are evaluating these solutions in the hope that spray cooling of the wall's exterior can be eliminated.

Activation. Activation calculations are used to determine afterheat and biological-hazard potential (bhp) for safety analyses, dose rate for establishing maintenance procedures, and the feasibility and cost of waste disposal. The energy-dependent fluxes determined by ALICE were used to energy-average the ACTL cross-section library⁵¹ for each spatial region using the ORLIB code.⁵³ The Li₂O zones were spatially averaged together, because mixing will eliminate any spatial flux-distribution effects. The wall zones were also spatially averaged, because they have similar neutron spectral shapes. The cross sections were used in the FORIG code^{54,55} to determine activity as a function of time. We assumed 30 y of flux at 3000 MW_t and a 70% capacity factor.

Activation of the Li₂O, steel, SiC, aluminum, and reflector are shown in Table 7-6 at various times during and after operation. During operation, the ¹⁶N in Li₂O

Table 7-6. Activation (kCi) for Cascade components.

Wall	During operation		After shutdown			
	1 mo	30 y	1 d	1 mo	50 y	500 y
Steel						
Li ₂ O	221 000	183 000	636	170	6.3	4.7
Steel	10 500	17 200	8 530	6 250	8	4
Reflector	6 580	11 100	5 800	5 030	1.7	1.6
Total	238 000	211 000	15 000	11 500	16	10
SiC/Al						
Li ₂ O	223 000	183 000	796	443	6	5
SiC	2 210	2 210	0.73	0.34	0.002	0.0017
Al	6 990	6 980	170	0.58	0.013	0.00024
Reflector	5 050	10 900	6 220	5 350	1.8	1.6
Total	237 000	203 000	7 190	5 790	8	7
SiC/steel						
Li ₂ O	225 000	184 000	798	444	6.4	4.7
SiC	1 800	1 800	0.65	0.31	0.0018	0.0014
Steel	5 920	10 600	5 620	4 300	5.5	2.5
Reflector	5 110	11 200	6 470	5 590	1.8	1.6
Total	238 000	208 000	12 900	10 300	13.7	8.8

Cascade Reactor

dominates the activity for all three configurations. After shutdown, the ^{55}Fe in the reflector and steel wall quickly take over. After several decades, ^{14}C in the Li_2O and ^{93}Mo in the reflector and steel dominate. All three wall configurations have similar activities (within a factor of 2) at all times during and after shutdown.

Activity after shutdown has plateaus at 1 d to 1 mo and 50 to 500 y. Thus, if the reactor has safety or maintenance problems 1 d after shutdown, waiting an additional month probably will not help. Also, if the

Table 7-7. Decay heat and surface dose rate for Cascade components.

Wall	Decay heat		Surface dose rate ^a		
	Shutdown ^b (kW)	1 d after shutdown (kW)	After shutdown		
			Shutdown (R/h)	1 d (R/h)	1 mo (R/h)
Steel					
Li ₂ O	7 620	0.2	7 050 000	324	185
Steel	118	13.8	222 000	25 400	19 800
Reflector	73.4	4.8	87 400	7 170	6 120
Total	7 820	18.8			
SiC/Al					
Li ₂ O	7 630	0.39	7 050 000	328	186
SiC	37.7	0.007	75 400	15	14
Al	125	4.7	278 000	14 100	0.33
Reflector	75.5	6	87 500	6 860	5 850
Total	7 870	11.1			
SiC/steel					
Li ₂ O	7 690	0.39	7 120 000	330	187
SiC	30.7	0.006	61 200	14	13
Steel	67.5	7.8	139 000	15 500	11 900
Reflector	76.9	6.0	89 000	6 720	5 720
Total	7 870	14.2			

^aNo totals are given because mutual shielding precludes simultaneous exposure to all components.

^bAfter 30 y.

Table 7-8. Inhalation bhp (km³) for Cascade components.

Wall	Shutdown ^a	1 d after shutdown
Steel		
Li ₂ O	6.50×10^6	1.55×10^5
Steel	2.31×10^6	1.87×10^6
Reflector	1.21×10^6	9.12×10^5
Total	1.00×10^7	2.54×10^6
SiC/Al		
Li ₂ O	6.51×10^6	3.16×10^5
SiC	7.67×10^4	1.01×10^3
Al	3.19×10^5	3.40×10^4
Reflector	1.18×10^6	9.46×10^5
Total	8.09×10^6	1.30×10^6
SiC/steel		
Li ₂ O	6.57×10^6	3.19×10^5
SiC	6.30×10^4	9.34×10^2
Steel	1.30×10^6	1.05×10^6
Reflector	1.18×10^6	9.43×10^5
Total	9.11×10^6	2.31×10^6

^aAfter 30 y.

components are not cheaply disposable 50 y after shutdown, waiting centuries probably will not help. These issues are further explored below, using the afterheat (decay heat after shutdown), inhalation bhp, surface-gamma dose rate, and shallow burial index as figures of merit. These figures of merit are discussed in more detail in the last two articles of this section and in Refs. 56 and 57.

Afterheat is important shortly after shutdown; thus, the shutdown and 1-d results summarized in Table 7-7 are most significant. Decay heat at shutdown is ~8 MW, and it quickly drops afterward. At 1 d, it is in the 11- to 19-kW range; the detailed output indicates that most of the decrease is in the first minute after shutdown. Because thermal radiation can remove a few megawatts, the decay heat can be easily removed by thermal radiation with no active cooling systems. The heat transport from the hot Li_2O through the cooler wall could be much more significant than the wall decay heat. However, a preliminary calculation⁵⁸ indicates that thermal radiation can also remove this heat.

Surface dose rate is important at times when maintenance could be performed; thus, the 1-d and 1-mo results shown in Table 7-7 are most significant. Levels above ~100 R/h probably preclude hands-on maintenance. The Li_2O must be drained before maintenance begins because of its activity. The steel wall, steel cables, and reflectors all require remote maintenance. The SiC/Al wall could be maintained with hands-on techniques if it were removed from the surrounding reflector region.

The inhalation bhp is the volume of air required to dilute vaporized material to breathing standards. If the plume occupies some tens of cubic kilometers, then Table 7-8 shows that, at 1 d after shutdown, a release fraction of $\sim 10^{-5}$ (20-km³ plume/2 × 10⁶ km³ bhp) of the reflector, steel wall, or cable steel would be tolerable. This is a factor of 10 to 100 better than some other ICF designs.⁵⁹ (See also the article on "Neutronics and Activation Analyses" in the discussion of Pulse-Star later in this section.) A release fraction of about 10⁻⁴ of the Li_2O would also be tolerable, as would 10⁻³ of the aluminum cables, or 1% of the SiC. Of course, the release

fraction depends on the accident scenario and on the major contributing nuclides, and thus these estimates are overly simple.

The shallow burial index was calculated for each reactor component (Table 7-9) by summing the ratios of the specific activities (in Ci/m³) to the 10CFR61 limits.⁶⁰ To arrive at the totals for the reactor, we used the total activity and total material volume to compute the specific activity for each nuclide. Thus, the reactor totals are much less than the arithmetic sum of their components. The Li₂O and reflector could be buried separately in all three cases, as could the SiC and the aluminum cables. The steel wall and cables cannot be buried alone. However, if they are buried as a unit with the Li₂O and reflector, only minor dilution is required. This dilution could be easily accomplished by including elements of the biological shield that surrounds the reactor. The addition of limits for other nuclides (such as ⁹³Mo) would increase the index; however, we believe shallow burial is still feasible.

The author has proposed an increase in the 0.2-Ci/m³ limit for ⁹⁴Nb, because that limit was set by assuming no surviving barriers or dilution, and the construction of homes in the (unrecognized) waste.⁵⁷ If the proposed limit of 1100 Ci/m³ is used, all three configurations could be buried with no dilution.

Summary. Cascade is a feasible reactor concept from the viewpoints of tritium breeding and activation. For the current design, the rotating reactor vessel will need external spray cooling during operation, but decay heat can be removed by thermal radiation alone after shutdown. The allowable release fraction of activated materials during an accident is considerably higher than in other designs. Remote maintenance will be required for steel walls and cables, and for the reflector. However, hands-on maintenance of a SiC/Al wall would be feasible after the wall is moved away from the reflector. This is the first fusion reactor with a hands-on maintainable first wall. Finally, disposal of activated reactor components by shallow land burial is feasible.

Author: J. A. Blink

Major Contributors: W. R. Meier and J. H. Pitts

Wall	Nuclides ^a				Total
	¹⁴ C	⁶³ Ni	⁹⁴ Nb	⁹⁹ Tc	
Steel					
Li ₂ O	0.513	0.011	0.0028	4.6 × 10 ⁻⁴	0.527
Steel	—	1.44	43.9	5.1	50.5
Reflector	0.0065	2 × 10 ⁻⁷	0.086	0.136	0.228
Total	0.387	0.0435	1.096	0.156	1.682
SiC/Al					
Li ₂ O	0.511	0.012	0.0029	4.7 × 10 ⁻⁴	0.525
SiC	0.0053	—	—	—	0.0053
Al	—	8.2 × 10 ⁻⁴	—	—	8.2 × 10 ⁻⁴
Reflector	0.0073	—	0.067	0.149	0.22
Total	0.380	0.0089	0.017	0.033	0.440
SiC/steel					
Li ₂ O	0.513	0.011	0.0029	4.7 × 10 ⁻⁴	0.527
SiC	0.0044	—	—	—	0.0044
Steel	—	1.99	41.05	4.25	47.2
Reflector	0.0076	—	0.068	0.143	0.218
Total	0.382	0.0346	0.562	0.088	1.067

^aOnly nuclides listed in 10CFR61 were considered.

Pulse*Star Reactor

Pulse*Star is an ICF reactor concept that uses a LiPb coolant in a pool geometry (Fig. 7-20). LiPb does not release great quantities of chemical energy in a fire, and the pool geometry makes it easier to safely transport the extremely dense fluid. The compact geometry and good neutronics of LiPb yield a net electric power density 29 times higher than that in an LWR containment building.

Pulse*Star is designed to operate at 5 Hz, using 600-MJ fusion pulses driven by a 2-MJ laser (gain = 300). The x rays and debris from the fusion pulses are absorbed by a wet screen canopy that acts as a first wall. The 4-m-radius screen is rewetted between pulses by LiPb droplets traveling through the 1.5-m-thick mist region between the screen and the 10-cm-thick, bell-jar reactor vessel. The liquid droplets, which occupy 50 vol% of this region are formed either by response to isochoric heating or by nozzles that penetrate the bell jar. Using liquid droplets reduces the shock transmitted to the reactor vessel and enhances the cooling and condensation of the LiPb ablated from the screen. The heated LiPb in the mist region falls into a hot pool within the bell jar. A steel-lined concrete neutron reflector reduces the liquid-metal inventory in this pool. Four pumps then remove the hot LiPb, force it through eight steam

Table 7-9. Shallow burial index at 50 y after shutdown. A value below 1.0 allows legal burial.

Fig. 7-20. Pulse*Star ICF reactor: (a) artist's concept; (b) top view of LiPb pool.

generators, and reinject it into the cold outer pool, completing the coolant flow circuit.

Corrosion considerations set an upper LiPb temperature limit of $\sim 500^\circ\text{C}$, and ductile-to-brittle transition considerations set a lower limit of $\sim 400^\circ\text{C}$. This 100°C temperature range and the plant's thermal power output dictate a $22\text{-m}^3/\text{s}$ LiPb flow rate. The low flow rate requires only four pumps, and the heat can be removed with eight steam generators; all fit easily into the 3-m-wide annular pool surrounding the bell jar, as shown in Fig. 7-20.

Two safety issues have been raised concerning the use of LiPb in a fusion reactor. First, tritium is almost totally insoluble in LiPb, leading to higher tritium pressures than in lithium systems. This means that special attention must be paid to designing steam generators and vacuum pumps that ensure adequate tritium containment. Second, lead activates into both long- and short-lived alpha emitters. Because these alpha emitters are similar to actinides produced in fission plants, they are of concern. The following two articles address these safety issues. The third article presents preliminary designs for the LiPb pumps and steam generators.

Author: J. A. Blink

Neutronics and Activation Analyses

As in our study of the Cascade reactor reported earlier in this section, we also ran neutronics and activation calculations on the Pulse*Star reactor design.⁶¹

Neutronics Model. Because the Pulse*Star reactor is not spherically symmetric, we performed a three-dimensional neutronics calculation using the ALICE⁶² continuous-energy, Monte Carlo, neutron-photon transport code. ALICE is a modification of TART,⁶³ which treats the resonance region in more detail. The geometry used in the calculation is shown in Fig. 7-21. Two beam ports of 4 deg half-angle cones are included, as are the beam port walls, defined by 4.27-deg half-angle cones. One of the 12 immersed steam generators and LiPb pumps was included, offset 90 deg from the beam ports. The material compositions used in the calculation are listed in Ref. 61. Because the neutronics model is an input to activation calculations, we paid particular attention to impurities in lithium, lead, and steel.

Within the reactor pool there is a total volume of 3970 m^3 . This is composed of 385 m^3 of central void, 5.6 m^3 of beam-port

void, 4.1 m³ of steel in the screen, 56 m³ of steel in the bell jar, 0.8 m³ of steel in the beam-port walls, 15 m³ steel in the pumps and steam generators, 148 m³ in the hot-pool reflector, 528 m³ of half-density LiPb in the mist region, and 2830 m³ of full-density LiPb. The total LiPb volume is 3090 m³.

Neutronics Results. The neutronics calculation used 18 000 neutron histories and produced reasonable standard deviations (a few percent) out to the bell jar and in the pool. The statistics in the steam generator (100% std dev), beam-port wall (23% std dev), outer cylindrical wall (100% std dev), and top wall (100% std dev) were marginal. No neutrons leaked through the bottom or sides, and only three neutrons leaked through the top. A total of 73 neutrons leaked through the beam port (19% std dev) or about 0.4%. This leakage is ~ 1.7 times the port area, much less than the multiplier of ~ 6 for a loop geometry. The leakage in Pulse*Star is lower because the port (at an 8.5-m radius) presents a very small angle to reflected neutrons compared to the port (at ~ 6 -m radius) in a loop reactor. The neutron energy leaked through the ports is 0.03 MeV/fusion, or about 0.25% of the neutron energy leaving the target. This leakage is roughly the port area, indicating that most excess leakage neutrons (0.7) are thermal.

A total of 22.198 MeV (including alpha energy) was deposited per fusion, leading to a fusion multiplication factor (thermal energy/fusion energy) of 1.263, or a blanket multiplication factor (thermal energy minus alpha energy, all divided by 14.08 MeV) of 1.330. Almost 90% of the energy was deposited directly in the LiPb. The tritium-breeding ratio was 1.22, including a contribution of 0.21 from the pool regions.

If a 5-Hz pulse rate and 42% thermal efficiency are assumed (these values are slightly higher than the preliminary ones used for steam generator and pump designs, but they appear feasible), the net electric power in Pulse*Star is 1426 MW. The net power density in the energy-absorbing volume (including the central void) is ~ 1.5 MW/m³, reasonably close to the 5.6 MW/m³ in a PWR pressure vessel. More important, the power density in the

pool is 29 times higher than that in an LWR containment building. This is a spectacular improvement over previous ICF and MFE reactor designs that had containment volumes larger than an LWR.

Activation Calculations. The ALICE energy-dependent fluxes were used to energy-average the ACTL cross-section library⁶² for each spatial region using the ORLIB code.⁶⁴ The LiPb regions were also spatially averaged, because mixing will eliminate any spatial flux-distribution effects. The cross sections were used in the FORIG code^{65,66} to determine activity as a function of time. Thirty years of flux at 4.33 Hz and 80% capacity was assumed for all reactor components except the screen, which is replaced every five years.

Our detailed calculations show that activated target debris is of little consequence if

Fig. 7-21. Neutronics model of Pulse*Star used in calculation.

Table 7-10. Activity (kCi) of Pulse*Star components. ▼

Component	During operation		After shutdown			
	1 mo	30 y	1 d	1 mo	50 y	500 y
LiPb	2 240 000	2 240 000	55 900	17 700	77	10
Screen	383 000	1 490 000	620 000	558 000	33	8
Bell jar	305 000	429 000	197 000	117 000	154	93
Lower bell jar	14 400	19 800	8 770	5 300	12	9
Beam port wall	4 730	8 280	4 230	2 890	2.0	0.96
Reflector	2 820	6 260	4 300	3 760	1.0	0.92
Steam generator ^a	42	51	17.5	8.7	0.03	0.02
Total ^b	2 950 000	4 190 000	890 000	705 000	277	122

^aActivation is for only one of the 12 steam generators and pumps in Pulse*Star; 10 vol% of the 0.5-m-radius steam generator was considered to be steel. The neutron flux statistics in this region were so sparse that the activation levels should be viewed as quite uncertain compared to those of the other components.

^bIncludes activation of all 12 pumps and steam generators.

low-activating materials are chosen for the target design; hence, only the reactor components are considered here. Activation of the reactor components modeled is shown in Table 7-10 at various times during and after operation. During operation, ^{209}Pb in the LiPb dominates the activity, but after shutdown, ^{55}Fe in the screen quickly takes over. After several decades, the ^{93}Mo in the thicker bell jar dominates. Total activity is reduced by a factor of 4.7 the first day after shutdown. In the next month, it is further reduced only by a factor of 1.3. By 50 y after shutdown, it is reduced by a factor of 15 000; the reduction from 50 to 500 y is only an additional factor of 2. These two plateaus—1 d to 1 mo and 50 to 500 y—could be quite important for two reasons. First, if the reactor has safety or maintenance problems 1 d after shutdown, waiting an additional month probably will not help. Second, if the components are not cheaply disposable 50 y after shutdown, waiting centuries also will not help. These reasons are further explored below, using the afterheat (decay heat after shutdown), inhalation bhp, surface-gamma dose rate, and shallow burial index for Pulse*Star as figures of merit. These figures of merit are

Table 7-11. Decay heat and surface dose rate for Pulse*Star components.

Component	Decay heat		Surface dose rate ^a		
	Shutdown ^b (kW)	1 d after shutdown (kW)	Shutdown ^b (R/h)	After shutdown	
				1 d (R/h)	1 mo (R/h)
LiPb	3 470	381	685	603	381
Screen	11 500	575	16 000 000	1 040 000	898 000
Bell jar	2 920	380	921 000	95 400	70 800
Lower bell jar	143	16	74 000	5 080	3 090
Beam port wall	53	6.8	1 630 000	183 000	142 000
Reflector	35	1.8	2 380	29	14
Steam generator	0.45	0.04	3 980	223	134
Total	18 100	1 360	—	—	—

^aNo totals are given for surface dose rate, because mutual shielding precludes simultaneous exposure to all components.

^bAfter 30 y.

Table 7-12. Inhalation bhp (km^3) for Pulse*Star components.

Component	Shutdown ^a	1 d after shutdown
LiPb	4.07×10^8	3.92×10^8
Screen	2.02×10^8	1.15×10^8
Bell jar	5.31×10^7	4.16×10^7
Lower bell jar	2.09×10^6	1.49×10^6
Beam port wall	1.05×10^6	8.45×10^5
Reflector	3.39×10^5	2.22×10^5
Steam generator	5.66×10^3	3.83×10^3
Total	6.66×10^8	5.51×10^8

^aAfter 30 y.

discussed in more detail in the last two articles of this section and in Refs. 67 and 68.

Afterheat is important shortly after shutdown; thus, the shutdown and 1-d results are most significant (Table 7-11). At shutdown, the LiPb accounts for 3.5 MW and the screen for 11.5 MW of the total 18.1 MW. One day later, the total has fallen to 1.4 MW, with the screen/bell jar (0.96 MW) still dominating. Several megawatts can be removed steadily by passive radiation or by a simple free-convection loop. The heat capacity of the LiPb allows several hours without heat removal before the submerged steel components lose their heat treatments.⁶⁹

Surface dose rate is important at times when maintenance could be performed; thus, the 1-d and 1-mo results are most significant (Table 7-11). Levels above ~ 100 R/h probably preclude hands-on maintenance; the bell jar, screen, steam generators, and pumps must therefore be maintained remotely. This conclusion could change for the pumps and steam generators for four reasons. First, the tube sheets (i.e., top and bottom plates) and pump impellers are at an elevation away from the midplane, and they will be activated much less than the height-averaged calculation indicates. Second, the statistics in this region were poor; hence, the spectral shape and collapsed cross sections are not trustworthy. Third, it would be easy to move these components out ~ 0.5 m without enlarging the pool, or even farther at little extra cost. Fourth, the 1-mo calculations are close to the hands-on maintenance limit of ~ 100 R/h for small radiation-source volumes.

The inhalation bhp is the volume of air required to dilute vaporized material to breathing standards. If the plume occupies some tens of cubic kilometers—the plume size beyond which evacuation would be difficult—then Table 7-12 shows that at 1 d after shutdown, only a small fraction of $\sim 10^{-7}$ (20 km^3 plume/ $2 \times 10^8 \text{ km}^3$ bhp) of either the screen or LiPb could be allowed to vaporize and be released. Of course, the release fraction depends on the accident scenario and on the major contributing nuclide, and thus this statement is overly simple.

We also calculated the shallow burial index for each reactor component (Table 7-13) by summing the ratios of the specific

Nuclide ^a	LiPb	Screen ^b	Bell jar	Lower bell jar	Beam port wall	Steam generator	Reflector	Total
¹⁴ C	9.3×10^{-14}	1.19×10^{-4}	6.43×10^{-5}	3.87×10^{-6}	1.42×10^{-4}	6.91×10^{-8}	1.16×10^{-3}	6.62×10^{-5}
⁵⁹ Ni	4.85×10^{-6}	0.0676	0.0338	5.42×10^{-3}	0.0655	1.28×10^{-4}	—	8.57×10^{-4}
⁶³ Ni	3.78×10^{-4}	7.73	1.70	0.126	3.01	4.11×10^{-3}	1.55×10^{-11}	0.0716
⁹⁰ Sr	6.03×10^{-12}	1.22×10^{-5}	8.28×10^{-11}	—	6.73×10^{-8}	—	—	8.28×10^{-8}
⁹⁴ Nb	0.133	103.1	47.0	17.7	62.2	0.855	5.29×10^{-6}	1.45
⁹⁹ Tc	1.23×10^{-9}	14.7	11.7	1.87	20.6	0.0848	0.0185	0.237
²¹⁰ Bism	5.6×10^{-3}	—	—	—	—	—	—	5.3×10^{-3}
Total	0.139	128	60.4	19.7	85.8	0.944	0.0197	1.76

^aOnly nuclides in 10CFR61 were included; using this list ignores 59% of the waste activity at 50 y after shutdown.

^bThe screen column includes only one of the six screens included in the total column.

activities (in Ci/m³) to the 10CFR61 limits.⁷⁰ To determine the totals for the entire reactor, we used the total activity and total material volume to compute the specific activity for each nuclide. Thus, the totals are much less than the arithmetic sum of the components. The ⁹⁴Nb from the 80-ppm ⁹³Nb impurity, and the molybdenum in the steel, dominate the results at 50 y after shutdown, the anticipated burial time. The 25-ppm ⁹³Nb impurity in lithium produces only 8% of the total ⁹⁴Nb, but the LiPb increases the volume by a factor of 50. The overall shallow burial index is 1.76, and it could be decreased to 1.0 (the legal limit for shallow burial) by adding 2700 m³ of grout to the volume. However, the addition to 10CFR61 of limits for other nuclides such as ⁹³Mo, ⁹³Nb^m, and ²⁰⁸Pb could push Pulse*Star further over the limit.

If the LiPb is recycled to the next reactor, the buried volume is only ~103 m³, and the undiluted shallow burial index is 56.5, 45.8 of which comes from ⁹⁴Nb. In Ref. 67, the author proposed an increase in the current 0.2-Ci/m³ limit for ⁹⁴Nb, because that limit was set by assuming no surviving barriers coupled with construction of homes in the (unrecognized) waste. If the proposed limit of 1100 Ci/m³ is used, the shallow burial index falls to 10.7, with ⁹⁹Tc accounting for ~76% and ⁶³Ni for ~23%. Then, only about 1000 m³ of grout would be required to satisfy the regulation.

Implications of the Activation Calculations. Afterheat due to the LiPb and steel is not a problem, because its low levels allow a reasonable reaction time. Waste disposal in shallow burial sites will also be feasible, although attention must be given to minimizing niobium and other impurities in both the steel and the LiPb coolant. The major remaining issue is the safety

implications of the activated lead in the coolant. The quantity of ²¹⁰Po produced is significant; we will evaluate its releasability and, if necessary, the feasibility of reducing ²¹⁰Po production by continuously removing the intermediate nuclide ²⁰⁹Pb from the LiPb.

Author: J. A. Blink

Tritium Containment

Tritium extraction through the Pulse*Star chamber will be primarily through the vacuum system, because tritium is almost insoluble in Li₁₇Pb₈₃. During 1983, we estimated the upper and lower limits on chamber pressure.⁷¹ The lower limit of ~10⁻³ Torr is influenced by design constraints on the vacuum system. The upper limit of 5 × 10⁻³ Torr is influenced by tritium-release limits from a spill and by tritium leakage into the steam generator water. These constraints mandate an upper pressure limit that is much lower than the ~10⁻¹ Torr set by laser beam propagation.

Lower Pressure Limit. Because the bell-jar nozzles in Pulse*Star finely divide the LiPb coolant each time it enters the vacuum chamber, we can reasonably assume equilibrium concentrations of dissolved hydrogen isotopes within the LiPb. These equilibrium concentrations ($C_{H-isotope}$) are a function of the square root of the hydrogen-isotope pressures within the chamber:

$$\sqrt{P_{H-isotope}} = K_S C_{H-isotope} \quad (29)$$

where K_S is Sievert's constant in units of the square root of pressure divided by the hydrogen-isotope concentration in atomic parts per million (appm).

Table 7-13. Shallow burial index for Pulse*Star components at 50 y after shutdown. The index cannot legally exceed 1.0 at burial.

The principal sources of gas pressure in the chamber are unburned DT mixtures from the target, helium from the fusion reaction, tritium and helium from the reaction of neutrons with lithium, and lead vapor pressure from the LiPb coolant. To avoid pumping an excessive quantity of lead vapor, we must maintain the chamber pressure at least 20 times above the LiPb vapor pressure of 1.5×10^{-5} Torr at 500°C . Hence, the chamber pressure should not be below 3×10^{-4} Torr. If a lower pressure is desired, a lower LiPb temperature must be used (LiPb vapor pressure drops⁷² an order of magnitude every 55°C).

There is, however, another practical limit on the lowest chamber pressure that can be maintained. Each 600-MJ fusion pulse consists of 21.3×10^{19} fusion reactions. As Table 7-14 shows, 5.16 molecules are injected per fusion (774.5 molecules/150 fusions). Thus, each 600-MJ pulse injects 0.011×10^{23} molecules or 0.0018 g/mole of gas, which corresponds to ~ 8200 liters at 10^{-3} Torr and 500°C . This volume must be pumped every 0.2 s to maintain steady-state

conditions. Geometric constraints on vacuum system designs indicate that this requirement is probably 10 times too severe and that the lower chamber pressure limit is $\sim 10^{-3}$ Torr.

Upper Pressure Limit. The tritium content of the liquid LiPb places an upper limit on pressure. This limit is found by restricting the tritium released from a LiPb spill that subsequently cools, and by considering tritium diffusion through the steam generator tubes into the water loop.

From Table 7-14, the ratio of hydrogen-isotope pressure to total pressure is 0.57. We can use Sievert's constant and the fact (shown in Table 7-14) that the ratio $T_2/(T_2 + D_2)$ is 0.6 to obtain the tritium dissolved in the liquid LiPb at 500°C . Then, knowing Sievert's constant as a function of temperature, we can solve for the tritium released during a spill with consequent cool-down or heat-up of the spilled liquid. However, we must first review the widely varying values for Sievert's constant and decide which values to use.

In 1980, we summarized⁷² the properties of LiPb solutions, including deuterium solubility data in LiPb alloys determined by Ihle.⁷³ Extrapolation of these data to $\text{Li}_{17}\text{Pb}_{83}$ was so variable that we were unable to predict whether tritium solubility in the liquid increased or decreased with temperature. To resolve the discrepancy, Chan and Veleckis carefully measured the data⁷⁴ from which Fig. 7-22 was constructed. These data indicate that the temperature dependence of Sievert's constant changes sign at $\text{Li}_{40}\text{Pb}_{60}$, and that Sievert's constant for $\text{Li}_{17}\text{Pb}_{83}$ increases as temperature decreases. Therefore, if $\text{Li}_{17}\text{Pb}_{83}$ spills and cools, tritium will be released into the gas space.

The Chan-Veleckis data are for protium (^1H) rather than for deuterium in LiPb, but we can convert the data to deuterium. For equal temperatures, kinetic energy is constant:

$$m_d v_d^2 = m_p v_p^2 \rightarrow \frac{v_d}{v_p} = \sqrt{\frac{m_p}{m_d}}, \quad (30)$$

where m is mass, v is velocity, and the subscripts denote deuterium (d), protium (p), and—in Eq. (32) below—tritium (t). Pressure

Table 7-14. History of 1000 fuel atoms in Pulse*Star.

	Preburn target atoms	Postburn target atoms ^a	Atoms created in LiPb ^b	Total molecules
Deuterium	500	350	—	175
Tritium	500	350	183	266.5
Helium		150	183	333
Total	1000	850	366	774.5

^a30% burnup.

^bTritium-breeding ratio = 1.22.

Fig. 7-22. Sievert's constant for $\text{Li}_{17}\text{Pb}_{83}$ as a function of temperature, constructed from the Chan-Veleckis semi-empirical model.⁷⁴

is proportional to momentum:

$$\frac{P_d}{P_p} \approx \frac{m_d v_d}{m_p v_p} = \sqrt{\frac{m_d}{m_p}} \quad (31)$$

Thus

$$\frac{P_d}{P_p} = \sqrt{2} \quad \text{and} \quad \frac{P_t}{P_p} = \sqrt{3} \quad (32)$$

Considerable uncertainty still exists, however, over hydrogen's solubility in $\text{Li}_{17}\text{Pb}_{83}$, as shown in Fig. 7-23. The Chan-Veleckis data can be extrapolated to pure lead, and the results do not correspond to either the Opie and Grant data⁷⁵ or to the ultra-high minimum values allowed by Maatsch⁷⁶ for Sievert's constant. Additional recent work on $\text{Li}_{17}\text{Pb}_{83}$ (Refs. 77 and 78) and on tritium⁷⁹ gives much lower values for Sievert's constant than those of Chan and Veleckis. Still, the Chan-Veleckis work gives conservative answers, and it was done so systematically that we used their data in our computations and recommend that others do the same.

In determining the upper pressure limit in the Pulse*Star reactor, we first considered LiPb spills. If the entire LiPb inventory spills into an outer containment building and cools, molecular tritium (T_2) will be released until an equilibrium is reached between gaseous tritium in the building and tritium dissolved in the spilled liquid. We considered only the case where cooling occurs. (If decay heating dominates the heat-removal process, much less tritium will be released.) The LiPb inventory in Pulse*Star is between 3400 and 4050 m^3 , depending on the design. The amount of tritium released into the containment building "ullage" (volume above the liquid) is a function of the amount of liquid spilled, the ullage volume, and the reactor vacuum level before the spill. Figure 7-24 shows the tritium released from a 4050- m^3 spill of LiPb at 500°C that cools to its freezing point. The worst case plotted (10^8 m^3 of ullage and 3×10^{-2} Torr of hydrogen-isotope pressure in the reactor vessel) results in over 65 g of tritium being released (>650 kCi). However, a spill of the entire LiPb volume is highly unlikely, because draining it into sealed storage tanks is a quick operation. For smaller spills,

Fig. 7-23. Variation in calculations of Sievert's constants for $\text{Li}_{17}\text{Pb}_{83}$ and pure lead.

Fig. 7-24. Tritium released from a 4050- m^3 LiPb spill as a function of ullage volume and reactor vacuum pressure.

the curves have the same general shape, and the tritium-release fractions are quite reasonable.

Because the amount of LiPb spilled may be a difficult parameter to control, and because the ullage may be more than 10^8 m^3 if the outer containment building is breached, the reactor vacuum pressure should be set with tritium-release limits in mind. The tritium-release curves indicate that this pressure should be $\sim 5 \times 10^{-3}$ Torr or lower. Thus, the probable pressure range for the Pulse*Star reactor is between 1 and 5×10^{-3} Torr. The most tritium that

could be released from a cooled spill of the total LiPb inventory would then be between 12 and 27 g.

Under steady-state conditions, 50 g/d of tritium diffuses into the steam generator water from the LiPb coolant.⁸⁰ This calculation is based on 4700 m² of 4.75-mm-thick ferritic steel tubes, no tritium barrier, a 50°C thermal gradient, and an upstream tritium pressure of 1.8×10^{-3} Torr. If the steam generator tubes are made from two 2.375-mm-thick concentric tubes with the entire thermal gradient in an unpumped gap between the two walls, the results do not change appreciably.⁸⁰ The 50 g/d diffusion level represents about 3% of the tritium injected into the LiPb daily.

Pulse*Star will recover and recycle tritium from the gas removed from the reactor chamber by the vacuum system. The tritium that diffuses through the steam generator walls may be recoverable from the water or by using flowing He + O₂ in the gap between the walls. If this tritium is recovered, it may be held until it decays, or it may be recycled into the target factory.

Summary. Vacuum system constraints place a lower limit on the Pulse*Star chamber pressure of $\sim 10^{-3}$ Torr. Higher pressures ease vacuum design constraints at the expense of increasing both tritium diffusion into the steam generator water and the tritium released during an accidental LiPb spill. At a chamber pressure of 5×10^{-3} Torr, about 3% of the tritium flow is lost to the water loop and about 20 g of tritium could be released from a spill of the entire LiPb inventory.

Although tritium diffusion through the steam generators does not prevent tritium self-sufficiency, it is four to five orders of magnitude above the allowable site-release levels. Hence, tritium diffusion must be drastically decreased by barriers, or water clean-up systems must be considered; both alternatives are being studied.

Our study places an upper bound of 20 g on the tritium released during a spill, but design features such as low ullage volumes and the quick recovery of spilled LiPb could significantly reduce the tritium released; such design features are being studied. In addition, the impact of large tritium releases has been studied by others, and

we are integrating their results into our design procedure.

Authors: J. A. Blink (LLNL) and N. J. Hoffman (Energy Technology Engineering Center)

Heat-Transfer Loop and Plant Layout

The hot- and cold-leg temperatures of the Li₁₇Pb₈₃ coolant in Pulse*Star have been determined by material-strength criteria. The pumps and piping are 2-1/4 Cr-1 Mo steel, which has an upper temperature limit of 500°C and a lower limit of 400°C. The lower limit is dictated by the danger of radiation-induced embrittlement in the steel below 350°C. To produce 3300 MW, from a 100°C temperature difference, the flow rate through the heat exchanger system must be 22 m³/s. Dividing the flow into four parallel heat-transfer loops results in a flow rate of 5.5 m³/s through each loop. Four LiPb properties are required to size the pumps and steam generators for Pulse*Star: the heat capacity (0.061 kJ/kg·K) and viscosity (0.00209 kg/m·s) are averaged values for lithium and lead; the density (9400 kg/m³) and thermal conductivity (11.6 W/m·K) are measured values.⁸¹

The feedwater temperature for the steam generators is 230°C, and the exiting steam temperature is 460°C at 15.8 MPa. These are design conditions applicable to a commercial fusion power plant with liquid-metal-heated steam generators. The electrical conversion efficiency is $\sim 39\%$, the optimum conversion efficiency of an LMFBF plant.

Pump Selection. A pump is required in each of the four heat-transfer loops to force the liquid LiPb through the steam generators and to maintain the pool level above the evacuated reaction chamber. The high density and low heat capacity of LiPb make the pumping power a significant portion of the power consumed within the plant. This emphasizes the importance of pump design. The pump characteristics and dimensions for Pulse*Star are summarized in Table 7-15. Of the total developed head of 14 m, 12.5 m is due to the height of the

Table 7-15. LiPb pump characteristics and dimensions for Pulse*Star.

Parameter	Value
Number of pumps	4
Flow per pump (m ³ /s)	5.5
Developed head (m)	14
Developed pressure (MPa)	1.3
NPSH available (m)	4
NPSH required (m)	3.4
Impeller diameter (m)	1.2
Pump speed (rpm)	278
Net fluid power per pump (MW)	7.1
Electrical power per pump (MW _e)	8.9
Pump heat absorbed by liquid (MW _l)	8.4

pool surface over the surface of the catch basin below the reaction chamber. The remaining head is due to the pressure drop through the steam generators.

We chose a mechanical rather than an electromagnetic pump because the latter's efficiency would be adversely affected by the high electrical resistivity of LiPb, and because a mechanical pump avoided cooling for immersed electrical coils. A mechanical pump is expected to operate at an 80% efficiency. The drive system will be above the pool with power transmitted down a shaft to the impeller.

Because the liquid-metal pumps have unusually low head and high flow requirements, we chose a mixed-flow pump. A mixed-flow impeller has flow behavior between axial and radial designs. Work on low net-positive suction head (NPSH) pumps for liquid-metal applications has generated a guideline that the NPSH available should be greater than three times the dynamic head at the impeller entrance. With a flow velocity of 4.7 m/s (based on a 1.2-m-diam pipe), an NPSH of 3.4 m at the pump inlet is needed.

To assure an adequate design margin to prevent pump cavitation, we designed an NPSH of 4 m into the plant. A suction-specific speed of 12 000 was used to calculate a pump speed of 278 rpm. A mixed-flow impeller should have a specific speed of about 5000 to 8000 (English units).⁸² To produce the required head within this range, two impellers are staged in series, with the pump shaft passing down the center of the discharge pipe.

The required electrical pumping power is calculated to be 8.9 MW_e for each pump, based on an 80% efficiency. The total pumping power for the LiPb recirculation system in Pulse*Star is 36 MW_e, which is less than 3% of the net plant output.

Steam Generator Design. The major requirement for the steam generators is that they be located within the coolant pool, close to the reaction chamber. Their design can be optimized by finding the most cost-effective combination of pumping power and heat-exchanger area. To keep pumping power low, the steam generator tubes must be widely spaced. This reduces the shell-side heat-transfer coefficient, resulting in an

increased tube surface area. Close spacing would decrease the heat transfer area and increase the pumping power. Our objective was to minimize the initial cost of the steam generators and the total cost of pumping power during the plant lifetime.

We investigated two configurations for the steam generators: horizontal crossflow steam generators located above the reactor at the top of the pool, and vertical counterflow steam generators located on the periphery of the pool. The tube design for both configurations is a facsimile of the double-wall tube in the Westinghouse breeder-reactor steam generator. The tubes have an outer diameter of 20.6 mm and an inner diameter of 11.1 mm. Helium is the third fluid, located in grooves between the inner and outer tube walls.

In the crossflow design, four two-pass, U-tube steam generators are located at the top of the pool. The LiPb from the pump discharge flows down across the tubes and into the reaction chamber (Fig. 7-25). The shell-side heat-transfer coefficient, h_s , for liquid-metal crossflow through staggered tube banks is⁸³

$$h_s = \frac{k}{D} \sqrt{\left(\frac{\phi_1}{D}\right) \left(1 - \frac{D}{P}\right)} \times (6.19 + 0.2665 \text{Pe}^{0.653}) \quad (33)$$

where k is the thermal conductivity of LiPb, D is the tube diameter, ϕ_1 is the unit hydrodynamic potential at the rear stagnation point of a rod (as plotted in Re. 54 and 85), P is the tube pitch (tube center-to-center distance), and Pe is the Peclet number ($\text{Pe} = D\nu\rho C_p/k$). Varying the tube pitch varies the Peclet number and D/P , and thus h_s .

Fig. 7-25. Top view of Pulse*Star reactor with crossflow steam generators.

Fig. 7-26. $1/P/D$ vs steam-generator capital and present-value pump operating costs.

The overall heat-transfer coefficient is a combination of shell- and tube-side coefficients and the tube-wall resistivity. The tube-side and wall resistivities for the double-wall tubes were measured by Westinghouse as 7.1×10^{-5} and 1.8×10^{-4} $\text{m}^2 \cdot \text{K}/\text{W}$, respectively.⁸⁶ To transfer 3300 MW_t of heat, 16 600, 16 950, and 17 160 m^2 of surface area are required for P/D ratios of 1.2, 1.5, and 2.0, respectively. The

pressure loss is more sensitive to P/D , with values of 1.65, 0.32, and 0.15 m for the same ratios.

The pressure drop through the shell side of the steam generator is

$$\Delta P = \frac{2fN_R v^2}{g} \quad (34)$$

where ΔP is the head (m), N_R is the number of tube rows, v is the coolant velocity (m/s), g is gravitational acceleration, and f is the friction factor given by⁸⁷

$$f = \left[0.23 + \frac{0.11}{\left(\frac{P}{D} - 1 \right)^{1.08}} \right] \text{Re}^{-0.15} \quad (35)$$

where Re is the Reynolds number.

The steam-generator and pumping costs as functions of P/D are plotted in Fig. 7-26. A P/D of 1.5 gives the lowest total cost.

In the countercurrent design, eight vertical steam generators are located between the reactor shell and the pool wall.

Fig. 7-27. Compact layout for Pulse*Star power plant.

Table 7-16. Comparison of crossflow and counterflow steam-generator designs.

	Crossflow	Counterflow
Number of steam generators	4	8
Total heat exchanger area (m ²)	16 960	14 720
LiPb pool height (m)	20.5	17.5
LiPb pool diameter (m)	17	17
Total required pumping power (MW)	40.1	35.5
Pressure drop through 1 steam generator (m)	0.3	1.5
Steam generator capital and operating cost (\$million)	103.1	94.6
LiPb volume in pool (m ³)	4 090	3 400

The LiPb is pumped to the top of the pool, where the flow from each pump is split and travels down through two heat exchangers. The feedwater is pumped to the bottom of the heat exchanger through a central conduit and flows upward. Differential expansion of the tubes is accounted for by attaching a slip fitting to the bottom tubesheet.

The heat-transfer coefficient for in-line flow of liquid metals through tube banks is⁸³

$$h_s = \left(\frac{k}{D_e} \right) \left[6.66 + 3.126 \left(\frac{P}{D} \right) + 1.184 \left(\frac{P}{D} \right)^2 + 0.0155 (\psi Pr)^{0.86} \right], \quad (36)$$

where D_e is the equivalent shell-side diameter (m); ψ is a diffusivity ratio given by⁸³

$$\psi = 1 - \frac{1.82}{\left(\frac{\epsilon_m}{\nu} \right)^{1.4} Pr}, \quad (37)$$

where Pr is the Prandtl number ($Pr = C_p \mu / k$), and (ϵ_m / ν) is a function of Re , as plotted in Refs. 84 and 85.

The pressure drop through the shell side of an unbauffed, countercurrent heat

exchange is

$$\Delta P = \left(\frac{0.315}{Re^{0.25}} \right) \left(\frac{L}{D_e} \right) \left(\frac{v^2}{2g} \right), \quad (38)$$

where L is the tube length. This correlation uses a standard friction factor for turbulent flow in a pipe.

The heat exchanger area is more sensitive to P/D in the counterflow steam generator, with values of 13 680, 14 720, and 17 200 m² for P/D ratios of 1.2, 1.5, and 2.0. The pressure loss is also more sensitive, with values of 27.1, 1.48, and 0.09 m for the same P/D ratios. The steam-generator and pumping costs for the counterflow design are also plotted in Fig. 7-26. A P/D of 1.5 again results in the lowest total cost. This corresponds to a steam generator diameter of 1.6 m, which easily fits between the reactor bell jar and the pool wall [see Fig. 7-20(b) in "Pulse*Star Reactor"].

Table 7-16 compares optimized designs for the crossflow and counterflow steam

generators. The crossflow design requires 15% more heat transfer area, mainly because of the 0.87 correction factor for the log-mean temperature difference of the heat exchangers.⁸⁸ Because of their horizontal placement above the reactor, the crossflow steam generators also require a 3-m greater pool depth and a slightly greater pumping power due to their higher static head. The capital and operating costs for the crossflow design are ~9% higher than for the counterflow design. Additional savings for the counterflow design can be realized by reducing pool depth and LiPb volume. Thus we have chosen counterflow steam generators for Pulse*Star.

Plant Layout. A schematic of one possible power plant layout is shown in Fig. 7-27. The layout is designed to provide an efficient, compact arrangement of components around the liquid-metal pool. The LiPb pool and drain tanks are recessed below grade for structural support and radiological shielding. The laser beam ports are also below grade, to be in line with the center of the reaction chamber. To avoid extensive excavation, all other plant areas are above ground. The turbine generator building is next to the reactor to minimize steam runs. The control room is adjacent to the reactor and the turbine generator to minimize the instrumentation and control wiring

between the major plant systems and buildings. A large maintenance area is adjacent to the reactor pool to facilitate removal and service of the pool components. Diesel generators that provide auxiliary power are located close to the control room to reduce cable routing. The target factory and vacuum pumping areas are located near the laser beam ports in the reactor.

We compared the size of the Pulse*Star plant with that of a 1000-MW_e LMFBR plant.⁸⁹ The Pulse*Star plant occupies a floor space of ~12 m²/MW_e; the figure is ~18 m²/MW_e for an LMFBR plant. Pulse*Star's smaller size results from a more compact reactor building and the elimination of separate steam-generator buildings. By having three reactors share the laser, target factory, and maintenance area, we can reduce the size of the Pulse*Star plant to 8 m²/MW_e.

A plant power balance for Pulse*Star is shown in Fig. 7-28. The 4.3-Hz pulse rate was selected to match the initially specified power of 3300 MW_t. A 5-Hz pulse rate will produce higher thermal and electrical outputs. Except for the laser system and LiPb primary pumps, the in-plant recirculated power requirements are based on the power requirements of an LMFBR plant, scaled by the plant's gross thermal output. The Pulse*Star system efficiency (net electric

Fig. 7-28. Pulse*Star power balance.

power/fusion power $\times$ system multiplier) is 34%, which is 110% of the HYLIFE efficiency and is comparable to that of a fission reactor.

Conclusions. The pool concept for an ICF reactor appears attractive. We have developed a compact configuration for steam generators and LiPb recirculation pumps that fits well around the periphery of the Pulse*Star reaction chamber. Conceptual designs of the pumps and steam generators are within the range of existing technology. The recirculated power is shared equally by the laser system and the balance of plant. The recirculating power is reasonable, the plant net efficiency is higher than in previous designs, and the net output exceeds 1000 MW_e. The high density of the LiPb fluid does not increase pumping power to unacceptable levels.

Authors: K. A. Murray and M. W. McDowell (Energy Technology Engineering Center), and J. A. Blink (LLNL)

Neutronics Studies of a Suppressed-Fission ICF Hybrid

Fusion-fission hybrid reactors have been classified into two categories: fast-fission and suppressed-fission.⁹⁰⁻⁹³ In a fast-fission hybrid blanket, the DT fusion plasma is surrounded by a region of ²³⁸U or ²³²Th. The 14-MeV fusion neutrons cause fast-fission reactions in these materials and thereby increase the number of neutrons available for breeding fissile fuel and tritium.

Suppressed-fission hybrids depend on nonfissioning neutron multipliers such as lead, beryllium, or ⁷Li to increase the number of neutrons available for breeding. [⁷Li can be considered a neutron multiplier here because the ⁷Li(n,n'⁺T) α reaction gives one of the desired products, tritium, and yet does not deplete the number of available neutrons.] Two basic configurations have been proposed for suppressed-fission blankets: a two-zone blanket and a one-zone, homogeneous blanket. In the two-zone blanket, neutron multiplication and tritium breeding occur in the first or

inner zone, while fissile breeding occurs in the second or outer zone. The first zone also moderates the neutrons, thus minimizing fast fission in the fertile zone. In the one-zone blanket, fast fissioning is suppressed by fertile dilution; i.e., the concentration of fertile material is low (<5 vol%). The majority of the blanket consists of neutron-multiplying, neutron-moderating, and tritium-breeding materials.

During 1983, we studied a new concept for a one-zone, suppressed-fission hybrid.⁹⁴ The breeding blanket consists of a beryllium neutron multiplier, metallic-thorium fertile fuel, and a liquid-lithium coolant. The concept is to suppress fission in the bred fissile fuel through competition for neutrons with ⁶Li tritium-breeding reactions. We found that the total fission rate in the proposed beryllium-lithium-thorium (Be/Li/Th) blanket (with 0.5 at.% ²³³U) goes through a maximum at ~5 vol% thorium. At ~30 vol% thorium, the fast fission rate in thorium is nearly as great as the fission rate in the bred ²³³U, but the total fission rate is less than in a blanket with only 2 vol% thorium.

We also found that there are significant resonance effects in this type of blanket. We compared neutronics results from TART⁹⁵ with those from ALICE.^{96,97} Both are Monte Carlo codes using the LLNL Evaluated-Nuclear-Data Library (ENDL)⁹⁸ of cross sections. ALICE, however, uses a multiband method that gives a better treatment of resonance self-shielding effects. For the Be/Li/Th blanket investigated, ALICE predicts a lower fissile-breeding rate, a higher parasitic capture rate, and a higher fission rate in ²³³U than TART. All these factors reduce the suppressed-fission hybrid performance. These findings are consistent with those of Taczanowski.⁹⁹

Chamber Concept

The hybrid chamber is shown in Figs. 7-29 and 7-30. The fusion target is surrounded by an array of beryllium columns. These 8-m-high columns are made up of individual blocks, 67 cm high and 40 cm in diameter, held together with steel fastening bands. A dense array of coolant channels is machined into each block. The lithium coolant is supplied at the top of the

Fig. 7-29. An ICF hybrid design that uses a beryllium multiplier, metallic-thorium fertile fuel, and a liquid-lithium coolant.

chamber, traverses through the beryllium columns, and exits at the bottom of the chamber. Each channel can accommodate a steel-clad, metallic-thorium fuel pin. The columns are attached to top and bottom plates so as to tolerate radiation-induced swelling and the vibrations resulting from each microexplosion.

A thin (~ 10 cm thick) lithium-fall region protects the beryllium columns from direct exposure to the x rays and debris emitted by the target. While shown as a continuous curtain, the protective fall could be composed of two rows of small-diameter jets.

The fusion energy yield per pulse is 400 MJ, and the pulse repetition rate is 2 Hz, giving a fusion power of 800 MW. This chamber design is adapted from a

concept developed by TRW in conjunction with LLNL for a pure fusion reactor.

One-Dimensional Neutronics

We used a spherical model of the blanket for our neutronics scoping studies and ran calculations on ALICE, the multiband version of TART. The model consisted of a 14.1-MeV neutron source distributed in a region of compressed DT ($\rho r = 3 \text{ g/cm}^2$), a 10-cm-thick lithium region at a radius of 1.9 m, and the breeding blanket between radii of 2 to 4 m. The breeding blanket was at 50% normal density to account for the fact that the array of beryllium columns has a 50% packing fraction. Consistent with

Fig. 7-30. Cross section of the hybrid reactor.

cooling and structural requirements, the breeding-blanket composition (by volume fraction) was 20% lithium, 2% iron, and 78% beryllium plus thorium.

We investigated blanket performance as a function of the thorium volume fraction. As this fraction was increased, the beryllium fraction was decreased to maintain the total of the two at 78 vol%. The ^6Li concentration in lithium was adjusted to achieve a tritium-breeding ratio of between 1.0 and 1.1. In this way one can obtain the maximum fissile-breeding rate, while allowing the fusion reactor to be self-sufficient in tritium. In all cases the thorium contained 0.5 at. % ^{233}U to represent an average fissile enrichment, assuming a discharge enrichment of $\sim 1\%$.

Figure 7-31 shows the total breeding ratio, $T + F$, and the fusion energy multiplication factor, M_f , as a function of the thorium volume percent. The tritium-breeding ratio, T , is the sum of $^6\text{Li}(n,T)\alpha$ and $^7\text{Li}(n,n'T)\alpha$ reactions per DT neutron. The net fissile-breeding ratio, F , is

given by

$$F = N[^{232}\text{Th}(n,\gamma)] - N[^{233}\text{U}(n,f)] - N[^{233}\text{U}(n,\gamma)] \quad (39)$$

where N is the number of reactions per DT neutron. The fusion energy multiplication factor is

$$M_f = \frac{E_n + 3.5}{17.6} \quad (40)$$

Fig. 7-31. Total breeding levels off and energy multiplication decreases beyond 5 vol% thorium.

Neutronics Studies of a Suppressed-Fission ICF Hybrid

Table 7-17. ${}^6\text{Li}$ enrichment and ratio of ${}^{233}\text{U}$ to ${}^6\text{Li}$ as a function of the thorium volume percent.

Thorium (vol%)	${}^6\text{Li}/\text{Li}$ (at.%)	${}^{233}\text{U}/{}^6\text{Li}$
2	0.08	0.44
6	0.20	0.53
13	1.20	0.19
20	3.00	0.12
30	6.00	0.07

Fig. 7-32. Increasing the thorium fraction reduces parasitic capture in iron and fission in ${}^{233}\text{U}$.

where E_n is the total neutron-induced energy deposited per DT neutron. Note that E_n includes the neutron energy deposited in the compressed-DT target and the energy resulting from fission reactions in the blanket.

As shown in Fig. 7-31, the breeding ratio increases rapidly with increasing thorium content and then levels off beyond ~5 vol% thorium. The fusion energy multiplication factor peaks at ~5 vol% thorium and then falls, rapidly at first, then gradually, as the thorium content increases.

The peak in the energy multiplication is related to the fission-rate variation in the breeding blanket. Figure 7-32 shows the number of ${}^{232}\text{Th}(n,f)$, ${}^{233}\text{U}(n,f)$, total fission, and $\text{Fe}(n,\gamma)$ reactions per DT neutron. The ${}^{232}\text{Th}(n,f)$ rate increases linearly with increasing thorium content as expected. The ${}^{233}\text{U}(n,f)$ rate initially increases very rapidly but then falls beyond ~5 vol% thorium. The total fission rate also goes through a maximum, decreases, and then

begins to level out at ~30 vol% thorium, where the number of thorium fast-fission reactions is approaching the number of ${}^{233}\text{U}(n,f)$ reactions.

The reason for this drop in the ${}^{233}\text{U}$ fission rate can be understood if we examine how the ${}^6\text{Li}$ content varies with thorium volume percent. Recall that the ${}^6\text{Li}$ concentration was varied to give a tritium-breeding ratio of ~1. Table 7-17 gives the atomic percent of ${}^6\text{Li}$ in lithium and also the ratio of ${}^{233}\text{U}$ atoms to ${}^6\text{Li}$ atoms in the blanket. The ${}^6\text{Li}$ fraction increases monotonically with thorium volume percent, but the ratio ${}^{233}\text{U}/{}^6\text{Li}$ reaches a peak at ~5 vol% thorium. Therefore, the competition between neutron absorption in the ${}^6\text{Li}(n,T)\alpha$ and ${}^{233}\text{U}(n,f)$ reactions governs the noted behavior of the total fission rate and M_f . Also note that the number of parasitic capture reactions in iron drops off rapidly as the thorium volume percent increases. This is another benefit of higher ${}^6\text{Li}$ concentrations.

Previous one-zone, suppressed-fission blanket designs had low thorium fertile-fuel contents, thus operating to the left of the peak in M_f (or in total fissions). In this study, we propose a blanket that has ~30 vol% thorium and operates to the right of this peak.

The fissile production rate per unit of thermal power is sometimes used as a figure of merit for the hybrid, because it is proportional to the thermal support ratio (i.e., the fission power fueled by the hybrid divided by the hybrid power). For our spherical model, the specific fissile production rate increases from about $1.1 \text{ g}({}^{233}\text{U})/\text{kW}_t\cdot\text{y}$ at 2 vol% thorium to about $2.0 \text{ g}({}^{233}\text{U})/\text{kW}_t\cdot\text{y}$ at 30 vol% thorium.

Two-Dimensional Neutronics

Our next step was to run calculations on a two-dimensional model of the hybrid chamber. This cylindrical model has a 0.1-m-thick lithium fall at $r = 1.9 \text{ m}$, followed by a 1.3-m-thick breeding blanket, an 0.3-m-thick graphite reflector at $r = 4.0 \text{ m}$, and then a 5-cm-thick steel vessel wall. Above the target DT (or = 3 g/cm^2) is a 2-cm-thick steel top plate, a 1-m-thick lithium zone representing the

Fig. 7-33. One scheme to make the fissile enrichment more uniform across the blanket is to include rotating arrays of beryllium columns.

inlet plenum, and the 5-cm-thick steel vessel wall. Below the target is a 1-m-thick lithium outlet plenum and the steel vessel wall.

The breeding blanket is 20 vol% lithium (5 at.% ^6Li), 48 vol% beryllium, 2 vol% iron, and 30 vol% thorium (0.5 at.% ^{233}U). The 1.3-m-thick breeding blanket is divided into six zones to reveal how the neutronic performance varies as a function of depth into the blanket. The inner two zones are each 13 cm thick, while the outer four are each 26 cm thick. All breeding-blanket zones are at 50% normal density to account for the 50% packing fraction of the beryllium.

The tritium-breeding ratio is 1.08 and the net fissile-breeding ratio is 0.97, for a total breeding ratio of 2.05. The total number of fissions per DT neutron is 0.122, with thorium contributing 0.046 and uranium giving 0.076. The neutron energy deposited is 41.1 MeV, which gives a fusion energy multiplication factor of 2.53. About 85% of the fissile breeding occurs in the inner half of the blanket. As a result, the enrichment rate at the blanket's inner edge is a factor of 3 higher than the average for the blanket.

The heavy-metal inventory of this blanket is ~ 300 t. With a fissile-breeding ratio of ~ 1 and a fusion power of 800 MW, the discharge enrichment after one full power year is 1.0%. After the decay of all the ^{233}Pa , the final enrichment is 1.13%.

A fuel management scheme may be desirable to level out the fissile concentration across the breeding blanket. One scheme,

which would eliminate the need for removing fuel from the reactor and repositioning it, is to rotate the fuel continuously from the inside to the outside of the blanket, as shown in Fig. 7-33. A group of seven beryllium columns forms a single rotating array. There are 10 such arrays. The seven columns are attached to disks at the top and bottom of the chamber. These disks rotate slowly, bringing the outer blocks toward the inside and vice versa. The regions between the rotating disks have columns in fixed positions.

Two additional calculations were carried out to determine how the neutronic performance varies as the fissile content is increased from zero to the discharge enrichment of 1%. We find that the tritium- and fissile-breeding ratios are essentially constant over the life cycle. The total fission rate increases from 0.045 to 0.20 as the ^{233}U enrichment increases from 0 to 1%. As a result, the fusion energy multiplication factor increases from 1.7 to 3.3. If the fusion power were held constant at 800 MW, the thermal power would increase from 1360 to 2640 MW_t over the one-year enrichment period.

Resonance Self-Shielding

As previously noted, we found significant differences between results obtained with TART and ALICE. ALICE gives more

Table 7-18. ALICE and TART results for our two-dimensional model.

General Studies and Reactor Development Technologies

Parameter	ALICE	TART
Reactions/DT neutron		
T_6^a	0.812	0.526
T_7^b	0.268	0.275
T	1.080	0.801
Th (n, γ)	1.056	1.285
U (n, γ)	0.011	0.038
Fe (n, γ)	0.009	0.004
Th (n,f)	0.046	0.046
U (n,f)	0.076	0.034
Fissions	0.122	0.080
E_n (MeV)	41.1	33.2
M_f	2.53	2.09

^a ^6Li (n,T) α reactions.

^b ^7Li (n,n' γ) α reactions.

accurate results when resonance absorption is a feature of the cross sections. This is the case for the $^{232}\text{Th}(n,\gamma)$, $^{233}\text{U}(n,f)$, $^{233}\text{U}(n,\gamma)$, and $\text{Fe}(n,\gamma)$ reactions.

Table 7-18 compares the results of TART and ALICE for our two-dimensional model. TART gives lower values for $^6\text{Li}(n,T)\alpha$, $\text{Fe}(n,\gamma)$, and $\text{U}(n,f)$ and higher values for $\text{Th}(n,\gamma)$ and $\text{U}(n,\gamma)$. As a result, TART predicts a tritium-breeding ratio of less than one and a significantly smaller neutron energy deposition. To get this ratio greater than one with TART would require a higher ^6Li concentration. This would further reduce the $\text{U}(n,f)$ rate and give an even lower energy multiplication factor.

From the standpoint of a suppressed-fission hybrid, results obtained with TART make the Be/Li/Th blanket concept look more attractive than it really is (i.e., TART underestimates the fission rate). However, the degree of discrepancy with other hybrid blanket concepts cannot be predicted a priori.

Summary

We have studied an ICF hybrid that uses a beryllium-neutron multiplier, metallic-thorium fuel, and a lithium coolant. We found that increasing the thorium content to 30 vol% actually reduces the energy multiplication factor. The ^6Li enrichment required to achieve a tritium-breeding ratio of 1 increases with the increasing fertile-fuel fraction. At high enrichments, the ^6Li effectively competes for neutrons and reduces both parasitic capture in iron and fission in the bred fissile fuel. The large heavy-metal inventory also permits a fuel residence time on the order of one year for the hybrid. We

also found that resonance self-shielding effects are important for this type of blanket, and we recommend that ALICE be used instead of TART for hybrid calculations.

Author: W. R. Meier

Major Contributors: D. H. Berwald, T. C. McCarville, and J. D. Gordon (TRW)

General Studies and Reactor Development Technologies

Cost Goals for ICF Reactors

One of the primary objectives of our work on new ICF reactor (ICFR) concepts is to develop one that will be economically competitive as an electric power producer. During the past year, we evaluated the allowable cost for an economically competitive ICFR and investigated the sensitivity of this cost to design and system parameters.

Definitions. We use the term *reactor* to describe the fusion chamber and the associated balance of plant required for heat transfer and thermal-to-electric power conversion. Each reactor produces $\sim 1.1 \text{ GW}_e$. *Power plant* refers to the entire system, including the reactor, driver, and target factory.

A single driver may be used to run more than one reactor by switching the beams from one chamber to another. In this way an ICFR power plant may produce several gigawatts. Similarly, a single target factory could produce targets for several reactors. Sharing these facilities lowers their effective capital costs (i.e., the cost per unit of electrical power produced).

For our cost comparisons, we chose the PWR as the base power plant. The cost of electricity (COE) from a PWR power plant has three components: capital, operation and maintenance, and fuel cycle. The relative proportions of these components are 0.69, 0.13, and 0.18, respectively.¹⁰⁰ The COE from an ICFR power plant also has three components: driver-plus-reactor capital, operation and maintenance, and targets. We define *economically competitive* to mean that the COE due to the capital cost of the ICF reactor and driver equals the COE due to the capital cost of a PWR. Under this

definition, the COE from the two plants would be equal if the cost of producing targets to fuel the ICFR plus its operation and maintenance costs were as high as the cost of the PWR fuel cycle plus operation and maintenance costs. While these numbers are rather uncertain for an ICFR, we anticipate they will be lower than the comparable PWR costs. In this sense, our definition of economic competitiveness is conservative.

We define *allowable reactor cost* as the difference between the reactor-plus-driver cost and the effective driver cost. Thus the allowable reactor cost increases when the driver is shared by several reactors. We assume that the direct capital cost of the driver is \$200 million at 1 MJ and that it scales as $E_d^{0.8}$, where E_d is the driver energy.

Important Parameters. The COE due to the reactor-plus-driver capital cost is proportional to capital cost divided by the net electric power produced. That is,

$$\text{COE} \sim \frac{C_r + C_d/N}{(P_f M_f \eta_p - P_d/\eta_d G) \alpha} \quad (41)$$

where

C_r = reactor capital cost,

C_d = driver capital cost,

N = number of reactors driven by a single driver,

P_f = fusion power of the reactor,

M_f = fusion energy multiplication factor,

η_p = plant efficiency (fraction of thermal power converted to electric power and available for sale or for operating the driver),

η_d = driver efficiency (ratio of beam energy on target to electric energy input to driver),

G = target gain, and

α = plant availability factor.

The second term in the denominator is the electric power consumed by the driver for each reactor it drives.

Equation (41) shows that there are several ways to reduce the capital portion of the COE. The cost of the reactor could be reduced by making it safer, simpler, and more compact. The effective cost of the driver could be reduced by innovation and/or by sharing the driver between several reactors. The driver power requirements could be reduced by increasing driver efficiency and/or increasing target

gain. We would like to increase the fusion-to-electric power conversion by increasing the energy multiplication factor and/or by increasing plant efficiency. Finally, we should seek to increase the plant availability by keeping the design simple and building in redundancy of critical components. Unfortunately, improvements in one area may result in penalties in another. For example, increasing the plant efficiency by designing for high-temperature operation or by utilizing direct conversion will probably result in a more expensive reactor.

In our study, we varied the following parameters to determine their impact on ICFR costs:

1. Number of reactors sharing a single driver: $N = 1, 2, \text{ or } 4$.
2. Plant efficiency: $\eta_p = 33.4, 40, \text{ or } 50\%$.
3. Driver efficiency (assumed to be independent of driver energy and pulse repetition rate): $\eta_d = 5, 10, \text{ or } 20\%$.
4. Target gain: $G = G_0/2, G_0, \text{ or } 2G_0$, where the reference gain, G_0 , is $185E_d^{0.7}$.

The selected reference target-gain relation is midway between the published conservative and optimistic gain bands.¹⁰¹ The reference gain curve passes through the point $G = 300$ at $E_d = 2$ MJ. This point has been suggested as a reasonable goal in the development of high-gain targets.¹⁰²

The relevant characteristics of the PWR to which we make our comparison are as follows¹⁰⁰:

- Thermal power, 3412 MW_t.
- Electrical power, 1139 MW_e.
- Plant efficiency, 33.4%.
- Direct capital cost, \$903 million.
- Direct cost per kW_e, \$793.
- Direct cost per kW_t, \$265.

These and all other costs are in 1982 dollars. We assume that the electric power produced by the ICFR is the same as that of the PWR (i.e., $P_f M_f \eta_p = 1139$ MW_e). A typical value of 1.15 was used for M_f . The net power for the ICFR, however, is less because of the electric power recirculated to the driver. Thus the capital cost for the ICF reactor plus driver must also be less, if the COE due to capital cost is to be the same as for the PWR.

Results. From the standpoint of chamber design, the chamber pulse repetition rate (or rep-rate) is an important variable. Figure 7-34 shows the allowable reactor cost as a function of chamber rep-rate for one, two, and four reactors with a single driver. Note

that the chamber rep-rate, R , is related to the driver energy E_d by

$$R = \frac{P_f}{E_d G(E_d)} \quad (42)$$

The fusion power is fixed for a given value of η_p :

$$P_f = \frac{1139}{M_f \eta_p} \quad (43)$$

Fig. 7-34. Sharing the driver and increasing the chamber rep-rate increases the allowable reactor cost.

Fig. 7-35. Allowable reactor cost as a function of chamber rep-rate and the number of reactors sharing a driver.

Fig. 7-36. Allowable reactor cost as a function of chamber rep-rate and plant efficiency.

In all three cases, the allowable reactor cost increases with increasing chamber rep-rate. This is because in going to higher rep-rates (and thus to lower driver energies), the driver cost is decreasing faster than the net power production. As the number of reactors sharing the driver increases, the allowable reactor cost increases and becomes less sensitive to the chamber rep-rate. With four reactors sharing a single driver, the allowable reactor cost is about \$150 million less than the cost of a PWR.

Perhaps a more appropriate comparison than direct capital cost is the capital cost per unit of thermal power. A more efficient plant requires less thermal power to produce a given amount of electrical power and therefore could conceivably be smaller and less costly.

Figure 7-35 shows the allowable ICFR capital cost per kilowatt normalized to the PWR cost of \$265/kWt. In this case, because the assumed ICFR plant efficiency is higher than the PWR efficiency (40 vs 33.4%), it takes only 2848 MWt to generate 1139 MWt, with the ICFR compared to 3412 MWt with the PWR. As a result, for four reactors sharing a single driver ($N = 4$), the allowable ICFR cost approaches that of the PWR. Note that the $N = 4$ case can also be viewed as a two-reactor power plant with a driver that costs only \$100 million/MJ. Likewise, $N = 1$ corresponds to a two-reactor plant with a \$400 million/MJ driver. Thus, Fig. 7-35 also reveals sensitivity to the actual driver cost.

Figure 7-36 shows the sensitivity to plant efficiency of the allowable ICFR cost per kilowatt. Note that this figure assumes a 10% driver efficiency and two reactors sharing a driver. As expected, increasing the plant efficiency has significant leverage in increasing the allowable reactor cost. An ICFR with a 50% plant efficiency could cost 10 to 20% more per kilowatt than a PWR and still be economically competitive by our definition. We have worked on two chamber concepts that can approach these high efficiencies by operating at high

temperatures (~ 1200 K) or by utilizing direct conversion schemes.^{103,104}

Figure 7-37 shows the sensitivity of allowable reactor cost to target gain. At a chamber rep-rate of 5 Hz, a factor-of-2 increase in target gain over G_0 results in a 10% increase in the allowable cost; a factor-of-2 decrease results in an 11% decrease. Note that the allowable cost is a weak function of the chamber rep-rate beyond 2 to 3 Hz.

Figure 7-38 shows the sensitivity of allowable reactor cost to driver efficiency. Again at 5 Hz, a factor-of-2 increase in efficiency produces a +5% change; a factor-of-2 decrease produces a -11% change. As in Fig. 7-37, the curves flatten out (become relatively independent of chamber rep-rate) beyond a few hertz. While a factor-of-2 decrease in either target gain or driver efficiency has a relatively small effect on the allowable reactor cost, a factor-of-2 decrease in both is significant: the allowable reactor cost (for $N = 2$) decreases by $\sim 31\%$.

We also examined the effects of combining our optimistic and pessimistic parameters. The pessimistic case has a plant efficiency equal to that of the PWR (33.4%) and a 5% efficient driver dedicated to a single reactor. For the reference gain (G_0), the allowable reactor cost is only 40 to 50% of the PWR cost. In this case the allowable cost is quite sensitive to target performance: increasing the gain by a factor-of-2 increases the allowable cost by 50%. There is little hope of reaching these cost goals.

The optimistic case has a plant efficiency of 50% and a 20% efficient driver shared by four reactors. Here the allowable reactor cost is ~ 1.3 times the PWR cost for the reference gain. A factor-of-2 increase in target gain gives allowable costs from 1.2 to 1.4 times the PWR cost.

From our discussion of normalized ICFR costs, we see that the allowable reactor cost increases with increasing chamber rep-rate. From the viewpoint of reactor design, however, high rep-rates are more difficult to obtain. For fluid-wall concepts,¹⁰⁵ time must be allowed between pulses for reestablishing the protective flow patterns. In addition, all concepts require time to pump out the chamber to allow for beam propagation for the following pulse. Because of these considerations, we asked ourselves how high the chamber rep-rate

Fig. 7-37. Allowable reactor cost as a function of chamber rep-rate and target gain.

Fig. 7-38. Allowable reactor cost as a function of chamber rep-rate and driver efficiency.

Table 7-19. Variation in chamber rep-rate for an allowable ICFR cost that is 95% of the maximum, allowable in the range of 1 to 10 Hz. ▼

Number of reactors per driver, N	Plant efficiency, η_p (%)	Driver efficiency, η_d (%)	Chamber rep-rate, R (Hz)
1	40	10	7.0
2	40	10	4.0
4	40	10	1.5
2	33.4	10	4.3
2	40	10	4.0
2	50	10	3.7
2	40	5	2.2
2	40	10	4.0
2	40	20	4.5

must be to have an allowable reactor cost that is within 5% of the maximum allowable in the range investigated.

Table 7-19 shows the chamber rep-rate at which the allowable cost is 95% of the maximum in the range of 1 to 10 Hz. For the nominal case of 40% plant efficiency and a 10% efficient driver for two reactors, the chamber rep-rate required is 4.0 Hz. The required rep-rate decreases with increasing

plant efficiency and with more reactors sharing the driver. On the other hand, it increases with increasing driver efficiency. The required rep-rate is most sensitive to the number of reactors sharing the driver and least sensitive to plant efficiency. The implication is that chamber rep-rates in the range of 2 to 5 Hz are adequate even though the maximum allowable reactor cost occurs at somewhat higher rep-rates.

Conclusions. We have estimated the allowable reactor cost for an economically competitive ICF power plant. We find that the cost per kW_e is 10 to 20% less than for a PWR for our nominal parameters of 40% plant efficiency and a 10% efficient driver for two reactors. Over the range of parameters investigated, the allowable reactor cost per kW, ranged up to 1.4 times that of a PWR. Increasing plant efficiency has the greatest leverage for increasing the allowable cost. Sharing the driver between several reactors also offers significant advantages. We find that a factor-of-2

change in driver efficiency or target gain has a smaller impact on the allowable cost. A factor-of-2 decrease in both gain and driver efficiency, however, leads to a significant reduction in the allowable reactor cost. For the nominal parameters, a chamber rep-rate greater than ~4 Hz is desirable.

Author: W. R. Meier

Development Facilities for ICF Reactor Commercialization

In a previous study,^{10b} we stated our belief that three integrated facilities will be needed to commercialize inertial confinement fusion. These facilities consist of an engineering test facility (ETF) acting in conjunction with a target physics demonstration facility (TPDF), an experimental power reactor (EPR), and a demonstration power reactor (DEMO). These facilities were characterized in terms of

Fig. 7-39. Relative cost of electricity as a function of fusion gain.

chamber-pulse-rate/driver-efficiency combinations in Ref. 106. In this work we report the results of our extended investigation of these combinations.

Target Gain and Reactor Design

Curves. To characterize each development facility, we must consider potential commercial plants in terms of their design points. Then, each development facility can be tailored to satisfy its design goals and to minimize costs by generating a minimum of fusion power.¹⁰⁷ The commercial plant design points are a function of the chamber pulse rep-rate, the driver efficiency, and the target-gain vs driver-energy relationship.

Monser has shown¹⁰⁸ that because the capital cost of the fusion reactor dominates the COE, for a given thermal efficiency η_t = (gross electric power/gross thermal power),

$$\text{COE} \approx \frac{\eta_t}{\eta_s} = \left(1 - \frac{f_a}{\eta_t} - \frac{1}{\eta_d G M \eta_t} \right)^{-1} \quad (44)$$

here η_s is the system efficiency (net electric power/reactor thermal power), η_d is the driver efficiency, f_a is the fraction of the thermal power required by nondriver auxiliary systems, G is the target gain ($\eta_d G$ = driver efficiency · target gain = fusion gain), and M is the system multiplication factor. The COE is plotted in Fig. 7-39 as a function of the fusion gain. The curve is a region of rapidly changing slope (a "knee") when the fusion gain is between 15 and 40. As the fusion gain is reduced to below ~15, the COE increases rapidly due to excessive recirculated power to the driver. Conversely, if the fusion gain is increased to above ~40, the reduction in COE becomes less significant.

Although Fig. 7-39 shows that the product of driver efficiency and target gain is a key variable, we need to consider the two separately. For this article, we use two driver efficiencies: 25% for a generic heavy-ion beam (HIB), and 5% for a generic short-wavelength laser (SWL). However, a more thorough analysis would consider lasers with higher efficiencies, HIBs with lower efficiencies, and the variation of each driver's efficiency with driver energy and pulse rate.¹⁰⁹

Target gain can be related to COE from Fig. 7-39 for each driver. Relating target gain

Fig. 7-40. Reactor-design curves and target-gain bands vs driver energy.

to driver energy is more difficult; indeed, it is the prime objective of both the theoretical and experimental portions of the ICF program. In Fig. 7-40, we show optimistic and conservative predictions¹¹⁰ for driver energy vs target gain that apply to both SWLs and HIBs. The optimistic band represents probable upper bounds to what might be achieved. The conservative band is predicted by current design codes when all known deleterious effects are included. We expect that multimegajoule experiments will produce conservative gains after some initial experimentation, with performance rising slowly as design and fabrication techniques are refined. Since the final gain curve will not be available until at least a few years after the TPDF operates, we must retain the flexibility in our development program to adjust to the TPDF results.

By using the driver efficiency and pulse repetition rate as parameters, we can plot

reactor design curves on the same graph as the target-gain bands (Fig. 7-40). The reactor design curves are shown for constant values of driver efficiency and chamber power. Unless otherwise noted, each reactor design curve begins at its upper left with a fusion gain of 40 and ends at its lower right with a fusion gain of 15. The net electric power per chamber is 1 GW_e except for the curve labeled 5 GW_e. The curves assume that $M = 1.15$, $f_a = 0.03$, $\eta_i = 39\%$, and that 40% of f_a is recovered as thermal power. The curves have been plotted using the limits from Fig. 7-39; hence, we have assumed that the plant capital cost does not change when driver type, pulse rate, and driver energy change. If complete reactor cost analyses were used, the curves would shift with respect to each other. However, the shifts are presently unknown, and they are not expected to be greater than a factor of 2.

Several trends are evident from Fig. 7-40:

- For a given driver efficiency, going to lower rep-rates at the same fusion gain (or COE) means higher driver energies. Consequently, less optimistic target gains are required.
- Alternatively, for a given driver efficiency, increasing the rep-rate at constant driver energy can accommodate low target performance until the lower fusion gain limit is reached. Then, low target performance must be countered by increasing driver energy and decreasing rep-rate.
- For a given rep-rate, increasing driver efficiency reduces the demand on target physics. However, the higher efficiency

must be obtained without greatly increasing the driver cost. Some of the difference in driver efficiency between SWLs and HIBs can be accommodated by operating at opposite sides of the fusion-gain knee shown in Fig. 7-39.

- Increased chamber power, at the same rep-rate relaxes the target physics, as can be seen by comparing the design curves for 1- and 5-GW_e chambers operating at 10 Hz with a 5%-efficient driver.

Selection of Reactor Design Points.

Choosing a design point with a low fusion gain can greatly reduce the required target gain. However, if the driver energy is increased excessively to compensate for the lower target gain, the driver cost can begin to dominate the COE and the equation plotted in Fig. 7-39 is no longer valid; this is particularly true for the low rep-rate designs. Because the cost of both reactor and driver systems is still quite uncertain, we limited the driver energy to a maximum of 5 MJ, and then selected operating points for five potential commercial designs (Table 7-20 and Fig. 7-40). The driver efficiency, chamber rep-rate, and net electric power are the independent variables that distinguish the five options. Using these values, a combination of driver energy and target gain was chosen to produce reasonable target physics requirements, fusion energy gain, and recirculated power. Finally, the plant power balance was completed. Our operating points used the minimum target physics and stayed above a fusion gain of 15 and below a 5-MJ driver energy. The points were chosen as examples; other choices would also be valid.

The 10-Hz, HIB-driven chamber design point is below the conservative gain band, and its fusion gain of 16.3 is about as low as is permitted. The 2-Hz, HIB-driven chamber design point is about 10% of the way between the conservative and optimistic bands; its fusion gain of 62 is in the diminishing-returns region. The 1-Hz, SWL-driven chamber design point has a target gain that is about 30% of the way between the conservative and optimistic bands; its fusion gain of 26.3 is in the middle of the knee of Fig. 7-39. The 5-Hz, SWL-driven chamber design point has a target gain that is also about 30% of the way between the gain bands, but its 2-MJ driver energy is

Table 7-20. Potential commercial facility operating points.

	Driver type: Driver rep-rate:	HIB High	HIB Low	Laser Low	Laser Medium	Laser High
Independent variables						
Driver efficiency (%)		25	25	5	5	5
Chamber rep-rate (Hz)		10	2	1	5	10
Net electric power (MW)		1046	1008	1000	1015	5076
Driver energy-target physics						
Driver energy (MJ)		4.5	5	5	2	5
Target gain		65	250	525	290	290
Target yield (MJ)		293	1250	2625	580	1450
Fusion energy gain		16.3	62.5	26.3	14.5	14.5
Recirculating power fraction (%)		21	11	16	23	23
Power balance						
Fusion power (MW)		2925	2500	2625	2900	14 500
Thermal power (MW)		3405	2910	3056	3376	16 878
Gross electric power (MW)		1328	1135	1192	1316	6582
Balance-of-plant power (MW)		102	87	92	101	506
Driver input power (MW)		180	40	100	200	1000

possibly below the threshold energy of the conservative gain band. To develop this target, the TPDF laser should have energy considerably above the conservative threshold; the fusion gain of 14.5 for this design is as low as permitted.

The 10-Hz, SWL-driven chamber requires target gains above the optimistic gain band if the chamber power is limited to 1 GW_e net. By increasing the chamber power to 5 GW_e, we can relax the requirement on target-gain physics; the design point is only 10% of the way between the conservative and optimistic gain bands. The fusion gain of 14.5 is as low as permitted.

Because either HIB or laser drivers can be shared among several chambers, multigigawatt plants will be possible using up to four of the chambers listed for each option. Modularized driver construction with some redundant components could make the system operate more like several 1-GW_e plants than like a single, multigigawatt plant from the viewpoint of grid reliability.

Characteristics of the Development Facilities. Using the commercial plant operating points given in Table 7-20, we specified characteristics of the three development facilities: TPDF/ETF, EPR, and DEMO. The TPDF driver size was specified at or above the commercial level to allow proof testing of reactor targets on TPDF. The ETF was specified with the constraint that the TPDF driver was to be used. EPR and DEMO were sized for fusion powers increasing to the commercial level in steps no larger than ~4. Once the fusion power was set for each facility, the scale-down of fusion yield and rep-rate from commercial to DEMO to EPR was determined. Fusion power was reduced by reducing both factors. Commercial rep-rate was not used in EPR and DEMO because the resulting low yield would make a geometrically scaled-down reactor chamber behave differently (because response is a function of both distance and time, and the time scale does not change when the dimensions are altered). Similarly, a commercial yield was not used in EPR and DEMO to allow study of pulsed reactor issues and to reduce chamber costs. However, EPR and DEMO should have the flexibility to increase either yield or rep-rate for limited periods.

	TPDF	ETF	EPR	DEMO	Commercial
Independent variables					
Driver efficiency (%)			25	25	25
Chamber rep-rate (Hz)		0.5-10 (2 pulses)	2	5	10
Net electric power (MW)			96 ^a	476	1046
Driver energy-target physics					
Driver energy (MJ)	4.5	2.25 (2 pulses)	4.5	4.5	4.5
Target gain	65	20	55	60	65
Target yield (MJ)	293	45	158	270	293
Fusion energy gain			8.8	15	16.3
Recirculating power (%)			33	22	21
Power balance					
Fusion power (MW)			315	1350	2925
Thermal power (MW)			367	1571	3405
Gross electric power (MW)			143 ^a	613	1328
Balance-of-plant power (MW)			11	47	102
Driver input power (MW)			36	90	180

^aElectric power may not be generated because of its low value in a low-availability experimental facility.

Table 7-21 shows a possible development plan for a HIB-driven, 10-Hz chamber. The required commercial target gain is demonstrated in TPDF; that facility will operate until the milestone is reached. The ETF add-on to TPDF will test several reactor concepts and allow selection of a preferred option for EPR. Off-line driver development will be geared to produce results before the EPR decision point; hence, the EPR will be built to integrate the plant subsystems for a single driver-reactor concept. The selected driver-reactor combination will probably be the only one carried forward, unless it develops insurmountable technical problems.

The increasing thermal power in going from EPR to DEMO to a commercial plant is accomplished by increasing both the rep-rate and the target gain in each step. However, this will be done only for reactor-development considerations; the commercial target-gain and driver-rep-rate demonstrations will be accomplished before EPR construction begins. A higher rep-rate or target gain will be possible in each facility for limited periods; thus short experiments can test the feasibility of the facility to follow.

Similar development scenarios were determined for an HIB-driven, 2-Hz chamber, and for SWL-driven, 1-, 5-, and 10-Hz chamber options (Ref. 107). The increasing target yield and rep-rate in succeeding development facilities are shown in Fig. 7-41 for the five commercial options.

Table 7-21. Development scenario to support a HIB driven, 10-Hz chamber option.

Fig. 7-41. Fusion yield and rep-rate for the development facilities required to support five commercial reactor options.

The development plan was devised to allow the TPDF/ETF to have the flexibility to test target gains and reactor concepts in support of all five scenarios. However, our development plan example limits EPR and DEMO to initially test only the preferred concept, based on TPDF/ETF experiments and off-line driver development data. If the development plan is adjusted to allow testing of all five driver/rep-rate combinations in EPR, either the EPR-to-DEMO and DEMO-to-commercial steps will be excessively large, or the EPR will be extremely expensive.

Summary. The ICF program needs to build a TPDF as soon as possible because demonstration of high gain is critical for both military and civilian applications. The TPDF could be upgraded to do ETF-type reactor experiments at little additional cost, and these experiments could support reactor development for a wide range of driver-efficiency and chamber-rep-rate combinations. The TPDF and ETF experiments will provide the target-physics and reactor-concept feasibility data required to select the appropriate combination of commercial driver efficiency and chamber rep-rate. Once these two characteristics are specified, the EPR and DEMO characteristics can also be specified to provide a smooth scale-up in power from facility to facility. In following this procedure for five potential combinations of driver efficiency and chamber rep-rate, we noted that if problems with target gain develop, the ICF program should move toward higher driver efficiency rather than toward higher rep-rate.

Author: J. A. Blinks

Major Contributor: M. J. Monsler
(KMS Fusion)

Disassembly of Isochorically Heated Liquids

In several ICF reactor concepts, the fusion neutrons will be suddenly deposited in liquid metals having free surfaces within the vacuum chamber. The neutron energy deposition will be so fast that the liquid will have little opportunity to expand during the deposition time; hence, the deposition can be considered to be at constant volume, or isochoric; i.e., the deposition time is much less than the time for a sound wave to transit the liquid region. The suddenly increased temperature and energy content of the liquid produces a sudden increase in pressure. Because the free surfaces are exposed to effectively zero pressure, tensile relief waves propagate inward from the surfaces. Tension causes liquids to cavitate, and this cavitation leads to fragmentation in liquids with available expansion volume. The fragments are accelerated by the pressure difference across the tensile wave.

Fragmentation of liquid metals inside ICF reactor vessels has two impacts on reactor design. First, the fragments eventually collide with the structure, causing a pressure load that can be calculated from continuum-mechanics computer codes. Second, the fragment size determines the liquid surface area available to interact thermally with superheated gases or plasma produced by the absorption of the short-range fusion x-rays and debris. Since the cooling and condensation times depend on the cooling surface area, it is possible that the maximum chamber rate will be set by the degree of fragmentation.

ICF reactors with free liquid in the reaction chamber include:

- **HYLIFE**¹¹¹—20-cm-diam lithium jets at radii of 0.5 to 2 m.
- **HIBALL**¹¹²—3- to 10-cm diam porous woven SiC "socks" filled with Li₁₇Pb₈₃ at radii of 5 to 7 m.
- **The Los Alamos Modified Wetted Wall (MWW)**^{111,113}—a 1- to 8-cm-thick lithium film flowing on a 2-m-radius wall.
- **EAGLE**¹¹⁴—a 1-m-thick spray of 0.5-mm-diam lithium droplets in a gas-protected chamber near the 4-m-radius wall.
- **Pulse-Star**¹¹⁵—a spray of Li₁₇Pb₈₃ between two spherical shells at radii of 4 and 5.5 m. The drops could be formed either

during the injection process or by disassembly induced by isochoric heating.

In some of these reactor concepts, the liquid-metal residence time is less than the interpulse time; in others, it is many times longer. The liquid-metal geometries include free spheres, free cylindrical jets, cylindrical jets confined by porous tubes (with some liquid film on the exterior surface), and liquid film on the inside surface of a sphere. Two different liquid metals have been suggested, and the fusion yields vary greatly. Thus, any analytic or empirical theory of isochoric heating-induced fragmentation must span quite a large parameter space.

Neutronic codes have been used to calculate the time- and space-dependent neutron energy-deposition levels in several of these reactors. In Fig. 7-42, we show the temporal neutron energy-deposition profile for the HYLIFE reactor. The deposition time τ_{dep} in any particular region is less than $1 \mu\text{s}$ because the profile is spatially integrated over the entire blanket. Similar results are expected for the other reactor geometries. The time for relief of the induced pressure is the liquid thickness D (or diameter) divided by the sound speed. As shown in Table 7-22, the sound speeds, c , of lithium and $\text{Li}_{17}\text{Pb}_{83}$ are 4.5 and 1.8 km/s, respectively. The heating is isochoric if

$$\frac{D}{c} \gg \tau_{\text{dep}} = 0.1 \mu\text{s} \quad (45)$$

That is, for isochoric heating the minimum diameter of the liquid, D_{min} , is 0.5 mm for lithium and 0.2 mm for $\text{Li}_{17}\text{Pb}_{83}$.

In Fig. 7-43, we show the spatial neutron energy-deposition profile in the HYLIFE reactor for an 1800-MJ fusion pulse. The neutron energy deposition drops to 35% of the peak value across the inner jets, and it is a factor of ~ 15 lower in the outer jets. Using the leading edge value of $\sim 800 \text{ kJ/kg}$, we can calculate the pressure rise P from the Grüneisen equation of state,

$$P = \rho I, \quad (46)$$

where ρ is the density (485 and 9320 kg/m^3 for lithium and $\text{Li}_{17}\text{Pb}_{83}$), Γ is the Grüneisen parameter (1.0 and 2.0 for lithium and $\text{Li}_{17}\text{Pb}_{83}$), and I is the specific energy deposited, in J/kg . For HYLIFE, the peak neutron-induced pressure is $\sim 400 \text{ MPa}$ or 4 kbar.

Fluid	Sound speed, c (m/s)	Density, ρ (kg/m^3)	Grüneisen parameter, Γ	Surface tension, $\gamma(\text{J/m}^2)$
Lithium	4500	485	1.0	0.35
$\text{Li}_{17}\text{Pb}_{83}$	1800	9320	2.0	0.435

Table 7-22. Fluid properties of free liquids in reaction chambers.

Fig. 7-42. Neutron energy deposition in the HYLIFE lithium-jet array as a function of time.

Fig. 7-43. Neutron energy-deposition profile in the HYLIFE lithium-jet array for 1800-MJ fusion pulses.

The neutron energy deposition per unit volume scales roughly with $\phi Y/r^2$, where Y is fusion yield, r is the blanket radius, and ϕ is the volume fraction of absorbing material. To scale from lithium to $\text{Li}_{17}\text{Pb}_{83}$, the volumetric deposition is increased by another factor of ~ 6 . Using this formula and published neutronic results for the other reactor geometries, we calculated induced pressures and relief times for each geometry (Table 7-23).

The jump-off velocity, v , of the expanding liquid can be calculated by considering the pressure drop across a moving rarefaction wave. Ahead of the wave, the fluid is at rest, or it has a velocity of $(-c)$ in the frame of reference of the wave. Behind the wave, the fluid has a velocity of $(v - c)$ in the same frame. The momentum equation is

$$P - (\rho c)(v - c) = 0 + (-\rho c)(-c), \quad (47)$$

or

$$P = \rho v^2. \quad (48)$$

Reactor	Yield (MJ)	Pulse rate (Hz)	Fluid	Thickness, R (cm)	Relief time (μ s)	Specific e. ergy, e (J/g)	Pressure (MPa)	Jump-off velocity (m/s)
HYLIFE (inner jet)	1800	1.5	Li	20	44	800	400	180
HYLIFE (outer jet)	1800	1.5	Li	20	44	50	25	11
HIBALL (inner jet, ignoring SiC sock)	400	5	$\text{Li}_{17}\text{Pb}_{83}$	3	17	0.5	9	0.5
MWW	150	10	Li	1-8	2-16	3.5	1.8	0.8
EAGLE	300	3	Li	0.05	0.1	3.7	1.9	0.9
Pulse*Star	600	5	$\text{Li}_{17}\text{Pb}_{83}$	1	6	1	19	1.1

Table 7-23. Isochoric $\blacktriangle$ heating in several ICF reactors.

Fig. 7-44. Pressure-volume diagram for lithium. The isochoric heating and isentropic expansion back to the saturated liquid line are shown for an inner HYLIFE jet.

For HYLIFE, the 400-MPa pressure results in a 180-m/s jump-off velocity. Table 7-23 also shows the jump-off velocities for the other reactor concepts.

Thermodynamic Breakup Theory. Some insight into the breakup can be obtained by examining a pressure-volume diagram for the liquid metal. Figure 7-44 is a P - V diagram for lithium, constructed using Young's soft-sphere model for the equation of state.¹¹⁶ The figure shows only a portion of the subcooled liquid region. The saturated-liquid line is very shallow in this density range; it is almost indistinguishable from the abscissa. The saturated-liquid line

eventually rises to the critical pressure of 114 MPa at $0.011 \text{ m}^3/\text{kg}$ (91 kg/m^3).

The initial condition, before the fusion pulse, is on the saturated-liquid line at 500°C (485 kg/m^3). The fusion neutrons add a maximum of 800 kJ/kg (HYLIFE inner-jet inner edge in Table 7-23) in a time short compared to the material response time. The liquid temperature and pressure quickly rise to 735°C and 380 MPa with no density change. As the relief wave moves into the jet, an adiabatic (isentropic) expansion begins. When the saturation line is reached, several isotherms have been crossed, and the liquid temperature has

Fig. 7-45. Continuation of the isotherms into the two-phase region. Isentropic expansion is shown until the spinodal is encountered in the lower right corner.

fallen to 695°C. During the expansion, the internal energy falls by $\sim 16 \text{ kJ/kg}$ from the value after the isochoric heating. Using Eqs. (46) and (48) to obtain the specific kinetic energy K/m , we have

$$\frac{K}{m} = \frac{1}{2} v^2 = \frac{1}{2} \left(\frac{P}{\rho c} \right)^2 = \frac{1}{2} \left(\frac{\Gamma c}{c} \right)^2 \quad (49)$$

With the values for Γ , c , and c from Tables 7-22 and 7-23, we have $\sim 15.2 \text{ kJ/kg}$ of specific kinetic energy produced in the HYLIFE inner jets by the tensile wave, without considering any phase change.

Hence, most of the lost internal energy is converted to kinetic energy.

We can continue to follow the expansion on the P - V diagram to obtain an upper bound on the vaporization contribution. The isotherms can be extended into the two-phase region, forming van der Waals loops, portions of which are shown in Fig. 7-45. The lines formed by the maxima and minima of the collection of isotherm extensions are known as the spinodal lines. States between either of these lines and the two-phase dome are metastable, and they eventually separate into liquid and

vapor regions by the processes of nucleation and growth. States between the spinodal lines are absolutely unstable because $(\partial P/\partial V)_T > 0$, which is an unphysical situation. These states decompose to liquid- and vapor-rich regions by the extremely fast process of spinodal decomposition; the regions then further decompose to pure liquid and vapor by nucleation and growth.

Suppose that the isentropic expansion continues all the way to the spinodal line; the intersection is shown in the lower right corner of Fig. 7-45. The specific volume has increased by $\sim 50\%$, and the vapor volume will be $\sim 0.001 \text{ m}^3/\text{kg}$ -liquid because $\sim 0.002 \text{ m}^3/\text{kg}$ -liquid is still occupied by liquid. If the vapor is in equilibrium with 735°C liquid, the vapor pressure is $\sim 120 \text{ Pa}$ (0.9 Torr) and the vapor specific volume is $\sim 9690 \text{ m}^3/\text{kg}$ -vapor. The quotient of these two specific volumes is $\sim 1.0 \times 10^{-7}$ (100 μg -vapor/kg-liquid), an extremely low vaporization fraction. Since vaporization requires $\sim 20.7 \text{ MJ/kg}$ -vapor, only 2.1 J/kg -liquid are required for vaporization. This is about 7000 times less energy than the kinetic energy produced by the hydrodynamic expansion of a pure (nonvaporizing) liquid. Hence, it appears that the hydrodynamic expansion creates space that can be easily filled by vapor, rather than sudden vaporization causing a violent expansion. Spinodal decomposition would be a sudden vaporization, but it would not be violent because it requires a prior expansion of 50% before it can occur.

It is unlikely that the spinodal line would be reached in a liquid-lithium ICF reactor. The tension in the liquid at the spinodal line $\sim 1500 \text{ MPa}$, much greater than the tensile strength of solid lithium or the tensions induced in extremely clean water during carefully controlled experiments. In an ICF reactor, the entrained gas, target debris, and corrosion products will provide an abundance of potential nucleation sites, allowing a more gradual phase change by nucleation and growth.

Hydrodynamic Breakup Theory. An early theory of stable drop size was proposed¹¹⁷ in 1979. In this model, the pressure was balanced against the surface tension γ (Table 7-22), producing a fragment size a of

$$a = \frac{2\gamma}{\rho\Gamma c} \quad (50)$$

Using the values of γ , ρ , Γ , and $c = 800 \text{ kJ/kg}$ from Tables 7-22 and 7-23, we find $a = 1.8 \text{ nm}$ for the HYLIFE leading edge.

A better estimate of the stable drop size can be made by considering that the pressure is relieved in the time τ required for a sound wave to propagate across the drop, that is, $\tau = 2a/c$. During that time, the outer surface can only move a distance $v\tau$. Because breakup is not expected during such a small dilation, the pressure to be resisted by surface tension is the dynamic pressure, $\rho v^2/2$. If we use this pressure, the fragment size is

$$a = \frac{4\gamma}{\rho v^2} = \frac{4\gamma}{\rho} \left(\frac{c}{\Gamma c} \right)^2 \quad (51)$$

Using the HYLIFE parameters again, we find $a = 90 \text{ nm}$, which is a more reasonable, but still very small, value.

However, we are not seeking the drop size that will not fragment; instead, we want to know how a larger drop breaks up. If there is no fragmentation, the surface forces can reverse the expansion velocity and eventually dissipate all the kinetic energy as heat. If there is fragmentation, the fragments have velocities, so not all of the kinetic energy is contained by the surface forces and subsequently dissipated as heat. The problem then is to specify what fraction of the kinetic energy (or dynamic pressure) must be balanced by the surface forces. Grady¹¹⁸ began work on this problem, and Glenn¹¹⁹ later extended it.

Grady begins by using the fragment size as the independent variable and splitting the kinetic energy into the motion of the fragment center of mass and the dilation about the fragment center of mass. This dilation (or local) kinetic energy is added to the surface energy, and the sum is minimized with respect to the fragment size.

Glenn found that fragment size is

$$a = \left[\frac{5\gamma}{\rho} \left(\frac{Rc}{\Gamma c} \right)^2 \right]^{1/3} \quad (52)$$

Here, R is the radius at a sphere edge. If the sphere size is chosen to avoid fragmentation ($a = R$), Glenn's formula agrees well with the preceding theory. For example, with HYLIFE loading, the stable ($a = R$) drop size is 114 nm , within 25% of the value given by the simple theory.

Reactor	Fluid	Thickness, R (cm)	Fragment radius, a (cm)	Total added energy density, e (J/kg)	Fragment kinetic- energy density (J/kg)	Dilation kinetic- energy density (J/kg)
HYLIFE (inner jet)	Li	20	0.11	800 000	15 800	1
HYLIFE (outer jet)	Li	20	0.65	50 000	61.6	0.16
HIBALL (inner jet, ignoring SIC sock)	Li ₁₇ Pb ₈₃	3	0.9	500	0.14	0.013
MWW	Li	4	1.4	3 500	0.22	0.081
EAGLE	Li	0.05	0.07 ^a	3 700	0 ^a	0.34 ^a
Pulse-Star	Li ₁₇ Pb ₈₃	1	0.27	1 000	0.58	0.042

^aNo breakup by hydrodynamic dilation predicted.

In Table 7-24, we show the predicted fragment size for the various reactor geometries described in Table 7-23. The table also compares the local dilational kinetic energy to the kinetic energy of the fragment centers of mass. Almost all the kinetic energy is fragment kinetic energy; only a small fraction is local dilational kinetic energy.

The Grady-Glenn theory gave the first reasonable estimates of fragment size. However, the theory is too simple in that it considers only the lowest final energy state; it does not attempt to find the path between the initial and optimum final state. There could be potential barriers; that is, the path between the states might pass through a state of higher energy. In that case, the optimum state predicted by Grady and Glenn might not be reached during the time of interest.

Isochorically heated drops could begin to break up in two ways. In the first, the growing relief wave could cause cavitation at some depth into the drop, and an onion-skin fragment would result. Since the skin area increases as it expands outward, it will stretch and break, perhaps by some surface instability mechanism. The drop size might then be about the onion-skin thickness. (This mode is expected for films on surfaces, such as the MWW or the liquid on the outsides of the HIBALL socks.) The second breakup mode would be to fragment the drop into wedges. Since the circumference increases 2π times faster than the radius during expansion, this mode seems more likely. If the wedges are longer than they are wide, they could either collapse into spheres or fragment further because of velocity gradients or surface waves.

Breakup by either of these initial modes could change the local kinetic-energy density and, thus, require revised calculations using the Grady-Glenn theory. The

modes could also lead to further nonspherical breakup.

Molecular Dynamics Experiments.

To gain insight into the physics of how isochorically heated liquid lithium disassembles into drops, we are modeling small regions on an atomic scale as nodes connected by nonlinear springs in two dimensions. This technique is known as molecular dynamics. Our simple molecular dynamics model uses the Lennard-Jones 6-12 potential for the atomic interactions.

Using arrays of a few hundred atoms, we have modeled solid (crystalline) states in which the atoms vibrate around fixed points, liquid states in which atoms exhibit extensive interactions and diffuse throughout the region, and gaseous states in which atoms exhibit little interaction. We have also observed the phenomena of freezing, melting, vaporization, vapor pressure, and surface tension. Other workers have used molecular dynamics to obtain transport properties and to model systems of macroscopic particles.

Our experiments begin with a perfect hexagonal crystal at a density appropriate to a saturated liquid that is near neither the critical point nor the triple point. Although the initial temperature is zero, the attractive bonds begin to collapse the crystal symmetrically. We enhance the collapse with a slowly converging outer boundary. Time exposures of the atomic paths show the crystal melting. The wall motion is stopped before the density is reduced beyond the desired value. (Some wall convergence is required because a circle enclosing a hexagon contains more area than the hexagon itself.)

After the wall motion is halted, the atomic velocities are adjusted to the desired temperature before isochoric heating. Then the region is allowed to equilibrate. The

Table 7-24. Predicted fragmentation in several ICF reactors.

Fig. 7-46. Time exposure of atomic trajectories near the end of the fragmentation process. The lines indicate the trajectories, and the dots indicate the final atomic positions.

wall remains during this period to prevent loss of atoms to vapor.

After the equilibration period, the wall is removed, and the atomic velocities are adjusted to account for the isochoric heating. If no heat is added, a 271-atom system does not fragment, although some atoms are ejected as vapor. When the equivalent of 800 kJ/kg is added, ejection of several clusters of 5 to 10 atoms occurs (Fig. 7-46). A number of single and double atoms are also ejected as vapor. The large, elongated cluster in Fig. 7-46 looks as if it will break into several more clusters, but it later becomes a single drop. Because the 271-atom system almost breaks, we believe that a somewhat larger system will definitely break, contrary to the prediction of the three-dimensional Grady-Glenn theory using lithium properties.

The cluster temperatures and rotational energies in the molecular dynamics experiments are consistent with equipartition. The cluster translational energies are somewhat higher than predicted by Eq. (49).

Summary. The sudden-vaporization approach to drop fragmentation appears to be unfruitful. Spinodal decomposition cannot cause sudden vaporization until considerable expansion volume has been created (hydrodynamically). The Grady-Glenn theory gives a reasonable method to estimate

fragment size, but other methods are needed to determine the influence of the fragmentation dynamics. Molecular dynamics is being used to study the fragmentation of small drops. We will attempt to extend the molecular-dynamics method to larger drops and the Grady-Glenn theory to smaller drops and two dimensions.

Author: J. A. Blink

Major Contributors: W. G. Hoover, R. A. Haas, and J. Hovingh

Blanket Optimization Techniques

Eighty percent of the energy released in the DT fusion reaction in an ICF reactor is in the form of high-energy neutrons. One approach to utilizing this energy is to design a thick blanket to absorb these neutrons, breed replacement tritium, and provide heat at a temperature suitable for producing electric power. In this approach, optimum blanket design is a key element in the reactor design. Finding the optimum design, however, is complicated by the nonlinear interdependence of design and performance variables and by the fact that modifications often produce conflicting effects, improving one parameter while degrading another.

The blanket-optimization problem consists of four principal elements:

1. The figure of merit for the particular reactor design (e.g., the COE).
2. A technique for estimating the neutronic performance as a function of selected independent design variables.
3. Constraints on both the design variables and the neutronic performance.
4. A nonlinear programming technique to optimize the figure of merit subject to these constraints.

In this article we describe the technique used for estimating the neutronic performance of fusion reactor blankets. The heart of the technique is variational interpolation,^{120,121} a particular application of a broader category of variational methods. Background information and related applications are discussed in Refs. 122-126.

Variational Interpolation. In our study of the neutronics of fusion reactor blankets, we are primarily interested in integral properties that can be written as linear functionals of the neutron flux. Examples

include tritium production, nuclear heating, displacement damage, helium and hydrogen production, and neutron activation rates. Consider, for example, a reaction rate, R . The reaction rate can be expressed as the inner product of the macroscopic reaction cross section, Σ , and the neutron flux distribution, ϕ :

$$R = \langle \Sigma, \phi \rangle \quad (53)$$

The inner product notation $\langle \rangle$ indicates a sum over all discrete, independent variables (e.g., various regions of the blanket) and an integral over all continuous, independent variables (e.g., space, energy, and direction).

The neutron flux is the solution of the time-independent Boltzmann transport equation,

$$L\phi = S \quad (54)$$

in which L is the transport operation such that

$$L\phi = \Omega \cdot \nabla \phi + \Sigma_T \phi - \int_{\Omega'} \int_{E'} \Sigma_T f \phi \, d\Omega' \, dE' \quad (55)$$

where

$\phi = \phi(\mathbf{r}, \Omega, E)$, the angular flux distribution at position $\mathbf{r}$, direction Ω , and energy E ,

$\Sigma_T = \Sigma_T(\mathbf{r}, E)$, the total macroscopic cross section,

$f = f(\mathbf{r}, \Omega' \rightarrow \Omega, E' \rightarrow E)$ is the transfer probability function such that $\Sigma_T(\mathbf{r}, E)f$ is the total probability of neutron transfer from Ω', E' to Ω, E , and

$S = S(\mathbf{r}, \Omega, E)$ is the external or independent neutron source.

Let x be a variable that characterizes the blanket. We can estimate the reaction rate, $R(x)$, using the flux, $\phi_{a'}$, calculated for some reference system, $x = x_{a'}$, in several ways. If we define

$$\delta\phi_a = \phi - \phi_a \quad (56)$$

then

$$R(x) = \langle \Sigma, \phi \rangle = \langle \Sigma, \phi_a \rangle + \langle \Sigma, \delta\phi_a \rangle \quad (57)$$

The estimate

$$R(x) \approx \langle \Sigma, \phi_a \rangle \quad (58)$$

is therefore in error to first order in $\delta\phi_a$.

A variational estimate of R yields a result accurate to second order¹²⁶ in $\delta\phi_a$ and $\delta\phi_{a'}$, where

$$\delta\phi_a^* = \phi^* - \phi_a^* \quad (59)$$

The adjoint flux, ϕ^* , is the solution to

$$L^* \phi^* = S^* \quad (60)$$

where the adjoint operator L^* is defined by¹²⁷

$$\langle \phi^*, L\phi \rangle = \langle L^* \phi^*, \phi \rangle \quad (61)$$

For our purposes, we let $S^* = \Sigma$ so that we can relate the adjoint flux to the reaction rate by

$$\langle \phi^*, L\phi \rangle = \langle \Sigma, \phi \rangle = R \quad (62)$$

We use a variational interpolation method of estimating R that is derived from the Schwinger (fractional) functional,¹²¹

$$R_S(\phi_a^*, \phi_{a'}; x) = \frac{\langle \phi_a^*, S(x) \rangle \langle \Sigma(x), \phi_a \rangle}{\langle \phi_a^*, L(x)\phi_a \rangle} \quad (63)$$

The interpolation form uses the forward flux calculated at $x = x_a$ and the adjoint flux calculated for a different reference system, $x = x_{b'}$, with $x_b > x_a$:

$$R_{SI}(\phi_b^*, \phi_{a'}; x) = \frac{\langle \phi_b^*, S(x) \rangle \langle \Sigma(x), \phi_a \rangle}{\langle \phi_b^*, L(x)\phi_a \rangle} \quad (64)$$

Applying Variational Interpolation to Fusion Reactor Blankets. For a fusion reactor blanket, the external neutron source is fixed in most cases of interest (i.e., a fixed fusion power level). Therefore,

$$L(x_a)\phi_a = L(x_b)\phi_b = S \quad (65)$$

If we further assume that the perturbation is linear in x , then $L(x)$ can be written as

$$L(x) = L(x_a) + \delta L(x) \quad (66)$$

where

$$\delta L(x) = \left(\frac{x - x_a}{x_b - x_a} \right) [L(x_b) - L(x_a)] \quad (67)$$

Substituting Eq. (66) into Eq. (64) and using Eq. (61) and Eq. (65) allows us to rewrite Eq. (64) in terms of forward fluxes only:

$$R_{SI}(\phi_b, \phi_a; x) = \frac{\langle \Sigma(x), \phi_a \rangle \langle \Sigma_b, \phi_b \rangle}{D} \quad (68)$$

where

$$D = \langle \Sigma_b, \phi_b \rangle + \left(\frac{x - x_a}{x_b - x_a} \right) [\langle \Sigma_b, \phi_a \rangle - \langle \Sigma_b, \phi_b \rangle] \quad (69)$$

The advantages of this formulation have been pointed out by Cheng.¹²¹ In particular, since only forward fluxes (ϕ_a and ϕ_b) are involved, no adjoint calculations are required. Calculating ϕ_a and ϕ_b allows us to interpolate in x on the functional $\langle \Sigma(x), \phi \rangle$ for any $\Sigma(x)$. That is, we can estimate several different reaction or damage rates based on the results of two transport calculations. This is in contrast to Eq. (64), where an adjoint calculation is required for each $\Sigma(x)$, i.e., S^* , of interest as indicated by Eq. (60).

Returning to Eq. (68), note that at $x = x_a$, $R_{SI} = \langle \Sigma_a, \phi_a \rangle$, which is exact. R_{SI} is also exact at $x = x_b$. Hence, Eq. (68) provides the exact solution at the two reference points and an estimate for other values of x .

Variations in Blanket Composition. Of particular interest in the study of fusion reactor blankets is how various reaction rates change as the isotopic composition of the blanket (or in a region of the blanket) is changed. For reactions with isotope j , the macroscopic cross section is

$$\Sigma_j = N_j \sigma_j = \frac{f_j \rho N_A \sigma_j}{M} \quad (70)$$

where

N_j = number density of isotope j ,
 σ_j = microscopic reaction cross section with isotope j ,

f_j = atom fraction of isotope j in the blanket material,

ρ = material density,

N_A = Avogadro's number, and

M = effective atomic weight of the material.

To investigate the effects of varying the fraction of isotope j in the blanket, we set $x = f_j$. Hence the macroscopic cross sections for reactions with isotope j are proportional to x . As a result, varying x gives linear perturbations in L , and Eq. (68) can thus be applied.

In this case, we can simplify (68) by noting that

$$\Sigma(x) = \frac{x \Sigma_a}{x_a} \quad (71)$$

Substituting gives

$$R_{SI}(\phi_b, \phi_a; x) = \left(\frac{x}{x_a} \right) \left[\frac{\langle \Sigma_a, \phi_a \rangle \langle \Sigma_b, \phi_b \rangle}{D} \right] \quad (72)$$

where

$$D = \langle \Sigma_b, \phi_b \rangle + \left(\frac{x - x_a}{x_b - x_a} \right) \left[\left(\frac{x_b}{x_a} \right) \langle \Sigma_a, \phi_a \rangle - \langle \Sigma_b, \phi_b \rangle \right] \quad (73)$$

In terms of the reaction rates obtained from the transport calculations at the two reference points, i.e., $R_a = \langle \Sigma_a, \phi_a \rangle$ and $R_b = \langle \Sigma_b, \phi_b \rangle$, Eq. (72) becomes

$$R_{SI}(R_a, R_b; x) = \frac{\left(\frac{x}{x_a} \right) R_a R_b}{D} \quad (74)$$

where

$$D = R_b + \left(\frac{x - x_a}{x_b - x_a} \right) \left[\left(\frac{x_b}{x_a} \right) R_a - R_b \right] \quad (75)$$

Equation (74) gives exact values for the reaction rate and a nonlinear interpolation for intermediate values of x . The three exact values are:

$R_{SI} = 0$ at $x = 0$,

$R_{SI} = R_a$ at $x = x_a$, and

$R_{SI} = R_b$ at $x = x_b$.

Note that Eq. (74) has a singularity at

$$x = x_a \equiv x_b - \frac{(x_b - x_a)R_b}{\left(\frac{x_b}{x_a}\right)R_a - R_b} \quad (76)$$

Since negative atom fractions are meaningless, this is of concern only if $x_a \geq 0$, which occurs if $R_a \geq R_b$. For $R_a < R_b$, Eq. (74) is well behaved and gives a nonlinear estimate of $R(x)$, which increases monotonically with x . If the reference-point calculations give $R_a \geq R_b$, then Eq. (74) is inadequate, and a three-point variational interpolation is required.

If the blanket is composed of only two isotopes, the atom fraction of each is related to x by $f_1 = x$ and $f_2 = 1 - x$. Then for isotope-2, Eq. (71) becomes

$$\Sigma(x) = \frac{(1-x)\Sigma_a}{1-x_a} \quad (77)$$

and Eq. (74) becomes

$$R_{s12}(R_a, R_b; x) = \frac{(1-x)R_b R_b}{(1-x_a)D} \quad (78)$$

where

$$D = R_b + \left(\frac{x-x_a}{x_b-x_a}\right) \left[\frac{(1-x_b)R_a}{1-x_a} - R_b \right] \quad (79)$$

Note that R_a and R_b are the reaction rates with isotope 2 derived from the reference-point calculations at $x = x_a$ and $x = x_b$, respectively.

If the atom fraction of an isotope is unaffected by the variation of x , then

$$\Sigma(x) = \Sigma_a = \Sigma_b = \text{constant} \quad (80)$$

In this case Eq. (74) becomes

$$R_{s1c}(R_a, R_b; x) = \frac{R_a R_b}{D} \quad (81)$$

where

$$D = R_b + \left(\frac{x-x_a}{x_b-x_a}\right) (R_a - R_b) \quad (82)$$

One example where this holds is in a blanket containing three or more isotopes with one or more specified as constant. In particular, for three isotopes we could have $f_1 = x$, $f_2 = 1 - C - x$, and $f_3 = C = \text{constant}$. Another example where Eq. (81) holds is in blanket regions other than the region where x is varied. In these examples, the variational estimate reflects the change in the flux arising from a change in x .

Variations in More than One Design Variable

Until now, we have considered the variation of a single design variable. In practice, it is desirable to simultaneously evaluate the effects of variations in several variables. By successive application of the variational interpolation procedure, we can estimate the neutronic parameters of interest at any point in an n -dimensional space, based on a limited number of reference-point transport calculations.

Consider the problem of estimating reaction rates as a function of two design variables, x_1 and x_2 . To do this using the variational interpolation scheme requires four reference-point transport calculations. These four points are most conveniently chosen to define a rectangle in Cartesian coordinates, as shown in Fig. 7-47. The reference points are denoted (x_{1a}, x_{2a}) , (x_{1b}, x_{2a}) , (x_{1a}, x_{2b}) , and (x_{1b}, x_{2b}) . The corresponding reaction rates are denoted R_{aa} , R_{ba} , R_{ab} , and R_{bb} .

The procedure used to estimate the reaction rate at some point (x_{1c}, x_{2c}) is illustrated in Fig. 7-48. We first interpolate on x_1 with $x_2 = x_{2a}$ to find the result R_{ca} . We then interpolate on x_1 with $x_2 = x_{2b}$ to find the result R_{cb} . The two results are then used to interpolate on x_2 with $x_1 = x_{1c}$ to find the result R_{cc} . Thus, three interpolations are required to estimate the reaction rate.

In terms of the previous expressions for the Schwinger interpolation, the successive-interpolation approach can be written as

$$\begin{aligned} R_{cs1}(R_{aa}, R_{ba}, R_{ab}, R_{bb}; x_1, x_2) \\ = R_{s1}[R_{s1}(R_{aa}, R_{ba}; x_1), \\ R_{s1}(R_{ab}, R_{bb}; x_1); x_2] \quad (83) \end{aligned}$$

(Clearly we could just as easily interpolate on x_2 first and then on x_1 .)

To extend the approach to a third variable, x_3 , would require four additional

Fig. 7-47. The four reference points for transport calculations are chosen on a rectangular grid.

Fig. 7-48. To find the result at an intermediate point requires three interpolations.

reference-point transport calculations. In this case, seven interpolations are required to estimate the reaction rate. Generalizing, to span n -space with this technique requires 2^n reference-point transport calculations and $(2^n - 1)$ interpolations for each reaction of interest at each point of interest.

Example Problem. To illustrate how the variational interpolation method is used to estimate reaction rates as a function of blanket design variables, we will use a modified HYLIFE chamber design, which we are currently evaluating. In HYLIFE,¹²⁸⁻¹³³ a thick liquid-lithium blanket surrounds the fusion target and protects the first structural wall from neutron damage. With a 1-m-thick blanket of natural lithium (7.42 at.% ^6Li and 92.58 at.% ^7Li), HYLIFE achieves a tritium-breeding ratio of 1.75.¹³⁴

The modified chamber design is an attempt to increase the energy deposited in the chamber per fusion reaction. The approach is to reduce the tritium-breeding ratio to near 1.0 and capture the excess neutrons in ^{55}Mn . Each such capture adds 11.0 MeV to the energy balance. The

$^{55}\text{Mn}(n,\gamma)$ reaction releases 7.3 MeV, while the decay of the product, ^{56}Mn , releases an additional 3.7 MeV. (The half-life for the β' decay is 2.6 h.)

There are two simple ways to reduce the tritium-breeding ratio: one is to reduce the fraction of ^6Li in the lithium blanket; the other is to reduce the effective blanket thickness.¹³⁵ Reducing the blanket thickness is advantageous in that it reduces the lithium flow rate, resulting in savings in pumping power and capital costs for liquid-metal pumps and piping. There is a lower limit to this reduction, however, set by the maximum allowable radiation-damage rate in the chamber wall. For HYLIFE, this rate was limited such that the wall could maintain its structural integrity for the 30-year life of the power plant.¹³⁶ Another constraint is the need for a tritium-breeding ratio greater than 1.0 to assure a self-sufficient fuel cycle. Therefore, we need to calculate the reaction rates as a function of the two design variables: the ^6Li fractions and the effective blanket thickness.

To illustrate, consider the ^6Li tritium-breeding reaction as a function of the ^6Li fraction in lithium, x_1 , and the effective blanket thickness, x_2 . Note that the actual zone thickness of the lithium blanket is held constant, while the packing fraction of liquid-lithium jets within that zone is varied. This is in essence a variation of the material density and thus qualifies as a linear perturbation, as Eq. (70) shows.

The four reference points for the transport calculations were

$$x_1 = 0.5 \text{ and } 7.42 \text{ at.}\% \text{ } ^6\text{Li at } x_2 = 0.75 \text{ m}$$

and

$$x_1 = 0.5 \text{ and } 7.42 \text{ at.}\% \text{ } ^6\text{Li at } x_2 = 1.25 \text{ m.}$$

The calculations were run on TART,¹³⁷ a Monte Carlo code that utilizes ENDF.¹³⁸

The results for the interpolation in x_1 for the two values of x_2 are plotted in Fig. 7-49. T_6 denotes the number of $^6\text{Li}(n,\gamma)$ reactions per DT neutron. The interpolation passes through the reference points, and $T_6 = 0$ at $x_1 = 0$. The points on these curves are then used to interpolate in x_2 to give T_6 as a function of x_1 and x_2 . The

resulting isobreeding lines for the ${}^6\text{Li}$ contribution are shown in Fig. 7-50.

The same procedure is carried out for ${}^7\text{Li}(n,n'\alpha)$ reactions. Adding T_6 and T_7 gives the tritium-breeding ratio as a function of x_1 and x_2 . The isobreeding lines for the tritium-breeding ratio are shown in Fig. 7-51. At low ${}^6\text{Li}$ concentrations, the tritium-breeding ratio is primarily a function of the ${}^6\text{Li}$ fraction and increases rapidly as that fraction increases. With more than 4 at. % ${}^6\text{Li}$, the tritium-breeding ratio is primarily a function of blanket thickness.

As a check on the accuracy of this successive-interpolation technique, we ran an additional transport calculation at $x_1 = 4$ at. % and $x_2 = 1$ m. The resulting tritium-breeding ratio was 1.57, which is only slightly higher than the 1.55 obtained by successive interpolation.

The type of information presented in Figs 7-50 and 7-51 is then coupled to a code that optimizes a specified figure of merit for the reactor design subject to constraints on the design variables and neutronic performance. Successive interpolation greatly reduces the number of neutronics calculations required in our efforts to optimize blanket design.

Author: W. R. Meier

Dependence of Displacement Damage on Blanket Design

Future fusion reactors will have to cope with the damaging effects of high-energy neutrons. One of the primary radiation-damage mechanisms in structural steels is the displacement of atoms from their lattice positions.¹³⁹ Displaced atoms leave vacancies that can conglomerate to form voids within the steel, and this leads to void swelling. After some total amount of damage, expressed in displacements per atom (dpa), void swelling deforms the steel or degrades its properties enough that it loses its integrity. Currently there is insufficient data to set absolute damage limits for fusion reactors. It is known, however, that ferritic steels are less susceptible to displacement damage than austenitic steels,¹⁴⁰ and a damage limit of ~ 200 dpa

Fig. 7-49. Variational estimate of tritium breeding from ${}^6\text{Li}$ as a function of the ${}^6\text{Li}$ fraction for 0.75- and 1.25-m-thick blankets.

Fig. 7-50. Isobreeding lines for ${}^6\text{Li}$ reactions as a function of the ${}^6\text{Li}$ fraction and the lithium blanket thickness.

Fig. 7-51. Isobreeding lines for total tritium breeding as a function of the ${}^6\text{Li}$ fraction and lithium blanket thickness.

was recently suggested¹⁴¹ as a reasonable estimate for high-chromium ferritic steels. A low-alloy ferritic steel, 2-1/4 Cr-1 Mo, has been specified in several ICF reactor designs¹⁴² because of its low cost and resistance to liquid-metal corrosion and radiation damage.

During 1983, we investigated how the design of the neutron blanket affects the displacement damage to the first structural wall (FSW) of an ICF reactor. We examined

two basic blanket configurations:

- The steel wall is directly exposed to the fusion neutrons.
- The steel wall is protected by a 1-m-thick inner blanket of lithium.

The second configuration represents a HY LIFE design shown to have displacement damage rates an order of magnitude lower than those of unprotected wall designs.¹⁴²⁻¹⁴⁴ We varied the two basic configurations to see how the dpa rate changes as the result of:

1. Adding a lithium blanket outside the FSW.

Fig. 7-52. Displacement cross section for iron used in calculations.

Fig. 7-53. Displacement damage calculated for four unprotected and four protected iron walls. ▽

2. Adding a neutron reflector (graphite) outside the FSW.
3. Changing the position of the inner lithium blanket relative to the FSW.

We also examined the effects of neutron moderation in the compressed-DT target for the two configurations.

Calculations. The displacement damage rate, R , is calculated as follows:

$$R = S \sum_i \sigma_i \Phi_i \text{ dpa/y} \quad (84)$$

where

- S = neutron source (n/y),
- σ_i = energy-dependent displacement cross section (b),
- Φ_i = energy-dependent neutron fluence (n/cm² per source neutron),
- i = energy group index for the multigroup calculation.

The source of DT neutrons is related to the fusion power, P_f , by

$$S = 11.2 \times 10^{24} P_f \quad (85)$$

where P_f is in MW.

In our calculations, we used the displacement-damage cross section for iron shown in Fig. 7-52. This cross section, calculated by Doran and Graves,¹⁴⁵ is somewhat higher than a previously published version.^{146,147} It is based on an effective displacement energy of 40 eV, which is recommended for iron. For low-energy neutrons, the displacement cross section varies as $E^{-0.5}$ from a value of 17 b at 0.025 eV (Ref. 146).

The 50-energy-group structure shown in Fig. 7-52 is compatible with the output from TART, the multigroup neutron-transport code used to calculate the neutron fluence.¹⁴⁸ All of the neutronics calculations were carried out in a one-dimensional, spherical geometry. In all cases, the FSW was a 2-cm-thick shell of iron ($\rho = 7.86 \text{ g/cm}^3$) located 5 m from the neutron source.

Figure 7-53 defines the eight blanket configurations studied: four unprotected and four protected iron walls. Case 1 is simply a 14.1-MeV point source without a

tritium-breeding blanket or neutron reflector. In cases 2 through 8, the 14.1-MeV neutron source is uniformly distributed throughout a region of compressed DT which has a ρ of 3 g/cm^3 . Case 3 adds a 1-m-thick lithium blanket ($\rho = 0.49 \text{ g/cm}^3$) outside the FSW. Natural lithium (7.42 at.% ^6Li and 92.58 at.% ^7Li) is used. In case 4, the 1-m-thick lithium blanket is replaced by a 30-cm-thick graphite reflector ($\rho = 1.7 \text{ g/cm}^3$). Case 5 has a 2-m-thick lithium blanket between the target and the FSW. This region is at one-half normal density (0.245 g/cm^3) to represent the 50% packing fraction of lithium jets within the HYLIFE chamber. This gives an effective thickness of 1 m of lithium protection. The inner radius of the lithium region is 0.5 m. Case 6 adds a 1-m-thick lithium blanket at full density outside the FSW. Case 7 replaces the outer lithium blanket with a 30-cm-thick graphite reflector. In case 8, the protective inner lithium blanket is moved outward so that its inner radius is 2.5 m and its outer radius is 4.5 m.

Results. Figure 7-54 compares the displacement damage rates for the four unprotected walls. These results are based on a fusion power of 3000 MW, a 5-m radius FSW, and a 100% capacity factor (i.e., the results are per full-power year).

For a 14.1-MeV source incident on an unprotected, unreflected FSW (case 1), the displacement damage rate is 35.6 dpa/y. The moderating effects of the compressed DT in the fusion target (case 2) only slightly reduce the damage rate. Adding a lithium blanket outside the FSW (case 3) raises the dpa rate by 50%. This increase is the result of two factors: first, fusion neutrons are scattered back into the wall from the lithium region; second, neutrons emitted by $^7\text{Li}(n,n'\text{T})\alpha$ reactions can also impinge on the FSW. Comparing cases 3 and 4, we see that the graphite reflector results in a slightly higher damage rate, indicating that more neutrons are directed back at the FSW.

Figure 7-55 compares the damage rates for the four protected walls. Note that in all cases there is nearly an order of magnitude reduction from the unprotected walls. Cases 5, 6, and 7 show the same trends as cases 2, 3, and 4. In particular, adding a lithium blanket outside the FSW increases the damage rate by 45%. Substituting a graphite

Fig. 7-54. Displacement damage rates in the unprotected iron walls.

Fig. 7-55. Displacement damage rates in the protected iron walls.

reflector for the outer lithium blanket (case 7) gives an even higher dpa rate.

Case 8 demonstrates an interesting effect. Moving the inner lithium blanket closer to the FSW reduces the damage rate by nearly 30%. One reason for this is that moving the blanket closer to the wall results in a larger effective thickness for neutrons that are scattered at least once. This was discussed and illustrated in Ref. 149. Another reason is that neutrons that are reflected back through the FSW are more likely to be absorbed or further moderated by the inner lithium blanket. In other words, a neutron reentering the fusion chamber is more likely to hit lithium than to traverse the vacuum chamber and strike the wall again. In particular, for a neutron reentering the chamber, the solid-angle fraction eclipsed by the 2.5-m-radius lithium blanket (case 7) is only 13%, whereas that eclipsed by the 4.5-m-radius blanket (case 8) is ~56%.

While moving the blanket outward reduces displacement damage, it would also significantly increase the lithium flow rate in a chamber such as HYLIFE. In the Cascade chamber, a solid-granule breeding blanket is held against the inside of the FSW by centrifugal force.¹⁵⁶ Assuming that the trends observed for lithium hold for breeding-blanket materials such as Li_2O and LiAlO_2 , the Cascade blanket will effectively minimize the displacement damage rate in the rotating chamber wall.

Summary. We examined the dependence of displacement damage in the FSW on the configuration of the fusion chamber blanket. We find that:

1. The compressed DT in the ICF target reduces the dpa rate by less than 10%.
2. Adding a lithium blanket outside the FSW increases the displacement damage rate by ~50%. This is true for both the unprotected and protected FSW.
3. Adding a graphite reflector outside the FSW increases the dpa rate by 66% for an unprotected wall and 73% for a protected wall.
4. Placing the equivalent of 1 m of lithium between the fusion target and the FSW decreases the damage rate by nearly an order of magnitude.
5. Moving the protective, inner blanket closer to the FSW significantly reduces the damage rate.

Author: W. R. Meier

Susceptibility of 2-1/4 Cr-1 Mo Steel to Liquid-Metal Embrittlement

Introduction. An ongoing investigation that provides the basis for this article considers the embrittlement of the pressure-vessel steel to be used in the construction of liquid-metal-cooled fusion-energy reactors. Several proposed designs for ICF reactors utilize liquid lithium contaminated by 1-at.% lead ($\text{Li}_{99}\text{Pb}_1$) as a combined heat-transfer medium, tritium-breeding medium, and radiation shield. Containment of the liquid metal requires thorough understanding of any tendency for either liquid lead or lithium to embrittle conventional pressure-vessel steels.

The embrittlement of normally ductile metals subjected to stress and exposed to

liquid-metal environments has been known for nearly 70 years.¹⁵¹ Liquid-metal-induced embrittlement (LMIE) of metals has been investigated for many ferrous-metal-liquid-metal systems, although a consistent experimental procedure yielding easily correlatable data for separate systems has yet to be developed. A series of experiments performed at the Colorado School of Mines showed that the LMIE of 2-1/4 Cr-1 Mo can indeed occur. As shown below, the intensity of the LMIE phenomenon varies with the microstructural condition of the steel. Several experimental tests have been used to characterize the LMIE phenomenon, including tensile elongation, ultimate tensile strength, and fracture-surface appearance. Although there are no standard embrittlement tests, these three were chosen because they appear to characterize the embrittlement.

Characterization of LMIE. A survey of the literature on LMIE of several liquid/solid metal couples (both ferrous and nonferrous) reveals some striking similarities. Most of these effects have been observed in tests performed in this study.

1. A sharp decrease in ductility occurs at or near the melting temperature of the liquid metal. As temperature drops, tensile elongation decreases to a minimum value near the melting temperature of the liquid metal bath. (Certain metal couples even show solid-metal-induced embrittlement¹⁵² in which one metal embrittles another at a temperature just below the melting point of the embrittling metal.)
2. The ductility minimum extends over a narrow temperature range, and the ductility increases sharply at temperatures even nearer the melting point of the liquid. The width of this temperature range varies with both the liquid/solid system and the imposed strain rate.¹⁵²
3. A transition of the fracture mode from transgranular or quasi-cleavage fracture to distinctly intergranular fracture is observed when the testing temperature is in the embrittlement susceptibility range.
4. Intimate solid/liquid contact is required to adequately supply the crack tip with the embrittling metal. Crack propagation stops if the supply of the embrittling agent ceases, as in freezing.^{153,154}
5. High-strength materials are more severely embrittled. In practice, reducing the effects of embrittlement reduces the

strength; this is true whether the reduction is achieved by heat treatment or by choosing a more ductile material.

The Embrittlement Mechanism. Among the proposed mechanisms for LMIE is the absorption-induced reduction-in-tensile-cohesion model of Stoloff and Johnston¹⁵⁴ and Westwood and Kamdar.¹⁵⁵ The model proposes that chemisorption of the liquid metal results in a reduced bond strength of atoms at stress-concentration locations in the solid, such as at a crack tip or at the head of dislocation pileups near surface obstacles. Extensive reviews of this model have been presented by Kamdar¹⁵⁶ and Old.¹

Roberts¹⁵⁷ proposed that the embrittling liquid enhanced the dissolution of material at the crack tip, thus causing crack extension and propagation.¹⁵⁸ This hypothesis is not consistent with the observation of an incubation period. Furthermore, the dissolution theory suggests that base-metal solubility in the embrittling metal is required for embrittlement, but solubility is in fact not always observed.

Other, less thoroughly developed embrittlement theories are sufficient to predict the initiation and propagation of a crack once the embrittling metal has reached the crack tip. Decreased tensile elongation of the embrittled metal with increasing test temperatures within the embrittlement-susceptibility range is thought to be due to thermally activated wetting. This type of wetting occurs just above the melting temperature of the liquid. Localized embrittlement occurs as soon as the liquid metal reaches the crack tip. A sharp return in ductility at higher temperatures may reflect an increased ability of the metal to plastically deform, leading to a relief in the stress and blunting of the crack tip.

Experimental Procedure. Coupons of 2-1/4 Cr-1 Mo steel plate were heat treated according to the specifications given in Table 7-25. The temper parameter TP in the table is defined as

$$TP = T \log t + 20 \quad (86)$$

where T is the tempering temperature in kelvins and t is the tempering time in hours. Material austenitized at 1300°C and oil-quenched (Heat Treatment 1 or H.T. 1) has been shown in early work at the Colorado School of Mines to simulate the

Process	Condition	Comments	TP (K)
H.T. 1	Austenitized at 1300°C and oil-quenched	Simulates as-welded heat-affected zone	
H.T. 2	H.T. 1, plus 740°C temper for 10 h	Represents PWHT ^a at maximum ASME code-allowable temperature	21 300
H.T. 3	H.T. 1, plus 695°C-temper for 121 h	Represents PWHT at minimum code-allowable temperature at approximately same TP as H.T. 2	21 400
H.T. 4	H.T. 1, plus 695°C temper for 15.7 h	Represents minimum PWHT	20 500

^aPWHT = post-weld heat treatment.

lithium-corrosion susceptibility near the heat-affected zone of a weld in 2-1/4 Cr-1 Mo steel. As-quenched material tempered at 740°C for 10 h (H.T. 2) simulates a weld given a severe post-weld heat treatment at the maximum temperature. Commercial fabricators would be concerned that such a heat treatment might excessively soften the surrounding base plate. Tempering as-quenched material for 121 h at 695°C (H.T. 3) produces approximately the same temper parameter as the H.T. 2 process, but at the minimum code-allowable temperature. Tempering as-quenched material for 15.7 h at 695°C (H.T. 4) simulates a minimum post-weld heat treatment. These tempering heat treatments give microstructures primarily consisting of $M_{23}C_6$ carbides homogeneously distributed in ferrite laths. A given temper parameter is often considered to produce unique mechanical properties, but changing the tempering temperature for a fixed temper parameter should alter the carbide distribution and affect the LMIE susceptibility.

Round tensile specimens were machined from the heat-treated coupons and were then tested in tension to failure at a constant engineering strain rate of 0.002 min^{-1} while submerged in a liquid-metal bath maintained at constant temperature. Mechanical properties for the steel in the liquid-metal environment were determined from the tensile test load-tension data.

Round tensile specimens of 2-1/4 Cr-1 Mo steel were subjected to pure lead, pure lithium, and $\text{Li}_{99}\text{Pb}_1$ liquids. The experiments with pure lead indicated that the lead was not wetting the steel, and these experiments were repeated using ZnCl_2 as a fluxing agent to force wetting. In this portion of the investigation, two heat treatments were

Table 7-25. Experimental heat treatments for 2-1/4 Cr-1 Mo steel LMIE studies.

Fig. 7-56. Ductility of the heat-affected zone of 2-1/4 Cr-1 Mo steel in liquid metals.

compared: the oil-quenched condition (H.T. 1) and the quenched-and-tempered (740°C) condition (H.T. 2).

The ductility results for each heat treatment and all of the four liquids are shown in Fig. 7-56. From the figure, a number of observations can be made.

First, steel specimens tested in pure molten lead have mechanical properties that change with temperature in the manner expected for nonembrittled, unwetted steel. The tensile elongation of the tempered steel increases with increasing temperature, reflecting the reduction of the base-metal flow stress. The as-quenched steel exhibits significant ductility throughout the temperature range.

Secondly, when a fluxing agent, $ZnCl_2$, is added to the lead bath, the steel is wetted by the lead, and liquid-lead embrittlement is observed. The ductility of the as-quenched steel decreases by nearly 70% throughout the temperature range, and the

tempered steel exhibits a ductility minimum near 400°C, somewhat above the 328°C melting point of lead.

Next, a sharp decrease in ductility is observed in pure lithium, for both heat treatments. The as-quenched material exhibits no elongation, and the tempered steel exhibits a sharp decrease in ductility near 250°C, somewhat above the 181°C melting point of lithium. This ductility trough illustrates the effect of embrittlement by lithium, as described earlier. The smaller embrittlement trough near 450°C has not been satisfactorily explained, but additional data confirm its existence.

Finally, both heat treatments also give rise to a sharp decrease in ductility in $Li_{99}Pb_1$. The as-quenched material exhibits no elongation, and the tempered steel exhibits a sharp decrease in ductility near 250°C, somewhat above the 177°C melting point of $Li_{99}Pb_1$. For the tempered steel, the embrittlement was somewhat less severe for

Fig. 7-57. Strength of the heat-affected zone of 2-1/4 Cr-1 Mo steel in liquid metals.

Li₉₉Pb₁ than for pure lithium, and there was no high-temperature embrittlement trough.

The measurements of ultimate tensile strength for both heat treatments and all four liquids are shown in Fig. 7-57.

For the nonwetting lead, both the as-quenched and the tempered steels show the expected decrease in strength with increasing temperature. The oil-quenched steel is considerably stronger than the tempered steel. By comparison,¹⁵⁹ spray-quenched 2-1/4 Cr-1 Mo steel exhibits an intermediate 810-MPa ultimate tensile strength at 300°C that decreases to 670 MPa at 600°C. The steel strength levels for both heat treatments were not significantly altered by liquid lead with a ZnCl₂ wetting agent.

The strength characteristics of the steel were significantly altered by pure liquid lithium and by Li₉₉Pb₁. The tempered-steel strength decreased in the embrittlement temperature range. The as-quenched steel strength fell well below the

tempered-steel values through most of the temperature range. Early fracture caused by the complete loss of ductility resulted in loss of strength.

The change in fracture modes is illustrated in SEM fractographs of fractured steel (Fig. 7-58). Three tempered steel samples were immersed in liquid Li₉₉Pb₁ baths at three temperatures: slightly above the liquid-metal melting point, at the ductility minimum, and at a higher temperature where embrittlement is insignificant. In the absence of LMIE, it is expected that the fracture mode will change continuously from entirely brittle, transgranular failure at very low temperatures, to a quasi-cleavage fracture at intermediate temperatures, and finally to a completely ductile tearing at higher temperatures. However, as shown in Fig. 7-58, LMIE of 2-1/4 Cr-1 Mo steel by Li₉₉Pb₁ causes intergranular fracture. Even at temperatures below the temperature of minimum ductility (190°C),

Fig. 7-58. Embrittlement of the heat-affected zone of 2-1/4 Cr-1 Mo steel by liquid $\text{Li}_{99}\text{Pb}_1$.

Fig. 7-59. Effect of tempering on embrittlement of the heat-affected zone of 2-1/4 Cr-1 Mo steel by liquid $\text{Li}_{99}\text{Pb}_1$.

partial intergranular fracture is observed [Fig. 7-58(a)]. At the temperature of minimum ductility (234°C), LMIE resulted in completely brittle, intergranular failure, as shown in Fig. 7-58(b). Only at temperatures (~450°C) well above the temperature of minimum ductility is the dimpled fracture surface characteristic of ductile rupture observed [Fig. 7-58(c)]. Similar fractographs were produced for exposure of tempered steel in a pure lithium bath.

The similarity of the lithium and $\text{Li}_{99}\text{Pb}_1$ data and fractographs indicates that lithium was the primary embrittling agent in the $\text{Li}_{99}\text{Pb}_1$ metal, and that the embrittling agent weakened intergranular bonds. Compositional or microstructural differences present at prior austenitic grain boundaries are likely to play an important role in the LMIE phenomenon; however, these effects are neither characterized nor understood.

The concluding investigation of this study was an evaluation of how variations in the temperature and time of tempering affected the LMIE-susceptibility of 2-1/4 Cr-1 Mo steel when subjected to liquid $\text{Li}_{99}\text{Pb}_1$. Tensile specimens for all four heat treatments specified in Table 7-25 are compared in Fig. 7-59, where percent elongation is plotted against test temperature. A low tempering temperature was found to be superior to a high temperature in reducing the LMIE susceptibility. Tempering at 695°C for only 15.7 h produced less LMIE effect than tempering at 740°C for 10 h. Tempering at 695°C for 121 h (a temper parameter equivalent to 740°C for 10 h) produced the least LMIE effect observed. Microdeformation mechanisms at the crack tip can be very sensitive to microstructure and are probably responsible for the LMIE differences shown in Fig. 7-59. An investigation of the subtle differences in microstructure caused by

the heat treatments specified in Table 7-25 is in progress.

Summary. This work has led to the following observations:

- LMIE occurs in the heat-affected zones near welds in 2-1/4 Cr-1 Mo steel through the action of both lithium and lead.
- Lithium is the primary embrittling element in the $\text{Li}_{99}\text{Pb}_1$ alloy.
- Embrittlement of 2-1/4 Cr-1 Mo steel by lithium results in intergranular fracture; the mechanistic details of the embrittlement are still being studied.
- Embrittlement susceptibility of 2-1/4 Cr-1 Mo steel weld heat-affected zones by liquid $\text{Li}_{99}\text{Pb}_1$ can be reduced by heat treatment. Temper parameters in excess of 21 000 K may be required.
- For a given TP, low tempering temperatures are superior to high temperatures in reducing LMIE susceptibility.

Authors: B. A. Mullineaux, B. A. Eberhard, and G. R. Edwards (Colorado School of Mines); J. A. Blink (LLNL)

Major Contributor: J. Hovingh (LLNL)

Waste Disposal

A major goal of current fusion design efforts is to find low-cost techniques for fusion waste disposal that ensure occupational and public health.¹⁶⁰ Thus, material selection is based on considerations of waste management, occupational exposure, and accidental-release hazard, as well as on technical performance. It is likely that fusion wastes will be regulated under the same institutional conditions as for fission wastes. The NRC regulation 10CFR61, "Licensing Requirements for Land Disposal of Radioactive Waste,"¹⁶¹ is a logical starting point for examining the effects of materials selection on fusion waste management.

We believe that the fundamental bases for 10CFR61 are excellent and that they are pertinent to fusion as well as to fission wastes. Fusion waste will be more completely characterized than many of the fission waste streams, and surrogate nuclides need not be used to set limits for a variety of other nuclides (in contrast to the 10CFR61 procedure for fission products).

Most of the fusion activation is embodied in the first wall. Dilution of that material waste with the lower activity blanket, shield, and other structure, as well as with grout, will reduce the average activity (which must meet the NRC limit), to 2 to 3 orders of magnitude below the peak (first wall) activity. This dilution is neither explicitly allowed nor forbidden in 10CFR61, but the implication of the regulation is that it will be allowed. Disposal of nuclear submarine steel under the regulation has used the diluted volume to calculate the specific activity (in Ci/m^3) of waste.¹⁶²

The NRC has made provision for reconsideration of waste limits in section 61.58 of the regulation. This section allows alternate limits if the performance objectives of the regulation (dose limits to intruders and the public) can be met. Thus, it is appropriate to consider revision of the limits for fusion waste. To this end, we used the INVERSI and INVERSW codes¹⁶³ (developed by the NRC for 10CFR61) to determine the allowable concentrations of one nuclide, ^{94}Nb , in undiluted fusion reactor waste. Our logic can be applied to other nuclides expected in fusion waste, including nuclides not explicitly considered by 10CFR61.

Scenarios Used to Establish Concentration Limits. In 10CFR61, several scenarios are used to ensure that the dose to an intruder or the general public is below set limits. These scenarios include wells dug by an intruder on the site, near the site boundary, or at a greater distance; and home construction or agriculture by an intruder on the site. Each scenario considers dose through one or more pathways, including direct gamma, air (inhalation of a dust cloud, deposition on garden plants, and gamma from the cloud), food, and water. Six organs and the whole body are considered, and the most restrictive limit is used.

In some cases, the scenarios are modified. For example, if the waste is stable, an intruder building a home 100 y after dump closure is assumed to recognize the waste; this reduces his exposure time from 500 to 6 h. Another example is the elimination of the wind/water (but not the intruder) erosion scenarios if the dump site is a hot-waste facility; the most expensive option in 10CFR61.

General Studies and Reactor Development Technologies

Table 7-26. Concentration limits^a (Ci/m³) for ⁹⁴Nb in fusion wall waste disposed in a hot-waste facility.

Scenario	Hot-waste facility
Intruder well	4×10^8
Boundary well	5×10^{12}
Population well	7×10^{17}
Intruder construction	1.3
Intruder agriculture	1.1
Erosion—water	none
Erosion—wind	none
Intruder erosion—water	4×10^4
Intruder erosion—wind	1×10^8
Accident—container	1×10^7
Accident—fire	none
Hot-waste facility limit	1.1

^aThe assumptions used in calculating the limits are listed in Ref. 160.

The result of applying the NRC's codes to ⁹⁴Nb is a limit of 1.1 Ci/m³ (Table 7-26). The 10CFR61 limit for ⁹⁴Nb is 0.2 Ci/m³. We attribute the difference to our use of a hot-waste facility and our specification of stable, structural waste.

Implications of the Limits. For ferritic steels in both ICF and MFE geometries, ⁹⁴Nb dominates the waste at 50 y after shutdown. This nuclide is produced from niobium (thermal neutron activation of natural ⁹³Nb) and molybdenum (fast neutron reactions on ⁹⁴Mo and ⁹⁶Mo, which account for ~1/4 of natural molybdenum). If the 0.2-Ci/m³ limit on ⁹⁴Nb is applied, niobium must be limited to ~0.8 weight ppm (wppm) for undiluted first-wall waste for both ICF and MFE.^{164,165} The molybdenum limit would be ~60 wppm for MFE and ~700 wppm (0.07 wt%) for ICF.^{164,165} (See also following article.) If both molybdenum and niobium are present, the limits must be reduced proportionately, i.e.,

$$\frac{C_{\text{Mo}}}{L_{\text{Mo}}} + \frac{C_{\text{Nb}}}{L_{\text{Nb}}} < 1, \quad (87)$$

where C is the concentration and L is the limit.

These limits pose a problem for fusion wall designers. It appears to be difficult to produce any steel with a niobium concentration below ~10 wppm. Further, steel designers would like the freedom to include up to 1 wt% Mo (10 000 wppm) in the steel to improve its structural properties. Although the molybdenum limit can be circumvented by removing the worst of the seven natural isotopes, there is only one natural isotope of niobium.

The conflict between design requirements and the 10CFR61 limits led us to examine the scenarios used to determine the limits. Table 7-26 shows that for ⁹⁴Nb in a hot-waste facility, the only significant scenarios are intruder construction and agriculture. In the construction scenario, an intruder digs a 3-m-deep foundation on the dump site for a house at a specified time after the waste was buried. The lower 1 m of the hole is assumed to be waste in some cases. The excavated earth is spread over a 25-m-radius circle around the house. In the agriculture scenario, a family garden is planted and maintained after the construction scenario is complete. We note that, for a hot-waste facility, the 10CFR61 limit would be raised to 1.1 Ci/m³. The published limit, 0.2 Ci/m³, was derived for a minimal facility with fewer barriers.

Waste Classification Procedure for the Intruder Agriculture and Construction Scenarios. The NRC has developed a logical procedure for determining the disposal configuration (regular, layered, or in a hot-waste facility) for the intruder-dominated nuclides. This procedure is illustrated in Fig. 7-60 for physically stable waste. Using the INVERSI code, we determined the specific activity limit for each of the tests depicted. The tests represent increasingly expensive treatments of the waste, and thus allow near-surface disposal of wastes with increasingly higher concentrations of the radioactive isotopes.

Before beginning the test procedure, we must first determine that the waste is physically stable and segregated from physically unstable waste (such as filter cartridges). Fusion first walls are clearly stable, and therefore the use of modified tests at early times (100 y) is permitted. However, the 500- and 1000-y tests allow no credit for waste stability.

The first test is the "Regular-Modified Test at IIC Years," where IIC is the time after closure when institutional control of the dump site is lost and the site markings are assumed to be no longer recognizable. The regulation uses 100 y as the value of IIC. The test assumes that an inadvertent intruder begins to excavate a home foundation at the dump. After 6 h of exposure, he recognizes that he is in waste rather than earth and terminates construction. The test

Fig. 7-60. Limit setting procedure for the construction and agricultural scenarios.

result is the concentration of each isotope (in Ci/m^3) that will result in the maximum permissible dose to the intruder, 500 mR/y.

The second test is the "Regular-Standard Test at 500 Years." In this test, the waste is assumed to have decomposed and mixed with the surrounding earth. The intruder spends 500 h in contact with the waste, not recognizing that he is constructing the house in an old radioactive waste dump. The test assumes that the bottom 1 m of the 3-m-deep excavation is waste, even if thicker engineered covers were installed at dump closure. It also gives no credit for layering the waste.

The third test, the "Layered-Modified Test at 100 Years," again assumes that the home builder spends 6 h in the excavation before noticing that this is an old dump. However, the waste disposal site is different in that the highest activity waste has been placed at the bottom of the dump with the less active waste above it. This procedure results in the intruder being subjected to less radiation than in the regular (unlayered) case. Thus, the limit from this scenario will be higher than for the regular modified test.

The "Hot Waste Facility Test at 100 Years" assumes that the waste is layered, grouted, and enclosed in a concrete vault with at least 5 m of cover. The concrete vault has 0.33-m-thick concrete sides and bottom and a 1-m-thick concrete top. The cover is usually an engineered barrier of clay and other materials. A final 1 m of fill

completes the facility. The direct gamma pathway of the agricultural scenario (in which the homebuilder has a family garden in the distributed waste) is not considered in the hot-waste-facility test. For the direct gamma pathway of the construction scenario, a reduction factor of 7×10^{-7} is applied to account for shielding, and a factor of 10^{-2} is applied to account for grouting and reduced accessibility. The effective removal of the radioactive waste from the biosphere by the hot-waste facility results in a limit that is higher than any of the other situations.

Finally, the "Regular Test at 1000 Years" assumes that the dump has had enough degradation that it is indistinguishable from the surroundings. The direct gamma pathway has a reduction factor of 1/10 to account for dilution during decomposition. While the additional time may allow short-lived material to decay, the test is difficult to meet for long-lived isotopes because of the lack of discovery (500 h of exposure during excavation), coupled with ignoring the engineered cover, the concrete vault, and the layering of the waste.

The results of applying the various intruder tests to hot-waste-facility waste containing ^{94}Nb are shown in Table 7-27. As Fig. 7-60 shows, a hot-waste facility must meet only the final two tests. For ^{94}Nb , the direct gamma exposure is the dominant pathway for both the construction and agricultural scenarios. The regular standard test at 1000 y sets the limits. This test assumes

Table 7-27. Intruder test results (^{94}Nb limits in Ci/m^3).

General Studies and Reactor Development Technologies

Test	Construction			Agriculture			
	Direct gamma	Air	Combined	Direct gamma	Air	Food	Combined
Regular modified, 100 y	103	2×10^8	103	none ^a	none ^a	none ^a	none ^a
Layered modified, 100 y	1.2×10^6	2×10^8	1.2×10^6	none ^a	none ^a	none ^a	none ^a
Regular standard, 500 y	1.2	2×10^6	1.2	1.0	6×10^6	4×10^7	1.0
Hot waste facility, 100 y	2×10^7	2×10^7	1×10^7	none ^a	none ^a	none ^a	none ^a
Regular standard, 1000 y	1.3	2×10^5	1.3	1.1	6×10^5	2×10^7	1.1

^aNo agriculture takes place due to discovery of the waste.

that waste identity is lost and that the waste is within 2 m of the surface (so intruders building homes excavate it). For the hot-waste facility, 1 m of earth, 5 m of engineered intruder barrier, 1 m of concrete cap, the lower activity layers and the grout filler are all ignored in this test. Since we have assumed a hot-waste facility for fusion waste, the Table 7-27 limit of $1.1 \text{ Ci}/\text{m}^3$ ^{94}Nb would replace the 10CFR61 limit of $0.2 \text{ Ci}/\text{m}^3$.

If the fusion wall (a few cubic meters of volume) is diluted by the remainder of the blanket and shield, the volume increase is about a factor of 1000. Thus, the limit for first walls distributed in other fusion waste in a hot-waste facility would be $1100 \text{ Ci}/\text{m}^3$ waste. This increased limit meets the performance objectives of 10CFR61 and thus should be justifiable under the provisions of section 61.58; however, it has not been explicitly authorized in the regulation.

Alternatively, if as little as 1 m of shielding remains intact at the time of discovery that a dump is being excavated, the dose will be more than three orders of magnitude less; i.e., the limit for undiluted waste could also be set three orders of magnitude higher, to $1100 \text{ Ci}/\text{m}^3$. However, the fundamental assumption of 10CFR61 is that all barriers will eventually be breached and that home construction in the waste will begin. Thus, assumption of any barrier retention or of the recognizability of the waste is contrary to the premise of the regulation. We believe, however, that the barriers can be marked at various depths with extremely durable signs that would warn a potential intruder that waste is stored below. With a limit of $1100 \text{ Ci}/\text{m}^3$, a steel containing 5% molybdenum and 0.4% niobium would be acceptable.

Discussion. Some comments are necessary on the likelihood that 10CFR61 limits

will be changed. Because the NRC methodology used in setting the limits produces a ^{94}Nb limit of $1.1 \text{ Ci}/\text{m}^3$ for a hot-waste facility, we believe that relaxation of the current $0.2 \text{ Ci}/\text{m}^3$ limit to this level is quite likely. Similarly, the dilution argument is not prohibited by the regulation, and it appears to be considered valid in the submarine waste example. On the other hand, assuming that any barriers remain intact at long times is directly counter to the regulation's scenarios, and it will be very difficult to make changes based on this assumption, even if they appear to be technically valid. Finally, if fusion waste does not meet 10CFR61 limits, it will always be possible to separate the hot waste (first wall) and dispose of it in the more expensive high-level-waste (geologic) repositories.

Author: J. A. Blink

Major Contributors: D. W. Dorn and R. C. Maninger

Steel Type and Impurity Effects on Activation

In 1982, we compared¹⁶⁶ the activation of PCA (stainless steel) with several ferritic steels. In this article, we report on our extension¹⁶⁷ of the previous study. We have added a new steel (UCVS-1) to the three previously considered ferritic steels (HT-9 and both thermally and niobium-stabilized 2-1/4 Cr-1 Mo). An additional figure of merit (surface dose rate) is added, and the potential for improvement by isotopic tailoring is evaluated. All of these calculations were done for both MFE and ICF geometries.

Activation Geometry. Four ferritic steels (Table 7-28) were compared using two simplified one-dimensional fusion reactor

geometries. In the MFE geometry, a 14.1-MeV point source was centered inside a 5-m-radius vessel. The 2-cm-thick steel first wall was surrounded by a tritium-breeding blanket and neutron reflector consisting of 78 cm lithium, 18 cm graphite, 2 cm B₄C, and a 10-cm-thick steel outer wall. In the ICF geometry, the neutron source was in a compressed plasma ($\rho = 3 \text{ g/cm}^3$), the region between 0.5 and 2.0 m was filled with lithium at 50% of the natural density, and the same wall and reflector were used as in the MFE reactor. Using more accurate geometries would not significantly affect this study which compares materials. For an estimate of the effects of using more accurate geometries, see Ref. 168.

Calculation Procedure. The TART neutronics code¹⁶⁹ was used to calculate the energy dependent neutron flux in the first wall. Then, the ORLIB code¹⁷⁰ was used with the flux as a weighting fraction to collapse the energy dependent ACTL cross-section library.¹⁷¹ (ACTL was expanded in 1981 by R. Howerton to allow treatment of fusion-reactor structural materials.) The resulting energy-averaged cross sections for each geometry were used in a series of activation calculations using the FORIG code.^{172,173} A separate calculation was done for each of the 16 elements in the steels (Table 7-28), using both the MFE and the ICF fluxes. In the MFE calculations, the neutron flux was on continuously for 5 y at an average of 70% of 1 GW_e power; in the ICF calculations, 30 y of operation was assumed. Finally, the results were assembled for each of the eight cases (four steels, in both MFE and ICF geometries) by linearly combining the results of the elemental activation calculations, using the steel composition as the weighting function.

Figures of Merit. Four figures of merit were used to compare the steels. The first two, afterheat and biological hazard potential, deal with accident safety. The third evaluates the potential feasibility for shallow land-burial disposal of the activated wall. The final figure of merit, surface dose rate, influences the hands-on maintainability of the structure.

Afterheat is the heat generated by residual radioactivity after the reactor is shut down. If the decay heat is less than a few megawatts (in these geometries), radiation-cooling alone could remove the afterheat¹⁷⁴,

Element	HT-9	UCVS-1	2-1/4 Cr-1 Mo	Nb-stabilized 2-1/4 Cr
C	0.19	0.11	0.07	0.08
N	---	0.015	---	---
Al	---	0.043	0.01	---
Si	0.31	0.3	0.35	0.2
P	0.021	0.007	0.005	0.03
S	0.008	0.015	0.013	0.03
Ti	---	0.003	---	---
V	0.31	1.5	---	---
Cr	11.8	2.46	2.14	1.9
Mn	0.57	0.3	0.46	0.3
Fe	84.741	95.137	96.002	95.79
Ni	0.48	0.05	---	---
Cu	---	0.04	---	---
Nb	---	---	---	0.8
Mo	1.04	0.02	0.95	0.87
W	0.53	---	---	---

at higher afterheat levels, emergency wall-cooling systems will be required. Because afterheat is most important shortly after shutdown, we present the results at 1 min and 10 h after shutdown.

The inhalation bhp is the amount of air required to dilute the radioactive nuclides to breathing standards if the entire structure is vaporized and released. Typical bhp's are $\sim 10^{17} \text{ m}^3$, which are extremely large numbers. However, it is unrealistic to expect the entire wall to be vaporized and released from the containment building; release fractions of order 10^{-6} have been predicted in some fusion accident analyses.¹⁷⁵ Also, if the wall is in contact with heated air, some alloying elements (e.g., molybdenum) may be preferentially oxidized from the structure prior to vaporization. Our simple comparison assumes that the hazard is due to vaporization and release of 10^{-6} of the activated structure, and that there is no preferential release of oxidizable alloying elements. Since release of the wall is least improbable shortly after shutdown, we present the results at 1 and 10 h after shutdown.

To evaluate the potential for disposal of the activated structure by shallow land burial (as low-level waste), we have used the procedure¹⁶⁷ of calculating the amount of water required to dilute the activated (and dissolved) steel to drinking standards (radioactive ingestion bhp), and then comparing that dilution volume to the water volume required to dilute clean, unactivated steel to drinking standards (using Environmental Protection Agency chemical limits). When the ratio of the two dilution volumes falls to 1.0, the radioactive and chemical hazards of the steel are similar. A more

Table 7-28. Compositions of the four steels (wt%).

current approach is described in the previous article. Since utilities could afford to hold decommissioned structures for decades, but not centuries, we present the waste disposal ratio at times of 50 and 100 y after plant shutdown. For the MFE geometry, the disposal ratios include the total ingestion bhp's for all six inner walls used during a 30-y plant lifetime. However, the denominator of the ratio is the chemical ingestion bhp for only a single wall (to allow direct comparison of the ICF and MFE radioactive wastes to a single chemical standard).

The final figure of merit is a very rough calculation of the dose rate to an unprotected person at the wall surface, using the procedure suggested by Dorn.^{17b} Because maintenance cannot be delayed indefinitely, we present the dose rates at 1 d and 1 mo after shutdown. Nuclear power industry experience indicates that

values above 100 R/h will preclude hands-on maintenance.

Results for the Four Steels. The four figures of merit are presented in Table 7-29 for the four steels in both MFE and ICF geometries. For afterheat, bhp, and dose rate, the differences between the steels are not significant, but the ICF geometry is 5 to 10 times better than the MFE geometry.

For waste disposal, the difference between geometries is much less significant. In fact, if each of the six MFE walls were compared separately to the single ICF wall, the MFE wall hazard would be reduced by a factor of ~6, and it would be slightly less hazardous than the ICF wall. The waste disposal results vary significantly among the four steels. The UCVS-1 steel is acceptable in either geometry, while the other three ferritic steels are within an order of magnitude of being acceptable. For

Table 7-29. Activation of the four ferritic steels.

	Afterheat (MW)		bhp ^a (km ³)		Waste disposal index		Dose rate (kR/h)	
	1 min	10 h	1 h	10 h	50 y	100 y	1 d	1 mo
<u>ICF geometry</u>								
HT-9	1.07	0.149	19.7	18.2	2.04	1.95	110	68.3
UCVS-1	0.689	0.116	17.9	16.6	0.138	0.104	104	66.0
2-1/4 Cr-1 Mo	0.687	0.127	18.9	17.5	1.65	1.63	107	67.0
Nb Stab 2-1/4 Cr	0.681	0.148	25.2	23.7	5.09	5.07	153	91.5
<u>MFE geometry</u>								
HT-9	5.55	0.942	180	168	5.89	5.60	809	711
UCVS-1	5.72	0.945	157	145	0.798	0.592	799	704
2-1/4 Cr-1 Mo	5.73	0.951	162	149	4.75	4.70	802	728
Nb Stab 2-1/4 Cr	5.64	0.942	226	211	8.52	8.42	826	715

^aBased on a release fraction of 10^{-6} .

Table 7-30. Top nuclide contributors to the figures of merit (in percent of the totals).

	Afterheat (MW)		bhp (km ³)		Waste disposal index		Dose rate (kR/h)	
	1 min	10 h	1 h	10 h	50 y	100 y	1 d	1 mo
<u>ICF geometry</u>								
HT-9	Fe 45.0	Fe 63.6	Fe 76.6	Fe 77.5	Mo 88.7	Mo 91.8	Fe 79.9	Fe 82.4
	Cr 36.0	W 15.9	W 6.8	W 6.6	Ni 11.2	Ni 8.0	Ni 7.7	Ni 9.8
UCVS-1	Fe 78.7	Fe 91.4	Fe 95.0	Fe 95.8	Cu 55.4	Cu 50.0	Fe 94.9	Fe 95.8
	Mn 9.3	Mn 4.9	Mn 2.6	Mn 2.6	Mo 25.2	Mo 33.1	Mn 2.1	Mn 2.9
2-1/4 Cr-1 Mo	Fe 79.5	Fe 84.3	Fe 90.5	Fe 91.5	Mo 99.9	Mo 99.9	Fe 93.1	Fe 95.2
	Mn 14.3	Mo 8.1	Mo 5.1	Mo 4.9	Fe 0.1	Fe 0.1	Mo 3.7	Mn 4.4
Nb Stab 2-1/4 Cr	Fe 80.0	Fe 67.7	Fe 67.9	Fe 67.5	Nb 70.3	Nb 70.4	Fe 64.9	Fe 69.5
	Mn 9.4	Nb 22.0	Nb 25.9	Nb 27.0	Mo 30.0	Mo 29.6	Nb 31.2	Nb 28.2
<u>MFE geometry</u>								
HT-9	Fe 85.6	Fe 81.6	Fe 68.9	Fe 67.9	Mo 87.3	Mo 90.7	Fe 78.5	Fe 82.4
	Cr 7.6	Mn 6.8	Cr 14.6	Cr 15.7	Ni 11.9	Ni 8.3	Mn 9.9	Mn 10.6
UCVS-1	Fe 93.4	Fe 91.4	Fe 88.5	Fe 87.6	Cu 70.7	Cu 65.2	Fe 89.2	Fe 93.5
	Mo 1.7	Mn 3.6	Mn 4.0	Mn 4.2	Mo 12.4	Mo 16.5	Mn 5.3	Mn 5.6
2-1/4 Cr-1 Mo	Fe 94.1	Fe 91.7	Fe 79.6	Fe 86.6	Mo 98.7	Mo 98.7	Fe 89.7	Fe 91.2
	Mn 2.3	Mn 5.4	Mn 5.9	Mn 6.3	Fe 1.3	Fe 1.3	Mn 8.1	Mn 8.3
Nb Stab 2-1/4 Cr	Fe 95.2	Fe 92.4	Fe 56.6	Fe 60.7	Mo 50.5	Mo 50.5	Fe 86.8	Fe 92.7
	Mn 1.5	Mn 3.6	Nb 30.3	Nb 31.7	Nb 48.8	Nb 48.8	Nb 6.2	Mn 5.5

General Studies and Reactor Development Technologies

Scheme	Wt% of element retained	Afterheat		bhp		Waste disposal		Dose rate	
		1 min	10 h	1 h	10 h	50 y	100 y	1 d	1 mo
Remove ⁵⁴ Fe	94.4	10.1 3.2	63.6 49.2	82.4 62.4	87.8 68.4			96.7 86.7	97.5 87.5
Retain ⁵⁷ Fe only	2.14	81.3 80.5	81.4 97.6	92.9 99.0	92.6 99.4			98.6 99.8	98.5 99.9
Remove ⁹² Mo	85.8					90.0 25.5	89.3 25.6		
Remove ⁹⁴ Mo	90.9					-0.5 58.0	-0.5 58.1		
Remove ⁹² Mo and ⁹⁴ Mo	76.7					99.6 97.2	99.3 97.4		
Keep ⁹⁷ Mo and ⁹⁸ Mo	34.3					99.8 99.9	99.8 99.9		

Table 7-31. Reduction of the iron or molybdenum contributions due to various isotopic tailoring schemes. In each vertical pair of figures, the percent reduction for ICF is shown above the MFE reduction.

comparison, the PCA stainless steel used in the STARFIRE tokamak design would be another order of magnitude worse.¹⁰⁰

For afterheat, the shutdown ICF levels of ~1 MW are acceptable without emergency wall cooling while the levels of ~6 MW for MFE are marginal. For bhp, the ICF levels of ~20 km³ are probably acceptable because evacuation to distances of several kilometers is feasible. However, MFE levels of ~200 km³ are marginal because evacuation to several tens of kilometers would be much more difficult. But, because we arbitrarily assumed a 10⁻⁶ release fraction, these numbers could shift to make both geometries either acceptable or unacceptable, depending on the release fraction. For dose rate, values of ~10⁵ to 10⁶ R/hr will require remote maintenance in both geometries.

Improvements Possible by Isotopic Tailoring. The principal parent elements contributing to the figures of merit are shown in Table 7-30, which lists the top two contributors in each case. The top contributor elements are Fe, Mo, Cu, and Nb. Additional elements in the second rank are Cr, W, Ni, Mn, and V. Neither manganese nor niobium can be tailored, because they consist of only one isotope each.

The dominance of iron is discouraging because tailoring of more than 85% of the steel is likely to be too expensive. Yet, if more than 10% effects are desired, iron offers the only chance in 48 of the 64 cases calculated (four figures of merit at two times each, for two geometries and four steels). The 48 cases are the top contributors in afterheat, bhp, and dose rate for all steels and geometries. The next best element to consider tailoring is molybdenum, with 16 opportunities, all in the waste disposal

category. Also in the waste disposal category, copper offers four chances for improvement, and nickel offers two chances. In the afterheat and bhp areas, chromium offers three chances and tungsten offers one.

The activation calculations were rerun for the isotopes of iron and molybdenum. This information was used to evaluate various schemes of isotopic tailoring, as shown in Table 7-31. Combining Tables 7-29, 7-30, and 7-31 allows a specific comparison between the various steels. In most cases, a given scheme helps all the steels in both ICF and MFE geometries; however, removing ⁹⁴Mo alone will help only for MFE. Schemes that utilize a large fraction of the natural element will be less expensive; therefore, removing ⁵⁴Fe is much more cost effective than concentrating ⁵⁷Fe, even though the result is not quite as great. Similarly, it is better to remove ⁹²Mo and ⁹⁴Mo than to concentrate ⁹⁷Mo and ⁹⁸Mo.

The schemes reduce the iron and molybdenum contributions by one to three orders of magnitude; further reductions in these elements would be ineffective unless other alloying elements are also tailored. For afterheat, bhp, and waste disposal, the reductions would be enough to satisfy the suggested limits on the figures of merit. However, for surface dose rate, the reduction is still inadequate to allow hands-on maintenance of the wall.

Summary. We compared the activation of four ferritic walls for both ICF and MFE reactors. The comparison criteria were afterheat, inhalation bhp and surface dose rate at early times after shutdown, and waste-disposal feasibility 50 to 100 y after shutdown. In general, the softer ICF flux leads to a significant advantage. The four steels are roughly equal, except

that UCVS-1 is superior from the waste-disposal viewpoint. The early performance (afterheat, bhp, and dose rate) can be improved by one to three orders of magnitude by isotopic tailoring of iron, but such tailoring would be very expensive and would not solve the remote-maintenance

problem. For waste disposal, molybdenum is the best element to tailor, and molybdenum tailoring would allow disposal of all four steels by shallow land burial within 50 y after shutdown.

Author: J. A. Blink

1. J. H. Pitts, "Cascade: A Centrifugal-Action Solid-Breeder Reaction Chamber," *Nucl. Technol./Fusion* 4 (2), Pt. 2, 967 (1983).
2. J. H. Pitts, "Shock Propagation Through the Cascade ICF Reactor Li_2O -Pebble Blanket," *Trans. Am. Nucl. Soc.* 45, 184 (1983).
3. J. H. Pitts, *Mechanical and Thermal Design of the Cascade Reactor*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-89273 (1983).
4. J. H. Pitts, "Cascade: A Centrifugal-Action Solid-Breeder Reaction Chamber," *Nucl. Technol./Fusion* 4 (2), Pt. 2, 967 (1983).
5. O. R. Walton, *Granular Flow: Considerations in the Design of a Cascade Solid Breeder Reaction Chamber*, Lawrence Livermore National Laboratory, Livermore, Calif., UCID-19903 (1983).
6. S. B. Savage and M. Sayed, "Gravity Flow of Cohesionless Granular Materials in Wedge-Shaped Hoppers," *Mechanics Applied to the Transport of Bulk Materials*, S. C. Cowin, Ed. (American Society of Mechanical Engineers, New York, N.Y., 1979), AMD-Vol. 31, pp. 1-24.
7. J. H. Pitts, *Cascade: A Centrifugal-Action Solid Breeder Reaction Chamber*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-88208 (1983).
8. C. E. Brennen, K. Sieck, and J. Paslaski, "Hydraulic Jumps in Granular Material Flow," *Power Technol.* 35, 31 (1983).
9. O. R. Walton, *Granular Flow: Considerations in the Design of a Cascade Solid Breeder Reaction Chamber*, Lawrence Livermore National Laboratory, Livermore, Calif., UCID-19903 (1983).
10. C. L. Hwang, *Mixing and Segregation in Flowing Powders*, Ph.D. thesis, Pennsylvania State University, University Park, Pa. (1978).
11. R. Hogg, Pennsylvania State University, University Park, Pa., private communication (1975).
12. S. Savage, McGill University, Montreal, Que., Can., private communication (1982).
13. B. A. Wills, *Mineral Processing Technology* (Pergamon Press, New York, N.Y., 1982), 2nd ed.
14. G. Hawkins, *Particle Advanced Technology Center*, Columbia, Md., private communication (1983).
15. J. H. Pitts, "Cascade: A Centrifugal-Action Solid-Breeder Reaction Chamber," *Nucl. Technol./Fusion* 4 (2), Pt. 2, 967 (1983).
16. C. E. Johnson, "Recent Advances in the Development of Solid Breeder Blanket Materials," *Proc. 3rd Top. Meet. Fusion Reactor Materials*, Albuquerque, N. Mex., 1983 (to be published).
17. J. H. Norman and G. R. Hightower, "Measurements of the Activity Coefficient of LiOH Dissolved in Li_2O (s) for an Evaluation of Li_2O as a Tritium Breeder Material," *Proc. 3rd Top. Meet. Fusion Reactor Materials*, Albuquerque, N. Mex., 1983 (to be published).
18. M. K. Booker, J. P. Strizak, and C. R. Brinkman, *Analysis of the Continuous Cycling Fatigue Behavior of 2.25 Cr 1 Mo Steel*, Oak Ridge National Laboratory, Oak Ridge, Tenn., ORNL-5593 (1979).
19. ASME, *Boiler and Pressure Vessel Code*, Code Case N-47-17, 1980.
20. J. H. Pitts, "A Consistent HYLIFE Wall Design That Withstands Transient Loading Conditions," *Proc. 4th Top. Meet. Technol. of Controlled Nuclear Fusion*, King of Prussia, Pa., 1980 (U.S. Dept. of Energy, Washington, D. C., 1981), vol. 1, pp. 1174-1181.
21. E. Schlünder, "Particle Heat Transfer," *Proc. 7th Intern. Heat Transfer Conf., München*, 1982 (Hemisphere Publishing Corp., Washington, D.C., 1982), vol. 1, pp. 195-211.
22. L. A. Glenn and D. A. Young, "Dynamic Loading of the Structural Wall in a Lithium-Fall Fusion Reactor," *Nucl. Engr. & Design* 54, 1 (1979).
23. F. S. Aschner, *Planning Fundamentals of Thermal Power Plants* (Wiley, New York, N.Y., 1978) p. 460.
24. J. H. Pitts, "Cascade: A Centrifugal-Action Solid-Breeder Reaction Chamber," *Nucl. Technol./Fusion* 4 (2), Pt. 2, 967 (1983).
25. R. Waszink, Westinghouse, Pensacola, Fla., private communication (February 1983).
26. E. U. Schlünder, "Particle Heat Transfer," *Proc. 7th Intern. Heat Transfer Conf., München*, 1982 (Hemisphere Publishing Corp., Washington, D.C., 1982), vol. 1, pp. 195-211.
27. J. Duderstadt and G. Moses, *Inertial Confinement Fusion* (Wiley, New York, N.Y., 1982), p. 8.
28. D. Coffin, Los Alamos National Laboratory, Los Alamos, N. Mex., private communications (June 1983).
29. Marks Standard Handbook for Mechanical Engineers, T. Baumeister, Ed. (McGraw-Hill, New York, N.Y., 1967), 7th ed., pp. 10-52 to 10-82.
30. R. W. Conn et al., *Solase: A Laser Fusion Reactor Study*, University of Wisconsin, Madison, Wis., UWFD-220 (1977).
31. J. H. Pitts, "Cascade: A Centrifugal-Action Solid-Breeder Reaction Chamber," *Nuclear Technol./Fusion* 4 (2), Pt. 2, 967 (1983).
32. J. R. Welty, C. E. Wicks, and R. E. Wilson, *Fundamentals of Momentum, Heat, and Mass Transfer* (Wiley, New York, N.Y., 1969), p. 318.
33. M. J. Monsler, J. Hovingh, D. L. Cook, T. G. Frank, and G. A. Moses, "An Overview of Inertial Fusion Reaction Design," *Nucl. Technol./Fusion* 1, 316 (1981).
34. W. E. Mason, Jr., TACO - A Finite Element Heat Transfer Code, Lawrence Livermore National Laboratory, Livermore, Calif., UCID-17980, rev. 1 (1980).
35. J. O. Hallquist, *NIKE2D - A Vectorized Implicit, Finite Deformation, Finite Element Code for Analyzing the Static and Dynamic Response of 2-D Solids*, Lawrence Livermore National Laboratory, Livermore, Calif., UCID-19677 (1983).
36. J. H. Pitts, "Cascade: A Centrifugal-Action Solid-Breeder Reaction Chamber," *Nucl. Technol./Fusion* 4 (2), Pt. 2, 967 (1983).
37. K. Okuno and H. Kudo, "Thermal Release of Tritium Produced in Sintered Li_2O Pellets," *J. Nucl. Mater.* 116, 82 (1983).
38. T. Tanifuji, K. Noda, S. Nasu, and K. Uchida, "Tritium Release from Neutron-Irradiated Li_2O : Constant Rate Heating Measurements," *J. Nucl. Mater.* 95, 108 (1980).
39. R. H. Clemmer et al., "The TRIO-01 Experiment: In-Situ Tritium Recovery Results," *3rd Top. Meet. Fusion Reactor Materials*, Albuquerque, N. Mex., September 1983 (to be published).

References

40. S. Nasu et al., "A Preliminary In-Pile Test of Tritium Release from Li_2O Pellets," *J. Nucl. Mater.* **101**, 220 (1981).
41. C. E. Kessel, Jr., *TRIT3 Program Manual: A Fortran Code for the Prediction of Tritium Inventories and Permeation Rates at Steady-State in Solid Breeder Blankets for Fusion Reactors*, GA Technologies, San Diego, Calif. (to be published).
42. S. Nasu et al., "Tritium Release from Neutron-Irradiated Li_2O Pellets," *J. Nucl. Mater.* **68**, 261 (1977).
43. J. R. Powell, R. H. Wiswall, and E. Wirsing, *Tritium Recovery from Fusion Blankets Using Solid Breeder Compounds*, Brookhaven National Laboratory, Upton, N. Y., BNL-20563 (1975).
44. W. R. Meier, Lawrence Livermore National Laboratory, Livermore, Calif., private communication (June 1983).
45. M. A. Abdou et al., *Blanket Comparison and Selection Study*, Argonne National Laboratory, Batavia, Ill., ANL/FPP-83-1, vol. II (1983), pp. XI-8 to XI-13.
46. J. R. Hofmann, "Transient Pressurization and Transport of a Two-Component Gas Within Interconnected Porosity of Solid Mixed-Oxide Fuel," *Nucl. Sci. and Eng.* **53**, 73 (1978).
47. N. W. Gregory and R. H. Mohr, "The Equilibrium $2\text{LiOH}(s) = \text{Li}_2\text{O}(s) + \text{H}_2\text{O}(g)$," *J. Am. Chem. Soc.* **77**, 2142 (1955).
48. J. H. Norman and G. R. Hightower, "Measurements of the Activity Coefficient of LiOH Dissolved in $\text{Li}_2\text{O}(s)$ for the Evaluation of Li_2O as a Tritium Breeding Material," *Proc. 3rd Top. Meet. Fusion Reactor Materials*, Albuquerque, N. Mex., September 1983 (to be published).
49. J. A. Blink, "Tritium Breeding and Activation in the Cascade Reactor," *10th IEEE Symp. Fusion Engineering*, Philadelphia, Pa., Dec. 5-9, 1983. Also, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-89292 (1983).
50. J. H. Pitts, "Cascade: A Centrifugal-Action Solid-Breeder Reaction Chamber," *Nucl. Technol./Fusion* **4** (2), Pt 2, 967 (1983).
51. R. J. Howerton, *A Status Report for Data Files and Monte Carlo Transport Codes Maintained by the Physical Data Group of LLNL*, Lawrence Livermore National Laboratory, Livermore, Calif., UCID-19907 (1983).
52. E. F. Plechaty and J. R. Kimlinger, *TARTNP: A Coupled Neutron-Photon Monte Carlo Transport Code*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-50400, vol. 14 (1976).
53. J. A. Blink, R. E. Dye, and J. R. Kimlinger, *ORIBR: A Computer Code that Produces One-Energy Group, Time- and Spatially-Averaged Neutron Cross Sections*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-53262 (1981).
54. J. A. Blink, *FORIG: A Modification of the ORIGEN2 Isotope Generation and Depletion Code for Fusion Problems*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-53263 (1981).
55. A. G. Croff, *A User's Manual for the ORIGEN2 Computer Code*, Oak Ridge National Laboratory, Oak Ridge, Tenn., ORNL/TM-7175 (1980).
56. J. A. Blink, *Neutron Activation of Four Ferritic Steels*, Lawrence Livermore National Laboratory, Livermore, Calif., UCID-19823 (1981).
57. J. A. Blink, D. W. Dorn, and R. C. Maninger, *Disposal of Activated Fusion Wall Materials*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-89511 (1983).
58. J. H. Pitts, Lawrence Livermore National Laboratory, Livermore, Calif., private communication (November 1983).
59. J. Blink and N. J. Hoffman, "Tritium Control and Activation in the Pulse-Star Reactor," *10th IEEE Symp. Fusion Engineering*, Philadelphia, Pa., Dec. 5-9, 1983. Also, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-89291 (1983).
60. Title 10, Code of Federal Regulations, Pt. 61, (10CFR61), published in *The Federal Register*, Dec. 27, 1982.
61. J. A. Blink, "Tritium Control and Activation in the Pulse-Star Reactor," *10th IEEE Symp. Fusion Engineering*, Philadelphia, Pa., Dec. 5-9, 1983 (to be published).
62. R. J. Howerton, *A Status Report for Data Files and Monte Carlo Transport Codes Maintained by the Physical Data Group of LLNL*, Lawrence Livermore National Laboratory, Livermore, Calif., UCID-19907 (1983).
63. E. F. Plechaty and J. R. Kimlinger, *TARTNP: A Coupled Neutron-Photon Monte Carlo Transport Code*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-50400, vol. 14 (1976).
64. J. A. Blink, R. E. Dye, and J. R. Kimlinger, *ORIBR: A Computer Code that Produces One-Energy Group, Time- and Spatially-Averaged Neutron Cross Sections*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-53262 (1981).
65. J. A. Blink, *FORIG: A Modification of the ORIGEN2 Isotope Generation and Depletion Code for Fusion Problems*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-53263 (1981).
66. A. G. Croff, *A User's Manual for the ORIGEN2 Computer Code*, Oak Ridge National Laboratory, Oak Ridge, Tenn., ORNL/TM-7175 (1980).
67. J. A. Blink, *Neutron Activation of Four Ferritic Steels*, Lawrence Livermore National Laboratory, Livermore, Calif., UCID-19823 (1983).
68. J. A. Blink, D. W. Dorn, and R. C. Maninger, "Disposal of Activated Fusion Wall Materials," *3rd Top. Meet. Fusion Reactor Materials*, Albuquerque, N. Mex., Sept. 1983. Also, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-89511 (1983).
69. J. A. Blink, *Afterheat Due to PbLi in a Pool Type Reactor*, Internal Memorandum, Lawrence Livermore National Laboratory, Livermore, Calif., E&MA-83-022 (1983).
70. Title 10, Code of Federal Regulations, Pt. 61, (10CFR61), published in *The Federal Register*, Dec. 27, 1982.
71. J. A. Blink and N. J. Hoffman, "Tritium Control and Activation in the Pulse-Star Reactor," *10th IEEE Symp. Fusion Engineering*, Philadelphia, Pa., Dec. 5-9, 1983 (to be published).
72. N. J. Hoffman, A. Damell, and J. A. Blink, *Properties of Lead-Lithium Solutions*, *Proc. 4th Top. Meet. Technol. Controlled Nucl. Fusion*, King of Prussia, Pa., 1980 (U. S. Department of Energy, Washington, D.C., 1981), CONF-801011.
73. H. R. Hile, A. Neubert, and C. H. Wu, "The Activity of Lithium and the Solubility of Deuterium

- Lithium Lead Alloys," *Proc. 10th Symp. Fusion Technol., Padova, Italy* (Pergamon, New York, N.Y., 1979), p. 639.
74. Y. C. Chan and E. Velecki, "A Thermodynamic Investigation of Dilute Solutions of Hydrogen in Liquid Lithium-Lead Alloys," *Proc. 3rd Top. Meet. Fusion Reactor Materials, Albuquerque, N. Mex., September 1983* (to be published).
75. W. R. Opie and N. J. Grant, "Solubility of Hydrogen in Molten Lead," *J. Metals* 3 (3), 244 (1951).
76. W. Hofmann and J. Maatsch, *Erzmetall* 10, 313 (1957).
77. G. Pierini, A. M. Polcaro, P. F. Ricci, and A. Viola, *Proc. 12th Symp. Fusion Technol., Jülich (FRG), September 13-17, 1982* (Pergamon, New York, N.Y., 1983).
78. C. H. Wu, *J. Nucl. Mater.* 114, 30 (1983).
79. C. H. Wu and A. J. Blair, *Proc. 12th Symp. Fusion Technol., Jülich (FRG), September 13-17, 1982* (Pergamon, New York, N.Y., 1983).
80. K. Wilson, Sandia National Laboratories, Livermore, Calif., private communication (Nov. 21, 1983).
81. N. J. Hoffman, A. J. Darnell, and J. A. Blink, "Properties of Lead-Lithium Solutions," 4th Top. Meet. Technol. Controlled Nucl. Fusion, King of Prussia, Pa., October 1980 (U.S. Department of Energy, Washington, D.C., 1981), CONF-80101.
82. D. G. Shepherd, *Principles of Turbomachinery* (MacMillan, New York, N. Y., 1956), p. 37.
83. O. E. Dwyer, *Liquid Metal Heat Transfer*, Brookhaven National Laboratory, Upton, N. Y., BNL-119-36R (1969), pp. 58 and 124.
84. M. W. McDowell, K. A. Murray, and J. A. Blink, "Pulse*Star Inertial Confinement Fusion Reactor: Heat Transfer Loop and Balance-of-Plant Considerations," 10th IEEE Symp. Fusion Engineering, Philadelphia, Pa., Dec. 5-9 1983 (to be published).
85. M. W. McDowell and K. A. Murray, *Pulse*Star Inertial Confinement Fusion Reactor: Heat Transfer Loop and Balance-of-Plant Considerations*, Lawrence Livermore National Laboratory, Livermore, Calif., UCID-15598 (1984).
86. R. Waszink, Westinghouse, Pensacola, Fla., private communication (February 1983).
87. M. S. Peter and K. D. Timmerhaus, *Plant Design and Economics for Chemical Engineers* (McGraw-Hill, New York, N.Y., 1980), p. 654.
88. *Chemical Engineers Handbook*, R. H. Perry and C. H. Chilton, Eds. (McGraw-Hill, New York, N.Y., 1973) 5th ed., pp. 10-24.
89. Rockwell International and Burns and Roe, *Liquid Metal Fast Breeder Reactor Conceptual Plant Design: 1000 MW*, U.S. Energy Research and Development Administration, Washington, D. C., FBR-73-3 (1977).
90. J. A. Maniscalco, D. H. Berwald, R. B. Campbell, R. W. Moir, and J. D. Lee, "Recent Progress in Fusion-Fission Hybrid Reactor Design Studies," *Nucl. Technol./Fusion* 1, 419 (1981).
91. M. Z. Youssef and R. W. Conn, *A Survey of Fusion-Fission System Designs and Nuclear Analyses*, University of Wisconsin, Madison, Wis., UWFD-308 (1979).
92. J. D. Lee and R. W. Moir, "Fission-Suppressed Blankets for Fissile Fuel Breeding Fusion Reactors," *J. Fusion Energy* 1 (3), 299 (1981).
93. D. H. Berwald et al., *Fission-Suppressed Hybrid Reactor: The Fusion Breeder*, Lawrence Livermore National Laboratory, Livermore, Calif., UCID-19638 (1992).
94. W. R. Meier, *A High Performance, Suppressed-Fission ICF Hybrid*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-89274 (1983).
95. E. F. Plechaty and J. R. Kimlinger, *TARTNP: A Coupled Neutron-Photon Monte Carlo Transport Code*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-50400, vol. 14 (1975).
96. D. E. Cullen, *Bondarenko Self-Shielded Neutron Cross Sections and Multiband Parameters Derived from the LLNL Evaluated-Nuclear-Data Library (ENDL)*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-50400, vol. 20 (1978).
97. E. F. Plechaty and D. E. Cullen, *Resonance Self-Shielding Calculations Using the Probability Table Method*, Lawrence Livermore National Laboratory, Livermore, Calif., UCID-17230 (1976).
98. E. F. Plechaty, D. E. Cullen, R. J. Howerton, and J. R. Kimlinger, *Tabular and Graphical Presentation of 75 Neutron-Group Constants Derived from the LLNL Evaluated-Nuclear-Data Library (ENDL)*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-50400, vol. 16, rev. 2 (1978).
99. S. Taczanowski, "Neutronic Considerations for Direct Enrichment and Spent Fuel Regeneration in Hybrids Without Reprocessing," *Proc. 3rd Intern. Conf. Emerging Nucl. Energy Systems, Helsinki, Finland, June 6-9, 1983* (to be published).
100. *Energy Economic Data base, Phase V Update Report*, United Engineers and Constructors, Philadelphia, Pa., UE&C-ANL-82115 and COO-0818-1 (1982).
101. *1981 Laser Program Annual Report*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-50014-1-81 (1982), pp. 3-1 to 3-2.
102. J. H. Nuckolls, "The Feasibility of Inertial Confinement Fusion," *Phys. Today* 35 (9), 24 (1982).
103. J. H. Pitts, "Cascade: A Centrifugal-Action Solid-Breeder Reaction Chamber," *Nucl. Technol./Fusion* 4 (2), Pt. 2, 967 (1983).
104. G. P. Lasche, "Inertial Confinement Fusion with Direct Electric Generation by Magnetic Flux Compression," *Nucl. Technol./Fusion* 4 (2), Pt. 2, 866 (1983).
105. M. J. Monsler and W. R. Meier, "A Conceptual Design Strategy for Liquid-Metal-Wall Inertial Fusion Reactors," *Nucl. Eng. & Design* 63, 289 (1981).
106. *1982 Laser Program Annual Report*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-50021-82 (1983), pp. 8-55 to 8-58.
107. J. A. Blink, *Development Facilities Required to Support Inertial Fusion Reactor Commercialization*, Internal Memorandum, Lawrence Livermore National Laboratory, Livermore, Calif., E&MA 83-017 (Feb. 7, 1984).
108. M. J. Monsler, J. Hovingh, D. L. Cook, T. G. Frank, and G. A. Moses, "An Overview of Inertial Fusion Reactor Design," *Nuclear Technol./Fusion* 1 (2), 302 (1981).
109. W. R. Meier, *Heavy-Ion-Beam-Driven Efficiency as a Function of Pulse Repetition Rate and Beam Energy*, Internal Memorandum, Lawrence Livermore National Laboratory, Livermore, Calif., E&MA 80-174 (Oct. 9, 1980).

References

110. 1981 *Laser Program Annual Report*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-50021-81 (1982), pp. 3-1 to 3-2.
111. M. J. Monsler et al., "An Overview of Inertial Fusion Reactor Design," *Nucl. Technol./Fusion* **1**, 302 (1981).
112. B. Badger et al., *HIBALL - A Conceptual Heavy Ion Beam Driven Fusion Reactor Study*, University of Wisconsin, Madison, Wis., UWFD-450 (1981).
113. J. H. Pendergrass, T. G. Frank, and I. O. Bohachevsky, "A Modified Wetted-Wall Inertial Fusion Reactor Concept," *Proc. 4th Top. Meet. Technol. Controlled Nucl. Fusion, King of Prussia, Pa., 1981* (U.S. Dept. of Energy, Washington, D.C., 1981), pp. 1131-1142.
114. W. O. Allen et al., *Phase I: Engineering Test Reactor Goal Specification, Final Report*, Electric Power Research Institute, Palo Alto, Calif., EPRI-1527 (1982).
115. J. A. Blink and N. J. Hoffman, "Tritium Control and Activation in the Pulse-Star Reactor," *10th IEEE Symp. Fusion Eng.*, Philadelphia, Pa., Dec. 5-9, 1983 (to be published).
116. J. A. Blink, *Lithium Equation-of-State*, Lawrence Livermore National Laboratory, Livermore, Calif., UCID-19890 (1983).
117. J. Hovingh, S. L. Thomson, and J. A. Blink, "Fluid Mechanics Considerations for Liquid Wall Inertially Confined Fusion Reactors," *Proc. 8th Symp. Eng. Problems of Fusion Research, San Francisco, Calif., 1979* (IEEE, Piscataway, N.J., 1979), vol. 4, pp. 2038-2041.
118. D. E. Grady, "Local Inertial Effects in Dynamic Fragmentation," *J. Appl. Phys.* **53**, 1 (1982).
119. L. A. Glenn, *On the Fragmentation of Condensed Material by Isochoric Heating and Release*, Lawrence Livermore National Laboratory, Livermore, Calif., UCID-19737 (1983).
120. E. T. Cheng and R. W. Conn, "A Multipoint Interpolation Method Based on Variational Principles for Functionals of the Solution to Linear Equations," *J. Math. Phys. (N.Y.)* **17** (5), 683 (1976).
121. E. T. Cheng and R. W. Conn, "The Influence of Design Variations on Controlled Thermonuclear Reactor Blanket Neutronic Performance Using Variational Techniques," *Nucl. Sci. Eng.* **62**, 601 (1977).
122. W. M. Stacey, Jr., "Variational Estimates and Generalized Perturbation Theory for the Ratios of Linear and Bilinear Functionals," *J. Math. Phys. (N.Y.)* **13** (8), 1119 (1972).
123. R. W. Conn and W. M. Stacey, Jr., "Variational Methods for Controlled Thermonuclear Reactor Blanket Studies," *Nucl. Fusion* **13**, 185 (1973).
124. S. A. W. Gerstl and W. M. Stacey, Jr., "A Class of Second-Order Approximate Formulations of Deep Penetration Radiation Transport Problems," *Nucl. Sci. Eng.* **51**, 339 (1973).
125. E. T. Cheng, "Application of Generalized Variational Principles to Controlled Thermonuclear Reactor Neutronics Analysis," *Nucl. Sci. Eng.* **74**, 147 (1980).
126. W. M. Stacey, Jr., *Variational Methods in Nuclear Reactor Physics* (Academic Press, New York, N.Y., 1974).
127. G. I. Bell and S. Glasstone, *Nuclear Reactor Theory* (Van Nostrand Reinhold, New York, N.Y., 1970), p. 255.
128. *Laser Program Annual Report - 1978*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-50021-78 (1979), p. 9-1.
129. *Laser Program Annual Report - 1979*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-50021-79 (1980), p. 8-1.
130. 1980 *Laser Program Annual Report*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-50021-80 (1981), p. 9-1.
131. 1981 *Laser Program Annual Report*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-50021-81 (1982), p. 8-1.
132. M. J. Monsler et al., "Electric Power from Laser Fusion: The HYLIFE Concept," *Proc. 13th IEEE Conf., San Diego, Calif. Aug. 20-25, 1978* (Society of Automotive Engineers, Warrendale, Pa., 1978), p. 2164.
133. M. J. Monsler and W. R. Meier, "A Conceptual Design Strategy for Liquid-Metal-Wall Inertial Fusion Reactors," *Nucl. Eng. Des.* **63**, 289 (1981).
134. W. R. Meier, "Two-Dimensional Neutronics Calculation for the HYLIFE Converter," *Nucl. Technol.* **52**, 22 (1981).
135. W. R. Meier, "Tritium Breeding Management in the HYLIFE Chamber," *Nucl. Technol.* **52**, 170 (1981).
136. W. R. Meier and W. B. Thompson, "Conceptual Design Considerations and Neutronics of Lithium-Fall Laser Fusion Target Chambers," *Proc. 3rd Top. Mtg. Technol. of Controlled Nucl. Fusion, Santa Fe, N. Mex., 1978* (CONF-780508), p. 297.
137. E. F. Plechaty and J. R. Kimlinger, *TARTNP: A Coupled Neutron Photon Monte Carlo Transport Code*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-50400, vol. 14 (1976).
138. E. F. Plechaty, D. E. Cullen, R. J. Howerton, and J. R. Kimlinger, *Tabular and Graphical Presentation of 175 Neutron-Group Constants Derived from the LLNL Evaluated Nuclear Data Library (ENDL)*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-50400 vol. 16, rev. 2 (1978).
139. G. L. Kulcinski, "Radiation Damage to Materials in DT Fusion Reactors," *Proc. 5th Conf. Plasma Phys. and Controlled Nucl. Fusion Research, Tokyo, Japan, Nov. 11-15, 1974*. Also, University of Wisconsin, Madison, Wis., UWFD-116 (1974).
140. F. A. Smidt, Jr., "Swelling Behavior of Commercial Ferritic Alloys, EM-12 and HT-9, as Assessed by Heavy Ion Bombardment," *irradiation Effects on the Microstructure and Properties of Metals*, American Society for Testing and Materials, ASTM STP 611 (1976), pp. 227-241.
141. E. E. Bloom, "Candidate First Wall Materials for Fusion Reactors," *Proc. 10th Symp. Fusion Eng., Philadelphia, Pa., Dec. 5-9, 1983* (to be published).
142. M. J. Monsler et al., "An Overview of Inertial Fusion Reactor Design," *Nucl. Technol./Fusion* **1**, 302 (1981).
143. W. R. Meier and W. B. Thompson, "Conceptual Design Considerations and Neutronics of Lithium-Fall Laser Fusion Target Chambers," *Proc. 3rd ANS Top. Meet. Technol. of Controlled Nucl. Fusion, Santa Fe, N. Mex., May 9-11, 1978*, p. 297.

144. H. I. Avci and G. L. Kulcinski, "The Effect of Liquid Metal Protection Schemes in Inertial Confinement Fusion Reactors," *Nucl. Technol.* **44**, 333 (1979).
145. D. G. Doran and N. J. Graves, "Neutron Displacement Damage Cross Sections for Structural Metals," *Irradiation Effects on the Microstructure and Properties of Metals*, American Society for Testing and Materials, ASTM STP 611 (1976), pp. 463-482.
146. D. G. Doran, *Nucl. Sci. Eng.* **49**, 130 (1972).
147. D. G. Doran, *Nucl. Sci. Eng.* **52**, 398 (1973).
148. E. F. Plechaty and J. R. Kimlinger, *TARTNP: A Coupled Neutron Photon Monte Carlo Transport Code*, Lawrence Livermore Laboratory, Livermore, Calif., UCRL-50400, vol. 14 (1976).
149. W. R. Meier, "Tritium Breeding Management in the High Yield Lithium Fusion Energy Chamber," *Nucl. Tech.* **52**, 170 (1981).
150. H. Pitts, "Cascade: A Centrifugal-Action Solid-Breeder Reaction Chamber," *Nucl. Technol./Fusion* **4** (2), Pt. 2, 967 (1983).
151. A. K. Huntington, *J. Inst. Met.* **11**, 108 (1914).
152. J. C. Lynn, W. R. Warke, and P. Gordon, *Mater. Sci. Engr.* **18**, 51 (1975).
153. C. F. Old and P. Trevena, *Met. Sci. J.* **13**, 591 (1979).
154. N. S. Stoloff and T. L. Johnston, *Acta Met.* **11**, 251 (1963).
155. A. R. C. Westwood and M. H. Kamdar, *Phil. Mag.* **8**, 77 (1963).
156. M. H. Kamdar, *Mechanisms of Embrittlement and Brittle Fracture in Liquid Metals*, National Technical Information Service, Springfield, Va., NTIS-AD-A037 (1977).
157. C. F. Old, "Liquid Metal Embrittlement of Nuclear Materials," *J. Nucl. Mat.* **92**, 2 (1980).
158. W. M. Robertson, *Trans. Metal. Soc. AIME* **236**, 1478 (1966).
159. R. G. Dunn, G. F. Whitely, and W. Fairhurst, *Molybdenum's Place in the Pressure Vessel Field*, Climax Molybdenum Co., AMAX Div., Greenwich, Conn. (1969).
160. J. A. Blink, D. W. Dorn, and R. C. Maninger, "Disposal of Activated Fusion Materials," 3rd Top. Meet. on Fusion Reactor Materials, Albuquerque, N. Mex. (Sept. 19-22, 1982), to be published in *J. Nucl. Mater.*
161. Title 10, Code of Federal Regulations, Part 61 (10CFR61), published in *The Federal Register* (Dec. 27, 1982).
162. Draft Environmental Impact Statement on the Disposal of Decommissioned and Detueled Navy Submarine Reactor Plants, U.S. Navy (Dec. 1982).
163. Draft Environmental Impact Statement for 10CFR61, Nuclear Regulatory Commission, Washington, D.C., NUREG 0782, Appendices G and H (1981).
164. N. Giononi, J. Blink, and N. Hoffman, "Selection of Alloy Steel Type for Fusion Power Plant Applications in the 450-500°C Range," Top. Conf. Ferritic Alloys for Use in Nucl. Energy Technol., Snowbird, Utah (June 19-23, 1983). See also following article, "Steel Type and Impurity Effects on Activation."
165. F. W. Wiffen and R. T. Santoro, "Control of Activation Levels to Simplify Waste Management of Fusion Reactor Ferritic Steel Components," Top. Conf. Ferritic Alloys for Use in Nucl. Energy Technologies, Snowbird, Utah (June 19-23, 1983).
166. J. A. Blink and G. P. Lasche, "The Influence of Steel Type on The Activation and Decay of Fusion Reactor First Walls," *Nucl. Technol./Fusion*, **4** (Pt. 3), 1146 (1983).
167. J. A. Blink, *Neutron Activation of Four Ferritic Steels*, Lawrence Livermore National Laboratory, Livermore, Calif., UCID-19823 (1983).
168. G. P. Lasche and J. A. Blink, "The Dependence of Neutron Induced Radioactivity in Fusion Reactors on Geometric Design Parameters," *Nucl. Technol./Fusion* **4** (Pt. 3), 823 (1983).
169. E. F. Plechaty and J. R. Kimlinger, *TARTNP: A Coupled Neutron-Photon Monte Carlo Transport Code*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-50400, vol. 14 (1976).
170. J. A. Blink, R. E. Dye, and J. R. Kimlinger, *ORLIB: A Computer Code That Produces One-Energy Group, Time- and Spatially-Averaged Neutron Cross Sections*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-53262 (1981).
171. M. A. Gardner and R. J. Howerton, *ACTL: Evaluated Neutron Activation Cross Section Library—Evaluation Techniques and Reaction Index*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-50400, vol. 18 (1978).
172. J. A. Blink, *FORIG: A Modification of the ORIGEN2 Isotope Generation and Depletion Code for Fusion Problems*, Lawrence Livermore National Laboratory, Livermore, Calif., UCRL-53263 (1981).
173. A. G. Croff, *A User's Manual for the ORICEN2 Computer Code*, Oak Ridge National Laboratory, Oak Ridge, Tenn., ORNL-1M-7175 (1980).
174. J. A. Blink, *HYLIFE First Wall Cooling After Shut-down*, Lawrence Livermore National Laboratory, Livermore, Calif., internal memo, E&MA 83-036 (1983).
175. S. J. Piet, EG&G-Idaho, Idaho Falls, Idaho, private communication (April 26, 1983).
176. D. W. Dorn and R. C. Maninger, "Issues in Radioactivity for Fusion Energy—Remote Maintenance Rating," 10th Symp. Fusion Eng., Philadelphia, Pa., Dec. 5-9, 1983 (to be published).

Subject Index

Italic listings are article titles; section numbers appear in boldface and precede page numbers.

0.35- μ m-wavelength light

1-1, 3 to 4; 2-2, 4 to 5, 8, 10, 12 to 13, 15 to 17, 19, 31, 33, 38 to 39, 47 to 48; 3-32 to 34; 5-7 to 8, 33, 43, 46; 6-61, 68 to 69, 77, 79

0.53- μ m-wavelength light

1-1 to 4; 2-2, 4 to 5, 8, 10, 12 to 13, 15 to 17, 19, 31, 33, 47 to 49, 51, 53, 58; 3-1, 26, 30, 34, 40 to 42; 5-1, 10, 23 to 25, 28, 33, 39, 45, 47, 54 to 55; 6-60 to 61, 66, 69 to 70, 85, 88

1.05- μ m-wavelength light

2-17; 5-25; 6-3 to 4, 48 to 49, 78

1.06- μ m-wavelength light

1-2 to 3; 2-2 to 4, 8, 12 to 13, 15, 19, 31, 33, 36, 38 to 39, 49, 51, 61; 3-26, 32, 34, 40 to 42; 5-9 to 11, 15, 33, 40 to 41, 44, 50; 6-18, 31, 48 to 49, 50 to 61, 69 to 71, 74, 88, 94

$\omega/2$

3-32, 34

ω_{pe} Instability in Current Experiments, The

3-32

1 ω frequency

See 1.05 μ m, 1.06 μ m, Fundamental frequency

2 ω frequency

See 0.53 μ m, Second harmonic

2 ω_{pe} (two-plasmon decay) instability

3-26 to 27, 32 to 35, 37, 5-10, 41, 57 to 58

2-1/4 Cr-1 Mo steel

7-16, 69, 71

3 ω frequency

See 0.35 μ m, Third harmonic

3 $\omega/2$ frequency

3-32 to 34

4 ω frequency

2-8, 13; 6-67, 69, 77

15-cm amplifier

6-3 to 4, 65 to 66, 84, 98

20.8-cm amplifier

6-4

31.5-cm amplifier

6-4, 6 to 11, 78

46-cm amplifier

6-4 to 7, 57, 78

A

Ablation pressure

1-1

ABR

See Diagnostics, for targets

Absorption

bremsstrahlung 1-3; inverse bremsstrahlung 3-30, 37, 51, 54; 5-40, 46

Activation calculations

7-26 to 27, 30, 77, 79

Activation, Steel Type and Impurity Effects on

7-76

ACTL cross-section library

7-27, 31, 77

Advanced lasers

For advanced solid state systems see 6-91 to 96; for other laser systems see 6-96 to 102

Afterheat

in activation calculations 7-27 to 28, 32 to 33, 77 to 80

ALICE neutron-photon transport code

7-26 to 27, 30 to 31, 41 to 42, 45 to 46

Ammonium dideuterium phosphate (AD*P)

6-86

Ammonium dihydrogen phosphate (ADP)

6-86

Amplified spontaneous emission (ASE)

2-25 to 26, 56; 6-5, 8, 11, 17, 30, 57 to 58, 98 to 99, 113

Amplifiers

multipass systems 6-13, 57 to 60, 84, 86, 88, 90; Nd:glass 6-7 to 8, 19; Novette power amplifier 5-28; Rapier-B 6-2, 96 to 100; rod 2-23 to 25, 49, 55; storage efficiency 1-4; 15-cm 6-3 to 4, 65 to 66, 84, 98; 20.8-cm 2-55 to 58, 6-4; 31.5-cm 2-1, 50, 56; 6-4, 6 to 11, 78; 46-cm 2-1, 12, 50, 56; 6-4 to 7, 57, 78; See also Solid state amplifiers 6-2 to 18

Angular detuning

6-70 to 74, 77, 87

Apodization

2-5, 10, 12, 37

Antireflective (AR) coatings

2-13, 6-1, 8, 31 to 32, 34 to 35, 81, 84, 96; P-etch 6-35; physical-vapor deposition 6-32; porous-silica 6-32

Argon-ion laser

6-63, 94

Argus laser system

1-1; 2-50; 3-1, 26, 32 to 33; 5-40, 43 to 44; 6-69

Atomic and Molecular Theory

6-102

Atomic Physics, LASNEX and

3-50

Automated x-ray image analysis system

(AXIAS)

4-16

Axisymmetric phase-mix damping

3-21, 24

B

Baby Optics Diamond Turning Machine

(BODTM)

4-2

Balance-of-Plant Design for the Cascade

Reactor

7-16

Beat-Wave Heating in Inhomogeneous Plasma

Irradiated with 0.53- and 1.06- μ m Light

3-40

Belt conveyor for solid-granule transport

7-19 to 20

Beryllium

columns, in reactor 7-41 to 42, 45; beryllium-lithium-thorium blanket 7-41, 46

Blackbody Pumping and Depumping of

Nd:Glass

6-29

Blanket, in fusion reactors

breeding 7-2, 41 to 45, 67 to 68, 77; thickness 7-8, 13, 24 to 25, 60 to 65

Blanket Design, Dependence of Displacement Damage on

7-65

Blanket Optimization Techniques for Fusion

Reactors

7-60

Bohm-Gross shift

5-44

Borosilicate glass

BK-3, BK-7, and BK-10 6-39 to 40

Branching ratios

6-102 to 103

Bremsstrahlung absorption

1-3; inverse 3-30, 37, 51, 54; 5-40, 46

Brillouin scattering

3-35; 5-10, 40 to 41, 43 to 44, 46

Broad-Band Energy Extraction in Laser Glass

6-48

Bulk diffusion

inventory in Li₂O 7-23 to 25

C

CAMAC data files

5-32

Capacitor bank

2-24

Cascade fusion reactor

1-5; 7-3 to 29

Cascade Reactor for Fusion Energy

7-3

Cassegrainian collector

5-43; See also Nova System Design and Performance 2-2

Chamber Design of the Cascade reactor

7-12

Chamber Shape Experiments for the Cascade reactor

7-4

Charge injection device (CID)

2-33 to 34

Charged-Particle Fusion

3-13

Clear-fused quartz flashlamp

6-24

CO₂ laser

3-34 to 35

Coatings

antireflective 6-1, 8, 31 to 32, 34 to 35, 81, 84, 96; for neutron diode 5-28; P-etch 6-35; Neutral solution process (NSP) 2-2, 10, 15, 17 to 18, 49; physical-vapor deposition 6-32; porous-silica 6-32; sol-gel 1-4, 2-5, 6-32 to 34; target 4-10 to 13; ultraviolet 1-4

Collinear scattering

3-105

Collisional

excitation 5-46, 48, 50; recombination 5-46, 50; redistribution of radiation 6-102

Compact Amplifiers

6-13

Computer codes

3-1; ALICE 7-26 to 27, 30 to 31, 41 to 42, 45 to 46; BASIC 6-1, 11, 32, 43, 77, 88; ENDL 7-41, 64; FORG 7-27, 31, 77; HIPPO 3-15; ICARCON 4-7;

Subject Index

INVERSI 7-73 to 74; INVERSW 7-73;
LASNFX 1-3, 3-1, 19 to 21, 31, 33, 36
to 37, 49 to 55, 62, 64, 66, 2-56; 5-41 to
42, 45; MALAPROP 1-3, 2-3 to 5, 50,
56, 61, 6-77; REFLECT2 5-5; SANDYL
6-96; SOCKITTOME 5-36 to 37;
SUPERSTRUCTURE 6-115; UFO 5-36
to 37; UNSPEC 5-37; VODEP 6-5;
XCSSPARI 5-35; XCSSPROC1 5-35;
models and simulations, GAINPK 6-6 to
7; OREMP 3-41; ORLIB 7-27; TART
7-26; ZOHRAR 3-33 to 34

Computers
LSI-11/23 2-27, 29, 46, 54; 6-3, 9, 20;
VAX 5-31 to 37

Concentration Quenching in Laser Glass
6-40

Controls change request (CCR) for Novette
2-54 to 55

Configuration interaction (CI)
6-91, 93, 109 to 111, 114 to 114

Connective Raman Backscattering: Effects of
En, Noise, Collisions, and Landau Damping,
Thresholds of
3-28

Convergence ratio (CR)
3-11 to 12

Cost of electricity (COE)
7-46 to 47, 50 to 52, 60

Cost of reactor
7-46 to 50

CR-39 plastic film
5-2

Critical density (n_c)
3-28, 30 to 35, 37 to 40, 52, 54 to 55,
5-18, 40 to 41, 56 to 59; for quarter-
critical density ($1/4n_c$) see 3-27, 32 to 35;
5-18, 40 to 42, 44 to 45, 57 to 58

Cryogenic
apparatus for target experiments 4-4 to
5; DT 1-4, 4-5, 7; fuel 1-3; pylon 4-6
to 7; shroud 4-6; target 2-40, 53 to 54;
4-5 to 8

Cryogenic DT, Thermophysical Properties of
4-7

Cryogenic Target Support System, Novette
4-5

Cryogenics in target fabrication
4-4

Crystal arrays
2-8 to 12, 58, 62; see frequency
conversion 2-2 to 4, 8 to 11

Crystal orientation measurement system
(COMS)
2-16

Crystal spectrograph
5-20 to 22, 31, 34 to 36, 50, 55

Crystal Spectrography, Time-Resolved
5-20

Crystalline Lasers, High-Average-Power, Solid-
State
6-94

Crystals, Theory of the Electronic Structure of
Impurities in
6-91

CSLOW data library
5-35

Cyclops laser
6-62

D

Damage
displacement damage to first structural
wall 7-61, 65 to 68; laser damage to
optical materials and coatings 2-5, 15 to
17, 19, 49, 62; 6-35 to 39

Damage in Conversion: Cells in Laser Systems
6-60

Dante Spectrometer, Design of Nova
5-6

Dante x-ray spectrometers
5-4 to 7, 37

Dante X-Ray Spectrometers, Novette
5-4

Data Management and Analysis for Fusion
Experiments
5-31

DECNET inter-computer network
5-32

Demonstration power reactor (DEMO)
7-50, 53 to 54

Detuning
6-70 to 74, 77, 79, 87, 102 to 103

Deuterium-tritium (DT)
1-4 to 5; 3-2, 5 to 11, 26, 42, 64 to 65;
4-1, 5, 7 to 10, 16; 7-18, 20, 26, 34, 41
to 46, 60, 64, 66 to 68; polarized fuel 3-5
to 6

Diagnostics
For fail-safe 5-50 to 54; for fusion
experiments 5-1 to 31; for Nova 2-8,
30 to 39; for soft x-rays 5-6 to 7, 37; for
targets 1-3, 2-37 to 38, 4-1 to 2, 10,
5-42 to 46

Diagnostics in Fusion Experiments
5-1

Diamond turning machine
4-14

Diffusion
in granular solids 7-11; of tritium 7-36;
bulk diffusion inventory in Li_2O 7-23 to
25

Direct-Implicit Plasma Simulations
3-46

Disk targets in fusion experiments
5-9, 23 to 24, 40 to 46

Displacement damage
to first structural wall 7-61, 65 to 68

Dissociative attachment
6-107

Dissociative Electron Attachment, Resonance-
Resonance Coupling in
6-107

DISSPLA graphics library
5-37 to 38

Driver efficiency
7-47 to 54

E

EAGLE ICF reactor
7-54, 56, 59

Edge claddings
6-7 to 12, 15

Electron density contours
6-106

Electron laser facility (ELF)
6-100 to 101

Electronic Shotbook for fusion experiments
data
5-37

Electrooptic switching
6-80

Energy and Military Applications
1-5

Energy conversion
target output to electrical 1-5

Energy Technology Engineering Center
(ETEC)
7-16, 18, 20, 36, 41

Engineering test facility (ETF)
7-50, 53 to 54

Equation of state (EOS)
SESAME 3-56; Thomas-Fermi based
3-55 to 59

Etalons
See Long-Pulse Oscillator and Slicer
Development 2-51 to 52

Etched Nuclear Tracks on laser optical
materials
6-34

Evaluated-Nuclear-Data Library (ENDL) of
cross sections
7-41, 64

Evolution of Rayleigh-Taylor Instabilities in
Multilayered Fluids
3-2

Experimental power reactor (EPR)
7-50, 53 to 54

Experimental test accelerator (ETA)
6-101

Experiments for Fusion
5-39

Exploding-foil experiments
5-40 to 46

F

Fail-safe diagnostic
5-50 to 54

Fast-neutron diodes
5-1

Fast-refreeze process for DT fuel in fusion
targets
4-5, 8

Filamentation
3-26, 30, 35 to 38; length 3-38 to 40

Filamentation by Angular Spread, Stabilization
of
3-35

Film intrinsic stress
4-11

First-layer granules
7-20 to 22; See also Granules

First structural wall (FSW)
7-64 to 68; displacement damage to
7-61, 65 to 68

Fixed-point iteration
3-62 to 64

Fizeau interferometer
See Crystal arrays

Flashlamp reflectors
full-involute 6-13 to 14; half-involute
6-13 to 14; vacuum-deposited silver
6-8; Winston-Rabl 6-7 to 8

- Flashlamps**
clear-fused quartz 6-24; in fusion experiments 5-40, 47, 50 to 51; plasma 6-18, 22, 29; research and development on 6-18 to 31; xenon 6-28 to 31
- Flashlamps in laser amplifiers**
6-18
- Flashlamps, Large-Bore in laser amplifiers**
6-18
- Flashlamps, Transient Response of**
6-23
- Fluorescence-line narrowing (FLN)**
6-48, 50
- FORIG computer code**
7-27, 31, 77
- Fourth-harmonic generation (4HG)**
6-67, 69
- Free-Electron Lasers: Experiments at the Electron-Laser Facility**
6-100
- Free-Electron Lasers (FEL): Experiments on the Rapier-B Amplifier**
6-2, 100 to 102
- Frequency conversion**
research and development on 2-50, 61 to 62, 6-1, 88, 90; *See also Nova Frequency Conversion and Focusing System* 2-8 to 14; *Frequency-Conversion Systems* 6-60 to 79; *Fundamental frequency*, *Second harmonic*, *Third harmonic*, *Fourth-harmonic generation*
- Frequency Conversion, High-Dynamic-Range**
6-69
- Frequency Conversion in MALAPROP, Incorporation of**
6-77
- Frequency-Conversion Systems in laser research and development**
6-60
- Frequency-conversion systems, KDP Growth in**
6-69
- Fresnel reflection**
2-5, 10; 6-8, 81
- Fresnel zone plate (FZP)**
5-19 to 20
- Front end processor (FEP)**
2-26 to 30
- Fuel pin Measurements, CR-39 Neutron Spectrometer for**
5-2
- Fuel-layering techniques**
4-8 to 10
- Fuel-Layering Techniques, Developments in**
4-8
- Fundamental frequency**
1.06 μm (1ω) 1-2 to 3; 3-26, 32, 34, 40 to 42; 5-9 to 11, 15, 33, 40 to 41, 44, 50; research and development for 6-18, 31, 48 to 49, 60 to 61, 69 to 71, 74, 88, 94; 1.05 μm 5-25, 6-3 to 4, 48, 49, 78
- Funding, Laser Fusion Program**
1-5 to 6
- Fused silica**
2-2, 5, 8; 6-1, 32, 34 to 35, 79, 82, 84, 96, 98; *see also Nova lens positioner* 2-13 to 14, *Optical Components for Nova* 2-13 to 17
- Fusion**
gain 7-50 to 53; power plant and reactor, *see Reactor concepts and Reactor development*; waste 7-73, 76
- Fusion experiments analysis facility (FEAF)**
5-31 to 38
- Fusion Target Fabrication (FTF) Program**
4-1
-
- G**
- Gain**
in fusion targets 7-50 to 53
- GAINPK computer model**
6-6 to 7
- Gamma-Ray Streak Camera**
5-8
- Gradient**
scalelength 3-29, 33; stabilization threshold 3-34
- Grady-Glenn theory**
7-59 to 60
- Granular Flow Analysis in the Cascade Reactor**
7-8
- Granule Thermal-Stress Analysis in the Cascade Reactor**
7-20
- Granules**
7-1, 3 to 4, 7 to 25; ceramic 7-1, 3 to 4, 7 to 8, 22; diffusion in 7-11; first-layer 7-20 to 22; flow 7-1, 4, 8, 11 to 12; melting and vaporization 7-20, 22; *See also Li₂O granules*
- Graphite (neutron) reflector**
7-26, 44, 67 to 68
- Grating spectrometer**
5-43
-
- H**
- H₂S, Photodissociation of**
6-103
- Harmonic**
cell design 1-4; conversion 1-1 to 2, 4; 2-49 to 50, 6-1, 74, 84 to 88; *See also Frequency conversion*, *Second harmonic*, *Third harmonic*, *Fourth-harmonic generation*
- Harmonic-Conversion Switch, Plasma-Electrode**
6-84
- HC-56 (halocarbon index-matching liquid)**
6-60 to 61
- Heat**
decay in activation calculations 7-28 to 29, 32, 35, 77; transfer from granules to reactor wall 7-13, 15, 17, 22, 46; treatment 7-2, 32, 69 to 73
- Heat-Transfer Loop and Plant Layout for the Pulse-Star Reactor**
7-36
- Heavy-ion fusion (HIF)**
3-21
- He-Ne laser**
4-6; 6-21, 82
- HIBALL ICF reactor**
7-54, 56, 59
- High-Temperature Experiment in Heavy-Ion**
- fusion, Studies for the**
3-13
- Highly reflective (HR) coatings for targets**
6-29, 107
- HIPPO computer code**
3-15
- Hoop stress**
7-13 to 15
- Hot Electrons by Laser-Excited Plasma Waves, Reheating of**
3-42
- Hot (suprathermal) electrons**
1-3; 3-26, 32 to 36, 42, 44; 5-10, 22, 40, 42 to 45, 57 to 58
- Hot-waste facility in fusion reactors**
7-74 to 76
- Hughes Research Laboratory (HRL)**
4-16 to 18, 21 to 22
- Hydrocarbon film deposition**
4-10 to 11
- Hydrodynamic Models for Simulation of Pinched Beams, Testing**
3-21
- Hydrodynamics model**
breakup theory 7-58 to 59; simulation of pinched beams 3-21 to 25
- HYLIFE ICF reactor**
3-14; 7-1, 13, 18, 26, 41, 54 to 59, 64 to 68
-
- I**
- ICARCON computer code**
4-7
- ICF reactor**
cost 7-46 to 50; *see also Reactor concepts, and Reactor development*
- ICF Reactor Commercialization, Development Facilities for**
7-50
- ICF Reactors, Cost Goals for**
7-46
- Imaging 20- to 100-keV X-Ray Emissions from Laser Fusion Targets**
5-22
- Index-matching fluid (IMF)**
1-2, 4; 6-1, 61 to 63, 67, 81, 84; HC-56 6-60 to 61
- In-flight aspect ratio (IFAR)**
3-10 to 13
- Inner-Shell Excitation, Hole Localization in**
6-109
- Instabilities**
Bremsstrahlung absorption 1-3; *inverse Bremsstrahlung absorption* 3-30, 37, 51, 54; 5-40, 46; *ion-acoustic decay* 5-58 to 59; *oscillating two-stream* 5-58 to 59; *Rayleigh-Taylor* 3-2, 4 to 5, 12 to 13, 36, 56; 4-13 to 15, 5-12; *Richtmeyer-Meshkov* 5-12; *stimulated Brillouin scattering* 5-10, 40 to 41, 43 to 44, 46; *stimulated Raman scattering* 3-26 to 28, 5-45 to 46, 40 to 45; 6-1, 61 to 69; *target-plasma and plasma instabilities* 5-10, 12 to 13, 41 to 42, 44 to 45; *thresholds* 1-3 to 4; *two-photon absorption* 6-39 to 40; *two-plasmon decay* ($2\omega_p$) 3-26 to 27, 32 to 35, 37; 5-10, 41, 57 to 58
- Interfacial tension in targets**
4-8 to 9

Subject Index

Inverse bremsstrahlung absorption
3-30, 37, 51, 54, 5-40, 46; *See also*
Bremsstrahlung absorption

INVERSI computer code
7-73 to 74

INVERSW computer code
7-73

Ionac calibration facility
5-4 to 5

Ion-acoustic-decay (IAD) instability
5-57 to 58

*Ion-Acoustic-Decay Instability in a Finite-
Length System, Electron Heating from the*
5-58

Ion-beam
drill 4-1, 16 to 21; ICF applications
3-13 to 16, 19

Ion-Beam Drill in target fabrication
4-16

Ion Beam Drilling Experiments on Metal Foils
4-18

Ion drilling
4-16 to 21

Irradiation
beryllium-disk 3-34; CH-foil 3-34; gold
disk 3-33 to 34

Isobarically Heated Liquids, Disassembly of
7-54

K

KDP
1-4 to 5; 2-2 to 5, 8, 10, 12 to 13, 15 to
16, 37, 39, 49, 50, 61 to 62; 5-25; 6-1, 60
to 71, 73, 75 to 77, 79 to 88

KD*P
6-70, 73 to 74, 76, 80 to 82, 84, 86, 88

Kidder model of high-gain ICF targets
3-5 to 7

KMS Fusion
1-4; 3-28; 4-8; 5-53; 7-16, 54

KRF laser
6-96, 99 to 100

L

Lamp check diagnostic (LCD)
2-30

Landau damping
5-58

Laser
argon ion 6-63, 94; CO₂ 3-34; Cyclops
6-62; extreme ultraviolet 5-28 to 31, 40;
free-electron 6-2, 100 to 102; KrF 6-96,
99 to 100; large-aperture 1-3; Nd:YAG
6-2, 40, 86 to 87, 94; soft x-ray 1-1;
x-ray 1-1 to 2, 5-1 to 2, 21, 29, 34, 39
to 40, 46 to 54; *See also* Argus, Nova,
Novette

Laser-Damage Studies on Optical Materials
6-35

Laser Data, Processing of
5-32

*Laser Glass in laser research and
development*
6-40

Laser glass types
BK-3, BK-7, and BK-10 6-39 to 40; LG-
750 6-4, 7, 12, 30 to 31, 42 to 43, 46, 48,

50, 57 to 58; LHG-8 6-4, 46, 57; Q-88
6-49; ED-2 6-18, 22, 42, 49; LG-56
6-48 to 49, 55; LG-650 6-42 to 43, 9 to
50, 57 to 58; LG-660 6-7, 11, 42 to 43;
LG-670 6-42 to 43

Laser Glasses, Comparison of
6-56

Laser Operations and Systems
1-2

Laser-plasma interaction
1-1 to 2; 5-1, 31, 39 to 46, 50

Laser-Plasma Interaction Experiments, Novette
5-41

Laser-Produced Plasma, Line Diagnostics of
5-55

Laser Research and Development
1-4

Laser Systems, Other
6-96

Laser Targets
3-2

LASNEX hydrodynamics computer code
1-3; 3-1, 19 to 21, 31, 33, 36 to 37, 49 to
55, 62, 64, 66; 2-56; 5-41 to 42, 45

Li₂O granules
7-4, 8, 10, 12 to 13, 15, 17 to 18, 20 to 29

LIOT
7-12, 23 to 25

LiPb
inventory 7-35 to 36; pumps 7-2, 30, 37

Li₇Pb₃₁
7-1, 33 to 36, 54 to 55

*Liquid-Metal Embrittlement, Susceptibility of
2-1/4 Cr 1 Mo Steel to*
7-68

Long pulse oscillator
See Nova Front End Subsystem 2-21

Los Alamos Modified Wetted Wall (MWV)
ICF reactor
7-54

M

MALAPROP propagation computer code
1-3; 2-3 to 5, 50, 56, 61; 6-77

Master oscillator
2-49, *See also* Novette Laser System 2-55
to 69

Master-oscillator power amplifier (MOPA)
6-1, 59 to 60

Master-oscillator room (MOR)
2-1, 5 to 7, 20 to 26, 28, 34 to 35, 41, 53;
See also Nova Front End Subsystem

Materials and Coatings for targets
4-10

*Measurement Development for targets and
materials*
4-15

Metaphosphate glass
LG-750 6-4, 7, 12, 30 to 31, 42 to 43, 46,
48, 50; LHG-8 6-4, 46, 57, Q-88 6-49

Meyer-Ter-Vehn (MTV) model of high-gain
ICF targets
3-5 to 9

Microchannel plate (MCP)
electron multiplier 5-9 to 10, 23, 42, 53;
glass 5-16; intensifier tubes 5-14

Mirrors
2-16 to 17, 43

Microspots
4-10 to 14

*Model for Gain vs Laser Energy in Direct-
Drive Targets*
3-9

Molecular Collisions, Resonances in
6-106

Molecular Collisions, IDHF Theory of
6-105

Molecular photodissociation
6-103

*Multi-group Radiation-Matter Coupling
Iteration Convergence Acceleration for*
3-62

Multipass amplifier systems
6-13, 57 to 60, 84, 86, 88, 90
Multipass Cavities, Subharmonic Injection into
6-88

N

Naval Research Laboratory (NRL)
3-15, 35 to 36

Nd:glass
6-5 to 6; amplifier 6-7 to 8, 19;
pumping 6-22 to 24, 29 to 30

Nd:YAG laser
6-2, 40, 86 to 87, 94

Near-Axis Imaging System, Novette
5-18

Neutral solution process (NSP)
2-2, 15, 17 to 19; *See* coatings

Neutron Diode, 50-ps
5-28

*Neutronics and Activation Analyses for the
Cascade Reactor*
7-25

*Neutronics and Activation Analyses for the
Pulse-Star Reactor*
7-30

Newton iteration
3-62

Noncollinear Phase-Matching Schemes
6-74

Nova
alignment controls for 2-34 to 35; crystal
arrays for 2-8 to 11; diagnostic sensor
packages for 2-36 to 39; future ICF
experiments 5-1, 6 to 8, 12, 53; lens
positioner for 2-13 to 14; present status
summary for 2-6 to 7; project
management for *See* Nova Project
Management 2-47 to 48; project
milestones 2-6; pulse shaping for *See*
Nova Front End Subsystem 2-20 to 24;
spaceframe 2-42; research and
development for 6-1 to 11, 13, 17, 31 to
32, 42, 46, 57, 61 to 62, 72, 74, 76 to 79;
summary of achievements 1-1 to 7;
system design for 2-2 to 5; target
chamber of 2-39 to 41; target design for
3-1; target fabrication for 4-1, 5; *See*
also Table of Contents for Section 2

Nova Amplifiers, Gain Performance of
6-2

*Nova Beam Alignment and Diagnostics
Systems*
2-30

Nova Frequency Conversion and Focusing System

2-8

Nova Front End Subsystem

2-20

Nova Integrated Controls

2-45

Nova Mechanical Systems

2-41

Nova Power Systems Energy Storage

2-24

Nova Project Management

2-47

Nova System Activation

2-5

Nova System Design and Performance

2-2

Nova Target Systems

2-39

Novanet

2-27, 35, 46 to 47, 54

Novette

design of 2-49 to 50; frequency conversion of 2-61 to 62; fusion experiments program 5-1 to 59; integration of Novette/Nova controls 2-51 to 55; performance of 2-55 to 62; research and development for 6-1, 3, 8, 60 to 62, 67, 69, 76 to 77, 113; summary of achievements 1-1 to 5; target design for 3-1 to 70; target fabrication for 4-1 to 2; See also *Novette Laser System* 2-48 to 62

Novette Integrated Test System (ITS)

4-4 to 5

Novette Laser System

2-48

O*Optical coatings*

sol-gel 1-4, 6-32 to 34; ultraviolet 1-4; See also *Antireflective coatings*, *Highly reflective coatings* and *Optical Components for Nova* 2-14

Optical Components for Nova

2-14

Optical Materials, Surfaces, and Coatings in laser research and development

6-31

Optical streak CCD camera and data files

5-13, 31 to 35

Optical switches

6-1, 79 to 91

Optical Switches, Large-Aperture

6-79

Optical and X-Ray Signals, Streaked Recording of Simultaneous

5-10

Optical/x-ray (OX) streak camera

5-7, 10 to 12, 42, 44 to 45

ORACLE Data Base Management System, Applications of the

5-32

ORACLE data base management system (DBMS)

5-31 to 33, 38

OREMP electromagnetic particle code

3-41

ORLIB computer code

7-27

Oscillating, two-stream (OTS) instability

5-57 to 58

Oscillator controls

2-26 to 28

P*Parylene deposition*

4-11

P-e'ch

6-35

Phase-mix damping

3-21, 24

Photodiodes

theory of 3-48 to 50

Photodiodes, Theory of

3-48

Photomissile Materials for 0.35- μ m

Wavelength

5-7

Plasma

density 5-20, 41, 45, 50, 55; in flashlamps 6-18, 22, 29; polymerization 4-1 to 2, 11 to 13; shutter 2-26; temperature 5-20, 40, 49, 56

Plasma Physics

3-25

Pockels cell

2-23, 25 to 26; 6-2, 79 to 84, 86

Pockels Cell, Plasma-Electrode

6-79

Polymer fuel containers

1-3

Ponderomotive

force 3-35, 41; self focusing 3-37

Porous-silica coatings

6-32

*Potassium dideuterium phosphate (KD*P)*

6-70, 73 to 74, 76, 80 to 82, 84, 86, 88

Potassium dihydrogen phosphate (KDP)

1-4 to 5; 6-1, 60 to 71, 73, 75 to 77, 79 to 88

Precision Engineering Research Lathe

4-2

Precision Engineering Research Lathe (PERL)

4-2 to 3

Pressure of resonance states

3-59 to 62

Pressurized-water reactor (PWR)

7-2, 46 to 50

Production of Targets

4-2

Program Facilities

1-6

Program Resources

1-5

Propagation codes

1-3

Pulsed-power controls

2-28 to 29

*Pulse*Star ICF reactor*

1-5; 7-54; bell jar vessel 7-1, 29 to 33, 39

*Pulse*Star Reactor for Fusion Energy*

7-29

Q*Quarter-critical density (1/4n_c)*3-27, 32 to 35; 5-18, 40 to 42, 44 to 45, 57 to 58; See also *Critical density***R***Radiation, Collisional Redistribution of*

6-102

Random phase shifter (RPS)

3-35

*Raman*See *Spontaneous Raman scattering* and *Stimulated Raman scattering**Raman Scattering in KDP and Index Matching Fluids, Spontaneous and Stimulated*

6-61

Rapier-B amplifier

6-2, 96 to 100

*Rare-Gas-Halide Lasers: Experiments on the**Rapier-B Amplifier*

6-96

Ray tracing

3-1

Rayleigh-Taylor (R-T) instability

3-2, 4 to 5, 12 to 13, 36, 56, 4-13 to 15

Rayleigh-Taylor Targets

4-13

Reactive-ion etching (RIE)

4-11 to 12

Reactor concepts

1-3, 7-54; Cascade 1-5, 7-3 to 29; EAGLE 7-54, 56, 59; HIBALL 7-54, 56, 59; HY LIFE 3-14, 7-1, 13, 18, 26, 41, 54 to 59, 64 to 68; Los Alamos Modified Wetted Wall (MWW) 7-54, Pulse*Star 1-5; Starburst 1-5

Reactor cost

7-46 to 50

Reactor development

belt conveyor for solid-granule transport 7-19 to 20; beryllium columns in hybrid reactors 7-41 to 42, 45; blanket thickness 7-8, 13, 24 to 25, 60 to 65; breeding blanket in hybrid reactor chamber 7-2, 41 to 45, 67 to 68, 77; centrifugal acceleration 7-4, 7, 9; centrifugal force 7-3, 7 to 8, 13 to 15, 68; cone experiments 7-6 to 8; Coriolis force 7-4, 8, 10

Reactor Development Technologies, General Studies and

7-46

Rectangular-Slab Amplifiers, Parasitic Oscillations in

6-11

REFLECT2 computer code

5-5

Reflectors

neutron (graphite) 7-26, 44 67 to 68

Resonance States, Thermal Occupation and Pressure of

3-59

Resonant photoexcitation

5-40, 46, 48, 51, 53 to 54

Richtmeyer-Meshkov instability

5-12

Subject Index

Rockwell International
4-2

S

SANDYL, computer code
6-96

Scanning electron microscopy (SEM)
4-16 to 17, 20 to 21, 5-20; 7-71

Scanning ion probe
4-16 to 21

Scattering Amplitudes Using Complex Basis
Function
6-111

Second harmonic (0.53 μm , 2 ω)
1-1 to 4; 3-1, 26, 30, 34, 40 to 42; 5-1,
10, 23 to 25, 28, 33, 39, 45, 47, 54 to 55;
research and development for 6-60 to
61, 66, 67 to 70, 85, 88; second-harmonic
generation (SHG) 6-14, 67, 69, 75 to 76,
79, 84 to 86, 90

Secondary-ion mass spectrometer (SIMS)
4-16 to 17

Seeded Microspot X-Ray Diagnostic Targets
4-11

Seeded polymer
4-13

Self focusing
length 3-37 to 40; ponderomotive 3-37;
relativistic 3-37; thermal 3-26, 37 to 38,
40

SESAME equation of state
3-56

Shiva
1-2; 2-48 to 49; 3-1, 32 to 34

Short-pulse oscillator
See Nova Front End Subsystem 2-20 to 22,
25

Silicate glass
ED-2 6-18, 22, 42, 49; LG-56 6-48 to 49,
55; LG-650 6-42 to 43, 49 to 50, 57 to 58
LG-660 6-7, 11, 42 to 43; LG-670 6-42
to 43

Silicone Edge Claddings
6-8

Simple Models of High-Gain Targets—
Comparisons and Generalizations
3-5

Six-Segment Amplifier in Solid State
Amplifiers
6-7

Sinusoidal targets
4-13 to 15

SOCKITTOME Octopus computer code
5-36 to 37

Soft X-Ray Laser Experiments, Novette
5-46

Soft x-ray laser targets
4-1 to 2

Soft X-Ray Lasers, Atomic Physics of
Laboratory
6-113

Solarization
6-39 to 40

Sol-Gel Coatings
1-4; 6-32 to 34; damage threshold for
2-5; See also Optical Components for Nova
2-19

Sol-Gel Coatings on laser optical materials
6-32

Solid-Granule Transport System for the
Cascade reactor
7-18

Solid-State Amplifiers in laser research and
development
6-2

Solid-State Systems, Advanced, in laser
research and development
6-91

Spaceframe
for Nova 2-42

Spatial filters
2-43

Spatially discriminating streak camera
(SDSS)
5-4, 6, to 7

Spectral hole burning
6-48, 53, 58

Spectral Measurements Above 80 Å, Time-
Resolved
5-28

Spectrometers
1-2; 5-1, 2, 4 to 7, 22, 35, 43

Sputtering efficiency
4-18 to 20

Starburst ICF reactor
1-5

Steam generator design
counterflow 7-37, 39 to 40; crossflow
7-37 to 38, 40

Steffenson acceleration
3-63

Stimulated Brillouin scattering (SBS)
5-10, 40 to 41, 43 to 44, 46

Stimulated Compton scattering
3-30

Stimulated Raman scattering (SRS)
2-8, 12 to 13; 3-26 to 28; 5-10, 40 to
45; 6-1, 61 to 69

Stimulated Raman Scattering in an
Inhomogeneous Plasma with an Obliquely
Incident Pump
3-26

Streak cameras
optical/x-ray (OX) 5-7, 10 to 12, 42, 44
to 45; spatially discriminating streak
camera (SDSS) 5-4, 6 to 7; x-ray streak
camera (XSC) 5-8 to 10, 20 to 22, 28 to
31

Streaked Crystal Spectrograph, Processing of
Data from the
5-34

Stress birefringence
2-11 to 12

SUPERSTRUCTURE computer code
6-115

Suppressed-Fission ICF Hybrid, Neutronics
Studies of a
7-41

Suprathermal electrons
1-3; 3-26, 32 to 36, 42, 44; 5-10, 22,
40, 42 to 45, 57 to 58

Suprathermal x-ray spectrum
5-23

T

Target

alignment 1-3; anti-mix layer 3-2 to 4;
coating 4-10 to 13; diagnostic targets
1-3, 4-1 to 2, 10; direct drive of 1-2 to
4; high-gain 1-1, 3; high-Z 1-3 to 4;
implosion 1-3; shell and plasma
instabilities 5-10, 12 to 13, 41 to 42, 44 to
45; surface analysis 4-16; See also
following entries

Target and Material Measurements
4-15

Target Assembly and Microfabrication
4-2

Target chamber
See Target Systems

Target Design
1-2 to 3, 3-1 to 70

Target Design, Target Fabrication, and Fusion
Experiments
1-3

Target Fabrication
1-3, 6; 4-1 to 22; direct precision
machining 4-14, Fusion Target
Fabrication (FTF) Program 4-1;
hydrocarbon film deposition 4-11; ion
milling 4-14; liquid molding 4-15;
masking 4-11 to 13; micromachining
4-1 to 2, 15, 21; microspot definition
4-11 to 13; parallel deposition 4-11;
replication from precision-machined mold
4-14; solution casting 4-11, 14

Target plane imager
2-33

Target positioner
2-40 to 41

Target Systems
2-39 to 41

Target types
ball-in-film 4-2 to 3; ball-on-stalk 4-3;
disk 4-1 to 3, 14, 5-9, 23 to 24, 40 to
46; x-ray laser 1-3; 4-1 to 3

TART neutronics computer code
7-26

Thermal gradient
inside glass spheres 4-8

Thermal Whole-Beam Self Focusing, 2-D
LASNEX Ray-Trace Calculations of
3-36

Thermal-Lensing Studies: Two-Photon
Absorption and Solarization
6-39

Thermo-mechanical Properties of laser glass
6-44

Thermonuclear neutrons
3-51

Thermonuclear Neutrons, Energy Spectrum of
3-64

Thermophysical properties of cryogenic DT
4-5, 7

Third harmonic (0.35 μm , 3 ω)
1-1, 3 to 4; 3-32 to 34; 5-7 to 8, 33,
43, 46; research and development for
6-61, 68 to 69, 77, 79; third-harmonic
generation (THG) 6-67 to 68

Thomas-Fermi-Based Equation-of-State Package
3-55

Three-Dimensional Laser Ray Tracing on an r - z Langrangian Mesh

3-51

Tritium

breeding 7-41, 60 to 61, 68; breeding ratio 7-18, 43 to 46, 64 to 65; diffusion 7-36

Tritium Containment

7-30, 33 to 36

Tritium Containment in the Pulse-Star reactor

7-33

Tritium inventory and recovery

7-22 to 25

Tritium Inventory and Recovery in the Cascade reactor

7-22

Two-photon absorption (2PA)

6-39 to 40

Two plasmon decay ($2\omega_p$)

3-26 to 27, 32 to 35, 37; 5-10, 41, 57 to 58

Two-Plasmon Instability and Electron Heating, Measurements of the

5-37

Type I KDP

6-75 to 76, 80, 84, 86 to 88. See also KDP

Type II KDP

6-63 to 64; See also KDP

U

UFO Code on the LEAF VAX Computer

5-36

Ultraposphates in Laser Glass

6-46

Ultraviolet coatings

1-4

Uniform liquid hydrogen (fuel) layer

4-8

University of Illinois

4-8, 10

University of Rochester

3-35 to 36; 4-2

UNSPEC computer code

5-37

V

Vacuum system

control system for 2-40; design of 2-40

VAX computer

2-27 to 30, 33, 45 to 47, 51, 54; 5-31 to 37

VODEP computer code

6-5

W

Waste

hot-waste facility 7-74 to 76; fusion

7-73, 76

Waste Disposal in Fusion Reactors

7-73

X

XCSSPAR1 computer code

5-35

XCSSPROC1 computer code

5-35

Xenon flashlamp

6-28 to 31

X-Ray Backlighting, Co-Propagating Beam on Novette for

5-24

X-ray Calibration and Standards Laboratory

5-6

X-ray diagnostics

5-6 to 7, 37, 53 to 55

X-Ray Framing Application: Design of Ultrafast Gated Intensifier for Laser Fusion

5-12

X-ray laser

1-1 to 2; 5-1 to 2, 21, 29, 34, 37-40, 46 to 54; targets 1-3, 4-1 to 3

X-Ray Laser "Fail-Safe" Diagnostics: Performance and Results

5-53

X-ray microscope

5-29, 42, 50

X-Ray Phase Lens

5-19

X-ray spectroscopy

1-2; 5-20, 37, 40 to 42, 50, 55; Dante 5-4 to 7, 37

X-ray streak camera (XSC)

5-8 to 10, 20 to 22, 28 to 31

XUV (extreme ultraviolet) laser

5-28 to 31, 40

Z

Z (atomic number)

high-Z targets 1-3 to 4; low-Z fuel 1-2

ZOHAR particle code

3-33 to 34

Zone plates

5-19 to 20

Author Index

All authors and major contributors are listed. Page numbers indicate where articles begin, not where authors' names are found.

A

Afeyan, B. (I.I.E., Rochester) 3-26
 Albritton, I. R. 3-42
 Altbaum, I. P. 1-5
 Auerbach, J. M. 5-31, 32, 34, 36, 37

B

Bailey, D. S. 3-37
 Bangert, R. O. 3-13
 Beeman, B. A. 2-36
 Bender, D. A. 2-36
 Benzel, D. M. 2-36
 Bernat, I. P. 4-4, 5, 8
 Berwald, D. H. (FRW) 7-41
 Blink, I. A. 7-1, 12, 16, 18, 25, 29, 30, 33, 36, 50, 54, 68, 73, 76
 Bliss, F. S. 2-30
 Bolt, R. D. 2-36
 Boyd, I. K. 3-21
 Boyd, R. D. 2-31, 34
 Bradley, I. P. 6-18, 23
 Brown, N. I. 2-14
 Browning, D. E. 2-20
 Bryan, J. B. 4-14
 Buksa, J. J. (University of New Mexico) 7-22
 Burch, R. E. 2-24
 Bushman, J. E. 5-8
 Buzzard, J. (University of Illinois) 4-8

C

Campbell, D. E. 5-28
 Campbell, E. M. 3-28, 32, 5-1, 39, 41, 53
 Campbell, I. H. 4-18
 Carey, R. W. 5-31, 32, 37
 Carmighia, C. K. (Optical Coating Laboratory, Inc.) 6-35
 Ceglio, N. M. 5-1, 19, 28
 Chaffee, P. H. 2-20
 Chen, V. K. 4-10
 Christie, D. I. 2-48
 Ciardo, D. R. 4-14
 Clouser, R. W. 4-14
 Cook, W. M. 5-12
 Cordial, C. S. 4-10
 Corley, R. A. 4-10
 Cormier, J. L. 7-4
 Craig, G. D. 3-48
 Creedon, R. L. (GA Technologies) 7-12

D

Darling, D. H. 4-5, 8
 DeGroot, A. J. 2-45, 48
 DeGroot, J. S. (UC Davis) 5-57, 58
 Demaret, R. D. 2-34
 DeMartini, B. J. 5-32, 37
 Denavit, J. 3-48
 Dorn, D. W. 7-73
 Douglas, D. 2-34

Drake, R. P. 3-28, 32; 5-10, 41
 Dreifuerst, G. R. 2-24
 Darius, R. C. 2-14; 7-4
 Duffy, J. M. 2-45

E

Eberhard, B. A. (CSM) 7-68
 Edwards, G. R. (CSM) 7-68
 Eimerl, D. 6-69, 74, 77, 88
 Ellis, R. J. 5-10
 Estabrook, K. G. 3-28, 37, 40; 5-41
 Evans, J. H. 2-36

F

Fahy, R. F. (Lincoln Laboratory) 6-94
 Fertig, R. A. 2-34
 Ford, D. H. 6-2
 Frick, J. A. 2-41; 5-24; 6-7
 Friedman, A. 3-46, 51, 62

G

Gee, W. F. 6-96; 7-4
 Godwin, R. O. 2-1
 Goldfarb, I. 6-79, 84, 96
 Gordon, J. D. (TRW) 7-41
 Greiner, G. J. 4-10, 16, 18
 Griffith, L. V. 4-7
 Griffith, R. L. 5-8
 Gritton, D. G. 2-30, 48

H

Haas, R. A. 6-84; 7-54
 Hagen, W. F. 6-13, 48, 56
 Hale, C. P. 5-7, 8
 Halsey, W. G. 4-15, 16
 Haney, K. I. 2-39
 Harwood, Jr., G. W. 2-31
 Hatcher, C. W. 4-2, 14; 5-41, 53
 Hawryluk, A. M. Editor, Section 4; 5-19
 Hazi, A. U. 6-107, 113
 Hendricks, C. D. Editor, Section 1
 Henesian, M. A. 6-61, 79, 84
 Hermes, G. L. 2-34
 Hernandez, A. 4-16
 Hewett, D. W. 3-46
 Hildum, J. S. 2-48
 Hill, J. R. 2-45
 Hindley, T. W. 6-39
 Hockett, I. P. 2-31
 Hoffman, N. J. (ETEC) 7-33
 Hogan, W. J. Editor, Section 7; 7-1
 Holcomb, F. J. 2-47
 Holloway, F. W. 2-34, 45
 Holloway, R. W. 2-24; 6-2
 Holzrichter, J. F. 1-1, 2
 Hoover, W. G. 7-54
 Hopper, R. W. 6-34
 Horton, R. L. 2-39
 Hovingh, J. 7-54, 68

Hsieh, E. J. 4-14
 Hugenberger, R. E. 2-36
 Hurley, C. A. 2-41
 Hurst, K. A. 2-39

I

Ives, B. H. 4-18

J

Jacobson, G. F. 4-16
 Jameson, G. T. 4-14
 Jancaitis, K. S. 6-11, 18, 23, 29, 96
 Johnson, B. C. 2-8; 6-60
 Johnson, D. 2-24
 Jones, W. A. 2-31

K

Kalibjian, I. R. 5-12
 Kauffman, R. L. 3-20, 28, 41, 46, 55
 Kilkenny, J. D. 3-55
 Kim, K. (University of Illinois) 4-8
 Kimmions, J. F. 4-10
 Kopp, R. (Los Alamos) 3-13
 Krammen, I. E. 2-45, 48
 Krueger, W. L. 3-25, 28, 32, 37, 40; 5-41
 Krupke, W. E. 6-94
 Kuizenga, D. J. 2-20; 5-24
 Kulander, K. C. 6-102, 103, 105, 106
 Kumpf, D. C. 2-39
 Kushner, M. I. 6-23

L

Lane, S. M. 5-2
 Langdon, A. B. 3-32, 35, 46
 Lasinski, B. F. 3-28, 32; 5-41
 Lee, F. P. (LBL) 3-13
 Lee, G. C. 6-7
 Lee, P. H. Y. 5-7
 Lee, R. W. 5-55
 Leipelt, G. R. 5-8, 22, 28, 53
 Lenz, D. J. 4-10
 Lerche, R. A. 5-31, 37
 Letts, S. A. 4-10; 5-41
 Levine, H. E. (MIT) 7-20
 Lindl, J. D. 3-1, 5, 9
 Lopert, R. 6-61
 Lowdermilk, W. H. 6-1
 Lutz, J. A. 6-79, 84, 96
 Lynch, J. E. 6-40, 46

M

Manes, K. R. 2-48; 5-24
 Maninger, R. C. 7-73
 Marchi, F. T. 2-14
 Mark, J. W-K. 3-13, 21
 Martos, A. 2-41
 Masquelier, D. A. 6-18, 96
 Matthews, D. L. 5-46, 53, 55
 Maya, I. (GA Technologies) 7-12, 20, 22
 Mazuch, C. M. 4-10

Author Index

McCardle, R. A. 1-6
McCarthy, P. L. 4-10
McCarville, T. C. (TRW) 7-41
McCurdy, C. W. (The Ohio State University) 6-111
McDowell, M. W. (ETEC) 7-16, 18, 36
McGee, R. A. 6-60
McGuigan, D. L. 2-34, 45
Mead, W. C. 3-32
Medeck, H. 5-7, 8, 10, 28
Meier, W. R. 7-12, 25, 41, 46, 60, 65
Merrill, R. 2-20
Merritt, B. T. 2-24; 6-2
Michel, C. 7-4
Mihalka, A. M. 2-24
Mikaelian, K. O. 3-2
Milam, D. 6-31, 32, 35, 60
Milanovich, F. P. 6-61
Miller, J. I. 4-5
Mizuno, K. (UC Davis) 5-57, 58
Mok, L. (University of Illinois) 4-8
Monsler, M. J. (KMS Fusion) 7-12, 50
More, J. M. 3-55, 59
Moulton, P. F. (Lincoln Laboratory) 6-94
Mullineaux, B. A. (CSM) 7-68
Murray, J. E. 6-2, 8
Murray, J. R. Editor, Section 6; 6-96
Murray, K. A. (ETEC) 7-12, 16, 18, 36
Myers, B. 4-5
Myers, D. W. 2-36

N

Nelson, D. G. 5-4, 6, 22, 28

O

Oicles, J. A. 2-24
Orel, A. E. 6-23, 109
Orzechowski, T. J. 6-100
Ozarski, R. G. 2-30, 36

P

Padilla, E. 2-24
Pasha, R. A. 6-79, 84
Patton, H. G. 2-41
Peluso, S. E. 6-35, 60, 61
Phillips, D. W. 3-28, 32; 5-18, 24, 41
Phillips, G. E. 5-10
Pitts, J. H. 7-3, 4, 8, 12, 16, 18, 20, 22, 25
Pitzer, R. M. 6-91
Poli, R. J. 6-8
Powell, H. T. 6-2, 7, 8, 18, 23, 29
Price, R. H. 5-12
Prior, J. A. 2-36
Prosnitz, D. 6-100

R

Rainer, F. 6-32, 35
Rajak, K. 6-91
Rambo, P. W. (UC Davis) 5-58
Randall, C. J. 3-28

Rash, M. K. 2-45
Rasmussen, P. S. 4-5
Rebentrost, F. (Max-Planck Institute) 6-102
Reed, R. K. 4-5
Rescigno, T. N. 6-109, 111
Richards, J. B. 2-48; 5-10
Rienecker, F. 2-39
Ries, E. A. 5-6
Riley, M. O. 6-8
Roberts, D. H. 4-5
Rodriguez, S. B. 2-39
Roelant, D. F. (University of Michigan) 7-22
Rogers, J. (UC Davis) 5-57
Rosen, M. D. 3-5, 9
Ryon, R. W. 4-10

S

Sanchez, J. J. 4-5
Saunders, R. L. 2-31
Schattenburg, M. 5-19
Schenz, R. F. 4-16, 18
Schlicht, M. 2-45
Schlitt, I. G. (Pulse Sciences, Inc.) 6-96
Schultz, K. R. (GA Technologies) 7-12, 20, 22
Schwab, T. L. 2-39
Sclippa, G. C. 2-14
Seka, W. 3-28
Seppala, L. G. 2-8, 14, 31, 36; 5-18
Severyn, J. R. 2-45; 4-5
Sharp, W. M. 3-21
Shelton, R. T. 2-34
Simmons, W. W. 2-2, 20, 48; 6-77
Slivinsky, V. W. Editor, Section 5; 5-28, 53
Smart, J. A. 2-24, 25
Smith, W. L. 6-61
Snyder, K. 2-24
Speck, D. R. 2-5
Staggs, M. C. 6-32, 35
Stevens, G. (Hughes Research Lab) 4-16
Stokowski, S. E. 2-14; 6-40, 44, 46, 69
Stowers, I. F. 4-1
Summers, M. A. 2-8; 6-60
Suski, G. J. Editor, Section 2
Suter, L. J. Editor, Section 3
Swift, C. D. 2-31

T

Taylor, J. W. 2-45
Thiessen, A. R. 3-5
Thomas, I. M. 6-32
Thomas, S. W. 5-8, 12
Thombley, R. L. 2-24
Thompson, C. E. 6-18
Thompson, D. C. 4-2
Tillotson, T. M. 4-10
Tisell, K. G. 5-4, 6
Tiszauer, D. H. 6-105

Trenholme, J. B. 6-7
Trevino, J. C. 4-14
Turner, R. E. 3-28, 32; 5-10, 41
Tuttle-Hart, T. (Optical Coating Laboratory, Inc.) 6-35

U

Uroff, D. J. 5-24

V

VanArsdall, P. J. 2-34
Vercimak, C. L. 6-60, 61
Vergino, M. D. 6-7
Victoria, J. (EG&G) 5-24

W

Wakakuwa, M. M. 5-37
Walkur, C. M. 4-10
Wallerstein, F. P. 2-14
Walmer, C. D. 2-14
Walton, O. K. 7-4, 8
Wang, C. L. 5-8, 22, 28, 41, 53
Ward, J. 4-18
Weber, M. J. 6-39, 40
Whistler, W. T. 2-14
White, J. T. 6-39
Whitham, K. 2-24
Whitten, B. L. 6-113
Wiedwald, J. D. 5-8, 12
Wilcox, R. B. 2-20, 24, 48
Wilder, J. G. 6-32
Wilkinson, L. L. 2-36
Williams, E. A. 3-26, 42; 5-41
Williams, J. D. 2-8; 6-60
Wintemute, J. D. 2-31
Winter, N. W. 6-91
Wirtenson, G. R. 2-14
Witt, L. A. 4-10
Woo, W. (UC Davis) 5-57, 58
Woods, B. W. 6-2, 8

Y

Yancey, R. L. (Kaiser Engineers) 7-4
Yu, S. S. (SIAC) 3-21

Z

Zaragoza, G. J. 2-34
Ze, F. 5-37
Zimmelman, G. B. 3-51, 55, 64

Scientific Editor Charles D. Hendricks

Publication Editors M. Louise Rufer
Peter W. Murphy

Editors Randall S. Frost
Derek P. Hendry
Robert D. Kirvel
Howard L. Lentzner
Kevin E. Lewis
Judith K. Prono

Editorial Assistants Sharilyn A. Tindell
Lori A. Brown

Lead Designer L. William Fulmer

Designer Brett S. Clark

Layout and Art Coordinator Wilma T. Leon

Layout Gary L. Alexander

Lead Compositor Susan S. Gray

Compositor Tracy L. Dwinell

Lead Proofreader Delores J. Mason

Proofreaders Yvernia Hatcher
Jill S. Silvers

Printing Coordinators Melvin G. Moura
Margaret E. Sylvester

DISCLAIMER

This document was prepared as an account of work sponsored by an agency of the United States Government. Neither the United States Government nor the University of California nor any of their employees, makes any warranty, express or implied, or assumes any legal liability or responsibility for the accuracy, completeness, or usefulness of any information, apparatus, product, or process disclosed, or represents that its use would not infringe privately owned rights. Reference herein to any specific commercial products, process, or service by trade name, trademark, manufacturer, or otherwise, does not necessarily constitute or imply its endorsement, recommendation, or favoring by the United States Government or the University of California. The views and opinions of authors expressed herein do not necessarily state or reflect those of the United States Government thereof, and shall not be used for advertising or product endorsement purposes.

Work performed under the auspices of the U.S. Department of Energy by Lawrence Livermore National Laboratory under Contract W-7405-Eng-48.