ANL/ACL-92/3

ANL/ACL--92/3

DE93 003217

Argonne National Laboratory
9700 South Cass Avenue
Argonne, Illinois 60439

THE STABILITY OF CALIBRATION STANDARDS FOR
ICP/AES ANALYSIS: SIX-MONTH STUDY

Edmund A. Huff and Doris R. Huff

Analytical Chemistry Laboratory
Chemical Technology Division

May 1992

MASTER

NO REPRODUCTION OF THIS DOCUMENT IS PERMITTED

ed

TABLE OF CONTENTS

	<u>Page</u>
ABSTRACT	1
INTRODUCTION	1
EXPERIMENTAL	2
RESULTS AND DISCUSSION	3
CONCLUSION	6
ACKNOWLEDGMENTS	7
REFERENCES	7

LIST OF TABLES

<u>No.</u>	<u>Title</u>	<u>Page</u>
1.	Instrumentation	8
2.	Operating Conditions	9
3.	Detection Limits	10
4.	Summary of Standard Solutions	11
5.	Weight Loss by Diffusion	13
6.	Signal-to-Background Ratio vs. Acidity	14
7.	Stability Data for Solution QF01	15
8.	Stability Data for Solution QF02	16
9.	Stability Data for Solution QF03	17
10.	Stability Data for Solution QF04	18
11.	Stability Data for Solution QF05	19
12.	Stability Data for Solution QF06	20
13.	Stability Data for Solution QF07	21
14.	Stability Data for Solution QF08	22
15.	Stability Data for Solution QF09	23
16.	Stability Data for Solution QF10	24
17.	Stability Data for Solution QF11	25
18.	Stability Data for Solution QF12	26
19.	Stability Data for Solution QF13	27

LIST OF TABLES (Contd)

<u>No.</u>	<u>Title</u>	<u>Page</u>
20.	Stability Data for Solution QF14	28
21.	Stability Data for Solution QF15	29
22.	Stability Data for Solution QF16	30
23.	Stability Data for Solution QF17	31
24.	Stability Data for Solution QF18	32
25.	Stability Data for Solution QF19	33
26.	Stability Data for Solution QF20	34
27.	Stability Data for Solution QF21	35
28.	Stability Data for Solution QF22	36
29.	Stability Data for Solution QF23	37
30.	Stability Data for Solution QF24	38
31.	Stability Data for Solution QI01	39
32.	Stability Data for Solution QI02	40
33.	Stability Data for Silver Solutions	41
34.	Stability Data for Tin Solutions	41
35.	Stability Data for Solutions QSi	42
36.	Stability Data for Solutions QAs	44
37.	Stability Data for Niobium Solutions	45
38.	Stability Data for Tungsten Solutions	45

LIST OF FIGURES

<u>No.</u>	<u>Title</u>	<u>Page</u>
1.	Signal-to-Background Ratio vs. Acidity	46
2.	Relative Standard Deviation for Aluminum.....	47
3.	Relative Standard Deviation for Beryllium	47
4.	Relative Standard Deviation for Chromium	48
5.	Relative Standard Deviation for Nickel	48
6.	Relative Standard Deviation for Lead	49
7.	Relative Standard Deviation for Tin	49

THE STABILITY OF CALIBRATION STANDARDS FOR ICP/AES ANALYSIS: SIX-MONTH STUDY

by

Edmund A. Huff and Doris R. Huff

ABSTRACT

The stability of instrument calibration standards for Inductively Coupled Plasma/Atomic Emission Spectrometric (ICP/AES) analysis was studied over a six-month period. Data were obtained as functions of analyte concentration, acid type, and acidity. The impact of acid concentration on signal-to-background ratios (S/B) was also assessed. The results show that analytes maintain their integrity over extended periods with appropriate inorganic acid preservatives. Thus, frequent standard preparations become unnecessary to obtain valid analytical data.

INTRODUCTION

Inductively Coupled Plasma/Atomic Emission Spectrometry (ICP/AES) is used extensively in the Analytical Chemistry Laboratory (ACL) to characterize diverse analytical samples. This method simultaneously measures the concentrations of multiple cations in solutions. The working standards used for instrument calibration are generally in the 5-20 $\mu\text{g/mL}$ range. They are prepared from certified single or multielement stock solutions by serial dilution and are preserved with inorganic acids. The accuracy and precision of analytical measurements are greatly affected by uncertainties in standard stabilities as functions of preservative type, analyte content, and acid concentration. As a consequence, a systematic study was initiated to assess the effect of these parameters on data quality, standard procurement needs, and waste minimization.

EXPERIMENTAL

Apparatus

The ICP/AES measurements were performed on a spectrometer system that incorporated a 48-channel polychromator and a computer-controlled scanning monochromator (Instruments S.A., Inc., Edison, N.J.). Both instruments were focused on a single plasma excitation source. Table 1 lists the instrumentation and Table 2 details the operating conditions for this study. Typical experimental detection limits (DLs) and their corresponding wavelengths are given in Table 3.

Reagents

Acids. Intra-Analyzed HCl and HNO₃ (Baker Chemical Co.) were used in blank and standard preparations. Water for dilutions (≥ 18 Mohm-cm) was obtained from a Sybron/Barnstead ion exchange system.

Standards. Certified standard stock solutions were purchased from the National Institute of Standards and Technology (NIST), SPEX Industries, and Baker Chemical Co. Test solutions and calibration standards were prepared by making serial dilutions of the concentrates and adjusting to the appropriate acidity. This information is summarized in Table 4.

Procedures

The plasma system was optimized with respect to maximum signal-to-background (S/B) ratios as a function of the vertical viewing aperture. The instrument was calibrated at this position using the two-point calibration procedure with standards and blanks prepared in a 2% acid (HCl or HNO₃) medium. Working standards were prepared on the day of the analytical run and verified against previously used calibration solutions. An agreement of $\pm 3\%$ for this comparative analysis was considered to provide a valid calibration curve for the subsequent analytical run.

Since instrument calibration standards were prepared at a single acid concentration (2%), information on the effect of acidity on signal intensity was needed for data assessment of the diverse analytical matrices. As a consequence, S/B ratios were determined as a function of HCl concentration, using a 1- μ g/mL multielement standard and a corresponding blank.

The stability of standards is also affected by the possible loss of analyte by a diffusion and/or evaporation mechanism through plastic containers. Since stock and working standards are stored in polyethylene bottles, the diffusion effect was studied by weight-loss measurements from high-density polyethylene (HDPE) and Teflon containers. Bottles were filled with acid solutions, and the change in weight as a function of time was established gravimetrically.

RESULTS AND DISCUSSION

Diffusion

Table 5 summarizes solution loss data by a diffusion mechanism as a function of time at normally used acid concentrations for ICP/AES analysis. These losses are essentially linear with time and are functions of container type and acid preservative. Solutions with 2% HNO₃ maintained their weight best in HDPE, while those in Teflon showed the lowest weight loss from HCl-HNO₃ mixtures. Although vaporization losses from repeated sampling of bottles and diffusion effects from partially filled containers were not studied, the data suggest that within the accuracy range of ICP/AES (1-10%), changes via these mechanisms would be in an acceptable range for valid analytical data.

S/B Ratios

One of the objectives of this investigation was to determine the stability of standards as a function of acidity. The analyte response as a function of this parameter is of interest, since instrument calibration is generally performed at a 2% acid concentration. Table 6 summarizes

S/B ratios for 18 elements in HCl solutions. The data show that the presence of mineral acids decreases the emission intensities of atomic and ionic lines for most elements, which is consistent with published results.¹ It has been postulated that these phenomena are related to the physical state of the plasma at lower acidities (<1M) and to the reduced rate of sample uptake at higher acid concentrations (>1M). A few, very sensitive elements (i.e., Be, Ca, Mg, and Sr) exhibit an increase in S/B ratios, an observation inconsistent with the above mechanism. Representative data for other elements are plotted in Figure 1 and tend to suggest a linear decrease in the S/B ratio as a function of acidity. These results show that for best accuracy, calibration standards and blanks should be matched with respect to acid concentration. Furthermore, for the purpose of this study, only relative analyte changes should be considered in stability assessments, since instrument calibration was performed at a single acid concentration.

Stability Data

Tables 7 through 30 summarize stability results for 18 elements as a function of acid type, acidity, and analyte concentration. Short- (two months), intermediate- (four months), and overall-term (six months) concentration data and their standard deviations (SD) are provided. In addition, relative standard deviations (RSDs) were calculated and are listed for the study period. In general, all analytes are stable for six months irrespective of preservative (HCl, HNO₃) or cation concentration. As expected, deviations from the mean decrease with increasing element content, consistent with favorable statistics at higher signal intensities. At low analyte concentration (0.2 µg/mL), elements with relatively high DLs have RSDs exceeding 10%, a result that can be reconciled with greater signal fluctuations. The unexpected stability of Sn in HNO₃ solutions can be explained by the presence of HCl in the stock standard used to prepare the test mixtures.

Tables 31 and 32 present data on an Instrument Calibration Verification Standard (ICV) provided by the U.S. EPA for the Contract Laboratory Program (CLP). This solution, prepared as directed, is used to assess the performance of the ICP/AES system for the analysis of environmental samples. The results clearly show that this material is stable over at least a six month period, and weekly re-preparations are redundant for valid analytical data and are in fact detrimental to the current trend of waste minimization activities. The Ag data are shown as QIC listing in Table 33 for the same solution, with acceptable results.

Table 33 provides a summary on the stability of Ag at various HNO_3 concentrations. All solutions were stable over the study period within the arbitrary $\pm 5\%$ RSD guidelines. (Deterioration of Ag standards can usually be related to the introduction of chloride ions during wash-out procedures.)

Table 34 summarizes the results for Sn. The data support the conclusion that an HCl- HNO_3 mixture provided the best shelf life. Standards prepared in HNO_3 deteriorated severely within one week, presumably through the precipitation of insoluble stannic acid.

Table 35 collates stability data for B, Mo, Si, and Ti as a function of concentration and acidity. The short-term results are acceptable for all analytes. However, it was noted that the measured Si results were increasing with time, an observation inconsistent with stability expectations. Further examination revealed that the 2 mg/mL Si stock standard (2% HCl), used for the preparation of the calibrating solution, contained a precipitate. Once this condition was corrected, the measured Si values for the test matrices returned to an acceptable range. In retrospect, the stability of Si standards should have been examined in the presence of HF, a medium known to promote complexation reactions and thus avoid precipitation due to hydrolysis.

The As, Sb, and Se stability data are tabulated in Table 36. The As and Se results are acceptable under the listed experimental conditions. However, like the Si data, the Sb results,

while having good precision in the short-term domain, exhibit an increase in concentration with time. The instability of the Sb stock solution (2% HCl) was the source of this discrepancy; the test solution values returned to the acceptance range once a fresh starting standard was prepared.

Tables 37 and 38 present Nb and W data, respectively, as a function of element concentration and preservative. The results verify that HF is required to stabilize these elements in calibration standards to ensure longer shelf-lives.

Figures 2 through 7 graphically summarize experimental RSDs of standards as a function of analyte concentration for selected elements over the six-month study period. The data show that for elements with relatively high DLs (Al,Pb,Sn), the variability at low concentrations (0.2 $\mu\text{g/mL}$) is relatively high (10-20%), a result expected at the low signal intensities. The reproducibility improves (2-3%) as the analyte concentration increases. The observed anomalous result for Sn in 1% HNO_3 is due to precipitation in this medium. Elements with moderate and low DLs (Be,Cr,Ni) exhibit a reasonably constant RSD (2%) over the concentration range studied, an observation consistent with expected fluctuations of relatively high signal intensities.

CONCLUSION

The results on the stability of calibration standards for ICP/AES analysis show that a suitable acid preservative enables analyte solutions (5-20 $\mu\text{g/mL}$) to maintain their integrity over at least a six-month period. For the wide range of elements studied, a mixture of 5% HCl + 1% HNO_3 appears to provide the best compromise for acceptable shelf lives for most cations. However, for those situations where the relatively high acidities are impractical or undesirable, 2% HCl or 2% HNO_3 stabilizes analytes satisfactorily, provided any chemical incompatibilities (i.e., precipitation, hydrolysis) are resolved.

The variations in S/B ratios as a function of acidity mandate that for best accuracy and precision, calibration standards and blanks should be matrix-matched (i.e., same acid type and concentration, same major constituents). However, for analytical data normally reported in the accuracy range of 3-10%, approximate acid concentrations are sufficient for satisfactory results.

Changes in analyte concentrations due to evaporation and diffusion through plastic containers do not appear to make a significant contribution to inaccuracies, compared to the more pronounced effects introduced by the excitation and sample transport system. However, partially filled bottles, increased frequency of sampling, and environmental storage conditions could have a considerably larger impact on standard stabilities than those observed in this study.

This study clearly shows that the frequency of calibration standard preparations could be lowered significantly for ICP/AES analyses at moderate accuracies (3-10%). As a consequence, benefits in lower costs for standard preparation and waste disposal would be realized, in addition to better allocation of analytical effort.

ACKNOWLEDGMENTS

The financial support of the Support Development Fund Committee at ANL has made this investigation possible. The development of computer software programs for data compilations and reductions by J. E. Kulaga is gratefully acknowledged.

REFERENCES

1. J. E. Yoshimura, H. Suzuki, S. Yamazaki, and S. Toda, "Interference by Mineral Acids in Inductively Coupled Plasma Atomic Emission Spectrometry," *Analyst* 115, 167 (1990).

Table 1. Instrumentation**A. Spectrometers**

Polychromator	Instruments S.A., Inc., Model J-Y48P 1-m Paschen-Runge with 2550 grooves/mm holographic concave grating, 20- μ m entrance and 50- μ m exit slits. Hamamatsu R300 and R306 photomultiplier tubes, 48 channels.
Monochromator	Instruments S. A., Inc., Model J-Y38 1-m Czerny-Turner with Spectra-Link controller. 2400 grooves/mm holographic plane grating, 50- μ m entrance and exit slits; Hamamatsu R955 photomultiplier tube.

B. Power and Nebulizer System

Generator	Plasma-Therm Model HFP-2500, 27.12 MHz with 3-turn copper load coil.
Nebulizer	Instruments S. A., Inc., concentric with a Pt-Ir capillary tube.
Torch	Instruments S. A., Inc., demountable in a Mermet-Trassey configuration.
Spray Chamber	Instruments S. A., Inc., double-barrel, constructed from Ryton.

C. Computer System

Computer	Digital Equipment Corp. PDP-11/73 with 768K byte of memory, three RL-02 disks.
Terminals	Digital Equipment Corp. DEC Writer III hard copy and DEC VT-340 video with graphics.
Software	<ol style="list-style-type: none"> 1. Instruments S. A., Inc. -supplied analytical program run by the DEC RSX-11M operating system. 2. ANL-developed report generation and quality assurance programs.

Table 2. Operating Conditions

Forward R.F. Power	1.10 kW
Reflected Power	<5 W
Argon Flow Rates	
Outer Gas	14 L/min
Intermediate Gas	0.4 L/min
Sheath Gas	0.5 L/min
Nebulizer Gas	0.6 L/min
Sample Uptake Rate	2.8 mL/min
Observation Height	16 mm above load coil
Integration Times	
Polychromator	10 sec
Monochromator	1 sec

Table 3. Detection Limits

<u>Element</u>	<u>Wavelength (nm)</u>	<u>DL (µg/L)</u> ^a	<u>Element</u>	<u>Wavelength (nm)</u>	<u>DL (µg/L)</u> ^a
Polychromator					
Ag	328.068	1.9	Mo	202.030	14.
Al	308.215	38.	Ni	231.604	9.3
As	193.695	52.	Pb	220.353	90.
B	208.959	11.	Sb	206.833	54.
Ba	233.527	1.3	Se	196.026	58.
Be	313.042	0.4	Si	251.611	14.
Ca	393.366	1.3	Sn	189.926	30.
Cd	226.502	3.4	Sr	407.771	0.5
Co	228.616	36.	Ti	334.941	2.2
Cr	267.716	4.6	Tl	190.801	24.
Cu	324.754	2.1	V	292.402	2.9
Fe	238.204	2.1	Zn	213.856	0.5
Mg	279.553	1.1	Zr	343.823	2.7
Mn	257.610	1.5			
Monochromator					
Nb	309.418	8.9			
W	224.875	7.9			

^a Detection Limit (DL) = 3 x the standard deviation of the baseline noise.

Table 4. Summary of Standard Solutions

<u>File Name</u>	<u>Conc. (µg/mL)</u>	<u>Elements</u>	<u>Acidity</u>
QF01	0.2	Spex A ^a	1% HNO3 + 0.002% HCl
QF02	0.2	Spex A	2% HNO3 + 0.002% HCl
QF03	0.2	Spex A	5% HNO3 + 0.002% HCl
QF04	0.2	Spex A	1% HCl
QF05	0.2	Spex A	2% HCl
QF06	0.2	Spex A	5% HCl
QF07	0.2	Spex A	5% HCl + 1% HNO3
QF08	1.0	Spex A	1% HNO3 + 0.01% HCl
QF09	1.0	Spex A	2% HNO3 + 0.01% HCl
QF10	1.0	Spex A	5% HNO3 + 0.01% HCl
QF11	1.0	Spex A	1% HCl
QF12	1.0	Spex A	2% HCl
QF13	1.0	Spex A	5% HCl
QF14	1.0	Spex A	5% HCl + 1% HNO3
QF15	4.0	Spex A	1% HNO3 + 0.04% HCl
QF16	4.0	Spex A	2% HNO3 + 0.04% HCl
QF17	4.0	Spex A	5% HNO3 + 0.04% HCl
QF18	4.0	Spex A	1% HCl
QF19	4.0	Spex A	2% HCl
QF20	4.0	Spex A	5% HCl
QF21	4.0	Spex A	5% HCl + 1% HNO3
QF22	10.0	Spex A	2% HNO3 + 0.1% HCl
QF23	10.0	Spex A	2% HCl
QF24	10.0	Spex A	5% HCl + 1% HNO3
QI01	1:10 dil.	ICV-1 ^b	2% HNO3
QI02	1:10 dil.	ICV-1	5% HCl + 1% HNO3
QIC1	1:10 dil.	ICV-1	2% HNO3
QIC2	1:10 dil.	ICV-1	5% HCl + 1% HNO3
QA01	0.2	Ag	1% HNO3
QA02	0.2	Ag	2% HNO3
QA03	0.2	Ag	5% HNO3
QA04	1.0	Ag	1% HNO3
QA05	1.0	Ag	2% HNO3
QA06	1.0	Ag	5% HNO3
QA07	4.0	Ag	1% HNO3
QA08	4.0	Ag	2% HNO3
QA09	4.0	Ag	5% HNO3
QA10	10.0	Ag	2% HNO3

^a Spex A: Al, Ba, Be, Ca, Cd, Co, Cr, Cu, Fe, Mg, Mn, Ni, Pb, Sn, Sr, V, Zn, Zr

^b ICV-1: Ag, Al, Ba, Be, Ca, Cd, Co, Cr, Cu, Fe, Mg, Mn, Ni, Pb, V, Zn

Table 4. Summary of Standard Solutions (contd)

<u>File Name</u>	<u>Conc. (µg/mL)</u>	<u>Elements</u>	<u>Acidity</u>
QSi1	0.5; 1.0	B, Mo, Ti; Si	2% HNO ₃
QSi2	0.5; 1.0	B, Mo, Ti; Si	2% HCl
QSi3	0.5; 1.0	B, Mo, Ti; Si	5% HCl + 1% HNO ₃
QSi4	2.5; 5.0	B, Mo, Ti; Si	2% HNO ₃
QSi5	2.5; 5.0	B, Mo, Ti; Si	2% HCl
QSi6	2.5; 5.0	B, Mo, Ti; Si	5% HCl + 1% HNO ₃
QSi7	10.0; 20.0	B, Mo, Ti; Si	2% HNO ₃
QSi8	10.0; 20.0	B, Mo, Ti; Si	2% HCl
QSi9	10.0; 20.0	B, Mo, Ti; Si	5% HCl + 1% HNO ₃
QAs1	5.0	As, Sb, Se	2% HNO ₃
QAs2	5.0	As, Sb, Se	2% HCl
QAs3	5.0	As, Sb, Se	5% HCl + 1% HNO ₃
QAs4	20.0	As, Sb, Se	2% HNO ₃
QAs5	20.0	As, Sb, Se	2% HCl
QAs6	20.0	As, Sb, Se	5% HCl + 1% HNO ₃
QSn1	5.0	Sn	2% HNO ₃
QSn2	5.0	Sn	2% HCl
QSn3	5.0	Sn	5% HCl + 1% HNO ₃
QSn4	20.0	Sn	2% HNO ₃
QSn5	20.0	Sn	2% HCl
QSn6	20.0	Sn	5% HCl + 1% HNO ₃
QW01	1.0	W	2% HCl
QW02	1.0	W	2% HCl + 0.5% HF
QW03	5.0	W	2% HCl
QW04	5.0	W	2% HCl + 0.5% HF
QW05	20.0	W	2% HCl
QW06	20.0	W	2% HCl + 0.5% HF
QNb1	1.0	Nb	2% HCl
QNb2	1.0	Nb	2% HCl + 0.5% HF
QNb3	5.0	Nb	2% HCl
QNb4	5.0	Nb	2% HCl + 0.5% HF
QNb5	20.0	Nb	2% HCl
QNb6	20.0	Nb	2% HCl + 0.5% HF

Table 5. Weight Loss by Diffusion

Material	Acid	% Weight Loss		
		3 months	6 months	9 months
HDPE ^a	2% HNO ₃	0.05	0.08	0.12
HDPE	2% HCl	0.09	0.16	0.23
HDPE	5% HCl + 1% HNO ₃	0.08	0.13	0.19
FEP ^b	2% HNO ₃	0.18	0.29	0.42
FEP	2% HCl	0.07	0.12	0.17
FEP	5% HCl + 1% HNO ₃	0.02	0.05	0.07

^a HDPE: High-Density Polyethylene

^b FEP: Fluorinated Ethylene Propylene (Teflon)

Table 6. Signal-to-Background Ratio vs. Acidity

Element	Signal-to-Background Ratio					% Change	
	1%HCl	2% HCl	5% HCl	10% HCl	20%HCl	1-5% HCl	1-20% HCl
Al	1.63	1.64	1.70	1.60	1.60	+4.3	-1.8
Ba	6.50	6.44	6.35	6.20	6.04	-2.3	-7.1
Be	83.4	84.5	88.4	90.7	91.2	+6.0	+9.3
Ca	14.58	15.47	19.13	18.86	18.08	+31.2	+24.0
Cd	7.75	7.65	7.60	7.43	7.26	-1.9	-6.3
Co	2.99	2.98	2.96	2.94	2.84	-1.0	-3.4
Cr	4.36	4.30	4.28	4.25	4.14	-1.8	-5.0
Cu	5.38	5.29	5.31	5.23	5.22	-1.3	-3.0
Fe	5.02	4.86	4.80	4.69	4.62	-4.4	-8.0
Mg	30.1	29.9	38.7	41.2	38.4	+28.6	+27.6
Mn	17.75	17.57	17.45	17.05	16.64	-1.7	-6.3
Ni	2.69	2.71	2.70	2.65	2.59	-0.4	-3.7
Pb	1.57	1.54	1.56	1.53	1.51	-0.6	-3.8
Sn	1.68	1.65	1.61	1.67	1.60	-4.2	-4.8
Sr	63.8	64.8	68.9	70.3	71.2	+8.0	+11.6
V	4.23	4.21	4.19	4.12	4.03	-0.9	-4.7
Zn	18.97	18.90	19.17	18.33	17.43	+1.0	-8.1
Zr	4.21	4.21	4.26	4.19	4.14	+1.1	-1.7

Table 7. Stability Data for Solution QF01

Element	<u>03/08/91-05/08/91</u> n=6		<u>05/16/91-09/13/91</u> n=12		<u>03/08/91-09/13/91</u> n=18		
	<u>Ave1</u>	<u>SD1</u>	<u>Ave2</u>	<u>SD2</u>	<u>Ave</u>	<u>SD</u>	<u>RSD</u>
Al	0.197	0.014	0.208	0.023	0.204	0.020	10.0
Ba	0.197	0.003	0.200	0.004	0.199	0.004	1.9
Be	0.194	0.002	0.196	0.002	0.195	0.002	1.0
Ca	0.179	0.007	0.172	0.004	0.174	0.005	3.0
Cd	0.198	0.003	0.202	0.004	0.201	0.004	1.8
Co	0.198	0.011	0.202	0.010	0.201	0.010	5.2
Cr	0.202	0.006	0.204	0.005	0.203	0.005	2.6
Cu	0.201	0.007	0.203	0.004	0.202	0.005	2.6
Fe	0.198	0.005	0.200	0.005	0.199	0.005	2.5
Mg	0.194	0.001	0.195	0.003	0.195	0.003	1.3
Mn	0.201	0.006	0.202	0.002	0.202	0.004	1.9
Ni	0.197	0.009	0.203	0.007	0.201	0.008	3.8
Pb	0.184	0.024	0.211	0.032	0.202	0.030	14.6
Sn	0.231	0.040	0.229	0.025	0.230	0.031	13.4
Sr	0.198	0.003	0.200	0.003	0.199	0.003	1.5
V	0.204	0.008	0.206	0.004	0.205	0.006	2.8
Zn	0.197	0.004	0.199	0.003	0.198	0.003	1.7
Zr	0.182	0.006	0.177	0.005	0.179	0.005	3.0

Standard concentration: 0.2 µg/mL
Acidity: 1% HNO₃ + 0.002% HCl

Table 8. Stability Data for Solution QF02

<u>Element</u>	<u>03/08/91-05/08/91</u> n=6		<u>05/16/91-09/13/91</u> n=12		<u>03/08/91-09/13/91</u> n=18		
	<u>Ave1</u>	<u>SD1</u>	<u>Ave2</u>	<u>SD2</u>	<u>Ave</u>	<u>SD</u>	<u>RSD</u>
Al	0.197	0.014	0.202	0.022	0.200	0.020	9.8
Ba	0.199	0.004	0.201	0.003	0.200	0.003	1.7
Be	0.196	0.002	0.198	0.002	0.197	0.002	0.9
Ca	0.190	0.007	0.183	0.005	0.185	0.006	3.1
Cd	0.200	0.003	0.203	0.003	0.202	0.003	1.5
Co	0.202	0.008	0.204	0.004	0.203	0.006	2.8
Cr	0.203	0.005	0.207	0.005	0.206	0.005	2.4
Cu	0.201	0.007	0.203	0.004	0.202	0.005	2.6
Fe	0.199	0.004	0.202	0.005	0.201	0.005	2.3
Mg	0.197	0.002	0.198	0.002	0.198	0.002	1.0
Mn	0.203	0.007	0.204	0.002	0.204	0.004	2.1
Ni	0.197	0.010	0.203	0.007	0.201	0.008	4.0
Pb	0.192	0.025	0.198	0.038	0.196	0.034	17.5
Sn	0.244	0.037	0.237	0.035	0.239	0.036	14.9
Sr	0.200	0.004	0.202	0.003	0.201	0.003	1.7
V	0.206	0.009	0.207	0.004	0.207	0.006	3.0
Zn	0.197	0.005	0.200	0.003	0.199	0.004	1.9
Zr	0.196	0.007	0.198	0.005	0.197	0.006	2.9

Standard concentration: 0.2 µg/mL
Acidity: 2% HNO₃ + 0.002% HCl

Table 9. Stability Data for Solution QF03

Element	03/08/91-05/08/91 n=6		05/16/91-09/13/91 n=12		03/08/91-09/13/91 n=18		
	Ave1	SD1	Ave2	SD2	Ave	SD	RSD
Al	0.181	0.013	0.170	0.034	0.174	0.029	16.6
Ba	0.195	0.003	0.195	0.006	0.195	0.005	2.7
Be	0.192	0.003	0.194	0.004	0.193	0.004	1.9
Ca	0.209	0.006	0.202	0.006	0.204	0.006	2.9
Cd	0.195	0.005	0.196	0.006	0.196	0.006	2.9
Co	0.196	0.009	0.195	0.011	0.195	0.010	5.3
Cr	0.200	0.005	0.200	0.008	0.200	0.007	3.6
Cu	0.195	0.007	0.195	0.007	0.195	0.007	3.6
Fe	0.195	0.003	0.195	0.007	0.195	0.006	3.1
Mg	0.194	0.001	0.195	0.004	0.195	0.003	1.7
Mn	0.200	0.006	0.200	0.004	0.200	0.005	2.4
Ni	0.190	0.003	0.189	0.011	0.139	0.009	4.8
Pb	0.178	0.013	0.170	0.034	0.173	0.029	16.7
Sn	0.234	0.043	0.215	0.021	0.221	0.030	13.6
Sr	0.198	0.004	0.200	0.004	0.199	0.004	2.0
V	0.202	0.009	0.200	0.006	0.201	0.007	3.6
Zn	0.190	0.004	0.192	0.004	0.191	0.004	2.1
Zr	0.206	0.008	0.203	0.007	0.204	0.007	3.6

Standard concentration: 0.2 µg/mL
Acidity: 5% HNO₃ + 0.002% HCl

Table 10. Stability Data for Solution QF04

Element	<u>03/08/91-05/08/91</u> n=6		<u>05/16/91-09/13/91</u> n=12		<u>03/08/91-09/13/91</u> n=18		
	<u>Ave1</u>	<u>SD1</u>	<u>Ave2</u>	<u>SD2</u>	<u>Ave</u>	<u>SD</u>	<u>RSD</u>
Al	0.202	0.009	0.215	0.023	0.211	0.019	9.2
Ba	0.199	0.002	0.201	0.004	0.200	0.003	1.7
Be	0.198	0.003	0.199	0.002	0.199	0.002	1.2
Ca	0.177	0.006	0.170	0.004	0.172	0.005	2.8
Cd	0.202	0.004	0.204	0.004	0.203	0.004	2.0
Co	0.204	0.009	0.206	0.006	0.205	0.007	3.5
Cr	0.204	0.005	0.207	0.005	0.206	0.005	2.4
Cu	0.204	0.007	0.207	0.004	0.206	0.005	2.5
Fe	0.200	0.004	0.202	0.006	0.201	0.005	2.7
Mg	0.197	0.002	0.199	0.003	0.198	0.003	1.4
Mn	0.203	0.007	0.204	0.003	0.204	0.005	2.3
Ni	0.199	0.005	0.205	0.009	0.203	0.008	3.9
Pb	0.198	0.018	0.202	0.041	0.201	0.035	17.5
Sn	0.234	0.047	0.237	0.030	0.236	0.037	15.5
Sr	0.200	0.004	0.202	0.002	0.201	0.003	1.4
V	0.207	0.009	0.209	0.005	0.208	0.007	3.2
Zn	0.202	0.006	0.204	0.003	0.203	0.004	2.1
Zr	0.188	0.006	0.187	0.005	0.187	0.005	2.9

Standard concentration: 0.2 µg/mL
Acidity: 1% HCl

Table 11. Stability Data for Solution QF05

Element	03/08/91-05/08/91 n=6		05/16/91-09/13/91 n=12		03/08/91-09/13/91 n=18		
	Ave1	SD1	Ave2	SD2	Ave3	SD	RSD
Al	0.199	0.011	0.215	0.017	0.210	0.015	7.3
Ba	0.197	0.003	0.200	0.003	0.199	0.003	1.5
Be	0.197	0.002	0.200	0.002	0.199	0.002	1.0
Ca	0.175	0.007	0.169	0.004	0.171	0.005	3.0
Cd	0.199	0.005	0.205	0.003	0.203	0.004	1.9
Co	0.197	0.007	0.208	0.005	0.204	0.006	2.8
Cr	0.201	0.005	0.206	0.004	0.204	0.004	2.1
Cu	0.201	0.006	0.204	0.003	0.203	0.004	2.1
Fe	0.198	0.004	0.202	0.004	0.201	0.004	2.0
Mg	0.194	0.002	0.197	0.002	0.196	0.002	1.0
Mn	0.201	0.006	0.203	0.002	0.202	0.004	1.9
Ni	0.199	0.007	0.204	0.005	0.202	0.006	2.8
Pb	0.195	0.013	0.208	0.023	0.204	0.020	9.9
Sn	0.235	0.038	0.236	0.032	0.236	0.034	14.5
Sr	0.198	0.004	0.201	0.002	0.200	0.003	1.4
V	0.204	0.009	0.208	0.003	0.207	0.006	2.8
Zn	0.199	0.005	0.203	0.002	0.202	0.003	1.6
Zr	0.195	0.007	0.198	0.003	0.197	0.005	2.4

Standard concentration: 0.2 µg/mL
Acidity: 2% HCl

Table 12. Stability Data for Solution QF06

<u>Element</u>	<u>03/08/91-05/08/91</u> n=6		<u>05/16/91-09/13/91</u> n=12		<u>03/08/91-09/13/91</u> n=18		
	<u>Ave1</u>	<u>SD1</u>	<u>Ave2</u>	<u>SD2</u>	<u>Ave</u>	<u>SD</u>	<u>RSD</u>
Al	0.176	0.010	0.183	0.016	0.181	0.014	7.9
Ba	0.179	0.003	0.183	0.003	0.182	0.003	1.7
Be	0.182	0.003	0.185	0.002	0.184	0.002	1.3
Ca	0.158	0.007	0.152	0.004	0.154	0.005	3.4
Cd	0.183	0.004	0.186	0.003	0.185	0.003	1.8
Co	0.182	0.007	0.185	0.006	0.184	0.006	3.5
Cr	0.181	0.004	0.187	0.003	0.185	0.003	1.8
Cu	0.182	0.006	0.185	0.003	0.184	0.004	2.3
Fe	0.180	0.004	0.184	0.004	0.183	0.004	2.2
Mg	0.176	0.002	0.179	0.002	0.178	0.002	1.1
Mn	0.183	0.006	0.185	0.002	0.184	0.004	2.1
Ni	0.179	0.005	0.183	0.006	0.182	0.006	3.1
Pb	0.181	0.024	0.186	0.024	0.184	0.024	13.0
Sn	0.227	0.033	0.211	0.031	0.216	0.032	14.6
Sr	0.180	0.004	0.184	0.002	0.183	0.003	1.5
V	0.186	0.008	0.188	0.003	0.187	0.005	2.8
Zn	0.180	0.005	0.183	0.002	0.182	0.003	1.8
Zr	0.188	0.007	0.192	0.003	0.191	0.005	2.5

Standard concentration: 0.2 µg/mL
Acidity: 5% HCl

Table 13. Stability Data for Solution QF07

Element	<u>03/08/91-05/08/91</u> n=6		<u>05/16/91-09/13/91</u> n=12		<u>03/08/91-09/13/91</u> n=18		
	<u>Ave1</u>	<u>SD1</u>	<u>Ave2</u>	<u>SD2</u>	<u>Ave</u>	<u>SD</u>	<u>RSD</u>
Al	0.187	0.008	0.185	0.019	0.186	0.016	8.7
Ba	0.191	0.002	0.196	0.002	0.194	0.002	1.0
Be	0.194	0.003	0.197	0.002	0.196	0.002	1.2
Ca	0.179	0.007	0.174	0.005	0.176	0.006	3.3
Cd	0.193	0.004	0.198	0.003	0.196	0.003	1.7
Co	0.193	0.006	0.196	0.009	0.195	0.008	4.2
Cr	0.195	0.004	0.199	0.003	0.198	0.003	1.7
Cu	0.195	0.005	0.196	0.003	0.196	0.004	1.9
Fe	0.193	0.004	0.195	0.004	0.194	0.004	2.1
Mg	0.190	0.001	0.193	0.002	0.192	0.002	0.9
Mn	0.195	0.006	0.198	0.002	0.197	0.004	1.9
Ni	0.188	0.007	0.195	0.007	0.193	0.007	3.6
Pb	0.183	0.018	0.188	0.029	0.186	0.026	13.9
Sn	0.239	0.040	0.225	0.038	0.230	0.039	16.8
Sr	0.193	0.004	0.197	0.002	0.196	0.003	1.4
V	0.198	0.008	0.200	0.003	0.199	0.005	2.6
Zn	0.190	0.004	0.194	0.002	0.193	0.003	1.5
Zr	0.197	0.006	0.199	0.003	0.198	0.004	2.1

Standard concentration: 0.2 µg/mL
Acidity: 1% HNO₃ + 5% HCl

Table 14. Stability Data for Solution QF08

Element	<u>03/08/91-05/08/91</u> n=6		<u>05/16/91-09/13/91</u> n=12		<u>03/08/91-09/13/91</u> n=18		
	Ave1	SD1	Ave2	SD2	Ave	SD	RSD
Al	0.996	0.019	1.011	0.028	1.006	0.025	2.5
Ba	0.993	0.010	0.998	0.015	0.996	0.014	1.4
Be	0.970	0.006	0.972	0.012	0.971	0.010	1.1
Ca	0.989	0.012	0.991	0.013	0.990	0.013	1.3
Cd	0.994	0.017	1.002	0.012	0.999	0.014	1.4
Co	1.003	0.021	1.016	0.020	1.012	0.020	2.0
Cr	1.005	0.014	1.014	0.015	1.011	0.015	1.5
Cu	1.000	0.032	1.008	0.013	1.005	0.021	2.1
Fe	0.997	0.011	1.006	0.016	1.003	0.015	1.4
Mg	0.997	0.023	1.005	0.012	1.002	0.017	1.6
Mn	0.993	0.025	1.004	0.012	1.000	0.017	1.7
Ni	0.993	0.011	1.001	0.018	0.998	0.016	1.6
Pb	0.995	0.038	0.992	0.045	0.993	0.043	4.3
Sn	1.180	0.160	1.144	0.049	1.156	0.101	8.7
Sr	0.984	0.013	0.983	0.013	0.983	0.013	1.3
V	1.012	0.022	1.013	0.011	1.013	0.016	1.5
Zn	0.984	0.019	0.985	0.011	0.985	0.014	1.4
Zr	0.973	0.017	0.961	0.011	0.965	0.013	1.4

Standard concentration: 1.0 µg/mL
Acidity: 1% HNO₃ + 0.01% HCl

Table 15. Stability Data for Solution QF09

Element	<u>03/08/91-05/08/91</u> n=6		<u>05/16/91-09/13/91</u> n=12		<u>03/08/91-09/13/91</u> n=18		
	Ave1	SD1	Ave2	SD2	Ave	SD	RSD
Al	0.978	0.018	0.982	0.042	0.981	0.036	3.7
Ba	0.980	0.012	0.984	0.021	0.983	0.018	1.9
Be	0.958	0.009	0.960	0.017	0.959	0.015	1.5
Ca	0.986	0.015	0.988	0.018	0.987	0.017	1.7
Cd	0.981	0.019	0.988	0.020	0.986	0.020	2.0
Co	0.994	0.022	1.000	0.024	0.998	0.023	2.3
Cr	0.993	0.013	1.000	0.023	0.998	0.020	2.0
Cu	0.984	0.031	0.993	0.019	0.990	0.024	2.4
Fe	0.988	0.017	0.994	0.024	0.992	0.022	2.2
Mg	0.988	0.022	0.994	0.017	0.992	0.019	1.9
Mn	0.981	0.025	0.933	0.018	0.989	0.021	2.1
Ni	0.982	0.014	0.985	0.023	0.984	0.020	2.1
Pb	0.974	0.042	0.972	0.052	0.973	0.049	5.0
Sn	1.160	0.110	1.118	0.056	1.132	0.078	6.9
Sr	0.973	0.014	0.974	0.018	0.974	0.017	1.7
V	0.998	0.022	1.004	0.020	1.002	0.021	2.1
Zn	0.962	0.018	0.966	0.013	0.965	0.015	1.5
Zr	0.994	0.011	0.994	0.020	0.994	0.018	1.8

Standard concentration: 1.0 µg/mL
Acidity: 2% HNO₃ + 0.01% HCl

Table 16. Stability Data for Solution QF10

Element	03/08/91-05/08/91 n=6		05/16/91-09/13/91 n=12		03/08/91-09/13/91 n=18		
	Ave1	SD1	Ave2	SD2	Ave	SD	RSD
Al	0.942	0.017	0.942	0.028	0.942	0.025	2.6
Ba	0.964	0.011	0.968	0.014	0.967	0.013	1.4
Be	0.939	0.009	0.944	0.010	0.942	0.010	1.0
Ca	0.997	0.018	1.002	0.010	1.000	0.013	1.3
Cd	0.960	0.018	0.967	0.012	0.965	0.014	1.5
Co	0.974	0.021	0.984	0.015	0.981	0.017	1.8
Cr	0.976	0.013	0.985	0.014	0.982	0.014	1.4
Cu	0.960	0.030	0.973	0.009	0.969	0.019	1.9
Fe	0.969	0.014	0.975	0.015	0.973	0.015	1.5
Mg	0.971	0.022	0.980	0.008	0.977	0.014	1.5
Mn	0.966	0.027	0.978	0.009	0.974	0.017	1.8
Ni	0.962	0.011	0.963	0.018	0.963	0.016	1.7
Pb	0.941	0.039	0.940	0.046	0.940	0.044	4.7
Sn	1.130	0.140	1.124	0.048	1.126	0.090	8.0
Sr	0.962	0.015	0.967	0.011	0.965	0.012	1.3
V	0.982	0.021	0.987	0.010	0.985	0.015	1.5
Zn	0.927	0.018	0.932	0.007	0.930	0.012	1.3
Zr	0.994	0.018	0.996	0.009	0.995	0.013	1.3

Standard concentration: 1.0 µg/mL
Acidity: 5% HNO₃ + 0.01% HCl

Table 17. Stability Data for Solution QF11

Element	<u>03/08/91-05/08/91</u> n=6		<u>05/16/91-09/13/91</u> n=12		<u>03/08/91-09/13/91</u> n=18		
	<u>Ave1</u>	<u>SD1</u>	<u>Ave2</u>	<u>SD2</u>	<u>Ave</u>	<u>SD</u>	<u>RSD</u>
Al	1.002	0.022	1.003	0.027	1.003	0.025	2.5
Ba	0.993	0.011	0.991	0.013	0.992	0.012	1.2
Be	0.979	0.006	0.973	0.010	0.975	0.009	0.9
Ca	0.989	0.017	0.981	0.011	0.984	0.013	1.4
Cd	0.997	0.018	1.001	0.014	1.000	0.015	1.5
Co	1.004	0.021	1.010	0.019	1.008	0.020	2.0
Cr	1.004	0.015	1.006	0.014	1.005	0.014	1.4
Cu	1.011	0.038	1.012	0.010	1.012	0.023	2.3
Fe	1.001	0.015	1.001	0.018	1.001	0.017	1.7
Mg	1.002	0.025	1.003	0.013	1.003	0.018	1.8
Mn	0.995	0.029	0.999	0.012	0.998	0.019	1.9
Ni	0.993	0.020	0.992	0.020	0.992	0.020	2.0
Pb	0.996	0.052	0.993	0.039	0.994	0.044	4.4
Sn	1.190	0.120	1.144	0.046	1.159	0.079	6.8
Sr	0.987	0.017	0.981	0.012	0.983	0.014	1.4
V	1.011	0.024	1.009	0.013	1.010	0.017	1.7
Zn	0.992	0.019	0.991	0.009	0.991	0.013	1.3
Zr	0.981	0.014	0.969	0.011	0.973	0.012	1.2

Standard concentration: 1.0 µg/mL
Acidity: 1% HCl

Table 18. Stability Data for Solution QF12

Element	<u>03/08/91-05/08/91</u> n=6		<u>05/16/91-09/13/91</u> n=12		<u>03/08/91-09/13/91</u> n=18		
	Ave1	SD1	Ave2	SD2	Ave	SD	RSD
Al	0.984	0.020	0.994	0.032	0.991	0.029	2.9
Ba	0.980	0.012	0.988	0.017	0.985	0.016	1.6
Be	0.973	0.009	0.973	0.014	0.973	0.013	1.3
Ca	0.974	0.014	0.976	0.015	0.975	0.015	1.5
Cd	0.989	0.019	0.996	0.016	0.994	0.017	1.7
Co	0.992	0.023	1.003	0.021	0.999	0.022	2.2
Cr	0.991	0.012	1.001	0.018	0.998	0.016	1.6
Cu	0.996	0.034	1.003	0.013	1.001	0.022	2.2
Fe	0.989	0.017	0.994	0.019	0.992	0.018	1.8
Mg	0.990	0.023	0.997	0.016	0.995	0.019	1.9
Mn	0.982	0.027	0.993	0.015	0.989	0.020	2.0
Ni	0.985	0.020	0.987	0.020	0.986	0.020	2.0
Pb	0.979	0.035	0.978	0.041	0.978	0.039	4.0
Sn	1.170	0.130	1.138	0.048	1.149	0.085	7.4
Sr	0.974	0.014	0.975	0.015	0.975	0.015	1.5
V	0.997	0.021	1.005	0.018	1.002	0.019	1.9
Zn	0.980	0.018	0.982	0.012	0.981	0.014	1.5
Zr	1.000	0.014	0.999	0.016	0.999	0.015	1.5

Standard concentration: 1.0 µg/mL
Acidity: 2% HCl

Table 19. Stability Data for Solution QF13

Element	<u>03/08/91-05/08/91</u> n=6		<u>05/16/91-09/13/91</u> n=12		<u>03/08/91-09/13/91</u> n=18		
	<u>Ave1</u>	<u>SD1</u>	<u>Ave2</u>	<u>SD2</u>	<u>Ave</u>	<u>SD</u>	<u>RSD</u>
Al	0.954	0.019	0.958	0.029	0.957	0.026	2.7
Ba	0.959	0.013	0.967	0.015	0.964	0.014	1.5
Be	0.969	0.010	0.968	0.013	0.968	0.012	1.2
Ca	0.956	0.015	0.961	0.014	0.959	0.014	1.5
Cd	0.969	0.020	0.979	0.016	0.976	0.017	1.8
Co	0.969	0.022	0.988	0.024	0.982	0.023	2.4
Cr	0.969	0.013	0.980	0.015	0.976	0.014	1.5
Cu	0.969	0.028	0.981	0.013	0.977	0.019	2.0
Fe	0.964	0.016	0.973	0.017	0.970	0.017	1.7
Mg	0.967	0.021	0.977	0.014	0.974	0.017	1.7
Mn	0.960	0.024	0.974	0.014	0.969	0.018	1.9
Ni	0.960	0.013	0.966	0.022	0.964	0.019	2.0
Pb	0.960	0.041	0.952	0.047	0.955	0.045	4.7
Sn	1.150	0.130	1.112	0.027	1.125	0.078	7.0
Sr	0.955	0.013	0.958	0.013	0.957	0.013	1.4
V	0.976	0.021	0.982	0.015	0.980	0.017	1.8
Zn	0.952	0.016	0.956	0.013	0.955	0.014	1.5
Zr	0.991	0.016	0.993	0.014	0.992	0.015	1.5

Standard concentration: 1.0 µg/mL
Acidity: 5% HCl

Table 20. Stability Data for Solution QF14

Element	<u>03/08/91-05/08/91</u> n=6		<u>05/16/91-09/13/91</u> n=12		<u>03/08/91-09/13/91</u> n=18	
	<u>Ave1</u>	<u>SD1</u>	<u>Ave2</u>	<u>SD2</u>	<u>Ave</u>	<u>RSD</u>
Al	0.978	0.018	0.982	0.042	0.981	3.7
Ba	0.980	0.012	0.984	0.021	0.983	1.9
Be	0.958	0.009	0.960	0.017	0.959	1.5
Ca	0.986	0.015	0.988	0.018	0.987	1.7
Cd	0.981	0.019	0.988	0.020	0.986	2.0
Co	0.994	0.022	1.000	0.024	0.998	2.3
Cr	0.993	0.013	1.000	0.023	0.998	2.0
Cu	0.984	0.031	0.993	0.019	0.990	2.4
Fe	0.988	0.017	0.994	0.024	0.992	2.2
Mg	0.988	0.022	0.994	0.017	0.992	1.9
Mn	0.981	0.025	0.993	0.018	0.989	2.1
Ni	0.982	0.014	0.985	0.023	0.984	2.1
Pb	0.974	0.042	0.972	0.052	0.973	5.0
Sn	1.160	0.110	1.118	0.056	1.132	6.9
Sr	0.973	0.014	0.974	0.018	0.974	1.7
V	0.998	0.022	1.004	0.020	1.002	2.1
Zn	0.962	0.018	0.966	0.013	0.965	1.5
Zr	0.994	0.011	0.994	0.020	0.994	1.8

Standard concentration: 1.0 µg/mL
Acidity: 1% HNO₃ + 5% HCl

Table 21. Stability Data for Solution QF15

Element	03/08/91-05/08/91 n=6		05/16/91-09/13/91 n=12		03/08/91-09/13/91 n=18		
	Ave1	SD1	Ave2	SD2	Ave	SD	RSD
Al	3.980	0.058	4.037	0.070	4.018	0.066	1.6
Ba	3.950	0.045	4.008	0.057	3.989	0.053	1.3
Be	3.940	0.019	3.946	0.050	3.944	0.042	1.1
Ca	3.990	0.043	4.013	0.039	4.005	0.040	1.0
Cd	3.950	0.039	3.972	0.060	3.965	0.054	1.4
Co	3.960	0.050	4.008	0.070	3.992	0.064	1.6
Cr	4.000	0.056	4.048	0.060	4.032	0.059	1.5
Cu	3.970	0.075	4.000	0.046	3.990	0.057	1.4
Fe	3.980	0.050	4.037	0.055	4.018	0.053	1.3
Mg	4.000	0.041	4.018	0.053	4.012	0.049	1.2
Mn	3.960	0.032	3.993	0.054	3.982	0.048	1.2
Ni	3.962	0.034	3.988	0.063	3.979	0.055	1.4
Pb	3.995	0.051	4.016	0.087	4.009	0.077	1.9
Sn	4.053	0.147	2.095	1.116	2.748	0.915	33.3
Sr	3.977	0.060	4.003	0.045	3.994	0.050	1.3
V	3.972	0.042	3.985	0.056	3.981	0.052	1.3
Zn	3.947	0.045	3.997	0.052	3.980	0.050	1.3
Zr	3.938	0.043	2.782	0.684	3.167	0.559	17.7

Standard concentration: 4.0 µg/mL
Acidity: 1% HNO₃ + 0.04% HCl

Table 22. Stability Data for Solution QF16

<u>Element</u>	<u>03/08/91-05/08/91</u> n=6		<u>05/16/91-09/13/91</u> n=12		<u>03/08/91-09/13/91</u> n=18	
	<u>Ave1</u>	<u>SD1</u>	<u>Ave2</u>	<u>SD2</u>	<u>Ave</u>	<u>SD</u> <u>RSD</u>
Al	3.955	0.048	4.004	0.045	3.988	0.046 1.2
Ba	3.943	0.047	3.988	0.046	3.973	0.046 1.2
Be	3.922	0.033	3.932	0.042	3.929	0.039 1.0
Ca	3.992	0.036	4.013	0.042	4.006	0.040 1.0
Cd	3.935	0.055	3.950	0.054	3.945	0.054 1.4
Co	3.967	0.075	3.982	0.058	3.977	0.064 1.6
Cr	3.985	0.054	4.033	0.049	4.017	0.051 1.3
Cu	3.942	0.062	3.977	0.036	3.965	0.046 1.2
Fe	3.977	0.050	4.023	0.057	4.008	0.055 1.4
Mg	3.992	0.044	4.003	0.046	3.999	0.045 1.1
Mn	3.948	0.037	3.977	0.043	3.967	0.041 1.0
Ni	3.957	0.041	3.978	0.054	3.971	0.050 1.3
Pb	3.957	0.069	3.983	0.068	3.974	0.068 1.7
Sn	4.158	0.122	4.058	0.106	4.091	0.112 2.7
Sr	3.972	0.056	3.993	0.042	3.986	0.047 1.2
V	3.965	0.048	3.983	0.043	3.977	0.045 1.1
Zn	3.898	0.033	3.949	0.042	3.932	0.039 1.0
Zr	4.005	0.019	3.982	0.040	3.990	0.034 0.9

Standard concentration: 4.0 µg/mL
Acidity: 2% HNO₃ + 0.04% HCl

Table 23. Stability Data for Solution QF17

Element	<u>03/08/91-05/08/91</u> n=6		<u>05/16/91-09/13/91</u> n=12		<u>03/08/91-09/13/91</u> n=18		
	<u>Ave1</u>	<u>SD1</u>	<u>Ave2</u>	<u>SD2</u>	<u>Ave</u>	<u>SD</u>	<u>RSD</u>
Al	3.830	0.058	3.858	0.050	3.849	0.053	1.4
Ba	3.848	0.058	3.888	0.052	3.875	0.054	1.4
Be	3.818	0.052	3.837	0.050	3.831	0.051	1.3
Ca	3.943	0.046	3.968	0.052	3.960	0.050	1.3
Cd	3.828	0.065	3.833	0.061	3.831	0.062	1.6
Co	3.862	0.087	3.886	0.060	3.878	0.070	1.8
Cr	3.890	0.060	3.940	0.057	3.923	0.058	1.5
Cu	3.832	0.060	3.868	0.044	3.856	0.050	1.3
Fe	3.878	0.062	3.920	0.068	3.906	0.066	1.7
Mg	3.898	0.051	3.909	0.048	3.905	0.049	1.3
Mn	3.858	0.049	3.880	0.050	3.873	0.050	1.3
Ni	3.862	0.059	3.866	0.065	3.865	0.063	1.6
Pb	3.807	0.057	3.824	0.085	3.818	0.077	2.0
Sn	3.997	0.138	3.958	0.078	3.971	0.102	2.6
Sr	3.902	0.060	3.928	0.048	3.919	0.052	1.3
V	3.882	0.062	3.898	0.050	3.893	0.054	1.4
Zn	3.732	0.034	3.779	0.046	3.763	0.042	1.1
Zr	3.950	0.038	3.940	0.049	3.943	0.046	1.2

Standard concentration: 4.0 µg/mL
Acidity: 5% HNO₃ + 0.04% H₂Cl

Table 24. Stability Data for Solution QF18

Element	03/08/91-05/08/91 n=6		05/16/91-09/13/91 n=12		03/08/91-09/13/91 n=18		
	Ave1	SD1	Ave2	SD2	Ave	SD	RSD
Al	4.000	0.063	4.008	0.064	4.005	0.064	1.6
Ba	3.948	0.061	3.965	0.076	3.959	0.071	1.8
Be	3.957	0.026	3.924	0.065	3.935	0.055	1.4
Ca	3.990	0.044	3.975	0.064	3.980	0.058	1.5
Cd	3.957	0.054	3.938	0.077	3.944	0.070	1.8
Co	3.980	0.064	3.969	0.081	3.973	0.076	1.9
Cr	3.987	0.063	4.007	0.074	4.000	0.071	1.8
Cu	4.003	0.086	3.997	0.054	3.999	0.066	1.7
Fe	3.983	0.059	3.998	0.089	3.993	0.080	2.0
Mg	4.007	0.052	3.987	0.065	3.994	0.061	1.5
Mn	3.957	0.046	3.956	0.067	3.956	0.061	1.5
Ni	3.963	0.041	3.955	0.080	3.958	0.069	1.8
Pb	3.972	0.053	3.971	0.110	3.971	0.095	2.4
Sn	4.117	0.116	4.013	0.130	4.048	0.126	3.1
Sr	3.985	0.073	3.970	0.062	3.975	0.066	1.7
V	3.972	0.055	3.958	0.062	3.963	0.060	1.5
Zn	3.980	0.054	3.991	0.067	3.987	0.063	1.6
Zr	3.978	0.019	3.919	0.058	3.939	0.049	1.2

Standard concentration: 4.0 µg/mL
Acidity: 1% HCl

Table 25. Stability Data for Solution QF19

<u>Element</u>	<u>03/08/91-05/08/91</u> n=6		<u>05/16/91-09/13/91</u> n=12		<u>03/08/91-09/13/91</u> n=18		
	<u>Ave1</u>	<u>SD1</u>	<u>Ave2</u>	<u>SD2</u>	<u>Ave</u>	<u>SD</u>	<u>RSD</u>
Al	3.970	0.066	3.995	0.072	3.987	0.070	1.8
Ba	3.928	0.067	3.968	0.073	3.955	0.071	1.8
Be	3.965	0.044	3.955	0.068	3.958	0.061	1.5
Ca	3.977	0.046	3.985	0.072	3.982	0.065	1.6
Cd	3.948	0.065	3.953	0.077	3.951	0.073	1.9
Co	3.958	0.083	3.988	0.070	3.978	0.075	1.9
Cr	3.970	0.070	4.008	0.076	3.995	0.074	1.9
Cu	3.982	0.076	3.995	0.065	3.991	0.069	1.7
Fe	3.970	0.066	4.004	0.086	3.993	0.080	2.0
Mg	3.988	0.057	3.991	0.069	3.990	0.065	1.6
Mn	3.942	0.056	3.959	0.066	3.953	0.063	1.6
Ni	3.957	0.065	3.964	0.080	3.962	0.075	1.9
Pb	3.943	0.075	3.977	0.092	3.966	0.087	2.2
Sn	4.020	0.153	4.081	0.122	4.061	0.133	3.3
Sr	3.965	0.071	3.977	0.068	3.973	0.069	1.7
V	3.957	0.066	3.965	0.064	3.962	0.065	1.6
Zn	3.953	0.047	3.988	0.067	3.976	0.061	1.5
Zr	4.003	0.031	3.977	0.064	3.986	0.055	1.4

Standard concentration: 4.0 µg/mL
Acidity: 2% HCl

Table 26. Stability Data for Solution QF20

Element	<u>03/08/91-05/08/91</u> n=6		<u>05/16/91-09/13/91</u> n=12		<u>03/08/91-09/13/91</u> n=18		
	<u>Ave1</u>	<u>SD1</u>	<u>Ave2</u>	<u>SD2</u>	<u>Ave</u>	<u>SD</u>	<u>RSD</u>
Al	3.880	0.054	3.898	0.072	3.892	0.067	1.7
Ba	3.867	0.058	3.903	0.071	3.891	0.067	1.7
Be	3.967	0.048	3.946	0.069	3.953	0.063	1.6
Ca	3.918	0.039	3.935	0.075	3.929	0.065	1.7
Cd	3.902	0.072	3.895	0.080	3.897	0.077	2.0
Co	3.908	0.076	3.911	0.079	3.910	0.078	2.0
Cr	3.907	0.053	3.945	0.074	3.932	0.068	1.7
Cu	3.910	0.060	3.923	0.065	3.919	0.063	1.6
Fe	3.905	0.061	3.937	0.085	3.926	0.078	2.0
Mg	3.925	0.046	3.926	0.070	3.926	0.063	1.6
Mn	3.875	0.042	3.893	0.067	3.887	0.060	1.5
Ni	3.882	0.064	3.896	0.073	3.891	0.070	1.8
Pb	3.862	0.068	3.886	0.097	3.878	0.088	2.3
Sn	4.035	0.127	3.982	0.123	4.000	0.124	3.1
Sr	3.907	0.055	3.925	0.069	3.919	0.065	1.7
V	3.892	0.059	3.900	0.067	3.897	0.064	1.7
Zn	3.865	0.033	3.896	0.064	3.886	0.056	1.4
Zr	3.968	0.033	3.946	0.066	3.953	0.057	1.4

Standard concentration: 4.0 µg/mL
Acidity: 5% HCl

Table 27. Stability Data for Solution QF21

Element	<u>03/08/91-05/08/91</u> n=6		<u>05/16/91-09/13/91</u> n=12		<u>03/08/91-09/13/91</u> n=18	
	<u>Ave1</u>	<u>SD1</u>	<u>Ave2</u>	<u>SD2</u>	<u>Ave</u>	<u>SD</u> <u>RSD</u>
Al	3.832	0.046	3.841	0.098	3.838	0.084 2.2
Ba	3.830	0.044	3.863	0.086	3.852	0.077 2.0
Be	3.913	0.049	3.890	0.094	3.898	0.082 2.1
Ca	3.895	0.036	3.905	0.087	3.902	0.074 1.9
Cd	3.837	0.066	3.833	0.095	3.834	0.086 2.2
Co	3.873	0.082	3.867	0.093	3.869	0.089 2.3
Cr	3.867	0.047	3.904	0.093	3.892	0.081 2.1
Cu	3.860	0.059	3.874	0.081	3.869	0.074 1.9
Fe	3.868	0.061	3.894	0.100	3.885	0.089 2.3
Mg	3.885	0.044	3.886	0.084	3.886	0.073 1.9
Mn	3.840	0.042	3.853	0.084	3.849	0.072 1.9
Ni	3.853	0.059	3.848	0.093	3.850	0.083 2.2
Pb	3.825	0.066	3.816	0.103	3.819	0.092 2.4
Sn	4.052	0.154	3.953	0.124	3.986	0.135 3.4
Sr	3.883	0.055	3.896	0.083	3.892	0.075 1.9
V	3.858	0.058	3.863	0.080	3.861	0.073 1.9
Zn	3.782	0.033	3.812	0.080	3.802	0.068 1.8
Zr	3.927	0.037	3.907	0.081	3.914	0.070 1.8

Standard concentration: 4.0 µg/mL
Acidity: 1% HNO₃ + 5% HCl

Table 28. Stability Data for Solution QF22

Element	<u>03/08/91-05/08/91</u> n=6		<u>05/16/91-09/13/91</u> n=12		<u>03/08/91-09/13/91</u> n=18		
	<u>Ave1</u>	<u>SD1</u>	<u>Ave2</u>	<u>SD2</u>	<u>Ave</u>	<u>SD</u>	<u>RSD</u>
Al	9.97	0.21	10.09	0.13	10.05	0.16	1.6
Ba	9.83	0.07	9.92	0.16	9.89	0.13	1.4
Be	9.82	0.09	9.85	0.15	9.84	0.14	1.4
Ca	10.12	0.13	10.20	0.16	10.17	0.15	1.5
Cd	9.92	0.12	9.97	0.17	9.95	0.15	1.6
Co	10.06	0.16	10.13	0.18	10.11	0.17	1.7
Cr	9.93	0.05	10.05	0.17	10.01	0.14	1.4
Cu	10.06	0.20	10.15	0.12	10.12	0.15	1.5
Fe	9.94	0.15	10.00	0.21	9.98	0.19	1.9
Mg	10.01	0.13	10.08	0.14	10.06	0.14	1.4
Mn	9.96	0.12	10.06	0.15	10.03	0.14	1.4
Ni	9.80	0.14	9.81	0.14	9.81	0.14	1.4
Pb	9.79	0.11	9.84	0.16	9.82	0.15	1.5
Sn	8.49	0.34	8.60	0.26	8.56	0.29	3.4
Sr	10.03	0.13	10.05	0.15	10.04	0.14	1.4
V	10.04	0.09	10.13	0.15	10.10	0.13	1.3
Zn	9.79	0.10	9.90	0.14	9.86	0.13	1.3
Zr	10.18	0.04	10.12	0.13	10.14	0.11	1.0

Standard concentration: 10.0 µg/mL
 Acidity: 2% HNO₃ + 0.1% HCl

Table 29. Stability Data for Solution QF23

Element	<u>03/08/91-05/08/91</u> n=6		<u>05/16/91-09/13/91</u> n=12		<u>03/08/91-09/13/91</u> n=18		
	<u>Ave1</u>	<u>SD1</u>	<u>Ave2</u>	<u>SD2</u>	<u>Ave</u>	<u>SD</u>	<u>RSD</u>
Al	9.87	0.18	9.96	0.20	9.93	0.19	1.9
Ba	9.73	0.08	9.77	0.22	9.76	0.19	1.9
Be	9.86	0.12	9.82	0.19	9.83	0.17	1.7
Ca	9.97	0.11	10.02	0.22	10.0	0.19	1.9
Cd	9.88	0.14	9.90	0.22	9.89	0.20	2.0
Co	9.96	0.20	10.01	0.23	9.99	0.22	2.2
Cr	9.84	0.08	9.92	0.23	9.89	0.19	2.0
Cu	10.05	0.19	10.07	0.22	10.06	0.21	2.1
Fe	9.83	0.16	9.86	0.24	9.85	0.22	2.2
Mg	9.92	0.13	9.97	0.22	9.95	0.19	2.0
Mn	9.87	0.14	9.92	0.20	9.90	0.18	1.8
Ni	9.72	0.15	9.70	0.20	9.71	0.18	1.9
Pb	9.76	0.17	9.70	0.19	9.72	0.18	1.9
Sn	8.53	0.45	8.51	0.28	8.52	0.35	4.1
Sr	9.92	0.10	9.90	0.20	9.91	0.17	1.7
V	9.95	0.09	9.97	0.20	9.96	0.17	1.7
Zn	9.82	0.12	9.87	0.18	9.85	0.16	1.6
Zr	10.12	0.04	10.01	0.19	10.05	0.16	1.6

Standard concentration: 10.0 µg/mL
Acidity: 2% HCl

Table 30. Stability Data for Solution QF24

Element	<u>03/08/91-05/08/91</u> n=6		<u>05/16/91-09/13/91</u> n=12		<u>03/08/91-09/13/91</u> n=18		
	<u>Ave1</u>	<u>SD1</u>	<u>Ave2</u>	<u>SD2</u>	<u>Ave</u>	<u>SD</u>	<u>RSD</u>
Al	9.55	0.20	9.69	0.15	9.64	0.17	1.7
Ba	9.47	0.12	9.58	0.15	9.54	0.14	1.5
Be	9.75	0.19	9.77	0.17	9.76	0.18	1.8
Ca	9.77	0.14	9.89	0.17	9.85	0.16	1.6
Cd	9.62	0.22	9.71	0.17	9.68	0.19	1.9
Co	9.70	0.25	9.78	0.19	9.75	0.21	2.2
Cr	9.58	0.12	9.72	0.18	9.67	0.16	1.7
Cu	9.74	0.18	9.84	0.14	9.81	0.15	1.6
Fe	9.59	0.21	9.66	0.19	9.64	0.20	2.0
Mg	9.68	0.18	9.77	0.16	9.74	0.17	1.7
Mn	9.61	0.16	9.73	0.15	9.69	0.15	1.6
Ni	9.47	0.21	9.49	0.16	9.48	0.18	1.9
Pb	9.40	0.20	9.45	0.18	9.43	0.19	2.0
Sn	8.42	0.49	8.34	0.23	8.37	0.34	4.1
Sr	9.72	0.14	9.77	0.16	9.75	0.15	1.6
V	9.71	0.14	9.80	0.16	9.77	0.15	1.6
Zn	9.42	0.15	9.53	0.14	9.49	0.14	1.5
Zr	9.93	0.09	9.93	0.15	9.93	0.13	1.3

Standard concentration: 10.0 µg/mL
Acidity: 1% HNO₃ + 5% HCl

Table 31. Stability Data for Solution QI01

Element	Actual Conc. ($\mu\text{g/mL}$)	03/08/91-05/08/91		05/16/91-09/13/91		03/08/91-09/13/91		
		Ave1	SD1	Ave2	SD2	Ave	SD	RSD
		n=5		n=8		n=13		
Al	2.010	1.970	0.023	1.984	0.050	1.979	0.042	2.1
Ba	1.960	1.968	0.022	1.978	0.034	1.974	0.030	1.5
Be	0.483	0.481	0.010	0.481	0.008	0.481	0.009	1.8
Ca	48.700	46.330	0.450	46.770	0.944	46.650	0.840	1.8
Cd	0.502	0.487	0.015	0.492	0.010	0.490	0.012	2.5
Co	0.502	0.499	0.022	0.510	0.017	0.506	0.019	3.8
Cr	0.510	0.481	0.011	0.488	0.011	0.485	0.011	2.3
Cu	0.515	0.529	0.013	0.532	0.010	0.531	0.011	2.1
Fe	2.040	1.942	0.041	1.950	0.042	1.947	0.042	2.1
Mg	24.600	24.070	0.350	23.990	0.419	24.010	0.380	1.6
Mn	0.504	0.514	0.012	0.519	0.009	0.517	0.010	2.0
Ni	0.480	0.486	0.021	0.487	0.017	0.487	0.019	3.8
Pb	4.800	4.692	0.102	4.723	0.099	4.711	0.100	2.1
V	0.488	0.531	0.015	0.536	0.012	0.534	0.013	2.5
Zn	3.100	2.798	0.036	2.814	0.050	2.808	0.045	1.6

ICV-1: 1:10 dilution
Acidity: 2% HNO₃

Table 32. Stability Data for Solution QI02

Element	Actual Conc. ($\mu\text{g/mL}$)	03/08/91-05/08/91 n=5		05/16/91-09/13/91 n=8		03/08/91-09/13/91 n=13		
		Ave1	SD1	Ave2	SD2	Ave	SD	RSD
Al	2.010	1.864	0.038	1.984	0.050	1.938	0.046	2.4
Ba	1.960	1.894	0.030	1.978	0.034	1.946	0.033	1.7
Be	0.483	0.478	0.014	0.481	0.008	0.480	0.011	2.2
Ca	48.700	44.800	0.700	44.800	0.700	45.420	1.020	2.2
Cd	0.502	0.473	0.017	0.492	0.010	0.485	0.013	2.7
Co	0.502	0.484	0.021	0.510	0.017	0.500	0.019	3.7
Cr	0.510	0.462	0.014	0.488	0.011	0.478	0.012	2.6
Cu	0.515	0.504	0.014	0.532	0.010	0.521	0.012	2.2
Fe	2.040	1.876	0.048	1.950	0.042	1.922	0.044	2.3
Mg	24.600	23.270	0.510	23.990	0.419	23.300	0.490	2.1
Mn	0.504	0.494	0.013	0.519	0.009	0.509	0.011	2.1
Ni	0.480	0.465	0.023	0.487	0.017	0.479	0.020	4.1
Pb	4.800	4.486	0.096	4.723	0.099	4.632	0.098	2.1
V	0.488	0.510	0.018	0.536	0.012	0.526	0.015	2.8
Zn	3.100	2.664	0.056	2.814	0.050	2.756	0.052	1.9

ICV-1: 1:10 dilution
Acidity: 1% HNO₃ + 5% HCl

Table 33. Stability Data for Silver Solutions

File ^a	<u>03/08/91-05/08/91</u>			<u>03/08/91-09/13/91</u>		
	Ave^b	SDⁿ⁼⁶	RSD¹	Ave^b	SDⁿ⁼¹⁷	RSD
QA01	0.205	0.006	2.9	0.204	0.008	3.9
QA02	0.199	0.006	3.0	1.200	0.007	3.5
QA03	0.190	0.005	2.6	0.189	0.007	3.7
QA04	1.042	0.017	1.6	1.039	0.018	1.7
QA05	1.022	0.016	1.6	1.023	0.017	1.7
QA06	0.990	0.012	1.2	0.990	0.015	1.5
QA07	4.057	0.057	1.4	4.036	0.056	1.4
QA08	4.002	0.072	1.8	3.998	0.060	1.5
QA09	3.882	0.065	1.7	3.886	0.052	1.3
QA10	10.18	0.16	1.6	10.06	0.19	1.9
QIC1	0.516	0.016	3.1	0.520	0.017	3.3
QIC2	0.489	0.015	3.1	0.492	0.018	3.6

^a Acid concentrations: QA01, QA04, QA07 = 1% HNO₃
 QA02, QA05, QA08, QA10, QIC1 = 2% HNO₃
 QA03, QA06, QA09 = 5% HNO₃
 QIC2 = 5% HCl + 1% HNO₃

^b Silver concentrations, µg/mL

Table 34. Stability Data for Tin Solutions

File ^a	<u>03/08/91-05/08/91</u>			<u>03/08/91-09/13/91</u>		
	Ave^b	SDⁿ⁼⁶	RSD¹	Ave^b	SDⁿ⁼¹⁷	RSD
QSn2	4.278	0.099	2.3	4.282	0.082	1.9
QSn3	4.825	0.101	2.1	4.852	0.111	2.3
QSn5	19.85	0.58	2.9	18.16	2.18	12.0
QSn6	19.12	0.36	1.9	19.12	0.34	1.8

^a Acid concentrations: QSn2, QSn5 = 2% HCl
 QSn3, QSn6 = 5% HCl + 1% HNO₃

^b Tin concentrations, µg/mL

Table 35. Stability Data for Solutions QSi

Element	03/08/91-05/08/91			03/08/91-09/13/91		
	Ave1^a	SD1ⁿ⁼⁶	RSD1	Ave^a	SDⁿ⁼¹⁷	RSD
QSi1: 2% HNO₃						
B	0.493	0.009	1.8	0.492	0.016	3.2
Mo	0.497	0.019	3.8	0.494	0.021	4.2
Si	1.048	0.027	2.6	1.045	0.055	5.3
Ti	0.503	0.015	1.0	0.507	0.005	1.0
QSi2: 2% HCl						
B	0.508	0.011	2.2	0.516	0.017	3.3
Mo	0.511	0.017	3.3	0.517	0.019	3.7
Si	1.030	0.014	1.3	1.035	0.053	5.1
Ti	0.498	0.004	0.8	0.504	0.007	1.4
QSi3: 5% HCl + 1% HNO₃						
B	0.518	0.019	3.7	0.522	0.021	4.0
Mo	0.505	0.015	3.0	0.509	0.017	3.3
Si	1.006	0.042	4.2	1.029	0.059	5.7
Ti	0.487	0.007	1.4	0.494	0.012	2.4
QSi4: 2% HNO₃						
B	2.437	0.016	0.7	2.405	0.072	3.0
Mo	2.452	0.018	0.7	2.446	0.040	1.6
Si	5.205	0.107	2.0	5.137	0.241	4.7
Ti	2.495	0.024	1.0	2.495	0.067	2.7
QSi5: 2% HCl						
B	2.533	0.027	1.1	2.519	0.103	4.1
Mo	2.487	0.043	1.7	2.489	0.045	1.8
Si	5.128	0.052	1.0	5.159	0.215	4.2
Ti	2.480	0.019	0.8	2.505	0.048	1.9

Table 35. Stability Data for Solutions QSi (contd)

Element	<u>03/08/91-05/08/91</u>			<u>03/08/91-09/13/91</u>		
	Ave1^a	SD1ⁿ⁼⁶	RSD1	Ave^a	SDⁿ⁼¹⁷	RSD
QSi6: 5% HCl + 1% HNO₃						
B	2.543	0.037	1.4	2.536	0.103	4.1
Mo	2.462	0.043	1.7	2.468	0.045	1.8
Si	5.017	0.075	1.5	5.076	0.215	4.2
Ti	2.422	0.028	1.1	2.461	0.048	1.9
QSi7: 2% HNO₃						
B	9.83	0.09	0.9	9.78	0.12	1.2
Mo	9.89	1.60	1.6	9.95	0.15	1.5
Si	20.77	0.51	2.4	21.06	0.76	3.6
Ti	10.06	0.08	0.8	10.13	0.17	1.7
QSi8: 2% HCl						
B	10.03	0.12	1.2	10.01	0.11	1.1
Mo	9.89	0.11	1.1	9.87	0.15	1.5
Si	20.32	0.34	1.7	20.68	0.81	3.9
Ti	9.81	0.10	1.0	9.92	0.16	1.6
QSi9: 5% HCl + 1% HNO₃						
B	10.01	0.11	1.1	10.03	0.12	1.2
Mo	9.69	0.08	0.8	9.73	0.14	1.4
Si	19.82	0.26	1.3	20.25	0.75	3.7
Ti	9.67	0.10	1.0	9.78	0.19	1.9

^a Concentrations, µg/mL

Table 36. Stability Data for Solutions QAs

Element	<u>03/08/91-05/08/91</u>			<u>03/08/91-09/13/91</u>		
	Ave1^a	SD1	RSD1	Ave^a	SD	RSD
QAs1: 2% HNO₃						
As	5.058	0.054	1.1	5.087	0.083	1.6
Sb	5.183	0.080	1.5	5.418	0.373	6.9
Se	4.893	0.047	1.0	4.862	0.101	2.1
QAs2: 2% HCl						
As	4.973	0.082	1.6	5.008	0.101	2.0
Sb	5.083	0.093	1.8	5.313	0.384	7.2
Se	4.855	0.107	2.2	4.803	0.124	2.6
QAs3: 5% HCl + 1% HNO₃						
As	4.917	0.067	1.4	4.961	0.115	2.3
Sb	5.020	0.126	2.5	5.249	0.368	7.0
Se	4.790	0.080	1.7	4.774	0.097	2.0
QAs4: 2% HNO₃						
As	20.22	0.25	1.2	20.37	0.40	2.0
Sb	20.72	0.45	2.2	21.65	1.51	7.0
Se	20.17	0.26	1.3	20.16	0.39	1.9
QAs5: 2% HCl						
As	19.77	0.25	1.3	19.95	0.41	2.0
Sb	20.22	0.39	1.9	21.14	1.47	7.0
Se	19.83	0.19	1.0	19.82	0.38	1.9
QAs6: 5% HCl + 1% HNO₃						
As	19.33	0.31	1.6	19.61	0.38	1.9
Sb	19.72	0.42	2.1	20.72	1.45	7.0
Se	19.45	0.28	1.4	19.49	0.32	1.6

^a Concentrations, µg/mL

Table 37. Stability Data for Niobium Solutions

File ^a	<u>04/29/91-06/24/91</u>			<u>04/29/91-10/21/91</u>		
	Ave1 ^b	SD1	RSD1	Ave ^a	SD	RSD
QNb1	0.918	0.020	2.2	0.895	0.027	3.0
QNb2	0.970	0.024	2.5	0.971	0.023	2.4
QNb3	4.820	0.187	3.9	4.440	0.419	9.4
QNb4	5.037	0.089	1.8	4.981	0.093	1.9
QNb5	20.00	0.26	1.3	20.00	0.32	1.6
QNb6	20.13	0.33	1.6	20.11	0.33	1.6

^a Acid concentrations: QNb1, QNb3, QNb5 = 2% HCl

QNb2, QNb4, QNb6 = 2% HCl + 0.5% HF

^b Niobium concentrations, µg/mL

Table 38. Stability Data for Tungsten Solutions

File ^a	<u>04/29/91-06/24/91</u>			<u>04/29/91-10/21/91</u>		
	Ave1 ^b	SD1	RSD1	Ave ^b	SD	RSD
QW01	0.961	0.013	1.4	0.960	0.023	2.4
QW02	0.987	0.027	2.7	0.987	0.037	3.7
QW03	4.842	0.041	0.8	4.320	0.739	17.1
QW04	4.953	0.052	1.0	4.950	0.056	1.1
QW05	19.93	0.07	0.4	17.20	3.49	20.3
QW06	20.17	0.15	0.7	20.09	0.21	1.0

^a Acid concentrations: QW01, QW03, QW05 = 2% HCl

QW02, QW04, QW06 = 2% HCl + 0.5% HF

^b Tungsten concentrations, µg/mL

Figure 1. Signal-to-Background Ratio vs. Acidity

Figure 2. Relative Standard Deviation for Aluminum

Figure 3. Relative Standard Deviation for Beryllium

Figure 4. Relative Standard Deviation for Chromium

Figure 5. Relative Standard Deviation for Nickel

Figure 6. Relative Standard Deviation for Lead

Figure 7. Relative Standard Deviation for Tin

Distribution for ANL/ACL-92/3Internal:

I. Ahmad	M. L. Dietz	S. D. Kent
O. O. Ajayi	J. D. Ditmars	C. E. Klotz
I. Ambats	H. Drucker	A. J. Kras
D. V. Applegate	F. J. Dudek	A. R. Krauss
J. G. Asbury	B. D. Dunlap	J. R. Krsul
R. W. Atcher	A. J. Dvorak	M. Lachman
M. D. Atella	W. L. Ebert	L. L. Lamoureux
S. W. Ballou	A. G. Engelkemeir	D. Larsen
A. Bandopadhyaya	M. D. Erickson	S. L. Lawrence
R. W. Bane	A. M. Essling	K. H. Leong
D. A. Bass	W. D. Fairman	L. G. LeSage
J. K. Bates	I. M. Fox	M. A. Lewis
J. E. Battles	F. Y. Fradin	P. C. Lindahl
C. A. Bebrich	N. W. Golchert	M. J. Lineberry
N. J. Beskid	A. J. Goldman	C. D. Livengood
M. H. Bhattacharyya	H. T. Goodspeed	C. D. O. Loureiro
S. K. Bhattacharyya	D. G. Graczyk	J. F. Marchaterre
N. Bhatti	D. W. Green	F. Markun
M. Blander	D. M. Gruen	V. A. Maroni
D. H. Bomkamp	W. H. Gunther	J. J. Marr
A. S. Boparai	J. E. Harmon	F. J. Martino
S. S. Borys	W. Harrison	J. R. McCreary
D. L. Bowers	D. A. Haugen	H. F. McFarlane
B. S. Brown	R. R. Heinrich	C. A. Melendres
F. A. Cafasso	J. E. Helt	V. A. Mendez (10)
S. E. Carpenter	W. F. Henning	D. E. Moncton
Y. I. Chang	J. E. Herceg	L. R. Morss
M. Clapper-Gowdy	R. L. Herriford	T. P. Mulcahey
R. B. Clare	I. Hlohowskyj	K. M. Myles
S. M. Clark	E. P. Horwitz	H. W. Myron
J. H. Cleland	H.-S. Huang	Z. Nagy
J. T. Collins	E. Huberman	B. W. Nashold
H. W. Conner	D. R. Huff (10)	L. A. Neimark
J. S. Crain	E. A. Huff (10)	P. A. Nelson
E. J. Croke	E. Y. Hwang	R. D. Nixon
J. C. Cunnane	M. J. Janik	J. W. O'Kelley
L. W. Deitrich	K. J. Jensen	Y. Orechwa
J. C. Demirgian	D. O. Johnson	K. J. Parish
C. B. Dennis	P. L. Johnson	J.-H. Park
M. H. Derbidge	S. N. Kakar	D. C. Parzyck
D. R. Diercks	T. F. Kassner	T. L. Patton

C. V. Pearson
 D. R. Pedersen
 M. J. Pellin
 M. Petrick
 K. C. Picel
 D. S. Poa
 R. B. Poeppel
 A. J. Policastro
 Y. L. Qian
 A. C. Raptis
 E. G. Rauh
 D. M. Ray
 G. T. Reedy
 C. A. Reilly
 M. J. Robinet
 B. J. Roop
 L. E. Ross
 C. S. Sabau
 N. A. Sacco-Gibson
 N. F. Sather
 W. W. Schertz
 A. Schriesheim
 W. J. Shack
 G. K. Shenoy
 J. G. Sleeth
 D. L. Smith

D. L. Smith
 F. P. Smith
 L. L. Smith
 R. K. Smither
 C. T. Snyder
 R. W. Springer
 B. Srinivasan
 R. E. Stajdohar
 V. C. Stamoudis
 M. J. Steindler
 C. M. Stevens
 L. M. Stock
 W. E. Streets
 K. Sugano
 T. G. Surles
 W. M. Swift
 T. A. Taiwo
 B. S. Tani
 J. D. Taylor
 L. B. TenKate
 T. TenKate
 R. J. Teunis
 C. E. Till
 D. Tomasko
 Z. Tomczuk

R. E. Toohey
 A. Travelli
 Y. Tsai
 J. P. Unik
 R. A. Valentin
 G. F. Vandegrift
 R. Venkatadri
 K. S. Vorres
 D. C. Wade
 L. C. Walters
 D. W. Warren
 D. P. Weber
 R. W. Weeks
 M. L. Wesely
 J. M. Williams
 R. E. Winans
 R. D. Wolson
 J. L. Woodring
 J. J. Wray
 R. A. Wynveen
 J. S. Yaeger
 R. K. Yoo
 ANL Patent Dept.
 ANL Contract File
 TIS Files (3)

External:

DOE-OSTI (2)

ANL Library (2)

Manager, Chicago Operations Office, DOE

T. A. Baillieul, DOE-CH

J. C. Haugen, DOE-CH

A. L. Taboas, DOE-CH-AAO

Chemical Technology Division Review Committee Members:

S. Baron, Brookhaven National Laboratory, Upton, NY

D. L. Douglas, Consultant, Bloomington, MN

N. Jarrett, Noel Jarrett Associates, Lower Burrell, PA

J. G. Kay, Drexel University, Philadelphia, PA

J. Stringer, Electric Power Research Institute, Palo Alto, CA

J. B. Wagner, Arizona State University, Tempe, AZ

R. G. Wymer, Oak Ridge National Laboratory, Oak Ridge, TN

A. Alford-Stevens, U.S. Environ. Protection Agency, Cincinnati, OH

T. Beasley, U. S. Environmental Measurements Lab., New York, NY

J. Bennett, EG&G Idaho, Inc., Idaho Falls, ID
 R. E. Berkeley, U.S. Environ. Protection Agency, Research Triangle Park, NC
 C. D. Bingham, New Brunswick Laboratory, Argonne, IL
 D. Bogen, USDOE, New York, NY
 N. F. Christopher, Martin Marietta Energy Systems, Piketon, OH
 G. Clark, Reynolds Electric & Engineering, Las Vegas, NV
 C. E. Coffey, Westinghouse Savannah River Co., Aiken, SC
 P. T. Cunningham, Los Alamos National Laboratory, Los Alamos, NM
 J. L. Daniel, Battelle Pacific Northwest Laboratory, Richland, WA
 H. J. Dewey, Los Alamos National Laboratory, Los Alamos, NM
 D. A. Dodd, Westinghouse Hanford Co., Richland, WA
 J. G. Donnon, Savannah River Plant, Aiken, SC
 J. G. Dorsey, Martin Marietta Energy Systems, Oak Ridge, TN
 L. P. Duffy, USDOE, Environ. Restoration & Waste Manage., Washington, DC
 P. A. Duhamel, USDOE, Health and Environmental Research, Germantown, MD
 C. Elly, U. S. Environmental Protection Agency, Chicago, IL
 D. J. Fauth, Analytical Lab West Valley Nuclear Services, West Valley, NY
 C. Fellers, Mound Laboratory, Miamisburg, OH
 R. B. Fitts, Oak Ridge National Laboratory, Oak Ridge, TN
 C. W. Frank, USDOE, Office of Technology Development, Washington, DC
 V. Gill, Westinghouse, Cincinnati, OH
 J. Goldstein, USDOE, Health and Environmental Research, Germantown, MD
 T. C. Greengard, Rocky Flats Plant, Golden, CO
 J. Harness, USDOE, Morgantown Energy Technology Center, Morgantown, WV
 D. L. Herman, Westinghouse Materials Company of Ohio, Cincinnati, OH
 D. Hunter, Rockwell International, Golden, CO
 W. R. Jacobsen, Westinghouse Savannah River Co., Aiken, SC
 W. Killian, Ferris State University, Big Rapids, MI
 P. Krey, USDOE, Environmental Measurements Labs, New York, NY
 W. R. Laing, Oak Ridge National Laboratory, Oak Ridge, TN
 J. M. Latkovich, Battelle Pacific Northwest Laboratory, Richland, WA
 R. G. Lewis, U.S. Environ. Protection Agency, Research Triangle Park, NC
 L. C. Lewis, Westinghouse Idaho Nuclear Company, Idaho Falls, ID
 D. Lillian, USDOE, Office of Technology Development, Washington, DC
 R. M. Lugar, EG&G Idaho, Inc., Idaho Falls, ID
 C. Marcelo, New Brunswick Laboratory, Argonne, IL
 W. A. McClenny, U.S. Environ. Protection Agency, Research Triangle Park, NC
 S. Messenger, Morton Arboretum, Lisle, IL
 R. W. Morrow, Martin Marietta Energy Systems, Oak Ridge, TN
 J. D. Mulik, U.S. Environ. Protection Agency, Research Triangle Park, NC
 R. Norman, Gas Research Institute, Chicago, IL
 R. Ohlinger, Westinghouse Materials Company of Ohio, Cincinnati, OH
 R. D. Oldham, New Brunswick Laboratory, Argonne, IL
 J. Phillips, Los Alamos National Laboratory, Los Alamos, NM
 J. D. Pleil, U.S. Environ. Protection Agency, Research Triangle Park, NC
 F. Podosek, Washington University, St. Louis, MO

J. W. Pyper, Lawrence Livermore National Laboratory, Livermore, CA
J. M. Ramirez, Allied-Signal, Inc., Kansas City, MO
C. Ross, U.S. Environmental Protection Agency, Chicago, IL
R. Scott, USDOE, Environmental Quality Assurance and Quality Control, Washington, DC
J. Q. Searcy, Sandia National Laboratories, Albuquerque, NM
W. D. Shults, Oak Ridge National Laboratory, Oak Ridge, TN
J. Smith, University of Chicago, Chicago, IL
J. R. Stetter, Transducer Research, Inc., Palos Hills, IL
F. D. Stevenson, USDOE, Basic Energy Sciences, Germantown, MD
C. Stroup, Westinghouse Hanford Company, Richland, WA
N. Trahey, National Institute of Standards and Technology, Gaithersburg, MD
R. Villarreal, Los Alamos National Laboratory, Los Alamos, NM
H. A. Vincent, U.S. Environmental Protection Agency, Las Vegas, NV
W. M. Walsh, Health Research Institute, Naperville, IL
W. C. Weimer, Battelle Pacific Northwest Laboratory, Richland, WA
R. P. Whitfield, USDOE, Environmental Restoration, Washington, DC
C. G. Whitworth, Jr., MSE, Inc., Butte, MT
R. J. Wingender, Gurnee, IL
P. D. Wooldridge, Paducah Gaseous Diffusion Plant, Paducah, KY
J. Zarret, EG&G Rocky Flats Plant, Golden, CO

END

**DATE
FILMED**

2 / 2 / 93