WSRC-RP-93-294

VISCOSITY MEASUREMENTS ON CLEAR LIQUIDS (U)

by D. D. Walker

Westinghouse Savannah River Company
Savannah River Site
Aiken, South Carolina 29808

WSRC-RP--93-294

DE93 009893

Other Authors:

This paper was prepared in connection with work done under Contract No. DE-AC09-89SR18035 with the U. S. Department of Energy. By acceptance of this paper, the publisher and/or recipient acknowledges the U. S. Government's right to retain a nonexclusive, royalty-free license in and to any copyright covering this paper, along with the right to reproduce and to authorize others to reproduce all or part of the copyrighted paper.

MASTER

DISTRIBUTION OF THIS DOCUMENT IS UNLIMITED

DISCLAIMER

This report was prepared as an account of work sponsored by an agency of the United States Government. Neither the United States Government nor any agency thereof, nor any of their employees, makes any warranty, express or implied, or assumes any legal liability or responsibility for the accuracy, completeness, or usefulness of any information, apparatus, product, or process disclosed, or represents that its use would not infringe privately owned rights. Reference herein to any specific commercial product, process, or service by trade name, trademark, manufacturer, or otherwise does not necessarily constitute or imply its endorsement, recommendation, or favoring by the United States Government or any agency thereof. The views and opinions of authors expressed herein do not necessarily state or reflect those of the United States Government or any agency thereof.

This report has been reproduced directly from the best available copy.

Available to DOE and DOE contractors from the Office of Scientific and Technical Information, P.O. Box 62, Oak Ridge, TN 37831; prices available from (615) 576-8401, FTS 626-8401.

Available to the public from the National Technical Information Service, U.S. Department of Commerce, 5285 Port Royal Rd., Springfield, VA 22161.

Keywords: In-tank process

Retention time: permanent

cc: G. K. Georgetown, 703-H
W. B. Vanpelt, 241-152H
D. J. McCabe, 676-T
D. B. Burns, 676-T
J. N. Brooke, 241-120H
SRTC Records (4)
February 9, 1993

TO: D. L. FISH, 773-A

FROM: D. D. WALKER, 773-A *DD Walker*

VISCOSITY MEASUREMENTS ON CLEAR LIQUIDS (U)

Summary

A procedure for measuring the viscosity of clear, aqueous, liquids has been developed to support the In-Tank Precipitation (ITP) cold chemical testing program. Viscosity measurements on the filtrate going to the benzene strippers are required in the evaluation of the stripper performance. The measured viscosities have been shown to be accurate to within $\pm 3.1\%$ (95% confidence level),¹ which is within the $\pm 5\%$ requested by Waste Management.² Further explanatory and supporting data are given below for T. B. Edwards' report "Statistical Support for Viscosity Study (U)," SCS-ASG-930008, January 22, 1993.¹

Introduction

During the ITP cold chemical testing program, the efficiency of the benzene strippers will be measured and evaluated. Since the stripping efficiency is partially dependent upon the dynamic viscosity of the liquid phase, this property must be measured on samples taken during the test program. A procedure to measure the dynamic viscosity of salt solutions was developed from standard American Society of Testing and Materials (ASTM) methods.³ The SRS procedure differs from the ASTM procedure and, therefore, a test program was initiated to determine its accuracy and precision. The results of these statistically designed tests are reported elsewhere,¹ but supporting information on the experimental procedures, standards, and equipment are given below.

Viscometers

The viscometers used in these studies were routine Cannon-Fenske type glass viscometers for use with clear liquids. They were obtained from two manufacturers: Viscometer #125 from International Research Glassware, Kenilworth, NJ; and Viscometers #428 and #430 from Cannon Instrument Co., State College, Pa. The

viscometers were Size #50 which are used for liquids with dynamic viscosities between 1 and 4 cP.

A certified calibration from the manufacturer was provided with each viscometer, but these calibrations were not used. This type of viscometer is normally used to measure viscosities of oils and is calibrated with mineral oil standards. Due to the difference in surface tension between mineral oil and water, the certified calibration is not accurate for the ITP salt solutions. Water was used in this study as the calibration standard.

Procedure for Measuring Viscosity

To measure the kinematic viscosity of a liquid (in centistokes, cSt), one first measures the time required for a certain volume of liquid to drain through a capillary tube in the viscometer. Using the viscometer constant, the kinematic viscosity is calculated from:

$$n = k \cdot t \quad \text{where} \quad \begin{array}{l} n = \text{kinematic viscosity (cSt)} \\ k = \text{viscometer constant (cSt/sec)} \\ t = \text{efflux time (sec)}. \end{array}$$

The viscometer constant is measured using a standard solution similar to the solutions of interest. In this case, deionized water was used as the standard.

The dynamic viscosity (in centipoise, cP) is calculated from the kinematic viscosity by multiplying by the density (in g/mL):

$$N = n \cdot D \quad \text{where} \quad \begin{array}{l} N = \text{dynamic viscosity (cP)} \\ D = \text{density (g/mL)}. \end{array}$$

The density of the sample is measured using volumetric glassware.

A procedure for measuring viscosities has been written and approved for use within Interim Waste Technology.⁴ This procedure was used for the measurements reported here and in Reference 1.

The major differences between the IWT procedure and the ASTM procedure are the required accuracy of the thermometer, the temperature stability within the water bath, and the purity of water used as the calibration standard. The ASTM method requires a thermometer accurate to $\pm 0.02^\circ\text{C}$, whereas the thermometers used in the IWT procedure can only be read to $\pm 0.1^\circ\text{C}$ and are calibrated to $\pm 0.2^\circ\text{C}$. The ASTM method requires that the temperature range between the top and bottom of the viscometer be no greater than 0.01°C . Temperature stability of the water bath used in these tests was measured and found to vary by as much as 0.15°C between the top and bottom of the viscometer. The ASTM method for calibrating viscometers⁵ requires distilled or deionized water, freshly distilled on the day of use. Deionized water was used, but it was not routinely redistilled before use (see below).

Viscometer Calibration with Water

The viscometers were calibrated with water at 40°C . The

TABLE I. Efflux Times for Water from Various Sources

<u>Source</u>	<u>Efflux time (sec)</u>
Deionized (Lab B-103)	212.68 $\pm$.10
Domestic water (Lab B-103)	212.40 $\pm$.05
Deionized (Lab B-130)	212.47 $\pm$.15
Deionized/boiled (Lab B-130)	212.54 $\pm$.11
Deionized/distilled (Lab B-130)	212.34 $\pm$.11

The $\pm$ -range shown is the standard deviation of three measurements. All measurements were made by the same operator using one viscometer.

kinematic viscosity of water at 40°C is 0.6580 cSt.⁶ The water used in the tests was deionized. However, several sources and treatments of water were tested, including freshly distilled water, and no large differences in efflux time were observed (see Table I). Deionized water from the deionizing train in Lab B-130, Bldg. 773-A, was used for all tests in the statistical evaluation of the procedure.

All viscosity measurements were made with the thermometer reading 40.0°C ($\pm$ 0.1°C). The emergent-stem correction for the thermometers at 40°C was calculated to be +0.28°C, but this correction was not applied to the readings. A 0.28°C temperature measurement bias causes a 0.5% decrease in the viscosity of water at 40°C.⁶ The effect of random temperature errors or fluctuations during the measurement on the efflux time was determined for water (Figure 1). A 0.1°C temperature increment will change the efflux time by 0.18%.

The amount of sample used in each viscometer is shown in Table II. It is necessary to accurately reproduce the volume of liquid in the viscometer each time that it is filled with sample. The procedure requires that the amount of sample be within $\pm$ 0.1 mL of the amount used to calibrate the viscometer. This is accomplished by measuring the weight of sample instead of measuring the volume. The reproducibility of filling can be maintained to within $\pm$ 0.03 grams using an analytical balance and eyedropper (one drop of water from an eyedropper weighs about 0.03 grams). The effect on efflux time of placing different amounts of sample into the viscometer is shown

TABLE II. Sample Volumes for Viscometers

<u>Viscometer</u>	<u>Sample Volume (mL)</u>
125	7.05
U428	6.75
U430	6.75

FIGURE 1. Effect of Temperature on Efflux Time

FIGURE 2. Effect of Sample Size on Efflux Time

TABLE III. Comparison of Viscometer Constants to Certified Calibrations (40°C)

<u>Viscometer</u>	Constant (cSt/sec)		
	SRS <u>(water)</u>	SRS <u>(corrected)</u>	Vendor <u>(mineral oil)</u>
125	0.003109	0.003121	0.003117
U428	0.004002	0.004024	0.004049
U430	0.003971	0.003992	0.004023

in Figure 2. A 0.1 gram error in filling the viscometer will change the efflux time by 0.14%.

Comparison of Viscometer Constants to Certified Calibrations

Each viscometer was purchased with a certified calibration from the vendor. However, the viscometers were recalibrated to determine the viscometer constants with water at 40°C. A surface tension correction was applied to each determination with water so that a direct comparison to the vendor calibration could be made. The corrected viscometer constants determined at the Savannah River Site agreed with the vendor calibrations within 0.8% (Table III).¹ Values for the density and surface tension of water and mineral oil were found in the literature.⁶

Viscosity Measurements on Standard Compounds

The dynamic viscosities of three common chemical compounds (ethanol, isopropanol, and n-butanol) were measured to check the accuracy of the viscometer calibrations and technique (Table IV). The measured kinematic viscosities agreed within 1.1% with the values reported in the literature.⁷ The three liquids were purchased as American Chemical Society reagent-grade chemicals and their purity was not verified. The entries in Table IV include corrections for differences in surface tension.

TABLE IV. Densities and Viscosities of Standard Compounds (40°C)

Compound	Density (g/mL)		Kinematic Visc. (cP)	
	Measured ^a	Literature ⁷	Measured ^a	Literature ⁷
ethanol	0.7725 (7)	0.7720	0.828 (10)	0.826
isopropanol	0.7685 (9)	0.7686	1.342 (4)	1.327
n-butanol	0.7948 (8)	0.7954	1.791 (20)	1.778

^aMeasured values are the average of six measurements by different operators using different volumetric flasks and viscometers. The number in parentheses is the standard deviation of the last digit(s). See Reference 1 for further statistical analysis.

Viscosity of Salt Solutions

Salt solutions similar in composition to those expected in the ITP cold runs were prepared and the kinematic viscosities measured. There were no unusual problems or changes in random error compared to water or the standard compounds.¹ The compositions of the salt solutions are listed in Table V and the measured viscosities are listed in Table VI. The compositions of the salt solutions used in this study are similar to the "high hydroxide" composition reported previously.⁸ The current and previous measurements agree within 5% when the previous work is extrapolated to 40°C.

Calibration of Volumetric Flasks

Five 25-mL volumetric flasks were calibrated at 40°C using water. All were Class B flasks per ASTM standards E288, E542, and E694, with a tolerance of ± 0.06 mL at 20°C. The density of water at 40°C was assumed to be 0.9922 g/mL.⁷ The five flasks were calibrated twelve times each. The calculated volumes at 40°C were all 25.00 mL within ± 0.05 mL. The volumes were subsequently assumed to be 25.00 mL at 40°C and no corrections based on the calibrations were applied.

TABLE V. Composition of Simulated Salt Solution

<u>Component</u>	<u>Concentration (molar)</u>		
	<u>SS#92-16</u>	<u>SS#92-17</u>	<u>SS#92-18</u>
NaOH	3.3	1.65	0.33
NaAlO ₂	0.40	0.20	0.040
NaNO ₃	1.4	0.70	0.14
NaNO ₂	0.40	0.20	0.040
Na ₂ CO ₃	0.030	0.015	0.0030
Na ₂ SO ₄	0.040	0.020	0.0040

TABLE VI. Density and Kinematic Viscosity of Salt Solutions (40°C)

<u>Solution</u>	<u>Density^a</u>	<u>Kinematic Visc.^a</u>
	<u>(g/mL)</u>	<u>(cP)</u>
SS#92-16	1.229(2)	2.101(10)
SS#92-17	1.118(1)	1.126(5)
SS#92-18	1.019(1)	0.722(3)

^aAverage of 9 measurements. The number in parentheses is the standard deviation of the last digit.

Quality Assurance

This work was performed according to the following program plan: D. D. Walker, "Viscosity Measurements Program Plan (U)," SRT-LWP-92-026, September 29, 1992. The data are recorded in Laboratory Notebooks WSRC-NB-92-132, pp 154-156, WSRC-NB-91-206, pp 36-119, and WSRC-NB-92-381, pp 55-56. All temperatures were measured with Category 1 calibrated thermometers (WP-560 or WP-830). Efflux times were measured using a Category 1 calibrated stopwatch (WP-559). All weighings were made on analytical balances (IW-0003 and IW-0004) which were checked daily with Category 1 calibrated weight sets (WP-800 or WP-823).

References

1. T. B. Edwards, "Statistical Support for Viscosity Study (U)," SCS-ASG-93008, January 22, 1993.
2. D. D. Walker, "Viscosity Measurements Program Plan (U)," SRT-LWP-92-026, September 29, 1992.
3. a. ASTM Method D445-88, "Standard Test Method for Kinematic Viscosity of Transparent and Opaque Liquids (and the Calculation of Dynamic Viscosity)."
b. ASTM Method D446-89a, "Standard Specifications and Operating Instructions for Glass Capillary Kinematic Viscometers."
4. D. D. Walker, "Measuring the Viscosity of Transparent Liquids," Manual #L12.1, Interim Waste Technology Procedure #IWT-OP-067, Revision 0, dated December 9, 1992.
5. ASTM Method D2162-86, "Standard Test Method for Basic Calibration of Master Viscometers and Viscosity Oil Standards."
6. *CRC Handbook of Chemistry and Physics*, 61st Edition, Robert C. Weast, Editor, CRC Press, Boca Raton, 1980, p F-51.
7. Clarence J. West, *International Critical Tables of Numerical Data, Physics, Chemistry, and Technology*, McGraw-Hill, New York, 1933.
8. D. D. Walker and G. K. Georgeton, "Viscosity and Density of Simulated Salt Solutions (U)," WSRC-RP-1088, October 19, 1989.

END

DATE
FILMED

5 / 7 / 93