**FINAL REPORT TO THE OFFICE OF BASIC ENERGY SCIENCES
U. S. DEPARTMENT OF ENERGY
Office of Energy Research, Chemical Sciences Division**

REPORTING INVESTIGATOR: Horacio A. Mottola
Department of Chemistry
Oklahoma State University
Stillwater, OK 74078-0447

DOE/ER/13346--T2

DE93 013225

REPORTING PERIOD: March 1, 1985 to February 28, 1991.

[This is a report of work performed under the support of grant DE-FG05-85ER13346,
DOE Reporting Classification Code KC-03-02-02-0]

GENERAL TITLE OF RESEARCH: "Unsegmented Continuous-Flow Sample
Processing and Electrochemical Detection and Determination of Gaseous Species".

CONTENTS

- 1. List of Publications Generated by Work under Grant
DE-FG05-85ER13346.**
- 2. Narrative Describing the Contributions to Chemical Analysis that
Emanated from Research Supported by Grant DE-FG05-85ER13346.**
- 3. Concluding Remarks.**

DISCLAIMER

This report was prepared as an account of work sponsored by an agency of the United States Government. Neither the United States Government nor any agency thereof, nor any of their employees, makes any warranty, express or implied, or assumes any legal liability or responsibility for the accuracy, completeness, or usefulness of any information, apparatus, product, or process disclosed, or represents that its use would not infringe privately owned rights. Reference herein to any specific commercial product, process, or service by trade name, trademark, manufacturer, or otherwise does not necessarily constitute or imply its endorsement, recommendation, or favoring by the United States Government or any agency thereof. The views and opinions of authors expressed herein do not necessarily state or reflect those of the United States Government or any agency thereof.

MASTER

DISTRIBUTION OF THIS DOCUMENT IS UNLIMITED

**1. LIST OF PUBLICATIONS GENERATED BY WORK UNDER GRANT
DE-FG05-85ER133446.**

1.A. Published Papers in Which DOE Support is Acknowledged

1. M. D. Luque de Castro, M. Valcarcel, F. N. Albahadily, and H. A. Mottola, "Electrochemical Behavior of Iron-1,10-Phenanthroline Complexes at Carbon Paste Electrodes", *J. Electroanal. Chem.* **1987**, 219, 139.
2. F. W. Nyasulu and H. A. Mottola, "Amperometric Determination of Nitrogen Dioxide in Air Samples by Flow Injection and Reaction at a Gas-Liquid Interface", *J. Autom. Chem.* **1987**, 9, 46.
3. F. N. Albahadily and H. A. Mottola, "Improved Response of Carbon-Paste Electrodes for Electrochemical Detection in Flow Systems by Pretreatment with Surfactants", *Anal. Chem.* **1987**, 59, 958.
4. H. A. Mottola and H. B. Mark, Jr., "Kinetic Determinations and Some Kinetic Aspects of Analytical Chemistry" (1988 Biennial Fundamental Reviews), *Anal. Chem.* **1988**, 60, 181R.
5. F. W. M. Nyasulu and H. A. Mottola, "Electrochemical Behavior of 5-Amino-1,10-phenanthroline and Oxidative Electropolymerization of tris[5-Amino-1,10-Phenanthroline]iron(II)", *J. Electroanal. Chem.* **1988**, 239, 175.
6. M. Bonakdar, J. Yu, and H. A. Mottola, "Continuous-Flow Performance of Electrodes Modified with Immobilized Fe(II)/Fe(III) Centers: Amperometric Response to N₂O(g), NO(g), and NO₂(g)", *Talanta* **1989**, 36, 219.
7. C. Hynes, M. Bonakdar, and H. A. Mottola, "Carbon Paste Electrode Chemically Modified by Direct Admixing of tris[4,7-Diphenyl-1,10-Phenanthroline]iron(II)", *Electroanalysis* **1989**, 1, 155.
8. M. Bonakdar, J. L. Vilchez, and H. A. Mottola, "Bioamperometric Sensors for Phenol Based on Carbon Paste Electrodes" *J. Electroanal. Chem.* **1989**, 266, 47.
9. M. Bonakdar and H. A. Mottola, "Electrocatalysis at Chemically Modified Electrodes: Detection /Determination of Redox Gaseous Species in Continuous-Flow Systems", *Anal. Chim. Acta* **1989**, 224, 305.
10. P. A. Nader, S. S. Vives, and H. A. Mottola, "Studies with a Sulfite Oxidase-Modified Carbon Paste Electrode for Detection/Determination of Sulfite Ion and SO₂(g) in Continuous-Flow Systems", *J. Electroanal. Chem.* **1990**, 284, 323.
11. H. A. Mottola, "Some Incidental Kinetic Aspects in Analytical Chemistry", *Analyst (London)* **1990**, 115, 679.

12. G. Sun, P. W. Geno, and H. A. Mottola, "Chemical Species Produced in the Reaction Between Ethanediol (Glyoxal) and 5-Amino-1,10-phenanthroline and Their Iron(II) Complexes: Preparation and Some Spectrometric Studies" *Anal. Chim. Acta* **1991**, 242, 233.
13. G. Sun and H. A. Mottola, "Chemical Species Produced in the Reaction Between Ethanediol (Glyoxal) and 5-Amino-1,10-Phenanthroline and Their Iron(II) Complexes", *Anal. Chim. Acta* **1991**, 242, 241.

1.B. Papers in press

1. D. W. Harak and H. A. Mottola, "Electrostatically Immobilized Hexacyanoferrate Ions as Redox Mediators in Biochemical Sensing: Controlled Release and Cyclic Voltammetric Behavior", *Biosensors and Bioelectronics*, in press.
2. P. R. A. Nader, P. I. Ortiz, and H. A. Mottola, "Polymer-Coated Electrode Based on the Electropolymerization of Resole Prepolymer Mixtures", *Anal. Chim. Acta*, in press.

1.C. Papers in preparation

1. P. I. Ortiz, P. R. A. Nader, and H. A. Mottola, "Amperometric Response of Phenolic Materials at a Polymer-Coated Electrode: Application to Detection after HPLC Separation".
2. J. Yu and H. A. Mottola, "Effect of Counter-Ion Diffusion in the Amperometric Detection of Gaseous Species at a Polymer-Coated Electrode".
3. C. J. Hynes and H. A. Mottola, "Surface Tension Effect in Chemically Modified Carbon Paste Electrodes Pretreated with Surfactants: Effect of Precipitating Counterion".
4. P. R. A. Nader and H. A. Mottola, "Semiempirical Series Process Model for Description of Transient Signals Obtained in Unsegmented Continuous-Flow Processing".

2. NARRATIVE DESCRIBING THE CONTRIBUTIONS TO CHEMICAL ANALYSIS THAT EMANATED FROM RESEARCH SUPPORTED BY GRANT DE-FG05-85ER13346.

The overall purpose of the research performed with DOE support under Grant No. DE-FG05-85ER13346 was the development of chemical sensors for detection/determination of mainly gaseous species in continuous-flow systems, their characterization, and their analytical application. The work started as a comprehensive study of chemically modified electrodes based on metal-complex redox couples of the 1,10-

phenanthroline family and similar chelating ligands. These studies were latter expanded to include biochemical sensors based on incorporation of enzymes in the conducting surface.

The basic observations that motivated these studies can be listed as: (a) the iron(II)/iron(III) complexes with ligands containing the 1,10-phenanthroline moiety can undergo oxidation or reduction (chemically or electrochemically) by changing the oxidation state of the central metal ion only without affecting the organic part of the molecule, (b) incorporation on or into conducting surfaces should result in promising sensors for redox species, (c) the chemical reaction of the Fe(II)/Fe(III) centers with $\text{SO}_2(\text{g})$ was known to be rapid enough in homogeneous solution and was expected to occur with most redox gaseous pollutants, and (d) the predominantly laminar flow conditions used in unsegmented sample/reagent(s) processing should be conducive to maintaining electrical contact in the working electrode/reference electrode/counter electrode set-up. These observations provided the incentive to initiate the work by exploring the electrochemistry in solution of the iron complexes of 1,10-phenanthroline itself since limited information was available (M. D. Luque de Castro et al., 1987, Ref. 1 above). The next step was to study in depth three basic ways to produce chemically modified surfaces:

- * direct admixing of relatively insoluble salts of the iron complexes with graphite to obtain chemically modified carbon paste preparations.
- * covalent binding of the complex cations to chemically modified graphite and pasting liquids.
- * oxidative electropolymerization of some derivative complex cations of the same family of ligands.

Work involving covalent binding of 5-amino-1,10-phenanthroline to oxidized graphite surfaces (F. N. Albahadily, Ph.D. Dissertation, Oklahoma State University, May 1988) by treatment of the latter with thionyl chloride to effect acyl chloride functions on the

surface and final reaction with the ligand resulted in rather poor surface coverage and did not produce encouraging electrodes for analytical application. On the other hand, direct admixing of complexes with graphite and oxidative electropolymerization resulted in the development of useful surfaces for detection in continuous-flow systems.

Three publications have resulted from work on direct admixing into carbon paste preparations (Refs. 7, 12, and 13 above). Direct incorporation of tris[4,7-diphenyl-1,10-phenanthroline]iron(II) perchlorate offers low loss of the complex by leaching and produces electrodes useful for continuous detection for a few hours (Ref. 7). With the hope of linking two phenanthroline units via aggregates of a dialdehyde molecule, so as to obtain a flexible tetradentate ligand and highly insoluble metal complexes, the reaction between 5-amino-1,10-phenanthroline and ethanedial (glyoxal) was investigated. A mixture of ligands with molecular weights 339, 371, 480, and 524 u, was obtained. Several properties of the ligands and their iron complexes as well as spectrometric characterization (UV-visible, infrared, and liquid secondary ion mass spectrometry) were reported in Ref. 12. Results of the electrochemical characterization and analytical application in continuous-flow systems were reported in the publication listed as Ref. 13 above. The iron complexes with the mixture of ligands obtained are highly insoluble in most liquid media, and this is of relevance to their application in flow injection procedures and in liquid chromatography. Direct admixing of these complexes to carbon paste preparations results in sensing surfaces useful for amperometric monitoring under continuous-flow operation for a relatively long time (more than 45 days).

Several fundamental aspects of the oxidative electropolymerization of tris[5-amino-1,10-phenanthroline]iron(II) were studied to ascertain the potential use of the electrogenerated polymeric film in continuous-flow sensing (Ref. 5 above, and Yu and Mottola under "Papers in Preparation"). The favorable charge transfer characteristics of the films

encouraged their application for amperometric detection of oxides of nitrogen, $\text{SO}_2(\text{g})$, $\text{H}_2\text{S}(\text{g})$, and $\text{Cl}_2(\text{g})$ (Refs. 6 and 9 above). These polymeric films provide very competitive limits of detection, particularly for $\text{NO}_2(\text{g})$ [about 2 ppb, v/v]. These films are also very stable even if stored in open air; they can be used in continuous-flow procedures for relatively long periods of time without noticeable deterioration (3 months at least), and are reasonably rugged.

Two bioamperometric sensors prepared by direct admixing of enzymes have been developed for the detection of phenol (Ref. 8 above) and $\text{SO}_2(\text{g})$ (ref. 10 above). These two sensors lead to the realization that if a redox mediator is needed for detection (e.g. hexacyanoferrate ions) the implementation of the biosensor is dictated by the role of the mediator. Consequences of this realization are described in a manuscript to be published in *Biosensors and Bioelectronics* (Harak and Mottola, listed under "Papers in Press" above). If the redox mediator participates in a reaction with a product of the enzyme-catalyzed reaction, direct co-immobilization of redox mediator and enzyme can be implemented; this is the case of tyrosinase. If, on the other hand, as sulfite oxidase requires, direct reaction with the enzyme to restore the catalytic cycle is required, the redox mediator needs to be present in the carrier stream of the continuous-flow system. A controlled-release cell was developed for convenient implementation of the latter case. The detection of phenol directed our attention to an attempt to exploit the oxidative electropolymerization of phenol to deposit protective films on glassy carbon. These studies are summarized in one of the manuscripts in press (Abu Nader, Ortiz, and Mottola, listed under "Papers in Press" above). Oxidative electropolymerization of resole prepolymer mixtures yielded films mimicking the surface conditions of activated glassy carbon but protecting the surface from fouling in the detection of phenolic materials (Ortiz, Nader,

and Mottola listed under "Papers in Preparation"). The utility of electrodes coated with a polymeric film from resole mixtures was illustrated with their use in high-performance liquid chromatography (Ortiz, Nader, and Mottola, listed under "Papers in Preparation").

A semiempirical model, based on the assumption that the following series process:

is responsible for the transient signals obtained in continuous-flow detection with the developed sensors, has been explored (Nader and Mottola listed under "Papers in Preparation"). The simple model assumed above was blended with a third rate constant (k_3) describing a dynamic deviation in the earlier stages of the recording of the signal.

3. CONCLUDING REMARKS

The above narrative, in succinct manner, documents and highlights the research performed under Grant DE-FG05-85ER13346. The PI considers it of interest to recall some of the comments of reviewers of the first proposal that led to funding:

Reviewer # 3 said: "The concern that I have is that I fear that it will not prove possible to immobilize iron-orthophenanthroline complex without affecting its ability to undergo rapid and reversible oxidation." The published results of the research performed with DOE support has demonstrated that the Reviewer's concerns were totally unfounded.

Reviewer # 4 said: "I conclude, sadly, in fact, that the PI does not appreciate the complexity, given that naive statements and presumptions are made, and some crucial aspects ignored..." "...the reactions of sample types such as NO_x , SO_2 , etc ...are $> 1e$ and a simple-minded approach to them with $1e$ reagents like ferroin is very likely to fail, i.e., no successful mediation rate.....Secondly, the PI does not appreciate that the carbon paste idea on p 9 will fail, almost certainly, because of low anion mobility in the precipitates..." All these statements have been proven by the work of the PI, his students, and postdoctoral fellows to be totally wrong. The PI is glad and grateful that those in the Office of Basic Energy Sciences of DOE gave the chance to do the research and prove these points.

REPORT TO THE OFFICE OF BASIC ENERGY SCIENCES, U.S DEPARTMENT OF
ENERGY, OFFICE OF ENERGY RESEARCH, CHEMICAL SCIENCES DIVISION

REPORTING INVESTIGATOR: Horacio A. Mottola
Department of Chemistry
Oklahoma State University
Stillwater, OK 74078-0447
Phone: (405) 744-5936
FAX: 405 744 7673

REPORTING PERIOD: March 1, 1985-August 15, 1989

[This is a report of work performed under the support of Grant
No. DE-FG05-85ER13346; DOE Reporting Classification Code
KC-03-02-02-0]

CONTENT

I. REVIEW OF ACCOMPLISHMENTS UNDER GRANT DE-FG05-85ER13346 DURING
REPORTING PERIOD AND WORK PRESENTLY UNDER DEVELOPMENT

II. MANUSCRIPTS SUBMITTED AND ACCEPTED FOR PUBLICATION DURING THE
REPORTING PERIOD

III. ORAL PRESENTATIONS DURING THE FUNDING PERIOD

I.1. REVIEW OF ACCOMPLISHMENTS UNDER GRANT DE-FG05-85ER13346
DURING REPORTING PERIOD.

This is a short overview of accomplishments under DOE support. With the hope of providing a better perspective to evaluate the new research proposed, this report encompasses the four and a half years of support that started on March 1, 1985. The goals of the original proposal ["Unsegmented Continuous-Flow Sample Processing and Electrochemical Detection and Determination of Gaseous Species] can be summarized by the following three statements quoted from its Abstract:

"A continuous-flow, unsegmented, all-gas carrier system and/or a segmented liquid/gas interface system will be assembled for sample introduction and transport to the detection/determination point".

"A regenerable electrode probe based on the redox characteristics of the iron(II) and iron(III) complexes with 1,10-phenanthroline and related ligands will serve as a detection device".

"Amperometric/coulometric current measurements will provide the analyte signal for quantitative purposes".

As can be seen in the narrative that follows, the work has progressed in a systematic manner following the general ideas of the original proposal to lead to systems fulfilling the goals of the three statements listed above.

The initial studies were design to obtain basic information on the electrochemical behavior of the Fe(II)/Fe(III) complexes of 1,10-phenanthroline. This was necessary because available information was insufficient to permit the development of probes with immobilized redox mediators. Results of these studies were published [see Luque de Castro, Valcárcel, Albahadily, and Mottola (1987) in list of manuscripts].

Early in the research, the use of an all-gas or a liquid/gas/liquid system was addressed and it was decided that the liquid/gas/liquid system afforded conveniences absent in the all-gas system; consequently it was adopted for further studies. As part of these studies a paper on NO₂(g) determination was published [see Nyasulu and Mottola (1987) in list of manuscripts]. This work also showed that amperometric detection with unmodified carbon paste electrodes but aided by the Fe(III)/Fe(II) phenanthroline complexes allowed reaching detection limits that, although not as low as desired, were low enough that NO₂(g) could be determined at ppm levels without need for pretrapping or preconcentration steps.

The choice of electrode material is critical when electrochemical detection is incorporated into continuous-flow systems and carbon paste electrodes have been given preference because they provide relatively quiet signals under continuous flow. Chemical modification (by blending of redox mediators into the paste) improves selectivity but many times also produces slower electron transfer. This drawback was observed in carbon paste electrodes modified by incorporation of

tris[1,10-phenanthroline]iron(II)/(III) perchlorate salts. In the belief that part of the problem was the failure of the carbon surface to wet well because of the presence of an oily film generated during the smoothing process, a surfactant pretreatment was explored in conjunction with a study of the impedance problems associated with carbon pastes. This part of the work was published in Analytical Chemistry [see Albahadily and Mottola (1987) in the list of manuscripts]. The surfactant treatment can improve performance of electrochemical detection (e.g. allow a higher sampling rate in certain determinations [1]), but due to surface tension artifacts can damage performance in thin-layer cells. Several aspects of the surfactant treatment effect are being studied [C. J. Hynes, unpublished results].

After these studies a systematic preparation, characterization, and application of chemically modified electrodes based on metal complex redox couples of the 1,10-phenanthroline family and similar ligands was undertaken. Three basic ways of producing modified surfaces are involved in these studies: (a) direct admixing of a relatively insoluble salt (in aqueous media) of a complex cation with carbon pastes, (b) covalent binding of the complex cation to chemically modified graphite and pasting liquids, and (c) oxidative electropolymerization of some derivative complex cations of the same ligand family.

As example of direct admixing to a carbon paste the work with tris[4,7-diphenyl-1,10-phenanthroline]iron(II) can be cited. These studies have been published [see Hynes, Bonakdar, and Mottola (1989) in the list of manuscripts]. An example of

surfaces modified by oxidative electropolymerization is the incorporation of polymeric films of tris[5-amino-1,10-phenanthroline]iron(II) on the surface of glassy carbon electrodes [see Nyasulu and Mottola (1988) in the list of manuscripts]. These two types of electrodes have been evaluated with respect to their response to oxides of nitrogen in air [see Bonakdar, Yu, and Mottola (1989) in the list of manuscripts] and the electrocatalysis taking place in the detection of $\text{NO}_2(\text{g})$, $\text{SO}_2(\text{g})$, $\text{H}_2\text{S}(\text{g})$, and $\text{Cl}_2(\text{g})$ has been reported [see Bonakdar and Mottola (1989) in the list of manuscripts]. The polymer-coated glassy carbon electrode, offering competitive rates of charge transfer across the film and requiring little configurational changes, appears best, particularly for $\text{NO}_2(\text{g})$ detection/determination since it exhibits competitive limits of detection when incorporated to thin-layer cells. Detection of ozone, $\text{O}_3(\text{g})$, however, and not totally unexpectedly was less successful. Although distinct signals were recorded in the 0- to 0.2-V potential window vs. Ag/AgCl as reference, and good reproducibility was observed in the first injections of air containing $\text{O}_3(\text{g})$ in ppm (v/v) concentrations, the response of the polymer-coated glassy carbon electrode deteriorated with repetitive injections. This may well be due to destruction of the polymeric film since ozone is known to attack double bonds. Such destruction was observed when relatively high concentrations of $\text{O}_3(\text{g})$ were injected.

Since the limits of detection for $\text{SO}_2(\text{g})$ have not been as low as desired, the possibility of developing bioelectrochemical sensors incorporating an immobilized enzyme together with a redox

mediator into the carbon paste was explored. On the basis of experience collected in work not directly related to the DOE funding [2], a sensor for phenol was developed [see Bonakdar, Vilchez, and Mottola (1989) in the list of manuscripts] to collect data for the development of a sensor utilizing sulfite oxidase for $\text{SO}_2(\text{g})$ detection/determination.

Direct chemical attachment of the iron complex to the graphite was attempted by a) anchoring 5-amino-1,10-phenanthroline before reacting with iron, and b) reacting 5-amino-1,10-phenanthroline with iron(II) before chemical immobilization. In both cases the graphite was previously heated at $400\text{ }^\circ\text{C}$ in a current of air [which provided $\text{O}_2(\text{g})$] for 24 h with the hope to increase the density of carboxylic acid functions on the surface of the graphite. This enrichment is at the expenses of other oxygen-containing groups on the surface (e.g. aldehydes, ketones, and quinones). The $-\text{COOH}$ groups were then modified with SOCl_2 to give acyl chlorides, which in turn were reacted with the 5-amino group of the 1,10-phenanthroline moiety. Success of the modification process was confirmed by square wave voltammetry. Relatively weak signals, however, were observed; electrodes thus chemically modified cannot compete with the polymer-coated glassy carbon or the carbon paste ones modified by direct admixing. Details of these attempts can be found in the dissertation of a student who has been associated with the DOE project [3].

Because of the successful oxidative electropolymerization of tris[5-amino-1,10-phenanthroline]iron(II) perchlorate, and with the idea to seek different redox voltage windows for amperometric

sensing in continuous-flow, the oxidative electropolymerization of the Co(II) and Cu(II) complexes in acetonitrile was attempted. The process for cobalt is considerably more complicated and less efficient than the oxidative electropolymerization of the iron(II) complex. The literature records a report on the cyclic voltammetric behavior of the electropolymerized Co(II) complex [4]; however, duplication of these observations has been unsuccessful in our hands. Equally unsuccessful were the attempts to electropolymerize the Cu(II) complex.

I.2. WORK CURRENTLY UNDER DEVELOPMENT

I.2.a. Preparation of New Ligands of the 1,10-Phenanthroline Family

The intention of preparing modified binders containing the immobilized Fe(II)/Fe(III) centers, since earlier attempts with hexadecanoic acid did not meet with success, led to recent efforts motivated by the PI's experience in enzyme immobilization [5]. The products of the reaction between ethanedial (glyoxal) and 5-NH₂-1,10-phenanthroline are being investigated. The initial hope was to obtain gel-type materials (glyoxal in aqueous solution does not actually exist as a monomeric (CHO)₂ unit but as a mixture of higher molecular weight hydrated polymers as confirmed by fast atom bombardment mass spectrometry, the results of which are illustrated in Figure 1). The hope was to use these gels in place of the oily binders conventionally utilized in carbon paste preparations. The different experimental conditions tested so far, however, gave as

FIGURE 1. Fast atom bombardment mass spectrum of a 40% w/v aqueous solution of glyoxal, $(\text{CHO})_2$. Peaks correspond to molecular ions, M^+ . Glycerol used as solvent. Background solvent peaks have been subtracted. The absence of a peak at a m/z value of 59 indicates that monomeric unhydrated glyoxal if present in solution is in very low concentrations.

The peaks suggest the presence of hydrated/dehydrated trimeric, tetrameric, and hexameric species.

final products highly insoluble solids. Both, ligands and complexes (with iron) have been prepared and the materials offered interesting properties for use as redox modifiers incorporated into carbon paste preparations. These properties can be summarized in the following points:

1. Very low solubility in aqueous media; actually it has been difficult to find a suitable solvent for characterization studies. The most successful solvent has been a 1.0 M H_2SO_4 + 1.0 M acetic acid solution in dimethyl sulfoxide. The problem of finding a suitable solvent has made difficult application of size exclusion (gel permeation) chromatography, considered in the hope of obtaining information on molecular weight distribution.

2. Very good complexation characteristics with iron(II), iron(III), and Cu(II) ions. Nickel(II), cobalt(II), cadmium(II), and manganese(II), however, did not seem to complex well or their complexes do not show electrochemical reactivity in the potential window afforded by carbon.

3. Well defined oxidation and reduction peaks for the iron complex when incorporated to a carbon paste (see Figure 2), with electrochemical behavior resembling quasireversibility.

4. Very good surface complexation of iron(II) and/or iron(III) ions by the ligand(s) when incorporated to a carbon paste electrode as evidence by the cyclic voltammetric behavior observed when the electrode is cycled in supporting electrolyte (Figure 2).

5. The iron superficially complexed by the ligand can be removed by dipping in a solution of

FIGURE 2. Cyclic voltammograms for: (A) iron(II) complex of the product of the reaction between glyoxal and 5-amino-1,10-phenanthroline. Scan rate: 100 mV.s^{-1} , supporting electrolyte: $0.10 \text{ M H}_2\text{SO}_4$, 10% w/w complex in carbon paste; (B) chemically modified carbon paste electrode containing 10% w/w of the ligand resulting from the reaction between glyoxal and 5-amino-1,10-phenanthroline after 24-h contact with a 0.050 M FeSO_4 aqueous solution. Scan rate: 100 mV.s^{-1} , supporting electrolyte: $0.10 \text{ M H}_2\text{SO}_4$.

of ethylenediamine-N,N,N',N'-tetraacetate (EDTA) of pH 7.00.

Re-complexing of iron by subsequent dipping in an iron containing solution is successful.

6. When the complex(es) are incorporated to a carbon paste electrode, a successful sensing unit for redox species is obtained (Figure 3 shows the sensor response to sulfite ion concentration). This suggests that the material can be useful for chemical sensors to detect/determine gaseous redox species.

The ligand and the complexes are being characterized by:

1. Elemental analysis
2. Cyclic voltammetry
3. UV-VIS and IR spectrophotometry
4. Thermogravimetric behavior
5. Fast atom bombardment mass spectrometry

Information collected up to this point suggests a mixture of oligomeric units as shown in the mass spectrum of Figure 4. Experimental conditions for the preparation of these materials have been given in the Report of period October 20, 1987 to October 20, 1988. Also in that report can be found thermogravimetric data. Figure 5 shows UV-VIS spectra and Table I summarizes elemental analysis data collected so far.

An increase in interest on metal-organic materials containing the 1,10-phenanthroline moiety has been noted in the literature. Sauvage et al. [6], for instance, have reported the first synthesis of compounds of considerable theoretical interest whose molecules are tied in cloverleaf knots. The 1,10-phenanthroline

FIGURE 3. Hydrodynamic voltammograms: (A) 5% w/w modified carbon paste with the iron complex of the product(s) of the reaction between glyoxal and 5-amino-1,10-phenanthroline, (B) unmodified carbon paste. Supporting electrolyte: 0.10 M phosphate buffer of pH 4.00. Sample injected: 1.00×10^{-5} M Na_2SO_3 .

Linear response to sulfite concentration, at + 0.800 V vs. 3 M Ag/AgCl, was observed between 1.00×10^{-5} and 1.00×10^{-2} M sulfite.

FIGURE 4. Fast atom bombardment mass spectrum of the product of the reaction of glyoxal with 5-amino-1,10-phenanthroline. Solvent: 1% trifluoroacetic acid in 4-nitrobenzyl alcohol. Solvent peaks subtracted.

The peaks are tentatively attributed to oligomeric species with the following assignments of M^+ peaks:

339+1 : $\text{Phen-N=CH-(CH=CH)}_6\text{-COH}$

371+1 : $\text{Phen-N=CH(CO-CH=CH-CO)}_2\text{H}$

480+1 : $\text{Phen-N=CH(CH}_2\text{CH=CHCO)CH=N-Phen}$

524+1 : $\text{Phen-N=CH(CO(CH}_2\text{)}_4\text{CO)CH=N-Phen}$

Phen = 1,10-phenanthroline moiety

TABLE I. Summary of elemental analysis data for the product of the reaction between glyoxal and 5-amino-1,10-phenanthroline in presence and absence of Fe(II).^{a, b}

LIGAND(S) [reaction carried out in absence of Fe(II)]. Percent w/w

CARBON	57.5, 57.69, 57.50
NITROGEN	9.81, 9.84, 9.78
HYDROGEN	3.1, 3.2, 2.78, 2.73, 2.83
OXYGEN	17.79, 17.39, 17.59

COMPLEX(ES) [reaction carried out in presence of Fe(II)]. Percent w/w

CARBON	48.03, 47.92
NITROGEN	8.30 , 8.33
HYDROGEN	3.51 , 3.49
IRON	5.23 , 5.16

^a Carbon, hydrogen, nitrogen, and oxygen determinations by CONOCO Inc. (Ponca City, OK), Multichem Labs. (Lowell, MA), and Guelph Chemical Labs. (Guelph, Ontario, Canada)

^b Iron determination performed by the 1,10-phenanthroline method [E. B. Sandell, "Colorimetric Determination of Trace Metals", 3rd. Ed., Interscience, New York, 1959, pp. 541-542] after organic matter destruction by repeated treatment with refluxing nitric/sulfuric acid mixtures in a Kjeldahl flask.

FIGURE 5. Absorption spectra (0.020 mg/mL in dimethyl sulfoxide 1.0 M in H_2SO_4 and 1.0 M in acetic acid) of: (A) glyoxal, (B) 5-amino-1,10-phenanthroline, (C) product(s) of the reaction between glyoxal and 5-amino-1,10-phenanthroline, (D) iron(II) complex of the product(s) of the reaction between glyoxal and 5-amino-1,10-phenanthroline.

moiety plays a key role in the synthetic model. The use of synthetic complexes to determine the principles governing site-specific recognition of deoxyribonucleic acid (DNA) have been highlighted recently [7]; this work focuses on complexes of the 1,10-phenanthroline family of ligands [8]. Although the goals of these studies differ from ours, the enhanced relevance of complexes of the 1,10-phenanthroline family of ligands provides additional incentive to our efforts. When these studies are completed a paper describing synthesis, characterization, and application of the prepared materials will be submitted for publication in a refereed journal.

I.2.b. Studies with Glassy Carbon Surfaces Coated with a Polymeric Film Prepared from Monomeric Units of Tris[5-amino-1,10-phenanthroline]iron(II)

The effects of counterion mobility on the cyclic voltammetric and hydrodynamic amperometric response [to $\text{NO}_2(\text{g})$] at the polymer-coated glassy carbon surfaces have been described in the report covering the October 20, 1987 to October 20, 1988 period. Additional work not reported yet has been directed to a) entrapment in the polymer film of an enzyme molecule during the electropolymerization process, b) characterization of the surfaces by scanning electron microscopy (SEM), c) growing films suitable for X-ray studies, and d) testing of amperometric detection of gaseous phenolic species.

Attempts to entrap the enzyme tyrosinase [polyphenol oxidase, EC 1.10.3.1] during the electropolymerization process failed to

provide a surface with enhanced response for phenol. Tyrosinase is one of the enzymes reported as retaining activity in nonaqueous solvents and has been used in chloroform for phenol detection [9]; consequently it was chosen as a model enzyme for the entrapment in the acetonitrile medium used for the electropolymerization process. The polymeric tris[5-amino-1,10-phenanthroline]iron(II) film by itself, however, shows a very sensitive response to phenol in aqueous solution (limit of detection close to 6 ng/mL) and this has prompted us to consider the determination of phenol in the gas phase. This work is presently under development, as are the attempts to grow films for X-ray studies.

Surface characterization by SEM is shown in Figure 6.

I.2.c. Studies with a Sulfite Oxidase/Hexacyanoferrate Modified Carbon Paste Electrode .

These studies have been completed in the part that corresponds to optimization of working conditions with sulfite ion in solution; now the work is focused on the detection/determination of $\text{SO}_2(\text{g})$ in air samples. As soon as the SO_2 work is completed a manuscript will be submitted for publication in a refereed journal.

The sulfite oxidase-catalyzed oxidation of sulfite ions in presence of hexacyanoferrate(III) is believed to proceed according to the following sequence of reactions:

CP3 x2500

A

PCP3 x2500

B

FIGURE 6. Scanning electron micrographs (2500X

amplification of (A) wax-based carbon paste surface before depositing the polymeric film, (B) same surface after oxidative electropolymerization of tris[5-amino-1,10-phenanthroline]iron(II) perchlorate.

(C) elemental scanning of the surface showing the chlorine (from ClO_4^-) and iron peaks as predominant.

in which E = sulfite oxidase. Reactions 1 to 3 are more rapid than reactions 4 or 5. Reaction 5 is affected by anions [10] and this poses a problem in amperometric detection in continuous flow because of the need for a supporting electrolyte in the carrier solution. Detection of $\text{SO}_2(\text{g})$ occurs by sulfite formation according to:

with pK_a 's values for the acids involved of about 1.8 and 7.2, respectively [11] and, consequently, at the optimum pH for enzyme activity (between 7.00 and 7.50) the sulfite is present as SO_3^{2-} and HSO_3^- in concentration ratios going from 0.9 to 1.2.

Several chemical parameters affect the chemical steps illustrated above. The hydrogen ion concentration, anion(s)

present in the system and their concentration, redox mediator concentration, and ethylenediamine-N,N,N',N'-tetraacetate concentration (added for preservation of sulfite solutions) can all affect the response to sulfite resulting from the electrical signal generated via:

Because of this, several strategies have been explored. Incorporation of the immobilized enzyme (on crosslinked agarose gel activated with 1,1'-carbonyldiimidazole) and immobilized redox mediator [on poly(4-vinylpyridine)] directly into the carbon paste did not produce satisfactory sensors. The problem associated with this approach was a continuous decrease of signal height with repetitive injections. This is due to a slow "bleeding" of the hexacyanoferrate mediator because of the relatively high pH which the enzyme operation requires. Since the hexacyanoferrate ions can be successfully immobilized on anion exchange resins [11] capable of operating in the entire 1 to 14 pH range, Dowex 1-X8 resin was reduced to powder and saturated with hexacyanoferrate(III) and then mixed with immobilized enzyme in a carbon paste preparation. No bleeding of mediator was observed and the electrochemistry of the Fe(II)/Fe(III) couple was clearly observed by cyclic voltammetry. The resulting sensor, however, failed to show any response to sulfite. It is apparent that for the enzyme catalysis be effective, the hexacyanoferrate must be in intimate contact with the sulfite and enzyme active site (relevance of reactions in eqs 3 and 5) and immobilization

of the hexacyanoferrate does not seem to provide an advantageous situation for chemical encounter.

Because of these observations two different strategies have been studied in detail:

(1) incorporation of the hexacyanoferrate in the carrier supporting electrolyte solution,

(2) incorporation of immobilized hexacyanoferrate(III) [on poly(4-vinylpyridine)] in the form of a solid pre-reservoir able to release at a constant rate the redox mediator (Figure 7 graphically illustrates this strategy).

Both these systems are now being evaluated for $\text{SO}_2(\text{g})$ detection/determination.

ACKNOWLEDGEMENT

The following individuals have been and/or are associated with the project and have been supported with funds from the DOE grant: F.W. Nyasulu, M. Bonakdar, and Paul A. Mader (postdoctoral research associates), F. N. Albahadily, C. J. Hynes, Jianbo Yu, Sun Gang (research assistants, graduate students working for the Ph.D degree in chemistry). Several visiting scholars have also been incidentally associated with the work reported here but received no financial support from the DOE grant: Drs. M. Valcarcel, M. D. Luque de Castro (University of Cordoba, Spain), Dr. Jose Luis Vilchez (University of Granada, Spain), and Dr. Salvador Sagrado Vives (University of Valencia, Spain).

FIGURE 7. Schematic illustration of the "bleeding" cell components.

a. Flow in; b, flow out; b', flow out to reference; c, flow out to auxiliary electrode and eventually waste. b and b' are interconnected by 40 mm of 1-mm i.d. Teflon tubing.

S, Spacer (Teflon, 1 mm in thickness), SR, solid reservoir filled with hecyanoferrate(III) electrostatically immobilized on poly(4-vinylpyridine). Blocks A and B both have a SR unit 10 mm long, 4.5 mm wide and 1 mm deep.

CP, Plain carbon paste electrode used to test background signal; EMCP, enzyme-modified carbon paste (the enzyme is immobilized on crosslinked agarose).

References

1. Rios, A.; Luque de Castro, M. D.; Valcarcel, M.; Mottola, H. A. Anal. Chem. 1987, 59 , 666.
2. Zachariah, K.; Mottola, H. A. Anal. Lett. 1989, 22 , 1145.
3. Albahadily, F. N., "Development and Appliation of Chemically Modified Graphite Electrodes for Use in Continuous-Flow Systems", Ph.D. Dissertation, Oklahoma State University, Stillwater, OK, 1989.
4. De Gregori, I.; Bedioui, F.; Devynch, J. J. Electroanal. Chem. 1987, 238 , 197.
5. Mottola, H. A. Analyst (London) 1987, 112 , 719.
6. Dietrich-Buchecker, C. O.; Sauvage, J.-P. Angew. Chem. Int. Ed. Engl. 1988, 101 , 189.
7. Baum, R. M. Chem. Eng. News 1989, June 12, pp. 22-25.
8. Pyle, A. M.; Long, E. C.; Barton, J.K. J. Am. Chem. Soc. 1989, 111 , 4520.
9. Hall, G. F.; Best, D. J.; Turner, A. P. F. Anal. Chim. Acta 1988, 213 , 112.
10. Cohen, H. J.; Fridovich, I. J. Biol. Chem. 1971, 246 , 359.
11. Bonakdar, M.; Vilchez, J. L.; Mottola, H. A. J. Electroanal. Chem. , in press.

II. MANUSCRIPTS SUBMITTED AND ACCEPTED FOR PUBLICATION DURING THE REPORTING PERIOD

Copy of manuscripts generated by the work supported by grant DE-FG05-85ER13346 published in refereed journals, and in which the DOE support is acknowledged, follow. These manuscripts comprise:

1. Luque de Castro, M.D., Valcarcel, M.; Albahadily, F. N.; Mottola, H. A. "Electrochemical Behavior of Iron-1,10-phenanthroline Complexes at a Carbon Paste Electrode", J. Electroanal. Chem. 1987, 219 , 139-151.
2. Nyasulu, F. W.; Mottola, H. A., "Amperometric Determination of Nitrogen Dioxide in Air Samples by Flow Injection and Reaction at a Gas-Liquid Interface", J. Autom. Chem. 1987, 9 , 46-49.
3. Albahadily, F. N.; Mottola, H. A., "Improved Response of Carbon-Paste Electrodes for Electrochemical Detection in Flow Systems by Pretreatment with Surfactants", Anal. Chem. 1987, 59 , 958-962.
4. Nyasulu, F. W. M.; Mottola, H. A., "Electrochemical Behavior of 5-Amino-1,10-Phenanthroline and Oxidative Electropolymerization of Tris[5-Amino-1,10-Phenanthroline]iron(II)", J. Electroanal. Chem. 1988, 239 , 175-186.
5. Hynes, C. J.; Bonakdar, M.; Mottola, H. A., "Carbon Paste Electrode Chemically Modified by Direct Admixing of Tris[4,7-Diphenyl-1,10-Phenanthroline]iron(II)", Electroanalysis 1989, 1 , 155-160.
6. Bonakdar, M.; Yu, J.; Mottola, H. A., "Continuous-Flow Performance of Carbon Electrodes Modified with Immobilized Fe(II)/Fe(III) Centers. Amperometric Response to N₂O, NO, and NO₂", Talanta 1989, 36 , 219-225.
7. Bonakdar, M.; Mottola, H. A., "Electrocatalysis at Chemically Modified Electrodes: Detection/Determination of Redox Gaseous Species in Continuous-Flow Systems", Anal. Chim. Acta , in press.
8. Bonakdar, M.; Vilchez, J. L.; Mottola, H. A., "Bioamperometric Sensors for Phenol Based on Carbon Paste Electrodes", J. Electroanal. Chem. , in press.
9. Mottola, H. A.; Perez-Bendito, D.; Mark, H. B., Jr., "Kinetic Determinations and Some Kinetic Aspects of Analytical Chemistry", Anal. Chem. 1988, 60 , 181R-200R. [Although this is not a report of research work but a review, the DOE support allowed the

PI's involvement in its preparation and due acknowledgement is given].

END

**DATE
FILMED**

6 / 8 / 93