DOE/CE/15361--T1

RUTGERS

Department of Mechanical and Aerospace Engineering

DOE/CE/15361--T1

DE93 005056

Final Report for

DOE Grant No. DE-FG01-89CE15361 entitled

Accurate Measure by Weight of Liquids in Industry

Submitted by

Professor Michael R. Muller

REC'D OSTI
DEC 21 1992

DISCLAIMER

This report was prepared as an account of work sponsored by an agency of the United States Government. Neither the United States Government nor any agency thereof, nor any of their employees, makes any warranty, express or implied, or assumes any legal liability or responsibility for the accuracy, completeness, or usefulness of any information, apparatus, product, or process disclosed, or represents that its use would not infringe privately owned rights. Reference herein to any specific commercial product, process, or service by trade name, trademark, manufacturer, or otherwise does not necessarily constitute or imply its endorsement, recommendation, or favoring by the United States Government or any agency thereof. The views and opinions of authors expressed herein do not necessarily state or reflect those of the United States Government or any agency thereof.

December 12, 1992

MASTER

DISTRIBUTION OF THIS DOCUMENT IS UNLIMITED

Table of Contents

1.0 Introduction	1
2.0 Literature Review	1
2.1 Basic techniques currently employed in industry.	1
2.1.1 Volumetric Designs.....	1
2.1.1.1 Piston Packed Metering Pump	2
2.1.1.2 Packed Plunger Metering Pump.....	3
2.1.1.3 Diaphragm Metering Pumps	3
2.1.2 Gravimetric Designs	4
2.1.2.1 Loss-in-Weight Systems	4
2.1.2.2 Developments in Gravimetric Metering	5
2.1.2.3 Advantages of Gravimetrics.....	5
2.2 Competitive Position of "Horak" Device	6
3.0 Design of Apparatus (see Figure 1 and Figure 2)	7
3.1 Dispensing Tanks	9
3.3 AC Solenoid Valves	11
3.4 Temperature Monitoring	11
3.5 Temperature Compensating Plunger System	11
3.6 Receiving Bucket/Weighing Scale.....	11

3.7 Fail-Safe Measures	12
3.8 Computer interface and electronics.....	12
4.0 System Operation for General Apparatus	12
5.0 Progress Timeline	14
6.0 Results	19
7.0 Error Analysis	23
7.1 Level Sensors	
1. aeration	23
2. liquid temperature	24
3. surface turbulence	24
4. tank dimensions.....	24
5. calibration.....	25
7.2 Control Valves	25
1. valve seat wear	25
2. normally closed position	25
7.3 RTD's	25
1. long time constant	25
2. calibration.....	26
7.4 Displacement Plunger	26
1. time needed for displacement	26
7.5 Water	26

1. aeration.....	26
2. condensation.....	26
3. expansion of tubing	27
7.6 Computer	27
1. delay time	27
8. Conclusion	28
9. Control Program.....	30
10. References	67

Table of Figures

Figure 1	Original Horak Design.....	11
Figure 2	Prototype Design	12
Figure 3	Dispensing Tank Side View	12
Figure 4	Dispensing Tank Front View	13
Figure 5	Standard Deviation of Baseline Data Tank A	27
Figure 6	Standard Deviation of Baseline Data Tank B.....	27
Figure 7	Standard Deviation of Dispensed Weight	29
	Tank A (with temperature compensation)	29
Figure 8	Standard Deviation of Dispensed Weight	29
	Tank B (without temperature compensation).....	29
Figure 9	Increase In Random Error Due to Change.....	31
	In Water Temperature	31
Figure 10	Main control panel	43
Figure 11	Stripped Main Panel	44
Figure 12	Main Sequence frame 0	45
Figure 13	Main Sequence frame1 subsequence 0.....	46
Figure 14	Main Sequence frame1 subsequence 1	47
Figure 15	Main Sequence frame1 subsequence 2.....	48

Figure 16 Main Sequence frame1 subsequence 3	49
Figure 17 Main Sequence frame1 subsequence 4	50
Figure 18 Main Sequence frame1 subsequence 5	51
Figure 19 Main Sequence frame1 subsequence 6	52
Figure 20 Main Sequence frame1 subsequence 7	53
Figure 21 Main Sequence frame1 subsequence 8	54
Figure 22 Main Sequence frame1 subsequence 9	55
Figure 23 Main Sequence frame2	56
Figure 24 Main Sequence frame3 subsequence 0	57
Figure 25 Main Sequence frame3 subsequence 1	58
Figure 26 Main Sequence frame3 subsequence 2	59
Figure 27 Main Sequence frame3 subsequence 3	60
Figure 28 Main Sequence frame3 subsequence 4	61
Figure 29 Main Sequence frame3 subsequence 5	62
Figure 30 Main Sequence frame3 subsequence 6	63
Figure 31 Main Sequence frame3 subsequence 7	64
Figure 32 Main Sequence frame3 subsequence 8	65
Figure 33 Main Sequence frame3 subsequence 9	66
Figure 34 Main Sequence frame3 subsequence 10	67
Figure 35 Main Sequence frame3 subsequence 11	68
Figure 36 Main Sequence frame3 subsequence 12	69

Figure 37 Main Sequence frame3 subsequence 13	70
Figure 38 Main Sequence frame3 subsequence 14	71
Figure 39 Main Sequence frame3 subsequence 15	72
Figure 40 Main Sequence frame3 subsequence 16	73
Figure 41 Main Sequence frame3 subsequence 17	74
Figure 42 Main Sequence frame3 subsequence 18	75
Figure 43 Main Sequence frame3 subsequence 19	76
Figure 44 Main Sequence frame 3 subsequence 20	77
Figure 45 Main Sequence frame 4	78

1.0 Introduction

Rutgers University and the Project Director (Professor Muller) received authorization to proceed with this research project on March 16, 1989.

Currently, most liquids are sold by weight rather than volume. Most industries now use metered pumps to fill their containers with fluid, this uses a representative volume for the required weight. Because fluid density is a function of temperature, it is impossible to specify an accurate volume when temperature varies, and usually more liquid is put in the container than is necessary to avoid underselling.

This research's focus was to build a prototype of a computerized liquid dispensing system, designed by Vladimir Horak. This liquid metering system is based on the concept of altering the representative volume to account for temperature changes in the liquid to be dispensed. This is actualized by using a measuring tank and a temperature compensating displacement plunger. By constantly monitoring the temperature of the liquid, the plunger can be used to increase or decrease the specified volume to more accurately dispense liquid with a specified mass.

In order to put the device being developed into proper engineering perspective, an extensive literature review was undertaken on all areas of industrial metering of liquids with an emphasis on gravimetric methods.

Over the course of the project several students were involved whose contribution to the work was substantial. These include:

Norma Cernadas
Philip Debrito
Ann Marie Holden
Peter P. Polomski

The project director wishes to thank them all for their hard work and enthusiastic assistance.

2.0 Literature Review

2.1 Basic techniques currently employed in industry

Techniques currently employed in industry to meter liquids make use of one of two basic measuring principles: volumetrics and gravimetrics. These two categories are differentiated by the units of quantity in which they deliver controlled flow rates of the metered fluid, volume or mass.

2.1.1 Volumetric Designs

Volumetric systems which dispense a required quantity from a controlled flow of liquid that are most commonly employed in industry are the positive displacement liquid pumps. There are various models presently marketed that proportion and monitor liquid flow delivery; however, their design does not emphasize accurate generation of desired volumes of fluids. These devices will not be addressed in detail since they offer no basis of comparison with the gravimetric units.

Positive displacement pumps divide the flow of liquids into distinct known volumes which are based on the physical dimensions of the meter (Liptak, 1982). Any positive displacement pump can be used to meter fluids due to its volumetric transfer mode. In practice, however, only those which have minimal internal and/or external leakage can provide the precision and accuracy required in a metering operation. Reciprocating metering pumps represent the most efficient and versatile example of this class.

Reciprocating metering pumps are accurate to within $\pm 1\%$ of setpoint and discharge a definite amount of liquid with each stroke of a piston or plunger. They are used for positive control of liquid flows ranging from 1 ml/hr to 20 gpm and are used against (or to generate) pressures as high as 30,000 psi (Poynton, 1983). Their capacity is a function of the diameter of the plunger, the effective length of the plunger stroke and the rate or speed of stroking

(Poynton, 1983). The power end used to generate the reciprocating movement can be a drive source (i.e. electric motor) or electromagnetic and pneumatic power ends which create motion using electromagnets and pneumatic pistons, respectively. Subdivisions based upon the configuration of the displacement chamber or liquid end separate these metering pumps into three basic types: piston packed seal, plunger packed seal, and diaphragm seal, mechanical and hydraulic inclusive.

2.1.1.1 Piston Packed Metering Pump

The piston packed pump operates using the piston as the measuring component of the pump which is designed to displace a known volume of fluid as it reciprocates within the pump. The forward travel of the piston reduces the internal volume of the liquid chamber, displacing the metered liquid out of the discharge valve. The reverse travel decreases the pressure within the liquid chamber by enlarging the internal volume of the chamber. A packing is used to effect a dynamic seal with the inside diameter of the cylinder and a static seal with the outside diameter of the piston (McCabe, Lanckton, & Dwyer, 1984).

The piston packed pump can be designed to be double acting, providing a discharge of fluid into a system on both the forward and backward thrust of the piston. This option allows for capacities ranging from several cubic centimeters per hour to 1200 gph. In spite of their low cost and production efficiency with accuracy's to 1% over 10:1 range, these devices are seriously limited by costly packing leakage and required maintenance when subjected to corrosive or abrasive fluids (McCabe, et al., 1984).

2.1.1.2 Packed Plunger Metering Pump

Operationally similar to the piston pump, the packed plunger pump has its packing installed in a stationary gland inside the cylinder. The reciprocating action of the pump generates a dynamic seal between the outer diameter of the plunger and the inside of the packing, and a static seal is formed between the outside diameter of the packing and the inner diameter of the stuffing box. This mode of operation prevents backflow, ensuring liquid movement from suction port to discharge port, through the displacement chamber, and allows

for the pump to operate with a pressure capability to 50,000 psi and with accuracy's to 1% over 15:1 range (McCabe, et al., 1984).

Under the idealized conditions of a proper fit in the displacement chamber, the plunger displaces a volume of liquid equal to that of the cylinder having the same diameter plunger, and a length corresponding to the stroke distance. James P. Poynton, an authority on reciprocating metering pumps, enumerates the following limiting factors which may cause deviations from this theoretical value:

1. excessive time associated with check valve operations,
2. fluid compression caused by high discharge pressures,
3. entrapped gases,
4. elastic deformation of the pump head,
5. leakage through defective piping and seals.

Additionally, wear produced by the reciprocating plunger decreases the volumetric efficiency of the pump and corrosive or abrasive fluids cannot be metered.

2.1.1.3 Diaphragm Metering Pumps

The diaphragm metering pump actuates a flexible member to transmit a pulsating force to the pumped fluid without allowing the leakage which occurs in the aforementioned reciprocating pumps, and thereby increasing the metering accuracy and eliminating contamination of the process liquids. These pumps are categorized by the type of liquid end, of which the most common design is the disc-diaphragm liquid end. Further breakdown generates a categorization based on the actuation of the diaphragm, mechanical or hydraulic.

The mechanical diaphragm pump operates by employing the same power side and capacity control previously described for the other types of metering pumps. In place of a piston rod or plunger, however, this pump employs a connecting rod fastened to the center of the diaphragm (McCabe, et al., 1984). This design feature provides a relatively low cost method of metering high-viscosity liquids, abrasive slurries and chemicals without the economic and processing disadvantages prevalent in other systems. The large axial movement of the diaphragm limits mechanical diaphragm pumps in pressure bellow 150 psig

(Poynton, 1980) and in capacity (Bryan, 1986). The determined accuracy is within a 5% range and with as much as 10% zero shift with pressure change from minimum to maximum (McCabe, et al., 1984).

The hydraulically balanced diaphragm pump is a hybrid design that incorporates the principle advantages of the previously discussed metering pump types, sharing the principles of power source and capacity control. The dissimilarity arises in the actuation of the diaphragm: the measuring piston or plunger reciprocates within a precisely sized cylinder at an established stroke length, displacing a volume of hydraulic liquid, not the product liquid (McCabe, et al., 1984). An automatic compensation system replaces any hydraulic oil lost during even the slightest leakage and regulates pressure within the hydraulic system. These built-in safety features allow for accuracy's of 1% deviation and capacities from a few cubic centimeters per hour to 1500 gph and pressures to 5000 psi. Aside higher initial costs, the hydraulic metering pump is subjected to a zero shift of 5-10% per 1000 psig and is limited to moderate-viscosity fluids.

2.1.2 Gravimetric Designs

Gravimetric metering designs currently employed in industry utilize a mode of operation ensuring liquid movement from a controlled material flow, through a feeder, to a weighting device. There are various instruments which operate using gravimetrics as the metering principle, most of which were developed to handle solids or materials containing particles. Emphasis is placed on the loss-in-weight feeder, a feeder type representative of the currently most advanced techniques for liquid metering systems, and on their advantages over volumetric systems.

2.1.2.1 Loss-in-Weight Systems

In loss-in-weight systems the weight of material in the hopper is counterbalanced by a retracting poise on a scale beam. The weight sensing section is this material hopper or supply tank which is supported by or suspended from one or more load cells. The weight signal generated by the load cell system is applied as a measured signal to a weight rate controller. Deviations between the weight of the liquid in the tank and the programmed

weight is continuously monitored and the flow rate of the fluid exiting from the hopper is regulated to compensate for any difference (Liptak, 1982).

When the hopper weight achieves the filled status, the differential gap controller terminates the filling process and begins the feeding cycle by transferring control of the material exit flow regulator to the loss-in-weight control system. Liptak (Liptak, 1982) notes that the feeding system is operating on a volumetric rather than gravimetric basis during the filling cycle; hence filling is accomplished as rapidly as possible. The editor adds that it is desirable to design the system such that the filling cycle is a small portion of the total cycle time.

2.1.2.2 Developments in Gravimetric Metering

A summary of developments in gravimetric metering, with special emphasis on loss-in-weight systems, appeared in the October 1986 issue of Plastics Technology. New computer-controlled gravimetric feeders reportedly have long term accuracy better than 0.05%, considerably higher in accuracy than volumetric devices. Microprocessor controls allow for the integration of the feeders with computer or master-control systems, creating processes involving various feeders which can be equipped with modifications, such as agitation mechanisms, without adversely affecting the weighing device that controls feed rates. The feed rate of these systems ranges from 2-1800 lb/hr, a spread that offers industry a versatility not present in volumetric units.

2.1.2.3 Advantages of Gravimetrics

Metering systems have undergone significant changes over the last several years with the most significant improvements and innovations in the area of gravimetric devices. Improvements in mechanical and electronic reliability, combined with inherent accuracy advantages, have increased the popularity of gravimetrics. These techniques offer greater reliability than their volumetric counterparts which require extrapolating mass flow measurements from volumetric flow, a complicated process necessitating the application of various corrections based on parameters related to density, such as pressure, temperature and specific gravity (Foust, Wenzel, Clump, Maus, & Anderson, 1980). These measurement and calculation steps significantly increase the margin of potential error.

The trend to gravimetric meters follows the emphasis on quality control. Gravimetric metering systems provide industry with a greater degree of accuracy that can make a substantial difference in the percentage of material lost in a cycle of operation and the quantity of dollars saved, factors which can justify the added expense of such units (Warren, 1988). Supplementing volumetric equipment with accessories designed to minimize the meter inaccuracies can nearly double the cost of the system. This increase, although it may not surpass the expense of the bulkier mass-based unit, would still not produce a device matching gravimetric efficiency. In fact, the economic factors which have placed gravimetrics out of the general reach of industry are being controlled as the advent of new technologies in this area generates a more competitive atmosphere.

2.2 Competitive Position of "Horak" Device

A prototype instrument for the computerized measurement of liquid quantities by weight was developed using the patented technologies of Mr. Vladimir Horak. This apparatus was designed to provide highly accurate mass metering with a target accuracy of $\pm 0.05\%$. The competitiveness of the "Horak" design is established by theoretical advantages over currently employed volumetric and gravimetric liquid metering devices.

The Horak device, which can be referred to as a static metering pump, allowed for the metering and dispensing of liquid quantities without problems inherent in positive displacement pumps. Since the apparatus was independent of input conditions of the liquid, it was not affected by inaccuracies due to fluctuations in operating speed caused by input power fluctuations or by variations in pressure of the liquid input line. Leakage due to wear of components and seal failure that can compromise safety and efficiency, particularly with the handling of hazardous liquids and fumes, would be eliminated.

The plunger displacement of a positive displacement pump is typically small in comparison with the volume of fluid metered in a given dose. This results in the requirement of a relatively large number of cycles in order to develop the specified dose. The inaccuracies present in a cycle is cumulative through the process resulting in appreciable error in the volume delivered. The Horak device avoids this disadvantage by metering a

volume of liquid which is independent of a previously batched volume and has been compensated for any deviation in density.

Computer automation, a design feature of the Horak device, provides an increase in quality control. This allows for communication into remote computers for inventory control and automatic generation of statistical process control records. Microprocessors on volumetric devices can provide these same advantages, but the accuracy of such data may still be less reliable because of its inability to adjust for density variation (Warren, 1988).

Disadvantages present in the apparatus become apparent with a comparison to gravimetric units. The Horak device adjusts the internal capacity of a volumetric liquid dispensing cavity to compensate for a decrease in density and expansion of liquid at ambient temperatures above the temperature of maximum density. This increases the margin of potential error because there is a need to measure and compensate for the affects on the parameters related to density.

3.0 Design of Apparatus

Figure 1 Original Horak Design

Figure 1 is the original design of the Horak system. The prototype (see Figure 2) does not incorporate the high pressure air supply because of the insignificant density change of water due to pressure and the inability of the acrylic tanks to withstand high pressures. Bi-conical plungers were replaced with cylindrical plungers for ease of programming and setup. Bottom mounted RTD's replaced the above design in order to obtain temperature values at center of water mass to avoid any ambient temperature affects. Separate water supply cutoff ball valves were substituted for the single cut-off valve. The bottom draining design for the receiving tank was abandoned because of the contact with the scale which would cause false

weight reading due to vibrations caused by the Solenoid valve opening and closing-- a siphon was substituted.

Figure 2. Prototype Design

Figure 3. Dispensing Tank Side View

Figure 4. Dispensing Tank Front View

The basic system consists of two tanks for liquid measurement and a bucket to collect liquid and check the weight. The tanks were made of 0.25 thick clear plastic. Because the tanks were clear, system checks were easily made. Each tank had a diameter of 12 inches

and a height of 18.5 inches. The tanks were mounted side by side on a table above the draining bucket. All drainage was done by gravity.

3.2 Level Sensing

Mounted in the walls of each tank were two Sensall Ultrasonic level sensors, a upper level and lower level sensor. Each level sensor of 1/32 inches and was separated by a vertical distance of approximately 12 inches.

3.3 AC Solenoid Valves

For each dispensing tank there was a drainage valve and a filling valve, both were AC controlled solenoid valves. The seals of the valves were metal on metal and needed to be checked periodically for any pitting or particles which prevented a proper seal. A fifth solenoid valve was present in for drainage of the receiving tank. An in-line dirt and rust filter was installed to remove particles larger than 5 microns from the water before it reached the solenoid valves.

3.4 Temperature Monitoring

Mounted through the bottom of each tank was an RTD temperature sensor. Because the liquid entered the tank at approximately 2.0 inches above the bottom of the receiving tank, the RTD temperature sensors had the earliest possible indication of temperature change. Each RTD was installed in a heavy metal casing which limited the response time of the sensor, but because the RTD's were mounted low, accurate readings were possible in the final prototype.

3.5 Temperature Compensating Plunger System

Mounted into the top of each tank was the temperature compensating plunger system, comprised of a simple design. A stepper motor mounted outside the tank, drove a cylindrical plunger via a threaded rod threaded through the lid of each receiving tank. The plunger was approximately 2.75 inches in diameter and 4 inches in height.

3.6 Receiving Bucket/Weighing Scale

The drainage bucket was located below and in front of the two tank dispensing tanks. The bucket was set on top of Sartorius scale used to measure the mass of the dispensed liquid. The scale had a maximum capacity of 29000 grams and an accuracy of ± 0.2 grams. Water was dispensed into the bucket via 1/2 I.D. Tygon™ tubing; water was emptied from the receiving bucket using siphoning.

3.7 Fail-Safe Measures

As a fail-safe measure all equipment is wired to a ground-fault detector (GFD) circuit with a sensitivity of 0.005 amps. GFD sensors were mounted at the top of dispensing tanks and receiving bucket. In case of power failure, all valves were normally closed, therefore, all valves would close to prevent overflows.

3.8 Computer interface and electronics

The computing requirements for the various operations involved in dispensing the liquid could have been easily handled by any of a host of available microcomputers. The speed of command execution is much faster than any of the other operations. The selection of

computer for interface with the equipment was determined by ease of use and available graphics. An MacintoshII si was selected.

The various sensors and control of valves were achieved using 16-bit A/D converters and D/A converters. A National Instruments MIO-16X data acquisition board was attached to the computer which had ample interfacing capabilities to handle all of the operations needed to evaluate the prototype.

Auxiliary electronics included operational amplifiers, opto-isolators, and various status lights and indicators.

4.0 System Operation for General Apparatus

A process of sequential electrical signals governed the operation of the metering system, providing an automation necessary which optimized accuracy and efficiency. Software specially developed for this metering device allows for the synchronization of the commands controlled by the personal computer and the electrical board which has the capacity to deliver these signals.

The liquid metered was conducted to an aperture at the lower side of the vessel via tubing which, in turn, was connected to a solenoid valve. This delivered the fluid until there was an attenuation of electricity, placing it in the normally closed status. The fluid was introduced to the container through the lower side, minimizing turbulence that can generate noise for the level sensors, masking the true signal. The piping was necessarily placed below the desired liquid level to allow for the filling of the tanks.

Two lines, with hot and cold adjustments, fed the input tubing to allow for the testing of the metering system with the liquid at various temperatures.

The solenoid valve allowed the flow of the liquid to pass until the upper ultrasonic level sensors relayed to the computer the filled status of the dispensing tank. These level measuring instruments are securely attached to the vessels, eliminating errors incurred by contamination and displacement. The height at which the sensors were secured was predetermined for the mass of the fluid occupying a given volume at conditions taken as the datum.

The temperature of the contained liquid was monitored by a RTD attached to the vessels. These devices continuously and directly transmitted temperature data to the computer which was programmed to signal any variation that would affect the properties of the fluid.

Deviations from the standardized temperature alert the computer to activate the plunger, a fluid displacing device that will allow correction for density change of the liquid. The plunger will move during the filling of the vessel to give the metering of a complete mass of material. A threaded rod is mechanically raised and lowered by a stepper motor driven by computer-controlled power. Information on the volume change of the metered liquid resulting from a certain change in temperature and the rotations of the rod were required to adjust the plunger to account for these fluctuations were inputted to the microprocessor.

The signal from the upper ultrasonic sensors completed the filling cycle and initiated the dispensing process. The fluid flows out the base of the container via tubing which, in turn, is connected to a solenoid valve. This valve terminated the release of the stream when signaled by the lower ultrasonic sensor, completing the first dispensing sequence. The location of the second sensor allowed for the corrected volume of the desired mass.

The complete metering operation involved two vessels to optimize time efficiency: the filling process of one corresponds to the dispensing process of the other.

5.0 Progress Timeline

July 1989 - September 1989

At this stage of development , a research endeavor was underway for prospective suppliers for the Plexiglas to construct the vessel walls and the end caps of each vessel of the dispensing tanks. In the interim, an IBM compatible AST computer was acquired by the University. Preliminary software development was being performed in order to make independent evaluations of accuracy and repeatability of the level sensors, the RTD's and the control valves. Evaluating repeatability and accuracy of each instrument allowed us to determine general uncertainties of the systems accuracy. Once the proper material for the dispensing vessel was obtained, a precise but lengthy machining of the plastic was performed in order minimize the probability of leaks which may increase systematic error during final testing stages.

October 1989 - December 1989

At this stage of development, the dispensing vessels had been constructed and mounted on an elevated platform to promote increased gravitational effects which allow for rapid draining of dispensing vessels and receiving tank. Decreasing turn-around times during testing is desired in that allows for greater number of data points to be obtained in a given time period-- a greater number of data points results in a more precise evaluation of the system's accuracy. Level sensors and RTD's have been mounted in the tanks. Water supply and drainage lines had been attached; receiving tank had been installed. Control valves had been mounted. Custom made computer interface boards, linking the computer to individual components were constructed. Development of the main computer program to control and monitor system operations began.

January 1990 - March 1990

System software development continued. In order to promote increased accuracy in system operation, baseline measurements were taken in order to determine the preferred configuration of the software. These baseline measurements were taken at constant

temperature and constant pressure without volume displacement. After each baseline measurement was taken the software was adjusted accordingly in order to promote the most efficient operation of the system from a software execution standpoint. Decreasing computer reaction times to changing system conditions was important in order to favor high repeatability over a set of acquired data.

April 1990 - June 1990

After extensive software modifications and baseline measurement analysis, it was determined that the current scale was not precise enough to accurately determine the system's repeatability. Measuring the weight of the dispensed liquid is the crucial determinate in evaluating the systems capability of precisely measuring the proper volume of liquid. Inability to accurately and precisely weigh dispensed liquid resulted in a need for a higher accuracy, higher precision scale. A Satorious model IB31000-P was acquired by the University. This scale has an accuracy of one tenth of one gram and a maximum capacity of 31,000 grams. The new scale allowed for a considerable reduction in computer reaction time due to the scale's ability to interface directly with the computer without any intermediate hardware. In order to have direct scale to computer interface, it was necessary to rewrite the entire software application in GWBasic. This resulted in an extensive, additional time period needed for program configuration and debugging,.

July 1990 - September 1990

At this time, application software became functional, and preliminary testing began. Tests were conducted at constant temperature, and constant pressure without volume displacement to obtain baseline figures. Adjustments to software were made accordingly to optimize accuracy of the system. Lengthy turn-around times between data points resulted in extensive periods of time between software and hardware adjustments. Systematic error analysis proved to be prolonged and arduous due to need for the system to be operating for long periods of time in order to acquire enough data points of precisely determine the effect of the previous hardware or software modification.

October 1990 - December 1990

At this stage of development, systematic error was greatly reduced and testing with fluid displacement was ready to begin. Fluid displacement was accomplished by a plastic cylinder mounted atop of each tank via a threaded rod. The threaded rod was threaded through the lid of the tank and attached to a stepper motor. The stepper motor was mounted on vertical shaft attached to the lid of the tank. The threaded rod would be turned a specific amount, raising or lowering the plastic cylinder and the stepper motors a unit. The specific amount in which the cylinder was lowered or raised, represented a specific change in volume of water to be displaced. This displaced volume is determined during individual fillings of the dispensing tank by the temperature of the incoming water filling the tank. The determination of the amount of displacement of volume needed for high repeatability of the mass of dispensed liquid is crucial to success of the Horak system; therefore, proper execution of the determined displacement is crucial.

Due to rotational torques generated by the stepper motors, and relatively high weights of the stepper motors, precision displacement of the fluid required that careful design of stepper motor mountings be thoroughly studied. Several mounting configurations for the stepper motors were tested in order to determine which mounting produced the most precise fluid displacement. After extensive examination the aforementioned design was implemented.

January 1991-March 1991

At this point in development, stepper motor software support was ready for debugging. Several tests were run at a constant temperature to determine the system's ability to displace liquid in such a fashion that implementation of stepper motors would increase the accuracy and precision in which the liquid was dispensed. Extensive software modifications were required to make the stepper motors react quickly enough to changing conditions. Arduous testing was performed before an acceptable repeatability and precision of stepper motor operation was reached.

April 1991 - June 1991

At this point of development, varying temperature testing began. Tests were conducted at temperatures varying from 60 degrees Fahrenheit to 140 degrees Fahrenheit.

Several difficulties in the hardware's ability to adapt to high temperature operation were encountered. Dissolved gas bubbles, present during high temperature operations condensed on the surface of the level sensors, resulting in false triggering causing the system to fail or produce useless data. Several attempts to solve this problem proved futile; it was decided that testing would be restricted to a temperature range of 60 degrees Fahrenheit to 100 degrees Fahrenheit. High liquid temperatures also caused condensation on the upper walls and lid of the dispensing vessels caused erroneous values. This was avoided when temperatures were restricted between 60 degrees Fahrenheit to 100 degrees Fahrenheit.

July 1991 - September 1991

At this point of development, tests were conducted at constant temperatures between 60 degrees Fahrenheit to 100 degrees Fahrenheit. Analysis of systematic error was conducted and corrections were implemented. Several tests were executed in attempt to maximize the accuracy of the system by adjusting control software configurations.

October 1991 - December 1991

At this point of development, it had been established that a maximum accuracy of 0.2% was achievable by the current system at constant liquid temperatures. It was decided that, to understand with more depth, the reasons the system was unable to achieve a higher accuracy, testing at varying temperatures was to be performed. It is desirable that the system be able to react to changing temperatures during operations because any systematic errors or random errors would be amplified; therefore, more easily detected and corrected. Tests were conducted to evaluate the system ability to react to changing temperatures between 60 degrees Fahrenheit and 100 degrees Fahrenheit. Water temperatures were altered to a random temperature between a random number of cycles. After extensive trials, it was determined that the computer (AST with 8 bit digital interface) was unable to process control information with adequate rapidity to adjust displacement of plunger. This was noted by a resolute bi-modal nature in the acquired data. This bi-modal nature occurred in cycles adjacent to points where water temperature was changed. A large differential in temperature between cycles would result in system requiring more than one cycle to adjust accordingly.

It was also determined that relatively slow reaction time by the computer to level sensor triggering caused a significant random error from one cycle to the subsequent cycle.

January 1992 -March 1992

To overcome the long time constant involved with the reaction time of the AST computer, the computer control system was changed to a 16 bit Macintosh based system running LabView2 software. This system is able to process incoming information from devices with much greater speed than the previous AST system. The LabVIEW2 software enables us to graphically view data acquisition in real-time, which allows us to prevent excessive faulty data acquisition, which in turn, increases the number of tests. Excellent progress has been made in the past few months in regard to customizing the LabVIEW2 application for our particular system. Baseline data has been recorded using constant temperature and without displacement. Data was also

..acquisition initiated using displacement in one tank and no displacement in the alternate tank.

April 1992 -July 1992

Further software refinement occurred in order to increase efficiency of computer operation, Final testing was completed using Tank A with temperature compensation and Tank B without temperature compensation.

6.0 Results

The following results are from final accuracy data obtained using a MacintoshII si with a National Instruments MIO-16X digital/analog data acquisition board.

Figure 5. Standard deviation of baseline data Tank A

Figure 6. Standard deviation of baseline data Tank B

Figure 1 and Figure 2 represent the baseline accuracy of Horak prototype at without temperature compensation, at a constant temperature of 65°F for 735 dispensing cycles per tank. Figure 1 shows an overall accuracy of 0.21% resulting in Dispensing Tank A with a 2σ corresponding to 42.84 grams out of approximately 20,000 grams total dispensed mass. Figure 2 show a similar behavior to Tank A; Dispensing Tank B shows an overall accuracy of 0.25% with 2σ corresponding to 50 grams out of a total dispensed volume of approximately 20,000 grams.

Standard Deviation of Dispensed Weight
Tank A (with temperature compensation)

Figure 7 Standard deviation of dispensed weight Tank A (with temperature compensation)

Figure 8 Standard deviation of dispensed weight Tank B (without temperature compensation).

Figure 3 and Figure 4 represent the Horak prototype's ability to displace the correct amount of water to maintain a constant dispensed mass. Seven hundred data dispensing cycles were performed with roughly equal number of cycles of hot water (140°F) and cold water (65°F). Figure 3 shows Tank A, where the temperature compensation mechanism was functioning, has an overall accuracy of 0.21% with a 2σ corresponding to 41.8 grams out of a total dispensed volume of approximately 20,000 grams; we note no bi-modal nature is present because hot water and cold water were dispensed at the same mean mass. Dispensing Tank B, where the temperature compensating mechanism is not functional, shows a clear bi-modal nature. An overall accuracy of 0.41% with 2σ corresponding to 71.14 grams out of a total dispensed volume of approximately 20,000 grams. We note by visual inspection of Figure 3 that high temperature water has a greater standard deviation than that of the cold water distribution. This error is due to the level sensors inability to function accurately at higher temperatures, resulting in a greater standard deviation.

Figure 9 Increase in random error due to change in water temperature.

Figure 5 illustrates the system component's inability to handle high temperatures. Figure 5 represents a sample of data where the system had operated with high temperature water for 140 cycles, then water temperature was changed to cold and operated till the 201 cycle. A dramatic decrease of overall error occurred. Under hot water condition the system functioned with an accuracy of 0.31% as opposed to the later, cold water cycles where error reduced to 0.09%. This remarkable error is due component error and is not inherent in the Horak design.

7.0 Error Analysis

The original design, proposed by Vladimir Horak stated an overall accuracy of 0.05%. This accuracy could not be achieved with this prototype due to inaccuracies in individual components of the system. The following error analysis discusses these inaccuracies.

7.1 Level Sensors:

1. A large source of error in the test system was the level sensor's inability to perform accurately in the presence of aerated water. Hot water inputted into the system was susceptible to aeration due to inherent qualities of the of the building's boiler used to heat the water, the piping and the valves used to transport the water to test system.

The hot water was more likely to suspend small gaseous bubbles during its transport to the test system due to turbulence caused by expansion fittings, partially opened valves, pipe/hose interfaces and filtration. Small gaseous bubbles affected measurement in two crucial ways. First, small gaseous bubbles clung to the level sensors, affecting phase change detection of the piezoelectric crystals (the mechanism within the level sensors that detect the water level and forced the change of state of the level sensors). Second, several small gaseous bubbles (approximately 2 mm in diameter) displaced a certain volume of fluid which results in erroneous level detection. These small gaseous bubbles were caused by turbulent flow within the water filter. Several bubbles would be rising from the water inlet at the bottom of the tank to the surface of the water inside the dispensing tank during the filling stage. Any bubbles rising inside the tank during triggering of the top level sensor, would then rise to the top of the tank **after** the tank has stopped filled. The end result being a reduction in the volume of the water to be dispensed by the certain volume of the bubbles present in the tank at the time of the top level sensor triggering.

It should also be noted that it has been determined that liquid foaming and incorrect mounting may contribute to level sensor error (Groetsch, 1461). These effects, however, were not significant in our prototype.

2. Data analysis has shown significantly lower system accuracy at higher liquid temperatures. This increase in error may be explained by increased water aeration at higher

operating temperatures and heat effects on the level sensor electronic circuitry and the piezoelectric crystals.

3. Surface turbulence due to vibration and tank filling results in additional inaccuracies. Surface waves may cause early triggering of level sensors despite a built-in half second delay. Excessive surface turbulence causes multi-triggering of the level sensor. Tank filling rate was reduced to minimized to reduce surface waves; however, time necessary to complete one filling increased.

4. Dimensions of the tank have another crucial effect on level detection accuracy. Each level sensor has a claimed level detection repeatability $1/32$ inches. Contribution of this error to the total accuracy error of the system varies as function of dispensing tank diameter to height ratio. If the tank has a relatively high diameter to height ratio, the volume of water which contributes to the error

$1/32$ in x cross sectional area of tank

becomes a significantly high percentage of the total tank volume. However, if the tank diameter to height ration is relatively low In consequent designs of the Horak, a balance between dispensing tank dimensions and rate of filling should be taken into account as per application.

5. Calibration difficulties arose when attempting to adjust the level sensor gauges for optimum performance in both hot water and cold water. Signal strength of level sensors (gain to the piezoelectric crystals) needed to be adjusted in order to compensate for aeration of the hot water. When adjustments were made to the gain, tank level at which the level sensor changes state changes, resulting in an different total volume to be dispensed, many early data sets were invalidated due to this difficulty.

7.2 Control Valves

1. Leaking in the solenoid valves caused invalidation of a number of data sets. Leaking occurred due to contamination and wearing of the valve seats over time. When leaking was discovered, it was necessary to remove and disassemble the solenoid valve.
2. An inherent difficulty that contributed to error associated with the solenoid valves is the reliance of the valve's closed position on differential pressure and gravity to hold the plunger closed. During slow filling, differential pressure between upstream and downstream from the valve gradually built up over a few seconds. During the build up pressure to close the valve, fluid was able to pass through the valve until the plunger became fully seated.

7.3 RTD's

1. An inherent complication caused by RTD's is its large time constant. Slow reaction times in temperature detection resulted in a need for adjustments to the system. Initially, it was necessary to install a pause during the filling of the dispensing tank, in order to allow the delayed temperature value to come within a reasonable amount of the true temperature before filling was completed. At later stages of development, this pause was eliminated when the plunger displacement function was incorporated into the system.
2. Difficulty in calibration of RTD's arose due to the nature of the RTD circuitry. The temperature difference between maximum and minimum operating temperature was approximately 75 °F. This relatively small temperature differential resulted in a small voltage range (3.5 Volts) outputted from the RTD amplification circuitry. The maximum resolution of the computer hardware for temperature detection of only ± 1.5 °F without excessive hysteresis.

7.4 Displacement Plunger

1. During initial installation of the displacement mechanism (plunger), it was discovered that excessive time was needed to displace the plunger for a given temperature difference. Initial programming executed displacement during the entire filling process; this often resulted in over filling of the dispensing tank due to displacement occurring during the time when the upper level sensor was to trigger.

Extensive software modifications were made to implement a timer during the first cycle of each test to detect the filling time for the dispensing tank. Once the a filling time was determined, the dispensing tank fills to 80% capacity without enabling plunger displacement. Once the dispensing tank is 80% full, filling is paused while the plunger displaces. After displacing is complete, the system completes filling.

7.5 Water

1. Aeration of hot water caused two significant additions to error. An increase in level sensor hysteresis occurred during hot water tests. Over dispensing error increased from 0.15% for cold water to over 0.30% for hot water dispensing. This was caused by false triggering of the level sensor as an effect of increased gaseous presence in the detection gap. Secondly, gaseous bubbles suspended in the water account a certain displacement of water, which results in an increase in overall error.

2. Water condensation on inside walls and inside of the top cover introduce additional error to the system. Water droplets dripping down from the top and sliding down the inner tank walls during hot water cycles account for a slight increase in system inaccuracy.

3. For ease of system maintenance and repair, Tygon tubing was used for draining and filling feeds to the dispensing tanks and receiving tank. It should be noted that error may be involved in contraction and expansion of tubing during filling and emptying.

7.6 Computer

1. Software support development became a major focus of the Horak System due to the intricate nature of the system's control procedure. Several systems were tested-- final testing was done using a MacintoshII-Si using a National Instruments MIO-16X digital/analog interface board supported by National Instruments LabView2 software.

In order to monitor and install flexibility into the Lab View2 program, several mathematical, communication, and graphical procedures were added to the control software. Many of these procedures were introduced to perform concurrently (i.e. real time monitoring of receiving tank weight and real time monitoring of temperature levels would be graphically represented on the computer screen). Since the a computer can only perform one task at a time, the computer sequentially executes a number of tasks that are programmed to be executed concurrently, forming a task sequence. When large amounts of tasks were programmed to be executed concurrently, the time required to execute one turn-around of this task sequence becomes significantly long relative to total error of the system.

A crucial task performed by the Horak control software was the detection of level sensor status (triggered or not triggered). Efficient execution of this task directly effected the accuracy of the system. If the task sequence containing the task of checking the level sensor status was relatively long, a remarkable random error becomes associated with the level detection. The time required to complete one turn-around of the task sequence was equal to the maximum time needed to detect a change in state in the level sensor; the minimum time required is virtually zero. The error associated with the computer detection of the state of the level sensor was dependent at what point in the task sequence the computer was relative to the point where it checked the level sensor. If the level sensor state changed the moment after the computer checked the state of the level sensor, then a maximum error occurred because the computer had to execute the entire task sequence before it detected the change in state of the level sensor. Conversely, if the computer was at the point in the task sequence just before the level sensor was checked a minimum error occurred. This resulted in a random length delay between the change of state of the level

sensor and the detection of the change of the state of the level sensor by the computer, which, in turn, affected the reaction time of the solenoid valve controlling the filling/dispensing.

This level sensor detection error was significant due to volume of liquid that was associated with this error. Three factors determine the size of random error-- rate of filling/emptying of the dispensing tank, cross-sectional area of dispensing tank and delay time in detection of the level sensor status. Since the delay time was not constant, any volume of water inputted/dispensed during delay time contributes to the overall error of the system

As the rate of filling/emptying of the dispensing tank increased, a greater the volume inputted/dispensed from the tank during the delay time.

$$\text{Volume} = \text{flow rate} * \text{delay time}$$

As we see from the equation: the longer the delay time, the greater the error associated with the input/dispensing accuracy.

The error associated with the delay time also increased as a function of increasing cross sectional area of the dispensing tank. Any volume inputted/dispensed from the system was a function of the diameter² of the dispensing tank.

$$\text{Volume} = \frac{\pi}{4} * \text{diameter}^2$$

Conclusions

The temperature compensating plunger mechanism of the Horak system performed as expected despite overall inaccuracies of the system. Comparing Figure 1 with Figure 3, it is observed that system accuracy is similar at constant temperature and changing temperature with compensation for Tank A. Figure 2 and Figure 4 (Tank B) show the significant differential in mass from hot water to cold water.

The prototype design has low accuracy (0.21%) compared to the desired accuracy of 0.05%. This increased error is due to the inadequacies of individual components of the system not to a flaw in the Horak design. The system was able to perform with accuracy's as high as 0.09% (see Figure 5) but only under cold water conditions. Improvements can be made to the system by choosing components that perform better under given conditions.

Improvements, such as level sensors that can operate well in aerated liquid and dynamic temperature conditions will detect liquid level more precisely to reduce error. Improving liquid input in the receiving tanks to reduces surface turbulence and aeration will reduce level detection error. Control valves that are more reliable and require less maintenance will prevent leaking errors. Using thermocouples instead of RTD's will greatly improve the time long time constant for temperature detection and using displacement mechanisms that react quicker to changes in temperature will improve the turn-around time for one dispensing cycle; therefore, improving efficiency of the system. Preventing aeration of liquid, controlling condensation inside the tank will improve level detection error. A computer control system designed for each application individually will improve overall accuracy's by reacting more rapidly to changes in the system, such as level sensor state detection, temperature detection and plunger displacement.

Control Program

In the following pages is the listing of the main computer control program used in the Horak design prototype. The language used to code the operation is a symbolic language called LabVIEW2. This language uses symbols as opposed to text as an interface with the programmer. This language was used for its ease of programmability, versatility and ability to perform real-time graphics in order to monitor the data of the system as it is running.

The program consists of five main sequence frames (0 through 4) and subsequence frames in each of the main frames. In the following pages is listed the frames of the main program (main Virtual Instrument or VI); other single function oriented sub-VI's are used but are not listed to simplify the description of the sequence of events that control the Horak design prototype.

Each frame of the main sequence and sub sequences are denoted at the top of each frame. Symbols placed within the field of the sequence are performed. Lines running between symbols are referred to as wires. Symbols connected by wires are linked in operation and symbols not linked together are meant to run concurrently but remain independent.

Figure 10 Main Panel of Control Panel .

This is the unmodified control panel which contains real time graphics that slow the operation of the computer. The tank represents the real time weight of the receiving tank, the two thermometers represent the temperatures of Tank A and Tank B and the level sensor graphic in the upper right hand corner of the control panel represents the state of the level sensor (covered or uncovered).

Each of these real time graphics are updated every iteration of the computer. The computer is unable to move to the next task until it has completed the previous task. The time needed to generate these graphics are significantly longer than the time needed to perform a mathematical calculation. These graphics are removed in the following figure.

Figure 11 Stripped Main Panel. Stripped Down Control Panel.

The real time graphics are removed in the stripped down version of the previous panel.

Figure 12 Main Sequence frame 0. 1st Sequence of control panel.

These commands create an new data file, reset Tank A filling timer and close all valves.

Figure 13 Main Sequence frame1 subsequence 0.

This picture depicts the second sequence frame of the main program.. In this sequence, a subsequence represents the timing of filling of Tank A. The first frame of this sequence opens the outlet valve for the receiving tank which empties receiving tank and shows a sub-VI (subroutine) called "Scale", which retrieves the current weight of the water in the dispensing tank. The subsequence is not advance until the weight of the Scale reads 6000 grams

Figure 14 Main Sequence frame1 subsequence 1.

The second frame of Tank A timer sequence closes the outlet valve and empties the remainder of any of water in Tank A until the lower level sensor in Tank A via the sub-VI named Level Sensor.

Figure 15 Main Sequence frame1 subsequence 2.

This frame of the Tank A timing subsequence resets the plunger to its reference position at 150F

Figure 16 Main Sequence frame1 subsequence 3.

This frame of the Tank A timing subsequence empties the receiving tank again.

Figure 17 Main Sequence frame1 subsequence 4.

This frame of the Tank A timing subsequence closes the output valve of the receiving tank.

Figure 18 Main Sequence frame1 subsequence 5.

This frame of the Tank A timing subsequence begins a counter to calculate the time for filling of Tank A. This is written to a file a file created in the previous sequence.

Figure 19 Main Sequence frame1 subsequence 6.

This frame of the Tank A timing subsequence fills Tank A from the bottom level sensor to the top level sensor . The subsequence is not advanced until the water level has reached the top level sensor-- this allows for the compensates for the change water input rate from one test to the next.

Figure 20 Main Sequence frame1 subsequence 7.

This frame of the Tank A timing subsequence records the stop time of the counter.

Figure 21 Main Sequence frame1 subsequence 8.

This frame of the Tank A timing subsequence empties Tank A and fills Tank B to prepare the tanks for the first cycle of the test.

Figure 22 Main Sequence frame1 subsequence 9.

This frame of the Tank A timing subsequence ends Tank A timing sequence.

Figure 23 Main Sequence frame2.

This frame of the main sequence calculates the time for filling Tank A from the Timer file and converts into seconds for usage by the following testing frame of the main sequence.

Figure 24 Main Sequence frame3 subsequence 0.

This frame represents the main testing of the Horak Prototype. Several sub-VI's are installed to calculate pertinent values such as temperature ("temp" sub-VI), plunger position ("Plunger" sub-VI).

In the center of this frame is the subsequence which controls the current operation the of the system. In the first frame of this subsequence, the histogram plots on the front panel are being updated to reflect the current data for the current test.

Figure 25 Main Sequence frame3 subsequence 1.

This subsequence frame empties the receiving tank by opening the outlet valve corresponding to the receiving tank.

Figure 26 Main Sequence frame3 subsequence 2.

Stabilizing water in receiving tank before initial tank weight is recorded.

Figure 27 Main Sequence frame3 subsequence 3.

Recording the Date and Time before emptying tank B.

Figure 28 Main Sequence frame3 subsequence 4.

Recording the weight of the stabilizing tank before emptying tank B.

Figure 29 Main Sequence frame3 subsequence 5.

Emptying Tank B and filling Tank A for 80% of the time of one entire filling.

Figure 30 Main Sequence frame3 subsequence 6.

This frame of the subsequence represents the displacement periods of the plunger in Tank A. The "Plunger" sub-VI sets a certain criteria for the displacement-- not until this criteria is satisfied does program advance to the next frame of the subsequence.

Figure 31 Main Sequence frame3 subsequence 7.

This subsequence frame advances the increment command to the next frame.

Figure 32 Main Sequence frame3 subsequence 8.

This subsequence frame completes the filling of Tank A and emptying of Tank B.

Figure 33 Main Sequence frame3 subsequence 9.

Recording temperature of Tank A.

Figure 34 Main Sequence frame3 subsequence 10.

This subsequence frame records the weight of the receiving tank after Tank B has been completely emptied.

Figure 35 Main Sequence frame3 subsequence 11.

This subsequence frame performs an "Extra Empty Cycle" for Tank B which is used to lower the water level below the lower level sensor in Tank B to prevent the accumulation of air bubbles on the lower level sensor.

Figure 36 Main Sequence frame3 subsequence 12.

This subsequence frame empties the receiving tank.

Figure 38 Main Sequence frame3 subsequence 14.

The function of this subsequence frame is to record the weight of the receiving tank before emptying Tank A.

Figure 39 Main Sequence frame3 subsequence 15.

This subsequence dispenses the water in Tank A and, concurrently fills Tank B. No plunger displacement occurs during this part of the cycle because no plunger is operating in Tank B

Figure 40 Main Sequence frame3 subsequence 16.

This subsequence frame records the temperature of Tank B as well as creates a ten second pause to stabilize any water ripples in the receiving tank before the weight is recorded into the data file.

Figure 41 Main Sequence frame3 subsequence 17.

The weight of the water dispensed from Tank A is recorded in this subsequence frame.

Figure 42 Main Sequence frame3 subsequence 18.

An extra empty cycle of Tank A is performed for 7.5 seconds in order to prevent any air bubbles to accumulate on the lower level sensor.

Figure 43 Main Sequence frame3 subsequence 19.

An additional subsequence is added in order to give the operator the option of stopping the control program after the current cycle has been completed.

Figure 45 Main Sequence frame 4.

If the "STOP" button or "Stop at EOF" (end of cycle) is pressed the main sequence is advance to the this frame (frame 4) and the control program closes all the solenoid valves and stops executing.

References

Bryan, R. and. N., Bill. (1986). Guidelines for Installing and Maintaining Reciprocating Metering Pumps. Plant Engineering, 56-59.

Foust, A., Wenzel, L., Clump, C., Maus, L., & Anderson, L. B. (1980). Principles of Unit Operations (second ed.). New York: John Wiley and Sons.

Liptak, B. (1982). Instrument Engineers' Handbook. Radnor, Pennsylvania: Chilton Book Company, 1003.

McCabe, R., Lanckton, P., & Dwyer, W. (1984). Metering Pump Handbook . New York: Industrial Press Inc.

Poynton, J. P. (1980). Selecting Diaphragm Metering Pumps. Machine Design, , 88-93.

Poynton, J. P. (1983). Metering Pumps: Selection and Application . New York: Marcel Dekker.

Warren, L. M. (1988). Gravimetrics: The New Heavyweight in Metering/Feeding. Plastics Technology, 34(11), 62-69.

END

**DATE
FILMED**

2 / 2 / 93