Date originally : DE93 007029

In support of continuing development of Computer Simulated Images of Radiopharmaceutical Distributions in Anthropomorphic Phantoms (CSIRDAP)

Progress to Date on the Project (first 3 years)

The initial development of Computer Simulated Images of Radiopharmaceuticals in Anthropomorphic Phantoms (CSIRDAP) over the first 3 years has been very successful. All components of the simulation have been coded, made operational and debugged (see Appendix for publications describing them in detail). These components include:

1. Input and display program for outlining patient anatomy.

Within the framework of CSIRDAP development an extensive program was written to read GE magtapes containing CT (or MRI) patient images. These images are displayed on the microVax 3500 workstation and can be manipulated by thresholding and changing color tables. The operator can outline the organs and store these contours on disk. These contours are used to define the fully 3-dimensional internal structures of the human body.

2. High resolution anthropomorphic phantom.

We have constructed an anatomically correct human geometry, which can be used to store radioisotope concentrations in 51 various internal organs. Each organ is associated with an index number which references to its attenuating characteristics (composition and density).

DISCLAIMER

This report was prepared as an account of work sponsored by an agency of the United States Government. Neither the United States Government nor any agency thereof, nor any of their employees, makes any warranty, express or implied, or assumes any legal liability or responsibility for the accuracy, completeness, or usefulness of any information, apparatus, product, or process disclosed, or represents that its use would not infringe privately owned rights. Reference herein to any specific commercial product, process, or service by trade name, trademark, manufacturer, or otherwise does not necessarily constitute or imply its endorsement, recommendation, or favoring by the United States Government or any agency thereof. The views and opinions of authors expressed herein do not necessarily state or reflect those of the United States Government or any agency thereof.

MASTER

DISTRIBUTION OF THIS DOCUMENT IS UNLIMITED
for

A list of the organs and their index number is given below:

0 - Void outside of phantom	18 - Small Intestine
1 - Skin/Body fat	19 - Colon/Large Intestines
2 - Brain (to be expanded)	20 - Pancreas
3 - Spinal Chord	21 - Adrenals
4 - Skull	22 - Fat (now included in one)
5 - Spine	23 - Blood Pool (all vessels)
6 - Rib Cage and Sternum	24 - Gas Volume (bowel)
7 - Pelvis	25 - Fluid Volume (bowel)
8 - Long Bones	26 - Bone Marrow
9 - Skeletal Muscle	27 - Lymph Nodes
10 - Lungs	28 - Thyroid
11 - Heart	29 - Trachea
12 - Liver	30 - Diaphragm
13 - Gall Bladder	31 - Spleen
14 - Left and Right Kidney	32 - Urine
15 - Bladder	33 - Feces
16 - Esophagus	34 - Testes
17 - Stomach	35 - Prostate

Organ Numbers for the Head

36 - Sinuses	51 - _____
37 - Jaw	52 - _____
38 - Optic Nerve	53 - _____
39 - _____	54 - _____
40 - Ear Bones	55 - _____
41 - Thalamus	56 - _____
42 - Corpus collosum	57 - _____
43 - Cerebellum	58 - _____
44 - Temporal Lobes	59 - Trachea
45 - Ventricles	60 - Cartilage
46 - _____	61 - _____
47 - Eye	62 - Nerves
48 - Striatum	63 - _____
49 - White matter	64 - _____
50 - Superior Sagittal Sinus	

2. Physical Constants of Radiation

Physical attenuation constants up to photon energies of 4 MeV are stored in separate tables for photoelectric, Compton and pair production interactions. The Klein-Nishina equation is available for generating stochastically determined scattering angles.

3. Generation of Decay Occurrences

By using conventional random number generation techniques, occurrences of radioactive decay sites are statistically modeled to correlate with the concentrations of radio-agent in the phantom. The amount of radioactivity in each organ is directly related to the probability that the next event will originate from that organ.

4. Histories of Gamma Rays

Histories of modes of interactions, mean path between interactions, and scatter angles are generated by mapping random numbers to the physical constants and scattering formulas. Calculation of gamma-ray histories are based on paths calculated through the digitized voxel phantom. Decisions made concerning mode of gamma-ray interaction and/or angle of scatter are based on the atomic composition and density of the organ found in the path of the radiation. The geometry of following 3-dimensional vectors and calculating interactions makes this portion of CSIRDAP the most sophisticated (and compute intensive) component.

5. Model of Nuclear Camera

The model of the nuclear camera includes: collimator acceptance angle, septal penetration, detector counting efficiency and intrinsic resolution as well as any image processing. Given a full geometric description of the collimator and detector system, the clinical image can be calculated and offers the possibility of easily "changing collimators" and/or "trying other cameras or techniques", e.g. SPECT, to get the best image.

6. Variance Reduction

Correctly treating 3-dimensional vectors is very compute intensive. As anticipated in the original grant proposal, methods of variance reduction have been applied for acceptable turn-around times. A description of these techniques have been published as a direct result of developing these increased efficiency algorithms [2,3]. With these variance reduction techniques, we can generate approximately 35 image counts per second per camera, running a VAX 3500.

7. Comparison to Clinical Studies

Comparisons to clinical studies were conducted by closely analyzing the emerging energy spectrum of the simulated and measured spectra in the clinical environment [4]. Realistic images have been publicly presented at a medically oriented conference [5].

The current status of CSIRDAP (last 6 months of original proposal) is in the generation of high quality planar images. Lesion detectability will be analyzed for several theoretical distributions of investigational radiopharmaceuticals by the end of this third year of funding.

8. Improved reconstruction algorithms

Additionally, as stated in the original grant proposal, one of the anticipated applications of the simulation would be to serve as an "ideal" acquisition for improving reconstruction algorithms. Ms. Mindy Lee has made good strides in using the output from CSIRDAP to improve nuclear medicine reconstructions [6,7]. She anticipates to be completed with her thesis by the end of 1992.

Extended Applications of CSIRDAP

CSIRDAP developments have already found fruitful applications in other

institutions. The phantom itself has been received by several institutions. Copies of the phantom have been made onto magtape and have been sent to the following groups:

Dr. Keith Eckerman
Health and Safety Research Division
Oak Ridge National Lab
P.O. Box 2008
Oak Ridge, Tennessee 37831-6383

Dr. David Weber
Medical Department
Brookhaven National Lab
Upton, New York 11973

Dr. Thomas Lewellen
Department of Radiology
University of Washington Medical Center
Seattle, Washington 98195

Dr. Bruce Barber
Arizona Health Sciences Center
Division of Nuclear Medicine
Tucson, AZ 85724

We are in the process of making the phantom available to:

Dr. Michael King
Department of Nuclear Medicine
University of Massachusetts Medical Center
55 Lake Avenue North

TABLE OF CONTENTS

- A1 High Resolution Anthropomorphic Phantom for Monte Carlo Analysis of Internal Radiation Sources
- A9 Voxel Based Monte Carlo Calculations of Nuclear Medicine Images and Applied Variance Reduction Techniques
- A20 Effect of Variance Reduction Techniques on Monte Carlo Simulated Energy Spectra
- A22 Monte Carlo Determination of Emerging Energy Spectra for Diagnostically Realistic Radiopharmaceutical Distributions
- A26 Bayesian Reconstruction of Functional Images Using Registered Anatomical Images as Priors
- A40 Locations of Discontinuities as Priors for Improved Bayesian Reconstructions from Projections

*Proofs & proofs removed
and cycled separately.*

END

**DATE
FILMED**

3 / 12 / 93