



Lawrence Berkeley Laboratory

UNIVERSITY OF CALIFORNIA

Physics, Computer Science & Mathematics Division

MASTER


To be presented at the First IEEE Computer Society
International Symposium on Medical Imaging and Image
Interpretation, Berlin, W. Germany, October 26-28, 1982;
and to be published in the Proceedings

THE IMAGING PERFORMANCE OF A MULTIWIRE PROPORTIONAL- CHAMBER POSITRON CAMERA

Victor Perez-Mendez, Alberto Del Guerra,
Walter R. Nelson, and Kwok C. Tam

August 1982

DISTRIBUTION OF THIS DOCUMENT IS UNLIMITED



LEGAL NOTICE

This book was prepared as an account of work sponsored by an agency of the United States Government. Neither the United States Government nor any agency thereof, nor any of their employees, makes any warranty, express or implied, or assumes any legal liability or responsibility for the accuracy, completeness, or usefulness of any information, apparatus, product, or process disclosed, or represents that its use would not infringe privately owned rights. Reference herein to any specific commercial product, process, or service by trade name, trademark, manufacturer, or otherwise, does not necessarily constitute or imply its endorsement, recommendation, or favoring by the United States Government or any agency thereof. The views and opinions of authors expressed herein do not necessarily state or reflect those of the United States Government or any agency thereof.

THE IMAGING PERFORMANCE OF A MULTIWIRED PROPORTIONAL-CHAMBER POSITRON CAMERA^(*)Victor Perez-Mendez^(1,2), Alberto Del Guerra^{(1,*),} Walter R. Nelson⁽³⁾, and Kwok C. Tan⁽⁴⁾

(1) Lawrence Berkeley Laboratory, Berkeley, CA. 94720; (2) Department of Radiology, UCSF, San Francisco, CA. 94143; (3) Stanford Linear Accelerator Center, Stanford, CA. 94305; (4) General Electric Co., Schenectady, NY 12345.

ABSTRACT

A new design - fully three dimensional - Positron Camera is presented, made of six MultiWire Proportional Chamber modules arranged to form the lateral surface of a hexagonal prism. A true coincidence rate of 56000 c/s is expected with an equal accidental rate for a 400 μ CI activity uniformly distributed in a $\sim 3\ell$ water phantom. A detailed Monte Carlo program has been used to investigate the dependence of the spatial resolution on the geometrical and physical parameters. A spatial resolution of 4.8 mm FWHM has been obtained for a ^{18}F point-like source in a 10 cm radius water phantom. The main properties of the limited angle reconstruction algorithms are described in relation to the proposed detector geometry.

In this paper we propose a new design for a large area positron camera and discuss its expected imaging performance, which has been evaluated both by experimental measurements and by Monte Carlo simulation. The proposed large area concentric detector is conceived to measure a sensitive volume of ~ 3 liters. The detector is based on a MWPC with lead glass drift space converters.

In the next section a description of the camera is given and a brief summary of the experimental results is reported. Section 3 deals with the Monte Carlo simulation of the detector and the results are presented in section 4. Limited angle reconstruction algorithms will be used to produce tomographical imaging of the real data. The main features of such algorithms, as applied to this detector, are described in section 5. Finally, some concluding remarks are given in the last section.

2. The MWPC Positron Camera

1. Introduction

The technique of tomographical imaging with x-rays^(1,2) has now become an essential tool of any major medical center. Although tomographical imaging originated in practice in early nuclear medicine laboratories^(3,4), its application lagged behind, due to technological difficulties. In fact, only after CT-scanners came into operation has research in this field become appreciable and have the first positron tomography cameras been developed in USA⁽⁵⁻⁹⁾ and European⁽¹⁰⁾ laboratories. A number of positron tomographs have been constructed and are now used for medical research. Most of these are single ring scintillator detectors, using either sodium iodide^(11,12) or bismuth germanate crystals^(13,14). More recently cesium fluoride crystals have been used⁽¹⁵⁻¹⁷⁾ because of their fast time response. It is usually necessary to measure more than one slice of the object: this can be done either by a series of parallel strip images taken sequentially, or by making large area detectors or multi-ring detectors (which are much more expensive to build than single section tomographs).

A schematic drawing of the proposed camera is shown in Fig. 1. Six modules (each $50 \times 50 \text{ cm}^2$) are arranged to form the lateral surface of a hexagonal prism. Following the annihilation of the positron within the target, two γ -rays are produced, each with an energy of 511 keV and opposite to the other within a few mrad. A good event results when both gammas are detected "in coincidence" in opposite modules. Each module consists of a standard MWPC ($45 \times 45 \text{ cm}^2$ active area), with a 2 cm thick lead glass tube converter on each side (Fig. 1).

2.1 Principle of Operation of a Dense Drift Space MWPC

The detection of 511 keV γ -rays with a MWPC requires the use of a high density, high Z converter with large surface to volume ratio. We have developed a converter made of glass capillarity (0.9 mm inner diameter, 0.096 mm wall thickness) of high lead content (80% PbO by weight, glass density of 6.2 g/cm^3), fused to form honeycomb matrices⁽¹⁸⁾. The lead glass matrices are treated in a H_2 reduction process to form a uniform resistive layer on the inner walls of each tube. The Compton- or photo-electron produced by the photon interacting within the converter has a finite range dependent upon its energy. If it reaches the gas region within the tube, a number of primary ionization electrons are produced. A voltage difference applied between the ends of the tubes then drifts these primary electrons along the electric field lines within the tube towards the cathode planes and into the avalanche region of the chamber.

(*) On leave of absence from: Istituto di Fisica dell' Università, Pisa, Italy.

(+*) This work was supported by the Director, Office of Energy Research, Office of High Energy and Nuclear Physics, Division of High Energy Physics of the U.S. Department of Energy under Contract #DE-AC03-76SF00098.




Fig. 1 - Proposed Positron Camera made of six modules arranged to form a hexagonal prism: plan view of the camera (TOP), cross view of a single module (BOTTOM).

2.2 Experimental Results

We have built two $50 \times 50 \text{ cm}^2$ modules and some experimental measurements have been taken. For convenience we have extensively tested a smaller module ($15 \times 15 \text{ cm}^2$ active area) equipped with one 1 cm thick converter. The chamber itself is a standard gas filled MWPC with the two cathode planes at 90° to each other in order to have both x and y localization. Position readout is achieved by means of fast delay lines (8 ns/cm) coupled to the cathode

planes (19). The experimental details have already been reported (20-22). Table 1 briefly summarizes the results obtained with a standard argon-methane (70-30) gas mixture at 2 atm for 511 keV γ -rays incident perpendicular onto the module plane. In order to extrapolate these results to converters of different thickness we assume that the efficiency is proportional to their interaction probability, whereas the time resolution is determined by the transit time of the electrons within the converter tubes. For the same electron drift velocity in the gas (i.e. the same value of the reduced electric field), the transit time is then proportional to converter thickness. The expected performance of the Positron Camera module is also presented in Table 1. The gas pressure will be kept at 2 atm in order to take advantage of the self quenching streamer regime (23,24).

2.3 Count Rates of the Tomograph

It is well known that given a pure positron emitting source and two opposite γ -ray detectors, both the single rate and the true coincidence rate are proportional to the source strength. If the solid angle fraction (ϵ_Ω) subtended at the source is the same for both detectors and if the efficiency (ϵ) for 511 keV γ -rays is the same and constant within the solid angle, one gets

$$N_1 = N_2 = (2S)\epsilon_\Omega\epsilon \quad (1a)$$

$$T = S (2\epsilon_\Omega^2)\epsilon^2 \quad (1b)$$

$$A = N_1 N_2 \quad (1c)$$

where N_1 and N_2 are the single rate of the detectors, S is the source strength, T and A on the true and accidental coincidence rates respectively, and τ is the resolving time of coincidence.

It is conventional to express the figure-of-merit parameter of a positron camera by giving the ratio ϵ^2/τ . One obtains the following relation, which applies to a module pair:

$$T^2/A = \epsilon^2/2\tau \quad (2)$$

For a given set of detector parameters T/A is inversely proportional to the source strength:

$$\frac{T}{A} = \frac{1}{S} \frac{1}{4 \epsilon_\Omega^2 \tau} \quad (3)$$

For a signal to noise ratio

$$\frac{T}{A} = 1 \quad (4)$$

Table 1
Specification and Performance of the Test Module and of the Positron Camera Module

| | Test Module | Positron Camera Module |
|--|----------------|------------------------|
| MWPC active area (cm^2) | 15×15 | 45×45 |
| Converter thickness (cm) | 1 | 2 + 2 |
| Gas pressure (atm) | 2.0 | 2.0 |
| Efficiency (%) | 4.5 | 15.0 |
| Time resolution (ns) | 130 | 200 (**) |
| Spatial resolution FWHM | | |
| -along the coordinate parallel to the snode wires (mm) | 1.3 | |
| -along the other coordinate (mm) | 2.0 | 2.0 |

(**) Using a 20% faster gas mixture according to ref. 20.

the value of the source strength is given by

$$S|_{T/A=1} = \frac{1}{4 \frac{E_D}{E} \tau} \quad (5)$$

For our detector, one gets a value of $\sim 400 \mu\text{Ci}$, which represents the maximum source strength we can use for a signal to noise ratio ≥ 1 . Using condition (4), equation (2) becomes

$$T|_{T/A=1} = c^2/2\tau \quad (6)$$

which gives the maximum true coincidence rate per module pair compatible with condition (4). The expected count rates in air and with a 10 cm radius water phantom are listed in Table 2. The effect of the absorbing medium is equivalent to decreasing the efficiency of each detector (as determined by using the Monte Carlo program described in section 3). Hence, the net result of the absorption and scattering in the phantom is to decrease the statistics, while keeping the same signal to noise ratio.

3. Monte Carlo Technique and Problem Model

3.1 The EGS Code: General Considerations

To study the spatial resolution properties of the tomograph, a general electromagnetic radiation transport code called EGS (Electron-Gamma-Shower)⁽²⁵⁾ was implemented for the problem at hand. EGS is currently being used to solve a variety of problems in accelerator, high-energy, and medical physics⁽²⁶⁾. In particular, EGS is capable of treating electrons, positrons and photons with kinetic energy as low as 10 keV (electrons and positrons) and 1 keV (photons). The transport can take place in any of one hundred different elements or in any mixture or compound of these elements. It is left to the user to construct his own geometry and to score a particular answer.

For the problem at hand, the hexagonal 3-dimensional geometry of the positron camera was simulated. In what follows, we will limit ourselves to the results obtained for a point-like source at the center of a 10 cm radius water phantom. The simulation may be subdivided into three main modes of operation:

- generation, transport, and annihilation of the positron,
- transport (including interaction) of the annihilation quanta within the phantom and the detector,

- scoring of coincidence events and production of spatial resolution histograms.

As many as 200 positrons per second are generated and "followed" up to the scoring point on the IBM-3081 (i.e. 5 ns/positron).

3.2 The Positron Mode

We have used a form for the theoretical beta spectrum given by Konopinski and Rose⁽²⁷⁾. The energy spectra of the radioisotopes were generated using the Fermi functions tabulation of Fano⁽²⁸⁾, corrected for the screening effect. The spectra were introduced into EGS in the form of look-up Tables.

Once the positron has been generated both in position and in direction and its energy has been sampled according to the above scheme, it is followed within the phantom until it reaches a lower kinetic energy cut-off (10 keV in our case), when it is forced to annihilate at rest. In addition to Bhabha scattering, multiple scattering and continuous energy loss, EGS also considers annihilation in flight as a discrete Monte Carlo process. Depending on the source, between 1 and 5 per cent of the positrons have been found to annihilate in flight, consistent with theory⁽²⁹⁾. Because of the thermal motion of the orbital electrons two photon annihilation at rest is not perfectly collinear in the laboratory. This non-collinearity is accounted for in the present Monte Carlo study by using a fit to the data by Colombino et al.⁽³⁰⁾, who measured an angular distribution of 8.5×10^{-3} rad (FWHM) in water at 22°C. The in flight angle of the annihilation quanta, on the other hand, is accounted for in EGS by means of kinematics. The position, energy and direction information for the annihilation quanta are used as input to the subsequent two-gamma simulation phase. Other particles (e.g., bremsstrahlung, and delta rays) are discarded in the phantom immediately upon production.

3.3 The Two-Gamma Mode

During the transport of the annihilation quanta within the phantom, all charged particles that are generated are immediately discarded. If the photon emerges from the phantom with an energy greater than the cut-off energy (100 keV), it is further transported through the hexagon detector geometry. Following an interaction in the lead converter, the Compton- or photo-electron is assumed to be detected by the MWPC provided its kinetic energy is greater than

Table 2
Count Rates for the Proposed Tomograph for a Point-like Source of 400 μCi ($T/A = 1$)

| | In Air | 10 cm radius water phantom(*) |
|--|-------------------|-------------------------------|
| Single rate per module (c/s) | 375×10^3 | 250×10^3 |
| True coincidence rate per module pair (c/s) | 56×10^3 | 18.5×10^3 |
| True coincidence rate for the system (c/s) | 168×10^3 | 56×10^3 |
| Total coincidence rate for the system: T+A (c/s) | 336×10^3 | 112×10^3 |

(*) As calculated by the Monte Carlo described in section 3.

200 keV. This value has been arbitrarily chosen, according to the range energy tables⁽³¹⁾, for the wall thickness of the glass tube. If more than one such electron is produced then the one closest to the wire plane is selected.


Each event can be displayed on a graphic device and the history of the two photons can be visualized. An example is presented in Figure 2, showing two orthogonal views of the tomograph. The photons are shown as solid and dotted lines. The scattering information related to each of the two photons are printed in the two corresponding boxes, where the various interactions are indicated. For each interaction (#) various quantities are provided (S, D, T, GE1, EKE, GE2, and TD). S is the sector where the interaction took place (1-6); it assumes the value of 0 if the interaction took place in the phantom itself. D indicates the converter (first or second); T identifies the type of interaction (C for Compton, P for photoelectric). GE1 and GE2 are the photon energies before and after the interaction; EKE

is the kinetic energy of the electron. TD is the distance of the interaction point from the end of the tube. In the example presented in Figure 2, the first photon (solid line and solid box) interacts twice in sector 3: the first one produces a Compton electron of 307 keV and the second a photoelectron of 116 keV. This photon is assumed to be detected at the point where the first interaction occurred, because it is the only one over the prestablished detection threshold. The second photon (dotted line and dotted box) also makes two interactions: the first one produces a Compton electron of 209 keV in the first converter of sector 6; the second produces a photoelectron of 215 keV in the second detector of the same sector. In this case the two electrons are both above the threshold of 200 keV, but only the nearest interaction point to the wire plane (the second one) is retained as detection position.

Subsequently the intrinsic spatial resolution of the MWPC module is introduced into the overall simulation. The real detector is

| # | S | D | T | GE1 | EKE | GE2 | TD(CM) |
|---|---|---|---|-----|-----|-----|--------|
| 1 | 3 | 1 | C | 511 | 307 | 204 | 0.038 |
| 2 | 3 | 1 | P | 204 | 116 | | 0.105 |

| # | S | D | T | GE1 | EKE | GE2 | TD(CM) |
|---|---|---|---|-----|-----|-----|--------|
| 1 | 6 | 1 | C | 511 | 209 | 302 | 1.146 |
| 2 | 6 | 2 | P | 302 | 215 | | 0.089 |



XBL 828-11166

Figure 2 - Typical Graphic display of a simulated event.

able to distinguish in which converter the interaction took place⁽³²⁾, but it is not able to tell where along the glass tube it occurred, generating a parallax error. This is accounted for in our simulation, by translating the actual detection point to the middle of the converter. Then the two coordinates on the wire plane are sampled according to gaussian distributions, whose FWHM has been experimentally determined (see table 1). Finally the spatial cuts, which take into account the missing angle between two adjacent modules, are applied. (Only the events inside the active area of the detector are kept.)

3.4 Spatial Resolution Histograms

To study the spatial resolution of the Positron Camera a point-like positron source was simulated on the x-z plane (see Figure 2) at the origin. A coincidence event is defined when both photons are detected in opposite modules. For each coincidence event the line which connects the two detection points is computed. The intersection of this line with the x-z plane is calculated and its distance from the source position is accumulated in a two-dimensional histogram. Finally, a profile along any direction on the plane across the origin gives the spatial resolution distribution of the Positron Camera.

4. Monte Carlo Results

4.1 Contributions to the Spatial Resolution

Several factors contribute to the spatial resolution of the system: positron range, two-gamma non-collinearity, Compton scattering in the phantom and intrinsic detector resolution. With our Monte Carlo simulation it has been possible to study the absolute contribution of each to the overall spatial resolution; the results, which have obtained for a 10 cm radius water phantom, are now discussed in turn.

Positron Range - The radioisotopes studied are listed in Table 3 together with their main parameters and the maximum path length⁽³³⁾. The range distributions are non-gaussian in shape and have a very long tail, especially for the high energy positron emitters. The effect of the positron range is best described by giving both the FWHM and the FW(0.1)M of the spatial distributions obtained in the simulation. Also included in Table 3 are the radii of the spheres which contain 50% (r_{50}) and 95% (r_{95}) of the annihilation points. These numbers have been found to be in

good agreement with experimental measurements⁽³⁴⁾. In general r_{95} is found to be $\sim 1/2$ of the maximum path length.

Two-gamma non-collinearity - To a first approximation, an angular spread at the detector plane ($\Delta\theta$) will cause a positional deviation at the image plane of $\sim R\Delta\theta/2$, where R is the distance between the two planes. From the Monte Carlo results we have found a FWHM of 2.2 mm and a FW(0.1)M of 7.1 mm. The contribution from annihilation in flight, which is isotope dependent, is negligible ($< 1\%$ for ^{11}C).

Compton scattering in the phantom - Contrary to a crystal-type detector, the efficiency of a dense drift space MWPC diminishes with decreasing energy of the incident photon^(35,36). For this reason, Compton scattering is much less important here than for scintillation cameras. For a 10 cm radius water phantom both single and coincidence rates are reduced by a factor 1.5 and 3, respectively (see Table 2). Furthermore, approximately one third of the coincidence events are uniformly distributed within the phantom, whereas the smearing of the spatial resolution is negligible (< 0.1 mm FWHM).

Detector response - To a first approximation parallax error has a triangular distribution at the detector plane with a FWHM of $t\sqrt{A/R}$, where t is the converter thickness, A is the area of the module and R is the distance from the source. The positional deviation at the image plane will then be ~ 1.5 mm ($t = 2$ cm, $A = 45 \times 45$ cm², $R = 50\sqrt{3}/2$ cm). From the Monte Carlo simulation we obtain a FWHM of 2.4 mm and a FW(0.1)M of 4.8 mm. The intrinsic spatial resolution of the MWPC for both x- and y-coordinates ($\Delta x = \Delta y = 2.0$ mm FWHM, see Table 1) introduces an additional contribution of $\sim 1/\sqrt{2} (\Delta x^2 + \Delta y^2)^{1/2} \sim 2.0$ mm FWHM.

Overall spatial resolution of the system - The various contributions and the overall spatial resolution of the system are given in Table 4 for a ^{11}C point-like source at a center of a 10 cm radius water phantom. The overall spatial resolution values, which are obtained by adding the FWHM values in quadrature are smaller than the calculated Monte Carlo values, as expected for the following reasoning: one cannot simply add non-gaussian distributions in quadrature. The Monte Carlo simulation, however, performs the convolution integration correctly. Similarly, for the other radioisotopes, the overall spatial resolution ranges from 4.8 mm (FWHM) for ^{18}F to 8.2 mm (FWHM) for ^{82}Rb .

Table 3

Positron Emitters Characteristics, Range and Contribution to the Spatial Resolution

| | ^{18}F | ^{11}C | ^{15}O | ^{38}K | ^{82}Rb |
|--|-----------------|-----------------|-----------------|-----------------|------------------|
| Mean life (m) | 109.8 | 20.38 | 2.03 | 7.61 | 1.25 |
| End-point of the β^+ spectrum (MeV) | 0.635 | 0.961 | 1.723 | 2.630 | 3.335 |
| Maximum path length in water ⁽³³⁾ (mm): | 2.40 | 4.09 | 8.14 | 12.93 | 16.57 |
| Range contribution to the spatial resolution: | | | | | |
| FWHM (mm) | 0.3 | 0.4 | 0.7 | 1.2 | 2.0 |
| FW(0.1)M (mm) | 1.2 | 1.6 | 3.3 | 6.0 | 8.5 |
| r_{50} in water (mm) | 0.31 | 0.60 | 1.47 | 2.76 | 3.91 |
| r_{95} in water (mm) | 0.98 | 1.72 | 3.75 | 6.50 | 8.86 |

Table 4
Contribution to the Positron Camera Spatial Resolution

| | FWHM(mm) |
|--------------------------------|----------|
| Positron range | 0.4 |
| Two gamma non-collinearity | 2.2 |
| Detector response: | |
| - Parallax error | 2.4 |
| - Intrinsic spatial resolution | 2.5 |
| Overall spatial resolution | 5.5 |

5. Radioisotope Distribution Reconstruction:
Limited Angle Data

The radioisotope distribution within the object that we determine from the coincidence pair distributions measured in the camera is obtained by a Fourier Inversion technique⁽³⁷⁾. Additionally, in order to achieve the best reconstruction with data subject to statistical fluctuations, we use a smoothing algorithm due to Phillips⁽³⁸⁾.

The Fourier Inversion method as such would be satisfactory if the camera were truly a six-sided box with no gaps in detection efficiency at the junctions of the detector sides. From a practical point of view, for simplicity in construction we allow an estimated gap of $\sim 5^\circ$ at the vertices of the hexagonal detector slots (Figure 1) -- amounting to a total missing angle of $\sim 30^\circ$.

It can be shown that the solid angle in front and back perpendicular to the axis of the hexagon contributes only to a loss of detection efficiency, since annihilation γ -rays going in those directions are not recorded. It does not contribute to any artifacts in the reconstruction so long as the object is smaller than the length of the sensitive area of the hexagon. On the other hand, if no provision is made for adequate treatment of the 30° azimuthal angle missing information, the quality of the reconstruction suffers and line artifacts will be created.

We summarize below the limited angle Fourier Inversion algorithm that we have developed and will use in this camera.

Let $\phi_0(r)$ be the point response of the detector system. In what follows we ensure that

$\phi_0(r)$ is space invariant - i.e. that every point within the object has the same detection efficiency - by restricting the cone of directions which we accept for the cone reconstruction.

Then we can write the following convolution

$$\phi(r) = \int \phi(r') \phi_0(r-r') d^3r' \quad (7)$$

where $\phi(r)$ is the measured distribution of counts, and $\phi(r')$ is the unknown radioisotope distribution within the object. From the Fourier Convolution theorem we can write:

$$\phi(p) = \phi(p) \phi_0(p); \quad \phi(p) = \phi(p) / \phi_0(p) \quad (8)$$

$$\phi(r) = \int \phi(p) e^{-2\pi i p r} d^3p \quad (9)$$

If we use the Phillips smoothing algorithm equation (8) becomes

$$\phi(p) = \frac{\phi(p)}{\phi_0(p) + \gamma(2\pi)^{-4} p^4 / \phi_0(p)} \quad (10)$$

The parameter γ is an adjustable parameter that depends on the statistical noise level. Increasing γ makes the solution smoother but eliminates fine details of the image. In practice we constrain γ to be consistent with the spatial resolution determined by the desired pixel size on the reconstruction.

Note that in equation (10) $\phi(p)$ is not defined for Fourier components in the missing angles, i.e. the directions in the Fourier space corresponding to those in which the camera has gaps. Using the Phillips algorithm, Eq. (10), leads to the result that $\phi(p) = 0$ for the missing angles in the Fourier plane. This permits the integral in (9) to be calculated unambiguously; however, the result is a poor approximation to the real distribution and creates a number of line artifacts for each point in the object⁽³⁹⁾.

The method for approximating the missing Fourier components outlined below depends on the fact that the object source distribution is in a finite volume and has a finite maximum. From mathematical principles it is well known that, for such distributions, knowledge of the Fourier components in a limited region determines their values over all space, since the Fourier distribution in this case is a complete function. The computational algorithm is based on a similar algorithm due to Gerchberg⁽⁴⁰⁾ and




Figure 3 - The computer iteration scheme.

applied to electrical signals by Papoulias⁽⁴¹⁾. Figure (3) shows the computer iteration scheme used for this purpose. On a test object simulation in our camera, for a circular phantom of 20 cm diameter reconstructed in $6 \times 6 \times 10 \text{ mm}^3$ voxels, and using a count of 400 counts/voxel we obtained an appreciable reduction in the error of the reconstruction compared with the test object simulation.

6. Concluding Remarks

The tomograph we propose has an intrinsic multislice capability: fields of view of 20-30 cm can be easily accommodated. Furthermore, the solid angle coverage (~50%) and the possibility of detecting off-plane photon pairs allows for less activity, thus reducing the dose delivered to the patient. In 3 min as many as 1×10^6 true coincidence events may be collected per slice, if a source of 400 μCi is uniformly distributed in a typical head phantom (Table 5), and as many as 10 simultaneous slices may be obtained for a total of 10×10^6 coincidence events. Approximately the same number of events are collected in the same time with the three-plane NEUROCAT Tomograph⁽⁴²⁾ with a higher activity source. The signal to noise ratio for the MWPC solution is worse than for NEUROCAT, but a better spatial resolution (5-6 mm FWHM) may be claimed, and systems with resolution better than 7 mm FWHM are now required for both brain and heart imaging⁽⁴³⁾.

Table 5

True Coincidence Rates for an Activity of
 $0.1 \mu\text{Ci/ml}$ of $^{111}\text{In}(\text{Tl-A})$ in a 10 cm Radius,
 10 cm Long Water Phantom

| | |
|--|-----------|
| Total True coincidence rate | 56000 c/s |
| Number of simultaneous slices (1 cm thick) | 10 |
| True coincidence rate per slice | 5600 c/s |
| Spatial resolution (FWHM) | 5.5 mm |
| Number of voxels per slice ($0.6 \times 0.6 \times 1 \text{ cm}^3$) | ~870 |
| True coincidence rate per voxel | 400 c/s |
| Compton distributed noise | ~1/3 |

References

- (1) G.N. Hounsfield, "Computerized Transverse axial scanning (Tomography): Part I. Description of System", Brit. J. Radiol., Vol. 46, pp. 1016-1022, 1973.
- (2) A.M. Cormack, "Reconstruction of Densities from their Projections, with Applications in Radiological Physics", Phys. Med. Biol., Vol 18, pp. 195-207, 1973.
- (3) D.E. Kuhl and R.Q. Edwards, "Image separation radioisotope scanning", Radiology, Vol. 80, pp. 653-661, 1963.
- (4) D.E. Kuhl and R.Q. Edwards, "The Mark IV system for quantitative reconstruction of brain radioactivity", J. Nucl. Med., Vol. 16, Abstract, p. 543, 1975.
- (5) C.A. Burnham and G.L. Brownell, "A multicrystal positron camera," IEEE Trans. Nucl. Sci., Vol. NS-19, No. 3, 201-205, 1972.

- (6) M.M. Ter-Pogossian, M. Phelps, and E.J. Hoffman, "A positron emission transaxial tomograph for nuclear imaging (PETT)," Radiology, Vol. 114, pp. 89-98, 1975.
- (7) C.J. Tompson, Y.L. Yamamoto, and E. Meyer, "A positron imaging system for the measurement of regional cerebral blood flow," in Proc. Soc. Photo-Opt. Instrum. Eng., Vol. 96, pp. 263-268, 1976.
- (8) Z.H. Cho, J.K. Chan, and L. Eriksson, "Circular ring transverse axial positron camera for 3-dimensional reconstruction of radio-nuclide distributions," IEEE Trans. Nucl. Sci., Vol. NS-23, pp. 613-622, 1976.
- (9) S.E. Derenzo, T.F. Budinger, J.L. Cahoon, R.H. Huesman, and H.G. Jackson, "High resolution computed tomography of positron emitters," IEEE Trans. Nucl. Sci., Vol. NS-24, pp. 544-558, 1977.
- (10) C. Bohm, L. Eriksson, M. Bergstrom, J. Litton, R. Sundman, and M. Singh, "A Computer assisted ring detector positron camera system for reconstruction tomography of the brain," IEEE Trans. Nucl. Sci., Vol. NS-25, pp. 624-637, 1978.
- (11) N.A. Mullani, M.M. Ter-Pogossian, C.S. Higgings, J.T. Hood, and D.C. Ficke, "Engineering aspects of PETT V," IEEE Trans. Nucl. Sci., Vol. NS-26, pp. 2703-2706, 1979.
- (12) L. Eriksson, Chr. Bohm, M. Bergstrom, K. Ericson, T. Greitz, J. Litton, and L. Widen, "One year experience with a high resolution ring - detector positron camera system: present status and future plans," IEEE Trans. Nucl. Sci., Vol. NS-27, pp. 435-444, 1980.
- (13) S.E. Derenzo, T.F. Budinger, R.H. Huesman, J.L. Cahoon and T. Vuletic, "Imaging properties of a Positron Tomograph with 280 BGO crystals," IEEE Trans. Nucl. Sci., Vol. NS-28, pp. 81-89, 1981.
- (14) E. Hoffman, M. Phelps, S. Huang, D. Plummer and D. Kuhl, "Evaluating the Performance of Multiplane Positron Tomographs designed for Brain Imaging," IEEE Trans. Nucl. Sci., Vol. NS-29, pp. 469-473, 1982.
- (15) R. Allemann, C. Grasslet, and J. Vacher, "Potential advantages of Cesium Fluoride Scintillator for a time-of-flight positron camera," J. Nucl. Med., Vol. 21, pp. 153-155, 1980.
- (16) D. Ficke, D. Beecher, C. Hoffman, J. Hood, J. Markham, N. Mullani, and M. Ter-Pogossian, "Engineering Aspects of PETT VI," IEEE Trans. Nucl. Sci., Vol. NS-29, pp. 474-478, 1982.
- (17) M. Yamamoto, D.C. Ficke, M. Ter-Pogossian, "Performance study of PETT VI, a Positron computed Tomograph with 288 Cesium Fluoride Detectors," IEEE Trans. Nucl. Sci., Vol. NS-29, pp. 529-533, 1982.
- (18) G.K. Lum, M.I. Green, V. Perez-Hendez and K.C. Tam, "Lead Oxide Glass Tubing Converters for Gamma Detection in MWPC," IEEE Trans. Nucl. Sci., Vol. NS-27, pp. 157-165, 1980.
- (19) P. Lecomte, V. Perez-Hendez and G. Stoker, "Electromagnetic Delay Lines in Spark, Proportional and Drift Chamber Applications," Nucl. Instr. and Meth., Vol. 153, pp. 543-551, 1978.
- (20) G.K. Lum, V. Perez-Hendez, and B. Sleaford, "Gamma-Ray Detection with Pb Glass

- Converters in MWPC: Electron Conversion Efficiency and Time Resolution", IEEE Trans. Nucl. Sci., Vol. NS-28, pp. 821-824, 1981.
- (21) A. Del Guerra, C.B. Lim, G.K. Lum, D. Ortendahl, and V. Perez-Mendez, "Medical positron imaging with a dense drift space Multi-Wire Proportional Chamber", Lawrence Berkeley Laboratory, LBL-14043 (March 1982), and IEEE Trans. Med. Imag., Vol. TH1-1, 1982, in press.
- (22) A. Del Guerra, V. Perez-Mendez, G. Schwartz, and S. Sleaford, Lawrence Berkeley Laboratory, LBL-14044 (March 1982), and Proceedings of the "Int. Conf. on Applications of Physics to Medicine and Biology", Trieste (Italy), 30 March - 3 April 1982, Ed. by: G. Alberi, Z. Bajzer, and P. Baza, World Scientific Publishing Co., Singapore, 1982, in press.
- (23) T.A. Mulera, and V. Perez-Mendez, "Observation of Large Saturated Pulses in Wire Chambers with Argon-Carbon Dioxide Mixtures", Lawrence Berkeley Laboratory, LBL-14003, (1982), and to be published in Nucl. Instr. and Methods, 1982.
- (24) T.A. Mulera, A. Del Guerra, V. Perez-Mendez, and G. Schwartz, "Large Signal Production in Wire Chambers Filled with Noble Gas-Carbon Dioxide and Noble-Gas Hydrocarbon Mixtures", Lawrence Berkeley Laboratory, LBL-14412 (1982), and to be presented at the 1982 IEEE NS-Symposium, Washington, D.C., October 20-22, 1982.
- (25) R. L. Ford and W. R. Nelson, "The EGS Code system: Computer program for the Monte Carlo Simulation of Electromagnetic Cascade Showers (Version 3)", Stanford Linear Accelerator Center, SLAC-210, June 1978.
- (26) W.R. Nelson, and T.M. Jenkins, editors, "Computer Techniques in Radiation Transport and Dosimetry", Plenum Press, New York, 1980.
- (27) E. J. Konopinski, and M.E. Rose, "The Theory of Nuclear β -decay", in Alpha-, Beta-, Gamma-Ray Spectroscopy", ed. K. Siegbahn, North-Holland Publ. Co., Amsterdam, 1965, pp. 1327-1364.
- (28) U. Fano "Tables for the Analysts of Beta Spectra", National Bureau of Standards, Applied Mathematics Series, Vol. 13, 1952.
- (29) W. Heitler, "The Quantum Theory of Radiation", Clarendon Press, Oxford, 1954, pp. 270-271.
- (30) P. Colombino, B. Piscella, and L. Trossi, "Study of Positronium in Water and Ice from 22 to -144°C by Annihilation Quanta Measurements", Nuovo Cimento, Vol. 38, pp. 707-723, 1965.
- (31) L. Katz, and A.S. Penfold, "Range-theory Relations for Electrons and the Determination of Beta-Ray End-Point Energies by Absorption", Rev. Mod. Phys., Vol 24, pp. 28-44, 1952.
- (32) A. Del Guerra, V. Perez-Mendez, G. Schwartz, and W.R. Nelson, "Design Considerations for a High Spatial Resolution Positron Camera", Lawrence Berkeley Laboratory Report, LBL-14414 (1982), and to be presented at the 1982 IEEE NS-Symposium, Washington D.C., October 20-22, 1982.
- (33) H.J. Berger and S.M. Seltzer, "Table of Energy Losses and Ranges of Electrons and Positrons", National Aeronautics and Space Administration Report, NASA - SP - 3012, 1964.
- (34) S.E. Derenzo, "Precision Measurement of Annihilation Point Spread Distributions for Medically Important Positron Emitters", in Proceedings on the "5th International Conference on Positron Annihilation", Sendai, Japan, The Japan Institute of Metals, pp. 819-823, 1979.
- (35) A.P. Jeavons, G. Charpak, and R.J. Stubbs, "The High-Density MultiWire Drift Chamber", Nucl. Instr. and Meth., Vol. 124, pp. 491-503, 1975.
- (36) D. Chu, K.C. Tam, V. Perez-Mendez, C.B. Lim, D. Lambert, and S.N. Kaplan, "High-Efficiency Collimator - Converters for Neutral Particle Imaging with MWPC", IEEE Trans. Nucl. Sci., Vol. NS-23, pp. 634-639, February 1976.
- (37) K.C. Tam and V. Perez-Mendez, "Tomographical Imaging with Limited-Angle Input", J. Opt. Soc. Am., Vol. 71, pp. 582-592, 1981.
- (38) D.L. Phillips, "A Technique for Numerical Solution of Certain Integral Equations", J. Assoc. Comp. Mach. Vol. 9, pp. 84-97, 1962.
- (39) K.C. Tam and V. Perez-Mendez, "Limited Angle Three Dimensional Reconstructions Using Fourier Transform Iterations and Radon Transform Iterations", Opt. Eng., Vol. 20, pp. 586-589, 1981.
- (40) R.W. Gerchberg, "Super Resolutions Through Energy Error Reduction", Opt. Acta, Vol. 21, pp. 709-720, 1974.
- (41) A. Papoulis, "The Fourier Integral and Its Applications", McGraw-Hill, New York, 1962, p. 44.
- (42) E.J. Hoffman, M.E. Phelps, and S.C. Huang, "Performance Evaluation of a Positron Tomograph Designed for Brain Imaging", Department of Radiological Sciences, UCLA, School of Medicine, Report, 1982.
- (43) T.F. Budinger, S.E. Derenzo, R.H. Buesman, and J.L. Cahoon, "Positron Emission Tomography: Instrumentation Perspectives", in Proceedings of the "Int. Workshop on Physics and Engineering in Medical Physics", March 15-18, 1982, Pacific Grove, California, IEEE Computer Society Press, pp. 3-13, 1982.

This report was done with support from the Department of Energy. Any conclusions or opinions expressed in this report represent solely those of the author(s) and not necessarily those of The Regents of the University of California, the Lawrence Berkeley Laboratory or the Department of Energy.

Reference to a company or product name does not imply approval or recommendation of the product by the University of California or the U.S. Department of Energy to the exclusion of others that may be suitable.

TECHNICAL INFORMATION DEPARTMENT
LAWRENCE BERKELEY LABORATORY
UNIVERSITY OF CALIFORNIA
BERKELEY, CALIFORNIA 94720