SRI International

Technical Progress Report • February 1993

NOVEL MASS SPECTROMETRIC INSTRUMENT FOR GASEOUS AND PARTICULATE CHARACTERIZATION AND MONITORING

Covering the period from September 25, 1992 through December 31, 1992

Prepared by:

Michael J. Coggiola

SRI Project 4019

Prepared for:

U. S. Department of Energy
Morgantown Energy Technology Center
ATTN: Contractor Reports Receipt Coordinator
Mail Stop F07
P. O. Box 880
Morgantown, WV 26507-0880

Contract No. DE-AC-21-92MC29116

MASTER

DISTRIBUTION OF THIS DOCUMENT IS UNLIMITED

333 Ravenswood Avenue • Menlo Park, CA 94025-3493 • (415) 326-6200 • FAX: (415) 326-5512 • Telex: 334486

CONTENTS

EXECUTIVE SUMMARY	1
INTRODUCTION	3
BACKGROUND.....	4
RESULTS AND DISCUSSION.....	7
CONCLUSIONS.....	15

LIST OF FIGURES

Figure 1.	Total ion chromatogram for the volatilization of powdered cadmium from a filament within the ion trap	11
Figure 2.	Mass spectrum recorded during the volatilization of the powdered cadmium showing the measured isotopic distribution	12
Figure 3.	Total ion chromatogram for the volatilization of powdered zinc from a filament within the ion trap.....	13
Figure 4.	Mass spectrum recorded during the volatilization of the powdered zinc showing the measured isotopic distribution	14

LIST OF TABLES

Table 1.	Comparison of measured and standard isotopic abundances for cadmium	10
Table 2.	Comparison of measured and standard isotopic abundances for zinc	10

EXECUTIVE SUMMARY

Under contract DE-AC21-92MC29116, SRI International will develop a unique new instrument that will be capable of providing real-time (<1 minute), quantitative, chemical characterization of gaseous and particulate pollutants generated from DOE waste cleanup activities. The instrument will be capable of detecting and identifying volatile organic compounds, polynuclear aromatic hydrocarbons, heavy metals, and transuranic species released during waste cleanup activities. The instrument will be unique in its ability to detect and quantify in real-time these diverse pollutants in both vapor and particulate form.

The instrument to be developed under this program will consist of several major components: (1) an isokinetic sampler capable of operating over a wide range of temperatures (up to 500 K) and flow rates; (2) a high pressure to low pressure transition and sampling region that efficiently separates particles from vapor-phase components for separate, parallel analyses; (3) two small mass spectrometers, one optimized for organic analysis using a unique field ionization source and one optimized for particulate characterization using thermal pyrolysis and electron-impact ionization (EI); and (4) a powerful personal computer for control and data acquisition.

The function of the sampler is to obtain a sufficient and representative quantity of air pollutants from a wide range of waste cleanup activities, both vapor and particulate. In addition, the sampler must deliver the vapors and particles to the detection instrument with minimal material loss. The pressure transition region serves multiple functions: (1) it focuses the particles into a "beam" toward the center of the gas flow by momentum separation, (2) it efficiently separates the particulate beam from the vapor stream, allowing independent chemical analyses of particles and vapors, and (3) it provides the first stage of pressure reduction between the ambient atmosphere and the lower pressure (1-2 Pa) region of an ion trap mass spectrometer.

Chemical analysis of the volatile organic species and the polynuclear aromatic hydrocarbons is typically accomplished by first enriching their concentration using a dimethylsilicone rubber membrane separator and then analyzing their composition with field ionization mass spectrometry (FIMS). FIMS is a soft ionization method that is ideally suited to the chemical characterization of complex mixtures of organic compounds. Also available will be an alternative direct inlet of vapors bypassing the membrane and use of EI, for comparisons with the membrane inlet and field ionization. Use of EI will permit monitoring of various small inorganic molecules as well, such as NO_x and SO_x .

Use of a direct sampling method and differential vacuum pumping allows particulate pollutants to be readily introduced into another ion trap mass spectrometer. Volatilization of particles on a heated surface within the volume of the ion trap followed by the complete and quantitative mass spectrometric analysis of the resulting vapors will yield the detailed information required to monitor waste cleanup activity. Based on conservative estimates of the instrument design and performance parameters, a single 0.1- μm -diameter particle will provide a signal level sufficient for complete analysis. All these operations are under the control of the data acquisition computer.

The overall goal of this effort is to develop and demonstrate an instrument that can provide real-time, part-per-billion detection sensitivity for important pollutant species generated by DOE waste cleanup activities. Initially, the instrument will be developed for targeted use in conjunction with the K-1435 Toxic Substances Control Act (TSCA) incinerator at the Oak Ridge National Laboratory K-25 site. Ultimately, the instrument will be designed to operate in the field at any cleanup site, located close to the stack or process vent, providing the plant operations personnel with real-time information and alarm capabilities. In addition, this instrument will be very broadly applicable for cleanup or sampling, for example, any time contaminated soil is moved or disturbed. Because of the versatility and analytical power of this instrument, commercialization for a wide variety of applications in monitoring and control will be important.

The instrument relies on both commercially available technology, including a simple laser scattering particle detector, an ion trap mass spectrometer and a membrane permeator, and unique technologies developed by our organization in the area of field ionization mass spectrometry, and high-temperature sampling. These technologies are now at a Maturity Level of II. The instrument is projected to be easily transportable with minimal utility requirements, making it ideal for deployment at a wide variety of DOE waste sites. Moreover, the straightforward design and use of proven technologies will result in an instrument of very high reliability and sensitivity that can be ready for actual field testing and use within two to three years.

INTRODUCTION

The goals of the Office of Technology Development as outlined by the director of the Environmental Restoration and Waste Management Program (EM), Leo Duffy, in 1991 appropriations hearings include conducting an aggressive technology development program for waste management, waste minimization, waste treatment, storage and disposal of waste. This will be done through the development of more effective remediation technology to reduce occupational and public exposure. We are developing a method to monitor airborne emissions from the TSCA plant at Oak Ridge which will provide real-time, accurate, and inexpensive data on the emission of hazardous organic and inorganic chemicals in both gaseous and particulate form, and to assist DOE in maintaining its ambitious schedules and overcoming significant scientific limitations of its current monitoring technologies.

In May 1988 under the Comprehensive Environmental Response, Compensation, and Liability Act (CERCLA), Oak Ridge National Laboratory (ORNL) established a timetable remedial investigation/feasibility study, and a Federal Facility Agreement is being negotiated with the Environmental Protection Agency and the Tennessee Department of Health and Environment. As a result, work has begun on inactive waste sites at ORNL to excavate and isolate wastes in compliance with current standards. In addition, ORNL must comply with the Resource Conservation and Recovery Act (RCRA), which regulates generation, transportation, treatment, storage, and disposal of hazardous wastes and corrective action of releases to the environment of hazardous wastes from active facilities.

Environmental restoration activities at DOE-ORO sites and facilities, such as K-25 and Y-12, are conducted to meet location, contaminant, and action-specific requirements imposed by statutes and regulations and to implement guidance and agreements. These activities typically follow a phased approach with public involvement throughout the process.

A key responsibility of the Oak Ridge Operations Office is for the environmental remediation and restoration activities at the Oak Ridge, Y-12 and K-25 facilities. Support of both the Department of Energy's and the State of Tennessee's commitment to environmental issues is a vital concern of the Oak Ridge site's workforce and management structure. Through advances in technology, the environmental remediation and restoration efforts may be accomplished at less cost, and with a decreased impact on the surrounding communities.

BACKGROUND

The Oak Ridge K-25 site is located 13 miles west of Oak Ridge. Until the summer of 1985, when the gaseous diffusion process was shut down, the mission of the K-25 site was ^{235}U enrichment of uranium hexafluoride (UF_6) for eventual use as a fuel in nuclear reactors. Typical facilities included the feed system and separate systems for collecting the product and the waste (tails) stream. The purging of contaminants from the product stream required the use of large amounts of alumina and sodium fluoride. A number of compounds were used for the clarification of coolant water, and hexavalent chromium, zinc, and phosphate were used to inhibit corrosion of heat transfer equipment. Support facilities produced solutions that required concentration and recovery with some residual that was discarded and buried. Laboratory sample residues and obsolete chemical reagents, spent chemicals, and contaminated equipment were candidates for discard. Sludges from chemical wastes contain hazardous materials, including chromium, copper, lead, mercury, naphthalene, and zinc. Wastes buried at the K-25 site include low-level radioactive solid waste, mixed chemical waste, and radioactive and nonradioactive classified materials. Uranium was released into the air, surface water, and on-site land disposal.

Waste management activities at the K-25 site are increasing. Low-level radioactive waste from other DOE-Oak Ridge Operations (ORO) sites are now being placed in interim storage facilities in the K-25 site building vaults until final disposition strategy is identified. Also, polychlorinated biphenyl (PCB) wastes began arriving from other DOE-ORO sites in 1987 for future incineration in the new K-1435 Toxic Substances Control Act incinerator.

Making an analogy, it is well established that heavy metals are present in the stack emissions of coal-fired power plants. Considerable research in this area has identified every naturally occurring element in coal fly ash, with concentrations ranging from percents to parts-per-million and below. In the case of coal combustion, these heavy metals are present in the feedstock, whereas in the process stream of a waste cleanup facility, many heavy metals will be present at high concentrations as part of the original waste material. Although a number of the naturally occurring elements are of minor concern, many of the heavy metals, such as mercury, lead, cadmium, arsenic, chromium, and uranium, pose significant health risks to humans. Additionally, the transuranic elements such as plutonium, are of particular concern due to their extreme toxicity as respirable particles. Many other elements, both radioactive and nonradioactive, also pose health and safety risks. Included in this category are particles containing beryllium, iodine, and cesium.

By analogy with the known distribution of heavy metals within airborne particles generated by coal-fired power plants, it is reasonable to conclude that significant levels of heavy metals and hazardous inorganic contamination can be unknowingly discharged from process vent stacks in the particulate form. To maintain a high degree of safety and control over cleanup activities at waste sites containing these materials, a monitoring instrument is required that can provide real-time, quantitative measurements of levels of important elements at any point within the process. Particular attention must be given to effluent discharge to the atmosphere via vents and stacks to prevent any release of hazardous by-products of the cleanup activities.

In addition to the heavy metal and inorganic emissions, a very wide range of VOCs will be produced and emitted from any waste cleanup plants, including PAHs, heterocyclic nitrogen and sulfur species, and polychlorinated biphenyls (PCBs). Many of these VOCs are known or suspected carcinogens whose levels must also be controlled. Unlike the metals and inorganic species, many of the organic compounds found in the gaseous effluent of waste cleanup facilities will include materials not initially present in the waste stream. For example, many PAHs are formed during the incomplete combustion of simple organic wastes. As with metals and inorganics, the organic pollutants can be released in both the vapor and particulate forms. Thus, a versatile real-time instrument for determining VOC levels is required to allow plant operators to monitor and regulate stack emissions of these hazardous species.

Any real-time waste cleanup process monitor should be capable of simultaneously measuring numerous important chemical species, both organic and inorganic, present as vapors and particles. Although very sophisticated monitoring technologies are currently under development that make use of optical, electrochemical, and microsensor methods, none yet offers quantitative chemical characterization of both vapor phase and particulate air pollutants. Moreover, most of these sophisticated sensors are capable of detecting only one or two specific chemical species, thereby requiring that many separate sensors be integrated to perform the necessary monitoring and control functions. Such multisensor systems will often be technologically more complex and less reliable than a single, multifunctional instrument, as well as more costly to develop, manufacture, operate, and maintain.

The most universal analytical technique available for monitoring a broad range of chemical species is mass spectrometry. If mass spectrometry is to be used, however, the sampled materials must be vaporized and ionized before analysis. The ionization method must be carefully chosen to prevent preferential, or nonstatistical, ionization while providing adequate and stable signal levels. Even before the effluent can be ionized, it must be sampled and transported from the stack or process stream to the ionizer. In the case of a waste cleanup plant such as the K-1435 TSCA

incinerator at the DOE-Oak Ridge Operations site, the sampled effluent will consist of vapors and particles. Thus, the development of an appropriate sampler will be a critical aspect of this effort. Most of the heavy metals, for example, will not appear in vapor form, but will be incorporated into particles of varying sizes. Not only must these particles be efficiently sampled, they must also be separated from the vapor component and then completely vaporized to provide the requisite composition information. At the same time, the sampled vapors, including the organic species of interest, must be isolated and efficiently detected. No single *instrumental* technique (FTIR, ICP-AES, ICP-MS, GC, GC-MS, LIF) is capable of simultaneously detecting both organic and inorganic species in vapor and particulate forms. However, a single *instrument* consisting of two closely integrated detectors that share a common sampler can provide this critical analytical capability.

RESULTS AND DISCUSSION

The goals of this research program are to develop, demonstrate, and field test a real-time monitoring instrument capable of providing quantitative chemical information on the vapor and particulate emissions produced during the incineration of toxic waste. The Phase I effort is divided into five main tasks and subtasks as follows:

Task 1: NEPA Documentation

Task 2: Instrument Development

Task 2.1: Sampler Development

Task 2.2: Vapor Instrument Development

Task 2.3: Particulate Instrument Development

Task 3: Instrument Testing

Task 3.1: Single Metal Species Testing

Task 3.2: Multiple Metal Species Testing

Task 4: System Design

Task 5: Project Management

During the initial period of performance, the primary effort was directed toward providing the DOE with the requisite information relating to the NEPA documentation. In addition, a number of reports and project initiation documents were prepared in accordance with the requirements of the contract. Because no laboratory effort could be undertaken prior to completion and approval of the NEPA documentation, only a limited amount of effort was expended on the project, primarily in the form of more extensive literature surveys, and project startup meetings within SRI.

On November 25, 1992, a letter from Randolph L. Kesling, Contracting Officer, Morgantown Energy Technology Center, verified that this project was granted a categorical exclusion for the Phase I effort under the DOE National Environmental Policy Act. A project startup meeting was therefore scheduled for December 15, 1992, at Morgantown.

At the project kickoff meeting, both Michael J. Coggiola, SRI's project leader; and Clyde Witham, leader of the sampler development task were present. DOE representatives included William Huber, the COTR from DOE/METC, Ronald Roth, DOE/METC contracts, and Al Tardiff

from DOE headquarters. Additional DOE/METC technical personnel attended as well. Although a technical representative from ORNL was scheduled to attend, they were unable to do so.

Bill Huber began the meeting with introductions, followed by an overview of METC presented by Bob Bedick. Mike Coggiola then reviewed the technical aspects of the project, including background information, goals, and schedule. Ron Roth reviewed the contractual details relating to the phased acquisition of this program. During the open discussion period, Al Tardiff from DOE headquarters described several parallel, related projects currently underway at other DOE facilities with which we will need to coordinate our efforts to avoid duplication. It was further agreed that a visit to the TSCA incinerator site at ORNL was necessary for both SRI and DOE representatives. No formal plans for this trip were made during the meeting, however, it was agreed to visit ORNL early in the year.

The week following the kickoff meeting at METC, the principal task leaders met at SRI to allocate resources between the tasks (in the form of establishing project subaccounts), and to begin assigning and scheduling personnel. SRI was closed between December 24, 1992 and January 3, 1993, and work on tasks 2.1, 2.2, and 2.3 began in early 1993.

During the month of January, the primary effort was devoted to initiating the design for a particulate sampler under task 2.1. Previously used particulate samplers were examined for their adaptability to this project, and details regarding the specific sampling conditions that will prevail at the ORNL test site were obtained. Although the exact sampling requirements (particle size, gas flows, temperatures, physical sizes and limitations, etc.) are not as yet clearly defined, a preliminary sampler design has been started based on the most likely operating conditions. This design is also being integrated with the instrument design to ensure proper sample transport, to minimize losses, and to satisfy the mass spectrometers vacuum and gas load restrictions. Because the sampler design is flexible with regard to gas flows, orifice sizes and transport distances, modifications can easily be made as detailed conditions become available. In any case, the Phase I sampler design may need to be modified slightly from the field prototype to accommodate the laboratory test facilities that will be used at SRI.

Development of the vapor instrument under task 2.2 is actually broader in scope because it involves adapting the ion trap mass spectrometer to operate in larger, more accessible vacuum system that will be used for both vapor and particulate testing. Using many existing vacuum components, including turbomolecular pumped vacuum chambers and suitable vacuum feedthrus for electrical and sample connections, one of our laboratory ion trap detectors is currently being modified and installed in the test chamber. Initially, the ion trap will be tested using directly introduced vapor samples and conventional electron impact ionization to verify that it operates

correctly in the new vacuum housing. Following those tests, modifications will begin to incorporate one of SRI's volcano field ionization devices as an external ionization source.

The major effort under task 2.3 is to design, modify, and test another ion trap detector to operate with an internal pyrolysis filament for particulate analysis. This second ion trap will be installed in the same vacuum chamber as the vapor instrument described above, and it will share the same electronics and computer system. During this first phase project, the two ion traps will not be operable simultaneously, so vapor and particulate tests must be done sequentially rather than concurrently. We have initiated a design study for the heated pyrolysis filament to determine its optimum size, shape, material, and most critically, its placement within the ion trap volume. As a first step, we have fabricated a simple filament loop from 0.013 cm diameter Chromel wire and inserted it into one of the existing 0.11 cm diameter holes in the exit end cap of the standard ITD. This arrangement allows us to investigate not only the effects of thermal pyrolysis of materials preloaded onto the filament, but also the electrical behavior of the ion trap with this perturbation present.

Our first tests involved placing small amounts of either cadmium or zinc on the wire loop, then slowly heating the wire while recording mass spectra. Figure 1 shows a typical total ion current (TIC) recorded as a function of time (and therefore as a function of temperature since the filament was heated linearly with time) for powdered cadmium. The single sharp peak at about 2 minutes is due entirely to the volatilization of cadmium from the filament. Examination of the mass spectra recorded during this vapor evolution (Figure 2) clearly shows that all eight cadmium isotopes were readily detected. Moreover, the isotopic ratios listed in Table I agree reasonably well with the standard values. Similar results were obtained for zinc, as shown in Figures 3 and 4 and Table II.

While the use of simple wire loop is useful to establish some characteristics, the actual filament must be significantly larger to intercept the beam of particulates from the sampler. It must also be designed such that particles do not simply bounce off of the filament without being pyrolyzed. Our initial design will focus on a Woods-horn shape, probably fabricated from tantalum foil. Once built, this filament will be tested to determine its heating and cooling characteristics, and its particle collection/accommodation properties. Eventually, it will be installed in the ion trap with suitable modifications, and tested.

Table 1. Comparison of measured and standard isotopic abundances for cadmium.

Mass	Measured	Standard
106	0.52	1.25
108	0.52	0.89
110	12.3	12.49
111	13.0	12.8
112	25.2	24.13
113	12.6	12.22
114	28.6	28.73
116	7.22	7.49

Table 2. Comparison of measured and standard isotopic abundances for zinc.

Mass	Measured	Standard
64	48.4	48.63
66	28.0	27.9
67	4.9	4.1
68	18.7	18.75

Comment: POWDERED Cd ON RHENIUM FIL. 0 TO 1 A HE=2.0 TURN
Scan Range: 1 - 300 Scan: 1 Int = 930 0 0:00

Figure 1. Total ion chromatogram for the volatilization of powdered cadmium from a filament within the ion trap.

Comment: POWDERED Cd ON RHENIUM FIL. 0 TO 1 A HE=2.0 TURN
Average of: 119 to 123 Minus: 105 to 113

Figure 2. Mass spectrum recorded during the volatilization of the powdered cadmium showing the measured isotopic distribution.

Comment: POWDERED Zn ON RHENIUM FIL. 0 TO 1 A HE=1.5 TURN
Scan Range: 1 - 233 Scan: 1 Int = 1653 @ 0:01

Figure 3. Total ion chromatogram for the volatilization of powdered zinc from a filament within the ion trap.

Comment: POWDERED Zn ON RHENIUM FIL. 0 TO 1 A HE=1.5 TURN
Average of: 77 to 80 Minus: 64 to 71

Figure 4. Mass spectrum recorded during the volatilization of the powdered zinc showing the measured isotopic distribution.

CONCLUSION

At the completion of the first three months of this research effort, progress is on schedule as proposed in both the original research proposal and the Management Plan. The main critical path milestone during the first quarter was the granting of a NEPA exclusion, thus allowing initiation of the laboratory research effort. SRI has allocated the appropriate technical staff to begin work on tasks 2.1, 2.2, and 2.3. Weekly staff meetings are attended by all personnel so that intertask communication and cooperation is maximized.

During the second quarter of this program, it is anticipated that preliminary designs will be completed for the sampler and its vacuum interface with the ion trap detectors. In addition, at least one of the ion trap detectors will be installed in the vacuum test stand, and its operation verified. Considerable progress will also be made toward designing and fabrication of the pyrolysis filament required for particulate analysis.

END

DATE
FILMED

7 / 15 / 93