Assessment Report for Hanford Analytical Services Quality Assurance Plan

L. H. Taylor

Date Published
November 1994

Prepared for the U.S. Department of Energy
Office of Environmental Restoration and
Waste Management

Westinghouse
Hanford Company

P.O. Box 1970
Richland, Washington

Hanford Operations and Engineering Contractor for the
U.S. Department of Energy under Contract DE-AC06-87RL10930

Approved for Public Release

MASTER

DISTRIBUTION OF THIS DOCUMENT IS UNLIMITED

jm

LEGAL DISCLAIMER

This report was prepared as an account of work sponsored by an agency of the United States Government. Neither the United States Government nor any agency thereof, nor any of their employees, nor any of their contractors, subcontractors or their employees, makes any warranty, express or implied, or assumes any legal liability or responsibility for the accuracy, completeness, or any third party's use or the results of such use of any information, apparatus, product, or process disclosed, or represents that its use would not infringe privately owned rights. Reference herein to any specific commercial product, process, or service by trade name, trademark, manufacturer, or otherwise, does not necessarily constitute or imply its endorsement, recommendation, or favoring by the United States Government or any agency thereof or its contractors or subcontractors. The views and opinions of authors expressed herein do not necessarily state or reflect those of the United States Government or any agency thereof.

This report has been reproduced from the best available copy.
Available in paper copy and microfiche.

Available to the U.S. Department of Energy
and its contractors from
Office of Scientific and Technical Information
P.O. Box 62
Oak Ridge, TN 37831
(615) 576-8401

Available to the public from the U.S. Department of Commerce
National Technical Information Service
5285 Port Royal Road
Springfield, VA 22161
(703) 487-4650

Printed in the United States of America

DISCLM-1.CHP (1-91)

RELEASE AUTHORIZATION

Document Number: WHC-EP-0822

Document Title: Assessment Report for Hanford Analytical Services
Quality Assurance Plan

Release Date: November 18, 1994

**This document was reviewed following the
procedures described in WHC-CM-3-4 and is:**

APPROVED FOR PUBLIC RELEASE

WHC Information Release Administration Specialist:

Kara Broz

November 18, 1994

DISCLAIMER

Portions of this document may be illegible in electronic image products. Images are produced from the best available original document.

**ASSESSMENT REPORT
for
HANFORD ANALYTICAL SERVICES
QUALITY ASSURANCE PLAN (HASQAP)**

This report documents the assessment results of DOE/RL-94-55, *Hanford Analytical Services Quality Assurance Plan*, against the following DOE Orders:

- 5700.6C, *Quality Assurance*
- 5480.19, *Conduct of Operations*
- 5480.20, *Training*

The assessment was conducted using the Requirement and Self-Assessment Database (RSAD), which contains mandatory and non-mandatory DOE Order statements for the above Orders.

The results of the assessment are summarized for each of the three Orders vs. the HASQAP in Attachments A1, A2, and A3.

For the assessment of DOE 5700.6C, the HASQAP was evaluated for compliance against those statements designated as "mandatory" for contractors in the RSAD database. See Attachment B for the Assessment Data Collection (ADC) Forms for the RSAD statements evaluated.

For the assessment of DOE 5480.19, both "mandatory" and "non-mandatory" statements in the RSAD database were addressed on the Assessment Data Collection Forms because all but two of the statements were designated as "non-mandatory" for contractors.

However, DOE 5480.19 (Conduct of Operations) was written to provide guidance for facility operations primarily for DOE production and test facilities. The difficulty of correlating much of the guidance in the Order to analytical laboratory services resulted in the determination of many of the areas and activities addressed by the Order as being not applicable to analytical laboratory services.

Where correlations to analytical laboratory activities could be made, the HASQAP was assessed to those parts of the Order. Due to the heavy emphasis and slant toward "operations" of production and test facilities, it was necessary to make the following interpretations or equivalencies to assess the intent of the guidance as applied to analytical laboratory activities:

- (1) "operations" and
"operations activities" $\Rightarrow$ "analytical laboratory
activities"
- (2) "operator" $\Rightarrow$ "laboratory analyst"
- (3) "operations
supervisor/manager" $\Rightarrow$ "laboratory
supervisor/manager"

See Attachment C for the ADC forms for the DOE 5480.19 RSAD statements evaluated.

For the assessment of DOE 5480.20, the HASQAP was evaluated for compliance against those statements designated as "mandatory" for contractors in the RSAD database. See Attachment D for the (ADC) Forms for the RSAD statements evaluated.

WHC-EP-0822
Attachment A1
DOE 5700.6C vs. HASQAP

FINDINGS

The HASQAP was found to not adequately comply with nine out of the twenty-five requirement statements assessed. See Attachment B of this report for the Assessment Data Collection forms that provide the details of the noncompliances. The nine noncompliances are summarized as follows:

- (1) 4 noncompliances concern DOE administrative functions (e.g. approval, concurrence, responsibilities).
- (2) 2 noncompliances address DOE mandates concerning records disposition (i.e. DOE 1324.2A and NARA GRS).
- (3) 3 noncompliances deal with requirement statements not being adequately addressed. 2 of the noncompliances were specific to the procurement area, and 1 was specific to management systems.

RECOMMENDATIONS

- (1) Determine the need or appropriateness to include or reference in the HASQAP the statements concerning DOE administrative functions.
- (2) Revise the HASQAP to reference DOE records disposition requirement documents.
- (3) Revise HASQAP to address the 3 noncompliances concerning procurement and management systems.

CONCLUSIONS

The HASQAP is in substantial compliance with DOE 5700.6C. The findings identified are:

- (1) judged to be minor,
- (2) primarily of an administrative nature,
- (3) easily addressable and correctable, and
- (4) not indicative of any programmatic weaknesses.

The Comparison Matrix on page 2 of this attachment illustrates the overall compliance of the HASQAP with the provisions of DOE 5700.6C.

WHC-EP-0822
Attachment A1
DOE 5700.6C vs. HASQAP

5700.6C

	a. General	1. Program	2. Personnel Training and	3. Quality Improvement	4. Documents and	5. Work	6. Design	7. Procurement	8. Inspection and Acceptance	9. Management Assessment	10. Independent
HASQAP											
Policy Statement		Y									
1.0 Introduction	Y										
2.0 Organization and Responsibilities	N	Y/N									
3.0 Personnel Qualification and Training			Y								
4.0 QA Objectives									Y		
5.0 Systems Quality Assurance					Y	Y	Y				
6.0 Sample Custody and Handling						Y					
7.0 Calibration						Y			Y		
8.0 Laboratory Procedures					Y	Y					
9.0 Data Collection, Reduction and Reporting						+					
10.0 Records	N				Y						
11.0 Quality Control						+					
12.0 Procedures to Assess Data Quality						+					
13.0 Audits	Y									Y	Y
14.0 Preventive Maintenance						Y			Y		
15.0 Corrective Action				Y					Y		
16.0 Quality Assurance Reporting				Y							
17.0 Data Assessment and Validation						+					
18.0 Procurement Control						Y		Y			

KEY: Y = Meets the Order requirements.
 N = Does not meet the Order requirements.
 + = Exceeds the Order requirements or provides more detail.
 X = Conflicting requirements.
 Y/N = Partially meets the Order requirements.

WHC-EP-0822
Attachment A2
DOE 5480.19 vs. HASQAP

FINDINGS

The HASQAP was found to not adequately comply with forty-one of the eighty-three requirement statements assessed. See Attachment C of this report for the Assessment Data Collection forms that provide the details of the noncompliances. The forty-one noncompliances are summarized as follows:

- (1) 8 noncompliances concern specific personnel protection requirements that are not detailed in the HASQAP.
- (2) 11 noncompliances address requirements for the investigation of abnormal events that are not detailed in the HASQAP.
- (3) 7 noncompliances deal with requirements for shift logs that are not detailed in the HASQAP.
- (4) 5 noncompliances concern requirements for monitoring operating performance that are not detailed in the HASQAP.
- (5) 3 noncompliances concern operating practices (safety planning, safety practices, and status practices) that are not detailed in the HASQAP.
- (6) 3 noncompliances concern notification requirements that are not detailed in the HASQAP.
- (7) 2 noncompliances concern use of the graded approach and an implementation matrix for the Order that are not detailed in the HASQAP.
- (8) 1 noncompliance requires physical security to DOE 5630.11 that is not covered in the HASQAP.
- (9) 1 noncompliance concerns periodic reviews of procedures that are not detailed in the HASQAP.

WHC-EP-0822
Attachment A2
DOE 5480.19 vs. HASQAP

RECOMMENDATIONS

Using the graded approach:

- (1) Incorporate the applicable DOE 5480.19 requirements (or their intent as applied to analytical laboratory services) into the HASQAP

OR

- (2) Develop an implementation listing or matrix for inclusion in the HASQAP that specifies those DOE 5480.19 sections/requirement statements (or their intent as applied to analytical laboratory services) that laboratory management must implement.

CONCLUSIONS

The large number of noncompliances may be somewhat misleading due to the assessment's methodology of conservatively trying to correlate the requirement statements of DOE 5480.19 to analytical laboratory operations.

Nevertheless, in order to bring the HASQAP into compliance with the applicable requirements of DOE 5480.19, one of the recommendations above should be considered.

The Comparison Matrix on page 3 of this attachment is indicative of the relative incompatibility of Order 5480.19 requirements written for production or test facilities and the HASQAP written for analytical laboratory services.

WHC-EP-0822
Attachment A2
DOE 5480.19 vs. HASQAP

5480.19

	5. Requirements	I. Ops Organization and Administration	II. Shift Routines and Operating Practices	V. Control of On-shift	VI. Investigation of Abnormal	VII. Notifications	VIII. Control of Equipment	XI. Logkeeping	XVI. Operations
HASQAP									
Policy Statement		Y							
1.0 Introduction	N								
2.0 Organization and Responsibilities	N	Y/N	N					N	
3.0 Personnel Qualification and Training				Y					
4.0 QA Objectives									
5.0 Systems Quality Assurance									
6.0 Sample Custody and Handling									
7.0 Calibration							Y		
8.0 Laboratory Procedures									Y
9.0 Data Collection, Reduction and Reporting						Y/N			
10.0 Records									
11.0 Quality Control									
12.0 Procedures to Assess Data Quality									
13.0 Audits									
14.0 Preventive Maintenance									
15.0 Corrective Action					Y/N				
16.0 Quality Assurance Reporting									
17.0 Data Assessment and Validation									
18.0 Procurement Control									

WHC-EP-0822
Attachment A2
DOE 5480.19 vs. HASQAP

KEY: Y = Meets the Order requirements.
N = Does not meet the Order requirements.
+ = Exceeds the Order requirements.
X = Conflicting requirements.
Y/N = Partially meets the Order requirements.

WHC-EP-0822
Attachment A3
DOE 5480.20 vs. HASQAP

FINDINGS

The HASQAP was found to not adequately comply with twenty-eight out of the forty-nine requirement statements assessed. See Attachment D of this report for the Assessment Data Collection forms that provide the details of the noncompliances. The twenty-eight noncompliances are summarized as follows:

- (1) 4 noncompliances concern specific GET training requirements that are not detailed in the HASQAP.
- (2) 3 noncompliances address specific management and supervisor training requirements that are not detailed in the HASQAP.
- (3) 5 noncompliances deal with specific entry-level requirements for specific job classifications that are not detailed in the HASQAP.
- (4) 8 noncompliances concern specifics for exceptions or exemptions to training requirements, extensions to training requirements, and alternatives to educational and experience requirements that are not detailed in the HASQAP.
- (5) 2 noncompliances concern specific requirements for records that are not detailed in the HASQAP.
- (6) 3 noncompliances concern specific time limits for maintaining qualifications, re-certifying, and continuing training that are not detailed in the HASQAP.
- (7) 2 noncompliances concern specific requirements for PRA training and safety-related systems training for technicians that are not detailed in the HASQAP.
- (8) 1 noncompliance concerns a Training Implementation Matrix for Order 5480.20 that is not detailed in the HASQAP.

RECOMMENDATIONS

- (1) Incorporate the applicable requirements of DOE 5480.20 into the HASQAP
- OR
- (2) Develop an implementation matrix or listing for inclusion in the HASQAP that specifies those sections/ requirement statements of DOE 5480.20 that laboratory management must implement.

CONCLUSIONS

Based on the number of noncompliances with DOE 5480.20, one might conclude that the HASQAP is seriously deficient in this area. However, The HASQAP's section on Personnel Qualification and Training is fairly

WHC-EP-0822
Attachment A3
DOE 5480.20 vs. HASQAP

comprehensive; but the HASQAP's level of specifics and detail for personnel selection, qualification, and training is substantially less than that delineated in DOE 5480.20.

In order to bring the HASQAP into compliance with the applicable specifics of DOE 5480.20, one of the recommendations above should be considered.

The Comparison Matrix on page 3 of this attachment is indicative of the difference in the level of detail between the two documents.

WHC-EP-0822
Attachment A3
DOE 5480.20 vs. HASQAP

5480.20

	I.3 Subcontractor Personnel	I.4 Personnel Selection Requirements	I.5 Qualification Process	I.6 Certification Process	I.7 Training	I.11 Exceptions to Training	I.12 Extension	I.13 Alternatives to Educational	I.14 Alternatives to Experience	I.16 Record	IV.2 Entry Level	IV.4 Staffing
HASQAP												
Policy Statement												
1.0 Introduction												
2.0 Organization and Responsibilities												
3.0 Personnel Qualification and Training	Y	Y/N	Y/N	Y/N	Y/N	Y/N	N	N	N	Y/N	N	N
4.0 QA Objectives												
5.0 Systems Quality Assurance												
6.0 Sample Custody and Handling												
7.0 Calibration												
8.0 Laboratory Procedures												
9.0 Data Collection, Reduction and Reporting												
10.0 Records												
11.0 Quality Control												
12.0 Procedures to Assess Data Quality												
13.0 Audits												
14.0 Preventive Maintenance												
15.0 Corrective Action												
16.0 Quality Assurance Reporting												
17.0 Data Assessment and Validation												
18.0 Procurement Control	Y											

WHC-EP-0822
Attachment A3
DOE 5480.20 vs. HASQAP

KEY: Y = Meets the Order requirements.
N = Does not meet the Order requirements.
+ = Exceeds the Order requirements or provides more
detail.
X = Conflicting requirements.
Y/N = Partially meets the Order requirements.

ATTACHMENT B

ASSESSMENT DATA COLLECTION FORMS

for

HASQAP ASSESSMENT

to

DOE ORDER 5700.6C
(Quality Assurance)

ASSESSMENT DATA COLLECTION FORM

Statement Number:	9.a.(1)	Mandatory (Y/N):	Y
<p>Statement:</p> <p>REQUIREMENTS. General. Senior management shall be responsible for Quality Assurance Program (QAP) implementation, assessment, and improvement. Departmental Elements and M&O contractors shall develop their QAPs by applying the quality assurance criteria specified in subparagraph 9b. The quality assurance program shall include a discussion of how the criteria will be satisfied taking into consideration the risk associated with the work. Appropriate standards shall be used, wherever applicable, to develop and implement QAPs.</p>			

1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): YES

2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):

HASQAP:

1.0 Introduction

This document, the *Hanford Analytical Services Quality Assurance Plan* (HASQAP), is issued by the U.S. Department of Energy, Richland Operations Office (RL). The HASQAP establishes quality requirements in response to U.S. Department of Energy (DOE) Order 5700.6C, *Quality Assurance* (10 CFR 830.120, "Quality Assurance Requirements"). The HASQAP is designed to meet the needs of the RL for controlling the analytical chemistry services provided by laboratory operations.

2.2 Structure, Responsibility, and Authority

Senior management shall be responsible for establishing the scope, implementing, assessing, and continually improving an effective quality system.

3.0 COMMENTS:

The concept of risk and/or graded-approach is not discussed in the HASQAP. This may be appropriate if the HASQAP is intended to apply only to regulatory-based analytical services.

ASSESSMENT DATA COLLECTION FORM

Statement Number:	9.a.(3)	Mandatory (Y/N):	Y
<p>Statement:</p> <p>REQUIREMENTS. General. Contractors shall obtain DOE approval of new QAPs prior to commencing work. Contractors working to existing quality assurance programs shall ensure that their QAPs meet the requirements of this Order and shall re-submit their QAPs, together with an implementation plan, to the Lead Program Secretarial Officers (PSOs) for approval within 180 days after this Order becomes effective. Quality assurance programs shall be regarded as approved by DOE 180 days after submittal, including any modifications which have been made or directed by DOE during this period, or receipt of a letter, whichever occurs first.</p>			

- 1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): NO
- 2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):
- Requirement is not addressed in the HASQAP.
- 3.0 CORRECTIVE ACTION:

ASSESSMENT DATA COLLECTION FORM

Statement Number:	9.a.(4)	Mandatory (Y/N):	Y
Statement: REQUIREMENTS. General. DOE contractors may, at any time, make changes to their DOE-approved QAPs. Changes made over the previous year to DOE-approved QAPs shall be submitted annually to the Lead PSO for review. All changes shall be regarded as approved by DOE 90 days after submittal, including any modifications which have been made or directed by DOE during this period, or receipt of a letter, whichever occurs first. The submittals shall identify the changes, the pages affected, the reason for the changes, and the basis for concluding the revised program continues to satisfy the requirements of this Order. Changes made to correct spelling, punctuation, or other editorial items do not require explanation.			

- 1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): NO
- 2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):
- Requirement is not addressed in the HASQAP.
- 3.0 CORRECTIVE ACTION:

ASSESSMENT DATA COLLECTION FORM

Statement Number: 9.a.(5)	Mandatory (Y/N): Y
Statement: REQUIREMENTS. General. Assessments of DOE and contractor work shall be conducted based on the DOE-approved quality assurance program.	

1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): YES

2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):

Requirement is met by HASQAP Section 13.0, "Audits."

ASSESSMENT DATA COLLECTION FORM

Statement Number:	9.a.(6)	Mandatory (Y/N):	Y
Statement: REQUIREMENTS. General. Programs, guides, and policies which implement this Order shall meet the requirements of DOE 1324.2A. (The Records and Reports Branch, AD-241.1, is available for any needed assistance regarding records management.)			

- 1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): NO
- 2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):
Requirement is not addressed in the HASQAP.
- 3.0 CORRECTIVE ACTION:

ASSESSMENT DATA COLLECTION FORM

Statement Number:	9.b.(1)(a)	Mandatory (Y/N):	Y
Statement: REQUIREMENTS. Quality Assurance Criteria. Management. Criterion 1--Program. Organizations shall develop, implement, and maintain a written Quality Assurance Program. The QAP shall describe the organizational structure, functional responsibilities, levels of authority, and interfaces for those managing, performing, and assessing adequacy of work. The QAP shall describe the management system, including planning, scheduling, and cost control considerations.			

1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): NO

2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):

Section 2.0, "Organization and Responsibility," adequately addresses most of the statement. However, the following sentence in the statement is not addressed in the HASQAP:

"The QAP shall describe the management system, including planning, scheduling, and cost control considerations."

3.0 CORRECTIVE ACTION:

ASSESSMENT DATA COLLECTION FORM

Statement Number:	9.b.(1)(b)	Mandatory (Y/N):	Y
Statement: REQUIREMENTS. Quality Assurance Criteria. Management. Criterion 2--Personnel Training and Qualification. Personnel shall be trained and qualified to ensure they are capable of performing their assigned work. Personnel shall be provided continuing training to ensure that job proficiency is maintained.			

- 1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): YES
- 2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):
- HASQAP Section 3.0, "Personnel Qualification and Training,"
satisfactorily addresses the statement.

ASSESSMENT DATA COLLECTION FORM

Statement Number: 9.b.(1)(c)	Mandatory (Y/N): Y
<p>Statement:</p> <p>REQUIREMENTS. Quality Assurance Criteria. Management. Criterion 3--Quality Improvement. The organization shall establish and implement processes to detect and prevent quality problems and to ensure quality improvement. Items and processes that do not meet established requirements shall be identified, controlled, and corrected. Correction shall include identifying the causes of problems and preventing recurrence. Item reliability, process implementation, and other quality-related information shall be reviewed and the data analyzed to identify items and processes needing improvement.</p>	

1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): YES

2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):

HASQAP Section 15.0, "Corrective Action," and Section 16.0, "Quality Assurance Reporting," satisfactorily address the statement.

ASSESSMENT DATA COLLECTION FORM

Statement Number:	9.b.(1)(d)	Mandatory (Y/N):	Y
Statement: REQUIREMENTS. Quality Assurance Criteria. Management. Criterion 4--Documents and Records. Documents shall be prepared, reviewed, approved, issued, used, and revised to prescribe processes, specify requirements, or establish design. Records shall be specified, prepared, reviewed, approved, and maintained.			

1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): YES

2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):

The statement is satisfactorily addressed by:

HASQAP subsection 5.2, "Administrative Systems"

HASQAP Section 10.0, "Records"

HASQAP Section 8.0, "Laboratory Procedures"

ASSESSMENT DATA COLLECTION FORM

Statement Number:	9.b.(2)(a)	Mandatory (Y/N):	Y
<p>Statement:</p> <p>REQUIREMENTS. Quality Assurance Criteria. Performance. Criterion 5--Work Processes. Work shall be performed to established technical standards and administrative controls. Work shall be performed under controlled conditions using approved instructions, procedures, or other appropriate means. Items shall be identified and controlled to ensure their proper use. Items shall be maintained to prevent their damage, loss, or deterioration. Equipment used for process monitoring or data collection shall be calibrated and maintained.</p>			

1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): YES

2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):

This statement is met by the following HASQAP sections:

- 8.0 Laboratory Procedures
- 6.0 Sample Custody and Handling
- 5.1 Software Systems
- 5.2 Administrative Systems
- 5.3 Technical Systems
- 5.4 Physical Facilities Systems
- 7.0 Calibration
- 14.0 Preventive Maintenance
- 9.0 Data Collection, Reduction, and Reporting
- 11.0 Quality Control
- 12.0 Procedures to Assess Data Quality
- 17.0 Data Assessment and Validation
- 18.0 Procurement Control

ASSESSMENT DATA COLLECTION FORM

Statement Number:	9.b.(2)(b)	Mandatory (Y/N):	Y
Statement: REQUIREMENTS. Quality Assurance Criteria. Performance. Criterion 6--Design. Items and processes shall be designed using sound engineering/scientific principles and appropriate standards. Design work, including changes, shall incorporate applicable requirements and design bases. Design interfaces shall be identified and controlled. The adequacy of design products shall be verified or validated by individuals or groups other than those who performed the work. Verification and validation work shall be completed before approval and implementation of the design.			

1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): YES

2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):

Classical design activities (i.e. design of physical systems and components) are not applicable to analytical laboratory operations. However, the design of software, administrative, and technical processes is covered under HASQAP Section 5.0, "Systems Quality Assurance."

ASSESSMENT DATA COLLECTION FORM

Statement Number: 9.b.(2)(c)	Mandatory (Y/N): Y
Statement: REQUIREMENTS. Quality Assurance Criteria. Performance. Criterion 7--Procurement. The organization shall ensure that procured items and services meet established requirements and perform as specified. Prospective suppliers shall be evaluated and selected on the basis of specified criteria. The organization shall ensure that approved suppliers can continue to provide acceptable items and services.	

1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): YES

2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):

The statement is met by HASQAP Section 18.0, "Procurement Controls."

ASSESSMENT DATA COLLECTION FORM

Statement Number:	9.b.(2)(d)	Mandatory (Y/N):	Y
Statement: REQUIREMENTS. Quality Assurance Criteria. Performance. Criterion 8--Inspection and Acceptance Testing. Inspection and acceptance testing of specified items and processes shall be conducted using established acceptance and performance criteria. Equipment used for inspections and tests shall be calibrated and maintained.			

1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): YES

2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):

The portion of the statement concerning acceptance and performance criteria is met by HASQAP subsection 4.1, "Data Quality Objectives."

The portion of the statement concerning calibrated and maintained testing equipment is met by HASQAP Section 7.0, "Calibration", and Section 14.0, "Preventive Maintenance."

ASSESSMENT DATA COLLECTION FORM

Statement Number:	9.b.(3)	Mandatory (Y/N):	Y
<p>Statement:</p> <p>REQUIREMENTS. Quality Assurance Criteria. Performance. Assessment. (a) Criterion 9--Management Assessment. Management at all levels shall periodically assess the integrated quality assurance program and its performance. Problems that hinder the organization from achieving its objectives shall be identified and corrected. (b) Criterion 10--Independent Assessment. Planned and periodic independent assessments shall be conducted to measure item quality and process effectiveness and to promote improvement. The organization performing independent assessments shall have sufficient authority and freedom from the line organization to carry out its responsibilities. Persons conducting independent assessments shall be technically qualified and knowledgeable in the areas assessed.</p>			

1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): YES

2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):

The statement is met by HASQAP Section 13.0, "Audits."

ASSESSMENT DATA COLLECTION FORM

Statement Number:	10.a.(9)[2]	Mandatory (Y/N):	Y
Statement: RESPONSIBILITIES AND AUTHORITIES. Lead Program Secretarial Officers...have the following responsibilities: Contracting Officers, based on the advice of the Procurement Request Originator or other designated individual, shall apply applicable provisions of this Order to awards falling within its scope.			

- 1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): NO
- 2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):
Statement is not addressed in the HASQAP.
- 3.0 CORRECTIVE ACTION:

ASSESSMENT DATA COLLECTION FORM

Statement Number: Att.I.II. [2]	Mandatory (Y/N): Y
Statement: GUIDANCE FOR DEVELOPING AND IMPLEMENTING QUALITY ASSURANCE PROGRAMS. In those cases where an organization believes certain criteria or parts of criteria do not apply to their work, the Lead PSO must concur.	

- 1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): NO
- 2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):
Statement is not addressed in the HASQAP.
- 3.0 CORRECTIVE ACTION:

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Att. I. II. A. 1. b.	Mandatory (Y/N):	Y
<p>Statement:</p> <p>GUIDANCE FOR DEVELOPING AND IMPLEMENTING QUALITY ASSURANCE PROGRAMS. Guidance for developing and using the 10 quality assurance criteria is as follows: Management. Criterion 1--Program. Senior management should retain and exercise the responsibility for the scope and implementation of an effective QAP. Line management is responsible for the achievement of quality. Each individual is responsible for the quality of his/her work.</p>			

1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): YES

2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):

Statement is met by the HASQAP "Policy Statement" and subsection 2.2, "Structure, Responsibility, and Authority."

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Att.I.II.A.4.a.(1)	Mandatory (Y/N):	Y
Statement: GUIDANCE FOR DEVELOPING AND IMPLEMENTING QUALITY ASSURANCE PROGRAMS. Guidance for developing and using the 10 quality assurance criteria is as follows: Management. Criterion 4--Documents and Records. Documents. A process should be established and implemented to control preparation, review, approval, issuance, use, and revision of documents that establish policies, prescribe work, specify requirements, or establish design.			

1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): YES

2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):

Statement is met by HASQAP subsection 5.2, "Administrative Systems."

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Att. I, II, A.4.b. (4)	Mandatory (Y/N):	Y
Statement: GUIDANCE FOR DEVELOPING AND IMPLEMENTING QUALITY ASSURANCE PROGRAMS. Guidance for developing and using the 10 quality assurance criteria is as follows: Management Criterion 4--Documents and Records. Records. The National Archives and Records Administration (NARA) exercises final authority for approving the disposition of Government records. Use of the GRS, which is published by the NARA, and the DOE unique schedules approved by the NARA are mandatory.			

- 1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): NO
- 2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):
Statement is not addressed in the HASQAP.
- 3.0 CORRECTIVE ACTION:

ASSESSMENT DATA COLLECTION FORM

Statement Number: Att.I.II.B.1.b.	Mandatory (Y/N): Y
<p>Statement:</p> <p>GUIDANCE FOR DEVELOPING AND IMPLEMENTING QUALITY ASSURANCE PROGRAMS. Guidance for developing and using the 10 quality assurance criteria is as follows: Performance. Criterion 5--Work Processes. Identification and Control of Items.</p> <p>(1) Processes should be established and implemented to identify, control, and maintain items.</p> <p>(2) Identification of items should be maintained to ensure appropriate traceability.</p> <p>(3) Processes should be established and implemented to control consumables and items with limited shelf life, prevent the use of incorrect or defective items, and control samples.</p>	

1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): YES

2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):

Statement is met by HASQAP subsection 5.4, "Physical Facilities Systems," Section 6.0, "Sample Custody and Handling," subsection 7.2, "Specifications of Standards Used for Calibration," and subsection 18.2.5, "Acceptance of Items or Services."

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Att. I, II, B, 3, g.	Mandatory (Y/N):	Y
Statement: GUIDANCE FOR DEVELOPING AND IMPLEMENTING QUALITY ASSURANCE PROGRAMS. Guidance for developing and using the 10 quality assurance criteria is as follows: Performance. Criterion 7--Procurement. Before a procured item is used or placed in service, procurement specification, inspection, and test requirements are to be satisfied and nonconformances properly dispositioned.			

1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): NO

2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):

HASQAP subsections 18.2.5 and 18.2.6 do not adequately address the following parts of the statement:

"Before a procured item is used or placed in service ... inspection, and test requirements are to be satisfied and nonconformances properly dispositioned."

3.0 CORRECTIVE ACTION:

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Att. I. II. B. 3. i	Mandatory (Y/N):	Y
<p>Statement:</p> <p>GUIDANCE FOR DEVELOPING AND IMPLEMENTING QUALITY ASSURANCE PROGRAMS. Guidance for developing and using the 10 quality assurance criteria is as follows: Performance. Criterion 7--Procurement. The quality of purchased items and services should be verified at intervals to a degree consistent with the item's or service's complexity, risk, quantity, and frequency of procurement.</p>			

- 1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): NO
- 2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):
- Statement is not addressed in the HASQAP.
- 3.0 CORRECTIVE ACTION:

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Att. I, II, B, 4.b. (5)	Mandatory (Y/N):	Y
<p>Statement:</p> <p>GUIDANCE FOR DEVELOPING AND IMPLEMENTING QUALITY ASSURANCE PROGRAMS.</p> <p>Guidance for developing and using the 10 quality assurance criteria is as follows: Performance. Criterion 8-Inspection and Acceptance Testing. Acceptance Testing. Retesting of items or processes to determine that they meet acceptance criteria is required after deficiencies are corrected.</p>			

1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): YES

2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):

For analytical laboratory testing the intent of this statement is met by the data quality objective (DQO) process described in HASQAP subsection 4.1, "Data Quality Objectives," and HASQAP Section 15.0, "Corrective Action."

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Att.I.II.C.1.b.	Mandatory (Y/N):	Y
<p>Statement:</p> <p>GUIDANCE FOR DEVELOPING AND IMPLEMENTING QUALITY ASSURANCE PROGRAMS.</p> <p>Guidance for developing and using the 10 quality assurance criteria is as follows: Assessment. Criterion 9-Management Assessment. Senior management should retain overall responsibility for management assessments. Direct participation by senior management during management assessments is essential. This process should involve all levels of management, as appropriate.</p>			

1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): YES

2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):

This statement is met by HASQAP subsection 13.1, "Management System Audits."

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Att.I.II.C.2.b.	Mandatory (Y/N):	Y
Statement: GUIDANCE FOR DEVELOPING AND IMPLEMENTING QUALITY ASSURANCE PROGRAMS. Guidance for developing and using the 10 quality assurance criteria is as follows: Assessment. Criterion 10-Independent Assessment. Personnel performing independent assessments should act in a management advisory function. Their responsibilities are to monitor work performance, identify abnormal performance and precursors of potential problems, identify opportunities for improvement, report results to a level of management having the authority to effect corrective action, and verify satisfactory resolution of problems.			

1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): YES

2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):

Statement is met by HASQAP Section 13.0, "Audits."

WHC-EP-0822

ATTACHMENT C

ASSESSMENT DATA COLLECTION FORMS

for

HASQAP ASSESSMENT

to

DOE ORDER 5480.19
(Conduct of Operations)

ASSESSMENT DATA COLLECTION FORM

Statement Number:	5.b.	Mandatory (Y/N):	Y
Statement: A graded approach shall be used in the application of the guidelines provided in Attachment 1 to assure that the depth of detail required and the magnitude of resources expended for operations are commensurate with each facility's programmatic importance and potential environmental, safety, and/or health impact.			

1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): NO

2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):

The HASQAP does not address the graded approach for application of the guidelines of DOE 5480.19 for analytical laboratory "operations."

3.0 CORRECTIVE ACTION:

ASSESSMENT DATA COLLECTION FORM

Statement Number:	5.c.	Mandatory (Y/N):	Y
<p>Statement:</p> <p>Conformance with the requirements of this Order shall be documented. However, it is not necessary to develop a separate manual or plan. As a minimum, a document (e.g., a matrix) shall be prepared in coordination with the Head of the Field Element and the cognizant Program Secretarial Officer(s) that:</p> <p>(1) Indicates whether a specific guideline applies to a facility;</p> <p>(2) Indicates where and how each of the guidelines (Attachment 1) of this Order are applied within the contractor's existing policies and procedures; and</p> <p>(3) Identifies any deviations or exemptions from the guidelines.</p> <p>This document shall, as a minimum, be approved by the Head of the Field Element.</p>			

- 1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): NO
- 2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):
- Not addressed in the HASQAP.
- 3.0 COMMENTS::
- This assessment may meet the intent of this statement.
- 4.0 CORRECTIVE ACTION:

ASSESSMENT DATA COLLECTION FORM

Statement Number:	6.a.(1)	Mandatory (Y/N):	Y
Statement: RESPONSIBILITIES AND AUTHORITIES. Program Secretarial Officers (PSOs) shall: Provide direction to field elements for the conduct of operations for facilities under their cognizance.			

N/A

Statement Number:	6.a.(2)	Mandatory (Y/N):	Y
Statement: RESPONSIBILITIES AND AUTHORITIES. Program Secretarial Officers (PSO's) shall: Ensure the preparation, review, and approval of contractor documentation implementing the requirements of this Order for programs under their cognizance;			

N/A

Statement Number:	6.a.(3)	Mandatory (Y/N):	Y
Statement: RESPONSIBILITIES AND AUTHORITIES. Program Secretarial officers (PSO's) shall: Ensure that Program Managers oversee the conduct of operations at those facilities under their cognizance, in accordance with the requirements of this Order.			

N/A

ASSESSMENT DATA COLLECTION FORM

Statement Number: 6.b.(1)	Mandatory (Y/N): Y
Statement: RESPONSIBILITIES AND AUTHORITIES. Assistant Secretary for Nuclear Energy (NE-1) shall: Develop, promulgate, and maintain policies necessary to implement and sustain effective DOE-wide conduct of operations of facilities and associated equipment/systems;	

N/A

Statement Number: 6.b.(2)	Mandatory (Y/N): Y
Statement: RESPONSIBILITIES AND AUTHORITIES. Assistant Secretary for Nuclear Energy (NE-1) shall: Develop, promulgate, and maintain guidance materials and conduct training and workshops, as necessary, for line management to implement the above policies and procedures;	

N/A

Statement Number: 6.b.(3)	Mandatory (Y/N): Y
Statement: RESPONSIBILITIES AND AUTHORITIES. Assistant Secretary for Nuclear Energy (NE-1) shall: Monitor appraisal reports relative to conduct of operations at DOE facilities to assess implementation of this Order to identify needed requirements.	

N/A

ASSESSMENT DATA COLLECTION FORM

Statement Number:	6.b.(4)	Mandatory (Y/N):	Y
Statement: RESPONSIBILITIES AND AUTHORITIES. Assistant Secretary for Nuclear Energy (NE-1) shall: Provide interpretation to the requirements of this Order upon request.			

N/A

Statement Number:	6.c.	Mandatory (Y/N):	Y
Statement: RESPONSIBILITIES AND AUTHORITIES. Assistant Secretary for Environment, Safety and Health (EH-1), acting as the independent element responsible for environment, occupational safety, and health oversight of line management for the Department, shall monitor and audit all aspects of the implementation of this Order related to the EH functional areas of responsibility.			

N/A

Statement Number:	6.d.	Mandatory (Y/N):	Y
Statement: RESPONSIBILITIES AND AUTHORITIES. Director of the Office of Nuclear Safety (NS-1), acting as the independent element responsible for nuclear safety oversight of line management for the Department, shall monitor and audit all aspects of the implementation of this Order related to nuclear safety.			

N/A

ASSESSMENT DATA COLLECTION FORM

Statement Number: 6.e.(1)	Mandatory (Y/N): Y
Statement: RESPONSIBILITIES AND AUTHORITIES. Heads of Field Elements shall: Ensure that adequate contractor plans, procedures, and programs are in place and assess the effectiveness of their implementation at sites under their jurisdiction, consistent with the provisions of this Order;	

N/A

Statement Number: 6.e.(2)	Mandatory (Y/N): Y
Statement: RESPONSIBILITIES AND AUTHORITIES. Heads of Field Elements shall: Ensure that DOE Facility Representatives are assigned responsibility for a major facility or group of lesser facilities, and oversee the day-to-day conduct of operations at these facilities in accordance with the requirements of this Order and the direction received from the Program Manager;	

N/A

Statement Number: 6.e.(3)	Mandatory (Y/N): Y
Statement: RESPONSIBILITIES AND AUTHORITIES. Heads of Field Elements shall: Approve documentation prepared by the contractor to demonstrate conformance to the guidelines in Attachment 1.	

N/A

Statement Number: 6.g.	Mandatory (Y/N): N
Statement: RESPONSIBILITIES AND AUTHORITIES. Heads of Power Administrations: ...shall determine the appropriate operations management program for their facilities, which will include consideration of appropriate parts of the criteria set forth by this Order.	

N/A

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch. I.C.1.[1]	Mandatory (Y/N):	N
<p>Statement:</p> <p>OPERATIONS ORGANIZATION AND ADMINISTRATION. GUIDELINES. Operations Policies. Procedures or other definitive documentation should specify policies that are to be applied for operations. These policies should specify goals and the means to achieve those goals. These documents should also provide for the types of controls necessary to implement policies as discussed in this and other chapters of the guidelines. Operations procedures should support facility and DOE guidance for operations. Responsibilities for implementing these policies, including the responsibility of shift personnel, if applicable, should be clearly defined. Operations personnel should clearly understand their authority, responsibility, accountability, and interfaces with other groups.</p>			

1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): YES

2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):

The HASQAP in its entirety fulfills this statement.

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch. I.C.1.[2]	Mandatory (Y/N):	N
Statement: OPERATIONS ORGANIZATION AND ADMINISTRATION GUIDELINES, Operations Policies, Physical security should be in accordance with DOE 5630.11.			

- 1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): NO
- 2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):
Not addressed in the HASQAP.
- 3.0 CORRECTIVE ACTION:

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch. I.C.2.	Mandatory (Y/N):	N
Statement: OPERATIONS ORGANIZATION AND ADMINISTRATION. GUIDELINES. Resources. The operations supervisor for DOE facilities should be provided with sufficient resources in materials and personnel to accomplish assigned tasks without requiring excessive overtime by the operations staff. These resources should include technical personnel needed to support the operations. A long-range staffing plan that anticipates personnel losses should be developed and implemented.			

1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): YES

2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):

The intent of this statement is met by the HASQAP "Policy Statement" and Section 2.0, "Organization and Responsibility."

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch. I.C.3.[1]	Mandatory (Y/N):	N
<p>Statement:</p> <p>OPERATIONS ORGANIZATION AND ADMINISTRATION. GUIDELINES. Monitoring of Operating Performance.</p> <p>As described in Chapter VI, operating problems should be documented and evaluated. Based on assessments of these problems, corrective actions should be taken to improve the performance of the operations department performance. Additionally, frequent direct observation of operations activities by supervisors and managers is essential to performance of monitoring operations.</p>			

1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): YES

2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):

This statement is met by the HASQAP Section 15.0, "Corrective Action."

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch. 1.C.3.[2]	Mandatory (Y/N):	N
Statement: OPERATIONS ORGANIZATION AND ADMINISTRATION GUIDELINES. Monitoring of Operating Performance, Safety, Environment, and Operating goals should be used as a management tool for involving cognizant groups or individuals in improving operating performance and for measuring operating effectiveness.			

1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): NO

2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):

Not specifically addressed in the HASQAP.

3.0 CORRECTIVE ACTION:

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch. I.C.3.[3]	Mandatory (Y/N):	N
<p>Statement:</p> <p>OPERATIONS ORGANIZATION AND ADMINISTRATION. GUIDELINES. Monitoring of Operating Performance</p> <p>Operations goals in areas such as the following should be established:</p> <ul style="list-style-type: none"> - minimizing the unavailability of safety systems; - minimizing personnel errors; - As-Low-As Reasonably-Achievable (ALARA); - minimizing lost facility capability; - minimizing the number of unscheduled facility shutdowns per year; - timely completion of scheduled surveillance; - minimizing the amount of overtime; - achieving and maintaining complete staffing and training of shift positions; - minimizing waste; and - minimizing the number of lighted annunciators. 			

1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): NO

2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):

Although many of the "operations goals" listed in the statement are questionable for laboratory operations, the HASQAP does not address specific laboratory operations goals.

3.0 CORRECTIVE ACTION:

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch. I.C.3.[4]	Mandatory (Y/N):	N
<p>Statement:</p> <p>OPERATIONS ORGANIZATION AND ADMINISTRATION. GUIDELINES. Monitoring of Operating Performance. Goals should be auditable, measurable, realistic, and challenging. Meeting goals should require a definite set of actions or an action plan. The action plan should be developed with input from personnel involved in conducting operations, reviewed by the operations supervisor at DOE facilities, and approved by management. The progress toward completing the action plan and achieving goals should be monitored periodically. If results show a significant variance from the desired progress in achieving goals, management should review the action plan to ensure that it is adequate and is being executed. An audit of performance relative to operating goals should be provided to facility management and DOE. This summary should include an explanation of performance and actions planned to improve future performance. Operating and safety goals should be set and used as motivators for improvement.</p>			

1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): NO

2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):

The HASQAP does not specifically address operating goals and action plans.

3.0 CORRECTIVE ACTION:

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch. I.C.3.[5]	Mandatory (Y/N):	N
<p>Statement:</p> <p>OPERATIONS ORGANIZATION AND ADMINISTRATION. GUIDELINES. Monitoring of Operating Performance. ...</p> <p>Line managers and supervisors should perform routine observations of personnel performing operating activities. Deficiencies identified should be documented, trended, and corrected. Also, other groups, such as quality assurance personnel, should periodically review and assess operation performance.</p>			

1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): NO

2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):

The statement is partially met:

- (1) "Line managers and supervisors should perform routine observations of personnel performing operating activities."

The HASQAP does not specifically address routine observations by line managers and supervisors.

- (2) "Deficiencies identified should be documented, trended, and corrected."

This statement is met by HASQAP Section 15.0, "Corrective Action."

- (3) "Also, other groups, such as quality assurance personnel, should periodically review and assess operation performance."

This statement is met by HASQAP Section 13.0, "Audits."

3.0 CORRECTIVE ACTION:

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch. I.C.4.	Mandatory (Y/N):	N
<p>Statement:</p> <p>OPERATIONS ORGANIZATION AND ADMINISTRATION. GUIDELINES.</p> <p>Accountability. Workers and their supervisors should be held accountable for operating performance. Personnel involved in significant or frequent violations of operating practices should be counseled, retrained, and disciplined, as appropriate. Supervisor performance appraisals and promotions should include an assessment of operating performance.</p>			

1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): NO

2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):

The statement is partially met:

- (1) "Workers and their supervisors should be held accountable for operating performance."

The intent of this statement is met by the HASQAP "Policy Statement" and Section 2.0, "Organization and Responsibility."

- (2) "Personnel involved in significant or frequent violations of operating practices should be counseled, retrained, and disciplined, as appropriate. Supervisor performance appraisals and promotions should include an assessment of operating performance."

The specifics of this statement are not addressed in the HASQAP.

3.0 CORRECTIVE ACTION:

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch. I.C. 5.	Mandatory (Y/N):	N
Statement: OPERATIONS ORGANIZATION AND ADMINISTRATION. GUIDELINES. Management Training. Formalized supervisory and management training should be incorporated into training programs.			

1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): YES

2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):

This statement is met by HASQAP Section 3.5, "Management Training."

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch. I.C.6.	Mandatory (Y/N):	N
Statement: OPERATIONS ORGANIZATION AND ADMINISTRATION. GUIDELINES. Planning for Safety. Facility guidance should exist which describes safety planning requirements for all operational activities. The guidance should explain the role of safety analysis reviews, job safety analyses, and the handling of safety matters. All operations personnel should understand the safety planning requirements.			

1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): NO

2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):

The HASQAP does not address planning for safety.

3.0 CORRECTIVE ACTION:

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch. II.C. 1.	Mandatory (Y/N):	N
<p>Statement:</p> <p>SHIFT ROUTINES AND OPERATING PRACTICES GUIDELINES</p> <p>Status Practices.</p> <p>The operator responsible for the facility should be promptly notified of all changes in facility status, abnormalities, or difficulties encountered in performing assigned tasks. Similarly, the operator should notify the shift supervisor (or cognizant manager for research and test facilities) of any unexpected situations.</p>			

1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): NO

2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):

The statement is not addressed in the HASQAP.

3.0 CORRECTIVE ACTION:

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch. II.C. 2.	Mandatory (Y/N):	N
<p>Statement:</p> <p>SHIFT ROUTINES AND OPERATING PRACTICES. GUIDELINES.</p> <p>Safety Practices.</p> <p>Operations personnel should adhere to the requirements of the facility industrial safety program. Proper hearing, eye, head, foot, and respiratory protection should be worn in designated areas to reduce the potential for injury. Similarly, ladders or other approved means should be used to access equipment located in the overhead when permanent steps or catwalks are not available, thus minimizing the potential for accidents. Operators should not routinely climb or walk on facility components and insulation, because this could result in personnel injury or damage to equipment. Operators should exercise appropriate precautions when entering or working in or around energized panels or equipment.</p>			

- 1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): NO
- 2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):
- Safety practices are not addressed in the HASQAP.
- 3.0 CORRECTIVE ACTION:

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch. II.C. 3.	Mandatory (Y/N):	N
-------------------	--------------	------------------	---

Statement:

SHIFT ROUTINES AND OPERATING PRACTICES. GUIDELINES.

Operator Inspection Tours.

...should be of sufficient detail to ensure that the status of equipment is known. Each operator should conduct a thorough tour of all areas within his/her responsibility, making appropriate equipment inspections at designated times at least once per shift. However, the operations supervisor may designate specific rooms to be inspected less frequently because of adverse radiological or equivalent personnel safety conditions, or more frequently if problems have been encountered. In these cases, the operations supervisor (or cognizant manager for research test facilities) should specify an alternate inspection schedule. Plant security concerns should not override operator safety assessment duties. A tour normally should be made early in the shift, before the operator attends to other duties, so that he/she can become familiar with the condition and status of equipment for which he/she is responsible. During the tour, equipment should be inspected to ensure that it is operating properly or, in the case of standby equipment, this it is fully operable. In addition, the following activities should be conducted in conjunction with the tour:

- The status of equipment (i.e., operating, standby, work in progress, or out-of-service) should be determined so that the operator will be best able to respond to problems he/she may face during his/her shift.
- Components...should be inspected for abnormal or unusual conditions. Unexpected conditions...should be reported to the control room so that supervisors will be aware of the conditions and be able to direct repairs, troubleshooting, or additional operator action, as necessary.
- Equipment panel alarm light bulbs and annunciators should be periodically checked to ensure satisfactory operation of visual and audible abnormal condition indicators.
- Each operator should inspect all areas for which he/she is responsible and note any deficiencies that may be present. Operators should take appropriate action to correct or report deficiencies noted during tours. Equipment deficiencies should also be documented in accordance with the facility maintenance work request system.

N/A

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch. I.L.C. 4.	Mandatory (Y/N):	N
-------------------	---------------	------------------	---

Statement:

SHIFT ROUTINES AND OPERATING PRACTICES. GUIDELINES. Round/Tour Inspection Sheets. ...should be developed and approved by the operations supervisor (or cognizant managers for test and research facilities). They should include areas located within the particular shift position and important parameters for equipment. Where appropriate, equipment parameters should include maximum/minimum values or expected operating ranges to enable operators to recognize abnormal readings quickly. Safety limits derived from Technical Specifications or Operational Safety Requirements should be highlighted. Equipment should be listed on round sheets in the same order that it would be encountered during a normal tour of the operating station, and the round sheets should include a narrative section.

Operators should use the narrative section to document major evolutions, causes of abnormal conditions, and actions taken to correct abnormal conditions. A narrative log book may be substituted for the narrative section on the round sheet. Data should be recorded on round sheets at the times specified by the operations supervisor. When round sheet data is not obtained within one hour of the specified time, the actual time the data was obtained should be noted on the round sheet.

Parameters exceeding the specified maximum/minimum values should be circled or otherwise highlighted on the round sheet and promptly reported to the control room and/or the cognizant operations manager. The causes of abnormal indications should be promptly investigated with supervisors becoming involved as appropriate. The round sheet data should be reviewed by a supervisor each shift to identify trends or abnormal readings and to verify that data has been properly recorded.

Operator rounds should be periodically monitored by supervisory personnel to ensure that comprehensive tours continue to be conducted, including, as necessary, periodic inspections of equipment and areas not listed on the round sheets.

N/A

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch. II.C. 5. [1]	Mandatory (Y/N):	N
Statement: SHIFT ROUTINES AND OPERATING PRACTICES. GUIDELINES. Personnel Protection. Operations personnel should be appropriately qualified to follow good personnel protection practices to maintain personnel exposure as low as reasonably achievable (ALARA) to radiation (DOE 5480.11), chemicals, electromagnetic fields, toxic materials, or other personnel hazards.			

- 1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): NO
- 2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):
- Personnel protection is not addressed in the HASQAP.
- 3.0 CORRECTIVE ACTION:

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch. II.C. 5. a.	Mandatory (Y/N):	N
Statement: SHIFT ROUTINES AND OPERATING PRACTICES. GUIDELINES. Personnel Protection. In particular, operations personnel should observe the following requirements: Operators should adhere to all posted personnel protection requirements and observe proper practices and precautions while in controlled areas.			

- 1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): NO
- 2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):
- Personnel protection is not addressed in the HASQAP.
- 3.0 CORRECTIVE ACTION:

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch. II.C. 5. b.	Mandatory (Y/N):	N
Statement: SHIFT ROUTINES AND OPERATING PRACTICES. GUIDELINES. Personnel Protection. In particular, operations personnel should observe the following requirements. Operators should correctly utilize appropriate monitoring instruments when required.			

- 1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): NO
- 2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):
Personnel protection is not addressed in the HASQAP.
- 3.0 CORRECTIVE ACTION:

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch. II.C. 5. c.	Mandatory (Y/N):	N
Statement: SHIFT ROUTINES AND OPERATING PRACTICES. GUIDELINES. Personnel Protection. In particular, operations personnel should observe the following requirements: Operators should be cognizant of their own exposure levels and take appropriate action to minimize exposures.			

- 1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): NO
- 2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):
Personnel protection is not addressed in the HASQAP.
- 3.0 CORRECTIVE ACTION:

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch. II.C. 5. d.	Mandatory (Y/N):	N
Statement: SHIFT ROUTINES AND OPERATING PRACTICES. GUIDELINES. Personnel Protection. In particular, operations personnel should observe the following requirements: Operators should be knowledgeable of the proper use of radiation work permits, safe work permits, or inhalation limits, where applicable.			

- 1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): NO
- 2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):
Personnel protection is not addressed in the HASQAP.
- 3.0 CORRECTIVE ACTION:

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch. II.C. 5. e.	Mandatory (Y/N):	N
Statement: SHIFT ROUTINES AND OPERATING PRACTICES. GUIDELINES. Personnel Protection. In particular, operations personnel should observe the following requirements: Operators should promptly report protection deficiencies and hazards to the control personnel and/or appropriate protection personnel. In addition, operators should take appropriate immediate actions to reduce or correct the hazards.			

- 1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): NO
- 2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):
- Personnel protection is not addressed in the HASQAP.
- 3.0 CORRECTIVE ACTION:

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch. II.C. 5. f.	Mandatory (Y/N):	N
Statement: SHIFT ROUTINES AND OPERATING PRACTICES, GUIDELINES. Personnel Protection. In particular, operations personnel should observe the following requirements: Appropriate protection personnel should be informed prior to evolutions or activities that have a potential to significantly change conditions in the facility.			

- 1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): NO
- 2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):
Personnel protection is not addressed in the HASQAP.
- 3.0 CORRECTIVE ACTION:

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch. II.C. 5. [2]	Mandatory (Y/N):	N
Statement: SHIFT ROUTINES AND OPERATING PRACTICES. GUIDELINES. Personnel Protection. Operations supervisory personnel should periodically review exposure trends of operating personnel under their supervision. Emphasis should be placed on determining the adverse factors that contribute to personnel exposures and minimizing those factors to keep exposures as low as reasonably achievable.			

- 1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): NO
- 2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):
Review of exposure trends is not specifically addressed in the HASQAP.
- 3.0 CORRECTIVE ACTION:

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch. II.C. 6.	Mandatory (Y/N):	N
<p>Statement:</p> <p>SHIFT ROUTINES AND OPERATING PRACTICES. GUIDELINES. Response to Indications. Operators should believe instrument readings and treat them as accurate unless proven otherwise. Ignoring an unusual reading because the operator believes an instrument is faulty can cause abnormal conditions to be undetected. In general, operators should check other indications, if possible, when unexpected readings are observed. Prompt action should be taken to investigate the cause of abnormal or unexpected indications so that prompt corrective action can occur. When malfunctioning or inaccurate instruments are discovered, they should be appropriately identified to prevent subsequent confusion and instrument and control personnel should be notified to effect repairs. In situations of operator doubt, operators should be instructed to achieve facility, personnel, and environmental safety above facility production.</p>			

1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): YES

2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):

The intent of this statement is met by HASQAP Section 7.0, "Calibration."

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch. II.C. 7.	Mandatory (Y/N):	N
<p>Statement:</p> <p>SHIFT ROUTINES AND OPERATING PRACTICES. GUIDELINES. Resetting Protective Devices. When protective devices trip (e.g., circuit breakers, fuses, reactor protection channels where multichannel logic exists), an attempt should be made to understand the cause of the trip before the device is reset. Normally, before action is taken, an operator should ensure no abnormal condition exists that would preclude reset. However, because the consequences of inappropriately resetting protective devices vary considerably, good judgment and specific guidance are necessary in this area. The operations management should provide the appropriate guidance so that tripped protective devices will be properly addressed. Facility trips and unplanned forced shutdown require a thorough investigation in accordance with the guidance of Chapter VI.</p>			

N/A

Statement Number:	Ch. II.C. 8.	Mandatory (Y/N):	N
<p>Statement:</p> <p>SHIFT ROUTINES AND OPERATING PRACTICES. GUIDELINES. Load Changes. The shift supervisor, the control room lead operator, or the cognizant manager for a test and research facility should approve all power or process rate changes because these persons are held accountable for safe operation. Additionally, they will probably be the persons most knowledgeable of problems that occur as a result of load changes. However, the operator could decrease load or rate without approval, if necessary, to respond to a facility emergency situation in accordance with the facility's emergency procedures.</p>			

N/A

Statement Number:	Ch. II.C. 9.[1]	Mandatory (Y/N):	N
<p>Statement:</p> <p>SHIFT ROUTINES AND OPERATING PRACTICES. GUIDELINES. Authority to Operate Equipment. The overall operation of the facility should be directed by the operations supervisor for a large DOE facility and by the cognizant manager for a test and research facility.</p>			

N/A

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch. II.C. 9.[2]	Mandatory (Y/N):	N
Statement: SHIFT ROUTINES AND OPERATING PRACTICES. GUIDELINES. Authority to Operate Equipment. Operations management should ensure that only trained and qualified personnel operate plant equipment.			

N/A

Statement Number:	Ch. II.C. 9.[3]	Mandatory (Y/N):	N
Statement: SHIFT ROUTINES AND OPERATING PRACTICES. GUIDELINES. Authority to Operate Equipment. In general, the operator and the operations supervisor should be aware of all activities affecting equipment.			

N/A

Statement Number:	Ch. II.C. 9.[4]	Mandatory (Y/N):	N
Statement: SHIFT ROUTINES AND OPERATING PRACTICES. GUIDELINES. Authority to Operate Equipment. The operations supervisor should specify those general activities that may normally be performed without informing the supervisor and should amplify these specifications as appropriate.			

N/A

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch. II.C.10.	Mandatory (Y/N):	N
Statement: SHIFT ROUTINES AND OPERATING PRACTICES. GUIDELINES. Shift Operating Bases. ...An operating base should be established for each shift position. Each operating base should be equipped with appropriate office equipment for the operator to maintain necessary procedures and references and to conduct administrative duties, and necessary communication equipment should be available at the operating base. Shift turnovers should be conducted within those facility areas assigned to the operations department and typically at the operating base.			

N/A

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch. II.C.11.	Mandatory (Y/N):	N
<p>Statement:</p> <p>SHIFT ROUTINES AND OPERATING PRACTICES. GUIDELINES. Potentially Distractive Written Material and Devices. Written material that does not relate to operation and entertainment devices (such as radios, televisions, tape players, and computer games) should be prohibited from use by on-duty operations personnel in order to minimize distractions from their responsibilities.</p>			

N/A

Statement Number:	Ch. III.C.1.	Mandatory (Y/N):	N
<p>Statement:</p> <p>CONTROL AREA ACTIVITIES FOR DOE FACILITIES. GUIDELINES. Control Area Access. ...should be limited to those persons on official business only. The "at-the-controls" area of the control room should be clearly identified, and its boundary should be understood by all persons who are granted access to the control room. Access to the "at-the-controls" area should be restricted to persons who need to be in the area. Entry into this area should be granted by designated individuals, and persons who might need to enter this area should know who can grant access.</p>			

N/A

Statement Number:	Ch. III.C.2.	Mandatory (Y/N):	N
<p>Statement:</p> <p>CONTROL AREA ACTIVITIES FOR DOE FACILITIES. GUIDELINES. Professional Behavior. ...should be displayed in the control area at all times. Only activities essential to supporting operation and activities authorized by management should be conducted in the control area. Potentially distracting activities (such as radio listening, game playing, and horseplay) should be prohibited. Non-job-related discussions should be minimized so as not to interfere with conduct of the shift or monitoring of key parameters.</p>			

N/A

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch. III.C.3.[1]	Mandatory (Y/N):	N
<p>Statement:</p> <p>CONTROL AREA ACTIVITIES FOR DOE FACILITIES. GUIDELINES. Monitoring the Main Control Panels. Operators should be alert and attentive to control panel indications and alarms. Control panel indications should be monitored frequently, and prompt action should be taken to determine the cause of and correct abnormalities. Emphasis should be placed on closely monitoring and trending to detect problem situations early. Operator response to alarms should be timely, and actions should be taken to address and correct the alarm causes. All reasonable action should be taken to clear alarming conditions.</p>			

N/A

Statement Number:	Ch. III.C.3.[2]	Mandatory (Y/N):	N
<p>Statement:</p> <p>CONTROL AREA ACTIVITIES FOR DOE FACILITIES. GUIDELINES. Monitoring the Main Control Panels. The number of evolutions affecting control panel indications that are performed concurrently should be limited so that the operators' ability to detect and respond to abnormal conditions will not be compromised.</p>			

N/A

Statement Number:	Ch. III.C.3.[3]	Mandatory (Y/N):	N
<p>Statement:</p> <p>CONTROL AREA ACTIVITIES FOR DOE FACILITIES. GUIDELINES. Monitoring the Main Control Panels. If computer or automated systems are in place, there should be an appropriate backup to those systems.</p>			

N/A

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch. III.C.4.	Mandatory (Y/N):	N
<p>Statement:</p> <p>CONTROL AREA ACTIVITIES FOR DOE FACILITIES. GUIDELINES. Control Operator Ancillary Duties. Duties assigned to operators should not interfere with their ability to monitor facility parameters. Activities such as preparation of tagging orders, reviews of operating procedures, required reading, and review of maintenance work activities should not comprise a major portion of these operators' shift responsibilities. The administrative workload of operators responsible for monitoring and operating the control board should be minimized.</p>			

N/A

Statement Number:	Ch. III.C.5.	Mandatory (Y/N):	N
<p>Statement:</p> <p>CONTROL AREA ACTIVITIES FOR DOE FACILITIES. GUIDELINES. Operation of Control Area Equipment. Only persons specifically authorized by the administrative procedures of the operations department should operate control area equipment. When trainees operate this equipment, they should be supervised and controlled by the operator who normally would perform the operations.</p>			

N/A

Statement Number:	Ch. IV.C.1.	Mandatory (Y/N):	N
<p>Statement:</p> <p>COMMUNICATIONS. GUIDELINES. Emergency Communications Systems. Methods should be implemented to ensure all facility personnel are promptly alerted to facility emergencies. When personnel are working in areas where the public address system or emergency signals cannot be heard, alternate methods for alerting these persons should be utilized.</p> <p>Emergency communications systems should be periodically tested to ensure that they are functional. Control areas should have the capability of overriding other users of the public address system for emergency announcements.</p>			

N/A

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch. IV.C.2.	Mandatory (Y/N):	N
<p>Statement:</p> <p>COMMUNICATIONS. GUIDELINES. Public Address System. Use of the facility public address system (page) should be administratively controlled to ensure it retains its effectiveness in contacting plant personnel. Excessive use of the public address system for paging of personnel and unnecessary announcements should be avoided because excessive use can reduce the impact of important announcements and can be distracting. Facility telephones and other point-to-point communications channels should be used in lieu of the public address system whenever practical. Consideration should be given to dedicating certain paging system channels to specific groups or functions.</p>			

N/A

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch. IV.C.3.	Mandatory (Y/N):	N
Statement: COMMUNICATIONS. GUIDELINES. Contacting Operators. Methods should be implemented to ensure that control areas can quickly contact on-shift operators or supervisors. ...However, to avoid operator confusion, distinction should be made between a routine and emergency notifications.			

N/A

Statement Number:	Ch. IV.C.4.	Mandatory (Y/N):	N
Statement: COMMUNICATIONS. GUIDELINES. Radios. Portable radios can be an effective means of providing mobile point-to-point communications and may be used for this purpose. However, radio usage should not be allowed in areas where electronic interference with plant equipment may result. Areas where radio use is prohibited should be delineated. Instructions regarding frequencies (channels) and postings should be provided. Consideration should be given to dedicating certain radio channels to specific groups or functions...			

N/A

Statement Number:	Ch. IV.C.5.	Mandatory (Y/N):	N
Statement: COMMUNICATIONS. GUIDELINES. Abbreviations and Acronyms. Only abbreviations and acronyms obtained from an approved list should be used in facility communications. Both written and spoken terms should be prescribed in the list.			

N/A

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch. IV.C.6.	Mandatory (Y/N):	N
Statement: COMMUNICATIONS. GUIDELINES. Oral Instructions and Informational Communications. Oral instructions should be clear and concise. In all communications, the sender and intended receiver should be readily identifiable. Instructions involving the operation of equipment should be repeated by the receiver to the extent necessary for the sender to ensure the instructions are correctly understood.			

N/A

Statement Number:	Ch. V.C.1.	Mandatory (Y/N):	N
Statement: CONTROL OF ON-SHIFT TRAINING. GUIDELINES. Adherence to Training Programs. On-shift training should be conducted in accordance with training programs that specifically identify items the trainee must accomplish on shift. The knowledge requirements for each item should be defined as well as what the trainee must do (perform, simulate, observe, or discuss). Both the instructor and the trainee should understand what is required for each training item.			

1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): YES

2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):

The intent of this statement is met by HASQAP Section 3.0, "Personnel Qualification and Training." However, the HASQAP does not contain the specificity of this statement.

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch. V.C.2.	Mandatory (Y/N):	N
Statement: CONTROL OF ON-SHIFT TRAINING. GUIDELINES. On-shift Instructor Qualification. On-shift training should be conducted by qualified operators. ...The on-shift instructors should be specifically selected, taking into account communication skills, technical knowledge, and ability to provide trainees with hands-on experience.			

1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): YES

2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):

The intent of this statement is met by HASQAP Section 3.0, "Personnel Qualification and Training." However, the HASQAP does not contain the specificity of this statement.

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch. V.C.3.	Mandatory (Y/N):	N
<p>Statement:</p> <p>CONTROL OF ON-SHIFT TRAINING. GUIDELINES. Qualified Operator Supervision and Control of Trainees. Whenever trainees operate equipment, a qualified operator serving as an on-shift instructor should observe the trainee in order to ensure the trainee does not make an error that could adversely impact the facility. Until the trainee has demonstrated reasonable proficiency in an operation, he/she should discuss the procedure steps, cautions, and notes with the instructor. Trainees should also demonstrate actions to be performed by pointing to the control switch, valve, breaker, etc., that will be manipulated.</p> <p>On-shift instructors should not become complacent with trainees. Just because a trainee has performed a task once does not mean he is aware of all problems that could occur. The instructor should always monitor the trainee closely and remain in a position to intervene or assume control, if necessary.</p> <p>When trainees record equipment parameters on official round sheets (as opposed to practice round sheets) or logs, the on-shift instructor should verify that the recorded information is correct. In addition, the trainee and on-shift instructor should discuss any out-of-specification readings and the consequences of allowing such trend to continue.</p>			

1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): YES

2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):

The intent of this statement is met by HASQAP Section 3.0, "Personnel Qualification and Training." However, the HASQAP does not contain the specificity of this statement.

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch. V.C.4.	Mandatory (Y/N):	N
Statement: CONTROL OF ON-SHIFT TRAINING. GUIDELINES. Operator Qualification Program Approval. The operator qualification program should be approved by the operations supervisor, and changes to the program should be coordinated with the training department. For operating positions requiring certification, qualifications should be based on one-to-one instruction at that station.			

1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): YES

2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):

The intent of this statement is met by HASQAP Section 3.0, "Personnel Qualification and Training."

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch. V.C.5.	Mandatory (Y/N):	N
Statement: CONTROL OF ON-SHIFT TRAINING. GUIDELINES. Training Documentation. Completion of the operator qualification program should be formally documented. Classroom requirements and written exam results should be documented by training department instructors. On-shift training and system checkouts should be documented by on-shift instructors.			

1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): YES

2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):

The intent of this statement is met by HASQAP Section 3.7, "Training Records."

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch. V.C.6.	Mandatory (Y/N):	N
<p>Statement:</p> <p>CONTROL OF ON-SHIFT TRAINING. GUIDELINES. Suspension of Training. Trainee operation of equipment should be immediately suspended during unanticipated or abnormal events, accident conditions, or whenever the operations personnel or on-shift instructor believes suspension is necessary to ensure safe and reliable facility operation. During abnormal or accident conditions, trainees should provide assistance at the discretion of the qualified operator.</p>			

N/A

Statement Number:	Ch. V.C.7.	Mandatory (Y/N):	N
<p>Statement:</p> <p>CONTROL OF ON-SHIFT TRAINING. GUIDELINES. Maximum Number of Trainees. ...allowed to simultaneously participate in any particular training evolution needs to be considered. Consideration should be given to training effectiveness and to the potential for adverse effects on the facility. A maximum limit for the trainee-to-instructor ratio will ensure that the trainee is provided with the most effective instruction and will ensure that the instructor is not distracted by having too many trainees at once. Shift supervisors should ensure that established limits are observed.</p>			

N/A

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch. VI.C.1.[1]	Mandatory (Y/N):	N
Statement: INVESTIGATION OF ABNORMAL EVENTS. GUIDELINES. Events Requiring Investigation. Events that occur in the facility and adversely affect operations, personnel safety, or DOE requirements (DOE 5000.3A) should receive a thorough investigation. The criteria for when to perform an event investigation should be clearly established. Specific events requiring investigation should be listed for supervisory use, along with criteria for use in deciding what "near miss" situations should receive review.			

1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): NO

2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):

HASQAP Section 15.0, "Corrective Action," emphasizes data quality rather than abnormal events.

3.0 CORRECTIVE ACTION:

ASSESSMENT DATA COLLECTION FORM

Statement Number: Ch. VI.C.1.[2]	Mandatory (Y/N): N
----------------------------------	--------------------

Statement:

INVESTIGATION OF ABNORMAL EVENTS. GUIDELINES: Events Requiring Investigation. The following conditions and situations should require an investigation:

- a. Design limits are violated (Tech Specs, OSR, SAR, or other limits).
- b. Facility system performance is unusual, abnormal or unexplained.
- c. Facility safety conditions are abnormal or unexplained.
- d. Safety or system features are improperly positioned.
- e. Reportability to DOE or other agencies (e.g., EPA) is appropriate.
- f. An unplanned shutdown or significant loss of operation occurs.
- g. A procedural violation or personnel error occurs that caused or could have caused serious personnel injury or equipment damage or could have affected facility safety.
- h. Equipment failure occurs that could affect facility capability or safety.
- i. Radiological or toxic material limits are exceeded or radioactive or toxic material is lost/released.
- j. Actual or attempted sabotage is suspected.
- k. Chemistry or process parameters are out of specification or indicate unexplained trends.
- l. A department head or the facility safety review committee deems an investigation is appropriate.
- m. Loss of Special Nuclear Material.
- n. Repetitive problems occur.

The above list is not intended to be all-inclusive. At the discretion of the operations supervisor (or other appropriate department head), other specific events should receive a formal investigation.

- 1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): NO
- 2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):
- This statement (or its intent) is not met by the HASQAP.
- 3.0 CORRECTIVE ACTION:

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch. VI.C.1.[3]	Mandatory (Y/N):	N
Statement: INVESTIGATION OF ABNORMAL EVENTS. GUIDELINES. Events Requiring Investigation. "Near miss" situations should also receive a formal review at the discretion of the responsible supervisor.			

- 1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): NO
- 2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):
Statement not addressed in the HASQAP.
- 3.0 CORRECTIVE ACTION:

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch. VI.C.2.	Mandatory (Y/N):	N
<p>Statement:</p> <p>INVESTIGATION OF ABNORMAL EVENTS. GUIDELINES. Investigation Responsibility. The operations supervisor or another manager should be responsible for event investigations. He/she may delegate specific investigations or portions of investigations to other personnel. However, the overall responsibility for the consistency and thoroughness of event investigations should be the responsibility of the appropriate manager.</p>			

- 1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): NO
- 2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):
- Statement not addressed in the HASQAP.
- 3.0 CORRECTIVE ACTION:

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch. VI.C.3.	Mandatory (Y/N):	N
Statement: INVESTIGATION OF ABNORMAL EVENTS. GUIDELINES. Investigator Qualification. The credibility of the investigative process is highly dependent upon the knowledge and experience of the individuals performing the investigation. ...Additionally, investigators should not have a bias or a vested interest in the results of the investigation. Investigators should be trained in facility systems and operations and other major disciplines appropriate for the event under investigation. Additionally, investigators should be trained in techniques for conducting an investigation.			

- 1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): NO
- 2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):

Statement not addressed in the HASQAP.
- 3.0 CORRECTIVE ACTION:

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch. VI.C.4.	Mandatory (Y/N):	N
<p>Statement:</p> <p>INVESTIGATION OF ABNORMAL EVENTS. GUIDELINES. Information to be Gathered. Sufficient data must be collected to allow the event investigators to perform a reconstruction and analysis of the event. An individual should be assigned responsibility for collecting the required information and assembling the information for review. It is important to collect the necessary information as soon as possible after the event. ...Information should be gathered in the following areas:</p> <ul style="list-style-type: none"> a. Initial facility conditions. b. Statements of operators and personnel involved in the event (this should be permanently recorded). c. Pertinent computer printouts (post-trip log sequence of events) and strip charts. d. Pertinent documentation...as required to establish conditions prior to and during the event. <p>...The collection of data should not interfere with the continued operation of the facility. Temporary interference should occur only if necessary to understand the event.</p>			

- 1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): NO
- 2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):
- Statement not addressed in the HASQAP.
- 3.0 CORRECTIVE ACTION:

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch. VI.C.5.	Mandatory (Y/N):	N
Statement: INVESTIGATION OF ABNORMAL EVENTS. GUIDELINES. Event Investigation. Upon completion of the data collection, a structured review of the abnormal event should occur. The format of the investigation will depend upon the significance of the event. Each event investigation should include the following steps, with emphasis placed on each individual step depending upon the severity or potential consequences of the event.			

- 1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): NO
- 2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):
- Statement not addressed in the HASQAP.
- 3.0 CORRECTIVE ACTION:

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch. VI.C.5.a.	Mandatory (Y/N):	N
<p>Statement:</p> <p>INVESTIGATION OF ABNORMAL EVENTS. GUIDELINES. Event Investigation. Event Reconstruction. The abnormal event should be reconstructed using the collected information...</p> <p>If the sequence of events printout is not available (or not applicable) or if the abnormal event was not of sufficient magnitude or nature to impact the sequence of events recorder, then a chronological listing of events should be developed.</p>			

- 1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): NO
- 2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):
- Statement not addressed in the HASQAP.
- 3.0 CORRECTIVE ACTION:

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch. VI.C.5.b.	Mandatory (Y/N):	N
<p>Statement:</p> <p>INVESTIGATION OF ABNORMAL EVENTS. GUIDELINES. Event Investigation. Event Analysis and Evaluation. Once the facts of the event have been established, the event should be analyzed to determine the response of equipment and involved personnel. During the analysis, emphasis should be placed on determining the proper response of systems, comparison of actual and expected response, adequacy of procedures, and factors affecting human performance. When available, the event should be compared with previous event investigations of similar events or transients. During the analysis, a safety evaluation should be performed to ascertain the proper response of equipment and to identify detrimental effects on facility equipment. If the event was a process shutdown, the acceptability of restart may be determined at this time. In some instances, however, steps 5c and 5d below may also be required prior to restart. Supervisors need to be sure no further corrective action is required prior to restart.</p>			

- 1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): NO
- 2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):
- Statement not addressed in the HASQAP.
- 3.0 CORRECTIVE ACTION:

ASSESSMENT DATA COLLECTION FORM

Statement Number: Ch. VI.C.5.c.	Mandatory (Y/N): N
Statement: INVESTIGATION OF ABNORMAL EVENTS. GUIDELINES. Event Investigation. Root-Cause Determination. The root cause of the event should be determined.	

1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): YES

2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):

Root cause analysis is provided for in HASQAP Section 15.4, "Root Cause Analysis."

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch. VI.C.5.d.	Mandatory (Y/N):	N
Statement: INVESTIGATION OF ABNORMAL EVENTS. Event investigation. Corrective Action Determination. Appropriate corrective action should be established for each event investigation, and specific personnel should be assigned responsibilities for the corrective action. The final approval for corrective action should be made by the facility manager.			

- 1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): YES
- 2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):
Corrective Action is addressed in HASQAP Section 15.0, "Corrective Action."

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch. VI.C.6.	Mandatory (Y/N):	N
<p>Statement:</p> <p>INVESTIGATION OF ABNORMAL EVENTS. GUIDELINES. Investigative Report... should be prepared in a timeframe determined by the responsible authority. The report should include a description of the event (including pertinent conditions), a discussion of the impact of the event, root cause, the lessons learned, and the proposed corrective action(s). The report should include positive aspects of the event... The investigative report should be approved by the facility manager and reviewed by appropriate supervisors, managers, and the safety review committee.</p> <p>It is important that the lessons learned from an event investigation be shared with all appropriate personnel who could benefit from the lessons learned.</p>			

1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): YES

2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):

The intent of this statement is met by HASQAP Section 15.0, "Corrective Action."

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch. VI.C.7.	Mandatory (Y/N):	N
Statement: INVESTIGATION OF ABNORMAL EVENTS. GUIDELINES. Event Training. In-house events should be evaluated by the operations supervisor to determine if the event should be included in the training program for operations personnel. In some cases, due to the severity or possible safety consequences of some events, it is appropriate to train shift operations personnel on the event immediately. A mechanism should exist so that appropriate shift personnel could be immediately trained on an event when they next report for work.			

- 1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): NO
- 2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):
Statement is not addressed by the HASQAP.
- 3.0 CORRECTIVE ACTION:

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch. VI.C.8.	Mandatory (Y/N):	N
Statement: INVESTIGATION OF ABNORMAL EVENTS. GUIDELINES. Event Trending. Patterns of deficiencies such as operator errors or inadequate procedures should be trended. A periodic summary report of events, causes, and trends should be submitted to department heads, the facility manager, and appropriate managers. Department heads should ensure training programs include appropriate material from the summary report.			

1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): YES

2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):

The intent of this statement is met by HASQAP Section 15.3, "Trend Analysis."

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch. VI.C.9	Mandatory (Y/N):	N
<p>Statement:</p> <p>INVESTIGATION OF ABNORMAL EVENTS. GUIDELINES. Sabotage. Acts of known or suspected sabotage are a special case of event investigations. If an act of sabotage is discovered or suspected, it is important to begin an investigation immediately and to accomplish the following:</p> <ul style="list-style-type: none"> a. Determine the condition of the affected system(s) and ensure the operability of all safety-related systems; b. Decide if continued operation is justified or if systems are available to support safe facility shutdown; and c. Minimize the impact of discovered acts of sabotage and deter future acts of sabotage. 			

1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): NO

2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):

Statement is not addressed by the HASQAP.

3.0 CORRECTIVE ACTION:

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch. VII.C.1.	Mandatory (Y/N):	N
<p>Statement:</p> <p>NOTIFICATIONS. GUIDELINES. Notification Procedures. Procedures should be developed to address appropriate notifications and should include the following elements:</p> <ul style="list-style-type: none"> a. Specific responsibilities for notifications; b. Identification of events and conditions requiring notifications; c. Identification of primary and alternate personnel to be notified for various situations; d. Establishment of time requirements for notifications that are consistent with the facility emergency plan; and e. Definition of recordkeeping requirements that documents the reason for notifications, the time of notifications, and the person notified. 			

1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): YES

2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):

The intent of this statement is met by HASQAP subsection 9.3.2, "Immediate Reporting," for analytical laboratory operations.

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch. VII.C.2.	Mandatory (Y/N):	N
Statement: NOTIFICATIONS. GUIDELINES. Notification Responsibility. The operations supervisor should ensure that all appropriate personnel receive notification when required. The actual notification of specific individuals or agencies may be accomplished by other individuals.			

1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): NO

2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):

The specifics of this statement are not addressed by HASQAP subsection 9.3.2, "Immediate Reporting."

3.0 CORRECTIVE ACTION:

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch. VII.C.3.	Mandatory (Y/N):	N
Statement: NOTIFICATIONS. GUIDELINES. Names and Phone Numbers. Names of primary and alternate contacts and current phone numbers and page codes should be readily available to the person assigned to make the notifications.			

1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): NO

2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):

The specifics of this statement are not addressed by HASQAP subsection 9.3.2, "Immediate Reporting."

3.0 CORRECTIVE ACTION:

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch. VII.C.4.	Mandatory (Y/N):	N
Statement: NOTIFICATIONS. GUIDELINES. Documentation. All notifications should be documented. Fill-in-the-blank forms for different types of situations might be useful as a checklist and for providing necessary documentation. In any case, a formal record of notifications should be maintained.			

1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): NO

2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):

The specifics of this statement are not addressed by HASQAP subsection 9.3.2, "Immediate Reporting."

3.0 CORRECTIVE ACTION:

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch. VII.C.5.	Mandatory (Y/N):	N
Statement: NOTIFICATIONS. GUIDELINES. Communication Equipment. Adequate communication equipment should be maintained in the main control area to meet the objectives of this chapter.			

N/A

Statement Number:	Ch. VIII.C. 1.	Mandatory (Y/N):	N
Statement: CONTROL OF EQUIPMENT AND SYSTEM STATUS. GUIDELINES. Status Change Authorization and Reporting. The operations supervisor is responsible for maintaining proper configuration and should authorize status changes to major equipment and systems... Authorization of status changes to equipment and systems of lesser importance may be delegated by the operations supervisor to other cognizant shift positions. ...The shift supervisor should be advised periodically of changes in status of equipment and systems so assigned. Since the operators must be aware of equipment and system status, the supervisor should ensure that all changes in status are communicated to these persons. Typically, the control area operator is in the line-of-information flow to and from the supervisor. Changes in the status of facility equipment and systems should be reported to the governing stations...or to the individual (or his relief) who authorized the change. Typically, changes in status of safety-related equipment and systems should be authorized by the supervisor and reported to the control area.			

N/A

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch. VIII.C. 2.	Mandatory (Y/N):	N
<p>Statement:</p> <p>CONTROL OF EQUIPMENT AND SYSTEM STATUS. GUIDELINES. Equipment and System Alignment. Prior to first placing the equipment or system into operation, individual components for facility equipment and systems should be properly aligned or checked for proper alignment. An initial alignment of valves, switches, and breakers establishes a baseline configuration against which further operations may be measured. ...Alignment checklists should be used to guide the operator in establishing the correct component positions. ...The supervisor should review and approve completed alignment checklists. The need for a complete alignment of equipment and systems should be based on the level of control that has been maintained over the status of the components. ...In addition, safety-related equipment and systems should be functionally tested in accordance with surveillance requirements in the technical specifications/operational safety requirements following maintenance and before the equipment or system is considered capable of performing its design function. Records of equipment and system alignments should be retained for reference by the operating shift. Administrative controls that analyze and document deviations from the reference alignment should be established.</p>			

N/A

Statement Number:	Ch. VIII.C. 3.	Mandatory (Y/N):	N
<p>Statement:</p> <p>CONTROL OF EQUIPMENT AND SYSTEM STATUS. GUIDELINES. Equipment Locking and Tagging. Locks and Tags should be used on those components that require special administrative control for safety or other reasons. ...Additionally, Locks and Tags should alert the operator of the importance of the component and remind him/her that special controls over repositioning are to be maintained. In this respect, all personnel should receive training regarding their responsibilities concerning the manipulation of locked or tagged controls.</p>			

N/A

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch. VIII.C. 4.	Mandatory (Y/N):	N
<p>Statement:</p> <p>CONTROL OF EQUIPMENT AND SYSTEM STATUS. GUIDELINES. Operational Limits Compliance. Administrative controls should be established to document compliance with requirements of operational limits (limiting conditions for operation). The operations supervisor should be cognizant of action taken to comply with operational limit requirements and should ensure that the actions taken are appropriate and correct or that they mitigate any adverse consequences to the facility. Logs, status sheets, turnover checklists, or other appropriate documentation should reflect the entry conditions and actions that are taken in response to operational limits requirements. Appropriate operating personnel should be apprised of limiting conditions for operations and actions for which they may be responsible. Responsible personnel should periodically review the limiting conditions for operation and action statements in effect to ensure that the required actions are met.</p>			

N/A

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch. VIII.C. 5.	Mandatory (Y/N):	N
Statement: CONTROL OF EQUIPMENT AND SYSTEM STATUS. GUIDELINES. Equipment Deficiency Identification and Documentation. Equipment deficiencies should be noted by facility operating personnel and identified in the work control system for correction. Methods that identify deficient equipment to operating personnel should be established.			

1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): YES

2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):

The intent of this statement is met by HASQAP Section 14.0, "Preventive Maintenance."

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch. VIII.C. 6.	Mandatory (Y/N):	N
Statement: CONTROL OF EQUIPMENT AND SYSTEM STATUS. GUIDELINES. Work Authorization and Documentation. The operations supervisor or his/her designee should authorize all shift activities (including Maintenance) on equipment that is important to safety, that affects operations, or that changes control indications or alarms. This authorization should be in writing on the document controlling the work. Documentation of the status of work in progress should be available in the control area for review by operating personnel.			

N/A

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch. VIII.C. 7.	Mandatory (Y/N):	N
Statement: CONTROL OF EQUIPMENT AND SYSTEM STATUS. GUIDELINES. Equipment Post-Maintenance Testing and Return to Service. Equipment should be tested following maintenance to demonstrate that it is capable of performing its intended function. The testing should include performance of all functions that may have been affected by the maintenance. The testing should also verify that the maintenance performed served to correct the original problem and that no new problems were introduced. Any testing following maintenance should be specified on the maintenance work order or accompanying documentation... The operations supervisor should ensure that testing appropriately proves equipment operability.			

1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): YES

2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):

The intent of the statement is met by HASQAP Section 7.0, "Calibration."

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch. VIII.C. 8.	Mandatory (Y/N):	N
<p>Statement:</p> <p>CONTROL OF EQUIPMENT AND SYSTEM STATUS. GUIDELINES. Alarm Status. The status of control panel and/or local panel alarms should be readily available to appropriate operating personnel. Information that should be available includes alarms that are totally disabled, alarms with individual inputs disabled, alarms with temporarily changed setpoints, alarms that are normally lighted during power operation, and multiple input alarms that do not reflash when more than one input is activated. Appropriate actions should be taken to monitor equipment parameters for abnormal conditions that would be masked by deficient or non-reflashing alarms.</p>			

N/A

Statement Number:	Ch. VIII.C. 9.	Mandatory (Y/N):	N
<p>Statement:</p> <p>CONTROL OF EQUIPMENT AND SYSTEM STATUS. GUIDELINES. Temporary Modification Control. Administrative control systems should be established for installation of temporary modifications... Prior to modification, these controls should provide for communicating the installation of temporary modifications to the design authority to allow for technical oversight and an evaluation of the impact on current design activities, and approval of the design modification. These control systems should make provisions for safety reviews, installation approval, independent verification of correct installation and removal, documentation of the modification, update of operating procedures and documents, training, marking of installed modifications, and periodic audits of outstanding modifications.</p>			

N/A

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch. VIII.C.10.	Mandatory (Y/N):	N
<p>Statement:</p> <p>CONTROL OF EQUIPMENT AND SYSTEM STATUS. GUIDELINES. Distribution and Control of Equipment and System Documents. A system should be established to ensure that the operations personnel receive and utilize the latest revisions of engineering drawings and specifications. Operations personnel should be made aware of all changes to these documents. The document distribution should include all operations related activities such as procedures review groups, maintenance groups, facility safety analysis groups, and testing groups.</p>			

N/A

The following statements concerning "LOCKOUTS AND TAGOUTS" are
N/A:

Statement Number:	Ch. IX.C. 1.	Mandatory (Y/N):	N
<p>Statement:</p> <p>LOCKOUTS AND TAGOUTS. GUIDELINES. Lockout/Tagout Use. Locks and Tags should be placed on controls when for safety or other special administrative reasons controls must be established. ...Locks may be built into the switch or be external locks...that may be affixed and or removed when necessary. ...Other devices...may also constitute a lesser degree of Lockout. ...Tags should be placed on the control that is tagged out, or as close as possible to indicate clearly the condition.</p>			

Statement Number:	Ch. IX.C. 2.a.	Mandatory (Y/N):	N
<p>Statement:</p> <p>LOCKOUTS AND TAGOUTS. GUIDELINES. Lockout and Tagout Implementation. If a device (e.g., for energy or toxic material isolation) has the capability of being locked out, it should be locked out.</p>			

Statement Number:	Ch. IX.C. 2.b.	Mandatory (Y/N):	N
<p>Statement:</p> <p>LOCKOUTS AND TAGOUTS. GUIDELINES. Lockout and Tagout Implementation. If the isolating device cannot be locked out, it should be tagged out.</p>			

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch.IX.C. 2.c.	Mandatory (Y/N):	N
<p>Statement:</p> <p>LOCKOUTS AND TAGOUTS. GUIDELINES. Lockout and Tagout Implementation. New equipment design and major modifications to existing equipment should be designed to provide the capability of being locked out.</p>			
Statement Number:	Ch.IX.C. 3.a.	Mandatory (Y/N):	N
<p>Statement:</p> <p>LOCKOUTS AND TAGOUTS. GUIDELINES. Protective Materials and Hardware. Locks, tags, chains, wedges, key blocks, adapter pins, self-locking fasteners, or other hardware should be provided for isolating, securing, or blocking machines or equipment from energy sources.</p>			
Statement Number:	Ch.IX.C. 3.b.(1)	Mandatory (Y/N):	N
<p>Statement:</p> <p>LOCKOUTS AND TAGOUTS. GUIDELINES. Protective Materials and Hardware. Lockout and Tagout devices should be singularly identified (i.e., should be the only devices used for controlling energy and should not be used for other purposes) and should meet the following requirements:</p> <p>Durable.</p> <ul style="list-style-type: none"> - Should be capable of withstanding the environment to which they are exposed for the maximum period of time for exposure is expected, - Should be constructed to minimize deterioration of the tag or the message when exposed to weather conditions, - Should not deteriorate when used in corrosive environments where acid and/or alkali chemicals are handled and/or stored. 			
Statement Number:	Ch.IX.C. 3.b.(2)	Mandatory (Y/N):	N
<p>Statement:</p> <p>LOCKOUTS AND TAGOUTS. GUIDELINES. Protective Materials and Hardware. Lockout and Tagout devices should be singularly identified (i.e., should be the only devices used for controlling energy and should not be used for other purposes) and should meet the following requirements:</p> <p>Standardized. Lockout and Tagout devices should be standardized within the facility in at least one of the following criteria: color, shape, or size. In addition, Tagout devices should utilize standardize print and format.</p>			

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch.IX.C. 3.b.(3)	Mandatory (Y/N):	N
<p>Statement:</p> <p>LOCKOUTS AND TAGOUTS. GUIDELINES. Protective Materials and Hardware. Lockout and Tagout devices should be singularly identified (i.e., should be the only devices used for controlling energy and should not used for other purposes) and should meet the following requirements: Substantial. Lockout devices should be substantial enough to prevent removal, without the use of excessive force or unusual techniques, such as the use of bolt cutters or other metal cutting tools.</p>			
Statement Number:	Ch.IX.C. 3.b.(4)	Mandatory (Y/N):	N
<p>Statement:</p> <p>LOCKOUTS AND TAGOUTS. GUIDELINES. Protective Materials and Hardware. Lockout and Tagout devices should be singularly identified (i.e., should be the only devices used for controlling energy and should not used for other purposes) and should meet the following requirements: Identifiable. Tagout/Lockout devices should indicate the identity of the personnel and the organization applying the device.</p>			
Statement Number:	Ch.IX.C. 3.b.(5)	Mandatory (Y/N):	N
<p>Statement:</p> <p>LOCKOUTS AND TAGOUTS. GUIDELINES. Protective Materials and Hardware. Lockout and Tagout devices should be singularly identified (i.e., should be the only devices used for controlling energy and should not used for other purposes) and should meet the following requirements: Warning. Tagout devices should warn against hazardous conditions if the machine or equipment is energized and should include a legend such as the following: Do Not Start, Do Not Open; Do Not Close; Do Not Energize; Do Not Operate.</p>			
Statement Number:	Ch.IX.C. 4.	Mandatory (Y/N):	N
<p>Statement:</p> <p>LOCKOUTS AND TAGOUTS. GUIDELINES. Lockout/Tagout Program. ...should be established consisting of procedures to control potentially hazardous energy and materials and personnel training. This program should ensure that potentially hazardous energy or toxic material sources are isolated and rendered inoperative during servicing or maintenance or in any case where unexpected energizing, startup, or release of stored energy or toxic material can cause injury.</p>			

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch.IX.C. 5.	Mandatory (Y/N):	N
<p>Statement:</p> <p>LOCKOUTS AND TAGOUTS. GUIDELINES. Procedures for Lockout/Tagout. ...should be developed, documented, validated, and utilized for control of potentially hazardous energy or material. Procedures should clearly and specifically state the scope, purpose, authorization, rules, and techniques of the Lockout/Tagout program.</p>			
Statement Number:	Ch.IX.C. 5.a.	Mandatory (Y/N):	N
<p>Statement:</p> <p>LOCKOUTS AND TAGOUTS. GUIDELINES. Procedures for Lockout/Tagout. ...should include, but not be limited to, the following:</p> <ol style="list-style-type: none"> (1) Specific statement of intent of use; (2) Specific procedural steps for isolating, blocking and securing machines or equipment for hazardous energy or material; (3) Specific procedural steps for the placement, removal, and transfer of the Lockout/Tagout device(s); and (4) Specific requirements to test machines and to determine and verify the effectiveness of Lockout, Tagout or other control measures. 			
Statement Number:	Ch.IX.C. 5.b.	Mandatory (Y/N):	N
<p>Statement:</p> <p>LOCKOUTS AND TAGOUTS. GUIDELINES. Procedures for Lockout/Tagout. It is not necessary to document the required procedures for a particular machine or equipment, when all of the following elements exist.</p> <ol style="list-style-type: none"> (1) The machine or equipment has no potential for stored or residual energy or reaccumulation of stored energy after shutdown which could endanger personnel. (2) The machine or equipment has a single energy source which can be readily identified and isolated. (3) The isolation and locking out of that energy source will completely de-energize and de-activate the machine or equipment. (4) The machine or equipment is isolated from that energy source and locked out during servicing or maintenance. (5) A single lockout device will achieve a locked out condition. (6) The lockout device is under the exclusive control of the authorized personnel performing servicing of maintenance. (7) The servicing or maintenance does not create hazards to other personnel. (8) The employer, in utilizing the exception, has had no accidents involving the unexpected activation or re-energization of the machine or equipment during servicing or maintenance. 			

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch.IX.C. 5.c.	Mandatory (Y/N):	N
<p>Statement:</p> <p>LOCKOUTS AND TAGOUTS. GUIDELINES. Procedures for Lockout/Tagout. Documentation of Lockout/Tagout. Lockout/Tagout placement, activation and removal should be recorded including any information relevant to their occurrence. This record should be maintained by the shift supervisor or appropriate manager to ensure accuracy and completeness. The Lockout/Tagout record should be reviewed periodically by operations and maintenance personnel as guidance.</p>			
Statement Number:	Ch.IX.C. 6.a.	Mandatory (Y/N):	N
<p>Statement:</p> <p>LOCKOUTS AND TAGOUTS. GUIDELINES. Application of Lockout/Tagout. The established procedure for the application of energy control (implementing lockout/tagout) should cover the elements and actions in the following sequence: Preparation for Shutdown. Before a machine is turned off, personnel should be informed of the type and magnitude of the energy, the hazards and the methods of control.</p>			
Statement Number:	Ch.IX.C. 6.b.	Mandatory (Y/N):	N
<p>Statement:</p> <p>LOCKOUTS AND TAGOUTS. GUIDELINES. Application of Lockout/Tagout. The established procedure for the application of energy control (implementing Lockout/Tagout) should cover the elements and actions in the following sequence: Machine or Equipment Shutdown. The equipment should be shutdown using the procedures required by this chapter. An orderly shutdown should be utilized to avoid any added hazard.</p>			
Statement Number:	Ch.IX.C. 6.c.	Mandatory (Y/N):	N
<p>Statement:</p> <p>LOCKOUTS AND TAGOUTS. GUIDELINES. Application of Lockout/Tagout. The established procedure for the application of energy control (implementing Lockout/Tagout) should cover the elements and actions in the following sequence: Equipment Isolation. All Lockout/Tagout devices needed for control are physically located and operated in such a manner as to isolate the equipment from the energy source.</p>			

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch. IX.C. 6.d.	Mandatory (Y/N):	N
<p>Statement:</p> <p>LOCKOUTS AND TAGOUTS. GUIDELINES. Application of Lockout/Tagout. The established procedure for the application of energy control (implementing Lockout/Tagout) should cover the elements and actions in the following sequence: Affixing Locks/Tags. Lockout or Tagout devices should be affixed to each isolation device by qualified personnel in a manner that will hold the devices in a "safe" or "off" position or clearly indicate that operation of the device is prohibited.</p>			

Statement Number:	Ch. IX.C. 6.e.	Mandatory (Y/N):	N
<p>Statement:</p> <p>LOCKOUTS AND TAGOUTS. GUIDELINES. Application of Lockout/Tagout. The established procedure for the application of energy control (implementing Lockout/Tagout) should cover the elements and actions in the following sequence: Stored Energy. Upon application of Lockout or Tagout devices, all potentially hazardous stored or residual energy should be relieved, disconnected, restrained, or otherwise rendered safe. If stored energy might re-accumulate to a hazardous level, verification of isolation should be continued until the servicing or maintenance is completed.</p>			

Statement Number:	Ch. IX.C. 6.f.	Mandatory (Y/N):	N
<p>Statement:</p> <p>LOCKOUTS AND TAGOUTS. GUIDELINES. Application of Lockout/Tagout. The established procedure for the application of energy control (implementing Lockout/Tagout) should cover the elements and actions in the following sequence: Verification of Installation. Prior to starting work on equipment that has been locked or tagged out, personnel should verify that isolation and de-energization have been accomplished.</p>			

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch.IX.C. 6.g.(1)	Mandatory (Y/N):	N
<p>Statement:</p> <p>LOCKOUTS AND TAGOUTS. GUIDELINES. Application of Lockout/Tagout. The established procedure for the application of energy control (implementing Lockout/Tagout) should cover the elements and actions in the following sequence: Release from Lockout/Tagout. Before Lockout or Tagout devices are removed and energy restored to the equipment, procedures should be followed and actions taken by personnel to ensure the following: Equipment/workspace. The machine or equipment is operationally intact. The area has been inspected to ensure that nonessential items have been removed. The individual authorizing tag removal should specify the final component position and, when appropriate, the sequence in which components should be positioned. In addition, the need to check the positioning of other components that were not locked or tagged, but are related to the operation of the Lockout/Tagout component, should be determined at this time and appropriate instructions should be developed. These checks should be done when necessary, to ensure that components within the Locked/Tagout boundaries are correctly aligned to support operation.</p>			

Statement Number:	Ch.IX.C. 6.g.(2)	Mandatory (Y/N):	N
<p>Statement:</p> <p>LOCKOUTS AND TAGOUTS. GUIDELINES. Application of Lockout/Tagout. The established procedure for the application of energy control (implementing Lockout/Tagout) should cover the elements and actions in the following sequence: Release from Lockout/Tagout. Before Lockout or Tagout devices are removed and energy restored to the equipment, procedures should be followed and actions taken by personnel to ensure the following: Personnel. The workspace should be checked to ensure that all personnel have been positioned safely or removed from the area. Before Lockout/Tagout devices are removed and equipment energized, affected personnel should be notified that equipment will be energized.</p>			

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch.IX.C. 6.g.(3)	Mandatory (Y/N):	N
<p>Statement:</p> <p>LOCKOUTS AND TAGOUTS. GUIDELINES. Application of Lockout/Tagout.</p> <p>The established procedure for the application of energy control (implementing Lockout/Tagout) should cover the elements and actions in the following sequence: Release from Lockout/Tagout. Before Lockout or Tagout devices are removed and energy restored to the equipment, procedures should be followed and actions taken by personnel to ensure the following: Lockout/Tagout Device Removal. Each Lockout/Tagout device should be removed from each energy-isolating device by the person who applied the device. He/she should reposition the components in the sequence specified and check the positioning as deemed necessary. The removal of Locks or Tags should be documented. Tags should be returned to the operations supervisor or authorizing manager. The supervisor/manager should be satisfied that all tags and safety devices associated with this Lockout/Tagout are removed and are accountable.</p> <p>When the person who applied the Lockout/Tagout device is not available to remove it, that device may be removed under the direction of the appropriate supervisor/manager, provided that specific procedures and training for such removal have been developed and incorporated into the facility program. The supervisor should demonstrate that the specific procedure provides equivalent safety to the removal of the device by the authorized personnel who applied it. The specific procedure should include the following elements:</p> <ul style="list-style-type: none"> - Verification by the supervisor/manager that the person who applied the device is not available. - Assurance that all reasonable efforts are made to inform him/her that the device has been removed. - Ensuring that all relevant personnel have this knowledge before he/she resumes work at the facility. 			

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch.IX.C. 7.	Mandatory (Y/N):	N
<p>Statement:</p> <p>LOCKOUTS AND TAGOUTS. GUIDELINES. Testing or Positioning of Equipment or Components. Temporary removal of Lockout/Tagout devices should be discouraged. However, situations in which lockout/tagout devices must be temporarily removed and the equipment energized, the following sequence of operations should be followed:</p> <ol style="list-style-type: none"> 1. Clear the equipment of tools and materials. 2. Remove personnel from the equipment area. 3. Remove the Lockout/Tagout device as procedures indicate. 4. Energize and proceed with testing or positioning. 5. De-energized all systems and reapply the Lockout/tagout. 			

Statement Number:	Ch.IX.C. 8.	Mandatory (Y/N):	N
<p>Statement:</p> <p>LOCKOUTS AND TAGOUTS. GUIDELINES. Periodic Inspections. ...should be conducted by authorized personnel, supervisor, or appropriate manager, to determine whether procedures are being followed and to correct any deviations or inadequacies observed. Inspections should include a review of the responsibilities of personnel and supervisors. The supervisor or appropriate manager should certify that the periodic inspections have been performed, documenting the equipment and procedures involved, dates of inspection, personnel participating in the inspections, and personnel performing the inspections.</p>			

Statement Number:	Ch.IX.C. 9.	Mandatory (Y/N):	N
<p>Statement:</p> <p>LOCKOUTS AND TAGOUTS. GUIDELINES. Caution Tags. ...should not be used for personnel protection (i.e., caution tags should not be used where it is appropriate to use a Lockout or Tagout device). Administration of caution tags could, however, be accomplished as part of the Lockout/Tagout program or it could be covered separately. The use of caution tags should be restricted to those situations in which a component or system is functional, but when some precaution or item(s) of information is necessary prior to operation.</p>			

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch.IX.C. 9.a.	Mandatory (Y/N):	N
<p>Statement:</p> <p>LOCKOUTS AND TAGOUTS. GUIDELINES. Caution Tags. In addition, the program should include the following elements: Caution tags should be uniquely identifiable and different in appearance from other station tags. The following information should be included on the tags as an aid in administer the program:</p> <ol style="list-style-type: none"> (1) caution tag number; (2) component name and number; (3) effective date; (4) precaution or information applicable to the particular; situation and/or component or system; and (5) signature of authorizing individual and organization. 			

Statement Number:	Ch.IX.C. 9.b.	Mandatory (Y/N):	N
<p>Statement:</p> <p>LOCKOUTS AND TAGOUTS. GUIDELINES. Caution Tags. In addition, the program should include the following elements: Situations that require special operator or maintenance precautions or amplifying information should be brought to the attention of the shift supervisor or appropriate manager. The supervisor or manager should ensure that issuing a caution tag is necessary and that the tag is not being used in place of more appropriate administrative action (such as a temporary procedure change, placing an operator aide, use of the work control system, or issuance of a safety Lockout or Tagout). The supervisor review should also ensure that any instruction contained in the caution tag does not deviate from established facility procedures, technical specifications, or OSR's.</p>			

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch.IX.C. 9.c.	Mandatory (Y/N):	N
Statement: LOCKOUTS AND TAGOUTS. GUIDELINES. Caution Tags. In addition, the program should include the following elements: A record of all active caution tags and associated amplifying information should be available to the appropriate personnel. This record and associated tags should be reviewed periodically by qualified personnel. This review should verify the continued need and applicability for each caution tag and ensure that the caution tag index accurately reflects all active caution tags. This review should be documented. Any caution tags remaining in an active status for extended periods (e.g., longer than three months, as appropriate) should be brought to the attention of the operations supervisor or manager. The operations supervisor or manager should determine what action is needed to resolve the continued use of the caution tag.			

Statement Number:	Ch.IX.C. 9.d.	Mandatory (Y/N):	N
Statement: LOCKOUTS AND TAGOUTS. GUIDELINES. Caution Tags. In addition, the program should include the following elements: Caution Tag Placement. ...should be placed in such a way that they do not interfere with or obscure indications, switches, or other control devices but are readily apparent to an individual prior to the operation of the tagged device. Placement of caution tags should be documented.			

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch.IX.C.10.a.	Mandatory (Y/N):	N
<p>Statement:</p> <p>LOCKOUTS AND TAGOUTS. GUIDELINES. Training and Communication. Training should be provided and documented to ensure that the purpose and function of the Lockout/Tagout program is understood by all personnel and that they have the knowledge and skills required for safe application, use, and removal of Lockouts and Tagouts. Training should include the following:</p> <p>(1) Recognition of applicable hazardous energy sources, type and magnitude of energy or materials involved and methods to isolate and control.</p> <p>(2) Personnel involved should be instructed in the purpose and use of the Lockout/Tagout procedures.</p> <p>(3) All personnel who may enter the area should be trained to recognize the Lockout/Tagout and to understand the procedure and prohibition relating to attempts to restart or re-energize machines or equipment.</p>			

Statement Number:	Ch.IX.C.10.b.	Mandatory (Y/N):	N
<p>Statement:</p> <p>LOCKOUTS AND TAGOUTS. GUIDELINES. Training and Communication. When Tagout systems are used, personnel should be trained in the limitations of tags:</p> <p>(1) Tags are essentially warning devices and do not provide physical restraint on those devices as are provided by a lock.</p> <p>(2) When a Tag is attached, it is not to be removed without authorization of the person responsible for it, it should not be bypassed, ignored, or otherwise defeated.</p> <p>(3) Tags must be legible and understandable by all personnel.</p> <p>(4) Tags and their means of attachment should be able to withstand the environmental conditions encountered in the workplace.</p> <p>(5) Tags may evoke a false sense of security and their meaning needs to be understood.</p> <p>(6) Tags must be securely attached so they cannot be inadvertently or accidentally detached during use.</p>			

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch.IX.C.10.c.	Mandatory (Y/N):	N
<p>Statement:</p> <p>LOCKOUTS AND TAGOUTS. GUIDELINES. Training and Communication. When Lockout systems are used, personnel should be trained in the limitations of Locks, especially concerning the following:</p> <p>(1) Operation of the facility may be hindered (this is significant when local component operations are necessary, such as remote shutdown or remote control).</p> <p>(2) Locks and chains installed on small instrument line isolation valves can contribute to seismic loading and may not have been considered during safety analysis.</p>			

Statement Number:	Ch.IX.C.10.d.	Mandatory (Y/N):	N
<p>Statement:</p> <p>LOCKOUTS AND TAGOUTS. GUIDELINES. Retraining. ...should be provided for all authorized or affected personnel whenever there is a change in job, procedures, machines, equipment or processes that present new hazards or periodic inspection determines it is needed.</p>			

Statement Number:	Ch.IX.C.11.	Mandatory (Y/N):	N
<p>Statement:</p> <p>LOCKOUTS AND TAGOUTS. GUIDELINES. Lockout or Tagout Implementation. Lockout or Tagout should only be accomplished by authorized, qualified personnel.</p>			

Statement Number:	Ch.IX.C.12.	Mandatory (Y/N):	N
<p>Statement:</p> <p>LOCKOUTS AND TAGOUTS. GUIDELINES. Notification of Personnel. A supervisor or appropriate manager should notify affected personnel of the application and removal of Lockout/Tagout devices. Notification should be given before the devices are applied and after they are removed.</p>			

Statement Number:	Ch.IX.C.13.	Mandatory (Y/N):	N
<p>Statement:</p> <p>LOCKOUTS AND TAGOUTS. GUIDELINES. Outside Contractors. Whenever outside servicing personnel are to be engaged in activities covered by the scope of the Lockout/Tagout program, the facility and the outside contractor should inform each other of their respective Lockout/Tagout procedures and should ensure that their personnel are aware of the changes.</p>			

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch.IX.C.14.	Mandatory (Y/N):	N
<p>Statement:</p> <p>LOCKOUTS AND TAGOUTS. GUIDELINES. Group Lockouts or Tagouts. When servicing or maintenance is performed by a crew, craft, department, or other group, they should utilize a procedure that will provide for equivalent safety provided by the personal Lockout or Tagout device.</p>			
Statement Number:	Ch.IX.C.15.	Mandatory (Y/N):	N
<p>Statement:</p> <p>LOCKOUTS AND TAGOUTS. GUIDELINES. Shift or Personnel Changes. Specific procedures should be utilized during shift of personnel changes to ensure that continuity of Lockout or Tagout protection, including provision for the orderly transfer of Lockout or Tagout devices between personnel or shifts.</p>			

The following statements concerning "INDEPENDENT VERIFICATION" are
N/A:

Statement Number:	Ch.X.C.1.	Mandatory (Y/N):	N
<p>Statement:</p> <p>INDEPENDENT VERIFICATION. GUIDELINES. Components Requiring Independent Verification. Components that are critical to ensure safe and reliable operation should receive an independent verification of their position when circumstances warrant. These components should be identified explicitly in facility procedures or other official documents so that unnecessary interpretation of requirements will be minimized. ... independent verification should not be limited to safety-related system components.</p>			

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch.X.C.1.a.	Mandatory (Y/N):	N
<p>Statement:</p> <p>INDEPENDENT VERIFICATION. GUIDELINES. Components Requiring Independent Verification. Each facility should use an accepted safety analysis method (e.g., fault-free analysis, Probability Risk Analysis) and/or expert opinion to determine its own facility specific list of systems and components, which should receive independent verification in accordance with the following guidelines: Safety-Related Systems. All components in systems that have safety-related functions should be considered for independent verification. However, if the following criteria have been met, some components might not need independent verification:</p> <ol style="list-style-type: none"> (1) Mispositioning would not affect system performance. (2) The mispositioning would be known immediately to an operator. (3) Significant radiation exposure would be received by the person(s) performing the independent verification. <p>In this case, alternate means for independent verification that do not involve radiation exposure, such as observing process parameters (Section C.3), should be considered. Variances from independent verification requirements in accordance with the above criteria should be approved by the operations supervisor.</p>			

Statement Number:	Ch.X.C.1.b.	Mandatory (Y/N):	N
<p>Statement:</p> <p>INDEPENDENT VERIFICATION. GUIDELINES. Components Requiring Independent Verification. Each facility should use an accepted safety analysis method (e.g., fault-tree analysis, Probability Risk Analysis) and/or expert opinion to determine its own facility specific list of systems and components, which should receive independent verification in accordance with the following guidelines: Systems Not Related to Safety. Independent verifications should be performed for certain components that are not related to safety which, if mispositioned, could lead to unplanned shutdowns, challenges to safety systems, or (through interfaces with safety-related systems) cause radioactive or toxic material release.</p>			

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch.X.C.2.[1]	Mandatory (Y/N):	N
<p>Statement:</p> <p>INDEPENDENT VERIFICATION. GUIDELINES. Occasions Requiring Independent Verification. Components should receive independent verification when the equipment they serve must be available and when a reasonable possibility exists that the components might have been mispositioned.</p>			

Statement Number:	Ch.X.C.2.[2]	Mandatory (Y/N):	N
<p>Statement:</p> <p>INDEPENDENT VERIFICATION. GUIDELINES. Occasions Requiring Independent Verification. The following guidelines describe situations where independent verification would be appropriate:</p> <p>a. Returning Equipment to Service Following Maintenance or Testing. Independent verifications should be performed to ensure that systems are properly aligned when equipment is returned to service. After maintenance activities, it is appropriate to verify independently the positions of components that have been tagged to establish an isolation boundary for the work. Also, additional components that could have been mispositioned during maintenance should be selected for verification on a case-by-case basis by the individual authorizing equipment for return to service. ...Selected components should therefore be independently verified during or after surveillance tests.</p> <p>b. Removing Equipment From Service. To ensure that only the specific items of equipment intended to be removed from service are affected by Tagouts, independent verifications should be performed when equipment is removed from service.</p> <p>c. System Lineups. ...When system lineups are performed, the considerations of Section C1, Components Requiring Independent Verifications, should be applied to determine the need for independent verification.</p> <p>d. Periodic Checks During Facility Operation. In order to verify that all associated equipment is fully functional, DOE facilities should perform routine, periodic verifications of certain critical components during normal operations.</p>			

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch.X.C.3.	Mandatory (Y/N):	N
<p>Statement:</p> <p>INDEPENDENT VERIFICATION. GUIDELINES. Verification Techniques. Operators should be trained in the appropriate techniques for verifying the positions of all facility components. In addition, a reference document should be provided explaining what an operator should do when verifying a component position. ...Each facility should use its most experienced personnel to develop these instructions, using applicable manufacturer's recommendations where available. The instructions should describe the techniques for verifying manual valves, motor- and air-operated valves, solenoid-operated valves, circuit breakers, blank flanges, removable links and fuses, and the availability of control power.</p>			

Statement Number:	Ch.X.C.3.a.	Mandatory (Y/N):	N
<p>Statement:</p> <p>INDEPENDENT VERIFICATION. GUIDELINES. Verification Techniques. The following general guidelines apply to component verification techniques: Independence. Independent verifications should be conducted in a manner such that each check constitutes an actual identification of the component and a determination of both its required and actual positions. To be independent, the integrity of the checks must be maintained by minimizing interaction between the personnel operating components and those performing the independent verifications.</p>			

Statement Number:	Ch.X.C.3.b.	Mandatory (Y/N):	N
<p>Statement:</p> <p>INDEPENDENT VERIFICATION. GUIDELINES. Verification Techniques. The following general guidelines apply to component verification techniques: Remote Position Indicators. ...independent verification should be performed locally unless ALARA or other overriding factors preclude. In this case, independent verification may be done at a remote position indicator.</p>			

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch.X.C.3.c.	Mandatory (Y/N):	N
<p>Statement:</p> <p>INDEPENDENT VERIFICATION. GUIDELINES. Verification Techniques. The following general guidelines apply to component verification techniques: Process Parameters. ...care must be exercised when using process parameters because alternate flow paths or other factors could cause them to be misleading indicators of component position. ...process parameters normally should not be used as the only indication of a component's position. Facility procedures should specify where and when process parameters are acceptable indicators of component position.</p>			

Statement Number:	Ch.X.C.3.d.	Mandatory (Y/N):	N
<p>Statement:</p> <p>INDEPENDENT VERIFICATION. GUIDELINES. Verification Techniques. The following general guidelines apply to component verification techniques: Throttled Valves. Verifying throttled valves by shutting and reopening the valve by a prescribed number of turns should be avoided because this practice has the potential to create valve mispositionings. Instead, position indicators, scribe marks, or other officially recognized and designated indicators should be used to determine throttled valve positions. When shutting and reopening a throttled valve are necessary to determine its position, having a verifier observe a valve operator's actions would be preferable to having both persons independently shut and reopen the valve.</p>			

Statement Number:	Ch.X.C.3.e.	Mandatory (Y/N):	N
<p>Statement:</p> <p>INDEPENDENT VERIFICATION. GUIDELINES. Verification Techniques. The following general guidelines apply to component verification techniques: Surveillance Testing. In some cases, surveillance testings may be substituted for normal verification techniques. ...However, surveillance tests frequently will not serve to verify the position of all components that are important to subsequent system's operation. ...Surveillance testing therefore should not be used as independent verification, unless it can be shown that the test conclusively proves the position of the components in question. The applicability of surveillance testing to satisfy independent verification requirements should be approved beforehand by the operations supervisor.</p>			

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch.X.C.3.f.	Mandatory (Y/N):	N
Statement: INDEPENDENT VERIFICATION. GUIDELINES. Verification Techniques. The following general guidelines apply to component verification techniques: Operations Self-Appraisal and Verification. Programmatic operations appraisals should be independently conducted and/or verified to ensure that environment, safety, and health considerations, and operations functions (such as training) are being conducted in accordance with established operational criteria.			

The above statements concerning "INDEPENDENT VERIFICATION" are
N/A:

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch.XI.C.1.	Mandatory (Y/N):	N
<p>Statement:</p> <p>LOGKEEPING. GUIDELINES. Establishment of Operating Logs. Narrative logs should be established at all key shift positions. As a minimum, a narrative log should be maintained by the operations supervisor or the control area operator. For those shift positions that are manned on a part-time basis (e.g., radwaste operator, makeup plant, etc.) a log should be established to ensure that pertinent information is passed from operator to operator. A narrative section should be provided on round sheets when a narrative log is not used for a particular shift position.</p>			

- 1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): NO
- 2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):
- The HASQAP does not address shift logs.
- 3.0 CORRECTIVE ACTION:

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch.XI.C.2.	Mandatory (Y/N):	N
Statement: LOGKEEPING. GUIDELINES. Timeliness of Recordings. Information should be promptly recorded in the logs.			

- 1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): NO
- 2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):
- The HASQAP does not address shift logs.
- 3.0 CORRECTIVE ACTION:

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch.XI.C.3.[1]	Mandatory (Y/N):	N
<p>Statement:</p> <p>LOGKEEPING. GUIDELINES. Information to be Recorded. The operations supervisor should provide written guidance to define thoroughly the type and scope of entries for each log and the format for making entries. The following information should be recorded in at least one station log, although any one log might not contain all these items:</p> <ul style="list-style-type: none"> a. Facility mode or condition changes...; b. Criticalities and appropriate critical data (for DOE reactors); c. Abnormal facility configurations; d. Status changes to safety-related and other major facility equipment; e. Occurrence of any reportable events; f. Initiation and completion of surveillance tests; g. Entering and exiting operational limit actions; h. Security incidents; i. Out-of-specification chemistry or process results; and j. Shift reliefs. <p>To aid in event reconstruction, as much significant information as possible should be logged during emergencies and abnormal or unexpected events. However, logkeeping should not take precedence over controlling and monitoring the facility.</p>			

- 1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): NO
- 2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):
- The HASQAP does not address the intent of this statement.
- 3.0 CORRECTIVE ACTION:

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch.XI.C.4.	Mandatory (Y/N):	N
Statement: LOGKEEPING. GUIDELINES. Legibility. Log entries should be made in a manner such that they can be easily read and understood. Additionally, the log entries should be readily reproducible with standard photocopy machines. For these reasons, black, waterproof ink is recommended.			

- 1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): NO
- 2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):
This statement is not addressed in the HASQAP.
- 3.0 CORRECTIVE ACTION:

ASSESSMENT DATA COLLECTION FORM

Statement Number: Ch. XI, C. 5.	Mandatory (Y/N): N
<p>Statement:</p> <p>LOGKEEPING. GUIDELINES. Corrections. A standardized method for correcting any erroneous entries should be established. The method chosen should not obscure the incorrect entry. Log entries should not be erased or covered up. Corrections can be made by placing a single line through the incorrect entry and writing the correct entry in a nearby space. All corrections should be initialed and dated.</p>	

- 1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): NO
- 2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):
- The HASQAP does not address corrections to logs.
- 3.0 CORRECTIVE ACTION:

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch.XI.C.6.	Mandatory (Y/N):	N
Statement: LOGKEEPING. GUIDELINES. Log Review. Control area logs should be reviewed periodically by the operations supervisor and logs kept by operators outside the control area should be reviewed by the control area operator or an appropriate supervisor. These reviews should ensure that entries are accurate and adequate.			

- 1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): NO
- 2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):
- The HASQAP does not address the intent of this statement.
- 3.0 CORRECTIVE ACTION:

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch.XI.C.7.	Mandatory (Y/N):	N
<p>Statement:</p> <p>LOGKEEPING. GUIDELINES. Care and Keeping of Logs. Management should provide written guidance on the disposition of completed logs. This guidance should address the following:</p> <p>a. Making the logs available for operators who return after periods of absence from their regular duties;</p> <p>b. Storing the completed logs to ensure preservation for the expected life of the facility; and</p> <p>c. Retrieving stored logs should this become necessary.</p>			

- 1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): NO
- 2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):
- The HASQAP does not address this statement.
- 3.0 CORRECTIVE ACTION:

ASSESSMENT DATA COLLECTION FORM

The following statements concerning "OPERATIONS TURNOVER" are
N/A:

Statement Number:	Ch.XII.C.1.a.	Mandatory (Y/N):	N
Statement: OPERATIONS TURNOVER. GUIDELINES. Turnover Checklists. As a minimum, supervisory positions should have a turnover checklist to be used in the turnover process. Equipment operator checklists..., or other formal documents, should provide for noting major components status (e.g., operating pumps, trains in service, etc.), abnormal lineups, valid alarms on all pertinent local control panels, and surveillances or evolutions planned or in progress. Provisions should also be made for documenting that a review of round sheets and narrative logs has been performed by the oncoming operator.			

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch.XII.C.1.b.	Mandatory (Y/N):	N
-------------------	---------------	------------------	---

Statement:

OPERATIONS TURNOVER. GUIDELINES. Turnover Checklists. As a minimum, supervisory positions should have a turnover checklist to be used in the turnover process. Operator Checklists. ...or other documents reviewed at shift change should provide for recording vital information about the following facility status. Examples may include:

- (1) Facility power level, test status, or equivalent;
- (2) Key temperatures limit status;
- (3) Key pressures limit status;
- (4) Important tank levels;
- (5) Safety equipment status;
- (6) Operational limits in effect;
- (7) Other important parameters;
- (8) Required chemistry or process sample times;
- (9) Maintenance, surveillances, tests, or evolutions (in progress or planned);
- (10) Problems experienced with equipment and major equipment out of service;
- (11) When specific planned actions are required;
- (12) Changes in radiological or hazardous materials conditions; and
- (13) Temporary procedure changes in effect.

Operator checklists may require documenting reviews of certain facility status documents. These may include the operator's own operating (narrative) log, the Tagout (clearance) log, the temporary modification and jumper logs, the out-of-service annunciator log, the operational limit action log, the shift order books, and required reading since the last shift. Each operator turnover checklist should be provided with enough space for the offgoing operator to list other important information his relief should have.

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch.XII.C.1.c.	Mandatory (Y/N):	N
<p>Statement:</p> <p>OPERATIONS TURNOVER. GUIDELINES. Turnover Checklists. As a minimum, supervisory positions should have a turnover checklist to be used in the turnover process. Operations Supervisory Checklists. ...or other documents reviewed at shift change should provide vital information about facility status. Examples may include:</p> <ol style="list-style-type: none"> (1) Facility status; (2) Evolutions (completed, in progress, or planned); (3) Controlled key status; (4) Abnormal lineups or conditions; (5) Changes in radiological or hazardous substance conditions; (6) Surveillance tests planned or in progress; (7) Maintenance planned or in progress; and (8) Waste management status. <p>Operations supervisory checklists should require documenting review of certain status documents. These can include the supervisor's own narrative log, the status log, the equipment status checklist, the emergency safeguards equipment list, shift order books, and required reading since the last shift.</p>			

Statement Number:	Ch.XII.C.2.	Mandatory (Y/N):	N
<p>Statement:</p> <p>OPERATIONS TURNOVER. GUIDELINES. Document Review. Oncoming operators and supervisors should review documents specified on their checklists prior to assuming responsibility for their shift position. Document review should be as intensive as necessary for the oncoming personnel to understand important history, present status, and planned events. Normally, narrative log entries for the previous 24-hour period or since his/her last shift should be reviewed. Status logs should be reviewed so that the operator and supervisors are familiar with all currently active entries, with an emphasis on what has changed since he/she last had the shift.</p>			

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch.XII.C.3.	Mandatory (Y/N):	N
<p>Statement:</p> <p>OPERATIONS TURNOVER. GUIDELINES. Control Panel Walkdown. Walkdowns of appropriate control panels should be conducted by each shift watchstander. ...Oncoming operations supervisors should walk down the main control panels before, during, or shortly after shift turnover. Oncoming and offgoing control area operators should walk down their unit's main control panel together. Equipment operators should review all pertinent local control panels, (e.g., fuel supplies, lubrication levels, makeup water treatment, and ventilation) during their tour early in the shift.</p>			

Statement Number:	Ch.XII.C.4.	Mandatory (Y/N):	N
<p>Statement:</p> <p>OPERATIONS TURNOVER. GUIDELINES. Discussion and Exchange of Responsibility. At a time when facility conditions are stable, the offgoing supervisor or operator should explain all items noted on the turnover checklist, and the oncoming operator or supervisor should ask any pertinent questions. When all operations personnel are satisfied that the oncoming crew is fully cognizant of the plant conditions, the oncoming operators and supervisors should state that he/she is assuming responsibility for the shift position with an entry into the log.</p>			

Statement Number:	Ch.XII.C.5.	Mandatory (Y/N):	N
<p>Statement:</p> <p>OPERATIONS TURNOVER. GUIDELINES. Shift Crew Briefing. ...should be conducted by the operations supervisor as required and may be conducted after he has accepted responsibility for the shift. The briefing should include a review of the status, problems with equipment, and evolutions in progress or planned during the shift. In addition to shift operators, personnel from support groups (e.g., contractors, vendors, and test groups) should also attend briefings when their activities can directly affect operations.</p>			

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch.XII.C.6.	Mandatory (Y/N):	N
Statement: OPERATIONS TURNOVER. GUIDELINES. Reliefs Occurring During the Shift. ... (e.g., exchange of control supervisory function) should have a turnover to ensure that the oncoming person is at least as knowledgeable of the conditions as he would have been had a complete shift turnover process been conducted.			

The above statements concerning "OPERATIONS TURNOVER" are
N/A:

ASSESSMENT DATA COLLECTION FORM

The following statements concerning "OPERATIONS ASPECTS OF FACILITY CHEMISTRY AND UNIQUE PROCESSES" are N/A:

Statement Number:	Ch.XIII.C.1.[1]	Mandatory (Y/N):	N
Statement: OPERATIONS ASPECTS OF FACILITY CHEMISTRY AND UNIQUE PROCESSES. GUIDELINES. Operator Responsibilities. The operations supervisor should define each operator's specific responsibilities with respect to process control.			

Statement Number:	Ch.XIII.C.1.[2]	Mandatory (Y/N):	N
Statement: OPERATIONS ASPECTS OF FACILITY CHEMISTRY AND UNIQUE PROCESSES. GUIDELINES. Operator Responsibilities. In general, operators should monitor chemistry or process parameters using control room instruments and other instruments related to equipment under operations control.			

Statement Number:	Ch.XIII.C.1.[3]	Mandatory (Y/N):	N
Statement: OPERATIONS ASPECTS OF FACILITY CHEMISTRY AND UNIQUE PROCESSES. GUIDELINES. Operator Responsibilities. Operators should be able to recognize out-of-specification conditions or adverse trends and be familiar with corrective actions associated with chemistry or process problems.			

Statement Number:	Ch.XIII.C.1.[4]	Mandatory (Y/N):	N
Statement: OPERATIONS ASPECTS OF FACILITY CHEMISTRY AND UNIQUE PROCESSES. GUIDELINES. Operator Responsibilities. Operators should consult and be advised by chemistry or technical process department personnel.			

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch. XIII.C.2.	Mandatory (Y/N):	N
<p>OPERATIONS ASPECTS OF FACILITY CHEMISTRY AND UNIQUE PROCESSES. GUIDELINES. Operators should be knowledgeable about aspects of facility processes and safety that affect operation and should be able to analyze off-normal situations and take appropriate action to correct the causes(s) of problems. Examples of the types of concepts and processes with which operations personnel should be familiar include the following:</p> <p>a. Concepts of pH, conductivity, cation conductivity, and dissolved oxygen and the effect these parameters have on the facility.</p>			

Statement Number:	Ch. XIII.C.3.	Mandatory (Y/N):	N
<p>OPERATIONS ASPECTS OF FACILITY CHEMISTRY AND UNIQUE PROCESSES. GUIDELINES. Operator Response to Process Problems. Each operator should be capable of correctly interpreting the chemistry or process parameters that he is responsible for and be able to provide appropriate, timely, corrective action when required.</p>			

Statement Number:	Ch. XIII.C.4.	Mandatory (Y/N):	N
<p>OPERATIONS ASPECTS OF FACILITY CHEMISTRY AND UNIQUE PROCESSES. GUIDELINES. Communication Between Operations and Process Personnel. If possible, operators should inform appropriate process personnel prior to commencing evolutions that could potentially affect facility processes or require action from support technicians. Operators should receive reports from the process department on key day-to-day process results and problems. When process-related problems occur, operations and process-personnel should work closely to provide prompt corrective action.</p>			

The above statements concerning "OPERATIONS ASPECTS OF FACILITY CHEMISTRY AND UNIQUE PROCESSES" are N/A:

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch.XIV.C.1.	Mandatory (Y/N):	N
<p>Statement:</p> <p>REQUIRED READING. GUIDELINES. File Index. A list of the types of documents to be included in the required-reading file should be maintained. The list should include the following documents:</p> <ul style="list-style-type: none"> a. Procedure changes; b. Equipment design changes; c. Related industry and in-house operating experience information; and d. Other information necessary to keep operations department personnel aware of current facility activities. <p>Reading material should be screened to ensure that only appropriate material is placed in the file.</p>			

N/A

Statement Number:	Ch.XIV.C.2.	Mandatory (Y/N):	N
<p>Statement:</p> <p>REQUIRED READING. GUIDELINES. Reading Assignments. A method should be in place to designate which documents need to be read by the individuals filling each position, and the reading file should be readily available to those individuals.</p>			

N/A

Statement Number:	Ch.XIV.C.3.	Mandatory (Y/N):	N
<p>Statement:</p> <p>REQUIRED READING. GUIDELINES. Required Dates for Completion of Reading. ...should be designated for...each document. This date would be determined based on the nature of the material. Personnel should complete their reading assignments by the required date. Certain documents may be designated for "immediate reading." These should be read before assuming responsibility for a shift position.</p>			

N/A

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch.XIV.C.4.	Mandatory (Y/N):	N
Statement: REQUIRED READING. GUIDELINES. Documentation. Completion of reading should be documented, and the documentation should be retained.			

N/A

Statement Number:	Ch.XIV.C.5.	Mandatory (Y/N):	N
Statement: REQUIRED READING. GUIDELINES. Review. The reading file should be periodically reviewed to ensure that all department personnel complete readings by the required dates. Material that has been read by all appropriate personnel should be removed from the active file. A "dead" file may be maintained as a reference.			

N/A

Statement Number:	Ch.XV.C.1.	Mandatory (Y/N):	N
Statement: TIMELY ORDERS TO OPERATORS. GUIDELINES. Content and Format. Information such as special operations, administrative directions, special data-collection requirements, plotting process parameters, and other similar short-term matters can be included in operator orders. Examples of orders could include instructions on the need for and performance of specific evolutions or tests; it could also include work priorities, announcements of policy information, and administrative information. Orders should be clearly written, dated, and maintained in the control room. Information and policies intended as permanent should be incorporated into appropriate administrative procedures. The operator orders program should not be used to change operating procedures, because the changes noted in the operator orders might be missed by a procedure user. Additionally, operator orders may not receive the approval appropriate for a procedural change. Therefore, information intended to supplement operating procedures should be promptly incorporated into the appropriate procedure by a procedure change or revision (see Chapter XVI).			

N/A- Laboratory Analysts work to procedures, not temporary orders.

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch.XV.C.2.	Mandatory (Y/N):	N
<p>Statement:</p> <p>TIMELY ORDERS TO OPERATORS. GUIDELINES. Issuing, Segregating, and Reviewing Orders. Operator orders should be issued by the operations supervisor or his/her designee whenever necessary to communicate instructions to the shift personnel. Operator orders should be segregated into daily and long-term orders in order to facilitate review by shift personnel. Daily orders that are postponed or prolonged should have daily review or updates. Reviews of long-term orders may not be required on a daily basis but should be made periodically and when changes occur. Appropriate operations personnel should review the orders early in the shift and document their review by initialing the log or notebook.</p>			

N/A- Laboratory Analysts work to procedures, not temporary orders.

Statement Number:	Ch.XV.C.3.	Mandatory (Y/N):	N
<p>Statement:</p> <p>TIMELY ORDERS TO OPERATORS. GUIDELINES. Removal of Orders. Orders that are no longer applicable or are outdated should be promptly removed or cancelled. The operations supervisor should periodically review the operator orders to ensure that only applicable and current orders remain effective.</p>			

N/A- Laboratory Analysts work to procedures, not temporary orders.

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch.XVI.C.1.	Mandatory (Y/N):	N
Statement: OPERATIONS PROCEDURES. GUIDELINES. Procedure Development. To ensure consistency among operations procedures, the methods for developing new procedures, including procedure formats, should be clearly defined. Administrative procedures and/or writers' guides should direct the development and review process for procedures.			

1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): YES

2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):

8.2 Procedure Requirements

Each page at a minimum shall have a unique code, title, and revision number traceable to the date issued. Each page shall carry the code and revision, at a minimum. The following information is recommended for analytical and technical procedures:

- title.
- scope: parameters measured, range, matrix, expected precision, and accuracy.
- referenced documents: title, author(s), year published, publisher, document code.
- terminology used.
- summary of method.
- interferences.
- apparatus and instrumentation.
- reagents and materials.
- hazards and precautions [Material Safety Data Sheet (MSDS) references].
- sample preparation.
- apparatus and instrumentation set-up.
- calibration and standardization.

ASSESSMENT DATA COLLECTION FORM

- procedural steps.
- demonstration of statistical control or quality control criteria.
- calculations.
- assignment of uncertainty.

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch.XVI.C.2.a.	Mandatory (Y/N):	N
<p>Statement:</p> <p>OPERATIONS PROCEDURES. GUIDELINES. Procedure Content. To provide uniformity in operations procedures, the content of procedures should conform to prescribed guidelines. The procedure aspects described below should be followed when developing operations procedures: The scope and applicability of individual procedures should be readily apparent. Procedures with single-unit applicability should be distinctively identified to avoid confusion with sister-unit procedures. In addition, to enhance rapid retrieval, emergency procedures should be distinguishable from other procedures. Color coding could be used for these purposes.</p>			

1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): YES

2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):

The intent of the above statement as it applies to analytical services is met in HASQAP Section 8.2, "Procedure Requirements." (See Statement Number Ch.XVI.C.1 above.)

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch.XVI.C.2.b.	Mandatory (Y/N):	N
<p>Statement:</p> <p>OPERATIONS PROCEDURES. GUIDELINES. Procedure Content. To provide uniformity in operations procedures, the content of procedures should conform to prescribed guidelines. The procedure aspects described below should be followed when developing operations procedures: Procedures should incorporate appropriate information from applicable source documents, such as the facility design documents, safety analysis documents, and vendor technical manuals.</p>			

1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): YES

2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):

The intent of the above statement as it applies to analytical services is met in HASQAP Section 8.2, "Procedure Requirements." (See Statement Number Ch.XVI.C.1 above.)

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch.XVI.C.2.c.	Mandatory (Y/N):	N
<p>Statement:</p> <p>OPERATIONS PROCEDURES. GUIDELINES. Procedure Content. To provide uniformity in operations procedures, the content of procedures should conform to prescribed guidelines. The procedure aspects described below should be followed when developing operations procedures: Prerequisites and initial conditions should be detailed. Careful consideration should be given to the location of this information within the procedure in order to help ensure that the intent of the procedure is understood. In addition, any hoses, tools, or other temporary testing equipment should be verified operable, calibrated, or inspected and in good condition where possible, before implementing any test procedure, to ensure that they function as expected during the test. These verifications should be identified in the prerequisite section, with completion sign-offs required. "Hold" points (requiring independent verification and/or approval) should be clearly delineated.</p>			

1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): YES

2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):

The intent of the above statement as it applies to analytical services is met in HASQAP Section 8.2, "Procedure Requirements." (See Statement Number Ch.XVI.C.1 above.)

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch.XVI.C.2.d.	Mandatory (Y/N):	N
<p>Statement:</p> <p>OPERATIONS PROCEDURES. GUIDELINES. Procedure Content. To provide uniformity in operations procedures, the content of procedures should conform to prescribed guidelines. The procedure aspects described below should be followed when developing operations procedures: Definitions used in the procedure should be explained.</p>			

1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): YES

2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):

The intent of the above statement as it applies to analytical services is met in HASQAP Section 8.2, "Procedure Requirements." (See Statement Number Ch.XVI.C.1 above.)

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch.XVI.C.2.e.	Mandatory (Y/N):	N
Statement: OPERATIONS PROCEDURES. GUIDELINES. Procedure Content. To provide uniformity in operations procedures, the content of procedures should conform to prescribed guidelines. The procedure aspects described below should be followed when developing operations procedures: Procedures should be easily understood, and actions should be clearly stated.			

1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): YES

2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):

The intent of the above statement as it applies to analytical services is met in HASQAP Section 8.2, "Procedure Requirements." (See Statement Number Ch.XVI.C.1 above.)

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch.XVI.C.2.f.	Mandatory (Y/N):	N
<p>Statement:</p> <p>OPERATIONS PROCEDURES. GUIDELINES. Procedure Content. To provide uniformity in operations procedures, the content of procedures should conform to prescribed guidelines. The procedure aspects described below should be followed when developing operations procedures: Procedures should contain only one action per step.</p>			

1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): YES

2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):

The intent of the above statement as it applies to analytical services is met in HASQAP Section 8.2, "Procedure Requirements." (See Statement Number Ch.XVI.C.1 above.)

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch.XVI.C.2.g.	Mandatory (Y/N):	N
Statement: OPERATIONS PROCEDURES. GUIDELINES. Procedure Content. To provide uniformity in operations procedures, the content of procedures should conform to prescribed guidelines. The procedure aspects described below should be followed when developing operations procedures: Procedures should contain sufficient but not excessive detail. The skill level, experience, and training of the users should be considered.			

1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): YES

2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):

The intent of the above statement as it applies to analytical services is met in HASQAP Section 8.2, "Procedure Requirements." (See Statement Number Ch.XVI.C.1 above.)

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch.XVI.C.2.h.	Mandatory (Y/N):	N
<p>Statement:</p> <p>OPERATIONS PROCEDURES. GUIDELINES. Procedure Content. To provide uniformity in operations procedures, the content of procedures should conform to prescribed guidelines. The procedure aspects described below should be followed when developing operations procedures: Warnings, notes, and cautions should be easily identifiable and should not contain action statements.</p>			

1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): YES

2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):

The intent of the above statement as it applies to analytical services is met in HASQAP Section 8.2, "Procedure Requirements." (See Statement Number Ch.XVI.C.1 above.)

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch.XVI.C.2.1.	Mandatory (Y/N):	N
Statement: OPERATIONS PROCEDURES. GUIDELINES. Procedure Content. To provide uniformity in operations procedures, the content of procedures should conform to prescribed guidelines. The procedure aspects described below should be followed when developing operations procedures. Warnings and cautions should precede the step to which they apply. Warnings, notes, and cautions should appear on the same page as the step to which they apply.			

1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): YES

2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):

The intent of the above statement as it applies to analytical services is met in HASQAP Section 8.2, "Procedure Requirements." (See Statement Number Ch.XVI.C.1 above.)

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch.XVI.C.2.j.	Mandatory (Y/N):	N
<p>Statement:</p> <p>OPERATIONS PROCEDURES. GUIDELINES. Procedure Content. To provide uniformity in operations procedures, the content of procedures should conform to prescribed guidelines. The procedure aspects described below should be followed when developing operations procedures: Procedures should be technically and administratively accurate (i.e., the instructions and information should be correct; referenced documents should be correctly identified; and necessary instructions should be present to guide the user when transferring between procedures).</p>			

1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): YES

2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):

The intent of the above statement as it applies to analytical services is met in HASQAP Section 8.2, "Procedure Requirements." (See Statement Number Ch.XVI.C.1 above.)

ASSESSMENT DATA COLLECTION FORM

Statement Number: Ch.XVI.C.2.k.	Mandatory (Y/N): N
<p>Statement: OPERATIONS PROCEDURES. GUIDELINES. Procedure Content. To provide uniformity in operations procedures, the content of procedures should conform to prescribed guidelines. The procedure aspects described below should be followed when developing operations procedures: Individual sign-offs should be provided for selected critical steps. One sign-off should not be applied to more than one action.</p>	

1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): YES

2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):

The intent of the above statement as it applies to analytical services is met in HASQAP Section 8.2, "Procedure Requirements." (See Statement Number Ch.XVI.C.1 above.)

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch.XVI.C.2.1.	Mandatory (Y/N):	N
<p>Statement:</p> <p>OPERATIONS PROCEDURES. GUIDELINES. Procedure Content. To provide uniformity in operations procedures, the content of procedures should conform to prescribed guidelines. The procedure aspects described below should be followed when developing operations procedures: Limits and/or tolerances for operating parameters should be specified and should be consistent with the readable accuracy of instrumentation. Operators should not be required to perform mental arithmetic to determine if a specified parameter is acceptable.</p>			

1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): YES

2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):

The intent of the above statement as it applies to analytical services is met in HASQAP Section 8.2, "Procedure Requirements." (See Statement Number Ch.XVI.C.1 above.)

ASSESSMENT DATA COLLECTION FORM

Statement Number: Ch.XVI.C.2.m.	
Mandatory (Y/N): N	
Statement: OPERATIONS PROCEDURES. GUIDELINES. Procedure content. To provide uniformity in operations procedures, the content of procedures should conform to prescribed guidelines. The procedure aspects described below should be followed when developing operations procedures: Acceptance criteria for surveillance or test procedures should be easily discerned, including tolerances and units. If calculations are needed to compare data to acceptance criteria, the calculations should be clearly explained.	

1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): YES

2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):

The intent of the above statement as it applies to analytical services is met in HASQAP Section 8.2, "Procedure Requirements." (See Statement Number Ch.XVI.C.1 above.)

ASSESSMENT DATA COLLECTION FORM

Statement Number: Ch.XVI.C.2.n.	Mandatory (Y/N): N
Statement: OPERATIONS PROCEDURES. GUIDELINES. Procedure Content. To provide uniformity in operations procedures, the content of procedures should conform to prescribed guidelines. The procedure aspects described below should be followed when developing operations procedures: Sequence of procedural steps should conform to the normal or expected operational sequence.	

1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): YES

2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):

The intent of the above statement as it applies to analytical services is met in HASQAP Section 8.2, "Procedure Requirements." (See Statement Number Ch.XVI.C.1 above.)

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch.XVI.C.2.o.	Mandatory (Y/N):	N
Statement: OPERATIONS PROCEDURES. GUIDELINES. Procedure Content. To provide uniformity in operations procedures, the content of procedures should conform to prescribed guidelines. The procedure aspects described below should be followed when developing operations procedures: Procedures should be developed with consideration for the human-factor aspects of their intended use. For example, references to components should exactly match drawing and label-plate identifiers, units should be the same as those marked on applicable instrumentation, and charts and graphs should be easily read and interpreted. Important factors (such as operating limits, warnings, cautions, etc.) should be highlighted.			

1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): YES

2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):

The intent of the above statement as it applies to analytical services is met in HASQAP Section 8.2, "Procedure Requirements." (See Statement Number Ch.XVI.C.1 above.)

ASSESSMENT DATA COLLECTION FORM

Statement Number: Ch.XVI.C.2.p.	Mandatory (Y/N): N
<p>Statement:</p> <p>OPERATIONS PROCEDURES. GUIDELINES. Procedure Content. To provide uniformity in operations procedures, the content of procedures should conform to prescribed guidelines. The procedure aspects described below should be followed when developing operations procedures: Emergency operating procedures should provide guidance in responding to single and multiple casualties.</p>	

1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): YES

2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):

The intent of the above statement as it applies to analytical services is met in HASQAP Section 8.2, "Procedure Requirements." (See Statement Number Ch.XVI.C.1 above.)

ASSESSMENT DATA COLLECTION FORM

Statement Number: Ch.XVI.C.2.g.	
Statement: OPERATIONS PROCEDURES, GUIDELINES. Procedure Content. To provide uniformity in operations procedures, the content of procedures should conform to prescribed guidelines. The procedure aspects described below should be followed when developing operations procedures: Portions or steps of other procedures that are used or referred to when performing a procedure should be specifically identified within the procedure so that operators will not be confused when transferring between procedures.	
Mandatory (Y/N):	N

1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): YES

2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):

The intent of the above statement as it applies to analytical services is met in HASQAP Section 8.2, "Procedure Requirements." (See Statement Number Ch.XVI.C.1 above.)

ASSESSMENT DATA COLLECTION FORM

Statement Number: Ch.XVI.C.2.r.	Mandatory (Y/N): N
<p>Statement:</p> <p>OPERATIONS PROCEDURES. GUIDELINES. Procedure Content. To provide uniformity in operations procedures, the content of procedures should conform to prescribed guidelines. The procedure aspects described below should be followed when developing operations procedures: Component or system shutdown and restoration requirements following shutdown or a surveillance or test activity should be specific and controlled by the procedure.</p>	

1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): YES

2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):

The intent of the above statement as it applies to analytical services is met in HASQAP Section 8.2, "Procedure Requirements." (See Statement Number Ch.XVI.C.1 above.)

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch. XVI.C.3.	Mandatory (Y/N):	N
<p>Statement:</p> <p>OPERATIONS PROCEDURES. GUIDELINES. Procedure Changes and Revisions. ...The review and approval process for each procedure change or revision should be documented. ...Procedure changes and revisions should conform to the following practices:</p> <p>a. Procedure changes intended for use more than one time should be documented in a location readily available for operator reference. To avoid the possibility of error, these changes should also be referenced in procedure copies used by operators.</p> <p>b. Appropriate procedure changes and revisions should be initiated when procedure inadequacies or errors are noted.</p> <p>c. Procedure revisions should be initiated when a change has been outstanding for an extended period (e.g., greater than 6 months) or when a procedure has been affected by several changes (e.g., more than five). All currently effective procedure changes should normally be incorporated when the procedure is revised.</p> <p>d. Procedure revisions should be implemented concurrently with modifications. Procedure updates required by temporary modifications should be handled as procedure "change" and implemented concurrently with the temporary modification installation.</p> <p>e. Important information regarding changed or revised procedures should be communicated to appropriate operations personnel via the required reading system (Chapter XIV), a pre-shift briefing, or a similar method.</p> <p>f. Documentation of the reason for key procedure steps should be maintained and reviewed when implementing changes or revisions that alter these steps.</p> <p>g. The review process should involve verification and validation of the procedure using walkthroughs or similar methods.</p>			

1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): YES

2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):

The intent of the above statement as it applies to analytical services is met in HASQAP Section 8.4, "Categories of Changes," and Section 8.5, "Change Control."

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch.XVI.C.4.	Mandatory (Y/N):	N
<p>Statement:</p> <p>OPERATIONS PROCEDURES. GUIDELINES. Procedure Approval. Operating procedures should be approved by the operations supervisor. In addition, procedures that affect safety-related equipment and emergency procedures should be reviewed by the facility safety review committee or by another appropriate review mechanism. Procedure revisions should receive the same depth of review and level of approval as the initial versions. New and revised procedures should be approved prior to use.</p> <p>Changes in Operations procedure that do not affect the intent of operations procedure should be approved by two individuals; one should be a qualified operator, and the other should be a member of facility management. For this purpose, management could be interpreted to mean the operations supervisor or a more senior individual within the operating organization. Within 2 weeks, these procedure changes should be concurred with by the individuals who would normally approve a revision or the initial version of the procedure. Changes that alter the intent of a procedure should receive the same approval as a new or revised procedure.</p>			

1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): YES

2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):

The intent of the above statement is met by HASQAP subsection 5.2, "Administrative Systems," and HASQAP subsection 8.5, "Change Control."

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch.XVI.C.5.	Mandatory (Y/N):	N
<p>Statement:</p> <p>OPERATIONS PROCEDURES. GUIDELINES. Procedure Review. New and revised operations procedures should be reviewed prior to issuance and at periodic intervals to ensure that the information and instructions are technically accurate and that appropriate human-factor considerations have been included. The frequency of subsequent reviews should be specified; it may vary with the type and complexity of the activity involved and with time as a given plant reaches operational maturity. Applicable procedures should be reviewed after an unusual incident... During reviews, procedures should be compared to source documents to verify their accuracy. In addition, new procedures should be validated by walk-throughs in the facility or by operation on a facility-specific simulator to ensure workability.</p>			

1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): NO

2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):

The intent of the above statement as it applies to review of procedures for analytical services is not addressed in the HASQAP.

3.0 CORRECTIVE ACTION:

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch.XVI.C.6.[1]	Mandatory (Y/N):	N
Statement: OPERATIONS PROCEDURES. GUIDELINES. Procedure Availability. A controlled copy of all operations procedures should be maintained in the control area for operator reference, and selected controlled procedures should be maintained at other appropriate locations. For example, controlled procedures for facility shutdown from outside the control area should be maintained at the remote shutdown location(s).			

1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): YES

2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):

The intent of the above statement as it applies to availability of procedures for analytical services is addressed in HASQAP Section 8.1.

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch.XVI.C.6.[2]	Mandatory (Y/N):	N
Statement: OPERATIONS PROCEDURES. GUIDELINES. Procedure Availability. Working copies of controlled procedures should be available for use during evolutions. However, since these documents have only a limited lifespan, working copies should be controlled and a system should be in place to ensure that outdated procedures are not used by mistake and that working copies are replaced according to approved procedures.			

1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): YES

2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):

The intent of the above statement as it applies to availability of procedures for analytical services is addressed in HASQAP Section 8.1.

ASSESSMENT DATA COLLECTION FORM

Statement Number: Ch.XVI.C.6.[3]	Mandatory (Y/N): N
<p>Statement:</p> <p>OPERATIONS PROCEDURES. GUIDELINES. Procedure Availability. Controlled annunciator response procedure information should be easily accessible to the operators responsible for responding to alarms. Some facilities can provide annunciator response procedures at local control panels. If this is not done, annunciator response procedures should be provided at an alternate location convenient to the equipment operator.</p>	

N/A

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch.XVI.C.7.[1]	Mandatory (Y/N):	N
<p>Statement:</p> <p>OPERATIONS PROCEDURES. GUIDELINES. Procedure Use. Facility operation should be conducted in accordance with applicable procedures that reflect the facility design basis. The requirements for use of procedures should be clearly defined and understood by all operators. If procedures are deficient, a procedure change should be initiated. In exception to this policy, operators may take whatever action is necessary during emergency conditions to place the facility in a safe condition, and to protect equipment, personnel, and public safety without first initiating a procedure change.</p>			

1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): YES

2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):

The intent of the above statement as it applies to use of procedures for analytical services is addressed in HASQAP Section 8.0.

ASSESSMENT DATA COLLECTION FORM

Statement Number: Ch.XVI.C.7.[2]	Mandatory (Y/N): N
<p>Statement:</p> <p>OPERATIONS PROCEDURES. GUIDELINES. Procedure Use. Operators should have procedures with them and follow them in a step-by-step manner when the procedures contain sign-offs for the various activities. In addition, procedures should be referenced during infrequent or unusual evolutions when the operator is not intimately familiar with the procedure requirements or when errors could cause significant adverse impact to the facility. Operators need not reference emergency procedures during the performance of immediate actions since these actions, are committed to memory; however, the emergency procedure immediate action instructions should be reviewed after the actions are performed, thus, verifying, that all required actions have been taken.</p>	

1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): YES

2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):

The intent of the above statement as it applies to use of procedures for analytical services is addressed in HASQAP Section 8.1.

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch.XVII.C.1.	Mandatory (Y/N):	N
<p>Statement:</p> <p>OPERATOR AID POSTINGS. GUIDELINES. Operator Aid Development. Any facility employee could develop an operator aid; however, it should be approved prior to posting. ...all facility personnel should be informed of the importance of controlling posted information and the procedure to be followed when posting information.</p>			

N/A

Statement Number:	Ch.XVII.C.2.	Mandatory (Y/N):	N
<p>Statement:</p> <p>OPERATOR AID POSTINGS. GUIDELINES. Approval. The operations supervisor, or a higher authority, should approve all operator aids. The person approving an operator aid should ensure that the aid is necessary and correct. Operator aids that alter procedures should not be approved. Instead, appropriate procedures should be changed to incorporate the necessary information.</p>			

N/A

Statement Number:	Ch.XVII.C.3.	Mandatory (Y/N):	N
<p>Statement:</p> <p>OPERATOR AID POSTINGS. GUIDELINES. Posting. Operator aids should not be posted in such a way as to obscure instruments or controls. The aids should be located in close proximity to the area of their expected use. The operator aids should, when possible, be protected (laminated) and should be securely fastened to the equipment to which they refer.</p>			

N/A

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch.XVII.C.4.	Mandatory (Y/N):	N
<p>Statement:</p> <p>OPERATOR AID POSTINGS. GUIDELINES. Use of Operator Aids. Operator aids should be viewed as a convenience to the individual using them, not a requirement. ...Operator aids may supplement approved procedures, but they should not be used in lieu of approved procedures.</p>			

N/A

Statement Number:	Ch.XVII.C.5.	Mandatory (Y/N):	N
<p>Statement:</p> <p>OPERATOR AID POSTINGS. GUIDELINES. Documentation. A listing of all approved operator aids should be maintained along with a copy of each aid posted in the facility. This list should be used during periodic reviews of operator aids in order to help ensure that the posted aids remain current, and to provide a reference copy should an operator aid be missing during the review. The list should be kept in the control area, the operations supervisor's office, or other appropriate location. References from which operator aids were derived, control numbers (including revision), and dates of approval should be provided.</p>			

N/A

Statement Number:	Ch.XVII.C.6.	Mandatory (Y/N):	N
<p>Statement:</p> <p>OPERATOR AID POSTINGS. GUIDELINES. Review. The posted operator aids should be reviewed periodically to ensure they are still correct and necessary. In addition, an operator should audit the control area listing to ensure that only currently posted aids are recorded on the listing. Operator aids no longer posted should be removed from the listing, and missing aids should be replaced. Some operator aids are graphs, curves, or other portions of approved procedures. To help ensure that these types of operator aids are the most current versions, they should be updated when the "parent" procedure is revised... During routine facility inspections, operations personnel should review operator aids to ensure that they are approved. Unapproved postings should be submitted for approval or removed.</p>			

N/A

ASSESSMENT DATA COLLECTION FORM

The following statements concerning "EQUIPMENT AND PIPING LABELING" are N/A:

Statement Number:	Ch. XVIII.C.1.	Mandatory (Y/N):	N
<p>Statement:</p> <p>EQUIPMENT AND PIPING LABELING GUIDELINES. Components Requiring Labeling. The following components should be labeled:</p> <ul style="list-style-type: none">a. Valves;b. Major equipment (e.g., tanks, pumps, and compressors);c. Switches;d. Circuit breakers (4.16KV, 480V, 120VAC/DC, etc.);e. Fuse blocks or fuse locations;f. Instruments and gauges;g. Busses and motor control centers;h. Cabinets (including internal components such as relays, terminals, etc.);i. Room doors;j. Emergency equipment (such as fire alarm stations, sound powered phone headsets, etc.); andk. Fire protection systems.			

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch.XVIII.C.2.	Mandatory (Y/N):	N
<p>Statement:</p> <p>EQUIPMENT AND PIPING LABELING. GUIDELINES. Label Information. ...should be consistent with the information found in facility procedures, valve lineup sheets, and piping and instrument diagrams. Abbreviations and nomenclature used should be standardized and should be understood by facility personnel. Labels should be permanent, securely attached, and have distinguishable, easy-to-read information. Each component label should list a unique component number. The accompanying component noun name or description and power supply, if applicable, should also be provided. Labels may be color coded for unit, system, and/or train designation. If color coding is used, the colors should be applied consistently and have only one meaning per color or combination.</p> <p>Piping should be labeled to indicate the fluid contained and the normal flow direction. OSHA color coding for pipes could be used, and pipes containing potentially radioactive fluids, toxic fluids, or explosive gases should be uniquely marked.</p> <p>Labels should be made from materials that are compatible with their particular application. For example, chloride-free labels should be used on stainless steel piping, and temperature-tolerant labels should be used on hot components. Adhesives used for label attachment should also be verified for compatibility.</p>			

Statement Number:	Ch.XVIII.C.3.	Mandatory (Y/N):	N
<p>Statement:</p> <p>EQUIPMENT AND PIPING LABELING. GUIDELINES. Label Placement. Labels should be placed on or as near as practicable to the equipment to be labeled. The label should be oriented in a manner that is easy to read and so that the correct component is easy to identify. For example, labels for switches, indications, and breakers on control or power panels should be placed closer to the identified component than any other component so that the label clearly identifies the correct component. Label placement should not interfere with equipment operation or obscure indicators.</p>			

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch.XVIII.C.4.a.	Mandatory (Y/N):	N
<p>Statement:</p> <p>EQUIPMENT AND PIPING LABELING. GUIDELINES. Replacing Labels. Identifying Lost or Damaged Labels. Procedures should be established to ensure that misplaced or damaged labels are replaced. ...Operators should also look for missing or damaged labels during their normal tours and during other routine activities. They should become especially observant for situations where a black marking pen has been used for labeling (and then sometimes crossed out and re-labeled) and should replace these informal labels with the proper labels.</p>			

Statement Number:	Ch.XVIII.C.4.b.	Mandatory (Y/N):	N
<p>Statement:</p> <p>EQUIPMENT AND PIPING LABELING. GUIDELINES. Replacing Labels. Providing New Labels. A method should exist to help ensure the prompt identification and correction of labeling deficiencies. ...a temporary label should be attached to the component until the replacement label can be made. The replacement label should be approved by the operations supervisor (or higher authority)... The attachment of temporary and/or replacement labels to correct components should be verified. The new label should meet the guidelines for label information as addressed in paragraph C2 of this chapter.</p>			

The above statements concerning "EQUIPMENT AND PIPING LABELING" are N/A:

ATTACHMENT D

ASSESSMENT DATA COLLECTION FORMS

for

HASQAP ASSESSMENT

to

DOE ORDER 5480.20
(Training)

ASSESSMENT DATA COLLECTION FORM

Statement Number:	10.a.	Mandatory (Y/N):	Y
<p>Statement:</p> <p>IMPLEMENTATION SCHEDULE. The requirements of this Order shall be implemented in accordance with the following: A Training Implementation Matrix shall be submitted to the cognizant field organization by November 8, 1991. The Training Implementation Matrix shall contain a schedule for implementing the requirements of this Order. The matrix should be based on the status of existing compliance and shall include the time-frame for incremental implementation and full implementation of the requirements;</p>			

N/A

Statement Number:	10.b.	Mandatory (Y/N):	Y
<p>Statement:</p> <p>IMPLEMENTATION SCHEDULE. The requirements of this Order shall be implemented in accordance with the following: From the issue date of this Order until the Training Implementation Matrix is approved, all DOE reactor and non-reactor nuclear facilities shall continue to comply with the requirements contained in DOE 5480.5 and DOE 5480.6, as appropriate to the facility;</p>			

N/A

Statement Number:	10.c.	Mandatory (Y/N):	Y
<p>Statement:</p> <p>IMPLEMENTATION SCHEDULE. The requirements of this Order shall be implemented in accordance with the following: Upon approval of the Training Implementation Matrix, personnel subsequently appointed to functional positions shall meet the education and experience requirements of this Order;</p>			

N/A

Statement Number:	10.d.	Mandatory (Y/N):	Y
<p>Statement:</p> <p>IMPLEMENTATION SCHEDULE. The requirements of this Order shall be implemented in accordance with the following: For facilities for which DOE 5480.18 is applicable, implementation of the training program content requirements of this Order may, for positions requiring accreditation, coincide with the implementation of DOE 5480.18.</p>			

N/A

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch.I. 2.	Mandatory (Y/N):	Y
<p>Statement:</p> <p>GENERAL REQUIREMENTS. TRAINING ORGANIZATION REQUIREMENTS. The operating contractor shall establish an organization(s) within the line management organization that is responsible for the training of operating organization personnel. In some cases (e.g., Category B reactors, low-hazard non-reactor nuclear facilities, or less complex, small facilities) this may be integrated into the operating organization and may not necessarily be officially designated as a training organization. The duties, responsibilities, qualifications, and authority of training organization personnel shall be documented, and managerial responsibilities and authority clearly defined. This organization may include subcontracted personnel who conduct training activities.</p>			

N/A

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch. I. 3.	Mandatory (Y/N):	Y
-------------------	-----------	------------------	---

Statement:

GENERAL REQUIREMENTS. SUBCONTRACTOR PERSONNEL QUALIFICATION REQUIREMENTS. The operating organization shall establish qualification criteria for subcontractor personnel who replace personnel in the operating organization. Subcontractor personnel shall meet the qualification requirements for the job function to be performed prior to active involvement in facility activities. For subcontractor personnel who do not meet the requirements, work activities on safety-related structures, systems, and components identified in the facility Safety Analysis Report shall be supervised by a person who meets the qualification criteria established by the operating organization for conduct of such activities. In addition, the operating organization shall ensure that subcontractor and temporary personnel who perform specialized activities such as radiation protection, maintenance, in-service inspection, radiography, and welding are qualified to perform their assigned tasks. This assurance shall be considered adequate with proper documentation of at least one of the following:

- a. The satisfactory result of an audit of subcontractor records which relate to qualification of the subcontractor personnel being considered for assignment by the operating organization; or
- b. Operating organization's previous verification (within 2 years) of the ability of the subcontractor employee to perform assigned tasks safely and efficiently; or
- c. Successful completion by the subcontractor employee of those segments of the operating organization's qualification program which are considered pertinent to accomplishment of the task to be performed.

1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): YES

2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):

The intent of this statement is met by HASQAP Section 3.0, "Personnel Qualification and Training," and subsection 18.2, "Control of Subcontractors."

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch.I. 4.a.	Mandatory (Y/N):	Y
Statement: GENERAL REQUIREMENTS. PERSONNEL SELECTION REQUIREMENTS. The operating contractor shall have a process for selection and assignment of personnel into the operating organization. This process should consider factors such as background, experience, and education and may involve a selection test. Selection should be based on the ability to meet job performance requirements.			

1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): YES

2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):

This statement is met by HASQAP subsection 3.2, "Personnel Selection."

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch.I. 4.b.	Mandatory (Y/N):	Y
Statement: GENERAL REQUIREMENTS. PERSONNEL SELECTION REQUIREMENTS. ...In those cases where an individual does not meet the literal experience requirements, consideration may be given to the collective experience of the operating organization. Individuals who do not meet the experience requirements for a position may be assigned to that position providing the overall operating organization is considered balanced and strong and that DOE approval is obtained on a case-by-case basis.			

- 1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): NO
- 2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):
Statement not addressed by the HASQAP.
- 3.0 CORRECTIVE ACTION:

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch.I. 5.	Mandatory (Y/N):	Y
Statement: GENERAL REQUIREMENTS. QUALIFICATION PROCESS REQUIREMENTS. ...Personnel at DOE reactor and non-reactor nuclear facilities shall possess qualifications which provide reasonable assurance that their decisions and actions will ensure that assigned responsibilities are conducted properly and safely.			

1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): YES

2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):

This statement is met by HASQAP subsection 3.1.1, "Initial Qualification," and subsection 3.1.2, "Maintaining Qualification."

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch. I. 5. a.	Mandatory (Y/N):	Y
<p>Statement:</p> <p>GENERAL REQUIREMENTS. QUALIFICATION PROCESS REQUIREMENTS. Operating organizations shall establish written procedures which clearly define qualification requirements for personnel in each functional level based on the criteria contained in this Order. The shift in relative importance of managerial and technical competence shall be considered by management in establishing these requirements. The need for specific knowledge, skills, and abilities differ for each level in the organization. At the higher functional level, managerial competence is the dominant need, whereas technical competence is the dominant need at other functional levels.</p>			

1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): YES

2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):

This statement is met by HASQAP subsection 3.1, "Qualification," which states: "Qualification requirements shall be established and re-evaluated by laboratory management on a scheduled basis for specific job categories, including managers, analysts, technicians, and data reviewers."

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch.I. 5.b.	Mandatory (Y/N):	Y
Statement: GENERAL REQUIREMENTS. QUALIFICATION PROCESS REQUIREMENTS. Qualification may be granted only after assuring that all requirements (including training and examinations as required) and other specified requirements...have been satisfactorily completed.			

1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): YES

2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):

This statement is met by HASQAP subsection 3.1, "Qualification," and subsection 3.1.1, "Initial Qualification."

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch.I. 5.c.	Mandatory (Y/N):	Y
Statement: GENERAL REQUIREMENTS. QUALIFICATION PROCESS REQUIREMENTS. Qualification shall be valid for a maximum of two years (unless revoked for cause) at which time the person shall be requalified in accordance with paragraph 10 of this Chapter.			

1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): NO

2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):

The HASQAP does not address specific time limits for maintaining qualifications.

3.0 CORRECTIVE ACTION:

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch. I. 6.a.	Mandatory (Y/N):	Y
Statement: GENERAL REQUIREMENTS, CERTIFICATION PROCESS REQUIREMENTS, The program leading to certification shall be governed by written procedures which include requirements for documented assessment of the person's qualifications through examinations and operational evaluations.			

1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): YES

2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):

The intent of this statement is met by HASQAP Section 3.0, "Personnel Qualification and Training."

3.0 COMMENTS:

The term "certification" does not appear in Section 3.0 until the last bullet in subsection 3.7 where it is used to define "formal qualification."

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch.I. 6.b.	Mandatory (Y/N):	Y
<p>Statement:</p> <p>GENERAL REQUIREMENTS. CERTIFICATION PROCESS REQUIREMENTS.</p> <p>Certification may be granted only after assuring that all qualification requirements (including written and oral examinations and operational evaluations) and other specified requirements...have been satisfactorily completed, and management has assured that the person is capable of safely performing all functions of the position.</p> <p>Satisfactory completion of qualifications which result in certification shall be verified by a person or group other than the person or group that provided the training or the candidate's immediate supervisor.</p> <p>Certification shall be valid for a maximum of two years (unless revoked for cause) at which time the person shall be recertified in accordance with paragraph 10 of this Chapter.</p>			

1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): NO

2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):

The following portion of the statement is not addressed in the HASQAP:

Satisfactory completion of qualifications which result in certification shall be verified by a person or group other than the person or group that provided the training or the candidate's immediate supervisor. Certification shall be valid for a maximum of two years (unless revoked for cause) at which time the person shall be recertified in accordance with paragraph 10 of this Chapter.

3.0 CORRECTIVE ACTION:

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch. I. 6.C.	Mandatory (Y/N):	Y
Statement: GENERAL REQUIREMENTS. CERTIFICATION PROCESS REQUIREMENTS. Reactor operators, senior reactor operators, and shift supervisors at DOE Category A and B reactors and fissionable material handlers and fissionable material handler supervisors at DOE non-reactor nuclear facilities shall be certified. For all other operators and their immediate supervisors, the operating organization shall identify in the Training Implementation Matrix any additional positions that will be subject to certification...			

N/A

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch.I. 7.a.	Mandatory (Y/N):	Y
Statement: GENERAL REQUIREMENTS. TRAINING REQUIREMENTS. General. Training for operations and maintenance personnel should be based on analyzed needs such as would result from a needs analysis or job analysis. Training for technical support personnel should be based on the results of an assessment of position responsibilities. Training programs shall consist of a combination of classroom-type and on-the-job training, and include simulator and laboratory training as it applies to the position. Classroom-type training includes lectures, seminars, computer-based, and structured self-study training activities.			

1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): YES

2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):

This statement is met by HASQAP Section 3.0, "Personnel Qualification and Training."

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch.I. 7.a.(1)	Mandatory (Y/N):	Y
<p>Statement:</p> <p>GENERAL REQUIREMENTS. TRAINING REQUIREMENTS. General. Qualification programs shall be reviewed by contractor facility management and kept up to date to reflect changes to the facility, procedures, regulations, and quality assurance requirements as well as applicable industry operating experience. The concept of training personnel as a team, stressing team communications and interaction, shall be used where job functions require team solutions and activities.</p>			

1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): YES

2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):

This statement is met by HASQAP subsection 3.1, "Qualification,"

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch. I. 7.a.(2)	Mandatory (Y/N):	Y
Statement: GENERAL REQUIREMENTS. TRAINING REQUIREMENTS. General. A Training Implementation Matrix which defines and describes the application of the selection, qualification, and training requirements of this Order shall be prepared by the operating organization. The Matrix shall clearly define the organization, planning, and administration of the qualification program and set forth the responsibility, authority, and methods for conducting training. Suitable justification for exceptions shall be included in the Matrix for any requirement not implemented. This Matrix shall be submitted to the Head of the Field Organization for approval. At some sites with several facilities, a combined Training Implementation Matrix may be submitted.			

- 1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): NO
- 2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):
- This statement is not addressed by the HASQAP.
- 3.0 CORRECTIVE ACTION:

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch.I. 7.b.	Mandatory (Y/N):	Y
<p>Statement:</p> <p>GENERAL REQUIREMENTS. TRAINING REQUIREMENTS. Training Process. Initial and continuing training programs shall be implemented to ensure that operating organization personnel are qualified to perform job requirements. This should be achieved by using a systematic approach to training such as performance-based training. The basic elements of a performance-based training program include the following:</p> <p>(1) Establishment of prerequisite standards of education, skills, and knowledge required for entry into the training program;</p> <p>(2) A systematic analysis of the job to be performed. A job or task analysis should be conducted by the operating organization to identify training which provides the necessary skills and knowledge so that assigned tasks can be effectively performed. Initial and continuing training programs should be based on, and traceable to, the analysis. The analysis should include normal and emergency duties, and place emphasis on the role played by each member of the operating organization (and maintenance and technical support organizations as required) in assuring safe operation. Because of varied complexity and scope of job functions, the degree of analysis necessary to determine skill and knowledge requirements may vary...;</p> <p>(3) Design and development of training programs based on job performance requirements and standards;</p> <p>(4) Implementation of training programs which contain instruction appropriate to job performance; and</p> <p>(5) Evaluation of trainees' ability to meet job performance requirements, and evaluation and revision of training programs.</p>			

1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): YES

2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):

This statement is met by HASQAP Section 3.0, "Personnel Qualification and Training."

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch.I. 7.c.	Mandatory (Y/N):	Y
Statement: GENERAL REQUIREMENTS. TRAINING REQUIREMENTS. Initial Training Requirements. An initial training program shall be established for operating organization personnel to develop or enhance their knowledge, skills, and ability to perform job assignments. Personnel in training shall not independently make decisions or take actions that could affect facility safety, nor shall personnel in training be placed in such positions. However, they may independently perform specific tasks or job assignments for which they are qualified.			

1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): YES

2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):

This statement is met by HASQAP subsection 3.6.1, "Initial Training Requirements."

ASSESSMENT DATA COLLECTION FORM

Statement Number: Ch.I. 7.d.	Mandatory (Y/N): Y
<p>Statement:</p> <p>GENERAL REQUIREMENTS. TRAINING REQUIREMENTS. Continuing Training Requirements. Continuing training programs shall be designed and implemented to maintain and enhance the proficiency of operating organization personnel who perform functions associated with safety-related structures, systems, and components identified in the facility Safety Analysis Report.</p>	

1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): YES

2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):

This statement is met by HASQAP subsection 3.6.2, "Continuing Training Requirements."

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch.I. 7.d.(1)	Mandatory (Y/N):	Y
Statement: GENERAL REQUIREMENTS. TRAINING REQUIREMENTS. Continuing Training Requirements. These programs shall be structured commensurate with specific position needs, and shall be administered on a biennial cycle.			

1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): NO

2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):

The portion of the statement that states "shall be administered on a biennial cycle" is not addressed in HASQAP subsection 3.6.2.

3.0 COMMENTS:

Administering a continuing training program per the statement implies a different meaning than appraising performance and evaluating personnel as stated in subsection 3.6.2.

4.0 CORRECTIVE ACTION:

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch.I. 7.d.(2)	Mandatory (Y/N):	Y
Statement: GENERAL REQUIREMENTS. TRAINING REQUIREMENTS. Continuing Training Requirements. Periodic written and oral examinations and/or operational evaluations shall be administered and documented throughout the cycle on material included in the training programs.			

- 1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): YES
- 2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):
- The intent of this statement is met by HASQAP subsection 3.6.2, "Continuing Training Requirements."
- 3.0 CORRECTIVE ACTION:

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch.I. 7.d.(3)	Mandatory (Y/N):	Y
Statement: GENERAL REQUIREMENTS TRAINING REQUIREMENTS. Continuing Training Requirements. Training and examination covering abnormal facility procedures and emergencies shall be required at least annually for certified operations personnel.			

1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): NO

2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):

Abnormal facility procedures and emergencies as a part of the continuing training program is not specifically addressed in the HASQAP.

3.0 CORRECTIVE ACTION:

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch.I. 7.d.(4)	Mandatory (Y/N):	Y
<p>Statement:</p> <p>GENERAL REQUIREMENTS. TRAINING REQUIREMENTS. Continuing Training Requirements. Continuing training programs for certified operations personnel shall consist of preplanned classroom-type training, on-the-job training, and operational evaluations on a regular and continuing basis, and shall include, as a minimum, the following as related to job performance:</p> <p>(a) Training in significant facility system and component changes, applicable procedure changes, applicable industry operating experience, selected fundamentals with emphasis on seldom used knowledge and skills necessary to assure safety, and other training as needed to correct identified performance problems;</p> <p>(b) Drills shall be conducted in the facility to enable personnel and operating teams to maintain proficiency in their ability to respond to abnormal or accident situations;</p> <p>(c) Instruction in the use of facility systems to control or mitigate accidents. Such training shall include both classroom-type training and training conducted in the facility;</p> <p>(d) Training, as applicable to the position, in the following subjects where examinations and experience (industry and facility-specific) or other evidence indicates emphasis in scope and depth of coverage is needed:</p> <ol style="list-style-type: none"> 1 Theory and principles of facility operation; 2 General and specific facility operating characteristics; 3 Facility instrumentation and control; 4 Facility protection systems; 5 Engineered Safety Features; 6 Normal, abnormal, and emergency procedures; 7 Radiation control and safety; and 8 Technical Specifications/Operational Safety Requirements. 			

1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): Yes

2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):

The intent of this statement is met by HASQAP Section 3.0, "Personnel Qualification and Training."

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch.I. 7.d.(5)	Mandatory (Y/N):	Y
Statement: GENERAL REQUIREMENTS. TRAINING REQUIREMENTS. Continuing Training Requirements. Training personnel are exempt from the provisions of 7d for the area of primary administrative responsibility.			

- 1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): NO
- 2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):
- Not addressed in the HASQAP.
- 3.0 CORRECTIVE ACTION:

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch.I. 7.e.(1)	Mandatory (Y/N):	Y
<p>Statement:</p> <p>GENERAL REQUIREMENTS. TRAINING REQUIREMENTS. General Employee Training (GET) Requirements. All persons employed either full- or part-time in DOE reactor and non-reactor nuclear facilities shall be trained in the following areas commensurate with their job duties:</p> <ul style="list-style-type: none"> (a) General description of facilities; (b) Job related policies, procedures, and instructions; (c) Radiological health and safety program; (d) Facility emergency plans; (e) Industrial safety/hygiene program; (f) Fire protection program; (g) Security program; and (h) Quality assurance program. 			

1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): NO

2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):

GET training is not addressed in the HASQAP.

3.0 CORRECTIVE ACTION:

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch. I. 7.e.(2)	Mandatory (Y/N):	Y
Statement: GENERAL REQUIREMENTS. TRAINING REQUIREMENTS. General Employee Training (GET) Requirements. Visitors, contracted personnel, and temporary personnel shall be under continuous escort while at the facility unless they have been trained in appropriate areas from the above list to the extent necessary to ensure safe execution of their duties.			

- 1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): NO
- 2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):
GET training is not addressed in the HASQAP.
- 3.0 CORRECTIVE ACTION:

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch.I. 7.e.(3)	Mandatory (Y/N):	Y
<p>Statement:</p> <p>GENERAL REQUIREMENTS. TRAINING REQUIREMENTS. General Employee Training (GET) Requirements. For persons requiring long-term (i.e., more than 1-2 weeks) access, understanding of the information provided by the GET program shall be evaluated by administering a written examination. The examination should cover areas selected for training and should be of sufficient difficulty to ensure the person has adequate knowledge to work independently at the facility. Persons who do not pass this examination shall not be permitted access without a continuous escort.</p>			

- 1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): NO
- 2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):
- GET training is not addressed in the HASQAP.
- 3.0 CORRECTIVE ACTION:

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch.I. 7.e.(4)	Mandatory (Y/N):	Y
Statement: GENERAL REQUIREMENTS. TRAINING REQUIREMENTS. General Employee Training (GET) Requirements. GET shall be included in continuing training programs for all facility personnel, and examinations administered at least every two years.			

- 1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): NO
- 2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):
- GET training is not addressed in the HASQAP.
- 3.0 CORRECTIVE ACTION:

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch.I. 7.f.	Mandatory (Y/N):	Y
<p>Statement:</p> <p>GENERAL REQUIREMENTS. TRAINING REQUIREMENTS. Probabilistic Risk Assessment (PRA) Training Requirements. For those facilities for which a PRA has been performed, initial and continuing training programs for operations and technical support personnel shall include training on the principal results of the PRA. This training shall address the following:</p> <p>(1) The importance of facility systems in preventing damage or severe accidents;</p> <p>(2) Locations of all significant amounts of radioactive and other hazardous materials, and measures to prevent its release; and</p> <p>(3) The importance of maintaining operational limits and conditions, and the consequences of violating those limits.</p>			

- 1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): NO
- 2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):
- PRA training is not addressed in the HASQAP.
- 3.0 CORRECTIVE ACTION:

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch.I. 7.g. (1)	Mandatory (Y/N):	Y
Statement: GENERAL REQUIREMENTS. TRAINING REQUIREMENTS. Technician and Maintenance Personnel Training Requirements. All technicians and maintenance personnel shall be qualified to perform the tasks associated with their specialty, or work under the direct supervision of personnel qualified to perform the activity or task.			

1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): YES

2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):

This statement is met by HASQAP Section 3.0, "Personnel Qualification and Training."

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch.I. 7.g.(2)	Mandatory (Y/N):	Y
<p>Statement:</p> <p>GENERAL REQUIREMENTS. TRAINING REQUIREMENTS. Technician and Maintenance Personnel Training Requirements. Training on safety-related systems identified in the facility Safety Analysis Report shall be conducted for personnel who perform work on those systems/components. Included in this category are systems having a direct impact on the safe operation of the facility. ...System training shall, at a minimum, include the following elements:</p> <p>(a) Purpose of the system;</p> <p>(b) General description of the system including major components, relationship to other systems, and all safety implications associated with working on the system; and</p> <p>(c) Related industry and facility-specific experience.</p>			

- 1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): NO
- 2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):
- Statement is not addressed in the HASQAP.
- 3.0 CORRECTIVE ACTION:

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch.I. 7.h.(1)	Mandatory (Y/N):	Y
Statement: GENERAL REQUIREMENTS. TRAINING REQUIREMENTS. Technical Support Personnel Training Requirements. Training shall be provided to entry-level personnel who provide technical support to the operating organization.			

1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): YES

2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):

This statement is met by HASQAP subsection 3.6, "Laboratory Training."

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch.I. 7.h.(2)	Mandatory (Y/N):	Y
-------------------	---------------	------------------	---

Statement:

GENERAL REQUIREMENTS. TRAINING REQUIREMENTS. Technical Support Personnel Training Requirements. The contractor shall develop a list of specific technical support personnel positions that may have direct impact on employee, facility, or public safety. Training in the following facility-specific subject areas shall be included as appropriate to the position:

- (a) Facility organization;
- (b) Facility fundamentals;
 - 1 Heat transfer, fluid flow, and thermodynamics
 - 2 Electrical science
 - 3 Nuclear physics
 - 4 Chemistry/chemistry controls
 - 5 Process controls
- (c) Facility systems, components, and operations;
- (d) Simulator training;
- (e) Environment, Safety, and Health Orders;
- (f) Codes and standards overview;
- (g) Facility document system;
- (h) Safety Analysis Reports and Technical Specifications/Operational Safety Requirements;
- (i) Nuclear criticality control;
- (j) Material, maintenance, and modification control;
- (k) ALARA and radwaste reduction program; and
- (l) QA/QC practices.

1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): YES

2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):

The intent of this statement is met by HASQAP subsection 3.6, "Laboratory Training."

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch.I. 7.i.	Mandatory (Y/N):	Y
Statement: GENERAL REQUIREMENTS. TRAINING REQUIREMENTS. Management and Supervisory Training Requirements. In addition to the training specified in paragraph 7h, managers and first line supervisors shall receive training in the following as appropriate to their job responsibilities. Supervisory skills and management training need not be subject to examination as part of initial training, nor categorically repeated in their continuing training programs.			

1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): NO

2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):

The specifics of the statement are not addressed in subsection 3.5, "Management Training," of the HASQAP.

3.0 CORRECTIVE ACTION:

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch.I. 7.i.(1)	Mandatory (Y/N):	Y
<p>Statement:</p> <p>GENERAL REQUIREMENTS, TRAINING REQUIREMENTS, Management and Supervisory Training Requirements, Supervisory Skills Training. ...shall include:</p> <ul style="list-style-type: none"> (a) Leadership; (b) Interpersonal communication; (c) Command responsibilities and limits; (d) Motivation of personnel; (e) Problem analysis and decision making; (f) Fitness for duty procedures; and (g) Administrative policies and procedures. 			

1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): NO

2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):

The specifics of the statement are not addressed in subsection 3.5, "Management Training," of the HASQAP.

3.0 CORRECTIVE ACTION:

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch.I. 7.i.(2)	Mandatory (Y/N):	N
<p>Statement:</p> <p>GENERAL REQUIREMENTS. TRAINING REQUIREMENTS. Management and Supervisory Training Requirements. Management Training. ...should include:</p> <p>(a) Supervisory skills training;</p> <p>(b) Quality assurance and quality control;</p> <p>(c) Facility security and emergency plans;</p> <p>(d) Purchasing;</p> <p>(e) Material storage;</p> <p>(f) Facility modifications;</p> <p>(g) Nuclear, industrial, and radiation safety;</p> <p>(h) Environmental issues; and</p> <p>(i) Budgeting.</p>			

1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): NO

2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):

The specifics of the statement are not addressed in subsection 3.5, "Management Training," of the HASQAP.

3.0 CORRECTIVE ACTION:

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch.I. 8.	Mandatory (Y/N):	Y
<p>Statement:</p> <p>GENERAL REQUIREMENTS. OPERATOR AND SUPERVISOR EXAMINATION REQUIREMENTS. Written and oral examinations and operational evaluations shall be prepared and administered to demonstrate that certified operators and supervisors possess the required knowledge and skills. Written examinations and operational evaluations shall be prepared and administered for the qualification of other operators and supervisors. For certified Category A reactor facility personnel, the oral examination shall be separate from the operational evaluation. Operational evaluations and oral examinations may be combined for certified Category B reactor and non-reactor nuclear facility personnel.</p>			

N/A

Statement Number:	Ch.I. 8.a.	Mandatory (Y/N):	Y
<p>Statement:</p> <p>GENERAL REQUIREMENTS. OPERATOR AND SUPERVISOR EXAMINATION REQUIREMENTS. Written procedures shall be established for written and oral examinations and operational evaluations (as required). These procedures shall address examination/evaluation development, approval, security, administration, and maintenance of examination question banks.</p>			

N/A

Statement Number:	Ch.I. 8.b.	Mandatory (Y/N):	Y
<p>Statement:</p> <p>GENERAL REQUIREMENTS. OPERATOR AND SUPERVISOR EXAMINATION REQUIREMENTS. Oral examinations may be conducted as a one-on-one walkthrough or by an oral board or committee consisting of personnel identified by contractor facility management. The oral examination content shall be tailored to evaluate the candidate's operational knowledge (initial/continuing training program subjects) and organizational awareness to determine how the individual will function in an operating environment. Oral examination questions, the candidate's response, and the evaluation by the examiner(s) of that response shall be documented.</p>			

N/A

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch.I. 9.	Mandatory (Y/N):	Y.
<p>Statement:</p> <p>GENERAL REQUIREMENTS. OPERATOR AND SUPERVISOR REEXAMINATION REQUIREMENTS. Reexaminations for certified and qualified operators and supervisors shall include subjects in which the person is expected to be proficient and emphasize those subjects covered by the continuing training program. The contractor shall administer biennial written and oral examinations and operational evaluations, or administer periodic examinations throughout the cycle that cover all continuing training program subjects/elements. Operational evaluations and oral examinations may be combined for Category B reactor and non-reactor nuclear facility personnel.</p>			

N/A

Statement Number:	Ch.I.10.	Mandatory (Y/N):	Y
<p>Statement:</p> <p>GENERAL REQUIREMENTS. REQUALIFICATION REQUIREMENTS. Employees shall not be allowed to function as qualified/certified operators and supervisors if they have not completed all of the requalification program elements within two years. If an operator or supervisor fails a requalification examination, or shows serious job performance deficiencies which indicate that he or she may perform in an unsafe manner, the person shall be removed from activities requiring qualification.</p>			

N/A

Statement Number:	Ch.I.10.a.	Mandatory (Y/N):	N
<p>Statement:</p> <p>GENERAL REQUIREMENTS. REQUALIFICATION REQUIREMENTS. Requalification may be regained after completing remedial training designed to correct the deficiency(s) and a reexamination is administered. In addition, recertification of operators and supervisors shall be based on the following:</p> <p>(1) A review of individual operating performance during the past certification period by either line management, by a committee, or by a person designated by management; and</p> <p>(2) A current medical examination as required by Chapter II paragraph 3, Chapter III paragraph 3, or Chapter IV paragraph 3.</p>			

N/A

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch.I.10.b.	Mandatory (Y/N):	Y
<p>Statement:</p> <p>GENERAL REQUIREMENTS. REQUALIFICATION REQUIREMENTS. When a certified operator or supervisor has been absent from certification duties for greater than 3 months, but less than 12 months, selected retraining (including written and oral examinations and operational evaluations) shall be given as deemed necessary prior to reassignment to certification duties. The certification base date remains the same as it was before the absence. However, if the absence is greater than 12 months, comprehensive written and oral examinations and operational evaluations (as required of initial candidates) shall be given to determine weak areas. Retraining and reexamination shall be required in areas of weakness, and upon successful completion, a new certification date may be established.</p>			

N/A

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch.I.11.	Mandatory (Y/N):	Y
Statement: GENERAL REQUIREMENTS. EXCEPTIONS TO TRAINING REQUIREMENTS. Initial training programs are developed for persons with entry-level knowledge and skills. Some candidates may already possess the necessary knowledge and skills for their job, and may be excepted from areas of the training program on the basis of prior education, experience, and training. Proficiency testing is the preferred method of excepting persons from specific areas of training. In all cases, the requisite examinations to establish qualification shall be completed.			

1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): YES

2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):

The statement is met by HASQAP subsection 3.6.1, "Initial Training Requirements," and subsection 3.1.1, "Initial Qualification."

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch.I.11.a.	Mandatory (Y/N):	Y
<p>Statement:</p> <p>GENERAL REQUIREMENTS, EXCEPTIONS TO TRAINING REQUIREMENTS: The operating organization shall establish procedures and criteria to administer and document exceptions to initial and continuing training program requirements. The name of the person and the specific subject for which the exception is requested, along with justification for the exception, shall be included as part of the documentation. In all cases, the operating organization shall ensure that sufficient facility-specific instruction is provided to enable the candidate to perform job requirements. The operating organization shall submit the procedure(s) which detail the criteria for granting an exception to the field organization for approval.</p>			

1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): NO

2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):

Although the HASQAP addresses the statement in a general sense, the specific details of the statement are not met concerning procedures, criteria, documentation, and field organization approval.

3.0 CORRECTIVE ACTION:

ASSESSMENT DATA COLLECTION FORM

Statement Number: Ch.I.11.b.	Mandatory (Y/N): Y
Statement: GENERAL REQUIREMENTS. EXCEPTIONS TO TRAINING REQUIREMENTS. Any exception from certification or qualification requirements shall be approved by contractor management.	

- 1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): NO
- 2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):
- Statement not addressed in the HASQAP.
- 3.0 CORRECTIVE ACTION:

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch.I.12.a.	Mandatory (Y/N):	Y
<p>Statement:</p> <p>GENERAL REQUIREMENTS. EXTENSION REQUIREMENTS. The operating organization shall establish an administrative procedure which addresses extensions to ensure timely completion of requirements associated with certification or qualification. This procedure should include as a minimum:</p> <p>(1) Responsibility for approval of the extension;</p> <p>(2) Length of extension;</p> <p>(3) Explanation of circumstances that prevented the person from completing the requirements; and</p> <p>(4) Description of the operational and/or schedular situation which necessitated the extension.</p>			

- 1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): NO
- 2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):
- Extensions are not addressed in the HASQAP.
- 3.0 CORRECTIVE ACTION:

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch.I.12.b.	Mandatory (Y/N):	Y
Statement: GENERAL REQUIREMENTS. EXTENSION REQUIREMENTS. Extensions of certification of operators and supervisors shall be approved by the field organization. Extensions of qualification of other personnel shall be approved by contractor facility management.			

- 1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): NO
- 2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):
Extensions are not addressed in the HASQAP.
- 3.0 CORRECTIVE ACTION:

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch. I.13.	Mandatory (Y/N):	Y
<p>Statement:</p> <p>GENERAL REQUIREMENTS. ALTERNATIVES TO EDUCATIONAL REQUIREMENTS. Educational criteria are described as either baccalaureate, associate degree, or high school diploma. In each case, the type of degree (diploma) required is a function of the person's responsibilities. Persons who do not possess the formal educational requirements specified shall not be automatically eliminated where other factors provide sufficient assurance of their abilities to fulfill the duties of a specific position. These factors shall be evaluated on a case-by-case basis and approved and documented by the operating organization. The following are examples that may be considered in making the evaluation of an acceptable alternative to the educational requirements:</p> <ul style="list-style-type: none"> a. General Education Development (GED) test for a high school diploma; b. Professional engineers license or completion of Engineer in Training (EIT) examination for a baccalaureate or associate degree requirement; c. Completion of technical portions of an engineering, engineering technology, or related science program may substitute for the baccalaureate or associate degree program. Successful completion shall be determined by a transcript or other certification by an accredited institution. ...The courses shall be in appropriate technical subjects relevant to the position to be filled; and d. Related experience may substitute for education at the rate of six semester credit hours for each year of experience up to a maximum of 60 credit hours. 			

1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): NO

2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):

Alternatives to education requirements are not addressed in the HASQAP.

3.0 CORRECTIVE ACTION:

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch.I.14.	Mandatory (Y/N):	Y
<p>Statement:</p> <p>GENERAL REQUIREMENTS. ALTERNATIVES TO EXPERIENCE REQUIREMENTS.</p> <p>Experience in design, construction, and operational training may be considered applicable nuclear experience and should be evaluated on a case-by-case basis.</p> <p>a. Where course work is related to job assignments, post-secondary education may be substituted. Formal education shall not be allowed to substitute for more than 50 percent of the experience requirement unless otherwise stated in Chapters II, III, or IV.</p> <p>b. Job-related training in the position sought may qualify as equivalent to nuclear experience on a one-for-one basis for up to a maximum of two years.</p>			

1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): NO

2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):

Alternatives to experience requirements are not addressed in the HASQAP.

3.0 CORRECTIVE ACTION:

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch.I.15.	Mandatory (Y/N):	Y
Statement: GENERAL REQUIREMENTS. LIMITATIONS FOR OVERTIME WORKED. Administrative procedures shall be developed and implemented to limit the working hours of operating organization personnel who perform safety-related functions...			

N/A

Statement Number:	Ch.I.15.a.	Mandatory (Y/N):	Y
Statement: GENERAL REQUIREMENTS. LIMITATIONS FOR OVERTIME WORKED. Adequate shift coverage shall be maintained without heavy use of overtime. However, in the event that unforeseen problems require substantial amounts of overtime to be used..., the following requirements shall be applied on a temporary basis: (1) A person shall not be permitted to work more than 16 consecutive hours, excluding shift turnover time; (2) A person shall not be permitted to work more than 24 hours in any 48 hour period, excluding shift turnover time; (3) A person shall not be permitted to work more than 72 hours in any 7-day period, excluding shift turnover time; (4) A person shall not be permitted to work more than 14 consecutive days without having two consecutive days off; and (5) Except during extended shutdown periods, brief pre- or post-shift training sessions, or weather-related occurrences, the use of overtime shall be considered on an individual basis and not for the entire shift crew.			

N/A

Statement Number:	Ch.I.15.b.	Mandatory (Y/N):	Y
Statement: GENERAL REQUIREMENTS. LIMITATIONS FOR OVERTIME WORKED. Any deviation from the overtime requirements shall be authorized on a case-by-case basis in advance by the Plant/Facility Manager or his or her designee in accordance with established procedures and with documentation of the basis for granting the deviation. Individual overtime shall be reviewed monthly by the Plant/Facility Manager or his or her designee to assure that excessive hours have not been worked.			

N/A

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch. I. 15.c.	Mandatory (Y/N):	N
Statement: GENERAL REQUIREMENTS. LIMITATIONS FOR OVERTIME WORKED. If a person is required to work in excess of 12 continuous hours, his or her duties should be carefully selected. It is preferable that this person not be assigned any task that could possibly endanger the safe operation of the facility.			

N/A

ASSESSMENT DATA COLLECTION FORM

Statement Number: Ch.I.16.	Mandatory (Y/N): Y
Statement: GENERAL REQUIREMENTS. RECORD REQUIREMENTS. Contractors shall develop and implement administrative procedures that specify requirements for the maintenance of training, qualification, and certification records for operating organization personnel.	

1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): YES

2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):

Statement is met by HASQAP subsection 3.7, "Training Records."

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch.I.16.a.	Mandatory (Y/N):	Y
<p>Statement:</p> <p>GENERAL REQUIREMENTS. RECORD REQUIREMENTS. Qualification and certification of personnel shall be documented in an easily auditable format. Individual record documentation shall include:</p> <p>(1) Education, experience, and employment history and most recent health evaluation summary (e.g., NRC Form 396);</p> <p>(2) Training programs completed and qualification/certification achieved;</p> <p>(3) Latest completed checklists, graded written examinations (with answers corrected as necessary or examination keys), simulator examinations (where applicable), and operational evaluations used for qualification/certification. The record should include an evaluation of the knowledge and performance of the operator/supervisor during operational evaluations;</p> <p>(4) Lists of questions asked and the examiner's overall evaluation of the operator/supervisor's responses on oral examinations;</p> <p>(5) Correspondence relating to exceptions to training requirements and extensions of qualification/certification;</p> <p>(6) Records of qualification for one-time-only special tests or operations; and</p> <p>(7) Attendance records for required training courses or sessions.</p>			

1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): NO

2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):

All specifics of the statement are not addressed in HASQAP subsection 3.7.

3.0 CORRECTIVE ACTION:

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch.I.16.b.	Mandatory (Y/N):	Y
<p>Statement:</p> <p>GENERAL REQUIREMENTS. RECORD REQUIREMENTS. A historical record that documents initial qualification or certification, and applicable information from the above list that verifies the most recent qualification or certification shall be retained in individual records. Superseded information should be handled in accordance with the procedures contained in DOE 1324.2A, RECORDS DISPOSITION.</p>			

1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): NO

2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):

The following portion of the statement is not addressed in the HASQAP:

Superseded information should be handled in accordance with the procedures contained in DOE 1324.2A, RECORDS DISPOSITION.

3.0 CORRECTIVE ACTION:

ASSESSMENT DATA COLLECTION FORM

The following Chapter II statements are specific to DOE Category A reactor facilities and are, therefore, **N/A** to analytical laboratories.

Statement Number:	Ch. II. 2. a.	Mandatory (Y/N):	Y
<p>Statement:</p> <p>GENERAL REQUIREMENTS. ENTRY-LEVEL REQUIREMENTS. Managers. ...are responsible for nuclear safety, operational efficiency and reliability, control of onsite emergencies, and any other activities necessary to safeguard the health and safety of the workforce, the general public, and the environment. Operational responsibilities include prioritizing and assessing facility activities including modifications, and overseeing the operating organization. Administrative responsibilities include maintenance of a qualified staff, budgets, maintaining employee performance, administering disciplinary actions consistent with company policies, public information, and coordination with corporate offices...</p>			

Statement Number:	Ch. II. 2. a. (1)	Mandatory (Y/N):	Y						
<p>Statement:</p> <p>GENERAL REQUIREMENTS. ENTRY-LEVEL REQUIREMENTS. Plant Manager.</p> <p>(a) Education: Baccalaureate in engineering or related science</p> <p>(b) Experience:</p> <table border="0"> <tr> <td>Nuclear</td> <td>6 years</td> </tr> <tr> <td>Supervisory or Management</td> <td>4 years</td> </tr> <tr> <td>Onsite</td> <td>6 months</td> </tr> </table> <p>(c) Special Requirements:</p> <ol style="list-style-type: none"> 1 Three years of the required nuclear experience may be power plant experience; and 2 The Plant Manager shall hold, or have held, a senior reactor operator certification for a similar Category A reactor plant (or equivalent) or have been certified at an appropriate simulator. Plant Managers who have an assistant holding a senior reactor operator certification need not meet this special requirement. 				Nuclear	6 years	Supervisory or Management	4 years	Onsite	6 months
Nuclear	6 years								
Supervisory or Management	4 years								
Onsite	6 months								

Statement Number:	Ch. II. 2. a. (2)	Mandatory (Y/N):	Y				
<p>Statement:</p> <p>CATEGORY A REACTOR PERSONNEL. ENTRY-LEVEL REQUIREMENTS. Operations Manager.</p> <p>(a) Education: Baccalaureate in engineering or related science</p> <p>(b) Experience:</p> <table border="0"> <tr> <td>Nuclear</td> <td>4 years</td> </tr> <tr> <td>Onsite</td> <td>6 months</td> </tr> </table> <p>(c) Special Requirements:</p> <ol style="list-style-type: none"> 1 One year of the required nuclear experience may be power plant experience; and 2 The Operations Manager shall hold a senior reactor operator certification at the time of appointment to the position. 				Nuclear	4 years	Onsite	6 months
Nuclear	4 years						
Onsite	6 months						

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch.II.2.a.(3)	Mandatory (Y/N):	Y
<p>Statement:</p> <p>CATEGORY A REACTOR PERSONNEL. ENTRY-LEVEL REQUIREMENTS. Maintenance Manager.</p> <p>(a) Education: Baccalaureate in engineering or related science</p> <p>(b) Experience:</p> <p style="padding-left: 40px;">Nuclear 4 years</p> <p style="padding-left: 40px;">Onsite 6 months</p> <p>(c) Special Requirements:</p> <p style="padding-left: 20px;">1 Two years of the required nuclear experience may be power plant experience; and</p> <p style="padding-left: 20px;">2 The Maintenance Manager shall be familiar with nondestructive testing and have an understanding of electrical, pressure vessel, and piping codes and standards.</p>			

Statement Number:	Ch.II.2.a.(4)	Mandatory (Y/N):	Y
<p>Statement:</p> <p>CATEGORY A REACTOR PERSONNEL. ENTRY-LEVEL REQUIREMENTS. Technical Manager.</p> <p>(a) Education: Baccalaureate in engineering or related science</p> <p>(b) Experience:</p> <p style="padding-left: 40px;">Nuclear 4 years</p> <p style="padding-left: 40px;">Onsite 6 months</p> <p>(c) Special Requirements:</p> <p style="padding-left: 20px;">1 One year of the required nuclear experience may be power plant experience; and</p> <p style="padding-left: 20px;">2 The Technical Manager shall hold, or have held, a senior reactor operator certification for a similar Category A reactor plant (or equivalent) or have been certified at an appropriate simulator.</p>			

Statement Number:	Ch.II.2.a.(5)	Mandatory (Y/N):	Y
<p>Statement:</p> <p>CATEGORY A REACTOR PERSONNEL. ENTRY-LEVEL REQUIREMENTS. Training Manager.</p> <p>(a) Education: Baccalaureate including courses in education and technical subjects</p> <p>(b) Experience:</p> <p style="padding-left: 40px;">Job related 4 years</p> <p style="padding-left: 40px;">which shall include,</p> <p style="padding-left: 40px;">Nuclear 2 years</p> <p>(c) Special Requirements:</p> <p style="padding-left: 20px;">1 The Training Manager shall have training in educational techniques if not included in baccalaureate course material; and</p> <p style="padding-left: 20px;">2 If the Training Manager does not hold a senior reactor operator certification, another person who holds a senior reactor operator certification shall be responsible to the Training Manager for the content and conduct of the certified operator training program.</p>			

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch.II.2.b.	Mandatory (Y/N):	Y
Statement: CATEGORY A REACTOR PERSONNEL. ENTRY-LEVEL REQUIREMENTS. Supervisors. ...duties include ensuring that work is performed in compliance with procedures, policies, and industrial safety practices.			

Statement Number:	Ch.II.2.b.(1)	Mandatory (Y/N):	Y
Statement: CATEGORY A REACTOR PERSONNEL. ENTRY-LEVEL REQUIREMENTS. Supervisors. Shift Supervisor. (a) Education: High School Diploma (b) Experience: Nuclear 4 years (c) Special Requirements: 1 Two years of the required nuclear experience may be power plant experience; and 2 The shift supervisor shall hold and maintain senior reactor operator certification.			

Statement Number:	Ch.II.2.b.(2)	Mandatory (Y/N):	Y
Statement: CATEGORY A REACTOR PERSONNEL. ENTRY-LEVEL REQUIREMENTS. Supervisors. Senior Reactor Operator. (a) Education: High School Diploma (b) Experience: Power Plant 4 years Nuclear 2 years Onsite 6 months (c) Special Requirements: 1 Two years of the power plant experience may be fulfilled by academic or related technical training; 2 If the candidate for senior reactor operator does not possess a baccalaureate in engineering or equivalent, the candidate shall have one year experience as a reactor operator at the reactor for which certification is sought; 3 Candidates for senior reactor operator with a baccalaureate in engineering or equivalent shall participate in reactor plant operations at power levels of at least 20 percent for at least six weeks, while assigned in the control room as a reactor operator candidate; and 4 Candidates for senior reactor operator with a baccalaureate in engineering or equivalent shall perform all control manipulations that a reactor operator candidate would perform.			

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch.II.2.b.(3)	Mandatory (Y/N):	Y						
<p>Statement:</p> <p>CATEGORY A REACTOR PERSONNEL. ENTRY-LEVEL REQUIREMENTS. Supervisors. Qualified Supervisor.</p> <p>(a) Education: High School Diploma</p> <p>(b) Experience:</p> <table> <tr> <td>Job related</td> <td>4 years</td> </tr> <tr> <td>Nuclear</td> <td>1 year</td> </tr> <tr> <td>Onsite</td> <td>3 months</td> </tr> </table>				Job related	4 years	Nuclear	1 year	Onsite	3 months
Job related	4 years								
Nuclear	1 year								
Onsite	3 months								

Statement Number:	Ch.II.2.c.	Mandatory (Y/N):	Y
<p>Statement:</p> <p>CATEGORY A REACTOR PERSONNEL. ENTRY-LEVEL REQUIREMENTS. Technical Support Personnel. ...have expertise in mechanical, electrical, instrumentation and control, chemistry, radiation protection, training, safety, quality assurance, or reactor engineering. Unless otherwise stated, the basic education requirement is a baccalaureate in engineering or related science; the experience requirement is 2 years job-related, of which 1 year shall be nuclear experience.</p>			

Statement Number:	Ch.II.2.c.(1)	Mandatory (Y/N):	Y								
<p>Statement:</p> <p>CATEGORY A REACTOR PERSONNEL. ENTRY-LEVEL REQUIREMENTS. Technical Support Personnel. Reactor Engineering.</p> <p>(a) Education: Baccalaureate in engineering or related science</p> <p>(b) Experience:</p> <table> <tr> <td>Job related</td> <td>4 years</td> </tr> <tr> <td>which shall include,</td> <td></td> </tr> <tr> <td>Nuclear</td> <td>2 years</td> </tr> <tr> <td>Onsite</td> <td>6 months</td> </tr> </table> <p>(c) Special Requirement: Nuclear experience shall be in such areas as reactor physics, core measurements, core heat transfer, and core physics testing programs.</p>				Job related	4 years	which shall include,		Nuclear	2 years	Onsite	6 months
Job related	4 years										
which shall include,											
Nuclear	2 years										
Onsite	6 months										

Statement Number:	Ch.II.2.c.(2)	Mandatory (Y/N):	Y								
<p>Statement:</p> <p>CATEGORY A REACTOR PERSONNEL. ENTRY-LEVEL REQUIREMENTS. Technical Support Personnel. Instrumentation and Control.</p> <p>(a) Education: Associate Degree in engineering or related science</p> <p>(b) Experience:</p> <table> <tr> <td>Job related</td> <td>2 years</td> </tr> <tr> <td>which shall include,</td> <td></td> </tr> <tr> <td>Nuclear</td> <td>1 year</td> </tr> <tr> <td>Onsite</td> <td>6 months</td> </tr> </table>				Job related	2 years	which shall include,		Nuclear	1 year	Onsite	6 months
Job related	2 years										
which shall include,											
Nuclear	1 year										
Onsite	6 months										

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch. II. 2. c. (3)	Mandatory (Y/N):	Y
<p>Statement:</p> <p>CATEGORY A REACTOR PERSONNEL. ENTRY-LEVEL REQUIREMENTS. Technical Support Personnel. Chemistry and Radiochemistry.</p> <p>(a) Education: Baccalaureate in chemistry or related science</p> <p>(b) Experience:</p> <p>Job related 2 years</p> <p>which shall include,</p> <p>Nuclear 1 year</p> <p>Onsite 6 months</p> <p>(c) Special Requirement: One year of nuclear experience shall be in radiochemistry.</p>			

Statement Number:	Ch. II. 2. c. (4)	Mandatory (Y/N):	Y
<p>Statement:</p> <p>CATEGORY A REACTOR PERSONNEL. ENTRY-LEVEL REQUIREMENTS. Technical Support Personnel. Radiation Protection.</p> <p>(a) Education: Baccalaureate in a science or engineering subject, including formal training in radiation protection.</p> <p>(b) Experience:</p> <p>Job related 4 years</p> <p>which shall include,</p> <p>Nuclear 3 years</p> <p>Onsite 6 months</p> <p>(c) Special Requirement. The three years nuclear experience shall be professional-level.</p>			

Statement Number:	Ch. II. 2. c. (5)	Mandatory (Y/N):	Y
<p>Statement:</p> <p>CATEGORY A REACTOR PERSONNEL. ENTRY-LEVEL REQUIREMENTS. Technical Support Personnel. Preoperational Testing Engineer.</p> <p>(a) Education: Baccalaureate in engineering or related science</p> <p>(b) Experience:</p> <p>Nuclear 1 year</p> <p>(c) Special Requirements:</p> <p>1 The required nuclear experience may be power plant experience; and</p> <p>2 These persons shall be knowledgeable of test program administration and the design and operational performance requirements of the system and equipment being tested and its interaction with other systems.</p>			

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch.II.2.c.(6)	Mandatory (Y/N):	Y
<p>Statement:</p> <p>CATEGORY A REACTOR PERSONNEL. ENTRY-LEVEL REQUIREMENTS. Technical Support Personnel. Startup Testing Engineer.</p> <p>(a) Education: Baccalaureate in engineering or related science</p> <p>(b) Experience:</p> <p style="padding-left: 40px;">Nuclear 2 years</p> <p>(c) Special Requirements:</p> <p style="padding-left: 20px;">1 One year of the required nuclear experience may be power plant experience; and</p> <p style="padding-left: 20px;">2 This person shall be knowledgeable of test program administration, the system design and operational requirements, and expected plant operational characteristics during the test.</p>			

Statement Number:	Ch.II.2.c.(7)	Mandatory (Y/N):	Y
<p>Statement:</p> <p>CATEGORY A REACTOR PERSONNEL. ENTRY-LEVEL REQUIREMENTS. Technical Support Personnel. Training Coordinator.</p> <p>(a) Education: High School Diploma</p> <p>(b) Experience:</p> <p style="padding-left: 40px;">Nuclear 2 years</p> <p style="padding-left: 40px;">Onsite 6 months</p> <p>(c) Special Requirement: The required nuclear experience may be power plant experience.</p>			

Statement Number:	Ch.II.2.c.(8)	Mandatory (Y/N):	Y
<p>Statement:</p> <p>CATEGORY A REACTOR PERSONNEL. ENTRY-LEVEL REQUIREMENTS. Technical Support Personnel. Training Instructor.</p> <p>(a) Education: High School Diploma</p> <p>(b) Experience: Consistent with the material being presented.</p> <p>(c) Special Requirements:</p> <p style="padding-left: 20px;">1 Instructors who provide instruction on the reactor plant simulator to certified personnel shall hold, or have held, a senior reactor operator certification for a similar Category A reactor plant (or equivalent) or have been certified on the reactor plant simulator;</p> <p style="padding-left: 20px;">2 Instructors who are responsible for instruction of subjects such as technical specifications, operating practice, and control manipulations shall have received senior reactor operator (or equivalent) training; and</p> <p style="padding-left: 20px;">3 Instructors shall have demonstrated knowledge of instructional techniques through training or experience and be qualified by the Training Manager for the material being presented.</p>			

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch.II.2.c.(9)	Mandatory (Y/N):	Y
<p>Statement:</p> <p>CATEGORY A REACTOR PERSONNEL. ENTRY-LEVEL REQUIREMENTS. Technical Support Personnel. Shift Technical Advisor.</p> <p>(a) Education: Baccalaureate in engineering or related science</p> <p>(b) Experience:</p> <p style="padding-left: 40px;">Nuclear 1 year</p> <p style="padding-left: 40px;">Onsite 6 months</p> <p>(c) Special Requirement: This person shall be knowledgeable of control room instruments and controls and be assigned to advise the responsible shift supervisor concerning abnormal plant operating conditions.</p>			

Statement Number:	Ch.II.2.d.	Mandatory (Y/N):	Y
<p>Statement:</p> <p>CATEGORY A REACTOR PERSONNEL. ENTRY-LEVEL REQUIREMENTS. Operators, Technicians, and Maintenance Personnel. ...Persons in training or apprentice positions shall not be considered operators, technicians, or maintenance personnel, but may perform work in the specific tasks for which qualification has been achieved. Persons in training or apprentice positions may perform tasks under the immediate supervision of a person qualified to perform that specific task.</p>			

Statement Number:	Ch.II.2.d.(1)	Mandatory (Y/N):	Y
<p>Statement:</p> <p>CATEGORY A REACTOR PERSONNEL. ENTRY-LEVEL REQUIREMENTS. Operators, Technicians, and Maintenance Personnel. Qualified Auxiliary Operators.</p> <p>(a) Education: High School Diploma</p> <p>(b) Experience: Qualified auxiliary operators whose actions could affect the quality of structures, systems, and components important to safety shall have 1 year of nuclear experience.</p> <p>(c) Special Requirement: The required nuclear experience may be power plant experience.</p>			

Statement Number:	Ch.II.2.d.(2)	Mandatory (Y/N):	Y
<p>Statement:</p> <p>CATEGORY A REACTOR PERSONNEL. ENTRY-LEVEL REQUIREMENTS. Operators, Technicians, and Maintenance Personnel. Reactor Operators.</p> <p>(a) Education: High School Diploma</p> <p>(b) Experience:</p> <p style="padding-left: 40px;">Nuclear 3 years</p> <p>(c) Special Requirements:</p> <ol style="list-style-type: none"> 1. Two years of the required nuclear experience may be power plant experience; and 2. Six months of the required experience shall be experience as a qualified auxiliary operator at the reactor for which certification is sought. 			

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch.II.2.d.(3)	Mandatory (Y/N):	Y
Statement: CATEGORY A REACTOR PERSONNEL. ENTRY-LEVEL REQUIREMENTS. Operators, Technicians, and Maintenance Personnel. Technicians. (a) Education: High School Diploma (b) Experience: Job related experience 3 years			

Statement Number:	Ch.II.2.d.(4)	Mandatory (Y/N):	Y
Statement: CATEGORY A REACTOR PERSONNEL. ENTRY-LEVEL REQUIREMENTS. Operators, Technicians, and Maintenance Personnel. Maintenance Personnel. (a) Education: Journeyman level (b) Experience: Related maintenance experience 3 years			

Statement Number:	Ch.II.3.	Mandatory (Y/N):	Y
Statement: CATEGORY A REACTOR PERSONNEL. MEDICAL EXAMINATION REQUIREMENTS. A medical examination shall be given to prospective employees and a reexamination shall be given at least every two years to certified operators and supervisors to verify health and physical fitness to safely perform their assigned tasks. Certified operators and supervisors must also be cleared by medical examination prior to returning to work following any serious illness or injury which keeps the person from performing their duties for a period exceeding one month. Medical examination requirements shall be in accordance with ANSI/ANS 3.4-1983, "Medical Certification and Monitoring of Personnel Requiring Operator Licenses for Nuclear Power Plants," and NRC Regulatory Guide 1.134, Rev. 2, 4-87, "Medical Evaluation of Nuclear Power Plant Personnel Requiring Operator Licenses." Medical examination requirements for other operating organization personnel shall be in accordance with the procedures of the operating contractor.			

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch.II.4.	Mandatory (Y/N):	Y
<p>Statement:</p> <p>CATEGORY A REACTOR PERSONNEL. ENGINEERING EXPERTISE ON SHIFT REQUIREMENTS. The operating organization shall ensure that the operating shift possesses adequate engineering and accident assessment expertise. This may be accomplished by designating a Shift Technical Advisor (STA) for each shift, or by combining the shift supervisor or an on-shift senior reactor operator with the STA position. If the combined approach is utilized, the designated STA shall meet the following qualifications:</p> <p>a. Currently certified as a senior reactor operator; and</p> <p>b. Successful completion of the STA training requirements in subparagraph 7a, and one of the following educational requirements:</p> <p>(1) Baccalaureate in engineering;</p> <p>(2) Professional engineer's license;</p> <p>(3) Baccalaureate in engineering technology including course work in the physical, mathematical, or engineering sciences; or</p> <p>(4) Baccalaureate in a physical science including course work in the physical, mathematical, or engineering sciences.</p>			

Statement Number:	Ch.II.5.	Mandatory (Y/N):	Y
<p>Statement:</p> <p>CATEGORY A REACTOR PERSONNEL. STAFFING REQUIREMENTS. The operating organization shall ensure that minimum staffing requirements are met for safe and reliable operations. Any temporary deviation from these requirements must be justified by a facility-specific analysis and approved by the Head of the Field Organization.</p>			

Statement Number:	Ch.II.5.a.(1)	Mandatory (Y/N):	Y
<p>Statement:</p> <p>CATEGORY A REACTOR PERSONNEL. STAFFING REQUIREMENTS. Facility Staffing. Overall staffing shall be based on the facility Safety Analysis Report and Technical Specifications. Facility operation shall be contingent upon meeting the criteria contained in these documents.</p>			

Statement Number:	Ch.II.5.a.(2)	Mandatory (Y/N):	Y
<p>Statement:</p> <p>CATEGORY A REACTOR PERSONNEL. STAFFING REQUIREMENTS. Facility Staffing. A list of facility personnel by name, title, and work and home telephone numbers shall be readily available in the control room for use by the operating crew. The list shall include:</p> <p>(a) Management personnel;</p> <p>(b) Radiation safety personnel; and</p> <p>(c) Technical support personnel.</p>			

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch.II.5.b.(1)	Mandatory (Y/N):	Y
<p>Statement:</p> <p>CATEGORY A REACTOR PERSONNEL. STAFFING REQUIREMENTS. Control Room Staffing. A certified reactor operator or senior reactor operator shall be at the controls in the control room at all times when the reactor is fueled.</p>			

Statement Number:	Ch.II.5.b.(2)	Mandatory (Y/N):	Y
<p>Statement:</p> <p>CATEGORY A REACTOR PERSONNEL. STAFFING REQUIREMENTS. Control Room Staffing. A certified senior reactor operator shall be in the control room at all times when the reactor is operating or in an operational mode other than cold shutdown or refueling as defined by the Technical Specifications.</p>			

Statement Number:	Ch.II.5.b.(3)	Mandatory (Y/N):	Y
<p>Statement:</p> <p>CATEGORY A REACTOR PERSONNEL. STAFFING REQUIREMENTS. Control Room Staffing. A certified senior reactor operator, or senior reactor operator limited to fuel handling, shall be present at the facility to directly supervise any alteration of the core including fuel loading or transfer. This person shall have no other concurrent duties.</p>			

Statement Number:	Ch.II.5.b.(4)	Mandatory (Y/N):	Y
<p>Statement:</p> <p>CATEGORY A REACTOR PERSONNEL. STAFFING REQUIREMENTS. Control Room Staffing. The operator at the controls of a Category A reactor should have an unobstructed view of, and access to, the operational control panels, including instrumentation displays and alarms, to be able to initiate prompt corrective action when necessary on receipt of indication (instrument response or alarm) of a changing condition.</p>			

Statement Number:	Ch.II.5.b.(5)	Mandatory (Y/N):	Y
<p>Statement:</p> <p>CATEGORY A REACTOR PERSONNEL. STAFFING REQUIREMENTS. Control Room Staffing. Administrative procedures shall be established that define and outline, preferably with sketches, the specific areas where the senior reactor operator and reactor operator must remain to be considered "in the control room" and "at the controls," respectively. These procedures should require the senior reactor operator in the control room to be in sight of, or in audible range of, the reactor operator at the controls or in audible range of the control room annunciators.</p>			

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch. II. 6.	Mandatory (Y/N):	Y
<p>Statement:</p> <p>CATEGORY A REACTOR PERSONNEL. SIMULATOR REQUIREMENTS. DOE production reactors shall have a full-scope simulator that meets the requirements contained in ANSI/ANS 3.5-1985, "Nuclear Power Plant Simulators for use in Operator Training," and Regulatory Guide Positions C.5 and C.6 contained in Nuclear Regulatory Commission Regulatory Guide 1.149, Revision 1, of April 1987, "Nuclear Power Plant Simulation Facilities for Use in Operator Licensing Examinations," for reactor operator and senior reactor operator training.</p>			

Statement Number:	Ch. II. 6. a.	Mandatory (Y/N):	Y
<p>Statement:</p> <p>CATEGORY A REACTOR PERSONNEL. SIMULATOR REQUIREMENTS. If a full-scope simulator is to be utilized for more than one production reactor, the differences between the simulator and reactor shall be identified and documented by the operating organization and shall be approved by the Head of the Field Organization. These differences should not be so significant that they have an adverse impact on the ability of the simulator to meet the requirements and guidance of ANSI/ANS 3.5-1985.</p>			

Statement Number:	Ch. II. 6. b.	Mandatory (Y/N):	Y
<p>Statement:</p> <p>CATEGORY A REACTOR PERSONNEL. SIMULATOR REQUIREMENTS. The need for a full-scope or partial-task simulator for DOE Category A test and research reactors shall be based on evaluations conducted by the operating organization. These evaluations shall include determination of the ability to adequately provide training in the facility covering all operator actions, where timely operator action must be taken to bring the reactor to, or maintain it in a safe state. The evaluation should also assess the ability to provide adequate training in normal operations, anticipated transients, and accident conditions. Adequate training may, however, be achieved by actual plant maneuvers, drills, partial-task simulators, or combinations of these. Training conducted in the facility shall not lead to or have the potential for safety concerns. The determination of the need for a simulator, for other than production reactors, shall be approved by the Head of the Field Organization and the cognizant PSO.</p>			

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch.II.7.a.	Mandatory (Y/N):	Y
<p>Statement:</p> <p>CATEGORY A REACTOR PERSONNEL. SPECIFIC TRAINING REQUIREMENTS. Shift Technical Advisor (STA). ... training shall include:</p> <ol style="list-style-type: none"> (1) Accidents analyzed in the facility Safety Analysis Report (SAR) and the consequences of these accidents; (2) Thermodynamics/fluid flow, reactor physics, system engineering, nuclear instrumentation, process computer, and facility response; (3) The duties, responsibilities, and authorities of the STA; (4) Performance of control manipulations on the simulator (for those facilities having simulators), by actual facility maneuvers or drills, or combinations of these; (5) Response to and analysis of facility transients and accidents; and (6) The relationship of accident conditions to offsite consequences and protective action strategies. 			

Statement Number:	Ch.II.7.b.	Mandatory (Y/N):	Y
<p>Statement:</p> <p>CATEGORY A REACTOR PERSONNEL. SPECIFIC TRAINING REQUIREMENTS. Fuel Handling Operations. All fuel handling operations shall be performed by or under the direct supervision of a person certified to perform the required functions. The requirements below are not necessary if fuel handling is performed by persons qualified for such under regular reactor operator and senior reactor operator certification programs:</p> <ol style="list-style-type: none"> (1) A specific qualification program shall be established for fuel handling operators and supervisors. The program shall include training for their assigned tasks; and (2) The qualification program for fuel handling operators and supervisors shall consist of initial and continuing training. Training and examination may be limited to that needed for fuel handling safety, the impact of fuel handling on safety, and actions to be taken during abnormal and emergency conditions. 			

Statement Number:	Ch.II.7.c.	Mandatory (Y/N):	Y
<p>Statement:</p> <p>CATEGORY A REACTOR PERSONNEL. SPECIFIC TRAINING REQUIREMENTS. Control Manipulations. Reactor operator and senior reactor operator candidates shall perform a minimum of five significant reactivity manipulations...for initial reactor operator and senior reactor operator qualification. Additional control manipulations shall be based on the job analysis. The control manipulations list shall specify which control manipulations are to be performed annually and which are to be performed biennially as determined by the analysis. Following initial qualification, the senior reactor operator need only direct control manipulations to satisfy this requirement.</p>			

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch.II.7.d.	Mandatory (Y/N):	Y
<p>Statement:</p> <p>CATEGORY A REACTOR PERSONNEL. SPECIFIC TRAINING REQUIREMENTS. Reactor Operator Written Examinations.shall contain a representative selection of questions on knowledge, skills, and abilities identified from learning objectives developed from the job analysis and from information in facility Safety Analysis Reports (SAR), system description manuals, operating procedures, Technical Specifications, and Unusual Occurrence Reports (UOR). The written examination shall be consistent with the job analysis, and include representative sampling from:</p> <ol style="list-style-type: none"> (1) Fundamentals of reactor theory, including fission process, neutron multiplication, source effects, control rod effects, criticality indications, reactivity coefficients, and poison effects; (2) General design features of the core, including core structure, fuel elements, control rods, core instrumentation, and coolant flow; (3) Mechanical components and design features of the primary system; (4) Secondary coolant and auxiliary systems that affect the facility; (5) Facility operating characteristics and reasons for these characteristics during steady state and transient conditions, including chemistry, cause and effects of temperature, pressure and reactivity changes, effects of load changes (if applicable), and operating limits; (6) Reactivity control mechanisms and instrumentation, including design, components, and functions; (7) Control and safety systems, including design, components, functions, instrumentation, signals, interlocks, failure modes, and automatic and manual features; (8) Emergency systems, including components, capacity, and functions; (9) Shielding, isolation, and containment/confinement design features, including access limitations; (10) Operating procedures, including administrative, normal, abnormal, and emergency; (11) Radiation monitoring systems, including purpose, operation, alarms, and survey equipment; (12) Radiological safety principles and procedures; (13) Radioactive materials and effluents, including procedures and equipment available for handling and disposal; (14) Principles of heat transfer, thermodynamics and fluid mechanics; and (15) Use of installed plant systems for the control and mitigation of an accident in which the core may be severely damaged. 			

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch.II.7.e.	Mandatory (Y/N):	Y
<p>Statement:</p> <p>CATEGORY A REACTOR PERSONNEL. SPECIFIC TRAINING REQUIREMENTS: Senior Reactor Operator Written Examinations. ...shall contain, consistent with the job analysis, the following topics in addition to those required for reactor operators:</p> <ol style="list-style-type: none"> (1) Conditions and limitations for facility operations; (2) Operating limitations in the Technical Specifications and their bases; (3) Procedures required to obtain authority for design and operating changes in the facility; (4) Radiation hazards that may arise during normal and abnormal situations, including maintenance activities and various contamination conditions; (5) Assessment of conditions and selection of appropriate procedures during normal, abnormal, and emergency situations; (6) Procedures and limitations for initial core loading, alterations in core configuration, control rod programming, and determination of various internal and external effects on core reactivity; and (7) Fuel handling facilities and procedures. 			

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch. II.7.f.	Mandatory (Y/N):	Y
-------------------	-------------	------------------	---

Statement:

CATEGORY A REACTOR PERSONNEL. SPECIFIC TRAINING REQUIREMENTS.

Operational Evaluations. ...shall include a facility walkthrough or a combination walkthrough and simulator examination (for those DOE facilities having simulators), and may include system and/or component operation. Operational evaluations, to the extent possible and consistent with the job analysis, shall require the candidate to demonstrate an understanding of, and the ability to perform the actions necessary to:

- (1) Perform prestartup procedures including operation of controls associated with equipment which could affect reactivity;
- (2) Manipulate the console controls as required to operate between shutdown and designated power levels;
- (3) Identify annunciators and condition-indicating signals and perform appropriate remedial actions;
- (4) Identify instrumentation systems and the significance of the instrument readings;
- (5) Observe and safely control the operating behavior characteristics of the facility;
- (6) Perform control manipulations to obtain desired operating results during normal, abnormal, and emergency situations;
- (7) Safely operate heat removal systems, including the primary coolant, emergency coolant, and decay heat removal systems and explain relationships between proper operation of these systems to the operation of the facility;
- (8) Safely operate auxiliary and emergency systems including controls of facility equipment that could affect reactivity or release radioactive or other hazardous materials to the environment;
- (9) Demonstrate or describe the use and function of radiation monitoring systems, including fixed radiation monitors and alarms, portable survey instruments, and personnel monitoring equipment;
- (10) Demonstrate knowledge of significant radiation hazards, including permissible levels in excess of those authorized, and the ability to perform other procedures to reduce excessive radiation levels and to guard against personnel exposure;
- (11) Demonstrate knowledge of the emergency plan, including, as appropriate, reactor operator or senior reactor operator responsibility to decide whether the plan should be executed and assigned duties under the plan;
- (12) Demonstrate knowledge and ability, as appropriate to the assigned position, to assume the responsibilities associated with safe operation of the facility; and
- (13) Demonstrate the ability to function within the control room team in such a way that procedures are adhered to and Technical Specifications are not violated.

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch.II.7.g.(1)	Mandatory (Y/N):	Y
<p>Statement:</p> <p>CATEGORY A REACTOR PERSONNEL. SPECIFIC TRAINING REQUIREMENTS. Operating Crew/Shift Training. Reactor Operators. ...trainees shall be assigned to an operating crew full-time for a minimum of 3 months shift training with no concurrent duties that are not related to the operation of the facility. During this period, under the observation and control of a certified reactor operator, the trainee shall manipulate the facility controls and perform the same duties as a certified reactor operator.</p>			
Statement Number:	Ch.II.7.g.(2)	Mandatory (Y/N):	Y
<p>Statement:</p> <p>CATEGORY A REACTOR PERSONNEL. SPECIFIC TRAINING REQUIREMENTS. Operating Crew/Shift Training. Senior Reactor Operator. ...trainees shall be assigned to an operating crew full-time for a minimum of 3 months shift training with no concurrent duties that are not related to the operation of the facility. During this period, under the observation and control of a certified senior reactor operator, the trainee shall supervise the manipulation of the facility controls and perform the same duties as a certified senior reactor operator.</p>			
Statement Number:	Ch.II.8.	Mandatory (Y/N):	Y
<p>Statement:</p> <p>CATEGORY A REACTOR PERSONNEL. OPERATOR PROFICIENCY REQUIREMENTS. In order to maintain proficiency, all certified reactor operators and senior reactor operators must actively perform job functions associated with their certification.</p> <p>a. To maintain proficiency, operators shall perform certification duties:</p> <ol style="list-style-type: none"> (1) During 5 eight-hour shifts per calendar quarter; or (2) During 3 twelve-hour shifts per calendar quarter. <p>b. If the proficiency requirement above is not met, certification shall be suspended and the person shall not be assigned certification duties until certification is regained. Certification may be regained by performing certified operator duties under the direct supervision of a certified operator for a minimum period of 24 hours and must include a complete tour of the facility and all shift turnover procedures.</p>			

The above Chapter II statements are specific to DOE Category A reactor facilities and are, therefore, **N/A** to analytical laboratories.

ASSESSMENT DATA COLLECTION FORM

The following Chapter III statements are specific to DOE Category B reactor facilities and are, therefore, **N/A** to analytical laboratories.

Statement Number:	Ch.III.2.a.[1]	Mandatory (Y/N):	Y
<p>Statement:</p> <p>CATEGORY B REACTOR PERSONNEL. ENTRY-LEVEL REQUIREMENTS. Managers. ...are responsible for nuclear safety, operational efficiency and reliability, control of onsite emergencies, and any other activities necessary to safeguard the health and safety of the workforce, the general public, and the environment. Operational responsibilities include prioritizing and assessing facility activities including modifications, and overseeing the operating organization. Administrative responsibilities include maintenance of a qualified staff, budgets, maintaining employee performance, administering disciplinary actions consistent with company policies, public information, and coordination with corporate offices.</p>			

Statement Number:	Ch.III.2.a.[2]	Mandatory (Y/N):	Y
<p>Statement:</p> <p>CATEGORY B REACTOR PERSONNEL. ENTRY-LEVEL REQUIREMENTS. Managers. This functional level typically includes the Plant/Facility Manager or Director, the Operations Manager, the Maintenance Manager, the Training Manager, and the Technical Manager. Prior to assuming the duties of the assigned position, persons at the manager level shall meet the following requirements:</p> <p>(1) Education: Baccalaureate in engineering or related science</p> <p>(2) Experience: Nuclear 6 years</p> <p>(3) Special Requirements:</p> <p>(a) Education or experience that is job-related may be substituted for a degree on a case-by-case basis. The degree may fulfill 4 of the 6 years of nuclear experience required on a one-for-one time basis. Experience acquired at a nuclear power, test, research, or production reactor, or a critical facility counts on a one-for-one time basis;</p> <p>(b) Managers shall receive facility-specific training based upon a comparison of the individual's background and abilities with the responsibilities and duties of the position; and</p> <p>(c) The Training Manager shall have a baccalaureate including courses in education and technical subjects (baccalaureate need not be in engineering or related science).</p>			

Statement Number:	Ch.III.2.b.[1]	Mandatory (Y/N):	Y
<p>Statement:</p> <p>CATEGORY B REACTOR PERSONNEL. ENTRY-LEVEL REQUIREMENTS. Supervisors. ...duties include ensuring that work is performed in compliance with procedures, policies, and industrial safety practices.</p>			

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch.III.2.b.[2]	Mandatory (Y/N):	Y
<p>Statement:</p> <p>CATEGORY B REACTOR PERSONNEL. ENTRY-LEVEL REQUIREMENTS. Supervisors. This functional level typically includes the reactor or shift supervisor and may be combined with the senior reactor operator position at smaller facilities. Prior to assuming the duties of the assigned position, supervisors shall meet the following requirements:</p> <p>(1) Education: High School Diploma</p> <p>(2) Experience: Nuclear 3 years</p> <p>(a) Experience acquired at critical facilities counts on a one-for-one time basis; and</p> <p>(b) Full-time academic training may be substituted on a one-for-one basis for 2 of the 3 years of required nuclear experience.</p> <p>(3) Special Requirement: The reactor or shift supervisor shall be certified as a senior reactor operator.</p>			

Statement Number:	Ch.III.2.c.	Mandatory (Y/N):	Y
<p>Statement:</p> <p>CATEGORY B REACTOR PERSONNEL. ENTRY-LEVEL REQUIREMENTS. Operators. ...At Category B reactor facilities this position includes the reactor operator and senior reactor operator.</p> <p>(1) Education: High School Diploma</p>			

Statement Number:	Ch.III.2.d.	Mandatory (Y/N):	Y
<p>Statement:</p> <p>CATEGORY B REACTOR PERSONNEL. ENTRY-LEVEL REQUIREMENTS. Technicians.</p> <p>(1) Experience: Job related 1 year</p>			

Statement Number:	Ch.III.2.e.	Mandatory (Y/N):	Y
<p>Statement:</p> <p>CATEGORY B REACTOR PERSONNEL. ENTRY-LEVEL REQUIREMENTS. Maintenance Personnel.</p> <p>(1) Experience: Maintenance related 1 year</p>			

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch.III.2.f.	Mandatory (Y/N):	Y				
<p>Statement:</p> <p>CATEGORY B REACTOR PERSONNEL. ENTRY-LEVEL REQUIREMENTS. Technical Support Personnel. ...have expertise in mechanical, electrical, instrumentation and control, chemistry, radiation protection, training, safety, quality assurance, or reactor engineering. For personnel assigned to equivalent positions, Category B reactor facilities should use the education and experience requirements contained in Chapter II, Category A Reactor Personnel, subparagraph 2c. For positions for which there is no equivalent, the education and experience requirements are as listed below. Education and experience requirements are intended to apply to supervisory positions or positions with authority to review and concur, and not to entry-level positions.</p> <p>(1) Education: Baccalaureate in engineering or related science</p> <p>(2) Experience:</p> <table> <tr> <td>Job related</td> <td>2 years</td> </tr> <tr> <td>Nuclear</td> <td>1 year</td> </tr> </table>				Job related	2 years	Nuclear	1 year
Job related	2 years						
Nuclear	1 year						

Statement Number:	Ch.III.3.	Mandatory (Y/N):	Y
<p>Statement:</p> <p>CATEGORY B REACTOR PERSONNEL. MEDICAL EXAMINATION REQUIREMENTS. A medical examination shall be given to prospective employees and a reexamination shall be given at least every two years to certified operators and supervisors to verify health and physical fitness to safely perform their assigned tasks. Certified operators and supervisors must also be cleared by medical examination prior to returning to work following any serious illness or injury which keeps the person from performing their duties for a period exceeding one month. Medical examinations shall be in accordance with requirements contained in ANSI/ANS 15.4-1988. Medical examination requirements for other operating organization personnel shall be in accordance with the procedures of the operating contractor.</p>			

Statement Number:	Ch.III.4.	Mandatory (Y/N):	Y
<p>Statement:</p> <p>CATEGORY B REACTOR PERSONNEL. STAFFING REQUIREMENTS. The operating organization shall ensure that minimum staffing requirements are met for safe and reliable operations. Any temporary deviation from these requirements must be justified by facility-specific analysis and approved by the Head of the Field Organization.</p>			

Statement Number:	Ch.III.4.a.(1)	Mandatory (Y/N):	Y
<p>Statement:</p> <p>CATEGORY B REACTOR PERSONNEL. STAFFING REQUIREMENTS. Facility Staffing. Overall staffing shall be based on the facility Safety Analysis Report and Technical Specifications. Facility operation shall be contingent upon meeting the criteria contained in these documents.</p>			

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch.III.4.a.(2)	Mandatory (Y/N):	Y
<p>Statement:</p> <p>CATEGORY B REACTOR PERSONNEL. STAFFING REQUIREMENTS. Facility Staffing. A list of facility personnel by name, title, and work and home telephone numbers shall be readily available in the control room for use by the operating crew. The list shall include:</p> <ul style="list-style-type: none"> (a) Management personnel; (b) Radiation safety personnel; and (c) Technical support personnel. 			

Statement Number:	Ch.III.4.b.	Mandatory (Y/N):	Y
<p>Statement:</p> <p>CATEGORY B REACTOR PERSONNEL. STAFFING REQUIREMENTS. Control Room Staffing.</p> <ul style="list-style-type: none"> (1) A certified reactor operator or senior reactor operator shall be at the controls at all times during the operation of the reactor. (2) A certified senior reactor operator shall be present at the reactor facility or readily available on call at all times during operation and refueling, and shall be present in the control room during start-up and approach to power, recovery from an unplanned or unscheduled shutdown or significant reduction in power, or as otherwise prescribed by the Technical Specifications. 			

Statement Number:	Ch.III.5.	Mandatory (Y/N):	Y
<p>Statement:</p> <p>CATEGORY B REACTOR PERSONNEL. SPECIFIC TRAINING REQUIREMENTS. The qualification program shall include classroom-type and on-the-job training to assure familiarity with all required aspects of reactor operation, including anticipated transients and accident conditions. Where construction precludes on-the-job training, practical experience at similar reactors, training on simulators, and other appropriate training is acceptable.</p>			

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch.III.5.a.	Mandatory (Y/N):	Y
<p>Statement:</p> <p>CATEGORY B REACTOR PERSONNEL. SPECIFIC TRAINING REQUIREMENTS. Senior Reactor Operator Training.</p> <p>(1) Senior reactor operator training shall be sufficiently comprehensive to develop the knowledge and skills commensurate with the position and cover areas which are fundamental to the candidate's job duties.</p> <p>(2) Initial and continuing training shall include topics from the following categories as applicable to the facility and the requirements of the job:</p> <p>(a) Reactor facility design and operating characteristics, including features of facility design, design and operating characteristics and limitations, safety and emergency systems, experiment and test facilities, engineered safety features, and shielding;</p> <p>(b) Principles of reactor facility operation, including principles of reactor operation, radiological protection, effects of experiments, basic reactor theory, and heat transfer, fluid flow, and thermodynamics, as necessary, for the specific design of the reactor;</p> <p>(c) Instrumentation and control, including nuclear instruments, process instruments, control systems, radiation monitoring systems and survey equipment, experiment and test facility instrumentation, and manipulation of reactivity controls;</p> <p>(d) Procedures and Technical Specifications, including normal, abnormal, emergency, radiological control, and administrative procedures, and operational limitations;</p> <p>(e) Radioactive materials handling, including Special Nuclear Material and radioactive materials hazards, handling, disposal, and safe practices; and</p> <p>(f) Advanced theory and operation, including reactivity effects during experimental and maintenance activities, fuel handling, fuel burnup and reactivity worth, alterations in core configuration, Technical Specification bases, and administrative responsibilities associated with the facility and appropriate for the senior reactor operator's level of responsibility.</p>			

Statement Number:	Ch.III.5.b.	Mandatory (Y/N):	Y
<p>Statement:</p> <p>CATEGORY B REACTOR PERSONNEL. SPECIFIC TRAINING REQUIREMENTS. Reactor Operator Training. ...should take into account the previous experience, training, and level of responsibility of the candidate. Initial and continuing training shall include the categories contained in subparagraphs 5a(2)(a) through (d) and other categories and topics which are applicable to the facility and to the requirements of the job.</p>			

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch.III.5.c.	Mandatory (Y/N):	Y
<p>Statement:</p> <p>CATEGORY B REACTOR PERSONNEL. SPECIFIC TRAINING REQUIREMENTS. Fuel Handling Operations. All fuel handling operations shall be performed by or under the direct supervision of a person certified to perform the required functions. The requirements below are not necessary if fuel handling is performed by persons qualified for such under regular reactor operator and senior reactor operator certification programs.</p> <p>(1) A specific qualification program shall be established for fuel handling operators and supervisors. The program shall include training for their assigned tasks.</p> <p>(2) The qualification program for operators and supervisors shall consist of initial and continuing training. Training and examination may be limited to that needed for fuel handling safety, the impact of fuel handling on safety, and actions to be taken during abnormal and emergency conditions.</p>			

Statement Number:	Ch.III.5.d.	Mandatory (Y/N):	Y
<p>Statement:</p> <p>CATEGORY B REACTOR PERSONNEL. SPECIFIC TRAINING REQUIREMENTS. Control Manipulations. Reactor operator and senior reactor operator candidates shall perform control manipulations for initial qualification and on a biennial basis. The list of control manipulations should be derived from a systematic analysis of the position and include reactivity manipulations in any combination of reactor startups, reactor shutdowns, or significant reactivity changes... Additional control manipulations important to the safe operation of the reactor should be included based on the results of the analysis. Following initial qualification, the senior reactor operator need only direct control manipulations to satisfy this requirement.</p>			

Statement Number:	Ch.III.6.	Mandatory (Y/N):	Y
<p>Statement:</p> <p>CATEGORY B REACTOR PERSONNEL. OPERATOR PROFICIENCY REQUIREMENTS. In order to maintain proficiency, all certified reactor operators and senior reactor operators must actively perform job functions associated with their certification for at least four hours per calendar quarter.</p> <p>a. If this requirement is not met, certification shall be suspended and the person shall not be assigned certification duties until certification is regained. Certification may be regained by performing certified operator duties under the direct supervision of a certified operator for a minimum period of six hours.</p> <p>b. If the reactor is not operated frequently enough to meet these requirements, certification shall be reinstated prior to reactor operation by administering written and oral examinations to ensure adequate operational knowledge.</p>			

ASSESSMENT DATA COLLECTION FORM

The above Chapter III statements are specific to DOE Category B reactor facilities and are, therefore, N/A to analytical laboratories.

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch.IV.2.a.[1]	Mandatory (Y/N):	Y
<p>Statement:</p> <p>NON-REACTOR NUCLEAR FACILITY PERSONNEL. ENTRY-LEVEL REQUIREMENTS. Managers. ...are responsible for nuclear safety, operational efficiency and reliability, control of onsite emergencies, and any other activities necessary to safeguard the health and safety of the workforce, the general public, and the environment. Operational responsibilities include prioritizing and assessing facility activities including modifications, and overseeing the operating organization. Administrative responsibilities include maintenance of a qualified staff, budgets, maintaining employee performance, administering disciplinary actions consistent with company policies, public information, and coordination with corporate offices.</p>			

1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): NO

2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):

Specific entry-level requirements for personnel are not addressed in the HASQAP.

3.0 CORRECTIVE ACTION:

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch. IV.2.a. [2]	Mandatory (Y/N):	Y
<p>Statement:</p> <p>NON-REACTOR NUCLEAR FACILITY PERSONNEL ENTRY-LEVEL REQUIREMENTS. Managers. This functional level typically includes the Plant/Facility Manager or Director, the Operations Manager, the Maintenance Manager, the Training Manager, and the Technical Manager. Prior to assuming the duties of the assigned position, persons at the manager level shall meet the following requirements:</p> <p>(1) Education: Baccalaureate in engineering or related science</p> <p>(2) Experience: Nuclear 4 years</p> <p>(3) Special Requirements:</p> <p>(a) Education or experience that is job-related may be substituted for a degree on a case-by-case basis. The degree may fulfill 3 of the 4 years of nuclear experience required on a one-for-one time basis;</p> <p>(b) Managers shall receive facility-specific training based upon a comparison of the individual's background and abilities with the responsibilities and duties of the position; and</p> <p>(c) The Training Manager shall have a baccalaureate including courses in education and technical subjects (baccalaureate need not be in engineering or related science).</p>			

1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): NO

2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):

Specific entry-level requirements for personnel are not addressed in the HASQAP.

3.0 CORRECTIVE ACTION:

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch.IV.2.b.	Mandatory (Y/N):	Y.
<p>Statement:</p> <p>NON-REACTOR NUCLEAR FACILITY PERSONNEL. ENTRY-LEVEL REQUIREMENTS. Supervisors. This functional level describes those persons who are responsible for the quantity and quality of work and who direct the actions of operators, technicians, or maintenance personnel. Their duties include ensuring that work is performed in compliance with procedures, policies, and industrial safety practices. Prior to assuming the duties of the assigned position, supervisors shall meet the following requirements:</p> <p>(1) Education: High School Diploma</p> <p>(2) Experience: Nuclear 3 years</p> <p>(3) Special Requirement: Full-time academic training may be substituted on a one-for-one basis for 2 of the 3 years of required nuclear experience.</p>			

1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): NO

2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):

Specific entry-level requirements for personnel are not addressed in the HASQAP.

3.0 CORRECTIVE ACTION:

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch. IV.2.c.	Mandatory (Y/N):	Y
<p>Statement:</p> <p>NON-REACTOR NUCLEAR FACILITY PERSONNEL. ENTRY-LEVEL REQUIREMENTS. Operators. . . are persons responsible for performing operations associated with safety systems, operating support systems which could affect safety systems, or conducting activities with radioactive materials. Duties may include manipulating facility controls, monitoring parameters, and operating equipment in facility safety systems. Operators include fissionable material handlers, tritium facility operators, chemical process operators, waste tank operators, and enrichment facility operators.</p> <p>(1) Education: High School Diploma</p>			

N/A

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch.IV.2.d.	Mandatory (Y/N):	Y
<p>Statement:</p> <p>NON-REACTOR NUCLEAR FACILITY PERSONNEL. ENTRY-LEVEL REQUIREMENTS. Technicians, ...are persons responsible for specific maintenance activities or analytical laboratory work. Their tasks include equipment maintenance, troubleshooting, repair, testing, instrument calibration, inspections, and data surveys. Technicians interpret and verify field data accumulated from tests such as radiation surveys, instrumentation systems tests, liquid and gaseous analysis, and calibration of electronic circuits.</p> <p>(1) Experience: Job related 1 year</p>			

1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): NO

2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):

Specific entry-level requirements for personnel are not addressed in the HASQAP.

3.0 CORRECTIVE ACTION:

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch.IV.2.e.	Mandatory (Y/N):	Y
Statement: NON-REACTOR NUCLEAR FACILITY PERSONNEL. ENTRY-LEVEL REQUIREMENTS. Maintenance personnel ... are responsible for the maintenance and repair of mechanical and electrical equipment. (1) Experience: Maintenance related 1 year			

N/A

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch.IV.2.f.	Mandatory (Y/N):	Y				
<p>Statement:</p> <p>NON-REACTOR NUCLEAR FACILITY PERSONNEL. ENTRY-LEVEL REQUIREMENTS. Technical Support Personnel. ...are responsible for supervision and performance of technical support functions for the operating organization. Personnel involved in surveillance, testing, analyzing plant data, planning modifications, program review, and technical problem resolution in their area of expertise are also included. They have expertise in mechanical, electrical, instrumentation and control, chemistry, radiation protection, training, safety, or quality assurance. For personnel assigned to equivalent positions, non-reactor nuclear facilities should use the education and experience requirements contained in Chapter II, Category A Reactor Personnel, subparagraph 2c. For positions for which there is no equivalent, the education and experience requirements are as listed below. Education and experience requirements are intended to apply to supervisory positions or positions with authority to review and concur, and not to entry-level positions.</p> <p>(1) Education: Baccalaureate in engineering or related science</p> <p>(2) Experience:</p> <table> <tr> <td>Job related</td> <td>2 years</td> </tr> <tr> <td>Nuclear</td> <td>1 year</td> </tr> </table>				Job related	2 years	Nuclear	1 year
Job related	2 years						
Nuclear	1 year						

1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): NO

2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):

Specific entry-level requirements for personnel are not addressed in the HASQAP.

3.0 CORRECTIVE ACTION:

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch.IV.3.	Mandatory (Y/N):	Y
<p>Statement:</p> <p>NON-REACTOR NUCLEAR FACILITY PERSONNEL. MEDICAL EXAMINATION REQUIREMENTS. For each type of operation, the operating organization shall determine the physical demands imposed upon certified personnel by the job tasks that are required to perform both routine and emergency functions. A medical examination shall be given to prospective employees and a reexamination shall be given at least every two years to certified operators, fissionable material handlers, and supervisors to verify health and physical fitness to safely perform their assigned tasks. Certified operators, fissionable material handlers, and supervisors must also be cleared by medical examination prior to returning to work following any serious illness or injury which keeps the person from performing their duties for a period exceeding one month. Medical examination requirements for other operating organization personnel shall be in accordance with the procedures of the operating contractor.</p>			

N/A

Statement Number:	Ch.IV.4.	Mandatory (Y/N):	Y
<p>Statement:</p> <p>NON-REACTOR NUCLEAR FACILITY PERSONNEL. STAFFING REQUIREMENTS. The operating organization shall ensure that minimum staffing requirements are met for safe and reliable operations. Any temporary deviation from these requirements must be justified by facility-specific analysis and approved by the Head of the Field Organization.</p>			

N/A

Statement Number:	Ch.IV.4.a.	Mandatory (Y/N):	Y
<p>Statement:</p> <p>NON-REACTOR NUCLEAR FACILITY PERSONNEL. STAFFING REQUIREMENTS. Overall staffing shall be based on the facility Safety Analysis Report and Operational Safety Requirements. Facility operation shall be contingent upon meeting the criteria contained in these documents.</p>			

N/A

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch.IV.4.b.	Mandatory (Y/N):	Y
<p>Statement:</p> <p>NON-REACTOR NUCLEAR FACILITY PERSONNEL. STAFFING REQUIREMENTS. A list of facility personnel by name, title, and work and home telephone numbers shall be readily available in the control room or other appropriate central location within the facility for use by the operating crew. The list shall include:</p> <ul style="list-style-type: none">(1) Management personnel;(2) Radiation safety personnel; and(3) Technical support personnel.			

- 1.0 STATEMENT SATISFIED BASED ON IMPLEMENTING EVIDENCE (Y/N): NO
- 2.0 GENERAL DESCRIPTION OF EVIDENCE (DESCRIPTION OF NONCOMPLIANCE OR DEFICIENCY IF APPLICABLE):
- Statement is not addressed in the HASQAP.
- 3.0 CORRECTIVE ACTION:

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch.IV.5.a.	Mandatory (Y/N):	Y
<p>Statement:</p> <p>NON-REACTOR NUCLEAR FACILITY PERSONNEL. SPECIFIC TRAINING REQUIREMENTS. Operators. Individual operator training shall be sufficiently comprehensive to cover areas which are fundamental to the candidate's assigned tasks.</p> <p>(1) A core of subjects such as industrial safety, instrumentation and control, basic physics, chemistry, industry operating experience, and major facility systems shall be established. The topics within these subject areas should be selected based on the results of a systematic evaluation of training needs such as would result from a job analysis.</p> <p>(2) On-the-job and classroom-type training shall be included to assure that personnel are familiar with all aspects of their positions. Such training shall include but not be limited to:</p> <ul style="list-style-type: none"> (a) Normal procedures; (b) Emergency actions; (c) Radiation control practices; (d) Location and functions of pertinent safety systems and equipment; (e) Configuration control procedures; (f) Procedures for making changes or alterations in operations; and (g) Operational Safety Requirements. 			

N/A

Statement Number:	Ch.IV.5.b.	Mandatory (Y/N):	Y
<p>Statement:</p> <p>NON-REACTOR NUCLEAR FACILITY PERSONNEL. SPECIFIC TRAINING REQUIREMENTS. Fissionable Materials Handlers. ...candidates shall be trained in the following subject areas in addition to that required in subparagraph 5a above to the extent applicable to the position:</p> <p>(1) Instrumentation and control, including types of instruments and control systems, principles of operation, and consequences of malfunctions;</p> <p>(2) Facility operating characteristics, including principal features, operating parameters, and operating limits of the facility (to include auxiliary systems); and</p> <p>(3) Principles of nuclear facility operation, including the processes involved and technical terminology for the chemical, physical, and metallurgical reactions and criticality safety principles, controls, and specifications.</p>			

N/A

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch.IV.5.c.	Mandatory (Y/N):	Y
<p>Statement:</p> <p>NON-REACTOR NUCLEAR FACILITY PERSONNEL. SPECIFIC TRAINING REQUIREMENTS. Supervisors. The supervisor training program shall include the categories and on-the-job training specified for operators and fissionable material handlers to the extent to which they are applicable. This training shall be of increased depth to reflect the added responsibility of the supervisor position. In addition, the supervisor training program shall include the following to the extent that it is applicable to job requirements:</p> <ol style="list-style-type: none"> (1) Design, control, and operating limitations for the facility, including instrumentation characteristics and adjustment, facility operation, facility console control mechanisms, and control room manipulations; (2) Procedures for making design and operating changes, including changes in operating procedures; (3) Radiation hazards which may arise during the performance of experiments; (4) Nuclear and radiation theory, including details of the fission process, neutron multiplication, source effects, and neutron poison effects; (5) Specific operating characteristics of the facility, such as causes and effects of temperature and pressure changes; (6) Procedures and limitations involved in initial equipment loading, alterations in fissionable material configuration, and determination of various internal and external effects on criticality safety; (7) Procedures, equipment, and facilities available for handling and disposing of radioactive materials and effluents; (8) Functions, assignments, and responsibilities of the maintenance organization as related to facility safety; and (9) Applicable portions of the facility Safety Analysis Report. 			

N/A

ASSESSMENT DATA COLLECTION FORM

Statement Number:	Ch.IV.6.	Mandatory (Y/N):	Y
<p>Statement:</p> <p>NON-REACTOR NUCLEAR FACILITY PERSONNEL. OPERATOR, FISSIONABLE MATERIALS HANDLER, AND SUPERVISOR PROFICIENCY REQUIREMENTS. In order to maintain proficiency, fissionable material handlers, fissionable material handler supervisors, and any other operators and/or supervisors identified by the operating organization to be certified, must actively participate in the job functions associated with their certification. The operating organization shall establish procedures which define requirements and frequency... necessary to maintain an active status. If established proficiency requirements are not met, certification shall be suspended and the person shall not be assigned certification duties until certification is regained. Certification may be regained by performing their duties under the direct supervision of a certified person in accordance with established procedures. If the facility is infrequently operated, certification shall be reinstated prior to facility operation by administering written and oral examinations to ensure adequate operational knowledge.</p>			

N/A

DISTRIBUTION

Number of copiesONSITE

8	<u>U.S. Department of Energy- Richland Operations Office</u>
	R. P. Carter S7-55
	C. E. Clark A5-15
	J. W. Day A5-19
	L. P. Jackson S7-55
	C. K. Kasch A5-55
	K. K. Kawabata S7-55
	T. M. Stram S7-55
	Public Reading Room A1-65
2	<u>Bechtel Hanford, Inc.</u>
	J. H. Kessner H4-23
	C. Stacey H4-23
1	<u>Hanford Environmental Health Foundation</u>
	J. K. Samuels B1-61
4	<u>Pacific Northwest Laboratory</u>
	A. G. King P7-22
	K. J. Kuhl-Klinger P7-27
	J. M. Latkovich P7-27
	C. D. Taylor B1-40
14	<u>Westinghouse Hanford Company</u>
	M. L. Bell T6-16
	J. L. Deichman H4-19
	D. G. Farwick H4-16
	R. R. Grabbe S3-28
	V. W. Hall H4-19
	E. J. Kosiancic T6-16
	R. P. Marshall T6-14
	H. K. Meznarich T6-16
	S. L. Moist R3-56
	P. M. Morant H4-25
	C. T. Narquis T6-50
	L. H. Taylor H4-25
	Central Files L8-04
	O.S.T.I. (2) L8-07