

**DISTRIBUTION SHEET**

To DISTRIBUTION	From ANALYTICAL SERVICES	Page 1 of 2	
		Date:	11/17/94
Project Title/Work Order WHC-SD-WM-DP-076, Rev. 0 45-Day Deliverable for Tank 241-BX-105 Auger Samples, Risers 2 and 6		EDT NO.:	EDT-140744
		ECN NO.:	N/A

Name	MSIN	Text With all Attach	Addendum 1A Only	EDT/ECN Only
------	------	-------------------------	---------------------	-----------------

ONSITE

Pacific Northwest Laboratory

S. F. Bobrowski	K7-28			X
S. J. Harris	K7-22			X
P. G. Heasler	K7-34			X
S. G. McKinley	P7-22			X
J. L. Scott	K1-22			X

U.S. Department of Energy, RL

J. M. Clark	S7-54			X
J. Noble-Dial	S7-54			X

Westinghouse Hanford Company

J. N. Appel	G3-21			X
H. Babad	S7-30			X
K. E. Bell	T6-06			X
D. R. Bratzel	S7-31			X
R. J. Cash	R2-78			X
C. DeFig-Price	X3-71			X
J. L. Deichman	H4-19			X
S. J. Eberlein	S7-31			X
C. E. Golberg	H5-49			X
J. M. Henderson	S6-12			X
D. C. Hetzer	S6-30			X
J. G. Hill	K7-97	X		
L. Jensen	T6-07	X		
T. J. Kelly	S7-30			X
N. W. Kirch	R2-11			X
J. G. Kristofzski	T6-06			X
P. M. Morant	H4-19			X
A. F. Noonan	B1-40	X		
A. D. Rice	T6-06			X
P. Sathyanarayana	R2-12			X
R. D. Schreiber	R2-12			X
B. C. Simpson	R2-12			X
M. J. Sutey	T4-07			X
D. A. Turner	R2-78			X
J. A. Voogd	R4-01			X
O. S. Wang	R2-78			X
E. T. Weber	G6-08			X
W. I. Winters	T6-50			X
Central Files	L8-04	2		
EDMC	H6-08	X		
LTIC	T6-06			X
OSTI	L8-07	2		

## **DISCLAIMER**

**Portions of this document may be illegible in electronic image products. Images are produced from the best available original document.**

**DISTRIBUTION SHEET CONTINUED**

To DISTRIBUTION	From ANALYTICAL SERVICES	Page 2 of 2
		Date: 11/17/94
Project Title/Work Order WHC-SD-WM-DP-076, Rev. 0 45-Day Deliverable for Tank 241-BX-105 Auger Samples, Risers 2 and 6		EDT NO.: EDT-140744
		ECN NO.: N/A

Name	MSIN	Text With all Attach	Addendum 1A Only	EDT/ECN Only
<u>Westinghouse Hanford Company</u> TFIC (Tank Farm Information Center)	R1-20			X

OFFSITE

<u>Washington State Department of Ecology</u> Single-Shell Tank Unit Manager S. E. McKinney P.O. Box 47600 Olympia, Washington 98504-7600				X
<u>Environmental Protection Agency</u> Single-Shell Tank Unit Manager D. R. Sherwood 712 Swift Boulevard, Suite 5 Richland, Washington 99352				X

NOV 22 1994

55 Sta.S

ENGINEERING DATA TRANSMITTAL

Page 1 of 1

1. EDT 140744

2. To: (Receiving Organization) Distribution	3. From: (Originating Organization) Program Support	4. Related EDT No.: N/A
5. Proj./Prog./Dept./Div.: Tank 241-BX-105/Waste Management/PS/AS	6. Cog. Engr.: Kevin E. Bell	7. Purchase Order No.: N/A
8. Originator Remarks: This document is being released into the Supporting Document System for retrievability purposes.		9. Equip./Component No.: N/A
11. Receiver Remarks: For Release.		10. System/Bldg./Facility: N/A
		12. Major Assm. Dwg. No.: N/A
		13. Permit/Permit Application No.: N/A
		14. Required Response Date: 11/18/94

15. DATA TRANSMITTED					(F)	(G)	(H)	(I)
(A) Item No.	(B) Document/Drawing No.	(C) Sheet No.	(D) Rev. No.	(E) Title or Description of Data Transmitted	Approval Designator	Reason for Transmittal	Originator Disposition	Receiver Disposition
1	WHC-SD-WM-DP-076	N/A	0	45-Day Deliverable for Tank 241-BX-105 Auger Samples, Risers 2 and 6	Q	2	1	

16. KEY					
Approval Designator (F)		Reason for Transmittal (G)		Disposition (H) & (I)	
E, S, Q, D or N/A (see WHC-CM-3-5, Sec.12.7)		1. Approval	4. Review	1. Approved	4. Reviewed no/comment
		2. Release	5. Post-Review	2. Approved w/comment	5. Reviewed w/comment
		3. Information	6. Dist. (Receipt Acknow. Required)	3. Disapproved w/comment	6. Receipt acknowledged

17. SIGNATURE/DISTRIBUTION (See Approval Designator for required signatures)											
(G)	(H)	(J) Name	(K) Signature	(L) Date	(M) MSIN	(J) Name	(K) Signature	(L) Date	(M) MSIN	Reason	Disp.
2	1	Cog.Eng. K. E. Bell	<i>Kevin Bell</i>	11/21/94		TL-04					
2	1	Cog. Mgr. J. G. Kristofzski	<i>J.G. Kristofzski</i>	11/22/94		TL-04					
2	1	QA J. C. Langford	<i>J.C. Langford</i>	11/21/94		TL-03					
		Safety									
		Env.									

18. A.E. Young <i>A.E. Young</i> 11-18-94 Signature of EDT Date Originator	19. _____ Authorized Representative Date for Receiving Organization	20. <i>J.G. Kristofzski</i> 11/22/94 Cognizant Manager Date	21. DOE APPROVAL (if required) Ctrl. No. <input type="checkbox"/> Approved <input type="checkbox"/> Approved w/comments <input type="checkbox"/> Disapproved w/comments
---	---	--	---



## RELEASE AUTHORIZATION

**Document Number:** WHC-SD-WM-DP-076, REV.0

**Document Title:** 45-Day Deliverable for Tank 241-BX-105 Auger Samples, Risers 2 and 6

**Release Date:** November 22, 1994

**This document was reviewed following the procedures described in WHC-CM-3-4 and is:**

**APPROVED FOR PUBLIC RELEASE**

**WHC Information Release Administration Specialist:**

By Jamara Kuhn, for  
C. Willingham

November 22, 1994

**TRADEMARK DISCLAIMER.** Reference herein to any specific commercial product, process, or service by trade name, trademark, manufacturer, or otherwise, does not necessarily constitute or imply its endorsement, recommendation, or favoring by the United States Government or any agency thereof or its contractors or subcontractors.

This report has been reproduced from the best available copy. Available in paper copy and microfiche. Printed in the United States of America. Available to the U.S. Department of Energy and its contractors from:

U.S. Department of Energy  
Office of Scientific and Technical Information (OSTI)  
P.O. Box 62  
Oak Ridge, TN 37831  
Telephone: (615) 576-8401

Available to the public from:

U.S. Department of Commerce  
National Technical Information Service (NTIS)  
5285 Port Royal Road  
Springfield, VA 22161  
Telephone: (703) 487-4650

**SUPPORTING DOCUMENT**

1. Total Pages |||

2. Title

45-Day Deliverable for Tank 241-BX-105 Auger Samples, Risers 2 and 6

3. Number

WHC-SD-WM-DP-076

4. Rev No.

0

5. Key Words

45-Day Deliverable, Tank 241-BX-105, BX-105, Auger Samples, Riser 2, Riser 6

6. Author

Name: Kevin E. Bell

*Kevin E. Bell*  
Signature

Organization/Charge Code 8E480/MDR2D

CW 11/22/94  
**APPROVED FOR PUBLIC RELEASE**

7. Abstract

N/A

8. RELEASE STAMP

NOV 22 1994  
55  
SLA.5



WHC-SD-WM-DP-076, REV. 0

**ANALYTICAL SERVICES**

**Project:**

**SINGLE SHELL TANK  
WASTE CHARACTERIZATION  
45-DAY SAFETY SCREENING FOR TANK  
241-BX-105 AUGER SAMPLES, RISERS 2 & 6**

**Date Printed:**

**NOVEMBER 16, 1994**

**DISCLAIMER**

This report was prepared as an account of work sponsored by an agency of the United States Government. Neither the United States Government nor any agency thereof, nor any of their employees, makes any warranty, express or implied, or assumes any legal liability or responsibility for the accuracy, completeness, or usefulness of any information, apparatus, product, or process disclosed, or represents that its use would not infringe privately owned rights. Reference herein to any specific commercial product, process, or service by trade name, trademark, manufacturer, or otherwise does not necessarily constitute or imply its endorsement, recommendation, or favoring by the United States Government or any agency thereof. The views and opinions of authors expressed herein do not necessarily state or reflect those of the United States Government or any agency thereof.

**MASTER**

**DISTRIBUTION OF THIS DOCUMENT IS UNLIMITED**

TABLE OF CONTENTS

Narrative . . . . .	1
Sample Data Summary . . . . .	7
BX-105 Immediate Notification Documentation . . . . .	10
Sampling and Custody Data . . . . .	14
Undigested Sample Analyses-Direct . . . . .	22
Differential Scanning Calorimetry	
WL# 109 (DSC) . . . . .	23
WL# 111 (DSC) . . . . .	29
WL# 112 (DSC) . . . . .	35
WL# 110 (DSC) . . . . .	39
Thermal Gravimetric Analyses	
WL# 106 (TGA) . . . . .	43
WL# 157 (TGA) . . . . .	49
WL# 104 (TGA) . . . . .	57
WL# 105 (TGA) . . . . .	63
Fusion Digestion Analyses . . . . .	68
Total Alpha/Total Beta Analyses	
WL# 139 . . . . .	69
WL# 114 . . . . .	83

This document consists of pages 1 through 107, plus page 6.1 and 63.1.

**WHC-SD-WM-DP-076, REV. 0**

**NARRATIVE**

From: Program Support 9457804  
Phone: 373-1604 T6-06  
Date: November 16, 1994  
Subject: 45-DAY DELIVERABLE FOR TANK 241-BX-105 AUGER SAMPLES; RISERS 2 & 6

TO: D. R. Bratzel R2-18

cc: H. Babad S7-31  
C. Defigh-Price X3-71  
J. L. Deichman H4-19  
N. W. Kirch R2-11  
R. D. Schrieber R2-12  
JGK File/LB

References: (1) WHC-SD-WM-TP-239, Rev. 0, "Tank 241-BX-105, Tank Characterization Plan," dated August 11, 1994, Westinghouse Hanford Company, Richland, WA 99352

This is the final report for the fiscal year 1995 BX-105 auger sample characterization effort. Included are copies of the differential scanning calorimetry (DSC) and thermogravimetric analysis (TGA) scans as requested in Reference 1. Also included is a copy of any immediate notification documentation, chain of custody forms, the hot cell work plan, extruded segment [auger] description sheets, and total alpha data.

#### BX-105 Analytical Summary

Two auger samples from single-shell tank 241-BX-105 (BX-105) were extruded, broken down, and analyzed for DSC, TGA, and total alpha as prescribed in Reference 1 except where noted below. Analytical results were tracked and reported using the laboratory information management system known as LabCore. In this system, each sample is assigned a unique sample number.

The TGA percent moisture results are below the safety criteria limit of 17% in samples from risers 2 and 6. Verbal and written notifications were made as prescribed in Reference 1. There are no DSC exotherms associated with any of the samples. Total alpha results are from 69 to 423 times below the action limit of 41 uCi/g. In some cases, the tank characterization plan (TCP) accuracy and precision criteria are not met. If a re-run was not performed when a TCP quality control limit was not met, reasons for not performing the re-run are provided.

Sample Receipt and Extrusion

BX-105 Riser 2

The 20 inch auger sample 94-AUG-008 was removed from riser 2 of tank BX-105 on 9/30/94 at 1230 hours. The sample was shipped to the 222-S Laboratory on 10/3/94 at 1045 hours and received at the laboratory on 10/3/94 at 1140 hours. Loading and extrusion of the sample in the hot cell took place on 10/6/94. No problems were noted in extruding the auger sample.

A total of 60.00 g of sample was collected. Sample material was present on all flutes of the auger, ranging in color from medium brown on flutes 1 through 7, a mixture of brown and off-white material on flutes 7 through 9, and gray and white material on flutes 9 through 19. There was very little material on flutes 1 through 7. A 20 inch auger has 19 flutes. The flutes are numbered such that flute 1 begins at the auger shaft and flute 19 ends at the tip of the auger bit. The spaces between the flutes were not filled, rather the sample adhered to the flutes and auger shaft. Although no drainable liquid was present, some creamy, grey, mud-like material fell onto the auger tray during extrusion and some remained on the tip of the auger. No liner liquid was present, but a portion of the liner was coated with what appeared to be the same material that fell onto the extrusion tray. The visible moisture content of this auger sample was variable. The tray material was runny while other material appeared dry and would flake off the auger when being subsampled.

Subsamples of this auger sample were taken as follows. Sample S94T000139 (139) for DSC/TGA analysis with a mass of 4.90 g was removed from flutes 15 and 16 approximately five minutes after extrusion. This subsample was grayish to white in color and appeared moist with a paste-like consistency. Some flaking occurred during removal. A second subsample (15.10 g) was collected for DSC/TGA analysis from the soft, brown, mud-like material on the auger tray. Some material for archiving was removed from the vial containing this subsample. The vial was later broken during sample loadout from the hotcell. Because the amount of time between breakage and discovery of the break was unknown, it was decided that the remaining sample had been compromised for its intended purpose and it was discarded. Two subsamples for safety screening were taken from the upper half (flutes 1-9) and lower half (flutes 9-19) of the auger.

The subsamples taken originally and described above are not necessarily the samples used for analysis. For ALARA reasons, smaller aliquots are usually loaded out of the hot cell. See Table 1 for additional subsampling information. Columns 3 and 4 of Table 1 respectively give the vial or jar number, in which the subsample was originally placed, and the weight of that subsample. If a smaller aliquot was subsequently removed from the original jar or vial, the identity of the new container and the aliquot weight are recorded in columns 5 and 6.

BX-105 Riser 6

The 20 inch auger sample 94-AUG-009 was removed from riser 6 of tank BX-105 on 10/5/94 at 1100 hours. The sample was shipped to the 222-S Laboratory on 10/6/94 at 1015 hours and received at the laboratory on 10/6/94 at 1130 hours. Loading and extrusion of the sample in the hot cell took place on 10/7/94. No problems were noted in extruding the auger sample.

A total of 319.25 g of sample was collected, the majority of which was grayish-white, thick, pasty material on flutes 11-15, although all flutes of the auger contained some sample. Approximately 5 mL of liner liquid was present and was not retained. Also, a small amount of material fell onto the auger tray and was not retained. The description of material on flutes 1-11 and 16-19 matches that of material on flutes 11-15 except that the material on flutes 11-15 was crusty. This may have been due to sample drying in the hot cell.

The following subsamples were taken from this auger. Sample 146 (5.81 g) for DSC/TGA analysis was removed from flutes 11-15 approximately five minutes after extrusion. The two subsamples for the safety screening analysis were removed from the upper half (flutes 1-10) and lower half (flutes 11-19) of the auger. See Table 1 for additional subsampling information.

Analytical Results

The safety screening analytical results are presented in Table 2. The table includes the LabCore sample number, which may be cross-referenced with Table 1. It also includes the upper or lower action limits as defined in the TCP. The limit selected for immediate notification is highlighted in greybar. A lower density greybar is used to indicate where action limits were exceeded. Column 2 of Table 2 indicates the sample preparation used, if any. As shown, total alpha analyses are marked with "F" indicating a fusion preparation was performed on the sample before analysis.

TGA (Moisture Content)

Weight percent water by TGA was performed under a nitrogen atmosphere using procedure LA-560-112, Rev. A-1. Analytical results are below the notification limit of 17% for all subsamples in riser 2, and the upper half and lower half subsamples of riser 6. Copies of the immediate notifications are reproduced in a following attachment. Sample 139, removed for TGA analysis shortly after extrusion, has an average moisture content 3.9 percentage points lower than sample 143, which was also taken from the lower half of the auger 30 to 45 minutes later. The TGA scan for sample 139 was integrated to approximately 90° C, while the 143 sample was integrated to approximately 200° C, therefore the percent moisture for the 139 sample is probably biased low relative to sample 143. Nevertheless, there does not appear to be an extreme amount of sample drying during extrusion. Results



for samples from riser 6 also support this conclusion. Sample 148 from the lower half of the auger contained an average of 15.48% water, and sample 146 taken 5 minutes after extrusion from the same location contained an average of 18.87% water.

Standard recoveries are very good ranging from 97.4% to 99.5%. The duplicate precision acceptance criteria of 10% was achieved with samples 146 and 148, but was not achieved with the remaining samples. Sample 143 was re-run with some improvement in precision. The sample was not re-run a second time. The other samples were not re-run because both sample and duplicate results are below the notification limit and there was reason to believe the disparities are most likely due to sample heterogeneity.

#### DSC (Energetics Content)


Analyses for DSC were performed under a nitrogen atmosphere using procedure LA-514-113, Rev B-1. No exotherms are observed for any of the samples run, therefore no exotherms are calculated on a dry weight basis. Because the DSC action limit is associated with the calculated value, Table 2 does not show the action limit for the DSC analyses. Samples 139 and 146 were run up to 450° C and not 500° C as mandated by the TCP because the samples appeared similar to those from tank SY-103 that damaged the DSC sensor. No exotherms are expected beyond 450° C since the baseline was decreasing prior to reaching 450° C and none of the other samples from BX-105 displayed exotherms.

#### Total Alpha

Analyses for total alpha were performed on an alpha proportional counter according to procedure LA-508-101, Rev. D-2. Standard recoveries are very good, demonstrating that the instrument is in control and functioning properly. All spike recoveries are outside the 90-110% range however. No spike was performed on sample 150 since Reference 1 requires a spike once per matrix and samples 149 and 150 appear to be of the same matrix. Re-runs were performed on samples 142 and 144. The spike recovery improved for 142, but worsened for 144. However, the total alpha values are very similar between the two runs. The TCP limit of  $\pm 10\%$  for relative percent difference (RPD) on duplicate runs was not met for samples 144 and 150. No further re-runs were performed for two reasons: 1) The chemist noted that the poor spike recoveries are due to a high amount of dissolved solids. This could also effect the RPD on duplicate runs. When the fused sample is dried on the planchet, a relatively large amount of solids will attenuate the alpha

detection by self-absorption. 2) The highest total alpha result found was still 69 times below the action limit.

Sincerely,



Kevin E. Bell,  
BX-105 Project Coordinator

keb

- Attachments
1. Summary Data Report (3 pages)
  2. Immediate Notification Documentation (4 pages)
  3. Sampling and Custody Data (8 pages)
  4. Analyses for DSC, TGA, and Total Alpha (86 pages)

Table 1. Subsampling and Sample Load-Out Information

Riser	Subsample Location	Original Vial or Jar	wt (g)	Transferred to vial or jar	Load-out wt (g)	LabCore Sample #
2	flutes 15, 16	vial 6171	4.90	no transfer	4.90	S94T000139
2	tray material	vial 6175	15.10	vial 6175 <sup>1</sup>	7.0	S94T000140
2	flutes 1-9	vial 6173	2.33	no transfer	2.33	S94T000141/142
2	flutes 9-19	jar 6109	37.67	vial 6178	5.78	S94T000143/144
2	archive	jar 6109 and vial 6175	<sup>2</sup>	vial 6168	26.42	S94T000151
6	flutes 11-15	vial 6170	5.81	no transfer	5.81	S94T000146
6	flutes 11-19	jar 6119	260.85	vial 6177	7.70	S94T000148/150
6	flutes 1-10	jar 6124	52.59	vial 6176	5.82	S94T000147/149
6	archive	jar 6119 and jar 6124 <sup>3</sup>	50.60 +11.21	vial 7019	61.71	S94T000152

<sup>1</sup> Vial 6175 was broken while being loaded out of the hot cell. However, 7.10 g of sample had been transferred into vial 6109 before the breakage.

<sup>2</sup> Following removal of an aliquot for analysis from vial 6109, 7.10 g of sample from vial 6175 was added and the material homogenized before being transferred into vial 6168.

<sup>3</sup> Added to vial 7019 was 50.60 g from jar 6119 and 11.21 g from jar 6124.

61

**WHC-SD-WM-DP-076, REV. 0**

**SAMPLE DATA SUMMARY**

14-nov-1994 07:53:10

A-0002-0

Table 2. Analytical Summary Table for BX-105 Auger Samples.  
BX-105

CORE NUMBER: 94-AUG-008

SEGMENT #: 1

SEGMENT PORTION: Immediate Sampling (to check moisture loss)

Sample#	R	A#	Analyte	Unit	Action Limits		Standard/%	Prep Blk	Result	Duplicate	Average	RPD/%	Spk Rec/%	Det Limit	Count	Err/%
					Lower	Upper										
S94T000139			% Water by TGA using Mettler	%	17.000	n/a	99.50	n/a	9.86	11.53	10.70	15.6	n/a	0.010		n/a
S94T000139			DSC Exotherm using Mettler	Joules/g	n/a	n/a	100.9	n/a	0	0	0.000	n/a	n/a	0.000		n/a

SEGMENT PORTION: L Lower Half of Segment

Sample#	R	A#	Analyte	Unit	Action Limits		Standard/%	Prep Blk	Result	Duplicate	Average	RPD/%	Spk Rec/%	Det Limit	Count	Err/%
					Lower	Upper										
S94T000143			% Water by TGA using Mettler	%	17.000	n/a	98.87	n/a	15.74	13.43	14.59	15.8	n/a	0.010		n/a
S94T000143			DSC Exotherm using Mettler	Joules/g	n/a	n/a	100.2	n/a	0	0	0.000	n/a	n/a	0.000		n/a
S94T000144		F	Alpha of Digested Solid	uCi/g	n/a	41.000	97.90	<4.00e-03	3.37e-1	2.03e-1	2.70e-01	49.6	70.10	0.007		9.0

SEGMENT PORTION: U Upper Half of Segment

Sample#	R	A#	Analyte	Unit	Action Limits		Standard/%	Prep Blk	Result	Duplicate	Average	RPD/%	Spk Rec/%	Det Limit	Count	Err/%
					Lower	Upper										
S94T000141			% Water by TGA using Mettler	%	17.000	n/a	98.23	n/a	13.76	7.24	10.50	62.1	n/a	0.010		n/a
S94T000141			DSC Exotherm using Mettler	Joules/g	n/a	n/a	100.9	n/a	0	0	0.000	n/a	n/a	0.000		n/a
S94T000142		F	Alpha of Digested Solid	uCi/g	n/a	41.000	97.90	<4.00e-03	5.96e-1	5.83e-1	5.89e-01	2.21	77.80	0.007		6.8

=> Limit violated

=> Selected Limit

WHC-SD-WM-DP-076, REV. 0

Table 2. Analytical Summary Table for BX-105 Auger Samples.  
BX-105

CORE NUMBER: 94-AUG-009

SEGMENT #: 1

SEGMENT PORTION: Immediate Sampling (to check moisture loss)

Sample#	R A# Analyte	Unit	Action Limits		Standard/%	Prep Blk	Result	Duplicate	Average	RPD/%	Spk Rec/%	Det Limit	Count	Err/%
			Lower	Upper										
S94T000146	% Water by TGA using Mettler	%	17.000	n/a	99.50	n/a	18.97	18.77	18.87	1.06	n/a	0.010		n/a
S94T000146	DSC Exotherm using Mettler	Joules/g	n/a	n/a	100.9	n/a	0	0	0.000	n/a	n/a	0.000		n/a

SEGMENT PORTION: L Lower Half of Segment

Sample#	R A# Analyte	Unit	Action Limits		Standard/%	Prep Blk	Result	Duplicate	Average	RPD/%	Spk Rec/%	Det Limit	Count	Err/%
			Lower	Upper										
S94T000148	% Water by TGA using Mettler	%	17.000	n/a	99.00	n/a	16.18	14.78	15.48	9.04	n/a	0.010		n/a
S94T000148	DSC Exotherm using Mettler	Joules/g	n/a	n/a	104.0	n/a	0	0	0.000	n/a	n/a	0.000		n/a
S94T000150	F Alpha of Digested Solid	uCi/g	n/a	41.000	96.33	<2.00e-03	1.32e-2	9.69e-3	1.10e-02	30.7	n/a	0.004		34.6

SEGMENT PORTION: U Upper Half of Segment

Sample#	R A# Analyte	Unit	Action Limits		Standard/%	Prep Blk	Result	Duplicate	Average	RPD/%	Spk Rec/%	Det Limit	Count	Err/%
			Lower	Upper										
S94T000147	% Water by TGA using Mettler	%	17.000	n/a	97.40	n/a	4.91	5.54	5.225	12.1	n/a	0.010		n/a
S94T000147	DSC Exotherm using Mettler	Joules/g	n/a	n/a	100.2	n/a	0	0	0.000	n/a	n/a	0.000		n/a
S94T000149	F Alpha of Digested Solid	uCi/g	n/a	41.000	96.33	<2.00e-03	1.15e-2	1.17e-2	1.20e-02	1.72	84.10	0.005		42.4

=> Limit violated

=> Selected Limit

**WHC-SD-WM-DP-076, REV. 0**

**BX-105 IMMEDIATE NOTIFICATION  
DOCUMENTATION**

WHC-SD-WM-DP-076, REV. 0

[32] From: Kevin E Bell at ~WHC225 11/9/94 4:46PM (1918 bytes: 23 ln)  
To: Cherri DeFigh-Price at ~WHC79, David R Bratzel at ~WHC268, Harry Babad at  
~WHC142, B C (Brad) Carpenter at ~WHC140, Ruth D Schreiber at ~WHC163,  
John L Deichman at ~WHC321, Nicholas W (Nick) Kirch at ~WHC140  
cc: Kevin E Bell, John G Kristofzski at ~WHC168, Don B Hardy at ~WHC28,  
Susan J Eberlein at ~WHC163, Andrew D Rice at ~WHC168  
Subject: SAFETY LIMIT EXCEEDED ON ~~C-103~~ BX-105 *N/S 11/9/94*

----- Message Contents -----

As required by the BX-105 TCP (WHC-SD-WM-TP-239, Rev. 0), an immediate notification was made at 1130 hours on 11/9/94 by D. Hardy, the 222-S on-duty Shift Manager, to S. Waltari, the East Tank Farms Shift Manager, regarding thermal gravimetric analysis (TGA) measurements indicating <17% water in auger samples from tank BX-105. Also as required, a phone call was made to S. Eberlein of the Characterization Program at 1338 hours on 11/9/94 relaying the information below. This cc:Mail is the required follow-up written notification of the initial, verbal notifications.

- 1) Tank BX-105, Riser 2, Upper half of auger: %water by TGA yields 13.76% for sample and 7.24% for duplicate sample.
- 2) Tank BX-105, Riser 6, Upper half of auger: %water by TGA yields 4.91% for sample and 5.54% for duplicate sample.
- 3) Tank BX-105, Riser 6, Lower half of auger: %water by TGA yields 16.18% for sample and 14.78% for duplicate sample.

Please address any questions to K. Bell @ 373-1629.



[21] From: Kevin E Bell at ~WHC225 11/10/94 1:06PM (2071 bytes: 27 ln)  
To: Harry Babad at ~WHC142, David R Bratzel at ~WHC268, Cherri DeFigh-Price at ~WHC79, John L Deichman at ~WHC321, Vernon W Hall at ~WHC321, Thomas J (Tom) Kelley at ~WHC396, Nicholas W (Nick) Kirch at ~WHC140, John G Kristofzski at ~WHC168, Andrew D Rice at ~WHC168, Ruth D Schreiber at ~WHC163

cc: Kevin E Bell

WHC-SD-WM-DP-076, REV. 0

Subject: SAFETY LIMIT EXCEEDED ON BX-105

----- Message Contents -----

As required by the BX-105 TCP (WHC-SD-WM-TP-239, Rev. 0), an immediate notification was made at 850 hours on 11/10/94 by C. Clark, the 222-S on-duty Shift Manager, to the East Tank Farms Shift Manager, regarding thermal gravimetric analysis (TGA) measurements indicating <17% water in an auger sample from tank BX-105. Also as required, a phone call was made to D. Bratzel of the Characterization Program at 1128 hours on 11/10/94 relaying the information below. This cc:Mail is the required follow-up written notification of the initial, verbal notifications.

1) Tank BX-105, Riser 2, Immediate Sample to Check Moisture Loss: %water by TGA yields 9.86% for sample and 11.53% for duplicate sample.

This sample was taken from flutes 15 and 16 approximately 5 minutes after extrusion. There was no exotherm by DSC associated with this sample. The auger sample contained no drainable liquid and the sample was pasty rather than crystalline.

Note that the subject of yesterday's immediate written notification for BX-105 was incorrectly titled for C-103.

Please address any questions to K. Bell @ 373-1629.

[13] From: Kevin E Bell at ~WHC225 11/14/94 7:16AM (2451 bytes: 33 ln)  
To: Harry Babad at ~WHC142, David R Bratzel at ~WHC268, Cherri DeFigh-Price at  
~WHC79, John L Deichman at ~WHC321, Vernon W Hall at ~WHC321,  
Thomas J (Tom) Kelley at ~WHC396, Nicholas W (Nick) Kirch at ~WHC140,  
John G Kristofzski at ~WHC168, Andrew D Rice at ~WHC168, Ruth D Schreiber at  
~WHC163, B C (Brad) Carpenter at ~WHC140  
cc: Kevin E Bell  
Subject: SAFETY LIMIT EXCEEDED ON BX-105 AND B-102      **WHC-SD-WM-DP-076, REV. 0**

----- Message Contents -----

As required by the BX-105 TCP (WHC-SD-WM-TP-239, Rev. 0) and the B-102 TCP (WHC-SD-WM-TP-226, Rev. 0), an immediate notification was made at 1402 hours on 11/11/94 by C. Clark, the 222-S on-duty Shift Manager, to the East Tank Farms Shift Manager, regarding thermal gravimetric analysis (TGA) measurements indicating <17% water in auger samples from tanks BX-105 and B-102. This cc:Mail is the required follow-up written notification of the initial, verbal notification.

1) Tank BX-105, Riser 2, Lower half of auger: %water by TGA yields 15.74% for sample and 13.43% for duplicate. There was no exotherm by DSC associated with this sample. The auger sample contained no drainable liquid and the sample was pasty rather than crystalline.

2) Tank B-102, Riser 1, Immediate sample to check moisture loss: %water by TGA yields 16.98% for sample. The duplicate result was not below the notification limit. This sample was taken approximately 5 minutes after extrusion. There is an exotherm of 168.4 J/g in the DSC associated with this sample. This exotherm has not yet been calculated on a dry weight basis.

3) Tank B-102, Riser 1, Lower half of auger: %water by TGA yields 13.07% for sample and 15.15 for duplicate. The DSC for this sample has not been reported at the time of writing.

Please address any questions to K. Bell @ 373-1629.

**WHC-SD-WM-DP-076, REV. 0**

**SAMPLING AND CUSTODY DATA**

CHAIN-OF-CUSTODY RECORD FOR AUGER SAMPLING

COPY

(1) Shipment Number 52098 (2) Sample Number 94-AUG-008 (3) Supervisor Bob Praznik  
 (4) Tank BX105 (5) Riser 2 (6) Cask Serial Number C 1029

Radiation Survey Data:		(7) FIELD	(31) LABORATORY	(8) Shipment Description	
Over Top Dose Rate	<u>5 m/hr</u>	<u>5 m/hr</u>	<u>5 m/hr</u>	A. Work Package Number	<u>2E-94-00570-W</u>
Side Dose Rate	<u>7 m/hr</u>	<u>7 m/hr</u>	<u>7 m/hr</u>	B. Cask Seal Number	<u>3679</u>
Bottom Dose Rate	<u>4.5 m/hr</u>	<u>4 m/hr</u>	<u>4 m/hr</u>	C. Date and Time Sample	<u>9-30-94 / 12:30pm</u>
Smearable Contamination	<u>&lt;20</u>	<u>&lt;20</u>	<u>&lt;20</u>	Removed from Tank	<u>20%</u>
	(Alpha)	(Alpha)	(Alpha)	D. Expected Liquid Content	<u>0%</u>
	<u>&lt;1000</u>	<u>&lt;1000</u>	<u>&lt;1000</u>	E. Expected Solid Content	<u>200 m/hr</u>
	(Beta-Gamma)	(Beta-Gamma)	(Beta-Gamma)	F. Dose Rate Through Drill String	<u>20"</u>
RCT*	<u>[Signature]</u>	<u>[Signature]</u>	<u>[Signature]</u>	G. Expected Sample Length	
	(Signature)	(Signature)	(Signature)		
	<u>10-3-94</u>				

(9) INFORMATION (Include statement of laboratory tests to be performed.)

15

(10) Field Comments

(32) Laboratory Comments

WHC-SD-WM-DP-076, REV. 0

(11) Point of Origin <u>BX105 Riser #2</u>	(12) Destination <u>2225 LAB</u>	(13) Sender Name (Sign and PRINT) <u>PRAZNIK [Signature]</u>	(14) Date/Time <u>10-3-94/10:45am</u>	(15) Sender Comments
(17) Relinquished By (Sign and PRINT) <u>[Signature] P A KATSEL</u>	(18) Received By (Sign and PRINT) <u>Cham E. Byrd Charles E. Byrd</u>	(19) Date/Time <u>10-3-94 11:15</u>	(20) Receiver Comments	
(21) Relinquished By (Sign and PRINT) <u>Cham E. Byrd Charles E. Byrd</u>	(22) Received By (Sign and PRINT) <u>[Signature]</u>	(23) Date/Time <u>10-3-94 11:40</u>	(24) Receiver Comments	
(25) Relinquished By (Sign and PRINT)	(26) Received By (Sign and PRINT)	(27) Date/Time	(28) Receiver Comments	

(16) Seal Intact Upon Release? <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No	(29) Seal Intact Upon Receipt? <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No	(30) Seal Data Consistent with this Record?
Shipment No. <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No	Cask Seal No. <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No	Sample No. <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No

### CHAIN-OF-CUSTODY RECORD FOR AUGER SAMPLING

COPY

(1) Shipment Number 200W-08-TF (2) Sample Number 94-AUG-009 (3) Supervisor Bob Praznik  
 (4) Tank BX-105 (5) Riser 6 (6) Cask Serial Number C-1042

<p><b>Radiation Survey Data:</b></p> <table style="width: 100%;"> <tr> <th style="width: 50%;">(7) FIELD</th> <th style="width: 50%;">(31) LABORATORY</th> </tr> <tr> <td>Over Top Dose Rate <u>1.05 mR/hr</u></td> <td><u>1.05 mR/hr</u></td> </tr> <tr> <td>Side Dose Rate <u>9 mR/hr</u></td> <td><u>9 mR/hr</u></td> </tr> <tr> <td>Bottom Dose Rate <u>3.5 mR/hr</u></td> <td><u>3.5 mR/hr</u></td> </tr> <tr> <td>Smearable Contamination (Alpha) <u>520</u></td> <td><u>520</u></td> </tr> <tr> <td>(Beta-Gamma) <u>1900</u></td> <td><u>1900</u></td> </tr> <tr> <td>RCT* <u>[Signature]</u></td> <td>RCT* <u>[Signature]</u></td> </tr> <tr> <td style="text-align: center;">(Signature)</td> <td style="text-align: center;">(Signature)</td> </tr> </table>	(7) FIELD	(31) LABORATORY	Over Top Dose Rate <u>1.05 mR/hr</u>	<u>1.05 mR/hr</u>	Side Dose Rate <u>9 mR/hr</u>	<u>9 mR/hr</u>	Bottom Dose Rate <u>3.5 mR/hr</u>	<u>3.5 mR/hr</u>	Smearable Contamination (Alpha) <u>520</u>	<u>520</u>	(Beta-Gamma) <u>1900</u>	<u>1900</u>	RCT* <u>[Signature]</u>	RCT* <u>[Signature]</u>	(Signature)	(Signature)	<p><b>(8) Shipment Description</b></p> <p>A. Work Package Number <u>2E-94-00570-W</u></p> <p>B. Cask Seal Number <u>3681</u></p> <p>C. Date and Time Sample Removed from Tank <u>10-5-94 / 1100hrs</u></p> <p>D. Expected Liquid Content <u>20%</u></p> <p>E. Expected Solid Content <u>80%</u></p> <p>F. Dose Rate Through Drill String <u>200 mR/h</u></p> <p>G. Expected Sample Length <u>20"</u></p>
(7) FIELD	(31) LABORATORY																
Over Top Dose Rate <u>1.05 mR/hr</u>	<u>1.05 mR/hr</u>																
Side Dose Rate <u>9 mR/hr</u>	<u>9 mR/hr</u>																
Bottom Dose Rate <u>3.5 mR/hr</u>	<u>3.5 mR/hr</u>																
Smearable Contamination (Alpha) <u>520</u>	<u>520</u>																
(Beta-Gamma) <u>1900</u>	<u>1900</u>																
RCT* <u>[Signature]</u>	RCT* <u>[Signature]</u>																
(Signature)	(Signature)																

(9) INFORMATION (Include statement of laboratory tests to be performed.)

16

(10) Field Comments	(32) Laboratory Comments
---------------------	--------------------------

WHC-SD-WM-DP-076, REV. 0

(11) Point of Origin <u>BX105-R6</u>	(12) Destination <u>222S LAB</u>	(13) Sender Name (Sign and PRINT) <u>R.S. PRAZNIK</u>	(14) Date/Time <u>10-6-94 / 10:50</u>	(15) Sender Comments
(17) Relinquished By (Sign and PRINT) <u>R.S. PRAZNIK</u>		(18) Received By (Sign and PRINT) <u>James R. [Signature]</u>		(19) Date/Time <u>10-6-94 / 10:20</u>
(21) Relinquished By (Sign and PRINT) <u>[Signature]</u>		(22) Received By (Sign and PRINT) <u>Sandra Cobb</u>		(20) Receiver Comments
(25) Relinquished By (Sign and PRINT)		(26) Received By (Sign and PRINT)		(23) Date/Time <u>10-26-94 / 1130</u>
(25) Relinquished By (Sign and PRINT)		(26) Received By (Sign and PRINT)		(24) Receiver Comments
(25) Relinquished By (Sign and PRINT)		(26) Received By (Sign and PRINT)		(27) Date/Time
(25) Relinquished By (Sign and PRINT)		(26) Received By (Sign and PRINT)		(28) Receiver Comments

(16) Seal Intact Upon Release? <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No	(29) Seal Intact Upon Receipt? <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No	Shipment No. <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No	Cask Seal No. <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No	Sample No. <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No
---	---	---	--	---

WHC-SD-WM-DP-076, REV. 0

Tank 241-BX-105  
Hot Cell Work Plan

I. Overview

- A. Waste tank 241-BX-105 sampling will consist of 2 auger samples. Two 20 inch augers will be used to sample risers 2 and 6. This is an exception to the TCP referenced below.
- B. This hot cell work plan (HCWP) is based upon Tank 241-BX-105 Tank Characterization Plan (WHC-SD-WM-TP-239 Rev. 0). If discrepancies occur between the HCWP and TCP, the TCP will take precedence, except where noted above. For specifics on QA/QC, refer to section 4.0 of the TCP.
- C. Auger samples will be loaded into 1-E2 hot cell in accordance with procedure LO-160-101, **Core Segment Receipt and Preparation**. The auger sample will be prepared for the 222-S Laboratory in accordance with procedure LO-160-103, **Core Segment Extrusion Process and Sample Preparation** and this work plan.

II. General Comments

- A. The cognizant scientist may deviate from this hot cell work plan should unforeseen circumstances arise. All deviations shall 1) be recorded in the laboratory notebook WHC-N-1028 and 2) relayed to the project coordinator.
- B. According to the TCP, Hot Cell operations will separate drainable liquid (if any) from sample. Hot Cell operations will isolate dry crust material (if any). Hot Cell operations will separate remaining sample into two equal subsamples.

III. Pre-job Preparation

1. Hot cell technicians will ensure that vials are tare weighed.
2. Hot cell technicians will apply labels with next available jar number. Clear tape will be used to cover and protect the labels on jars. Use appropriate outer covering on jars to minimize decontamination efforts when loaded out of the hot cell.
3. Hot Cell technicians will log all appropriate information about jars into logbook WHC-N-754. Record the date, jar number, jar size, tare weight, tank number, and customer I.D. after the jar contains the sample.

## WHC-SD-WM-DP-076, REV. 0

4. Hot Cell technicians will log all appropriate information on archive samples in logbook WHC-N-755. Record the date, jar number, jar size, tare weight, tank number, and customer I.D. after the jar contains the sample.
5. Check out video equipment and ensure battery is charged.
6. Obtain new Super-VHS tape and label 241-BX-105.

### IV. Sample Preparation, Extrusion and Breakdown

#### A. Sample Preparation

1. Perform section A (Preparation of 1E Laboratory for Cask Disassembly) of procedure LO-160-101.
2. Perform section B (Cask Receipt and Preparation of Sampler for Load-in into 1E-2 Hot Cell) of procedure LO-160-101.

Note: Contact Kevin Bell, Norton G. McDuffie (373-2653) and John Johnson before performing the next step.

3. Perform section C (Loading of Sampler into 1E-2) of procedure LO-160-101.

#### B. Sample Extrusion

1. Perform section I (Authorization and Preliminary Setup) of procedure LO-160-103.

NOTE: Position the auger liner so drainable liquid will not be lost during the removal of the auger.


2. Remove the rubber stopper.
3. Using the auger extraction tool, remove auger sample and sleeve from the liner.
4. Place the auger assembly on the auger stand.
5. Disconnect auger extraction tool from auger assembly.


NOTE: If video camera is available, video the process of removing the sleeve.

6. Remove sleeve from the auger sample.
7. Video sample on the auger.
8. Subsample auger as follows:

WHC-SD-WM-DP-076, REV. 0

- a. Inspect the top portion of the auger sample for a hard dry layer. If present, separate the hard dry layer and retain as a subsample. Place subsample in an appropriate size vial or jar.
  - b. Separate remaining auger sample into equal subsamples. Place each subsample into an appropriate size jar or vial.
9. Collect the liner liquid into an appropriate size collection vessel or jar.
  10. Clean auger, auger stand, and hot cell as required before processing the next sample.
  11. Prior to submitting samples to the laboratory for analyses, perform the following steps:
    - a. Homogenize each subsample individually
    - b. If any, filter the aqueous sample thru a 0.45 micron filter. Retain liquid subsample for safety screening analyses.
    - c. Remove aliquots of each sample according to the TCP.
    - d. Have project coordinator generate laboratory numbers and labels for all samples to be submitted to the laboratory.

Cognizant Scientist  Date 10/2/94

Project Coordinator  Date 10/6/94



Hot Cell Temp - 81.2 °F  
 % Humidity - 20 %

WHC-SD-WM-DP-076, REV. 0

Oct. 06, 1994


**Tank 241-BX-105**  
**Extruded Segment Description Sheet**

20 grams = 19.98  
 500 grams = 499.90

Sample 94-AUG-008 Riser 2 Core Auger Segment \_\_\_\_\_

**General Description of Sample:**

- (1) 1037. used auger extraction tool to remove auger from liner. observed no liner liquid.
- (2) Subsampled for DSC/TGA vial 6171# from Flutes 16# & 15# net weight @ 5.90 grams
- (3) vial #6175 - 15.10g very soft mud from tray # 7.90 grams #08 10-06-94
- (4) vial #6189 - Flutes 9 to Flute 10 net wt = 37.67 grams (125ml) white flaky material. #08 10-6-94
- (5) vial #6173 Flutes 9 to 1 net wt = 2.33g (2ml) Brown Flaky material.



PROJECT: BX-105 | S94T000138  
 GROUP: 9400003 | PRESERVATIVE: NONE  
 CUST ID: AUGER 008, RISER 2  
 ANALYSIS: DLIQV01, DLIQW01, RST.G/ML,  
 EXTRUD01, LLIQW01, NOPEBOOK, ORGVOL01,  
 SLDVOL01, SLDWT-01

Photo

Sampler Efficiency

Drainable Liquid Density

Volume of Sampler N/A

Total Weight NA

%Volume of Air ↓

Total Volume ↓

%Volume of Liquid ↓

Density ↓

%Volume of Solids ↓

Turbidity ↓

Comments: Auger Sampler

Comments: NONE

Solids

Liner Liquid

Drainable Liquid

Color Brown/white

Collection Jar 6122

Collection Jar NA

Consistency crusty to creamy

Gross Wt. ↓

Gross Wt. ↓

Homogeneity inhomogenous

Tare Wt. ↓

tare Wt. ↓

Texture crusty to moist.

Net Wt. NONE

Net Wt. ↓

Penetrometer NA

No. Phases NA

No. Phases ↓

Cognizant Scientist: Allen B. Barry

Oct 6, 1994

Reviewed By: Keith Fuller

Oct 07, 1994 13

TEMPERATURE INSIDE HOTCELL: 80.2 °F  
% Humidity : 20.0 %

WHC-SD-WM-DP-076, REV. 0


### Tank 241-BX-105 Extruded Segment Description Sheet

20grams - 20.01  
500grams - 499.70

Sample 94-AUG-009 Riser -6 Core AUGER Segment \_\_\_\_\_

**General Description of Sample:**

1. 1335 used auger extraction tool to remove AUGER from Liner.
2. 1340 remove sleeve from auger. White milky material, paste like consistency starting from Flutes 19 through 11. (Flutes 1-11 thin white translucent cover)
3. 6170<sup>#</sup> vial for DSC/TGA - 5.8 grams. 1350 hrs. Flutes (5) through (5).
4. 6119<sup>#</sup> jar (250ml) - 260.85 grams (Tare weighed w/o cap) Flutes (19) through (11).
5. 6124<sup>#</sup> jar (125ml) - 52.59 grams (Tare weighed w/o cap) Flutes (11) through (1).
6. 6122<sup>#</sup> jar (250ml) less than 5ml of Liner liquid, Did not retain.



Photo

Sampler Efficiency

Drainable Liquid Density

Volume of Sampler NA

Total Weight NA

%Volume of Air ∩

Total Volume ∩

%Volume of Liquid ∩

Density ∩

%Volume of Solids ∩

Turbidity ∩

Comments: Auger Sampler

Comments: NONE

PROJECT: BX-105 | S947000145  
GROUP: 94000004 | PRESERVATIVE: NONE

CUST ID: AUGER 009, RISER 6

ANALYSIS: DL1QV01, DL1QW01, EST.G/ML,  
EXTRU01, LLIQW01, NOTEBOOK, ORGVOLO1,  
SLDVOLO1, SLDWT-01

Solids

Liner Liquid

Drainable Liquid

Color white

Collection Jar 6122

Collection Jar NA

Consistency cream (damp)

Gross Wt. < 5ml

Gross Wt. ∩

Homogeneity Homogenous

Tare Wt. \_\_\_\_\_

tare Wt. ∩

Texture smooth

Net Wt. N/A

Net Wt. ∩

Penetrometer NA

No. Phases \_\_\_\_\_

No. Phases \_\_\_\_\_

Cognizant Scientist: Archie B. Bony

Oct 07 94

Reviewed By: Keith Fuller 21

**WHC-SD-WM-DP-076, REV. 0**

**UNDIGESTED SAMPLE ANALYSES-DIRECT**

August 5

OP

# LABCORE Data Entry Template for Worklist# 109

Analyst: KLV Instrument: DSC01 Method: LA-514-113 B-1 K11/14/94

Worklist Comment: Please use N2 purge. JMF **WHC-SD-WM-DP-076, REV. 0**

Seg	Type	Sample#	Rep	Al	Test	Matrix	Actual	Found	DL	Unit
1	STD				DSC-01	SOLID	<u>28.45</u>	<u>28.7</u>	<u>100.9</u> N/A	Joule
2	SAMPLE	S94T000139	0		DSC-01	SOLID	<u>N/A</u>	<u>0</u> <u>100.</u>	<u>g.m.F. 10/31/94</u>	Joule
3	DUP	S94T000139	0		DSC-01	SOLID	<u>0</u>	<u>0</u>	N/A	Joule
4	SAMPLE	S94T000146	0		DSC-01	SOLID	<u>N/A</u>	<u>0</u>		Joule
5	DUP	S94T000146	0		DSC-01	SOLID	<u>0</u>	<u>0</u>	N/A	Joule

Final page for worklist # 109

  
Analyst Signature

10/29/94  
Date

Std # 12N14A (Indivm)


Extend and approved. J.M. Luyze 11/1/94

Data Entry Comments:  
S94T000139 has 2 endotherms 459.5J/g at 117.5°C, 498.0J/g at 283.1°C

Duplicate: 386.8J/g at 116.1°C and 516.9J/g at 291.3°C

Sample S94T000146 has 2 endotherms of 507.9J/g at 122.3°C and 547.2J/g at 283.7°C; Duplicate has 441.4J/g at 122.9°C and 470.0J/g at 285.8°C

Units shown for QC (SPK) may not reflect the actual units.

  
10/31/94



S94T000139 N2

14.575 mg

File: 00014.001

Ident: 0.0


DSC METTLER

29-Oct-94

222-S Laboratory


Rate: 10.0 °C/min

WHC-SD-WM-DP-076, REV. 0



File: 00016.001 DSC 00111.FIN 29-Oct-94  
Weight: 0.0 222-S Laboratory

0017000239 (Dup) 1.2  
11.277 mg Rate: 10.0 °C/min



100. 200. 300. °C

S94T000146 N2

19.808 mg

File: 00018.001

DSC METTLER

29-Oct-94


Rate: 10.0 °C/min

Ident: 0.0

222-S Laboratory


50. mW

exo





S94T000146 (DUP) N2  
14.161 mg  
Rate: 10.0 °C/min  
File: 00020.001 DSC METTLER 29-Oct-94  
Ident: 0.0 222-S Laboratory



NOV. 5-94

OP

# LABCORE Data Entry Template for Worklist# 111

Analyst: KL V Instrument: DSC01 \_\_\_\_\_ Method: LA-514-113 B-1 KN 11/14/94  
 Worklist Comment: Please run using N2 purge.JMF WHC-SD-WM-DP-076, REV. 0

Seg	Type	Sample#	Rep	Al	Test	Matrix	Actual	Found	DL	Unit
1	STD				DSC-01	SOLID	28.45	28.5	100.2% Rec. N/A	Joul
2	SAMPLE	S94T000143	0		DSC-01	SOLID	N/A	0		Joul
3	DUP	S94T000143	0		DSC-01	SOLID	0	0	N/A	Joul
4	SAMPLE	S94T000147	0		DSC-01	SOLID	N/A	0		Joul
5	DUP	S94T000147	0		DSC-01	SOLID	0	0	N/A	Joul

### Final page for worklist # 111

  
 Analyst Signature

10/30/94  
 Date


Std # 12N14A (Indium)

Entered and approved. J.M. Luyze 11/1/94

Data Entry Comments:

S94T 000143 has two endotherms of 487.9 J/g at 131.3°C and 523.5 J/g at 289.0°C; Duplicate has 472.2 J/g at 109.5°C and 625.5 J/g at 289.1°C;  
 S94T 000147 has two endotherms of 217.3 J/g at 106.0°C and 732.1 J/g at 286.7°C;  
 Duplicate has 256.4 J/g at 107.8°C and 742.0 J/g at 284.6°C  
 Units shown for QC (SPK) may not reflect actual units.


Page: 1

  
 11/1/94

DSC STD  
6.504 mg

Rate: 10.0 °C/min

File: 00028.001 DSC METTLER 30-Oct-94  
Ident: 0.0 222-S Laboratory



WHC-SD-WM-DP-076, REV. 0

Integration  
Delta H 185 mJ  
Peak 160.8 °C  
-7.8 mW

SIGNATURE BELOW REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT  
COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 30 TO 34.

120. 140. 160. 180. °C

*Handwritten signature*  
10/30/94


S94T000143 N2  
13.767 mg

File: 00029.001  
Ident: 0.0

DSC METTLER  
222-S Laboratory

30-Oct-94


WHC-SD-WM-DP-076, REV. 0



File: 00031.001 DSC METTLER 30-Oct-94  
Ident: 0.0 222-S Laboratory

S94T000143 (DUP) N2  
11.367 mg

Rate: 10.0 °C/min




S94T000147 N2  
13.601 mg

File: 00035.001  
Ident: 0.0


DSC METTLER 30-Oct-94  
222-S Laboratory

Rate: 10.0 °C/min



S94T000147 (DUP) N2  
12.447 mg  
Rate: 10.0 °C/min  
File: 00036.001 DSC METTLER 30-Oct-94  
Ident: 0.0 222-S Laboratory

WHC-SD-WM-DP-076, REV. 0



August 9-10

00A  
1

# LABCORE Data Entry Template for Worklist# 112

Analyst: KLV Instrument: DSC01 Method: LA-514-113 B-1 KV 11/14/94

Worklist Comment: Please use N2 purge.JMF WHC-SD-WM-DP-076, REV. 0

Seg	Type	Sample#	Rep	AI	Test	Matrix	Actual	Found	DL	Unit
1	STD				DSC-01	SOLID	<u>27.45</u>	<u>29.6</u>	<u>104.0% Rec</u> N/A	Joule
2	SAMPLE	S94T000148	0		DSC-01	SOLID	<u>N/A</u>	<u>0</u>		Joule
3	DUP	S94T000148	0		DSC-01	SOLID	<u>0</u>	<u>0</u>	N/A	Joule
4	SAMPLE	S94T000190	0		DSC-01	SOLID	<u>N/A</u>	<u>168.4</u>		Joule
5	DUP	S94T000190	0		DSC-01	SOLID	<u>168.4</u>	<u>144.1</u>	N/A	Joule

Final page for worklist # 112

  
Analyst Signature

10/31/94  
Date

Std # 12N14A (Indium)

Entered and approved 11/1/94. J. M. Luye

Data Entry Comments:


S94T000148 has no exotherms but has two endotherms of 661.19 J/g at 118.8°C and 667.49 J/g at 285°C; duplicate has endotherms of 693.99 J/g at 108.7°C and 669.59 J/g at 285°C. S94T000190 also has two endotherms of 669.29 J/g at 115.40°C and 284.29 J/g at 287.9°C ~~and~~ the exotherm is at 381.3°C; duplicate has <sup>endotherms of</sup> 609.69 J/g at 122.1°C and 237.29 J/g at 290.0°C, and the exotherm at 379.6°C.

Units shown for Q (SPK) may not reflect the actual units.

Jerry Luye  
11/1/94



SIGNATURE BELOW REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT  
COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 36 TO 38




*Di Vito 10/31/94*

S94T000148 N2  
10.166 mg

File: 00040.001  
Ident: 0.0

DSC METTLER 31-Oct-94  
222-S Laboratory


Rate: 10.0 °C/min



S94T000148 (DUP) N2  
10.710 mg  
Rate: 10.0 °C/min

File: 00042.001 DSC METTLER 31-Oct-94  
Ident: 0.0 222-S Laboratory

WHC-SD-WM-DP-076, REV. 0



170 g 8 OP

# LABCORE Data Entry Template for Worklist# 110

Analyst: DWS Instrument: DSC01 \_\_\_\_\_ Method: LA-514-113 B-1 <sup>11/14/94</sup>

Worklist Comment: Please use N2 purge. JMF

WHC-SD-WM-DP-076, REV. 0

Seg Type	Sample#	Rep Al	Test	Matrix	Actual	Found	DL 100.9% Rec.	Unit
1 STD			DSC-01	SOLID	28.45	28.7	N/A	Joule
2 SAMPLE	S94T000140	0	DSC-01	SOLID	N/A	NO Sample		Joule
3 DUP	S94T000140	0	DSC-01	SOLID	no sample		N/A	Joule
4 SAMPLE	S94T000141	0	DSC-01	SOLID	N/A	0		Joule
5 DUP	S94T000141	0	DSC-01	SOLID	0	0	N/A	Joule

## Final page for worklist # 110

  
Analyst Signature

LAD 11-16-94  
~~10-11-94~~ 11-1-94  
Date

Std # 12N14A (Ind. um)

Entered and approved 11/1/94. JMF

Data Entry Comments:

S94T000141 has no exotherms but has two endotherms  
of 330.09 J/g at 116.9°C and 465.49 J/g at 289.4°C; Duplicate has 432.87 J/g  
at 125.6°C and 569.67 J/g at 284.4°C. JMF 11/1/94

Units shown for QC (SPK) may not reflect the actual units.

Page: 1

SIGNATURE BELOW REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT  
COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 40 TO 42 .

DSC STD  
6.565 mg

Rate: 10.0 °C/min

File: 00022.001  
Ident: 0.0

DSC METTLER 30-Oct-94  
222-S Laboratory

exo

5. mW

40

Integration  
Delta H 188 mJ  
28.7 J/g  
Peak 161.2 °C  
-7.2 mW

WHC-SD-WM-DP-076, REV. 0

120.

140.

160.

180. °C

*Healco Diag*  
11-11-94

S94T0000141

16.082 mg


File: 00023.001 DSC METTLER 30-Oct-94

Ident: 0.0 222-S Laboratory

Rate: 10.0 °C/min

EXO  $\Delta$

10. mW



Integration  
 Delta H 7485 mJ  
 465.4 J/g  
 Peak 289.4°C  
 -21.5 mW

Integration  
 Delta H 5307 mJ  
 330.0 J/g  
 Peak 116.9°C  
 -20.9 mW

WHC-SD-WM-DP-076, REV. 0

S94T0000141 DUP

20.795 mg

Rate: 10.0 °C/min

File: 00024.001

DSC METTLER

30-Oct-94


Ident: 0.0

222-S Laboratory

exo

20. mW

42



WHC-SD-WM-DP-076, REV. 0

2000 9 29

1012

# LABCORE Data Entry Template for Worklist# 106

Analyst: KLV Instrument: TGA01 Method: LA-560-112 <sup>KV 11/14/94</sup>  
 Worklist Comment: Please use N2 purge. JMF <sup>B+A-1 KV 11/14/94</sup>  
 WHC-SD-WM-DP-076, REV. 0

Seg	Type	Sample#	Rep	AI	Test	Matrix	Actual	Found	DL	Unit
1	STD	4208A			TGA-01	SOLID	59.19	57.67	97.49% Rec	%
2	SAMPLE	S94T000143	0		TGA-01	SOLID	N/A	14.63%		%
3	DUP	S94T000143	0		TGA-01	SOLID	14.63	18.73	N/A	%
4	SAMPLE	S94T000147	0		TGA-01	SOLID	N/A	4.91		%
5	DUP	S94T000147	0		TGA-01	SOLID	4.91	5.54	N/A	%

Final page for worklist # 106

  
 Analyst Signature

10/30/94  
 Date

Entered and approved 11/1/94 JMF.

Data Entry Comments:

S94T000143 sample has 2<sup>nd</sup> wt loss of 24.76% at 277.0°C; Duplicate has 20.55% at 279.0°C

S94T000147 sample has 2<sup>nd</sup> wt loss of 29.22% at 273.0°C; Duplicate has 27.98% at 277.0°C

Units shown for QC (SPK) may not reflect the actual units.



TGA STD  
19.284 mg

Rate: 10.0 °C/min

File: 00026.001

TG METTLER  
222-S Laboratory

10-30-94  
~~27-Oct-94~~  
LAD 11-16-94

SIGNATURE BELOW REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT  
COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 44 TO 48.

Step Analysis

Height-11.12 mg

-57.67 %

ResiC. 8.16 mg


42.33 %

Dpeak 81.0 °C

WHC-SD-WM-DP-076, REV. 0

44

5. mg



Kelley 10/30/94

S94T000143 N2

12.687 mg

Rate: 10.0 °C/min

File: 00030.001

TG

METTLER

Ident: 0.0

222-S Laboratory

10-30  
~~27-Oct-94~~  
LAD 11-16-98


15

2. mg

Step Analysis  
Height -1.86 mg  
-14.63 %  
ResiC. 10.83 mg  
85.37 %  
Dpeak 89.0 °C

Step Analysis  
Height -3.14 mg  
-24.76 %  
ResiC. 7.62 mg  
60.03 %  
Dpeak 277.0 °C

WHC-SD-WM-DP-076, REV. 0




LSD 11-16 94  
27-Oct-94  
10-30

File: 00032.001 TG METTLER  
Ident: 0.0 222-S Laboratory

S94T000143 (DUP) N2  
17.983 mg Rate: 10.0 °C/min

WHC-SD-WM-DP-076, REV. 0



Step Analysis  
Height -3.70 mg  
-20.55 %  
Resic. 10.75 mg  
59.79 %  
Dpeak 279.0 °C

Step Analysis  
Height -3.37 mg  
-18.73 %  
Resic. 14.62 mg  
81.27 %  
Dpeak 109.0 °C

S94T000147 N2

11.705 mg

Rate: 10.0 °C/min

File: 00034.001


TG

METTLER

Ident: 0.0

222-S Laboratory

10-30  
~~27-Oct-94~~  
LAD  
11-16-94



S94T000147 (DUP) N2

14.314 mg

Rate: 10.0 °C/min


File: 00037.001

TG METTLER

Ident: 0.0

222-S Laboratory

10-30  
~~28 Oct-94~~  
11-16-94  
LAD



WHC-SD-WM-DP-076, REV. 0

# LABCORE Data Entry Template for Worklist# 157

DA  
9

Analyst: DWS Instrument: TGA01 \_\_\_\_\_ Method: LA-560-112 Rev A-1  
 WHC-SD-WM-DP-076, REV. 0  
 Worklist Comment: This is a rerun for a 45 day report. please rush asap lad

Seg	Type	Sample#	Rep	Al	Test	Matrix	Actual	Found	DL	Unit
1	STD	<del>48</del> <sup>2</sup> N8A			TGA-01	SOLID	59.19	58.52	N/A	%
2	SAMPLE	S94T000143	0		TGA-01	SOLID	N/A	15.74		%
3	DUP	S94T000143	0		TGA-01	SOLID	15.74	13.43	N/A	%

## Final page for worklist # 157

Doreen W. Smith  
 Analyst Signature

11-10-94  
 Date  
 11-10-94

approved Jenn Suye 11/11/94

Data Entry Comments:  
S94T000143 has a second weight loss step of  
20.21% at 277°C and 20.90% at 275°C in the duplicate.

762 STD  
1.520 mg

29mF 11/10/94  
48N8A

Rate: 15.0 °C/min

File: 00081.001

Ident: 0.0

TC PHILLIP  
222-2 Laboratory

07-107-91  
11/10/94  
9mF

SIGNATURE BELOW REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT  
COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 50 TO 52.

Step Analysis

Height -8.50 mg

-58.52 %

ResiC. 6.03 mg


41.48 %

Dpeak 79.0 °C

BEST AVAILABLE COPY

WHC-SD-WM-DP-076, REV. 0

James L. Lyle  
for Dave Smith  
11/10/94



50

50

200

300

400

500

37.801 mg

37.801 mg

37.801 mg

11-10  
KAD  
11-16-94


Step Analysis  
Height -5.01 mg  
-15.74 %  
ResiC. 26.79 mg  
84.26 %  
Dpeak 97.0 °C

Step Analysis  
Height -6.47 mg  
-20.33 %  
ResiC. 20.21 mg  
63.54 %  
Dpeak 277.0 °C

WHC-SD-WM-DP-076, REV. 0

BEST AVAILABLE COPY

51





29 17 0007 43 (76 1)  
18.823 mg

Time: 10.0 min

170:00073.001  
Date: 0.0

16 1-17-11-13  
277-8 Laboratory


11-10  
LAD  
11-16-94

### Step Analysis

Height -2.53 mg  
-13.43 %  
ResiC. 16.29 mg  
86.57 %  
Dpeak 79.0 °C

### Step Analysis

Height -3.93 mg  
-20.90 %  
ResiC. 12.19 mg  
64.76 %  
Dpeak 275.0 °C



BEST AVAILABLE COPY

52

WHC-SD-WM-DP-076, REV. 0

# LABCORE Data Entry Template for Worklist# 107

108

Analyst: KLV Instrument: TGA01 Method: LA-560-112 A-1 KV 11/14/94  
 WHC-SD-WM-DP-076, REV. 0

Worklist Comment: Please use N2 purge. JMF

Seg	Type	Sample#	Rep	AI	Test	Matrix	Actual	Found	DL	Unit
1	STD	42N8A			TGA-01	SOLID	59.19	58.61	99.0% N/A	%
2	SAMPLE	S94T000148	0		TGA-01	SOLID	N/A	16.18		%
3	DUP	S94T000148	0		TGA-01	SOLID	16.18	14.78	N/A	%
4	SAMPLE	S94T000190	0		TGA-01	SOLID	N/A	16.98		%
5	DUP	S94T000190	0		TGA-01	SOLID	16.98	17.82	N/A	%

**Final page for worklist # 107**

  
 Analyst Signature

10/31/94  
 Date


Entered and approved. J. M. Foye 11/1/94

Data Entry Comments:

S94T000148 Bluey, gooey Ivory color, has 2<sup>nd</sup> wt loss 23.64% at 275°C; dup  
 23.53% at 273°C. S94T000190 Rocky road ice cream consistency, has 2<sup>nd</sup>  
 wt loss of 24.01% at 279°C; dup 23.78% at 281°C - J. M. Foye

Units shown for QC (SPK) may not reflect the actual units.





S94T000148 (DUP) N2

15.364 mg

Rate: 10.0 °C/min

File: 00043.001

TG


METTLER

10-31  
28-Oct-94

Ident: 0.0

222-S Laboratory

LAD  
11-16-94



WHC-SD-WM-DP-076, REV. 0

Page 8 10A

# LABCORE Data Entry Template for Worklist# 104

Analyst: KLK Instrument: TGA01 Method: LA-560-112 A-1 KV 11/14/94

Worklist Comment: Please use N2 purge. JMF

WHC-SD-WM-DP-076, REV. 0

Seg	Type	Sample#	Rep	AI	Test	Matrix	Actual	Found	DL	Unit
1	STD	42N8A			TGA-01	SOLID	59.19 <del>58.89</del> <small>JMF 10/28/94 JMF</small>	58.89	99.5% N/A	%
2	SAMPLE	S94T000139	0		TGA-01	SOLID	N/A	9.86		%
3	DUP	S94T000139	0		TGA-01	SOLID	9.86	11.53	N/A	%
4	SAMPLE	S94T000146	0		TGA-01	SOLID	N/A	18.97		%
5	DUP	S94T000146	0		TGA-01	SOLID	18.97	18.77	N/A	%

Final page for worklist # 104

  
Analyst Signature

10/29/94  
Date

*Extend and approved 11/1/94*

Data Entry Comments:

S94T000139 has second weight loss step 26.41% at 277.0°C; Duplicate  
25.36% at 277.0°C.

S94T000146 has second weight loss step 21.62% at 279°C, dup 21.97% at 275°C

Units shown for QC (SPK) may not reflect the actual units.



S94T000139 N2

12.420 mg

Rate: 10.0 °C/min


File: 00015.001

TG METTLER

Ident: 0.0

222-S Laboratory

29  
20 Oct-94  
LAD  
11-16-94





S94T000139 (DUP) N2

12.801 mg

Rate: 10.0 °C/min

File: 00017.001

TG METTLER

Ident: 0.0

222-S Laboratory

29  
26-Oct-94  
LAD  
11/16/94

50

2. mg

Step Analysis

Height -1.48 mg

-11.53 %

ResiC. 11.33 mg

88.47 %

Dpeak 47.0 °C

Step Analysis

Height -3.25 mg


-25.36 %

ResiC. 7.92 mg

61.87 %

Dpeak 277.0 °C

WHC-SD-MM-DP-076, REV. 0



S94T000146 N2

15.534 mg

Rate: 10.0 °C/min

File: 00019.001


TG

METTLER

29  
27-Oct-94  
LAD  
11-16-94

Ident: 0.0

222-S Laboratory



WHC-SD-WM-DP-076, REV. 0


61

29  
27-Oct-94  
LAD  
11-16-94

File: 00021.001 TG METTLER  
Ident: 0.0 222-S Laboratory

S94T000146 (DUP) N2  
Rate: 10.0 °C/min  
14.230 mg

WHC-SD-WM-DP-076, REV. 0



Step Analysis  
Height -3.13 mg  
ResiC. 8.26 mg  
Dpeak 275.0 °C

Step Analysis  
Height -2.67 mg  
ResiC. 11.56 mg  
Dpeak 47.0 °C

Aug 8

OK

# LABCORE Data Entry Template for Worklist# 105

Analyst: DWS Instrument: TGA01 Method: LA-560-112 A-1 <sup>KV</sup> 11/14/94

Worklist Comment: Please use N2 purge. JMF

WHC-SD-WM-DP-076, REV. 0

Seg	Type	Sample#	Rep	AI	Test	Matrix	Actual	Found	DL	Unit
1	STD	42N8A			TGA-01	SOLID	59.19	58.14	N/A	%
2	SAMPLE	S94T000140	0		TGA-01	SOLID	N/A	sample broken		%
3	DUP	S94T000140	0		TGA-01	SOLID	sample broken		N/A	%
4	SAMPLE	S94T000141	0		TGA-01	SOLID	N/A	13.76		%
5	DUP	S94T000141	0		TGA-01	SOLID	13.76	7.24	N/A	%

### Final page for worklist # 105

*DWS*  
Analyst Signature

11-1-94  
Date

*Entered & approved 11/3/94 J. M. Luyk*

Data Entry Comments: S94T000141 has a second weight loss of 17.27% at 275°C  
its duplicate has 22.02% at 279.0°C. J. M. Luyk 11/3/94

# LABCORE Data Entry Template for Worklist# 165

Analyst: JMF Instrument: DSC01 Method: LA-514-113

Worklist Comment: Dry DSC BX-105. JMF

WHC-SD-wm-DP-076, Rev. 0

Seg	Type	Sample#	Rep	AI	Test	Matrix	Actual	Found	DL	Unit
1	SAMPLE	S94T000139	0		DSC-02	SOLID	N/A	0		Joule
2	DUP	S94T000139	0		DSC-02	SOLID	0	0	N/A	Joule
3	SAMPLE	S94T000146	0		DSC-02	SOLID	N/A	0		Joule
4	DUP	S94T000146	0		DSC-02	SOLID	0	0	N/A	Joule
5	SAMPLE	S94T000141	0		DSC-02	SOLID	N/A	0		Joule
6	DUP	S94T000141	0		DSC-02	SOLID	0	0	N/A	Joule
7	SAMPLE	S94T000143	0		DSC-02	SOLID	N/A	0		Joule
8	DUP	S94T000143	0		DSC-02	SOLID	0	0	N/A	Joule
9	SAMPLE	S94T000147	0		DSC-02	SOLID	N/A	0		Joule
10	DUP	S94T000147	0		DSC-02	SOLID	0	0	N/A	Joule
11	SAMPLE	S94T000148	0		DSC-02	SOLID	N/A	0		Joule
12	DUP	S94T000148	0		DSC-02	SOLID	0	0	N/A	Joule

**Final page for worklist # 165**

*James Luge*  
Analyst Signature

11/16/94  
Date

*Entered & reviewed 11/16/94. JMF*

Data Entry Comments:

---



---

# LABCORE Data Entry Template for Worklist# 105

Analyst: \_\_\_\_\_ Instrument: TGA01 \_\_\_\_\_ Method: LA-560-112  
Worklist Comment: Please use N2 purge. JMF **WHC-SD-WM-DP-076, REV. 0**

Seg	Type	Sample#	Rep	AI	Test	Matrix	Actual	Found	DL	Unit
1	STD				TGA-01	SOLID			N/A	%
2	SAMPLE	S94T000141	0		TGA-01	SOLID	N/A			%
3	DUP	S94T000141	0		TGA-01	SOLID			N/A	%

**Final page for worklist # 105**

\_\_\_\_\_  
Analyst Signature

\_\_\_\_\_  
Date

Data Entry Comments:

*See attached work list for data. JMF  
4/13/94*

Units shown for QC (SPK) may not reflect the actual units.

Page: 1

TGA STD  
12.930 mg

Rate: 10.0 °C/min

File: 00066.001  
Ident: 0.0


TG METTLER  
222-S Laboratory

11-1  
~~30-Oct-94~~  
LAD  
11-16-94

SIGNATURE BELOW REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT  
COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 65 TO 67.

Step Analysis  
Height -7.52 mg  
-58.14 %  
ResiC. 5.41 mg  
41.86 %  
Dpeak 71.0 °C

WHC-SD-WM-DP-076, REV. 0



65


Leslie Dian 11-16-94

11-1  
30-Oct-94  
LAD  
11-16-94

S94T0000141 N2  
17.848 mg  
File: 00067.001 TG METTLER  
Ident: 0.0 222-S Laboratory

Rate: 10.0 °C/min

WHC-SD-WM-DP-076, REV. 0



Step Analysis  
Height -3.08 mg  
-17.27 %  
Resid. 12.29 mg  
68.87 %  
Dpeak 275.0 °C

Step Analysis  
Height -2.46 mg  
-13.76 %  
Resid. 15.39 mg  
86.24 %  
Dpeak 63.0 °C


100. 200. 300. 400. °C



File: 00068.001 TG METTLER 30-Get-94  
Ident: 0.0 222-S Laboratory LAD 11-16-94

S94T0000141 (DUP) N2  
19.651 mg Rate: 10.0 °C/min

WHC-SD-WM-DP-076, REV. 0



Step Analysis  
Height -4.33 mg  
-22.02 %  
Resic. 13.90 mg  
70.75 %  
Dpeak 279.0 °C

Step Analysis  
Height -1.42 mg  
-7.24 %  
Resic. 18.23 mg  
92.76 %  
Dpeak 53.0 °C

**WHC-SD-WM-DP-076, REV. 0**

**FUSION DIGESTION ANALYSES**

August 8  
08

# LABCORE Data Entry Template for Worklist# 139

Analyst: JIS Instrument: AB00 16 Method: LA-508-101 D-2  
 WHC-SD-WM-DP-076, REV. 0

Worklist Comment: Use a sample size of .050 mL and a SPK volume of .100 mL.

Seg	Type	Sample#	Rep	AI	Test	Matrix	Actual	Found	DL	Unit
h2	1	BLNK-PREP			@ALPHA01 ALPHA01	SOLID			N/A	uCi/g
	1	BLNK-PREP			@ALPHA01 ALPHA01E	SOLID			N/A	% Ct.
3,4	2	STD			@ALPHA01 ALPHA01	SOLID			N/A	uCi/g
	2	STD			@ALPHA01 ALPHA01E	SOLID			N/A	% Ct.
5,6	3	SAMPLE	S94T000142	0	F	@ALPHA01 ALPHA01	SOLID	N/A		uCi/g
	3	SAMPLE	S94T000142	0	F	@ALPHA01 ALPHA01E	SOLID	N/A		% Ct.
7,8	4	DUP	S94T000142	0	F	@ALPHA01 ALPHA01	SOLID		N/A	uCi/g
	4	DUP	S94T000142	0	F	@ALPHA01 ALPHA01E	SOLID		N/A	% Ct.
9,10	5	SPK	S94T000142	0	F	@ALPHA01 ALPHA01	SOLID		N/A	uCi/g
11,12	6	SAMPLE	S94T000144	0	F	@ALPHA01 ALPHA01	SOLID	N/A		uCi/g
	6	SAMPLE	S94T000144	0	F	@ALPHA01 ALPHA01E	SOLID	N/A		% Ct.
13,14	7	DUP	S94T000144	0	F	@ALPHA01 ALPHA01	SOLID		N/A	uCi/g
	7	DUP	S94T000144	0	F	@ALPHA01 ALPHA01E	SOLID		N/A	% Ct.
15,16	8	SPK	S94T000144	0	F	@ALPHA01 ALPHA01	SOLID		N/A	uCi/g

**Final page for worklist # 139**

JIS  
 Analyst Signature  
 M/S

11-02-94  
 Date  
 11/3/94

**BEST AVAILABLE COPY**

**Data Entry Comments:**

Low SPK recoveries are due to high dissolved solids - No rework  
needed. High RPD for sample S94T000144 is acceptable due to  
low sample activity. 11/3/94

Units shown for QC (SPK) may not reflect the actual units.

Page: 1

11-18-94

LA-508-101 D-1 D-2

AT: LIQUIDS/SOLIDS		STANDARD	REPLICATE
Type	DETECTOR NUMBER	16	
STANDARD	DISH SIZE 1, 2, or 5 (MS)	2	
Work List	TOTAL COUNTS (TC)	1277	1346
139	COUNT TIME in MINUTES (CT)	30	
AT or TB 2	BACKGROUND in cpm (BKG)	.1	
AT	SAMPLE SIZE in mL (SS)	<del>8.50</del> 10mL	
Test Code	DILUTION FACTOR (DF)	1	
Alpha 01	DIGEST DILUTION FACTOR (DDF)	1	
Matrix	EFFICIENCY FACTOR (EFF)		
LIQUID	Le, Rmax, or Rs (SAMPLE RATE) as APPROPRIATE		
Sample #	STANDARD BOOK 161852	9.54E-3	
Instrument Code			
WB27806			
Analyst			
DVS			
Date			
11-2-94			
Time			
1500			

WHC-SD-WM-DP-076, REV. 0

D-2 LAD 11-16-94

AT: LIQUIDS/SOLIDS		BLANK	REPLICATE
Type	DETECTOR NUMBER	16	
BLANK	DISH SIZE 1, 2, or 5 (MS)	2	
	TOTAL COUNTS (TC)	5	4
	COUNT TIME in MINUTES (CT)	30	
AT or TB 2	BACKGROUND in cpm (BKG)	.1	
AT 142/144	SAMPLE SIZE in mL (SS)	.050	
	DILUTION FACTOR (DF)	1	
	DIGEST DILUTION FACTOR (DDF)	2.1175	
Matrix	EFFICIENCY FACTOR (EFF)		
<del>LIQUID</del> Solid	Le, Rmax, or Rs (SAMPLE RATE) as APPROPRIATE		

Data Entry by: <i>Joem Van Sant</i>	Date: 11-2-94
Data Entry by: <i>M. C. ...</i>	Date: 11/3/94

508101\_C

5,6

LA-508-101 D-1		LIQUIDS/SOLIDS	SAMPLE	REPLICATE
Type	DETECTOR NUMBER		16	
SAMPLE	DISH SIZE 1, 2, or 5 (MS)		2	
Work List	TOTAL COUNTS (TC)		844	930
139	COUNT TIME in MINUTES (CT)		30	
AT or TB?	BACKGROUND in cpm (BKG)		.1	
AT	SAMPLE SIZE in mL (SS)		.050	
Test Code	DILUTION FACTOR (DF)		1	
Alpha 01	DIGEST GRAMS of SOLIDS/L (Dg/L)		2.1175	
Matrix	EFFICIENCY FACTOR (EFF)			
SOLID	I.e. Rmax, or Rs, (SAMPLE RATE) as APPROPRIATE			
Sample #				
594T000142				
Instrument Code				
WB 27806				
Analyst				
JJS				
Date				
11-2-94				
Time				
1500				

WHC-SD-WM-DP-076, REV. 0

7,8

2AD  
11-10-94

LA-508-101 D-1		AT: LIQUIDS/SOLIDS	DUPLICATE	REPLICATE
Type	DETECTOR NUMBER		16	
DUPLICATE	DISH SIZE 1, 2, or 5 (MS)		2	
	TOTAL COUNTS (TC)		859	882
	COUNT TIME in MINUTES (CT)		30	
AT or TB?	BACKGROUND in cpm (BKG)		.1	
AT	SAMPLE SIZE in mL (SS)		.050	
	DILUTION FACTOR (DF)		1	
	DIGEST GRAMS of SOLIDS/L (Dg/L)		2.1235	
Matrix	EFFICIENCY FACTOR (EFF)			
SOLID	I.e. Rmax, or Rs, (SAMPLE RATE) as APPROPRIATE			

Data Entry by: <i>Jm [Signature]</i>	Date: 11-2-94
Data Entry by: <i>M [Signature]</i>	Date: 11-3-94

508101\_C

9,10

LA-508-101 D <sup>2</sup> AT: SPIKED SAMPLE		SPIKE	REPLICATE
Type	DETECTOR NUMBER	16	
SPIKE	DISH SIZE 1, 2, or 5 (MS)	2	
Work List	TOTAL COUNTS (TC)	43264	37915
139	COUNT TIME in MINUTES (CT)	30	
AT or TB?	BACKGROUND in cpm (BKG)	.1	
AT	SAMPLE VOLUME in mL (Spiked Vial) (SS)	.050	
Test Code	SAMPLE DILUTION FACTOR (Spiked Vial) (DF)	1	
Alpha 01	DIGEST GRAMS of SOLIDS/L (Dg/L)	2.1175	
Matrix	SPIKE VOLUME in mL (SVol)	.100	
solid	SPIKE DILUTION FACTOR (SDF)	1	
Sample #	SPIKE VALUE in $\mu$ Ci/L (SVal)	36.4	
S94T000142	INSTRUMENT EFFICIENCY FACTOR (EFF)		
Instrument Code	SAMPLE + SPIKE $\mu$ Ci/g (S+S)		
WB27806	AVERAGE or MAXIMUM $\mu$ Ci/g from FORM C	5.96E-1	
Analyst	SPIKE BOOK #	94843	36.4
JWS			
Date			
11-2-94			
Time			
1500			

WHC-SD-WM-DP-076, REV. 0

Data Entry by: Cheryl Vance Date: 11-2-94  
 508101\_X  
M. B. Brown 11-3-94

LA-508-101 D <sup>2</sup> TB: SPIKED SAMPLE		SPIKE	REPLICATE
Type	DETECTOR NUMBER		
SPIKE	DISH SIZE 1, 2, or 5 (MS)		
	TOTAL COUNTS (TC)		
	COUNT TIME in MINUTES (CT)		
	BACKGROUND in cpm (BKG)		
	SAMPLE VOLUME in mL (Spiked Vial) (SS)		
	SAMPLE DILUTION FACTOR (Spiked Vial) (DF)		
	DIGEST GRAMS of SOLIDS/L (Dg/L)		
	SPIKE VOLUME in mL (SVol)		
	SPIKE DILUTION FACTOR (SDF)		
	SPIKE VALUE in $\mu$ Ci/L (SVal)		
	INSTRUMENT EFFICIENCY FACTOR (EFF)		
	SAMPLE + SPIKE $\mu$ Ci/g (S+S)		
	AVERAGE or MAXIMUM $\mu$ Ci/g from FORM C		
	SPIKE BOOK #		

Data Entry by: \_\_\_\_\_ Date: \_\_\_\_\_  
 Data Entry by: \_\_\_\_\_ Date: \_\_\_\_\_

D-2 LAD 11-16-94

11,12

LA-508-101 D-1	AT: LIQUIDS/SOLIDS	SAMPLE	REPLICATE
Type	DETECTOR NUMBER	16	
SAMPLE	DISH SIZE 1, 2, or 5 (MS)	2	
Work List	TOTAL COUNTS (TC)	487	522
139	COUNT TIME in MINUTES (CT)	30	
AT or TB?	BACKGROUND in cpm (BKG)	.1	
AT	SAMPLE SIZE in mL (SS)	.050	
Test Code	DILUTION FACTOR (DF)	1	
Alpha O\	DIGEST GRAMS of SOLIDS/L (Dg/L)	2.1215	
Matrix	EFFICIENCY FACTOR (EFF)		
SOLID	LC, Rmax, or Rs (SAMPLE RATE) as APPROPRIATE		
Sample #			
594T000144			
Instrument Code			
WB27806			
Analyst			
JVS			
Date			
11-2-94			
Time			
1500			

WHC-SD-WM-DP-076, REV. 0

LAD 11-16-94

13,14

LA-508-101 D-1	AT: LIQUIDS/SOLIDS	DUPLICATE	REPLICATE
Type	DETECTOR NUMBER	16	
DUPLICATE	DISH SIZE 1, 2, or 5 (MS)	2	
	TOTAL COUNTS (TC)	298	312
	COUNT TIME in MINUTES (CT)	30	
AT or TB?	BACKGROUND in cpm (BKG)	.1	
AT	SAMPLE SIZE in mL (SS)	.050	
	DILUTION FACTOR (DF)	1	
	DIGEST GRAMS of SOLIDS/L (Dg/L)	2.1190	
Matrix	EFFICIENCY FACTOR (EFF)		
SOLID	LC, Rmax, or Rs (SAMPLE RATE) as APPROPRIATE		

Data Entry by: *[Signature]* Date: 11-2-94  
 Data Entry by: *[Signature]* Date: 11-3-94

508101\_C



LA-508-101 D<sup>2</sup> AT: SPIKED SAMPLE

15/16

		SPIKE	REPLICATE
Type	DETECTOR NUMBER	16	
SPIKE	DISH SIZE 1, 2, or 5 (MS)	2	
Work List	TOTAL COUNTS (TC)	39274	33219
139	COUNT TIME in MINUTES (CT)	30	
AT or TB?	BACKGROUND in cpm (BKG)	.1	
AT	SAMPLE VOLUME in mL (Spiked Vial) (SS)	.050	
Test Code	SAMPLE DILUTION FACTOR (Spiked Vial) (DF)	1	
Alpha 01	DIGEST GRAMS of SOLIDS/L (Dg/L)	2.1215	
Matrix	SPIKE VOLUME in mL (SVol)	.100	
solid	SPIKE DILUTION FACTOR (SDF)	1	
Sample #	SPIKE VALUE in $\mu$ Ci/L (SVal)	36.4	
594000144	INSTRUMENT EFFICIENCY FACTOR (EFF)		
Instrument Code	SAMPLE + SPIKE $\mu$ Ci/g (S+S)		
WB27806	AVERAGE or MAXIMUM $\mu$ Ci/g from FORM C	3.37E-1	
Analyst	SPIKE BOOK #	94843	
JWS			
Date			
11-2-94			
Time			
1500			

WHC-SD-WM-DP-076, REV. 0

Data Entry by: *JWS* Date: 11-2-94

*M. B. ...*  
508101\_X

11-3-94

LA-508-101 D<sup>2</sup> TB: SPIKED SAMPLE

		SPIKE	REPLICATE
Type	DETECTOR NUMBER		
SPIKE	DISH SIZE 1, 2, or 5 (MS)		
	TOTAL COUNTS (TC)		
	COUNT TIME in MINUTES (CT)		
	BACKGROUND in cpm (BKG)		
	SAMPLE VOLUME in mL (Spiked Vial) (SS)		
	SAMPLE DILUTION FACTOR (Spiked Vial) (DF)		
	DIGEST GRAMS of SOLIDS/L (Dg/L)		
	SPIKE VOLUME in mL (SVol)		
	SPIKE DILUTION FACTOR (SDF)		
	SPIKE VALUE in $\mu$ Ci/L (SVal)		
	INSTRUMENT EFFICIENCY FACTOR (EFF)		
	SAMPLE + SPIKE $\mu$ Ci/g (S+S)		
	AVERAGE or MAXIMUM $\mu$ Ci/g from FORM C		
	SPIKE BOOK #		

Data Entry by: \_\_\_\_\_ Date: \_\_\_\_\_  
Data Entry by: \_\_\_\_\_ Date: \_\_\_\_\_

508101\_X

74



PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

WHC-SD-WM-DP-076, REV. 0

AT : LA-508-101 (D-2) LIQUIDS

		STANDARD	REPLICATE
Type	DETECTOR NUMBER	16	16
STANDARD	DISH SIZE 1, 2, or 5 (MS)	2	2
Work List	TOTAL COUNTS (TC)	1277	1346
139	COUNT TIME in MINUTES (CT)	30	30
AT or TB ?	BACKGROUND in cpm (BKG)	0.1	0.1
AT	SAMPLE SIZE in mL (SS)	10.000	10.000
Test Code	DILUTION FACTOR (DF)	1	1
ALPHA01	DIGEST DILUTION FACTOR (DDF)	1	1
Matrix	EFFICIENCY FACTOR (EFF)	0.2104	0.2104
LIQUID	Lc, Rmax, or Rs,(SAMPLE RATE) as APPROPRIATE	42.467	44.767
Sample #	Sample Concentration in $\mu\text{Ci/L}$	9.09E-03	
	Replicate Concentration in $\mu\text{Ci/L}$	9.58E-03	
Instrument Code			
WB27806	Average Concentration in $\mu\text{Ci/L}$	9.3380E-03	
Analyst			
JMV	$R_s$ (Sample Count Rate) = (TC / CT) - BKG		
Date	ALPHA TOTAL $\mu\text{Ci/L}$ = $R_s * 1000\text{mL/L} * DF * DDF / (EFF * SS * 2220000\text{dpm}/\mu\text{Ci})$		
11/02/94	ALPHA TOTAL $\mu\text{Ci/mL}$ = ALPHA TOTAL $\mu\text{Ci/L} / 1000\text{mL/L}$		
Time	Relative Counting Error = $[(\text{The Square Root of } TC + BKG * CT) / (TC - BKG * CT)] * 1.96 * 100$		
03:00 PM	Detection Levels and Less Than Values are determined from Procedure LA-508-002.		

v RESULTS v

ALPHA TOTAL	in $\mu\text{Ci/mL}$ (Average) =	9.34E-06	DETECTION LEVEL
RELATIVE COUNTING ERROR	=	5.5%	

Data Entry by: *M. Brown* Date: 11/03/94  
 Approved by: *[Signature]* Date: 11/3/94

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

WHC-SD-WM-DP-076, REV. 0

AT : LA-508-101 (D-2) SOLIDS

		BLANK	REPLICATE
Type	DETECTOR NUMBER	.16	.16
BLANK	DISH SIZE 1, 2, or 5 (MS)	2	2
Work List	TOTAL COUNTS (TC)	5	4
139	COUNT TIME in MINUTES (CT)	.30	.30
AT or TB ?	BACKGROUND in cpm (BKG)	0.1	0.1
AT	SAMPLE SIZE in mL (SS)	0.050	0.050
Test Code	DILUTION FACTOR (DF)	1	1
ALPHA01	DIGEST GRAMS of SOLIDS/L (Dg/L)	2.1175	2.1175
Matrix	EFFICIENCY FACTOR (EFF)	0.2104	0.2104
SOLID	Lc, Rmax, or Rs,(SAMPLE RATE) as APPROPRIATE	0.222	0.179
Sample #	Sample Concentration in $\mu\text{Ci/g}$	< 4.49E-03	
142,144	Replicate Concentration in $\mu\text{Ci/g}$	< 3.62E-03	
Instrument Code			
WB27806	Maximum Concentration in $\mu\text{Ci/g}$	< 4.4938E-03	
Analyst			
JMV	$R_s$ (Sample Count Rate) = (TC / CT) - BKG		
Date	ALPHA TOTAL $\mu\text{Ci/g}$ = $R_s * 1000\text{mL/L} * DF / (EFF * SS * Dg/L * 2220000\text{dpm}/\mu\text{Ci})$		
11/02/94			
Time	Relative Counting Error = $[(\text{The Square Root of } TC + BKG * CT) / (TC - BKG * CT)] * 1.96 * 100$		
03:00 PM	Detection Levels and Less Than Values are determined from Procedure LA-508-002.		

RESULTS

ALPHA TOTAL	in $\mu\text{Ci/g}$ (Maximum) =	< 4.49E-03	DETECTION LEVEL
LESS THAN Value was Determined from Rmax.			7.28E-03
RELATIVE COUNTING ERROR	=	500.0%	$\mu\text{Ci/g}$

Data Entry by: *M. B. ...* Date: 11/03/94  
 Approved by: *[Signature]* Date: 11/3/94

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

WHC-SD-WM-DP-076, REV. 0

AT : LA-508-101 (D-2) SOLIDS

		SAMPLE	REPLICATE
Type	DETECTOR NUMBER	16	16
SAMPLE	DISH SIZE 1, 2, or 5 (MS)	2	2
Work List	TOTAL COUNTS (TC)	844	930
139	COUNT TIME in MINUTES (CT)	30	30
AT or TB ?	BACKGROUND in cpm (BKG)	0.1	0.1
AT	SAMPLE SIZE in mL (SS)	0.050	0.050
Test Code	DILUTION FACTOR (DF)	1	1
ALPHA01	DIGEST GRAMS of SOLIDS/L (Dg/L)	2.1175	2.1175
Matrix	EFFICIENCY FACTOR (EFF)	0.2104	0.2104
SOLID	Lc, Rmax, or Rs,(SAMPLE RATE) as APPROPRIATE	28.033	30.900
Sample #	Sample Concentration in $\mu\text{Ci/g}$	5.67E-01	
S94T00142	Replicate Concentration in $\mu\text{Ci/g}$	6.25E-01	
Instrument Code			
WB27806	Average Concentration in $\mu\text{Ci/g}$	5.9585E-01	
Analyst			
JMV	$R_s$ (Sample Count Rate) = (TC / CT) - BKG		
Date	ALPHA TOTAL $\mu\text{Ci/g}$ = $R_s * 1000\text{mL/L} * DF / (EFF * SS * Dg/L * 2220000\text{dpm}/\mu\text{Ci})$		
11/02/94			
Time	Relative Counting Error = $[(\text{The Square Root of } TC + BKG * CT) / (TC - BKG * CT)] * 1.96 * 100$		
03:00 PM	Detection Levels and Less Than Values are determined from Procedure LA-508-002.		

v RESULTS v

ALPHA TOTAL in $\mu\text{Ci/g}$ (Average)	=	5.96E-01	DETECTION LEVEL
RELATIVE COUNTING ERROR	=	6.8%	

Data Entry by: <i>M. S. ...</i>	Date: 11/03/94
Approved by: <i>[Signature]</i>	Date: 11/3/94

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

WHC-SD-WM-DP-076, REV. 0

AT : LA-508-101 (D-2) SOLIDS

		SAMPLE	REPLICATE
Type	DETECTOR NUMBER	16	16
DUPLICATE	DISH SIZE 1, 2, or 5 (MS)	2	2
Work List	TOTAL COUNTS (TC)	859	882
139	COUNT TIME in MINUTES (CT)	30	30
AT or TB ?	BACKGROUND in cpm (BKG)	0.1	0.1
AT	SAMPLE SIZE in mL (SS)	0.050	0.050
Test Code	DILUTION FACTOR (DF)	1	1
ALPHA01	DIGEST GRAMS of SOLIDS/L (Dg/L)	2.1235	2.1235
Matrix	EFFICIENCY FACTOR (EFF)	0.2104	0.2104
SOLID	Lc, Rmax, or Rs,(SAMPLE RATE) as APPROPRIATE	28.533	29.300
Sample #	Sample Concentration in $\mu\text{Ci/g}$	5.75E-01	
S94T000142	Replicate Concentration in $\mu\text{Ci/g}$	5.91E-01	
Instrument Code			
WB27806	Average Concentration in $\mu\text{Ci/g}$	5.8308E-01	
Analyst			
JMV	$R_s$ (Sample Count Rate) = (TC / CT) - BKG		
Date	ALPHA TOTAL $\mu\text{Ci/g}$ = $R_s \cdot 1000\text{mL/L} \cdot \text{DF} / (\text{EFF} \cdot \text{SS} \cdot \text{Dg/L} \cdot 2220000\text{dpm}/\mu\text{Ci})$		
11/02/94			
Time	Relative Counting Error = $[(\text{The Square Root of } \text{TC} + \text{BKG} \cdot \text{CT}) / (\text{TC} - \text{BKG} \cdot \text{CT})] \cdot 1.96 \cdot 100$		
03:00 PM	Detection Levels and Less Than Values are determined from Procedure LA-508-002.		

v RESULTS v

ALPHA TOTAL	in $\mu\text{Ci/g}$ (Average)	=	5.83E-01	DETECTION LEVEL
				7.26E-03 $\mu\text{Ci/g}$
RELATIVE COUNTING ERROR		=	6.7%	

Data Entry by:	<i>[Signature]</i>	Date:	11/03/94
Approved by:	<i>[Signature]</i>	Date:	11/3/94

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

WHC-SD-WM-DP-076, REV. 0

AT : LA-508-101 (D-2) SPIKED SAMPLE		SPIKE	REPLICATE
Type	DETECTOR NUMBER	16	16
SPIKE	DISH SIZE 1, 2, or 5 (MS)	2	2
Work List	TOTAL COUNTS (TC)	43264	37915
139	COUNT TIME in MINUTES (CT)	30	30
AT or TB ?	BACKGROUND in cpm (BKG)	0.1	0.1
AT	SAMPLE VOLUME in mL (Spiked Vial) (SS)	0.050	0.050
Test Code	SAMPLE DILUTION FACTOR (Spiked Vial) (DF)	1	1
ALPHA01	DIGEST GRAMS of SOLIDS/L (Dg/L)	2.1175	2.1175
Matrix	SPIKE VOLUME in mL (SVol)	0.100	0.100
SOLID	SPIKE DILUTION FACTOR (SDF)	1	1
Sample #	SPIKE VALUE in $\mu\text{Ci/L}$ (SVal)	36.4	36.4
S94T000142	INSTRUMENT EFFICIENCY FACTOR (EFF)	0.2104	0.2104
Instrument Code	SAMPLE + SPIKE $\mu\text{Ci/g}$ (S+S)	2.92E+01	2.56E+01
WB27806	AVERAGE or MAXIMUM $\mu\text{Ci/g}$ from FORM C	5.9600E-01	
Analyst	$R_s \text{ (Sample Count Rate)} = (TC / CT) - BKG$		
JMV	$\text{SAMPLE + SPIKE } \mu\text{Ci/g} = R_s * 1000\text{mL/L} * DF / (EFF * SS * Dg/L * 2220000\text{dpm}/\mu\text{Ci})$		
Date	$\text{PERCENT SPIKE RECOVERY} = (((S+S \mu\text{Ci/g}) - \text{SAMPLE } \mu\text{Ci/g}) * ((SDF/SVol)/(DF/SS/Dg/L)))/(SVal)*100$		
11/02/94			
Time			
03:00 PM			

RESULT **AVG. PERCENT SPIKE RECOVERY = 77.8%**

Data Entry by: <i>[Signature]</i>	Date: 03-Nov-94
Approved by: <i>[Signature]</i>	Date: 11/3/94

Form 508101\_X Rev. 1.1 Page 1 of 1

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

WHC-SD-WM-DP-076, REV. 0

AT : LA-508-101 (D-2) SOLIDS

		SAMPLE	REPLICATE
Type	DETECTOR NUMBER	16	16
SAMPLE	DISH SIZE 1, 2, or 5 (MS)	2	2
Work List	TOTAL COUNTS (TC)	487	522
139	COUNT TIME in MINUTES (CT)	30	30
AT or TB?	BACKGROUND in cpm (BKG)	0.1	0.1
AT	SAMPLE SIZE in mL (SS)	0.050	0.050
Test Code	DILUTION FACTOR (DF)	1	1
ALPHA01	DIGEST GRAMS of SOLIDS/L (Dg/L)	2.1215	2.1215
Matrix	EFFICIENCY FACTOR (EFF)	0.2104	0.2104
SOLID	Lc, Rmax, or Rs, (SAMPLE RATE) as APPROPRIATE	16.133	17.300

Sample #	Sample Concentration in $\mu\text{Ci/g}$	3.26E-01
S94T000144	Replicate Concentration in $\mu\text{Ci/g}$	3.49E-01
Instrument Code		
WB27806	Average Concentration in $\mu\text{Ci/g}$	3.3739E-01

Analyst	
JMV	$R_s$ (Sample Count Rate) = (TC / CT) - BKG
Date	ALPHA TOTAL $\mu\text{Ci/g}$ = $R_s * 1000\text{mL/L} * DF / (EFF * SS * Dg/L * 2220000\text{dpm}/\mu\text{Ci})$
11/02/94	
Time	Relative Counting Error = $[(\text{The Square Root of } TC + BKG * CT) / (TC - BKG * CT)] * 1.96 * 100$
03:00 PM	Detection Levels and Less Than Values are determined from Procedure LA-508-002.

v RESULTS v

ALPHA TOTAL in $\mu\text{Ci/g}$ (Average)	=	3.37E-01	DETECTION LEVEL
			7.27E-03 $\mu\text{Ci/g}$
RELATIVE COUNTING ERROR	=	9.0%	

Data Entry by:	<i>[Signature]</i>	Date:	11/03/94
Approved by:	<i>[Signature]</i>	Date:	11/3/94



PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

WHC-SD-WM-DP-076, REV. 0

**AT : LA-508-101 (D-2) SOLIDS**

		SAMPLE	REPLICATE
Type	DETECTOR NUMBER	16	16
DUPLICATE	DISH SIZE 1, 2, or 5 (MS)	2	2
Work List	TOTAL COUNTS (TC)	298	312
139	COUNT TIME in MINUTES (CT)	30	30
AT or TB ?	BACKGROUND in cpm (BKG)	0.1	0.1
AT	SAMPLE SIZE in mL (SS)	0.050	0.050
Test Code	DILUTION FACTOR (DF)	1	1
ALPHA01	DIGEST GRAMS of SOLIDS/L (Dg/L)	2.119	2.119
Matrix	EFFICIENCY FACTOR (EFF)	0.2104	0.2104
SOLID	Lc, Rmax, or Rs,(SAMPLE RATE) as APPROPRIATE	9.833	10.300
Sample #	Sample Concentration in $\mu\text{Ci/g}$	1.99E-01	
S94T000144	Replicate Concentration in $\mu\text{Ci/g}$	2.08E-01	
Instrument Code			
WB27806	Average Concentration in $\mu\text{Ci/g}$	2.0342E-01	
Analyst			
JMV	$R_s$ (Sample Count Rate) = (TC / CT) - BKG		
Date	ALPHA TOTAL $\mu\text{Ci/g}$ = $R_s * 1000\text{mL/L} * DF / (EFF * SS * Dg/L * 2220000\text{dpm}/\mu\text{Ci})$		
11/02/94			
Time	Relative Counting Error = $[(\text{The Square Root of } TC + BKG * CT) / (TC - BKG * CT)] * 1.96 * 100$		
03:00 PM	Detection Levels and Less Than Values are determined from Procedure LA-508-002.		

v RESULTS v

ALPHA TOTAL	in $\mu\text{Ci/g}$ (Average)	=	2.03E-01	DETECTION LEVEL
RELATIVE COUNTING ERROR		=	11.5%	

Data Entry by:	<i>[Signature]</i>	Date:	11/03/94
Approved by:	<i>[Signature]</i>	Date:	11/3/94

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

WHC-SD-WM-DP-076, REV. 0

AT : LA-508-101 (D-2) SPIKED SAMPLE

		SPIKE	REPLICATE
Type	DETECTOR NUMBER	16	16
SPIKE	DISH SIZE 1, 2, or 5 (MS)	2	2
Work List	TOTAL COUNTS (TC)	39274	33219
139	COUNT TIME in MINUTES (CT)	30	30
AT or TB ?	BACKGROUND in cpm (BKG)	0.1	0.1
AT	SAMPLE VOLUME in mL (Spiked Vial) (SS)	0.050	0.050
Test Code	SAMPLE DILUTION FACTOR (Spiked Vial) (DF)	1	1
ALPHA01	DIGEST GRAMS of SOLIDS/L (Dg/L)	2.1215	2.1215
Matrix	SPIKE VOLUME in mL (SVol)	0.100	0.100
SOLID	SPIKE DILUTION FACTOR (SDF)	1	1
Sample #	SPIKE VALUE in $\mu\text{Ci/L}$ (SVal)	36.4	36.4
S94T000144	INSTRUMENT EFFICIENCY FACTOR (EFF)	0.2104	0.2104
Instrument Code	SAMPLE + SPIKE $\mu\text{Ci/g}$ (S+S)	2.64E+01	2.23E+01
WB27806	AVERAGE or MAXIMUM $\mu\text{Ci/g}$ from FORM C	3.3700E-01	

Analyst	
JMV	
Date	
11/02/94	
Time	
03:00 PM	

$R_s$  (Sample Count Rate) =  $(TC / CT) - BKG$   
 $SAMPLE + SPIKE \mu\text{Ci/g} = R_s * 1000\text{mL/L} * DF / (EFF * SS * Dg/L * 2220000\text{dpm}/\mu\text{Ci})$   
 $PERCENT SPIKE RECOVERY = (((S+S \mu\text{Ci/g} - SAMPLE \mu\text{Ci/g}) * ((SDF/SVol)/(DF/SS/Dg/L)))/SVal)*100$

RESULT **AVG. PERCENT SPIKE RECOVERY = 70.1%**

Data Entry by: <i>MJB</i>	Date: 03-Nov-94
Approved by: <i>A. J. [Signature]</i>	Date: 11/3/94



# LABCORE Data Entry Template for Worklist# 114

Analyst: KRM Instrument: AB00 15 Method: LA-508-101 D-2  
**WHC-SD-WM-DP-076, REV. 0**  
 Worklist Comment: Run 1 STD for worlist. Determine SS using ludlum detector

Seg	Type	Sample#	Rep	Al	Test	Matrix	Actual	Found	DL	Unit
1	1	BLNK-PREP			@ALPHA01 ALPHA01	SOLID			N/A	uCi/c
2	1	BLNK-PREP			@ALPHA01 ALPHA01E	SOLID			N/A	% Ct.
3	2	STD			@ALPHA01 ALPHA01	SOLID			N/A	uCi/c
4	2	STD			@ALPHA01 ALPHA01E	SOLID			N/A	% Ct.
5	3	SAMPLE	S94T000142	0	F	@ALPHA01 ALPHA01	SOLID	N/A		uCi/c
6	3	SAMPLE	S94T000142	0	F	@ALPHA01 ALPHA01E	SOLID	N/A		% Ct.
7	4	DUP	S94T000142	0	F	@ALPHA01 ALPHA01	SOLID		N/A	uCi/c
8	4	DUP	S94T000142	0	F	@ALPHA01 ALPHA01E	SOLID		N/A	% Ct.
9-10	5	SPK	S94T000142	0	F	@ALPHA01 ALPHA01	SOLID		N/A	uCi/c
11	6	SAMPLE	S94T000144	0	F	@ALPHA01 ALPHA01	SOLID	N/A		uCi/c
12	6	SAMPLE	S94T000144	0	F	@ALPHA01 ALPHA01E	SOLID	N/A		% Ct.
13	7	DUP	S94T000144	0	F	@ALPHA01 ALPHA01	SOLID		N/A	uCi/c
14	7	DUP	S94T000144	0	F	@ALPHA01 ALPHA01E	SOLID		N/A	% Ct.
15-16	8	SPK	S94T000144	0	F	@ALPHA01 ALPHA01	SOLID		N/A	uCi/c
17	9	BLNK-PREP			@ALPHA01 ALPHA01	SOLID			N/A	uCi/c
18	9	BLNK-PREP			@ALPHA01 ALPHA01E	SOLID			N/A	% Ct.
—	10	STD			@ALPHA01 ALPHA01	SOLID			N/A	uCi/c
—	10	STD			@ALPHA01 ALPHA01E	SOLID			N/A	% Ct.
19	11	SAMPLE	S94T000149	0	F	@ALPHA01 ALPHA01	SOLID	N/A		uCi/c
20	11	SAMPLE	S94T000149	0	F	@ALPHA01 ALPHA01E	SOLID	N/A		% Ct.

Data Entry Comments:  
Rerun for samples S94T000142, 144 requested due to poor SPK  
recovery  
LAD 11-16-94



# LABCORE Data Entry Template for Worklist# 114

Analyst: KRM Instrument: AB00 \_\_\_\_\_ Method: LA-508-101  
**WHC-SD-WM-DP-076, REV. 0**

Worklist Comment: Run 1 STD for worlist. Determine SS using ludlum detector

Seg	Type	Sample#	Rep	Al	Test	Matrix	Actual	Found	DL	Unit
21	12 DUP	S94T000149	0	F	@ALPHA01 ALPHA01	SOLID	_____	_____	N/A	uCi/g
22	12 DUP	S94T000149	0	F	@ALPHA01 ALPHA01E	SOLID	_____	_____	N/A	% Ct.
23-24	13 SPK	S94T000149	0	F	@ALPHA01 ALPHA01	SOLID	_____	_____	N/A	uCi/g
25	14 SAMPLE	S94T000150	0	F	@ALPHA01 ALPHA01	SOLID	N/A	_____	_____	uCi/g
26	14 SAMPLE	S94T000150	0	F	@ALPHA01 ALPHA01E	SOLID	N/A	_____	_____	% Ct.
27	15 DUP	S94T000150	0	F	@ALPHA01 ALPHA01	SOLID	_____	_____	N/A	uCi/g
28	15 DUP	S94T000150	0	F	@ALPHA01 ALPHA01E	SOLID	_____	_____	N/A	% Ct.

**Final page for worklist # 114**

  
 \_\_\_\_\_  
 Analyst Signature  


10-30-94  
 \_\_\_\_\_  
 Date  
 10/31/94

Data Entry Comments:

---



---



---

D-2 LAD 11-16-94

LA-508-101 B-1 AT: LIQUIDS/SOLIDS

		STANDARD	REPLICATE
Type	DETECTOR NUMBER	15	
STANDARD	DISH SIZE 1, 2, or 5 (MS)	2	
Work List	TOTAL COUNTS (TC)	1454	1477
114	COUNT TIME in MINUTES (CT)	30	
AT or TB?	BACKGROUND in cpm (BKG)	03	
AT	SAMPLE SIZE in mL (SS)	10	
Test Code	DILUTION FACTOR (DF)	1	
Alpha 01	DIGEST DILUTION FACTOR (DDF)	1	
Matrix	EFFICIENCY FACTOR (EFF)		
LIQUID	lc, Rmax, or Rs (SAMPLE RATE) as APPROPRIATE		
Sample #	STANDARD BOOK #	9,546-3	61852
Instrument Code			
WB26872			
Analyst			
KR MONTETH			
Date			
10-30-94			
Time			
06:00			

3-4

WHC-SD-WM-DP-076, REV. 0



D-2 11-16-94

LA-508-101 D-1 AT: LIQUIDS/SOLIDS

		BLANK	REPLICATE
Type	DETECTOR NUMBER	15	
BLANK	DISH SIZE 1, 2, or 5 (MS)	2	
	TOTAL COUNTS (TC)	18	20
	COUNT TIME in MINUTES (CT)	30	
AT or TB?	BACKGROUND in cpm (BKG)	03	
AT 142/144	SAMPLE SIZE in mL (SS)	100	
	DILUTION FACTOR (DF)	1	
	DIGEST DILUTION FACTOR (DDF)	2.1175	
Matrix	EFFICIENCY FACTOR (EFF)		
<del>LIQUID</del> Solid	lc, Rmax, or Rs (SAMPLE RATE) as APPROPRIATE		

1-2

85

Data Entry by: 	Date: 10-30-94
Data Entry by: 	Date: 10-31-94

5-6

LA-508-101 D-1 <sup>LAD</sup> AT: <sup>11-16-94</sup> LIQUIDS/SOLIDS

		SAMPLE	REPLICATE
Type	DETECTOR NUMBER	15	
SAMPLE	DISH SIZE 1, 2, or 5 (MS)	2	
Work List	TOTAL COUNTS (TC)	2098	1805
114	COUNT TIME in MINUTES (CT)	30	
AT or TB ?	BACKGROUND in cpm (BKG)	.3	
AT	SAMPLE SIZE in mL (SS)	.100	
Test Code	DILUTION FACTOR (DF)	1	
Alpha 01	DIGEST GRAMS of SOLIDS/L (Dg/L)	2.1175	
Matrix	EFFICIENCY FACTOR (EFF)		
SOLID	I.e. Rmax, or Rs, (SAMPLE RATE) as APPROPRIATE		
Sample #			
594800142			
Instrument Code			
WB26872			
Analyst			
KR MONTEITH			
Date			
10-30-94			
Time			
06:00			

WHC-SD-WM-DP-076, REV. 0

7-8

LA-508-101 D-1 <sup>DJ</sup> AT: <sup>11-16-94</sup> LIQUIDS/SOLIDS

		DUPLICATE	REPLICATE
Type	DETECTOR NUMBER	15	
DUPLICATE	DISH SIZE 1, 2, or 5 (MS)	2	
	TOTAL COUNTS (TC)	1938	2168
	COUNT TIME in MINUTES (CT)	30	
AT or TB ?	BACKGROUND in cpm (BKG)	.3	
AT	SAMPLE SIZE in mL (SS)	.100	
	DILUTION FACTOR (DF)	1	
	DIGEST GRAMS of SOLIDS/L (Dg/L)	2.1235	
Matrix	EFFICIENCY FACTOR (EFF)		
SOLID	I.e. Rmax, or Rs, (SAMPLE RATE) as APPROPRIATE		

Data Entry by: <i>[Signature]</i>	Date: 10-30-94
Data Entry by: <i>[Signature]</i>	Date: 10-31-94

508101\_C

LA-508-101 D<sup>2</sup>-AT: SPIKED SAMPLE

		SPIKE	REPLICATE
Type	DETECTOR NUMBER	15	
SPIKE	DISH SIZE 1, 2, or 5 (NS)	2	
Work List	TOTAL COUNTS (TC)	35230	34062
114	COUNT TIME in MINUTES (CT)	30	
AT or TB?	BACKGROUND in cpm (BKG)	.3	
AT	SAMPLE VOLUME in mL (Spiked Vial) (SS)	.100	
Test Code	SAMPLE DILUTION FACTOR (Spiked Vial) (DF)	1	
Alpha 01	DIGEST GRAMS of SOLIDS/L (Dg/L)	2.1175	
Matrix	SPIKE VOLUME in mL (SVol)	.100	
solids	SPIKE DILUTION FACTOR (SDF)	1	
Sample #	SPIKE VALUE in $\mu$ Ci/L (SVal)	36.4	
WB26872	INSTRUMENT EFFICIENCY FACTOR (EFF)		
Instrument Code	SAMPLE + SPIKE $\mu$ Ci/g (S+S)		
S94T000142	AVERAGE or MAXIMUM $\mu$ Ci/g from FORM C		
Analyst	SPIKE BOOK #	93B43	
KR MONTETH			
Date			
10-30-94			
Time			
06100			

WHC-SD-WM-DP-076, REV. 0

9-10

Data Entry by: \_\_\_\_\_ Date: \_\_\_\_\_

508101\_X

LA-508-101 D<sup>2</sup>-TB: SPIKED SAMPLE

		SPIKE	REPLICATE
Type	DETECTOR NUMBER		
SPIKE	DISH SIZE 1, 2, or 5 (NS)		
	TOTAL COUNTS (TC)		
	COUNT TIME in MINUTES (CT)		
	BACKGROUND in cpm (BKG)		
	SAMPLE VOLUME in mL (Spiked Vial) (SS)		
	SAMPLE DILUTION FACTOR (Spiked Vial) (DF)		
	DIGEST GRAMS of SOLIDS/L (Dg/L)		
	SPIKE VOLUME in mL (SVol)		
	SPIKE DILUTION FACTOR (SDF)		
	SPIKE VALUE in $\mu$ Ci/L (SVal)		
	INSTRUMENT EFFICIENCY FACTOR (EFF)		
	SAMPLE + SPIKE $\mu$ Ci/g (S+S)		
	AVERAGE or MAXIMUM $\mu$ Ci/g from FORM C		
	SPIKE BOOK #		

Data Entry by: *[Signature]* Date: 10-30-94

Data Entry by: *[Signature]* Date: 10-31-94

508101\_X

87

LAD 11-16-94

LA-508-101 B-102

AT: LIQUIDS/SOLIDS

SAMPLE REPLICATE

Type	DETECTOR NUMBER	15	
SAMPLE	DISH SIZE 1, 2, or 5 (MS)	2	
Work List	TOTAL COUNTS (TC)	987	1002
114	COUNT TIME in MINUTES (CT)	30	
AT or TB?	BACKGROUND in cpm (BKG)	.3	
AT	SAMPLE SIZE in mL (SS)	.100	
Test Code	DILUTION FACTOR (DF)	1	
Alpha 01	DIGEST GRAMS of SOLIDS/L (Dg/L)	2.1215	
Matrix	EFFICIENCY FACTOR (EFF)		
SOLID	I.e. Rmax, or Rs, (SAMPLE RATE) as APPROPRIATE		
Sample #			
S94T000044			
Instrument Code			
W826872			
Analyst			
KR MONTEITH			
Date			
10-30-94			
Time			
06:00			

WHC-SD-WM-DP-076, REV. 0

11-12

D-2 11-16-94

LA-508-101 B-1

AT: LIQUIDS/SOLIDS

DUPLICATE REPLICATE

Type	DETECTOR NUMBER	15	
DUPLICATE	DISH SIZE 1, 2, or 5 (MS)	2	
	TOTAL COUNTS (TC)	779	734
	COUNT TIME in MINUTES (CT)	30	
AT or TB?	BACKGROUND in cpm (BKG)	.3	
AT	SAMPLE SIZE in mL (SS)	.100	
	DILUTION FACTOR (DF)	1	
	DIGEST GRAMS of SOLIDS/L (Dg/L)	2.1190	
Matrix	EFFICIENCY FACTOR (EFF)		
SOLID	I.e. Rmax, or Rs, (SAMPLE RATE) as APPROPRIATE		

13-14

Data Entry by:

*[Signature]*

Date: 10-30-94

Data Entry by:

*[Signature]*

Date: 10-31-94

508101\_C

LA-508-101 D-2

SPIKE	DETECTOR NUMBER	1, 2, or 5	(NS)	2
Work List	TOTAL COUNTS		(TC)	47501
114	COUNT TIME in MINUTES		(CT)	30
AT of TB	BACKGROUND in cpm		(BKG)	03
AT	SAMPLE VOLUME in mL (Spiked Vial)		(SS)	.100
Test Code	SAMPLE DILUTION FACTOR (Spiked Vial)		(DF)	1
Alpha 01	DIGEST GRAMS of SOLIDS/L		(Dg/L)	2.1215
Matrix	SPIKE VOLUME in mL		(SVol)	.100
solid	SPIKE DILUTION FACTOR		(SDF)	1
Sample #	SPIKE VALUE in $\mu\text{Ci/L}$		(SVal)	36.4
5947000144	INSTRUMENT EFFICIENCY FACTOR		(EFF)	
Instrument Code	SAMPLE + SPIKE $\mu\text{Ci/g}$		(S+S)	
WB26872	AVERAGE or MAXIMUM $\mu\text{Ci/g}$ from FORM C			93843
Analyst	SPIKE BOOK #			93843
KR MONTETH				
Date				
10-30-94				
Time				
06:00				

WHC-SD-WM-DP-076, REV. 0

15-16

10-30-94

Data Entry by: \_\_\_\_\_ Date: \_\_\_\_\_  
508101\_X

LA-508-101 D-2<sup>2</sup>: SPIKED SAMPLE

Type	DETECTOR NUMBER	SPIKE	REPLICATE
SPIKE	DISH SIZE 1, 2, or 5 (NS)		
	TOTAL COUNTS (TC)		
	COUNT TIME in MINUTES (CT)		
	BACKGROUND in cpm (BKG)		
	SAMPLE VOLUME in mL (Spiked Vial) (SS)		
	SAMPLE DILUTION FACTOR (Spiked Vial) (DF)		
	DIGEST GRAMS of SOLIDS/L (Dg/L)		
	SPIKE VOLUME in mL (SVol)		
	SPIKE DILUTION FACTOR (SDF)		
	SPIKE VALUE in $\mu\text{Ci/L}$ (SVal)		
	INSTRUMENT EFFICIENCY FACTOR (EFF)		
	SAMPLE + SPIKE $\mu\text{Ci/g}$ (S+S)		
	AVERAGE or MAXIMUM $\mu\text{Ci/g}$ from FORM C		
	SPIKE BOOK #		

Data Entry by: *[Signature]* Date: 10-30-94  
Data Entry by: *[Signature]* Date: 11-31-94

508101\_X



LAD 11-16-94  
LA-508-101 D-1 D-2

AT: LIQUIDS/SOLIDS		STANDARD	REPLICATE
Type	DETECTOR NUMBER		
STANDARD	DISH SIZE 1, 2, or 5 (MS)		
Work List	TOTAL COUNTS (TC)		
	COUNT TIME in MINUTES (CT)		
AT or TB ?	BACKGROUND in cpm (BKG)		
AT	SAMPLE SIZE in mL (SS)		
Test Code	DILUTION FACTOR (DF)		
	DIGEST DILUTION FACTOR (DDF)		
Matrix	EFFICIENCY FACTOR (EFF)		
LIQUID	Ic, Rmax, or Rs, (SAMPLE RATE) as APPROPRIATE.		
Sample #	STANDARD BOOK #		
Instrument Code			
Analyst			
Date			
Time			

WHC-SD-WM-DP-076, REV. 0

17-18

AT: LIQUIDS/SOLIDS		BLANK	REPLICATE
Type	DETECTOR NUMBER	15	
BLANK	DISH SIZE 1, 2, or 5 (MS)	2	
	TOTAL COUNTS (TC)	3	4
	COUNT TIME in MINUTES (CT)	30	
AT or TB ?	BACKGROUND in cpm (BKG)	0.3	
AT 149 / 150	SAMPLE SIZE in mL (SS)	.100	
	DILUTION FACTOR (DF)	1	
	DIGEST DILUTION FACTOR (DDF)	2.2860	
Matrix	EFFICIENCY FACTOR (EFF)		
LIQUID - Solid	Ic, Rmax, or Rs, (SAMPLE RATE) as APPROPRIATE.		

Data Entry by:  Date: 10-30-94  
 Data Entry by:  Date: 10-31-94

508101\_C



LA-508-101-D-2 AT:

LIQUIDS/SOLIDS

SAMPLE REPLICATE

Type	DETECTOR NUMBER	(MS)	15	
SAMPLE	DISH SIZE 1, 2, or 5	(MS)	2	
Work List	TOTAL COUNTS	(TC)	42	59
114	COUNT TIME in MINUTES	(CT)	30	
AT or TB?	BACKGROUND in cpm	(BKG)	.3	
AT	SAMPLE SIZE in mL	(SS)	.100	
Test Code	DILUTION FACTOR	(DF)	1	
Alpha 01	DIGEST GRAMS of SOLIDS/L	(Dg/L)	2.2860	
Matrix	EFFICIENCY FACTOR	(EFF)		
SOLID	I.e., Rmax, or Rs, (SAMPLE RATE) as APPROPRIATE			
Sample #				
5947000149				
Instrument Code				
WB26872				
Analyst				
KR MONTETH				
Date				
10-30-94				
Time				
0600				

19-20

WHC-SD-WM-DP-076, REV. 0

LA-508-101-D-1 AT: <sup>D2 11/6/94</sup>

LIQUIDS/SOLIDS

DUPLICATE REPLICATE

Type	DETECTOR NUMBER	(MS)	15	
DUPLICATE	DISH SIZE 1, 2, or 5	(MS)	2	
	TOTAL COUNTS	(TC)	57	46
	COUNT TIME in MINUTES	(CT)	30	
AT or TB?	BACKGROUND in cpm	(BKG)	.3	
AT	SAMPLE SIZE in mL	(SS)	.100	
	DILUTION FACTOR	(DF)	1	
	DIGEST GRAMS of SOLIDS/L	(Dg/L)	2.2935	
Matrix	EFFICIENCY FACTOR	(EFF)		
SOLID	I.e., Rmax, or Rs, (SAMPLE RATE) as APPROPRIATE			

21-22

Data Entry by: *[Signature]* Date: 10-30-94  
 Data Entry by: *[Signature]* Date: 10/31/94

508101\_C

LA-508-101 D-<sup>2</sup> AT: SPIKED SAMPLE

		SPIKE	REPLICATE
Type	DETECTOR NUMBER	15	
SPIKE	DISH SIZE 1, 2, or 5 (MS)	2	
Work List	TOTAL COUNTS (TC)	47068	50883
114	COUNT TIME in MINUTES (CT)	30	
AT or TB?	BACKGROUND in cpm (BKG)	3	
AT	SAMPLE VOLUME in mL (Spiked Vial) (SS)	.100	
Test Code	SAMPLE DILUTION FACTOR (Spiked Vial) (DF)	1	
Alpha 01	DIGEST GRAMS of SOLIDS/L (Dg/L)	2.2860	
Matrix	SPIKE VOLUME in mL (SVol)	.100	
Solid	SPIKE DILUTION FACTOR (SDF)	1	
Sample #	SPIKE VALUE in $\mu$ Ci/L (SVal)	36.4	
S94T000149	INSTRUMENT EFFICIENCY FACTOR (EFF)		
Instrument Code	SAMPLE + SPIKE $\mu$ Ci/g (S+S)		
WB26872	AVERAGE or MAXIMUM $\mu$ Ci/g from FORM C		
Analyst	SPIKE BOOK #	43843	
KR/WATETH			
Date			
10-30-94			
Time			
06:00			

WHC-SD-WM-DP-076, REV. 0



23-24

Data Entry by: \_\_\_\_\_ Date: \_\_\_\_\_

508101\_X

LA-508-101 D-<sup>2</sup> TB: SPIKED SAMPLE

		SPIKE	REPLICATE
Type	DETECTOR NUMBER		
SPIKE	DISH SIZE 1, 2, or 5 (MS)		
	TOTAL COUNTS (TC)		
	COUNT TIME in MINUTES (CT)		
	BACKGROUND in cpm (BKG)		
	SAMPLE VOLUME in mL (Spiked Vial) (SS)		
	SAMPLE DILUTION FACTOR (Spiked Vial) (DF)		
	DIGEST GRAMS of SOLIDS/L (Dg/L)		
	SPIKE VOLUME in mL (SVol)		
	SPIKE DILUTION FACTOR (SDF)		
	SPIKE VALUE in $\mu$ Ci/L (SVal)		
	INSTRUMENT EFFICIENCY FACTOR (EFF)		
	SAMPLE + SPIKE $\mu$ Ci/g (S+S)		
	AVERAGE or MAXIMUM $\mu$ Ci/g from FORM C		
	SPIKE BOOK #		

Data Entry by:  Date: 10-30-94  
 Data Entry by:  Date: 10/31/94

508101\_X

D-2 LAB 11/10/94

LA-508-101 D-1 AT: LIQUIDS/SOLIDS

		SAMPLE	REPLICATE
Type	DETECTOR NUMBER	15	
SAMPLE	DISH SIZE 1, 2, or 5 (MS)	2	
Work List	TOTAL COUNTS (TC)	63	54
114	COUNT TIME in MINUTES (CT)	30	
AT or TB?	BACKGROUND in cpm (DKG)	.3	
AT	SAMPLE SIZE in mL (SS)	.100	
Test Code	DILUTION FACTOR (DF)	1	
Alpha 01	DIGEST GRAMS of SOLIDS/L (Dg/L)	2.3725	
Matrix	EFFICIENCY FACTOR (EFF)		
SOLID	1c, Rmax, or Its (SAMPLE RATE) as APPROPRIATE		
Sample #			
S94T000150			
Instrument Code			
W326872			
Analyst			
KRMONTETH			
Date			
10-30-94			
Time			
06:00			

WHC-SD-WM-DP-076, REV. 0

25-26

D-2 LAB 11-16-94 LAD

LA-508-101 D-1 AT: LIQUIDS/SOLIDS

		DUPLICATE	REPLICATE
Type	DETECTOR NUMBER	15	
DUPLICATE	DISH SIZE 1, 2, or 5 (MS)	2	
	TOTAL COUNTS (TC)	42	49
	COUNT TIME in MINUTES (CT)	30	
AT or TB?	BACKGROUND in cpm (DKG)	.3	
AT	SAMPLE SIZE in mL (SS)	.100	
	DILUTION FACTOR (DF)	1	
	DIGEST GRAMS of SOLIDS/L (Dg/L)	2.3725	
Matrix	EFFICIENCY FACTOR (EFF)		
SOLID	1c, Rmax, or Its (SAMPLE RATE) as APPROPRIATE		

27-28

Data Entry by: <i>M. G. Sullivan</i>	Date: 1-13-99
Data Entry by:	Date:

508101\_C

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

WHC-SD-WM-DP-076, REV. 0

AT : LA-508-101 (D-2) LIQUIDS

		STANDARD	REPLICATE
Type	DETECTOR NUMBER	15	15
STANDARD	DISH SIZE 1, 2, or 5 (MS)	2	2
Work List	TOTAL COUNTS (TC)	1454	1477
114	COUNT TIME in MINUTES (CT)	30	30
AT or TB ?	BACKGROUND in cpm (BKG)	0.3	0.3
AT	SAMPLE SIZE in mL (SS)	10.000	10.000
Test Code	DILUTION FACTOR (DF)	1	1
ALPHA01	DIGEST DILUTION FACTOR (DDF)	1	1
Matrix	EFFICIENCY FACTOR (EFF)	0.2380	0.2380
LIQUID	Lc, Rmax, or Rs, (SAMPLE RATE) as APPROPRIATE	48.167	48.933
Sample #	Sample Concentration in $\mu\text{Ci/L}$	9.12E-03	
	Replicate Concentration in $\mu\text{Ci/L}$	9.26E-03	
Instrument Code			
WB26872	Average Concentration in $\mu\text{Ci/L}$	9.1888E-03	
Analyst			
KRM	$R_s$ (Sample Count Rate) = $(TC / CT) - BKG$		
Date	ALPHA TOTAL $\mu\text{Ci/L}$ = $R_s * 1000\text{mL/L} * DF * DDF / (EFF * SS * 2220000\text{dpm}/\mu\text{Ci})$		
10/30/94	ALPHA TOTAL $\mu\text{Ci/mL}$ = ALPHA TOTAL $\mu\text{Ci/L} / 1000\text{mL/L}$		
Time	Relative Counting Error = $[ \{ (\text{The Square Root of } TC + BKG * CT) / (TC - BKG * CT) \} ] * 1.96 * 100$		
06:00 AM	Detection Levels and Less Than Values are determined from Procedure LA-508-002.		

RESULTS

ALPHA TOTAL	in $\mu\text{Ci/mL}$	(Average) =	9.19E-06	DETECTION LEVEL
RELATIVE COUNTING ERROR		=	5.2%	$\mu\text{Ci/mL}$

Data Entry by:	<i>M. Sp.../jt</i>	Date:	10/31/94
Approved by:	<i>[Signature]</i>	Date:	10/31/94

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

WHC-SD-WM-DP-076, REV. 0

AT : LA-508-101 (D-2) SOLIDS

		BLANK	REPLICATE
Type	DETECTOR NUMBER	15	15
BLANK	DISH SIZE 1, 2, or 5 (MS)	2	2
Work List	TOTAL COUNTS (TC)	18	20
114	COUNT TIME in MINUTES (CT)	30	30
AT or TB ?	BACKGROUND in cpm (BKG)	0.3	0.3
AT	SAMPLE SIZE in mL (SS)	0.100	0.100
Test Code	DILUTION FACTOR (DF)	1	1
ALPHA01	DIGEST GRAMS of SOLIDS/L (Dg/L)	2.1175	2.1175
Matrix	EFFICIENCY FACTOR (EFF)	0.2380	0.2380
SOLID	Lc, Rmax, or Rs,(SAMPLE RATE) as APPROPRIATE	0.300	0.367
Sample #	Sample Concentration in $\mu\text{Ci/g}$	2.68E-03	
142,144	Replicate Concentration in $\mu\text{Ci/g}$	3.28E-03	
Instrument Code			
WB28872	Average Concentration in $\mu\text{Ci/g}$	2.9794E-03	
Analyst			
KRM	$R_s$ (Sample Count Rate) = $(TC / CT) - BKG$		
Date	ALPHA TOTAL $\mu\text{Ci/g}$ = $R_s \cdot 1000\text{mL/L} \cdot DF / (EFF \cdot SS \cdot Dg/L \cdot 2220000\text{dpm}/\mu\text{Ci})$		
10/30/94			
Time	Relative Counting Error = $[(\text{The Square Root of } TC + BKG \cdot CT) / (TC - BKG \cdot CT)] \cdot 1.96 \cdot 100$		
06:00 AM	Detection Levels and Less Than Values are determined from Procedure LA-508-002.		

v RESULTS v

ALPHA TOTAL in $\mu\text{Ci/g}$ (Average)	=	2.98E-03	DETECTION LEVEL
RELATIVE COUNTING ERROR	=	113.2%	

Data Entry by: *[Signature]* Date: 10/31/94  
 Approved by: *[Signature]* Date: 10/31/94

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

WHC-SD-WM-DP-076, REV. 0

AT : LA-508-101 (D-2) SOLIDS

		SAMPLE	REPLICATE
Type	DETECTOR NUMBER	15	15
SAMPLE	DISH SIZE 1, 2, or 5 (MS)	2	2
Work List	TOTAL COUNTS (TC)	2098	1805
114	COUNT TIME in MINUTES (CT)	30	30
AT or TB ?	BACKGROUND in cpm (BKG)	0.3	0.3
AT	SAMPLE SIZE in mL (SS)	0.100	0.100
Test Code	DILUTION FACTOR (DF)	1	1
ALPHA01	DIGEST GRAMS of SOLIDS/L (Dg/L)	2.1175	2.1175
Matrix	EFFICIENCY FACTOR (EFF)	0.2380	0.2380
SOLID	Lc, Rmax, or Rs, (SAMPLE RATE) as APPROPRIATE	69.633	59.867
Sample #	Sample Concentration in $\mu\text{Ci/g}$	6.22E-01	
S94T000142	Replicate Concentration in $\mu\text{Ci/g}$	5.35E-01	
Instrument Code			
WB26872	Average Concentration in $\mu\text{Ci/g}$	5.7874E-01	
Analyst			
KRM	$R_s$ (Sample Count Rate) = (TC / CT) - BKG		
Date	ALPHA TOTAL $\mu\text{Ci/g}$ = $R_s \cdot 1000\text{mL/L} \cdot \text{DF} / (\text{EFF} \cdot \text{SS} \cdot \text{Dg/L} \cdot 2220000\text{dpm}/\mu\text{Ci})$		
10/30/94			
Time	Relative Counting Error = $[(\text{The Square Root of } \text{TC} + \text{BKG} \cdot \text{CT}) / (\text{TC} - \text{BKG} \cdot \text{CT})] \cdot 1.96 \cdot 100$		
06:00 AM	Detection Levels and Less Than Values are determined from Procedure LA-508-002.		

v RESULTS v

ALPHA TOTAL	in $\mu\text{Ci/g}$ (Average)	=	5.79E-01	DETECTION LEVEL
				4.98E-03 $\mu\text{Ci/g}$
RELATIVE COUNTING ERROR		=	4.6%	

Data Entry by:	<i>[Signature]</i>	Date:	10/31/94
Approved by:	<i>[Signature]</i>	Date:	10/31/94



PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

WHC-SD-WM-DP-076, REV. 0

**AT : LA-508-101 (D-2) SOLIDS**

		SAMPLE	REPLICATE
Type	DETECTOR NUMBER	15	15
DUPLICATE	DISH SIZE 1, 2, or 5 (MS)	2	2
Work List	TOTAL COUNTS (TC)	1938	2168
114	COUNT TIME in MINUTES (CT)	30	30
AT or TB ?	BACKGROUND in cpm (BKG)	0.3	0.3
AT	SAMPLE SIZE in mL (SS)	0.100	0.100
Test Code	DILUTION FACTOR (DF)	1	1
ALPHA01	DIGEST GRAMS of SOLIDS/L (Dg/L)	2.1235	2.1235
Matrix	EFFICIENCY FACTOR (EFF)	0.2380	0.2380
SOLID	Lc, Rmax, or Rs,(SAMPLE RATE) as APPROPRIATE	64.300	71.967
Sample #	Sample Concentration in $\mu\text{Ci/g}$	5.73E-01	
S94T000142	Replicate Concentration in $\mu\text{Ci/g}$	6.41E-01	
Instrument Code			
WB26872	Average Concentration in $\mu\text{Ci/g}$	6.0726E-01	
Analyst			
KRM	$R_s$ (Sample Count Rate) = (TC / CT) - BKG		
Date	ALPHA TOTAL $\mu\text{Ci/g}$ = $R_s \cdot 1000\text{mL/L} \cdot \text{DF} / (\text{EFF} \cdot \text{SS} \cdot \text{Dg/L} \cdot 2220000\text{dpm}/\mu\text{Ci})$		
10/30/94			
Time	Relative Counting Error = $[(\text{The Square Root of } \text{TC} + \text{BKG} \cdot \text{CT}) / (\text{TC} - \text{BKG} \cdot \text{CT})] \cdot 1.96 \cdot 100$		
06:00 AM	Detection Levels and Less Than Values are determined from Procedure LA-508-002.		

**v RESULTS v**

ALPHA TOTAL in $\mu\text{Ci/g}$ (Average)	=	6.07E-01	DETECTION LEVEL
RELATIVE COUNTING ERROR	=	4.5%	4.97E-03 $\mu\text{Ci/g}$

Data Entry by: *[Signature]* Date: 10/31/94  
 Approved by: *[Signature]* Date: 10/31/94

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

WHC-SD-WM-DP-076, REV. 0

**AT : LA-508-101 (D-2) SPIKED SAMPLE**

		SPIKE	REPLICATE
Type	DETECTOR NUMBER	15	15
SPIKE	DISH SIZE 1, 2, or 5 (MS)	2	2
Work List	TOTAL COUNTS (TC)	35230	34062
114	COUNT TIME in MINUTES (CT)	30	30
AT or IB ?	BACKGROUND in cpm (BKG)	0.3	0.3
AT	SAMPLE VOLUME in mL (Spiked Vial) (SS)	0.100	0.100
Test Code	SAMPLE DILUTION FACTOR (Spiked Vial) (DF)	1	1
ALPHA01	DIGEST GRAMS of SOLIDS/L (Dg/L)	2.1175	2.1175
Matrix	SPIKE VOLUME in mL (SVol)	0.100	0.100
SOLID	SPIKE DILUTION FACTOR (SDF)	1	1
Sample #	SPIKE VALUE in $\mu\text{Ci/L}$ (SVal)	36.4	36.4
S94T000142	INSTRUMENT EFFICIENCY FACTOR (EFF)	0.238	0.238
Instrument Code	SAMPLE + SPIKE $\mu\text{Ci/g}$ (S+S)	1.05E+01	1.01E+01
WB26872	AVERAGE or MAXIMUM $\mu\text{Ci/g}$ from FORM C	5.7900E-01	
Analyst			
KRM	Rs (Sample Count Rate) = (TC / CT) - BKG		
Date	SAMPLE + SPIKE $\mu\text{Ci/g}$ = $Rs * 1000\text{mL/L} * DF / (EFF * SS * Dg/L * 2220000\text{dpm}/\mu\text{Ci})$		
10/30/94	PERCENT SPIKE RECOVERY = $((S+S \mu\text{Ci/g} - \text{SAMPLE } \mu\text{Ci/g}) * ((SDF/SVol)/(DF/SS/Dg/L)))/(SVal)*100$		
Time			
06:00 AM			

RESULT **AVG. PERCENT SPIKE RECOVERY = 56.7%**

Data Entry by: <i>[Signature]</i>	Date: 31-Oct-94
Approved by: <i>[Signature]</i>	Date: 10/31/94

Form 508101\_X Rev. 1.1 Page 1 of 1



PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

WHC-SD-WM-DP-076, REV. 0

**AT : LA-508-101 (D-2) SOLIDS**

		SAMPLE	REPLICATE
Type	DETECTOR NUMBER	15	15
SAMPLE	DISH SIZE 1, 2, or 5 (MS)	2	2
Work List	TOTAL COUNTS (TC)	987	1002
114	COUNT TIME in MINUTES (CT)	30	30
AT or TB ?	BACKGROUND in cpm (BKG)	0.3	0.3
AT	SAMPLE SIZE in mL (SS)	0.100	0.100
Test Code	DILUTION FACTOR (DF)	1	1
ALPHA01	DIGEST GRAMS of SOLIDS/L (Dg/L)	2.1215	2.1215
Matrix	EFFICIENCY FACTOR (EFF)	0.2380	0.2380
SOLID	Lc, Rmax, or Rs,(SAMPLE RATE) as APPROPRIATE	32.600	33.100
Sample #	Sample Concentration in $\mu\text{Ci/g}$	2.91E-01	
S94T000144	Replicate Concentration in $\mu\text{Ci/g}$	2.95E-01	
Instrument Code			
WB26872	Average Concentration in $\mu\text{Ci/g}$	2.9306E-01	
Analyst			
KRM	$R_s$ (Sample Count Rate) = (TC / CT) - BKG		
Date	ALPHA TOTAL $\mu\text{Ci/g}$ = $R_s \cdot 1000\text{mL/L} \cdot \text{DF} / (\text{EFF} \cdot \text{SS} \cdot \text{Dg/L} \cdot 2220000\text{dpm}/\mu\text{Ci})$		
10/30/94			
Time	Relative Counting Error = $[(\text{The Square Root of } \text{TC} + \text{BKG} \cdot \text{CT}) / (\text{TC} - \text{BKG} \cdot \text{CT})] \cdot 1.96 \cdot 100$		
06:00 AM	Detection Levels and Less Than Values are determined from Procedure LA-508-002.		

**v RESULTS v**

ALPHA TOTAL in $\mu\text{Ci/g}$ (Average)	=	2.93E-01	DETECTION LEVEL
RELATIVE COUNTING ERROR	=	6.3%	$\mu\text{Ci/g}$

Data Entry by: <i>M. V. [Signature]</i>	Date: 10/31/94
Approved by: <i>[Signature]</i>	Date: 10/31/94

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

WHC-SD-WM-DP-076, REV. 0

AT : LA-508-101 (D-2) SOLIDS

		SAMPLE	REPLICATE
Type	DETECTOR NUMBER	15	15
DUPLICATE	DISH SIZE 1, 2, or 5 (MS)	2	2
Work List	TOTAL COUNTS (TC)	779	734
114	COUNT TIME in MINUTES (CT)	30	30
AT or TB?	BACKGROUND in cpm (BKG)	0.3	0.3
AT	SAMPLE SIZE in mL (SS)	0.100	0.100
Test Code	DILUTION FACTOR (DF)	1	1
ALPHA01	DIGEST GRAMS of SOLIDS/L (Dg/L)	2.119	2.119
Matrix	EFFICIENCY FACTOR (EFF)	0.2380	0.2380
SOLID	Lc, Rmax, or Rs,(SAMPLE RATE) as APPROPRIATE	25.667	24.167
Sample #	Sample Concentration in $\mu\text{Ci/g}$	2.29E-01	
S94T000144	Replicate Concentration in $\mu\text{Ci/g}$	2.16E-01	
Instrument Code			
WB26872	Average Concentration in $\mu\text{Ci/g}$	2.2255E-01	
Analyst			
KRM	$R_s$ (Sample Count Rate) = (TC / CT) - BKG		
Date	ALPHA TOTAL $\mu\text{Ci/g}$ = $R_s \cdot 1000\text{mL/L} \cdot \text{DF} / (\text{EFF} \cdot \text{SS} \cdot \text{Dg/L} \cdot 2220000\text{dpm}/\mu\text{Ci})$		
10/30/94			
Time	Relative Counting Error = $[(\text{The Square Root of } \text{TC} + \text{BKG} \cdot \text{CT}) / (\text{TC} - \text{BKG} \cdot \text{CT})] \cdot 1.96 \cdot 100$		
06:00 AM	Detection Levels and Less Than Values are determined from Procedure LA-508-002.		

v RESULTS v

ALPHA TOTAL	in $\mu\text{Ci/g}$ (Average)	=	2.23E-01	DETECTION LEVEL
RELATIVE COUNTING ERROR		=	7.4%	$\mu\text{Ci/g}$

Data Entry by: *[Signature]* Date: 10/31/94  
 Approved by: *[Signature]* Date: 10/31/94

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

WHC-SD-WM-DP-076, REV. 0

**AT : LA-508-101 (D-2) SPIKED SAMPLE**

		SPIKE	REPLICATE
Type	DETECTOR NUMBER	15	15
SPIKE	DISH SIZE 1, 2, or 5 (MS)	2	2
Work List	TOTAL COUNTS (TC)	47501	43668
114	COUNT TIME in MINUTES (CT)	30	30
AT or IB?	BACKGROUND in cpm (BKG)	0.3	0.3
AT	SAMPLE VOLUME in mL (Spiked Vial) (SS)	0.100	0.100
Test Code	SAMPLE DILUTION FACTOR (Spiked Vial) (DF)	1	1
ALPHA01	DIGEST GRAMS of SOLIDS/L (Dg/L)	2.1215	2.1215
Matrix	SPIKE VOLUME in mL (SVol)	0.100	0.100
SOLID	SPIKE DILUTION FACTOR (SDF)	1	1
Sample #	SPIKE VALUE in $\mu\text{Ci/L}$ (SVal)	36.4	36.4
S94T000144	INSTRUMENT EFFICIENCY FACTOR (EFF)	0.238	0.238
Instrument Code	SAMPLE + SPIKE $\mu\text{Ci/g}$ (S+S)	1.41E+01	1.30E+01
WB26872	AVERAGE or MAXIMUM $\mu\text{Ci/g}$ from FORM C	2.9306E-01	
Analyst			
KRM	$R_s$ (Sample Count Rate) = (TC / CT) - BKG		
Date	SAMPLE + SPIKE $\mu\text{Ci/g}$ = $R_s * 1000\text{mL/L} * DF / (EFF * SS * \text{Dg/L} * 2220000\text{dpm}/\mu\text{Ci})$		
10/30/94	PERCENT SPIKE RECOVERY = $((S+S \mu\text{Ci/g} - \text{SAMPLE } \mu\text{Ci/g}) * ((SDF/SVol)/(DF/SS/Dg/L)) / SVal) * 100$		
Time			
06:00 AM			

RESULT **AVG. PERCENT SPIKE RECOVERY = 77.3%**

Data Entry by: *[Signature]* Date: 31-Oct-94  
 Approved by: *[Signature]* Date: 10/31/94  
 Form 508101\_X Rev. 1.1

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

WHC-SD-WM-DP-076, REV. 0

**AT : LA-508-101 (D-2) SOLIDS**

		BLANK	REPLICATE
Type	DETECTOR NUMBER	15	15
BLANK	DISH SIZE 1, 2, or 5 (MS)	2	2
Work List	TOTAL COUNTS (TC)	3	4
114	COUNT TIME in MINUTES (CT)	30	30
AT or TB ?	BACKGROUND in cpm (BKG)	0.3	0.3
AT	SAMPLE SIZE in mL (SS)	0.100	0.100
Test Code	DILUTION FACTOR (DF)	1	1
ALPHA01	DIGEST GRAMS OF SOLIDS/L (Dg/L)	2.286	2.286
Matrix	EFFICIENCY FACTOR (EFF)	0.2380	0.2380
SOLID	Lc, Rmax, or Rs,(SAMPLE RATE) as APPROPRIATE	0.233	0.233
Sample #	Sample Concentration in $\mu\text{Ci/g}$	< 1.93E-03	
149,150	Replicate Concentration in $\mu\text{Ci/g}$	< 1.93E-03	
Instrument Code			
WB26872	Maximum Concentration in $\mu\text{Ci/g}$	< 1.9319E-03	
Analyst			
KRM	$R_s$ (Sample Count Rate) = (TC / CT) - BKG		
Date	ALPHA TOTAL $\mu\text{Ci/g}$ = $R_s \cdot 1000\text{mL/L} \cdot \text{DF} / (\text{EFF} \cdot \text{SS} \cdot \text{Dg/L} \cdot 2220000\text{dpm}/\mu\text{Ci})$		
10/30/94			
Time	Relative Counting Error = $[(\text{The Square Root of TC} + \text{BKG} \cdot \text{CT}) / (\text{TC} - \text{BKG} \cdot \text{CT})] \cdot 1.96 \cdot 100$		
06:00 AM	Detection Levels and Less Than Values are determined from Procedure LA-508-002.		

v RESULTS v

ALPHA TOTAL	in $\mu\text{Ci/g}$ (Maximum) =	< 1.93E-03	DETECTION LEVEL
LESS Than Value was Determined from Lc.			
RELATIVE COUNTING ERROR	=	141.3%	4.61E-03 $\mu\text{Ci/g}$

Data Entry by: *[Signature]* Date: 10/31/94  
 Approved by: *[Signature]* Date: 10/31/94

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

WHC-SD-WM-DP-076, REV. 0

**AT : LA-508-101 (D-2) SOLIDS**

		SAMPLE	REPLICATE
Type	DETECTOR NUMBER	15	15
SAMPLE	DISH SIZE 1, 2, or 5 (MS)	2	2
Work List	TOTAL COUNTS (TC)	42	59
114	COUNT TIME in MINUTES (CT)	30	30
AT or TB ?	BACKGROUND in cpm (BKG)	0.3	0.3
AT	SAMPLE SIZE in mL (SS)	0.100	0.100
Test Code	DILUTION FACTOR (DF)	1	1
ALPHA01	DIGEST GRAMS of SOLIDS/L (Dg/L)	2.286	2.286
Matrix	EFFICIENCY FACTOR (EFF)	0.2380	0.2380
SOLID	Lc, Rmax, or Rs,(SAMPLE RATE) as APPROPRIATE	1.100	1.667
Sample #	Sample Concentration in $\mu\text{Ci/g}$	9.11E-03	
S94T000149	Replicate Concentration in $\mu\text{Ci/g}$	1.38E-02	
Instrument Code			
WB26872	Average Concentration in $\mu\text{Ci/g}$	1.1453E-02	
Analyst			
KRM	$R_s$ (Sample Count Rate) = (TC / CT) - BKG		
Date	ALPHA TOTAL $\mu\text{Ci/g}$ = $R_s * 1000\text{mL/L} * DF / (EFF * SS * Dg/L * 2220000\text{dpm}/\mu\text{Ci})$		
10/30/94			
Time	Relative Counting Error = $[(\text{The Square Root of } TC + BKG * CT) / (TC - BKG * CT)] * 1.96 * 100$		
06:00 AM	Detection Levels and Less Than Values are determined from Procedure LA-508-002.		

▼ RESULTS ▼

ALPHA TOTAL	in $\mu\text{Ci/g}$ (Average)	=	1.15E-02	DETECTION LEVEL
RELATIVE COUNTING ERROR		=	42.4%	$\mu\text{Ci/g}$

Data Entry by: *[Signature]* Date: 10/31/94  
 Approved by: *[Signature]* Date: 10/31/94  
 Form 508101\_C Rev. 1.1 Page 1 of 1

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

WHC-SD-WM-DP-076, REV. 0

AT : LA-508-101 (D-2) SOLIDS

		SAMPLE	REPLICATE
Type	DETECTOR NUMBER	15	15
DUPLICATE	DISH SIZE 1, 2, or 5 (MS)	2	2
Work List	TOTAL COUNTS (TC)	57	46
114	COUNT TIME in MINUTES (CT)	30	30
AT or TB?	BACKGROUND in cpm (BKG)	0.3	0.3
AT	SAMPLE SIZE in mL (SS)	0.100	0.100
Test Code	DILUTION FACTOR (DF)	1	1
ALPHA01	DIGEST GRAMS of SOLIDS/L (Dg/L)	2.2935	2.2935
Matrix	EFFICIENCY FACTOR (EFF)	0.2380	0.2380
SOLID	Lc, Rmax, or Rs, (SAMPLE RATE) as APPROPRIATE	1.600	1.233
Sample #	Sample Concentration in $\mu\text{Ci/g}$	1.32E-02	
S94T000149	Replicate Concentration in $\mu\text{Ci/g}$	1.02E-02	
Instrument Code			
WB26872	Average Concentration in $\mu\text{Ci/g}$	1.1691E-02	
Analyst			
KRM	$R_s$ (Sample Count Rate) = (TC / CT) - BKG		
Date	ALPHA TOTAL $\mu\text{Ci/g}$ = $R_s \cdot 1000\text{mL/L} \cdot \text{DF} / (\text{EFF} \cdot \text{SS} \cdot \text{Dg/L} \cdot 2220000\text{dpm}/\mu\text{Ci})$		
10/30/94			
Time	Relative Counting Error = $[(\text{The Square Root of } \text{TC} + \text{BKG} \cdot \text{CT}) / (\text{TC} - \text{BKG} \cdot \text{CT})] \cdot 1.96 \cdot 100$		
06:00 AM	Detection Levels and Less Than Values are determined from Procedure LA-508-002.		

v RESULTS v

ALPHA TOTAL	in $\mu\text{Ci/g}$ (Average)	=	1.17E-02	DETECTION LEVEL
RELATIVE COUNTING ERROR		=	39.3%	4.60E-03 $\mu\text{Ci/g}$

Data Entry by:	<i>[Signature]</i>	Date:	10/31/94
Approved by:	<i>[Signature]</i>	Date:	10/31/94



PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

WHC-SD-WM-DP-076, REV. 0

**AT : LA-508-101 (D-2) SPIKED SAMPLE**

		SPIKE	REPLICATE
Type	DETECTOR NUMBER	15	15
SPIKE	DISH SIZE 1, 2, or 5 (MS)	2	2
Work List	TOTAL COUNTS (TC)	47068	50883
114	COUNT TIME in MINUTES (CT)	30	30
AT or TB ?	BACKGROUND in cpm (BKG)	0.3	0.3
AT	SAMPLE VOLUME in mL (Spiked Vial) (SS)	0.100	0.100
Test Code	SAMPLE DILUTION FACTOR (Spiked Vial) (DF)	1	1
ALPHA01	DIGEST GRAMS of SOLIDS/L (Dg/L)	2.286	2.286
Matrix	SPIKE VOLUME in mL (SVol)	0.100	0.100
SOLID	SPIKE DILUTION FACTOR (SDF)	1	1
Sample #	SPIKE VALUE in µCi/L (SVal)	36.4	36.4
S94T000149	INSTRUMENT EFFICIENCY FACTOR (EFF)	0.238	0.238
Instrument Code	SAMPLE + SPIKE µCi/g (S+S)	1.30E+01	1.40E+01
WB26872	AVERAGE or MAXIMUM µCi/g from FORM C	1.1453E-01	
Analyst			
KRM	Rs (Sample Count Rate) = (TC / CT) - BKG		
Date	SAMPLE + SPIKE µCi/g = Rs * 1000mL/L * DF / (EFF * SS * Dg/L * 2220000dpm/µCi)		
10/30/94	PERCENT SPIKE RECOVERY = (((S+S µCi/g - SAMPLE µCi/g) * ((SDF/SVol)/(DF/SS/Dg/L)))/(SVal)*100		
Time			
06:00 AM			

**RESULT** **AVG. PERCENT SPIKE RECOVERY = 84.1%**

Data Entry by: <i>M. S. Smith</i>	Date: 31-Oct-94
Approved by: <i>[Signature]</i>	Date: 10/31/94

Form 508101\_X Rev. 1.1 Page 1 of 1

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

WHC-SD-WM-DP-076, REV. 0

**AT : LA-508-101 (D-2) SOLIDS**

		SAMPLE	REPLICATE
Type	DETECTOR NUMBER	15	15
SAMPLE	DISH SIZE 1, 2, or 5 (MS)	2	2
Work List	TOTAL COUNTS (TC)	63	54
114	COUNT TIME in MINUTES (CT)	30	30
AT or TB?	BACKGROUND in cpm (BKG)	0.3	0.3
AT	SAMPLE SIZE in mL (SS)	0.100	0.100
Test Code	DILUTION FACTOR (DF)	1	1
ALPHA01	DIGEST GRAMS of SOLIDS/L (Dg/L)	2.3725	2.3725
Matrix	EFFICIENCY FACTOR (EFF)	0.2380	0.2380
SOLID	Lc, Rmax, or Rs,(SAMPLE RATE) as APPROPRIATE	1.800	1.500
Sample #	Sample Concentration in $\mu\text{Ci/g}$	1.44E-02	
S94T000150	Replicate Concentration in $\mu\text{Ci/g}$	1.20E-02	
Instrument Code	Average Concentration in $\mu\text{Ci/g}$	1.3163E-02	
WB26872			
Analyst			
KRM	$R_s$ (Sample Count Rate) = (TC / CT) - BKG		
Date	ALPHA TOTAL $\mu\text{Ci/g}$ = $R_s * 1000\text{mL/L} * DF / (EFF * SS * Dg/L * 2220000\text{dpm}/\mu\text{Ci})$		
10/30/94			
Time	Relative Counting Error = $[(\text{The Square Root of } TC + BKG * CT) / (TC - BKG * CT)] * 1.96 * 100$		
08:00 AM	Detection Levels and Less Than Values are determined from Procedure LA-508-002.		

v RESULTS v

ALPHA TOTAL in $\mu\text{Ci/g}$ (Average)	=	1.32E-02	DETECTION LEVEL
RELATIVE COUNTING ERROR	=	34.6%	

Data Entry by: <i>[Signature]</i>	Date: 10/31/94
Approved by: <i>[Signature]</i>	Date: 10/31/94



PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

WHC-SD-WM-DP-076, REV. 0

**AT : LA-508-101 (D-2) SOLIDS**

		SAMPLE	REPLICATE
Type	DETECTOR NUMBER	15	15
DUPLICATE	DISH SIZE 1, 2, or 5 (MS)	2	2
Work List	TOTAL COUNTS (TC)	42	49
114	COUNT TIME in MINUTES (CT)	30	30
AT or TB?	BACKGROUND in cpm (BKG)	0.3	0.3
AT	SAMPLE SIZE in mL (SS)	0.100	0.100
Test Code	DILUTION FACTOR (DF)	1	1
ALPHA01	DIGEST GRAMS of SOLIDS/L (Dg/L)	2.3775	2.3775
Matrix	EFFICIENCY FACTOR (EFF)	0.2380	0.2380
SOLID	Lc, Rmax, or Rs, (SAMPLE RATE) as APPROPRIATE	1.100	1.333
Sample #	Sample Concentration in $\mu\text{Ci/g}$	8.76E-03	
S94T000150	Replicate Concentration in $\mu\text{Ci/g}$	1.06E-02	
Instrument Code			
WB26872	Average Concentration in $\mu\text{Ci/g}$	9.6855E-03	
Analyst			
KRM	$R_s$ (Sample Count Rate) = (TC / CT) - BKG		
Date	ALPHA TOTAL $\mu\text{Ci/g}$ = $R_s \cdot 1000\text{mL/L} \cdot \text{DF} / (\text{EFF} \cdot \text{SS} \cdot \text{Dg/L} \cdot 2220000\text{dpm}/\mu\text{Ci})$		
10/30/94			
Time	Relative Counting Error = $[(\text{The Square Root of TC} + \text{BKG} \cdot \text{CT}) / (\text{TC} - \text{BKG} \cdot \text{CT})] \cdot 1.96 \cdot 100$		
06:00 AM	Detection Levels and Less Than Values are determined from Procedure LA-508-002.		

**v RESULTS v**

ALPHA TOTAL	in $\mu\text{Ci/g}$ (Average)	=	9.69E-03	DETECTION LEVEL
RELATIVE COUNTING ERROR		=	42.4%	$\mu\text{Ci/g}$

Data Entry by:	<i>[Signature]</i>	Date:	10/31/94
Approved by:	<i>[Signature]</i>	Date:	10/31/94