PNL--8464

DE93 005085

PERFORMANCE EVALUATION OF THE
ENRAF-NONIUS MODEL 872 RADAR GAGE

T. J. Peters
W. R. Park

Task Manager: J. W. Voyles
Project Manager: J. B. Coison

December 1992

Prepared for
Westinghouse Hanford Company and
U.S. Department of Energy
under Contract DE-AC06-76RLO 1830

Pacific Northwest Laboratory
Richland, Washington 99352

MASTER

dk
DISTRIBUTION OF THIS DOCUMENT IS UNLIMITED

SUMMARY

There are indications that the Enraf-Nonius Radar Gage installed in Tank 241-SY-101 (Tank 101-SY) may not be providing an accurate reading of the true surface level in the waste tank. The Pacific Northwest Laboratory (PNL) performed an initial study to determine the effect of the following items on the distance read by the gage:

1. Tank riser
2. Material permittivity and conductivity
3. Foam
4. Proportion of supernatant to solid material in the field of view of the instrument
5. Physical geometry of the supernatant and solid material changing in the field of view with respect to time
6. Varying water content in the solid material.

The results of the tests indicate that distance measured by the radar gage is affected by the permittivity, conductivity, and angle of the target surface. These parameters affect the complex input impedance of the signal received by the radar gage to measure the distance to the target. In Tank 101-SY, the radar gage is placed on top of a 12 in. diameter riser. The riser affects the field of view of the instrument, and a much smaller target surface is detected when the radar beam propagates through a riser. In addition, the riser acts as a waveguide, and standing waves are enhanced between the target surface and the radar gage. The result is a change in the level measured by the radar gage due to changing properties of the target surface even when the distance to the target does not change.

The test results indicate that the radar will not detect dry crust or foam. However, if the crust or foam is stirred so that it becomes wet, then the crust or foam becomes detectable. The level read using the radar gage decreased as the moisture in the crust or foam evaporated.

This particular level gage was designed for use in the petroleum industry to measure the liquid level in their storage tanks, which is an uniform and

highly reflective surface. It was not designed to be placed on a pipe, and it is recommended in the installation manual that a radius of 1.27 m (50 in.) on either side of the gage be free from large reflecting objects in order for it to operate properly. The installation manual states that an antenna unit designed to be roof-mounted cannot be interchanged with one designed to be well-mounted because settings for gain and filtering must be changed, which is done at the factory. It is recommended that Enraf-Nonius personnel be contacted to determine if the radar unit can be modified for the length of riser pipe that the radar beam propagates through. Finally, other methods to measure the distance to the surface should be examined, such as acoustic or optical probes.

CONTENTS

SUMMARY	iii
BACKGROUND.	1
TEST RESULTS.	3
TEST SET-UP	3
TARGET TRACKING	5
FIELD OF VIEW	7
EFFECT OF PERMITTIVITY AND CONDUCTIVITY	8
SIMULATION OF VARYING WATER CONTENT	10
FOAM.	10
PROPORTION OF SUPERNATANT TO SOLID MATERIAL IN THE FIELD-OF-VIEW.	12
PHYSICAL GEOMETRY	15
REPEATABILITY OF RESULTS	15
CONCLUSIONS	21
RECOMMENDATIONS	23
REFERENCES.	25
APPENDIX A - RAW DATA SHEETS	A.1
APPENDIX B - RAW REFLECTION DIAGRAMS	B.1

FIGURES

1	Experimental Set-Up	4
2	Reflection Diagrams - Glass Target With and Without a Riser in Front of the Antenna	6
3	Reflection Diagrams - Dry and Wet Glass Targets	9
4	Reflection Diagrams - Tracking of a Wet Glass Target Behind a Dry Glass Target	11
5	Change in Measured Distance vs. Displacement Parallel to the Magnetic Component	13
6	Change in Measured Distance vs. Displacement Parallel to the Electric Component	14
7	Detection of Small Targets with Riser	16
8	Change in Measured Distance vs. Angle of Simulant	17
9	Reflection Diagrams - Effect of Rotating 102 mm. (4 in.) Diameter Simulated Waste	18

TABLES

1	Repeatability of Results	19
---	------------------------------------	----

BACKGROUND

The surface level of the contents of waste Tank 241-SY-101 has been measured for several years using a Food Instrument Corporation (FIC) electrical contact probe. When the probe contacts a liquid surface, electrical contact is registered and the liquid level is determined by measuring the amount of cable that was extended into the tank when electrical contact occurred. The desired accuracy for measuring the surface level is ± 6.4 mm. (± 0.25 in.). The surface level is automatically recorded using the data acquisition system at predetermined times depending on the status of the tank.

One of the problems with measuring surface level using the contacting FIC probe is that crystals form on the probe as it is withdrawn from the liquid material and left hanging in the tank. This leads to inaccuracies in subsequent measurements made with this probe. Another problem with the present FIC probe is that it cannot accurately determine the level of solid materials if the surface dries and becomes nonconductive. Also, the weight of the probe eventually may pound a hole into the solid material, resulting in a false reading. Finally, because the dried sludge level can be so uneven and the salt cakes may shift side-to-side, the probe may produce false readings of the actual (average) level of material in the tanks. These problems generally stem from the fact that the FIC probe was designed to detect liquid surfaces and was not designed to detect solid surfaces.

A noncontact approach for measuring the surface level of the tank contents would solve several of the FIC problems discussed above. To address this problem, Westinghouse Hanford Company (WHC) personnel installed an Enraf-Nonius Radar Gage in Tank 101-SY in 1991. The Enraf radar level gage consists of an antenna unit mounted on the tank to be gaged; the field interface unit; the remote indicator, whose display indicates the distance measured by the radar gage; and the communication interface unit, used to obtain communication between a remote computer and the radar gage. A radar beam propagates from the antenna toward the target. Part of the beam is reflected from the target to the antenna where it is received. The reflected beam arriving at the antenna exhibits a phase shift with respect to the transmitted beam. This

kind of radar detection is known as a reflectometer radar because the radar transmitter and the target surface act like a resonant cavity. The value of the complex input impedance received at the antenna is a measure of the length of the radar beam return path (B.V. Enraf-Nonius Delft 1990, p. 9). Before installing the radar gage in the Tank 101-SY, WHC personnel tested the device using a target consisting of 80% urea and 12% sugar water. The testing indicated that the gage had an accuracy of ± 2 mm. (± 0.08 in.). Reflections from a simulated riser placed in front of the antenna were received by the gage, and it was necessary to program the radar gage to ignore any reflections less than 3.3 m (130 in.) away.

PNL personnel were requested to perform an initial study to determine the reason the radar gage was sometimes reading changes in the level that were not being detected by the FIC probe. It was found that the complex impedance used to measure the distance to the target is also a function of the target's permittivity, conductivity, and angle. For homogenous, highly reflective targets, such as the liquids for which this radar unit was designed, these other parameters will not change, so that the complex impedance becomes only a function of the distance the radar beam travelled. However, waste Tank 101-SY contents are not smooth, high reflective targets and the permittivity and conductivity of this material varies. As a result, level changes are being indicated by the radar gage that are not being detected by the FIC probe.

TEST RESULTS

TEST SET-UP

The Enraf-Nonius Model 872 Radar Gage was set up so the beam propagated horizontally across the length of a laboratory room at a height of approximately 1.32 m (52 in.) from the floor. Except for the field of view tests, in which the antenna was rotated 90°, all of the tests were done with the flange on the terminal box horizontal, which means the electric component of the radar waves was propagating horizontally and the magnetic component was propagating vertically. To simulate the riser, a 4.47 m (176 in.) galvanized steel heating duct, 305 mm (12 in.) in diameter, was placed in front of the antenna unit (see Figure 1a) and leveled using adjustable stands. Tests were also conducted with a 3 m (118 in.) length of heating duct, and the results were similar between the two lengths of simulated riser. Target distance was measured using an optical bench, and the target raised to the proper height using styrofoam stands. It was determined that the radar beam from this antenna unit is transmitted through styrofoam. For liquid targets, an aluminum reflector, set at 45°, was placed approximately 610 mm (24 in.) from the end of the duct so the radar beam would be reflected vertically downward to the floor (see Figure 1b). Smooth aluminum is a high reflector of the radar beam; and because it was set at a 45° angle, the reflector was not detectable. Targets were placed approximately 6.4 m (252 in.) from the front of the antenna unit. The distance measured using the radar gage was read on the meter attached to the remote indicator. The system was programmed to measure the distance from the bottom of a tank, which for these tests was set at 9.89 m (350 in.). It was also programmed to ignore any reflections occurring within 4.88 m (192 in.) of the antenna. The distances stated in this report are converted so that everything is measured from the face of the antenna unit. In addition, the signals being received at the radar gage were evaluated by examining the reflection diagrams, and some of these are shown in this report.

FIGURE 1a. Horizontal Experimental Set-Up

FIGURE 1b. Vertical Experimental Set-Up

FIGURE 1. Experimental Set-Up

TARGET TRACKING

The first test performed with the radar gage was to determine its accuracy. A thin glass target was moved from 6.25 m (246 in.) to 6.55 m (258 in.) in 25.4 mm (1 in.) increments. A glass target was used because glass has a relative permittivity (compared to empty space) of 6-10 (Weber 1969, p. 283). This is similar to the relative permittivity of sodium nitrate, which is 5.2, and sodium carbonate, which is 5 - 8 (Weast 1969, p. E-64), which are two major constituents in the waste tanks. Water, another major constituent, has a relative permittivity of approximately 80 at room temperature (Weast 1969, p. E-67). The amount of energy reflected from the target surface is a function of the permittivity and conductivity of the surface. The results of the first tests showed that the gage had a accuracy of about ± 2.54 mm (± 0.1 in.). The repeatability was about ± 1.27 mm (± 0.05 in.). Similar results were obtained when the simulated riser was not present. The distance to the glass target was recorded for a 24-hour period with the radar beam propagating through the riser. The measured distance did not vary more than ± 1.02 mm (± 0.04 in.).

In order to determine the time it takes for the radar gage to respond to changes in the surface level, a glass target was placed approximately one foot in front of another glass target. The first glass target was then removed. The response time was approximately one minute.

Testing indicated that the presence of the riser affects the measured target distance. Without moving a glass target placed at 6.4 m (252 in.) from the antenna, the distance to the target was read on the radar gage before and after the riser was placed in front of the antenna. As shown in Figure 2, the target distance changed by several millimeters (a few tenths of an inch). Note that the amplitude of the signal from the target, as well as the noise level, increased when the riser was placed in front of the antenna. The magnitude and the direction of the change in the distance measured by the radar gage when the riser is placed in front of the antenna is affected by the distance to the target and the material.

FIGURE 2a. Glass Target at 6.40 m (252 in.) with No Riser

FIGURE 2b. Glass Target at 6.40 m (252 in.) with a Riser

FIGURE 2. Reflection Diagrams - Glass Target With and Without a Riser in Front of the Antenna

The distance to the target that was measured using the radar gage depended not only on the distance to the target, but also on the location of the simulated riser. Moving the riser approximately 38.1 mm (1.5 in.) away from the face of the antenna, the distance to the glass target approximately 6.1 m (240 in.) away increased by approximately 3.18 mm (0.125 in.). Again, the magnitude of the change in the measured distance is a function of the distance to the target and the target material.

The reason the distance to the target determined using the radar gage is influenced by the riser is because the riser is acting as a waveguide and the radar beam is being reflected from the cylindrical walls of the simulated riser. The distance measured using the Enraf-Nonius Model 872 Radar Gage is determined by the standing wave created between the antenna and the target surface, and this standing wave is being influenced by the riser. This is demonstrated by the fact that moving the riser without moving the target will affect the distance measured by the radar gage. As will be shown in the rest of this report, the permittivity and conductivity of the target will also affect the distance measured by the radar gage.

FIELD OF VIEW

The riser affects the area that is detected using the gage. A test was conducted to determine the field-of-view of the gage when placed inside the simulated riser. The glass target (457 x 610 mm or 18 x 24 in.) was placed 6.4 m (252 in.) from the front face of the antenna unit, and an aluminum reflector (559 x 660 mm or 22 x 26 in.) was slowly moved to the center of the target. The position of the aluminum reflector when the gage first detected the aluminum plate was recorded. This usually showed up as an error signal on the remote indicator, i.e., the gage was becoming confused by the two targets. This was repeated with the antenna rotated 90° so the magnetic component was propagating horizontally. The results indicate that at 6.4 m (252 in.), after the beam propagates through the 4.47 m (176 in.) simulated riser, the target area seen by the radar gage is only 432 x 330 mm (17 x 13 in.), with the larger diameter being along the direction of the electric component of the radar beam. This would indicate that the radar beam is expanding at an angle of $\pm 1.9^\circ \times \pm 0.4^\circ$ from the edge of the simulated riser, i.e., the radius of the

beam at any position is 6 in. plus the product of the distance from the end of the riser and the tangent of the angle. The angle of expansion was similar when the length of the riser was decreased to approximately 3.05 m (120 in.). Without the riser, based on the manufacturer's data, the area detected using the radar gage is 1.35 x 0.89 m (53 in. x 35 in.) at a range of 6.4 m (252 in.), which is 12° x 8°. Thus, the area seen by the radar gage through the riser is 12% of its normal field of view. This means most of the energy in the radar beam is being affected by the riser.

EFFECT OF PERMITTIVITY AND CONDUCTIVITY

Two paper towels were placed on a glass target set at approximately 6.4 m (252 in.) from the radar antenna. The radar gage read 6.41 m (252.25 in.). Water was applied to the towels. This would simulate the top surface of the waste contents becoming wet. The gage read 6.44 m (253.47 in.), a change of 31 mm (1.22 in.) without any change in the movement of the target. This is caused by the change in the permittivity of the target. As stated previously, water has a relative permittivity of approximately 80, whereas glass has a relative permittivity of 6 to 10, and the reflectivity of a surface is a function of its permittivity and conductivity. Moving the glass draped with wet towels 152.4 mm (6 in.) closer to the antenna, the reading changed 177 mm (6.97 in.). Additional testing indicated that the magnitude and direction of the change when the wet towels are applied to the glass was a function of the distance to the target. This is due to the standing wave effect which is enhanced by the wave propagating through the riser. The distance measured by the radar gage as the water evaporated from the towels always returned to the distance measured by the radar gage before the water was applied to the towels. Figure 3 shows the reflection diagram for a dry and a wet glass target at 6.55 m (258 in.). Note that the amplitude of the signal decreased for the wet glass target, which could be caused by an increased scattering of the radar beam from the wet surface.

To determine if changing the conductivity of the target surface will change the distance measured by the radar, sodium nitrate was added to a tub of water. Salt water is known to have a higher conductivity than tap water. The distance measured by the radar gage decreased by 10.9 mm (0.43 in.)

FIGURE 3a. Dry Glass Target at 6.55 m (258 in.)

FIGURE 3b. Wet Glass Target at 6.55 m (258 in.)

FIGURE 3. Reflection Diagrams - Dry and Wet Glass Targets

at approximately 6.48 m (255 in.) after the sodium nitrate was added to the water.

SIMULATION OF VARYING WATER CONTENT

In order to simulate water moving inside the solid material, the glass plate was set at 6.40 m (252 in.), and the gage read 6.40 m (252.01 in.). Then, a wet glass plate was set at 6.65 m (262 in.). The gage read 6.41 m (252.48 in.), or a change of 11.9 mm (0.47 in.). The wet glass plate was moved to 6.55 m (258 in.) and the gage read 6.41 m (252.27 in.). Thus, it appears that water placed behind the glass plate, which had a permittivity similar to the dry waste tank contents, affects the gage reading. The test was repeated with the dry and wet plate separated by 610 mm (24 in.), as shown in Figure 4. Note that two peaks are shown between 5.99 m (236 in.) and 7.01 m (276 in.) in the reflection diagram when the dry and wet glass plate are separated by 610 mm (24 in.). The second peak has a lower amplitude. The peaks merge when the wet plate was brought up to the dry glass plate, and the distance that was measured using the radar gage increased by 5.1 mm (0.2 in.). The magnitude of the change in the measured distance is a function of the distance of the target from the antenna.

FOAM

A Tank 101-SY simulant waste (SYI-SIM-91A) was placed in the beam to determine if the radar beam would detect the top foamy, crusty layer. The simulant consisted of four layers--a bottom layer of white material about 19.1 mm (0.75 in.) thick, a second layer of off-white color about 19.1 mm (0.75 in.) thick, a liquid 57.2 mm (2.25 in.) thick, and a top layer about 12.7 mm (0.5 in.) thick. When placed in the radar beam, the radar did not detect the top layer, but read the distance to the top of the liquid level. Stirring the top layer caused the distance measured using the radar gage to decrease by 7.9 mm (0.31 in.). As the moisture in this layer evaporated, the distance slowly increased 2.5 mm (0.1 in.).

Because this test was done with a small 102 mm (4 in.) diameter target, another test was conducted using soap bubbles created in a swimming pool of

FIGURE 4a. Dry Glass Target at 6.10 m (240 in.),
Wet Glass Target at 6.70 m (264 in.)

FIGURE 4b. Dry Glass Target at 6.10 m (240 in.),
Wet Glass Target at 6.20 m (244 in.)

FIGURE 4. Reflection Diagrams - Tracking of a Wet
Glass Target Behind a Dry Glass Target

water. The swimming pool was larger than the field of view of the radar gage. Adding 25.4 mm (1 in.) of water to the pool, the radar gage read 6.49 m (255.65 in.). When soap was added to the water and bubbles created, the radar gage read 6.50 m (255.83 in.), or a change of 4.6 mm (0.18 in.). Just adding soap to the water without creating bubbles did not change the reading on the radar gage. It appears that the radar unit is not detecting the soap bubbles, but it is detecting the liquid level below the soap bubbles. The liquid level would be slightly displaced by the bubbles and this is apparently detected by the radar gage.

PROPORTION OF SUPERNATANT TO SOLID MATERIAL IN THE FIELD-OF-VIEW

In order to determine if a higher permittivity and conductivity material, such as the supernatant, at the edge of the field of view will cause the gage to read a larger reading than when it is directly underneath it, a test was conducted of moving a tub of water, 127 x 127 mm (5 x 5 in.) and approximately 57.2 mm (2.25 in.) high, in the field of view of the instrument. Glass was placed under the styrofoam tub so that the background area around the tub within the field of view of the radar gage would be constant. As discussed previously, water has a higher reflectivity than glass. The tub of water was moved in both the direction of the magnetic field and the direction of the electric field. The results are shown in Figures 5 and 6. The results indicate that an increase in the distance measured by the radar gage (on the order of a few millimeters or a few tenths of an inch) is obtained by varying the location of the highest reflector in the field of view. This is a greater change than would be expected from a simple trigonometric relationship, i.e., tangent of an angle. The test was repeated by moistening a paper towel, cut to the same size as the tub of water, and moving it directly on the glass plate to eliminate height variations between the two targets in the field of view of the instrument. The results, also shown in Figures 5 and 6, show a wider curve, indicating a smaller, but still significant, error in measuring the distance to the target with a higher reflector target near the edge of the field of view.

Because the field of view is smaller when the beam propagates through the riser, a smaller target will affect the reading. For example, the amplitude

FIGURE 5. Change in Measured Distance vs. Displacement Parallel to the Magnetic Component

FIGURE 6. Change in Measured Distance vs. Displacement Parallel to the Electric Component

of the signal from the four-inch diameter simulated waste tank contents was not high enough to be detected by the radar gage without the riser, but the signal had a high enough amplitude when the riser was placed in front of the antenna. This is seen in Figure 7. Thus, small areas of material which fall into the field of view of the radar beam and have different properties than the surrounding material will affect the distance measured by the radar gage because the beam is propagating through the riser.

PHYSICAL GEOMETRY

A test was conducted to determine the effect of angle of the target on the distance measured using the radar gage. A 102 mm (4 in.) diameter dish of simulated waste tank contents, consisting mostly of sodium nitrite and iron oxide, was placed in the center of the field of view of the output from the radar gage. The background consisted of the wall of the room located 6.91 m (272 in.) from the antenna. Water was added to the material until it had a consistency similar to peanut butter. The angle of the material was varied and the resulting change in the distance measured by the radar gage was recorded. The results are shown in Figure 8 for nominal distance to the material of 6.25 m (246 in.), 6.40 m (252 in.), and 6.55 m (258 in.), respectively. Figure 8 indicates that it is possible to obtain changes in the distance measured by the radar unit, on the order of a 25-50 mm, for small changes in the angle of the target. The effect is a function of the distance from the antenna. Figure 9 shows the reflection diagrams for this test. Note that the amplitude of the reflected signal decreases as the simulant is rotated until the amplitude falls below the threshold of detection, after which the radar gage detected the next signal above this threshold, which for this test was the wall of the room.

REPEATABILITY OF RESULTS

Table 1 is an example of the repeatability obtained with the radar gage during these tests. The distance to the target measured by the radar gage was determined at four positions for the dry and wet glass target (run 1), and then the target was moved back to the first position and the test repeated (run 2). From the results in Table 1, it is seen that the repeatability of

FIGURE 7a. Detection of Simulated Waste with No Riser

FIGURE 7b. Detection of Simulated Waste with Riser

FIGURE 7. Detection of Small Targets with Riser

FIGURE 8. Change in Measured Distance vs. Angle of Simulant

FIGURE 9a. Simulant at 6.40 m (252 in.) and 0°

FIGURE 9b. Simulant at 6.40 m (252 in.) and 5°

FIGURE 9c. Simulant at 6.40 m (252 in.) and 15°

FIGURE 9. Reflection Diagrams - Effect of Rotating 102 mm. (4 in.) Diameter Simulated Waste

Actual Position (mm) (in.)		Radar Gage Reading							
		Dry Glass Target				Wet Glass Target			
		Run 1		Run 2		Run 1		Run 2	
		(mm)	(in.)	(mm)	(in.)	(mm)	(in.)	(mm)	(in.)
6248	246	6246	245.92	6247	245.93	6280	247.25	6260	247.26
6401	252	6404	252.14	6405	252.16	6410	252.36	6405	252.15
6553	258	6563	258.40	6561	258.30	6574	258.80	6586	259.30
6706	264	6720	264.58	6729	264.93	6738	265.27	6713	264.28

TABLE 1. Repeatability of Results

the radar gage to determine the position of a smooth homogeneous target, such as the dry glass target, was 3.0 mm (0.12 in.). This is consistent with the data discussed in the Target Tracking section of the report. However, the repeatability of the radar gage to determine the position of the rough wet target was 10.9 mm (0.43 in.). It would be expected that the repeatability of the measurements will be less for the wet target than the dry target because, as shown in this report, the distance measured to the target is a function of the permittivity and conductivity of the target, and the conductivity and permittivity of the wet target is going to be a function of the amount of water on the target. In addition, due to the standing wave produced by the radar unit, the amount of change in the distance measured using the radar gage is a function of the distance to the target. This is also shown in Table 1. The change in the distance due to the target becoming wet varies depending on the position of the target from the radar gage. The distance measured using the radar gage is a function of the standing wave created between the target surface and the antenna of the radar gage. This standing wave is affected by many different surface parameters. Thus, it appears that it would not be possible to correct for the changes in the distance determined by the radar gage due to changes in the properties of the target material.

CONCLUSIONS

The results of these tests indicate that the distance measured by the Enraf-Nonius radar unit is dependent on many variables, including the conductivity and permittivity of the material within the field of view, the position of the highest reflector within the field of view, and the angle of the target. Placing the radar unit into a metal riser affects the performance of the instrument. The metal riser acts as a waveguide and resonator which affects the standing wave between the target surface and the antenna. The standing wave is used by the radar gage to determine the distance to the target surface. The riser greatly decreases the field of view of the radar antenna, thus allowing a small piece of material, that gets into the field of view and has different properties than the surrounding waste tank contents, to affect the distance measured by the radar gage. In addition, because the radar beam propagating through the riser is only detecting a small area, changes in the surface level due to the roughness of the surface will be measured using the radar gage.

The change in distance to a target surface, as measured by the radar gage, due to changes in the properties of the target material is a function of the distance to the target. Thus, it appears that it would not be possible to correct for the changes in the distance determined by the radar gage due to changes in the properties of the target material.

The radar gage did not detect dry crust or foam. However, if any moisture gets onto the crust or foam, the radar gage will detect the surface, and the level will change as this moisture evaporates from the crust or foam.

The radar gage does measure accurately (± 0.10 in.) for a smooth, homogenous, highly reflective target. Thus, it would be expected to obtain accurate readings for a waste tank that has a homogenous liquid surface.

RECOMMENDATIONS

- The manufacturer of the radar unit should be contacted to determine if it is possible to place the antenna above a metal pipe of the size and shape of the riser. The installation manual indicates that it may be possible to place the antenna above a metal tube, but it requires a change in factory settings in gain and filtering in order to make the unit operate properly. The radar unit is being used in a situation for which it was not designed to operate. It was designed to be used on petroleum storage tanks that have homogenous highly reflective contents, which is not what is present in the waste tanks at Hanford.
- Other methods to measure the distance to the surface, such as acoustic sensors, amplitude modulated laser, and optical time domain reflectometer, should be examined because the accuracy of the radar unit will still be affected by the different properties and roughness of the materials that will get into the field of view of the instrument.

REFERENCES

B. V. ENRAF-NONIUS DELFT. June 1990. Installation and Service Manual Series 872RLG Radar Level Gage, B.V. Enraf-Nonius Delft, The Netherlands, Version 0.0.

Weast, R. C., ed. 1969. Handbook of Chemistry and Physics. The Chemical Rubber Co., Cleveland, Ohio.

Weber, R. L., Manning, K. V., and White, M. W. 1969. College Physics. McGraw Hill, Inc., New York.

APPENDIX A

RAW DATA SHEETS

TARGET TRACKING

A. Move Glass Target in 1 in. Increments

<u>Target Location (in.)</u>	<u>Radar Gage Reading (in.)</u>	<u>Difference in Readings (in.)</u>
252	252.12	
246	246.31	
247	247.43	1.12
248	248.43	1.00
249	249.41	0.98
250	250.24	0.83
251	251.22	0.98
252	252.21	0.99
253	253.18	0.97
254	254.27	1.09
255	255.35	1.08
256	256.44	1.09
257	257.26	0.82
258	258.43	1.17

Average Difference $\pm$ Standard Deviation = 1.01 ± 0.10 in.

Difference Between Two Readings at 252 in. = 0.09 in.

B. Glass Target - No Riser

<u>Target Location (in.)</u>	<u>Radar Gage Reading (in.)</u>	<u>Difference in Readings (in.)</u>
246	246.01	
252	252.24	6.21
258	258.18	6.06

Average Difference $\pm$ Standard Deviation = 6.14 ± 0.11 in.

TARGET TRACKING

C. Variation with Time When Glass Target in One Location

<u>Date</u>	<u>Time</u>	<u>Radar Gage Reading (in.)</u>
Sept. 15, 1992	7:42 AM	251.67
	8:33	251.67
	9:30	251.69
	11:08	251.71
	12:06 PM	251.72
	12:47	251.73
	1:42	251.73
	2:19	251.74
	2:52	251.73
	3:20	251.73
	3:45	251.73
Sept. 16, 1992	7:30 AM	251.68
	8:49	251.69

Average Reading $\pm$ Standard Deviation = 251.71 $\pm$ 0.03 in.

D. Time to Find Target

Experimental Set-Up: One glass target at 21 ft, the other glass target at 20 ft. Remove glass target at 20 ft, record time to stabilize on glass target at 21 ft.

<u>Time (sec.)</u>	<u>Radar Gage Reading (in.)</u>
0	240.19
10	240.33
20	248.51
30	250.98
40	251.71
50	251.89
60	251.99
70	252.02
80	252.03
90	252.02
100	252.02
110	252.03
120	252.03

FIELD OF VIEW

A. Electric Component

<u>Aluminum Target Position (in.)</u>	<u>Radar Gage Reading (in.)</u>	<u>Meter Status</u>
20.5	252.40	HA
19.5	252.40	HA
18.5	252.38	HA
17.5	252.39	HA
16.5	252.39	HA
15.5	252.41	HA
14.5	252.41	HA
13.5	252.39	HA
12.5	252.39	HA
11.5	252.39	HA
10.5	252.39	HA
9.5	252.43	HA
8.5	252.40	HA
7.5	252.38	HA FF
6.5	252.48	HA FF
5.5	252.31	HA
4.5	252.02	HA FF
3.5	252.05	HA FF
2.5	252.28	HA FF
1.5	250.40	HA FF
0.5	246.49	HA FF
-0.5	244.01	HA FF
-1.5	243.64	HA FF
-2.5	243.76	HA FF
-3.5	243.70	HA FF
-4.5	243.40	HA FF
-5.5	243.68	HA FF
-6.5	243.91	HA FF
-7.5	243.99	HA FF
-8.5	207.13	HA FF
-9.5	204.38	HA FF
-10.5	169.63	HA FF

FIELD OF VIEW

B. Magnetic Component

<u>Aluminum Target Position (in.)</u>	<u>Radar Gage Reading (in.)</u>	<u>Meter Status</u>
19.5	252.54	HA
18.5	252.54	HA
17.5	252.53	HA
16.5	252.53	HA
15.5	252.54	HA
14.5	252.53	HA
13.5	252.49	HA
12.5	252.54	HA
11.5	252.61	HA
10.5	252.40	HA
9.5	252.40	HA
8.5	252.60	HA
7.5	252.87	HA
6.5	252.43	HA
5.5	251.56	HA FF
4.5	251.75	HA FF
3.5	252.00	HA FF
2.5	251.77	HA FF
1.5	247.78	HA FF
0.5	244.57	HA FF
-0.5	243.60	HA FF
-1.5	243.50	HA FF

PROPORTION OF SUPERNATANT TO SOLID MATERIAL IN THE FIELD-OF-VIEW

A. Electric Component

<u>Position</u> <u>(in.)</u>	<u>Target</u> <u>2.25 in. From Floor</u>	<u>Target</u> <u>On Floor</u>
	<u>Radar Gage</u> <u>Reading (in.)</u>	<u>Radar Gage</u> <u>Reading (in.)</u>
3	258.31	258.32
2	258.05	258.32
1	257.57	258.24
0	257.34	258.19
-1	257.25	258.06
-2	257.16	258.00
-3	257.17	258.01
-4	257.78	257.99
-5	259.52	257.99
-6	----	258.02

B. Magnetic Component

<u>Position</u> <u>(in.)</u>	<u>Target</u> <u>2.25 in. From Floor</u>	<u>Target</u> <u>On Floor</u>
	<u>Radar Gage</u> <u>Reading (in.)</u>	<u>Radar Gage</u> <u>Reading (in.)</u>
5	----	258.16
4	258.90	258.16
3	258.89	258.17
2	256.91	258.20
1	256.91	258.15
0	257.15	257.99
-1	257.38	257.98
-2	257.41	257.82
-3	257.41	257.78
-4	257.37	257.78
-5	257.28	257.84
-6	257.86	257.93
-7	259.58	257.97

PHYSICAL GEOMETRY

A. Target 246 in. from Antenna

<u>Angle</u>	<u>Radar Gage Reading (in.)</u>	<u>Difference in Readings (in.)</u>
7.5°	Wall	----
5°	247.06	0.92
0°	246.98	Base
-10°	246.89	-0.09
-12.5°	246.96	-0.02
-15°	Wall	----

B. Target 252 in. from Antenna

<u>Angle</u>	<u>Radar Gage Reading (in.)</u>	<u>Difference in Readings (in.)</u>
10°	Wall	----
7.5°	254.42	2.15
5°	252.83	0.56
0°	252.27	Base
-10°	253.18	0.91
-12.5°	Wall	----

C. Target 258 in. from Antenna

<u>Angle</u>	<u>Radar Gage Reading (in.)</u>	<u>Difference in Readings (in.)</u>
5°	Wall	----
2°	258.35	0.23
0°	258.03	Base
-5°	258.05	0.02
-10°	268.77	10.74

APPENDIX B

RAW REFLECTION DIAGRAMS

FIGURE B.1. Riser Next to Antenna

FIGURE B.2. Riser Moved 1.5 in. from Antenna

FIGURE B.3. Water

FIGURE B.4 Add Sodium Nitrate to Water

FIGURE B.5. Liquid Simulant

FIGURE B.6. Add Foam Crust to Simulant

FIGURE B.7. 1 in. Water in Pool

FIGURE B.8. Add Soap Bubbles

FIGURE B.9. Water at Center

FIGURE B.10. Water Move 2 in. Mag. Dir.

DISTRIBUTION

No. of
Copies

OFFSITE

- 2 DOE/Office of Scientific and Technical
Information

ONSITE

DOE Richland Field Office Office

R. F. Christensen, A4-02

- 6 Westinghouse Hanford Company

H. Babad
G. D. Johnson
N. W. Kirch
J. W. Lentsch
C. P. Schroeder
W. F. White

- 14 Pacific Northwest Laboratory

R. T. Allemann
J. B. Colson
M. R. Kreiter
T. J. Peters (3)
D. M. Strachan
J. W. Voyles
Publishing Coordination
Technical Report Files (5)

END

**DATE
FILMED**

3 / 16 / 93