PNL-7953
AD-940

PNL--7953

DE92 007207

HYDROGEN SAFETY PROJECT
CHEMICAL ANALYSIS SUPPORT TASK

WINDOW "C"
VOLATILE ORGANIC ANALYSIS

B. M. Gillespie
R. W. Stromatt
G. A. Ross
E. A. Hoppe

January 1992

Prepared for
the U.S. Department of Energy
under Contract DE-AC06-76RLO 1830

Pacific Northwest Laboratory
Richland, Washington 99352

MASTER

DMB

DISTRIBUTION OF THIS DOCUMENT IS UNLIMITED

SUMMARY

Core samples taken from Hanford double-shell waste tank 101-SY during Window "C" (after the May 1991 gas release event) were analyzed for volatile organics by staff of the Analytical Chemistry Laboratory at Pacific Northwest Laboratory. Westinghouse Hanford Company submitted these samples as Sample Delivery Groups 1, 2, 4, and 5.

The procedures used and the chain of custody were carefully documented. Results are reported on Contract Laboratory Program report forms as required by the Technical Project Plan 17667. The results will be used by the 101-SY Hydrogen Safety Project.

CONTRIBUTORS LIST

PROJECT MANAGEMENT OFFICE

RW Stromatt, Co-Project Manager
BM Gillespie, Co-Project Manager
BO Barnes, Quality Engineer
SM Panther
DR Sauer

HOT CELL OPERATIONS

RT Steele, Sub-Task Leader
JK Rau
FV Hoopes
SA Homi

ORGANIC CHEMISTRY ANALYSIS

RW Stromatt, Sub-Task Leader
EW Hoppe
GA Ross
DL Bellofatto
MJ Steele
WT Cobb

CERTIFICATION STATEMENT

I certify that this data package is in compliance with the terms and conditions of the TPP 17667 and QA Plan MCS-027 for completeness. Release of the data contained in this hard copy data package has been authorized by the Analytical Chemistry Laboratory Project Manager or the Project Manager's designee, as verified by the following signature.

B. M. Gillespie

B. M. Gillespie
101-SY ACL Project Manager

12-17-91

Date

CONTENTS

SUMMARY	iii
CONTRIBUTORS LIST	v
CERTIFICATION STATEMENT	vii
1.0 INTRODUCTION	1.1
2.0 VOLATILE ORGANIC DATA	2.1
2.1 LABORATORY ANALYST SIGNATURE LIST	2.1
2.2 SAMPLE DELIVERY GROUP 1	2.2
2.3 SAMPLE DELIVERY GROUP 2	2.22
2.4 SAMPLE DELIVERY GROUPS 4 AND 5	2.44
APPENDIX A - TEST INSTRUCTIONS	A.1
APPENDIX B - CHAIN OF CUSTODY	B.1
B.1 WESTINGHOUSE CHAIN OF CUSTODY AND SAMPLE ANALYSIS REQUEST FORMS	B.1
B.2 PNL CHAIN OF CUSTODY FORMS	B.15

TABLES

1.1 101-SY Sample Numbers	1.2
-------------------------------------	-----

1.0 INTRODUCTION

This data package contains the results obtained by Pacific Northwest Laboratory (PNL) staff in the characterization of samples for the 101-SY Hydrogen Safety Project. The samples were submitted for analysis by Westinghouse Hanford Company (WHC) under the Technical Project Plan (TPP) 17667 and the Quality Assurance Plan MCS-027. They came from a core taken during Window "C" after the May 1991 gas release event. The analytical procedures required for analysis were defined in the Test Instructions (TI) prepared by the PNL 101-SY Analytical Chemistry Laboratory (ACL) Project Management Office in accordance with the TPP and the QA Plan.

The samples (Table 1) were submitted with the appropriate WHC Chain of Custody (COC) and Sample Analysis Request Forms. The samples were delivered to the "B" hot cells in the 325 Building in the 300 Area.

The requested analysis for these samples was volatile organic analysis. The quality control (QC) requirements for each sample are defined in the Test Instructions for each sample. The QC requirements outlined in the procedures and requested in the WHC statement of work were followed.

The data in this package are reported on forms as per the Contract Laboratory Program (CLP) reporting format. Two appendices are provided: one for Test Instructions and one for Chain of Custody, Sample Analysis Request, and Sample Receipt Forms.

TABLE 1.1. 101-SY Sample Numbers

WHC Sample Number PNL ALO Sample Number

Sample Delivery Group #1

#424	91-6264
#420	91-6265
#421	91-6266
#563	91-6267

Sample Delivery Group #2

#578	91-6425
#579	91-6426
#580	91-6427
#581	91-6428

Sample Delivery Group #4

#531	91-7174
#529	91-7175

Sample Delivery Group #5

#533	91-7319
------	---------

2.0 VOLATILE ORGANIC DATA

This section contains volatile organic data from the analyses of Sample Delivery Groups 1, 2, 4, and 5 from Window "C." For each sample delivery group, a brief analysis report is provided along with the Contract Laboratory Program report forms required by the Technical Project Plan 17667. The Laboratory Analyst Signature List is presented below.

ORGANIC LABORATORY ANALYST SIGNATURE LIST

<u>Analyst</u> <u>Name</u>	<u>Initials</u>	<u>Written</u> <u>Name</u>	<u>Initials</u>
Robert W. Stromatt	RWS
Eric W. Hoppe	EWB
Gerald A. Ross	GAR
Diana L. Bellofatto	DLB
Marilyn J. Steele	MJS
William T. Cobb	WTC

SAMPLE DELIVERY GROUP 1

A summary report of the analyses of samples in Delivery Group 1 provides general information on the samples, sample preparation, analysis method, and location of data files, and definitions of codes used in the CLP report forms. The report forms are included after the summary report.

**ORGANIC COMPOUND ANALYSIS REPORT
VOLATILE COMPOUNDS**

SAMPLE ANALYSIS REPORTED

Analysis of four slurry samples for volatile organic compounds by gas chromatography/mass spectrometry (GC/MS) is the subject of this report.

SAMPLE DESCRIPTION

<u>Sample ID</u>	<u>ACL Lab Number</u>
420-2	91-6265
563-2	91-6267
424-2	91-6264
421-2	91-6266

The samples were received in the "B" Analytical Hot Cell with substantial headspace, non-septa type closures, and at ambient temperature on 6/28/91. These sample handling procedures are not consistent with proper analytical protocol for this analysis.

SAMPLE PREPARATION

GC/FID volatile hexadecane screens were performed and indicated a dilution of the sample was not necessary. Due to the radioactivity of the samples, 1 gram aliquots of the samples were transferred to 20ml pre-cleaned VOA vials, and accurately weighed. The samples were serially dissolved in 5mL of lab water then added to a purge vessel attached to a Tekmar Liquid Sample Concentrator.

For proper vendor software quantitation of the weighed samples, "soil" matrix must be selected; although a "water" blank is used to accurately represent the handling of the samples. The analysis followed EPA-CLP SOW 2/88 procedures for the analysis of volatile compounds.

ANALYSIS METHOD

• GC/MS procedure:	PNL-ALO-335.
• GC/MS instrumentation:	HP-5890/5970 GC/MS (WB46864).
• GC/MS location:	Lab 427, 325 building.
• GC screening analysis:	PNL-ALO-340.
• GC/FID instrumentation:	HP-5890 (WB60701).
• GC/FID location:	Lab 330, 325 building.

QUALITY CONTROL

Quality control procedures specified for this method were followed. While these forms are not all included in this report,

the quality assurance performance requirements are summarized below:

<u>Form Information</u>	<u>Comments</u>
2A Surrogate Recovery	Meets all requirements. (563-2 had 1 Surrogate high.)
3A MS/MSD Recovery	Analysis not requested.
4A Method Blank Summary	Meets all requirements.
5A Tune/Mass Calibration	Meets all requirements.
6A Initial Calibration	5 point calibration. Meets all requirements.
7A Daily Calibration	Meets all requirements.
8A Internal Standards	Meets all requirements.

DATA

The data and calibration are archived on magnetic tape in the 325 building, 327-A laboratory. The following is the list of pertinent files:

<u>File Name</u>	<u>Sample Number</u>	<u>Sample Analyzed</u>
>SA201		Mass Calibration/Tune Check
>SA202		Daily Calibration
>SA2B3		Method Blank
>SA205	91-6265	420-2
>SA207	91-6267	563-2
>SA204	91-6264	424-2
>SA206	91-6266	421-2

Lab data are also maintained as follows:

<u>Activity</u>	<u>LRB Number</u>	<u>Page Number</u>
GC/MS injection log	BNW-52907	80-81

RESULTS

CLP Target Compounds: The attached 1A Forms show that volatile target compounds, Acetone, and 2-Butanone, were detected above the Contract Required Quantitation Limits, (CRQL's). However, they were observed in the blank at a similar level, and their presence is probably due to building renovations. In addition, the levels observed are acceptable according to the CLP methods criteria. They are indicated by the "B" flag in the

sample. Chloromethane, Methylene Chloride, and Vinyl Acetate were observed in the blank, below the CRQL.

The following defines the qualifiers, Q-flags, in the Form 1's:

<u>"Q" Flag</u>	<u>Definition</u>
U	Indicates the compound was analyzed for but not detected, the U-flagged concentration number is the CRQL.
J	Indicates an estimated value for the target or tentatively identified compounds, spectra meet criteria but response is below the CRQL for the target compounds.
B	Compound was found in the blank.
X	Indicates compound was manually deleted because all requirements were not met.
D	Analysis was performed on a diluted sample.
E	Indicates that Quantitation was outside the calibration range.
A	Aldol condensation product.

Tentatively Identified Compounds: No non-EPA target compounds were observed in the blank. As shown in the attached 1E Forms, 2-Ethyl-1-hexanol and n-Alkanes were observed in the samples.

ANALYST D. R. R. DATE 9/18/91 REVIEW [Signature] DATE 9-18-91

1A
VOLATILE ORGANICS ANALYSIS DATA SHEET

EPA SAMPLE NO.

Lab Name: BATTELLE-PNL

Contract: -----

BLANK

Lab Code: ----- Case No.: ----- SAS No.: ----- SDG No.: -----

Matrix: (soil/water) WATER

Lab Sample ID: BLANK

Sample wt/vol: 5 (g/mL) mL

Lab File ID: >SA2B3

Level: (low/med) LOW

Date Received: 06/27/91

% Moisture: not dec.-----

Date Analyzed: 7/02/91

Column: (pack/cap) CAP

Dilution Factor: 1.0

CAS NO.	COMPOUND	CONCENTRATION UNITS: (ug/L or ug/Kg) ug/L	Q
74-87-3-----	Chloromethane	2.	JB
74-83-9-----	Bromomethane	10.	U
75-01-4-----	Vinyl Chloride	10.	U
75-00-3-----	Chloroethane	10.	U
75-09-2-----	Methylene Chloride	4.	JB
67-64-1-----	Acetone	13.	B
75-15-0-----	Carbon Disulfide	5.	U
75-35-4-----	1,1-Dichloroethene	5.	U
75-34-3-----	1,1-Dichloroethane	5.	U
540-59-0-----	1,2-Dichloroethene (total)	5.	U
67-66-3-----	Chloroform	5.	U
107-02-2-----	1,2-Dichloroethane	5.	U
78-93-3-----	2-Butanone	17.	B
71-55-6-----	1,1,1-Trichloroethane	5.	U
56-23-5-----	Carbon Tetrachloride	5.	U
108-05-4-----	Vinyl Acetate	3.	JB
75-27-4-----	Bromodichloromethane	5.	U
78-87-5-----	1,2-Dichloropropane	5.	U
10061-01-5-----	cis-1,3-Dichloropropene	5.	U
79-01-6-----	Trichloroethene	5.	U
124-48-1-----	Dibromochloromethane	5.	U
79-00-5-----	1,1,2-Trichloroethane	5.	U
71-43-2-----	Benzene	5.	U
10061-02-6-----	trans-1,3-Dichloropropene	5.	U
75-25-2-----	Bromoform	5.	U
108-10-1-----	4-Methyl-2-pentanone	10.	U
591-78-6-----	2-Hexanone	10.	U
127-18-4-----	Tetrachloroethene	5.	U
79-34-5-----	1,1,2,2-Tetrachloroethane	5.	U
108-88-3-----	Toluene	5.	U
108-90-7-----	Chlorobenzene	5.	U
100-41-4-----	Ethylbenzene	5.	U
100-42-5-----	Styrene	5.	U
133-02-7-----	Xylene (total)	5.	U

1E
VOLATILE ORGANICS ANALYSIS DATA SHEET
TENTATIVELY IDENTIFIED COMPOUNDS

EPA SAMPLE NO.

BLANK

Lab Name: BATTELLE-PNL

Contract: -----

Lab Code: -----

Case No.: -----

SAS No.: -----

SDG No.: -----

Matrix: (soil/water) WATER

Lab Sample ID: BLANK

Sample wt/vol: 5 (g/mL) mL

Lab File ID: >SA2B3

Level: (low/med) LOW

Date Received: 07/02/91

% Moisture: not dec.-----

Date Analyzed: 07/02/91

Column: CAP

Dilution Factor: 1.0

Number TICs found: 0

CONCENTRATION UNITS:
(ug/L or ug/Kg) ug/L

CAS NUMBER	COMPOUND NAME	RT	EST. CONC.	Q
1.				
2.				
3.				
4.				
5.				
6.				
7.				
8.				
9.				
10.				
11.				
12.				
13.				
14.				
15.				
16.				
17.				
18.				
19.				
20.				
21.				
22.				
23.				
24.				
25.				
26.				
27.				
28.				
29.				
30.				

1A
VOLATILE ORGANICS ANALYSIS DATA SHEET

EPA SAMPLE NO.

91-6264

Lab Name: BATTELLE-PNL

Contract: ---

Lab Code: ---

Case No.: ----- SAS No.: ----- SDG No.: -----

Matrix: (soil/water) SOIL

Lab Sample ID: 424-2 101-S

Sample wt/vol: 1.0994 (g/mL) G

Lab File ID: >SA204

Level: (low/med) LOW

Date Received: 06/27/91

% Moisture: not dec.----

Date Analyzed: 7/02/91

Column: (pack/cap) CAP

Dilution Factor: 1.0

CAS NO.	COMPOUND	CONCENTRATION UNITS:	
		(ug/L or ug/Kg)	ug/Kg
74-87-3-----	Chloromethane_____	46.	IU
74-83-9-----	Bromomethane_____	46.	IU
75-01-4-----	Vinyl Chloride_____	46.	IU
75-00-3-----	Chloroethane_____	46.	IU
75-09-2-----	Methylene_Chloride_____	14.	IJB
67-64-1-----	Acetone_____	56.	I B
75-15-0-----	Carbon Disulfide_____	23.	IU
75-35-4-----	1,1-Dichloroethene_____	23.	IU
75-34-3-----	1,1-Dichloroethane_____	23.	IU
540-59-0-----	1,2-Dichloroethene_(total)_____	23.	IU
67-66-3-----	Chloroform_____	23.	IU
107-02-2-----	1,2-Dichloroethane_____	23.	IU
78-93-3-----	2-Butanone_____	92.	I B
71-55-6-----	1,1,1-Trichloroethane_____	23.	IU
56-23-5-----	Carbon Tetrachloride_____	23.	IU
108-05-4-----	Vinyl Acetate_____	23.	IU
75-27-4-----	Bromodichloromethane_____	23.	IU
78-87-5-----	1,2-Dichloropropane_____	23.	IU
10061-01-5-----	cis-1,3-Dichloropropene_____	23.	IU
79-01-6-----	Trichloroethene_____	23.	IU
124-48-1-----	Dibromochloromethane_____	23.	IU
79-00-5-----	1,1,2-Trichloroethane_____	23.	IU
71-43-2-----	Benzene_____	23.	IU
10061-02-6-----	trans-1,3-Dichloropropene_____	23.	IU
75-25-2-----	Bromoform_____	23.	IU
108-10-1-----	4-Methyl-2-pentanone_____	46.	IU
591-78-6-----	2-Hexanone_____	46.	IU
127-18-4-----	Tetrachloroethene_____	23.	IU
79-34-5-----	1,1,2,2-Tetrachloroethane_____	23.	IU
108-88-3-----	Toluene_____	23.	IU
108-90-7-----	Chlorobenzene_____	23.	IU
100-41-4-----	Ethylbenzene_____	23.	IU
100-42-5-----	Styrene_____	23.	IU
133-02-7-----	Xylene (total)_____	23.	IU

1E
VOLATILE ORGANICS ANALYSIS DATA SHEET
TENTATIVELY IDENTIFIED COMPOUNDS

EPA SAMPLE NO.

91-6264

Lab Name: BATTELLE-PNL

Contract: ---

Lab Code: ---

Case No.: --- SAS No.: --- SDG No.: ---

Matrix: (soil/water) SOIL

Lab Sample ID: 424-2 101-S

Sample wt/vol: 1.0994 (g/mL) G

Lab File ID: >SA204

Level: (low/med) LOW

Date Received: 06/27/91

% Moisture: not dec. ---

Date Analyzed: 7/02/91

Column: CAP

Dilution Factor: 1.0

Number TICs found: 3

CONCENTRATION UNITS:
(ug/L or ug/Kg) ug/L

CAS NUMBER	COMPOUND NAME	RT	EST. CONC.	Q
1. 96479	Furan, tetrahydro-2-methyl-	10.10	25.	
2. 104767	1-Hexanol, 2-ethyl-	29.95	150.	
3.	n-Alkane	31.81	24.	
4.				
5.				
6.				
7.				
8.				
9.				
10.				
11.				
12.				
13.				
14.				
15.				
16.				
17.				
18.				
19.				
20.				
21.				
22.				
23.				
24.				
25.				
26.				
27.				
28.				
29.				
30.				

1A
VOLATILE ORGANICS ANALYSIS DATA SHEET

EPA SAMPLE NO.

91-6265

Lab Name: BATTELLE-PNL

Contract: ---

Lab Code: ---

Case No.: -----

SAS No.: -----

SDG No.: -----

Matrix: (soil/water) SOIL

Lab Sample ID: 420-2 101-S

Sample wt/vol: 1.2346 (g/mL) G

Lab File ID: >SA205

Level: (low/med) LOW

Date Received: 06/27/91

% Moisture: not dec. ----

Date Analyzed: 7/02/91

Column: (pack/cap) CAP

Dilution Factor: 1.0

CAS NO.	COMPOUND	CONCENTRATION UNITS: (ug/L or ug/Kg) ug/Kg	Q
74-87-3-----	Chloromethane	41.	IU
74-83-9-----	Bromomethane	41.	IU
75-01-4-----	Vinyl Chloride	41.	IU
75-00-3-----	Chloroethane	41.	IU
75-09-2-----	Methylene Chloride	20.	IU
67-64-1-----	Acetone	41.	IU
75-15-0-----	Carbon Disulfide	20.	IU
75-35-4-----	1,1-Dichloroethane	20.	IU
75-34-3-----	1,1-Dichloroethane	20.	IU
540-59-0-----	1,2-Dichloroethane (total)	20.	IU
67-66-3-----	Chloroform	20.	IU
107-02-2-----	1,2-Dichloroethane	20.	IU
78-93-3-----	2-Butanone	76.	B
71-55-6-----	1,1,1-Trichloroethane	20.	IU
56-23-5-----	Carbon Tetrachloride	20.	IU
108-05-4-----	Vinyl Acetate	20.	IU
75-27-4-----	Bromodichloromethane	20.	IU
78-87-5-----	1,2-Dichloropropane	20.	IU
10061-01-5-----	cis-1,3-Dichloropropene	20.	IU
79-01-6-----	Trichloroethene	20.	IU
124-48-1-----	Dibromochloromethane	20.	IU
79-00-5-----	1,1,2-Trichloroethane	20.	IU
71-43-2-----	Benzene	20.	IU
10061-02-6-----	trans-1,3-Dichloropropene	20.	IU
75-25-2-----	Bromoform	20.	IU
108-10-1-----	4-Methyl-2-pentanone	20.	IU
591-78-6-----	2-Hexanone	41.	IU
127-18-4-----	Tetrachloroethene	20.	IU
79-34-5-----	1,1,2,2-Tetrachloroethane	20.	IU
108-88-3-----	Toluene	20.	IU
108-90-7-----	Chlorobenzene	20.	IU
100-41-4-----	Ethylbenzene	20.	IU
100-42-5-----	Styrene	20.	IU
133-02-7-----	Xylene (total)	20.	IU

1E
VOLATILE ORGANICS ANALYSIS DATA SHEET
TENTATIVELY IDENTIFIED COMPOUNDS

EPA SAMPLE NO.

91-6265

Lab Name: BATTELLE-PNL

Contract: ---

Lab Code: ---

Case No.: -----

SAS No.: -----

SDG No.: -----

Matrix: (soil/water) SOIL

Lab Sample ID: 420-2 101-S

Sample wt/vol: 1.2346 (g/mL) G

Lab File ID: >SA205

Level: (low/med) LOW

Date Received: 06/27/91

% Moisture: not dec. ----

Date Analyzed: 7/02/91

Column: CAP

Dilution Factor: 1.0

CONCENTRATION UNITS:

Number TICs found: 4

(ug/L or ug/Kg) ug/Kg

CAS NUMBER	COMPOUND NAME	RT	EST. CONC.	Q
1. 104767	1-Hexanol, 2-ethyl-	29.96	150.	
2.	n-Alkane	31.07	38.	
3.	n-Alkane	31.82	63.	
4.	n-Alkane	32.67	26.	
5.				
6.				
7.				
8.				
9.				
10.				
11.				
12.				
13.				
14.				
15.				
16.				
17.				
18.				
19.				
20.				
21.				
22.				
23.				
24.				
25.				
26.				
27.				
28.				
29.				
30.				

1A
VOLATILE ORGANICS ANALYSIS DATA SHEET

EPA SAMPLE NO.

91-6266

Lab. Name: BATTELLE-PNL

Contract: -----

Lab Code: -----

Case No.: -----

SAS No.: -----

SDG No.: -----

Matrix: (soil/water) SOIL

Lab Sample ID: 421-2 101-S

Sample wt/vol: 1.0818 (g/mL) G

Lab File ID: >SA206

Level: (low/med) LOW

Date Received: 06/27/91

% Moisture: not dec. ----

Date Analyzed: 7/02/91

Column: (pack/cap) CAP

Dilution Factor: 1.0

CAS NO.	COMPOUND	CONCENTRATION UNITS:		Q
		(ug/L or ug/Kg)	ug/Kg	
74-87-3-----	Chloromethane	46.	1U	
74-83-9-----	Bromomethane	46.	1U	
75-01-4-----	Vinyl Chloride	46.	1U	
75-00-3-----	Chloroethane	46.	1U	
75-09-2-----	Methylene Chloride	23.	1U	
67-64-1-----	Acetone	110.	B	
75-15-0-----	Carbon Disulfide	23.	1U	
75-35-4-----	1,1-Dichloroethene	23.	1U	
75-34-3-----	1,1-Dichloroethane	23.	1U	
540-59-0-----	1,2-Dichloroethene (total)	23.	1U	
67-66-3-----	Chloroform	23.	1U	
107-02-2-----	1,2-Dichloroethane	23.	1U	
78-93-3-----	2-Butanone	46.	1U	
71-55-6-----	1,1,1-Trichloroethane	23.	1U	
56-23-5-----	Carbon Tetrachloride	23.	1U	
108-05-4-----	Vinyl Acetate	17.	JB	
75-27-4-----	Bromodichloromethane	23.	1U	
78-87-5-----	1,2-Dichloropropane	23.	1U	
10061-01-5-----	cis-1,3-Dichloropropene	23.	1U	
79-01-6-----	Trichloroethene	23.	1U	
124-48-1-----	Dibromochloromethane	23.	1U	
79-00-5-----	1,1,2-Trichloroethane	23.	1U	
71-43-2-----	Benzene	23.	1U	
10061-02-6-----	trans-1,3-Dichloropropene	23.	1U	
75-25-2-----	Bromoform	23.	1U	
108-10-1-----	4-Methyl-2-pentanone	46.	1U	
591-78-6-----	2-Hexanone	46.	1U	
127-18-4-----	Tetrachloroethene	23.	1U	
79-34-5-----	1,1,2,2-Tetrachloroethane	23.	1U	
108-88-3-----	Toluene	23.	1U	
108-90-7-----	Chlorobenzene	23.	1U	
100-41-4-----	Ethylbenzene	23.	1U	
100-42-5-----	Styrene	23.	1U	
133-02-7-----	Xylene (total)	23.	1U	

1E
VOLATILE ORGANICS ANALYSIS DATA SHEET
TENTATIVELY IDENTIFIED COMPOUNDS

EPA SAMPLE NO.

91-6266

Lab Name: BATTELLE-PNL

Contract: -----

Lab Code: -----

Case No.: -----

SAS No.: -----

SDG No.: -----

Matrix: (soil/water) SOIL

Lab Sample ID: 421-2 101-S

Sample wt/vol: 1.0818 (g/mL) G

Lab File ID: >SA206

Level: (low/med) LOW

Date Received: 06/27/91

% Moisture: not dec.-----

Date Analyzed: 7/02/91

Column: CAP

Dilution Factor: 1.0

CONCENTRATION UNITS:
(ug/L or ug/Kg) ug/Kg

Number TICs found: 5

CAS NUMBER	COMPOUND NAME	RT	EST. CONC.	Q
1. 104767	1-Hexanol, 2-ethyl-	29.94	230.	
2.	n-Alkane	30.20	38.	
3.	n-Alkane	31.06	480.	
4.	n-Alkane	31.80	870.	
5.	n-Alkane	32.64	287.	
6.				
7.				
8.				
9.				
10.				
11.				
12.				
13.				
14.				
15.				
16.				
17.				
18.				
19.				
20.				
21.				
22.				
23.				
24.				
25.				
26.				
27.				
28.				
29.				
30.				

1A
VOLATILE ORGANICS ANALYSIS DATA SHEET

EPA SAMPLE NO.

91-6267

Lab Name: BATTELIE-PNL

Contract: -----

Lab Code: -----

Case No.: -----

SAS No.: -----

SDG No.: -----

Matrix: (soil/water) SOIL

Lab Sample ID: 563-2 101-S

Sample wt/vol: 1.1253 (g/mL) G

Lab File ID: >SA207

Level: (low/med) LOW

Date Received: 06/27/91

% Moisture: not dec.-----

Date Analyzed: 7/02/91

Column: (pack/cap) CAP

Dilution Factor: 1.0

CAS NO.	COMPOUND	CONCENTRATION UNITS: (ug/L or ug/Kg) ug/Kg	Q
---------	----------	---	---

74-87-3-----	Chloromethane	44.	IU
74-83-9-----	Bromomethane	44.	IU
75-01-4-----	Vinyl Chloride	44.	IU
75-00-3-----	Chloroethane	44.	IU
75-09-2-----	Methylene Chloride	22.	IU
67-64-1-----	Acetone	60.	B
75-15-0-----	Carbon Disulfide	22.	IU
75-35-4-----	1,1-Dichloroethene	22.	IU
75-34-3-----	1,1-Dichloroethane	22.	IU
540-59-0-----	1,2-Dichloroethene (total)	22.	IU
67-66-3-----	Chloroform	22.	IU
107-02-2-----	1,2-Dichloroethane	22.	IU
78-93-3-----	2-Butanone	110.	B
71-55-6-----	1,1,1-Trichloroethane	17.	IJ
56-23-5-----	Carbon Tetrachloride	22.	IU
108-05-4-----	Vinyl Acetate	18.	IJB
75-27-4-----	Bromodichloromethane	22.	IU
78-87-5-----	1,2-Dichloropropane	22.	IU
10061-01-5-----	cis-1,3-Dichloropropene	22.	IU
79-01-6-----	Trichloroethene	22.	IU
124-48-1-----	Dibromochloromethane	22.	IU
79-00-5-----	1,1,2-Trichloroethane	22.	IU
71-43-2-----	Benzene	22.	IU
10061-02-6-----	trans-1,3-Dichloropropene	22.	IU
75-25-2-----	Bromoform	22.	IU
108-10-1-----	4-Methyl-2-pentanone	44.	IU
591-78-6-----	2-Hexanone	44.	IU
127-18-4-----	Tetrachloroethene	22.	IU
79-34-5-----	1,1,2,2-Tetrachloroethane	22.	IU
108-88-3-----	Toluene	22.	IU
108-90-7-----	Chlorobenzene	22.	IU
100-41-4-----	Ethylbenzene	22.	IU
100-42-5-----	Styrene	22.	IU
133-02-7-----	Xylene (total)	22.	IU

1E
VOLATILE ORGANICS ANALYSIS DATA SHEET
TENTATIVELY IDENTIFIED COMPOUNDS

EPA SAMPLE NO.

91-6267

Lab Name: BATTELLE-PNL

Contract: -----

Lab Code: -----

Case No.: -----

SAS No.: -----

SDG No.: -----

Matrix: (soil/water) SOIL

Lab Sample ID: ^{5 Data 7/24/91} 63-2 101-S

Sample wt/vol: 1.1253 (g/mL) G

Lab File ID: >SA207

Level: (low/med) LOW

Date Received: 06/27/91

% Moisture: not dec. ----

Date Analyzed: 7/02/91

Column: CAP

Dilution Factor: 1.0

Number TICs found: 7

CONCENTRATION UNITS:
(ug/L or ug/Kg) ug/Kg

CAS NUMBER	COMPOUND NAME	RT	EST. CONC.	Q
1.	Unknown	28.63	26.	
2.	n-Alkane	28.79	28.	
3. 104767	1-Hexanol, 2-ethyl-	29.96	173.	
4.	n-Alkane	30.21	24.	
5.	n-Alkane	31.07	47.	
6.	n-Alkane	31.82	88.	
7.	n-Alkane	32.66	34.	
8.				
9.				
10.				
11.				
12.				
13.				
14.				
15.				
16.				
17.				
18.				
19.				
20.				
21.				
22.				
23.				
24.				
25.				
26.				
27.				
28.				
29.				
30.				

2B
SOIL VOLATILE SURROGATE RECOVERY

Lab Name: BATTELLE-PNL

Contract: -----

Lab Code: -----

Case No.: -----

SAS No.: -----

SDG No.: -----

Level: (low/med) LOW

	EPA SAMPLE NO.	S1 (TOL)#	S2 (BFB)#	S3 (DCE)#	OTHER	TOT OUT
01	DAILY CAL	101	101	100		0
02	BLANK	98	98	105		0
03	91-6264	97	101	103		0
04	91-6265	97	101	104		0
05	91-6266	97	99	116		0
06	91-6267	98	98	122 *		1
07						
08						
09						
10						
11						
12						
13						
14						
15						
16						
17						
18						
19						
20						
21						
22						
23						
24						
25						
26						
27						
28						
29						
30						

QC LIMITS

S1 (TOL) = Toluene-d8 (81-117)

S2 (BFB) = Bromofluorobenzene (74-121)

S3 (DCE) = 1,2-Dichloroethane-d4 (70-121)

Column to be used to flag recovery values

* Values outside of contract required QC limits

D Surrogates diluted out

4A
VOLATILE METHOD BLANK SUMMARY

Lab Name: Battelle-PNL Contract: -----
 Lab Code: ----- Case No.: ----- SAS No.: ----- SDG No.: -----
 Lab File ID: >SA2B3 Lab Sample ID: BLANK
 Date Analyzed: 07/02/91 Time Analyzed: 14:06
 Matrix: (soil/water) WATER Level: (low/med) LOW
 Instrument ID: 5970 #2

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

	EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	TIME ANALYZED
01	91-6265	420-2 101-SY	>SA205	17:34
02	91-6267	563-2 101-SY	>SA207	18:54
03	91-6264	424-2 101-SY	>SA204	19:49
04	91-6266	421-2 101-SY	>SA206	20:39
05				
06				
07				
08				
09				
10				
11				
12				
13				
14				
15				
16				
17				
18				
19				
20				
21				
22				
23				
24				
25				
26				
27				
28				
29				
30				

COMMENTS:

5A
VOLATILE ORGANIC GC/MS TUNING AND MASS
CALIBRATION - BROMOFLUOROBENZENE (BFB)

Lab Name: BATTELLE-PNL

Contract: ----

Lab Code: ----

Case No.: ----

SAS No.: ----

SDG No.: ----

Lab File ID: >SA201

BFB Injection Date: 7/02/91

Instrument ID: 5970 #2

BFB Injection Time: 8:55

Matrix: (soil/water) SOIL Level: (low/med) LOW Column: (pack/cap) CAP

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% OF MASS 95	19.1
75	30.0 - 60.0% OF MASS 95	52.0
95	Base peak, 100% relative abundance	100.
96	5.0 - 9.0% of mass 95	6.9
173	Less than 2.0% of mass 174	0.0(0.0)1
174	Greater than 50.0% of mass 95	78.5
175	5.0 - 9.0% of mass 174	5.2(6.6)1
176	Greater than 95.0%, but less than 101.0% of mass 174	79.0(100.8)1
177	5.0 - 9.0% of mass 176	5.1(6.5)2

1-Value is % mass 174

2-Value is % mass 176

THIS TUNE APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

	EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
01	DAILY CAL	50 PPB STD	>SA202	7/02/91	10:38
02	BLANK	BLANK	>SA2B3	7/02/91	14:06
03	91-6265	420-2 101-SY	>SA205	7/02/91	17:34
04	91-6267	563-2 101-SY	>SA207	7/02/91	18:54
05	91-6264	424-2 101-SY	>SA204	7/02/91	19:49
06	91-6266	421-2 101-SY	>SA206	7/02/91	20:39
07					
08					
09					
10					
11					
12					
13					
14					
15					
16					
17					
18					
19					
20					
21					
22					

6A
VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: BATTELLE-PNL

Contract: -----

Lab Code: -----

Case No.: -----

SAS No.: -----

SDG No.: -----

Instrument ID: 5970 #2 Calibration Date(s): 07/01/91 07/02/91

Matrix: (soil/water) WATER Level: (low/med) LOW Column: (pack/cap) CAP

Min RRF for SPCC(%) = 0.300 (0.250 for Bromoform) Max %RSD for CCC(*) = 30.0%

LAB FILE ID:		RRF20 => RC704		RRF50 => RC705			
RRF100 => RC706		RRF150 => RC707		RRF200 => RC708			
COMPOUND		RRF20	RRF50	RRF100	RRF150	RRF200	% RSD
Chloromethane_#		.722	.654	.678	.667	.629	5.1#
Bromomethane		1.324	1.221	1.284	1.290	1.224	3.5
Vinyl_Chloride*		.779	.751	.811	.803	.762	3.3*
Chloroethane		.480	.456	.471	.433	.395	7.6
Methylene_Chloride		1.105	1.119	1.148	1.134	1.083	2.2
Acetone		.294	.190	.193	.188	.186	22.2
Carbon_Disulfide		2.427	2.476	2.606	2.747	2.598	4.9
1,1-Dichloroethane*		1.080	1.060	1.127	1.123	.959	6.4*
1,1-Dichloroethane#		1.869	1.841	1.981	2.024	1.938	3.9#
1,2-Dichloroethane_(total)_		1.392	1.352	1.466	1.490	1.415	4.5
Chloroform*		2.951	2.935	3.154	3.196	3.073	3.8*
1,2-Dichloroethane		1.958	1.819	1.965	1.992	1.935	3.5
2-Butanone		.157	.081	.085	.087	.082	33.2
1,1,1-Trichloroethane		2.854	2.879	3.111	3.146	3.051	4.5
Carbon_Tetrachloride		2.891	2.864	3.138	3.218	3.082	5.1
Vinyl_Acetate		1.479	1.260	1.371	1.435	1.378	6.0
Bromodichloromethane		2.896	2.933	3.220	3.295	3.192	5.8
1,2-Dichloropropene*		.286	.277	.285	.290	.279	1.8*
cis-1,3-Dichloropropene		.507	.493	.518	.530	.517	2.7
Trichloroethane		.466	.456	.467	.477	.462	1.7
Dibromochloromethane		.827	.781	.830	.859	.844	3.5
1,1,2-Trichloroethane		.366	.322	.333	.340	.335	4.8
Benzene		.691	.685	.709	.709	.686	1.7
trans-1,3-Dichloropropene		.473	.447	.476	.477	.483	2.9
Bromoform*		.642	.530	.577	.604	.598	7.0*
4-Methyl-2-pentanone		.364	.238	.238	.240	.246	21.0
2-Hexanone		.246	.154	.164	.163	.161	21.7
Tetrachloroethane		.535	.519	.534	.535	.516	1.8
1,1,2,2-Tetrachloroethane		.804	.571	.566	.596	.587	16.1
Toluene*		.609	.600	.607	.620	.601	1.3*
Chlorobenzene#		.935	.924	.900	.958	.910	2.4#
Ethylbenzene*		.392	.387	.391	.403	.387	1.6*
Styrene		.847	.839	.819	.854	.830	1.6
Xylene_(total)		.809	.795	.797	.818	.794	1.3
Toluene-d8		1.050	1.053	1.056	1.096	1.075	1.8
Bromofluorobenzene		.788	.787	.784	.803	.784	1.0
1,2-Dichloroethane-d4		1.701	1.613	1.697	1.801	1.734	4.0

7A
VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: BATTELLE-PNL

Contract: -----

Lab Code: -----

Case No.: -----

SAS No.: -----

SDG No.: -----

Instrument ID: 5970 #2 Calibration Date: 7/02/91 Time: 10:38

Lab File ID: >SA202 Init. Calib. Date(s): 07/01/91 07/02/91

Matrix: (soil/water) WATER Level: (low/med) LOW Column: (pack/cap) CAP

Min RRF50 for SPCC(*) = 0.300 (0.250 for Bromoform) Max %D for CCC(*) = 25.0%.

COMPOUND	RRF	RRF50	%D
Chloromethane	.670	.684	2.2
Bromomethane	1.269	1.299	2.4
Vinyl Chloride	.781	.828	5.9
Chloroethane	.447	.471	5.5
Methylene_Chloride	1.118	1.155	3.3
Acetone	.210	.202	3.9
Carbon_Disulfide	2.571	2.765	7.5
1,1-Dichloroethane	1.070	1.127	5.4
1,1-Dichloroethane	1.931	1.969	2.0
1,2-Dichloroethane_(total)	1.423	1.424	.1
Chloroform	3.062	3.166	3.4
1,2-Dichloroethane	1.934	1.957	1.2
2-Butanone	.099	.086	12.6
1,1,1-Trichloroethane	3.008	3.057	1.6
Carbon_Tetrachloride	3.038	3.090	1.7
Vinyl Acetate	1.385	1.318	4.8
Bromodichloromethane	3.107	3.179	2.3
1,2-Dichloropropane	.283	.289	1.9
cis-1,3-Dichloropropene	.513	.529	3.1
Trichloroethane	.466	.481	3.3
Dibromochloromethane	.828	.852	2.9
1,1,2-Trichloroethane	.339	.345	1.7
Benzene	.696	.718	3.2
trans-1,3-Dichloropropene	.471	.488	3.5
Bromoform	.590	.564	4.5
4-Methyl-2-pentanone	.265	.233	12.0
2-Hexanone	.178	.158	11.0
Tetrachloroethene	.528	.534	1.1
1,1,2,2-Tetrachloroethane	.625	.582	6.8
Toluene	.607	.631	3.9
Chlorobenzene	.925	.974	5.2
Ethylbenzene	.392	.397	1.2
Styrene	.838	.875	4.4
Xylene_(total)	.802	.831	3.6
Toluene-d8	1.066	1.118	4.9
Bromofluorobenzene	.789	.829	5.0
1,2-Dichloroethane-d4	1.709	1.719	.5

8A
VOLATILE INTERNAL STANDARD AREA SUMMARY

Lab Name: BATTELLE-PNL

Contract: -----

Lab Code: -----

Case No.: -----

SAS No.: -----

SDG No.: -----

Lab File ID (Standard): >SA202

Date Analyzed: 7/02/91

Instrument ID: 5970 #2

Time Analyzed: 10:38

Matrix: (soil/water) WATER Level: (low/med) LOW

Column: (pack/cap) CAP

	IS1 (BCM)	RT	IS2 (DFB)	RT	IS3 (CBZ)	RT
	AREA #		AREA #		AREA #	
12 HOUR STD	43116.	13.46	168210.	16.78	141698.	24.40
UPPER LIMIT	86232.		336420.		283396.	
LOWER LIMIT	21558.		84105.		70849.	
EPA SAMPLE NO.						
01 BLANK	40901.	13.32	170544.	16.72	144049.	24.37
02 91-6264	46698.	13.07	182398.	16.49	153042.	24.18
03 91-6265	45447.	13.16	178737.	16.57	151661.	24.24
04 91-6266	39088.	13.08	182441.	16.49	152539.	24.17
05 91-6267	40839.	13.11	185434.	16.53	156557.	24.20
06						
07						
08						
09						
10						
11						
12						
13						
14						
15						
16						
17						
18						
19						
20						
21						
22						

IS1 (BCM) = Bromochloromethane
IS2 (DFB) = 1,4-Difluorobenzene
IS3 (CBZ) = Chlorobenzene-d5

UPPER LIMIT = + 100%
of internal standard area.
LOWER LIMIT = - 50%
of internal standard area.

Column used to flag internal standard area values with an asterisk

SAMPLE DELIVERY GROUP 2

A summary report of the analyses of samples in Delivery Group 2 provides general information on the samples, sample preparation, analysis method, and location of data files. The summary report also defines codes used in the CLP report forms, which follow.

**ORGANIC COMPOUND ANALYSIS REPORT
VOLATILE COMPOUNDS**

SAMPLE ANALYSIS REPORTED

Analysis of four water samples for volatile organic compounds by gas chromatography/mass spectrometry (GC/MS) is the subject of this report.

SAMPLE DESCRIPTION

<u>Sample ID</u>	<u>ACL Lab Number</u>
578	91-6425
579	91-6426
580	91-6427
581	91-6428

The samples were received in the 325 Building "B" Analytical Hot Cell with substantial headspace, non-septa type closures, and at ambient temperature on 7/10/91. These sample handling procedures are not consistent with proper analytical protocol for this analysis.

SAMPLE PREPARATION

GC/FID volatile hexadecane screens were performed and indicated a dilution of the sample was not necessary. Five mL of each of the samples were added to a purge vessel attached to a Tekmar Liquid Sample Concentrator. The analysis followed EPA-CLP SOW 2/88 procedures for the analysis of volatile compounds.

ANALYSIS METHOD

- GC/MS procedure: PNL-ALO-335.
- GC/MS instrumentation: HP-5890/5970 GC/MS (WB46864).
- GC/MS location: Lab 427, 325 building.

- GC screening analysis: PNL-ALO-340.
- GC/FID instrumentation: HP-5890 (WB60701).
- GC/FID location: Lab 330, 325 building.

QUALITY CONTROL

Quality control procedures specified for this method were followed. While these forms are not all included in this report, the quality assurance performance requirements are summarized below:

Form Information

Comments

2A Surrogate Recovery

Meets all requirement

3A	MS/MSD Recovery	Analysis not requested.
4A	Method Blank Summary	Meets all requirements.
5A	Tune/Mass Calibration	Meets all requirements.
6A	Initial Calibration	5 point calibration. Meets all requirements.
7A	Daily Calibration	Meets all requirements.
8A	Internal Standards	Meets all requirements.

DATA

The data and calibration are archived on magnetic tape in the 325 building, 327-A laboratory. The following is the list of pertinent files:

<u>File Name</u>	<u>Sample Number</u>	<u>Sample Analyzed</u>
>SB801		Mass Calibration/Tune Check
>SB802		Daily Calibration
>SB803		Method Blank
>SB804	91-6425	578
>SB805	91-6426	579
>SB806	91-6427	580
>SB807	91-6428	581
>SB808	91-6428 DUP	581 DUP

Lab data are also maintained as follows:

<u>Activity</u>	<u>LRB Number</u>	<u>Page Number</u>
GC/MS injection log	BNW-52907	80-81

RESULTS

CLP Target Compounds: The attached 1A Forms show that volatile target compounds, Acetone, and 2-Butanone, were detected above the Contract Required Quantitation Limits, (CRQL's), in sample 578, but not in the other three samples, (579, 580, and 581). 2-Butanone and Vinyl Acetate were observed in the blank, and the other three samples, at similar levels, below the CRQL, and are "B-flagged".

The following defines the qualifiers, Q-flags, in the Form 1's:

<u>"Q" Flag</u>	<u>Definition</u>
-----------------	-------------------

- U Indicates the compound was analyzed for but not detected, the U-flagged concentration number is the CRQL.
- J Indicates an estimated value for the target or tentatively identified compounds, spectra meet criteria but response is below the CRQL for the target compounds.
- B Compound was found in the blank.
- X Indicates compound was manually deleted because all requirements were not met.
- D Analysis was performed on a diluted sample.
- E Indicates that Quantitation was outside the calibration range.
- A Aldol condensation product.

Tentatively Identified Compounds: As shown in the attached 1E Forms, no non-EPA target compounds were observed in the blank or the samples.

ANALYST 924 Rorr DATE 9/18/91 REVIEW [Signature] DATE 9-18-91

1A
VOLATILE ORGANICS ANALYSIS DATA SHEET

EPA SAMPLE NO.

Lab Name: BATTELLE-PNL

Contract: -----

MET BLANK

Lab Code: ----- Case No.: ----- SAS No.: ----- SDG No.: -----

Matrix: (soil/water) WATER

Lab Sample ID: LAB BLANK

Sample wt/vol: 5 (g/mL) mL

Lab File ID: >SB803

Level: (low/med) LOW

Date Received: 07/11/91

% Moisture: not dec.-----

Date Analyzed: 7/18/91

Column: (pack/cap) CAP

Dilution Factor: 1.0

CAS NO.	COMPOUND	CONCENTRATION UNITS: (ug/L or ug/Kg) ug/L	Q
74-87-3-----	Chloromethane	10.	IU
74-83-9-----	Bromomethane	10.	IU
75-01-4-----	Vinyl Chloride	10.	IU
75-00-3-----	Chloroethane	10.	IU
75-09-2-----	Methylene Chloride	5.	IU
67-64-1-----	Acetone	10.	IU
75-15-0-----	Carbon Disulfide	5.	IU
75-35-4-----	1,1-Dichloroethene	5.	IU
75-34-3-----	1,1-Dichloroethane	5.	IU
540-59-0-----	1,2-Dichloroethene (total)	5.	IU
67-66-3-----	Chloroform	5.	IU
107-02-2-----	1,2-Dichloroethane	5.	IU
78-93-3-----	2-Butanone	10.	IJB
71-55-6-----	1,1,1-Trichloroethane	5.	IU
56-23-5-----	Carbon Tetrachloride	5.	IU
108-05-4-----	Vinyl Acetate	3.	IJB
75-27-4-----	Bromodichloromethane	5.	IU
78-87-5-----	1,2-Dichloropropane	5.	IU
10061-01-5-----	cis-1,3-Dichloropropene	5.	IU
79-01-6-----	Trichloroethene	5.	IU
124-48-1-----	Dibromochloromethane	5.	IU
79-00-5-----	1,1,2-Trichloroethane	5.	IU
71-43-2-----	Benzene	5.	IU
10061-02-6-----	trans-1,3-Dichloropropene	5.	IU
75-25-2-----	Bromoform	5.	IU
108-10-1-----	4-Methyl-2-pentanone	10.	IU
591-78-6-----	2-Hexanone	10.	IU
127-18-4-----	Tetrachloroethene	5.	IU
79-34-5-----	1,1,2,2-Tetrachloroethane	5.	IU
108-88-3-----	Toluene	5.	IU
108-90-7-----	Chlorobenzene	5.	IU
100-41-4-----	Ethylbenzene	5.	IU
100-42-5-----	Styrene	5.	IU
133-02-7-----	Xylene (total)	5.	IU

1E
VOLATILE ORGANICS ANALYSIS DATA SHEET
TENTATIVELY IDENTIFIED COMPOUNDS

EPA SAMPLE NO.

MET BLANK

Lab Name: BATTELLE-PNL

Contract: -----

Lab Code: -----

Case No.: -----

SAS No.: -----

SDG No.: -----

Matrix: (soil/water) WATER

Lab Sample ID: LAB BLANK

Sample wt/vol: 5 (g/mL) mL

Lab File ID: >SB803

Level: (low/med) LDW

Date Received: 07/11/91

% Moisture: not dec.----

Date Analyzed: 7/18/91

Column: CAP

Dilution Factor: 1.0

Number TICs found: 0

CONCENTRATION UNITS:
(ug/L or ug/Kg) ug/L

CAS NUMBER	COMPOUND NAME	RT	EST. CONC.	Q
1.				
2.				
3.				
4.				
5.				
6.				
7.				
8.				
9.				
10.				
11.				
12.				
13.				
14.				
15.				
16.				
17.				
18.				
19.				
20.				
21.				
22.				
23.				
24.				
25.				
26.				
27.				
28.				
29.				
30.				

1A
VOLATILE ORGANICS ANALYSIS DATA SHEET

EPA SAMPLE NO.

Lab Name: BATTELLE-PNL

Contract: -----

91-6425

Lab Code: -----

Case No.: -----

SAS No.: -----

SDG No.: -----

Matrix: (soil/water) WATER

Lab Sample ID: 578

Sample wt/vol: 5 (g/mL) mL

Lab File ID: >SB804

Level: (low/med) LOW

Date Received: 07/11/91

% Moisture: not dec. ----

Date Analyzed: 7/18/91

Column: (pack/cap) CAP

Dilution Factor: 1.0

CAS NO.	COMPOUND	CONCENTRATION UNITS: (ug/L or ug/Kg) ug/L	Q
74-87-3-----	Chloromethane	10.	U
74-83-9-----	Bromomethane	10.	U
75-01-4-----	Vinyl Chloride	10.	U
75-00-3-----	Chloroethane	10.	U
75-09-2-----	Methylene Chloride	5.	U
67-64-1-----	Acetone	24.	U
75-15-0-----	Carbon Disulfide	5.	U
75-35-4-----	1,1-Dichloroethene	5.	U
75-34-3-----	1,1-Dichloroethane	5.	U
540-59-0-----	1,2-Dichloroethene (total)	5.	U
67-66-3-----	Chloroform	5.	U
107-02-2-----	1,2-Dichloroethane	5.	U
78-93-3-----	2-Butanone	14.	B
71-55-6-----	1,1,1-Trichloroethane	5.	U
56-23-5-----	Carbon Tetrachloride	5.	U
108-05-4-----	Vinyl Acetate	3.	JB
75-27-4-----	Bromodichloromethane	5.	U
78-87-5-----	1,2-Dichloropropane	5.	U
10061-01-5-----	cis-1,3-Dichloropropene	5.	U
79-01-6-----	Trichloroethene	5.	U
124-48-1-----	Dibromochloromethane	5.	U
79-00-5-----	1,1,2-Trichloroethane	5.	U
71-43-2-----	Benzene	5.	U
10061-02-6-----	trans-1,3-Dichloropropene	5.	U
75-25-2-----	Bromoform	5.	U
108-10-1-----	4-Methyl-2-pentanone	10.	U
591-78-6-----	2-Hexanone	10.	U
127-18-4-----	Tetrachloroethene	5.	U
79-34-5-----	1,1,2,2-Tetrachloroethane	5.	U
108-88-3-----	Toluene	1.	J
108-90-7-----	Chlorobenzene	5.	U
100-41-4-----	Ethylbenzene	5.	U
100-42-5-----	Styrene	5.	U
133-02-7-----	Xylene (total)	5.	U

1E
VOLATILE ORGANICS ANALYSIS DATA SHEET
TENTATIVELY IDENTIFIED COMPOUNDS

EPA SAMPLE NO.

91-6425

Lab Name: BATTELLE-PNL

Contract: -----

Lab Code: -----

Case No.: -----

SAS No.: -----

SDG No.: -----

Matrix: (soil/water) WATER

Lab Sample ID: 578

Sample wt/vol: 5 (g/mL) mL

Lab File ID: >SB804

Level: (low/med) LOW

Date Received: 07/11/91

% Moisture: not dec.-----

Date Analyzed: 07/18/91

Column: CAP

Dilution Factor: 1.00

Number TICs found: 0

CONCENTRATION UNITS:
(ug/L or ug/Kg) ug/L

CAS NUMBER	COMPOUND NAME	RT	EST. CONC.	Q
1.				
2.				
3.				
4.				
5.				
6.				
7.				
8.				
9.				
10.				
11.				
12.				
13.				
14.				
15.				
16.				
17.				
18.				
19.				
20.				
21.				
22.				
23.				
24.				
25.				
26.				
27.				
28.				
29.				
30.				

1A
VOLATILE ORGANICS ANALYSIS DATA SHEET

EPA SAMPLE NO.

91-6426

Lab Name: BATTELLE-PNL

Contract: -----

Lab Code: -----

Case No.: -----

SAS No.: -----

SDG No.: -----

Matrix: (soil/water) WATER

Lab Sample ID: 579

Sample wt/vol: 5 (g/mL) mL

Lab File ID: >SB805

Level: (low/med) LOW

Date Received: 07/11/91

% Moisture: not dec. ----

Date Analyzed: 7/18/91

Column: (pack/cap) CAP

Dilution Factor: 1.0

CAS NO.	COMPOUND	CONCENTRATION UNITS: (ug/L or ug/Kg) ug/L	Q
74-87-3	Chloromethane	10.	IU
74-83-9	Bromomethane	10.	IU
75-01-4	Vinyl Chloride	10.	IU
75-00-3	Chloroethane	10.	IU
75-09-2	Methylene Chloride	2.	IJ
67-64-1	Acetone	10.	IU
75-15-0	Carbon Disulfide	5.	IU
75-35-4	1,1-Dichloroethene	5.	IU
75-34-3	1,1-Dichloroethane	5.	IU
540-59-0	1,2-Dichloroethene (total)	5.	IU
67-66-3	Chloroform	5.	IU
107-02-2	1,2-Dichloroethane	5.	IU
78-93-3	2-Butanone	10.	IJB
71-55-6	1,1,1-Trichloroethane	5.	IU
56-23-5	Carbon Tetrachloride	5.	IU
108-05-4	Vinyl Acetate	3.	IJB
75-27-4	Bromodichloromethane	5.	IU
78-87-5	1,2-Dichloropropane	5.	IU
10061-01-5	cis-1,3-Dichloropropene	5.	IU
79-01-6	Trichloroethene	5.	IU
124-48-1	Dibromochloromethane	5.	IU
79-00-5	1,1,2-Trichloroethane	5.	IU
71-43-2	Benzene	5.	IU
10061-02-6	trans-1,3-Dichloropropene	5.	IU
75-25-2	Bromoform	5.	IU
108-10-1	4-Methyl-2-pentanone	10.	IU
591-78-6	2-Hexanone	10.	IU
127-18-4	Tetrachloroethene	5.	IU
79-34-5	1,1,2,2-Tetrachloroethane	5.	IU
108-88-3	Toluene	5.	IU
108-90-7	Chlorobenzene	5.	IU
100-41-4	Ethylbenzene	5.	IU
100-42-5	Styrene	5.	IU
133-02-7	Xylene (total)	5.	IU

1E
VOLATILE ORGANICS ANALYSIS DATA SHEET
TENTATIVELY IDENTIFIED COMPOUNDS

EPA SAMPLE NO.

91-6426

Lab Name: BATTELLE-PNL

Contract: -----

Lab Code: -----

Case No.: -----

SAS No.: -----

SDG No.: -----

Matrix: (soil/water) WATER

Lab Sample ID: 579

Sample wt/vol: 5 (g/mL) mL

Lab File ID: >SB805

Level: (low/med) LOW

Date Received: 07/11/91

% Moisture: not dec.-----

Date Analyzed: 7/18/91

Column: CAP

Dilution Factor: 1.0

Number TICs found: 0

CONCENTRATION UNITS:
(ug/L or ug/Kg) ug/L

CAS NUMBER	COMPOUND NAME	RT	EST. CONC.	Q
1.				
2.				
3.				
4.				
5.				
6.				
7.				
8.				
9.				
10.				
11.				
12.				
13.				
14.				
15.				
16.				
17.				
18.				
19.				
20.				
21.				
22.				
23.				
24.				
25.				
26.				
27.				
28.				
29.				
30.				

1A
VOLATILE ORGANICS ANALYSIS DATA SHEET

EPA SAMPLE NO.

91-6427

Lab Name: BATTELLE-PNL

Contract: -----

Lab Code: -----

Case No.: -----

SAS No.: -----

SDG No.: -----

Matrix: (soil/water) WATER

Lab Sample ID: 580

Sample wt/vol: 5 (g/mL) mL

Lab File ID: >SB806

Level: (low/med) LOW

Date Received: 07/11/91

% Moisture: not dec.-----

Date Analyzed: 7/18/91

Column: (pack/cap) CAP

Dilution Factor: 1.0

CAS NO.	COMPOUND	CONCENTRATION UNITS: (ug/L or ug/Kg) ug/L	Q
74-87-3-----	Chloromethane	10.	U
74-83-9-----	Bromomethane	10.	U
75-01-4-----	Vinyl Chloride	10.	U
75-00-3-----	Chloroethane	10.	U
75-09-2-----	Methylene Chloride	1.	J
67-64-1-----	Acetone	10.	U
75-15-0-----	Carbon Disulfide	5.	U
75-35-4-----	1,1-Dichloroethene	5.	U
75-34-3-----	1,1-Dichloroethane	5.	U
540-59-0-----	1,2-Dichloroethene (total)	5.	U
67-66-3-----	Chloroform	5.	U
107-02-2-----	1,2-Dichloroethane	5.	U
78-93-3-----	2-Butanone	10.	U
71-55-6-----	1,1,1-Trichloroethane	5.	U
56-23-5-----	Carbon Tetrachloride	5.	U
108-05-4-----	Vinyl Acetate	3.	JB
75-27-4-----	Bromodichloromethane	5.	U
78-87-5-----	1,2-Dichloropropene	5.	U
10061-01-5-----	cis-1,3-Dichloropropene	5.	U
79-01-6-----	Trichloroethene	5.	U
124-48-1-----	Dibromochloromethane	5.	U
79-06-5-----	1,1,2-Trichloroethane	5.	U
71-43-2-----	Benzene	5.	U
10061-02-6-----	trans-1,3-Dichloropropene	5.	U
75-25-2-----	Bromoform	5.	U
108-10-1-----	4-Methyl-2-pentanone	10.	U
591-78-6-----	2-Hexanone	10.	U
127-18-4-----	Tetrachloroethene	5.	U
79-34-5-----	1,1,2,2-Tetrachloroethane	5.	U
108-88-3-----	Toluene	5.	U
108-90-7-----	Chlorobenzene	5.	U
100-41-4-----	Ethylbenzene	5.	U
100-42-5-----	Styrene	5.	U
133-02-7-----	Xylene (total)	5.	U

1E
VOLATILE ORGANICS ANALYSIS DATA SHEET
TENTATIVELY IDENTIFIED COMPOUNDS

EPA SAMPLE NO.

91-6427

Lab Name: BATTELLE-PNL

Contract: -----

Lab Code: -----

Case No.: -----

SAS No.: -----

SDG No.: -----

Matrix: (soil/water) WATER

Lab Sample ID: 580

Sample wt/vol: 5 (g/mL) mL

Lab File ID: >SB806

Level: (low/med) LOW

Date Received: 07/11/91

% Moisture: not dec. ----

Date Analyzed: 7/18/91

Column: CAP

Dilution Factor: 1.0

Number TICs found: 0

CONCENTRATION UNITS:
(ug/L or ug/Kg) ug/L

CAS NUMBER	COMPOUND NAME	RT	EST. CONC.	Q
1.				
2.				
3.				
4.				
5.				
6.				
7.				
8.				
9.				
10.				
11.				
12.				
13.				
14.				
15.				
16.				
17.				
18.				
19.				
20.				
21.				
22.				
23.				
24.				
25.				
26.				
27.				
28.				
29.				
30.				

1A
VOLATILE ORGANICS ANALYSIS DATA SHEET

EPA SAMPLE NO.

91-6428

Lab Name: BATTELLE-PNL

Contract: -----

Lab Code: -----

Case No.: -----

SAS No.: -----

SDG No.: -----

Matrix: (soil/water) WATER

Lab Sample ID: 581

Sample wt/vol: 5 (g/mL) mL

Lab File ID: >SB807

Level: (low/med) LOW

Date Received: 07/11/91

% Moisture: not dec.-----

Date Analyzed: 7/18/91

Column: (pack/cap) CAP

Dilution Factor: 1.0

CAS NO.	COMPOUND	CONCENTRATION UNITS:	
		(ug/L or ug/Kg)	ug/L
			Q
74-87-3-----	Chloromethane	10.	IU
74-83-9-----	Bromomethane	10.	IU
75-01-4-----	Vinyl Chloride	10.	IU
75-00-3-----	Chloroethane	10.	IU
75-09-2-----	Methylene Chloride	1.	IJ
67-64-1-----	Acetone	10.	IU
75-15-0-----	Carbon Disulfide	5.	IU
75-35-4-----	1,1-Dichloroethene	5.	IU
75-34-3-----	1,1-Dichloroethane	5.	IU
540-59-0-----	1,2-Dichloroethene (total)	5.	IU
67-66-3-----	Chloroform	5.	IU
107-02-2-----	1,2-Dichloroethane	5.	IU
78-93-3-----	2-Butanone	10.	IU
71-55-6-----	1,1,1-Trichloroethane	5.	IU
56-23-5-----	Carbon Tetrachloride	5.	IU
108-05-4-----	Vinyl Acetate	3.	IJB
75-27-4-----	Bromodichloromethane	5.	IU
78-87-5-----	1,2-Dichloropropane	5.	IU
10061-01-5-----	cis-1,3-Dichloropropene	5.	IU
79-01-6-----	Trichloroethene	5.	IU
124-48-1-----	Dibromochloromethane	5.	IU
79-00-5-----	1,1,2-Trichloroethane	5.	IU
71-43-2-----	Benzene	5.	IU
10061-02-6-----	trans-1,3-Dichloropropene	5.	IU
75-25-2-----	Bromoform	5.	IU
108-10-1-----	4-Methyl-2-pentanone	10.	IU
591-78-6-----	2-Hexanone	10.	IU
127-18-4-----	Tetrachloroethene	5.	IU
79-34-5-----	1,1,2,2-Tetrachloroethane	5.	IU
108-88-3-----	Toluene	5.	IU
108-90-7-----	Chlorobenzene	5.	IU
100-41-4-----	Ethylbenzene	5.	IU
100-42-5-----	Styrene	5.	IU
133-02-7-----	Xylene (total)	5.	IU

1E
VOLATILE ORGANICS ANALYSIS DATA SHEET
TENTATIVELY IDENTIFIED COMPOUNDS

EPA SAMPLE NO.

91-6428

Lab Name: BATTELLE-PNL

Contract: -----

Lab Code: -----

Case No.: -----

SAS No.: -----

SDG No.: -----

Matrix: (soil/water) WATER

Lab Sample ID: 581

Sample wt/vol: 5 (g/mL) mL

Lab File ID: >SB807

Level: (low/med) LOW

Date Received: 07/11/91

% Moisture: not dec.-----

Date Analyzed: 7/18/91

Column: CAP

Dilution Factor: 1.0

Number TICs found: 0

CONCENTRATION UNITS:
(ug/L or ug/Kg) ug/L

CAS NUMBER	COMPOUND NAME	RT	EST. CONC.	Q
1.				
2.				
3.				
4.				
5.				
6.				
7.				
8.				
9.				
10.				
11.				
12.				
13.				
14.				
15.				
16.				
17.				
18.				
19.				
20.				
21.				
22.				
23.				
24.				
25.				
26.				
27.				
28.				
29.				
30.				

1A
VOLATILE ORGANICS ANALYSIS DATA SHEET

EPA SAMPLE NO.

Lab Name: BATTELLE-PNL

Contract: -----

91-6428 DUP

Lab Code: -----

Case No.: -----

SAS No.: -----

SDG No.: -----

Matrix: (soil/water) WATER

Lab Sample ID: 581 DUP

Sample wt/Vol: 5 (g/mL) mL

Lab File ID: >SB808

Level: (low/med) LOW

Date Received: 07/11/91

% Moisture: not dec. ----

Date Analyzed: 7/18/91

Column: (pack/cap) CAP

Dilution Factor: 1.0

CAS NO.	COMPOUND	CONCENTRATION UNITS: (ug/L or ug/Kg) ug/L	Q
74-87-3-----	Chloromethane	10.	U
74-83-9-----	Bromomethane	10.	U
75-01-4-----	Vinyl Chloride	10.	U
75-00-3-----	Chloroethane	10.	U
75-09-2-----	Methylene Chloride	5.	U
67-64-1-----	Acetone	10.	U
75-15-0-----	Carbon Disulfide	5.	U
75-35-4-----	1,1-Dichloroethene	5.	U
75-34-3-----	1,1-Dichloroethane	5.	U
540-59-0-----	1,2-Dichloroethene (total)	5.	U
67-66-3-----	Chloroform	5.	U
107-02-2-----	1,2-Dichloroethane	5.	U
78-93-3-----	2-Butanone	10.	U
71-55-6-----	1,1,1-Trichloroethane	5.	U
56-23-5-----	Carbon Tetrachloride	5.	U
108-05-4-----	Vinyl Acetate	3.	JB
75-27-4-----	Bromodichloromethane	5.	U
78-87-5-----	1,2-Dichloropropane	5.	U
10061-01-5-----	cis-1,3-Dichloropropene	5.	U
79-01-6-----	Trichloroethene	5.	U
124-48-1-----	Dibromochloromethane	5.	U
79-00-5-----	1,1,2-Trichloroethane	5.	U
71-43-2-----	Benzene	5.	U
10061-02-6-----	trans-1,3-Dichloropropene	5.	U
75-25-2-----	Bromoform	5.	U
108-10-1-----	4-Methyl-2-pentanone	10.	U
591-78-6-----	2-Hexanone	10.	U
127-18-4-----	Tetrachloroethene	5.	U
75-34-5-----	1,1,2,2-Tetrachloroethane	5.	U
108-88-3-----	Toluene	5.	U
108-90-7-----	Chlorobenzene	5.	U
100-41-4-----	Ethylbenzene	5.	U
100-42-5-----	Styrene	5.	U
133-02-7-----	Xylene (total)	5.	U

1E
VOLATILE ORGANICS ANALYSIS DATA SHEET
TENTATIVELY IDENTIFIED COMPOUNDS

EPA SAMPLE NO.

91-6428 DUP

Lab Name: BATTELLE-PNL

Contract: -----

Lab Code: -----

Case No.: -----

SAS No.: -----

SDG No.: -----

Matrix: (soil/water) WATER

Lab Sample ID: 581 DUP

Sample wt/vol: 5 (g/mL) mL

Lab File ID: >SB808

Level: (low/med) LOW

Date Received: 07/11/91

% Moisture: not dec.-----

Date Analyzed: 7/18/91

Column: CAP

Dilution Factor: 1.0

Number TICs found: 0

CONCENTRATION UNITS:
(ug/L or ug/Kg) ug/L

CAS NUMBER	COMPOUND NAME	RT	EST. CONC.	Q
1.				
2.				
3.				
4.				
5.				
6.				
7.				
8.				
9.				
10.				
11.				
12.				
13.				
14.				
15.				
16.				
17.				
18.				
19.				
20.				
21.				
22.				
23.				
24.				
25.				
26.				
27.				
28.				
29.				
30.				

2A
WATER VOLATILE SURROGATE RECOVERY

Lab Name: BATTELLE-PNL

Contract: -----

Lab Code: ----- Case No.: ----- SAS No.: ----- SDG No.: -----

	EPA SAMPLE NO.	S1 (TOL)*	S2 (BFB)*	S3 (DCE)*	OTHER	TOT TOT
01	DAILY CAL	97	96	100		0
02	MET BLANK	101	101	88		0
03	91-6425	101	100	90		0
04	91-6426	102	101	90		0
05	91-6427	105	102	91		0
06	91-6428	101	102	91		0
07	91-6428 DU	103	103	91		0
08						
09						
10						
11						
12						
13						
14						
15						
16						
17						
18						
19						
20						
21						
22						
23						
24						
25						
26						
27						
28						
29						
30						

QC LIMITS

S1 (TOL) = Toluene-d8 (88-110)
 S2 (BFB) = Bromofluorobenzene (86-115)
 S3 (DCE) = 1,2-Dichloroethane-d4 (76-114)

* Column to be used to flag recovery values

* Values outside of contract required QC limits

D Surrogates diluted out

4A
VOLATILE METHOD BLANK SUMMARY

Lab Name: BATTELLE-PNL Contract: -----

Lab Code: ----- Case No.: ----- SAS No.: ----- SDG No.: -----

Lab File ID: >SB803 Lab Sample ID: LAB BLANK

Date Analyzed: 7/18/91 Time Analyzed: 14:17

Matrix: (soil/water) WATER Level: (low/med) LOW

Instrument ID: 5970 #2

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

	EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	TIME ANALYZED
01	91-6425	578	>SB804	15:41
02	91-6426	579	>SB805	16:31
03	91-6427	580	>SB806	17:20
04	91-6428	581	>SB807	18:12
05	91-6428 DUP	581 DUP	>SB808	19:02
06				
07				
08				
09				
10				
11				
12				
13				
14				
15				
16				
17				
18				
19				
20				
21				
22				
23				
24				
25				
26				
27				
28				
29				
30				

COMMENTS:

5A
VOLATILE ORGANIC GC/MS TUNING AND MASS
CALIBRATION - BROMOFLUOROBENZENE (BFB)

Lab Name: BATTELLE-PNL

Contract: -----

Lab Code: -----

Case No.: -----

SAS No.: -----

SDG No.: -----

Lab File ID: >SB801

BFB Injection Date: 7/18/91

Instrument ID: 5970 #2

BFB Injection Time: 9:04

Matrix: (soil/water) WATER Level: (low/med) LOW Column: (pack/cap) CAP

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% OF MASS 95	20.8
75	30.0 - 60.0% OF MASS 95	48.5
95	Base peak, 100% relative abundance	100.
96	5.0 - 9.0% of mass 95	7.2
173	Less than 2.0% of mass 174	0.0(0.0)1
174	Greater than 50.0% of mass 95	90.5
175	5.0 - 9.0% of mass 174	7.6(8.4)1
176	Greater than 95.0%, but less than 101.0% of mass 174	88.0(97.3)1
177	5.0 - 9.0% of mass 176	6.1(7.0)2

1-Value is % mass 174 2-Value is % mass 176

THIS TUNE APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA	LAB	DATE	TIME
SAMPLE NO.	SAMPLE ID	FILE ID	ANALYZED
01	DAILY CAL	50 PPB STD	>SB802
02	MET BLANK	LAB BLANK	>SB803
03	91-6426	579	>SB805
04	91-6427	580	>SB806
05	91-6428	581	>SB807
06	91-6428 DUP	581 DUP	>SB808
07			
08			
09			
10			
11			
12			
13			
14			
15			
16			
17			
18			
19			
20			
21			
22			

6A
VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: BATTELLE-PNL

Contract: -----

Lab Code: -----

Case No.: -----

SAS No.: -----

SDG No.: -----

Instrument ID: 5970 #2 Calibration Date(s): 07/01/91 07/18/91

Matrix: (soil/water) WATER Level: (low/med) LOW Column: (pack/cap) CAP

Min RRF for SPOC(+) = 0.300 (0.250 for Bromoform) Max %RSD for C0C(*) = 30.0%

LAB FILE ID:	RRF20 => RC704	RRF50 => RC705
RRF100=>RC706	RRF150=>RC707	RRF200=>RC708

COMPOUND	RRF20	RRF50	RRF100	RRF150	RRF200	RRF	% RSD
Chloromethane_#	.722	.654	.678	.667	.629	.670	5.1#
Bromomethane_#	1.324	1.221	1.284	1.290	1.224	1.269	3.5
Vinyl_Chloride_*	.779	.751	.811	.803	.762	.781	3.3*
Chloroethane_#	.480	.456	.471	.433	.395	.447	7.6
Methylene_Chloride_#	1.105	1.119	1.148	1.134	1.083	1.118	2.2
Acetone_#	.294	.190	.193	.188	.186	.210	22.2
Carbon_Disulfide_#	2.427	2.476	2.606	2.747	2.598	2.571	4.9
1,1-Dichloroethene_*	1.080	1.060	1.127	1.123	.959	1.070	6.4*
1,1-Dichloroethane_#	1.869	1.841	1.981	2.024	1.938	1.931	3.9#
1,2-Dichloroethene_(total)_#	1.392	1.352	1.466	1.490	1.415	1.423	4.5
Chloroform_*	2.951	2.935	3.154	3.196	3.073	3.062	3.8*
1,2-Dichloroethane_#	1.958	1.819	1.965	1.992	1.935	1.934	3.5
2-Butanone_#	.157	.081	.085	.087	.082	.099	33.2
1,1,1-Trichloroethane_#	2.854	2.879	3.111	3.146	3.051	3.008	4.5
Carbon_Tetrachloride_#	2.891	2.864	3.138	3.218	3.082	3.038	5.1
Vinyl_Acetate_#	1.479	1.260	1.371	1.435	1.378	1.385	6.0
Bromodichloromethane_#	2.896	2.933	3.220	3.295	3.192	3.107	5.8
1,2-Dichloropropane_*	.286	.277	.285	.290	.279	.283	1.8*
cis-1,3-Dichloropropene_#	.507	.493	.518	.530	.517	.513	2.7
Trichloroethene_#	.466	.456	.467	.477	.462	.466	1.7
Dibromochloromethane_#	.827	.781	.830	.859	.844	.828	3.5
1,1,2-Trichloroethane_#	.366	.322	.333	.340	.335	.339	4.8
Benzene_#	.691	.685	.709	.709	.686	.696	1.7
trans-1,3-Dichloropropene_#	.473	.447	.476	.477	.483	.471	2.9
Bromoform_#	.642	.530	.577	.604	.598	.590	7.0#
4-Methyl-2-pentanone_#	.364	.238	.238	.240	.246	.265	21.0
2-Hexanone_#	.246	.154	.164	.163	.161	.178	21.7
Tetrachloroethene_#	.535	.519	.534	.535	.516	.528	1.8
1,1,2,2-Tetrachloroethane_#	.804	.571	.566	.596	.587	.625	16.1
Toluene_*	.609	.600	.607	.620	.601	.607	1.3*
Chlorobenzene_#	.935	.924	.900	.958	.910	.925	2.4#
Ethylbenzene_*	.392	.387	.391	.403	.387	.392	1.6*
Styrene_#	.847	.839	.819	.854	.830	.838	1.6
Xylene_(total)_#	.809	.795	.797	.818	.794	.802	1.3
Toluene-d8_#	1.050	1.053	1.056	1.096	1.075	1.066	1.8
Bromofluorobenzene_#	.788	.787	.784	.803	.784	.789	1.0
1,2-Dichloroethane-d4_#	1.701	1.613	1.697	1.801	1.734	1.709	4.0

7A
VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: BATTELLE-PNL

Contract: -----

Lab Code: -----

Case No.: -----

SAS No.: -----

SDG No.: -----

Instrument ID: 5970 #2 Calibration Date: 7/18/91 Time: 10:21

Lab File ID: >SB802 Init. Calib. Date(s): 07/01/91 07/18/91

Matrix: (soil/water) WATER Level: (low/med) LOW Column: (pack/cap) CAP

Min RRF50 for SPCC(%) = 0.300 (0.250 for Bromoform) Max %D for CCC(*) = 25.0%

COMPOUND	RRF	RRF50	%D
Chloromethane	.670	.689	2.9
Bromomethane	1.269	1.322	4.2
Vinyl Chloride	.781	.803	2.8
Chloroethane	.447	.482	7.8
Methylene_Chloride	1.118	1.145	2.4
Acetone	.210	.207	1.5
Carbon_Disulfide	2.571	2.789	8.5
1,1-Dichloroethane	1.070	1.161	8.6
1,1-Dichloroethane	1.931	2.077	7.6
1,2-Dichloroethane_(total)	1.423	1.512	6.3
Chloroform	3.062	3.326	8.6
1,2-Dichloroethane	1.934	2.130	10.2
2-Butanone	.099	.093	6.1
1,1,1-Trichloroethane	3.008	3.352	11.4
Carbon_Tetrachloride	3.038	3.410	12.2
Vinyl Acetate	1.385	1.353	2.3
Bromodichloromethane	3.107	3.486	12.2
1,2-Dichloropropane	.283	.288	1.5
cis-1,3-Dichloropropene	.513	.526	2.5
Trichloroethane	.466	.475	2.0
Dibromochloromethane	.828	.880	6.3
1,1,2-Trichloroethane	.339	.347	2.4
Benzene	.696	.703	1.0
trans-1,3-Dichloropropene	.471	.483	2.5
Bromoform	.590	.581	1.6
4-Methyl-2-pentanone	.265	.244	7.8
2-Hexanone	.178	.163	8.4
Tetrachloroethene	.528	.514	2.6
1,1,2,2-Tetrachloroethane	.625	.600	3.9
Toluene	.607	.615	1.3
Chlorobenzene	.925	.917	.9
Ethylbenzene	.392	.398	1.4
Styrene	.838	.812	3.2
Xylene_(total)	.802	.796	0.8
Toluene-d8	1.066	1.089	2.1
Bromofluorobenzene	.789	.820	3.8
1,2-Dichloroethane-d4	1.709	1.961	14.8

8A
VOLATILE INTERNAL STANDARD AREA SUMMARY

Lab Name: BATTELLE-PNL

Contract: -----

Lab Code: -----

Case No.: -----

SAS No.: -----

SDG No.: -----

Lab File ID (Standard): >SB802

Date Analyzed: 7/18/91

Instrument ID: 5970 #2

Time Analyzed: 10:21

Matrix: (soil/water) WATER Level: (low/med) LOW Column: (pack/cap) CAP

	IS1 (BCM)		IS2 (DFB)		IS3 (CBZ)	
	AREA #	RT	AREA #	RT	AREA #	RT
12 HOUR STD	33682.	13.41	142188.	16.79	121224.	24.42
UPPER LIMIT	67364.		284376.		242448.	
LOWER LIMIT	16841.		71094.		60612.	
EPA SAMPLE NO.						
01 MET BLANK	36822.	13.57	148656.	16.92	123445.	24.54
02 91-6425	38286.	13.52	158516.	16.90	132772.	24.52
03 91-6426	38719.	13.46	162298.	16.85	134561.	24.49
04 91-6427	38938.	13.45	166925.	16.83	135099.	24.49
05 91-6428	34751.	13.52	148545.	16.88	121676.	24.50
06 91-6428 DUP	36147.	13.50	156686.	16.85	125672.	24.49
07						
08						
09						
10						
11						
12						
13						
14						
15						
16						
17						
18						
19						
20						
21						
22						

IS1 (BCM) = Bromochloromethane
IS2 (DFB) = 1,4-Difluorobenzene
IS3 (CBZ) = Chlorobenzene-d5

UPPER LIMIT = + 100%
of internal standard area.
LOWER LIMIT = - 50%
of internal standard area.

Column used to flag internal standard area values with an asterisk

SAMPLE DELIVERY GROUPS 4 AND 5

A summary report of the analyses of samples in Delivery Groups 4 and 5 provides general information on the samples, sample preparation, analysis method, and location of data files. The summary report also defines codes used in the CLP report forms, which follow.

**ORGANIC COMPOUND ANALYSIS REPORT
VOLATILE COMPOUNDS**

SAMPLE ANALYSIS REPORTED

Analysis of three slurry samples for volatile organic compounds by gas chromatography/mass spectrometry (GC/MS) is the subject of this report.

SAMPLE DESCRIPTION

<u>Sample ID</u>	<u>ACL Lab Number</u>
531-2	91-7174
529-2	91-7175
533-2	91-7319

The samples were received in the 325 Building "B" Analytical Hot Cell with substantial head space, non-septa type closures, and at ambient temperature on 7/23/91. These sample handling procedures are not consistent with proper analytical protocol for this analysis.

SAMPLE PREPARATION

GC/FID volatile hexadecane screens were performed and indicated a dilution of the sample was not necessary. Due to the radioactivity of the samples, 1 gram aliquots of the samples were transferred to 20 mL pre-cleaned VOA vials, and accurately weighed. The samples were serially dissolved in 5mL of lab water, then added to a purge vessel attached to a Tekmar Liquid Sample Concentrator.

For proper vendor software quantitation of the weighed samples, "soil" matrix must be selected; although a "water" blank is used to accurately represent the handling of the samples. The analysis followed EPA-CLP SOW 2/88 procedures for the analysis of volatile compounds.

ANALYSIS METHOD

- GC/MS procedure: PNL-ALO-335 (on 7/30/91).
- GC/MS instrumentation: HP-5890/5970 GC/MS (WB46864).
- GC/MS location: Lab 427, 325 building.

- GC screening analysis: PNL-ALO-340.
- GC/FID instrumentation: HP-5890 (WB60701).
- GC/FID location: Lab 330, 325 building.

QUALITY CONTROL

Quality control procedures specified for this method were followed. While these forms are not all included in this report,

the quality assurance performance requirements are summarized below:

<u>Form Information</u>	<u>Comments</u>
2A Surrogate Recovery	Meets all requirement
3A MS/MSD Recovery	Analysis not requested.
4A Method Blank Summary	Meets all requirements.
5A Tune/Mass Calibration	Meets all requirements.
6A Initial Calibration	5 point calibration. Meets all requirements.
7A Daily Calibration	Meets all requirements.
8A Internal Standards	Meets all requirements.

DATA

The data and calibration are archived on magnetic tape in the 325 building, 327-A laboratory. The following is the list of pertinent files:

<u>File Name</u>	<u>Sample Number</u>	<u>Sample Analyzed</u>
>SC001		Mass Calibration/Tune Check
>SC002		Daily Calibration
>SC003		Method Blank
>SC004	91-7174	531-2
>SC005	91-7175	529-2
>SC006	91-7319	533-2

Lab data are also maintained as follows:

<u>Activity</u>	<u>LRB Number</u>	<u>Page Number</u>
GC/MS injection log	BNW-52907	82-83

RESULTS

CLP Target Compounds: The attached 1A Forms show that volatile target compounds, Acetone, and 2-butanone, were detected above the Contract Required Quantitation Limits, (CRQL's), in the samples. Vinyl Acetate was observed in the blank below the CRQL, and in the samples, at similar levels to the blank, and is "B-flagged".

The following defines the qualifiers, Q-flags, in the Form 1's:

<u>"Q" Flag</u>	<u>Definition</u>
U	Indicates the compound was analyzed for but not detected, the U-flagged concentration number is the CRQL.
J	Indicates an estimated value for the target or tentatively identified compounds, spectra meet criteria but response is below the CRQL for the target compounds.
B	Compound was found in the blank.
X	Indicates compound was manually deleted because all requirements were not met.
D	Analysis was performed on a diluted sample.
E	Indicates that Quantitation was outside the calibration range.
A	Aldol condensation product.

Tentatively Identified Compounds: As shown in the attached 1E Forms, no non-EPA target compounds were observed in the blank. 2-Ethyl-1-hexanol, n-Alkanes, and an Unknown were observed in the samples.

ANALYST W. R. R. R. DATE 9/18/91 REVIEW C. E. R. R. DATE 9-18-91

1A
VOLATILE ORGANICS ANALYSIS DATA SHEET

EPA SAMPLE NO.

MET BLANK

Lab Name: BATTELLE-PNL

Contract: -----

Lab Code: -----

Case No.: -----

SAS No.: -----

SDG No.: -----

Matrix: (soil/water) WATER

Lab Sample ID: LAB BLANK

Sample wt/vol: 5 (g/mL) mL

Lab File ID: >SC003

Level: (low/med) LOW

Date Received: 02/30/91

% Moisture: not dec.-----

Date Analyzed: 7/30/91

Column: (pack/cap) CAP

Dilution Factor: 1.0

CAS NO.	COMPOUND	CONCENTRATION UNITS: (ug/L or ug/Kg) ug/L	Q
74-87-3-----	Chloromethane	10.	IU
74-83-9-----	Bromomethane	10.	IU
75-01-4-----	Vinyl Chloride	10.	IU
75-00-3-----	Chloroethane	10.	IU
75-09-2-----	Methylene Chloride	5.	IU
67-64-1-----	Acetone	10.	IU
75-15-0-----	Carbon Disulfide	5.	IU
75-35-4-----	1,1-Dichloroethene	5.	IU
75-34-3-----	1,1-Dichloroethane	5.	IU
540-59-0-----	1,2-Dichloroethene (total)	5.	IU
67-66-3-----	Chloroform	5.	IU
107-02-2-----	1,2-Dichloroethane	5.	IU
78-93-3-----	2-Butanone	10.	IU
71-55-6-----	1,1,1-Trichloroethane	5.	IU
56-23-5-----	Carbon Tetrachloride	5.	IU
108-05-4-----	Vinyl Acetate	2.	IJB
75-27-4-----	Bromodichloromethane	5.	IU
78-87-5-----	1,2-Dichloropropane	5.	IU
10061-01-5-----	cis-1,3-Dichloropropene	5.	IU
79-01-6-----	Trichloroethene	5.	IU
124-48-1-----	Dibromochloromethane	5.	IU
79-00-5-----	1,1,2-Trichloroethane	5.	IU
71-43-2-----	Benzene	5.	IU
10061-02-6-----	trans-1,3-Dichloropropene	5.	IU
75-25-2-----	Bromoform	5.	IU
108-10-1-----	4-Methyl-2-pentanone	5.	IU
591-78-6-----	2-Hexanone	10.	IU
127-18-4-----	Tetrachloroethene	5.	IU
79-34-5-----	1,1,2,2-Tetrachloroethane	5.	IU
108-88-3-----	Toluene	5.	IU
108-90-7-----	Chlorobenzene	5.	IU
100-41-4-----	Ethylbenzene	5.	IU
100-42-5-----	Styrene	5.	IU
133-02-7-----	Xylene (total)	5.	IU

1E
VOLATILE ORGANICS ANALYSIS DATA SHEET
TENTATIVELY IDENTIFIED COMPOUNDS

EPA SAMPLE NO.

MET BLANK

Lab Name: BATTELLE-PNL

Contract: -----

Lab Code: -----

Case No.: -----

SAS No.: -----

SDG No.: -----

Matrix: (soil/water) WATER

Lab Sample ID: LAB BLANK

Sample wt/vol: 5.0 (g/mL) mL

Lab File ID: >SC003

Level: (low/med) LOW

Date Received: 07/23/91

% Moisture: not dec.-----

Date Analyzed: 7/30/91

Column: CAP

Dilution Factor: 1.0

Number TICs found: 0

CONCENTRATION UNITS:
(ug/L or ug/Kg) ug/L

CAS NUMBER	COMPOUND NAME	RT	EST. CONC.	Q
1.				
2.				
3.				
4.				
5.				
6.				
7.				
8.				
9.				
10.				
11.				
12.				
13.				
14.				
15.				
16.				
17.				
18.				
19.				
20.				
21.				
22.				
23.				
24.				
25.				
26.				
27.				
28.				
29.				
30.				

1A
VOLATILE ORGANICS ANALYSIS DATA SHEET

EPA SAMPLE NO.

91-7174

Lab Name: BATTELLE-PNL

Contract: -----

Lab Code: -----

Case No.: -----

SAS No.: -----

SDG No.: -----

Matrix: (soil/water) SOIL

Lab Sample ID: 531-2

Sample wt/vol: 0.9795 (g/mL) G

Lab File ID: >SC004

Level: (low/med) LOW

Date Received: 07/23/91

% Moisture: not dec.-----

Date Analyzed: 7/30/91

Column: (pack/cap) CAP

Dilution Factor: 1.0

CAS NO.	COMPOUND	CONCENTRATION UNITS:	
		(ug/L or ug/Kg)	ug/Kg
			Q
74-87-3-----	Chloromethane	51.	IU
74-83-9-----	Bromomethane	51.	IU
75-01-4-----	Vinyl Chloride	51.	IU
75-00-3-----	Chloroethane	51.	IU
75-09-2-----	Methylene Chloride	26.	IU
67-64-1-----	Acetone	51.	IU
75-15-0-----	Carbon Disulfide	26.	IU
75-35-4-----	1,1-Dichloroethene	26.	IU
75-34-3-----	1,1-Dichloroethane	26.	IU
540-59-0-----	1,2-Dichloroethene (total)	26.	IU
67-66-3-----	Chloroform	26.	IU
107-02-2-----	1,2-Dichloroethane	26.	IU
78-93-3-----	2-Butanone	51.	IU
71-55-6-----	1,1,1-Trichloroethane	26.	IU
56-23-5-----	Carbon Tetrachloride	26.	IU
108-05-4-----	Vinyl Acetate	15.	IJB
75-27-4-----	Bromodichloromethane	26.	IU
78-87-5-----	1,2-Dichloropropane	26.	IU
10061-01-5-----	cis-1,3-Dichloropropene	26.	IU
79-01-6-----	Trichloroethene	26.	IU
124-48-1-----	Dibromochloromethane	26.	IU
79-00-5-----	1,1,2-Trichloroethane	26.	IU
71-43-2-----	Benzene	26.	IU
10061-02-6-----	trans-1,3-Dichloropropene	26.	IU
75-25-2-----	Bromoform	26.	IU
108-10-1-----	4-Methyl-2-pentanone	51.	IU
591-78-6-----	2-Hexanone	51.	IU
127-18-4-----	Tetrachloroethene	26.	IU
79-34-5-----	1,1,2,2-Tetrachloroethane	26.	IU
108-88-3-----	Toluene	26.	IU
108-90-7-----	Chlorobenzene	26.	IU
100-41-4-----	Ethylbenzene	26.	IU
100-42-5-----	Styrene	26.	IU
133-02-7-----	Xylene (total)	26.	IU

1E
VOLATILE ORGANICS ANALYSIS DATA SHEET
TENTATIVELY IDENTIFIED COMPOUNDS

EPA SAMPLE NO.

91-7174

Lab Name: BATTELLE-PNL

Contract: -----

Lab Code: -----

Case No.: -----

SAS No.: -----

SDG No.: -----

Matrix: (soil/water) SOIL

Lab Sample ID: 531-2

Sample wt/vol: 0.9795 (g/mL) G

Lab File ID: >SC004

Level: (low/med) LOW

Date Received: 07/23/91

% Moisture: not dec.-----

Date Analyzed: 07/30/91

Column: CAP

Dilution Factor: 1.0

Number TICs found: 1

CONCENTRATION UNITS:
(ug/L or ug/Kg) ug/Kg

CAS NUMBER	COMPOUND NAME	RT	EST. CONC.	Q
1. 104767	1-Hexanol, 2-ethyl-	30.06	220.	
2.				
3.				
4.				
5.				
6.				
7.				
8.				
9.				
10.				
11.				
12.				
13.				
14.				
15.				
16.				
17.				
18.				
19.				
20.				
21.				
22.				
23.				
24.				
25.				
26.				
27.				
28.				
29.				
30.				

1A
VOLATILE ORGANICS ANALYSIS DATA SHEET

EPA SAMPLE NO.

91-7175

Lab Name: BATTELLE-PNL

Contract: -----

Lab Code: -----

Case No.: -----

SAS No.: -----

SDG No.: -----

Matrix: (soil/water) SOIL

Lab Sample ID: 529-2

Sample wt/vol: 1.0070 (g/mL) G

Lab File ID: >SC005

Level: (low/med) LOW

Date Received: 07/23/91

% Moisture: not dec. ----

Date Analyzed: 7/30/91

Column: (pack/cap) CAP

Dilution Factor: 1.0

CAS NO.	COMPOUND	CONCENTRATION UNITS:	
		(ug/L or ug/Kg)	ug/Kg
			Q
74-87-3-----	Chloromethane	50.	IU
74-83-9-----	Bromomethane	50.	IU
75-01-4-----	Vinyl Chloride	50.	IU
75-00-3-----	Chloroethane	50.	IU
75-09-2-----	Methylene_Chloride	8.	IJ
67-64-1-----	Acetone	50.	I
75-15-0-----	Carbon Disulfide	25.	IU
75-35-4-----	1,1-Dichloroethane	25.	IU
75-34-3-----	1,1-Dichloroethane	25.	IU
540-59-0-----	1,2-Dichloroethane_(total)	25.	IU
67-66-3-----	Chloroform	25.	IU
107-02-2-----	1,2-Dichloroethane	25.	IU
78-93-3-----	2-Butanone	86.	I
71-55-6-----	1,1,1-Trichloroethane	25.	IU
56-23-5-----	Carbon Tetrachloride	25.	IU
108-05-4-----	Vinyl Acetate	15.	IJB
75-27-4-----	Bromodichloromethane	25.	IU
78-87-5-----	1,2-Dichloropropane	25.	IU
10061-01-5-----	cis-1,3-Dichloropropene	25.	IU
79-01-6-----	Trichloroethene	25.	IU
124-48-1-----	Dibromochloromethane	25.	IU
79-00-5-----	1,1,2-Trichloroethane	25.	IU
71-43-2-----	Benzene	25.	IU
10061-02-6-----	trans-1,3-Dichloropropene	25.	IU
75-25-2-----	Bromoform	25.	IU
108-10-1-----	4-Methyl-2-pentanone	50.	IU
591-78-6-----	2-Hexanone	50.	IU
127-18-4-----	Tetrachloroethene	25.	IU
79-34-5-----	1,1,2,2-Tetrachloroethane	25.	IU
108-88-3-----	Toluene	25.	IU
108-90-7-----	Chlorobenzene	25.	IU
100-41-4-----	Ethylbenzene	25.	IU
100-42-5-----	Styrene	25.	IU
133-02-7-----	Xylene (total)	25.	IU

1E
VOLATILE ORGANICS ANALYSIS DATA SHEET
TENTATIVELY IDENTIFIED COMPOUNDS

EPA SAMPLE NO.

91-7175

Lab Name: BATTELLE-PNL

Contract: -----

Lab Code: -----

Case No.: -----

SAS No.: -----

SDG No.: -----

Matrix: (soil/water) SOIL

Lab Sample ID: 529-2

Sample wt/vol: 1.0070 (g/mL) G

Lab File ID: >SC005

Level: (low/med) LOW

Date Received: 07/23/91

% Moisture: not dec.-----

Date Analyzed: 07/30/91

Column: CAP

Dilution Factor: 1.0

Number TICs found: 1

CONCENTRATION UNITS:
(ug/L or ug/Kg) ug/Kg

CAS NUMBER	COMPOUND NAME	RT	EST. CONC.	Q
1. 104767	1-Hexanol, 2-ethyl-	30.04	97.	
2.				
3.				
4.				
5.				
6.				
7.				
8.				
9.				
10.				
11.				
12.				
13.				
14.				
15.				
16.				
17.				
18.				
19.				
20.				
21.				
22.				
23.				
24.				
25.				
26.				
27.				
28.				
29.				
30.				

1A
VOLATILE ORGANICS ANALYSIS DATA SHEET

EPA SAMPLE NO.

91-7319

Lab Name: BATTELLE-PNL

Contract: -----

Lab Code: -----

Case No.: -----

SAS No.: -----

SDG No.: -----

Matrix: (soil/water) SOIL

Lab Sample ID: 533-2

Sample wt/vol: 1.0525 (g/mL) G

Lab File ID: >SC006

Level: (low/med) LOW

Date Received: 07/25/91

% Moisture: not dec. ----

Date Analyzed: 7/30/91

Column: (pack/cap) CAP

Dilution Factor: 1.0

CAS NO.	COMPOUND	CONCENTRATION UNITS: (ug/L or ug/Kg) ug/Kg	Q
74-87-3	Chloromethane	48.	U
74-83-9	Bromomethane	48.	U
75-01-4	Vinyl Chloride	48.	U
75-00-3	Chloroethene	48.	U
75-09-2	Methylene Chloride	10.	J
67-64-1	Acetone	46.	J
75-15-0	Carbon Disulfide	24.	U
75-35-4	1,1-Dichloroethene	24.	U
75-34-3	1,1-Dichloroethane	24.	U
540-59-0	1,2-Dichloroethene (total)	24.	U
67-66-3	Chloroform	24.	U
107-02-2	1,2-Dichloroethane	24.	U
78-93-3	2-Butanone	68.	
71-55-6	1,1,1-Trichloroethane	24.	U
56-23-5	Carbon Tetrachloride	24.	U
108-05-4	Vinyl Acetate	48.	U
75-27-4	Bromodichloromethane	24.	U
78-87-5	1,2-Dichloropropane	24.	U
10061-01-5	cis-1,3-Dichloropropene	24.	U
79-01-6	Trichloroethene	24.	U
124-48-1	Dibromochloromethane	24.	U
79-00-5	1,1,2-Trichloroethane	24.	U
71-43-2	Benzene	24.	U
10061-02-6	trans-1,3-Dichloropropene	24.	U
75-25-2	Bromoform	24.	U
108-10-1	4-Methyl-2-pentanone	48.	U
591-78-6	2-Hexanone	48.	U
127-18-4	Tetrachloroethene	24.	U
79-34-5	1,1,2,2-Tetrachloroethane	24.	U
108-88-3	Toluene	24.	U
108-90-7	Chlorobenzene	24.	U
100-41-4	Ethylbenzene	24.	U
100-42-5	Styrene	24.	U
133-02-7	Xylene (total)	24.	U

1E
VOLATILE ORGANICS ANALYSIS DATA SHEET
TENTATIVELY IDENTIFIED COMPOUNDS

EPA SAMPLE NO.

91-7319

Lab Name: BATTELLE-PNL

Contract: -----

Lab Code: -----

Case No.: -----

SAS No.: -----

SDG No.: -----

Matrix: (soil/water) SOIL

Lab Sample ID: 533-2

Sample wt/vol: 1.0525 (g/mL) G

Lab File ID: >SC006

Level: (low/med) LOW

Date Received: 07/25/91

% Moisture: not dec.-----

Date Analyzed: 7/30/91

Column: CAP

Dilution Factor: 1.0

Number TICs found: 5

CONCENTRATION UNITS:
(ug/L or ug/Kg) ug/Kg

CAS NUMBER	COMPOUND NAME	RT	EST. CONC.	Q
1.	Unknown	28.80	45.	
2. 104767	1-Hexanol, 2-ethyl-	30.07	100.	
3. 1120214	Undecane	30.31	24.	
4. 112403	Dodecane	31.16	34.	
5. 629505	Tridecane	31.93	29.	
6.				
7.				
8.				
9.				
10.				
11.				
12.				
13.				
14.				
15.				
16.				
17.				
18.				
19.				
20.				
21.				
22.				
23.				
24.				
25.				
26.				
27.				
28.				
29.				
30.				

2A
WATER VOLATILE SURROGATE RECOVERY

Lab Name: BATTELLE-PNL

Contract: -----

Lab Code: ----- Case No.: ----- SAS No.: ----- SDG No.: -----

	EPA SAMPLE NO.	S1 (TOL)#	S2 (BFB)#	S3 (DCE)#	OTHER	TOT OUT
01	DAILY CAL	101	101	94		0
02	MET BLANK	95	97	90		0
03	91-7174	99	101	99		0
04	91-7175	96	99	97		0
05	91-7319	96	99	93		0
06						
07						
08						
09						
10						
11						
12						
13						
14						
15						
16						
17						
18						
19						
20						
21						
22						
23						
24						
25						
26						
27						
28						
29						
30						

QC LIMITS

S1 (TOL) = Toluene-d8 (88-110)
 S2 (BFB) = Bromofluorobenzene (86-115)
 S3 (DCE) = 1,2-Dichloroethane-d4 (76-114)

Column to be used to flag recovery values

* Values outside of contract required QC limits

D Surrogates diluted out

4A
VOLATILE METHOD BLANK SUMMARY

Lab Name: BATTELLE-PNL

Contract: -----

Lab Code: -----

Case No.: -----

SAS No.: -----

SDG No.: -----

Lab File ID: >SC003

Lab Sample ID: LAB BLANK

Date Analyzed: 7/30/91

Time Analyzed: 13:25

Matrix: (soil/water) WATER

Level: (low/med) LOW

Instrument ID: 5970 #2

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

	EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	TIME ANALYZED
01	91-7174	531-2	>SC004	16:03
02	91-7175	529-2	>SC005	17:01
03	91-7319	533-2	>SC006	18:00
04				
05				
06				
07				
08				
09				
10				
11				
12				
13				
14				
15				
16				
17				
18				
19				
20				
21				
22				
23				
24				
25				
26				
27				
28				
29				
30				

COMMENTS:

5A
VOLATILE ORGANIC GC/MS TUNING AND MASS
CALIBRATION - BROMOFLUOROBENZENE (BFB)

Lab Name: BATTELLE-PNL

Contract: -----

Lab Code: -----

Case No.: -----

SAS No.: -----

SDG No.: -----

Lab File ID: >SC001

BFB Injection Date: 7/30/91

Instrument ID: 5970 #2

BFB Injection Time: 10:33

Matrix: (soil/water) WATER Level: (low/med) LOW Column: (pack/cap) CAP

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% OF MASS 95	20.1
75	30.0 - 60.0% OF MASS 95	52.3
95	Base peak, 100% relative abundance	100.
96	5.0 - 9.0% of mass 95	5.8
173	Less than 2.0% of mass 174	0.0(0.0)1
174	Greater than 50.0% of mass 95	85.1
175	5.0 - 9.0% of mass 174	7.3(8.6)1
176	Greater than 95.0%, but less than 101.0% of mass 174	83.0(97.5)1
177	5.0 - 9.0% of mass 176	5.5(6.7)2

1-Value is % mass 174

2-Value is % mass 176

THIS TUNE APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

	EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
01	DAILY CAL	50 PPB CHECK	>SC002	7/30/91	11:43
02	MET BLANK	LAB BLANK	>SC003	7/30/91	13:25
03	91-7174	531-2	>SC004	7/30/91	16:03
04	91-7175	529-2	>SC005	7/30/91	17:01
05	91-7319	533-2	>SC006	7/30/91	18:00
06					
07					
08					
09					
10					
11					
12					
13					
14					
15					
16					
17					
18					
19					
20					
21					
22					

6A
VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: BATTELLE-PNL

Contract: -----

Lab Code: -----

Case No.: -----

SAS No.: -----

SDG No.: -----

Instrument ID: 5970 #2 Calibration Date(s): 07/01/91 07/30/91

Matrix: (soil/water) WATER Level: (low/med) LOW Column: (pack/cap) CAP

Min RRF for SPCC(%) = 0.300 (0.250 for Bromoform) Max %RSD for CCC(*) = 30.0%

LAB FILE ID:	RRF20 => RC704	RRF50 => RC705
RRF100 => RC706	RRF150 => RC707	RRF200 => RC708

COMPOUND	RRF20	RRF50	RRF100	RRF150	RRF200	RRF	% RSD
Chloromethane	.722	.654	.678	.667	.629	.670	5.1#
Bromomethane	1.324	1.221	1.284	1.290	1.224	1.269	3.5
Vinyl_Chloride	.779	.751	.811	.803	.762	.781	3.3*
Chloroethane	.480	.456	.471	.433	.395	.447	7.6
Methylene_Chloride	1.105	1.119	1.148	1.134	1.083	1.118	2.2
Acetone	.294	.190	.193	.188	.186	.210	22.2
Carbon_Disulfide	2.427	2.476	2.606	2.747	2.598	2.571	4.9
1,1-Dichloroethene	1.080	1.060	1.127	1.123	.959	1.070	6.4*
1,1-Dichloroethane	1.869	1.841	1.981	2.024	1.938	1.931	3.9#
1,2-Dichloroethene_(total)	1.392	1.352	1.466	1.490	1.415	1.423	4.5
Chloroform	2.951	2.935	3.154	3.196	3.073	3.062	3.8*
1,2-Dichloroethane	1.958	1.819	1.965	1.992	1.935	1.934	3.5
2-Butanone	.157	.081	.085	.087	.082	.099	33.2
1,1,1-Trichloroethane	2.854	2.879	3.111	3.146	3.051	3.008	4.5
Carbon_Tetrachloride	2.891	2.864	3.138	3.218	3.082	3.038	5.1
Vinyl_Acetate	1.479	1.260	1.371	1.435	1.378	1.385	6.0
Bromodichloromethane	2.896	2.933	3.220	3.295	3.192	3.107	5.8
1,2-Dichloropropane	.286	.277	.285	.290	.279	.283	1.8*
cis-1,3-Dichloropropene	.507	.493	.518	.530	.517	.513	2.7
Trichloroethene	.466	.456	.467	.477	.462	.466	1.7
Dibromochloromethane	.827	.781	.830	.859	.844	.828	3.5
1,1,2-Trichloroethane	.366	.322	.333	.340	.335	.339	4.8
Benzene	.691	.685	.709	.709	.686	.696	1.7
trans-1,3-Dichloropropene	.473	.447	.476	.477	.483	.471	2.9
Bromoform	.642	.530	.577	.604	.598	.590	7.0#
4-Methyl-2-pentanone	.364	.238	.238	.240	.246	.265	21.0
2-Hexanone	.246	.154	.164	.163	.161	.178	21.7
Tetrachloroethene	.535	.519	.534	.535	.516	.528	1.8
1,1,2,2-Tetrachloroethane	.804	.571	.566	.596	.587	.625	16.1
Toluene	.609	.600	.607	.620	.601	.607	1.3*
Chlorobenzene	.935	.924	.900	.958	.910	.925	2.4#
Ethylbenzene	.392	.387	.391	.403	.387	.392	1.6*
Styrene	.847	.839	.819	.854	.830	.838	1.6
Xylene_(total)	.809	.795	.797	.818	.794	.802	1.3
Toluene-d8	1.050	1.053	1.056	1.096	1.075	1.066	1.8
Bromofluorobenzene	.788	.787	.784	.803	.784	.789	1.0
1,2-Dichloroethane-d4	1.701	1.613	1.697	1.801	1.734	1.709	4.0

7A
VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:BATTELLE-PNL

Contract:-----

Lab Code: -----

Case No.: -----

SAS No.: -----

SDG No.: -----

Instrument ID: 5970 #2 Calibration Date: 7/30/91 Time: 11:43

Lab File ID: >SC002

Init. Calib. Date(s):07/01/91 07/30/91

Matrix:(soil/water) WATER Level:(low/med) LOW Column:(pack/cap) CAP

Min RRF50 for SPCC(#) = 0.300 (0.250 for Bromoform) Max %D for CCC(*) = 25.0%

COMPOUND	RRF	RRF50	%D
Chloromethane	.670	.661	1.3 #
Bromomethane	1.269	1.335	5.2
Vinyl Chloride	.781	.805	3.1 *
Chloroethane	.447	.505	12.9
Methylene_Chloride	1.118	1.130	1.1
Acetone	.210	.205	2.7
Carbon_Disulfide	2.571	2.785	8.3
1,1-Dichloroethene	* 1.070	1.113	4.0 *
1,1-Dichloroethane	# 1.931	2.026	4.9 #
1,2-Dichloroethene_(total)	1.423	1.483	4.0
Chloroform	* 3.062	3.321	8.5 *
1,2-Dichloroethane	1.934	2.039	5.5
2-Butanone	.099	.081	17.6
1,1,1-Trichloroethane	3.008	3.363	11.8
Carbon_Tetrachloride	3.038	3.346	10.1
Vinyl Acetate	1.385	1.353	2.3
Bromodichloromethane	3.107	3.359	8.1
1,2-Dichloropropane	* .283	.290	2.4 *
cis-1,3-Dichloropropene	.513	.543	5.8
Trichloroethene	.466	.503	7.9
Dibromochloromethane	.828	.906	9.3
1,1,2-Trichloroethane	.339	.356	4.9
Benzene	.696	.720	3.4
trans-1,3-Dichloropropene	.471	.481	2.1
Bromoform	# .590	.603	2.2 #
4-Methyl-2-pentanone	.265	.225	15.2
2-Hexanone	.178	.147	17.1
Tetrachloroethene	.528	.549	4.1
1,1,2,2-Tetrachloroethane	.625	.565	9.6
Toluene	* .607	.630	3.7 *
Chlorobenzene	# .925	.974	5.3 #
Ethylbenzene	* .392	.382	2.5 *
Styrene	.838	.866	3.3
Xylene_(total)	.802	.827	3.1
Toluene-d8	1.066	1.104	3.6
Bromofluorobenzene	.789	.826	4.6
1,2-Dichloroethane-d4	1.709	1.845	7.9

8A
VOLATILE INTERNAL STANDARD AREA SUMMARY

Lab Name: BATTELLE-PNL

Contract: -----

Lab Code: -----

Case No.: -----

SAS No.: -----

SDG No.: -----

Lab File ID (Standard): >SC002

Date Analyzed: 7/30/91

Instrument ID: 5970 #2

Time Analyzed: 11:43

Matrix: (soil/water) WATER Level: (low/med) LOW Column: (pack/cap) CAP

	IS1 (BCM)		IS2 (DFB)		IS3 (CBZ)	
	AREA #	RT	AREA #	RT	AREA #	RT
12 HOUR STD	37130.	13.44	148689.	16.81	128957.	24.44
UPPER LIMIT	74260.		297378.		257914.	
LOWER LIMIT	18565.		74345.		64479.	
EPA SAMPLE NO.						
01 MET BLANK	37287.	13.54	148137.	16.91	124120.	24.51
02 91-7174	40191.	13.52	164280.	16.89	137035.	24.49
03 91-7175	41441.	13.43	165651.	16.80	142217.	24.44
04 91-7319	41752.	13.50	164782.	16.88	139752.	24.53
05						
06						
07						
08						
09						
10						
11						
12						
13						
14						
15						
16						
17						
18						
19						
20						
21						
22						

IS1 (BCM) = Bromochloromethane
IS2 (DFB) = 1,4-Difluorobenzene
IS3 (CBZ) = Chlorobenzene-d5

UPPER LIMIT = + 100%
of internal standard area.
LOWER LIMIT = - 50%
of internal standard area.

Column used to flag internal standard area values with an asterisk

APPENDIX A

TEST INSTRUCTIONS

TEST INSTRUCTIONSTI-101SY-1
Page 1 of 2101-SY SAMPLE TEST INSTRUCTION
WINDOW C
FOR ORGANIC ANALYSIS

DATE PREPARED: June 27, 1991

PREPARED BY: Barb Gillespie

SAMPLE NUMBERS: 91-6264 (#424), 91-6265 (#420), 91-6266 (#421) and 91-6267 (#563)APPROVED BY: B. M. Gillespie
Project ManagerDATE: 6-27-91

CONTROLLING DOCUMENTS:

Project TPP: 101-SY Chemical Analysis Support Task TPP 17667

Project QAPP: Hydrogen Safety Project QA Plan MCS-027

Administrative Control Procedure: PNL-ALO-010

INTRODUCTION

This Test Instruction (TI) defines the scope of work to be completed on 101-SY tank samples that may be crust, slurry/sludge, saltcake and/or liquid samples and their accompanying blind and blank samples requiring Organic analysis.

This TI is based on the TPP 17667. Any deviations from the instructions contained in this TI will require prior approval from the Project Manager.

Upon sample receipt, samples and their extractants will be stored at $4 \pm 2^{\circ}\text{C}$, radioactive levels permitting.

All analyses are to be completed following the identified procedure and reported in the units of ug/L or ug/Kg. If deviations from the procedure must be made, the Project Manager must first be notified. Then the change is documented and this documentation must accompany the analytical data. All analytical data are returned to the ALO Project Support Office.

REQUESTED ANALYSES

<u>Requested Analysis</u>	<u>Procedure #</u>	<u>Task Leader</u>
Volatile Organics	PNL-ALO-335	Stromatt
<u>Screening Procedures</u>	<u>Procedure #</u>	<u>Task Leader</u>
Volatiles	PNL-ALO-332	Stromatt

101-SY SAMPLE TEST INSTRUCTION
WINDOW C
FOR ORGANIC ANALYSIS

SAMPLE PREPARATION INSTRUCTIONS

If samples aliquots must be transferred to other groups or out of the building, they must be transferred in accordance with PNL-ALO-010 (i.e. COC).

Respective representative sample aliquots, as spelled out in the procedures below, will be used for sample preparation as follows:

<u>Task</u>	<u>Preparation Procedure</u>	<u>Task Leader</u>
Volatiles	PNL-ALO-330	Stromatt

QUALITY CONTROL

The quality control requirements that require the CERCLA QC protocol outlined in the above procedures will be followed except no duplicate analyses or spike analyses are required by the WHC SOW or the TPP 17667.

Volatile sample analyses will be identified by the sample number only.

Data Reporting

<u>Task</u>	<u>Procedure</u>	<u>Task Leader</u>
Volatiles	PNL-ALO-350	Stromatt

101-SY SAMPLE TEST INSTRUCTION
WINDOW C
FOR ORGANIC ANALYSIS

DATE PREPARED: July 11, 1991

PREPARED BY: R. W. Stromatt

SAMPLE NUMBERS: 91-6424 (#451), 91-6425 (#578), 91-6426 (#579) 91-6427 (#580)
and 91-6428 (#581)

APPROVED BY: *R. W. Stromatt*
Project Manager

DATE: 7-11-91

CONTROLLING DOCUMENTS:

Project TPP: 101-SY Chemical Analysis Support Task TPP 17667

Project QAPP: Hydrogen Safety Project QA Plan MCS-027

Administrative Control Procedure: PNL-ALO-010

INTRODUCTION

This Test Instruction (TI) defines the scope of work to be completed on 101-SY tank samples that may be crust, slurry/sludge, saltcake and/or liquid samples and their accompanying blind and blank samples requiring Organic analysis.

This TI is based on the TPP 17667. Any deviations from the instructions contained in this TI will require prior approval from the Project Manager.

Upon sample receipt, samples and their extractants will be stored at $4 \pm 2^{\circ}\text{C}$, radioactive levels permitting.

All analyses are to be completed following the identified procedure and reported in the units of ug/L or ug/Kg. If deviations from the procedure must be made, the Project Manager must first be notified. Then the change is documented and this documentation must accompany the analytical data. All analytical data are returned to the ALO Project Support Office. Use of screen is optional for blanks.

REQUESTED ANALYSES

<u>Requested Analysis</u>	<u>Procedure #</u>	<u>Task Leader</u>
Volatile Organics	PNL-ALO-335	Stromatt
Total Organic Carbon	PNL-MA-597 Vol.7, Method 40.47	
<u>Screening Procedures</u>	<u>Procedures #</u>	<u>Task Leader</u>
Volatiles	PNL-ALO-332	Stromatt

101-SY SAMPLE TEST INSTRUCTION
WINDOW C
FOR ORGANIC ANALYSIS

SAMPLE PREPARATION INSTRUCTIONS

If samples aliquots must be transferred to other groups or out of the building, they must be transferred in accordance with PNL-ALO-010 (i.e. COC).

Respective representative sample aliquots, as spelled out in the procedures below, will be used for sample preparation as follows:

<u>Task</u>	<u>Preparation Procedure</u>	<u>Task Leader</u>
Volatiles	PNL-ALO-330	Stromatt

QUALITY CONTROL

The quality control requirements that require the CERCLA QC protocol outlined in the above procedures will be followed except no duplicate analyses or spike analyses are required by the WHC SOW or the TPP 17667.

Volatile sample analyses will be identified by the sample number only.

Data Reporting

<u>Task</u>	<u>Procedure</u>	<u>Task Leader</u>
Volatiles	PNL-ALO-350	Stromatt

101-SY SAMPLE TEST INSTRUCTION
WINDOW C
FOR ORGANIC ANALYSIS

ADDENDUM TO TI101-SY-2 DATED JULY 11, 1991

DATE PREPARED: July 16, 1991

PREPARED BY: B. M. Gillespie

SAMPLE NUMBERS: 91-6424, 91-6425, 91-6426, 91-6427 and 91-6428

APPROVED BY: B.M. Gillespie
Project Manager

DATE: 7-16-91

INSTRUCTIONS

Sample 91-6424 only requires TOC analysis. The other four samples only require VOA analysis.

101-SY SAMPLE TEST INSTRUCTION
WINDOW C
FOR ORGANIC ANALYSIS

DATE PREPARED: July 25, 1991

PREPARED BY: B. M. Gillespie

SAMPLE NUMBERS: 91-7174 (#531), 91-7175 (#529) and 91-7319 (#533)

APPROVED BY: B. M. Gillespie
Project Manager

DATE: 7-26-91

CONTROLLING DOCUMENTS:

Project TPP: 101-SY Chemical Analysis Support Task TPP 17667

Project QAPP: Hydrogen Safety Project QA Plan MCS-027

Administrative Control Procedure: PNL-ALO-010

INTRODUCTION

This Test Instruction (TI) defines the scope of work to be completed on 101-SY tank samples that may be crust, slurry/sludge, saltcake and/or liquid samples and their accompanying blind and blank samples requiring Organic analysis.

This TI is based on the TPP 17667. Any deviations from the instructions contained in this TI will require prior approval from the Project Manager.

Upon sample receipt, samples and their extractants will be stored at $4 \pm 2^{\circ}\text{C}$, radioactive levels permitting.

All analyses are to be completed following the identified procedure and reported in the units of ug/L or ug/Kg. If deviations from the procedure must be made, the Project Manager must first be notified. Then the change is documented and this documentation must accompany the analytical data. All analytical data are returned to the ALO Project Support Office. Use of screen is optional for blanks.

REQUESTED ANALYSES

<u>Requested Analysis</u>	<u>Procedure #</u>	<u>Task Leader</u>
Volatile Organics	PNL-ALO-335	Stromatt
<u>Screening Procedures</u>	<u>Procedure #</u>	<u>Task Leader</u>
Volatiles	PNL-ALO-332	Stromatt

101-SY SAMPLE TEST INSTRUCTION
WINDOW C
FOR ORGANIC ANALYSIS

SAMPLE PREPARATION INSTRUCTIONS

If samples aliquots must be transferred to other groups or out of the building, they must be transferred in accordance with PNL-ALO-010 (i.e. COC).

Respective representative sample aliquots, as spelled out in the procedures below, will be used for sample preparation as follows:

<u>Task</u>	<u>Preparation Procedure</u>	<u>Task Leader</u>
Volatiles	PNL-ALO-330	Stromatt

QUALITY CONTROL

The quality control requirements that require the CERCLA QC protocol outlined in the above procedures will be followed except no duplicate analyses or spike analyses are required by the WHC SOW or the TPP 17667.

Volatile sample analyses will be identified by the sample number only.

Data Reporting

<u>Task</u>	<u>Procedure</u>	<u>Task Leader</u>
Volatiles	PNL-ALO-350	Stromatt

APPENDIX B

CHAIN OF CUSTODY

APPENDIX B

CHAIN OF CUSTODY

B.1 WESTINGHOUSE CHAIN OF CUSTODY AND SAMPLE ANALYSIS REQUEST FORMS AND PNL SAMPLE RECEIPT FORMS

Sample analysis request forms (SAR) for samples #424, 420, 421 and 563 not received from Westinghouse Hanford Company.

Westinghouse Hanford
Company

CHAIN OF CUSTODY

Company Contact BEV CRAWFORD Telephone 3-1972
Sample Collected by K.J. PATTERSON Date 6/26/91 Time —
Sample Locations (101 sy)
Ice Chest No. — Field Logbook and Page No. —
Remarks —

Bill of Lading No. — Offsite Property No. —
Method of Shipment B-plant truck (Y.E. #5 pig)
Shipped to PNL (Rick Steele)
Possible Sample Hazards —

Sample Identification

#421 - SEG #3
#420 - SEG #6
#424 - SEG #14
#563 - VJ.

Chain of Possession

Relinquished by:

[Signature]

Received by:

[Signature]

Date/Time:

6:34pm 6/27/91

Relinquished by:

[Signature]

Received by:

[Signature]

Date/Time:

6-27-91 2220 hrs

Relinquished by:

Received by:

Date/Time:

Relinquished by:

Received by:

Date/Time:

Westinghouse Hanford Company	CHAIN OF CUSTODY
Company Contact <u>RED CRAWFORD</u>	Telephone <u>3-1972</u>
Sample Collected by <u>R. J. PATTERSON</u>	Date <u>6/26/91</u> Time <u>—</u>
Sample Locations <u>(101 S4)</u>	
Ice Chest No. <u>—</u>	Field Logbook and Page No. <u>—</u>
Remarks <u>—</u>	
Bill of Lading No. <u>—</u>	Offsite Property No. <u>—</u>
Method of Shipment <u>B. plant TRUCK</u>	<u>(Pic # G-10)</u>
Shipped to <u>P.N.L. (Rick Steele)</u>	
Possible Sample Hazards <u>—</u>	

Sample Identification

Chain of Possession

SAMPLE RECEIPT FORM

Delivered by: WHC B-PLANT OVERPACK Date/Time: 6-27-91 / 2220 hr

Received by: RT Steele / JKRAU

Customer Sample Number(s): 424 420 421 563

ALO Sample Number(s): 91-6264 91-6265 91-6266 91-6267

1. Customer Chain-of-Custody Form: Present ✓ Absent

2. Additional Shipping Forms (list):

OFFSITE RSR

3. Custody Seals on Shipping and/or Sample Containers and their Conditions.

Present ✓ Absent

If Present, Condition: Good

4. Sample Tag(s) ID Numbers if not Recorded on the Chain-of-Custody Record or on Sample Vial.

Notes: N/A

5. Condition of Shipping Container (i.e., broken container, dented, breached plastic bag, temperature of sample container as defined in Section 3.0 in PNL-ALO-051, etc.) OK ~ 24°C (Ambient)

6. Condition of Sample Vials. OK

7. Verification of Agreement or Nonagreement of Information on Receiving Documents. Agreement

8. Resolution of Problems or Discrepancies.
N/A

RETURN COMPLETED FORM TO PROJECT MANAGER

Westinghouse Hanford
Company

CHAIN OF CUSTODY

Company Contact Ben Crawford Telephone 3-1972

Sample Collected by Wayne Edmonson / KT Paterson 7/10/91 Time

Sample Locations 101-SY Corel and auger samples

Ice Chest No. Field Logbook and Page No.

Remarks Samples submitted for VOA and TOC analysis

Bill of Lading No. Offsite Property No.

Method of Shipment Doorstop carrier / overpack

Shipped to 325 Building Attn: Rick Steele - PNL

Possible Sample Hazards see RSR # 12103

Sample Identification

#451 - R9217 Auger sample from Riser 13A for TOC

#578 - Hot Cell blank #1 for VOA

#579 - Hot Cell blank #2 for VOA

#580 - Hot Cell blank #3 for VOA

#581 - Hot Cell blank #4 for VOA

Chain of Possession

Relinquished by:

B.A. Crawford

Received by:

T. E. Gomer

Date/Time:

7/10/91 6:56 pm

Relinquished by:

T. E. Gomer

Received by:

Rick Steele

Date/Time:

7/10/91 2050 hrs

Relinquished by:

Received by:

Date/Time:

Relinquished by:

Received by:

Date/Time:

REQUEST FOR SPECIAL ANALYSIS (RSA) to PNL

Sample Point 101-SY B.A. Crawford/ D.L. Hering		(2) Date/Time Issued 7/10/91 3:30pm	(3) Date/Time Required 	(4) Charge Code
(6) Number of Samples 5		(7) Customer I.D. #451, #578 #579, #580 #581		(5) Work Package W7A11
(10) Release RPT		(8) Laboratory I.D. 		(9) Requester Name/Phone B.A. Crawford 3-1972
				(11) Volume of Sample all in 20mL containers. (see below)
(12) Determination		(13) Expected Range		(14) OTR values Minimum Detection Level
#451 - (R9217 Auger sample) for TOC		200mR/hr		(15) wt. of sample Method submitted 2.31g
#578 - Hot Cell blank #1 for VOA		<5mR/hr		9.1g
#579 - Hot Cell blank #2 for VOA		<5mR/hr		11.7g
#580 - Hot Cell blank #3 for VOA		<5mR/hr		11.7g
#581 - Hot Cell blank #4 for VOA		<5mR/hr		11.9g
(16) Matrix (Other Metals or Anions Present)				
(17) Radioactivity Level (Actual ☐ Estimated ☐ Total Alpha _____ μ Ci/L Total Beta _____ μ Ci/L Total Gamma _____ μ Ci/L			(18) Additional Information (Measurement Uncertainty or Other Pertinent Information)	
(19) Estimated Cost			(20) Samples Received	
By _____ From _____ Date _____ Time _____			(21) Distribution of Final Results/Sample Disposal Instructions As per SOW	
Laboratory Manager B01-007			 	

SAMPLE RECEIPT FORM

Delivered by: WHL B-PIANT OVERPACK Date/Time: 2050 hrs 7/10/91
Received by: JK AAW RT STEELE
Customer Sample Number(s): #451 #574 #579 #580 #581
ALO Sample Number(s): 91-6424 91-6425 91-6426 91-6427 91-6428

1. Customer Chain-of-Custody Form: Present X Absent _____
2. Additional Shipping Forms (list):
OFFSITE RSR
RSA
3. Custody Seals on Shipping and/or Sample Containers and their Conditions.

Present X Absent _____

If Present, Condition: GOOD

4. Sample Tag(s) ID Numbers if not Recorded on the Chain-of-Custody Record or on Sample Vial.

Notes:

N/A

5. Condition of Shipping Container (i.e., broken container, dented, breached plastic bag, temperature of sample container as defined in Section 3.0 in PNL-ALO-051, etc.)

OK ~ 25°C (AMBIENT)

6. Condition of Sample Vials.

GOOD except LIDS HAVE NO SEPTA

7. Verification of Agreement or Nonagreement of Information on Receiving Documents.

Agreement

8. Resolution of Problems or Discrepancies.

N/A

RETURN COMPLETED FORM TO PROJECT MANAGER

Westinghouse Hanford
Company

CHAIN OF CUSTODY

101-SY
to PNL

Company Contact B.A. Crawford

Telephone 3-1972

Sample Collected by D.W. Edmonson

Date 7/23/91

Time 6:00 pm

Sample Locations _____

Ice Chest No. _____

Field Logbook and Page No. _____

Remarks Tamper Seal applied to N-55: Overpack on B-plant truck
Samples submitted for VOA analysis.

Bill of Lading No. _____

Offsite Property No. _____

Method of Shipment _____

Shipped to 325 Bldg Attn Rick Steele

Possible Sample Hazards see attached RSR

Sample Identification

2 Samples: #531 sludge samples 101-SY seg 19

529 " " 101-SY seg 22

Chain of Possession

Relinquished by:

B.A. Crawford

Received by:

B.A. Hoehle

Date/Time:

7/23/91 7:14 pm

Relinquished by:

B.A. Hoehle

Received by:

Rick S. Steele

Date/Time:

7-23-91 2158 hrs

Relinquished by:

Received by:

Date/Time:

Relinquished by:

Received by:

Date/Time:

	SAMPLE ANALYSIS REQUEST
--	--------------------------------

PART I: FIELD SECTION

Collector D. W. Edmonson Date Sampled — Time — hours
Company Contact B. A. Crawford Telephone () 373-1977

[illegible]

Field Information** Samples ~~was~~ contained in 20mL vials with screw caps. At the time of sampling no other vials were available (as per ^{previous} phone discussions with D. Bisenius).

Special Handling and/or Storage #531 reading ~0.0 Rad through sample wall
#529 reading 3.5 Rad.

Possible Sample Hazards Radioactive materials, caustic matrix.

PART II: LABORATORY SECTION

Given by _____ Title _____ Date _____

Analysis Required _____

B.9

****Use back of page for additional information relative to sample location.**

A-6000-406 (05/90)

SAMPLE RECEIPT FORM

Delivered by: B Plant overpack Truck Date/Time: 7-23-91 2158hrs

Received by: Steele / Rau / Homi

Customer Sample Number(s): # 531 + # 529

ALO Sample Number(s): 91-7174 + 91-7175

1. Customer Chain-of-Custody Form: Present X Absent _____

2. Additional Shipping Forms (list): COC

SAR
ORSR

3. Custody Seals on Shipping and/or Sample Containers and their Conditions.

Present X(?) Absent *

If Present, Condition: Good

4. Sample Tag(s) ID Numbers if not Recorded on the Chain-of-Custody Record or on Sample Vial.

Notes: NA

5. Condition of Shipping Container (i.e., broken container, dented, breached plastic bag, temperature of sample container as defined in Section 3.0 in PNL-ALO-051, etc.)

Good (31°C)

6. Condition of Sample Vials.

Good but wrong Lids

7. Verification of Agreement or Nonagreement of Information on Receiving Documents.

Agreed

8. Resolution of Problems or Discrepancies.

NA

* See JACK

RETURN COMPLETED FORM TO PROJECT MANAGER

Westinghouse Hanford
Company

CHAIN OF CUSTODY

Company Contact Ben Crawford Telephone 3-1972
Sample Collected by Wayne Edmonson Date _____ Time _____
Sample Locations 101-SY
Ice Chest No. _____ Field Logbook and Page No. 7/24
Remarks Samples submitted for TOC and Character VOA analyses
Bill of Lading No. _____ Offsite Property No. _____
Method of Shipment B-plant dig in N-55 Overpacks
Shipped to 325 Bldg Attn: Rick Steele
Possible Sample Hazards See RSR # 12221
7/24

Sample Identification

#533 - Sec 24-101-SY (for VOA analysis) Core 1
#443 - RSR 12221 Auger Sample

Chain of Possession

Relinquished by: <u>Bruce Hays</u>	Received by: <u>Mike Stanley</u>	Date/Time: <u>7-25-91 6:23</u>
Relinquished by: <u>Mike Stanley</u>	Received by: <u>Rick J. Steele</u>	Date/Time: <u>7-25-91 2:15</u>
Relinquished by:	Received by:	Date/Time:
Relinquished by:	Received by:	Date/Time:

A-6000-107 (04/70)

SAMPLE RECEIPT FORM

Delivered by: B-FLANT OVERPACE Truck Date/Time: 7-25-91 2150

Received by: J.K. RAW

Customer Sample Number(s): 533, 443

ALO Sample Number(s): 91-7319 & 91-7320

1. Customer Chain-of-Custody Form: Present X Absent _____
2. Additional Shipping Forms (list): ORSR
SAR
3. Custody Seals on Shipping and/or Sample Containers and their Conditions.
Present ✓ Absent _____

If Present, Condition: GOOD

4. Sample Tag(s) ID Numbers if not Recorded on the Chain-of-Custody Record or on Sample Vial.

Notes: N/A

5. Condition of Shipping Container (i.e., broken container, dented, breached plastic bag, temperature of sample container as defined in Section 3.0 in PNL-ALO-051, etc.) GOOD (28°C)

6. Condition of Sample Vials. OK

7. Verification of Agreement or Nonagreement of Information on Receiving Documents. AGREEMENT

8. Resolution of Problems or Discrepancies.

N/A

RETURN COMPLETED FORM TO PROJECT MANAGER

ALO CHAIN OF CUSTODY

<u>91-6264</u> ALO SAMPLE NUMBER	<u>GC/MS VOA</u> ANALYSIS REQUESTED	<u>Jar #424 (solid/sludge)</u> SAMPLE DESCRIPTION
SENDER <u>[Signature]</u>	<u>6/28/91</u> DATE	
RECEIVER <u>Diana L. Bellojatto</u>	<u>6/28/91</u> DATE	

<u>91-6265</u> ALO SAMPLE NUMBER	<u>GC/MS VOA</u> ANALYSIS REQUESTED	<u>Jar #420 (liquid)</u> SAMPLE DESCRIPTION
SENDER <u>[Signature]</u>	<u>6/28/91</u> DATE	
RECEIVER <u>Diana L. Bellojatto</u>	<u>6/28/91</u> DATE	

<u>91-6266</u> ALO SAMPLE NUMBER	<u>GC/MS VOA</u> ANALYSIS REQUESTED	<u>Jar #421 (liquid)</u> SAMPLE DESCRIPTION
SENDER <u>[Signature]</u>	<u>6/28/91</u> DATE	
RECEIVER <u>Diana L. Bellojatto</u>	<u>6/28/91</u> DATE	

<u>91-6267</u> ALO SAMPLE NUMBER	<u>GC/MS VOA</u> ANALYSIS REQUESTED	<u>Jar #563 (solid)</u> SAMPLE DESCRIPTION
SENDER <u>[Signature]</u>	<u>6/28/91</u> DATE	
RECEIVER <u>Diana L. Bellojatto</u>	<u>6/28/91</u> DATE	

Original - Project Management Office
Copy - Sender
Copy - Receiver

Applicable Test Instruction

TI-01SY-1

<u>91-6425</u> ALO SAMPLE NUMBER	<u>GC/MS VOA</u> ANALYSIS REQUESTED	<u>#578 HOT CELL BLANK</u> SAMPLE DESCRIPTION
SENDER _____	_____	DATE _____
RECEIVER _____	_____	DATE _____

<u>91-6426</u> ALO SAMPLE NUMBER	<u>GC/MS VOA</u> ANALYSIS REQUESTED	<u>#579 HOT CELL BLANK</u> SAMPLE DESCRIPTION
SENDER _____	_____	DATE _____
RECEIVER _____	_____	DATE _____

<u>91-6427</u> ALO SAMPLE NUMBER	<u>GC/MS VOA</u> ANALYSIS REQUESTED	<u>#580 HOT CELL BLANK</u> SAMPLE DESCRIPTION
SENDER _____	_____	DATE _____
RECEIVER _____	_____	DATE _____

<u>91-6428</u> ALO SAMPLE NUMBER	<u>GC/MS/VOA</u> ANALYSIS REQUESTED	<u>#581 HOT CELL BLANK</u> SAMPLE DESCRIPTION
SENDER _____	_____	DATE _____
RECEIVER _____	_____	DATE _____

Original - Project Management Office
Copy - Sender
Copy - Receiver

Applicable Test Instruction

TI-101SY-2

SEE ATTACHED FOR SIGNATURES

9/25/91 ALO
CHAIN OF CUSTODY

~~Chain of Custody No. TI-10154-2~~
 Chain of custody No. TI-10154-02A
 DS 9/25/91

<u>91-6425</u> SAMPLE NUMBER	<u>GC/MS VOA</u> ANALYSIS REQUESTED	<u>#578 Hot Cell BLANK</u> SAMPLE DESCRIPTION
SENDER <u>Rich I. Steele</u>		<u>7-15-91</u> DATE
RECEIVER <u>Darrell Ross</u>		<u>7/15/91 1140 am</u> DATE

<u>91-6426</u> SAMPLE NUMBER	<u>GC/MS VOA</u> ANALYSIS REQUESTED	<u>#579 Hot Cell BLANK</u> SAMPLE DESCRIPTION
SENDER <u>Rich I. Steele</u>		<u>7-15-91</u> DATE
RECEIVER <u>Darrell Ross</u>		<u>7/15/91 1140 am</u> DATE

<u>91-6427</u> SAMPLE NUMBER	<u>GC/MS VOA</u> ANALYSIS REQUESTED	<u>#580 Hot Cell BLANK</u> SAMPLE DESCRIPTION
SENDER <u>Rich I. Steele</u>		<u>7-15-91</u> DATE
RECEIVER <u>Darrell Ross</u>		<u>7/15/91</u> DATE

<u>91-6428</u> SAMPLE NUMBER	<u>GC/MS VOA</u> ANALYSIS REQUESTED	<u>#581 Hot Cell BLANK</u> SAMPLE DESCRIPTION
SENDER <u>Rich I. Steele</u>		<u>7-15-91</u> DATE
RECEIVER <u>Darrell Ross</u>		<u>7/15/91</u> DATE

Original - Project Management Office
 Copy - Sender
 Copy - Receiver

Applicable Test INSTRUCTION

TI-10154-2

ALO CHAIN OF CUSTODY

<u>91-7174</u> ALO SAMPLE NUMBER	<u>VOA ORGANICS</u> ANALYSIS REQUESTED	<u>#531</u> SAMPLE DESCRIPTION
SENDER <u>Rich L. Steele</u>		<u>7-26-91</u> DATE
RECEIVER <u>Diana L. Bellozatto</u>		<u>7-26-91</u> DATE

<u>91-7175</u> ALO SAMPLE NUMBER	<u>VOA ORGANICS</u> ANALYSIS REQUESTED	<u>#529</u> SAMPLE DESCRIPTION
SENDER <u>Rich L. Steele</u>		<u>7-26-91</u> DATE
RECEIVER <u>Diana L. Bellozatto</u>		<u>7-26-91</u> DATE

<u>91-7319</u> ALO SAMPLE NUMBER	<u>VOA ORGANICS</u> ANALYSIS REQUESTED	<u>#533</u> SAMPLE DESCRIPTION
SENDER <u>Rich L. Steele</u>		<u>7-26-91</u> DATE
RECEIVER <u>Diana L. Bellozatto</u>		<u>7-26-91</u> DATE

<u>ALO SAMPLE NUMBER</u>	<u>ANALYSIS REQUESTED</u>	<u>SAMPLE DESCRIPTION</u>
SENDER _____		_____ DATE
RECEIVER _____		_____ DATE

Original - Project Management Office
Copy - Sender
Copy - Receiver

Applicable Test Instruction

TI-101SY-3

DISTRIBUTION

No. of
Copies

OFFSITE

2 DOE/Office of Scientific and
Technical Information

ONSITE

4 Westinghouse Hanford Company

J. R. Jewett, T6-50 (4)

14 Pacific Northwest Laboratory

D. L. Baldwin, P7-22

B. M. Gillespie, P7-27 (5)

F. V. Hoopes, P7-22

R. W. Stromatt, P7-22

Publishing Coordination

Technical Report Files (5)

END

**DATE
FILMED**

3 / 24 / 92