1 of 2

VED
JAN 20 1994
OSTI

Measurement of Elemental Speciation by Liquid Chromatography - Inductively
Coupled Plasma Mass Spectrometry (LC-ICP-MS) with the Direct Injection
Nebulizer (DIN)

by Sam Shum

thesis submitted to Iowa State University

Ames Laboratory, U. S. DOE

Iowa State University

Ames, Iowa 50011

Date Transmitted: May 1993

Prepared for the U. S. Department of Energy

Under contract no. W-7405-eng-82.

DISCLAIMER

This report was prepared as an account of work sponsored by an agency of the United States Government. Neither the United States Government nor any agency thereof, nor any of their employees, makes any warranty, express or implied, or assumes any legal liability or responsibility for the accuracy, completeness, or usefulness of any information, apparatus, product, or process disclosed, or represents that its use would not infringe privately owned rights. Reference herein to any specific commercial product, process, or service by trade name, trademark, manufacturer, or otherwise does not necessarily constitute or imply its endorsement, recommendation, or favoring by the United States Government or any agency thereof. The views and opinions of authors expressed herein do not necessarily state or reflect those of the United States Government or any agency thereof.

MASTER

ds
DISTRIBUTION OF THIS DOCUMENT IS UNLIMITED

Measurement of elemental speciation by liquid chromatography - inductively coupled plasma mass spectrometry (LC-ICP-MS) with the direct injection nebulizer (DIN)

Sam Chiu-Kin Shum

In charge of major work: R. S. Houk
From the Department of Chemistry
Iowa State University

A new version of the direct injection nebulizer (DIN) is used to interface liquid chromatographic (LC) separations with element - selective detection by inductively coupled plasma - mass spectrometry (ICP-MS). The advantages of using LC-DIN-ICP-MS include little dead volume, low sample and solvent consumption, excellent plasma stability when nebulizing samples containing concentrated organic solvent, excellent absolute detection limits, excellent precision, superior chromatographic resolution, and reduced memory effects from memory-prone elements (e.g., Hg).

Various compounds containing arsenic (AsO_2^- , HAsO_4^{2-} , $\text{CH}_3\text{AsO}_3^{2-}$ and $(\text{CH}_3)_2\text{AsO}_2^-$), tin ($\text{CH}_3\text{Sn}^{3+}$, $(\text{CH}_3)_2\text{Sn}^{2+}$, $(\text{C}_2\text{H}_5)_2\text{Sn}^{2+}$ and $(\text{CH}_3)_3\text{Sn}^+$), mercury (Hg^{+2} , CH_3Hg^+ , $\text{C}_2\text{H}_5\text{Hg}^+$ and $\text{C}_6\text{H}_5\text{Hg}^+$) and lead (Pb^{+2} , $(\text{CH}_3)_3\text{Pb}^+$ and $(\text{C}_2\text{H}_5)_3\text{Pb}^+$) are separated by reversed-phase ion-pairing LC. Selenium species (SeO_3^{2-} and SeO_4^{2-}) are separated by ion chromatography (IC). Several metalloproteins containing Na, Fe, Cu, Zn, Cd, Ba, and Pb are separated by size-exclusion chromatography (SEC).

Some of these compounds are separated and measured in biological and environmental samples such as human urine and human serum. Detection limits are $\approx 3, 0.7, 1, 0.5, 5, 0.5, 15, 7,$ and 0.2 pg for Fe, Cu, Zn, As, Se, Cd, Sn, Hg, and Pb, respectively. Microbore (1 - 2 mm i.d.) packed columns are employed because the liquid flow rates used ($30 - 100 \mu\text{L min}^{-1}$) are compatible with the DIN.

Aerosol droplet sizes and velocities from a DIN are measured with radial and axial spatial resolution by phase doppler particle analysis (PDPA). The droplets on the central axis of the spray become finer and their size becomes more uniform when $\approx 20\%$ methanol is added to the usual aqueous solvent. This could explain why the analyte signal is a maximum at this solvent composition when the DIN is used for ICP-MS. Mean droplet velocities are 12 to 22 m s^{-1} with standard deviations of ± 4 to $\pm 7 \text{ m s}^{-1}$. The outer fringes of the aerosol plume tend to be enriched in large droplets. The Sauter mean diameter ($D_{3,2}$) and velocity of the droplets also vary substantially with axial position in the aerosol plume. These findings are valuable for improving the analytical performance of the DIN.

Measurement of elemental speciation by liquid chromatography - inductively coupled
plasma mass spectrometry (LC-ICP-MS) with the direct injection nebulizer (DIN)

by

Sam Chiu-Kin Shum

A Dissertation Submitted to the
Graduate Faculty in Partial Fulfillment of the
Requirements for the Degree of
DOCTOR OF PHILOSOPHY

Department: Chemistry
Major: Analytical Chemistry

Approved:

In Charge of Major Work

For the Major Department

For the Graduate College

Iowa State University
Ames, Iowa

1993

TABLE OF CONTENTS

GENERAL INTRODUCTION	1
 PAPER I. ELEMENTAL SPECIATION BY LIQUID CHROMATOGRAPHY - INDUCTIVELY COUPLED PLASMA MASS SPECTROMETRY WITH DIRECT INJECTION NEBULIZATION	 13
INTRODUCTION	14
EXPERIMENTAL SECTION	16
RESULTS AND DISCUSSION	22
CONCLUSION	28
LITERATURE CITED	29
 PAPER II. SPECIATION OF MERCURY AND LEAD COMPOUNDS BY MICROBORE COLUMN LIQUID CHROMATOGRAPHY - INDUCTIVELY COUPLED PLASMA MASS SPECTROMETRY WITH DIRECT INJECTION NEBULIZATION	 46
INTRODUCTION	47
EXPERIMENTAL SECTION	49
RESULTS AND DISCUSSION	54
LITERATURE CITED	65

PAPER III. SPATIALLY RESOLVED MEASUREMENTS OF SIZE AND VELOCITY DISTRIBUTIONS OF AEROSOL DROPLETS FROM A DIRECT INJECTION NEBULIZER	83
INTRODUCTION	84
EXPERIMENTAL SECTION	86
RESULTS AND DISCUSSION	89
CONCLUSION	101
LITERATURE CITED	102
PAPER IV. ELEMENTAL SPECIATION BY ANION EXCHANGE AND SIZE EXCLUSION CHROMATOGRAPHY WITH DETECTION BY INDUCTIVELY COUPLED PLASMA MASS SPECTROMETRY WITH DIRECT INJECTION NEBULIZATION	122
INTRODUCTION	123
EXPERIMENTAL SECTION	126
RESULTS AND DISCUSSION	130
CONCLUSION	135
LITERATURE CITED	136
SUMMARY	151
ADDITIONAL LITERATURE CITED	155
ACKNOWLEDGEMENTS	161

GENERAL INTRODUCTION

Since its introduction in 1980 (1), inductively coupled plasma mass spectrometry (ICP-MS) has become a widely used analytical technique for elemental analysis (2-14). Furthermore, ICP-MS has been used as an element-selective detector for chromatography to provide elemental speciation information (15-46). In elemental speciation, one wants to identify and quantify various chemical species that together comprise the total element concentration in a sample. This is important because the toxicity and biological importance of many trace elements depend greatly on their chemical forms and/or oxidation states. A good example is arsenic. Figure 1 shows six common arsenic species. As(III) and As(V) are the most toxic; administration of 21 μg of As in the form of As(III) and 41 μg of As in the form of As(V) causes 50% mortality in rats within 96 hours (47). Monomethylarsonic acid and dimethylarsinic acid are only moderately toxic (48), while arsenobetaine and arsenocholine are relatively nontoxic (49,50). In fact, biotransformation of inorganic arsenic to DMA and finally to arsenobetaine is the detoxification mechanism for many marine animals such as fish, shrimp and lobster (51-56). The end product is either excreted or stored.

Atomic spectroscopic detection methods such as atomic mass spectrometry, atomic emission spectrometry, atomic absorption spectrometry and atomic fluorescence generally measure only the total concentration or amount for the

As(III)

As(V)

Monomethylarsonic acid (MMA)

Dimethylarsinic acid (DMA)

Arsenobetaine

Arsenocholine

Figure 1. Six common arsenic compounds.

elements of interest. However, by coupling with chromatographic separations, atomic spectroscopic detection methods become species selective detection methods, i.e., speciation information can be obtained (15). Here various elemental species will first be separated by chromatographic methods and then observed one at a time by the detection system.

Inductively coupled plasma mass spectrometry (ICP-MS) is one of the most attractive detection systems for elemental speciation because of its extremely high sensitivity and low detection limits. So far, various separation schemes, including reversed-phase (RP) (16-20), reversed-phase ion-pairing (RP-IP) (21-29), ion chromatography (IC) (16,23,25,30-35), size exclusion chromatography (SEC) (16,36-43), supercritical fluid chromatography (SFC) (44) and gas chromatography (GC) (45,46) have been coupled with ICP-MS. Nevertheless, there is still substantial room for improvement including better chromatographic separations and lower detection limits.

An Overview of ICP-MS

The ICP is an atmospheric pressure discharge in a flowing gas stream (usually argon). Electrical power is added inductively to the plasma from a radio-frequency generator (57). The temperature of an ICP is in the range of 6000 to 8000 K (58,59). Due to the high temperature of the plasma, sample constituents travelling through the axial channel of the ICP will be vaporized, atomized and ionized. For most elements,

the degree of ionization in the plasma is almost 100% (14). Exceptions are those elements with high ionization energies (IE), such as Hg (IE = 10.5 eV, only 38% ionized).

Since the ICP operates at atmospheric pressure, ions formed in the plasma must be extracted into a vacuum system before they can be mass analyzed. Figure 2 shows the schematic diagram of an ICP and a typical ion extraction interface (14). Sample in the form of aerosol is injected into the plasma (Figure 2, B) through the axial channel of the torch (Figure 2, C). A metal cone, the sampler, with a circular orifice of about 1mm diameter is immersed into the normal analytical zone where ions are located (Figure 2, E). Ions and neutral gas flow from the ICP through the sampling orifice into the first stage of the vacuum chamber, which is evacuated by a mechanical pump to a pressure of about 1 torr. Due to the reduction in pressure, a supersonic jet forms behind the sampler. The central section of the jet flows through a second orifice called the skimmer. Behind the skimmer, several ion lenses are used to focus and direct ions into a quadrupole mass analyzer (3,60).

For most ICP-MS experiments, a pneumatic nebulizer is used to introduce liquid sample to the plasma. Figure 3 shows a schematic diagram of a concentric pneumatic nebulizer with a spray chamber (61). Liquid sample is introduced through a narrow tube, and at the exit end, the liquid is shattered into aerosol droplets by a stream of Ar gas. The spray chamber prevents large droplets from reaching the plasma and removes as much solvent as possible when cooled (62,63). However,

Figure 2. ICP and ion sampling interface. A = torch and load coil (HV = high voltage), B = induction region of ICP, C = a solution aerosol being injected into axial channel, D = initial radiation zone, E = normal analytical zone, F = nickel cone with sampling orifice in tip, G = skimmer cone, H = boundary layer of ICP gas deflected outside sampling orifice, I = expanding jet of C gas sampled from ICP, and J = ion lens elements. Reproduced from reference (14) with permission.

Figure 3. Concentric pneumatic nebulizer and spray chamber for introducing solution into an ICP. Reproduced from reference (61) with permission.

there are shortcomings of this type of nebulizer. First, only 1 - 3% of the sample reaches the plasma while most of the sample goes down to the drain. Second, the spray chamber has a large dead volume that can cause substantial chromatographic band broadening when used to couple LC to ICP-MS.

Direct Injection Nebulization

For the present work, a new version of the direct injection nebulizer (DIN) was used to couple LC to ICP-MS (27,29,64). Figure 4 shows the schematic diagram of a DIN and a regular DIN torch. The design and construction of the DIN have been discussed elsewhere (27,65). In brief, a DIN is composed of a nebulizer tip, a ceramic support tube, some swagelok fittings and a capillary sample transfer line. Typically, a 50 μ m internal diameter x 40 cm long capillary is used. A stream of pressurized Ar gas passes through the swagelok tee (Figure 4, C) and exits with high linear velocity at the end of the nebulizer tip. Aerosol is produced pneumatically at the tip of the nebulizer. When fully assembled, the direct injection nebulizer is placed inside a DIN torch. The nebulizer tip is positioned even with the inner tube of the torch and only 2 - 3 mm from the base of the plasma. With the DIN, 100% of the sample is introduced into the plasma and the use of a spray chamber is unnecessary.

In this dissertation results for the elemental speciation of more than 20 compounds of arsenic, tin, mercury, lead, selenium and metalloproteins are presented.

Figure 4. DIN and DIN torch. A = stainless steel nebulizer tip, B = ceramic support tube, C = swagelok tee and fittings, D = fused silica capillary for introducing liquid sample, E = entrance for the DIN, F = make-up gas inlet, G = auxilliary gas inlet, H = plasma gas inlet.

DIRECT INJECTION NEBULIZER

The development of the direct injection nebulizer, a new device used to couple microbore LC to ICP-MS, is emphasized.

Explanation of Dissertation Format

This dissertation is composed of four papers formatted for publication in four different journals. Each paper stands independent of the others as a complete scientific manuscript with accompanying tables, figures and literature cited. References cited in this general introduction and the summary are given after the summary.

In the first paper, a new version of the direct injection nebulizer (DIN) is used to interface liquid chromatographic (LC) separations with element - selective detection by inductively coupled plasma - mass spectrometry (ICP-MS). The DIN injects all the sample into the ICP and has a dead volume of less than 1 μ L. Charged species of As and Sn are separated as ion pairs on a micro-scale (1 mm i.d.), packed, reversed phase column. Detection limits are 0.2 - 0.6 pg As and 8 - 10 pg Sn. For methanol/water eluents, signal is highest at 25% methanol and stays within 25% of this maximum as the methanol fraction is varied from 20% to 80%. Compared to LC-ICP-MS with conventional nebulizers, the absolute detection limits and chromatographic resolution are substantially superior, and the dependence of analyte signal on solvent composition is somewhat less severe with the DIN.

In the second paper, various cationic species of Hg (Hg^{+2} , MeHg^+ , EtHg^+ ,

and PhHg^+) and Pb (Pb^{+2} , $(\text{Me})_3\text{Pb}^+$, and $(\text{Et})_3\text{Pb}^+$) are separated as ion pairs by reversed-phase liquid chromatography. Lead is detected at ≈ 0.2 pg and Hg at ≈ 7 pg by inductively coupled plasma - mass spectrometry (ICP-MS). A direct injection nebulizer (DIN) minimizes band broadening and yields good sensitivity by injecting all the sample into the plasma. Chromatographic conditions such as mobile phase composition and the compound used for ion pairing are selected based on the mutual requirements of chromatographic performance and detection sensitivity. The inorganic ions Pb^{+2} and Hg^{+2} can be monitored directly in human urine by simply diluting and injecting the liquid sample. The organolead and organomercury ions were not concentrated enough to be seen directly in urine, but good separations were obtained when these compounds were spiked into the sample.

In the third paper, aerosol droplet sizes and velocities from a direct injection nebulizer (DIN) are measured with radial and axial spatial resolution by phase doppler particle analysis (PDPA). The droplets on the central axis of the spray become finer and their size becomes more uniform when $\approx 20\%$ methanol is added to the usual aqueous solvent. This could explain why the analyte signal is a maximum at this solvent composition when the DIN is used for inductively coupled plasma - mass spectrometry (ICP-MS). Mean droplet velocities are 12 to 22 m s^{-1} with standard deviations of ± 4 to ± 7 m s^{-1} . The outer fringes of the aerosol plume tend to be enriched in large droplets. The Sauter mean diameter ($D_{3,2}$) and velocity of the droplets also vary substantially with axial position in the aerosol plume.

In the last paper, a direct injection nebulizer (DIN) is used with packed microbore columns for anion chromatography (AC) and size exclusion chromatography (SEC). Two Se species (SeO_3^{2-} and SeO_4^{2-}) are separated by AC with detection limits of 15 pg Se. The isotope ratio $^{74}\text{Se}/^{78}\text{Se}$ can be measured with a relative standard deviation of 0.5% on 25 ng of each Se species during the actual chromatographic separation. Proteins in human serum are separated by SEC without sample pretreatment. The metals present in each molecular weight fraction are determined by ICP-MS with detection limits of 0.5 - 3 pg metal. These absolute detection limits are 1 - 2 orders of magnitude better than those obtained with conventional nebulizers.

PAPER I

ELEMENTAL SPECIATION BY LIQUID CHROMATOGRAPHY -
INDUCTIVELY COUPLED PLASMA MASS SPECTROMETRY WITH DIRECT
INJECTION NEBULIZATION

INTRODUCTION

Atomic spectroscopic methods generally measure only the total concentration or amount of the element(s) of interest. However, these methods can provide speciation information when combined with chromatographic separations. Inductively coupled plasma mass spectrometry (ICP-MS) has demonstrated excellent sensitivity and selectivity as a chromatographic detector both in initial demonstration studies (1-7) and in several real applications. Examples of the latter include the measurement of (a) As species in fish tissue (7-9) and urine (10,11), (b) various elements in polychaetes (12), (c) Cd species in foods (13), and (d) species of Pb (14) and Au (15) in blood.

Although some speciation applications can be addressed by chromatographic separations with ICP-MS detection in its present state, there is still substantial room for improvement. Most liquid chromatography-ICP-MS (LC-ICP-MS) studies have employed conventional nebulizers, which introduce only 1-3% of the sample into the plasma and have large dead volumes that can cause band broadening. In fact, the nebulizer is generally recognized to be one of the weakest components of the entire ICP-MS apparatus.

Fassel and co-workers (16,17) addressed these problems for ICP emission spectrometry by using a direct injection nebulizer (DIN). A microconcentric pneumatic nebulizer is positioned inside the torch with its tip only 3 - 4 mm from the

base of the plasma. Virtually all of the sample is injected into the plasma, and the dead volume is only 2 μl or less. A new version of the DIN that is simpler to construct and more rugged was described recently by Wiederin et al. (18), as were some initial results pertaining to the use of the new DIN for LC separations (19). The work described in this paper represents a more thorough evaluation of the DIN as an interface between an LC column and an ICP-MS detector. Micro-scale LC separations in small, packed columns are studied because the liquid flow rates used ($\approx 30 \mu\text{l min}^{-1}$) are compatible with the DIN and the dead volume is low enough to prevent excessive band broadening. Conventional nebulizers would have problems in both these areas if used with such columns.

EXPERIMENTAL SECTION

ICP-MS Device and Conditions

The inductively coupled plasma mass spectrometer used was the Elan Model 250 (Perkin-Elmer Sciex, Thornhill, Ontario, Canada). Detailed descriptions of the instrumental system and conditions are given in Table 1. Various plasma and sampling conditions were optimized daily to maximize the signal from the analyte of interest.

DIN Construction and Characteristics

The DIN used in this work was modified somewhat from the one described previously (Fig. 1 of ref. 18). These modifications are summarized in Table 2. The fused silica capillary was narrower (30 compared with 50 μm i.d. for the previous DIN), which minimized dead volume and post-column band broadening. The outer tube of the nebulizer tip (i.e., the tube that surrounded the open end of the sample capillary) was ceramic and the support tube was stainless steel. The width of the annular gap between the inner capillary and the nebulizer tip was approximately 25 μm , as for the previous work (18).

Some operating characteristics of the 30 μm DIN were very different from those of the previous DIN (18). In the present work, the nebulizer gas flow rate that produced maximum ion signal was only 0.12 l min⁻¹. The total gas flow rate (i.e.,

nebulizer flow + make up flow = 0.42 l min^{-1}) was also much lower than that obtained with the previous DIN (1.30 l min^{-1}). Finally, the finer $30 \mu\text{m}$ capillary became plugged more readily. For example, nebulization of a 5 mmol dm^{-3} solution of sodium dodecylsulfate for only 15 min completely plugged the $30 \mu\text{m}$ capillary. Plugging was not a serious problem during the chromatography work; eluents containing metal salts were avoided because they also suppressed analyte signals by the usual matrix effects (20-26).

Arsenic and Alkyltin Separations

A schematic diagram of the separation system is given in Fig. 1. A digital high-performance liquid chromatography (HPLC) pump (SSI Model 222D) with a bioclean micro-flow pump head (Scientific Systems, State College, PA, USA) was employed. The solvent switching valve (Fig. 1, C) was a Rheodyne 9010 metal-free injector with a 3 ml PEEK injection loop. Solvents of different composition were loaded onto the 3 ml loop and then connected to the main stream. The solvent in a 3 ml loop lasted approximately 100 min. Samples were injected using a Rheodyne 7410 micro-injector with a $0.5 \mu\text{l}$ internal sample loop disc. The analytical column (Fig. 1, F) used in all studies was a 10 cm long, 1 mm i.d. metal-free GLT column packed with $5 \mu\text{m}$ diameter Inertsil ODS-2 material (Scientific Glass Engineering, Austin, TX, USA). A silica pre-column was placed between the pump and the analytical column to saturate the mobile phase with silica. The connections (Fig. 1, H₁ and H₂)

from the column inlet to the sample injector and the column outlet to the switching valve were 5 cm long, 50 μm i.d. polysil tubes (Scientific Glass Engineering). The switching valve was another Rheodyne 9010 injector with a 1 ml PEEK injection loop. Standard solutions for optimizing ICP-MS conditions were loaded onto the 1 ml loop and then injected into the nebulizer. The fused silica capillary of the DIN was connected to the switching valve. The outlet of the analytical column was directed into the DIN at all times except during optimization. Mobile phase flow rate was maintained at 30 $\mu\text{l min}^{-1}$. The dead volume of the system was $< 1 \mu\text{l}$.

The separation conditions for both arsenic and alkyltin speciation are summarized in Table 3. The analytical column was equilibrated with the ion pairing reagent and mobile phase prior to use. Several different combinations of organic modifier concentration, type and concentration of counter ion, pH, etc. were evaluated to optimize chromatographic performance. The conditions listed in Table 3 are those that yielded the best chromatographic resolution of the various sets tested.

The ICP-MS device was operated in multi-element monitoring mode (Table 4). Plasma and sampling conditions were optimized by introducing a solution composed of As and Sn at 250 $\mu\text{g l}^{-1}$ each. Backgrounds were determined by a blank injection. The count rates were smoothed with a ten-point Savitzky - Golay routine (27). Peak area was determined by summing all the counts under a peak. For this work, the detection limit is defined as the amount of the element necessary to give a peak area equal to three times the standard deviation of the background count rate at each mass.

The detection limits for each species were estimated from their calibration curves (log - log plot).

Effect of Solvent Composition on Ion Signal

Flow injection techniques were employed. The inlet of the fused silica capillary was connected to the first injector (Fig. 1, C); i.e., the analytical column was bypassed. Stock solutions of three easily ionized elements (Ga, Nd and Tl) and four moderately ionized elements (As, Se, Sn and Hg) at $200 \mu\text{g l}^{-1}$ each in various solvents were prepared. These solutions were then loaded onto the injection loop and nebulized. The pump and transfer line leading to the injector were flushed with the appropriate solvent blank after each change of solvent composition. A solution of Rh (i.e. an element in the middle of the mass range) at $200 \mu\text{g l}^{-1}$ was introduced to the plasma for optimization. The solution flow rate was maintained at $50 \mu\text{l min}^{-1}$.

Ion count rates for all seven elements were measured in multi-element monitoring mode (Table 4). Backgrounds were determined using approximately 30 points prior to injection of the sample solution. Count rates were determined by averaging about 30 points of steady-state signal, followed by background subtraction.

Effect of Liquid Flow Rate on Ion Signal

Instrumentation, experimental conditions and data acquisition were the same as those for the experiments carried out to determine the effect of solvent composition.

A solution of six elements (Be, As, Se, Sn, Nd and Hg) at $300 \mu\text{g l}^{-1}$ each in distilled, de-ionized water (DDW) was introduced into the plasma. The plasma and sampling conditions were readjusted at each flow rate being studied to obtain the maximum response for Rh.

Reagents and Samples

The mobile phases were prepared as follows. Ion pairing reagents Q7 (0.5 mol dm^{-3} heptatriethylammonium phosphate) and S7 (0.5 mol dm^{-3} sodium heptanesulfonate) were purchased from Regis Chemical Company (Morton Grove, IL, USA). For As speciation, mobile phases of Q7 were prepared by adding appropriate organic modifier (HPLC grade) to the ion-pairing reagent and then diluting with DDW. For alkyltin speciation, ammonium heptanesulfonate was prepared by passing S7 through a column filled with cation-exchange resin Dowex 50W - X8 in the ammonium form. The resulting stock solution of ammonium heptanesulfonate was then diluted with organic modifier (HPLC grade) and DDW. Nitric acid (5%) and ammonium hydroxide (1 mol dm^{-3}) were added to adjust the pH. All mobile phases were filtered through a $0.45\text{-}\mu\text{m}$ pore nylon filter and de-gassed under light vacuum from an aspirator for 20 min. After use, the column was flushed overnight with a solution of 50% methanol in water at a flow rate of $50 \mu\text{l min}^{-1}$.

Sodium arsenate (As^{V}), sodium arsenite (As^{III}), dimethylarsinic acid (DMAA), monomethyltin trichloride (MMT-TCl), dimethyltin dichloride (DMT-DCI),

trimethyltin chloride (TMT-Cl), and diethyltin dichloride (DET-DCI) were obtained from Alfa Products (Danvers, MA, USA) and were used without further purification. Stock solutions of arsenic and alkyltin species were prepared from the above analytical-reagent grade compounds in DDW. A previously made stock solution of monomethylarsonic acid (MMAA) containing $100 \mu\text{g l}^{-1}$ As was used.

Stock solutions of seven elements (Ga, As, Se, Sn, Nd, Hg and Tl) and six elements (Be, As, Se, Sn, Nd and Hg) were prepared by diluting 1000 ppm standards (PLASMACHEM Associates, Bradley Beach, NJ, USA) in solvents with the appropriate organic modifier. All other chemicals used were of analytical-reagent grade.

RESULTS AND DISCUSSION

Effect of Liquid Flow Rate on Ion Signal

Fig. 2 shows the effect of liquid flow rate on analyte signal with the present 30 μm DIN at very low flow rates. After each change of liquid flow rate, the plasma and sampling conditions (Table 1) were readjusted to yield maximum analyte signal. The ion signals of all six elements studied varied linearly with liquid flow rate if the ICP conditions were readjusted to compensate for the change in solvent load. For each element, the correlation coefficient for a plot of ion signal versus liquid flow rate was 0.994 - 0.999. The sensitivity that could be obtained after readjustment dropped with flow rate because the ICP-MS is, of course, a mass flow sensitive detector. Most of the present work was performed at a liquid flow rate of 30 $\mu\text{l min}^{-1}$. Previous work has shown that a 30 μm DIN provides useful analyte signals even at flow rates as low as 4 $\mu\text{l min}^{-1}$. Hence, ICP-MS detection could conceivably be used with very low flow rate separation methods such as open tubular LC or capillary zone electrophoresis (CZE). However, operation at such low liquid flow rates would require one of the new generation ICP-MS devices (28) with 100 times better sensitivity than that of the veteran instrument used in this work.

Effect of Solvent Composition on Ion Signal

In LC-ICP-MS, increasing the organic content of the mobile phase generally

causes problems such as a decrease in analyte sensitivity (29,30), an increase in the abundance of polyatomic ions containing carbon (29,31,32), and deposition of solid carbon on the sampling orifice (29,30). Nebulization of volatile solvents such as methanol can extinguish the plasma or make it unstable (33-37). For these reasons, most LC experiments with ICP detection have been isocratic with aqueous eluents containing only modest amounts of organic modifiers (e.g., 30% methanol or less). In this section, the effects of changing solvent composition on analyte signal with the DIN are described.

With the DIN, the plasma tolerated MeOH or acetonitrile (ACN) very well. No green C_2 emission was observed visually from the plasma unless the solvent contained at least 60% MeOH. The analyte signal did not drift perceptibly when 85% MeOH was nebulized for at least one hour. Presumably, oxygen from the H_2O and MeOH helped prevent deposition of carbon on the sampling cone (29,30,38). The plasma did not go out and remained stable when either 100% MeOH or 70% ACN were nebulized; however, carbon deposition on the sampling cone from these latter solvents caused substantial signal drift.

Fig. 3 shows the effect of solvent composition on ion signal for three easily ionized elements (Ga, Nd and Tl). The ion signal increased sharply for mixtures containing up to 25% MeOH and then decreased very gradually at MeOH fractions above 25%. Results from replicate experiments showed that the maximum analyte signal was in the range of 20 - 30% MeOH in water, and 15 - 25% ACN in water.

Normalized curves for the same elements in ACN were almost identical with those in Fig. 3 except that the decrease in the ion signal for ACN was more marked.

Presumably, the plasma was cooled more extensively by ACN than MeOH. For these easily ionized elements, the analyte signal varied only by $\pm 15\%$ as the solvent composition changed from 20% to 80% MeOH (Fig. 3). A reasonable plateau was found between 15 and 60% ACN.

Fig. 4 shows the effect of solvent composition on ion signal for four moderately ionized elements (As, Se, Sn and Hg). The general shape of the curves was similar to those in Fig. 3 except that the decreases at higher methanol concentrations were steeper in Fig. 4. The signals for these moderately ionized elements were more dependent on solvent composition than the signals for the easily ionized elements. Presumably this is because plasma cooling had more of an effect on the ionization efficiency for elements with high ionization energies. Nevertheless, analyte signal varied only by $\pm 25\%$ as the solvent composition changed from 20 to 80% MeOH (Fig. 4). For ACN, a reasonable plateau was found between 15% and 60% ACN. Thus mild solvent gradients (in HPLC) in this range should not induce extreme changes in analyte sensitivity with the DIN. This capability is an attractive feature because the species selectivity in LC-ICP-MS is based solely on retention, thus good chromatographic separations are necessary. Use of the DIN outside the torch in a spray chamber with desolvation may remove enough solvent to facilitate the use of solvent gradients, although band broadening may be substantial in the aerosol

transfer line between the nebulizer and the plasma.

Arsenic Separation

A typical chromatogram of a solution containing As^{III} , DMAA, MMAA, and As^{V} is shown in Fig. 5. Separation conditions are given in Table 3. The analytical figures of merit under these separation conditions are shown in Table 5. The peaks for both inorganic and organic As species were baseline resolved, compared with previous work (1,19) where either As^{III} and MMAA, or DMAA and MMAA could not be resolved. The pH was held at 6.0 to keep all arsenic species in their ionic forms. At a higher pH (≈ 7.0), the retention time for MMAA increased from 7.6 min to 9.6 min, while the retention time for As^{V} increased to > 30 min. Peak-area measurements indicated that the sensitivity (count rate per ng of As) was similar for the various forms of As (Table 5). The precision based on five replicate injections of approximately 0.5 ng (as As) of each species and measurement of peak areas was better than 4% RSD for all four species (Table 5). Calibration curves for each species based on injections of 0.05, 0.5, 1, 5 and 10 ng (as As) and peak areas were all linear with correlation coefficients of 0.9994 or better. Absolute detection limits were approximately 0.6 pg of As for all four forms (Table 5). These absolute detection limits are superior by factors of 50 to 500 over those obtained previously by LC-ICP-MS with conventional nebulizers (1,6-11). The improvement in absolute detection limits is expected because all of the sample reaches the plasma with the

DIN. Relative detection limits for As species are about 1 ppb. As the injected volume in the present work is very low ($0.5\ \mu\text{l}$), these relative detection limits are comparable to or slightly worse than those obtained with conventional nebulizers and much larger injections.

A step gradient can be used to elute the last As peak more rapidly (Fig. 6). The first three peaks were eluted with 5% methanol in water for 11 min. The As^{V} peak was then eluted by stepping the eluent up to 25% methanol in water. The step gradient was performed by storing the second eluent (25% methanol in water) in the 3 ml loop (Fig. 1, C) in front of the column and then injecting this loop 11 min after sample injection. Note that both the baseline and sensitivity of the fourth peak increased. The reproducibility of this separation is still approximately 4% RSD (five replicate injections) for all four species. The absolute detection limit for the fourth peak (As^{V}) is $\approx 0.2\ \text{pg}$. The improved detection limit is due to a great extent to the enhancement in sensitivity when the eluent is 25% MeOH (Fig. 4).

Alkyltin Separation

A typical chromatogram of a standard mixture of MMT-TCl, DMT-DCl, DET-DCl, and TMT-Cl is shown in Fig. 7. Pertinent conditions are given in Table 3. The analytical figures of merit for this separation method are given in Table 6. A reversed-phase cation pairing separation mode was used here. The cation pairing reagent was in the ammonium form rather than the common sodium form to avoid

matrix effects and clogging of the DIN. Ammonium heptanesulfonate yielded a better separation than ammonium dodecylsulfate. The pH was maintained at 3.1 to ensure that the alkyltins were present as ions in the mobile phase. A higher percentage of organic modifier (25% methanol) was used in this separation than for the As separations. All four alkyltins were separated, although previous work (39) reported difficulties in resolving MMT and DMT. MMT-TCl gives a broad and split peak, presumably because of an equilibrium among different forms of monomethyltin compound (5). For example, the ion pair species for MMT could be a mixture of $\text{CH}_3\text{SnCl}_2\text{R}$, $\text{CH}_3\text{SnClR}_2$ and CH_3SnR_3 , where R = heptanesulfonate. Peak-area measurements indicated that the sensitivity was similar for the various alkyltins, including MMT-TCl (Table 6). Absolute detection limits are approximately 10 pg for all forms of alkyltin (Table 6). These absolute detection limits are superior by factors of 5 - 150 over those obtained previously by LC-ICP-MS (4,5,39-41), but are inferior to recent Sn values obtained by supercritical fluid chromatography with ICP-MS detection (42). The relative detection limits of 16 - 20 ppb of Sn in Table 6 are poorer than the values of 0.2 - 0.4 ppb Sn in butyltin compounds reported by McLaren et al. (41). The elevated baseline in Fig. 7, which has also been observed previously (4,39,41), is probably due to Sn^+ although the precise source of this Sn contamination is uncertain.

CONCLUSION

The analytical merits of coupling micro-scale HPLC with ICP-MS with the 30 μm DIN for elemental speciation are demonstrated. Detection limits for As and Sn are improved by 1 - 2 orders of magnitude. The low dead volume ($< 1 \mu\text{l}$) associated with the DIN resulted in low extracolumn broadening, a vital factor for good resolution in micro-scale LC. In addition, the DIN allows the direct injection of HPLC eluents containing large amounts of organic modifier (up to 85% MeOH) into the ICP, which remains stable over a long period of time. With the DIN, the analyte signal varies only by $\pm 20\%$ as the eluent is changed from 20 to 80% MeOH in H_2O . Thus, mild solvent gradients within this range should not induce extreme changes in analyte sensitivity with the DIN. As the toxicological and biological importance of trace elements depends mainly on their chemical forms, more research efforts in developing methods to identify and quantify chemical species are expected. The excellent detection limits, low dead volume and potential for use with gradient elution make the DIN a significant improvement in LC-ICP-MS for elemental speciation.

LITERATURE CITED

- (1) Thompson, J. J., and Houk, R. S., *Anal. Chem.*, 1986, **58**, 2541.
- (2) Dean, J. R., Munro, S., Ebdon, L., Crews, H. M., and Massey, R. C., *J. Anal. At. Spectrom.*, 1987, **2**, 607.
- (3) Jiang, S. J., and Houk, R. S., *Spectrochim. Acta, Part B*, 1988, **43**, 405.
- (4) Suyani, H., Creed, J., Davidson, T., and Caruso, J., *J. Chromatogr. Sci.*, 1989, **27**, 139.
- (5) Suyani, H., Heitkemper, D., Creed, J., Caruso, J., *J. Appl. Spectrosc.*, 1989, **43**, 962.
- (6) Shibata, Y., and Morita, M., *Anal. Sci.*, 1989, **5**, 107.
- (7) Beauchemin, D., Siu, K. W. M., McLaren, J. W., and Berman, S. S., *J. Anal. At. Spectrom.*, 1989, **4**, 285.
- (8) Beauchemin, D., Bednas, M. E., Berman, S. S., McLaren, J. W., Siu, K. W. M., and Sturgeon, R. E., *Anal. Chem.*, 1988, **60**, 2209.
- (9) Shibata, Y., and Morita, M., *Anal. Chem.*, 1989, **61**, 2118.
- (10) Heitkemper, D., Creed, J., Caruso, J., and Fricke, F. L., *J. Anal. At. Spectrom.*, 1989, **4**, 279.
- (11) Sheppard, B. S., Shen, W.-L., Caruso, J. A., Heitkemper, D. T., and Fricke, F. L., *J. Anal. At. Spectrom.*, 1990, **5**, 431.

- (12) Mason, A. Z., Storms, S. D., and Jenkins, K. D., *Anal. Biochem.*, 1990, **186**, 187.
- (13) Crews, H. M., Dean, J. R., Ebdon, L., and Massey, R. C., *Analyst*, 1989, **114**, 895.
- (14) Gercken, B., and Barnes, R. M., *Anal. Chem.*, 1991, **63**, 283.
- (15) Matz, S. G., Elder, R. C., and Tepperman, K., *J. Anal. At. Spectrom.*, 1989, **4**, 767.
- (16) Lawrence, K. E., Rice, G. W., and Fassel, V. A., *Anal. Chem.*, 1984, **56**, 289.
- (17) LaFreniere, K. E., Fassel, V. A., and Eckels, D. E., *Anal. Chem.*, 1987, **59**, 879.
- (18) Wiederin, D. R., Smith, F. G., and Houk, R. S., *Anal. Chem.*, 1991, **63**, 219.
- (19) Houk, R. S., Shum, S. C. K., and Wiederin, D. R., *Anal. Chim. Acta*, 1991, **250**, 61.
- (20) Olivares, J. A., and Houk, R. S., *Anal. Chem.*, 1986, **58**, 20.
- (21) Ross, B.S., and Hieftje, G. M., *Spectrochim. Acta, Part B*, 1991, **46**, 1263.
- (22) Tan, S. H., and Horlick, G., *J. Anal. At. Spectrom.*, 1987, **2**, 745.
- (23) Beauchemin, D., McLaren, J. W., and Berman, S. S., *Spectrochim. Acta, Part B*, 1987, **42**, 467.
- (24) Gregoire, D. C., *Spectrochim. Acta, Part B*, 1987, **42**, 895.

- (25) Vandecasteele, C., Nagels, M., Vanhoe, H., and Dams, R., *Anal. Chim. Acta*, 1988, **211**, 91.
- (26) Crain, J. S., Houk, R. S., and Smith, F. G., *Spectrochim. Acta, Part B*, 1988, **43**, 1355.
- (27) Savitzky, A., and Golay, M. J. E., *Anal. Chem.*, 1964, **36**, 1627.
- (28) PMS 2000, Yokogawa Electric Co., or TS Sola, Turner Scientific.
- (29) Hutton, R. C., *J. Anal. At. Spectrom.*, 1986, **1**, 259.
- (30) Hausler, D., *Spectrochim. Acta, Part B*, 1987, **42**, 63.
- (31) Long, S. E., and Martin, T. D., *ICP Information Newsletter*, 1991, **16**, 460.
- (32) Wiederin, D. R., Houk, R. S., Winge, R. K., and D'Silva, A. P., *Anal. Chem.*, 1990, **62**, 1155.
- (33) Boorn, A. W., and Browner, R. F., *Anal. Chem.*, 1982, **54**, 1402.
- (34) Maessen, F. J. M., Seeverens, P. J. H., and Kreuning, G., *Spectrochim. Acta, Part B*, 1984, **39**, 1171.
- (35) Blades, M. W., and Caughlin, B. L., *Spectrochim. Acta, Part B*, 1985, **40**, 579.
- (36) Trussel, F. C., *Anal. Chem.*, 1985, **57**, 191R.
- (37) Kreuning, G., and Massen, F. J. M. J., *Spectrochim. Acta, Part B*, 1989, **44**, 367.
- (38) Browner, R. F., and Boorn, A. W., *Anal. Chem.* 1984, **56**, 786A.

- (39) Houk, R. S., and Jiang, S. J. in *Trace Metal Analysis and Speciation*, ed. Krull, I. S., Elsevier, Amsterdam, 1991, ch. 5.
- (40) Branch, S., Ebdon, L., Hill, S., and O'Neill, P., *Anal. Proc.*, 1989, **26**, 401.
- (41) McLaren, J. W., Siu, K. W. M., Lam, J. W., Willie, S. N., Maxwell, P. S., Palepu, A., Koether, M., and Berman, S. S., *Fresenius J. Anal. Chem.*, 1990, **337**, 721.
- (42) Shen, W.-L., Vela, N. P., Sheppard, B. S., and Caruso, J. A., *Anal. Chem.*, 1991, **63**, 1491.

Table 1. Instrument conditions and operating procedures

ICP-MS	Sciex Elan Model 250
ICP torch	Modified Sciex short torch : injector tube orifice diameter = 1 mm; 6 mm o.d. x 4 mm i.d. quartz T attached at torch base
Argon flow rates/l min ⁻¹ -	
Plasma	12*
Auxiliary	1*
Make-up	0.30* Regulated by mass flow controller
Nebulizer gas	0.1*
Sample flow rate	30 µl min ⁻¹ typical
Forward power	1.3 kW
Sampling position	20 mm above load coil, on centre*
Sampler	Copper, 1.0 mm diameter orifice
Skimmer	Nickel, 0.9 mm diameter orifice
Detector voltage	-4000 V
Ion lens setting -	
Bessel	-19.80 V
Plate	-11.00 V
Barrel	+5.42 V
Photon stop	-4.16 V
Operating pressures -	
Interface	1 Torr
Quadrupole chamber	3 x 10 ⁻⁵ Torr

* Typical values cited. These parameters were adjusted daily to optimize ion signal (see text).

Table 2. DIN modifications

Support tube*	Stainless steel (Alltech) 1.6 mm o.d. x 0.76 mm i.d. x 150 mm long
Nebulizer tip	Ceramic tube (Scientific Instrument Services, Inc.) 0.51 mm o.d. x 200 μ m i.d. x 15 mm long
Capillary	Fused silica (Polymicro Technologies Inc.) 150 μ m o.d. x 30 μ m i.d. x 450 mm long

* See Fig. 1 of ref. 18 for diagram identifying support tube, nebulizer tip and capillary.

Table 3. Chromatographic conditions

	<u>Arsenic Species</u>	<u>Alkyltin Species</u>
Column	Scientific Glass Engineering Metal free GLT microcolumn 1 mm i.d. x 100 mm long	Same
Stationary phase	Inertsil ODS-2	Same
Mobile phase	5% methanol in water 5 mmol dm ⁻³ heptyltriethylammonium phosphate pH = 6.0	25% methanol in water 5 mmol dm ⁻³ ammonium heptanesulfonate pH = 3.1
Sample flow rate	30 μ l min ⁻¹	Same
Injection volume	0.5 μ l	Same
Compounds	Arsenite (As ^{III}) Arsenate (As ^V) Monomethylarsonic acid (MMAA) Dimethylarsinic acid (DMAA)	Monomethyltin trichloride (MMT-TCI) Dimethyltin dichloride (DMT-DCI) Diethyltin dichloride (DET-DCI) Trimethyltin chloride (TMT-Cl)
Isotopes monitored	m/z = 75	m/z = 120

Table 4. Data acquisition

Data acquisition - Ion signal versus flow rate and ion signal versus eluent composition	Multi-element monitoring mode, low resolution setting; three measurements per peak spaced $m/z = 0.1$ about peak top; dwell time at each position 20 ms, with total measurement time of 0.3 s																		
	<table> <tr> <th>Elements</th><th>m/z</th></tr> <tr> <td>Be</td><td>9</td></tr> <tr> <td>Ga</td><td>69</td></tr> <tr> <td>As</td><td>75</td></tr> <tr> <td>Se</td><td>78</td></tr> <tr> <td>Sn</td><td>120</td></tr> <tr> <td>Nd</td><td>144</td></tr> <tr> <td>Hg</td><td>202</td></tr> <tr> <td>Tl</td><td>205</td></tr> </table>	Elements	m/z	Be	9	Ga	69	As	75	Se	78	Sn	120	Nd	144	Hg	202	Tl	205
Elements	m/z																		
Be	9																		
Ga	69																		
As	75																		
Se	78																		
Sn	120																		
Nd	144																		
Hg	202																		
Tl	205																		
Arsenic and tin separations	Multi-element monitoring mode, low resolution setting; one measurement per peak; dwell time 20 ms, with total measure- ment time of 1 s																		

Table 5. Analytical figures of merit for separation of arsenic species (Fig. 5)

Compounds	As ^{III}	DMAA	MMAA	As ^V
Retention time/min	2.0	5.0	7.6	29.7
Sensitivity/counts ng ⁻¹ (x 10 ⁵)	2.4	2.2	2.6	2.3
RSD* (%)	3.5	3.8	3.0	3.2
Detection limits [†] -				
(a) Amount of As/pg	0.6	0.6	0.5	0.6
(b) Amount of As/μg l ⁻¹	1.2	1.2	1.0	1.2

* Relative standard deviation of peak area for 5 replicate injections of 0.5 ng (as As) of each species. See Table 3 for HPLC conditions.

† Detection limit defined as amount of As required to yield a net peak that was 3x the standard deviation of the background. Calibration curves and peak areas were used in these calculations.

Table 6. Analytical figures of merit for alkyltin separation (Fig. 7)

Compounds	MMT-TCI	DMT-DCI	DET-DCI	TMT-CI
Retention time/min	4.6	7.4	10.5	14.1
Sensitivity/counts ng ⁻¹ of Sn (x 10 ⁴)	4.7	5.5	4.8	5.0
Detection limits* -				
(a) Amount of Sn/pg	10	8	10	9
(b) amount of Sn/ μ g l ⁻¹	20	16	20	18

* Detection limit defined as amount of Sn required to yield a net peak that was 3x the standard deviation of the background. Peak areas were used in these calculations.

39

Fig. 1. Schematic diagram of the separation system: (A) HPLC pump; (B) pre-column; (C) solvent switching valve; (D) inlet filter; (E) micro-injector; (F) analytical column; (G) switching valve; and (H₁) and (H₂) polysil tubes

Fig. 2. Effect of liquid flow rate on ion signal for Be (▼), As (●), Se (▽), Sn (▲), Nd (■), and Hg (□)

Fig. 3. Effect of solvent composition on ion signal for Ga ($\square$), Nd ($\blacktriangledown$), and Tl ($\circ$)

Fig. 4. Effect of solvent composition on ion signal for As (Δ), Se ($\square$), Sn ($\blacktriangledown$), and Hg ($\circ$)

Fig. 5. Separation of four arsenic compounds, 0.5 ng of As for each species. HPLC conditions as in Table 3. Background, $\approx 200 \text{ counts s}^{-1}$

Fig. 6. Separation of four arsenic compounds with step gradient, 0.8 ng of As for each species. Eluent composition was 5 mmol dm⁻³ heptyltriethylammonium phosphate (Q7) in MeOH-H₂O (5 + 95) for the first 11 min, and then stepped to 5 mmol dm⁻³ Q7 in MeOH-H₂O (25 + 75). The pH was adjusted to 7 for both eluents. Background for first 11 min, ≈ 200 counts s⁻¹, and then increased to ≈ 400 counts s⁻¹ after 13 min

Fig. 7. Separation of four alkyltin compounds, 6 ng of Sn for each species. HPLC conditions as in Table 3. Background, ≈ 400 counts s⁻¹

PAPER II

**SPECIATION OF MERCURY AND LEAD COMPOUNDS BY MICROBORE
COLUMN LIQUID CHROMATOGRAPHY - INDUCTIVELY COUPLED PLASMA
MASS SPECTROMETRY WITH DIRECT INJECTION NEBULIZATION**

INTRODUCTION

The toxicological and biological roles of trace elements depend on their chemical forms and/or oxidation states (1-4). Thus, analytical methodology for measuring trace element speciation is necessary. The atomic spectroscopic methods that are usually used for elemental analysis generally do not distinguish the various species present for each element. However, speciation information can be obtained by coupling chromatographic separations with element-selective detection by atomic spectroscopy (5-10). Inductively coupled plasma mass spectrometry (ICPMS) is one of the most attractive detection systems for elemental speciation because of its excellent detection limits, its multielement capability, and its ability to measure isotope ratios (11-28).

In most experiments with plasma detection for liquid chromatography (LC) separations, the nebulizer is one of the weakest links in the whole process because it causes band broadening and loss of much of the sample (29). Several recent papers (30,31) describe the basic properties and analytical performance of a new version of the direct injection nebulizer (DIN), originally developed by Fassel and co-workers (32,33). The DIN is a microconcentric pneumatic nebulizer placed inside the ICP torch. The DIN has a low dead volume ($< 2 \mu\text{L}$) and produces a mist of fine droplets (1 - 10- μm diameter) (34). Rinse-out times for memory-prone elements such as Hg, I, and B are also reduced greatly compared to those obtained with

conventional nebulizers (30). When used for LC-ICPMS of As, Se, and Sn species, the DIN provides absolute detection limits that are superior by 1 - 2 orders of magnitude to those obtained with conventional nebulizers. In addition, the low dead volume and absence of a spray chamber minimize postcolumn band broadening and facilitate the use of microscale LC columns and liquid flow rates ($30 - 100 \mu\text{L min}^{-1}$) that are low enough for all the column effluent to be introduced into the plasma (24,31).

The present work reports the capabilities of LC-DIN-ICPMS for measuring charged species of Hg and Pb. The species chosen (Hg^{+2} , MeHg^+ , EtHg^+ , PhHg^+ , Pb^{+2} , $(\text{Me})_3\text{Pb}^+$, and $(\text{Et})_3\text{Pb}^+$), where Me = methyl, Et = ethyl, and Ph = phenyl, are known or potential toxins and are of substantial interest in medical and environmental sciences (1,35-39). Optimum chromatographic conditions for separating these species as ion pairs by reversed-phase LC are discussed. Finally, application of this methodology for the speciation of Hg and Pb directly in human urine with very little sample preparation is described.

EXPERIMENTAL SECTION

ICPMS and Direct Injection Nebulizer (DIN)

The inductively coupled plasma mass spectrometer used was the Elan Model 250 (Perkin-Elmer SCIEX, Thornhill, ON, Canada). The instrumental system and conditions are described in Table I. The plasma and sampling conditions indicated with an asterisk were optimized daily to maximize the signal from the analyte of interest.

The DIN used in this work was similar to one described previously (Figure 1 of ref. 30). In order to operate at a liquid flow rate of $100 \mu\text{L min}^{-1}$ while maintaining low backpressure ($< 750 \text{ psi}$), a $50\text{-}\mu\text{m-i.d.} \times 45\text{-cm-long}$ fused silica capillary was used. In addition, the $50\text{-}\mu\text{m}$ DIN plugged less easily than the $30\text{-}\mu\text{m}$ one used for the previous LC work (24). Tolerance to plugging was necessary for the analysis of urine, since the salt matrix was not removed beforehand. The width of the annular gap between the inner capillary and the nebulizer tip was $\approx 25 \mu\text{m}$, the same as in the previous work.

Instrumentation, Columns, and Mobile Phases

A digital LC pump (SSI Model 222D) with a bioclean microflow pump head (Scientific Systems, Inc., State College, PA) was employed. Samples were injected using a Rheodyne 7410 microinjector with a $2\text{-}\mu\text{L}$ internal sample loop disk. The

analytical column used in most of this work was a 5-cm-long, 1.6-mm-i.d. PEEK column packed with reversed-phase C_{18} material (CETAC Technologies, Omaha, NE). A similar but longer column (15 cm) was used for Hg speciation in urine. A precolumn filled with Adsorbosil silica (Alltech Association, Inc., Deerfield, IL) was placed between the pump and the analytical column to saturate the mobile phase with silica and to extend the lifetime of the analytical column. The outlet of the analytical column was connected through a switching valve (Rheodyne 9010) to the DIN with a narrow bore polysil tube (50- μ m-i.d. x 5-cm-long, Scientific Glass Engineering, Inc., Austin, TX). The narrowbore connecting capillary minimized extracolumn band broadening.

Standard solutions for optimizing ICPMS conditions were loaded to a 1-mL loop on the switching valve and then injected to the nebulizer. The fused silica capillary of the DIN was connected to the outlet of the switching valve. The outlet of the analytical column was directed to the DIN at all times except during optimization. A solution containing 5 mM ammonium pentanesulfonate in 20:80 v/v acetonitrile (ACN) - water served as the mobile phase. Separations were performed isocratically at a flow rate of 100 μ L min⁻¹.

Data Acquisition

New Elan 500 upgraded hardware and software have been installed on this ICPMS. Data were acquired by peak hopping in the multielement monitoring mode

using 0.5-s measurement time, a 20-ms dwell time, and 1 measurement per peak. The mass spectral resolution was 0.9 amu at 10% peak height. The most abundant isotopes of Hg^+ ($m/z = 202$) and Pb^+ ($m/z = 208$) were monitored. The ICP operating conditions and voltages applied to the ion lenses were optimized to provide maximum signal for inorganic Pb at $100 \mu\text{g L}^{-1}$ and inorganic Hg at $250 \mu\text{g L}^{-1}$ in 20:80 v/v ACN- H_2O .

Chromatograms were recorded in real time and stored on the hard disk of an IBM PS/2 Model 70 computer. These data (as ASCII files) were then processed using a spread sheet program. The raw count rates were first smoothed with a five-point Savitzky - Golay routine (40). Peak area was determined by summing all the count rates under each peak. The background was measured while nebulizing only the mobile phase by summing the total counts in the particular chromatographic region corresponding to each peak. For this work, the detection limit was defined as the amount of the element necessary to give a peak area equal to three times the standard deviation of the background count rate at each analyte mass.

Reagents and Samples

The mobile phases were prepared as follows. Ion-pairing reagents S5 (0.5 M sodium pentanesulfonate), S7 (0.5 M sodium heptanesulfonate) and S12 (0.5 M sodium dodecanesulfonate) were purchased from Regis Chemical Co. (Morton Grove, IL). Sodium salts in the eluent would be expected to cause plugging of the DIN (24),

plugging of the sampling orifice (41-43), and matrix effects (usually suppression) on the signal for Hg^+ and Pb^+ (42,44-48). Therefore, the ion-pairing reagents were converted to their ammonium salts by passing them through a column filled with cation-exchange resin Dowex 50W - X8 in the ammonium form. The nitrogen and hydrogen present in the ammonium cation did not deposit as solids on the sampler and did not cause serious matrix effects. The resulting stock solutions of ammonium salts of S5, S7 and S12 were then diluted with ACN (HPLC grade) and deionized water (DW). Nitric acid (5%) and ammonium hydroxide (1 M) were added to adjust the pH. All mobile phases were filtered through a 0.45- μm -pore nylon filter and degassed under light vacuum from an aspirator for 10 min. At the end of each working day, the LC system was flushed overnight with 75% methanol in water as instructed by the supplier.

Chloride salts of MeHg^+ , EtHg^+ , PhHg^+ , $(\text{Me})_3\text{Pb}^+$, and $(\text{Et})_3\text{Pb}^+$ were obtained from Alfa Products (Danvers, MA) and were used without further purification. Stock solutions of various Hg and Pb compounds at $\approx 300 \text{ mg L}^{-1}$ each were prepared from the above reagent-grade compounds in DW. Stock solutions of inorganic Hg and Pb at 100 mg L^{-1} each were prepared by diluting 1000 mg L^{-1} standards (PLASMACHEM Associates, Inc., Bradley Beach, NJ) in DW. All standards were prepared fresh daily by diluting stock solutions using DW. ACN was HPLC grade from Fisher Scientific, DW (resistance $\approx 18 \text{ M}\Omega$) was obtained from a Barnstead Nanopure-II system (Newton, MA), and ethylenediaminetetraacetic acid

(EDTA) was reagent grade from Fisher Scientific.

Two urine samples were used. First, a freeze-dried urine standard reference material (SRM 2670, National Institute of Standards and Technology, Gaithersburg, MD) was reconstituted as instructed and used without further dilution for Pb speciation in urine. Second, a 24-h human urine specimen was collected from the first author for Hg speciation.

RESULTS AND DISCUSSION

System Dead Volume and System Variance

The general advantages of microbore columns compared to conventional packed columns include high separation efficiencies, better detection limits, decreased solvent consumption and lower costs, compatibility with on-line detectors like MS, and the possibility of using exotic mobile phases or reagents (49,50). However, attention must be paid to the extracolumn volumes in connecting tubes, injector devices, and detectors; otherwise, the efficiency of a separation may be compromised by extracolumn broadening (51,52). This problem is especially important for microbore columns operated at low liquid flow rates.

In chromatography, peak broadening is often expressed in terms of peak variance (53). In the simplest terms, the total variance of a peak is

$$\sigma^2 = \sigma_{\text{col}}^2 + \sigma_{\text{sys}}^2 \quad \text{----- (1)}$$

where σ_{col}^2 represents the contribution of the analytical column and σ_{sys}^2 is the contribution of the rest of the system (injector, connecting tubing, detector, etc.) to the peak variance. Here σ_{sys}^2 is the juxtaposition of all broadening occurring outside the analytical column. For Gaussian peaks, σ can be calculated from the full width at half-maximum ($W_{1/2}$) of the peak:

$$W_{1/2} = 2.35 \sigma \quad \text{----- (2)}$$

In this section, both the system dead volume (μL) and σ_{sys}^2 (s^2) are estimated to

evaluate the contribution of the nebulizer to peak broadening relative to that from the chromatographic column.

The system dead volume was determined by adding known dead volumes and measuring the resulting effect on peak variance. Tubing of different sizes was used to connect the outlet of the analytical column to the switching valve. A 2- μL sample containing MeHg^+ ($500 \mu\text{g Hg L}^{-1}$) was introduced to the 5-cm-long column. The mobile phase was 20:80 v/v ACN- H_2O , and no pairing reagent was used. Under these conditions, MeHg^+ was not retained and eluted at the void time.

A plot of σ^2 vs added volume is shown in Figure 1. Extrapolation to $\sigma^2 = 0$ gave the system dead volume. The system dead volume was found to be 1.6 μL , which was close to the value of 2 μL estimated previously for the 50- μm DIN (30).

Two methods were used to estimate the system variance. First, σ_{sys}^2 was estimated from the y-intercept of the plot in Figure 1 and was found to be 4.1 s^2 . Second, σ_{sys}^2 was estimated using flow injection peaks. In this method, the analytical column was removed from the HPLC system, that is, the injector was connected directly to the switching valve (by a polysil tube) which was connected to the DIN. A 2- μL sample of MeHg^+ ($500 \mu\text{g Hg L}^{-1}$) and AuCl_4^{-1} ($100 \mu\text{g Au L}^{-1}$) was injected to the plasma. The carrier solvent contained S5 at 5 mM in 20:80 v/v ACN- H_2O . The standard deviations of these measurements were determined by five replicate injections.

The peak asymmetry factors measured at 10% peak height (54) were $1.3 \pm$

0.2 for MeHg^+ and 1.2 ± 0.1 for Au. Thus, the flow injection peaks were approximately Gaussian, and the $W_{1/2}$ values for these flow injection peaks could be used to calculate σ_{sys} and σ_{sys}^2 for MeHg^+ and Au respectively. The system variance was found to be $5.0 \pm 0.3 \text{ s}^2$ for MeHg^+ and $4.0 \pm 0.2 \text{ s}^2$ for Au. The chromatographic peak for MeHg^+ shown subsequently (e.g., Figure 2) had a variance of $\sigma^2 \approx 56 \text{ s}^2$. Thus, the estimated value of $\sigma_{\text{sys}}^2 = 4 - 5 \text{ s}^2$ indicated that most of the peak broadening was caused by the column, and less than 5% of the peak broadening was caused by the DIN.

Chromatographic Conditions

When coupling HPLC with ICPMS by the DIN, the separation conditions must be carefully selected so that the analytical performance of the nebulizer and plasma are not compromised. The composition and flow rate of the mobile phase are particularly important (24,30).

The effect of the ACN concentration (v/v) in the mobile phase on the peak shape was studied. Results for MeHg^+ are given in Figure 2. A 2- μL sample containing MeHg^+ , EtHg^+ , and PhHg^+ ($500 \mu\text{g Hg L}^{-1}$ for each species) was injected to the 5-cm-long column. The pairing reagent used was 5 mM S5. Figure 2 shows that the overall peak broadening and retention time of MeHg^+ were sensitive to the percentage of ACN in the mobile phase. With more ACN present, the MeHg^+ peak became sharper and retention time was reduced. In fact, $W_{1/2}$ increased by 150%

from 40% ACN to 10% ACN. Although not shown in Figure 2, the same effect was also found for EtHg^+ and, to a lesser extent, for PhHg^+ .

Figure 3 shows the effect of various ion-pairing reagents on the peak shape of MeHg^+ . Each reagent was present at 5 mM in 20:80 v/v ACN- H_2O . There were no significant differences in peak shapes obtained with S5 and S7. However, S12 broadened the peak greatly, so this reagent was not used further. When the concentration of S5 in 20% ACN (v/v) was varied from 5 to 20 mM, and the pH was varied from 3.0 to 6.5, no significant effect on peak shape was observed.

The effect of the percentage of ACN in the mobile phase on retention time for two trialkyllead and three organomercury compounds was also studied (Figures 4a,b). As expected, the retention decreased as the concentration of the organic modifier increased. In particular, the percentage of ACN exerted a stronger effect on the retention times of the compounds with the large organic groups (i.e., PhHg^+ and $(\text{Et})_3\text{Pb}^+$). In the separation of the organomercury compounds, even though good separations were achieved using 5% ACN, the peaks were broad (Figure 2). On the other hand, sharp peaks were achieved using 40% ACN, when each Hg species was injected individually, but the peaks from the mixture of the three Hg species could not be resolved.

The effect of flow rate and percentage of ACN on ion signal with the DIN has been described elsewhere (24). No green C_2 emission was observed visually from the plasma when the solvent contained less than 50% ACN. The plasma stayed on and

remained stable when 75% ACN was nebulized, but carbon deposition (55,56) on the sampling cone did cause substantial signal drift. In addition, maximum ion signal was attained when the DIN was operated at a liquid flow rate of $\approx 120 \mu\text{L min}^{-1}$ and when a solvent containing 15 - 25% ACN was nebulized. Fortunately, this mobile-phase composition also suppressed tailing of the chromatographic peak, as shown in curve b of Figure 2. Finally, at a liquid flow rate of $120 \mu\text{L min}^{-1}$, the rinse-out time for Hg in 2% HNO_3 was 15 s or less (30).

In summary, 5 mM S5 in 20:80 v/v ACN - water at $100 \mu\text{L min}^{-1}$ was chosen as the optimum mobile phase and liquid flow rate. These conditions represent a compromise between maximum sensitivity, reasonable rinse-out time, reasonable separation time, and resolution. The performance of plasma-based detection methods for LC is generally sensitive to the composition and flow rate of the mobile phase (6,11), and ICPMS with the DIN is no exception.

Speciation of Hg and Pb Compounds in a Test Mixture

An aqueous test mixture containing MeHg^+ , EtHg^+ , PhHg^+ , $(\text{Me})_3\text{Pb}^+$, and $(\text{Et})_3\text{Pb}^+$ was prepared. A separation of these species in a single injection is shown in Figure 5. The analytical figures of merit under these separation conditions are shown in Tables II and III.

The peaks for $(\text{Me})_3\text{Pb}^+$ and $(\text{Et})_3\text{Pb}^+$ were easily resolved, and the separation was completed in less than 3 min (Figure 5). Peak area measurements indicated that

the Pb sensitivity (total counts / pg Pb) was similar (i.e., within 5%) for the two forms (Table II). Precision based on five separate injections of ≈ 40 pg (as Pb) of each species and measurement of peak areas was better than 2% RSD for both species. Calibration curves for each species based on peak areas from injections of 0.04, 0.1, 0.4, and 1 ng (as Pb) were all linear with correlation coefficients of 0.9998 for $(\text{Me})_3\text{Pb}^+$ and 0.9993 for $(\text{Et})_3\text{Pb}^+$. Absolute detection limits were ≈ 0.2 pg Pb for both species (Table II). These absolute detection limits were superior by 1 - 4 orders of magnitude over those obtained previously by electrochemical detection (57), UV-vis absorption (58), ICPAES (25,59), and LC-ICPMS (25). Relative detection limits for Pb species were $\approx 0.1 \mu\text{g L}^{-1}$.

In the separation of organomercury species, previous work (27,60) reported a high background for Hg, presumably due to bleeding of species containing Hg from the column or to memory effects in the spray chamber or desolvation system of the nebulizer. For the present work, the Hg background was ≈ 100 counts/s, which was comparable to the background usually seen at $m/z \approx 200$ when the DIN is used with this particular ICPMS device (30). The PEEK column and tubing used and the ability of the DIN to operate without a spray chamber probably helped keep the Hg background at a reasonable level. However, the Hg peaks had longer tails than the Pb peaks, presumably due to some memory for Hg in either the column or the DIN. Despite this tailing, the peaks for MeHg^+ , EtHg^+ , and PhHg^+ were resolved adequately (Figure 5).

Peak area measurements indicated that sensitivity (total counts / pg Hg) was similar for all three forms of Hg (Table III). Precision based on five separate injections of ≈ 2 ng (as Hg) of each species and measurement of peak areas was $\approx 3\%$ RSD for all three species. Calibration curves for each species based on injections of 0.2, 0.5, 2, 5, and 10 ng (as Hg) and peak areas were all linear with correlation coefficients 0.9995 or better. Absolute detection limits were ≈ 7 pg of Hg for all three forms (Table III). These absolute detection limits were superior by factors of 5 - 1000 over those obtained previously by LC-cold vapor ICPMS (27), LC-ICPMS (27), and LC-ICPAES (61). The improvement in absolute detection limits is expected because all of the sample reaches the plasma with the DIN. Relative detection limits for all three forms were $\approx 4 \mu\text{g L}^{-1}$, which are comparable to those obtained with conventional nebulizers.

Pb Speciation in Freeze-Dried Urine

The proposed method was tested for measurement of Pb and Hg species in human urine. The NIST SRM 2670 freeze-dried urine (normal level) was used for experiments in Pb speciation. Inorganic Pb^{+2} was retained permanently on the column; presumably, the reversed-phase column was not completely end-capped, a problem noted previously (11). The Pb^{+2} was removed by flushing the column with EDTA, which yielded a substantial Pb signal that was well above the Pb background of ≈ 200 counts/s in EDTA. In order to elute the Pb^{+2} , excess EDTA was added to

the urine sample before injection to ensure a metal to ligand molar ratio of 1:50 or larger. Background was determined by injecting a blank containing the same amount of EDTA as in the urine sample.

Chromatograms for the separation of inorganic Pb and two trialkyllead species spiked into urine are given in Figure 6. Injection 1 shows the peak for Pb^{+2} in NIST SRM 2670 freeze-dried urine (normal level). Only the inorganic peak was found. Using a standard addition method, the sample was found to contain $10.3 \mu\text{g L}^{-1}$ of Pb, which is close to the suggested value ($10 \mu\text{g L}^{-1}$) from NIST.

Since the NIST SRM did not contain measurable levels of $(\text{Me})_3\text{Pb}^+$ and $(\text{Et})_3\text{Pb}^+$, these compounds were spiked into the NIST urine to test the suitability of the method for this separation in the presence of a difficult matrix. Injection 2 of Figure 6 shows the separation of all three Pb species. The retention times for the two trialkyllead species remained the same as in their separation from a simple aqueous test mixture (Table II). Peak area measurements showed that sensitivity was similar for all three forms and was comparable (within 10%) to those obtained for Pb species in the test mixture. The reproducibility of the peak areas was better than 3% RSD ($n = 3$). Absolute detection limits were 0.2 pg for all three Pb compounds in urine.

In injection 2, the peak height for Pb^{+2} was the same as that in injection 1. Also, reasonable chromatographic peaks were seen for $(\text{Me})_3\text{Pb}^+$ and $(\text{Et})_3\text{Pb}^+$. These observations showed that the use of EDTA to prevent permanent retention of Pb^{+2} did not disturb the relative amounts of the alkyllead ions or perturb their chromatographic

behavior. Undesired interconversions between different chemical forms of an element or instability of particular species are common problems in speciation experiments. The simple procedure for preparing the sample (i.e., mere dilution) also may help prevent problems from interconversion or instability of the species present (62).

Hg Speciation in 24-h Human Urine

A 24-h (63) urine specimen (≈ 2 L) from the first author was analyzed for Hg species. A chromatogram from one injection of the urine specimen is given in Figure 7. Only the inorganic mercury peak was found. Using a standard addition method, the specimen was found to contain $25 \mu\text{g L}^{-1}$ of inorganic mercury. The total mercury content of this urine sample was determined to be $28 \mu\text{g L}^{-1}$ by flow injection and standard addition, i.e., no chromatographic column was employed. None of the organomercury ions were present at sufficient concentrations to be observed directly in urine, as was the case for the organolead ions described in the preceding section. A preliminary preconcentration procedure would be required to see these species (64).

The chromatographic procedure described in the present work was evaluated further by the following spike experiment. Another urine aliquot was diluted by a factor of 10 with deionized water and spiked with 0.75 ng of Hg^{+2} and 2 ng (as Hg) of each of the alkylmercury species. The dilution minimized the matrix effect of Na on Hg signal. A longer column (15 cm) was also required to resolve the Hg^{+2} peak from that for MeHg^{+} .

Figure 8 shows the separation of these three Hg compounds in urine. Since a longer column was used, the retention times for MeHg^+ and EtHg^+ were almost twice as long as those reported for Hg speciation in the aqueous test mixture. Peak area measurements indicated that the sensitivity for Hg^{+2} was similar to that obtained for the Hg species in the simple aqueous solution. However, the sensitivities for MeHg^+ and EtHg^+ were reduced by a factor of 2 in the urine matrix, presumably due to matrix suppression from Na (≈ 1000 ppm Na). With the anionic pairing reagent in the eluent, Na^+ was weakly retained and eluted at ≈ 2 min after the sample was injected, i.e., after the Hg^{+2} peak but before the subsequent peaks for MeHg^+ and EtHg^+ . Thus, sensitivity for Hg^{+2} was not affected by the urine matrix, because the Na from the urine had not yet reached the plasma when Hg^{+2} eluted. In contrast, because of the tail on the Na^+ chromatographic peak, some Na was present in the plasma when MeHg^+ and EtHg^+ eluted. Hence, the Hg^+ signal from these species was diminished by a matrix effect caused by Na. For this separation, the precision was still 5% RSD or better ($n = 3$). Absolute detection limits were 7 - 18 pg of Hg for the three Hg compounds in urine.

As a final note, the reversed-phase ion-pairing separation reported here proved remarkably robust and resilient even when a difficult matrix (urine) was analyzed. Chromatographic retention times were not affected by repeated injections of the urine matrix. Chromatographic resolution (53) was only slightly poorer in the presence of the urine matrix than that obtained for the test solutions. For example, the resolution

between $(\text{Me})_3\text{Pb}^+$ and $(\text{Et})_3\text{Pb}^+$ peaks was 3.7 in Figure 5 (for the test solution) compared to 3.0 in Figure 6 (for the spiked urine matrix).

LITERATURE CITED

1. Goyer, R. A. In *Casarett and Doull's Toxicology: The Basic Science of Poisons*, 4th ed.; Casarett, L. J.; Amdur, M. O.; Doull, J.; Klaassen, C. D., Eds.; Pergamon Press: New York, 1991; Chapter 19.
2. Cappon, C. J. *LC/GC* **1988**, *6*, 584-599.
3. Batley, G. E.; Low, G. K.-C. In *Trace Element Speciation Analytical Methods and Problems*; Batley, G. E., Ed.; CRC Press, Inc.: Boca Raton, Florida, 1989; Chapter 9.
4. Gardiner, P. E. *J. Anal. At. Spectrom.* **1988**, *3*, 163-168.
5. Uden P. C., Ed. *Element - Specific Chromatographic Detection by Atomic Emission Spectroscopy*; American Chemical Society: Washington, D.C., 1992.
6. Krull, I. S., Ed. *Trace Metal Analysis and Speciation*; Elsevier: New York, 1991.
7. Keliher, P. N.; Ibrahim, H.; Gerth, D. J. *Anal. Chem.* **1990**, *62*, 184R-212R.
8. Harrison, R. M.; Rapsomanikis, S., Eds. *Environmental Analysis Using Chromatography Interfaced with Atomic Spectroscopy*; Ellis Horwood Ltd.: Chichester, U. K., 1989.
9. Keliher, P. N.; Gerth, D. J.; Snyder, J. L.; Wang, H.; Zhu, S. F. *Anal. Chem.* **1988**, *60*, 342R-368R.

10. Ebdon, L.; Hill, S.; Ward, R. W. *Analyst* **1987**, *112*, 1-16.
11. Thompson, J. J.; Houk, R. S. *Anal. Chem.* **1986**, *58*, 2541-2548.
12. Jiang, S. J.; Houk, R. S. *Spectrochim. Acta* **1988**, *43B*, 405-411.
13. Dean, J. R.; Munro, S.; Ebdon, L.; Crews, H. M.; Massey, R. C. *J. Anal. At. Spectrom.* **1987**, *2*, 607-610.
14. Matz, S. G.; Elder, R. C.; Tepperman, K. *J. Anal. At. Spectrom.* **1989**, *4*, 767-771.
15. Crews, H. M.; Dean, J. R.; Ebdon, L.; Massey, R. C. *Analyst* **1989**, *114*, 895-899.
16. Beauchemin, D.; Bednas, M. E.; Berman, S. S.; McLaren, J. W.; Siu, K. W. M.; Sturgeon, R. E. *Anal. Chem.* **1988**, *60*, 2209-2212.
17. Beauchemin, D.; Siu, K. W. M.; McLaren, J. W.; Berman, S. S. *J. Anal. At. Spectrom.* **1989**, *4*, 285-289.
18. Heitkemper, D.; Creed, J.; Caruso, J.; Fricke, F. L. *J. Anal. At. Spectrom.* **1989**, *4*, 279-284.
19. Sheppard, B. S.; Shen, W.-L.; Caruso, J. A.; Heitkemper, D. T.; Fricke, F. L. *J. Anal. At. Spectrom.* **1990**, *5*, 431-435.
20. Shibata, Y.; Morita, M. *Anal. Sci.* **1989**, *5*, 107-109.
21. Shibata, Y.; Morita, M. *Anal. Chem.* **1989**, *61*, 2116-2118.
22. Suyani, H.; Creed, J.; Davidson, T.; Caruso, J. J. *Chromatogr. Sci.* **1989**, *27*, 139-143.

23. Suyani, H.; Heitkemper, D.; Creed, J.; Caruso, J. *Appl. Spectrosc.* **1989**, *43*, 962-967.
24. Shum, S. C. K.; Neddersen, R.; Houk, R. S. *Analyst* **1992**, *117*, 571-575.
25. Al-Rashdan, A.; Heitkemper, D.; Caruso, J. *J. Chromatogr. Sci.* **1991**, *29*, 98-102.
26. Gercken, B.; Barnes, R. M. *Anal. Chem.* **1991**, *63*, 283-287.
27. Bushee, D. S. *Analyst* **1988**, *113*, 1167-1170.
28. Bushee, D. S.; Moody, J. R.; May, J. C. *J. Anal. At. Spectrom.* **1989**, *4*, 773-775.
29. Browner, R. F.; Boorn, A. W. *Anal. Chem.* **1984**, *56*, 786A-798A.
30. Wiederin, D. R.; Smith, F. G.; Houk, R. S. *Anal. Chem.* **1991**, *63*, 219-225.
31. Houk, R. S.; Shum, S. C. K.; Wiederin, D. R. *Anal. Chim. Acta* **1991**, *250*, 61-70.
32. Lawrence, K. E.; Rice, G. W.; Fassel, V. A. *Anal. Chem.* **1984**, *56*, 289-292.
33. LaFreniere, K. E.; Fassel, V. A.; Eckels, D. E. *Anal. Chem.* **1987**, *59*, 879-887.
34. Wiederin, D. R.; Houk, R. S. *Appl. Spectrosc.* **1991**, *45*, 1408-1412.
35. Cox, C.; Clarkson, T. W.; Marsh, D. O.; Amin-Zaki, L.; Tikriti, S.; Myers, G. G. *Environ. Res.* **1989**, *49*, 318-332.

36. Friberg, L.; Vostal, J. *Mercury in the Environment*; CRC Press: Cleveland, 1972.
37. Grandjean, P. *Biological Effects of Organolead Compounds*; CRC Press: Boca Raton, FL, 1984.
38. Mahaffey, K. R., Ed. *Dietary and Environmental Lead: Human Health Effects*; Elsevier Scientific: New York, N. Y., 1985.
39. Beockx, R. L. *Anal. Chem.* **1986**, *58*, 274A-288A.
40. Savitzky, A.; Golay, M. J. E. *Anal. Chem.* **1964**, *36*, 1627-1639.
41. Olivares, J. A.; Houk, R. S. *Anal. Chem.* **1986**, *58*, 20-25.
42. Jiang, S. -J.; Houk, R. S. *Anal. Chem.* **1986**, *58*, 1739-1743.
43. Douglas, D. J.; Kerr, L. A. *J. Anal. At. Spectrom.* **1988**, *3*, 749-752.
44. Tan, S. H.; Horlick, G. J. *Anal. At. Spectrom.* **1987**, *2*, 745-763.
45. Beauchemin, D.; McLaren, J. W.; Berman, S. S. *Spectrochim. Acta* **1987**, *42B*, 467-490.
46. Gregoire, D. C. *Spectrochim. Acta* **1987**, *42B*, 895-907.
47. Vandecasteele, C.; Nagels, M.; Vanhoe, H.; Dams, R. *Anal. Chim. Acta* **1988**, *211*, 91-98.
48. Crain, J. S.; Houk, R. S.; Smith, F. G. *Spectrochim. Acta* **1988**, *43B*, 1355-1364.
49. Novotny, M. *LC Mag.* **1985**, *3*, 876-886.

50. Guiochon, G.; Colin, H. In *Microcolumn High-Performance Liquid Chromatography*; Kucera, P., Ed.; Elsevier: New York, 1984; Chapter 1.
51. Saito, M.; Hibi, K.; Ishii, D.; Takeuchi, T. In *Introduction to Microscale High-Performance Liquid Chromatography*; Ishii, D., Ed.; VCH Publishers: New York, N. Y., 1988; Chapter 2.
52. Gluckman, J. C.; Novotny, M. In *Microcolumn Separations: Columns, Instrumentation, and Ancillary Techniques*; Novotny, M. V.; Ishii, D., Eds.; Elsevier: New York, 1985; pp. 57-72.
53. Fritz, J. S.; Schenk, G. H. *Quantitative Analytical Chemistry*; Allyn and Bacon, Inc.: Newton, MA, 1987; Chapter 21.
54. Kirkland, J. J.; Yau, W. W.; Stoklosa, H. J.; Dilks, C. H. *J. Chromatogr. Sci.* **1977**, *15*, 303-316.
55. Hutton, R. C. *J. Anal. At. Spectrom.* **1986**, *1*, 259-263.
56. Hausler, D. *Spectrochim. Acta* **1987**, *42B*, 63-73.
57. MacCrehan, W. A.; Durst, R. A.; Bellama, J. M. *Anal. Lett.* **1977**, *10*, 1175-1188.
58. Blaszkiewicz, M.; Baumhoer, G.; Neidhart, B. *Fresenius' Z. Anal. Chem.* **1984**, *317*, 221-225.
59. Ibrahim, M.; Nisamanepong, W.; Hass, H. L.; Caruso, J. A. *Spectrochim. Acta* **1985**, *40B*, 367-376.

60. Houk, R. S.; Jiang, S. J. In *Trace Metal Analysis and Speciation*; Krull, I. S., Ed.; Elsevier: New York, 1991; Chapter 5.
61. Krull, I. S.; Bushee, D. S.; Schleicher, R. G.; Smith, S. B., Jr. *Analyst* **1986**, *111*, 345-349.
62. Sheppard, B.; Caruso, J.; Heitkemper, D.; Wolnick, K. *Analyst*, **1992**, *117*, 971-976.
63. Jacobs, D. S.; Kasten, B. L., Jr.; Demott, W. R.; Wolfson, W. L. *Laboratory Test Handbook with DRG Index*; Lexi-Comp Inc.: Ohio, 1984; p.514.
64. Blaszkiewicz, M.; Baumhoer, G.; Neidhart, B. *Fresenius' Z. Anal. Chem.* **1986**, *325*, 129-135.

Table I. Instrument Conditions and Operating Procedures

ICPMS	Sciex Elan Model 250
ICP torch	modified Sciex short torch : injector tube orifice diameter = 1 mm; 6-mm-o.d. x 4-mm-i.d. quartz tee attached at torch base
argon flow rates (L min ⁻¹)	
outer	12*
auxiliary	1*
make-up	0.30* regulated by mass flow controller
nebulizer gas	0.4*
sample flow rate	100 μ L min ⁻¹ typical
forward power	1.4 kW*
sampling position	20 mm from load coil, on center*
sampler	Copper, 1.0-mm-diameter orifice
skimmer	Nickel, 0.9-mm-diameter orifice
detector voltage	-4000 V
ion lens setting	
bessel	-19.80 V
plate	-11.00 V
barrel	+5.42 V
photon stop	-7.46 V*
operating pressures	
interface	1.0 Torr
quadrupole chamber	3 x 10 ⁻⁵ Torr

* Typical values cited. These parameters were adjusted daily to maximize ion signal (see text) and differed slightly from day to day.

Table II. Analytical Figures of Merit for the Separation of Lead Species

	Pb^{+2}	$(\text{Me})_3\text{Pb}^+$	$(\text{Et})_3\text{Pb}^+$
(A) In Aqueous Sample (Figure 5)			
retention time (min)	-	0.9	2.0
sensitivity ^a (counts/pg of Pb)	-	750	730
RSD ^b (%)	-	1.7	2.0
detection limits ^c :			
(pg of Pb)	-	0.2	0.2
($\mu\text{g L}^{-1}$, ppb)	-	0.1	0.1
(B) In Urine (Figure 6)			
retention time (min)	0.6	0.9	2.0
sensitivity ^a (counts/pg of Pb)	850	800	750
RSD ^d (%)	2.8	1.7	2.5
detection limits ^c :			
(pg of Pb)	0.2	0.2	0.2
($\mu\text{g L}^{-1}$, ppb)	0.1	0.1	0.1

^a Sensitivity was calculated based on peak area and amount injected. ^b Relative standard deviation of peak area for five replicate injections of 40 pg (as Pb) of each species. See Figure 5 for LC conditions. ^c Detection limit defined as amount of Pb required to yield a net peak that was 3 times the standard deviation of background. Peak areas were used in these calculations. ^d Relative standard deviation of peak area for three replicate injections of SRM 2670 + 40 pg (as Pb) spike of each of the trialkyllead species. See Figure 6 for HPLC conditions.

Table III. Analytical Figures of Merit for the Separation of Mercury Species

	Hg ⁺²	MeHg ⁺	EtHg ⁺	PhHg ⁺
(A) In Aqueous Sample (Figure 5)				
retention time (min)	-	1.2	1.8	4.4
sensitivity ^a (counts/pg of Hg)	-	30	35	37
RSD ^b (%)	-	2.7	2.8	3.0
detection limits ^c :				
(pg of Hg)	-	7	7	6
(μg L ⁻¹ , ppb)	-	4	4	3
B) In Urine (Figure 8)				
retention time (min)	1.7	2.8	4.7	-
sensitivity ^a (counts/pg of Hg)	30	13	16	-
RSD ^d (%)	3.7	3.8	4.1	-
detection limits ^c :				
(pg of Hg)	7	18	16	-
(μg L ⁻¹ , ppb)	4	9	8	-

^a Sensitivity was calculated based on peak area and amount injected. ^b Relative standard deviation of peak area for five replicate injections of 2 ng (as Hg) of each species. See Figure 5 for LC conditions. ^c Detection limit defined as amount of Hg required to yield a net peak that was 3 times the standard deviation of background. Peak areas were used in these calculations. ^d Relative standard deviation of peak area for three replicate injections of diluted urine spiked with Hg⁺², MeHg⁺, and EtHg⁺. See Figure 7 for HPLC conditions.

Figure 1. Determination of system dead volume and system variance. The line represents a least squares fit. See text for conditions

Figure 2. Effect of ACN percentage on peak shape for MeHg^+ . Each peak is normalized to the same maximum value so that the peak widths can be compared readily. See text for conditions.

Figure 3. Effect of various ion pairing reagents on peak shape for MeHg^+ . Each peak is normalized to the same maximum value. See text for conditions.

Figure 4a. Effect of mobile phase composition on the retention time for $(\text{Me})_3\text{Pb}^+$ and $(\text{Et})_3\text{Pb}^+$. Column, 1.6-mm-i.d. x 5-cm-long; flow rate, $100\ \mu\text{L}\ \text{min}^{-1}$; mobile phase, 5 mM ammonium pentanesulfonate in various percentages of ACN (v/v) in water; injection volume, $2\ \mu\text{L}$; sample size, 40 pg (as Pb) of each species.

Figure 4b. Effect of mobile-phase composition on the retention time of MeHg⁺, EtHg⁺, and PhHg⁺. Conditions as in Figure 4a. Sample size: 2 ng (as Hg) of each species.

Figure 5. Separation of two trialkyllead and three organomercury species. Column, 1.6-mm-i.d. x 5-cm-long; flow rate, $100 \mu\text{L min}^{-1}$; mobile phase, 5 mM ammonium pentanesulfonate in 20:80 v/v ACN- H_2O (pH = 3.4); injection volume, $2 \mu\text{L}$; sample size, 40 pg (as Pb) for $(\text{Me})_3\text{Pb}^+$, 80 pg (as Pb) for $(\text{Et})_3\text{Pb}^+$, and 2 ng (as Hg) for each of the organomercury species.

Figure 6. Separation of Pb species in NIST SRM 2670 freeze-dried urine (normal level). Conditions as in Figure 5. EDTA was added to the sample at 10 mg L^{-1} before injection. Injection 1: NIST SRM 2670 freeze-dried urine (normal level). Injection 2: NIST SRM 2670 freeze-dried urine spiked with 40 pg (as Pb) of each of the trialkyllead species.

Figure 7. Hg^{+2} in a 24-h human urine specimen. Conditions as in Figure 5. Column: 1.6-mm-i.d. x 15-cm-long.

Figure 8. Separation of Hg species in a diluted 24-h human urine specimen. Conditions as in Figure 7. Sample: urine diluted by a factor of 10 with deionized water and spiked with 0.75 ng (as Hg) of Hg⁺², and 2 ng (as Hg) of each of the organomercury species.

PAPER III

SPATIALLY RESOLVED MEASUREMENTS OF SIZE AND
VELOCITY DISTRIBUTIONS OF AEROSOL DROPLETS FROM A
DIRECT INJECTION NEBULIZER

INTRODUCTION

A new version of the direct injection nebulizer (DIN) has been developed recently for sample introduction in ICP spectroscopy, particularly ICP-MS (1-5). The advantages of the DIN include low sample and solvent consumption rates, excellent plasma stability when samples containing high concentrations of organic solvents are nebulized, good absolute detection limits, and low memory effects from difficult elements such as B, I, and Hg (5). In addition, when used as an interface for liquid chromatography/ICP-MS, the DIN minimizes extra-column broadening of the chromatographic peaks (2,6,7).

A DIN is essentially a micro-concentric pneumatic nebulizer and produces an axi-symmetric conical spray. When used in ICP experiments, the DIN is usually positioned inside the torch with its tip only 3 - 4 mm from the base of the plasma. Thus, virtually all sample is sprayed into the axial channel of the plasma. The size and velocity distributions of the DIN spray probably influence the vaporization, atomization, excitation, and ionization processes in the plasma. Hence, characterization of these fundamental properties of the DIN aerosol should prove valuable for optimization and for improvement of analytical performance.

A survey of possible techniques for measuring size and velocity distributions led to the choice of the phase Doppler particle analysis (PDPA) method for this work. Theory and applications of the PDPA have been described by Bachalo et al. (8,9).

Montaser and co-workers have used the PDPA method to characterize aerosols produced by conventional nebulizers (10) and by ultrasonic nebulizers (11). In PDPA, size and velocity distributions of aerosol droplets are measured in situ with the use of dual-beam light scattering interferometry. In addition, spatially resolved information on both size and velocity distributions can be obtained, which could prove very instructive. The aerosol produced by the DIN is generally sent immediately into the ICP. Spatial variations in droplet properties (such as size or velocity distributions) as the droplets traverse the ICP could then influence the conversion of droplets into atoms and ions in ways not seen with conventional nebulizers, where the aerosol is generally mixed up in a spray chamber before it reaches the plasma. The present work addresses these questions. In addition, the effect of solvent composition on droplet size distribution is also discussed.

EXPERIMENTAL SECTION

DIN

The DIN used throughout this work was constructed as described previously (5,6). The sample transfer line was a fused-silica capillary (50 μm i.d. x 150 μm o.d. x 45 cm long). The width of the annular gap between the inner capillary and the nebulizer tip was $\approx 25 \mu\text{m}$. An LC pump (SSI Model 222D, Scientific System, Inc., State College, PA) with a bioclean micro-flow pump head maintained a typical liquid flow rate of 100 $\mu\text{L min}^{-1}$. A Rheodyne 9010 high-pressure injector with a 2-mL PEEK sample loop was used for sample injection. For this work, deionized water or mixtures containing various percentages by volume of methanol in water were nebulized. The argon nebulizer gas flow rate was adjusted by varying the back pressure with a gas regulator.

PDPA Measurements

This instrument was manufactured by Aerometrics, Inc. (Model P/DPA). Figure 1 shows a schematic diagram of the optical system. The laser was split into two equal-intensity beams focused to an intersection point inside the aerosol spray. A receiver consisting of focusing lenses and three precisely spaced photomultiplier tubes (PMTs) was located 30° off the axis of the transmitted beams. Droplets passing the intersection point of the beams scattered light independently from each beam. The

scattered light then interfered to form a fringe pattern in the plane of the receiver lens. To determine the drop velocity, one of the PMTs measured the temporal frequency of the scattered fringe pattern. This temporal frequency was a function of the velocity of the drop, the beam intersection angle, and the light wavelength. For measurement of the drop size, the spatial frequency of the scattered fringe pattern was measured with the use of all three PMTs. The spatial frequency was inversely proportional to the droplet diameter but also depended upon the laser wavelength, the beam intersection angle, the droplet refractive index, and the spacing between different detectors on the receiver. Refractive indexes for DI water and various sample mixtures were determined with a refractometer (Bausch and Lomb). The measured refractive index for water was 1.3315 and that for 50% methanol:water was 1.3399, so the change in refractive index with solvent composition was very slight and exerted little influence on the measured droplet sizes (8).

The DIN was mounted vertically on a translational stage with the spray pointed downward in the z-direction (Fig. 1). Unless specified otherwise, the beam intersection point was on center and 25 mm downstream from the nebulizer tip. In the subsequent discussion, this position is denoted $z = 25$ mm, $r = 0$ where z and r are the axial and radial positions, respectively. Spatial profiles were obtained by physically moving the DIN with the translation stage. A sample size of 10,000 validated aerosol droplets gave reproducible droplet size and velocity distributions. The detector voltage was -450 V for each of the three PMTs, unless indicated

otherwise.

The measured droplet size and velocity distributions were presented as histograms. Each histogram contained fifty discrete class bins covering the specified size or velocity range. No droplets larger than $40\ \mu\text{m}$ were detected, and the smallest size that could be measured by this instrument was $0.5\ \mu\text{m}$. Therefore, the $1 - 40\ \mu\text{m}$ size range was used for all measurements. Furthermore, because no droplets travelling faster than $50\ \text{m s}^{-1}$ were detected, the upper limit of the velocity range was chosen to be $50\ \text{m s}^{-1}$, while the lower limit was between -2 and $5\ \text{m s}^{-1}$, depending on the application. The sign on the measured velocity was considered positive if the droplets were moving axially away from the nebulizer, i.e., downward in Fig. 1.

Reagents and Samples

Deionized water (resistance $\approx 18\ \text{M}\Omega$) was obtained from a Barnstead Nanopure-II system; methanol (MeOH) was HPLC grade from Fisher Scientific. Mixtures of MeOH/H₂O (5 - 50% v/v) were prepared by dilution.

2 of 2

RESULTS AND DISCUSSION

Terminology and Definitions

Various representative diameters for aerosol droplet size measurements are reported in this work. Detailed mathematical definitions of these parameters have been given elsewhere (11,12), and only general descriptions of their physical significance are necessary here. The Sauter mean diameter ($D_{3,2}$) is the diameter of a drop whose volume-to-surface-area ratio is the same as that of the entire spray. Ten percent of the total liquid mass (or volume) is in drops smaller than the 10% mass diameter ($D_{0.1}$). Half of the total liquid mass is in drops smaller than the mass median diameter ($D_{0.5}$). Similarly, 90% of the total liquid mass is comprised of drops smaller than the 90% mass diameter ($D_{0.9}$).

Volume flux (F) is the total volume of the droplets passing a cross-sectional area in a given direction per unit time and is a vector quantity. In droplet velocity measurements, the mean (U) and the root mean square (rms) velocities are reported. The rms, or velocity variance, is the standard deviation of the velocity distribution.

Measurement of Aerosol Droplet Size Distribution by PDPA

Figure 2a shows the aerosol droplet size distribution (at $z = 25$ mm, $r = 0$) when deionized water is being nebulized. Normal DIN operating conditions were employed; i.e., the nebulizer gas and liquid flow rates were 0.4 L min^{-1} and $100 \mu\text{L}$

min^{-1} . The $D_{3,2}$, $D_{0.9}$, $D_{0.5}$, and $D_{0.1}$ values (see Table I) were found to be ≈ 14 , 28, 18, and $7 \mu\text{m}$, respectively.

The droplet size distribution shown in Fig. 2a is substantially larger than that reported by Wiederin and Houk (13), as are the characteristic diameter values. In addition, the droplet size distribution is also distinctly bimodal, whereas that reported in Ref. 13 is not. There are several possible reasons for these discrepancies. First, the DIN used previously had a narrower annular gap ($15 \mu\text{m}$ vs. $25 \mu\text{m}$). Second, different methods were used for measuring droplet sizes in the two studies (14-16). A forward-scattering spectrometer probe was used previously (13), which produces a line-of-sight optical average (17) of the droplet size distribution, in contrast to the spatially resolved size distribution provided by the PDPA. Finally, all laser scattering methods are subject to bias in the resulting droplet size distributions. In general, the smaller droplets are underrepresented relative to the larger ones (18). The size distributions produced by these two droplet measurement methods could be affected somewhat differently by these bias problems.

Using the PDPA method, Montaser and co-workers report droplet sizes for tertiary aerosols from five nebulization systems (10) and for desolvated aerosols from an ultrasonic nebulizer (11). Under normal operating conditions, their droplet diameters are ≈ 3 times smaller than those reported in the present work. As expected, the use of a spray chamber with a conventional nebulizer produces a finer aerosol than the DIN, but at the cost of poorer sample transport efficiency, increased

memory effects for memory-prone elements (Hg, I, and B), and poorer precision (1,3,5).

Effect of Solvent Composition on Aerosol Droplet Size Distribution

A curious effect was noted previously in chromatographic studies with this DIN. Some organic modifier, such as methanol or acetonitrile, is often necessary in the aqueous eluent to optimize the chromatographic separation. In general, the signal for atomic ions increases sharply (i.e., by 3x to 5x) as the fraction of MeOH is raised from zero to 25%. The signals for Ga^+ , Nd^+ , and Tl^+ then remain roughly constant as the MeOH fraction increases further to 80%, whereas the signals for Hg^+ , Sn^+ and Se^+ diminish gradually (7). Other workers using conventional nebulizers have also reported enhancements in atomic ion signals when the usual aqueous samples are spiked with modest amounts of organic solvents (19) or other organic matrices (20).

Changes in the droplet size distribution with solvent composition could contribute to these observed enhancements in ion signal. Droplet size distributions for various solvent compositions are shown in Figs. 2b - 2d. In each case the usual nebulizer gas and liquid flow rates of 0.4 L min^{-1} and $100 \mu\text{L min}^{-1}$ are used, and the droplets are probed at the same spatial position ($z = 25 \text{ mm}$, $r = 0$). As shown in Fig. 2, the "second" hump in the droplet size distribution (i.e., the one for the larger droplets) is attenuated as the MeOH concentration increases.

Table I summarizes the effects of the MeOH/water fraction on $D_{3,2}$, $D_{0.9}$, $D_{0.5}$,

and $D_{0.1}$. Note that both $D_{3,2}$ and $D_{0.9}$ decrease by $\approx 30\%$ and $D_{0.5}$ decrease by $\approx 40\%$ when the MeOH fraction increases from 0 to 30%. However, $D_{0.1}$ remains relatively constant ($7\ \mu\text{m}$ for water and $6\ \mu\text{m}$ for 20% MeOH), which suggests that the fraction of aerosol with diameters less than $6\ \mu\text{m}$ changes only slightly as the MeOH fraction increases. Thus, the net effect of nebulizing solvents containing up to $\approx 30\%$ MeOH is 1) elimination of large droplets, 2) production of finer droplets with diameters between 5 and $15\ \mu\text{m}$, and 3) reduction of the width of the droplet size distribution.

These findings can be attributed to changes in viscosity, surface tension, and liquid density of the various MeOH/H₂O mixtures (21-24), which in turn affect the droplet sizes of the ensuing aerosol. In general, viscosity forces tend to oppose the disintegration of liquids into drops and to resist further breakup of drops already formed, while surface tension forces tend to oppose distortion of the liquid surface. Browner and co-workers (21) suggest that, of the various solvent physical properties, surface tension exerts the principal influence on $D_{3,2}$. Others (23,24) show that $D_{3,2}$ increases as viscosity and surface tension increase and as liquid density decreases. In our experiment, viscosity increases while surface tension and liquid density decrease when the concentration of MeOH in water increases (25). The observed reduction in $D_{3,2}$ as MeOH fraction increases reinforces Browner's viewpoint (21).

When solvents containing more than 30% MeOH are nebulized, the resulting droplet size distributions are similar to Fig. 2d. In fact, $D_{3,2}$, $D_{0.9}$, $D_{0.5}$, and $D_{0.1}$ (see

Table I) differ by less than 5% when the solvent contains from 30% to 50% MeOH in water. Thus, a significant reduction in $D_{3,2}$, $D_{0,9}$, and $D_{0,5}$ is observed only when water containing up to 30% MeOH is nebulized. Increasing the concentration of MeOH further does not yield appreciably finer droplets. Thus, the observed enhancement in ion signal when solvents containing up to 20% of MeOH in water (7) are being nebulized may be due to the production of finer droplets in the aerosol. The gradual decrease in ion signal when solvents with $\geq 30\%$ MeOH are nebulized is presumably due to cooling of the plasma by the additional organic solvent.

Effect of Nebulizer Gas Flow Rate on Droplet Size Distribution

The effect of nebulizer gas flow rate on aerosol size distribution of a DIN spray was also studied. Here only deionized water was introduced to the DIN at $100 \mu\text{L min}^{-1}$, and nebulizer gas flow rates of $0.3 - 0.9 \text{ L min}^{-1}$ were employed. Detector voltages were -400 V for all three PMTs. The aerosol was probed at the usual position ($z = 25 \text{ mm}$, $r = 0$).

For various nebulizer gas flow rates, histograms of the aerosol droplet size distribution obtained were similar to the one shown in Fig. 2a. The $D_{3,2}$, $D_{0,9}$, $D_{0,5}$, and $D_{0,1}$ obtained were $\approx 16 - 17$, 28 , $21 - 24$, and $8 - 10 \mu\text{m}$, respectively. Contrary to expectation, these parameters do not change much with nebulizer gas flow rate. In general, increasing the nebulizer gas flow rate would be expected to increase both the nebulizer gas velocity and the gas/liquid mass flow ratio, resulting in the

formation of finer aerosol droplets (21,23,26). The lack of dependence of droplet sizes on nebulizer gas flow rate can be explained in terms of gas dynamics in nebulizers.

When the DIN is operated at a nebulizer gas flow rate $\geq 0.3 \text{ L min}^{-1}$, the ratio of the upstream pressure ($P_0 \geq 75 \text{ psi}$) to the downstream pressure ($P_d = 15 \text{ psi}$) is ≥ 5 . According to Sharp (27), the DIN should therefore behave as a choked nozzle. The maximum gas velocity from a choked nozzle cannot exceed the sonic velocity; i.e., the gas velocity obtainable at the tip of the DIN remains constant at the sonic velocity as long as $P_0/P_d \geq 5$. If the DIN behaves as a choked nozzle at all the gas flow rates tested, the gas velocity would then be the same in each case, and the droplet-size distribution would not vary much with the nebulizer gas flow rate.

Additional reinforcement comes from other studies of the effect of gas/liquid mass flow ratio on droplet size distributions from pneumatic nebulizers (23,26). According to these results from the literature, $D_{3,2}$ decreases sharply as the gas/liquid mass flow ratio increases to about five. A further increase in the ratio above a value of about five does not yield appreciably finer aerosol droplets (see Figure 6 in Ref. 23). For the DIN, the calculated gas/liquid mass flow ratio exceeds 5 at all of the flow rates used. Thus, increasing the gas/liquid mass flow ratio by increasing the nebulizer gas flow rate does not affect the aerosol droplet size greatly, as shown previously for the DIN (13). In contrast, a conventional concentric pneumatic nebulizer, of the type generally used with ICPs, uses a much higher liquid flow rate

(≈ 0.5 to 1.0 mL min^{-1}) and a somewhat higher gas flow rate (0.5 to 1.0 L min^{-1}) than the DIN. Such conventional nebulizers have a gas/liquid mass flow ratio of between one and three; hence they operate in a regime where an increase in gas flow rate does yield somewhat finer droplets (23,26).

Effect of Nebulizer Gas Flow Rate on Aerosol Droplet Velocity Distribution

Figures 3a - 3c show the velocity distributions (based on count percent) for aerosol droplets produced by the DIN operated at three different nebulizer gas flow rates (0.3 , 0.4 , and 0.9 L min^{-1}). As the nebulizer gas flow rate increases, the mean velocity and the width of the velocity distribution both increase. These trends are also shown in Fig. 4.

This type of DIN is generally operated at a gas flow rate of $\approx 0.4 \text{ L min}^{-1}$. The mean droplet velocity (at $z = 25 \text{ mm}$, $r = 0$) is $\approx 16 \text{ m s}^{-1}$ (Fig. 3b and Fig. 4) at this gas flow rate, which is on the low end of the range of values usually cited for flow velocity into the axial channel of the ICP (≈ 10 to 30 m s^{-1}) (28-30). Our reported velocity values are much slower than those estimated previously by Lawrence et. al. for the original DIN (31). The present work shows that the droplets from the DIN should reside in the axial channel for 2 - 3 ms before they reach the sampling orifice of the MS. This range of residence times is comparable to that seen with conventional nebulizers and torches, and there is no reason to believe that the DIN blows the sample through the plasma too quickly for proper atomization and

ionization.

Radial Profiles of Aerosol Droplet Size and Velocity Distributions

Figure 5 shows the variation in the value of $D_{3,2}$ with r . The nebulizer gas flow rate is 0.4 L min^{-1} , and two liquid flow rates (100 and $300 \mu\text{L min}^{-1}$) are used. The aerosol is probed at $z = 25 \text{ mm}$ downstream from the nebulizer tip.

At a liquid flow rate of $100 \mu\text{L min}^{-1}$, $D_{3,2}$ increases from $12 \mu\text{m}$ at the center ($r = 0$) to $23 \mu\text{m}$ at $r = 6 \text{ mm}$. For a liquid flow rate of $300 \mu\text{L min}^{-1}$, $D_{3,2}$ also increases from $11 \mu\text{m}$ at the center to $28 \mu\text{m}$ at $r = 7 \text{ mm}$. These results suggest that $D_{3,2}$ increases with distance from the center of the spray, as reported by others (32,33). The small aerosol droplets tend to lose their radial momentum component rapidly upon leaving the origin of the spray. These smaller droplets then travel axially with the gas flow. In contrast, the large aerosol droplets move further out into the fringes of the spray because of their larger inertia. This pattern sorts the aerosol droplets by size across the spray cross section.

Figure 6 shows the variation in the volume flux (F) as a function of r . At a liquid flow rate of $100 \mu\text{L min}^{-1}$, no significant change in F is observed for radial positions up to 5 mm from the center; F then increases sharply to $0.011 \text{ cm}^3 \text{ s}^{-1} \text{ cm}^{-2}$ at $r = 6 \text{ mm}$. About 60% of the total volume flux is included in the central region ($0 < r < 4 \text{ mm}$) of the spray. At the higher liquid flow rate ($300 \mu\text{L min}^{-1}$), F increases moderately from $0.0034 \text{ cm}^3 \text{ s}^{-1} \text{ cm}^{-2}$ at the center to $0.006 \text{ cm}^3 \text{ s}^{-1} \text{ cm}^{-2}$ at r

= 4 mm; F then increases sharply to $0.047 \text{ cm}^3 \text{ s}^{-1} \text{ cm}^{-2}$ at $r = 7 \text{ mm}$. If each curve in Fig. 6 is integrated from $r = 0$ to $r = 4 \text{ mm}$, the integrated volume flux is $\approx 0.022 \text{ cm}^3 \text{ s}^{-1} \text{ cm}^{-2}$ at either of the two liquid flow rates. Thus, the same integrated volume flux enters the plasma from the central region of the spray at the two liquid flow rates tested.

These results indicate that an increase in liquid flow rate produces a wider cone of spray. At $300 \mu\text{L min}^{-1}$, about 80% of the total volume flux is outside ($r > 4 \text{ mm}$) the central region of the spray. The Sauter mean diameters of these off-axis droplets are larger than $20 \mu\text{m}$ (see Fig. 5). Presumably, these large droplets are not properly atomized and ionized as they traverse the ICP. Furthermore, these large off-center droplets are probably along the outer rim of the axial channel, where they likely cool the plasma or otherwise interfere with the transfer of energy to the axial channel. These effects probably explain why a high liquid flow rate ($\geq 120 \mu\text{L min}^{-1}$) into the DIN causes a decrease in signal for analyte ions (5).

Figure 7 shows the radial profile of mean flow velocity. Note that similar profiles are observed at both liquid flow rates (100 and $300 \mu\text{L min}^{-1}$). The mean velocity decreases with distance from the center of the spray. A similar trend is commonly observed for sprays produced by low-flow air-assisted concentric nozzles (33) and is attributed to air entrainment and aerodynamic drag on the droplets. Finally, droplets produced at higher liquid flow rate (i.e., $300 \mu\text{L min}^{-1}$) move slightly more slowly at all locations.

Axial Profiles of Aerosol Droplet Size and Velocity Distributions

Figure 8 shows the variation in $D_{3,2}$ as a function of z . The nebulizer gas and liquid flow rates are 0.4 L min^{-1} and $100 \mu\text{L min}^{-1}$. The aerosol droplets are probed along the central axis ($r = 0$). The value of $D_{3,2}$ decreases by 40% from $\approx 14 \mu\text{m}$ (at $z = 2.5 \text{ mm}$) to $\approx 9 \mu\text{m}$ (at $z = 10 \text{ mm}$). Apparently, large droplets are still being shattered into smaller ones at axial positions up to $\approx 10 \text{ mm}$ from the nebulizer tip. The $D_{3,2}$ value then increases from $\approx 9 \mu\text{m}$ (at $z = 10 \text{ mm}$) to $\approx 12 \mu\text{m}$ (at $z = 25 \text{ mm}$), which indicates that some of the smaller droplets coalesce into large ones as they proceed further downstream. Sharp (27) has described this recombination phenomenon. This is not much of a problem when the DIN is used with the ICP, where the hot environment undoubtedly desolvates the droplets before they recombine appreciably.

The large $D_{3,2}$ reported at $z < 5 \text{ mm}$ could also be partly due to measurement error. One source of error is caused by coincidences, i.e., two or more droplets that are present simultaneously in the measuring volume. Near the nebulizer tip where the droplet number density is high, the probability of coincidences is also high. In general, coincidences shift the measured distributions to coarser particle sizes in comparison with the true distribution (34). Another source of error is the possible presence of nonspherical droplets such as liquid ligaments or oscillating droplets near the nebulizer tip ($z \leq 5 \text{ mm}$). Measurement errors of up to 45% on diameters have been reported from these problems (35). Furthermore, the increase in $D_{3,2}$ for

droplets at $z > 10$ mm could also be partly due to evaporation (32). Although evaporation decreases the size of all droplets, smaller droplets evaporate faster, and this pattern weights the average toward larger droplets. Therefore, $D_{3,2}$ increases as evaporation becomes more extensive further downstream in a spray. This problem should not be severe in the present work, because most of our measurements are performed along the central axis of the spray, where evaporation should not be extensive because the atmosphere there also contains plenty of solvent vapor.

Droplet size distributions close to the nebulizer tip ($z = 2.5$ mm) and somewhat further downstream ($z = 10$ mm) are compared in Fig. 9. The droplets are much larger and their size distribution is much broader closer to the nebulizer tip, as expected from the $D_{3,2}$ data shown previously (Fig. 8).

This large decrease in droplet size as axial position increases could be partly responsible for the following analytical observation. When this particular DIN is used for ICP-MS, the auxiliary gas flow rate must be quite high (≈ 1.4 L min⁻¹) to maximize the analyte ion signal (6). Powell et al., who use a somewhat different DIN, likewise report that the auxiliary gas flow rate should be 0.6 L min⁻¹ to maximize Hg⁺ signal (1). Traditionally, the auxiliary gas flow keeps the plasma from contacting and melting the inner tubes of the torch. This role is certainly important with the DIN. In addition, we propose that a high auxiliary gas flow rate also pushes the base of the plasma further downstream away from the DIN. The initially large droplets then have more opportunity to be shattered into smaller, more uniform ones,

either before they reach the plasma or in the upstream reaches of the axial channel. Thus, use of a high auxiliary gas flow rate delays the injection of droplets into the plasma, which may actually be desirable.

An axial profile of mean droplet velocity is given in Fig. 10. The nebulizer gas flow rate is 0.4 L min^{-1} and the liquid flow rate is $100 \mu\text{L min}^{-1}$. The two leftmost points, at axial positions of 2.5 and 5 mm, are probably not appreciably different, so the mean droplet velocity near the nebulizer tip is probably $15 - 17 \text{ m s}^{-1}$. The droplet velocity then quickly reaches a maximum of 25 m s^{-1} at $z = 10 \text{ mm}$, after which velocity drops smoothly to 15 m s^{-1} at $z = 25 \text{ mm}$.

This behavior is attributed to the following processes. Consider an individual droplet leaving the tip of the nebulizer. The droplet takes some time to be accelerated by the gas jet issuing from the nebulizer, so its velocity increases at first. The fast-moving gas from the nebulizer slows down as it encounters the stationary gas in the surrounding air (or the slower gas in the ICP), so the droplet also slows down as it moves further from the nebulizer tip.

CONCLUSION

The PDPA method provides a wealth of information concerning droplet sizes and velocities. These parameters change spatially within the spray issuing from the DIN, which could prove very important because the DIN aerosol goes directly into the ICP, without first being mixed up in a spray chamber. Addition of MeOH at up to 30% in water provides smaller droplets with a narrower size distribution, which probably accounts for the substantial increase in ion signal seen at this methanol composition (7). As indicated by Etkin et al. in their work with dried, monodisperse particulates (36), the spread of droplet or particle sizes is particularly important in ICP-MS, as this spread influences the range of axial positions in the plasma over which atomic ions are generated. If a given particle atomizes too far upstream, the resulting vapor cloud expands too much before it reaches the sampler, and some of the resulting ions are too far from the central axis to pass successfully through both the sampler and skimmer. Thus, the sensitivity in ICP-MS falls off if the sampler is positioned too far downstream from the initial radiation zone (37,38). The present work supports the premise of Jong et al. that monodisperse droplets are particularly advantageous for ICP-MS (36), perhaps more so than for the more-common emission measurements.

LITERATURE CITED

1. M. J. Powell, E. S. K. Quan, D. W. Boomer, and D. R. Wiederin, *Anal. Chem.* **64**, 2253 (1992).
2. R. S. Houk, S. C. K. Shum, and D. R. Wiederin, *Anal. Chim. Acta* **250**, 61 (1991).
3. F. G. Smith, D. R. Wiederin, R. S. Houk, C. B. Egan, and R. E. Serfass, *Anal. Chim. Acta* **248**, 229 (1991).
4. D. R. Wiederin, R. E. Smyczek, and R. S. Houk, *Anal. Chem.* **63**, 1626 (1991).
5. D. R. Wiederin, F. G. Smith, and R. S. Houk, *Anal. Chem.* **63**, 219 (1991).
6. S. C. K. Shum, H. Pang, and R. S. Houk, *Anal. Chem.* **64**, 2444 (1992).
7. S. C. K. Shum, R. Neddersen, and R. S. Houk, *Analyst* **117**, 577 (1992).
8. W. D. Bachalo and M. J. Houser, *Opt. Eng.* **23**, 583 (1984).
9. W. D. Bachalo, *Appl. Opt.* **19**, 363 (1980).
10. R. H. Clifford, I. Ishii, A. Montaser, and G. A. Meyer, *Anal. Chem.* **62**, 390 (1990).
11. R. H. Clifford, P. Sohal, H. Liu, and A. Montaser, *Spectrochim. Acta* **47B**, 1107 (1992).
12. A. H. Lefebvre, *Atomization and Sprays* (Hemisphere Publishing Corp., New York, 1989), Chap. 1.
13. D. R. Wiederin and R. S. Houk, *Appl. Spectrosc.* **45**, 1408 (1991).

14. L. G. Dodge, D. J. Rhodes, and R. D. Reitz, *Appl. Opt.* **26**, 2144 (1987).
15. T. A. Jackson and G. S. Samuelsen, *J. Eng. Gas Turbines Power* **108**, 196 (1986).
16. D. C. Hammond, Jr., *Appl. Opt.* **20**, 493 (1981).
17. American Society for Testing Materials (ASTM) Standard E799.
18. R. W. Sellens, "Phase-Doppler Measurements Near the Nozzle in a Low-Pressure Water Spray" in *Liquid Particle Size Measurement Techniques*, E. D. Hirleman, W. D. Bachalo, and P. G. Felton, Eds (ASTM STP 1083, American Society for Testing and Materials, Philadelphia, 1990), Vol. 2, p. 193.
19. P. Allain, L. Jaunault, Y. Murras, J. Mermet, and T. Delaporte, *Anal. Chem.* **63**, 1499 (1991).
20. G. F. Wallace, *Enhancement of Some Analyte Signals by Carbon Compounds in ICP-MS*, FACSS Conference (Philadelphia, PA, September 1992), Paper No. 458.
21. A. Canals, J. Wagner, and R. F. Browner, *Spectrochim. Acta* **43B**, 1321 (1988).
22. R. F. Browner and A. W. Boorn, *Anal. Chem.* **56**, 787A (1984).
23. G. E. Lorenzetto and A. H. Lefebvre, *AIAA Journal* **15**, 1006 (1977).
24. A. A. Rizkalla and A. H. Lefebvre, *ASMS J. Fluids Eng.* **97**, 316 (1975).
25. C. Robles, J. Mora, and A. Canals, *Appl. Spectrosc.* **46**, 669 (1992).
26. K. Y. Kim and W. R. Marshall, Jr., *AIChE Journal* **17**, 575 (1971).
27. B. L. Sharp, *J. Anal. At. Spectrom.* **3**, 613 (1988).

28. R. K. Winge, J. S. Crain, and R. S. Houk, *J. Anal. At. Spectrom.* **6**, 601 (1991).
29. M. T. Cicerone and P. B. Farnsworth, *Spectrochim. Acta* **44B**, 897 (1989).
30. R. M. Barnes and J. L. Genna, *Spectrochim. Acta* **36B**, 299 (1981).
31. K. E. Lawrence, G. W. Rice, and V. A. Fassel, *Anal. Chem.* **56**, 289 (1984).
32. R. D. Reitz, "Effect of Vaporization and Turbulence on Spray Drop-Size and Velocity Distributions" in *Liquid Particle Size Measurement Techniques*, E. D. Hirleman, W. D. Bachalo, and P. G. Felton, Eds (ASTM STP 1083, American Society for Testing and Materials, Philadelphia, 1990), Vol. 2, p. 225.
33. T. A. Jackson and G. S. Samuelsen, *Appl. Opt.* **26**, 2137 (1987).
34. J. Raasch and H. Umhauer, *Part. Charact.* **1**, 53 (1984).
35. T. A. Jackson, "Droplet Sizing Interferometry" in *Liquid Particle Size Measurement Techniques*, E. D. Hirleman, W. D. Bachalo, and P. G. Felton, Eds (ASTM STP 1083, American Society for Testing and Materials, Philadelphia, 1990), Vol. 2, p. 151.
36. B. Etkin, J. B. French, and R. Jong, *Monodisperse Dried Microparticulate Injection*, FACSS Conference (Philadelphia, PA, September 1992), Paper No. 218.
37. S. H. Tan and G. Horlick, *J. Anal. At. Spectrom.* **2**, 745 (1987).
38. M. A. Vaughan, G. Horlick, and S. H. Tan, *J. Anal. At. Spectrom.* **2**, 765 (1987).

Table I. Effect of MeOH percentage in water on $D_{3,2}$, $D_{0.9}$, $D_{0.5}$, and $D_{0.1}$ for the aerosol droplets produced by the DIN^a

% MeOH (v/v %)	$D_{3,2}$ (μm)	$D_{0.9}$ (μm)	$D_{0.5}$ (μm)	$D_{0.1}$ (μm)
0	14.1	28.1	18.3	6.7
5	12.0	27.3	14.4	5.9
10	12.2	29.7	14.8	5.9
20	9.5	18.6	10.3	5.9
30	8.8	18.6	9.6	5.1
40	8.6	17.8	9.4	5.1
50	9.2	19.4	10.0	5.1

- ^a The nebulizer gas flow rate was 0.4 L min^{-1} and the liquid flow rate was $100 \mu\text{L min}^{-1}$. PDPA detectors were operated at -450 V . The probe beam was located at $z = 25 \text{ mm}$, $r = 0$.

Figure 1. Schematic diagram of the optical system for the PDPA: (A) He-Ne laser; (B) beamsplitter; (C) DIN; (D) DIN spray; (E) beam intersection point; (F) focusing lenses; (G) PMTs.

Figure 2. The effect of methanol concentration in water on the aerosol droplet size distribution. Experimental conditions are given in the text. The methanol concentrations were: (a) 0% (Water), (b) 10% MeOH, (c) 20% MeOH, and (d) 30% MeOH.

Figure 2. (continued)

Figure 2. (continued)

Figure 2. (continued)

Figure 3. The effect of nebulizer gas flow rate on the aerosol droplet velocity distribution. Experimental conditions are given in the text. The nebulizer gas flow rates were: (a) 0.3, (b) 0.4, and (c) 0.7 L min⁻¹.

Figure 3. (continued)

Figure 3. (continued)

Figure 4. The effect of nebulizer gas flow rate on droplet velocity distribution. Error bars indicate $\pm$ one standard deviation.

Figure 5. Radial profile of Sauter mean diameter ($D_{3,2}$) at $z = 25$ mm for two liquid flow rates: (○) $100 \mu\text{L min}^{-1}$ and (●) $300 \mu\text{L min}^{-1}$.

Figure 6. Radial profile of droplet volume flux (F) at $z = 25$ mm for two liquid flow rates: (○) $100 \mu\text{L min}^{-1}$ and (●) $300 \mu\text{L min}^{-1}$.

Figure 7. Radial profile of droplet mean velocity at $z = 25$ mm for two liquid flow rates: ($\circ$) $100 \mu\text{L min}^{-1}$ and ($\bullet$) $300 \mu\text{L min}^{-1}$.

Figure 8. Axial profile of Sauter mean diameter ($D_{3,2}$) at $r = 0$.

Figure 9. Droplet size distributions measured at two different axial position: (a) $z = 2.5 \text{ mm}$, $r = 0$ and (b) $z = 10 \text{ mm}$, $r = 0$. Experimental conditions are given in the text.

Figure 9. (continued)

Figure 10. Axial profile of droplet mean velocity at $r = 0$.

PAPER IV

**ELEMENTAL SPECIATION BY ANION EXCHANGE AND SIZE EXCLUSION
CHROMATOGRAPHY WITH DETECTION BY INDUCTIVELY COUPLED
PLASMA MASS SPECTROMETRY WITH DIRECT INJECTION NEBULIZATION**

INTRODUCTION

The toxicological and biological roles of trace elements depend on their chemical forms and/or oxidation states (1-4). Thus, analytical methodology for measuring trace element speciation is necessary. Liquid chromatography (LC) has been coupled to inductively coupled plasma mass spectrometry (ICP-MS) to provide such speciation information (5-33). So far, various modes of LC including reversed-phase (RP) (6-10), reversed-phase ion-pairing (RP-IP) (11-19), anion chromatography (AC) (6,13,20-23), cation chromatography (CC) (15,24-25), and size exclusion chromatography (SEC) (6,26-33) have been combined with ICP-MS.

Fassel and co-workers originally described the direct injection nebulizer (DIN), which is a microconcentric pneumatic nebulizer placed inside the ICP torch (34). Recently, a new version of the DIN has been successfully used as an interface for LC-ICP-MS experiments (17,19,35) and has shown substantial improvement in terms of chromatographic resolution and detection limits, compared to those obtainable with conventional nebulizers. The DIN has a low dead volume (0.5 - 2 μL) (19) and produces a mist of fine droplets (Sauter mean diameter $\approx 15 \mu\text{m}$) (36). When used for LC-ICP-MS of As, Hg, Pb, and Sn species, the DIN provides absolute detection limits that are superior by 1 - 2 orders of magnitude to those obtained with conventional nebulizers (17,19). In addition, the low dead volume and absence of a spray chamber minimize postcolumn band broadening and facilitate the use of

microbore LC columns and liquid flow rates (30 - 100 $\mu\text{L min}^{-1}$) that are low enough for all the column effluent to be introduced into the plasma.

Metals are found in many forms in human blood serum, and the distribution of a metal in serum between amino acids, proteins, enzymes, etc. often has important pathological implications. For example, iron exists in porphyrins in hemoglobin, myoglobin, cytochromes, peroxidases and catalases, is associated with sulfur in ferredoxins, and in invertebrates is present as the oxygen-carrying pigment, hemerythrin (37). Speciation analysis of a biological sample with its complex organic matrix is a difficult task. Nevertheless, it is essential that analytical capabilities in this area be improved if we are to understand in more detail the role of trace elements in health and disease (38).

A case in point is selenium, which is an essential element. However, not all selenium species are considered useful. In general, both of the common inorganic selenium species (selenite and selenate) are toxic, while organoselenium species, in particular the Se-containing amino acids, are biologically useful. Furthermore, ICP-MS detection for chromatography should also permit isotope ratio measurements for individual elemental species for isotope tracer studies. Isotope tracer studies of the nutritional and medical role of Se and Se species have recently been of great interest (39,40).

The present work demonstrates the separation and measurement of various metalloproteins in human serum and two selenium species (SeO_3^{2-} and SeO_4^{2-}) by

SEC-DIN-ICP-MS and AC-DIN-ICP-MS, respectively. In addition, preliminary results of isotope ratio measurements on chromatographically separated species of Se are presented.

EXPERIMENTAL SECTION

HPLC-DIN-ICP-MS

The HPLC system was composed of an SSI Model 222D metal-free micro-flow pump (Scientific Systems, Inc., State College, PA), a Rheodyne 9010 metal-free high-pressure sample injector with a 2 μ L PEEK injection loop, and an appropriate analytical column. A 1.6-mm-i.d. x 10-cm-long anion-exchange column (MCANX1710, CETAC Technologies, Omaha, NE) was used for the separation of Se species. A 2.0-mm-i.d. x 25-cm-long size exclusion column (GPC 300, SynChrom, Inc., Lafayette, IN) was used for separation of metalloproteins.

The conditions for both separations are summarized in Table I. The analytical column was equilibrated with the mobile phase prior to use. Several different combinations of mobile phase concentration, type and concentration of counter ion, pH, etc. were evaluated to optimize chromatographic performance. The conditions listed in Table I are those that yielded the best chromatographic resolution of the various sets tested.

The design and the construction of the DIN have been discussed elsewhere (17,41). A 50- μ m-i.d. x 40-cm-long fused silica capillary was used to transport effluent from the column into the plasma. The 50- μ m DIN is more resistant to plugging than the 30- μ m one used previously (19). The width of the annular gap between the inner capillary and the nebulizer tip was $\approx 25 \mu$ m, as in the previous

work (17,19).

The ICP-MS used was the Elan Model 250 (Perkin-Elmer Sciex, Thornhill, ON, Canada). Table II summarizes the instrumental operating conditions of the ICP-MS device. The plasma and sampling conditions indicated with an asterisk were optimized daily to maximize the signal from the analyte of interest.

Data Acquisition

For the separation of metalloproteins, ten isotopes were measured during the chromatographic separation, as listed in Table I. To achieve compromise operating conditions over the whole mass range, the plasma, sampling conditions, and ion lens voltages were optimized to give maximum signal for $^{103}\text{Rh}^+$. The data were acquired in the multielement mode by peak hopping over ten m/z positions using a 0.2 s measurement time, a 10 ms dwell time, and 1 measurement per peak.

For the separation of SeO_3^{2-} and SeO_4^{2-} , two isotopes (^{74}Se and ^{78}Se) were monitored during the chromatographic separation. The data were acquired by peak hopping over the two isotopes using a 1 s measurement time, a 20 ms dwell time, and 1 measurement per mass peak. Several different combinations of measurement time (0.1 - 3 s), dwell time (1 - 100 ms), and measurements per mass peak (1 - 3) were evaluated for precision studies. The conditions reported for this work yielded the best precision of the various sets tested.

Chromatograms were recorded in real time and stored on the hard disk of an

IBM PS/2 Model 70 computer. These data (as ASCII files) were then processed using a spread sheet program. The raw count rates were first smoothed with a five-point Savitzky - Golay routine (42). Peak area was determined by summing all the count rates under each peak. The background was measured while nebulizing only the mobile phase by summing the total counts in the particular chromatographic region corresponding to each peak. For this work, the detection limit was defined as the amount of the element necessary to give a peak area equal to three times the standard deviation of the background count rate at each analyte mass.

Reagents and Samples

Deionized water (18 M Ω cm at 25 °C) obtained from a Barnstead Nanopure-II system (Newton, MA) was used throughout. For the separation of metalloproteins, a tris(hydroxymethyl)aminomethane/hydrochloric acid (Tris/HCl) buffer was used as the mobile phase. Eluent of high ionic strength, such as 0.1 M NaCl, was not used to avoid plugging of the DIN. A 0.1 M Tris/HCl solution was prepared by dissolving certified A.C.S. grade Tris (Fisher Scientific, Fair Lawn, NJ) in deionized water. The pH was adjusted to 6.9 with concentrated Ultrex II grade HCl (J. T. Baker, Inc., Phillipsburg, NJ). The relation between molecular weight (MW) and retention time for the size exclusion column was calibrated by monitoring the retention times of several pure protein standards including thyroglobulin (MW 6.7×10^5 g mol⁻¹), ferritin (MW 4.4×10^5), β -amylase (MW 2.0×10^5), alcohol dehydrogenase (MW 1.5

$\times 10^5$), and carbonic anhydrase (MW 2.9×10^4). These standards were obtained from Sigma Chemicals (St. Louis, MO). A human serum standard reference material (SRM 909a-1, National Institute of Standards and Technology, Gaithersburg, MD) was reconstituted in 10 mL of 0.1 M Tris/HCl buffer and used without further dilution.

For the separation of Se species, a mixture of 5 mM $(\text{NH}_4)_2\text{CO}_3$ / 5 mM NH_4HCO_3 was used as the mobile phase. Pure samples (Certified A.C.S. grade) of these compounds were obtained from Fisher Scientific. Carbonate and hydrogen carbonate in their ammonium form were chosen, rather than the usual sodium form, to avoid matrix effects (43-47) and clogging of the DIN (17,19).

RESULTS AND DISCUSSION

Calibration of the Size Exclusion Column

In this study, approximately 2×10^{-5} M solutions of each of the metalloprotein standards in 0.1 M Tris/HCl were prepared. Mass spectra ($m/z = 20 - 220$) of these metalloprotein standard solutions were obtained. The unique elements (i.e., those other than C, N, O, H, and Cl) present in each protein standard were identified and are given in Table III. For example, three metals (Fe, Cu, and Zn) were found to be associated with ferritin (48).

Next, the calibration curve for the size exclusion column was obtained. The retention times for each of the protein standards were determined by operating the ICP-MS device in the peak-hopping mode while monitoring the isotopes of interest. These results are also given in Table III. Relative standard deviations ($n = 3$) for the measured retention times were under 1.5%. Figure 1 shows a typical column calibration curve obtained by plotting $\log$ (molecular weight) vs retention time. The correlation coefficient equals 0.997. Note that the selective permeation (linear) range of this gel column was $\approx 4.5 - 5.75$ min and was rather small. In principle, this range can be extended by using either several columns in series, each containing gel of different pore size, or a mixed gel column (49).

Separation of Metalloproteins by SEC-DIN-ICP-MS

Chromatograms for the separation of metalloproteins are given in Figure 2. Pertinent separation conditions are given in Table I. Six metal-binding molecular weight fractions (> 650 , 300 , 130 , 85 , 50 , and 15 kDa) were observed. Table IV summarizes and sorts the metal contents of these molecular weight fractions. Four of these molecular weight fractions (> 650 , 300 , 130 , and 15 kDa) contained Pb. Gercken and Barnes (28), in a similar study, reported three molecular weight fractions (> 600 , 260 , and 140 kDa) of Pb species. The major Pb-containing molecular weight fraction (130 kDa) also contained Cd, Zn, Ba, Cu, and Na. One possible protein responsible for this molecular weight fraction is ceruloplasmin, a well known glycoprotein for metal storage and electron transfer (28,50). There was only one Fe-binding molecular weight fraction found at ≈ 85 kDa, which could be serum transferrin, a well known iron glycoprotein in physiological fluids (51). The proteins responsible for the other molecular weight fractions await identification by techniques such as immunological reactions or other types of MS.

At about 6.7 min after injection (Figure 2), a Pb, two Cd and two Ba peaks were observed. These peaks coeluted with Na from the sample. The Pb peak was determined to be inorganic lead (Pb^{2+}) by adding some Pb^{2+} to the sample. However, the addition of Cd^{+2} and Ba^{+2} to the sample yielded peaks at retention times longer than 8.5 min; that is, those Cd and Ba peaks found at about 6.7 min were not from Cd^{+2} and Ba^{+2} . In a previous study with the DIN (52), the

background across the upper end of the mass range ($m/z > 45$) shifted to higher values as the salt content of the solvent increased. Perhaps the unknown Cd and Ba peaks were artifacts from a background shift when the highly concentrated Na matrix eluted.

Elevated backgrounds for Zn, Ba, Cu, Fe, and Na were observed (Figure 2). The elevated Ba and Na background was caused by the presence of Ba and Na impurities in the buffer. The elevated ^{54}Fe , ^{63}Cu , and ^{64}Zn backgrounds were caused by the formation of $^{40}\text{Ar}^{14}\text{N}^+$, $^{40}\text{Ar}^{23}\text{Na}^+$, and $^{40}\text{Ar}^{24}\text{C}_2^+$ polyatomic ions resulting from the nitrogen, carbon and sodium in the buffer solution. The high background (at $m/z = 64$) could also be due to the formation SO_2^+ from S-containing amino acids in sample.

Estimated Detection Limits

The detection limits for Fe, Cu, Zn, Cd, and Pb in metalloproteins were estimated from their peak heights for the molecular weight fraction at 130k Da. The total concentrations of Fe, Cu, Zn, Cd, and Pb in the serum sample measured by standard addition were 113, 107, 163, 6.3, and $6.5 \mu\text{g L}^{-1}$, respectively. The appropriate amount of analyte responsible for each peak in the molecular weight fraction (130k Da) was determined by proportion.

Calculated detection limits are given in Table V. The absolute detection limits for Fe, Cu, Zn, Cd, and Pb in metalloproteins were 3, 0.7, 1, 0.5, and 0.5 pg,

respectively. These absolute detection limits are comparable to those obtained for the same metals in aqueous samples with the DIN and are superior by 1 - 2 orders of magnitude over those obtained previously by SEC-ICP-MS (Table V). The relative detection limits for Fe, Cu, Zn, Cd and Pb were 2, 0.4, 0.5, 0.3, and $0.3 \mu\text{g L}^{-1}$, respectively. These relative detection limits are comparable with those obtained previously by SEC-ICP-MS with conventional nebulizers (Table V).

Separation of Se Species by AC-DIN-ICP-MS

Figure 3 shows a typical chromatogram for the separation of SeO_3^{2-} and SeO_4^{2-} . Pertinent separation conditions are given in Table I. The amount of Se for each species used per injection was 5 ng. The analytical figures of merit for this separation method are given in Table VI. Peak area measurements indicated that the Se sensitivity (total counts / ng Se) was similar (i.e., within 5%) for the two forms of Se (Table VI). Precision based on five separate injections and measurement of peak areas was $\approx 3\%$ RSD for both species. Absolute detection limits calculated using ^{78}Se and peak area measurements were $\approx 15 \text{ pg}$ for both of the species. These absolute detection limits were superior by an order of magnitude over those obtained previously by LC-ICP-MS with ultrasonic nebulizer (11). Relative detection limits for both forms were $\approx 7 - 8 \mu\text{g L}^{-1}$, which are comparable to those obtained previously (11).

Isotope ratios for each selenium species were determined by measuring the

area of the appropriate chromatographic peaks for each isotope. Table VII summarizes the results on isotope ratio measurements of the two selenium species. When only 5 ng of Se for each Se species was used, the relative standard deviations obtainable was $\approx 2\%$. High RSDs obtainable was caused by the low total counts for ^{74}Se and ^{78}Se (≈ 80000 and 2800 counts for ^{78}Se and ^{74}Se , respectively). The RSD expected from counting statistics on the minor isotope (^{74}Se) would be $(\sqrt{(2800)}/2800)$ or 1.8%. This value is comparable to the 2% precision cited above for the isotope ratio, so counting statistics limited the precision in this case. When the amount of Se for each selenium species was increased to 25 ng, the RSDs improved to 0.3 and 0.5%. Therefore, the amount of sample injected had a significant effect on the precision in the present work, as expected because of the low abundance of ^{74}Se .

With conventional nebulizers, experience has shown that 1 - 10 μg of analyte is required for isotope ratio measurements with RSDs of $\pm 0.5\%$ or better, using either flow injection or continuous flow sample introductions (53-55). In the present work the amount of analyte used per injection was 25 ng. Suppose 10 injections were needed to evaluate the precision of the isotope ratio measurement. These injections would require a total of $\approx 0.25 \mu\text{g}$ of the analyte. The DIN is therefore attractive for isotope ratio measurements when the amount of analyte available is below $\approx 1 \mu\text{g}$.

CONCLUSION

The analytical merits of SEC-DIN-ICP-MS and AC-DIN-ICP-MS are demonstrated. The absolute detection limits for Se in a test mixture and for metals (Fe, Cu, Zn, Cd, and Pb) in protein fractions in human serum are improved by 1 - 2 orders of magnitude, relative to those obtained with conventional nebulizers. Metalloproteins were measured directly in human serum without preliminary extraction or preconcentration. The low dead volume ($< 2 \mu\text{L}$) associated with the DIN resulted in low extracolumn broadening and good chromatographic resolution.

The capability of using chromatographic peak areas for isotope ratio measurements on Se species is also demonstrated. A relative standard deviation of less than 0.5% is obtainable in cases where precision is not limited by counting statistics. The feasibility of applying the same method to measure isotope ratios for elements of interest in tough matrix such as urine, serum, and seawater warrants further investigation.

LITERATURE CITED

1. Goyer, R. A. In *Casarett and Doull's Toxicology: The Basic Science of Poisons*, 4th ed.; Casarett, L. J.; Amdur, M. O.; Doull, J.; Klaassen, C. D., Eds.; Pergamon Press: New York, 1991; Chapter 19.
2. Cappon, C. J. *LC/GC* **1988**, *6*, 584-599.
3. Batley, G. E.; Low, G. K.-C. In *Trace Element Speciation Analytical Methods and Problems*; Batley, G. E., Ed.; CRC Press, Inc.: Boca Raton, Florida, 1989; Chapter 6.
4. Gardiner, P. E. *J. Anal. At. Spectrom.* **1988**, *3*, 163-168.
5. Houk, R. S.; Jiang, S. J. In *Trace Metal Analysis and Speciation*; Krull, I. S., Ed.; Elsevier: New York, 1991; Chapter 5.
6. Elder, R. C.; Jones, W. B.; Tepperman, K. In *Element - Specific Chromatographic Detection by Atomic Emission Spectroscopy*; Uden, P. C., Ed.; ACS Symposium Series 479: Washington, DC, 1992, Chapter 18.
7. Braverman, D. S. *J. Anal. At. Spectrom.* **1992**, *7*, 43-46.
8. Suyani, H.; Heitkemper, D.; Creed, J.; Caruso, J. *Appl. Spectrosc.* **1989**, *43*, 962-967.
9. Bushee, D. S. *Analyst* **1988**, *113*, 1167-1170.
10. Bushee, D. S.; Moody, J. R.; May, J. C. *J. Anal. At. Spectrom.* **1989**, *4*, 773-775.

11. Thompson, J. J.; Houk, R. S. *Anal. Chem.* **1986**, *58*, 2541-2548.
12. Jiang, S. J.; Houk, R. S. *Spectrochim. Acta* **1988**, *43B*, 405-411.
13. Beauchemin, D.; Siu, K. W. M.; McLaren, J. W.; Berman, S. S. *J. Anal. At. Spectrom.* **1989**, *4*, 285-289.
14. Shibata, Y.; Morita, M. *Anal. Chem.* **1989**, *61*, 2116-2118.
15. Suyani, H.; Creed, J.; Davidson, T.; Caruso, J. J. *Chromatogr. Sci.* **1989**, *27*, 139-143.
16. Al-Rashdan, A.; Vela, N. P.; Caruso, J. A.; Heitkemper, D. T. *J. Anal. At. Spectrom.* **1992**, *7*, 551-555.
17. Shum, S. C. K.; Neddersen, R.; Houk, R. S. *Analyst* **1992**, *117*, 577-582.
18. Zhao, Z.; Jones, W. B.; Tepperman, K.; Dorsey, J. G.; Elder, R. C. *J. Pharm. Biomed. Anal.* **1992**, *10*, 279-287.
19. Shum, S. C. K.; Pang, H. M.; Houk, R. S. *Anal. Chem.* **1992**, *64*, 2444-2450.
20. Sheppard, B. S.; Caruso, J. A.; Heitkemper, D. T.; Wolnik, K. A. *Analyst* **1992**, *117*, 971-975.
21. Sheppard, B. S.; Shen, W.-L.; Caruso, J. A.; Heitkemper, D. T.; Fricke, F. L. *J. Anal. At. Spectrom.* **1990**, *5*, 431-435.
22. Heitkemper, D.; Creed, J.; Caruso, J.; Fricke, F. L. *J. Anal. At. Spectrom.* **1989**, *4*, 279-284.

23. Suzuki, S.; Tsuchihashi, H.; Nakajima, K.; Matsushita, A.; Nagao, T. *J. Chromatogr.* **1988**, *437*, 322-327.
24. Kawabata, K.; Kishi, Y.; Kawaguchi, O.; Watanabe, Y.; Inoue, Y. *Anal. Chem.* **1991**, *63*, 2137-2140.
25. Boomer, D. W.; Powell, M. J.; Hipfner, J. *Talanta* **1990**, *37*, 127-134.
26. Takatera, K.; Watanabe, T. *Anal. Sci.* **1992**, *8*, 469-474.
27. Owen, L. M. W.; Crews, H. M.; Hutton, R. C.; Walsh, A. *Analyst* **1992**, *117*, 649-655.
28. Gercke, B.; Barnes, R. M. *Anal. Chem.* **1991**, *63*, 283-287.
29. Takatera, K.; Watanabe, T. *Anal. Sci.*, **1991**, *7*, 695-698.
30. Mason, A. J.; Storms, S. D.; Jenkins, K. D. *Anal. Biochem.* **1990**, *186*, 187-201.
31. Crews, H. M.; Dean, J. R.; Ebdon, L.; Massey, R. C. *Analyst* **1989**, *114*, 895-899.
32. Matz, S. G.; Elder, R. C.; Tepperman, K. J. *Anal. At. Spectrom.* **1989**, *4*, 767-771.
33. Dean, J. R.; Munro, S.; Ebdon, L.; Crews, H. M.; Massey, R. C. *J. Anal. At. Spectrom.* **1987**, *2*, 607-610.
34. LaFreniere, K. E.; Fassel, V. A.; Eckels, D. E. *Anal. Chem.* **1987**, *59*, 879-887.

35. Houk, R. S.; Shum, S. C. K.; Wiederin, D. R. *Anal. Chim. Acta* **1991**, *250*, 61-70.
36. Shum, S. C. K.; Pang, H. M.; Johnson, S. K.; Houk, R. S. *Appl. Spectrosc.*, **1993**, in press.
37. Otsuka, S.; Yamanaka, T., Eds. *Metalloproteins Chemical Properties and Biological Effects*; Kodansha Ltd.: Tokyo, 1988; Chapter 5.
38. Florence, T. M. In *Trace Element Speciation Analytical Methods and Problems*; Batley, G. E., Ed.; CRC Press, Inc.: Boca Raton, Florida, 1989; Chapter 9.
39. Janghorbani, M.; Lynch, N. E.; Mooers, C. S.; Ting, B. T. G. *J. Nutr.*, **1990**, *120*, 190-199.
40. Janghorbani, M.; Young, V. R. In *Selenium in Biology and Medicine, Third International Symposium*; Combs, G. F., Jr.; Spallholz, J. E.; Lewander, O. A.; Oldfield, J., Eds.; Van Nostrand Reinhold: New York, 1987; pp. 450-471.
41. Wiederin, D. R.; Smith, F. G.; Houk, R. S. *Anal. Chem.* **1991**, *63*, 219-225.
42. Savitzky, A.; Golay, M. J. E. *Anal. Chem.* **1964**, *36*, 1627-1639.
43. Olivares, J. A.; Houk, R. S. *Anal. Chem.* **1986**, *58*, 20-25.
44. Jiang, S. -J.; Houk, R. S. *Anal. Chem.* **1986**, *58*, 1739-1743.
45. Douglas, D. J.; Kerr, L. A. *J. Anal. At. Spectrom.* **1988**, *3*, 749-752.
46. Tan, S. H.; Horlick, G. *J. Anal. At. Spectrom.* **1987**, *2*, 745-763.

47. Beauchemin, D.; McLaren, J. W.; Berman, S. S. *Spectrochim. Acta* **1987**, *42B*, 467-490.
48. Shinjo, S. In *Metalloproteins Chemical Properties and Biological Effects*; Otsuka, S.; Yamanaka, T., Ed.; Kodansha Ltd.: Tokyo, 1988; pp. 249-255.
49. Hunt, B. J. In *Size Exclusion Chromatography*; Hunt, B. J.; Holding, S. R., Ed.; Blackie: London, 1989; p. 10.
50. Nakamura, T. In *Metalloproteins Chemical Properties and Biological Effects*; Otsuka, S.; Yamanaka, T., Eds.; Kodansha Ltd.: Tokyo, 1988; pp. 291-307.
51. Brock, J. H. In *Metalloproteins*, Part 2; Harrison, P. M., Ed.; Macmillan: London, 1985; Chapter 5.
52. Wiederin, D. R.; Smyczek, R. E.; Houk, R. S. *Anal. Chem.* **1991**, *63*, 1626-1631.
53. Janghorbani, M.; Ting, B. T. G.; *J. Nutr. Biochem.* **1990**, *1*, 4-19.
54. Viczián, M.; Lásztity, A.; Barnes, R. M. *J. Anal. At. Spectrom.* **1990**, *5*, 293-300.
55. Date, A. R.; Cheung, Y. Y. *Analyst* **1987**, *112*, 1531-1540.

Table I. Chromatographic conditions

	Separation of selenium species	Separation of metalloproteins
Column	CETAC Technologies PEEK column 1.6-mm-i.d. x 100-mm-long	SynChrom, Inc. SynChropak GPC 300 2-mm-i.d. x 250-mm-long
Stationary phase	Anion exchange resin (5- μ m particles)	Silica gel (5- μ m particles, pore size = 300 Å)
Mobile phase	5 mM NH_4HCO_3 / 5 mM $(\text{NH}_4)_2\text{CO}_3$	0.1 M Tris/HCl (pH = 6.9)
Sample flow rate	100 $\mu\text{L min}^{-1}$	100 $\mu\text{L min}^{-1}$
Injection volume	2 μL	2 μL
Isotopes monitored	$m/z = {}^{74}\text{Se}, {}^{78}\text{Se}$	$m/z = {}^{23}\text{Na}, {}^{63}\text{Cu}, {}^{65}\text{Cu}, {}^{54}\text{Fe}, {}^{57}\text{Fe},$ ${}^{64}\text{Zn}, {}^{67}\text{Zn}, {}^{208}\text{Pb}, {}^{138}\text{Ba}, {}^{114}\text{Cd}$

Table II. Instrument Conditions and Operating Procedures

ICP torch	Modified Sciex short torch : injector tube orifice diameter = 1 mm; 6-mm-o.d. x 4-mm-i.d. quartz tee attached at torch base
argon flow rates (L min ⁻¹)	
outer	12*
auxiliary	1*
make-up	0.30* regulated by mass flow controller
nebulizer gas	0.4*
sample flow rate	100 μ L min ⁻¹ typical
forward power	1.4 kW*
sampling position	20 mm from load coil, on center*
sampler	Copper, 1.0-mm-diameter orifice
skimmer	Nickel, 0.9-mm-diameter orifice
detector voltage	-4000 V
ion lens setting	
bessel	-19.80 V
plate	-11.00 V
barrel	+5.42 V
photon stop	-7.46 V*
operating pressures	
interface	1.5 torr
quadrupole chamber	3 x 10 ⁻⁵ torr

* Typical values cited. These parameters were adjusted daily to maximize ion signal (see text) and differed slightly from day to day.

Table III. Estimated molecular weight values and retention times for protein standards separated by SEC-DIN-ICP-MS using 0.1 M Tris (pH 6.9) and elements associated with the individual protein standards.

Protein standards	Estimated molecular weight (Da) ^a	Isotope monitored	Mean retention time (min) ^b	RSD (%) ^b
Thyroglobulin	6.7 x 10 ⁵	¹²⁷ I	4.47	0.8
Ferritin	4.4 x 10 ⁵	⁵⁴ Fe, ⁶³ Cu, ⁶⁴ Zn	4.71	1.1
β-Amylase	2.0 x 10 ⁵	⁶³ Cu	4.94	1.5
Alcohol dehydrogenase	1.5 x 10 ⁵	⁶⁴ Zn, ¹¹⁴ Cd	5.15	1.3
Carbonic anhydrase	2.9 x 10 ⁴	⁶³ Cu, ⁶⁴ Zn	5.88	1.4

^aMolecular weight values were provided by Sigma.

^bn = 3 separate injections

Table IV. Six metal-binding molecular weight fractions determined in human serum

Metal-binding molecular weight fractions	Trace metals associated with each molecular weight fraction
> 650k	Pb, Cd, Zn, Cu
300k	Pb, Zn, Cu
130k	Pb, Cd, Zn, Ba, Cu, Na
85k	Fe
50k	Zn
15k	Pb, Zn

Table V. Estimated detection limits for metal ions in proteins

Metal	m/z	Detection limit			
		Present work ^a		Literature values	
		Absolute (pg)	Relative ($\mu\text{g L}^{-1}$)	Absolute (pg)	Relative ($\mu\text{g L}^{-1}$)
Fe	54	3	2	900 ^b	9 ^b
Cu	63	0.7	0.4	6 ^b , 270 ^c , 10 ^d	0.06 ^b , 2.7 ^c , 0.2 ^d
Zn	64	1	0.5	40 ^b , 63 ^c , 25 ^d	0.4 ^b , 0.63 ^c , 0.5 ^d
Cd	114	0.5	0.3	12 ^c , 25 ^d , 100 ^c	0.12 ^c , 0.5 ^d , 1 ^c
Pb	208	0.5	0.3	5 - 15 ^b	0.05 - 0.15 ^b

^a See text for calculation.

^b Ref. 28

^c Ref. 30

^d Ref. 32

^e Ref. 31

Table VI. Analytical figures of merit for the separation of selenium species

	SeO_3^{2-}	SeO_4^{2-}
Retention time (min)	3.4	9.0
Sensitivity ^a (counts/pg of Se)	15	14
RSD ^b (%)	2.8	2.9
Detection limits ^c :		
(pg of Se)	14	15
($\mu\text{g L}^{-1}$, ppb)	7	8

^a Sensitivity was calculated based on net peak area and amount injected.

^b Relative standard deviation of peak area for five replicate injections of 5 ng (as Se) of each species. See Table I for LC conditions.

^c Detection limit defined as amount of Se required to yield a net peak that was 3 times the standard deviation of background. Peak areas were used in these calculations.

Table VII. Isotope ratio measurements on two selenium species

Selenium Species	Amount (ng)	Determined Mean	$^{74}\text{Se}/^{78}\text{Se}$ RSD(%) [*]
SeO_3^{2-}	5	0.035	1.3
SeO_4^{2-}	5	0.033	1.7
SeO_3^{2-}	25	0.036	0.3
SeO_4^{2-}	25	0.036	0.5

Accepted ratio: 0.038

^{*}n = 5 separate injections

Figure 1. Calibration curve for the SEC column. See text for conditions.

Figure 2. Separation of metalloproteins in human serum by SEC-DIN-ICP-MS. All seven chromatograms were obtained from a single injection. See text for conditions.

Figure 3. Separation of SeO_3^{2-} and SeO_4^{2-} by AC-DIN-ICP-MS. See text for conditions.

SUMMARY

The primary goal of this work has been the development of the DIN for coupling LC with ICP-MS to provide elemental speciation information. In brief, the advantages of using LC-DIN-ICP-MS (27,29,64) include little dead volume, low sample and solvent consumption, excellent plasma stability when nebulizing samples containing concentrated organic solvent, excellent absolute detection limits, excellent precision, superior chromatographic resolution, and reduced memory effects from memory-prone elements (e.g., Hg, I, and B).

Table I summarizes all the compounds, separation modes, and detection limits described in this dissertation. More than 20 compounds of As, Se, Hg, Pb, Sn, and metalloproteins were separated and measured. Some of these compounds were present in biological and environmental samples, such as human urine and human serum, where no extraction or sample preparation was used that might disturb speciation. Various separation modes including RP-IP, IC, and SEC have been employed. Detection limits obtainable were $\approx 0.5 - 10$ pg depending on the element of interest. These detection limits are superior by at least 2 orders of magnitude than those obtained previously by LC-ICP-MS with a conventional nebulizer.

We also demonstrated the ability to introduce solvents containing more than 20% of most organic modifiers into the plasma (27). In fact, a matrix containing about 20 - 30% methanol yields the best ion signal for many elements tested.

Table I. A summary of various separations studied in this dissertation

Compounds ^a	Separation Mode	Detection Limits (pg)
As (As ^{III} , As ^V , MMAA, DMAA)	RP-IP	0.5
Se (Se ^{IV} , Se ^{VI})	IC	15
Hg (Hg ⁺² , MeHg ⁺ , EtHg ⁺ , PhHg ⁺)	RP-IP	10
Pb (Pb ⁺² , (Me) ₃ Pb ⁺ , (Et) ₃ Pb ⁺)	RP-IP	0.2
Metalloproteins	SEC	
Sn (TMT, DMT, DET, MMT)	RP-IP	8

^aAs^{III} = arsenite, As^V = arsenate, MMAA = monomethylarsonic acid, DMAA = dimethylarsinic acid, Se^{IV} = selenite, Se^{VI} = selenate, Me = methyl, Et = ethyl, Ph = phenyl, TMT = trimethyltin, DMT = dimethyltin, DET = diethyltin, and MMT = monomethyltin.

Narrowing of the aerosol droplet size distribution with organic solvent composition probably caused this enhancement in ion signal (66). In addition, results from the spatially resolved measurements of size and velocity distributions of aerosol droplets from a DIN support the importance of generating aerosol with narrow distribution of droplet sizes in ICP-MS (66,67).

Several possible areas for future research with the DIN are conceivable. First, there is still room for improvement in the design and construction of the DIN. The thickness and inner diameter (i.d.) of the inner capillary, as well as the width of the annular gap between the inner capillary and the nebulizer tip, are important parameters for the production of fine aerosol droplets with the DIN. A systematic study on the effect of these parameters on the aerosol droplet size distribution warrants investigation. Furthermore, the possibility of measuring size distribution for aerosol particles inside the ICP by light scattering methods should be explored.

The possible benefits of a tapered DIN are being studied in our laboratory. Here, the tip of the fused silica capillary is carefully etched into a sharp point. Presumably, this design will change the nebulizer gas flow pattern to enhance the interaction of the nebulizer gas with the sample liquid for the production of finer aerosol droplets.

So far, only deactivated fused silica capillaries have been used to construct the DIN. However, capillaries that are chemically modified with C_8 or C_{18} are said to be more inert, which presumably will give even lower memory effects with memory-

prone elements (Hg, I, and B) and metalloproteins. Thus, better chromatographic resolution should be obtainable for these elements.

Second, problems related to the solvent derived interferences in ICP-MS with the DIN need to be addressed (68-70). The DIN introduces more solvent than with other nebulizers; polyatomic ions such as ArO^+ , ArOH^+ , ArC^+ , ArN^+ , and ArCl^+ are substantial. Several possible ways to reduce polyatomic ions are 1) the combination of the DIN with a small spray chamber with proper desolvation by cooling (71), 2) the addition of xenon to the make-up gas of the DIN (72), and 3) the combination of the DIN with a new ICP-MS device equipped with an offset ion lens (73,74). These experiments are under way in our laboratory.

Finally, more speciation experiments can be investigated with the LC-DIN-ICP-MS system. Some of these speciation experiments involve 1) chromium in organic samples (75), 2) As in urine or sea-water (32), 3) packed microcolumn ($0.2 < \text{i.d.} < 0.5 \text{ mm}$), packed microcapillary column ($0.05 < \text{i.d.} < 0.2 \text{ mm}$), and open-tubular column ($0.01 < \text{i.d.} < 0.06 \text{ mm}$) (76), 4) solvent gradient to achieve faster and better chromatographic separation (27), and 5) on-line standard addition (77) or isotope dilution (78) for quantitation of chromatographic separated species.

ADDITIONAL LITERATURE CITED

1. R.S. Houk, V.A. Fassel, G.D. Flesch, H.J. Svec, A.L. Gray and C.E. Taylor, *Anal. Chem.*, 52 (1980) 2283.
2. D.W. Koppenaal, *Anal. Chem.*, 64 (1992) 320R.
3. K.E. Jarvis, A.L. Gray and R.S. Houk, *Handbook of Inductively Coupled Plasma Mass Spectrometry*, Blackie, London, 1992.
4. H.E. Taylor and J.R. Garbarino, in A. Montaser and D.W. Golightly (Eds.), *Inductively Coupled Plasma in Analytical Atomic Spectrometry*, 2nd Edition, VCH, New York, 1992, Ch. 14.
5. G. Holland and A.N. Eaton (Eds.), *Application of Plasma Source Mass Spectrometry*, Royal Soc. Chem., Cambridge, 1991.
6. K.E. Jarvis, A.L. Gray, J.G. Williams and I. Jarvis (Eds.), *Plasma Source Mass Spectrometry*, Royal Soc. Chem., Cambridge, 1990.
7. H.P. Longerich, G.A. Jenner, B.J. Fryer and S.E. Jackson, *Chem. Geol.*, 83 (1990) 105.
8. N. Imai, *Anal. Sci.*, 6 (1990) 389.
9. A.L. Gray, *Adv. Mass. Spectrom.*, 11B (1989) 1674.
10. D. Beauchemin, *Mikrochim. Acta*, 3 (1989) 273.
11. U. Siewers, *Mikrochim. Acta*, 3 (1989) 365.
12. D.C. Gregoire, *Prog. Anal. Spectrosc.*, 12 (1989) 432.

13. A.R. Date and A.L. Gray (Eds.), Applications of Inductively Coupled Plasma Mass Spectrometry, Blackie, London, 1989.
14. R.S. Houk, Anal. Chem. 58 (1986) 97A.
15. R.S. Houk and S.J. Jiang, in I.S. Krull (Ed.), Trace Metal Analysis and Speciation, Elsevier, Amsterdam, 1991, Ch. 5.
16. R.C. Elder, W.B. Jones and K. Tepperman, in P.C. Uden (Ed.), Element - Specific Chromatographic Detection by Atomic Emission Spectroscopy, ACS Symposium Series 479, Washington, DC, 1992, Ch. 18.
17. D.S. Braverman, J. Anal. At. Spectrom., 7 (1992) 43.
18. H. Suyani, D. Heitkemper, J. Creed and J. Caruso, Appl. Spectrosc. 43 (1989) 962.
19. D.S. Bushee, Analyst, 113 (1988) 1167.
20. D.S. Bushee, J.R. Moody and J.C. May, J. Anal. At. Spectrom., 4 (1989) 773.
21. J.J. Thompson and R.S. Houk, Anal. Chem., 58 (1986) 2541.
22. S.J. Jiang and R.S. Houk, Spectrochim. Acta, Part B, 43 (1988) 405.
23. D. Beauchemin, K.W.M. Siu, J.W. McLaren and S.S. Berman, J. Anal. At. Spectrom., 4 (1989) 285.
24. Y. Shibata and M. Morita, Anal. Chem., 61 (1989) 2116.
25. H. Suyani, J. Creed, T. Davidson and J. Caruso, J. Chromatogr. Sci., 27 (1989) 139.

26. A. Al-Rashdam, N.P. Vela, J.A. Caruso and D.T. Heitkemper, *J. Anal. At. Spectrom.*, 7 (1992) 551.
27. S.C.K. Shum, R. Neddersen and R.S. Houk, *Analyst*, 117 (1992) 571.
28. Z. Zhao, W.B. Jones, K. Tepperman, J.G. Dorsey and R.C. Elder, *J. Pharm. Biomed. Anal.*, 10 (1992) 279.
29. S.C.K. Shum, H.M. Pang and R.S. Houk, *Anal. Chem.*, 64 (1992) 2444.
30. B.S. Sheppard, J.A. Caruso, D.T. Heitkemper and K.A. Wolnik, *Analyst*, 117 (1992) 971.
31. B.S. Sheppard, W.-L. Shen, J.A. Caruso, D.T. Heitkemper and F.L. Fricke, *J. Anal. At. Spectrom.*, 5 (1990) 431.
32. D. Heitkemper, J. Creed, J. Caruso and F.L. Fricke, *J. Anal. At. Spectrom.*, 4 (1989) 279.
33. S. Suzuki, H. Tsuchihashi, K. Nakajima, A. Matsushita and T. Nagao, *J. Chromatogr.*, 437 (1988) 322.
34. K. Kawabata, Y. Kishi, O. Kawaguchi, Y. Watanabe and Y. Inoue, *Anal. Chem.*, 63 (1991) 2137.
35. D.W. Boomer, M.J. Powell and J. Hipfner, *Talanta*, 37 (1990) 127.
36. K. Takatera and T. Watanabe, *Anal. Sci.*, 8 (1992) 469.
37. L.M.W. Owen, H.M. Crews, R.C. Hutton and A. Walsh, *Analyst*, 117 (1992) 649.
38. B. Gercken and R.M. Barnes, *Anal. Chem.*, 63 (1991) 283.

39. K. Takatera and T. Watanabe, *Anal. Sci.*, 7 (1991) 695.
40. A.Z. Mason, S.D. Storms and K.D. Jenkins, *Anal. Biochem.*, 186 (1990) 187.
41. H.M. Crews, J.R. Dean, L. Ebdon and R.C. Massey, *Analyst*, 114 (1989) 895.
42. S.G. Matz, R.C. Elder and K. Tepperman, *J. Anal. At. Spectrom.*, 4 (1989) 767.
43. J.R. Dean, S. Munro, L. Ebdon, H.M. Crews and R.C. Massey, *J. Anal. At. Spectrom.*, 2 (1987) 607.
44. W.L. Shen, N.P. Vela, B.S. Sheppard and J.A. Caruso, *Anal. Chem.*, 63 (1991) 1491.
45. J.C. Van Loon, L.R. Alcock, W.H. Pinchin and J.B. French, *Spectrosc. Letters*, 19 (1986) 1125.
46. N.S. Chong and R.S. Houk, *Appl. Spectrosc.*, 41 (1987) 66.
47. R.J. Lewis and R.L. Tatken, *Registry of Toxic Effects of Chemical Substances*, Department of Health, Education and Welfare, Cincinnati, OH, 1978.
48. S.A. Peoples, in E.A. Woolson (Ed.), *Review of Arsenical Pesticides*, ACS Symposium Series 7, American Chemical Society, Washington, DC, 1974, pp 1-12.
49. M. Vahter, E. Marafante and L. Dencker, *Sci. Total Environ.*, 30 (1983) 197.
50. E. Marafante, M. Vahter and L. Dencker, *Sci. Total Environ.*, 34 (1984) 223.
51. E.A. Crecelius, *Environ. Health Perspect.*, 45 (1982) 165.

52. V. Foa, A. Colombi, M. Maroni, M. Buratti and G. Calzaferri, *Sci. Total Environ.*, 34 (1984) 241.
53. J.S. Edmonds, K.A. Francesconi, J.R. Cannon, C.L. Raston, B.W. Skelton and A.H. White, *Tetrahedron Lett.*, 18 (1977) 1543.
54. J.R. Cannon, J.S. Edmonds, K.A. Francesconi, C.L. Raston, J.B. Saunders, B.W. Skelton and A.H. White, *Aust. J. Chem.*, 34 (1981) 787.
55. J.F. Lawrence, P. Michalik, G. Tam and H.B.S. Conacher, *J. Agric. Food Chem.*, 34 (1986) 315.
56. H. Norin and A. Christakopoulos, *Chemosphere*, 11 (1982) 287.
57. D.J. Douglas, in A. Montaser and D.W. Golightly (Eds.), *Inductively Coupled Plasma in Analytical Atomic Spectrometry*, 2nd Edition, VCH, New York, 1992, Ch. 13.
58. T. Hasegawa, M. Umemoto, H. Haraguchi, C. Hsieh and A. Montaser, in A. Montaser and D.W. Golightly (Eds.), *Inductively Coupled Plasma in Analytical Atomic Spectrometry*, 2nd Edition, VCH, New York, 1992, Ch. 8.
59. J.S. Crain, F.G. Smith and R.S. Houk, *Spectrochim. Acta*, 45B (1990) 249.
60. P.E. Miller and M.B. Denton, *J. Chem. Ed.*, 63 (1986) 617.
61. R.S. Houk, in K.A. Gschneidner and L. Eyring (Eds.), *Handbook of Phys and Chem. of Rare Earths*, Vol. 13, North-Holland Physics Publishing, Amsterdam, 1989.
62. G. Zhu and R.F. Browner, *J. Anal. At. Spectrom.*, 3 (1988) 781.

63. R.C. Hutton and A.N. Eaton, *J. Anal. At. Spectrom.*, 2 (1987) 595.
64. R.S. Houk, S.C.K. Shum and D.R. Wiederin, *Anal. Chim. Acta*, 250 (1991) 61.
65. D.R. Wiederin, F.G. Smith and R.S. Houk, *Anal. Chem.*, 63 (1991) 219.
66. S.C.K. Shum, H.M. Pang, S.K. Johnson and R.S. Houk, *Appl. Spectrosc.*, 47 (1993) in press.
67. B. Etkin, J.B. French and R. Jong, Monodisperse Dried Microparticulate Injection, FACSS Conference, Philadelphia, PA, 1992, Paper No. 218.
68. S.H. Tan and G. Horlick, *Appl. Spectrosc.*, 40 (1986) 445.
69. D. Beauchemin, J.W. McLaren and S.S. Berman, *Spectrochim. Acta*, 42B (1987) 467.
70. S.H. Tan and G. Horlick, *J. Anal. At. Spectrom.*, 2 (1987) 745.
71. D.R. Wiederin, Ph.D. dissertation, Iowa State University, Ames, IA, 1991.
72. F.G. Smith, D.R. Wiederin and R.S. houk, *Anal. Chem.*, 63 (1991) 1458.
73. K. Hu, S. Clemons and R.S. Houk, *J. Am. Soc. Mass Spectrom.*, 4 (1993) 16.
74. K. Hu and R.S. Houk, *J. Am. Soc. Mass Spectrom.*, 4 (1993) 28.
75. R. Roehl and M.M. Alforque, *ICP Information Newsletter*, 16 (1991) 455.
76. M. Novotny, *Anal. Chem.*, 60 (1988) 500A.
77. D.R. Wiederin, R.E. Smyczek and R.S. Houk, *Anal. Chem.*, 63 (1991) 1626.
78. M. Viczián, A. Lásztity, X. Wang and R.M. Barnes, *J. Anal. At. Spectrom.*, 5 (1990) 125.

ACKNOWLEDGEMENTS

First, I would like to express my most sincere gratitude to Professor R. S. Houk for guidance, not only in this research but in matters both personal and professional. His wide latitude, tolerance, confidence, and support are greatly appreciated. In addition, he taught me a great deal about scientific writing.

Also helpful were the many hours of discussion with Professor James S. Fritz, Dr. Ho-Ming Pang, Dr. Daniel Wiederin, Dr. Fred Smith, and Dr. Ke Hu for which I am privileged to express my gratitude.

I also like to acknowledge Luis Alves, Hongsen Niu, Rocky Warren, Scott Clemons, Xiaoshan Chen, Shen Luan, Tonya Bricker, Lloyd Allen, Steve Johnson, Bill Wheeler, and Robert Neddersen for their support, cooperation, and friendship.

The donation of time and talent by my wife, Swee Chin, for the preparation of the final draft is appreciated. Only inadequately can I express my gratitude for her patience and understanding throughout the period of this endeavor. I thank my children, Benjamin and Jennifer, for their supports and cheers when I was down.

I like to thank the members of the Ames Seventh-Day Adventist Church for providing us, in the last four and a half years, the opportunities to serve the Lord, the church body, and the Ames community. Special thanks go to Charles and Nancy Dye; Robb and Christine Long; Bruce, Linda, Andrew and Megan Wilkinson; David, Linda, Rob and Jessie Anderson; Craig, Amy and Sam Abel; Bill and Treva

Martsching; Cody and Tracie Cameron; Nyla, Camden, Karen and Lance Hodges; Dave and Jill Koch; Gertrude, Donald and Nancy Miller; Kevin Voss; Debbie Lien; D'Joane McCorkle; Edward Osei; Buena Boyer; Dale, Maria, Daniel and Shanna Bivens for their prayers, supports, and friendship.

I also like to thank the Hong Kong gang, Che-Ting Chan, Stella Luk, Ho-Ming Pang, Kit-Sum Wong, Nai-Ho Cheung, Po-Lin Tang, Frankie Lam, Grace Ho, and Man-Kit Ho, for their friendship and all the wonderful parties and social gatherings that my wife and I were privileged to join.

I wish to express my appreciation to Delavan Inc. for the use of the PDPA apparatus. Special thanks go to Dr. Chien-Pei Mao for his useful comments and efforts in arranging and conducting the experiments. I thank CETAC Technologies and Serasep, Inc. for supplying some of the columns used in this work.

The Phillips Petroleum Graduate Research Fellowship awarded to me is gratefully acknowledged. This work was performed at the Ames Laboratory under contract No. W-7405-ENG-82 with the U. S. Department of Energy. The United States government has assigned DOE report No. IS-T 1644 to this dissertation.

Lastly, thanks go to my parents, Tung-Sang Shum and Wai-Lan Wong, and my in-laws, Yew-Seng Wong and Mei-Chan Lok, for their countless support and encouragement in my education. Regrettably my mom did not live long enough to share the joy with us.

DATE

FILMED

2 / 8 / 94

END