ANL/APS/LS-187
Rev. 1

APS BEAMLINE STANDARD COMPONENTS HANDBOOK

ANL/APS/LS--187-Rev.1
DE93 008925

Version 1.3

February 1993

Advanced Photon Source

Argonne National Laboratory, Argonne, Illinois 60439
operated by The University of Chicago
for the U.S. Department of Energy under Contract W-31-109-Eng-38

MASTER

DISTRIBUTION OF THIS DOCUMENT IS UNLIMITED

Argonne National Laboratory, with facilities in the states of Illinois and Idaho, is owned by the United States government, and operated by The University of Chicago under the provisions of a contract with the Department of Energy.

DISCLAIMER

This report was prepared as an account of work sponsored by an agency of the United States Government. Neither the United States Government nor any agency thereof, nor any of their employees, makes any warranty, express or implied, or assumes any legal liability or responsibility for the accuracy, completeness, or usefulness of any information, apparatus, product, or process disclosed, or represents that its use would not infringe privately owned rights. Reference herein to any specific commercial product, process, or service by trade name, trademark, manufacturer, or otherwise, does not necessarily constitute or imply its endorsement, recommendation, or favoring by the United States Government or any agency thereof. The views and opinions of authors expressed herein do not necessarily state or reflect those of the United States Government or any agency thereof.

Reproduced from the best available copy.

Available to DOE and DOE contractors from the
Office of Scientific and Technical Information

P.O. Box 62

Oak Ridge, TN 37831

Prices available from (615) 576-8401, FTS 626-8401

Available to the public from the
National Technical Information Service

U.S. Department of Commerce

5285 Port Royal Road

Springfield, VA 22161

Distribution Category: Atomic,
Molecular, and Chemical Physics
(UC-411)

ARGONNE NATIONAL LABORATORY
9700 South Cass Avenue
Argonne, Illinois 60439

ANL/APS/LS-187
Rev. 1

APS BEAMLINE STANDARD COMPONENTS HANDBOOK

Version 1.3

Ulrich Hahn, Deming Shu, and Tuncer M. Kuzay
Experimental Facilities Division

February 1993

work sponsored by

U.S. DEPARTMENT OF ENERGY
Office of Energy Research

ACKNOWLEDGMENTS

We would like to thank the Insertion Device, Optics, and Engineering and Construction Groups of the Experimental Facilities Division for their work on the design of the APS beamline "standard components." Specifically significant input to this document from Ercan Alp, Juan Barraza, Christa Brite, Joseph Chang, Howard Czerwinski, Steve Davey, Dean Haeffner, Nisy Ipe, Harold Moe, Mike Muscia, Susan Picologlou, and Thomas Sanchez is greatly appreciated. We thank the members of the Beamline Standardization and Modularization Committee for the time and effort they have contributed to the process of identifying the specifying the standard components. In addition, we would like to thank various users and CAT representatives who have contributed ideas and suggestions toward this effort.

Introduction

1. Vacuum Specifications

- 1.1 APS beam transport - windowless
- 1.1.1 APS beam transport - mirrors
- 1.2 APS beam transport - Be window

2. Beamline Shielding

- 2.0.1. Sources of radiation for beamline shielding
- 2.0.2. Shielding metrology
- 2.0.3. First optical enclosure (FOE) - white beam hutch
- 2.0.4. Pink beam hutch
- 2.0.5. Monochromatic beam hutch
- 2.0.6. White beam transport
- 2.0.7. Pink beam transport
- 2.0.8. Monochromatic beam transport
- 2.1. Typical first optical enclosure (FOE) - white beam hutch
- 2.2. Typical pink beam hutch
- 2.3. Typical monochromatic beam hutch
- 2.4. Typical white beam transport
- 2.5. Typical pink beam transport
- 2.6. Typical monochromatic beam transport
- 2.7. APS FOE and experimental hutches (added later)
- 2.8. APS shielded monochromatic transport and pump station (added later)

3. Filters

- 3.1. Filter assembly
- 3.2. Filter material

4. Windows

- 4.1. Be - windows
- 4.2. C - windows
- 4.3. Differential pump

5. Slits

- 5.1.1. White beam slits - insertion device
- 5.1.2. White beam slits - bending magnet
- 5.2. Monochromatic beam slits

6. Beam Shutters and Beam Stops

- 6.1. Monochromatic beam shutter - insertion device
- 6.2. White beam stop with integral shutter - insertion device
- 6.4. White beam stop with integral shutter - bending magnet (added later)

7. Mirror Chambers

8. Beam Position Monitors

- 8.1. White beam position monitor
- 8.2. Monochromatic beam position monitor

9. Supports

- 9.1. Kinematic mount support tables
  - 9.1.1. Economic style support table
    - 9.1.1.1. Vertical motion cone
    - 9.1.1.2. Vertical motion V
    - 9.1.1.3. Vertical motion flat
    - 9.1.1.4. Horizontal motion
  - 9.1.2. Standard style support table
    - 9.1.2.1. Standard vertical stage
    - 9.1.2.2. Standard horizontal stages
    - 9.1.2.3. Standard spherical coupling
    - 9.1.2.4. Standard V coupling
    - 9.1.2.5. Standard Flat coupling
  - 9.1.3. Precision style support table
    - 9.1.3.1. Precision vertical stage - "A"
    - 9.1.3.2. Precision vertical stage - "B"
    - 9.1.3.3. Precision vertical stage - "C"
    - 9.1.3.4. Precision horizontal stage - "A"
    - 9.1.3.5. Precision horizontal stage - "B"
    - 9.1.3.6. Precision horizontal stage - "C"
- 9.2. Beam position monitor support
  - 9.2.1. Beam position monitor vertical stage
  - 9.2.2. Beam position monitor horizontal stage
  - 9.2.3. Beam position monitor rotation stage
  - 9.2.4. Beam position monitor support column
  - 9.2.5. Beam position monitor kinematic mount cone
  - 9.2.6. Beam position monitor kinematic mount flat
  - 9.2.7. Beam position monitor kinematic mount V
- 9.3. Ion pump supports
  - 9.3.1. Inline 60 l/s ion pump support
- 9.4. Beam pipe supports
  - 9.4.1. 6" beam pipe support

10. Stepping Motor Drives

- 10.1. Stepping motors
- 10.2. Drivers/Controllers

11. Interlock - Interfaces

- 11.1. Switches
- 11.2. Solenoids
- 11.3. Interfaces - cables and connectors
  - 11.3.1. Ion pumps with power supplies
  - 11.3.2. Ion gauges with power supplies
  - 11.3.3. Vacuum valves

12. Actuators

- 12.1. Heavy load with stepping motor drive
- 12.2. Heavy load with pneumatic drive
- 12.3. Light load with stepping motor drive
- 12.4. Light load with pneumatic drive

13. Bellows

- 13.1. Welded bellows
- 13.2. Formed bellows

INTRODUCTION

It is expected that the members of the APS Collaborative Access Team (CAT) would like to concentrate their effort on designing specialized equipment related to their scientific programs rather than on routine or standard beamline components. Thus, an effort is made at the APS to identify standard and modular components of APS beamlines. Identifying standard components is a non-trivial task because these components should support diverse beamline objectives. To assist with this effort, the APS has obtained advice and help from a Beamline Standardization and Modularization Committee consisting of the following experts in beamline design, construction, and operation:

Richard Boyce	SSRL
Richard Hewitt	EXXON and CMCCAT
Tunch M. Kuzay (Chair)	APS
Richard Levesque	LLNL
Ed Melcer	LBL/APS
Dennis M. Mills	APS and SRI CAT
Tom Oversluizen	NSLS
Wilfried Schildkamp	University of Chicago and CARS CAT

A group of CAT representatives have also joined in this effort to lend their views on the subject.

Based on the charge given to this committee by Gopal Shenoy, Director of the Experimental Facilities of the Advanced Photon Source on January 22, 1990 (see Appendix A), the committee identified the following tasks:

- 1) to identify standard and modular components of an APS insertion device beamline,
- 2) to insure safety and quality in the beamline design concepts,
- 3) to develop conceptual and engineering designs for the components,
- 4) to review the designs,
- 5) to develop prototypes,
- 6) to disseminate the information to the user community and vendors.

(Tasks 2, 5 and 6 are carried out exclusively by the staff of the APS.)

In addition, the guidelines to this committee called for the following considerations:

- modular design of as many components as possible,
- vendor production design to reduce cost to users,
- safety of personnel and equipment,

- ALARA* design objective,
- Experiment Hall should be a Class IV restricted area (500 mR per year)
- engineered safety included in all the components to avoid inadvertent actions leading to hazards
- interlocks with adequate redundancies as well as visual and audible alarms to prevent human errors,
- spares for standard components will be available in the APS stockroom.

With these guidelines in mind, the staff of the Experimental Facilities Division identified various components thought to be standard items for beamlines, regardless of the specific scientific objective of a particular beamline. A generic beamline layout (see next page) formed the basis for this identification. This layout is based on a double-crystal (or multi-layer) monochromator as the first optical element, with the possibility of other elements to follow. Preliminary engineering designs were then made of the identified standard components. The Beamline Standardization and Modularization Committee has reviewed these designs and provided very useful input regarding the specifications of these components during many meetings.

This *Handbook* in its current version (1.3) contains descriptions, specifications, and preliminary engineering design drawings for many of the standard components. The design status and schedules have been provided wherever possible. In the near future, the APS plans to update engineering drawings of identified standard beamline components and complete the *Handbook*. The completed version of this *Handbook* will become available to both the CATs and potential vendors. Use of standard components should result in major cost reductions for CATs in the areas of beamline design and construction.

Because of the involved nature of the job at hand, we encourage CAT Directors to inform us of their specific needs as they progress in completing their beamline preliminary designs. The engineering drawings, in their current state of design, for the identified standard components are now available on the APS Design Exchange.

---

* ALARA ("as low as reasonable achievable") refers to DOE's policy of establishing a program for minimizing radiation exposure to personnel and equipment.

1. Vacuum specifications

- 1.1 APS beam transport - windowless
- 1.1.1 APS beam transport - mirrors
- 1.2 APS beam transport - Be window

1. Vacuum Specifications

The Storage Ring defines the vacuum conditions in the front end, which is the link to the beamlines. A pressure of 1×10^{-9} torr or better will be maintained in the Storage Ring during operation.

The pressure in the beamlines will strongly depend on the components that will be used to transport the beam to the experiment. The interface between beamline and front end is the front end valve (FEV) in the first optics enclosure (FOE). The pressure in the beamline immediately upstream of this valve must be $< 1 \times 10^{-8}$ torr in order to obtain permission to open the valve. The pressure at the safety shutters (SS2) in the front end must be $< 6 \times 10^{-9}$ torr. Figure 1 illustrates the situation. The trigger gauge downstream from the front end valve will have two set points. When the pressure exceeds 1×10^{-8} torr, the slow valve (SV) will close. If there is an accidental venting and if the pressure rises in less than 1 ms to more than 1×10^{-5} torr, the fast acting valve (FV) will be activated followed by the slow valve. Figure 1 also shows the setup for windowless operation with two differential pumps (DP), which allow a higher pressure limit for the beamline operation to be $< 1 \times 10^{-6}$ torr (instead of being $< 1 \times 10^{-8}$ torr in the absence of the differential pumps).

1.1 APS Beam Transport - Windowless

Beamlines under vacuum have to maintain a pressure $< 1 \times 10^{-6}$ torr. To have a windowless connection between the beamline and the front end, a differential pump will be needed. This achieves the needed pressure at the front end valve ($< 1 \times 10^{-8}$ torr) and an additional delay in shock wave propagation (in case of accidental venting) to protect the front end. For pumping, ion pumps are recommended. The vacuum of the beamline must be free of hydrocarbons. This will be verified (by the APS) upstream from the front end valve through residual gas analysis. The result of this analysis must be that the masses above 38 have a total pressure of less than 5×10^{-11} torr.

1.1.1 APS Beam Transport - Mirrors

All beam transports that are connected without windows to mirror chambers must be particularly free from hydrocarbons to avoid carbon contamination of the mirrors. It should be verified by the CAT that hydrocarbons contribute less than 1×10^{-3} to the total pressure of these beam transport components. This can be verified by the mass spectrum of a single component. The sum of the masses > 38 has to be less than 1×10^{-3} of the total pressure.

1.2 APS Beam Transport - Be Window

Beam transport that is separated from the front end by a Be window has to be evacuated to a pressure $< 1 \times 10^{-6}$ torr. To avoid carbon contamination of the window, oil free pumping of beamline components is required.

If the window terminates the vacuum part of the beamline, special measures must be taken to prevent contamination from the atmospheric side. Exposing the windows to air with the beam on will not be allowed. If the beamline design does not permit vacuum on one side of the window, we suggest one of the following measure be included: a short clean hydrocarbon-free He buffer (terminated by an Al foil) or a protective coating on the window. The mechanical and thermal design of the windows is discussed in section 4.

Figure 1

2.0. Beamline Shielding

- 2.0.1 Sources of Radiation for Beamline Shielding
- 2.0.2 Shielding Metrology
- 2.0.3 First Optics Enclosure (FOE) - White Beam Hutch
- 2.0.4 Pink Beam Hutch
- 2.0.5 Monochromatic Beam Hutch
- 2.0.6 White Beam Transport
- 2.0.7 Pink Beam Transport
- 2.0.8 Monochromatic Beam Transport
- 2.1. Typical First Optics Enclosure (FOE) - White Beam Hutch
- 2.2. Typical Pink Beam Hutch
- 2.3. Typical Monochromatic Beam Hutch
- 2.4. Typical White Beam Transport
- 2.5. Typical Pink Beam Transport
- 2.6. Typical Monochromatic Beam Transport
- 2.7. APS FOE and Experimental Hutches (section to be added later)
- 2.8. APS Shielded Monochromatic Transport and Pump Station (section to be added later)

Technical Analysis

APS has performed technical analysis of the shielding required under a variety of unusual operation conditions. These are summarized in sections 2.1 - 2.6

Design Stage and Schedule

The recommended shielding for various configurations of operation are now being evaluated from an engineering standpoint, and will be added to section 2.7 and 2.8. The results will lead to final shielding specifications for the hutches and the transports. This work is planned to be completed by May 1993.

2. Beamline Shielding

The shielding for the first optics enclosure (FOE), white beam, pink beam (reflected white beam), and monochromatic beam stations and transport lines shall be designed to keep the annual integrated dose equivalent at the shielding surface in occupied areas to less than 500 mrem. This design criteria is based on D.O.E. requirements (Radiation Control Manual, June 1982). For a 2000 hour operating year, this translates to a design criterion of 0.25 mrem/h. However, because it is highly unlikely that any user will be on the experimental floor at the shield surface for 2000 hours per year, and coupled with the fact that calculations were performed with conservative assumptions, annual dose equivalents received by individuals on the experimental floor should be well below 500 mrem.

2.01 Sources of Radiation for Beamline Shielding

Three sources of radiation should be considered for beamline shielding. They are:

- 1) gas bremsstrahlung
- 2) neutrons produced by gas bremsstrahlung
- 3) synchrotron radiation

Gas bremsstrahlung is produced by the interaction of the primary electron/positron beam with the residual gas in the Storage Ring vacuum chamber and is a major source of beam losses in storage rings. The interaction of the electron/positron with the gas molecules or ions takes place all around the Storage Ring, however, the photon intensity in the straight sections is very high. This is because in each interaction, the bremsstrahlung is produced in a narrow cone (characteristic emission angle = $1/\gamma = 73 \mu\text{radians}$ for the APS) around the same direction and each contribution adds up resulting in a narrow monodirectional photon beam of high intensity. This narrow beam of bremsstrahlung travels down the synchrotron beamline and may be of major concern in the shielding design of the beamlines. The problem is aggravated with poor vacuum conditions in the Storage Ring.

The photons may be energetic enough to propagate an electromagnetic shower in any thick targets that they strike, or they can scatter off thin targets. In either case, this results in contributions to dose outside the transverse shielding of the beam transport, the optical enclosures/experimental stations, and the back walls of these stations. Because these photons are much more energetic than synchrotron radiation, they require far greater thicknesses of shielding material. For example, about 2 inches of lead will reduce the gas bremsstrahlung by a factor of 10, whereas it takes only about 3 mm of lead to reduce the 90° scattered synchrotron radiation (section 2.1) by the same factor. Monte-Carlo calculations are currently underway to determine the impact of gas bremsstrahlung on the shielding design of the beamlines. If it is determined that gas bremsstrahlung is a major concern in the shielding design of beamlines, additional local shielding around the target/scatterer will be specified.

In addition, if the target with which the gas bremsstrahlung interacts is sufficiently thick and the photons have energies above the photo-neutron production cross section, neutrons can also be produced. Analysis of neutron production is also being carried out. A foot of concrete will reduce the neutron and bremsstrahlung dose equivalent rate by a factor of 10 and 5, respectively.

Synchrotron radiation traveling down the beamline can scatter off any limiting aperture or any device that it strikes (e.g., monochromators, mirrors, masks, shutters, etc.) Thus the synchrotron beamline needs to be shielded against the scattered radiation.

The next section describes the shielding methodology. It must be pointed out that beam stops will be required in the forward direction inside the enclosures/experimental stations, however, they will not be addressed in this document. Other aspects of beamline shielding such as

collimators for bremsstrahlung, treatment of joints, doors, feedthroughs for cables and pipes are also not addressed.

2.02

The "Photon" program^{1,2} developed at the National Synchrotron Light Source was used to determine the shielding for the beamlines. This is a computer program that calculates dose in the following sequence:

- 1) calculation of photon flux as a function of energy and vertical opening angle of the synchrotron beam,
- 2) attenuation by filters,
- 3) a scattering process,
- 4) conversion from photon flux to dose.

Compton scattering is the primary mechanism for scattering at the large angles necessary to strike a shielding wall at or near normal incidence (where the effective elastically shielding thickness is a minimum). Low energy photons, which are more efficiently scattered through large angles, will not penetrate the shielding walls and therefore will not contribute to the dose outside the shielding.

In "Photon," the scatterer is assumed to be an isotropic point source and the total angularly integrated Compton cross section is used to calculate the scatter of the incident synchrotron spectrum from a target. In reality, there will be no point source of scatter in the beamlines, hence, "Photon" will overestimate the scattered intensities compared to that from an extended source.

"Photon" does not take into account the polarization dependence of the scattering. This will overestimate the scattered intensities in the horizontal plane while underestimating the scattered intensities in the vertical plane.

No consideration is given in the "Photon" program for electron-photon beam interactions, finite source size, or horizontal beam distribution. It must be pointed out that "Photon" uses the narrow beam attenuation coefficient for determining the dose outside the shielding. This can

---

¹D. Chapman, N. Gmur, N. Lazarz, and W. Thomlinson, "Photon: A Program for Synchrotron Radiation Dose Calculations," Nucl. Instr. and Meth. A266 191-194 (1988).

²E. Brauer and W. Thomlinson, "Experimental Verification of "Photon." A Computer Program for use in X-Ray Shielding," BNL 39541, Brookhaven National Laboratory, New York.

underestimate the dose because there will be build up from scattered photons. No build-up factor is used in the program. However, experimental data indicate that "Photon" overestimates the dose in the polarization plane.³

Sections 2.1 to 2.6 outline the shielding parameters and specifications for the beamlines. A maximum positron energy of 7 GeV, maximum current of 300 mA, and a maximum vertical divergence of $4/\gamma$ were used for all calculations. Because the beamlines have to be shielded against scattered synchrotron radiation, "Photon" was run several times with different target materials (air, silicon, copper, and lead) and different target thicknesses to determine the material and thickness of the target that resulted in the highest scattered dose. Based on these runs, 30 cm of copper was chosen as the optimum target. Because enclosures/experimental stations will have optical elements such as mirrors, monochromators, safety shutters, and beam stops, all calculations for transverse shielding were based on using 30 cm of copper as the target (sections 2.1 - 2.3).

For beam transport pipes, a different approach is used. If ray traces indicate that the synchrotron beam could strike a high target (e.g., mask) or the beam pipe itself, during normal operating conditions or cases in of missteering, copper should be chosen as the optimum target (sections 2.4 - 2.6). If ray traces indicate that the synchrotron radiation could never strike a high-Z target in the transport pipe (this could be achieved with upstream collimation), 30 cm of air should be chosen as the optimum target. This would mean that no scattered radiation should be measured during normal operating conditions when the beam transport pipe is under good vacuum. However, if the vacuum in the pipe should be lost, then the synchrotron radiation could scatter off the air molecules. For the ID beamlines, calculations are done only for wigglers, which would result in more stringent specifications than for undulators. For wigglers, a critical energy of 32.6 keV, 32 poles, and a horizontal divergence of 2 mrad was used. For bending magnets, a critical energy of 19.5 keV and a horizontal divergence of 6 mrad was used. In addition, a safety factor of 2 for dose was incorporated into all the calculations.

The shielding specifications are based on model calculations and do not represent the final specifications for construction. After assessing the engineering needs, the APS will provide the final specifications.

The shielding specification in sections 2.1 to 2.6 can be used as long as the parameters used lie within the envelope of the specified parameters. The dose equivalent rates scale linearly with current, number of poles, and horizontal divergence. The dose equivalent rates follow an inverse square law with distance.

---

³ E. Brauer, "Health Physics Measurements Around the New F2 Wiggler Station at CHESS (Cornell University) ESRF Report.

The specifications of steel thicknesses for shielding are based on recommendations (for structural support) by a commercial supplier. The shielding effect of steel has also been taken into account.

2.03 First Optics Enclosure/White Beam Hatches

The shielding parameters are specified in section 2.1. The copper scatterer is assumed to be at a distance of 29.45 m from the source. The calculations do not depend on this distance as long as all the radiation is taken into account by integrating over the full vertical extent of the beam. The energy range of the incident spectrum is 1-800 keV. The first and second tables specify the shielding requirements for the ID and bending magnet, respectively. The first column in each table specifies the wall under consideration. The second column shows the distance between the scatterer and the wall. The third, fourth, and fifth columns show the thickness of steel, lead, and steel required for each wall. The sixth column shows the dose equivalent rate (DER) at the shield surface for which the hatch wall has been shielded. The last column shows the effective tenth value layer (TVL) for lead, i.e., the thickness required to reduce the DER by a factor of 10. The lateral wall has been shielded for 0.25 mrem/h because a high occupancy can be expected outside this wall.

The roof has been shielded for 2.5 mrem/h since it is not expected to be occupied. Conservative calculations for skyshine (i.e., x-rays reflected back from the atmosphere to the experimental floor) indicate that the dose equivalent rates will be at least two orders of magnitude lower. So, the contribution from skyshine to the experimental floor will be less than 0.025 mrem/h.

The back walls of the hatch are thicker than the lateral walls by a tenth-value layer, because "Photon" underestimates the scattered intensities in the vertical plane and to account for the forward scattered photons, which are more energetic than the laterally scattered ones. Inherent in this approach is the assumption that beam stops will be designed to stop both the forward directed gas bremsstrahlung and forward directed synchrotron radiation. Provisions should be made for additional shielding of the back walls, if beam stops do not intercept all the forward scattered radiation.

2.04 Pink Beam Hatches

It is difficult to generalize shielding for pink beam hatches, because the type of mirror used and the grazing angle of incidence can vary. Hence, the shielding specifications for pink beam hatches (section 2.2) has been given only as an example and is not to be used as a general case. In this example, a platinum mirror with a grazing angle of incidence of 0.15° and a cut-off energy of 32 keV, (reflectivity = 0.4) was used. However, for shielding calculations reflectivities of 1.0, 0.0164, and 0.0017 were used for photon energies in the range 1-50 keV, 50-80 keV, and 80-100 keV, respectively. Hence, the results are conservative.

2.05 Monochromatic Beam Hutches

The determination of shielding for monochromatic beam hutches is not straight forward because any energy and bandwidth can be chosen. In addition, one must shield for harmonics as well. The shielding specifications in section 2.3 assume a monochromatic beam with a 0.1% bandwidth in the energy range of 1-300 keV. The photons in the energy range of 200 - 300 keV contribute the highest dose, hence a photon energy in this range was chosen with a 0.1% bandwidth, and the shielding was determined. A half-value layer (HVL) of lead, that is, the thickness required to reduce the dose by a factor of 2, was added to the thickness determined earlier to account for the harmonics. These new values are reported in column 4.

2.06 White Beam Transport

As mentioned in section 2.06, the calculations were done for two scatterers, 30 cm of copper and 30 cm of air. The beam pipe diameter is four inches (section 2-4). If ray traces indicate that the beam may strike a high-Z target in the beam pipe, the shielding in the second row of the tables should be used, because this will limit the dose equivalent rate to 0.25 mrem/h. If ray traces indicate that only during missteering could the synchrotron beam could strike a high-Z target in the beam pipe, then the shielding in the third row should be used. A dose equivalent rate of 2.5 mrem/h at the shield surface can be tolerated because it will last only during missteering and not persist for 2000 hours.

If ray traces indicate that the synchrotron beam will not strike any high-Z material in the beam pipe, then technically no shielding is required for the beam pipe. However, it is possible to lose the vacuum in the beam pipe, and, if this happens, the synchrotron radiation could scatter off air molecules. In this case, the shielding in row 5 should be used.

2.07 Pink Beam Transport

A similar calculation was carried out for pink beam transport (section 2.5) was done for white beam transport. The pink beam transport shielding specifications are given only as an example and should not be used as a general case.

2.08 Monochromatic Beam Transport

A similar calculation was carried out for monochromatic beam transport (section 2.6) as for white beam transport.

2.1. Typical First Optical Enclosure (FOE) - White Beam Hutch

Shielding Parameters

The following parameters were used to determine the shielding for the walls and the roof of the FOE and the white beam stations at the ID beamlines.

Positron energy	7.0 GeV	
Positron current	300 mA	
Photon energy range	1 - 800 keV	
Vertical divergence	$\Psi = \pm 4/\gamma$	
Scatterer	300 mm Cu at a distance of 29.45 m from the source	
	<i>Insertion Device (ID)</i>	<i>Bending Magnet (BM)</i>
Critical energy	32.6 keV	19.5 keV
Number of poles	32	1
Horizontal divergence	2 mrad	6 mrad

Specifications

Insertion Device (ID)

Wall	Distance ¹ (m)	Shielding Material			DER ² (mrem/h)	TVL ³ (mm)
		steel (mm)	+lead (mm)	+steel (mm)		
Lateral	1.0	6.35	12.8	6.35	0.25	3.1
Roof	1.5	3.2	9.9	3.2	2.5	2.7
Back	1.0	6.35	15.9	6.35		

Bending Magnet (BM)

Wall	Distance (m)	Shielding Material			DER (mrem/h)	TVL (mm)
		steel (mm)	+lead (mm)	+steel (mm)		
Lateral	1.0	3.2	5.3	3.2	0.25	1.3
Roof	1.5	3.2	3.7	3.2	2.5	1.1
Back	1.0	3.2	6.6	3.2		

¹ Distance from scatterer to wall

² DER = Dose Equivalent Rate at the shielding surface

³ TVL = Tenth Value Layer (thickness of Pb required to reduce the dose equivalent rate by a factor of ten)

2.2. Typical Pink Beam Hutch

Shielding Parameters

The following parameters were used to determine the shielding for the walls and the roof of the pink beam stations at beamlines with reflected beams.

Positron energy	7.0 GeV
Positron current	300 mA
Photon energy range	1 - 50 keV (R=1), 50 - 80 keV (R=0.0164), 80 - 100 keV (R=0.0017)
	R = reflectivity of the mirror
Vertical divergence	$\Psi = \pm 4/\gamma$
Scatterer	300 mm Cu
	Insertion Device (ID)
Critical energy	32.6 keV
Number of poles	32
Horizontal divergence	2 mrad

Specification

Insertion Device (ID)

Wall	Distance ¹ (m)	Shielding Material			DER ² (mrem/h)	TVL ³ (mm)
		steel (mm)	+lead (mm)	+steel (mm)		
Lateral	1.0	3.2	4.1	3.2	0.25	1.0
Roof	1.5	3.2	2.9	3.2	2.5	0.9
Back	1.0	3.2	5.1	3.2		

¹ Distance from scatterer to wall

² DER = Dose Equivalent Rate at the shielding surface

³ TVL = Tenth Value Layer (thickness of Pb required to reduce the dose equivalent rate by a factor of ten)

2.3. Typical Monochromatic Beam Hutch

Shielding Parameters

The following parameters were used to determine the shielding for the walls and the roof of the monochromatic beam stations at beamlines with monochromatic beams.

Positron energy	7.0 GeV	
Positron current	300 mA	
Photon energy range	1 - 300 keV (0.1% bandwidth)	
Vertical divergence	$\Psi = \pm 4/\gamma$	
Scatterer	300 mm Cu	
	<i>Insertion Device (ID)</i>	<i>Bending Magnet (BM)</i>
Critical energy	32.6 keV	19.5 keV
Number of poles	32	1
Horizontal divergence	2 mrad	6 mrad

Specifications

Insertion Device (ID)

Wall	Distance ¹ (m)	Shielding Material			DER ² (mrem/h)	TVL ³ (mm)
		steel (mm)	+lead (mm)	+steel (mm)		
Front + Lateral	1.0	3.2	8.9	3.2	0.25	2.7
Roof	1.5	3.2	4.9	3.2	2.5	2.9
Back	1.0	3.2	11.6	3.2		

Bending Magnet (BM)

Wall	Distance (m)	Shielding Material			DER (mrem/h)	TVL (mm)
		steel (mm)	+lead (mm)	+steel (mm)		
Front + Lateral	1.0	3.2	3.5	3.2	0.25	1.1
Roof	1.5	3.2	2.0	3.2	2.5	1.0
Back	1.0	3.2	4.6	3.2		

¹ Distance from scatterer to wall

² DER = Dose Equivalent Rate at the shielding surface

³ TVL = Tenth Value Layer (thickness of Pb required to reduce the dose equivalent rate by a factor of ten)

2.4. Typical White Beam Transport

Shielding Parameters

The following parameters were used to determine the shielding for the white beam transport lines.

Positron energy	7.0 GeV	
Positron current	300 mA	
Photon energy range	1 - 800 keV	
Vertical divergence	$\Psi = \pm 4/\gamma$	
Ø of the beam pipe	4"	
	<i>Insertion Device (ID)</i>	<i>Bending Magnet (BM)</i>
Critical energy	32.6 keV	19.5 keV
Number of poles	32	1
Horizontal divergence	2 mrad	6 mrad

Specifications

Insertion Device (ID)

Scatterer	Shielding Material			DER (mrem/h)
	steel (mm)	+lead (mm)	+steel (mm)	
300 mm Cu	6.35	21.2	6.35	0.25
300 mm Cu	6.35	18.0	6.35	2.5
300 mm air	6.35	13.2	6.35	0.25
300 mm air	6.35	10.6	6.35	2.5

Bending Magnet (BM)

Scatterer	Shielding Material			DER (mrem/h)
	steel (mm)	+lead (mm)	+steel (mm)	
300 mm Cu	3.2	9.3	3.2	0.25
300 mm Cu	3.2	7.6	3.2	2.5
300 mm air	3.2	6.0	3.2	0.25
300 mm air	3.2	4.9	3.2	2.5

2.5. Typical Pink Beam Transport

Shielding Parameters

The following parameters were used to determine the shielding for the walls of the pink beam transport lines (beamlines with reflected beams).

Positron energy	7.0 GeV
Positron current	300 mA
Photon energy range	1 - 50 keV (R=1), 50 - 80 keV (R=0.0164), 80 -100 keV (R=0.0017) R = reflectivity of the mirror
Vertical divergence	$\Psi = \pm 4/\gamma$
Ø of the beam pipe	4"

Insertion Device (ID)

Critical energy	32.6 keV
Number of poles	32
Horizontal divergence	2 mrad

Specifications

Insertion Device (ID)

Scatterer	Shielding Material			DER (mrem/h)
	steel (mm)	+lead (mm)	+steel (mm)	
300 mm Cu	3.2	6.7	3.2	0.25
300 mm Cu	3.2	5.6	3.2	2.5
300 mm air	3.2	5.0	3.2	0.25
300 mm air	3.2	4.0	3.2	2.5

2.6. Typical Monochromatic Beam Transport

Shielding Parameters

The following parameters were used to determine the shielding for the walls of the monochromatic beam transport lines.

Positron energy	7.0 GeV	
Positron current	300 mA	
Photon energy range	1 - 300 keV (0.1% bandwidth)	
Vertical divergence	$\Psi = \pm 4/\gamma$	
Ø of the beam pipe	4"	
	<i>Insertion Device (ID)</i>	<i>Bending Magnet (BM)</i>
Critical energy	32.6 keV	19.5 keV
Number of poles	32	1
Horizontal divergence	2 mrad	6 mrad

Specifications

Insertion Device (ID)

Scatterer	Shielding Material			DER (mrem/h)
	steel (mm)	+lead (mm)	+steel (mm)	
300 mm Cu	6.35	15.5	6.35	0.25
300 mm Cu	6.35	12.8	6.35	2.5
300 mm air	3.2	9.1	3.2	0.25
300 mm air	3.2	6.2	3.2	2.5

Bending Magnet (BM)

Scatterer	Shielding Material			DER (mrem/h)
	steel (mm)	+lead (mm)	+steel (mm)	
300 mm Cu	3.2	6.3	3.2	0.25
300 mm Cu	3.2	5.2	3.2	2.5
300 mm air	3.2	4.4	3.2	0.25
300 mm air	3.2	3.3	3.2	2.5

3. Filters

- 3.1. Filter assembly
- 3.2. Filter material

Technical Analysis

The details of a filter (material and dimensions) are closely linked to the nature of the experiment being planned by the CATs. We suggest that the CATs design the filter material to meet their needs. It should be recognized that filter failure during operation can lead to damage of equipment in the beamline.

In section 3.1 we present a generic design for the filter assembly that the CATs could use in this preliminary design of the beamline. Section 3.2 provides filter properties essential for the selection of the material.

Design Schedule

The APS is pursuing detailed analysis of filter material including both thermal and mechanical considerations. This will be available in May 1993.

3. Filters

Filters in beamlines are primarily used to reduce the bandpass of the beam and thus the power load on the elements that interact with the beam (windows, monochromator crystals, etc.). For x-ray beamlines, a combination of high pass (filter foils) and low pass filters, such as a beam-deflecting mirror is very effective. These combinations are also able to suppress higher order reflections in the monochromatic beam.

3.1. Filter Assembly

General description

The filter assembly is designed to provide a modular beamline filtering system. The assembly consists of a housing with a selection of linear filter mounts. The assembly or the linear filter mounts can be placed in every vacuum segment of the user's beamline. For ease of replacement, the filter foils are mounted on standardized filter frames. The filter frames can carry up to 5 different filter foils. The filters are linearly moved into the beam by a light load actuator. The positions of the filters must be interlocked with the beam. This protects the filter frames from being hit by the beam. The filter frames are directly cooled. In the case of radiation cooling, the surrounding vacuum chamber is protected by water cooled-radiation shielding or is directly cooled by attached water cooling.

Specifications

Filter assembly tank

Two ultra high vacuum (UHV) compatible vacuum tanks with two 8" O.D. flanges for filter inserts from the top are under design. The length of one tank is 325 mm. The I.D. of the entrance and exit flange is 4". Ports for pumping (170 l/sec) and viewing the filter foils are supplied. Figure 2 shows the two tanks for a maximum of four assemblies.

Filter assembly

Two types of filter assemblies are in the design phase: a single insert type and a double insert type based on the APS light load actuator with a stepping motor drive. In five different places, up to two filter foils can be mounted on a water-cooled frame. The frame can be moved linearly to different filter slots. The sizes of the filters are 15 x 80 mm. The drive is supplied with interlock switches to prevent accidental irradiation of the frame. The filters can be positioned to within 0.1 mm. Figure 3 shows the water-cooled frame with a linear drive and with radiation shielding.

3.2. Filter Material

The choice of filter material is primarily determined by the transmission energy range, which is defined by the photo absorption coefficient of the material and the thickness of the filter. Figures 4 to 9 show transmission curves for different materials that can be used as x-ray high-pass filters. The filter materials for high power insertion device beamlines have to be carefully analyzed and chosen for the ability to be cooled. Both cooling schemes (conduction cooling and radiation cooling) have to be taken into account because filters often operate at high temperatures. At the APS, failure criteria and an analysis method for filter and window assemblies are under development.

Figure 2

APS BEAM LINE FILTER ASSY
2/20/93

Figure 3 APS BEAM LINE FILTER

Fraction of Transmitted Photons vs.

Thickness of Be Filter ($\rho=1.85\text{g/cm}^3$)

Figure 4

Fraction of Transmitted Photons vs.

Thickness of Diamond Filter ($\rho=3.51\text{g/cm}^3$)

Figure 5

Fraction of Transmitted Photons vs.

Thickness of Al Filter ($\rho=2.6989\text{g/cm}^3$)

Figure 6

Fraction of Transmitted Photons vs.

Thickness of Si Filter ($\rho=2.33\text{g/cm}^3$)

Figure 7

Fraction of Transmitted Photons vs.

Thickness of SiC Filter ($\rho=3.22\text{g/cm}^3$)

DAB
sic.dat, PHOTON
6 Aug 92

Figure 8

Fraction of Transmitted Photons vs.

Thickness of BN Filter ($\rho=2.25\text{g/cm}^3$)

Figure 9

4. Windows

- 4.1. Be - windows
- 4.2. C - windows
- 4.3. Differential pump

Technical Analysis

At APS, failure criteria and an analysis method for filter and window assemblies and their cooling schemes are under development. Reports on the analysis will be made available to all CATs by May 1993.

4. Windows

Windows are used to separate beamline sections with different vacuum or pressure conditions (UHV - HV - atmospheric pressure). For the transmission of the windows, the same design criteria are used as for the filters. The vacuum tightness of the windows ($<1 \times 10^{-10}$ torr/l/sec) together with the pressure difference at a transition from vacuum to atmosphere are additional complications of the window design. The main difficulty is introduced by the high heat load of the ID beams, which can introduce power loads in the kW range. In these cases, windows have to be protected by filters, which have to take a large amount of the power load. Such filter - window combinations severely limit the availability of radiation below 5 - 6 keV. For many CATs, we strongly recommend the use of windowless configurations on ID beamlines to retain the flexibility of using the entire energy spectrum.

It should also be pointed out that if the CAT's interest is primarily in the hard x-ray range (>10 keV), the use of Al windows may be appropriate and should alleviate many design problems of filters and windows.

At APS, failure criteria and an analysis method for filter and window assemblies and their cooling schemes are under development.

4.1. Be - windows

For windows low-Z materials are preferred. Beryllium is a very common window material. In conventional synchrotron applications better than 99.8% pure special grade Be of 250 μm (10 mil) thickness foil is utilized as the window material and is metallurgically bonded to a water-cooled copper Conflat flange. Several reputable US and European manufacturers of Be windows now exist from which one can purchase ready-made window assemblies with 125 to 250 μm thickness that are warranted to be vacuum tight. Windows are assumed to fail under several scenarios. These include severe vacuum incursions, an atmospheric shock wave being the worst possible case; oxidization of Be itself from the vacuum gas impurities; and, deterioration of the window bond itself. The most recognized and studied case of window failures in open literature (Asaoka et al., Rev. Sci. Instrum. Vol. 63, No. 1, pp 473-476, Jan 1992) is however the thermally induced failures under x-ray heating of the Be foil. Structural failure of the thin Be foil under heating and the resulting thermal stresses have been poorly understood and postulated. The conventional criterion that when the thermally induced shear stress becomes half of the foil's yield has underpredicted window failures. Depending on the usage and power source etcetera, the window failures may occur as sudden cracking under heat, gradual evaporative erosion, and elastic/plastic buckling. The latter scenario has been experimentally observed and we can verify it analytically.

Our current analyses and understanding of the thermal behaviour of the cooled, whitebeam APS front end Be window lead us to believe that the window should be protected by at least a 300 μm thick carbon (graphite) filter, should not be allowed to absorb more than 60 W total power, and the window temperature should not exceed 250°C.

Here we present a set of window designs for various uses.

1. A window suitable for white radiation from an APS wiggler (with $K=14$, 7 GeV and 100 mA). This can also be used for monochromatic radiation from APS IDs.
2. A double window to deliver both white and/or monochromatic radiation from either an APS ID or bending magnet source.
3. A window suitable for transmitting bending magnet white radiation with a horizontal width of approximately 110 mm.

WIG. WHITE & WIG/UND MONO. WINDOW

Figure 10

W3S1.DWG 02-19-1993

MONO. & WHITE DUAL WINDOW

W5S1.DWG 02-19-1993

Figure 11

BM WHITE BEAM WINDOW

W4S1.DWG 02-19-1993

Figure 12

4.2. C - windows

4.2.1. Carbon foil windows

Carbon foil windows consists mostly of a thin pyrolytic graphite foil that is mounted on a water-cooled copper block. These windows cannot handle pressure differences in the range of one atmosphere. Also, accidental venting will destroy these windows. They are mainly used as contamination barriers to separate UHV beamline sections from sections with relative bad vacuum conditions ($\sim 10^{-5}$ torr). A detailed analysis of the vacuum separation properties of thin carbon foils was carried out by Lagomarsino et al. [Nucl. Inst. Meth. A307 (1991) 309-311]. Figure 13 shows the design of such window for a high power wiggler beamline at HASYLAB; (1) is the carbon foil (50 mm x 80 mm and 130 μ m thick). The foil is clamped on a water-cooled copper block(2). The bypass valve (3) is used for pump down and venting.

Carbon foil window

Contamination barrier window for high power x-ray beamlines

- 1 Carbon foil (50 x 80 mm, 130 μ m thick)**
- 2 Water-cooled Cu - Block**
- 3 Vacuum bypass with valve**

Figure 13

4. 3. Differential Pump

Differential pumps make use of the good vertical collimation of the synchrotron radiation beam. It is possible to get several orders of magnitude pressure difference in the high vacuum range.

Figure 14 shows the standard APS design for windowless operation. With a minimal aperture of 10 x 78 mm, two 170 l/s ion pumps, and a total length of ~ 1 m, a pressure difference of two orders of magnitude is achieved. With its narrow aperture this device gives good performance in delay of shock waves in case of accidental venting.

A prototype differential pump similar to the one shown in figure 14 has been built and successfully tested by the APS. A pressure differential of more than two orders of magnitude has been reached with the aperture of 10 x 78 mm.

Figure 14

5. Slits

- 5.1.1. White beam slits - insertion device
- 5.1.2. White beam slits - bending magnet
- 5.2. Monochromatic beam slits

Design Status and Schedule

The design of slits and the choice of the material for their construction is closely related to their application. The CATs will have to participate closely in slit designs. For the present, the APS has put much emphasis on the power handling issue related to slit designs. These designs are not likely to meet all user needs. More work will be scheduled in this area as users provide more input to the APS during the beamline preliminary design phase.

In sections 5.1.1 and 5.1.2, the currently designed APS white beam slits for the insertion device and the bending magnet source are included along with specifications. The APS is currently working on its fabrication procedures to develop a prototype to be completed by the end of 1993.

In section 5.2 a monochromatic slit has been specified.

5.1.1. APS ID White Beam Slits

- Horizontal and vertical slits
- White beam compatible
- Four independent precision high load actuators
- Cooling for structural, vibrational, and thermal stability
- UHV compatible
- Provide for closed aperture

The APS white beam slits provide for precise aperture for the white synchrotron radiation. APS standard high load stepping linear actuator modules are used with horizontal and vertical slits. Each of the four masks in the slit assembly is independently movable. The beam intercepts the slits with a grazing incidence. Water-cooled copper foam removes the heat from the slit masks. The masks have been designed so that a pair of slits can be removed through a single vacuum port.

Specifications:

- Slit positional resolution:	2 μm
- Slit positional reproducibility:	5 μm
- Vertical aperture:	0 - 30 mm
- Horizontal aperture:	0 - 100 mm
- Grazing incidence angle:	3° (typical)
- Vacuum tank flanges:	6 inch
- Vacuum tank length:	< 1500 mm
- Actuator:	Standard APS heavy load stepping linear actuator
- Actuator drive:	Stepping motor with linear encoder
- Actuator maximum speed:	20 mm/min
- Encoder	Linear encoder

Figure 15

5.1.2. APS BM White Beam Slits

- Horizontal and vertical slits
- White beam compatible
- Four independent precision light load actuators
- Cooling for structural, vibrational, and thermal stability
- UHV compatible
- Provide for closed aperture

The APS white beam slits provide for precise aperture for the white synchrotron radiation. APS standard light load stepping linear actuator modules are used with horizontal and vertical slits. Each of the four masks in the slit assembly is independently movable. The beam intercepts the slits with a normal incidence. Water cooling removes the heat from the slit masks. The masks have been designed so that each one can be removed through a single vacuum port.

Specifications:

- Slit positional resolution:	10 μm
- Slit positional reproducibility:	25 μm
- Vertical aperture:	0 - 30 mm
- Horizontal aperture:	0 - 150 mm
- Vacuum tank flanges:	8 inch
- Vacuum tank length:	< 900 mm
- Actuator:	Standard APS light load stepping linear actuator
- Actuator drive:	Stepping motor with linear encoder
- Actuator maximum speed:	50 mm/min

ADVANCED PHOTON SOURCE

DRAFT

BM WHITE BEAM HORIZONTAL AND VERTICAL SLITS

5.2. APS Monochromatic Beam Slits

- Horizontal and vertical slits
- Monochromatic beam compatible
- ID and BM monochromatic beamline slits
- Four independent precision light load actuators
- Convective air cooling
- UHV compatible
- Provide for closed aperture

The APS monochromatic beam slits provide for precise aperture for the monochromatic synchrotron radiation. APS standard light load actuator modules are used with horizontal and vertical slits. Each of the four masks in the slit assembly is independently movable. The beam intercepts the slits with normal incidence. Convective air cooling removes the heat from the slit masks. The slits have been designed to be mounted on a standard 6-way cross.

Specifications:

- Slit positional resolution:	1 μm
- Slit positional reproducibility:	5 - 10 μm
ID slits:	
Vertical aperture:	0 - 30 mm
Horizontal aperture:	0 - 120 mm
Vacuum tank flanges:	8 inch
BM slits:	
Vertical aperture:	0 - 30 mm
Horizontal aperture:	0 - 120 mm
Vacuum tank flanges:	8 inch
- Actuators:	Standard APS light load stepping linear actuator
- Actuator Maximum speed:	20 mm/min

ADVANCED PHOTON SOURCE

DRAFT

ID MONOCHROMATIC BEAM HORIZONTAL AND VERTICAL SLITS

Figure 17 IDMONOS.DWG 01-20-92

6. Beam Shutters and Beam Stops

- 6.1. Monochromatic beam shutter - insertion device
- 6.2. White beam stop with integral shutter - insertion device
- 6.3. White beam stop with integral shutter - bending magnet (added later)

Design Considerations

All white beam stops discussed in sections 6.2 and 6.3 are combinations of water-cooled photon absorbers and heavy metal beam stops. Together photon absorber and beam stop completely stop all the radiation from any of the insertion device or bending magnet sources. In order to have a safe operationing scheme, the absorber and the stop will be locked with a "Kirk" key system parallel in its appropriate position.

Design Schedule

The design of the prototype is in an advanced stage and will be completed by May 1993.

6.1 Monochromatic Beam Shutter - Insertion Device

The monochromatic beam shutter is designed to handle monochromatic beam loads from APS insertion devices. In this design, the heat is conducted by a copper rod to a conventional aircooled radiator.

Assumptions

All APS beamlines are required to have two independent beam shutters, each capable of stopping the beam, so that there is less exposure than permitted according DOE guidelines (0.25 mrem/h). The following assumptions are made in this design specification:

1. Maximum energy transmitted	500 keV
2. Maximum critical energy of the ID	32.6 keV
3. Maximum total power	10 W
4. Maximum bandpass at 500 keV	0.1%
5. Expected radiation dose behind one shutter	0.25 mrem/h

These are extremely conservative assumptions for the design because the monochromators are unlikely to propagate full spectrum power in the higher harmonics.

Specifications

1. Material of the shutter	1/2" Cu + 1 5/8" (40 mm) heavy metal
2. Optical aperture	80 x 40 mm ²
3. Actuator	APS light load actuator with pneumatic drive (12.4)
4. Closing time	1 sec
5. Mounting flange	8" O.D.
6. Vacuum	UHV compatible

ADVANCED PHOTON SOURCE

Design Specifications for P4 ID White Beam Photon Absorber

- 1, Location : 31.10 m from APS ID Straight Section Center
  - 2, Optical Aperture Area : 75 mm (H) x 13 mm (V)
  - 3, Horizontal Maximum Acceptance : 2.57 mrad
  - 4, Vertical Maximum Acceptance : 0.42 mrad
  - 5, Closing Time : 1 sec
  - 6, Input Fringe O.D. : 8 inch
  - 7, Output Fringe O.D. : 14 inch
  - 8, Mask Grazing Incidence Angle : 3 degree
  - 9, Mask Material : OFHC with Glid-Cop Face Plate
  - 10, Water Cooling : Mesh-filled Tube
  - 11, Mask Motion Structure : Two-Point Suspension Hockey Stick
  - 12, Actuator : Pneumatic
  - End Point Reproducibility : 10 μ m
  - 13, Damper : Pneumatic
  - 14, Mask Length : 250 mm
  - 15, Device Fringe to Fringe Length : 1030 mm (with heavy metal stop)
  - 16, Vacuum : UHV Compatible
  - 17, Maximum Total Power : 5000 W
-

P4 ID MONO. PHOTON SHUTTER 1

P4S1.DWG 02-22-1993

Figure 18

Figure 19

7. Mirror Chambers

Mirror chambers are used to provide a stable, vibration- and force-free support of beam deflecting mirrors. The chamber must allow the very precise alignment of the mirror. Angle resolutions of the deflection angle of less than $1 \mu\text{rad}$ have to be achieved. The vacuum condition in the chamber ($<1 \times 10^{-9}$ torr) shall avoid carbon contamination of the reflecting surface. In combination with high power beamlines effective cooling has to be supplied. The chamber has to allow bending of the mirror for variable focusing.

To shift the photon energy cut off of the mirrors to high energies, very small reflection angles (e.g. 3 - 4 mrad) are needed. To get a sufficient beam acceptance of the mirror, long mirrors (> 1 m) are used. This strongly influences the chamber design. HASYLAB has had good experience with a chamber design, that makes use of a stiff central chamber frame. The mirror is supported by this frame. The stiffness of the frame withstands easy deformations introduced to the chamber by air pressure or thermal expansion without affecting the alignment of the mirror. The deformations are mainly carried by the cylindrical part of the chamber.

The mirror is aligned by aligning the whole chamber. The chamber movements are decoupled from the beamline by bellows, which allow transversal movements in the vertical and horizontal directions of ± 25 mm. Angular rotations around the beam axis have to be excluded to protect the bellows. This rotation can be easily excluded because of the very small deflection angle.

Figure 21 shows the cooled x - ray mirror chamber design for a 1-m-long cooled toroidal mirror at HASYLAB. Because these are well-tested designs, the APS plans to adopt these designs for the use of the CATs. An improved design drawing for such a mirror chamber will be available before September 1993. The support system for the mirror chamber is also a standard design and is discussed in section 9.1.3.

X - ray mirror chamber for high power beamlines at HASYLAB

Design criteria:

- mirror alignment by moving the whole mirror chamber
- linear and rotational movings mechanically decoupled
- rigid central chamber frame as mirror support
- friction-free rotation of the deflection angle (resolution < 1 μ rad)
- chamber movement decoupled from the beamline by formed bellows
- range of linear movement ± 25 mm

- 1 toroidal mirror (1000 x 130 x 130)
- 2 water cooling
- 3 water-cooled absorber (protection of the mirror face)
- 4 linear encoder
- 5 mirror support and aligning system
- 6 ion pump
- 7 bellows

Figure 20

8. Beam position monitors

- 8.1. White beam position monitor
- 8.2. Monochromatic beam position monitor

Design Schedule

A prototype of the white beam position monitor for the ID beams has been built and tested. The drawings are in section 8.1, along with specifications.

The monochromatic beam position monitors are of a lower priority for design, because many designs are already in use at various synchrotron facilities.

Figure 21

ID FE PBPM
(FIXED VERSION)

ID FE PBPM
(FOR UNDULATOR A)

ID FE PBPM
(FOR WIGGLER A)

Figure 22

9. Supports

- 9.1. Kinematic mount support tables
  - 9.1.1. Economic style support table
    - 9.1.1.1. Vertical motion cone
    - 9.1.1.2. Vertical motion V
    - 9.1.1.3. Vertical motion Flat
    - 9.1.1.4. Horizontal motion
  - 9.1.2. Standard style support table
    - 9.1.2.1. Standard vertical stage
    - 9.1.2.2. Standard horizontal stages
    - 9.1.2.3. Standard spherical coupling
    - 9.1.2.4. Standard V coupling
    - 9.1.2.5. Standard Flat coupling
  - 9.1.3. Precision style support table
    - 9.1.3.1. Precision vertical stage - "A"
    - 9.1.3.2. Precision vertical stage - "B"
    - 9.1.3.3. Precision vertical stage - "C"
    - 9.1.3.4. Precision horizontal stage - "A"
    - 9.1.3.5. Precision horizontal stage - "B"
    - 9.1.3.6. Precision horizontal stage - "C"
- 9.2. Beam position monitor support
  - 9.2.1. Beam position monitor vertical stage
  - 9.2.2. Beam position monitor horizontal stage
  - 9.2.3. Beam position monitor rotation stage
  - 9.2.4. Beam position monitor support column
  - 9.2.5. Beam position monitor kinematic mount cone
  - 9.2.6. Beam position monitor kinematic mount flat
  - 9.2.7. Beam position monitor kinematic mount V
- 9.3. Ion pump supports
  - 9.3.1. Inline 60 l/s ion pump support
- 9.4. Beam pipe supports
  - 9.4.1. 6" beam pipe support

Design Status

All supports have been fully designed. The prototypes of each of the supports are either already built or under construction. The tests will be completed by June 1993.

APS/XFD FE Support System

	FE Economic Support Table	FE Standard Support Table	FE Precision Support Table	FE PBPM Support
Max. Load (Kg)	1000	1000	1000	80
Slide Type (V)	N/A	Linear Rolling	Linear Rolling	Linear Rolling
Slide Type (H)	Regular Friction	Regular Friction	Linear Rolling	Linear Rolling
Travel Range (V) (H) (mm)	50	50	12.7	12.7
Motion (V) Resolution (um)	250	50	10	0.1
Motion (H) Resolution (um)	250	50	10	0.1
Motion (V) Repeatability (um)	400	100	50	2
Motion (H) Repeatability (um)	400	250	50	2
(V) Straightness of Trajectory (rad/5mm)	N/A	5 E -4	2 E -4	1 E -5
(H) Straightness of Trajectory (rad/5mm)	N/A	2 E -3	1 E -4	1 E -5
Basic Operating Mode	Manual	Manual	Stepping Motor	Stepping Motor
Optional Operating Mode	N/A	Stepping Motor	Manual	Manual

ADVANCED PHOTON SOURCE

Figure 23

3-POINT KINEMATIC MOUNTING

ECONOMICAL SUPPORT TABLE SPECIFICATIONS

Max. Load (Kg)	1000
Slide Type (V)	N/A
Slide Type (H)	Regular Friction
Travel Range (V) 0-0 (mm)	50
Motion (V) Resolution (um)	250
Motion (H) Resolution (um)	250
Motion (V) Repeatability (um)	400
Motion (H) Repeatability (um)	400
(V) Straightness of Trajectory (rad/5mm)	N/A
(H) Straightness of Trajectory (rad/5mm)	N/A
Basic Operating Mode	Manual
Optional Operating Mode	N/A

ADVANCED PHOTON SOURCE

TYPICAL MOUNTING HOLES, 3 PLACES

Figure 24

ECONOMICAL SUPPORT BASE WITH ECONOMICAL STYLE KINEMATIC MOUNT STAGES

ADVANCED PHOTON SOURCE

(6) ϕ 1/4-20
on ϕ 3.000 [76.20] CIRCLE

± 1.00 [25.40] TRAVEL

Figure 25

ECONOMICAL VERTICAL KINEMATIC MOUNT - CONE

ADVANCED PHOTON SOURCE

Figure 26

ECONOMICAL VERTICAL KINEMATIC MOUNT - V

ADVANCED PHOTON SOURCE

Figure 27

ECONOMICAL VERTICAL KINEMATIC MOUNT - FLAT

KNURLED KNOB

ECONOMICAL KINEMATIC MOUNT HORIZONTAL SLIDE

STANDARD SUPPORT TABLE SPECIFICATIONS

Max. Load (Kg)	1000
Slide Type (V)	Linear Rolling
Slide Type (H)	Regular Friction
Travel Range (V) 0-0 (mm)	50
Motion (V) Resolution (um)	50
Motion (H) Resolution (um)	50
Motion (V) Repeatability (um)	100
Motion (H) Repeatability (um)	250
(V) Straightness of Trajectory (rad/5mm)	5 E -4
(H) Straightness of Trajectory (rad/5mm)	2 E -3
Basic Operating Mode	Manual
Optional Operating Mode	Stepping Motor

ADVANCED PHOTON SOURCE

TYPICAL MOUNTING HOLES, 3 PLACES

SPECIFICATIONS:

1. LOAD CAPACITY: 2200 LBS [1000 kg]
2. DEGREES OF FREEDOM: 5
3. TRAVEL RANGE: $\pm 1"$ [25.4 mm]
4. MECHANICAL COUNTER RESOLUTION: .005"/COUNT
5. VERTICAL AND HORIZONTAL ADJUSTMENT: HAND KNOB
6. OPTIONAL MOTOR MOUNTS AVAILABLE

Figure 29 STANDARD SUPPORT BASE WITH STANDARD KINEMATIC MOUNT STAGES

1. TRAVEL 2 inch [50.80]
2. LOAD CAPACITY 1000 lb. [450 kg.]
3. MANUAL VERTICAL ADJUSTMENT :
4. MECHANICAL COUNTER RESOLUTION :
5. TURNS PER INCH OF TRAVEL : 20
6. MAX. OFFSET LOAD DISTANCE : 1 inch.

Figure 30 STANDARD KINEMATIC MOUNT VERTICAL STAGE

Figure 31

Figure 32

1. ALL DIMENSIONS SHOWN IN [ ] ARE MILLIMETERS.

[illegible][illegible]

Figure 33 STANDARD KINEMATIC MOUNT V COUPLING

Figure 34 STANDARD KINEMATIC MOUNT FLAT COUPLING

PRECISION SUPPORT TABLE SPECIFICATIONS

Max. Load (Kg)	1000
Slide Type (V)	Linear Rolling
Slide Type (H)	Linear Rolling
Travel Range (V) (H) (mm)	12.7
Motion (V) Resolution (μm)	10
Motion (H) Resolution (μm)	10
Motion (V) Repeatability (μm)	50
Motion (H) Repeatability (μm)	50
(V) Straightness of Trajectory (rad/5mm)	$2 \text{ E } -4$
(H) Straightness of Trajectory (rad/5mm)	$1 \text{ E } -4$
Basic Operating Mode	Stepping Motor
Optional Operating Mode	Manual

- SPECIFICATIONS:
1. LOAD CAPACITY: 2200 LBS [1000 kg]
  2. TRAVEL RANGE: ± 25 [6.35]
  3. DEGREES OF FREEDOM: 5
  4. REPEATABILITY: $50 \mu\text{m}$
  5. RESOLUTION: $10 \mu\text{m}$
  6. HORIZONTAL AND VERTICAL ADJUSTMENT: STEPPER MOTOR

Figure 35 SUPPORT BASE WITH PRECISION KINEMATIC MOUNT STAGES

- TECHNICAL SPECIFICATIONS:**
1. PRIOR TO ASSEMBLY, ALL MACHINED PARTS MUST BE FREE OF BURRS, CHIPS, OR OTHER DEFECTS. ALL PARTS MUST BE INSPECTED AND APPROVED BY QUALITY CONTROL PRIOR TO ASSEMBLY.
  2. PRIOR TO FITTING, ALL LAMINAR BEAMS MUST BE MARKED WITH A CROSSHATCH PATTERN FOR IDENTIFICATION.

1. FIT BEARING BALLS INTO THEIR BASE AND SECURE LIGHTLY WITH SCREWERS.
2. INSERT ROLLER COCKS AND CENTER ALONG BEARING BALLS.
3. ADJUST PRELOAD BY USING SET SCREWS UNTIL PLAT BEARS EVENLY ON BOTH BEARING BALLS TO APPROX. 12-15 LB. TENSION TO APPROX. 10-12 LB.
4. REMOVE SCREWS ON BEARING BALLS THOROUGHLY LEAVE A FEW SCREWS TO HOLD PLAT IN POSITION.

- ### **A SPECIFICATIONS**
1. LOAD CAPACITY: 1000 LBS (454 kg)
  2. TRAVEL RANGE: 2 1/4" (6.35)
  3. RESOLUTION: 10 μ m
  4. REPEATABILITY: 50 μ m
  5. STRAIGHTNESS OF TRAJECTORY: 1 in 4 mil PER 25 mm.

[illegible][illegible]

Figure 40

FE PBPM SUPPORT TABLE SPECIFICATIONS

Max. Load (Kg)	90
Slide Type (V)	Linear Rolling
Slide Type (H)	Linear Rolling
Travel Range (V) (H) (mm)	12.7
Motion (V) Resolution (um)	0.1
Motion (H) Resolution (um)	0.1
Motion (V) Repeatability (um)	2
Motion (H) Repeatability (um)	2
(V) Straightness of Trajectory (rad/5mm)	1 E -5
(H) Straightness of Trajectory (rad/5mm)	1 E -5
Basic Operating Mode	Stepping Motor
Optional Operating Mode	Manual

PATENT HOLD

This document copy, since it is transmitted in advance of patent clearance, by Argonne National Laboratory, is made available in confidence solely for use in performance of work under contracts with the U.S. Department of Energy. This document is not to be published nor its contents otherwise disseminated or used for purposes other than specified above before patent approval for such release or use has been secured, upon request, from the Argonne Patent Dept., Argonne National Laboratory, 9700 South Cass Avenue, Argonne, IL 60439.

Figure 42 PBPM VERTICAL STAGE ASSEMBLY

SPECIFICATIONS:

1. LOAD CAPACITY: 90 kg
2. TRAVEL RANGE: ± 5 mm
3. RESOLUTION: $0.1 - 0.5 \mu\text{m}$
4. REPEATABILITY: VERTICAL: $\pm 2 \mu\text{m}$
HORIZONTAL: $\pm 5 \mu\text{m}$
5. ANGULAR RESOLUTION: 5 ARCSECONDS
6. STRAIGHTNESS OF TRAJECTORY: 1×10^{-5} RAD/25 mm

Figure 43 PBPM SUPPORT TABLE ASSEMBLY

Figure 44 PBPM HORIZONTAL STAGE ASSEMBLY

Figure 45 PBPM ROTATION STAGE ASSEMBLY

Figure 46 PBPM SUPPORT COLUMN ASSEMBLY

SECTION A-A

Figure 47 PBPM KINEMATIC MOUNT CONE ASSEMBLY

Figure 48 PBPM KINEMATIC MOUNT FLAT ASSEMBLY

Figure 49 PBPM KINEMATIC MOUNT V ASSEMBLY

Figure 50

T1 BENDING MAGNET
FRONT END
INLINE ION PUMP CLIPP
ASSEMBLY

SECTION A-A

DETAIL "C"

Figure 51

T1 BENDING MAGNET
FRONT END
BEAM PIPE SUPPORT
ASSEMBLY

10. Stepping Motor Drives

- 10.1. Stepping motors (being developed)
- 10.2. Drivers/Controllers (being developed)

Design Status

The APS user organization has appointed a subcommittee to advise the APS in developing standard recommendations for stepping motors, drivers, and controllers. It is expected that the committee will provide their recommendation before the end of 1993 by working closely with the APS. The standardisation is also closely linked to anew software development work at the APS that uses Experimental Physics Interface Control System (EPICS).

As a starting point, the APS has used the stepping motor/drivers specified in section 10.1 - 10.2, which are used in all standard supports (section 9).

10.1. - 10.2. Stepping Motors - Drivers/Controllers

- Standardize stepping motor drives and controllers
- Control for multi-motor driven systems
- Open and/or closed loop control
- Acceleration and deceleration control
- Power fail-safe control with battery backup coordinate memory

The APS stepping motor driver/controller is designed to provide a user friendly control environment for multi-motor driven systems such as kinematic mounting tables or beamline slits. Control is to be either manual or through computer control.

Specifications:

- Stepping motor:	four phase, 400 half steps/rev, 3% accuracy M063-LE06-E: 3.36 V DC, 2.9 A M093-FD8107-E: 4.3V, 3.5A
- Maximum stepping rate:	40000 half steps/sec
- Able to interface with computer	
- Position display:	seven digit and sign with computer conversion
- Feedback:	open and/or closed loop with TTL incremental encoder signals
- Memory backup:	battery backup nonvolatile memory of control data
- Power fail safe	
- Switch status information to host:	open or closed of limits or home
- Connectors	(to be standardized)
- Cables	(to be standardized)

11. Interlock - Interfaces

11.1. Switches

11.2. Solenoids

11.3. Interfaces - cables and connectors

11.3.1. Ion pumps with power supplies

11.3.2. Ion gauges with power supplies

11.3.3. Vacuum valves

Design Status

Work will be completed by the end of 1993.

12. Actuators

- 12.1. Heavy load with stepping motor drive
- 12.2. Heavy load with pneumatic drive
- 12.3. Light load with stepping motor drive
- 12.4. Light load with pneumatic drive

Design Status

The design work on all the actuators is in advanced stage, and prototypes will be built and tested in 1993. The following sections include various design specifications.

12.1. APS Heavy Load Actuator with Stepping Motor Drive

SLAH-30-8, SLAH-75-8
(A2-81) (A2-82)

- Heavy load actuators
- For use with APS ID white beam slits
- Stepping motor driven with shaft encoder

A typical application for these actuators is to drive APS ID white beam slits. There are three different ranges of travel designed.

Specifications:

- Slide type:	Linear rolling
- Travel ranges:	30 & 75 mm
- Motion resolution:	2 μ m
- Motion reproducibility:	5 μ m
- Straightness:	2×10^{-4} rad/25 mm
- Vacuum load:	38 kg
- Maximum useful axial load:	100 kg
- Stepping motor:	Slo-Syn M092-LE06
- Actuator flange O. D.:	8 inch
- Feedthrough flange O. D.:	4.5 inch
- Maximum feedthrough diameter:	50 mm
- Redundant limit switches	
-Encoder:	Optical on motor shaft and/or linear on slide
-Maximum speed:	20 mm/min
-Vacuum:	UHV compatible

Figure 52

APS HEAVY LOAD STEPPING LINEAR ACTUATOR
SLAH-75-8
END VIEW
1/14/93
C. BRITE

Figure 53

12.2. APS Heavy Load Actuator with Pneumatic Drive

PLAH-33-10

- Heavy load actuators
- For use with APS ID and BM front end and beamline safety shutters

A typical application for the heavy load pneumatic actuators is to drive APS safety shutters.

Specifications:

- Stroke:	33 mm $\pm$ 0.25 mm
- Motion repeatability:	0.5 mm $\pm$ 0.1 mm
- Total axial load:	190 kg
- Vacuum load:	40 kg
- Maximum useful axial load:	70 kg
- Chamber flange O. D.:	8 inch
- Rod flange O. D.:	4.5 inch
- Redundant limit switches in each end position	Micro switch BZZRQ2X-A2
- Pneumatic power supply:	90 - 120 psi
- Maximum closing time:	0.5 - 1 sec
- Pneumatic damping:	in each end position
- Life time:	500k cycles
- Vacuum:	UHV compatible

APS HEAVY LOAD PNEUMATIC LINEAR ACTUATOR

PLAH-86-10

PLAH.DWG 01-18-92

Figure 54

SPECIFICATIONS

STROKE	33 mm
MOTION REPEATABILITY	0.5 mm
TOTAL AREA LOAD	220 kg
VACUUM LOAD	40 kg
MINIMUM URBTA AREA LOAD	150 kg
CHAMBER FLANGE O.D.	8 inch
ROD FLANGE O.D.	4.5 inch
REDUNDANT LIMIT SWITCHES	IN EACH END POSITION
PNEUMATIC POWER SUPPLY	80 - 120 psi
MAINTAIN CLOSING TIME	0.5 sec
PNEUMATIC DAMPING	IN EACH END POSITION
VACUUM	300 COMPATIBLE

NOTES:
1. THIS IS A UNIV ASSEMBLY, KEEP THE UNIV PARTS CLEAN DURING ASSEMBLY AND WASH UP FOR UNIV FILING WITH ALUMINUM POL.

2. DEVICE SHALL BE LEAK TESTED USING A MASS SPECTROMETER WITH MINIMUM SENSITIVITY FOR HELIUM OF 2X 10⁻¹⁰ STANDARD CC/SEC PER LEAK METER DIVISION, SUCH AS:

ALCATEL ASM-110TCL
VARIAN MGR 925 OR 936
VEECO MS-9, MS-80 OR MS-18
DU PONT CEC 24-120

CALIBRATION OF THE LEAK DETECTOR SENSITIVITY SHALL BE PERFORMED JUST PRIOR TO TESTING.

PHIAL TEST WILL CONSIST OF SURROUNDING THE CHAMBER (BAGGING) WITH HELIUM. THE CHAMBER WILL BE RE-CEDED BY A 2X DEFLECTION IN THE MOST SENSITIVE RANGE OF THE LEAK DETECTOR IS SENSED WITHIN 1 MIN.

3. ALL DIMENSIONS WITH ( ) ARE MILLIMETERS

REFERENCE SOURCE:

1. MICRO SWITCH
621 - ROUTE 63
BENSENVALE, IL 60106
PHONE: 840-3948

20	3/8-18 UNCG-1/2 LG HHCS DRILLED HO	SST	8
19	3/8-18 UNCG-1/2 LG HHCS DRILLED HO	SST	4
18	3/8-18 UNCG-1/2 LG HHCS DRILLED HO	SST	4
17	3/8-18 UNCG-1/2 LG HHCS DRILLED HO	SST	4
16	3/8-18 UNCG-1/2 LG HHCS DRILLED HO	SST	4
15	3/8-18 UNCG-1/2 LG HHCS DRILLED HO	SST	4
14	3/8-18 UNCG-1/2 LG HHCS DRILLED HO	SST	4
13	3/8-18 UNCG-1/2 LG HHCS DRILLED HO	SST	4
12	3/8-18 UNCG-1/2 LG HHCS DRILLED HO	SST	4
11	3/8-18 UNCG-1/2 LG HHCS DRILLED HO	SST	4
10	3/8-18 UNCG-1/2 LG HHCS DRILLED HO	SST	4
9	3/8-18 UNCG-1/2 LG HHCS DRILLED HO	SST	4
8	3/8-18 UNCG-1/2 LG HHCS DRILLED HO	SST	4
7	3/8-18 UNCG-1/2 LG HHCS DRILLED HO	SST	4
6	3/8-18 UNCG-1/2 LG HHCS DRILLED HO	SST	4
5	3/8-18 UNCG-1/2 LG HHCS DRILLED HO	SST	4
4	3/8-18 UNCG-1/2 LG HHCS DRILLED HO	SST	4
3	3/8-18 UNCG-1/2 LG HHCS DRILLED HO	SST	4
2	3/8-18 UNCG-1/2 LG HHCS DRILLED HO	SST	4
1	3/8-18 UNCG-1/2 LG HHCS DRILLED HO	SST	4

20	3/8-18 UNCG-1/2 LG HHCS DRILLED HO	SST	8
19	3/8-18 UNCG-1/2 LG HHCS DRILLED HO	SST	4
18	3/8-18 UNCG-1/2 LG HHCS DRILLED HO	SST	4
17	3/8-18 UNCG-1/2 LG HHCS DRILLED HO	SST	4
16	3/8-18 UNCG-1/2 LG HHCS DRILLED HO	SST	4
15	3/8-18 UNCG-1/2 LG HHCS DRILLED HO	SST	4
14	3/8-18 UNCG-1/2 LG HHCS DRILLED HO	SST	4
13	3/8-18 UNCG-1/2 LG HHCS DRILLED HO	SST	4
12	3/8-18 UNCG-1/2 LG HHCS DRILLED HO	SST	4
11	3/8-18 UNCG-1/2 LG HHCS DRILLED HO	SST	4
10	3/8-18 UNCG-1/2 LG HHCS DRILLED HO	SST	4
9	3/8-18 UNCG-1/2 LG HHCS DRILLED HO	SST	4
8	3/8-18 UNCG-1/2 LG HHCS DRILLED HO	SST	4
7	3/8-18 UNCG-1/2 LG HHCS DRILLED HO	SST	4
6	3/8-18 UNCG-1/2 LG HHCS DRILLED HO	SST	4
5	3/8-18 UNCG-1/2 LG HHCS DRILLED HO	SST	4
4	3/8-18 UNCG-1/2 LG HHCS DRILLED HO	SST	4
3	3/8-18 UNCG-1/2 LG HHCS DRILLED HO	SST	4
2	3/8-18 UNCG-1/2 LG HHCS DRILLED HO	SST	4
1	3/8-18 UNCG-1/2 LG HHCS DRILLED HO	SST	4

Figure 55

20	3/8-18 UNCG-1/2 LG HHCS DRILLED HO	SST	8
19	3/8-18 UNCG-1/2 LG HHCS DRILLED HO	SST	4
18	3/8-18 UNCG-1/2 LG HHCS DRILLED HO	SST	4
17	3/8-18 UNCG-1/2 LG HHCS DRILLED HO	SST	4
16	3/8-18 UNCG-1/2 LG HHCS DRILLED HO	SST	4
15	3/8-18 UNCG-1/2 LG HHCS DRILLED HO	SST	4
14	3/8-18 UNCG-1/2 LG HHCS DRILLED HO	SST	4
13	3/8-18 UNCG-1/2 LG HHCS DRILLED HO	SST	4
12	3/8-18 UNCG-1/2 LG HHCS DRILLED HO	SST	4
11	3/8-18 UNCG-1/2 LG HHCS DRILLED HO	SST	4
10	3/8-18 UNCG-1/2 LG HHCS DRILLED HO	SST	4
9	3/8-18 UNCG-1/2 LG HHCS DRILLED HO	SST	4
8	3/8-18 UNCG-1/2 LG HHCS DRILLED HO	SST	4
7	3/8-18 UNCG-1/2 LG HHCS DRILLED HO	SST	4
6	3/8-18 UNCG-1/2 LG HHCS DRILLED HO	SST	4
5	3/8-18 UNCG-1/2 LG HHCS DRILLED HO	SST	4
4	3/8-18 UNCG-1/2 LG HHCS DRILLED HO	SST	4
3	3/8-18 UNCG-1/2 LG HHCS DRILLED HO	SST	4
2	3/8-18 UNCG-1/2 LG HHCS DRILLED HO	SST	4
1	3/8-18 UNCG-1/2 LG HHCS DRILLED HO	SST	4

20	3/8-18 UNCG-1/2 LG HHCS DRILLED HO	SST	8
19	3/8-18 UNCG-1/2 LG HHCS DRILLED HO	SST	4
18	3/8-18 UNCG-1/2 LG HHCS DRILLED HO	SST	4
17	3/8-18 UNCG-1/2 LG HHCS DRILLED HO	SST	4
16	3/8-18 UNCG-1/2 LG HHCS DRILLED HO	SST	4
15	3/8-18 UNCG-1/2 LG HHCS DRILLED HO	SST	4
14	3/8-18 UNCG-1/2 LG HHCS DRILLED HO	SST	4
13	3/8-18 UNCG-1/2 LG HHCS DRILLED HO	SST	4
12	3/8-18 UNCG-1/2 LG HHCS DRILLED HO	SST	4
11	3/8-18 UNCG-1/2 LG HHCS DRILLED HO	SST	4
10	3/8-18 UNCG-1/2 LG HHCS DRILLED HO	SST	4
9	3/8-18 UNCG-1/2 LG HHCS DRILLED HO	SST	4
8	3/8-18 UNCG-1/2 LG HHCS DRILLED HO	SST	4
7	3/8-18 UNCG-1/2 LG HHCS DRILLED HO	SST	4
6	3/8-18 UNCG-1/2 LG HHCS DRILLED HO	SST	4
5	3/8-18 UNCG-1/2 LG HHCS DRILLED HO	SST	4
4	3/8-18 UNCG-1/2 LG HHCS DRILLED HO	SST	4
3	3/8-18 UNCG-1/2 LG HHCS DRILLED HO	SST	4
2	3/8-18 UNCG-1/2 LG HHCS DRILLED HO	SST	4
1	3/8-18 UNCG-1/2 LG HHCS DRILLED HO	SST	4

12.3 APS Light Load Actuator with Stepping Motor Drive

SLAL-30-8, SLAL-75-8, SLAL-150-8
(A1-81) (A1-82) (A1-83)

- Light load actuators
- For use with APS ID and BM monochromatic slits and BM white slits
- Stepping-motor driven with shaft encoder

A typical application for these actuators is to drive the APS monochromatic beam slits and BM white beam slits. There are three different ranges of travel designed.

Specifications:

- Slide type:	Linear rolling
- Travel ranges:	30, 75, and 150 mm
- Motion resolution:	1 μm
- Motion reproducibility:	5 μm
- Straightness:	2×10^{-4} rad/25 mm
- Vacuum load:	19 kg
- Maximum useful axial load:	10 kg
- Stepping motor:	Slo-Syn M061-LE06
- Actuator flange O. D.:	8 inch
- Feedthrough flange O. D.:	2.75 inch
- Maximum feedthrough diameter:	25 mm
- Redundant limit switches	
-Encoder:	Optical on motor shaft and/or linear on slide
-Maximum speed:	50 mm/min
-Vacuum:	UHV compatible

TOP VIEW

ELEVATION

END VIEW

ALPS LIGHT LOAD STEPPING LINEAR ACTUATOR
 SLAL-75-8
 1/26/93
 C. BRITE

Figure 56

APS LIGHT LOAD STEPPING LINEAR ACTUATOR
 SLAL-30-8
 1/27/93
 C. BRITE

Figure 57

12.4. APS Light Load Pneumatic Actuator with Pneumatic Drive

A3-81, A3-83

- **Light load actuator**
- **For use with APS monochromatic beam shutters and BM white beam shutters**

A typical application for the light load pneumatic actuators is to drive APS monochromatic beam shutters and BM white beam shutters.

Specifications:

- Stroke:	31 mm $\pm$ 0.25 mm & 70 mm $\pm$ 0.25 mm
- Motion repeatability:	0.1 mm
- Total axial load	90 kg
- Vacuum load:	17 kg
- Maximum useful axial load:	30 kg
- Actuator flange O. D.:	8 inch
- Feedthrough flange:	4.5 inch
- Maximum feedthrough diameter:	25 mm
- Bore size:	2.5 inch
- Rod size:	1.375 inch
- Bellows O. D.:	56 mm
- Bellows I. D.:	35.6 mm
- Redundant limit switches in each end position:	Micro switch BZZRQ2X-A2
- Pneumatic power supply:	60 - 100 psi
- Maximum closing time:	1 sec
- Pneumatic damping:	in each end position
- Solenoid:	24 V DC (power failure safe design)
- Life time:	500 k cycles
- Vacuum:	UHV compatible

A3-81, 31mm STROKE

A3-83, 70 mm STROKE

APS A3 Light Load Pneumatic Linear Actuators

Figure 58

A3S1.DWG 02-22-1993

13. Bellows

13.1. Welded bellows

13.2. Formed bellows

Design Status

The design work for welded bellows is complete, while that for formed bellows will be completed before May 1993. With the completion of the designs, these bellows can be procured from vendors. They will also be available from the APS stockroom.

REF. SOURCES

STANDARD BELLOWS CO.
375 TUNPIKE AVE.
WINDSOR LOCKS, CONN. 06096

WMD VACUUM PRODUCTS CORP.
23842 CABOT BLVD.
HAYWARD, CA. 94545-1651

NOTES:

1. WHEN MACHINING VACUUM PARTS, USE OF SILICONE AND SULPHUR-BASED CUTTING FLUIDS IS PROHIBITED. USE ONE OF THE FOLLOWING:
- A) CIMCOOL 5 STAR 49
  - B) TRIM SOL

2. ALL WELDS JOINING BELLOWES AND FLANGES ARE TO BE INTERNAL

3. BELLWOWS ASSEMBLY SHALL BE LEAK TESTED USING A MASS SPECTROMETER WITH MINIMUM SENSITIVITY FOR HELIUM OF 2×10^{-10} STANDARD CC/SEC PER LEAK METER DIVISION, SUCH AS

ALCATEL SAM-110TCL

Du PONT CEC 24-1208

VARIAN MS-9, MS-90 OR MS-18

- CALIBRATION OF THE LEAK DETECTOR SENSITIVITY SHALL BE PERFORMED JUST PRIOR TO TRESTING.

FINAL TEST WILL CONSIST OF SURROUNDING THE ASSEMBLY (BAGGING) WITH HELIUM. THE ASSEMBLY WILL BE REJECTED IF A 2% DEFLECTION ON THE MOST SENSITIVE RANGE OF THE LEAK DETECTOR IS SENSED WITHIN 1 MIN.

4. KEEP THE PART CLEAN, AND WRAP FOR UHV
PACKING WITH ALUMINUM FOIL.

DIMENSIONS IN { } ARE MILLIMETERS

Figure 59

3				STD. BELLOW. CAT. #497 - 400 - 7 - EE	WITH EXTENDED ENDS	1
2	WDC'S CAT. #100026			6" CF. ROTATABLE FLANGE		1
1	WDC'S CAT. #100026			6" CF. NONROTATABLE FLANGE		1
ITEM	QTY	DWG./PART NUMBER	NOMENCLATURE OR DESCRIPTION	MATERIAL	/ SPEC	QTY

UNLESS OTHERWISE SPECIFIED: ALL DIMENSIONS ARE IN INCHES		THIS DRAWING IS THE PROPERTY OF		ARGONNE NATIONAL LABORATORY	
DIMENSIONS ARE IN INCHES		A09231		PARTS LIST	
SIGNALS (1/3200) 4.25" 4.00 (0.125) 4.00 (0.125) SURFACE ROUGHNESS 125 /		DATE 7/27/73 DRAWN BY MUSCIA CHECKED BY DESIGNED MUSCIA RESPONSIBLE ENGINEER D. SHULMAN APPROVED/RELEASED PROJECT NO.		TITLE ADVANCED PHOTON SOURCE V2 - 82 WELDED BELLOW. ASSY. 6" - 6" CF. 4" I.D. X 210mm	
PLANT ALL DIMENSIONS SURFACE ROUGHNESS TO BE SURFACE FINISH TO BE DIMENSIONS & TOLERANCES IN DIMENSIONS & TOLERANCES IN DIMENSIONS & TOLERANCES IN		DATE 7/27/73 DRAWN BY MUSCIA CHECKED BY DESIGNED MUSCIA RESPONSIBLE ENGINEER D. SHULMAN APPROVED/RELEASED PROJECT NO.		SCALE HALF SHEET 1 of 1	
DIMENSIONS & TOLERANCES IN DIMENSIONS & TOLERANCES IN DIMENSIONS & TOLERANCES IN		DATE 7/27/73 DRAWN BY MUSCIA CHECKED BY DESIGNED MUSCIA RESPONSIBLE ENGINEER D. SHULMAN APPROVED/RELEASED PROJECT NO.		SCALE HALF SHEET 1 of 1	
DIMENSIONS & TOLERANCES IN DIMENSIONS & TOLERANCES IN DIMENSIONS & TOLERANCES IN		DATE 7/27/73 DRAWN BY MUSCIA CHECKED BY DESIGNED MUSCIA RESPONSIBLE ENGINEER D. SHULMAN APPROVED/RELEASED PROJECT NO.		SCALE HALF SHEET 1 of 1	

REVISIONS				
NO.	DATE	BY	CHANGE DESCRIPTION	WAS

WDC VACUUM PRODUCTS CORP.
23842 CABOT BLVD.
HAYWARD, CA 94545-1651

5 DIMENSIONS IN [ ] ARE MILLIMETERS

2 1 1

3	JDC'S CAT. #100026	STD. BELLOW'S CAT. #497 - 400 - 5 - EE	WITH EXTENDED ENDS	1
2	JDC'S CAT. #100026	6" CF. ROTATABLE FLANGE		1
1	JDC'S CAT. #110026	6" CF. NONROTATABLE FLANGE		1
ITEM	DWG/PART NUMBER	DESCRIPTION	MATERIAL / SPEC	QTY

UNLESS OTHERWISE SPECIFIED: ALL DIMENSIONS ARE IN INCHES TOLERANCES FRACTIONS DECIMALS ANGULAR ±.010 (±.000) ±.02° ±.020 (±.010) ±.05° ±.005 (±.012)		THIS DRAWING IS THE PROPERTY OF ARGONNE NATIONAL LABORATORY	
A09232 DRAWN BY MUSCIA CHECKED BY DESIGNER RESPONSIBLE ENGINEER MATERIAL SEE PARTS LIST		DATE 2/20/81 DATE 2/20/81 DATE 2/20/81 PROJECT NO. APPROVED/RELEASED DATE 2/20/81	
SURFACE ROUGHNESS $\sqrt{125}$ REMOVE ALL BURRS AND BREAK SHARP EDGES AND RAIL SURFACE FINISH TO BE R1A1 DIMENSIONS IN PARENTS IN DIMENSIONS IN PARENTS IN DIMENSIONS IN PARENTS IN		TITLE ADVANCED PHOTON SOURCE V2 - 83 WELDED BELLOW'S ASSY. 6" - 6" CF. 4" X 175mm	
		SCALE HALF	DRAWING NUMBER
		SHEET 1 of 1	4105090702-830000-00

REF. SOURCES

- STANDARD BELLOWS CO.
375 TUNPIKE AVE.
WINDSOR LOCKS, CONN. 06096
- MDC VACUUM PRODUCTS CORP.
23842 CABOT BLVD.
HAYWARD, CA. 94545-1651

NOTES:

- WHEN MACHINING VACUUM PARTS, USE OF SILICONE AND SULPHUR-BASED CUTTING FLUIDS IS PROHIBITED. USE ONE OF THE FOLLOWING:
A) CIMCOOL 5 STAR 49
B) TRIM SOL
- ALL WELDS JOINING BELLOWS AND FLANGES ARE TO BE INTERNAL
- BELLOWS ASSEMBLY SHALL BE LEAK TESTED USING A MASS SPECTROMETER WITH MINIMUM SENSITIVITY FOR HELIUM OF 2 X 10⁻¹⁰ STANDARD CC/SEC PER LEAK METER DIVISION, SUCH AS ALCATEL SAM-110TCL
DU PONT CEC 24-120B
VARIAN MS-9, MS-90 OR MS-18
CALIBRATION OF THE LEAK DETECTOR SENSITIVITY SHALL BE PERFORMED JUST PRIOR TO TRESTING.

FINAL TEST WILL CONSIST OF SURROUNDING THE ASSEMBLY (BAGGING) WITH HELIUM. THE ASSEMBLY WILL BE REJECTED IF A 2% DEFLECTION ON THE MOST SENSITIVE RANGE OF THE LEAK DETECTOR IS SENSED WITHIN 1 MIN.

KEEP THE PART CLEAN, AND WRAP FOR UHV PACKING WITH ALUMINUM FOIL.

DIMENSIONS IN [ ] ARE MILLIMETERS

ITEM	QTY	DESCRIPTION	EXTENDED ENDS
3	1	STD. BELLOWS CAT. #497 - 400 - 2 - EE	WITH EXTENDED ENDS
2	1	MDC'S CAT. #100026 6" CF. ROTATABLE FLANGE	
1	1	MDC'S CAT. #110026 6" CF. NONROTATABLE FLANGE	

UNLESS OTHERWISE SPECIFIED, ALL DIMENSIONS ARE IN INCHES		TOLERANCES		SURFACE FINISHES		MATERIALS		PROCESSING		INSPECTION		REVISIONS	
DIMENSIONS		FINISHES		TREATMENTS		MATERIALS		PROCESSING		INSPECTION		REVISIONS	
DIMENSIONS		FINISHES		TREATMENTS		MATERIALS		PROCESSING		INSPECTION		REVISIONS	
DIMENSIONS		FINISHES		TREATMENTS		MATERIALS		PROCESSING		INSPECTION		REVISIONS	

Figure 61

ARGONNE NATIONAL LABORATORY		PROJECT NO.		DATE		SCALE		SHEET		OF		TOTAL	
A09233		MUSCIA		2/7/73		1/2"		1		1		1	
DESIGNED BY		CHECKED BY		DATE		APPROVED		RELEASED		DATE		BY	
MUSCIA		MUSCIA		2/7/73		1/2"		1		1		1	
TITLE		ADVANCED PHOTON SOURCE		V2 - 84 WELDED BELLOWS ASSY.		6" - 6" CF. X 4" I.D. X 110mm		4105090702-840000-00		C		1	

NOTES:

1. WHEN MACHINING VACUUM PARTS, USE OF SILICONE AND SULPHUR-BASED CUTTING FLUIDS IS PROHIBITED. USE ONE OF THE FOLLOWING:
A) CIMCOOL 5 STAR 49
B) TRIM SOL

2. ALL WELDS JOINING BELLINGS AND FLANGES ARE TO BE INTERNAL

3. BELLINGS ASSEMBLY SHALL BE LEAK TESTED USING A MASS SPECTROMETER WITH MINIMUM SENSITIVITY FOR HELIUM OF 2 X 10⁻¹⁰ STANDARD CC/SEC PER LEAK METER DIVISION, SUCH AS ALCATEL SAM-110TCL
DU PONT CEC 24-120B
VARIAN MS-9, MS-90 OR MS-18

CALIBRATION OF THE LEAK DETECTOR SENSITIVITY SHALL BE PERFORMED JUST PRIOR TO TRESTING.

FINAL TEST WILL CONSIST OF SURROUNDING THE ASSEMBLY (BAGGING) WITH HELIUM. THE ASSEMBLY WILL BE REJECTED IF A 2% DEFLECTION ON THE MOST SENSITIVE RANGE OF THE LEAK DETECTOR IS SENSED WITHIN 1 MIN.

4. KEEP THE PART CLEAN, AND WRAP FOR UHV PACKING WITH ALUMINUM FOIL

5. DIMENSIONS IN [ ] ARE MILLIMETERS

REF. SOURCES

1. MDC VACUUM PRODUCTS CORP.
23842 CABOT BLVD.
HAYWARD, CA. 94545-1651
2. STANDARD BELLINGS CO.
375 TUNPIKE AVE.
WINDSOR LOCKS, CONN. 06096

ITEM	QTY	DESCRIPTION	MATERIAL / SPEC
3	1	STD. BELLINGS CAT. #497 - 400 - 7 - EE	WITH EXTENDED ENDS
2	1	MDC'S CAT. #100026	6" CF. ROTATABLE FLANGE
1	1	MDC'S CAT. #110026	6" CF. NONROTATABLE FLANGE

PARTS LIST

UNLESS OTHERWISE SPECIFIED: ALL DIMENSIONS ARE IN INCHES TOLERANCES: FRACTIONS DECIMALS ANGLES X .001 (1/1000) 0.001 0.001 X .005 (1/200) 0.005 0.005 X .010 (1/100) 0.010 0.010 X .020 (1/50) 0.020 0.020 X .050 (1/20) 0.050 0.050 X .100 (1/10) 0.100 0.100 X .150 (3/20) 0.150 0.150 X .200 (1/5) 0.200 0.200 X .250 (1/4) 0.250 0.250 X .300 (3/10) 0.300 0.300 X .375 (3/8) 0.375 0.375 X .500 (1/2) 0.500 0.500 X .750 (3/4) 0.750 0.750 X 1.000 (1) 1.000 1.000 X 1.500 (3/2) 1.500 1.500 X 2.000 (1) 2.000 2.000 X 2.500 (5/4) 2.500 2.500 X 3.000 (3) 3.000 3.000 X 4.000 (1) 4.000 4.000 X 5.000 (1) 5.000 5.000 X 6.000 (1) 6.000 6.000 X 7.000 (1) 7.000 7.000 X 8.000 (1) 8.000 8.000 X 9.000 (1) 9.000 9.000 X 10.000 (1) 10.000 10.000 X 12.000 (1) 12.000 12.000 X 15.000 (1) 15.000 15.000 X 20.000 (1) 20.000 20.000 X 25.000 (1) 25.000 25.000 X 30.000 (1) 30.000 30.000 X 35.000 (1) 35.000 35.000 X 40.000 (1) 40.000 40.000 X 45.000 (1) 45.000 45.000 X 50.000 (1) 50.000 50.000 X 55.000 (1) 55.000 55.000 X 60.000 (1) 60.000 60.000 X 65.000 (1) 65.000 65.000 X 70.000 (1) 70.000 70.000 X 75.000 (1) 75.000 75.000 X 80.000 (1) 80.000 80.000 X 85.000 (1) 85.000 85.000 X 90.000 (1) 90.000 90.000 X 95.000 (1) 95.000 95.000 X 100.000 (1) 100.000 100.000 X 105.000 (1) 105.000 105.000 X 110.000 (1) 110.000 110.000 X 115.000 (1) 115.000 115.000 X 120.000 (1) 120.000 120.000 X 125.000 (1) 125.000 125.000 X 130.000 (1) 130.000 130.000 X 135.000 (1) 135.000 135.000 X 140.000 (1) 140.000 140.000 X 145.000 (1) 145.000 145.000 X 150.000 (1) 150.000 150.000 X 155.000 (1) 155.000 155.000 X 160.000 (1) 160.000 160.000 X 165.000 (1) 165.000 165.000 X 170.000 (1) 170.000 170.000 X 175.000 (1) 175.000 175.000 X 180.000 (1) 180.000 180.000 X 185.000 (1) 185.000 185.000 X 190.000 (1) 190.000 190.000 X 195.000 (1) 195.000 195.000 X 200.000 (1) 200.000 200.000 X 205.000 (1) 205.000 205.000 X 210.000 (1) 210.000 210.000 X 215.000 (1) 215.000 215.000 X 220.000 (1) 220.000 220.000 X 225.000 (1) 225.000 225.000 X 230.000 (1) 230.000 230.000 X 235.000 (1) 235.000 235.000 X 240.000 (1) 240.000 240.000 X 245.000 (1) 245.000 245.000 X 250.000 (1) 250.000 250.000 X 255.000 (1) 255.000 255.000 X 260.000 (1) 260.000 260.000 X 265.000 (1) 265.000 265.000 X 270.000 (1) 270.000 270.000 X 275.000 (1) 275.000 275.000 X 280.000 (1) 280.000 280.000 X 285.000 (1) 285.000 285.000 X 290.000 (1) 290.000 290.000 X 295.000 (1) 295.000 295.000 X 300.000 (1) 300.000 300.000 X 305.000 (1) 305.000 305.000 X 310.000 (1) 310.000 310.000 X 315.000 (1) 315.000 315.000 X 320.000 (1) 320.000 320.000 X 325.000 (1) 325.000 325.000 X 330.000 (1) 330.000 330.000 X 335.000 (1) 335.000 335.000 X 340.000 (1) 340.000 340.000 X 345.000 (1) 345.000 345.000 X 350.000 (1) 350.000 350.000 X 355.000 (1) 355.000 355.000 X 360.000 (1) 360.000 360.000 X 365.000 (1) 365.000 365.000 X 370.000 (1) 370.000 370.000 X 375.000 (1) 375.000 375.000 X 380.000 (1) 380.000 380.000 X 385.000 (1) 385.000 385.000 X 390.000 (1) 390.000 390.000 X 395.000 (1) 395.000 395.000 X 400.000 (1) 400.000 400.000 X 405.000 (1) 405.000 405.000 X 410.000 (1) 410.000 410.000 X 415.000 (1) 415.000 415.000 X 420.000 (1) 420.000 420.000 X 425.000 (1) 425.000 425.000 X 430.000 (1) 430.000 430.000 X 435.000 (1) 435.000 435.000 X 440.000 (1) 440.000 440.000 X 445.000 (1) 445.000 445.000 X 450.000 (1) 450.000 450.000 X 455.000 (1) 455.000 455.000 X 460.000 (1) 460.000 460.000 X 465.000 (1) 465.000 465.000 X 470.000 (1) 470.000 470.000 X 475.000 (1) 475.000 475.000 X 480.000 (1) 480.000 480.000 X 485.000 (1) 485.000 485.000 X 490.000 (1) 490.000 490.000 X 495.000 (1) 495.000 495.000 X 500.000 (1) 500.000 500.000 X 505.000 (1) 505.000 505.000 X 510.000 (1) 510.000 510.000 X 515.000 (1) 515.000 515.000 X 520.000 (1) 520.000 520.000 X 525.000 (1) 525.000 525.000 X 530.000 (1) 530.000 530.000 X 535.000 (1) 535.000 535.000 X 540.000 (1) 540.000 540.000 X 545.000 (1) 545.000 545.000 X 550.000 (1) 550.000 550.000 X 555.000 (1) 555.000 555.000 X 560.000 (1) 560.000 560.000 X 565.000 (1) 565.000 565.000 X 570.000 (1) 570.000 570.000 X 575.000 (1) 575.000 575.000 X 580.000 (1) 580.000 580.000 X 585.000 (1) 585.000 585.000 X 590.000 (1) 590.000 590.000 X 595.000 (1) 595.000 595.000 X 600.000 (1) 600.000 600.000 X 605.000 (1) 605.000 605.000 X 610.000 (1) 610.000 610.000 X 615.000 (1) 615.000 615.000 X 620.000 (1) 620.000 620.000 X 625.000 (1) 625.000 625.000 X 630.000 (1) 630.000 630.000 X 635.000 (1) 635.000 635.000 X 640.000 (1) 640.000 640.000 X 645.000 (1) 645.000 645.000 X 650.000 (1) 650.000 650.000 X 655.000 (1) 655.000 655.000 X 660.000 (1) 660.000 660.000 X 665.000 (1) 665.000 665.000 X 670.000 (1) 670.000 670.000 X 675.000 (1) 675.000 675.000 X 680.000 (1) 680.000 680.000 X 685.000 (1) 685.000 685.000 X 690.000 (1) 690.000 690.000 X 695.000 (1) 695.000 695.000 X 700.000 (1) 700.000 700.000 X 705.000 (1) 705.000 705.000 X 710.000 (1) 710.000 710.000 X 715.000 (1) 715.000 715.000 X 720.000 (1) 720.000 720.000 X 725.000 (1) 725.000 725.000 X 730.000 (1) 730.000 730.000 X 735.000 (1) 735.000 735.000 X 740.000 (1) 740.000 740.000 X 745.000 (1) 745.000 745.000 X 750.000 (1) 750.000 750.000 X 755.000 (1) 755.000 755.000 X 760.000 (1) 760.000 760.000 X 765.000 (1) 765.000 765.000 X 770.000 (1) 770.000 770.000 X 775.000 (1) 775.000 775.000 X 780.000 (1) 780.000 780.000 X 785.000 (1) 785.000 785.000 X 790.000 (1) 790.000 790.000 X 795.000 (1) 795.000 795.000 X 800.000 (1) 800.000 800.000 X 805.000 (1) 805.000 805.000 X 810.000 (1) 810.000 810.000 X 815.000 (1) 815.000 815.000 X 820.000 (1) 820.000 820.000 X 825.000 (1) 825.000 825.000 X 830.000 (1) 830.000 830.000 X 835.000 (1) 835.000 835.000 X 840.000 (1) 840.000 840.000 X 845.000 (1) 845.000 845.000 X 850.000 (1) 850.000 850.000 X 855.000 (1) 855.000 855.000 X 860.000 (1) 860.000 860.000 X 865.000 (1) 865.000 865.000 X 870.000 (1) 870.000 870.000 X 875.000 (1) 875.000 875.000 X 880.000 (1) 880.000 880.000 X 885.000 (1) 885.000 885.000 X 890.000 (1) 890.000 890.000 X 895.000 (1) 895.000 895.000 X 900.000 (1) 900.000 900.000 X 905.000 (1) 905.000 905.000 X 910.000 (1) 910.000 910.000 X 915.000 (1) 915.000 915.000 X 920.000 (1) 920.000 920.000 X 925.000 (1) 925.000 925.000 X 930.000 (1) 930.000 930.000 X 935.000 (1) 935.000 935.000 X 940.000 (1) 940.000 940.000 X 945.000 (1) 945.000 945.000 X 950.000 (1) 950.000 950.000 X 955.000 (1) 955.000 955.000 X 960.000 (1) 960.000 960.000 X 965.000 (1) 965.000 965.000 X 970.000 (1) 970.000 970.000 X 975.000 (1) 975.000 975.000 X 980.000 (1) 980.000 980.000 X 985.000 (1) 985.000 985.000 X 990.000 (1) 990.000 990.000 X 995.000 (1) 995.000 995.000 X 1000.000 (1) 1000.000 1000.000 X 1005.000 (1) 1005.000 1005.000 X 1010.000 (1) 1010.000 1010.000 X 1015.000 (1) 1015.000 1015.000 X 1020.000 (1) 1020.000 1020.000 X 1025.000 (1) 1025.000 1025.000 X 1030.000 (1) 1030.000 1030.000 X 1035.000 (1) 1035.000 1035.000 X 1040.000 (1) 1040.000 1040.000 X 1045.000 (1) 1045.000 1045.000 X 1050.000 (1) 1050.000 1050.000 X 1055.000 (1) 1055.000 1055.000 X 1060.000 (1) 1060.000 1060.000 X 1065.000 (1) 1065.000 1065.000 X 1070.000 (1) 1070.000 1070.000 X 1075.000 (1) 1075.000 1075.000 X 1080.000 (1) 1080.000 1080.000 X 1085.000 (1) 1085.000 1085.000 X 1090.000 (1) 1090.000 1090.000 X 1095.000 (1) 1095.000 1095.000 X 1100.000 (1) 1100.000 1100.000 X 1105.000 (1) 1105.000 1105.000 X 1110.000 (1) 1110.000 1110.000 X 1115.000 (1) 1115.000 1115.000 X 1120.000 (1) 1120.000 1120.000 X 1125.000 (1) 1125.000 1125.000 X 1130.000 (1) 1130.000 1130.000 X 1135.000 (1) 1135.000 1135.000 X 1140.000 (1) 1140.000 1140.000 X 1145.000 (1) 1145.000 1145.000 X 1150.000 (1) 1150.000 1150.000 X 1155.000 (1) 1155.000 1155.000 X 1160.000 (1) 1160.000 1160.000 X 1165.000 (1) 1165.000 1165.000 X 1170.000 (1) 1170.000 1170.000 X 1175.000 (1) 1175.000 1175.000 X 1180.000 (1) 1180.000 1180.000 X 1185.000 (1) 1185.000 1185.000 X 1190.000 (1) 1190.000 1190.000 X 1195.000 (1) 1195.000 1195.000 X 1200.000 (1) 1200.000 1200.000 X 1205.000 (1) 1205.000 1205.000 X 1210.000 (1) 1210.000 1210.000 X 1215.000 (1) 1215.000 1215.000 X 1220.000 (1) 1220.000 1220.000 X 1225.000 (1) 1225.000 1225.000 X 1230.000 (1) 1230.000 1230.000 X 1235.000 (1) 1235.000 1235.000 X 1240.000 (1) 1240.000 1240.000 X 1245.000 (1) 1245.000 1245.000 X 1250.000 (1) 1250.000 1250.000 X 1255.000 (1) 1255.000 1255.000 X 1260.000 (1) 1260.000 1260.000 X 1265.000 (1) 1265.000 1265.000 X 1270.000 (1) 1270.000 1270.000 X 1275.000 (1) 1275.000 1275.000 X 1280.000 (1) 1280.000 1280.000 X 1285.000 (1) 1285.000 1285.000 X 1290.000 (1) 1290.000 1290.000 X 1295.000 (1) 1295.000 1295.000 X 1300.000 (1) 1300.000 1300.000 X 1305.000 (1) 1305.000 1305.000 X 1310.000 (1) 1310.000 1310.000 X 1315.000 (1) 1315.000 1315.000 X 1320.000 (1) 1320.000 1320.000 X 1325.000 (1) 1325.000 1325.000 X 1330.000 (1) 1330.000 1330.000 X 1335.000 (1) 1335.000 1335.000 X 1340.000 (1) 1340.000 1340.000 X 1345.000 (1) 1345.000 1345.000 X 1350.000 (1) 1350.000 1350.000 X 1355.000 (1) 1355.000 1355.000 X 1360.000 (1) 1360.000 1360.000 X 1365.000 (1) 1365.000 1365.000 X 1370.000 (1) 1370.000 1370.000 X 1375.000 (1) 1375.000 1375.000 X 1380.000 (1) 1380.000 1380.000 X 1385.000 (1) 1385.000 1385.000 X 1390.000 (1) 1390.000 1390.000 X 1395.000 (1) 1395.000 1395.000 X 1400.000 (1) 1400.000 1400.000 X 1405.000 (1) 1405.000 1405.000 X 1410.000 (1) 1410.000 1410.000 X 1415.000 (1) 1415.000 1415.000 X 1420.000 (1) 1420.000 1420.000 X 1425.000 (1) 1425.000 1425.000 X 1430.000 (1) 1430.000 1430.000 X 1435.000 (1) 1435.000 1435.000 X 1440.000 (1) 1440.000 1440.000 X 1445.000 (1) 1445.000 1445.000 X 1450.000 (1) 1450.000 1450.000 X 1455.000 (1) 1455.000 1455.000 X 1460.000 (1) 1460.000 1460.000 X 1465.000 (1) 1465.000 1465.000 X 1470.000 (1) 1470.000 1470.000 X 1475.000 (1) 1475.000 1475.000 X 1480.000 (1) 1480.000 1480.000 X 1485.000 (1) 1485.000 1485.000 X 1490.000 (1) 1490.000 1490.000 X 1495.000 (1) 1495.000 1495.000 X 1500.000 (1) 1500.000 1500.000 X 1505.000 (1) 1505.000 1505.000 X 1510.000 (1) 1510.000 1510.000 X 1515.000 (1) 1515.000 1515.000 X 1520.000 (1) 1520.000 1520.000 X 1525.000 (1) 1525.000 1525.000 X 1530.000 (1) 1530.000 1530.000 X 1535.000 (1) 1535.000 1535.000 X 1540.000 (1) 1540.000 1540.000 X 1545.000 (1) 1545.000 1545.000 X 1550.000 (1) 1550.000 1550.000 X 1555.000 (1) 1555.000 1555.000 X 1560.000 (1) 1560.000 1560.000 X 1565.000 (1) 1565.000 1565.000 X 1570.000 (1) 1570.000 1570.000 X 1575.000 (1) 1575.000 1575.000 X 1580.000 (1) 1580.000 1580.000 X 1585.000 (1) 1585.000 1585.000 X 1590.000 (1) 1590.000 1590.000 X 1595.000 (1) 1595.000 1595.000 X 1600.000 (1) 1600.000 1600.000 X 1605.000 (1) 1605.000 1605.000 X 1610.000 (1) 1610.000 1610.000 X 1615.000 (1) 1615.000 1615.000 X 1620.000 (1) 1620.000 1620.000 X 1625.000 (1) 1625.000 1625.000 X 1630.000 (1) 1630.000 1630.000 X 1635.000 (1) 1635.000 1635.000 X 1640.000 (1) 1640.000 1640.000 X 1645.000 (1) 1645.000 1645.000 X 1650.000 (1) 1650.000 1650.000 X 1655.000 (1) 1655.000 1655.000 X 1660.000 (1) 1660.000 1660.000 X 1665.000 (1) 1665.000 1665.000 X 1670.000 (1) 1670.000 16	

[illegible]

Figure 64

1. WHEN MACHINING VACUUM PARTS, USE OF SILICONE AND SULPHUR-BASED CUTTING FLUIDS IS PROHIBITED. USE ONE OF THE FOLLOWING:
  - A) CIMCOOL 5 STAR 49
  - B) TRIM SOL
2. ALL WELDS JOINING BELLOWES AND FLANGES ARE TO BE INTERNAL
3. BELLOWES ASSEMBLY SHALL BE LEAK TESTED USING A MASS SPECTROMETER WITH MINIMUM SENSITIVITY FOR HELIUM OF 2 X 10⁻¹⁰ STANDARD CC/SEC PER LEAK METER DIVISION, SUCH AS
ALCATEL SAM-110TCL
Du PONT CEC 24-120B
VARIAN MS-9, MS-90 OR MS-18
CALIBRATION OF THE LEAK DETECTOR SENSITIVITY SHALL BE PERFORMED JUST PRIOR TO TRESTING.
FINAL TEST WILL CONSIST OF SURROUNDING THE ASSEMBLY (BAGGING) WITH HELIUM. THE ASSEMBLY WILL BE REJECTED IF A 2% DEFLECTION ON THE MOST SENSITIVE RANGE OF THE LEAK DETECTOR IS SENSED WITHIN 1 MIN.
4. KEEP THE PART CLEAN, AND WRAP FOR UHV PACKING WITH ALUMINUM FOIL
5. DIMENSIONS IN [ ] ARE MILLIMETERS

NOTES:

1. WHEN MACHINING VACUUM PARTS, USE OF SILICONE AND SULPHUR-BASED CUTTING FLUIDS IS PROHIBITED. USE ONE OF THE FOLLOWING:
A) CIMCOOL 5 STAR 49
B) TRIM SOL

2. ALL WELDS JOINING BELLOWES AND FLANGES ARE TO BE INTERNAL

3. BELLOWES ASSEMBLY SHALL BE LEAK TESTED USING A MASS SPECTROMETER WITH MINIMUM SENSITIVITY FOR HELIUM OF 2 X 10⁻¹⁰ STANDARD CC/SEC PER LEAK METER DIVISION, SUCH AS ALCATEL SAM-110TCL
DU PONT CEC 24-1208
VARIAN MS-9, MS-90 OR MS-18

CALIBRATION OF THE LEAK DETECTOR SENSITIVITY SHALL BE PERFORMED JUST PRIOR TO TRESTING.

FINAL TEST WILL CONSIST OF SURROUNDING THE ASSEMBLY (BAGGING) WITH HELIUM. THE ASSEMBLY WILL BE REJECTED IF A 2% DEFLECTION ON THE MOST SENSITIVE RANGE OF THE LEAK DETECTOR IS SENSED WITHIN 1 MIN.

4. KEEP THE PART CLEAN, AND WRAP FOR UNH PACKING WITH ALUMINUM FOIL

5. DIMENSIONS IN [ ] ARE MILLIMETERS

REF. SOURCES

- 3 STANDARD BELLOWES CO.
375 TUNPIKE AVE.
WINDSOR LOCKS, CONN. 06096
- 2 MDC VACUUM PRODUCTS CORP.
23842 CABOT BLVD.
HAYWARD, CA. 94545-1651
- 1 NOR-CAL PRODUCTS, INC.
1967 S. OREGON ST.
YREKA, CA. 96097

ITEM	QTY	DESCRIPTION	MATERIAL / SPEC
3	1	STD. BELLOWES CAT. #750 - 600 - 5 - EE WITH EXTENDED ENDS	
2	1	MDC'S CAT. #100031 8" CF. ROTATABLE FLANGE	
1	1	N.C. #1200-00N 12" CF. NONROTATABLE FLANGE	

PARTS LIST

THIS DRAWING IS THE PROPERTY OF
ARGONNE NATIONAL LABORATORY

PROJECT NO. **20-56**
DATE **2/2/74**
BY **20-56**
CHECKED BY **20-56**
APPROVED/RELEASED **20-56**

DESIGNED BY **MUSCIA**
DATE **2/2/74**
BY **20-56**
CHECKED BY **MUSCIA**
DATE **2/2/74**
BY **20-56**

SCALE **HALF**
SIZE **C**
SHEET **1** OF **1**

SEE PARTS LIST

Figure 65

SYN	CHANGE DESCRIPTION	BY	CHKD	DATE

REF. SOURCES

2 STANDARD BELLOWS CO.
375 TURNPIKE AVE.
WINDSOR LOCKS, CONN. 06096

1 MDC VACUUM PRODUCTS CORP.
23842 CABOT BLVD.
HAYWARD, CA. 94545-1651

NOTES:

1. WHEN MACHINING VACUUM PARTS, USE OF SILICONE AND SULPHUR-BASED CUTTING FLUIDS IS PROHIBITED. USE ONE OF THE FOLLOWING:
A) CIMCOOL 5 STAR 49
B) TRIM SOL

2. ALL WELDS JOINING BELLOWS AND FLANGES ARE TO BE INTERNAL

3. BELLOWS ASSEMBLY SHALL BE LEAK TESTED USING A MASS SPECTROMETER WITH MINIMUM SENSITIVITY FOR HELIUM OF 2 X 10⁻¹⁰ STANDARD CC/SEC PER LEAK METER DIVISION, SUCH AS ALCATEL SAM-110TQL

Du PONT CEC 24-120B

VARIAN MS-9, MS-90 OR MS-18

CALIBRATION OF THE LEAK DETECTOR SENSITIVITY SHALL BE PERFORMED JUST PRIOR TO TRESTING.

FINAL TEST WILL CONSIST OF SURROUNDING THE ASSEMBLY (BAGGING) WITH HELIUM. THE ASSEMBLY WILL BE REJECTED IF A 2% DEFLECTION ON THE MOST SENSITIVE RANGE OF THE LEAK DETECTOR IS SENSED WITHIN 1 MIN.

4. KEEP THE PART CLEAN, AND WRAP FOR UHV PACKING WITH ALUMINUM FOIL.

5. DIMENSIONS IN [ ] ARE MILLIMETERS

ITEM	QTY/PART NUMBER	DESCRIPTION	MATERIAL / SPEC	QTY
3	STD. BELLOWS CAT. #750 - 600 - 5 - EE	WITH EXTENDED ENDS		1
2	MDC'S CAT. #100031	8" CF. ROTATABLE FLANGE		1
1	MDC'S CAT. #110031	8" CF. NONROTATABLE FLANGE		1

ARGONNE NATIONAL LABORATORY SO5254 DRAWN BY: MUSCIA CHECKED BY: MUSCIA DATE: 7/1/78 PROJECT NO.: 21247 APPROVED/RELEASED: D. Shaw 7/1/78		ADVANCED PHOTON SOURCE 212-93 WELDED BELLOWS ASSY. 8" - 8" CF. 6" I.D. X 200mm (P4105090702-930000-00)	
SEE PARTS LIST DO NOT SCALE DRAWING		SCALE: HALF SHEET 1 of 1	

Figure 66

SYN	CHANGE DESCRIPTION	BY	CHK.	DATE

REVISIONS

5. DIMENSIONS IN ( ) ARE MILLIMETERS

2 1 1

[illegible]

REF. SOURCES

2 STANDARD BELLWINDS CO.
375 TUNPIKE AVE.
WINDSOR LOCKS, CONN. 06096

1 MDC VACUUM PRODUCTS CORP.
23842 CABOT BLVD.
HAYWARD, CA. 94545-1651

NOTES:

1. WHEN MACHINING VACUUM PARTS, USE OF SILICONE AND SULPHUR-BASED CUTTING FLUIDS IS PROHIBITED. USE ONE OF THE FOLLOWING:
A) CIMCOOL 5 STAR 49
B) TRIM SOL

2. ALL WELDS JOINING BELLWINDS AND FLANGES ARE TO BE INTERNAL

3. BELLWINDS ASSEMBLY SHALL BE LEAK TESTED USING A MASS SPECTROMETER WITH MINIMUM SENSITIVITY FOR HELIUM OF 2 X 10⁻¹⁰ STANDARD CC/SEC PER LEAK METER DIVISION, SUCH AS ALCATEL SAM-110TCL
DU PONT CEC 24-1208
VARIAN MS-9, MS-90 OR MS-18

CALIBRATION OF THE LEAK DETECTOR SENSITIVITY SHALL BE PERFORMED JUST PRIOR TO TRESTING.

FINAL TEST WILL CONSIST OF SURROUNDING THE ASSEMBLY (BAGGING) WITH HELIUM. THE ASSEMBLY WILL BE REJECTED IF A 2% DEFLECTION ON THE MOST SENSITIVE RANGE OF THE LEAK DETECTOR IS SENSED WITHIN 1 MIN.

4. KEEP THE PART CLEAN, AND WRAP FOR UHV PACKING WITH ALUMINUM FOIL

5. DIMENSIONS IN [ ] ARE MILLIMETERS

Figure 68

PARTS LIST			
ITEM	QTY/PART NUMBER	NOMENCLATURE OR DESCRIPTION	MATERIAL / SPEC
3	STD. BELLWINDS CAT. #750 - 600 - 5 - EE	WITH EXTENDED ENDS	1
2	MDC'S CAT. #100031	8" CF. ROTATABLE FLANGE	1
1	MDC'S CAT. #110031	8" CF. NONROTATABLE FLANGE	1
THIS DRAWING IS THE PROPERTY OF			
ARGONNE NATIONAL LABORATORY			
S05256			
DESIGNED BY	MUSCIA	DATE	2/2/81
CHECKED BY	MUSCIA	DATE	2/2/81
APPROVED BY	MUSCIA	DATE	2/2/81
SEE PARTS LIST			
SCALE HALF 1/2			
SHEET 1 of 1			

ADVANCED PHOTON SOURCE
2-95 WELDED BELLWINDS ASSY.
8" - 8" I.D. X 230mm
(P4105090702-950000-00)

END

**DATE
FILMED**

5 / 3 / 93