NOTICE

**CERTAIN DATA
CONTAINED IN THIS
DOCUMENT MAY BE
DIFFICULT TO READ
IN MICROFICHE
PRODUCTS.**

**KINEMATICS OF COMPTON BACKSCATTERING
X-RAY SOURCE FOR ANGIOGRAPHY**

L.N. Blumberg

May 1992

NATIONAL SYNCHROTRON LIGHT SOURCE

**BROOKHAVEN NATIONAL LABORATORY
ASSOCIATED UNIVERSITIES, INC.**

Under Contract No. DE-AC02-76CH00016 with the

UNITED STATES DEPARTMENT OF ENERGY.

DISTRIBUTION OF THIS DOCUMENT IS UNLIMITED

DISCLAIMER

This report was prepared as an account of work sponsored by an agency of the United States Government. Neither the United States Government nor any agency thereof, nor any of their employees, nor any of their contractors, subcontractors, or their employees, makes any warranty, express or implied, or assumes any legal liability or responsibility for the accuracy, completeness, or usefulness of any information, apparatus, product, or process disclosed, or represents that its use would not infringe privately owned rights. Reference herein to any specific commercial product, process, or service by trade name, trademark, manufacturer, or otherwise, does not necessarily constitute or imply its endorsement, recommendation, or favoring by the United States Government or any agency, contractor or subcontractor thereof. The views and opinions of authors expressed herein do not necessarily state or reflect those of the United States Government or any agency, contractor or subcontractor thereof.

BNL--47503

DE92 014016

KINEMATICS OF COMPTON BACKSCATTERING X-RAY SOURCE FOR ANGIOGRAPHY

L.N. Blumberg

May 1992

National Synchrotron Light Source
Brookhaven National Laboratory
Upton, New York 11972

MASTER

DISTRIBUTION OF THIS DOCUMENT IS UNLIMITED

KINEMATICS OF COMPTON BACKSCATTERING X-RAY SOURCE FOR ANGIOGRAPHY

L. N. Blumberg
National Synchrotron Light Source
Brookhaven National Laboratory
Upton, NY 11973

April, 1992

ABSTRACT

Calculations of X-Ray production rates, energy spread, and spectrum of Compton-backscattered photons from a Free Electron Laser on an electron beam in a low energy (136-MeV) compact (8.5-m circumference) storage ring indicate that an X-Ray intensity of 34.6×10^7 X-Ray photons per 0.5-mm x 0.5-mm pixel for Coronary Angiography near the 33.169-keV iodine K-absorption edge can be achieved in a 4-msec pulse within a scattering cone of 1-mrad half angle. This intensity, at 10-m from the photon-electron interaction point to the patient is about a factor of 10 larger than presently achieved from a 4.5-T superconducting wiggler source in the NSLS 2.5-GeV storage ring and over an area about 5 times larger. The 2.2-keV energy spread of the Compton-backscattered beam is, however, much larger than the 70-eV spread presently attained from the wiggler source and use of a monochromator. The beam spot at the 10-m interaction point-to-patient distance is 20-mm diameter; larger spots are attainable at larger distances but with a corresponding reduction in X-Ray flux. Such a facility could be an inexpensive clinical alternative to present methods of non-invasive Digital Subtraction Angiography (DSA), small enough to be deployed in an urban medical center, and could have other medical, industrial and aerospace applications. Problems with the Compton backscattering source include laser beam heating of the mirror in the FEL oscillator optical cavity, achieving a large enough X-Ray beam spot at the patient, and obtaining radiation damping of the transverse oscillations and longitudinal emittance dilution of the storage ring electron beam resulting from photon-electron collisions without going to higher electron energy where the X-Ray energy spread becomes excessive for DSA.

INTRODUCTION

The Laser-Electron Gamma Source (LEGS) operating at the X-Ray Ring of the NSLS [1] at BNL suggests the possibility of using the same Compton backscattering mechanism as a source of lower energy photons in the X-Ray region to produce a narrow X-Ray beam of moderate energy spread for medical applications--in particular, for X-Ray imaging of coronary arteries (Angiography). The successful application of the Digital Subtraction Angiography (DSA) technique to coronary arteries, first at the Stanford Synchrotron Radiation Laboratory (SSRL) [2] and more recently at the BNL SMRF facility [3] at the NSLS X-Ray Ring, suggests that a more intense beam of high repetition rate covering a larger area of the coronary region may produce images of clinical interest. The present method utilizes the broad-band X-Ray spectrum emitted by high (2.5-GeV) energy electrons traversing a high field (4.5 T) superconducting wiggler magnet, and subsequent monochromatization by Bragg reflection from a Si crystal. The method employs digital subtraction of two images obtained by successive exposures of narrow strips (0.5-mm x 128-mm) of the coronary region to the ribbon beam of the wiggler with a narrow energy band (70 eV) immediately above and below the K-absorption edge of Iodine (33.169 keV) introduced as a contrast agent by various

injection. The high intensity of the beams, presently about 3×10^7 photons per $0.5\text{-mm} \times 0.5\text{-mm}$ pixel incident on a 256 pixel linear array of Lithium-drifted silicon solid state detectors in a 4-millisecond pulse[4] has produced pre-clinical quality images of arterial blockages. The use of venous injection, rather than introduction of the contrast agent with a catheter inserted into the arterial system and subsequently positioned in the aorta adjacent to the branch point of the coronary artery under study, is a less invasive procedure which is likely to reduce the risk of induced infarction while at the same time significantly reduce the radiation dose presently delivered to the patient (and the cardiologist) from a conventional X-Ray tube source essentially identical to the tube developed by Roentgen nearly 100 years ago. Disadvantages of the digital subtraction (Dichromography) venous injection method include the presence of iodinated blood within the heart chambers and other vessels in proximity to the heart during the irradiation as well as the large and costly high energy electron storage ring and associated high field wiggler magnet and complex beam-line monochromator system required. We seek a smaller, inexpensive, nearly monochromatic X-Ray source which would be affordable and could be accommodated within an urban medical center. Several attempts to design such a source based on present technology of a compact storage ring and very high field wiggler have been suggested[5,6]. In this and subsequent papers we will explore a more elegant, simple, and less costly approach using a low energy electron ring (or Linac) to produce low energy, infrared photons from an Infrared Free-Electron Laser (IR-FEL) which are then converted to the required higher energy X-Rays via the photon-electron Compton scattering process. In some sense this device could be regarded as an X-Ray laser source. A similar system for converting ultra-violet photons to gamma-rays using an in-ring UV-FEL has already been suggested by Luccio[7] to increase the energy and intensity of gamma-rays in the LEGS facility[1]. The disadvantage of low intensity inherent in the collision of electron-photon bunches can be offset by use of an IR-FEL producing photon bunches at $\sim 200\text{-MHz}$ rates rather than the $\sim 1\text{-Hz}$ rates from conventional infrared lasers such as CO_2 or the low average power (5 watts) of the higher energy ($\lambda_L = 0.5\text{-micron}$) CW Argon-ion laser. The IR-FEL is now a well-developed technology, as is the compact, low energy ($\sim 200\text{-MeV}$) electron storage ring or low emittance, high repetition-rate Linac ($\sim 50\text{-MeV}$), both recently developed at BNL[8,9]. A preliminary study of the IR-FEL-based Compton backscattering source has recently been performed by E.B. Blum[10] and has yielded very encouraging results on X-Ray intensity and identified several potentially serious problems, including the serious perturbation of the storage ring electron beam by the photon-electron scattering process. The purpose of this note is to first set forth the basic kinematic and intensity considerations and then address in a preliminary manner the problems of beam blow-up and mirror heating. The possibility of large and unacceptable energy and size blow-up of the electron beam during the 4-msec period of electron-photon collisions will require further study using Monte-Carlo techniques.

1. ENERGY/WAVELENGTH OF REQUIRED LASER PHOTONS vs. ELECTRON ENERGY

The kinematics of the photon-electron colliding beam Compton scattering process has been lucidly presented by Sandorfi, Levine and Thorn[1] in the NSLS LEGS facility proposal. They obtain, to good approximation, an expression for the energy of the backscattered photon in the lab system, E_x , as a function of lab total electron energy $E_e = \gamma mc^2$, lab energy of the incident laser photon E_L and lab scattering angle θ between the electron and scattered photon as

$$E_x(\theta) = 4\gamma^2 E_L / (1 + 4\gamma E_L / mc^2 + \gamma^2 \theta^2) \quad (1)$$

In Eq. (1) mc^2 is the rest energy of the electron. The scattering process is illustrated in Fig. 1 both in the lab and CM system, where the CM system (commonly referred to as the center-of-mass system but more precisely called the center-of-momentum system) is, in our case, to a very good approximation, the system in which the electron is at rest. For a head-on collision, $\theta = 0$ in Eq. (1) and the laser energy E_L required to produce a backscattered photon energy $E_x(0)$ at electron energy E_e is

$$E_L = (mc^2)^2 E_x(0) / [4E_e(E_e - E_x(0))] \quad (2)$$

or, correspondingly, since $E_L = hc/\lambda_L$ with $hc = 1.239842447$ eV-microns, the laser wavelength

$$\lambda_L = 4hcE_e(E_e - E_x(0)) / [(mc^2)^2 E_x(0)] \quad (3)$$

In Fig. (2) we show λ_L vs. E_e for $E_x(0) = 33.169$ keV, the K-absorption edge of iodine. Several points of interest are: $\lambda_L = 1.000\mu$ for $E_e = 41.81$ -MeV Linac energy considered by Blum[10], $\lambda_L = 10.6\mu$ at $E_e = 136.1$ -MeV for the CO₂ laser and the FEL energy used in this report, $\lambda_L = 22.900\mu$ for $E_e = 200$ -MeV energy of the NSLS SXLS Phase I compact storage ring, and $\lambda_L = 100\mu$ at $E_e = 417.92$ -MeV ring energy considered by Blum[10]. It is not clear at this point what the optimum electron energy should be -- a lower ring energy implies a smaller and less costly injector which, for the 136-MeV ring, could be a Microtron whereas the higher (~400-MeV) energy ring will have the important advantage of radiation damping of the transverse and longitudinal emittance blow-up caused by the electron-photon collisions.

II. ENERGY OF BACKSCATTERED PHOTONS vs. SCATTERING ANGLE θ .

For the desired energy of the backscattered X-Ray at $\theta = 0$, $E_x(0) = 33.169$ keV, and the desired electron beam energy E_e , Eq. (2) gives the required laser photon energy E_L . Then we can evaluate the variation of scattered photon energy $E_x(\theta)$ as a function of scattering angle θ . It is clear that $E_x(\theta)$ is a decreasing function of θ . In Fig. (3) we show this variation for the four electron energies $E_e = 41.81, 136.1, 200$ and 417.92 MeV. We see that E_x drops rapidly with θ at the higher electron energies, giving an energy spread much larger than the maximum ($\Delta E_x \approx 1$ -keV) which could be contemplated [11] for Digital Subtraction angiography without use of a monochromator. Values of the energy spreads $\Delta E_x = E_x(0) - E_x(\theta = \theta_c)$ for the half angle of the scattering cone $\theta = \theta_c$ defined by a collimator hole in the laser cavity mirror are given in Table I. Since the electron beam RMS angle spread σ_e' for a 200-MeV ring such as the SXLS Phase I ring [12] is about 0.1-milliradian, values of the half cone angle much less than $\theta_c \approx 0.5$ -mrad could not be realized and we could not consider a ring of energy much greater than about 200-MeV without use of a monochromator. Note that the RMS angle spread in the laser photon beam $\sigma_L' = .777/\gamma = 2.92$ -mrad at 136.1-MeV would not be of concern in this regard since, as Sandorfi, et.al.[1] point out, the scattered X-Ray energy is only weakly dependent on angle ϕ between the electron and incident laser photon.

TABLE I
Electron Energy E_e (MeV) Backscattered Photon Energy Spread ΔE_x (eV)

	θ_c		
	$\theta_c = 0.1$ -mrad	$\theta_c = 0.5$ -mrad	$\theta_c = 1$ -mrad
41.81	2	55	220
136.10	23	578	2197
200.00	51	1223	4405
417.92	220	4752	13293

III. CROSS-SECTION AND X-RAY RATES INTO CONE OF HALF ANGLE θ_c .

The LEGS proposal[1] gives for the CM differential Compton scattering cross-section at lab angles θ and ϕ_x of the scattered X-Ray

$$d\sigma/d\Omega' = r_0^2 (E'_x/E'_L)^2 [E'_x/E'_L + E'_L/E'_x - 2 \sin^2\theta' \cos^2\phi_x] / 2 \quad (4)$$

where r_0 = classical electron radius = $e^2/mc^2 = 2.81794092 \times 10^{-13}$ cm, E'_x is the CM energy of the scattered X-Ray, E'_L is the CM energy of the incident laser photon, and θ' is the CM angle between the incident electron and scattered photon direction corresponding to lab angle θ and is given by

$$\cos\theta' = \gamma(\cos\theta - \beta) / [\sin^2\theta + \gamma^2(\cos\theta - \beta)^2]^{1/2} \quad (5)$$

and the ratio of scattered to incident photon CM energies is given as

$$E'_x/E'_L = 1/[1 + 2\gamma E_L(1 + \cos\theta')/mc^2]. \quad (6)$$

The lab differential cross section is then

$$d\sigma/d\Omega = (d\sigma/d\Omega') (d\Omega'/d\Omega) = J(\theta) d\sigma/d\Omega' \quad (7)$$

where the Jacobian of the CM to lab solid angle transformation is

$$J(\theta) = \gamma[\sin^2\theta + \cos\theta(\cos\theta - \beta)] / [\sin^2\theta + \gamma^2(\cos\theta - \beta)^2]^{3/2} \quad (8)$$

The lab cross-section for scattering a photon into a cone of half angle θ_c is then, after integration over azimuthal angle ϕ_x ,

$$\sigma(\theta_c) = \int_0^{2\pi} \int_0^{\theta_c} J(\theta) d\sigma/d\Omega' \sin\theta d\theta d\phi_x = \pi r_0^2 \int_0^{\theta_c} (E'_x/E'_L)^2 [E'_x/E'_L + E'_L/E'_x - \sin^2\theta'] J(\theta) \sin\theta d\theta \quad (9)$$

Eq. (7) is evaluated in SUBROUTINE COMPUTE making use of Eqs. (4), (5), (6) and (8), and Eq. (9) is then evaluated in PROGRAM COMPSCAT, given in Appendix A, using Simpson's Rule integration in FUNCTION SIMP.

Also given in the LEGS proposal and evaluated by COMPSCAT is the rate of X-Ray photons scattered into the cone of half angle θ_c . In practical units the production rate is

$$dN_x/dt = 2.604 I_e P_L \sigma(\theta_c) L_L / E_L A \quad \text{sec}^{-1} \quad (10)$$

where I_e is the ring electron beam current in Amps, P_L is the peak laser bunch power in Watts, $\sigma(\theta_c)$ is the cross-section of Eq. (9) in millibarns, L_L is the length of the incident laser bunch in centimeters, E_L is the incident laser energy in eV, and A is the effective overlap area of the photon and electron bunch at the interaction point in cm² given by

$$A = 2\pi [\sigma_L^2 + \sigma_{ex}^2]^{1/2} [\sigma_L^2 + \sigma_{ey}^2]^{1/2} \quad (11)$$

Eq. (10) assumes that every electron bunch in the ring collides on each revolution with a photon bunch, and can be derived from the event rate for a colliding beam machine

$$dN/dt = \mathcal{L} \sigma \quad (12)$$

where the luminosity $\mathcal{L}$ is given by

$$\mathcal{L} = f N_L N_e / A_{eff} \quad (13)$$

with f the bunch collision frequency (the RF frequency of the ring in the present case), N_L the number of laser photons per bunch, N_e the number of electrons per bunch and, for round beams of Gaussian profile it can be shown[13] that $A_{eff} = 2\pi(\sigma_L^2 + \sigma_e^2)$. For both beams having the same RMS radius, $A_{eff} = 4\pi\sigma^2$. For the results of Table II I have used the theoretical beam properties of the SCLS Phase I machine scaled from $E_e = 200$ -MeV to 136.1-MeV. From the basic expression for the damped emittance given by Sands[14] and further elucidated by Helm, et. al. [15]

$$\mathcal{E}_e = (55/32\sqrt{3}) (hc/mc^2) [E_e/mc^2]^2 I_5 / (I_2 - I_4) \quad (14)$$

where the synchrotron radiation integrals are

$$\begin{aligned} I_2 &= \oint ds/\rho^2 \\ I_3 &= \oint ds/|\rho^3| \\ I_4 &= \oint (1-2\eta)\eta/\rho^3 ds \\ I_5 &= \oint [\beta_x \eta_x'^2 + 2\alpha_x \eta_x \eta_x' + \eta_x^2 (1 + \alpha_x^2)/\beta_x] / |\rho^3| ds \end{aligned} \quad (15)$$

the dependence of emittance on ring energy is E_e^2 and the 136.1-MeV emittance is $\mathcal{E}_e = 2.74 \times 10^{-8}$ m-rad scaled from the 200-MeV value[12] $\mathcal{E}_e = 5.92 \times 10^{-8}$ m-rad. With the further assumption of 100% coupling we have $\mathcal{E}_e = \mathcal{E}_{ex} + \mathcal{E}_{ey}$. I then get $\mathcal{E}_{ex} = \mathcal{E}_{ey} = 1.37 \times 10^{-8}$ m-rad at 136.1-MeV. The electron beam sizes are then obtained from the relations[12]

$$\begin{aligned} \sigma_{ex} &= [\mathcal{E}_{ex} \beta_x + (\frac{\sigma_E}{E})^2 \eta_x^2]^{\frac{1}{2}} = 6.929 \times 10^{-4} \text{ m} \\ \sigma_{ey} &= [\mathcal{E}_{ey} \beta_y]^{\frac{1}{2}} = 1.460 \times 10^{-4} \text{ m} \end{aligned} \quad (16)$$

where I have used the same betatron functions as pertain for the 200-MeV ring[16], $\alpha_x=0$, $\beta_x=2.479$ -m, $\eta_x=1.415$ -m, $\beta_y=1.555$ -m and $(\sigma_E/E)=6.9373 \times 10^{-4}$ scaled down to $(\sigma_E/E)=4.7208 \times 10^{-4}$ with the E_e dependence[15]

$$(\sigma_E/E)^2 = (55/32\sqrt{3}) (hc/mc^2) [E_e/mc^2]^2 [I_3/(2I_2 + I_4)] \quad (17)$$

For the laser bunch I used a round beam of the same area as the rectangular electron bunch and obtain

$$\sigma_L = [4\sigma_{ex}\sigma_{ey}/\pi]^{\frac{1}{2}} = 3.590 \times 10^{-4} \text{ m} \quad (18)$$

For the length of the laser bunch L_L in Eq. (10) I use the length of the electron bunch $L_e = 2\sigma_{eL}$ with $\sigma_{eL} = 1.6$ -cm at 200-MeV[12] scaled down to 136.1-MeV as $L_e = 2.178$ -cm from the relation

$$\sigma_{eL} = \alpha_c [C_g/J_g f]^{\frac{1}{2}} [E_e/mc^2] / \Omega_s \quad (19)$$

Finally, for the peak power in the laser bunch P_L I assume that the number of photons which can be obtained in an FEL bunch is the same as that contained in a CO_2 laser bunch of energy $E_{CO_2} = 100$ -mJ/bunch as quoted by Blum[10]. At a photon energy of $E_L = 0.116966269$ -eV (10.6 μ) I obtain $N_L = 5.3367 \times 10^{18}$ photons/bunch in a volume given by a cylinder of dimensions in Eqs. (18) and (19) which gives a peak power of $P_L = 1.375678$ -GW. For the electron beam current in Eq. (10) I use a value of $I_e = .5$ -Amps which is approximately the value attained in 6-bunch operation of the SCLS Phase I machine as summarized by Murphy[17] and presented here as Table III. This current gives 8.8516×10^{10} electrons in the 8.5-m circumference ring or, with 6 bunches in the ring, $N_e = 1.4753 \times 10^{10}$ electrons/bunch. It is likely that this number could be considerably larger with a

better injector than used by the SXLS.

In Table II we show results from program COMPSCAT for the X-Ray rate dN_x/dt and average number of X-Rays per 0.5-mm x 0.5-mm pixel delivered in a 4-msec pulse over a circular area defined by the collimator half angle θ_c . It is seen that the number of photons/pixel does not vary strongly with θ_c , whereas the energy spread in the X-Ray beam and area of the circle irradiated by the X-Rays increases with θ_c . At $\theta_c=1$ -mrad the number of photons/pixel at 10-m from the photon-electron interaction point is more than 10 times the $3 \cdot 10^7$ photons/pixel presently attained [4] in the X-17 NSLS wiggler beam, and over an area about 5 times as large as irradiated by the X-17 beam at 36-m from the interaction point. However, the energy spread in the Compton backscattered beam is 2.197-keV as compared to the 70-eV for the X-17 beam. At $d=36$ -m the Compton backscattered X-Ray circle would be 72-mm in diameter and the average number of photons/pixel drops to $2.67 \cdot 10^7$.

It should also be noted that Blum[10] made the additional important observation that the event rate would increase if a low-beta insertion were used in the FEL undulator straight-section (higher luminosity) since the bunch cross section area in the denominator of Eq.(10) would decrease, i.e. σ_{ex} and σ_{ey} of Eq.(16) would be smaller with smaller β_x and β_y . He used smaller β 's in his rate estimates than are used here; the SXLS values of $\beta_x=2.479$ -m and $\beta_y=1.555$ -m are not particularly small compared, for example, to the $\beta=10$ -cm used in modern colliding beam machines such as the Fermilab Tevatron. Other factors which may increase dN_x/dt are higher electron current I_e (the XLS achieved 1.2-Amps at 200-MeV), higher FEL photon energy content (the CO₂ laser used by the BNL-ATF facility was designed at Los Alamos for 250-mJ/pulse), and multiple photon-electron bunch collisions in the undulator straight-section (see Fig.8) if the oscillator cavity is many RF wavelengths long.

TABLE II
X-RAY PRODUCTION CROSS-SECTION, PHOTON RATES, AND ENERGY SPREAD INTO CONE OF HALF-ANGLE . PROGRAM COMPSCAT.

$E_L=1.16966269$ -eV $\lambda_L=10.6\mu$ $E_e=136.1$ -MeV $I_e=.5$ -A $N_L=5.3367E18$ /bunch
 $P_L=1.375678$ -GW $\sigma_{Lx}=1.089$ -cm $\sigma_{ex}=.6929$ -mm $\sigma_{ey}=.146$ -mm $\sigma_L=.359$ -mm

θ_c (mrad)	σ (mb)	dN_x/dt (sec ⁻¹)	ΔE_x (eV)	diam@10-m (mm)	$\overline{N_x}$ /pixel @10-m/4-msec
0.1	0.71	$1.240 \cdot 10^{12}$	23	2	$39.470 \cdot 10^7$
0.2	2.81	4.938	94	4	39.296
0.3	6.29	11.034	210	6	39.024
0.4	11.07	19.426	372	8	38.646
0.5	17.08	29.98	578	10	38.172
0.6	24.23	42.52	826	12	37.596
0.7	32.40	56.88	1114	14	36.950
0.8	41.49	72.82	1441	16	36.218
0.9	51.36	90.14	1802	18	35.422
1.0	61.88	108.62	2197	20	34.574
1.1	72.94	128.02	2622	22	33.678
1.2	84.40	148.14	3075	24	32.746
1.3	96.16	168.76	3551	26	31.786
1.4	108.10	189.72	4049	28	30.812
1.5	120.12	210.8	4566	30	29.822
1.6	132.15	232.0	5098	32	28.846
1.7	144.10	253.0	5644	34	27.866
1.8	155.91	273.6	6200	36	26.880
1.9	167.53	294.0	6763	38	25.924
2.0	178.91	314.0	7333	40	24.988

TABLE III
OPERATIONAL STATUS OF THE BNL SXLS PHASE I RING, MARCH 1992

ELECTRON ENERGY (MeV)	AV. CURRENT (mA)	MAX. CURRENT (mA)	# BUNCHES
200	900	1200	6
200	900	1200	3
200	300	450	2
200	400	500	1
160	600	762	6
120	NA	360	6
80	180	270	6

Maximum stored current of the machine is limited by the short lifetimes at low energy and the injector performance. The study of low energy injection is limited by the performance of the injection system.

IV. ENERGY SPECTRUM OF THE BACKSCATTERED X-RAYS

In Fig. (3) we have seen that the X-Ray energy decreases with increasing scattering angle θ . However, the lab X-Ray differential cross-section of Eq. (7) does not decrease strongly with θ as seen in the output of program COMPSCAT in Table IV; thus the photon yield increases with θ_c as seen in Table II. The weak dependence of lab cross-section on θ suggests that the number of X-Rays per unit energy interval is also relatively constant over the angle range θ_c accepted by the collimator. Thorn, et.al. [18] have obtained an approximate expression for the X-Ray spectrum as

$$d\sigma/dE_x = 2\pi r_0^2 E_L (2\gamma/(1+x^2))^2 [y + 1/y - (2x/(1+x^2))^2] / E_x^2 y^2 \quad (20)$$

$$\text{where } x \equiv \gamma\theta = [4\gamma E_L (\gamma/E_x - 1/mc^2) - 1]^{1/2}$$

$$\text{and } y \equiv 1 + 4\gamma E_L / mc^2 (1+x^2)$$

This expression is also evaluated by COMPSCAT and given in Table IV. The spectrum is seen to decrease by only 12% over the 1-mrad collimator acceptance angle θ_c . The spectrum is plotted in Fig. (4) together with the absorption coefficient for iodine, bone and water in the region near the iodine K-edge as computed by D. Chapman [11] in Fig. (5).

V. ENERGY SPREAD IN ELECTRON BEAM FROM COMPTON SCATTERING.

Blum [10] has made the important observation that the Compton scattering process will induce a large energy spread into the circulating electron beam of the ring such that a significant fraction of the electrons may exceed the energy acceptance of the ring and will be lost, i.e. the electrons will "random walk" out of the RF bucket after multiple scatterings. To estimate the magnitude of this effect I note from Table II that, with a $\theta_c = 1$ -mrad collimator opening, the scattered X-Ray rate is $1.086 \cdot 10^{14}$ /sec which is $4.344 \cdot 10^{14}$ photon-electron collisions during the 4-msec pulse. The total number of collisions which occurred during this time interval is larger since the integrated cross-section in the collimator cone is only 61.88-mb whereas the total scattering cross-section over the entire 4π -steradians, given by the Klein-Nishina formula [19] is about 665-mb. Bethe and Ashkin [20] give the Klein-Nishina total cross section result in the electron rest frame as

$$\sigma_{CT} = (3\sigma_0/8r) \{ [1-2(r+1)/r^2] \ln(2r+1) + 1/2 + 4/r - 1/2(2r+1)^2 \} \quad (21)$$

$$\text{where } r = E_L'/mc^2 \text{ and } \sigma_0 = \text{Thompson cross-section} = 8\pi r_0^2/3 = 665.2457 \text{ mb}$$

TABLE IV. CROSS-SECTION, ENERGY AND SPECTRUM OF X-RAYS. $E_e = 136.1\text{-MeV}$

PROGRAM COMPSCAT-13-APR-92 11:11:55 cross-section and backscattering rate, 33.

INPUT DATA

EL(eV) = 0.116925000 GAMMA=266.341000 THETAC(mrad) = 1.00 N = 40

RESULTS

THETA(MR)	THETAP(RAD)	DSDOMP(MB/SR)	DSDOM(MB/SR)	EG(KEV)	DSDEX(MB/KEV)
0.000	0.00000	0.7936905E+02	0.225208E+08	33.169	30.0767
0.025	0.01332	0.7936201E+02	0.225168E+08	33.168	30.0740
0.050	0.02663	0.7934091E+02	0.225049E+08	33.164	30.0660
0.075	0.03995	0.7930577E+02	0.224849E+08	33.156	30.0527
0.100	0.05326	0.7925663E+02	0.224570E+08	33.146	30.0341
0.125	0.06656	0.7919354E+02	0.224213E+08	33.133	30.0102
0.150	0.07986	0.7911656E+02	0.223777E+08	33.117	29.9810
0.175	0.09315	0.7902577E+02	0.223263E+08	33.098	29.9466
0.200	0.10644	0.7892129E+02	0.222672E+08	33.076	29.9070
0.225	0.11971	0.7880320E+02	0.222005E+08	33.051	29.8622
0.250	0.13297	0.7867164E+02	0.221263E+08	33.023	29.8124
0.275	0.14623	0.7852675E+02	0.220447E+08	32.992	29.7574
0.300	0.15947	0.7836867E+02	0.219557E+08	32.959	29.6975
0.325	0.17269	0.7819758E+02	0.218596E+08	32.923	29.6327
0.350	0.18590	0.7801364E+02	0.217565E+08	32.884	29.5629
0.375	0.19909	0.7781705E+02	0.216464E+08	32.842	29.4884
0.400	0.21227	0.7760802E+02	0.215296E+08	32.797	29.4092
0.425	0.22543	0.7738675E+02	0.214063E+08	32.750	29.3253
0.450	0.23857	0.7715347E+02	0.212765E+08	32.700	29.2369
0.475	0.25169	0.7690842E+02	0.211405E+08	32.647	29.1440
0.500	0.26478	0.7665185E+02	0.209984E+08	32.592	29.0468
0.525	0.27786	0.7638401E+02	0.208505E+08	32.533	28.9453
0.550	0.29091	0.7610517E+02	0.206969E+08	32.473	28.8396
0.575	0.30393	0.7581560E+02	0.205379E+08	32.409	28.7298
0.600	0.31693	0.7551559E+02	0.203736E+08	32.344	28.6161
0.625	0.32990	0.7520544E+02	0.202042E+08	32.275	28.4985
0.650	0.34284	0.7488543E+02	0.200300E+08	32.204	28.3772
0.675	0.35576	0.7455589E+02	0.198512E+08	32.131	28.2523
0.700	0.36864	0.7421713E+02	0.196679E+08	32.055	28.1239
0.725	0.38150	0.7386945E+02	0.194805E+08	31.977	27.9921
0.750	0.39432	0.7351320E+02	0.192892E+08	31.897	27.8571
0.775	0.40711	0.7314670E+02	0.190941E+08	31.814	27.7189
0.800	0.41987	0.7277628E+02	0.188955E+08	31.729	27.5778
0.825	0.43259	0.7239628E+02	0.186936E+08	31.642	27.4338
0.850	0.44527	0.7200905E+02	0.184887E+08	31.553	27.2870
0.875	0.45792	0.7161494E+02	0.182810E+08	31.461	27.1376
0.900	0.47053	0.7121428E+02	0.180706E+08	31.367	26.9857
0.925	0.48311	0.7080743E+02	0.178579E+08	31.272	26.8315
0.950	0.49564	0.7039473E+02	0.176430E+08	31.174	26.6751
0.975	0.50814	0.6997654E+02	0.174262E+08	31.074	26.5166
1.000	0.52059	0.6955322E+02	0.172077E+08	30.973	26.3561

Now, enter IE, PL, LL, SIGL, SIGXE, SIGYE (all real)

.5.1375678000. .2.178. .06 359. .06929. .0146

IE(amp), PL(watt), LL(cm), SIGL(cm), SIGXE(cm), SIGYE(cm) = 0.500 1375678000.00 2.1

SIGMA = 61.88 MB

DNGDT = 0.10865E+15 PHOTONS/SEC

ENTER HEADER(ABO) =

I evaluated Eq. (21) in program SIGCOMPT given in Appendix B for σ_{CT} vs. E_L' as shown in Fig. (6). In the electron rest frame (CM) $E_L' = 2\gamma E_L$ and at $E_e = 136.1\text{-MeV}$, $\gamma = 266.341$ and, for our $E_L = 1.16966269\text{-eV}$ laser photon, $E_L' = 62.306\text{-eV}$. At this energy $\sigma_{\text{CT}} = 665.08\text{-mb}$, i.e. the total cross-section is approximately σ_0 at low photon energy. Note also that the total cross-section is an invariant, i.e. its the same in the lab and CM systems. Thus the ratio $\sigma_{\text{CT}}/\sigma_0(\theta_e)$ is 10.75 and the total number of photon-electron collisions in 4-msec becomes $46.70 \cdot 10^{10}$. But, with an electron beam of $I_e = 5\text{-Amp}$ in the 8.5-m circumference ring, there are only $8.85 \cdot 10^{10}$ electrons. So we have on average about 53 collisions per electron in the 4-msec interval. If the average energy loss per collision is $E_x(\theta=0)/2$ where we are envisioning an X-Ray spectrum from 35.6-keV to 33.2-keV, then we would have an average energy loss of 17.8-keV/collision which becomes an average total energy loss of 939-keV/electron in 4-msec. This is indeed a large effect; therefore I have evaluated the contour of the RF bucket as given by Livingood[21] to see if the energy acceptance of the ring can accommodate this spread. The bucket equation is given as

$$\Delta E = \pm(E_s \Omega / h \omega_s T) \{2[\cos \phi + \phi \sin \phi_s + \cos \phi_s - (\pi - \phi_s) \sin \phi_s] / \cos \phi_s\} \quad (22)$$

where ϕ_s is the RF phase angle of the synchronous particle, E_s is the kinetic energy and ω_s the angular rotation frequency of the synchronous particle ($\omega_s = 2\pi f_R$), and ϕ is the RF phase angle of the off-energy particle of energy E with $\Delta E = E - E_s$. In Eq. (22) the angular frequency of small amplitude oscillations about the stable phase ϕ_s is given by

$$\Omega = [-\alpha V_m h \omega_s^2 T \cos \phi_s / 2\pi E_s]^{1/2} \quad (23)$$

where h is the harmonic number of the ring (number of RF buckets per revolution), V_m is the amplitude of the RF voltage, and

$$T = (\alpha - 1/\gamma_s^2) / \beta_s^2 \quad (24)$$

with α the momentum compaction factor $\alpha = (dL/L)/(dp/p)$ with L the ring circumference, p the particle momentum, and γ_s and β_s the usual relativistic factors $\gamma_s = 1 + E_s/mc^2 = [1 - \beta_s^2]^{-1/2}$. Note that we use the Courant-Snyder[22] definition of momentum compaction in Eq. (24); Livingood uses the reciprocal. Eqs. (22)-(24) are evaluated in Program BUCKET given in Appendix C. The results are plotted in Fig. (7) for the cases $E_e = 200\text{-MeV}$, $V_m = 20\text{-kV}$; $E_e = 136.1\text{-MeV}$, $V_m = 20\text{-kV}$; and $E_e = 136.1\text{-MeV}$, $V_m = 100\text{-kV}$. The 200-MeV case is also tabulated in Table V. We see that, for the SXLS case at $E_e = 200\text{-MeV}$, the maximum bucket size is $(\Delta E)_{\text{max}} = \pm 1.141\text{-MeV}$ and the stable phase angle (actually $\pi - \phi_s$) is 0.67° , in excellent agreement with the results of Murphy and Vignola in the SXLS Conceptual Design Report[12]. For the case we are considering at $E_e = 136.1\text{-MeV}$ with $V_m = 20\text{-kV}$ the calculation gives $(\Delta E)_{\text{max}} = \pm 948\text{-keV}$ which is indeed very close to the average energy loss per electron in 4-msec estimated above. The solution to the problem is to increase the peak RF voltage. At $V_m = 100\text{-kV}$ which is attainable with the SXLS Phase II cavity, the bucket size is $(\Delta E)_{\text{max}} = \pm 2.124\text{-MeV}$, which appears safely larger than the average energy loss expected from Compton scattering. The higher RF voltage will also be advantageous in increasing the beam Touschek lifetime which is also proportional to $V_m^{1/2}$. Another point to note is that the phase oscillation frequency for this case, $f_s = \Omega/2\pi$ is 528.3-kHz; thus the electron will make many phase oscillations during the 4-msec pulse and the scattering event is equally likely to occur when the particle energy is greater than the synchronous energy E_s . Thus each collision will not necessarily move the particle energy away from E_s and the energy losses will therefore not be additive.

0.8353	0.460	2.3060	1.044	3.7768	1.081	5.2475	0.533
0.8500	0.468	2.3208	1.047	3.7915	1.078	5.2622	0.525
0.8647	0.476	2.3355	1.051	3.8062	1.075	5.2769	0.517
0.8794	0.483	2.3502	1.054	3.8209	1.072	5.2916	0.509
0.8941	0.491	2.3649	1.057	3.8356	1.069	5.3063	0.501
0.9088	0.499	2.3796	1.060	3.8503	1.066	5.3210	0.493
0.9236	0.506	2.3943	1.063	3.8650	1.063	5.3358	0.484
0.9383	0.514	2.4090	1.067	3.8797	1.060	5.3505	0.476
0.9530	0.521	2.4237	1.069	3.8944	1.057	5.3652	0.468
0.9677	0.529	2.4384	1.072	3.9091	1.054	5.3799	0.459
0.9824	0.536	2.4531	1.075	3.9238	1.050	5.3946	0.451
0.9971	0.544	2.4678	1.078	3.9386	1.047	5.4093	0.442
1.0118	0.551	2.4825	1.081	3.9533	1.044	5.4240	0.434
1.0265	0.558	2.4972	1.083	3.9680	1.040	5.4387	0.425
1.0412	0.566	2.5119	1.086	3.9827	1.036	5.4534	0.416
1.0559	0.573	2.5267	1.089	3.9974	1.033	5.4681	0.407
1.0706	0.580	2.5414	1.091	4.0121	1.029	5.4828	0.398
1.0853	0.588	2.5561	1.094	4.0268	1.026	5.4975	0.390
1.1000	0.595	2.5708	1.096	4.0415	1.022	5.5122	0.380
1.1147	0.602	2.5855	1.098	4.0562	1.018	5.5269	0.371
1.1295	0.609	2.6002	1.101	4.0709	1.014	5.5417	0.362
1.1442	0.616	2.6149	1.103	4.0856	1.010	5.5564	0.353
1.1589	0.623	2.6296	1.105	4.1003	1.006	5.5711	0.343
1.1736	0.630	2.6443	1.107	4.1150	1.002	5.5858	0.334
1.1883	0.638	2.6590	1.109	4.1298	0.998	5.6005	0.324
1.2030	0.645	2.6737	1.111	4.1445	0.994	5.6152	0.315
1.2177	0.651	2.6884	1.113	4.1592	0.990	5.6299	0.305
1.2324	0.658	2.7031	1.115	4.1739	0.985	5.6446	0.295
1.2471	0.665	2.7178	1.116	4.1886	0.981	5.6593	0.285
1.2618	0.672	2.7326	1.118	4.2033	0.977	5.6740	0.274
1.2765	0.679	2.7473	1.120	4.2180	0.972	5.6887	0.264
1.2912	0.686	2.7620	1.121	4.2327	0.968	5.7034	0.253
1.3059	0.692	2.7767	1.123	4.2474	0.963	5.7181	0.242
1.3207	0.699	2.7914	1.124	4.2621	0.958	5.7329	0.231
1.3354	0.706	2.8061	1.126	4.2768	0.954	5.7476	0.220
1.3501	0.712	2.8208	1.127	4.2915	0.949	5.7623	0.208
1.3648	0.719	2.8355	1.128	4.3062	0.944	5.7770	0.196
1.3795	0.726	2.8502	1.130	4.3209	0.939	5.7917	0.183
1.3942	0.732	2.8649	1.131	4.3357	0.935	5.8064	0.170
1.4089	0.738	2.8796	1.132	4.3504	0.930	5.8211	0.156
1.4236	0.745	2.8943	1.133	4.3651	0.925	5.8358	0.142
1.4383	0.751	2.9090	1.134	4.3798	0.920	5.8505	0.126
1.4530	0.758	2.9238	1.135	4.3945	0.915	5.8652	0.108
1.4677	0.764	2.9385	1.136	4.4092	0.909	5.8799	0.087
1.4824	0.770	2.9532	1.136	4.4239	0.904	5.8946	0.061

ENTER NAME OF OUTPUT FILE OUTFILE(A12)=

'xls200.buck'

ENTER HEADER(AB0)=

It seems clear that a Monte-Carlo calculation will be required to access the dilution of longitudinal (energy) phase space as well as transverse phase space (and thus beam transverse dimensions σ_x and σ_y). It will also be necessary to design the physical aperture of the storage ring to accommodate the larger beam size resulting from the collisional energy spread estimated above. The energy spread gives rise to a horizontal size increase of $\Delta\sigma_x = \eta_x (\sigma_E/E)$ where η_x is the dispersion function and the effect is included in the expression for horizontal size σ_x in Eq. (16). For the SCLS the maximum η_x is only 1.4-m and the resulting beam size increase is well within the acceptance aperture. For the Angiography ring the energy spread (σ_E/E) is much larger. Another important point is that we will want $\eta_x = 0$ in the FEL undulator straight-section since a large beam, as previously noted, will decrease the collision rate from Eq. (10). Also, if we choose to change the X-Ray energy by changing the ring electron energy E_e it will be important that $\eta_x = 0$ in the FEL so that the $\theta = 0$ direction remains fixed.

VI. LASER BEAM HEATING OF THE FEL OSCILLATOR CAVITY MIRRORS.

Another important observation by Blum[23] is heat dissipation in the mirrors at each end of the optical cavity. The energy content of each 2.178-cm laser bunch used for the present calculation is, as previously noted, 0.1-Joule. The pulses are incident on the mirror at a 211.55-MHz rate. Thus, in 4-msec, the energy incident on a mirror is 84,620-J. It is therefore very important that the reflectivity of the mirrors be very high. At $E_e = 0.1$ -eV, Weaver[24] gives the normal incidence reflectivity of the best metallic reflectors, gold and silver, as $R = 0.995$. We would thus have an energy loss of 423-J per 4-msec pulse. If this energy is distributed over a circle determined by the angular divergence of the FEL laser photon beam of $\sigma'_L = .777/\gamma = 2.92$ -mrad for 136.1-MeV electrons (see discussion of angle spread in the laser beam in Section VII below), and the optical cavity half length is taken to be 4 RF wavelengths = 5.6684-m, then the RMS radius of the laser beam at the mirror is 1.653-cm and the RMS area is 8.583-cm². The average areal energy density absorbed by the mirror is then 49.3-J/cm². This appears to be an excessively high loss, particularly if we envision a pulse rate of ~ 10 per second to visualize the time dependence of the blood flow through the artery. We can reduce the absorbed energy density by decreasing the electron energy and thus making the laser beam angular divergence larger; this causes no penalty in X-Ray energy since, as pointed out by Sandozfi, et.al.[1] the scattered X-Ray kinematics is essentially independent of angle between the incident laser photon and the electron -- its unmeasurably small and in fact has been dropped from Eq. (1) and the cross-section expressions used here. However, the energy and transverse damping times of the electron beam increase cubically with decreasing electron energy E_e as seen from the expressions[15]

$$\tau_i = (30/J; c r_0) [mc^2/E_e]^3 / I_L \quad i=x,y,E \quad (25)$$

and, at 136.1-MeV are already excessively large to achieve any beam damping in the interval between X-Ray pulses if we desire to approach ~ 10 pulses/sec. Also, we would want to reinject electrons between X-Ray pulses and damp their oscillations before the next X-Ray pulse to restore losses. At 200-MeV the damping times are[12] $\tau_x = 91$ -msec, $\tau_y = 48$ -msec, and $\tau_E = 20$ -msec. Thus, we must explore possible means to improve the mirror reflectivity, either by going to higher laser wavelength (higher E_e) or cooling the mirrors to obtain lower electrical resistivity. The temperature coefficient of resistivity for Silver is quite large, .0038/ $^{\circ}\text{C}$ [24] so an appreciable improvement in reflect-

ivity could be realized at liquid nitrogen temperature. Indeed, Williams[25] has pointed out that the reflectivity will go to 100% for a superconducting mirror for laser photon energies below the superconducting energy gap, but it does not seem possible in the present application to provide the necessary super-insulation in the FEL optical cavity and still allow the infrared radiation to impinge on the mirror. Perhaps a high T_c superconducting film evaporated onto a suitable substrate could provide the needed reflectivity[25].

It should be noted that I am using straight line approximations for the laser spot size in the above estimate rather than the parabolic dependence of the beam envelope in a drift region[35]. Also, the expressions used for damping times, damped emittance, etc. in this report are based on single-particle beam dynamics theory and will undoubtedly be worse in a real machine at high current where space-charge, ion trapping and intrabeam scattering effects are important.

VII. ANGLE SPREAD OF FEL UNDULATOR BEAM. TUNABILITY OF THE X-RAY ENERGY

The wavelength of the spontaneous radiation from the FEL undulator is given by Krinsky[26] as

$$\lambda_n = \lambda_u (1 + K^2/2 + \gamma^2 \theta^2) / 2n\gamma^2 \quad n=1,2,3,4,5,\dots \quad (26)$$

where the undulator parameter K , dependent on length of the period of the undulator magnetic field λ_u and the peak strength of the assumed sinusoidal field variation along the undulator axis B_0 is given by(cgs)

$$K = e \lambda_u B_0 / 2\pi mc^2 \quad (27)$$

Only the odd harmonics $n=1,3,5,\dots$ are present in the forward direction ($\theta=0$) but all harmonics are present for non-zero θ . For the fundamental peak ($n=1$) Krinsky writes for the width of the forward peak

$$(\Delta\lambda/\lambda)_0 \approx 1/N_u \quad (28)$$

where N_u is the number of periods of the undulator field. If one accepts radiation in a cone of half-angle θ_u , however, Krinsky writes for the additional broadening of the line as

$$(\Delta\lambda/\lambda)_{add} = \gamma^2 \theta_u^2 (1 + K^2/2) \quad (29)$$

If one wants to preserve the narrow wavelength (energy) width of the peak in the presence of an angle spread which also changes the undulator wavelength via Eq. (26) it would be necessary to restrict the half-cone angle θ_u of the radiation to $\gamma\theta_u = [(1+K^2)/N_u]^{1/2}$. But since we are accepting a large spread in X-Ray energy from the angle variation of $E_x(\theta)$ in Eq. (1) of $\Delta E_x/E_x \approx 2.2\text{-keV}/34.3\text{-keV} \approx 6.4\%$ it is not unreasonable to accept an additional spread $\Delta E_x/E_x$ from $\Delta E_L/E_L$ ($\Delta E_x/E_x \approx \Delta E_L/E_L$ from Eq. (1)) of $\sim 1\%$. If we were to operate the undulator with the usual value of $K \approx 1$, then $\Delta\lambda_L/\lambda_L = \gamma^2 \theta_u^2 / 1.5 = .01$ which, at $E_e = 136.1\text{-MeV}$ and $\gamma = 266.341$, leads to $\theta_u = .46\text{-mrad}$. That's too small to give the large area circle of the laser beam on the mirror as required to reduce the absorbed energy in the mirror as discussed in Section VI above. We need $K \approx 10$ for the envisioned $N_u = 100$ undulator periods to attain $\gamma^2 \theta_u^2 \approx 1$. Such a large value of K should not be attained by increasing the period λ_u of the undulator field since we want large N_u to attain the large single-pass laser gain which is proportional to N_u^3 . This leaves only the possibility of increasing K by an increase in the undulator field B_0 which, for this application, may be possible since we desire a

pulsed undulator of short (4-msec) duration. A note of caution here is that in the present work I am using FEL laser beam properties based only on the undulator equations which consider only spontaneous radiation; I am not considering the properties of the stimulated emission photons which are the source of the gain we hope to achieve in the FEL. This question requires further study.

To attain $K \approx 10$ in Eq. (27) we must first establish what value of undulator wavelength λ_u is required to give the $\theta=0^\circ$ laser wavelength $\lambda_L=10.6\mu$ used as an example in this report. From Eqs. (26) and (27) we can, upon substitution of the expression for K in practical units $K=.934B_0(T)\lambda_u(\text{cm})$ obtain a cubic equation for λ_u as

$$.436178 B_0^2(T)\lambda_u^3 + \lambda_u - 2\gamma^2\lambda_L = 0 \quad (30)$$

The peak value of field B_0 which could be attained in a pulsed pole of high permeability iron surrounded by a conventional copper conductor can be estimated from experience with the H-10 ejection septum of the BNL AGS Fast Extracted Beam [32] where a current pulse of 20,000 Amps/cm is attained [33] in a thin, uncooled septum of ~ 0.5 -mm Cu. Using a solenoid approximation we have $H=2 \cdot 10^6$ -A/m = 25133 Oe. Using the magnetization curve [34] for the high permeability steel VITRANAMEL the magnetic induction is $B=46,000$ -Gauss [5], i.e. we still have a permeability $\mu=1.83$ at this field. Then Eq. (30) with $B_0=4.6$ -T yields $\lambda_u=2.535$ -cm, a small value compared to the $\lambda_u=5$ -cm suggested by Krinsky [26] or the $\lambda_u=6.5$ -cm used by Luccio [7]. But this is a pulsed application. The value of K given by this B_0 and λ_u is $K=10.8922$. If this value is substituted into Eq. (29) with $(\Delta\lambda/\lambda)_u=.01$ the undulator angle spread is $\gamma\theta_u=.77666$, i.e. $\theta_u=2.916$ -mrad. It is clear that such a strong undulator (wiggler?) will perturb the ring, principally from the well-known tune shift caused by the vertical fringe-field focussing of the array of dipoles comprising the undulator. I assume for purposes of this report that a first-order correction can be accomplished with pulsed quadrupoles at either end of the undulator, as noted also by Blum [23] and Murphy [37]. A strong undulator may also contribute to radiation damping, a very desirable property otherwise lacking in this ring. The energy dissipation per electron per pass through the undulator may be large compared to the 50-eV/electron/turn lost to synchrotron radiation in the ring bending magnets or the 77 eV/electron/turn average loss to Compton scattering. But the undulator is pulsed and we will also need damping of injected electrons between X-Ray pulses.

Another requirement of the Angiography method is that we be able to change the X-Ray energy band from above to below the iodine K-edge rapidly. This can be accomplished in either of two ways--change the laser energy of Eq. (1) by pulsing the undulator to a different field value B_0 or change the electron energy in the ring since the X-Ray energy depends quartically (see below) on electron energy E_e from Eq. (1). The latter method would be preferable since the electron energy in a ring can be changed rapidly via a change in the phase angle of the RF voltage in the ring RF cavity, or correspondingly change the RF frequency. The energy change occurs in relatively few cycles of the energy oscillator (synchrotron) frequency of the ring which, in our case is ~ 500 -KHz. Thus, the energy change could be accomplished in a small fraction of a millisecond. This method of energy change has been used successfully in the BNL AGS synchrotron for many years [27] to change the radial position of the electron beam. In our case, however, we desire that the electron beam position remain constant in the FEL straight section which implies that the momentum dispersion function $\eta_x=0$ in the FEL straight section, which is the case for the NSLS VUV

and X-Ray storage rings[28,29].

Another important point is that the X-Ray energy of Eq. (1) really varies as γ^4 since an additional factor of γ^2 is introduced by E_L via the energy dependence of the laser radiation in Eq. (26), i.e.

$$E_L = hc/\lambda_L = 2\gamma^2 hc/\lambda_u (1 + K^2/2) \quad \text{at } \theta=0 \quad (31)$$

which gives for the X-Ray energy at $\theta=0$

$$E_X(\theta=0) = 8hc\gamma^4/\lambda_u (1 + K^2/2) \quad (32)$$

From my estimate of K and λ_u above I get

$$E_X(\theta=0) = 6.48608 \cdot 10^{-9} \gamma^4 \text{ keV} \quad (33)$$

For example, to attain the $E_X = 90.526$ -keV of the ^{83}Bi K-absorption edge I would need $\gamma^4 = 1.39569663 \cdot 10^{10}$ which is an electron energy $E_e = 175.64$ -MeV, not a large energy change for a ring designed for 136.1-MeV. Just consider how large a storage ring would be to generate synchrotron radiation with a critical energy E_c that large! With $E_c(\text{keV}) = 0.665 E_e(\text{GeV}) B(T)$ [5,36] and with a $B=4.5$ -T wiggler, we would need $E_e = 5.5$ -GeV. Thus, Compton backscattering gives the possibility of exploring the use of elements heavier than ^{83}I for contrast agents at X-Ray energies where the absorption coefficient of water and bone is presumably much smaller than at 33-keV, thereby reducing the absorbed dose to the patient and perhaps eliminating altogether the need for digital subtraction. Eq. (33) also suggests a note of caution--the sensitivity of X-Ray energy to electron energy $E_e = \gamma mc^2$ is so large that we may have difficulty with tolerances on ring power supplies, RF frequency stability, etc. From Eq. (33) we get $dE_X/dE_e = 3.80506 \cdot 10^{-7} E_e^3(\text{MeV}) \text{ keV/MeV}$ which, at $E_e = 136.1$ -MeV gives $dE_X/dE_e = .96 \text{ keV/MeV}$. Thus, to maintain E_X to 0.1-keV we will have to hold the beam energy to 0.1-MeV, i.e. $\Delta E_e/E_e = 7.3 \cdot 10^{-4}$. This seems achievable but certainly not easy--we designed the NSLS storage rings with power supply tolerances of 1 part in 10^4 .

VIII. DISCUSSION AND CONCLUSIONS

The above calculations indicate that an intense X-Ray beam can be generated by Compton backscattering of infrared photons produced by an IR-FEL in a low energy electron storage ring. The major assumption is that the single pass gain and saturated power of the FEL will be sufficiently large such that the laser bunch will attain a 100-mJ energy content ($5.53 \cdot 10^{18}$ photons at $\lambda_L = 10.6\mu$) presently available from a CO_2 laser pulse but at a repetition rate of 211.55-MHz of the compact SXLS storage ring. Under this condition, and the further assumption that the theoretical electron beam emittance can be attained at the current recently demonstrated in the SXLS Phase I 200-MeV compact (8.5-m circumference) storage ring, the X-Ray intensity at the iodine K-edge energy of 33.169-keV can be as large as $1.09 \cdot 10^{14} / \text{sec}$ with a 2.2-keV energy spread in a cone of half angle $\theta_c = 1$ -mrad. This intensity is considerably larger than that originally attained at SSRL [2] with the SPEAR storage ring at 3-GeV or the parallel effort by Dix, et.al. [30] at the Deutsches Elektronen Synchrotron DESY ring at 3.5-GeV. The Compton backscattered intensity attainable is at least a factor of 10 greater than the present day values obtained at dedicated facilities at the NSLS X-Ray ring [3] and at the upgraded DESY wiggler beam [31]. A significant limitation may be the energy density of $\sim 50 \text{ J/cm}^2$ absorbed by the laser cavity mirror. We propose a study program directed at the

following topics.

1. Determine an optimum electron energy for the ring, which considers the need for beam damping between X-Ray pulses (higher energy), acceptable X-Ray energy spread in a beam of sufficiently large angle spread to give X-Ray spots of clinical interest (lower energy), and large angle spread in the laser beam to reduce the absorbed energy per unit area on the laser cavity mirror (lower E_e).
2. Perform a Monte-Carlo calculation to determine the extent of both transverse (x and y) and longitudinal (energy) emittance dilution as a function of duration of the X-Ray pulse.
3. Design a compact ring with a low beta insertion (smaller σ_{ex} and σ_{ey}) at the photon-electron interaction point, with zero dispersion in the undulator straight-section, and with smaller damping times (larger I_1) and emittance (also larger I_2 and tune ν) than the SXLS Phase I machine. A conceptual layout of such a ring is given in Fig. (8) which includes a large number of short, high field H-magnets and a large number of quadrupoles which together produce a high tune and large I_2 which result in smaller emittance [38].
4. Explore further the suggestion of Blum [10] for a high repetition rate Linac-driven IR-FEL.
5. Study the possibility of higher reflectivity of the optical cavity mirrors of the FEL.
6. Explore the design of a pulsed (laminated pole) short period undulator to attain maximal single pass gain while at the same time provide a large (4-mrad) half cone angle for the laser beam to minimize the absorbed energy per unit area of the laser cavity mirror.
7. Explore the possibility of grazing incidence reflection of X-Rays from an oscillating mirror to sweep a small cross-section X-Ray beam across the required coronary area of the patient.

IX. ACKNOWLEDGEMENTS

I am most grateful to Dr. E. B. Blum for the preliminary calculations he recently performed which demonstrated that the desired X-Ray intensity is indeed attainable by Compton backscattering. I also acknowledge with gratitude the many suggestions he made on problem areas which will have to be addressed. The present work also relies heavily on the lucid and rigorous presentation of the parameters and accelerator physics of the SXLS storage ring as calculated and documented by Dr. J. B. Murphy in the SXLS Conceptual Design report, as well as his considerable efforts with others, notably Dr. S. Kramer, in the commissioning of the machine and verification of the machine properties. I also thank Dr. Murphy for his careful reading of this work and constructive criticisms which have been incorporated into this report. I further wish to thank Dr. G. P. Williams for several informative discussions on reflectivity in the infrared. I also acknowledge my debt to Drs. A. Sandorfi, M. LeVine and C. Thorn for their excellent presentation of the physics of Compton backscattering in the LEGS proposal, a document which inspired my thinking about the possible application of that method to Angiography. Finally I acknowledge the crucial contribution of Dr. A. Luccio to that proposal with the idea of use of an FEL as the laser photon source.

X. REFERENCES

- [1]. A. M. Sandorfi, M. J. LeVine, C. E. Thorn, G. Giordano, and G. Matrone, The Fabrication of a Very High Energy Polarized Gamma Ray Beam Facility and A Program of Medium Energy Physics Research at the National Synchrotron Light Source. BNL Physics Dept. Proposal to the Dept. of Energy, (Sept. 1982). For a more complete study of the kinematics see

- L.N. Blumberg and S.I. Schlesinger, Kinematics of the Relativistic Two-Body Problem, Los Alamos unclassified report LAMS-1718 (1955).
- [2]. E. Rubinstein, J. C. Giacomini, H. J. Gordon, A. C. Thompson, G. Brown, R. Hofstadter, W. Thomlinson, and H. D. Zeman, Nucl. Instr. & Methods A291 (1990) 80 and references therein.
- [3]. W. Thomlinson, N. Gmur, D. Chapman, R. Garrett, N. Lazarz, H. Moulin, A. C. Thompson, H. D. Zeman, G. S. Brown, J. Morrison, P. Reiser, V. Padmanabahn, L. Ong, S. Green, J. Giacomini, H. Gordon, and E. Rubinstein, Rev. Sci. Instr. 63(1)1992, p. 625 and references therein.
- [4]. W. Thomlinson, private communication, 1992.
- [5]. L.N. Blumberg, Conceptual Design Study of an Intense X-Ray Source for Coronary Angiography, Stanford Univ. HEPL report (unpublished), June 1984; reissued as BNL informal Report BNL-47356 (April, 1992).
- [6]. H. Wiedemann, Pulsed X-Ray Source for Non-Invasive Digital Subtraction Angiography. Proc. U.S. National Accel. Conf., Vancouver, B.C., Canada, 1985. IEEE Trans. Nucl. Sci. Vol NS-32 (1985) 3308.
- [7]. A. Luccio, Status and Perspectives of the FEL Experiment at Brookhaven, Physics of Quantum Electronics Workshop, Sun Valley, ID, June, 1981; Free Electron Generators of Coherent Radiation, Physics of Quantum Electronics, Vol. 8, S.F. Jacobs and G. T. Moore, eds., Addison-Wesley Press, Reading, PA.
- [8]. J. B. Murphy, R. Biscardi, J. Bittner, L. N. Blumberg, E. Bozoki, E. Desmond, H. Halama, R. Heese, H. Hsieh, J. Keane, S. Kramer, R. Nawrocky, T. Romano, J. Rothman, J. Schuchman, M. Thomas, J. M. Wang, J. Krishnaswamy, W. Louie, and R. Rose, 1991 IEEE Particle Accel. Conf., San Francisco, CA, May 1991 IEEE Conference Record 91CH3038-7, Vol. 2 (1991) 1107.
- [9]. I. Ben-Zvi, The BNL Accelerator Test Facility and Experimental Program, 1991 IEEE Particle Accel. Conf., San Francisco, CA, May, 1991 IEEE Conference Record 91CH3038-7, Vol. 1 (1991) 550.
- [10]. E. B. Blum, Can an Inverse Compton Scattering Angiography Source be Built?, Informal BNL NSLS note, April 1, 1992 (unpublished).
- [11]. D. Chapman, private communication, 1992.
- [12]. J. B. Murphy and G. Vignola, Design Book, Superconducting X-Ray Lithography Source, Vol. I. Design of the Iron Magnet Storage Ring, BNL Informal Report, Jan. 1989 (unpublished).
- [13]. Superconducting Super Collider Conceptual Design, SSC-SR-2020, (March, 1986). SSC Central Design Group, LBL, Berkeley, CA. J. D. Jackson editor.
- [14]. M. Sands, The Physics of Electron Storage Rings, SLAC Report 121 (Nov. 1970).
- [15]. R. Helm, M. J. Lee, P. L. Morton and M. Sands, IEEE Trans. Nucl. Sci. NS-20 (1970) 900.
- [16]. L. N. Blumberg, J. B. Murphy and S. Sharma, Proc. 1991 IEEE Particle Accel. Conf., San Francisco, CA, May 1991, IEEE Conference Record 91CH3038-7, Vol. 4 (1991) 2661.
- [17]. J. B. Murphy, contained in presentation to DARPA/BNL Review Committee, March 26, 1992 (unpublished).
- [18]. C. E. Thorn, G. Giordano, D. C. Kistner, G. Matone, A. M. Sardonfi, C. Schaerf and C. S. Whisnant, The LEGS Electron Spectrometer for Tagging Backscattered Photons. Nucl. Instr. & Methods A285 (1989) 447.
- [19]. O. Klein and Y. Nishina, Z. Physik 52, 853 (1929).
- [20]. H. A. Bethe and J. Ashkin, Passage of Radiations Through Matter, Part II of Experimental Nuclear Physics, Vol. 1, Ed. E. Segre, J. Wiley & Sons, Inc., New York (1953) 322.
- [21]. J. J. Livingood, Principles of Cyclic Particle Accelerators, D. van Nostrand Co., Inc., Princeton, NJ (1961), Chapter 6, p. 92.
- [22]. E. D. Courant and H. S. Snyder, Ann. Phys., Vol. 3 (1958), Eq. 5.2
- [23]. E. B. Blum, private communication, April, 1992.
- [24]. CRC Handbook of Chemistry and Physics, 55th Edition, 1974-1975, CRC Press, Cleveland, OH. Ed. R. C. Weast, Section E, General Physical

Constants, E-84.

- [25]. G.P. Williams, private communication (April, 1992)
- [26]. S. Krinsky, An Undulator for the 700-MeV VUV Ring of the National Synchrotron Light Source, BNL report BNL-26199(1979). Some other relevant reports by Krinsky are: S. Krinsky, Wigglers for the National Synchrotron Light Source, BNL-24548(1978); S. Krinsky, M.L. Perlman and R.E. Watson, Characteristics of Synchrotron Radiation and of Its Sources, BNL-27678(1980); A. Luccio and S. Krinsky, Studies on the Undulator for the Free Electron Laser Experiment at the National Synchrotron Light Source, BNL-30078(1981); S. Krinsky, W. Thomlinson and A. van Steenbergen, An Overview of Undulators and Wigglers at the NSLS, BNL-31989(1982); S. Krinsky, Some Comments on the Design of Electron Storage Rings for Free Electron Lasers, BNL-33813(1983).
- [27]. G. Cottingham and E. C. Raka, private communications (April 1992).
- [28]. L. Blumberg, J. Bittner, J. Galayda, R. Heese, S. Krinsky, J. Schuchman, and A. van Steenbergen, National Synchrotron Light Source VUV Storage Ring Lattice, U.S. National Particle Accelerator Conference, San Francisco, CA, May 1979, Proc. IEEE Trans. Nucl. Sci. NS-26, No. 3 (1979) 3842. Also BNL Informal report BNL-25777(1979).
- [29]. S. Krinsky, L. Blumberg, J. Bittner, J. Galayda, R. Heese, J. C. Schuchman and A. van Steenbergen, Design Status of the 2.5-GeV National Synchrotron Light Source X-Ray Ring, U.S. National Particle Accel. Conf., San Francisco, CA, May, 1979, Proc. IEEE Trans. Nucl. Sci. Vol. NS-26, No. 3 (1979) 3806. Also BNL Informal report BNL-25774(1979).
- [30]. W. R. Dix, C. C. Gluer, W. Graeff, K. H. Mohne and W. Kupper, Non-Invasive Angiography Using Synchrotron Radiation, Deutsches Elektronen-Synchrotron report DESY SR-82-24, Hamburg, (Dec. 1982).
- [31]. W. R. Dix, W. Graeff, J. Heuer, M. Lohmann, B. Reime, R. Reumann, K. Engelke, C. Hamm, W. Kupper, and B. Kaempf, Coronary Angiography Using Synchrotron Radiation--Studies in Human Subjects with the System NIKOS II, Deutsches Elektronen-Synchrotron report DESY SR-91-01 (Sept. 1991).
- [32]. L. N. Blumberg, J. G. Cottingham, J. W. Glenn, J. J. Grisoli, M. Month and A. van Steenbergen, Fast Extraction of the AGS Beam, 1971 U.S. Particle Accel. Conf., Chicago, March 1-3, 1971, Proc. IEEE Trans. Nucl. Sci. Vol. NS-18, No. 3 (June 1971) 1009. Also see L. Blumberg, et al., BNL Accel. Dept. Internal report AGS-DIV 71-2(1971) and L. Blumberg, et al., Fast External Beam Operations Manual, BNL Accel. Dept. Internal report, (unpublished), 1974.
- [33]. A. Soukas, private communication (1973).
- [34]. J. Jackson, private communication (1983).
- [35]. A good discussion of beam envelope methods is contained in H. Ploss and L. Blumberg, Methods of Emittance Measurement in External Beams Using Ellipse Approximations, BNL Accel. Dept. Internal report AGS-DIV 68-4(1968) and L. N. Blumberg, M. G. Barton, J. D. Fox, J. W. Glenn and L. E. Repeta, Emittance Measurements in the AGS Slow External Beam, BNL Accel. Dept. Internal report AGS-DIV 69-12(1969) and Proc. VII-th Int. Conf. on High Energy Accelerators, Yerevan, USSR (Sept. 1969).
- [36]. J. B. Murphy, Synchrotron Light Source Data Book, BNL Informal report BNL-42333(May, 1990).
- [37]. J. B. Murphy, private communication (April, 1992).
- [38]. L. N. Blumberg and R. Z. Liu, Low Emittance Lattice for SPEAR, SSRL Internal Report SSRL-TN-83/01(1983). Also L. N. Blumberg, J. Harris, R. Stege, J. Cerino, R. Hettel, A. Hofmann, R. Z. Liu, H. Wiedemann and H. Winick, Proc. 1985 Particle Accel. Conf., Vancouver, BC, Canada, May 13-16, 1985, IEEE Trans. Nucl. Sci., Vol. NS-32, 3433(1985). See also J. Safraneck and H. Wiedemann, Low-Emittance in SPEAR, 1991 IEEE Particle Accelerator Conf., San Francisco, CA, May 1991, IEEE Conference Record 91CH3038-7, Vol. 2(1991) 1104.

LASER WAVELENGTH VS ELECTRON ENERGY FOR COMPTON BACKSCATTERING

X-RAY ENERGY VS COMPTON BACKSCATTERING ANGLE THEIA FOR EX(0)=33.169--

Wed Apr 22 13:53 1992

COMPTON SCATTERED X-RAY SPECTRUM, THETAC=1-mrad, Ee=136.1-MeV

Fig. 5

6keV energy range centered at 33.169keV (I K edge)

RF BUCKETS FOR SXLS PHASE I AT $E_e = 136.1$ AND 200-MeV

Fig. 8

CONCEPTUAL PLAN VIEW OF
ELECTRON STORAGE RING
WITH FEL UNDULATORS AND
OPTICAL CAVITY
(NOT TO SCALE)

PROGRAM COMPSCAT

```

C COMPUTES COMPTON BACKSCATTERING CROSS SECTION AND PHOTON FLUX INTO
C SOLID ANGLE DEFINED BY CONE OF HALF ANGLE THETAC. USES FUNCTION COMPUTE
C SANDORFI, LEVINE AND THORN, BNL PROPOSAL TO DOE (1982). NUMERICAL
C INTEGRATION BY FUNCTION SIMP WHICH CALLS FUNCTION SFUN FOR INTEGRAND.
C ENERGY SPECTRUM FROM FUNCTION DSDE. WRITES TO FILE IF NOPT.NE.0
C INPUT DATA:
C EL=ENERGY OF INCIDENT LASER PHOTONS (eV)
C GAMMA=ENERGY OF INCIDENT ELECTRON (IN UNITS OF MeC*C)
C THETAC=MAXIMUM LAB ANGLE OF PHOTON SCATTERING (MRAD)
C N=NUMBER OF INCREMENTS OF DELTA(THETAC)
C IE=ELECTRON BEAM CURRENT (AMPS)
C PL=LASER POWER (WATTS)
C LL=LENGTH OF LASER PULSE (CENTIMETERS)
C SIGL=RMS RADIUS OF LASER BEAM (CM)
C SIGXE=RMS WIDTH OF ELECTRON BEAM (CM)
C SIGYE=RMS HEIGHT OF ELECTRON BEAM (CM)
C
C COMPUTED QUANTITIES:
C THETAP=CM ANGLE OF SCATTERED PHOTON (RAD)
C THETA=LAB ANGLE OF SCATTERED PHOTON (MRAD)
C DSDOMP=CM DIFFERENTIAL CROSS SECTION (MB/SR)
C DSDOM=LAB DIFFERENTIAL CROSS SECTION (MB/SR)
C J=JACOBIAN OF CM TO LAB TRANSFORMATION
C EG=PHOTON LAB ENERGY (KEV)
C SIGMA=BACKSCATTERING CROSS-SECTION INTO THETAC CONE (MB)
C DNGDT=PHOTONS/SEC INTO CONE OF HALF ANGLE THETAC
C
      IMPLICIT REAL*8 (A-H,O-Z)
      REAL*8 IE,LL
      CHARACTER*10 DAY,HOUR
      CHARACTER*12 OUTFILE
      CHARACTER*80 HEADER
      COMMON EL,GAMMA
      EXTERNAL SFUN
C
      TWOPI=6.283185307179586D0
1    CALL DATE(DAY)
      CALL TIME(HOUR)
      WRITE(*,*) 'ENTER HEADER (A80)= '
      READ(*,*) HEADER
      WRITE(*,*) 'Please enter EL,GAMMA,THETAC,N,NOPT (EL=0 to exit) >'
      READ(*,*) EL,GAMMA,THETAC,N,NOPT
      IF(EL.EQ.0.0D0) CALL EXIT
      WRITE(*,20) DAY,HOUR,HEADER,EL,GAMMA,THETAC,N
20   FORMAT(1X,'PROGRAM COMPSCAT-',2A10.1x,A80//20X,'INPUT DATA' /
1     10X,'EL(eV)=',F12.9,' GAMMA=',F10.6,' THETAC(mrad)=',F6.2,
2     ' N =',I3//23X,'RESULTS'/' THETA(MR) THETAP(RAD)',
3     ' DSDOMP(MB/SR) DSDOM(MB/SR) EG(KEV) DSDEX(MB/KEV)')
C
      IF(NOPT.NE.0) THEN
        WRITE(*,*) 'ENTER NAME OF OUTPUT FILE(A12)= '
        READ(*,*) OUTFILE
        OPEN(1,FILE=OUTFILE,STATUS='NEW',FORM='FORMATTED')
      ENDIF
      DELTA=THETAC/DFLOTJ(N)
      THETA=0.0D0
30   THETAR=.001D0*THETA
      CALL COMPUTE(THETAR,THETAP,DSDOMP,DSDOM)
      EGG=EG(THETAR)

```

```

      DSIGDEX=DSDE(THETA)
      WRITE(*,40) THETA, THETAP, DSDOMP, DSDOM, EGG, DSIGDEX
40  FORMAT(F10.3,F12.5,E14.7,E13.6,F11.3,F14.4)
      IF (NOPT.NE.0) WRITE(1,40) THETA, THETAP, DSDOMP, DSDOM, EGG, DSIGDEX
      THETA=THETA+DELTA
      IF(THETA.LE.THETAC+DELTA/2.0D0) GOTO 30
      THETACR=.001D0*THETAC
      SIGMA=TWOPI*SIMP(0.0D0, THETACR, N, SFUN)

C
      WRITE(*,*) 'Now, enter IE, PL, LL, SIGL, SIGXE, SIGYE (all real))'
      READ(*,*) IE, PL, LL, SIGL, SIGXE, SIGYE
      WRITE(*,60) IE, PL, LL, SIGL, SIGXE, SIGYE
60  FORMAT(//' IE(amp), PL(watt), LL(cm), SIGL(cm), SIGXE(cm), SIGYE(cm)=',
1   F5.3, F14.2, F8.3, 3F8.5)
      A=TWOPI*DSQRT(SIGL*SIGL+SIGXE*SIGXE)*DSQRT(SIGL*SIGL+SIGYE*SIGYE)
      DNGDT=(2.604D0*IE*PL*SIGMA*LL)/(EL*A)
      WRITE(*,70) SIGMA, DNGDT
70  FORMAT(20X,'SIGMA =',F7.2,' MB'/
1   20X,'DNGDT =',E11.5,' PHOTONS/SEC')
      IF (NOPT.NE.0) CLOSE(1)
      CALL EXIT
      END

C *****
      FUNCTION EG(THETA)
      IMPLICIT REAL*8 (A-H, O-Z)
      COMMON EL, GAMMA
      EG=0.004D0*GAMMA*GAMMA*EL/
1   (1.0D0+(4.0D0*GAMMA*EL/5.110034D+5)+(GAMMA*THETA)**2)
      RETURN
      END

C *****
      SUBROUTINE COMPUTE(THETA, THETAP, DSDOMP, DSDOM)
      IMPLICIT REAL*8 (A-H, O-Z)
      COMMON EL, GAMMA
      BETA=DSQRT(GAMMA*GAMMA-1.0D0)/GAMMA
      TEMP1=DCOS(THETA)-BETA
      TEMP2=DSIN(THETA)**2
      COSTP=GAMMA*TEMP1/DSQRT(TEMP2+GAMMA*GAMMA*TEMP1**2)
      THETAP=DACOS(COSTP)
      EGPEGP=1.0D0/(1.0D0+(2.0D0*GAMMA*EL/5.110034D+5)*(1.0D0+COSTP))
      DSDOMP=39.703873D0*EGPEGP*EGPEGP
1   *(EGPEGP+(1.0D0/EGPEGP)-DSIN(THETAP)**2)
      FJ=GAMMA*(TEMP2+DCOS(THETA)*TEMP1)
1   /(TEMP2+(GAMMA*GAMMA)*TEMP1**2)**1.5
      DSDOM=FJ*DSDOMP
      RETURN
      END

C *****
      FUNCTION SIMP(A0, AMAX, N1, SFUN)
      IMPLICIT REAL*8 (A-H, O-Z)
      REAL*8 NEWSIMP
      N=N1
      ITER=0
      IF(MOD(N,2).NE.0) N=N+1
      OLDSIMP=0.0D0
1  DELTA=(AMAX-A0)/DFLOTJ(N)
      SUM=0.0D0
      X=A0
      DO 10 I=1,N+1
      IF(MOD(I,2).EQ.0) CONST=4.0D0

```

```

      IF (MOD(I,2).EQ.1) CONST=2.0D0
      IF ((I.EQ.1).OR.(I.EQ.N+1)) CONST=1.0D0
      SUM=SUM+CONST*SFUN(X)
C     WRITE(*,*) ' SIMP', AO, AMAX, X, CONST, SUM
10    X=X+DELTA
      NEWSIMP=SUM*DELTA/3.0D0
C     WRITE(*,*) ' SIMP ', OLDSIMP, NEWSIMP
      ITER=ITER+1
      IF (DABS(OLDSIMP-NEWSIMP)/NEWSIMP).LE.1.D-4) GOTO 20
      IF (ITER.GT.20) STOP ' NO CONVERGENCE AFTER 20 ITERATIONS'
      N=2*N
      OLDSIMP=NEWSIMP
      GOTO 1
20    SIMP=NEWSIMP
      N1=N
      RETURN
      END
C *****
      FUNCTION SFUN(THETA)
      IMPLICIT REAL*8 (A-H, O-Z)
      CALL COMPUTE(THETA, THETAP, DSCOMP, DSDOM)
      SFUN=DSDOM*DSIN(THETA)
      RETURN
      END
C *****
      FUNCTION DSDE(THETA)
C Computes cross section dsigma/dEx (mb/keV) per unit energy
C interval of Compton backscattered X-Ray of initial energy EL(eV)
C at angle theta(radians) relative to electron direction. Ee=electron
C LAB energy (eV)=GAMMA*mc**2. From Thorn, et. al., NIM A285(1989)447.
      IMPLICIT REAL*8 (A-H, O-Z)
      COMMON EL, GAMMA
      MCSQR=5.1099906D5
      TWOPIROSQ=4.989346153D2
      X=GAMMA*THETA
      Y=1.0D0+4.0D0*GAMMA*EL/(MCSQR*(1.0D0+X**2))
      TERM1=(2.0D0*GAMMA/(1.0D0+X**2))**2
      TERM2=Y+1.0D0/Y-(2.0D0*X/(1.0D0+X**2))**2
      DSDE=TWOPIROSQ*EL*1.0D-3*TERM1*TERM2/(Y*EG(THETA))**2
      RETURN
      END

```

```

PROGRAM SIGCOMPT
C Evaluates total Compton cross-section (cm**2) for photons of energy
C EGAMMA (MeV) from Klein-Nishina formula (see Halliday, p.171 or Bethe
C and Ashkin, Ed. Segre, Exptl. Nucl. Phys., p322. (1952). Writes file if NOUT=1
  IMPLICIT REAL*16 (A-H, O-Z)
  CHARACTER*10 DAY, HOUR
  CHARACTER*12 OUTFILE
  CHARACTER*80 HEADER
  DIMENSION EGAMMA(500), SCOMPT(500)
  EMASS=5.10999060-01
  WRITE(*,*) 'enter HEADER(80)= '
  READ(*,*) HEADER
  WRITE(*,*) 'enter N, NOUT, EGAMIN, DEGAM= '
  READ(*,*) N, NOUT, EGAMIN, DEGAM
  EGAMMA(1)=EGAMIN
  DO 10 J=1, N
    R=EGAMMA(J)/EMASS
    SCOMPT(J)=2.494670-25/R*((1.000-2.000*(R+1.000)/R**2)*QLOG(2.000
1 *R+1.000)+5.00-01+4.000/R-5.00-01/(2.000*R+1.000)**2)
10 EGAMMA(J+1)=EGAMMA(J)+DEGAM
C      print results
    CALL DATE(DAY)
    CALL TIME(HOUR)
    WRITE(*,20) DAY, HOUR, HEADER
20 FORMAT(1X,2A10,1X,A80//'      EGAMMA(MEV)  SIGCOMPT(CM2)'//)
    WRITE(*,30) (EGAMMA(J), SCOMPT(J), J=1, N)
30 FORMAT(2(1PE15.6))
    IF(NOUT.EQ.1) THEN
      WRITE(*,*) 'enter output file name(A12)= '
      READ(*,*) OUTFILE
      OPEN(1, FILE=OUTFILE, STATUS='NEW', FORM='FORMATTED')
      WRITE(1,30) (EGAMMA(J), SCOMPT(J), J=1, N)
      CLOSE(1)
    ENDIF
  CALL EXIT
END

```


PROGRAM BUCKET

```

C COMPUTES CONTOUR OF RF BUCKET. SYNCHROTRON FREQUENCY.ETC AT N
C POINTS. WRITES TO FILE IF NF=1.
C VM=PEAK RF CAVITY VOLTAGE (IN KILOVOLTS)
C ES=KINETIC ENERGY OF SYNCHRONOUS PARTICLE (IN MEV)
C FR=ROTATIONAL FREQUENCY OF SYNCHRONOUS PARTICLE (MHZ)
C HN=HARMONIC NUMBER
C DE=ENERGY GAIN PER TURN FROM RF CAVITY (IN KEV)
C ALPHA=MOMENTUM COMPACTION FACTOR
      IMPLICIT REAL*8 (A-H,O-Z)
      DIMENSION PHI(500), E(500)
      CHARACTER*10 DAY,HOUR
      CHARACTER*12 OUTFILE
      CHARACTER*80 HEADER
10  WRITE(*,*) 'ENTER HEADER(80)='
      READ(*,*) HEADER
      CALL DATE(DAY)
      CALL TIME(HOUR)
      WRITE(*,20) DAY,HOUR,HEADER
20  FORMAT(1X,2A10,1X,'PROGRAM BUCKET ',A80)
      WRITE(*,*) 'ENTER N,NF,VM(KV),ES(MeV),FR(MHz),HN,DE(keV),ALPHA='
      READ(*,*) N,NF,VM,ES,FR,HN,DE,ALPHA
      WRITE(*,25) VM,ES,FR,HN,DE,ALPHA
25  FORMAT(1X,'VM(KV)=' ,F5.1,' ES(MeV)=' ,F6.2,' FR(MHz)=' ,F8.5,
1    ' HN=' ,F3.0,' DE(keV)=' ,F7.4,' ALPHA=' ,1PE11.5)
      IF(VM.EQ.0.0D0) GOTO 99
      GAMMAS=1.0D0+ES/.51099906D0
      BETASQ=(GAMMAS**2-1.0D0)/GAMMAS**2
      GAMMA=(ALPHA-1.0D0/GAMMAS**2)/BETASQ
      PI=3.141592653589793D0
      SPHIS=DE/VM
      PHIS=PI-DASIN(SPHIS)
      CPHIS=DCOS(PHIS)
      FS=FR*DSQRT(-VM*HN*GAMMA*CPHIS/(2.0D3*PI*ES))
      PHIL=PI-PHIS
      PHIUI=PHIS+1.0D0
30  F=DCOS(PHIUI)+(PHIUI+PHIS-PI)*SPHIS+CPHIS
      PHIU=PHIUI+F/(DSIN(PHIUI)-SPHIS)
      IF(DABS(PHIU-PHIUI).LE.1.0D-5) GOTO 40
      PHIUI=PHIU
      GOTO 30
40  C=1.414213562D0*FS*ES/(HN*FR*GAMMA)
      EMAX=C*DSQRT((2.0D0*CPHIS+(2.0D0*PHIS-PI)*SPHIS)/CPHIS)
      WRITE(*,50) GAMMA,FS,PHIS,PHIL,PHIU,EMAX
50  FORMAT(1H,' GAMMA=' ,F7.5,' FS=' ,F7.5,' MHZ PHIS=' ,
1    F6.4,' RAD PHIL=' ,F5.4,' RAD' /1H,' PHIU=' ,F7.5,
2    ' RAD EMAX=' ,F6.3,' MEV')
      WRITE(*,60)
60  FORMAT(1X,4(' PHI(R) E(MEV) '))
      DPHI=(PHIU-PHIL)/DFLOTJ(N+1)
      PHI(1)=PHIL+DPHI
      N4=N/4
      DO 70 J=1,N
      E(J)=C*DSQRT(DABS(DCOS(PHI(J))+CPHIS+(PHI(J)+PHIS-PI)*SPHIS))
70  PHI(J+1)=PHI(J)+DPHI
      DO 90 L=1,N4
      I=L+N4
      J=L+2*N4
      K=L+3*N4
      WRITE(*,80) PHI(L),E(L),PHI(I),E(I),PHI(J),E(J),PHI(K),E(K)

```

```
80 FORMAT(1X,4(F6.4,F7.3,2X))
90 CONTINUE
  IF(NF.EQ.1) THEN
    WRITE(*,*) 'ENTER NAME OF OUTPUT FILE OUTFILE(A12)= '
    READ(*,*) OUTFILE
    OPEN(1,FILE=OUTFILE,STATUS='NEW',FORM='FORMATTED')
    WRITE(1,95) (PHI(J),E(J), J=1,N)
95  FORMAT(F7.5,1X,F9.5)
    DO 96 J=1,N
96  E(J)=-E(J)
    WRITE(1,95) (PHI(J),E(J), J=1,N)
    CLOSE(1)
  ENDIF
  GOTO 10
99 CALL EXIT
  END
```

END

**DATE
FILMED**

7 / 7 / 92