**ACID MEASUREMENTS VIA
NEAR-INFRARED SPECTROSCOPY (U)**

by

B. R. Buchanan *[Signature]* 5/15/91
Westinghouse Savannah River Company
Savannah River Laboratory
Aiken, SC 29808

OSTI
MAY 15 1992

A report for submission to OSTI
and release to general public.

DISCLAIMER

This report was prepared as an account of work sponsored by an agency of the United States Government. Neither the United States Government nor any agency thereof, nor any of their employees, makes any warranty, express or implied, or assumes any legal liability or responsibility for the accuracy, completeness, or usefulness of any information, apparatus, product, or process disclosed, or represents that its use would not infringe privately owned rights. Reference herein to any specific commercial product, process, or service by trade name, trademark, manufacturer, or otherwise does not necessarily constitute or imply its endorsement, recommendation, or favoring by the United States Government or any agency thereof. The views and opinions of authors expressed herein do not necessarily state or reflect those of the United States Government or any agency thereof.

This article was prepared in connection with work done under Contract No. DE-AC09-89SR18035 with the U.S. Department of Energy. By acceptance of this article, the publisher and/or recipient acknowledges the U. S. Government's right to retain a nonexclusive, royalty-free license in and to any copyright covering this article, along with the right to reproduce and to authorize others to reproduce all or part of the copyrighted article.

MASTER
[Signature]

DISTRIBUTION OF THIS DOCUMENT IS UNLIMITED

2/13/91

WSRC-TR-91-82

Page 1 of 21

ACID MEASUREMENTS VIA NEAR-INFRARED SPECTROSCOPY (U)

To: W. J. Blair, 703-F
M. T. Feller, 704-19F

From: B. R. Buchanan, 773A *BRB*

CC: C. W. Jenkins, 773A
P. E. O'Rourke, 773A
TIS Library (4)

Title Page	i
Table of Contents	ii
Distribution	1
Introduction	2
Experimental	2
Theory	3
Results	5
Sampling Systems	6
Future Work	7
Conclusion	7
Acknowledgements	7
Table 1	8
Figure 1: Block Diagram of NIR Instrument	9
Figure 2: Prediction of Verification set using Model 1	10
Figure 3: Prediction of Verification set using Model 2	11
Figure 5: Eigen Value Plot for Model 1	13
Figure 6: Eigen Value Plot for Model 2	14
Figure 7: Temperature Run using Models 1 and 2	15
Figure 8: Schematic of Dip Probe	16
Figure 9: Schematic of SwageLock [™] Tee cell	17
Figure 10: Schematic of Plexiglass sample cell	18
Figure 11: Temperature run using model 3	19
Figure 12: Prediction of Verification set using Model 3	20
References	21

then placed in an aluminum block with a heating element and heated to 75 °C. The block contained a lens assembly so data could be collected during the heating process.

Data was analyzed using a program, SRLMVA, written in Quick BASIC¹ at the Savannah River Laboratory. The program was compiled to run under IBM OS/2 version 1.1 using Microsoft's Professional BASIC version 7.0. The water scans were subtracted from the remaining samples and the result multiplied by -1. The absorbance value at 1071 nm was subtracted from each spectrum to compensate for baseline offsets.

Out of 33 samples that contained only nitric acid, a subset, consisting of 21 samples, was used to build a PCR² model. This model was designated Model 1. Model 1 was then used to predict the remaining samples. The verification set consisted of the remaining 12 samples.

A second model was developed using spectra from all the types of samples. This was done to compensate for the $\text{Al}(\text{NO}_3)_3$ interference. In Model 2, the calibration set consisted of 49 samples out of 72 total. The verification set consisted of the remaining 23 samples.

A third model was developed using the samples selected for model 2 and two of the three temperature run samples. It was hoped that this would help compensate for temperature differences.

Theory:

The Near-Infrared (NIR) region is by definition between 0.765 and 2.65 μm . This region consists mainly of the overtones and combinations of the fundamental vibrations, although there are some electronic transitions at the shorter wavelengths. Because the transitions are overtones of fundamental vibrations, it is not unreasonable to expect situations that affect the fundamental bands to also affect the NIR spectrum. Because acids in solution affect the hydrogen bonding of the system, which in turn affects the vibrational spectrum, it was reasonable to assume that we could correlate the effect with concentration.

¹Beginners All-purpose Symbolic Instruction Code.

²Principal Component Regression. A method of determining the eigenvectors of a system which can be used to reduce a spectrum to a concentration value. See references 9 and 10 for more details.

Several researchers have shown that species that affect hydrogen bonding cause changes in the NIR spectrum^{1,2,3,4,5,6,7}. Both cations and anions appear to affect the spectra of water. The differences are related to the ionic entropies of the system. Because entropy is temperature dependent, the NIR spectra also show a temperature dependence. It was shown that the hydrogen bonding would decrease as the temperature increased. Because we were correlating the concentration of acid to the NIR spectrum, which is temperature dependant, temperature will affect the calculated concentration unless it is modelled out.

There are three global methods, i.e., those that use all of the sensor responses, that are widely used in spectroscopy. One of these, Classical Least Squares (CLS), requires that the pure component spectra be available⁸. When only mixture spectra are available, Principle Component Regression models (PCR) or Partial Least Squares models (PLS) are often used. Both PCR and PLS models were developed to determine which model to use.

Principal Component Regression (PCR) is a method that finds the eigenvectors of a set of data^{9,10}. The mechanics of finding the eigenvectors will not be discussed here as they can be found in the cited references. Suffice it to say that each spectrum is mapped onto the eigenvectors and a least squares fit of the resulting eigenvalues to concentration is done. The advantages of using a full spectrum model are that the model is more immune to perturbations to the system and additional information is available. Partial Least Squares (PLS) is similar to PCR except that the information in the concentration vector is used to optimize the selection of the model vectors.

To develop a PCR or PLS model the usual methodology consists of using a set of standards, a calibration set, to build the model. Then the model is tested on a second set of standards, the verification set, to determine the optimal number of vectors. The verification set standards are not included in the development of the model, but are within the calibration space³. To test the robustness of the model, samples suspected to be outside the calibration space would need to be tested, but this test is usually separately performed.

We used several statistics to determine the optimal number

³Calibration space can be thought of as the region confined by the responses of the standards. For example, if the standards contained 0-5M HNO₃, a sample of 4M HNO₃ would be within the calibration space. A 10 M HNO₃ could fall outside the space and a sample containing Al(NO₃)₃ would be outside the calibration space.

of vectors to use and to test the predictions. The Standard Error of Prediction (SEP) is a measure of how far off the predicted samples were from the actual samples. At the optimal number of vectors it should be a minimum value. In addition to the SEP, a bias value can be used. The bias is an indication of a systematic offset of the predicted concentrations. A final statistic that we use is the F of Residuals. This statistic is a comparison of the average sum of squares of the calibration samples residuals to the sum of squares of the unknown sample's residuals. Standard tables of F values can be used to test whether the unknown's residuals are similar to the calibration set's residuals.

Results:

Model type 1 consisted of samples which contained only nitric acid. The optimal Standard Error of Prediction (SEP) for the PCR model was 0.0412 mols/liter at 2 vectors. Figure 2 shows the prediction of acid concentration of the verification set using model 1. The optimal SEP for the PLS model was 0.0427 mols/liter again at 2 vectors. The PCR model did slightly better, but was not significantly different.

Model type 2, the calibration set consisted of a subset of randomly selected samples out of the entire data set. The optimal SEP for the PCR model was 0.28588 mols/liter, Figure 3, while the SEP for the PLS model was 0.28833 mols/liter. Both models used 4 vectors. Prediction shows obvious bias.

The PCR model seemed to perform slightly better than the PLS model. The PCR model was variance scaled, weighting those samples that had a low signal-to-noise ratio (SNR) less than the higher SNR samples. The PLS model weights the samples with the concentration of acids regardless of the SNR of the samples spectrum. The SNR of samples with similar concentrations might have varied as a result of how the data was collected. The slight increase in performance of the PCR model could be because of the weighting to variance. However, the increased performance of the PCR model was not great enough to recommend it over PLS. A PCR model will be used in this study simply because it did perform slightly better.

When Model 1 was used to predict those samples that contained aluminum, a bias of the prediction developed. From Figure 4, it can be seen that the bias tends to follow the aluminum concentration. The effect of aluminum on the model is an apparent increase in predicted acid concentration. This result was not unexpected as aluminum is a small, highly charged cation, which could effect the hydrogen bonding of the water in a similar manner to the acid. Figure 5 shows that the Al samples

are not in the same space as the calibration set. Model 1 had trouble predicting acid concentration for those samples which contained Al. Fortunately, the effect of Al is not exactly the same as that of acid, as can be seen in Model 2.

Model 1 was also used in the prediction of H_2SO_4 samples. The actual acid concentration was the sum of the nitric and sulfuric acids added to the standard. If Model 1 is modelling acid concentration, the source of the acid concentration will be irrelevant. The "actual" versus the predicted concentration is presented in Table 1. It can be seen that Model 1 was able to predict the acid concentration of the samples that contained sulfuric acid.

Figure 6 shows the plot of eigenvalues with all samples mixed together. The eigenvalues for the samples containing nitric acid, sulfuric acid, and the verification and calibration sets all seem to occupy the same space on the plot of the eigenvalues. The samples containing aluminum nitrate seem to occupy their own space on the plot of the eigenvalues for model 2. The relatively good predictive ability of model 2 would support this.

As was expected, temperature has an adverse effect in predicting acid concentration. Because the prediction depends on the hydrogen bonding affects on the spectrum of the acid, the reduction of hydrogen bonding at the higher temperatures causes the predicted value to be greater than actual, Figure 7. This may be a result of the temperature affecting the bulk water more than the individual ion/water interactions of the solvent cages. Also, the increasing temperature could be affecting the equilibrium between the free acid and it's conjugate base. The effect of temperature on the nitric acid only model is much less than the all samples model. This is probably a combination of two effects. First, because the all samples model contains other species that might have varied with temperature, a shift in temperature would cause a much more pronounced effect. Second, and probably the greatest effect, is the calibration set for model 1 contained spectra that had been collected over 4 days, while model 2 contained spectra collected over 2 days. Therefore, model 1 contained much more temperature information than model 2 and appears be more immune to the temperature increase.

Sampling Systems:

There are several sampling configurations available. A dip probe that can be inserted in the tank has been developed at SRL. Figure 8 shows a schematic of this probe. Another configuration that has been tested at SRL is the use of a SwageLokTM Tee, diagramed in Figure 9. This cell can be placed in line or on a

slip stream. A sampling block can also be set up for discrete samples that have been pulled from the process. The block is diagramed in Figure 10. Any combination of these sampling configurations can be used because the instrument is multiplexed.

Future Work:

Plutonium has a characteristic spectrum in the Near-Infrared region. Because this will tend to interfere with the acid measurement, the effect of plutonium will have to be modelled out. The effect of other species that might also be present in the system would have to be investigated. It is known that uranyl nitrate has no effect on the acid measurement.

As was shown earlier, temperature affects the measurement of acid. The effect is not great until a temperature of about 35 °C is reached, Figure 11. This effect can be greatly reduced, but the introduction of the temperature data causes the prediction to worsen, Figure 12. The effect of temperature will have to be further investigated.

Conclusion:

We were able to predict acid concentration to ± 0.08 M HNO_3 . In the presence of Al^{+3} added as interferant, the prediction dropped to ± 0.29 mols/liter over a range of 0 to 9M HNO_3 . Temperature affects the prediction of acid adversely and would have to be modelled out or the sample cell thermostated prior to using this method.

Acknowledgments:

The author would like to acknowledge S. Boynton for allowing us to use her lab space and M. A. Sanders for allowing us to use his office space. Much of this work was done by Yasmin K. Ali under a Ruth Patrick Science Education Center program.

The information contained in this article was developed during the course of work done under Contract No. DE-AC09-89-SR18035 with the U. S. Department of Energy.

Table 1

Prediction of Sulfuric acid Samples using Model 1

Actual Mols/liter	Predicted Mols/liter
0.202452	0.246108 ⁴
0.202452	0.207180
0.767383	0.838004
0.309175	0.761956 ⁵
7.22274	7.011891
3.66439	4.345082
3.71124	3.635772
7.12001	6.955925
3.50501	3.473714
0.305576	0.967184

⁴The next three samples are sulfuric acid only.

⁵The rest of the samples are sulfuric and nitric acid.

Figure 1: Block Diagram of NIR Instrument

Figure 2: Prediction of Verification set using Model 1

Figure 3: Prediction of Verification set using Model 2

Figure 4: Prediction of $\text{Al}(\text{NO}_3)_3$ samples using Model 1

Prediction with Al Nitrate
Model 1: Only Nitric Acid

Figure 5: Eigenvalue Plot for Model 1

Figure 6: Eigenvalue Plot for Model 2

Figure 7: Temperature Run using Models 1 and 2

Prediction Temperature Run
7.09 M Acid - Nitric and Sulfuric

Figure 8: Schematic of Dip Probe

Figure 9: Schematic of SwageLock™ Tee cell

Figure 10: Schematic of Plexiglass sample cell

Actual dimensions are 3" l X 3" w X 2" h.

Figure 11: Temperature run using model 3

Figure 12: Prediction of Verification set using Model 3

References:

1. G. R. Choppin and N. J. Horning, *Spectrochimica Acta* 30A, 1615 (1974).
2. A. Grant, A. M. C. Davies, and T. Bilverstone, *Analyst* 114, 819 (1989).
3. G. R. Choppin and M. R. Violante, *J. Chem. Phys.* 56, 5899 (1972).
4. G. R. Choppin and J. R. Downey, Jr., *J. Chem. Phys.*, 56, 5899 (1972).
5. J. L. Hollenberg and J. B. Ifft, *J. Phys. Chem.* 86, 1938 (1982).
6. O. D. Bonner and G. B. Woolsey, *J. Phys. Chem.* 72, 899 (1968).
7. J. Paquette and C. Jolicoeur, *J. of Sol. Chem.* 6, 403 (1977).
8. D. M. Haaland and E. V. Thomas, *Anal. Chem.* 60, 1193 (1988).
9. T. Hirschfeld, J. B. Callis, and B. R. Kowalski, *Science* 226, 312 (1984).
10. I. A. Cowe and J. W. McNicol, *Appl. Spectrosc.* 39, 257 (1985).

END

DATE
FILMED

10 / 20 / 92