Collaborative Research on Fluidization
Employing Computer-Aided Particle Tracking

Quarterly Progress Report No. 5
For Period Oct. 1, 1989-Dec. 31, 1989

Contract DE-FC21-88MC25048

by

Michael M. Chen
Department of Mechanical and Industrial Engineering
University of Illinois at Urbana-Champaign
1206 West Green Street
Urbana, IL 61801

Submitted to

U.S. Department of Energy
Morgantown Energy Technology Center
Collins Ferry Road
Morgantown, WV 26505

RECEIVED

MAY 21 1993

OSTI

MASTER

Introduction

This report covers the fifth quarter of the subject contract. The primary objective of this quarter was debugging and testing of the hardware and software of the transportable computer-aided particle tracking apparatus to be used at the Morgantown Energy Technology Center. First year effort was devoted to system interconnection and software development. While the hardware interconnections were finished in the fourth quarter, the software development and debugging was completed in this quarter with successful initial testing. Two tracer particles were also made from 1/8th inch nylon particles, and they will be sent to METC soon.

Discussions

I. Software Development

The software used in the particle tracking facility is composed of four programs: RAWCAL.EXE which takes the raw data from a calibration, PD.EXE which creates the calibration tables from this raw data, MAIN.EXE which runs the actual data acquisition, and PROCESS.EXE which converts data from the MAIN.EXE program into tracer position data. All these programs have been successfully tested. Other data reduction programs had been written in Fortran for the DEC PDP-11 and can be easily adapted for the new system.

DISCLAIMER

I.I Program RAWCAL.EXE

This report was prepared as an account of work sponsored by an agency of the United States Government. Neither the United States Government nor any agency thereof, nor any of their employees, makes any warranty, express or implied, or assumes any legal liability or responsibility for the accuracy, completeness, or usefulness of any information, apparatus, product, or process disclosed, or represents that its use would not infringe privately owned rights. Reference herein to any specific commercial product, process, or service by trade name, trademark, manufacturer, or otherwise does not necessarily constitute or imply its endorsement, recommendation, or favoring by the United States Government or any agency thereof. The views and opinions of authors expressed herein do not necessarily state or reflect those of the United States Government or any agency thereof.

Program RAWCAL.EXE is used to take the raw calibration data for the detectors which relate photon count to distance between the detector and the tracer. Basically, the user inputs a position in three coordinates which specifies the fixed position of the tracer particle. The computer then takes count data over several counting periods. The counts for each detector are averaged and the distance from the particle to the specific detector is calculated. The distance and its corresponding count are stored for each detector. This is done a number of times. Enough data must be taken here to fit a polynomial curve from which the final look-up tables will be constructed.

Specifically program RAWCAL.EXE is a C-language shell with two C-language subroutines and an Assembly language interrupt handler. The main program takes care of initialization of the hardware as well as buffers and arrays into which data is read. The program prompts the user for the position of the tracer particle in three Cartesian coordinates (x,y,z). When these are given and verified, the interrupt generator is enabled by setting the computer's internal clock with a user specified counting period which is set at the beginning of the program. At each interrupt, the program jumps into the interrupt handler. Here internal functions and parameters are set and the actual data taking procedures are called. The LeCroy 4434 scaler is the module accessed. A command register within the 4434 is programmed with the commands to be executed and the channels to be read. Once all sixteen channels have been read, the interrupt routine returns control to the main program. This interrupt is preformed 32 times for each tracer location and the results are

averaged to give statistical mean counts. This averaging is necessary because photon counting is purely statistical. The average count is then stored with its respective detector-to-source distance, for later work. When a significant amount of data has been taken, the user may exit the program.

I.II Program PD.EXE

Program PD.EXE is used to compile the look-up tables for the position-calculation program (PROCESS.EXE). Many count versus distance data have been stored for each detector during the calibration process. From these a characteristic polynomial curve fit can be found for each detector, and it is from this polynomial that the final table is generated. The degree of the polynomial is chosen by the user and through a graphics routine the degree for the best curve fit may be deduced. In the table a distance is stored which corresponds to an integer count. This data is stored for each detector. The tracer position calculation program can then obtain a distance value for a detector from the location of the matching count in the table.

I.III Program MAIN.EXE

Program MAIN.EXE takes count data during an experiment and stores them in the computer. The basic process of this data acquisition program (MAIN.EXE) is the same as that of the calibration program (RAWCAL.EXE). It is also written in C and uses the same subroutines and interrupt handler as those in program RAWCAL.EXE. Here, initialization is done and the interrupt is enabled. The

counts are read into the computer and the interrupt routine ends. Two buffers of 1 kilobyte each are used to temporarily store the counts. When one buffer is full, its data are written on the hard disk of the computer while the other buffer is being filled. This process takes several hours in a typical experiment.

I.IV Program PROCESS.EXE

Program PROCESS.EXE converts the detector count data into tracer position data. From each count a distance is found from the look-up tables generated by program PD.EXE. From these distances and the detector locations the tracer particle position can be triangulated. Using sixteen detectors provides redundancy which is helpful in error reduction. Weighting factors are used in the calculation procedure so that those detectors which are closest to the tracer have the greatest emphasis in determining its location. This is done because the count versus distance data is more accurate and detailed in the short distance ranges.

II. Fabrication of Tracer Particles

Two tracer particles have been made matching the size and weight of the 1/8th inch nylon particles to be used at METC. The method involves drilling a 1 mm through hole in a nylon particle and placing a small quantity of scandium metal inside. Two red nylon particles were selected, and the scandium metals imbedded in them have approximately spherical shape of diameter 0.93 mm and of masses 1.340 mg and 1.260 mg. The scandium metals were first coated with polyurethane. After they are inserted in the nylon particles, the

holes are sealed with small nylon pieces glued by polyurethane. The completed tracer particles have same size but slightly heavier (by about 4%) than the normal nylon particles. These tracer particles will be shipped to METC soon.

END

**DATE
FILMED**

6 / 21 / 93