WSRC-RP-92-1215

**AT-LINE BENZENE MONITOR FOR MEASURING BENZENE
IN PRECIPITATE HYDROLYSIS AQUEOUS (U)**

by W. J. Jenkins

Westinghouse Savannah River Company
Savannah River Site
Aiken, South Carolina 29808

WSRC-RP--92-1215

DE93 008936

Other Authors:

This paper was prepared in connection with work done under Contract No. DE-AC09-89SR18035 with the U. S. Department of Energy. By acceptance of this paper, the publisher and/or recipient acknowledges the U. S. Government's right to retain a nonexclusive, royalty-free license in and to any copyright covering this paper, along with the right to reproduce and to authorize others to reproduce all or part of the copyrighted paper.

MASTER

DISTRIBUTION OF THIS DOCUMENT IS UNLIMITED

DISCLAIMER

This report was prepared as an account of work sponsored by an agency of the United States Government. Neither the United States Government nor any agency thereof, nor any of their employees, makes any warranty, express or implied, or assumes any legal liability or responsibility for the accuracy, completeness, or usefulness of any information, apparatus, product, or process disclosed, or represents that its use would not infringe privately owned rights. Reference herein to any specific commercial product, process, or service by trade name, trademark, manufacturer, or otherwise does not necessarily constitute or imply its endorsement, recommendation, or favoring by the United States Government or any agency thereof. The views and opinions of authors expressed herein do not necessarily state or reflect those of the United States Government or any agency thereof.

This report has been reproduced directly from the best available copy.

Available to DOE and DOE contractors from the Office of Scientific and Technical Information, P. O. Box 62, Oak Ridge, TN 37831; prices available from (615) 576-8401.

Available to the public from the National Technical Information Service, U. S. Department of Commerce, 5285 Port Royal Rd., Springfield, VA 22161.

**WESTINGHOUSE SAVANNAH RIVER COMPANY
SAVANNAH RIVER TECHNOLOGY CENTER**

WSRC-RP-92-1215

Keywords: DWPF, DWPT, PHA, PR,
Hydragard™, Benzene,
Liquid Sampler, PRBT,
Analytical, Process Control

Retention: Permanent

October 14, 1992

TO: DISTRIBUTION

FROM: W. John Jenkins

**AT-LINE BENZENE MONITOR FOR MEASURING
BENZENE IN PRECIPITATE HYDROLYSIS AQUEOUS (U)**

ATTACHMENTS: Benzene Monitor System Report (U), WSRC-TR-92-478.
Statistical Analysis of the At-line Benzene Monitor Study (U),
SCS-ASG-92-066.

SUMMARY

A highly accurate and repeatable at-line benzene monitor (ALBM) has been developed to measure the benzene concentration in precipitate hydrolysis aqueous (PHA) in the DWPF. This analyzer was conceived and jointly developed within SRTC by the Analytical Development and the Defense Waste Process Technology Sections with extensive support from the Applied Statistics Group and the TNX Operations Section. It is recommended that

- an ALBM specifically adapted to DWPF analytical requirements be used to measure benzene in PHA,
- calibrations be performed using a 10% methanol solution matrix (for standard stability), and
- based on experience gained in development at TNX, the services of ADS and ASG be employed to both adapt the ALBM to DWPF requirements and develop statistical control procedures.

BACKGROUND

To prevent possible formation of a flammable mixture in the Precipitate Reactor Bottoms Tank (PRBT), it is a DWPF Process Requirement (PR) that the benzene in the PHA be less than 30 ppm.^[1] However, there has been much concern about obtaining an accurate and repeatable measurement because of the extreme volatility of benzene and the physical

design of the Hydragard™ liquid sampler and sample vial.^[2] In order to assure the needed timing of an acceptable measurement technique, development was divided into two parts: 1) a "work-around sampler" for cold chemical runs (CCR),^[3] and 2) a permanent solution compatible with radioactive operation. After completion of the first portion, two analyzers were selected for further development; an adaptation of an on-line optical analyzer developed for use during the "work-around sampler" testing and a modified Hydragard™ using a solvent extraction technique.^[4] Because of the complexity and difficulty of the solvent extraction technique, the on-line analyzer (simplified to at-line use) was selected for final development and testing.

DISCUSSION

Analyzer - Summary of Attachment 1

The recommended analyzer employs absorption spectrophotometry adapted to remote measurement using fiber optics. Only the sample interface (optical cell) is contained within the shielded cell and is connected by electrical and fiber optic cables to the remaining components located in a clean area. The liquid phase benzene concentration is measured by recirculating air through the sample until the concentration in the vapor has equilibrated. The vapor benzene concentration is then measured by an ultraviolet-visible spectrophotometer.

During development and testing, both on the bench and on the full scale SRAT (Sludge Receipt and Adjustment Tank) at TNX, the analyzer has demonstrated the following advantages:

- ease of and methods to calibrate,
- operation using peanut vials filled from the existing Hydragard™ sampler,
- precision (~0.5% at 1 σ) better than the purge and trap method,
- direct measurement without pretreatment,
- analytical time of ~ 5 minutes (required once per 43 hour precipitate hydrolysis cycle), and
- production of no waste stream other than the sample itself, which is also required for other analyses.

The general steps in performing an analysis are:

- 1) The sample is collected using the existing Precipitate Reactor Hydragard™ sampler and placed in the ALBM sample holder.
- 2) The sample ID is input to the ALBM computer and the measurement sequence initiated. A reference measurement using air is automatically performed.
- 3) Using a lever action, the operator raises the vial into position when prompted - engaging the sample vial with the ALBM sample interface and beginning the sample analysis.

- 4) The sample is purged by the recirculating air and the absorption spectrum is measured after a set period of time. The spectrum is used to calculate the liquid phase concentration.

Because of the basic design, the liquid sample volume to analyzer gas volume ratio is very important. Operational experience at TNX confirmed that sample vial fill-volume is a very important parameter to be controlled. (That is, the analysis is incorrect if the vial is not sufficiently filled.) Visual and/or solution mass checks will be necessary to ensure the vial is sufficiently filled prior to analysis using the ALBM.

Statistics - Summary of Attachment 2

Bench scale and full scale testing and data analysis proceeded in six phases:

- Phase 1. Initial calibration using water solution matrix
Precision was good; but observed systematic bias was attributed to system modifications and changes in sample volumes after water calibration.
- Phase 2. First experimental design
A statistically designed experiment (fractional factorial) was prepared to investigate solution matrix (water, PHA), sample temperature (80°F, 140°F), optical cell temperature (85°C, 95°C), and benzene concentration (10 ppm, 50 ppm).
- Phase 3. Analysis of data from first test
Sample temperature and the solution matrix x standard concentration interaction were significant at the 5% level. The variability for repeated ALBM measurements was not the same across all samples. However, the largest standard deviation within repeated measurements was only 0.85 ppm.
- Phase 4. Second experimental design and data analysis
This experiment investigated the ALBM when calibrated with water solution standards and used to measure standards prepared with other solution matrices (PHA, 10% methanol). There was no evidence of statistically significant differences among the variances of the repeat measurements across all the standards. The largest standard deviation measured was 0.47 ppm. Calibration curves indicated a measurement bias on the PHA and 10% methanol standards when the ALBM was calibrated with the water solution matrix; the measured biases were 1.30 and 1.32, respectively. Stability measurements on the PHA and 10% methanol matrix standards indicated no statistically significant change in concentration over a two hour period. The 10% methanol matrix standard was used for Phase 6.
- Phase 5. Third experimental design and data analysis
Benzene spiking tests were performed using the full scale SRAT at TNX. Tank samples taken over time (the benzene concentration decreases with time as benzene evaporates from the tank) and standards prepared in the 10% methanol matrix were measured with the ALBM calibrated with water solution standards. The largest standard deviation for repeatability across all samples was only 0.57 ppm and there was no evidence of an inconsistency in the repeatability of the data.

The importance of sample vial fill-volume was clearly demonstrated when some exceptionally low benzene values were measured in partially filled vials.

Phase 6. Calibration using 10% methanol standards

The ALBM was calibrated using the 10% methanol matrix standards and the data on PHA standards collected in Phase 4 were reanalyzed. No inconsistencies in the repeatability of the data were indicated using the five measurements available for each standard. The largest standard deviation was only 0.28 ppm. Results from this analysis indicate the ALBM value for a sample with a true benzene concentration of 30 ppm would be greater than 28.5 ppm at a 99% confidence level.

REFERENCES

- [1] R. A. Jacobs and R. E. Eibling, WSRC-IM-91-2 PROCESS REQUIREMENTS PRECIPITATE HYDROLYSIS, DWPFPR-S03 Rev. 1 (U), January 27, 1992.
- [2] L. F. Landon, WSRC-RP-91-1064, BENZENE ANALYSIS OF AQUEOUS PRODUCT PRODUCED DURING PRECIPITATE HYDROLYSIS (U), October 23, 1992.
- [3] W. J. Jenkins, WSRC-RP-92-852, TECHNICAL BASIS FOR "WORK-AROUND" PR SAMPLER (U), June 29, 1992.
- [4] R. E. Eibling, WSRC-RP-92-566, DWPF BENZENE ANALYSIS: COLD CHEMICAL AND RADIOACTIVE OPERATION (U), April 24, 1992.

**WESTINGHOUSE SAVANNAH RIVER COMPANY
SAVANNAH RIVER TECHNOLOGY CENTER**

WSRC-TR-92-478

Keywords: Benzene, Analyzer,
DWPF, Fiber Optics
Spectrophotometry

Retention: Permanent

October 12, 1992

TO: W. J. Jenkins, 704-1T
FROM: 10/12/92
R.R. Livingston, 773-A
CC: Distribution

BENZENE MONITOR SYSTEM REPORT (U)

SUMMARY

Two systems for monitoring benzene in aqueous streams have been designed and assembled by the Savannah River Technology Center, Analytical Development Section(ADS). These systems were used at TNX to support sampling studies of the full-scale "SRAT/SME/PR" and to provide real-time measurements of benzene in Precipitate Hydrolysis Aqueous(PHA) simulant. This report describes the two ADS Benzene Monitor System (BMS) configurations, provides data on system operation, and reviews the results of scoping tests conducted at TNX. These scoping tests will allow comparison with other benzene measurement options being considered for use in the Defense Waste Processing Facility (DWPF) laboratory. A report detailing the preferred BMS configuration statistical performance during recent tests has been issued under separate title: Statistical Analyses of the At-line Benzene Monitor Study, SCS-ASG-92-066.

The current BMS design, called the At-line Benzene Monitor (ALBM), allows remote measurement of benzene in PHA solutions. We have demonstrated the ability to calibrate and operate this system using peanut vials from a standard Hydragard™ sampler. The equipment and materials used to construct the ALBM are similar to those already used in other applications by the DWPF lab. The precision of this system ($\pm 0.5\%$ Relative Standard Deviation(RSD) at 1 sigma) is better than the purge & trap - gas chromatography reference method currently in use. Both BMSs provide a direct measurement of the benzene that can be purged from a solution with no sample pretreatment. Each analysis requires about five minutes per sample, and the system operation requires no special skills or training. The analyzer's computer software can be tailored to provide desired outputs. Use of this system produces no waste stream other than the samples themselves (i.e. no organic extractants).

Derivative Classifier:

INTRODUCTION

In January 1992, the DWPF need for a benzene analysis method was first discussed with ADS. This measurement is needed to provide the benzene content of PHA before transfer from the Precipitate Reactor(PR) to the Precipitate Reactor Bottoms Tank(PRBT). The benzene concentration needs to be less than 30 parts per million(ppm) before transfer in order to guarantee that benzene vapor in the PRBT off gas system does not exceed the benzene Lower Explosive Limit(Reference 2). The sample from PR will contain approximately one weight percent suspended solids. This sample(PHA) will be at about 50 degrees centigrade (°C), have an acidic component of about 0.1 molar(M) formic acid, and a salt content of around 0.5 M(Reference 3). The PHA solids are black in color and prevent direct measurement of benzene in solution by absorption spectrophotometry.

ADS experience with absorption spectrophotometry has demonstrated its usefulness for remote process measurements(Reference 4). This technique is well adapted to remote measurement by use of fiber optics. To minimize the amount of space required in the shielded cells and the number of contaminated system components, only the sample interface is contained in the shielded cells. Electrical control lines and fiber optics provide the connection to system components located in the clean area.

In addition to the size requirements imposed by available equipment space inside the DWPF shielded cells, plumbing changes in the current DWPF sampling system are strictly limited. The first BMS delivered to TNX in mid March was an online analyzer designed to operate unattended in a continuous mode. This analyzer would make measurements based upon operator-selected sampling frequency and control the instrument's process interface. This first analyzer (Figure 1) worked satisfactorily to support initial mixing & sampling tests on the PR (Reference 5). This system was capable of meeting the size restrictions given by DWPF but would require plumbing changes. Several key problems were identified in this first system: 1) solenoid valve failures resulted from sample plugging and corrosion, 2) sample vessel filling and emptying were dependent on process flows and pressures that were difficult to control, and 3) quality control(QC) checks would require an additional valve and fluid pump.

The most recent design is called an At-Line Benzene Monitor(ALBM). This system: 1) requires no DWPF plumbing changes, 2) uses no valves, and 3) uses a standard DWPF sampling vial for each measurement (Figure 2). This new design allows for QC samples and recalibration when needed without any changes from the sample measurement configuration. However, process solutions must be manually sampled using existing laboratory sampling techniques. This sample is then transferred to the ALBM and measured from the sample container. User input is required to initiate analyzer measurement of each sample. Training requirements are minimal for the cell operators and lab analysts because the ALBM prompts the user at each step.

The ALBM described above was delivered to TNX in July 1992. This system was used for additional testing of the PR sampling system and documentation of benzene loss rate from the full scale SRAT/SME/PR. Similar results for benzene loss from PR were obtained from the ALBM system as from earlier use of the online analyzer(Reference 1). The online analyzer measured one sample every four minutes while the ALBM system required about six minutes per sample. The recent statistical studies conducted by Defense Waste Processing Technical(DWPT), Applied Statistics Group (ASG) and ADS show that the ALBM system is capable of excellent precision and accuracy even under process conditions. In summary, the ASG report documents that the ALBM is capable of making bias free benzene measurements in the needed concentration range. The statistical study also provides an evaluation of instrument calibration, a comparison of synthetic and process solutions, and documentation of process studies. With the current ALBM, the 95% confidence interval for a 30 ppm standard is 28.9 to 32.0 ppm.

DISCUSSION

Operation of the online and ALBM systems are based upon the same principles:

1. Volatile organic compounds can be purged from aqueous solutions using air or other gases.
2. For fixed volumes of liquid and gas, the vapor concentration of a volatile organic has an equilibrium value.
3. The equilibrium concentration of volatile organic in the confined space above a solution is related to the initial concentration of volatile organic in solution.

The benzene concentration of PHA cannot be measured directly in the sample solution by absorption spectrophotometry because not enough light can get through the sample solution (too black). To get around this problem, the ALBM uses a pump to recirculate air through the liquid sample in closed loop. The air containing purged benzene vapor is continuously pumped through an optical cell and back into the sample container where it bubbles through the sample solution (Figure 3). This process is called a "Recycling Purge". The UV spectrum of benzene is very unique(Figure 4) and can be readily identified for quantitation using an ultraviolet-visible (UV-Vis) spectrophotometer. The concentration of benzene in a sample is directly related to the amount of UV light absorbed at certain wavelengths as light passes through the optical cell. Figure 5 explains the principles used in absorption spectrophotometry for calculating component concentrations from absorption data.

We can monitor the change in benzene concentration versus time from start of purge sequence(i.e. measurements are started when the recirculating air pump is switched on). The change in benzene concentration with time is called a purge profile (Figure 6). The purge profile is affected by sample temperature, gas flow rate, salt content of sample and analyzer gas volume. The purge profile can be used to detect leaks in tubing or other analyzer components, because under normal operation the benzene vapor concentration will reach a maximum and stay at this

level until the system is vented. In a leaking system, the benzene vapor concentration will reach a maximum then decline at a rate determined by the leak size.

The ALBM uses the peanut vial sample container as the purge vessel. In this system, the purge time is set at three times the equilibrium time indicated by purge profiles for a given sample type. This 3X factor allows for minor differences in sample temperature, ionic strength and purge gas flow rate. The vapor volume of the ALBM system is about ten times the sample volume. This causes most of the benzene to be extracted from a typical PHA sample (~90% at 30 ppm). The ALBM's attainable sensitivity and detection limit are functions of 1) analyte extraction efficiency/sample volume relationship and 2) the optical cell path length.

The optical cell is maintained at a set temperature ($\pm 0.1^{\circ}\text{C}$) above the maximum sample temperature to prevent condensation of sample on the optical cell windows. Condensation on optical components causes scattering of light similar to fogging-up a bathroom mirror. Increasing optical cell temperature above the maximum dew point has a negative effect of reducing the benzene absorption signal. Observation of the online system benzene measurements plotted against time, before temperature control was stabilized ($\pm 5^{\circ}\text{C}$ swings), revealed a sawtooth pattern that correlates with optical cell temperature. As optical cell temperature increased, the measured benzene concentration decreased thus demonstrating an inverse relationship between absorption signal and optical cell temperature. Benzene absorption, in the expected concentration range (up to 100 ppm in solution) is linear with concentration (Figure 7). Because samples in this concentration range behave according to Beer's Law, increasing path length will give lower detection limits and greater sensitivity. Air flow through the optical cell is stopped for absorption measurements except when collecting a purge profile.

The pump used to move the air/benzene mix through the ALBM is a miniature diaphragm pump. The diaphragm and seals are made of viton and the other wetted components are Kynar. This pump operates on 6 - 12VDC for flow control. Pump speed is set to minimize transfer of sample solutions from the purge vessel into the sample vial interface which also serves as a vapor-liquid separator. Connections between the pump, optical cell, and sample vial interface are made with approximately 3/32 inch inside diameter (ID) Viton tubing (#14). This size allows excellent friction fit with 1/8 inch outside diameter (OD) stainless steel tubing. The optical cell (Figure 8) and sample vial interface (Figure 9) are manufactured from stainless steel and have Swagelok tube connections for analyzer plumbing. The purge tube is fitted with a 2 μM stainless steel frit supplied by Mott Metalurgical Supply Company to produce small air bubbles. This process reduces the time required for benzene to reach equilibrium in the vapor space.

The connections to the optical cell are 1/8 inch tubing for air/benzene flow and 1/2 inch Swagelok stainless fittings for fiber optic lens assemblies. The fiber optic lens assembly (Figure 10) provides a seal for the optical cell and allows light from the fiber optic to pass through the air/benzene mix. Light from a UV-Vis light source (located outside the cell) is focused onto the end of a fiber optic. This fiber optic is

run into the shielded cell and attached to the first lens assembly. Light from this fiber is collimated by the lens, travels through the optical cell and is refocused by a second lens onto another fiber optic. The second fiber optic is attached to a Diode Array Spectrophotometer(DAS) on the outside of the shielded cell. The DAS(Hewlett Packard 8452), IBM PC compatible computer, pump power supply & temperature control (supplied by SRTC Equipment Engineering Section), and light source are housed in a single cabinet outside the shielded cell. For UV measurement applications, the current spectrophotometer is limited to about 20 meters of fiber optic cable in each direction(i.e. a total of 40 M).

Cables from the optical cell temperature control and pump power are run into the shielded cell along with the fiber optics. The optical cell is heated with two 150 watt cartridge heaters that operate on 110 volt(v) alternating current(AC). The optical cell temperature is measured by a Resistance Temperature Detector(RTD) and controlled with a proportional temperature controller (Figure 11). Temperature is maintained independently without input from DAS or computer. The pump is switched on/off by using pin #25 of the DAS serial part. This pin of the DAS is software controlled, and allows +5v direct current(DC) supplied on another pin of the serial interface to complete a circuit to ground when closed. Completing this circuit closes a relay to supply current at selected voltage to the air pump (Figure 11).

An IBM Industrial Computer with HPIB (IEEE-488) interface controls the DAS. Any IBM PC compatible computer may be used although a 16MHz or faster machine with math coprocessor is recommended. Sample absorbance data and instrument control signals are passed back and forth via the HPIB interface. Sample measurements are made based upon software set-up parameters. These parameters include wavelength range, integration time, and number of scans for the spectrophotometer. The spectrophotometer uses this data to control a shutter on the UV-Vis light source.

The ALBM measures the absorption signal of benzene between a wavelength range of approximately 240 and 270 nanometers(nM). The light source is a deuterium lamp that has high UV output between 190 and 350 nM(Figure 12). This lamp is held in a housing that allows alignment of the lamp filament so that light can be aligned and focused with appropriate spot size on a fiber optic. The spectrophotometer shutter is mounted between the focusing optics and the fiber optic adapter. The lamp is provided with a power supply that provides very stable light intensity. This type lamp typically has a 1500 hour lifetime and a one year shelf life. Alignment and focusing of the light source is aided by software supplied with the ALBM.

The software used on this analyzer is a revision of code used for several years at SRTC. This program provides many options for system operation, but has been adapted for specific user interface needed by the ALBM. In general, this code (1) reads instrument setup parameters from disk, (2) requests necessary user input like sample identification, number of samples and timing information for start of analysis then (3) controls sample measurement sequence and (4) calculates benzene

concentration from absorption spectrum. Benzene concentration is calculated based upon a Principle Component Regression(PCR) model calculated during instrument calibration. The PCR model is essentially a multidimensional calibration curve that compensates for the effects of other sample constituents on benzene concentration.

Calibration of ALBM will not be needed on a frequent basis. Because absorption spectrophotometry uses a ratio technique, corrections are continuously made for changes in lamp intensity, plus fiber and other optical changes. A new calibration may be required to compensate for changes in system configuration like replacing spectrophotometer or optical cell. Other changes like lamp, fibers, pump, shutter or compuer should not require a new calibration. The process of building calibration models for prediction of unknown samples' benzene concentration includes data processing to enhance the benzene signal.

Because the benzene signal has relatively narrow peaks, a second derivative can be used to separate benzene signal from large slowly charging background signal (Figure 13). This improves the quality of benzene measurements especially in the presence of other organic vapors like formic acid. Since this type model uses many different wavelengths to compute benzene concentration, the computer is able to provide a warning signal when the measured spectra do not "look" like benzene. Compounds similar to benzene like toluene or phenol have very different absorption spectra(Figure 4) and would produce a large residual ratio. The residual signal is a goodness of fit statistic used to alert operators of possible interfering chemicals.

Samples are measured on the ALBM according to the following general steps:

- 1) Sample is collected using the normal Hydragard™ sampler then placed one at a time in the ALBM sample vial holder.
- 2) The ALBM computer is provided with the sample ID and told to initiate measurement sequence. The instrument begins by making a reference measusrment with air containing no benzene in the optical cell.
- 3) The operator is prompted to raise the sample holder using a lever action like found on the Hydragard™ sampler. This engages the sample vial with the ALBM sample vial interface, and the sample analysis is started.
- 4) The sample is purged by recirculating air through the sample. An absorption spectrum of benzene in air is measured after a set purge time, and this spectrum is used to calculate sample benzene concentration.

This same data collection sequence is followed for each sample. When standards are measured for instrument calibration, the absorption spectrum is saved on disk for later calculation of a calibration model. Software for creating calibration models also runs on the IBM PC or compatible computer. To maintain QC charts, a standard is substituted for a sample in the above sequence and analyzed in the same manner as a sample.

Normal operation of the ALBM should include analysis of QC standards. The most logical concentration of standard to assure excellent instrument performance is 30 ppm because this is the benzene level established for the safe transfer of PHA from PR to PRBT. Standards prepared in 10% methanol solution have been shown to have similar characteristics to standards prepared in PHA(Reference 1). The first step in preparation of these standards is dilution of pure benzene to approximately 1760 ppm in pure methanol(1 part in 50). This allows preparation of 30 ppm standard by addition of approximately 3.4 ml stock solution for 200 mL of standard. These standards are stable for at least two (2) hours when kept in peanut vials with viton solid septa and nearly zero head space. The 1760 ppm stock solution has been stored for up to 2 weeks in a freezer without noticeable change.

Operating experience and system evaluation conducted at TNX has provided excellent operational insight. The primary component of error that must be controlled by procedures is filling of the sample vial. Visual checks and/or solution mass checks should be used to make sure vials are filled within specified limits. Small variations in sample volume will not cause measurements to exceed control limits when peanut vials are completely filled. The fill volume is a parameter that must be matched with calibration because fill volume affects vapor/gas volume, total benzene available for purge and air extraction efficiency. Efforts to correct for these volume changes(i.e. half filled vials) have not been effective.

As currently used at TNX, the ALBM does not produce any waste other than the samples themselves. Therefore, this instrument has the added benefits of supporting waste minimization and the resulting cost savings over solvent extraction methods. The components selected for use in the shielded cells should be functional at high dose rates for reasonable lengths of time (up to 50 megarads for lenses). This system shows promise for being packaged with a very small footprint design that can be maintained for daily operation.

The next ALBM model (beta test unit) will incorporate features necessary for shield cell installation and maintenance. These fixtures include:

1. Individual components with weights < 12 lbs.
2. Quick disconnect fittings for power/control cables.
3. Sealed components for corrosion protection of motor windings, etc.
4. Longer pathlength optical cell with 1/2" diameter light path.
5. Wing nuts for access to in cell system components.
6. Customized software based upon DWPF operating requirements.
7. Use of a faster(386) PC compatible computer.

The changes in optical cell(#4 above) are desirable because: 1) The current system uses an air volume expander tube to improve air extraction efficiency. This additional volume is better used in the optical cell. 2) Increasing light path diameter will improve light throughput thus allowing a longer pathlength and faster measurements. 3) The increased path length will help improve analyzer sensitivity and detection limit. However, the pathlength will also have to be balanced with required length and quality of fiber optic cable to be used. To circumvent this

difficulty, a different spectrophotometer and detector could be substituted for the Hewlett Packard DAS. Use of this improved detector will allow extension of fiber optic beyond the present 20 meter limit.

Additional work should be done before completion of the ALBM Beta test model to optimize the system vapor volume. Some adjustment of this parameter is provided by the optical cell light path diameter and length; however, these dimensions must be balanced against the limits given for component size and weight that can be moved into the shielded cells. Optimum vapor volume can be measured using the current ALBM and then divided between the optical cell and sample interface during design of the beta test unit.

At present, we view use of this system like any other lab instrument for QA purposes. The software is documented by comments in source code and a manuscript user's manual (Reference 6), and the system reliability will be checked by repeat measurements of QC standards.

REFERENCES

1. Statistical Analyses of the At-line Benzene Monitor Study (U), SCS-ASG-92066, T.B. Edwards, September 1991.
2. SRAT and PRBT Vapor Flamability... WSRC-RP-90-0793, R.A. Jacobs, June 1991.
3. Vapor pressure of Benzene over Acid Hydrolysis Product WSRC-TR-91-165 D.D. Walker, April 1991.
4. Online Fiber-Optic Spectrophotometry DP-MS-88-186, D.R. VanHare et al, February 1989.
5. Technical Basis for "Work-Around" PR Sampler WSRC-RP-92-852, W.J. Jenkins, June 1992.
6. SRS Instrument Control and Data Reduction Programs Manual WSRC-TR-91-0103, B.R. Buchanan et al, July 1991.
7. Ultraviolet Spectra of Aromatic Compounds, Robert Friedel & Milton Orchin, John Wiley & Sons, 1951.
8. Laboratory Notebook, WSRC-NB-92-44, R.R. Livingston, starting March 1992

LIST OF FIGURES

<u>FIGURE #</u>	<u>TITLE</u>
1	Online Benzene Monitor System Components
2	Online Analyzer Sample Interface
3	Atline Benzene Monitor Sample Interface
4	UltraViolet Spectra of Atomic Compounds
5	Absorption Spectroscopy
6	Purge Profile for ALBM
7	Linearity of Purged Benzene
8	Optical Cell Drawing
9	Liquid-Vapor Separator/Sample Vial Interface
10	Fiber-Optic Lens Assembly
11	Temp. Pump Control Circuits
12	Deuterium Lamp Spectrum
13	Benzene Absorption Spectra - 2nd Derrivative Correction for Background Signal

FIGURE 1

ONLINE BENZENE MONITOR SYSTEM COMPONENTS

FIGURE 2

ONLINE ANALYZER SAMPLE INTERFACE

FIGURE 3

ATLINE BENZENE MONITOR SAMPLE INTERFACE

FIGURE 4
UltraViolet Spectra of Aromatic Compounds

FIGURE 5

Absorption Spectroscopy

$$A(\nu) = - \text{Log} (I(\nu) / I_0(\nu))$$

$$\text{Conc.} = A(\nu) * E(\nu) / \text{Pathlength}$$

Measure Intensity of light at many colors
without sample present -- $I_0(\nu)$
and with sample present -- $I(\nu)$

FIGURE 6

FIGURE 7

FIGURE 8
Optical Cell Drawing

Tue, Sep 29, 1992 7:43:36 AM

FIGURE 10
Fiber-Optic Lens Assembly

- (1). Mount 1/2" barrel of reducer in lathe - remove 1/4" fitting
- (2). Drill and tap reducer for 1/4-36 tap - install SMA adapter
- (3). Ream inside of reducer with #8 tapered ream (.3971-.5050)
- (4). Tap 1/2" reducer with 1/8" MPT tap
- (5). Smooth ID of reducer with 450 grit emory cloth
- (6). Srink TEFLON tubing on lens and trim
- (7). Press lens and tubing into reducer until flush
- (8). Heat assembly to 150 deg C - then press lens into reducer again
- (9). Repeat step #8 until lens no longer creeps during heating

FIGURE 11
Temperature & Pump Control Circuits

Temperature Control

Pump Control

FIGURE 12

FIGURE 13

INTER-OFFICE MEMORANDUM
Westinghouse Savannah River Company

SCS-ASG-92066
Task 92-022-0
Non-critical Data

September 16, 1992

To: **W. J. Jenkins**, 704-1T, Rm 214

From: **T. B. Edwards**, 773-42A, Rm 125 *BE*

cc: M. A. Baich, 704-T
J. T. Carter, 704-1T
R. E. Eibling, 704-T
C. W. Jenkins, 773-A
R. R. Livingston, 773-A
P. E. O'Rourke, 773-A
C. T. Randall, 704-T
ASG Files, 773-59A
M. R. Buckner, 773-A
C. E. Coffey, 773-A
E. W. Holtzscheiter, 773-A
L. F. Landon, 704-1T
J. C. Marek, 704-T
J. R. Pelfrey, 773-42A
C. P. Reeve, 773-42A

J. R. Pelfrey

J. R. Pelfrey, Manager

Date: 9-28-92

**STATISTICAL ANALYSES OF THE AT-LINE BENZENE MONITOR
STUDY (U)**

Summary

The experiments outlined in this report have provided an opportunity to gain valuable insight into the calibration and use of the At-line Benzene Monitor (ALBM). This report describes the statistical analyses completed in support of these experiments. Factors affecting the calibration process such as solution matrix and sample temperature have been investigated. Repeatability studies indicate a consistent set of ALBM values for each sample even with as few as five measurements. The process of preparing benzene standards was also

investigated. Standards prepared using a solution matrix of 10% methanol were found to be stable for least a two hour period, which is thought to be sufficient time for use during calibrations and calibration checks.

The ALBM, after being calibrated using 10% methanol standards, performed well in analyzing samples prepared using a Precipitate Hydrolysis Aqueous (PHA) solution matrix. A linear calibration curve was found to fit these data well. Such a calibration curve can be used to determine a lower prediction limit for ALBM measurements corresponding to a true benzene concentration of 30 parts per million (ppm). For the example of the 10% methanol calibration used in this study, a lower 99% limit for the ALBM value corresponding to a 30 ppm benzene in PHA standard is 28.5 ppm.

It is recommended that the Applied Statistics Group (ASG) work with the Analytical Development Section (ADS) to provide further statistical support for the ALBM. Statistical control charts or methodologies should be implemented as part of the procedures used for calibrations and calibration checks. Operational procedures can also benefit from the use of statistics. Control charts for repeatability of ALBM measurements of a given sample should be developed. A lower prediction limit on the ALBM measurement for a true benzene level of 30 ppm can be used to develop appropriate operational limits for the Precipitate Reactor (PR) based on the ALBM measurements.

Introduction

Over the course of the past few weeks, a great deal of work has been completed on the prototype At-line Benzene Monitor (ALBM) which is being developed by the Analytical Development Section (ADS) of SRTC for use at DWPF. Personnel from the Applied Statistics Group (ASG) and DWPT along with those from ADS have been working on the evaluation of this monitor and methods for its calibration and use.

This report traces through the statistical analyses that have been conducted to support this evaluation effort. These analyses were completed in several phases; each of which will be discussed below.

JMP for the Apple Macintosh personal computer and SAS-PC for the IBM personal computer were used during these analyses. Both of these statistical software packages are by SAS Institute Inc.

Statistical Analyses

Phase 1. Initial Calibration using Water Solution Matrix

A calibration of the monitor was completed by the Analytical Development Section (ADS) using a water solution matrix, samples at

82 degrees Fahrenheit ($^{\circ}$ F) and standards at four benzene concentrations (0, 5, 30, and 75 parts per million).

A set of seven standards were then prepared and analyzed by ADS. These standards were prepared using a water solution matrix; they were analyzed using the ALBM at low (80° F) and high (140° F) sample temperatures. Eleven ALBM measurements of the benzene in parts per million (ppm) were recorded for each standard. These data were provided to ASG for analysis and are given in Table 1.1 of Appendix A.

Repeatability of the Measurements

The repeatability of the measurements from the ALBM for each standard was investigated. The standard deviation of the repeated measurements for each standard at each sample temperature (which will be called a set) appears in Table 1.2 of Appendix A. Statistical tests that these variances are equal across all the sets, provided in Table 1.3 of Appendix A, indicate that there are significant differences among the sets. Measurements taken at the higher sample temperature are judged to have equal variances since the Prob>F values are greater than 0.05. Lower sample temperature seems to lead to more variability within the repeated measurements. However, the largest standard deviation within a set of repeat measurements is only 0.16 ppm.

Calibration Curve

Calibration curves, which were linear and quadratic in the standard concentration and which included sample temperature, were investigated. The average benzene values in ppm for the sets of repeated measurements were used in this part of the study. Table 1.4 in Appendix A provides the results from the quadratic fit. Each of the parameters in the model is seen to be statistically significant at the 5% level since their corresponding values in the Prob>|t| column are all less than 0.05. The root mean square error is 0.14 ppm. The root mean square error is a measure of the scatter about the calibration curve.

Predicted Benzene Measurements

Table 1.5 in Appendix A shows predicted benzene measurements and the associated prediction limits for standard concentration and sample temperature combinations using the quadratic calibration curve. These prediction intervals are at the 95% confidence level.

Thus, the interval (21.4, 22.2) is a 95% prediction interval for the ALBM value corresponding to a true benzene level of 30 ppm when the sample temperature is at 140° F. The bias seen here was attributed to system modifications and changes in sample volume after water calibration. This bias could have been corrected for within the ALBM and was, therefore, not considered a problem.

Phase 2. First Experimental Design

The next phase of the study involved an investigation into factors thought to have an effect on the calibration and measurement processes. These factors were identified as the solution matrix, sample temperature, optical cell temperature, and benzene concentrations of the standards. An experiment was needed to test for and estimate the effects of these factors as well as the following two-way interactions:

solution matrix x standard concentration,

solution matrix x sample temperature, and

standard concentration x sample temperature.

Two levels were selected for each of the factors and a replicated, fractional factorial design developed. The levels were:

	Low (-1)	High (1)
Solution Matrix:	Water	PHA
Sample Temperature:	80	140 degrees Fahrenheit
Optical Cell Temperature:	85	95 degrees Celsius
Benzene Concentration:	10	50 ppm

The experiment was randomized to support its being conducted over two days. The complete design is provided in Table 2.1 of Appendix A.

Phase 3. Analysis of Experimental Results

Standards were prepared and analyzed using the ALBM by ADS according to the fractional factorial design outlined in Phase 2. The ALBM calibration was the same as in the Phase 1 study. The resulting data, which are in Table 3.1 of Appendix A, were provided to ASG. The summary data by sample is given in Table 3.2 in Appendix A. A variable of interest for this study was the difference between the measured and the standard benzene concentrations (the last two

columns of Table 3.2 with standards -1 and 1 representing 10 and 50 ppm, respectively).

Significant Factors and Interactions

The data were analyzed using the JMP package and the main effects and interactions were estimated. The results are provided in Table 3.3 of Appendix A. Sample temperature and the solution matrix x standard concentration interaction are each significant at the 5% level since their respective values in the Prob>|t| column are less than 0.05.

Repeatability of Measurements

Table 3.4 of Appendix A shows that the variability of the repeated ALBM measurements is not the same across all samples as noted by the Prob>F values being less than 0.05. The largest standard deviation within the repeated measurements for a sample is only 0.85 ppm.

Phase 4. Second Experimental Design and Analysis

A second experiment was designed to investigate the behavior of the ALBM calibrated with a water solution matrix as in Phase 1 when used to analyze standards prepared with other solution matrices. This experiment involved standards prepared with both a 10% methanol solution matrix and with a precipitate hydrolysis aqueous (PHA) solution matrix. For the 10% methanol solution matrix, five standards were prepared at each of the following benzene concentrations (in ppm): 0, 8.5, 30, and 50. For the PHA solution matrix, 5 standards were prepared at 0 and 30 ppm and 4 standards were prepared at 50 ppm. This test was conducted by ADS and the resulting data are provided in Table 4.1 of Appendix A. The purpose of this study was to investigate the stability of the 10% methanol standards and to study the performance of water solution matrix calibrations in measuring standards prepared with these other solution matrices.

Repeatability of Measurements

Repeatability of the ALBM for each standard was investigated. Five repeat measurements were recorded for each standard. The variance for each set of five values was computed. Table 4.2 in Appendix A provides the results of a statistical test for equality of

these variances across all the standards. There is no evidence of statistically significant differences among the variances. The largest standard deviation within a set of repeated measurements is only 0.47 ppm.

Calibration Curve

Calibration curves were determined for the standards prepared with each of the two solution matrices using the summary data provided in Table 4.3 of Appendix A. Both linear and quadratic fits were performed. Tables 4.4 and 4.5 in Appendix A provide these results. These curves indicate that there is a bias in using the ALBM which has been calibrated for a water solution matrix to analyze standards prepared with 10% methanol or PHA solution matrices. This bias is seen in the estimated coefficient for the linear term for each fit being different from unity and is consistent across the two types of solution matrices. For the linear calibration curves, these estimated coefficients are 1.32 for 10% methanol and 1.30 for PHA.

Stability of Standards

Tables 4.6 and 4.7 in Appendix A provide the results of the analysis of the benzene standards over time for 10% methanol and PHA solution matrices, respectively. A linear regression of the ALBM measurements on time was determined for each standard benzene concentration for each of the two solution matrices. There is no significant linear relationship between these ALBM measurements and time at the 5% level for any benzene concentration. This is evident by the Prob>F values being greater than 0.05 for each of the regressions. Since there is no problem with stability over the two hour period covered in these data, the standards prepared with a 10% methanol solution matrix were thought to be suitable for use during calibrations. This type of calibration was conducted for Phase 6 discussed below.

Phase 5. Analysis of Tank Samples

An experiment was designed for spiking the Precipitate Reactor (PR) with benzene and then taking samples from the tank. Table 5.1 in Appendix A provides the data from this test. Samples taken from the PR along with standards prepared with a 10% methanol solution matrix were analyzed. Table 5.2 in Appendix A provides the summary data from this test. The repeatability of the measurements as well as the behavior of the measurements of the benzene in the tank over time are of interest. The ALBM was calibrated as in Phase 1.

During this study, the fact that sample volume is a key component to a quality ALBM measurement was reinforced. The benzene values for some samples were exceptionally low due to the vials being partially filled.

Repeatability of the Measurements

Table 5.3 in Appendix A shows the repeatability of the measurements by giving the standard deviation for measurements from each sample. There is no indication of an inconsistency in the repeatability of these data. The largest standard deviation for repeatability across all the samples is only 0.57 ppm.

Time Behavior of the Tank Samples

The natural logs of the benzene measurements were determined and regressed against the time each sample was taken from the PR in order to investigate loss of benzene from the tank. Table 5.4 in Appendix A provides these results. The estimated time coefficient for the natural log model is -0.119. The equation for the percent benzene loss is given by:

$$100 (1 - e^{-0.119 h}) \%$$

where h is in hours. Evaluated at h=1, the loss is estimated at 11.2%. From a previous study (references 1 and 2), the benzene loss in the PR had been estimated at about 11% per hour.

Phase 6. Calibration using 10% Methanol Standards

The ALBM was calibrated using the standards prepared with 10% methanol solution matrices. The PHA standards data collected in Phase 4 were then re-analyzed. These results are provided in Appendix A in Table 6.1. The purpose of this analysis was to investigate the performance of the ALBM and the corresponding calibration curve after the monitor was calibrated with the 10% methanol standards.

Repeatability of Measurements

Table 6.2 in Appendix A provides the analysis of the repeatability of these measurements. Five measurements are available for each

standard and there is no indication of an inconsistency in the repeatability of these data. The largest standard deviation is only 0.28 ppm.

Calibration Curve for PHA Standards

Tables 6.3 and 6.4 of Appendix A provide the results of the development calibration curves for these data. Both quadratic and linear fits were performed. The second degree coefficient for the quadratic fit is not statistically significant and thus, a linear fit is appropriate. Note that the bias that was evident in the calibration curve from the water solution matrix of Phase 4 has been eliminated. This can be seen in the estimate of the linear term (.98) which is near unity. Also, note that the root mean square error is only 0.67 ppm.

Predicted Benzene Measurements

Table 6.5 in Appendix A provides the predicted ALBM measurements along with their associated prediction limits at the 95% confidence level for standards of 0, 30 and 50 ppm. Thus, a 95% prediction interval for the ALBM measurement corresponding to a true value of 30 ppm is given by (28.9, 32.0). A 99% lower prediction interval for the ALBM value for a true benzene value of 30 ppm can also be determined using the standard error from Table 6.5 of 0.699 and a Student's t value of $t_{.01, 12} = 2.681$ as follows:

$$\text{Lower 99\% Limit} = 30.44 - (2.681 \times 0.699) = 28.5$$

Thus, the ALBM value for a sample with a true benzene level of 30 ppm would be greater than 28.5 ppm at a 99% confidence level.

Inverse Calibration Problem

The inverse calibration problem --- predicting the true benzene level for a given ALBM measured value --- was also investigated. Appendix B provides the details of this investigation and Table 6.6 of Appendix A lists the SAS-PC program which implements the approach.

The results are presented in Table 6.7 of Appendix A. An upper prediction limit for the true value, X, corresponding to a ALBM value, Y, at a 99% confidence level is provided for values of Y ranging from 4 to 40 ppm. For a ALBM value of 28 ppm an upper 99% prediction limit for the true benzene level would be given by 29.42 ppm.

This equation could be incorporated into the ALBM so that measurement results are given at the upper 99% confidence level.

Conclusions and Recommendations

Standards prepared with a solution matrix of 10% methanol were found to be stable over at least a two hour period, which is thought to be sufficient time for use during calibrations and calibration checks. The ALBM, after being calibrated using 10% methanol standards, performed well in analyzing samples prepared using a PHA solution matrix. A linear calibration curve was found to fit these data well. Such a calibration curve can be used to determine a lower prediction limit for the ALBM measurement corresponding to a true benzene level of 30 ppm.

It is recommended that the Applied Statistics Group (ASG) work with the Analytical Development Section (ADS) to provide further statistical support for the ALBM. Statistical control charts or methodologies should be implemented as part of the procedures used for calibrations and calibration checks. Operational procedures can also benefit from the use of statistics. Control charts for repeatability of ALBM measurements of a given sample should be developed. Lower prediction limits on the ALBM measurement for a true benzene level of 30 ppm can be used to develop appropriate operational limits for the Precipitate Reactor (PR) based on the ALBM measurements.

References

- [1] Jenkins, W. John, WSRC-RP-92-852, "TECHNICAL BASIS FOR 'WORK-AROUND' PR SAMPLER (U)", June 29, 1992.
- [2] Reeve, C.P., SCS-ASG-92043, "STATISTICAL ANALYSIS OF BENZENE LOSS DATA (U)", June 25, 1992.
- [3] Coleman, Hugh W. and W. Glenn Steele. **Experimentation and Uncertainty Analysis for Engineers**. John Wiley & Sons, New York, 1989.

Appendix A: Tables

Table 1.1:	Data from Initial Calibration Study
Table 1.2:	Initial Calibration Data Summarized by Vial
Table 1.3:	Tests for Equal Variances for Repeated Measurements
Table 1.4:	Quadratic Calibration Curve for Initial Data
Table 1.5:	Predicted Benzene Measurements and Limits
Table 2.1:	Fractional Factorial Experiment
Table 3.1:	Data from Designed Experiment
Table 3.2:	Summary Data from Designed Experiment
Table 3.3:	Analysis of Factor Effects and Significant Interactions
Table 3.4:	Tests for Equal Variances for Repeated Measurements
Table 4.1:	Data from Solution Matrix Study
Table 4.2:	Tests for Equal Variances for Repeated Measurements
Table 4.3:	Summary of Solution Matrix Data by Sample
Table 4.4:	Linear Calibration Fit for Solution Matrix Data
Table 4.5:	Quadratic Calibration Fit for Solution Matrix Data
Table 4.6:	Time Stability of 10% Methanol Preparation
Table 4.7:	Time Stability of PHA Preparation
Table 5.1:	Data from PR Study
Table 5.2:	Summary Data from PR Study
Table 5.3:	Tests for Equal Variances for Repeated Measurements
Table 5.4:	Time Study of PR Data
Table 6.1:	PHA Standards with 10% Methanol Calibration
Table 6.2:	Tests for Equal Variances for Repeated Measurements
Table 6.3:	Linear Calibration Fit for PHA Standards Data
Table 6.4:	Quadratic Calibration Fit for PHA Standards Data
Table 6.5:	Predicted Benzene Measurements
Table 6.6:	SAS Code for Inverse Prediction Problem
Table 6.7:	Inverse Prediction Problem

Table 1.1: Data from Initial Calibration Study

Soln Matrix (1- water, 0-air)
Y = benzene measurement (ppm)

Std Conc (ppm)

Rows	Sample	Soln Matrix	Vial	Sample Temp	Std Conc	Y
1	1	1	1	80	0	0.13157
2	2	1	1	80	0	0.1149
3	3	1	1	80	0	0.14221
4	4	1	1	80	0	0.14841
5	5	1	1	80	0	0.00846
6	6	1	1	80	0	0.19124
7	7	1	1	80	0	0.1668
8	8	1	1	80	0	0.0855
9	9	1	1	80	0	0.1518
10	10	1	1	80	0	0.14725
11	11	1	1	80	0	0.12764
12	12	1	1	120	0	0.29192
13	13	1	1	120	0	0.20395
14	14	1	1	120	0	0.22604
15	15	1	1	120	0	0.19484
16	16	1	1	120	0	0.15415
17	17	1	1	120	0	0.29638
18	18	1	1	120	0	0.32212
19	19	1	1	120	0	0.11533
20	20	1	1	120	0	0.23455
21	21	1	1	120	0	0.2381
22	22	1	1	120	0	0.20944
23	23	1	2	80	10	7.71864
24	24	1	2	80	10	7.60326
25	25	1	2	80	10	7.58204
26	26	1	2	80	10	7.60552
27	27	1	2	80	10	7.65123
28	28	1	2	80	10	7.64509
29	29	1	2	80	10	7.69033
30	30	1	2	80	10	7.51932
31	31	1	2	80	10	7.58928
32	32	1	2	80	10	7.59883
33	33	1	2	80	10	7.5579
34	34	1	2	140	10	8.24771
35	35	1	2	140	10	8.34416
36	36	1	2	140	10	8.35106
37	37	1	2	140	10	8.42275
38	38	1	2	140	10	8.25816
39	39	1	2	140	10	8.29113
40	40	1	2	140	10	8.3047
41	41	1	2	140	10	8.22375
42	42	1	2	140	10	8.34427
43	43	1	2	140	10	8.27869
44	44	1	2	140	10	8.31104
45	45	0	3	80	0	0.14594
46	46	0	3	80	0	0.26914
47	47	0	3	80	0	0.10971
48	48	0	3	80	0	0.22543
49	49	0	3	80	0	0.2128
50	50	0	3	80	0	0.36518
51	51	0	3	80	0	0.30561
52	52	0	3	80	0	0.11998
53	53	0	3	80	0	0.22879
54	54	0	3	80	0	0.1304
55	55	0	3	80	0	0.18948
56	56	1	4	80	10	8.06066

Table 1.1 (Con't): Data from Initial Calibration Study

Soln Matrix (1- water, 0-air)

Std Conc (ppm)

Y = benzene measurement (ppm)

Rows	Sample	Soln Matrix	Vial	Sample Temp	Std Conc	Y
57	57	1	4	80	10	7.99343
58	58	1	4	80	10	8.0026
59	59	1	4	80	10	7.94168
60	60	1	4	80	10	7.99827
61	61	1	4	80	10	7.91733
62	62	1	4	80	10	7.85888
63	63	1	4	80	10	7.8495
64	64	1	4	80	10	8.063
65	65	1	4	80	10	7.86198
66	66	1	4	80	10	7.94577
67	67	1	4	140	10	8.11827
68	68	1	4	140	10	8.29336
69	69	1	4	140	10	8.22784
70	70	1	4	140	10	8.24383
71	71	1	4	140	10	8.33885
72	72	1	4	140	10	8.16548
73	73	1	4	140	10	8.27367
74	74	1	4	140	10	8.29012
75	75	1	4	140	10	8.15097
76	76	1	4	140	10	8.11222
77	77	1	4	140	10	8.17492
78	78	1	5	80	30	21.23019
79	79	1	5	80	30	21.11263
80	80	1	5	80	30	21.16985
81	81	1	5	80	30	21.13117
82	82	1	5	80	30	21.07295
83	83	1	5	80	30	20.97327
84	84	1	5	80	30	20.92756
85	85	1	5	80	30	20.87691
86	86	1	5	80	30	20.9427
87	87	1	5	80	30	20.97706
88	88	1	5	80	30	20.81371
89	89	1	5	140	30	21.69709
90	90	1	5	140	30	21.70276
91	91	1	5	140	30	21.73048
92	92	1	5	140	30	21.81815
93	93	1	5	140	30	21.71635
94	94	1	5	140	30	21.74069
95	95	1	5	140	30	21.79634
96	96	1	5	140	30	21.7063
97	97	1	5	140	30	21.6774
98	98	1	5	140	30	21.64387
99	99	1	5	140	30	21.76217
100	100	1	6	80	50	33.18953
101	101	1	6	80	50	33.12656
102	102	1	6	80	50	33.15739
103	103	1	6	80	50	32.97876
104	104	1	6	80	50	32.96033
105	105	1	6	80	50	32.97533
106	106	1	6	80	50	32.92765
107	107	1	6	80	50	32.97259
108	108	1	6	80	50	32.78265
109	109	1	6	80	50	32.88462
110	110	1	6	80	50	32.63007
111	111	1	6	140	50	33.67671
112	112	1	6	140	50	33.63881

Table 1.1 (Con't): Data from Initial Calibration Study

Soln Matrix (1- water, 0-air)

Std Conc (ppm)

Y = benzene measurement (ppm)

Rows	Sample	Soln Matrix	Vial	Sample Temp	Std Conc	Y
113	113	1	6	140	50	33.73264
114	114	1	6	140	50	33.71717
115	115	1	6	140	50	33.61948
116	116	1	6	140	50	33.68389
117	117	1	6	140	50	33.63571
118	118	1	6	140	50	33.63924
119	119	1	6	140	50	33.67085
120	120	1	6	140	50	33.46386
121	121	1	6	140	50	33.54355
122	122	1	7	80	0	0.17847
123	123	1	7	80	0	0.2189
124	124	1	7	80	0	0.2409
125	125	1	7	80	0	0.188801
126	126	1	7	80	0	0.13019
127	127	1	7	80	0	0.09407
128	128	1	7	80	0	0.2267
129	129	1	7	80	0	0.2433
130	130	1	7	80	0	0.28347
131	131	1	7	80	0	0.27036
132	132	1	7	80	0	0.33717

Table 1.2: Initial Calibration Data Summarized by Vial

Y = benzene measurement (ppm)

Rows	Vial	Sample Temp	N	Mean(Std Conc)	Mean(Y)	Std(Y)
1		80	11	0	0.12670727	0.0483066
2	1	120	11	0	0.22607455	0.06156212
3	2	80	11	10	7.61467636	0.05769302
4	2	140	11	10	8.30703818	0.05658214
5	3	80	11	0	0.20931455	0.08144695
6	4	80	11	10	7.95391636	0.07690086
7	4	140	11	10	8.21723	0.07719101
8	5	80	11	30	21.0207273	0.13112975
9	5	140	11	30	21.7265091	0.05093625
10	6	80	11	50	32.9623164	0.16343534
11	6	140	11	50	33.6383555	0.07713798
12	7	80	11	0	0.21930282	0.06923594

Table 1.3 Tests for Equal Variances for Repeated Measurements

Level = Vial-Temp Combination

Tests that the Variances are Equal

Level	Count	Std Dev	MeanAbsDif to Mean	MeanAbsDif to Median
1-80	11	0.0483066	0.0324235	0.0311336
2-80	11	0.0576930	0.0448336	0.0423364
3-80	11	0.0814470	0.0638296	0.0646609
4-80	11	0.0769009	0.0633415	0.0629727
5-80	11	0.1311298	0.1114825	0.1078573
6-80	11	0.1634353	0.1138658	0.1131809
7-80	11	0.0692359	0.0520151	0.0520517

Test	F Ratio	DF Num	DF Den	Prob>F
O'Brien[.5]	3.0250	6	70	0.0109
Brown-Forsythe	2.7436	6	70	0.0188
Levene	3.2906	6	70	0.0066
Bartlett	3.8302	6	.	0.0008

Welch Anova testing Means Equal, allowing Std's Not Equal

F Ratio	DF Num	DF Den	Prob>F
109009.92	6	30.745	0.0000

Tests that the Variances are Equal

Level	Count	Std Dev	MeanAbsDif to Mean	MeanAbsDif to Median
1-120	11	0.0615621	0.0459450	0.0467155
2-140	11	0.0565821	0.0432889	0.0436527
4-140	11	0.0771910	0.0662345	0.0667236
5-140	11	0.0509363	0.0391426	0.0391327
6-140	11	0.0771380	0.0528767	0.0527527

Test	F Ratio	DF Num	DF Den	Prob>F
O'Brien[.5]	0.6917	4	50	0.6012
Brown-Forsythe	0.8204	4	50	0.5183
Levene	0.8222	4	50	0.5172
Bartlett	0.6592	4	.	0.6203

Welch Anova testing Means Equal, allowing Std's Not Equal

F Ratio	DF Num	DF Den	Prob>F
390949.28	4	24.837	0.0000

Table 1.4: Quadratic Calibration Curve for Initial Data

Y = benzene measurement (ppm)

Response: Y

Summary of Fit

Rsquare	0.999909
Root Mean Square Error	0.141959
Mean of Response	12.90891
Observations (or Sum Wgts)	11

Lack of Fit

Source	DF	Sum of Squares	Mean Square	F Ratio
Lack of Fit	3	0.07457026	0.024857	1.4953
Pure Error	4	0.06649525	0.016624	Prob>F
Total Error	7	0.14106551		0.3439

Parameter Estimates

Term	Estimate	Std Error	t Ratio	Prob> t
Intercept	-0.5934	0.15694	-3.78	0.0069
Std Conc^1	0.7617409	0.00975	78.10	0.0000
Std Conc^2	-0.002146	0.00019	-11.58	0.0000
Temp	0.0103795	0.00162	6.39	0.0004

Effect Test

Source	Nparm	DF	Sum of Squares	F Ratio	Prob>F
Poly(Std Conc,2)	2	2	1395.5196	34624.47	0.0000
Temp	1	1	0.8227	40.8229	0.0004

Table 1.5: Predicted Benzene Measurements and Limits

Y = benzene measurement (ppm)

Rows	Std Conc	Temp	Y	Predicted Y	Lower 95% Indiv Y	Upper 95% Indiv Y
1	0	80	0.1287	0.23696318	-0.1431978	0.61712417
2	10	80	7.6147	7.63981872	7.26996944	8.00966799
3	10	140	8.307	8.2625911	7.87677461	8.64840759
4	0	80	0.2093	0.23696318	-0.1431978	0.61712417
5	10	80	7.9539	7.63981872	7.26996944	8.00966799
6	10	140	8.2172	8.2625911	7.87677461	8.64840759
7	30	80	21.0207	21.1582073	20.7456085	21.5708061
8	30	140	21.7265	21.7809796	21.3883603	22.173599
9	50	80	32.9623	32.9601658	32.5381698	33.3821617
10	50	140	33.6384	33.5829382	33.1552619	34.0106144
11	0	80	0.2193	0.23696318	-0.1431978	0.61712417

Table 2.1: Fractional Factorial Experiment

Std Conc:	-1(10)	1(50)	ppm benzene
Soln Matrix:	-1(water)	1(PHA)	
Optical Cell Temp:	-1(85)	1(95)	degrees Celsius
Sample Temp:	-1(80)	1(140)	degrees F

Design 4 Factors

Run Sequence

- 1 - Std Conc
- 2 - Soln Matrix
- 3 - Optical Cell Temp
- 4 - Sample Temp

Sample	1	2	3	4
DAY 1				
18	1	-1	-1	1
12	1	1	-1	-1
4	1	1	-1	-1
28	1	1	-1	-1
1	-1	-1	-1	-1
27	-1	1	-1	1
9	-1	-1	-1	-1
25	-1	-1	-1	-1
13	-1	-1	1	1
29	-1	-1	1	1
22	1	-1	1	-1
31	-1	1	1	-1
24	1	1	1	1
5	-1	-1	1	1
16	1	1	1	1
23	-1	1	1	-1

DAY 2				
3	-1	1	-1	1
10	1	-1	-1	1
2	1	-1	-1	1
17	-1	-1	-1	-1
20	1	1	-1	-1
26	1	-1	-1	1
19	-1	1	-1	1
11	-1	1	-1	1
6	1	-1	1	-1
14	1	-1	1	-1
15	-1	1	1	-1
30	1	-1	1	-1
21	-1	-1	1	1
7	-1	1	1	-1
32	1	1	1	1
8	1	1	1	1

Table 3.1: Data from Designed Experiment

Meas Conc - benzene measurement in ppm

Row#	Day	Set	Sample	Soln	Matrix	Opt	Cell	Temp	Sample	Temp	Std	Conc	Meas Conc
1	1	1	18		-1		-1		1		1		65.50713
2	1	1	18		-1		-1		1		1		65.42088
3	1	1	18		-1		-1		1		1		65.57587
4	1	1	18		-1		-1		1		1		65.22071
5	1	1	18		-1		-1		1		1		65.25409
6	1	1	18		-1		-1		1		1		65.93105
7	1	1	18		-1		-1		1		1		65.01086
8	1	1	18		-1		-1		1		1		65.11868
9	1	1	18		-1		-1		1		1		64.98681
10	1	1	18		-1		-1		1		1		65.36665
11	1	1	18		-1		-1		1		1		65.70443
12	1	1	18		-1		-1		1		1		65.86087
13	1	1	18		-1		-1		1		1		65.58838
14	1	1	18		-1		-1		1		1		66.31398
15	1	1	18		-1		-1		1		1		65.97965
16	1	1	18		-1		-1		1		1		65.98921
17	1	1	18		-1		-1		1		1		66.57765
18	1	1	18		-1		-1		1		1		66.2322
19	1	1	18		-1		-1		1		1		66.14937
20	1	1	18		-1		-1		1		1		66.43178
21	1	1	18		-1		-1		1		1		66.45731
22	1	1	18		-1		-1		1		1		65.32878
23	1	1	18		-1		-1		1		1		65.85291
24	1	1	18		-1		-1		1		1		65.41803
25	1	1	18		-1		-1		1		1		65.72713
26	1	1	18		-1		-1		1		1		65.50383
27	1	1	18		-1		-1		1		1		66.32381
28	1	1	18		-1		-1		1		1		66.44702
29	1	1	18		-1		-1		1		1		65.78588
30	1	1	18		-1		-1		1		1		65.43088
31	1	1	18		-1		-1		1		1		65.12102
32	1	1	18		-1		-1		1		1		65.15324
33	1	1	18		-1		-1		1		1		64.7874
34	1	1	18		-1		-1		1		1		64.784
35	1	1	18		-1		-1		1		1		65.08777
36	1	1	18		-1		-1		1		1		64.74815
37	1	1	18		-1		-1		1		1		64.81644
38	1	1	18		-1		-1		1		1		65.14287
39	1	1	18		-1		-1		1		1		65.19273
40	1	1	18		-1		-1		1		1		64.90885
41	1	1	12		1		-1		-1		1		53.8893
42	1	1	12		1		-1		-1		1		53.4472
43	1	1	12		1		-1		-1		1		53.18092
44	1	1	12		1		-1		-1		1		53.45211
45	1	1	12		1		-1		-1		1		53.38778
46	1	1	12		1		-1		-1		1		53.38213
47	1	1	12		1		-1		-1		1		52.71671
48	1	1	12		1		-1		-1		1		52.90932
49	1	1	12		1		-1		-1		1		53.24338
50	1	1	12		1		-1		-1		1		53.02206
51	1	1	4		1		-1		-1		1		49.92213
52	1	1	4		1		-1		-1		1		50.62249
53	1	1	4		1		-1		-1		1		50.07381
54	1	1	4		1		-1		-1		1		50.3829
55	1	1	4		1		-1		-1		1		49.81109
56	1	1	4		1		-1		-1		1		50.34214
57	1	1	4		1		-1		-1		1		50.64531
58	1	1	4		1		-1		-1		1		50.41595
59	1	1	4		1		-1		-1		1		50.15979
60	1	1	4		1		-1		-1		1		49.88064
61	1	1	28		1		-1		-1		1		60.535
62	1	1	28		1		-1		-1		1		60.20721
63	1	1	28		1		-1		-1		1		60.16485
64	1	1	28		1		-1		-1		1		60.24607
65	1	1	28		1		-1		-1		1		60.12049
66	1	1	28		1		-1		-1		1		60.11395
67	1	1	28		1		-1		-1		1		60.1649
68	1	1	28		1		-1		-1		1		60.11412
69	1	1	28		1		-1		-1		1		60.1288
70	1	1	28		1		-1		-1		1		60.16407
71	1	1	1		-1		-1		-1		-1		8.38853

Meas Conc - benzene measurement in ppm

Rows	Day	Set	Sample	Soln	Matrix	Opt	Cell	Temp	Sample	Temp	Std	Conc	Meas	Conc
72	1	1	1	1	-1	-1	-1	-1	-1	-1	-1	-1	8.5601	
73	1	1	1	1	-1	-1	-1	-1	-1	-1	-1	-1	8.89346	
74	1	1	1	1	-1	-1	-1	-1	-1	-1	-1	-1	8.7215	
75	1	1	1	1	-1	-1	-1	-1	-1	-1	-1	-1	9.00631	
76	1	1	1	1	-1	-1	-1	-1	-1	-1	-1	-1	8.54238	
77	1	1	1	1	-1	-1	-1	-1	-1	-1	-1	-1	8.46378	
78	1	1	1	1	-1	-1	-1	-1	-1	-1	-1	-1	8.31673	
79	1	1	1	1	-1	-1	-1	-1	-1	-1	-1	-1	8.95782	
80	1	1	1	1	-1	-1	-1	-1	-1	-1	-1	-1	9.27343	
81	1	1	1	1	-1	-1	-1	-1	-1	-1	-1	-1	8.80787	
82	1	1	1	1	-1	-1	-1	-1	-1	-1	-1	-1	8.6658	
83	1	1	1	1	-1	-1	-1	-1	-1	-1	-1	-1	8.51648	
84	1	1	1	1	-1	-1	-1	-1	-1	-1	-1	-1	8.59957	
85	1	1	1	1	-1	-1	-1	-1	-1	-1	-1	-1	7.824	
86	1	1	1	1	-1	-1	-1	-1	-1	-1	-1	-1	7.85428	
87	1	1	1	1	-1	-1	-1	-1	-1	-1	-1	-1	8.54	
88	1	1	1	1	-1	-1	-1	-1	-1	-1	-1	-1	8.26882	
89	1	1	1	1	-1	-1	-1	-1	-1	-1	-1	-1	8.74594	
90	1	1	1	1	-1	-1	-1	-1	-1	-1	-1	-1	8.6214	
91	1	1	1	27	1	-1	-1	-1	-1	1	-1	-1	13.97453	
92	1	1	1	27	1	-1	-1	-1	-1	1	-1	-1	14.18084	
93	1	1	1	27	1	-1	-1	-1	-1	1	-1	-1	14.14089	
94	1	1	1	27	1	-1	-1	-1	-1	1	-1	-1	14.35478	
95	1	1	1	27	1	-1	-1	-1	-1	1	-1	-1	14.65954	
96	1	1	1	27	1	-1	-1	-1	-1	1	-1	-1	14.55844	
97	1	1	1	27	1	-1	-1	-1	-1	1	-1	-1	13.78791	
98	1	1	1	27	1	-1	-1	-1	-1	1	-1	-1	13.87851	
99	1	1	1	27	1	-1	-1	-1	-1	1	-1	-1	14.70357	
100	1	1	1	27	1	-1	-1	-1	-1	1	-1	-1	14.62744	
101	1	1	1	27	1	-1	-1	-1	-1	1	-1	-1	14.82586	
102	1	1	1	27	1	-1	-1	-1	-1	1	-1	-1	14.12041	
103	1	1	1	27	1	-1	-1	-1	-1	1	-1	-1	14.39711	
104	1	1	1	27	1	-1	-1	-1	-1	1	-1	-1	14.92893	
105	1	1	1	27	1	-1	-1	-1	-1	1	-1	-1	14.39683	
106	1	1	1	27	1	-1	-1	-1	-1	1	-1	-1	14.52909	
107	1	1	1	27	1	-1	-1	-1	-1	1	-1	-1	14.54977	
108	1	1	1	27	1	-1	-1	-1	-1	1	-1	-1	14.53087	
109	1	1	1	27	1	-1	-1	-1	-1	1	-1	-1	14.47897	
110	1	1	1	27	1	-1	-1	-1	-1	1	-1	-1	14.54258	
111	1	1	1	9	-1	-1	-1	-1	-1	-1	-1	-1	9.8291	
112	1	1	1	9	-1	-1	-1	-1	-1	-1	-1	-1	9.99831	
113	1	1	1	9	-1	-1	-1	-1	-1	-1	-1	-1	9.80708	
114	1	1	1	9	-1	-1	-1	-1	-1	-1	-1	-1	9.52149	
115	1	1	1	9	-1	-1	-1	-1	-1	-1	-1	-1	9.85905	
116	1	1	1	9	-1	-1	-1	-1	-1	-1	-1	-1	10.19914	
117	1	1	1	9	-1	-1	-1	-1	-1	-1	-1	-1	9.88226	
118	1	1	1	9	-1	-1	-1	-1	-1	-1	-1	-1	9.58401	
119	1	1	1	9	-1	-1	-1	-1	-1	-1	-1	-1	9.46888	
120	1	1	1	9	-1	-1	-1	-1	-1	-1	-1	-1	9.75795	
121	1	1	1	25	-1	-1	-1	-1	-1	-1	-1	-1	10.47376	
122	1	1	1	25	-1	-1	-1	-1	-1	-1	-1	-1	9.70951	
123	1	1	1	25	-1	-1	-1	-1	-1	-1	-1	-1	9.64923	
124	1	1	1	25	-1	-1	-1	-1	-1	-1	-1	-1	9.50317	
125	1	1	1	25	-1	-1	-1	-1	-1	-1	-1	-1	10.22151	
126	1	1	1	25	-1	-1	-1	-1	-1	-1	-1	-1	9.81538	
127	1	1	1	25	-1	-1	-1	-1	-1	-1	-1	-1	9.91916	
128	1	1	1	25	-1	-1	-1	-1	-1	-1	-1	-1	9.79961	
129	1	1	1	25	-1	-1	-1	-1	-1	-1	-1	-1	9.74352	
130	1	1	1	25	-1	-1	-1	-1	-1	-1	-1	-1	9.85058	
131	1	1	2	13	-1	-1	-1	-1	1	1	-1	-1	12.40559	
132	1	1	2	13	-1	-1	-1	-1	1	1	-1	-1	13.27381	
133	1	1	2	13	-1	-1	-1	-1	1	1	-1	-1	12.68278	
134	1	1	2	13	-1	-1	-1	-1	1	1	-1	-1	12.74402	
135	1	1	2	13	-1	-1	-1	-1	1	1	-1	-1	12.59328	
136	1	1	2	13	-1	-1	-1	-1	1	1	-1	-1	12.54363	
137	1	1	2	13	-1	-1	-1	-1	1	1	-1	-1	13.05847	
138	1	1	2	13	-1	-1	-1	-1	1	1	-1	-1	12.68122	
139	1	1	2	13	-1	-1	-1	-1	1	1	-1	-1	12.85703	
140	1	1	2	13	-1	-1	-1	-1	1	1	-1	-1	12.15153	
141	1	1	2	29	-1	-1	-1	-1	1	1	-1	-1	12.8099	
142	1	1	2	29	-1	-1	-1	-1	1	1	-1	-1	12.85327	

Table 3.1 (con't): Data from Designed Experiment

Meas Conc - benzene measurement in ppm

Rows	Day	Set	Sample	Soln	Matrix	Opt	Cell	Temp	Sample	Temp	Std	Conc	Meas Conc
143	1	2	29		-1		1	1			-1	12.36515	
144	1	2	29		-1		1	1			-1	13.05966	
145	1	2	29		-1		1	1			-1	12.9899	
146	1	2	29		-1		1	1			-1	13.36694	
147	1	2	29		-1		1	1			-1	12.93347	
148	1	2	29		-1		1	1			-1	12.71732	
149	1	2	29		-1		1	1			-1	12.98613	
150	1	2	29		-1		1	1			-1	12.77691	
151	1	2	22		-1		1	-1			1	49.48412	
152	1	2	22		-1		1	-1			1	49.44804	
153	1	2	22		-1		1	-1			1	48.84034	
154	1	2	22		-1		1	-1			1	48.97804	
155	1	2	22		-1		1	-1			1	48.72696	
156	1	2	22		-1		1	-1			1	48.87057	
157	1	2	22		-1		1	-1			1	48.80644	
158	1	2	22		-1		1	-1			1	49.2517	
159	1	2	22		-1		1	-1			1	48.86395	
160	1	2	22		-1		1	-1			1	48.87427	
161	1	2	31		1		1	-1			-1	11.77276	
162	1	2	31		1		1	-1			-1	11.84696	
163	1	2	31		1		1	-1			-1	11.86397	
164	1	2	31		1		1	-1			-1	11.43123	
165	1	2	31		1		1	-1			-1	11.88186	
166	1	2	31		1		1	-1			-1	11.52101	
167	1	2	31		1		1	-1			-1	11.74534	
168	1	2	31		1		1	-1			-1	12.42051	
169	1	2	31		1		1	-1			-1	11.91189	
170	1	2	31		1		1	-1			-1	11.89118	
171	1	2	24		1		1	1			1	45.8162	
172	1	2	24		1		1	1			1	45.77035	
173	1	2	24		1		1	1			1	45.7308	
174	1	2	24		1		1	1			1	45.8004	
175	1	2	24		1		1	1			1	45.19038	
176	1	2	24		1		1	1			1	45.49245	
177	1	2	24		1		1	1			1	46.07298	
178	1	2	24		1		1	1			1	46.21712	
179	1	2	24		1		1	1			1	45.98821	
180	1	2	24		1		1	1			1	45.78042	
181	1	2	5		-1		1	1			-1	12.55628	
182	1	2	5		-1		1	1			-1	12.64508	
183	1	2	5		-1		1	1			-1	12.83791	
184	1	2	5		-1		1	1			-1	12.99206	
185	1	2	5		-1		1	1			-1	13.07727	
186	1	2	5		-1		1	1			-1	13.38534	
187	1	2	5		-1		1	1			-1	12.85481	
188	1	2	5		-1		1	1			-1	12.82821	
189	1	2	5		-1		1	1			-1	13.31961	
190	1	2	5		-1		1	1			-1	13.19059	
191	1	2	5		-1		1	1			-1	12.75942	
192	1	2	5		-1		1	1			-1	12.5824	
193	1	2	5		-1		1	1			-1	13.00144	
194	1	2	5		-1		1	1			-1	12.94868	
195	1	2	5		-1		1	1			-1	12.7005	
196	1	2	5		-1		1	1			-1	12.57333	
197	1	2	5		-1		1	1			-1	13.12341	
198	1	2	5		-1		1	1			-1	12.83354	
199	1	2	5		-1		1	1			-1	13.11223	
200	1	2	5		-1		1	1			-1	13.03866	
201	1	2	16		1		1	1			1	45.82014	
202	1	2	16		1		1	1			1	45.80207	
203	1	2	16		1		1	1			1	45.41883	
204	1	2	16		1		1	1			1	45.72491	
205	1	2	16		1		1	1			1	46.13076	
206	1	2	16		1		1	1			1	45.48864	
207	1	2	16		1		1	1			1	45.99266	
208	1	2	16		1		1	1			1	45.58976	
209	1	2	16		1		1	1			1	45.81046	
210	1	2	16		1		1	1			1	45.2461	
211	1	2	23		1		1	-1			-1	11.40828	
212	1	2	23		1		1	-1			-1	11.17857	
213	1	2	23		1		1	-1			-1	11.58664	

Table 3.1 (con't): Data from Designed Experiment

Meas Conc - benzene measurement in ppm

Row#	Day	Set	Sample	Soln	Matrix	Opt	Cell	Temp	Sample	Temp	Std	Conc	Meas Conc
214	1	2	23		1			1		-1	-1	11.85876	
215	1	2	23		1			1		-1	-1	11.79036	
216	1	2	23		1			1		-1	-1	12.17793	
217	1	2	23		1			1		-1	-1	11.77773	
218	1	2	23		1			1		-1	-1	12.09013	
219	1	2	23		1			1		-1	-1	12.00399	
220	1	2	23		1			1		-1	-1	11.78834	
221	2	3	6		-1			1		-1	1	55.75312	
222	2	3	6		-1			1		-1	1	55.93447	
223	2	3	6		-1			1		-1	1	55.91771	
224	2	3	6		-1			1		-1	1	55.52485	
225	2	3	6		-1			1		-1	1	55.58602	
226	2	3	6		-1			1		-1	1	55.49733	
227	2	3	6		-1			1		-1	1	55.62831	
228	2	3	6		-1			1		-1	1	55.72814	
229	2	3	6		-1			1		-1	1	55.78778	
230	2	3	6		-1			1		-1	1	55.18705	
231	2	3	14		-1			1		-1	1	55.259	
232	2	3	14		-1			1		-1	1	55.39927	
233	2	3	14		-1			1		-1	1	55.14655	
234	2	3	14		-1			1		-1	1	55.4353	
235	2	3	14		-1			1		-1	1	54.92737	
236	2	3	14		-1			1		-1	1	55.20428	
237	2	3	14		-1			1		-1	1	55.2034	
238	2	3	14		-1			1		-1	1	55.3867	
239	2	3	14		-1			1		-1	1	54.83859	
240	2	3	14		-1			1		-1	1	54.77531	
241	2	3	15		1			1		-1	-1	15.04288	
242	2	3	15		1			1		-1	-1	15.48218	
243	2	3	15		1			1		-1	-1	15.28432	
244	2	3	15		1			1		-1	-1	15.21728	
245	2	3	15		1			1		-1	-1	15.69851	
246	2	3	15		1			1		-1	-1	15.48324	
247	2	3	15		1			1		-1	-1	15.17972	
248	2	3	15		1			1		-1	-1	15.25691	
249	2	3	15		1			1		-1	-1	15.3892	
250	2	3	15		1			1		-1	-1	15.12808	
251	2	3	30		-1			1		-1	1	53.09345	
252	2	3	30		-1			1		-1	1	53.10847	
253	2	3	30		-1			1		-1	1	52.57281	
254	2	3	30		-1			1		-1	1	53.18788	
255	2	3	30		-1			1		-1	1	53.01433	
256	2	3	30		-1			1		-1	1	53.05359	
257	2	3	30		-1			1		-1	1	52.95297	
258	2	3	30		-1			1		-1	1	52.82834	
259	2	3	30		-1			1		-1	1	52.54406	
260	2	3	30		-1			1		-1	1	52.80809	
261	2	3	21		-1			1		-1	-1	14.45727	
262	2	3	21		-1			1		-1	-1	14.45634	
263	2	3	21		-1			1		-1	-1	14.14779	
264	2	3	21		-1			1		-1	-1	14.68811	
265	2	3	21		-1			1		-1	-1	14.38313	
266	2	3	21		-1			1		-1	-1	14.28278	
267	2	3	21		-1			1		-1	-1	15.29121	
268	2	3	21		-1			1		-1	-1	14.47802	
269	2	3	21		-1			1		-1	-1	14.64821	
270	2	3	21		-1			1		-1	-1	14.73933	
271	2	3	7		1			1		-1	-1	14.89911	
272	2	3	7		1			1		-1	-1	15.01243	
273	2	3	7		1			1		-1	-1	14.60401	
274	2	3	7		1			1		-1	-1	15.34403	
275	2	3	7		1			1		-1	-1	15.49153	
276	2	3	7		1			1		-1	-1	14.92724	
277	2	3	7		1			1		-1	-1	15.38995	
278	2	3	7		1			1		-1	-1	15.02309	
279	2	3	7		1			1		-1	-1	14.99257	
280	2	3	7		1			1		-1	-1	15.22852	
281	2	3	32		1			1		-1	-1	59.54505	
282	2	3	32		1			1		-1	-1	59.24363	
283	2	3	32		1			1		-1	-1	59.4325	
284	2	3	32		1			1		-1	-1	58.89358	

Table 3.1 (con't): Data from Designed Experiment

Meas Conc - benzene measurement in ppm

Row#	Day	Set	Sample	Soin	Matrix	Opt	Cell	Temp	Sample	Temp	Std	Conc	Meas Conc
285	2	3	32		1			1		1		1	59.33782
286	2	3	32		1			1		1		1	59.73745
287	2	3	32		1			1		1		1	59.39429
288	2	3	32		1			1		1		1	59.81537
289	2	3	32		1			1		1		1	59.5536
290	2	3	32		1			1		1		1	59.73557
291	2	3	8		1			1		1		1	55.82822
292	2	3	8		1			1		1		1	55.38225
293	2	3	8		1			1		1		1	55.58532
294	2	3	8		1			1		1		1	55.98873
295	2	3	8		1			1		1		1	55.53044
296	2	3	8		1			1		1		1	58.52246
297	2	3	8		1			1		1		1	55.84894
298	2	3	8		1			1		1		1	58.18232
299	2	3	8		1			1		1		1	55.78974
300	2	3	8		1			1		1		1	58.21178
301	2	4	3		1			-1		1		-1	17.80189
302	2	4	3		1			-1		1		-1	18.089233
303	2	4	3		1			-1		1		-1	17.5444
304	2	4	3		1			-1		1		-1	18.18774
305	2	4	3		1			-1		1		-1	17.80864
306	2	4	3		1			-1		1		-1	18.73478
307	2	4	3		1			-1		1		-1	17.88721
308	2	4	3		1			-1		1		-1	18.03773
309	2	4	3		1			-1		1		-1	18.48981
310	2	4	3		1			-1		1		-1	18.08304
311	2	4	3		1			-1		1		-1	18.87783
312	2	4	3		1			-1		1		-1	18.89622
313	2	4	3		1			-1		1		-1	18.87725
314	2	4	3		1			-1		1		-1	18.01832
315	2	4	3		1			-1		1		-1	18.53259
316	2	4	3		1			-1		1		-1	18.50514
317	2	4	3		1			-1		1		-1	18.72227
318	2	4	3		1			-1		1		-1	17.88885
319	2	4	3		1			-1		1		-1	18.0477
320	2	4	3		1			-1		1		-1	18.818
321	2	4	10		-1			-1		1		1	82.88258
322	2	4	10		-1			-1		1		1	82.82082
323	2	4	10		-1			-1		1		1	82.78901
324	2	4	10		-1			-1		1		1	83.04218
325	2	4	10		-1			-1		1		1	82.33418
326	2	4	10		-1			-1		1		1	82.82558
327	2	4	10		-1			-1		1		1	82.88786
328	2	4	10		-1			-1		1		1	82.88148
329	2	4	10		-1			-1		1		1	82.81852
330	2	4	10		-1			-1		1		1	82.38559
331	2	4	2		-1			-1		1		1	82.52032
332	2	4	2		-1			-1		1		1	82.835
333	2	4	2		-1			-1		1		1	82.18902
334	2	4	2		-1			-1		1		1	81.99719
335	2	4	2		-1			-1		1		1	82.32707
336	2	4	2		-1			-1		1		1	83.12121
337	2	4	2		-1			-1		1		1	81.98117
338	2	4	2		-1			-1		1		1	82.05881
339	2	4	2		-1			-1		1		1	82.38204
340	2	4	2		-1			-1		1		1	82.10853
341	2	4	17		-1			-1		-1		-1	14.08959
342	2	4	17		-1			-1		-1		-1	14.58989
343	2	4	17		-1			-1		-1		-1	13.87809
344	2	4	17		-1			-1		-1		-1	13.13487
345	2	4	17		-1			-1		-1		-1	13.87808
346	2	4	17		-1			-1		-1		-1	14.18838
347	2	4	17		-1			-1		-1		-1	13.78074
348	2	4	17		-1			-1		-1		-1	14.10535
349	2	4	17		-1			-1		-1		-1	13.80728
350	2	4	17		-1			-1		-1		-1	13.28429
351	2	4	20		1			-1		-1		1	45.88282
352	2	4	20		1			-1		-1		1	45.53344
353	2	4	20		1			-1		-1		1	45.89272
354	2	4	20		1			-1		-1		1	45.8157
355	2	4	20		1			-1		-1		1	45.25933

Table 3.1 (con't): Data from Designed Experiment

Meas Conc - benzene measurement in ppm

Rows	Day	Set	Sample	Soin	Matrix	Opt	Cell	Temp	Sample	Temp	Std	Conc	Meas Conc
356	2	4	20		1			-1		-1	1		45.73041
357	2	4	20		1			-1		-1	1		45.94117
358	2	4	20		1			-1		-1	1		46.25527
359	2	4	20		1			-1		-1	1		45.99584
360	2	4	20		1			-1		-1	1		46.88133
361	2	4	20		1			-1		-1	1		45.15547
362	2	4	20		1			-1		-1	1		44.85314
363	2	4	20		1			-1		-1	1		44.22766
364	2	4	20		1			-1		-1	1		44.84261
365	2	4	20		1			-1		-1	1		43.88345
366	2	4	20		1			-1		-1	1		43.91394
367	2	4	20		1			-1		-1	1		44.57028
368	2	4	20		1			-1		-1	1		44.30128
369	2	4	20		1			-1		-1	1		44.35421
370	2	4	20		1			-1		-1	1		44.17833
371	2	4	26		-1			-1		1	1		58.74856
372	2	4	26		-1			-1		1	1		58.74421
373	2	4	26		-1			-1		1	1		59.0121
374	2	4	26		-1			-1		1	1		58.43946
375	2	4	26		-1			-1		1	1		58.22923
376	2	4	26		-1			-1		1	1		58.57983
377	2	4	26		-1			-1		1	1		58.52016
378	2	4	26		-1			-1		1	1		58.84409
379	2	4	26		-1			-1		1	1		58.63424
380	2	4	26		-1			-1		1	1		59.12725
381	2	4	19		1			-1		1	-1		15.58134
382	2	4	19		1			-1		1	-1		15.98402
383	2	4	19		1			-1		1	-1		15.84056
384	2	4	19		1			-1		1	-1		15.50854
385	2	4	19		1			-1		1	-1		16.1746
386	2	4	19		1			-1		1	-1		15.7239
387	2	4	19		1			-1		1	-1		14.98512
388	2	4	19		1			-1		1	-1		15.68838
389	2	4	19		1			-1		1	-1		15.42598
390	2	4	19		1			-1		1	-1		15.75082
391	2	4	11		1			-1		1	-1		15.88474
392	2	4	11		1			-1		1	-1		15.82855
393	2	4	11		1			-1		1	-1		16.06788
394	2	4	11		1			-1		1	-1		16.25082
395	2	4	11		1			-1		1	-1		15.21227
396	2	4	11		1			-1		1	-1		15.75995
397	2	4	11		1			-1		1	-1		16.16904
398	2	4	11		1			-1		1	-1		15.90315
399	2	4	11		1			-1		1	-1		15.98121
400	2	4	11		1			-1		1	-1		15.65042

Table 3.2: Summary Data from Designed Experiment

Meas Conc - benzene measurement in ppm
N = number of repeated measurements

Rows	Sample	N	Day	Set	Soln	Matrix	Opt	Cell Temp	Sample Temp	Std Conc	Meas Conc
1	1	20	1	1		-1		-1	-1	-1	8.579441
2	2	10	2	4		-1		-1	1	1	62.326636
3	3	20	2	4		1		-1	1	-1	18.2354222
4	4	10	1	1		1		-1	-1	1	50.231625
5	5	20	1	2		-1		1	1	-1	12.9183458
6	6	10	2	3		-1		1	-1	1	55.651556
7	7	10	2	3		1		1	-1	-1	15.089248
8	8	10	2	3		1		1	1	1	55.860817
9	9	10	1	1		-1		-1	-1	-1	9.770725
10	10	10	2	4		-1		-1	1	1	62.702824
11	11	10	2	4		1		-1	1	-1	15.868551
12	12	10	1	1		1		-1	-1	1	53.239089
13	13	10	1	2		-1		1	1	-1	12.697136
14	14	10	2	3		-1		1	-1	1	55.157577
15	15	10	2	3		1		1	-1	-1	15.314207
16	16	10	1	2		1		1	1	1	45.682635
17	17	10	2	4		-1		-1	-1	-1	13.883452
18	18	40	1	1		-1		-1	1	1	65.5548985
19	19	10	2	4		1		-1	1	-1	15.642104
20	20	20	2	4		1		-1	-1	1	45.15026
21	21	10	2	3		-1		1	1	-1	14.555019
22	22	10	1	2		-1		1	-1	1	48.994443
23	23	10	1	2		1		1	-1	-1	11.766073
24	24	10	1	2		1		1	1	1	45.783931
25	25	10	1	1		-1		-1	-1	-1	9.868563
26	26	10	2	4		-1		-1	1	1	58.697713
27	27	20	1	1		1		-1	1	-1	14.3933775
28	28	10	1	1		1		-1	-1	1	60.195906
29	29	10	1	2		-1		1	1	-1	12.885865
30	30	10	2	3		-1		1	-1	1	52.896494
31	31	10	1	2		1		1	-1	-1	11.817868
32	32	10	2	3		1		1	1	1	59.458886

Table 3.3: Analysis of Factor Effects and Significant Interactions

(Meas - Std) = Measured Benzene - Standard Benzene (ppm)

Response: (Meas - Std)

Summary of Fit

Rsquare	0.517072
Root Mean Square Error	3.853278
Mean of Response	4.089709
Observations (or Sum Wgts)	32

Parameter Estimates

Term	Estimate	Std Error	t Ratio	Prob> t
Intercept	4.089709	0.68117	6.00	0.0000
Soln Matrix	-0.731584	0.68117	-1.07	0.2935
Opt Cell Temp	-1.181578	0.68117	-1.73	0.0956
Sample Temp	1.7393011	0.68117	2.55	0.0174
Std Conc	0.7593717	0.68117	1.11	0.2760
Soln Mat*Std Conc	-2.167103	0.68117	-3.18	0.0040
Soln Mat*Sample T	-1.231711	0.68117	-1.81	0.0831
Std Conc*Sample T	0.4201608	0.68117	0.62	0.5432

Effect Test

Source	Nparm	DF	Sum of Squares	F Ratio	Prob>F
Soln Matrix	1	1	17.12689	1.1535	0.2935
Opt Cell Temp	1	1	44.67603	3.0089	0.0956
Sample Temp	1	1	96.80538	6.5199	0.0174
Std Conc	1	1	18.45265	1.2428	0.2760
Soln Mat*Std Conc	1	1	150.28273	10.1216	0.0040
Soln Mat*Sample T	1	1	48.54755	3.2697	0.0831
Std Conc*Sample T	1	1	5.64912	0.3805	0.5432

Table 3.4: Tests for Equal Variances for Repeated Measurements

Level = sample number

Tests that the Variances are Equal					
Level	Count	Std Dev	MeanAbsDif to Mean	MeanAbsDif to Median	
1	20	0.3501591	0.2498390	0.2498390	
2	10	0.3562821	0.2664920	0.2664920	
3	20	0.4269360	0.3734891	0.3687209	
4	10	0.2912556	0.2461330	0.2461330	
5	20	0.2430870	0.2011978	0.2011978	
6	10	0.2208027	0.1666840	0.1666840	
7	10	0.2591384	0.2154076	0.2021760	
8	10	0.3610873	0.2883984	0.2860230	
9	10	0.2220970	0.1678070	0.1678070	
10	10	0.2228253	0.1773460	0.1773460	
11	10	0.2946912	0.2057650	0.2057650	
12	10	0.2912574	0.2294692	0.2286110	
13	10	0.3178767	0.2289572	0.2260860	
14	10	0.2365512	0.1884976	0.1793330	
15	10	0.1965033	0.1552544	0.1492770	
16	10	0.2670397	0.2114770	0.2114770	
17	10	0.4252121	0.3064020	0.3064020	
18	40	0.5326480	0.4459944	0.4410425	
19	10	0.3202194	0.2182000	0.2182000	
20	20	0.8541054	0.7450440	0.7445230	
21	10	0.3175992	0.2289568	0.2135570	
22	10	0.2939213	0.2401062	0.2120510	
23	10	0.3066717	0.2249458	0.2181610	
24	10	0.2880503	0.1910510	0.1910510	
25	10	0.2835976	0.2017482	0.1875550	
26	10	0.2757237	0.2171290	0.2171290	
27	20	0.2934144	0.2314772	0.2221895	
28	10	0.1265808	0.0801124	0.0876600	
29	10	0.2588514	0.1813550	0.1813550	
30	10	0.2402546	0.2065372	0.1952420	
31	10	0.2712165	0.1875700	0.1875700	
32	10	0.2303150	0.1785220	0.1785220	
Test	F Ratio	DF Num	DF Den	Prob>F	
O'Brien(.5)	6.6105	31	368	0.0000	
Brown-Forsythe	5.1836	31	368	0.0000	
Levene	5.8768	31	368	0.0000	
Bartlett	3.8033	31	.	0.0000	
Welch Anova testing Means Equal, allowing Std's Not Equal					
F Ratio	DF Num	DF Den	Prob>F		
67786.638	31	114.25	0.0000		

Table 4.1: Data from Solution Matrix Study

Ben Std (ppm) Ben Meas (ppm) Meas Res = Std - Meas (ppm)
AB-Air Blank

Row	Sample	Soln Matrix	Ben Std	Ben Meas	Meas Res
1	AB		0	-0.84998	-0.84998
2	AB		0	-0.34419	-0.34419
3	AB		0	-0.32103	-0.32103
4	AB		0	-0.16603	-0.16603
5	AB		0	-0.39295	-0.39295
6	12	10% METH	30	39.77394	9.77394
7	12	10% METH	30	39.47152	9.47152
8	12	10% METH	30	39.37978	9.37978
9	12	10% METH	30	39.50319	9.50319
10	12	10% METH	30	39.91367	9.91367
11	16	10% METH	50	66.47563	16.47563
12	16	10% METH	50	66.35574	16.35574
13	16	10% METH	50	66.54255	16.54255
14	16	10% METH	50	66.37568	16.37568
15	16	10% METH	50	66.57366	16.57366
16	4	10% METH	0	0.14261	0.14261
17	4	10% METH	0	-0.14528	-0.14528
18	4	10% METH	0	-0.0323	-0.0323
19	4	10% METH	0	0.21277	0.21277
20	4	10% METH	0	0.27179	0.27179
21	7	10% METH	8.5	10.60352	2.10352
22	7	10% METH	8.5	10.92929	2.42929
23	7	10% METH	8.5	11.22902	2.72902
24	7	10% METH	8.5	10.87516	2.37516
25	7	10% METH	8.5	11.34597	2.84597
26	18	10% METH	50	66.09776	16.09776
27	18	10% METH	50	65.68311	15.68311
28	18	10% METH	50	65.89748	15.89748
29	18	10% METH	50	65.70853	15.70853
30	18	10% METH	50	65.81303	15.81303
31	6	10% METH	8.5	11.01632	2.51632
32	6	10% METH	8.5	10.94755	2.44755
33	6	10% METH	8.5	11.23273	2.73273
34	6	10% METH	8.5	11.68676	3.18676
35	6	10% METH	8.5	11.53977	3.03977
36	11	10% METH	30	39.57084	9.57084
37	11	10% METH	30	39.91191	9.91191
38	11	10% METH	30	40.15063	10.15063
39	11	10% METH	30	39.95545	9.95545
40	11	10% METH	30	39.95749	9.95749
41	3	10% METH	0	-1.04686	-1.04686
42	3	10% METH	0	-0.46477	-0.46477
43	3	10% METH	0	-0.55638	-0.55638
44	3	10% METH	0	0.07748	0.07748
45	3	10% METH	0	-0.14036	-0.14036
46	1	10% METH	0	-0.83631	-0.83631
47	1	10% METH	0	-0.76485	-0.76485
48	1	10% METH	0	-0.41941	-0.41941
49	1	10% METH	0	-0.42111	-0.42111
50	1	10% METH	0	-0.40282	-0.40282
51	17	10% METH	50	66.60383	16.60383
52	17	10% METH	50	66.41769	16.41769
53	17	10% METH	50	66.55096	16.55096
54	17	10% METH	50	66.45265	16.45265
55	17	10% METH	50	66.75877	16.75877
56	19	10% METH	8.5	10.79207	2.29207
57	19	10% METH	8.5	11.19322	2.69322
58	19	10% METH	8.5	11.25927	2.75927
59	19	10% METH	8.5	10.74262	2.24262
60	19	10% METH	8.5	11.02168	2.52168
61	14	10% METH	30	37.84676	7.84676
62	14	10% METH	30	38.25079	8.25079
63	14	10% METH	30	38.28799	8.28799
64	14	10% METH	30	38.47384	8.47384
65	14	10% METH	30	38.27759	8.27759
66	8	10% METH	8.5	10.54972	2.04972
67	8	10% METH	8.5	11.07576	2.57576

Table 4.1 (con't): Data from Solution Matrix Study

Ben Std (ppm) Ben Meas (ppm) Meas Res = Std - Meas (ppm)

Row	Sample	Soln Matrix	Ben Std	Ben Meas	Meas Res
68	10	10% METH	8.5	11.22459	2.72459
69	18	10% METH	8.5	11.02795	2.52795
70	18	10% METH	8.5	11.10517	2.60517
71	12	10% METH	0	-0.9049	-0.9049
72	12	10% METH	0	-0.43429	-0.43429
73	12	10% METH	0	-0.11271	-0.11271
74	12	10% METH	0	-0.10967	-0.10967
75	12	10% METH	0	-0.21274	-0.21274
76	19	10% METH	50	65.69588	15.69588
77	19	10% METH	50	65.90359	15.90359
78	19	10% METH	50	66.06059	16.06059
79	19	10% METH	50	65.96348	15.96348
80	19	10% METH	50	66.05424	16.05424
81	15	10% METH	30	40.1388	10.1388
82	15	10% METH	30	39.88678	9.88678
83	15	10% METH	30	39.60692	9.60692
84	15	10% METH	30	39.86271	9.86271
85	15	10% METH	30	39.66508	9.66508
86	13	10% METH	30	39.32112	9.32112
87	13	10% METH	30	39.23036	9.23036
88	13	10% METH	30	39.57061	9.57061
89	13	10% METH	30	39.14924	9.14924
90	13	10% METH	30	40.11661	10.11661
91	10	10% METH	8.5	10.27459	1.77459
92	10	10% METH	8.5	10.61563	2.11563
93	10	10% METH	8.5	10.7662	2.2662
94	10	10% METH	8.5	10.68682	2.18682
95	10	10% METH	8.5	10.40623	1.90623
96	20	10% METH	50	63.01056	13.01056
97	20	10% METH	50	63.12938	13.12938
98	20	10% METH	50	63.51512	13.51512
99	20	10% METH	50	63.14193	13.14193
100	20	10% METH	50	63.25729	13.25729
101	5	10% METH	0	-0.0734	-0.0734
102	5	10% METH	0	0.14545	0.14545
103	5	10% METH	0	0.3176	0.3176
104	5	10% METH	0	0.37807	0.37807
105	5	10% METH	0	0.80517	0.80517
106	35	10% METH	50	65.26662	15.26662
107	35	10% METH	50	65.48147	15.48147
108	35	10% METH	50	65.58974	15.58974
109	35	10% METH	50	65.87126	15.87126
110	35	10% METH	50	66.17568	16.17568
111	21	PHA	0	0.32828	0.32828
112	21	PHA	0	0.44313	0.44313
113	21	PHA	0	0.69627	0.69627
114	21	PHA	0	0.22993	0.22993
115	21	PHA	0	0.61789	0.61789
116	30	PHA	30	39.73798	9.73798
117	30	PHA	30	39.93182	9.93182
118	30	PHA	30	40.27683	10.27683
119	30	PHA	30	39.93504	9.93504
120	30	PHA	30	40.27122	10.27122
121	29	PHA	30	39.13848	9.13848
122	29	PHA	30	39.2536	9.2536
123	29	PHA	30	39.3772	9.3772
124	29	PHA	30	39.38414	9.38414
125	29	PHA	30	39.79552	9.79552
126	34	PHA	50	65.56908	15.56908
127	34	PHA	50	65.83411	15.83411
128	34	PHA	50	66.05009	16.05009
129	34	PHA	50	65.93581	15.93581
130	34	PHA	50	66.46484	16.46484
131	23	PHA	0	0.0615	0.0615
132	23	PHA	0	0.6856	0.6856
133	23	PHA	0	0.77612	0.77612
134	23	PHA	0	0.96831	0.96831

Table 4.1 (con't): Data from Solution Matrix Study

Ben Std (ppm) Ben Meas (ppm) Meas Res = Std - Meas (ppm)

Row	Sample	Soln Matrix	Ben Std	Ben Meas	Meas Res
135	23	PHA	0	1.36489	1.36489
136	33	PHA	50	64.68394	14.68394
137	33	PHA	50	64.72086	14.72086
138	33	PHA	50	65.11858	15.11858
139	33	PHA	50	65.25486	15.25486
140	33	PHA	50	65.50217	15.50217
141	22	PHA	0	0.52922	0.52922
142	22	PHA	0	0.74998	0.74998
143	22	PHA	0	0.58	0.58
144	22	PHA	0	1.61764	1.61764
145	22	PHA	0	0.95881	0.95881
146	28	PHA	30	40.13964	10.13964
147	28	PHA	30	40.342	10.342
148	28	PHA	30	40.1082	10.1082
149	28	PHA	30	40.62416	10.62416
150	28	PHA	30	40.0825	10.0825
151	27	PHA	30	41.82557	11.82557
152	27	PHA	30	42.14285	12.14285
153	27	PHA	30	41.87814	11.87814
154	27	PHA	30	42.03688	12.03688
155	27	PHA	30	42.56201	12.56201
156	31	PHA	50	63.40255	13.40255
157	31	PHA	50	63.91504	13.91504
158	31	PHA	50	63.7344	13.7344
159	31	PHA	50	63.91013	13.91013
160	31	PHA	50	64.04967	14.04967
161	25	PHA	0	0.78065	0.78065
162	25	PHA	0	0.77883	0.77883
163	25	PHA	0	1.06523	1.06523
164	25	PHA	0	0.84866	0.84866
165	25	PHA	0	1.23341	1.23341
166	24	PHA	0	-1.04448	-1.04448
167	24	PHA	0	-0.046686	-0.046686
168	24	PHA	0	-0.031499	-0.031499
169	24	PHA	0	-0.040468	-0.040468
170	24	PHA	0	-0.3085	-0.3085
171	26	PHA	30	40.49129	10.49129
172	26	PHA	30	40.50966	10.50966
173	26	PHA	30	40.52385	10.52385
174	26	PHA	30	40.94539	10.94539
175	26	PHA	30	40.98933	10.98933
176	32	PHA	50	66.32249	16.32249
177	32	PHA	50	66.23838	16.23838
178	32	PHA	50	66.22123	16.22123
179	32	PHA	50	66.40597	16.40597
180	32	PHA	50	66.25004	16.25004
181	00	10% METH	0	-0.69381	-0.69381
182	00	10% METH	0	-0.63082	-0.63082
183	00	10% METH	0	-0.56337	-0.56337
184	00	10% METH	0	0.06718	0.06718
185	00	10% METH	0	-0.2317	-0.2317

Table 4.2: Tests for Equal Variances for Repeated Measurements
Level = sample number

Tests that the Variances are Equal					
Level	Count	Std Dev	MeanAbsDif to Mean	MeanAbsDif to Median	
00	5	0.3210616	0.2625952	0.2455120	
1	5	0.2131175	0.1853440	0.1561260	
10	5	0.2038933	0.1675872	0.1686780	
11	5	0.2105883	0.1354336	0.1255220	
12	5	0.2250920	0.1883080	0.1735960	
13	5	0.3906464	0.2928176	0.2796760	
14	5	0.2304109	0.1522536	0.1349360	
15	5	0.2101151	0.1568464	0.1555300	
16	5	0.0972800	0.0791856	0.0903020	
17	5	0.1353113	0.0996160	0.1090260	
18	5	0.1676085	0.1261104	0.1376100	
19	5	0.1491355	0.1086568	0.1150500	
2	5	0.3345915	0.2517864	0.2433680	
20	5	0.1912069	0.1402792	0.1290040	
21	5	0.1944683	0.1551840	0.1741600	
22	5	0.4415941	0.3208760	0.3274420	
23	5	0.4749594	0.3181872	0.3353240	
24	5	0.4352526	0.3057307	0.2574462	
25	5	0.2007173	0.1662112	0.1612340	
26	5	0.2521995	0.2203648	0.1895920	
27	5	0.2929215	0.2106720	0.2214240	
28	5	0.2282379	0.1790240	0.1613800	
29	5	0.2482497	0.1622928	0.1589040	
3	5	0.4298405	0.3157904	0.3263940	
30	5	0.2361371	0.1947576	0.1762940	
31	5	0.2499700	0.1871064	0.1665340	
32	5	0.0765949	0.0612864	0.0561020	
33	5	0.3511683	0.2829456	0.2977020	
34	5	0.3285003	0.2293432	0.2426880	
35	5	0.3537926	0.2772128	0.2814240	
4	5	0.1740703	0.1429664	0.1464600	
5	5	0.3255196	0.2228424	0.2343320	
6	5	0.3221458	0.2629112	0.2958140	
7	5	0.2957485	0.2327224	0.2300880	
8	5	0.2601527	0.1787672	0.1563000	
9	5	0.2315646	0.1875416	0.2178680	
AB	5	0.2576452	0.1740576	0.1558060	
Test	F Ratio	DF Num	DF Den	Prob>F	
O'Brien[.5]	0.7899	36	148	0.7938	
Brown-Forsythe	0.6840	36	148	0.9084	
Levene	1.0035	36	148	0.4734	
Bartlett	0.9933	36	.	0.4799	
Warning: Small sample sizes. Use Caution.					
Welch Anova testing Means Equal, allowing Std's Not Equal					
F Ratio	DF Num	DF Den	Prob>F		
3693.9878	36	52.012	0.0000		

Table 4.3: Summary of Solution Matrix Data by Sample

Benzene Values in ppm

AB = air blank

Rows	Time (hrs)	Sample	Soln Matrix	N	Mean(Ben Std)	Mean(Ben Meas)	Std(Ben Meas)
1	• 00	10% METH		5	0	-0.410504	0.32106164
2	1.316 1	10% METH		5	0	-0.5689	0.21311747
3	2.179 10	10% METH		5	8.5	10.549894	0.20389328
4	0.613 11	10% METH		5	30	39.909224	0.21058825
5	0.000 12	10% METH		5	30	39.60842	0.22509201
6	2.091 13	10% METH		5	30	39.477588	0.39064643
7	1.620 14	10% METH		5	30	38.227394	0.23041092
8	2.003 15	10% METH		5	30	39.832058	0.21011513
9	0.105 16	10% METH		5	50	66.464692	0.09728004
10	1.419 17	10% METH		5	50	66.55678	0.13531133
11	0.421 18	10% METH		5	50	65.839982	0.16760846
12	1.910 19	10% METH		5	50	65.935556	0.14913547
13	1.811 2	10% METH		5	0	-0.354882	0.33459153
14	2.265 20	10% METH		5	50	63.210856	0.1912069
15	0.000 21	PHA		5	0	0.4631	0.1944683
16	0.555 22	PHA		5	0	0.88713	0.4415941
17	0.369 23	PHA		5	0	0.771284	0.47495942
18	1.079 24	PHA		5	0	-0.2943266	0.43525263
19	0.974 25	PHA		5	0	0.941556	0.20071729
20	1.178 26	PHA		5	30	40.691904	0.25219951
21	0.776 27	PHA		5	30	42.08909	0.29292151
22	0.660 28	PHA		5	30	40.2593	0.22823786
23	0.181 29	PHA		5	30	39.389788	0.24824972
24	0.714 3	10% METH		5	0	-0.426178	0.42984048
25	0.088 30	PHA		5	30	40.030578	0.23613707
26	0.875 31	PHA		5	50	63.802358	0.24996999
27	1.273 32	PHA		5	50	66.287622	0.07659488
28	0.459 33	PHA		5	50	65.056082	0.35116828
29	0.282 34	PHA		5	50	65.970786	0.32850032
30	2.444 35	10% METH		5	50	65.676954	0.35379256
31	0.212 4	10% METH		5	0	0.089918	0.17407026
32	2.350 5	10% METH		5	0	0.314578	0.32551962
33	0.518 6	10% METH		5	8.5	11.284626	0.32214593
34	0.326 7	10% METH		5	8.5	10.896592	0.29574848
35	1.715 8	10% METH		5	8.5	10.996638	0.28015271
36	1.526 9	10% METH		5	8.5	11.001772	0.23156456
37	• AB			5	0	-0.414836	0.25764518

Table 4.4: Linear Calibration Fit for Solution Matrix Data

Linear Fit Soln Matrix=10% METH					
Summary of Fit					
Rsquare		0.999325			
Root Mean Square Error		0.719706			
Mean of Response		29.28241			
Observations (or Sum Wgts)		22			
Analysis of Variance					
Source	DF	Sum of Squares	Mean Square	F Ratio	
Model	1	15334.782	15334.8	29605.13	
Error	20	10.360	0.5	Prob>F	
C Total	21	15345.142		0.0000	
Parameter Estimates					
Term	Estimate	Std Error	t Ratio	Prob> t	
Intercept	-0.208984	0.23005	-0.91	0.3745	
Mean(Ben Std)	1.3173822	0.00766	172.06	0.0000	

Linear Fit Soln Matrix=PHA					
Summary of Fit					
Rsquare		0.998702			
Root Mean Square Error		1.025262			
Mean of Response		33.31045			
Observations (or Sum Wgts)		14			
Analysis of Variance					
Source	DF	Sum of Squares	Mean Square	F Ratio	
Model	1	9707.1182	9707.12	9234.645	
Error	12	12.6140	1.05	Prob>F	
C Total	13	9719.7321		0.0000	
Parameter Estimates					
Term	Estimate	Std Error	t Ratio	Prob> t	
Intercept	0.8278121	0.43513	1.90	0.0814	
Mean(Ben Std)	1.2993054	0.01352	96.10	0.0000	

Table 4.5: Quadratic Calibration Fit for Solution Matrix Data

Polynomial Fit, degree=2 Soln Matrix=10% METH

Summary of Fit

Rsquare	0.999328
Root Mean Square Error	0.736734
Mean of Response	29.28241
Observations (or Sum Wgts)	22

Analysis of Variance

Source	DF	Sum of Squares	Mean Square	F Ratio
Model	2	15334.829	7667.41	14126.28
Error	19	10.313	0.54	Prob>F
C Total	21	15345.142		0.0000

Parameter Estimates

Term	Estimate	Std Error	t Ratio	Prob> t
Intercept	-0.249473	0.2729	-0.91	0.3721
Mean(Ben Std)	1.3265886	0.03232	41.04	0.0000
Mean(Ben Std)^2	-0.000183	0.00062	-0.29	0.7723

Polynomial Fit, degree=2 Soln Matrix=PHA

Summary of Fit

Rsquare	0.999091
Root Mean Square Error	0.896201
Mean of Response	33.31045
Observations (or Sum Wgts)	14

Analysis of Variance

Source	DF	Sum of Squares	Mean Square	F Ratio
Model	2	9710.8972	4855.45	6045.303
Error	11	8.8349	0.80	Prob>F
C Total	13	9719.7321		0.0000

Parameter Estimates

Term	Estimate	Std Error	t Ratio	Prob> t
Intercept	0.5537487	0.40079	1.38	0.1945
Mean(Ben Std)	1.3864347	0.04187	33.11	0.0000
Mean(Ben Std)^2	-0.001839	0.00085	-2.17	0.0529

Table 4.6: Time Stability of 10% Methanol Preparation

Mean(Ben Meas) By Time (hrs)

Fitting

- Linear Fit Mean(Ben Std)=0
- Linear Fit Mean(Ben Std)=8.5
- Linear Fit Mean(Ben Std)=30
- Linear Fit Mean(Ben Std)=50

Table 4.6 (con't): Time Stability of 10% Methanol Preparation

Linear Fit Mean(Ben Std)=0					
Summary of Fit					
Rsquare		0.046233			
Root Mean Square Error		0.421745			
Mean of Response		-0.18909			
Observations (or Sum Wgts)		5			
Analysis of Variance					
Source	DF	Sum of Squares	Mean Square	F Ratio	
Model	1	0.02586639	0.025866	0.1454	
Error	3	0.53360724	0.177869	Prob>F	
C Total	4	0.55947363		0.7284	
Parameter Estimates					
Term	Estimate	Std Error	t Ratio	Prob> t	
Intercept	-0.310279	0.36955	-0.84	0.4628	
Time (hrs)	0.0946142	0.24811	0.38	0.7284	
Linear Fit Mean(Ben Std)=8.5					
Summary of Fit					
Rsquare		0.55066			
Root Mean Square Error		0.20399			
Mean of Response		10.9659			
Observations (or Sum Wgts)		5			
Analysis of Variance					
Source	DF	Sum of Squares	Mean Square	F Ratio	
Model	1	0.15298493	0.152985	3.6765	
Error	3	0.12483594	0.041612	Prob>F	
C Total	4	0.27782088		0.1510	
Parameter Estimates					
Term	Estimate	Std Error	t Ratio	Prob> t	
Intercept	11.273011	0.18433	61.16	0.0000	
Time (hrs)	-0.245173	0.12787	-1.92	0.1510	

Table 4.6 (con't): Time Stability of 10% Methanol Preparation

Linear Fit Mean(Ben Std)=30

Summary of Fit

Rsquare	0.062711
Root Mean Square Error	0.764327
Mean of Response	39.41094
Observations (or Sum Wgts)	5

Analysis of Variance

Source	DF	Sum of Squares	Mean Square	F Ratio
Model	1	0.1172602	0.117260	0.2007
Error	3	1.7525885	0.584196	Prob>F
C Total	4	1.8698488		0.6845

Parameter Estimates

Term	Estimate	Std Error	t Ratio	Prob> t
Intercept	39.646719	0.62754	63.18	0.0000
Time (hrs)	-0.186316	0.41587	-0.45	0.6845

Linear Fit Mean(Ben Std)=50

Summary of Fit

Rsquare	0.278863
Root Mean Square Error	1.166376
Mean of Response	65.61414
Observations (or Sum Wgts)	6

Analysis of Variance

Source	DF	Sum of Squares	Mean Square	F Ratio
Model	1	2.1043131	2.10431	1.5468
Error	4	5.4417283	1.36043	Prob>F
C Total	5	7.5460413		0.2815

Parameter Estimates

Term	Estimate	Std Error	t Ratio	Prob> t
Intercept	66.566252	0.90156	73.83	0.0000
Time (hrs)	-0.667046	0.53634	-1.24	0.2815

Table 4.7: Time Stability of PHA Preparation

Mean(Ben Meas) By Time (hrs)

Table 4.7 (con't): Time Stability of PHA Preparation

Linear Fit Mean(Ben Std)=0

Summary of Fit

Rsquare	0.09062
Root Mean Square Error	0.560475
Mean of Response	0.553749
Observations (or Sum Wgts)	5

Analysis of Variance

Source	DF	Sum of Squares	Mean Square	F Ratio
Model	1	0.0939107	0.093911	0.2990
Error	3	0.9423974	0.314132	Prob>F
C Total	4	1.0363081		0.6226

Parameter Estimates

Term	Estimate	Std Error	t Ratio	Prob> t
Intercept	0.7597898	0.45258	1.68	0.1918
Time (hrs)	-0.345997	0.63281	-0.55	0.6226

Linear Fit Mean(Ben Std)=30

Summary of Fit

Rsquare	0.357431
Root Mean Square Error	0.933973
Mean of Response	40.49213
Observations (or Sum Wgts)	5

Analysis of Variance

Source	DF	Sum of Squares	Mean Square	F Ratio
Model	1	1.4556714	1.45567	1.6688
Error	3	2.6169174	0.87231	Prob>F
C Total	4	4.0725888		0.2869

Parameter Estimates

Term	Estimate	Std Error	t Ratio	Prob> t
Intercept	39.716198	0.73161	54.29	0.0000
Time (hrs)	1.3456797	1.0417	1.29	0.2869

Table 4.7 (con't): Time Stability of PHA Preparation

Linear Fit Mean(Ben Std)=50

Summary of Fit

Rsquare	0.003167
Root Mean Square Error	1.362764
Mean of Response	65.27921
Observations (or Sum Wgts)	4

Analysis of Variance

Source	DF	Sum of Squares	Mean Square	F Ratio
Model	1	0.0117996	0.01180	0.0064
Error	2	3.7142504	1.85713	Prob>F
C Total	3	3.7260501		0.9437

Parameter Estimates

Term	Estimate	Std Error	t Ratio	Prob> t
Intercept	65.177052	1.45151	44.90	0.0005
Time (hrs)	0.1414519	1.77458	0.08	0.9437

Table 5.1: Data from PR Study
Ben Meas in ppm

Rows	Sample	Ben Meas	Rows	Sample	Ben Meas	Rows	Sample	Ben Meas	Rows	Sample	Ben Meas
1	999	-0.19334	72	111	59.72482	143	121	35.8532	214	129	34.75734
2	999	-0.25969	73	111	60.33495	144	121	35.8546	215	129	34.85804
3	999	-0.2093	74	111	60.34708	145	121	35.8893	216	130	33.57825
4	999	-0.39797	75	111	59.8987	146	53	31.94838	217	130	33.77693
5	999	-0.30894	76	112	62.10652	147	53	32.05491	218	130	33.78544
6	101	69.10377	77	112	63.21442	148	53	32.04588	219	130	34.15701
7	101	68.54425	78	112	63.05088	149	53	31.81088	220	130	34.08804
8	101	69.19318	79	112	63.20584	150	53	32.53487	221	47	36.8354
9	101	69.38837	80	112	62.94489	151	38	37.25439	222	47	36.7548
10	101	69.24288	81	44	30.11154	152	38	37.1126	223	47	38.94451
11	102	67.28651	82	44	29.94359	153	38	37.38132	224	47	37.08758
12	102	67.05527	83	44	30.42192	154	38	37.67402	225	47	36.88148
13	102	68.05871	84	44	30.52904	155	38	37.354	226	50	38.46827
14	102	67.97137	85	44	30.83273	156	45	37.92014	227	50	38.82883
15	102	67.38458	86	113	52.96329	157	45	38.47528	228	50	38.86736
16	103	70.18487	87	113	53.19398	158	45	38.10764	229	50	39.02715
17	103	69.58895	88	113	53.22433	159	45	38.64293	230	50	39.27453
18	103	70.53442	89	113	52.94418	160	45	38.80235	231	131	33.12585
19	103	70.27133	90	113	53.16408	161	122	39.55858	232	131	33.5433
20	103	69.97591	91	114	53.21711	162	122	39.54527	233	131	33.55018
21	36	39.54118	92	114	53.67831	163	122	40.003255	234	131	33.98474
22	36	40.0981	93	114	53.91128	164	122	39.8779	235	131	36.8792
23	36	39.93218	94	114	54.22833	165	122	39.93522	236	132	30.88884
24	36	40.12999	95	114	53.98818	166	123	40.06279	237	132	30.87099
25	36	39.8402	96	115	53.10402	167	123	40.25023	238	132	31.32583
26	104	65.40551	97	115	52.86809	168	123	40.35887	239	132	30.82903
27	104	65.14808	98	115	52.83423	169	123	39.85371	240	132	31.05843
28	104	65.43729	99	115	53.28466	170	123	40.44098	241	133	30.44883
29	104	65.80562	100	115	53.60905	171	124	39.24843	242	133	30.75428
30	104	65.35538	101	42	32.67854	172	124	39.0882	243	133	30.79586
31	105	66.7257	102	42	32.83889	173	124	39.83021	244	133	31.20081
32	105	67.27786	103	42	32.4827	174	124	39.52884	245	133	31.162
33	105	67.07717	104	42	33.15945	175	124	39.83102	246	48	39.51342
34	105	67.47341	105	42	32.84108	176	40	35.80817	247	48	39.18823
35	105	67.38899	106	116	41.48799	177	40	35.77447	248	48	39.64759
36	106	63.82871	107	116	41.86876	178	40	35.23411	249	48	39.55811
37	106	64.34317	108	116	42.20533	179	40	35.85122	250	48	39.87181
38	106	64.822	109	116	41.9821	180	40	35.71376	251	998	0.00429
39	106	65.28719	110	116	42.44899	181	125	38.35815	252	998	-0.30439
40	106	65.02509	111	117	49.21764	182	125	38.53088	253	998	0.38045
41	41	35.80885	112	117	49.37881	183	125	38.09856	254	998	0.47052
42	41	36.58057	113	117	49.88229	184	125	38.75239	255	998	0.09337
43	41	36.34893	114	117	49.85855	185	125	38.59409			
44	41	36.88198	115	117	49.62275	186	126	38.48013			
45	41	36.22205	116	118	49.80809	187	126	38.61843			
46	107	66.28431	117	118	49.09934	188	126	38.53276			
47	107	66.43889	118	118	49.8015	189	126	39.33936			
48	107	66.62437	119	118	49.99126	190	126	39.15043			
49	107	66.98165	120	118	49.84129	191	127	36.97993			
50	107	66.40582	121	49	37.36262	192	127	37.32285			
51	108	64.73483	122	49	36.74314	193	127	37.48153			
52	108	64.64619	123	49	36.85221	194	127	37.81023			
53	108	64.57802	124	49	37.02195	195	127	37.55532			
54	108	64.99439	125	49	37.60116	196	51	31.28708			
55	108	65.02758	126	43	31.88927	197	51	31.40865			
56	109	65.26377	127	43	32.28064	198	51	31.63588			
57	109	65.32124	128	43	32.0104	199	51	31.91938			
58	109	65.3352	129	43	32.33823	200	51	31.88904			
59	109	65.07578	130	43	32.21884	201	37	37.22836			
60	109	65.02044	131	119	44.20918	202	37	37.1205			
61	39	38.98947	132	119	44.37271	203	37	37.21756			
62	39	38.85346	133	119	44.12404	204	37	37.04192			
63	39	39.11581	134	119	44.63036	205	37	37.72862			
64	39	39.17305	135	119	44.88643	206	128	35.40456			
65	39	39.24884	136	120	43.80605	207	128	35.59035			
66	110	64.00198	137	120	43.72208	208	128	35.74059			
67	110	65.00899	138	120	44.05705	209	128	36.20587			
68	110	65.13288	139	120	43.76207	210	128	36.14424			
69	110	64.70216	140	120	44.15276	211	129	33.91877			
70	110	64.6255	141	121	35.73044	212	129	34.26767			
71	111	59.79974	142	121	35.77391	213	129	34.68422			

Table 5.2: Summary Data from PR Study
Ben Meas in ppm

Rows	Sample	N	Mean(Ben Meas)	Elapsed Time
1	36	5	39.90773	.
2	37	5	37.267232	.
3	38	5	37.351266	.
4	39	5	39.092126	.
5	40	5	35.656346	.
6	41	5	36.362276	.
7	42	5	32.800128	.
8	43	5	32.147476	.
9	44	5	30.387764	.
10	45	5	38.369668	.
11	47	5	36.876754	.
12	48	5	39.551432	.
13	49	5	37.116216	.
14	50	5	38.892428	.
15	51	5	31.624002	.
16	53	5	32.078576	.
17	101	5	69.090446	0
18	102	5	67.552888	0.01666667
19	103	5	70.102696	0.03333333
20	104	5	65.449972	0.43333333
21	105	5	67.1901	0.45
22	106	5	64.657232	0.46666667
23	107	5	66.548608	0.81666667
24	108	5	64.794802	0.83333333
25	109	5	65.203286	0.85
26	110	5	64.694258	1.15
27	111	5	60.012654	1.16666667
28	112	5	62.904466	1.18333333
29	113	5	53.097968	2.48333333
30	114	5	53.806642	2.5
31	115	5	53.13561	2.51666667
32	116	5	42.000234	2.9
33	117	5	49.552028	2.91666667
34	118	5	49.708296	2.93333333
35	119	5	44.444544	4.15
36	120	5	43.900002	4.16666667
37	121	5	35.78029	4.18333333
38	122	5	39.783641	4.73333333
39	123	5	40.213512	4.75
40	124	5	39.46734	4.76666667
41	125	5	38.447974	5.13333333
42	126	5	38.823822	5.15
43	127	5	37.429972	5.16666667
44	128	5	35.8171188	6.15
45	129	5	34.496408	6.16666667
46	130	5	33.868734	6.18333333
47	131	5	33.7184	6.51666667
48	132	5	30.9542128	6.53333333
49	133	5	30.872276	6.55
50	998	5	0.124848	.
51	999	5	-0.273848	.

Table 5.3: Tests for Equal Variances for Repeated Measurements
Level = sample number

Tests that the Variances are Equal					
Level	Count	Std Dev	MeanAbsDif to Mean	MeanAbsDif to Median	
36	5	0.2366201	0.1736320	0.1871380	
37	5	0.2694537	0.1850352	0.1605120	
38	5	0.2065340	0.1342168	0.1351340	
39	5	0.1283955	0.1045288	0.1112400	
40	5	0.2668354	0.1881648	0.1888240	
41	5	0.3923570	0.2831992	0.3063060	
42	5	0.2489148	0.1756064	0.1682880	
43	5	0.1901793	0.1581128	0.1562000	
44	5	0.3846114	0.2881592	0.3027520	
45	5	0.3680264	0.3006224	0.3170300	
47	5	0.1685923	0.1348912	0.1249840	
48	5	0.2550485	0.1684856	0.1760880	
49	5	0.3583492	0.2925392	0.3076340	
50	5	0.2960972	0.2067296	0.2098220	
51	5	0.2878448	0.2289176	0.2719880	
53	5	0.2733939	0.1825176	0.1683180	
101	5	0.3199503	0.2184784	0.2024980	
102	5	0.4394901	0.3689216	0.3588740	
103	5	0.3611742	0.2650128	0.2738700	
104	5	0.2788651	0.1822592	0.1746500	
105	5	0.2957424	0.2281840	0.2330780	
106	5	0.5743693	0.4570336	0.4646980	
107	5	0.2696303	0.2035216	0.1873920	
108	5	0.2032646	0.1709464	0.1766800	
109	5	0.1454847	0.1241408	0.1235380	
110	5	0.4402532	0.3044144	0.3181660	
111	5	0.3065784	0.2626808	0.2668820	
112	5	0.4600604	0.3191784	0.2949240	
113	5	0.1335585	0.1153944	0.1081520	
114	5	0.3838860	0.2871456	0.2835980	
115	5	0.3183034	0.2409960	0.2668060	
116	5	0.3618431	0.2615408	0.2849820	
117	5	0.2538261	0.2030024	0.2007660	
118	5	0.3400090	0.2435824	0.1876600	
119	5	0.3134785	0.2510808	0.2694200	
120	5	0.1923923	0.1639224	0.1539280	
121	5	0.0735574	0.0588880	0.0663060	
122	5	0.2170701	0.1861808	0.1787930	
123	5	0.2029259	0.1642096	0.1787940	
124	5	0.2944308	0.2352200	0.2431940	
125	5	0.2421682	0.1772952	0.1713900	
126	5	0.3931618	0.3368584	0.3121140	
127	5	0.3069932	0.2228656	0.2273120	
128	5	0.3484201	0.2863426	0.3010848	
129	5	0.3940163	0.3233504	0.3004120	
130	5	0.2384693	0.1950328	0.1789700	
131	5	0.4819917	0.3743440	0.3420800	
132	5	0.2456329	0.1894538	0.1812012	
133	5	0.3125871	0.2472232	0.2401760	
998	5	0.3060914	0.2325096	0.2440300	
999	5	0.0828947	0.0638856	0.0707040	
Test	F Ratio	DF Num	DF Den	Prob>F	
O'Brien(.5)	0.9105	50	204	0.6442	
Brown-Forsythe	0.8688	50	204	0.7171	
Levene	1.2194	50	204	0.1712	
Bartlett	0.9517	50	.	0.5708	
Warning: Small sample sizes. Use Caution.					
Welch Anova testing Means Equal, allowing Std's Not Equal					
F Ratio	DF Num	DF Den	Prob>F		
25548.819	50	70.657	0.0000		

Table 5.4: Time Study of PR Data

Ln(Ben) By Elapsed Time

Fitting
— Linear Fit

Linear Fit

Summary of Fit

Rsquare	0.965824
Root Mean Square Error	0.051828
Mean of Response	3.874111
Observations (or Sum Wgts)	33

Analysis of Variance

Source	DF	Sum of Squares	Mean Square	F Ratio
Model	1	2.3532717	2.35327	876.0724
Error	31	0.0832710	0.00269	Prob>F
C Total	32	2.4365427		0.0000

Parameter Estimates

Term	Estimate	Std Error	t Ratio	Prob> t
Intercept	4.2502021	0.01558	272.73	0.0000
Elapsed Time	-0.119394	0.00403	-29.60	0.0000

Table 6.1: PHA Standards with 10% Methanol Calibration
Ben Meas in ppm

Row	Sample	Soln Matrix	Ben Std	Ben Meas
111	21	PHA	0	0.52471
112	21	PHA	0	0.65728
113	21	PHA	0	0.92435
114	21	PHA	0	0.90801
115	21	PHA	0	0.98803
116	30	PHA	30	30.09866
117	30	PHA	30	30.28685
118	30	PHA	30	30.46665
119	30	PHA	30	30.33013
120	30	PHA	30	30.47381
121	29	PHA	30	29.59862
122	29	PHA	30	29.59842
123	29	PHA	30	29.84984
124	29	PHA	30	29.89616
125	29	PHA	30	30.21155
126	34	PHA	50	50.13058
127	34	PHA	50	50.31184
128	34	PHA	50	50.32833
129	34	PHA	50	50.51561
130	34	PHA	50	50.68206
131	23	PHA	0	0.66721
132	23	PHA	0	0.90075
133	23	PHA	0	0.99622
134	23	PHA	0	0.9927
135	23	PHA	0	1.40823
136	33	PHA	50	49.3082
137	33	PHA	50	49.50616
138	33	PHA	50	49.66357
139	33	PHA	50	49.7508
140	33	PHA	50	49.63065
141	22	PHA	0	0.71252
142	22	PHA	0	0.6851
143	22	PHA	0	0.88332
144	22	PHA	0	1.24241
145	22	PHA	0	1.17701
146	28	PHA	30	30.08192
147	28	PHA	30	30.1141
148	28	PHA	30	30.27688
149	28	PHA	30	30.42872
150	28	PHA	30	30.42011
151	27	PHA	30	31.93026
152	27	PHA	30	31.96146
153	27	PHA	30	31.80143
154	27	PHA	30	32.00945
155	27	PHA	30	32.12294
156	31	PHA	50	48.5326
157	31	PHA	50	48.72649
158	31	PHA	50	48.89215
159	31	PHA	50	49.14115
160	31	PHA	50	49.19444
161	25	PHA	0	0.82569
162	25	PHA	0	0.99483
163	25	PHA	0	1.25912
164	25	PHA	0	1.08209
165	25	PHA	0	1.27337
166	24	PHA	0	0.46516
167	24	PHA	0	0.44898
168	24	PHA	0	0.6287
169	24	PHA	0	0.78075
170	24	PHA	0	0.79055
171	26	PHA	30	30.6975
172	26	PHA	30	30.83247
173	26	PHA	30	30.90329
174	26	PHA	30	30.96179
175	26	PHA	30	31.25516
176	32	PHA	50	50.54341
177	32	PHA	50	50.88185

Row	Sample	Soln Matrix	Ben Std	Ben Meas
178	32	PHA	50	50.73755
179	32	PHA	50	50.9556
180	32	PHA	50	50.93818

Table 6.2: Tests for Equal Variances for Repeated Measurements
Level = sample number

Ben Meas By Sample

Tests that the Variances are Equal

Level	Count	Std Dev	MeanAbsDif to Mean	MeanAbsDif to Median
21	5	0.1986767	0.1672648	0.1489460
22	5	0.2586307	0.2157104	0.2385200
23	5	0.2678972	0.1673624	0.1680020
24	5	0.1644588	0.1326064	0.1618420
25	5	0.1878573	0.1433800	0.1598460
26	5	0.2067737	0.1427464	0.1490960
27	5	0.1171488	0.0808696	0.0863800
28	5	0.1637769	0.1330688	0.1592080
29	5	0.2541876	0.1863984	0.1917980
30	5	0.1538889	0.1112080	0.1196460
31	5	0.2783719	0.2163432	0.2484320
32	5	0.1725602	0.1366768	0.1338220
33	5	0.1715514	0.1317568	0.1265860
34	5	0.2110805	0.1641208	0.1543480

Test	F Ratio	DF Num	DF Den	Prob>F
O'Brien[.5]	0.6407	13	56	0.8093
Brown-Forsythe	0.7279	13	56	0.7284
Levene	0.5938	13	56	0.8488
Bartlett	0.4295	13	.	0.9600

Warning: Small sample sizes. Use Caution.

Welch Anova testing Means Equal, allowing Std's Not Equal				
F Ratio	DF Num	DF Den	Prob>F	
39159.174	13	21.505	0.0000	

Table 6.3: Linear Calibration Fit for PHA Standards Data

Response: Mean(Ben Meas)

Summary of Fit

Rsquare	0.99902
Root Mean Square Error	0.673505
Mean of Response	25.53132
Observations (or Sum Wgts)	14

Lack of Fit

Source	DF	Sum of Squares	Mean Square	F Ratio
Lack of Fit	1	0.3972860	0.397286	0.8661
Pure Error	11	5.0460142	0.458729	Prob>F
Total Error	12	5.4433001		0.3720

Parameter Estimates

Term	Estimate	Std Error	t Ratio	Prob> t
Intercept	0.9774651	0.28584	3.42	0.0051
Mean(Ben Std)	0.9821543	0.00888	110.58	0.0000

Effect Test

Source	Nparm	DF	Sum of Squares	F Ratio	Prob>F
Mean(Ben Std)	1	1	5546.6052	12227.74	0.0000

Table 6.4: Quadratic Calibration Fit for PHA Standards Data

Polynomial Fit, degree=2

Summary of Fit

Rsquare	0.999091
Root Mean Square Error	0.677295
Mean of Response	25.53132
Observations (or Sum Wgts)	14

Analysis of Variance

Source	DF	Sum of Squares	Mean Square	F Ratio
Model	2	5547.0025	2773.50	6046.062
Error	11	5.0460	0.46	Prob>F
C Total	13	5552.0485		0.0000

Parameter Estimates

Term	Estimate	Std Error	t Ratio	Prob> t
Intercept	0.8886036	0.3029	2.93	0.0136
Mean(Ben Std)	1.0104048	0.03164	31.93	0.0000
Mean(Ben Std)^2	-0.000596	0.00064	-0.93	0.3720

Table 6.5: Predicted Benzene Measurements
Ben Meas in ppm

Rows	Sample	N	Mean(Ben Std)	Mean(Ben Meas)	Predicted Mean(Ben Meas)	Lower 95% Indiv Mean(Ben Meas)
15	21	5	0	0.800076	0.97746511	-0.6166693
16	22	5	0	0.940072	0.97746511	-0.6166693
17	23	5	0	0.993022	0.97746511	-0.6166693
18	24	5	0	0.622828	0.97746511	-0.6166693
19	25	5	0	1.08702	0.97746511	-0.6166693
20	26	5	30	30.930042	30.442093	28.920067
21	27	5	30	31.965108	30.442093	28.920067
22	28	5	30	30.264346	30.442093	28.920067
23	29	5	30	29.830518	30.442093	28.920067
25	30	5	30	30.33122	30.442093	28.920067
26	31	5	50	48.897366	50.0851783	48.4910439
27	32	5	50	50.811326	50.0851783	48.4910439
28	33	5	50	49.571876	50.0851783	48.4910439
29	34	5	50	50.393684	50.0851783	48.4910439

Rows	Upper 95% Indiv Mean(Ben Meas)	StdErr	Indiv Mean(Ben Meas)
15	2.57159954		0.73165169
16	2.57159954		0.73165169
17	2.57159954		0.73165169
18	2.57159954		0.73165169
19	2.57159954		0.73165169
20	31.9641191		0.69855648
21	31.9641191		0.69855648
22	31.9641191		0.69855648
23	31.9641191		0.69855648
25	31.9641191		0.69855648
26	51.6793128		0.73165169
27	51.6793128		0.73165169
28	51.6793128		0.73165169
29	51.6793128		0.73165169

Table 6.6: SAS Code for Inverse Prediction Problem

```
GOPTIONS DEVICE=PS2EGA;
.....;
* LINEAR CALIBRATION WITH CONSTANT :
* WEIGHTED OR UNWEIGHTED
* INPUT LO AND HI Y
* Developed by Steve Harris
.....;

DATA D0;
INPUT X Y;
W=1; * FOR VAR(E)= SIG2; TYPE='U';
Z=0.0; * W=1/X**Z; * FOR VAR(E)= SIG2*X**Z ; * TYPE='T1';
CARDS;
0 .800076
0 .940072
0 .993022
0 .6228286
0 1.08702
30 30.930042
30 31.965108
30 30.264346
30 29.830518
30 30.33122
50 48.89366
50 50.811326
50 49.571876
50 50.393684
;
PROC PRINT;
PROC REG DATA=D0 OUTEST=COEFF COVOUT;
MODEL Y=X/R COVB;
WEIGHT W;
PROC PRINT;
DATA D1; SET COEFF; N=1;
IF _TYPE_='PARMS' THEN DO;
MSE=_RMSE_**2;
B0=INTERCEP;
B1=X;OUTPUT;
END;
DATA D2; SET COEFF; N=1;
IF _TYPE_='COV' AND _NAME_='INTERCEP' THEN DO;
V_B0=INTERCEP;
C_01=X;OUTPUT;
END;
DATA D3; SET COEFF; N=1;
IF _TYPE_='COV' AND _NAME_='X' THEN DO;
```

Table 6.6 (con't): SAS Code for Inverse Prediction Problem

```

      V_B1=X;OUTPUT;
END;
DATA DZ;SET D0:N=1;IF _N_=1 THEN OUTPUT; KEEP Z TYPE N;
DATA D; MERGE D1 D2 D3 DZ;BY N;
      KEEP B0 B1 V_B0 V_B1 C_01 MSE Z TYPE;
PROC PRINT;
DATA D;SET D;
      LO_Y=4; HI_Y= 40;
      INC=(HI_Y-LO_Y)/18;

DO Y= LO_Y TO HI_Y BY INC;
      X=(Y-B0)/B1;
      C1= 1/B1**2;
      C2=(Y-B0)**2/B1**4;
      C3=(Y-B0)/B1**3;
      V_SYS=C1*V_B0 + C2*V_B1 + 2*C3*C_01;

      IF TYPE='T1' THEN V_RND=(MSE*(Y-B0)**Z)/B1**(Z+2);
      IF TYPE='U' THEN V_RND=MSE/B1**2;

      IF V_SYS LT 0 THEN V_SYS=.;
      IF V_RND LT 0 THEN V_RND=.;
      .....;
      STD_SYS=SQRT(V_SYS);
      PER_SYS=100*(STD_SYS/X);
      STD_RND=SQRT(V_RND);
      PER_RND=100*(STD_RND/X);
      STD =SQRT(V_RND+V_SYS);
      PER_STD=100*STD/X;
      IF X LT 0 THEN DO;
        PER_RND=.;
        PER_SYS=.;
        PER_STD=.;END;
      ..... t(.01,12) = 2.681 .....;
      XM=X-STD;
      XP=X+STD;
      U99XP=X+2.681*STD;
      OUTPUT;
END;
DROP C1 C2 C3;
PROC PRINT;
VAR Y X V_SYS V_RND STD PER_SYS PER_RND PER_STD U99XP;
      TITLE ***** HDS_FIN *****;
RUN;

```

Table 6.7: Inverse Prediction Problem

The following table is a subset of the output from the SAS program listed in Table 6.6.

Y = ALBM benzene value (ppm); X = True Benzene Value (ppm)

Y	X	Variance Systematic Error	Variance Random Error	Standard Deviation Total Error	Upper 99% Prediction Limit
4	3.08	0.073	0.47	0.74	5.05
6	5.11	0.066	0.47	0.73	7.08
8	7.15	0.060	0.47	0.73	9.10
10	9.19	0.054	0.47	0.72	11.13
12	11.22	0.049	0.47	0.72	13.16
14	13.26	0.045	0.47	0.72	15.19
16	15.30	0.41	0.47	0.72	17.21
18	17.33	0.038	0.47	0.71	19.25
20	19.37	0.036	0.47	0.71	21.28
22	21.40	0.035	0.47	0.71	23.31
24	23.44	0.034	0.47	0.71	25.35
26	25.48	0.034	0.47	0.71	27.38
28	27.51	0.034	0.47	0.71	29.42
30	29.55	0.035	0.47	0.71	31.46
32	31.59	0.037	0.47	0.71	33.50
34	33.62	0.040	0.47	0.71	35.54
36	35.6593	0.043	0.47	0.72	37.58
38	37.70	0.047	0.47	0.72	39.63
40	39.73	0.051	0.47	0.72	41.67

Appendix B: Inverse Calibration Problem

Let x be the true benzene level and y the ALBM value for a given sample. The linear calibration equation is given by

$$y = A + Bx + \epsilon \quad (1)$$

where A and B represent the unknown model parameters and ϵ represents the error term. The equation of the inverse calibration is given by

$$x = \frac{(y - A)}{B} - \frac{\epsilon}{B} \quad (2)$$

Estimates a and b of the parameters A and B , respectively, of equation (1) are determined from the data. Estimates of the variance of ϵ and the variances and covariance of the estimates a and b are also determined by fitting the data to equation (1). Let σ_{aa} , σ_{bb} , and σ_{ab} be the estimates of the variance of a , the variance of b , and the covariance of a and b , respectively.

There is a systematic error for estimating x using a and b in equation (2). The variance of this error may be estimated by a propagation of variance (reference 3) as follows:

$$s_{xx} \approx \left(\frac{\partial x}{\partial a}\right)^2 \sigma_{aa} + \left(\frac{\partial x}{\partial b}\right)^2 \sigma_{bb} + 2 \left(\frac{\partial x}{\partial a}\right) \left(\frac{\partial x}{\partial b}\right) \sigma_{ab} \quad (3)$$

where $\frac{\partial x}{\partial \cdot}$ represents the partial derivative of equation (2) evaluated at the mean level of the random variable.

Appendix B: Inverse Calibration Problem (con't)

These partials derivatives are:

$$\frac{\partial x}{\partial a} = -\frac{1}{b} \qquad \frac{\partial x}{\partial b} = \frac{-(y-a)}{b^2}$$

Substituting these partial derivatives into equation (3) gives

$$s_{xx} \approx \left(-\frac{1}{b}\right)^2 \sigma_{aa} + \left(\frac{-(y-a)}{b^2}\right)^2 \sigma_{bb} + 2 \left(-\frac{1}{b}\right) \left(\frac{-(y-a)}{b^2}\right) \sigma_{ab} \quad (4)$$

The variance of the random error for x is given by

$$r_{xx} \approx \frac{\sigma_{ee}}{b^2} \quad (5)$$

where σ_{ee} is the variance of the error term e estimated from fitting the data to equation (1).

The variance of the total error of x is given by the sum of the systematic error variance in equation (4) and the random error variance in equation (5):

$$\sigma_{xx} \approx s_{xx} + r_{xx}$$

A 99% upper prediction limit for x for a given value of y is

$$\frac{(y - a)}{b} + t_{.01, n-2} \sqrt{\sigma_{xx}}$$

where $t_{.01, n-2}$ is the critical value of Student's t with n-2 degrees of freedom and n equals the number of standards analyzed during the fit of the calibration curve for the ALBM.

END

DATE
FILMED

5/04/93