12/6-1-92950

PNL-8099
UC-940

Development and Validation of a Preparation and Cleanup Method for Hydrocarbon Containing Samples for the Analysis of Volatile Organic Compounds

E. W. Hoppe
R. B. Lucke

G. A. Ross
J. A. Campbell

April 1992

Prepared for the U.S. Department Energy
under Contract DE-AC06-76RLO 1830

Pacific Northwest Laboratory
Operated for the U.S. Department of Energy
by Battelle Memorial Institute

PNL-8099

DISCLAIMER

This report was prepared as an account of work sponsored by an agency of the United States Government. Neither the United States Government nor any agency thereof, nor Battelle Memorial Institute, nor any of their employees, makes **any warranty, expressed or implied, or assumes any legal liability or responsibility for the accuracy, completeness, or usefulness of any information, apparatus, product, or process disclosed, or represents that its use would not infringe privately owned rights.** Reference herein to any specific commercial product, process, or service by trade name, trademark, manufacturer, or otherwise does not necessarily constitute or imply its endorsement, recommendation, or favoring by the United States Government or any agency thereof, or Battelle Memorial Institute. The views and opinions of authors expressed herein do not necessarily state or reflect those of the United States Government or any agency thereof.

PACIFIC NORTHWEST LABORATORY
operated by
BATTELLE MEMORIAL INSTITUTE
for the
UNITED STATES DEPARTMENT OF ENERGY
under Contract DE-AC06-76RLO 1830

Printed in the United States of America

Available to DOE and DOE contractors from the
Office of Scientific and Technical Information, P.O. Box 62, Oak Ridge, TN 37831;
prices available from (615) 576-8401. FTS 626-8401.

Available to the public from the National Technical Information Service,
U.S. Department of Commerce, 5285 Port Royal Rd., Springfield, VA 22161.

PNL--8099

DE92 014072

DEVELOPMENT AND VALIDATION OF A PREPARATION AND CLEANUP
METHOD FOR HYDROCARBON CONTAINING SAMPLES FOR THE
ANALYSIS OF VOLATILE ORGANIC COMPOUNDS

E. W. Hoppe
R. B. Lucke
G. A. Ross
J. A. Campbell

April 1992

Prepared for
the U.S. Department of Energy
under Contract DE-AC06-76RLO 1830

Pacific Northwest Laboratory
Richland, Washington 99352

MASTER

EXECUTIVE SUMMARY

Samples obtained from the Hanford single shell tanks are contaminated with normal paraffin hydrocarbon (NPH) as hydrostatic fluid from the sampling process or can be native to the tank waste. The contamination is usually high enough that a dilution of up to several orders of magnitude may be required before the sample can be analyzed by the conventional purge and trap concentration followed by gas chromatography/mass spectrometry methodology. This can prevent detection and measurement of organic constituents that are present at lower concentration levels. To eliminate or minimize the problem, a sample cleanup method has been developed and validated and is presented in this document.

An aliquot (nominally 1 gram) of the sample is extracted with 5 mL of methanol/water (90:10) containing the surrogate compounds identified in the U.S. Environmental Protection Agency (EPA) Contract Laboratory Program (CLP) Statement of Work (SOW) 2/88. A 2-mL aliquot is syringe-pumped through a C₁₈ Sep-Pak^R cartridge and followed by 0.5-mL rinses until a final volume of 5 mL is obtained. One hundred μ L of the eluent is transferred to a purge vessel containing 4.9 mL of water. Internal standards are added to the purge vessel and purge and trap extracted per the EPA method or the identical procedure PNL-ALO-335, "GC/MS Analysis of Volatile Organic Compounds." Inherent with this method is a 625-fold dilution.

The effectiveness of this cleanup method was demonstrated for all of the volatile target list compounds in the CLP SOW. Both blanks and single shell tank samples spiked with all of the target compounds and NPH were used in this development and validation study. The NPH removal for samples and blank spikes is better than 99%. The recoveries were excellent for most of the target compounds in blank spikes and satisfactory for all compounds, although lower for the more volatile compounds. The blank spike residues were shown to store well for up to 12 days. The single shell tank samples were observed to have a substantial matrix effect on recovery of the gases, ketones, and saturated compounds that are halogenated. Advisory quality control limits from

tank waste data were established for the surrogate and matrix spike compounds. As more data is obtained using this cleanup method, required QC limits will be put in place, and a better definition of cleanup effectiveness can be established.

CONTENTS

EXECUTIVE SUMMARY	iii
INTRODUCTION	1
INITIAL DEVELOPMENT	3
FINAL DEVELOPMENT AND VALIDATION STUDY	4
EXPERIMENTAL	6
PROCEDURES	6
Initial Development	6
Final Development and Validation Study	7
ANALYSIS	8
Initial Development Instrumentation	8
Final Development and Validation Instrumentation	8
Calibration	9
RESULTS	11
INITIAL DEVELOPMENT	11
FINAL DEVELOPMENT AND VALIDATION STUDY	11
DISCUSSION	27
INITIAL DEVELOPMENT	27
FINAL DEVELOPMENT AND VALIDATION STUDY	28
CONCLUSIONS	33
REFERENCES	35
APPENDIX - PNL-ALO-121, Cleanup of Hydrocarbon Containing Samples for the Analysis of Volatile Organic Compounds	36

FIGURES

1	Closed System Used for the Volatile Cleanup Procedure	4
2	Chromatogram of Acetone, Xylenes and NPH Eluted from a C18 Cartridge Column	12
3	Amount of NPH Measured in Various Methanol/Water Solutions at 21°C	15
4	Percent Recovery Results Measured After Column Cleanup for Various Methanol/Water Solutions	15
5	Chromatographic Effects of the Presence of 100, 200, and 300 μ L of Methanol on Bromomethane	16
6	Chromatographic Effects of the Presence of 10, 100, 200, and 300 μ L of Methanol on Chloroform	17

TABLES

1	Percent Recovery Results of a Partial Target List Standard After Cleanup Using the Initial Closed System	13
2	Percent Recovery Results Obtained from Cartridge Cleanup with Varied Eluting Solvent Composition	14
3	Percent Recovery Results Obtained from a Blank Spike Extraction, Followed by the Cleanup Procedures Using 90% Methanol, 200 μ L for Analysis	18
4	Percent Recovery Results from Same Day and Day 12 GC/MS Analysis of a Blank Spike, After Cleanup, Stored at 4°C	19
5	Percent Recovery Results from Same Day and Day 12 GC/MS Analysis of Spiked SST U110 Sample, After Cleanup, Stored at 4°C	20
6	Comparison of Results Obtained from a Blank Using 200 μ L of Methanol to Blanks Using 100 μ L	21
7	Percent Recovery Results, RPD, and Matrix Effect for SST U110 Sample	23
8	Percent Recovery Results, RPD, and Matrix Effect for B110 Sample	24
9	Percent Recovery Results, RPD, and Matrix Effect for B110 Duplicate	25
10	Calculated Advisory Control Limits for SST Samples	26

INTRODUCTION

Analysis of samples for organic compounds as part of the single shell tank (SST) chemical characterization has been compromised by the presence of large amounts of normal paraffin hydrocarbons (NPH). NPH is made up of straight chain fully saturated hydrocarbons primarily in the C12 to C14 range and used as a hydrostatic fluid to keep the bore hole open as core sample segments are removed from the waste tanks. The use of NPH as a hydrostatic fluid leads to contamination of the core segments. In addition, many tank wastes contain substantial quantities of NPH which are native to the sample itself. The presence of the high NPH concentration in the samples, whether introduced by sampling or native to the sample, can lead to significant problems in the analyses performed by purge and trap sample introduction followed by gas chromatography/mass spectrometry (GC/MS). When performing GC/MS analysis, dilutions often up to several orders of magnitude were necessary to prevent overloading the analytical system. Thus, detection of many less prominent sample constituents was not possible.

Attempts to resolve the difficulties associated with high levels of NPH must be twofold. One approach is to attempt to find another suitable hydrostatic fluid that does not impact sample analyses or to identify a new sampling process that does not require a hydrostatic fluid. This would eliminate NPH as a contaminant and would allow determination of NPH native to the sample. The second approach is to develop a laboratory separations cleanup process that allows removal of native and contaminant NPH from SST samples without compromising the volatile organic analyses. This document deals with minimizing the NPH problems associated with the volatile analysis utilizing a sample extraction and cleanup process.

The approach for separation of NPH from the analytes of interest in the volatile fraction is based on selective retention of certain components within one phase while the other component is transferred to a second phase. Ideally, one would obtain a complete separation of NPH from the analytes of interest. However, since the separations are based on thermodynamic and kinetic processes that are dependent upon the chemical characteristics of the various

compounds and surfaces of the substrates, one must balance the complete removal of NPH against the losses of the analytes of interest. Thus, as the chemical characteristics of the analytes of interest approach those of NPH, primarily polarity in this case, the separation becomes more difficult. The approach then must be to remove adequate amounts of NPH and evaluate the recovery of the analytes of interest.

The cleanup method described in this document is employed following extraction of the waste or soil using a similar procedure to the medium level process provided by the U.S. Environmental Protection Agency (EPA). After the extract cleanup, purge and trap concentration is performed, followed by GC/MS analysis. The GC/MS procedures used at the Analytical Chemistry Laboratory at Pacific Northwest Laboratory (PNL)^(a) are derived directly from EPA Contract Laboratory Program (CLP) Statement of Work (SOW) 2/88, such as PNL method PNL-ALO-335. In the PNL extraction method, an aliquot of the sample, usually 1 gram, is extracted using 5 mL of methanol/water in which the surrogate compounds have been added. A 2-mL aliquot of the extract is pumped through a C₁₈ (octadecyl silane) cartridge column and followed by 0.5-mL rinses with methanol/ water (90:10) until a final volume of 5 mL is obtained. One hundred μ L of the eluent is transferred to a purge vessel containing 4.9 mL of water. Internal standards are added to the purge vessel and purge and trap extracted per the EPA method or the identical procedure PNL-ALO-335, "GC/MS Analysis of Volatile Organic Compounds." Inherent with the extraction and subsequent cleanup procedure is a 625-fold dilution.

The study described in this document consists of an initial development investigation performed by the Chemical Sciences Department followed by further development and validation experiments performed by the Analytical Chemistry Laboratory at PNL. The initial development involved choosing an appropriate cartridge column, designing the three-syringe closed-system cleanup process, and qualitatively assessing the removal of NPH and the recovery of a limited set of volatile compounds. The subsequent development and validation portion of the work consisted of performing the extraction and

(a) Operated for the U.S. Department of Energy by Battelle Memorial Institute under Contract DE-AC06-76RLO 1830.

testing the initial proposed cleanup procedure, based on the initial development study, and GC/MS analyses on blank-spikes with NPH. For these studies, the spike contained all the CLP SOW volatile organic target compounds and the NPH. Following further development and configuration, blank spikes also containing NPH and actual SST samples were processed to validate the method.

INITIAL DEVELOPMENT

The initial development did not attempt to assess an extraction procedure. It was envisioned that use of the EPA medium-level extraction would be employed and that recovery of the analytes would be tested during the validation portion of this study. The initial development concentrated on producing an acceptable cleanup method. The cleanup method would then require further refinement during the final development and validation study, particularly to assess its usability on actual SST waste samples.

The NPH cleanup method for the volatile organic compounds is based on the selective retention of NPH on a C_{18} column while the more polar volatile analytes are eluted with a methanol/water mixture. To initially assess the viability of C_{18} as the sorbent or stationary phase of choice, a preliminary study monitored elution of selected analytes and NPH from a C_{18} cartridge column. This study determined if the analytes could be separated from NPH obtaining adequately differing retention times on the column.

Following the initial assessment of the C_{18} column, retention of the NPH on the column was evaluated. NPH was measured by GC/MS following column elution with methanol/water (80:20) and then after elution with pure methanol.

To preclude the potential loss of the volatile analytes, a closed system, shown in Figure 1, was conceived that would minimize the exposure to air the solvent containing analytes. It is believed the apparatus would reduce the potential analyte loss and also minimize the potential adsorption of unwanted species from the surrounding ambient air during the cleanup process. The closed system apparatus was employed to assess the recovery of a partial list of the EPA volatile target compounds. This was evaluated by eluting a standard mix through the closed system using a C_{18} column and eluting with methanol/water (80:20).

FIGURE 1. Closed System Used for the Volatile Cleanup Procedure

FINAL DEVELOPMENT AND VALIDATION STUDY

Low recoveries of the non-polar (low water solubility) compounds was observed during the initial development analysis. Further development was required to improve these recoveries. Additionally, the initial studies did not include the complete target compound list, and assessment of all compounds is necessary to obtain an applicable completed method. Based on the initial development, several primary goals for the final development were established:

- Determine the methanol/water composition required to obtain acceptable full target list analyte recovery through the cleanup column.
- Assess the solubility of NPH in various methanol/water compositions to determine the quantity of NPH that will be potentially loaded onto the column.
- Assess the capability of the cleanup column to remove NPH in various methanol/water solvent systems.
- Determine the maximum amount of methanol containing the cleaned extractant that can be introduced into the purge and trap and GC/MS systems while maintaining satisfactory chromatographic performance.

- Determine the optimum set of conditions, based on the data obtained from this set objectives, which will complete the development of a cleanup method allowing for the maximum overall recovery of analytes and removal of NPH.
- Validate the finalized extraction and cleanup method by measuring recoveries of the full volatile target compound list by use of blank spikes and actual SST spiked samples.

These parameters for optimizing the cleanup procedure were investigated as follows:

To assess the analyte recovery, blank spikes containing the full volatile target compound list were loaded onto the C₁₈ cartridge cleanup column and eluted using 80, 85, and 90 percent methanol in water.

To determine the potential loading of NPH on to the cleanup column, a solubility study was performed by saturating various methanol/water compositions and measuring the NPH present.

To assess the ability of the cartridge column to remove NPH in various methanol/water solvent compositions, blanks were spiked with NPH and eluted through a cartridge column. The elution solvent composition was varied from 80 to 100 percent methanol at five different levels. The amount of NPH present in the column eluent was then measured.

It was also necessary to determine the maximum amount of methanol, and therefore the amount of cleaned extract analyzed, that can be introduced into the purge and trap and GC/MS systems. When substantial quantities of methanol are introduced into the analytical systems, the chromatography suffers. The effect is highly compound specific and the chromatographic difficulties usually manifest themselves as severe peak broadening, splitting, or retention time shifts due to GC column overloading. Various quantities of methanol were added to the purge vessel along with the target analyte standard to assess the effect on chromatographic performance.

Evaluation of the above experiments yielded a final set of parameters for use in the column cleanup procedure. These empirically derived parameters were verified experimentally by analyzing spiked blanks and measuring recoveries. Validation of the procedure for use on actual samples was performed by analyzing blank spikes and spiked SST samples.

the cartridge column and into the "volatile fraction" receiving syringe while the "eluent solvent" syringe valve is closed. The "sample" syringe valve was then closed, the "eluent solvent" syringe valve opened, and 4.0 mL of methanol/water (80:20) passed from the "eluent solvent" syringe into the receiving "volatile fraction" syringe. The valve for the "volatile fraction" syringe is closed and the contents are ready for analysis.

Final Development and Validation Study

To determine the methanol/water composition required to obtain acceptable analyte recovery, 5.0-mL methanol/water (80:20) blank spikes containing the full volatile target compound list were prepared using Supelco (Bellefonte, Pennsylvania) standards. One mL of the solution was loaded onto the column and eluted using 4.0 mL of methanol/water (80:20) and analyzed immediately afterward by GC/MS. The experiment was repeated using 85 and 90 percent methanol as the column eluent.

To determine the potential loading of NPH onto the column, an NPH solubility study was performed by saturating various methanol/water compositions and measuring the NPH present. Five-mL methanol/water solutions, ranging from 0 to 100 percent methanol, were saturated with NPH at 21°C. The solutions were saturated by the addition of up to 500 μ L of NPH, then vortexed for 1 minute and allowed to sit for 30 minutes. The lower 2.0 mL of the solutions were transferred to gas tight syringes through a 0.45 μ m nylon filter. The filtrate was immediately analyzed using a GC/FID to determine the NPH concentration.

To assess the retention of NPH on the C₁₈ column using various methanol/water solvent systems, 5.0-mL methanol/water (80:20) blanks were spiked with NPH to saturation. The column was then eluted with 4.0-mL solvent mixture and the eluent collected. The elution solvent composition was varied between 80 and 100 percent methanol at five different levels. The amount of NPH present in the column eluent was then immediately measured by GC/MS.

To determine the maximum amount of methanol that can be introduced into the purge and trap and GC/MS systems while maintaining satisfactory chromatographic performance, various quantities of methanol were added to the purge

vessel along with the target analyte standard to assess the effect on chromatographic performance. Peak width, retention time stability, and noise on the peak were assessed at 10, 20, 100, 200, and 300 μ L of methanol. The data system's ability to adequately integrate the peak correctly was also used as criteria for acceptability.

Evaluation of the above experiments provided a final set of parameters for use in the column cleanup procedure. These empirically derived parameters were verified experimentally by analyzing spiked blanks and measuring recovery. Validation of the procedure for use on actual samples was performed by analyzing blank spikes and spiked SST samples. The detailed procedure is given in the Appendix, as method PNL-ALO-121.

ANALYSIS

The instrumentation and its calibration, the reagents, and special apparatus used in these studies are described in the following discussion.

Initial Development Instrumentation

The gas chromatograph flame ionization detector (GC/FID) used for NPH analysis during this phase of development consisted of a Hewlett Packard 5880A gas chromatograph equipped with a DB5 (J&W Scientific, Folsom, California) 30 m by 0.25 mm I.D. capillary column 0.25 μ m film thickness.

Volatile analyte analysis was performed on an HP5880 GC and an HP5987 mass spectrometer with an RTE-6 data system. Sample introduction was performed with a Tekmar LSC-2 purge and trap instrument.

Final Development and Validation Instrumentation

The GC/FID described previously was used for a portion of this development phase also. The volatile analysis was performed using a Tekmar LSC-2 purge and trap as sample introduction into a Hewlett Packard GC/MS system consisting of an HP-5890 GC (using a DB624, 30-meter by 0.53 mm I.D., 3.0 μ m film thickness, capillary column), an HP-5790 MS, the HP-1000 series A system controller/data acquisition/data processor system, and associated peripherals.

Radioactive tank waste analysis was performed using a Tekmar LSC-3 purge and trap. Following the procedure, the trap was removed from the LSC-3 and replaced on the LSC-2 for desorption into the GC/MS for analysis.

Calibration

Initial Development Instrumentation

The GC/FID used for NPH analysis was calibrated using a single-level calibration. The mass spectrometer mass calibration and tune parameters were adjusted using PFTBA as a tuning compound. The linearity of the instrument was verified from an initial four-point calibration. A single-level continuing calibration was performed for a partial CLP volatile target list of compounds each day of operation.

Final Development and Validation Instrumentation

The GC/FID used for NPH analysis was calibrated using a single-level calibration. The mass spectrometer mass calibration and tune parameters were confirmed with a bromofluorobenzene analysis according to EPA CLP SOW 2/88 specifications. The linearity of the instrument was verified from an initial 5-point calibration. A single-level continuing calibration was performed for the CLP volatile target list compounds each day of operation.

Reagents

The methanol used in this work was "Purge and Trap" quality or equivalent. The target compounds for the spikes and instrument calibration were prepared from Supelco (Supelco, Inc., Bellefonte, Pennsylvania) Analytical Reference Standards for EPA Contract Laboratory Program Methods mixes. The NPH was obtained from Westinghouse Hanford Company. The area percentages, obtained from analysis by GC/MS, for the NPH constituents were:

dodecane - 18%

tridecane - 48%

tetradecane - 33%

pentadecane - 1+%

Additional information on the reagents is provided in the procedure in the Appendix.

Apparatus and Materials

The closed cleanup system apparatus was modified somewhat during the final development. The initial study used 10-mL gas-tight syringes; these were replaced with 5-mL syringes of the same manufacturer. The three-way valve, originally nylon, was replaced with a stainless steel valve, and finally replaced with a Hamilton HV 1-1 two-way valve for use on radioactive samples. These modifications are described further in the procedure in the Appendix.

RESULTS

The results of both the initial development and the final development and validation studies are described in the following discussion.

INITIAL DEVELOPMENT

To establish that the initial concept of using C_{18} to remove NPH was viable, the effluent from a cartridge column was monitored using ultraviolet (UV) and refractive index (RI) detection in an experiment described previously. As seen in Figure 2, the analytes, acetone and xylenes, are well resolved from the NPH.

The verification analysis by GC/MS showed that after eluting the column with 5.0 mL of methanol/water (80:20) no measurable amount of NPH was present. The same column, after elution with pure methanol, recovered almost all of the NPH.

Table 1 shows the results obtained after eluting a standard, containing the partial CLP target list, using the closed cleanup system and methanol/water (80:20) elution solvent.

FINAL DEVELOPMENT AND VALIDATION STUDY

Table 2 shows the analyte recoveries obtained from GC/MS when the composition of the eluting solvent was varied from 80 to 90 percent methanol. All spike recoveries are obtained from blanks or samples containing NPH unless noted.

Figure 3 shows the results of the NPH in methanol/water solubility study. The graph presents the percent of methanol, versus the GC/FID measured NPH concentration.

Figure 4 shows the results from GC/FID analysis obtained when NPH was eluted from C_{18} cartridge columns and the eluting solvent composition was varied between 80 and 100 percent methanol.

FIGURE 2. Chromatogram of Acetone, Xylenes and NPH Eluted from a C18 Cartridge Column

Figure 5 illustrates the effect that the amount of methanol in the purge vessel has on chromatography of bromomethane, a gas volatile compound, on a GC/MS system.

Figure 6 illustrates the effect that the amount of methanol in the purge vessel has on chromatography of chloroform on a GC/MS system.

Table 3 shows the analysis of a blank and the recoveries obtained from a blank spike before and after the cleanup procedure.

Table 4 shows recoveries from a blank spike analyzed the same day and 12 days after the cleanup preparation.

Table 5 shows recoveries from U110 tank sample spike analyzed the same day and 12 days after the cleanup preparation.

TABLE 1. Percent Recovery Results of a Partial Target List Standard After Cleanup Using the Initial Closed System

No.	Compound Name	80 Percent Methanol	
		STD	STD w/ NPH
1)	*Bromochloromethane	100.0	100.0
2)	Chloromethane	NA (a)	NA
3)	Bromomethane	332.8	97.2
4)	Vinyl Chloride	NA	NA
5)	Chloroethane	62.4	78.4
6)	Methylene Chloride	82.0	104.2
7)	Acetone	133.8	177.4
8)	Carbon Disulfide	10.2	17.2
9)	1,1-Dichloroethene	19.0	35.8
10)	1,1-Dichloroethane	34.0	64.4
11)	trans-1,2-Dichloroethene	24.4	48.4
12)	cis-1,2-Dichloroethene	NA	NA
13)	Chloroform	39.2	72.4
14)	#1,2-Dichloroethane-d4	NA	NA
15)	1,2-Dichloroethane	48.8	86.6
16)	2-Butanone	218.2	298.0
17)	1,1,1-Trichloroethane	12.4	22.8
18)	Carbon Tetrachloride	7.0	11.4
19)	Vinyl Acetate	103.4	132.6
20)	Bromodichloromethane	38.0	70.0
21)	*1,4-Difluorobenzene	100.0	100.0
22)	1,2-Dichloropropane	27.2	52.6
23)	cis-1,3-Dichloropropene	32.6	61.4
24)	Trichloroethene	12.0	20.8
25)	Dibromochloromethane	33.6	63.8
26)	1,1,2-Trichloroethane	39.8	70.8
27)	Benzene	26.0	49.2
28)	trans-1,3-Dichloropropene	NA	NA
29)	Bromoform	NA	NA
30)	*Chlorobenzene-d5	100.0	100.0
31)	4-Methyl-2-Pentanone	51.0	86.2
32)	2-Hexanone	28.8	51.8
33)	Tetrachloroethene	18.2	20.5
34)	1,1,2,2-Tetrachloroethane	26.2	54.6
35)	Toluene	10.0	17.6
36)	#Toluene-d8	NA	NA
37)	Chlorobenzene	10.4	19.8
38)	Ethylbenzene	9.2	5.2
39)	Styrene	8.2	7.6
40)	p-Xylene	NA	NA
41)	o-Xylene	2.0	0.0
42)	#Bromofluorobenzene	NA	NA

* Compound is an Internal Standard

Compound is a Surrogate Spike

(a) NA = Not analyzed

TABLE 2. Percent Recovery Results Obtained from Cartridge Cleanup with Varied Eluting Solvent Composition

No.	Compound Name	Percent Methanol		
		80	85	90
1)	*Bromochloromethane	100.0	100.0	100.0
2)	Chloromethane	NA (a)	119.9	70.3
3)	Bromomethane	97.2	111.1	78.7
4)	Vinyl Chloride	NA	81.6	59.2
5)	Chloroethane	78.4	89.2	73.6
6)	Methylene Chloride	104.2	104.7	96.6
7)	Acetone	177.4	180.3	133.0
8)	Carbon Disulfide	17.2	30.1	47.4
9)	1,1-Dichloroethene	35.8	47.8	60.5
10)	1,1-Dichloroethane	64.4	65.7	99.8
11)	trans-1,2-Dichloroethene	48.4	57.7	97.7
12)	cis-1,2-Dichloroethene	NA	91.6	104.6
13)	Chloroform	72.4	75.0	108.1
14)	#1,2-Dichloroethane-d4	NA	92.9	108.4
15)	1,2-Dichloroethane	86.6	106.0	109.0
16)	2-Butanone	298.0	301.5	107.6
17)	1,1,1-Trichloroethane	22.8	44.5	99.5
18)	Carbon Tetrachloride	11.4	32.5	96.4
19)	Vinyl Acetate	132.6	30.5	105.5
20)	Bromodichloromethane	70.0	87.1	108.3
21)	*1,4-Difluorobenzene	100.0	100.0	100.0
22)	1,2-Dichloropropane	52.6	74.7	99.3
23)	cis-1,3-Dichloropropene	61.4	85.9	104.9
24)	Trichloroethene	20.8	43.6	96.4
25)	Dibromochloromethane	63.8	80.3	94.9
26)	1,1,2-Trichloroethane	70.8	89.2	104.2
27)	Benzene	49.2	66.0	90.8
28)	trans-1,3-Dichloropropene	NA	75.4	92.3
29)	Bromoform	NA	69.7	91.9
30)	*Chlorobenzene-d5	100.0	100.0	100.0
31)	4-Methyl-2-Pentanone	86.2	108.0	104.5
32)	2-Hexanone	51.8	105.8	110.1
33)	Tetrachloroethene	20.5	15.5	113.8
34)	1,1,2,2-Tetrachloroethane	54.6	73.0	110.1
35)	Toluene	17.6	41.1	113.7
36)	#Toluene-d8	NA	105.9	114.2
37)	Chlorobenzene	19.8	40.8	115.5
38)	Ethylbenzene	5.2	24.2	114.5
39)	Styrene	7.6	36.2	112.6
40)	p-Xylene	NA	42.1	115.0
41)	o-Xylene	0.0	22.4	108.9
42)	#Bromofluorobenzene	NA	106.6	111.6

* Compound is an Internal Standard

Compound is a Surrogate Spike

(a) NA = Not analyzed

FIGURE 3. Amount of NPH Measured in Various Methanol/Water Solutions at 21°C

FIGURE 4. Percent Recovery Results Measured After Column Cleanup for Various Methanol/Water Solutions

FIGURE 5. Chromatographic Effects of the Presence of 100, 200, and 300 μ L of Methanol on Bromomethane

FIGURE 6. Chromatographic Effects of the Presence of 10, 100, 200, and 300 μ L of Methanol on Chloroform

TABLE 3. Percent Recovery Results Obtained from a Blank Spike Extraction, Followed by the Cleanup Procedures Using 90% Methanol, 200 μ L for Analysis

No.	Compound Name	High MeOH Blank	Extracted Blk Spike	Cleaned Blk Spike	Procedure Effect
1)	*Bromochloromethane	100.0	100.0	100.0	0.0
2)	Chloromethane	0.0	91.0	70.3	-20.7
3)	Bromomethane	0.0	98.0	78.7	-19.3
4)	Vinyl Chloride	0.0	91.9	59.2	-32.7
5)	Chloroethane	0.0	96.5	73.6	-22.9
6)	Methylene Chloride	0.0	90.3	96.6	6.4
7)	Acetone	29.9	72.1	133.0	60.9
8)	Carbon Disulfide	0.0	85.6	47.4	-38.2
9)	1,1-Dichloroethene	0.0	96.3	60.5	-35.9
10)	1,1-Dichloroethane	0.0	102.8	99.8	-3.0
11)	trans-1,2-Dichloroethene	0.0	104.7	97.7	-7.0
12)	cis-1,2-Dichloroethene	0.0	105.6	104.6	-1.0
13)	Chloroform	0.0	106.8	108.1	1.3
14)	#1,2-Dichloroethane-d4	104.5	107.8	108.4	0.6
15)	1,2-Dichloroethane	0.5	106.1	109.0	2.9
16)	2-Butanone	35.8	71.2	107.6	36.4
17)	1,1,1-Trichloroethane	0.0	102.5	99.5	-3.1
18)	Carbon Tetrachloride	0.0	101.3	96.4	-4.8
19)	Vinyl Acetate	0.0	101.7	105.5	3.8
20)	Bromodichloromethane	0.0	103.4	108.3	4.9
21)	*1,4-Difluorobenzene	100.0	100.0	100.0	0.0
22)	1,2-Dichloropropane	0.0	101.5	99.3	-2.1
23)	cis-1,3-Dichloropropene	0.0	102.2	104.9	2.7
24)	Trichloroethene	0.0	100.2	96.4	-3.9
25)	Dibromochloromethane	0.0	95.1	94.9	-0.2
26)	1,1,2-Trichloroethane	0.0	103.0	104.2	1.2
27)	Benzene	0.0	99.1	90.8	-8.3
28)	trans-1,3-Dichloropropene	0.0	90.8	92.3	1.5
29)	Bromoform	0.9	85.0	91.9	7.0
30)	*Chlorobenzene-d5	100.0	100.0	100.0	0.0
31)	4-Methyl-2-Pentanone	9.2	93.8	104.5	10.7
32)	2-Hexanone	11.6	92.0	110.1	18.1
33)	Tetrachloroethene	0.0	111.8	113.8	2.0
34)	1,1,2,2-Tetrachloroethane	3.4	102.3	110.1	7.8
35)	Toluene	0.0	110.2	113.7	3.5
36)	#Toluene-d8	107.0	111.6	114.2	2.5
37)	Chlorobenzene	0.0	110.8	115.5	4.7
38)	Ethylbenzene	0.0	109.1	114.5	5.5
39)	Styrene	0.0	111.5	112.6	1.1
40)	p-Xylene	0.0	109.8	115.0	5.2
41)	o-Xylene	0.0	105.4	108.9	3.6
42)	#Bromofluorobenzene	103.9	105.4	111.6	6.2

* Compound is an Internal Standard

Compound is a Surrogate Spike

TABLE 4. Percent Recovery Results from Same Day and Day 12 GC/MS
Analysis of a Blank Spike, After Cleanup, Stored at 4°C

No.	Compound Name	Blank Spike		Difference	Percent Difference
		day 1	day 12		
1)	*Bromochloromethane	100.0	100.0	0.0	0.0
2)	Chloromethane	81.9	77.6	-4.3	-5.2
3)	Bromomethane	83.3	81.6	-1.7	-2.0
4)	Vinyl Chloride	71.2	63.7	-7.5	-10.5
5)	Chloroethane	75.0	70.4	-4.6	-6.2
6)	Methylene Chloride	96.4	74.3	-22.1	-23.0
7)	Acetone	120.1	105.8	-14.3	-11.9
8)	Carbon Disulfide	56.7	50.8	-5.9	-10.4
9)	1,1-Dichloroethene	65.9	63.0	-2.9	-4.4
10)	1,1-Dichloroethane	90.2	77.8	-12.5	-13.8
11)	trans-1,2-Dichloroethene	89.8	86.0	-3.8	-4.3
12)	cis-1,2-Dichloroethene	93.2	88.5	-4.6	-5.0
13)	Chloroform	95.5	92.4	-3.1	-3.3
14)	#1,2-Dichloroethane-d4	93.7	97.3	3.6	3.9
15)	1,2-Dichloroethane	92.8	90.4	-2.4	-2.6
16)	2-Butanone	111.7	110.1	-1.7	-1.5
17)	1,1,1-Trichloroethane	89.6	85.3	-4.3	-4.8
18)	Carbon Tetrachloride	86.0	82.4	-3.7	-4.2
19)	Vinyl Acetate	84.3	77.8	-6.5	-7.8
20)	Bromodichloromethane	91.9	87.1	-4.8	-5.2
21)	*1,4-Difluorobenzene	100.0	100.0	0.0	0.0
22)	1,2-Dichloropropane	97.9	93.9	-4.0	-4.1
23)	cis-1,3-Dichloropropene	106.1	100.5	-5.6	-5.3
24)	Trichloroethene	98.2	93.4	-4.9	-5.0
25)	DiBromochloromethane	103.9	97.1	-6.8	-6.6
26)	1,1,2-Trichloroethane	98.9	96.0	-2.9	-2.9
27)	Benzene	93.7	92.3	-1.4	-1.5
28)	trans-1,3-Dichloropropene	88.7	85.4	-3.2	-3.6
29)	Bromoform	96.8	91.0	-5.8	-6.0
30)	*Chlorobenzene-d5	100.0	100.0	0.0	0.0
31)	4-Methyl-2-Pentanone	108.6	107.0	-1.6	-1.5
32)	2-Hexanone	99.9	101.8	1.9	1.9
33)	Tetrachloroethene	92.9	93.4	0.5	0.6
34)	1,1,2,2-Tetrachloroethane	104.1	104.2	0.1	0.1
35)	Toluene	103.1	130.9	27.8	26.9
36)	#Toluene-d8	104.7	107.5	2.8	2.7
37)	Chlorobenzene	101.0	100.5	-0.6	-0.5
38)	Ethylbenzene	98.0	98.7	0.7	0.7
39)	Styrene	99.9	99.0	-0.9	-0.9
40)	p-Xylene	96.7	95.0	-1.8	-1.8
41)	o-Xylene	97.3	97.6	0.3	0.3
42)	#Bromofluorobenzene	103.1	103.9	0.8	0.8

* Compound is an Internal Standard

Compound is a Surrogate Spike

TABLE 5. Percent Recovery Results from Same Day and Day 12 GC/MS Analysis of Spiked SST U110 Sample, After Cleanup, Stored at 4°C

No.	Compound Name	Tank Sample U110		Percent	
		Avg day 1	day 12	Difference	Difference
1)	*Bromochloromethane	100.0	100.0	0.0	0.0
2)	Chloromethane	57.8	39.8	-18.0	-31.1
3)	Bromomethane	62.2	6.3	-55.9	-89.9
4)	Vinyl Chloride	46.0	38.3	-7.7	-16.7
5)	Chloroethane	71.8	46.7	-25.1	-35.0
6)	Methylene Chloride	104.3	74.5	-29.8	-28.6
7)	Acetone	48.7	33.1	-15.6	-32.1
8)	Carbon Disulfide	0.2	0.6	0.4	159.9
9)	1,1-Dichloroethene	105.1	138.0	32.9	31.3
10)	1,1-Dichloroethane	98.7	79.7	-19.1	-19.3
11)	trans-1,2-Dichloroethene	90.5	74.6	-15.9	-17.5
12)	cis-1,2-Dichloroethene	101.2	80.3	-20.9	-20.6
13)	Chloroform	103.8	84.7	-19.1	-18.4
14)	#1,2-Dichloroethane-d4	93.0	59.0	-34.0	-36.5
15)	1,2-Dichloroethane	91.6	70.5	-21.0	-23.0
16)	2-Butanone	41.4	36.3	-5.0	-12.2
17)	1,1,1-Trichloroethane	98.5	81.7	-16.7	-17.0
18)	Carbon Tetrachloride	97.6	81.8	-15.7	-16.1
19)	Vinyl Acetate	0.3	0.0	-0.3	-100.0
20)	Bromodichloromethane	100.6	79.0	-21.7	-21.5
21)	*1,4-Difluorobenzene	100.0	100.0	0.0	0.0
22)	1,2-Dichloropropane	100.8	87.9	-12.9	-12.8
23)	cis-1,3-Dichloropropene	105.4	68.0	-37.4	-35.5
24)	Trichloroethene	186.2	172.6	-13.6	-7.3
25)	Dibromochloromethane	85.2	73.9	-11.3	-13.3
26)	1,1,2-Trichloroethane	45.5	1.3	-44.2	-97.1
27)	Benzene	98.0	84.8	-13.2	-13.5
28)	trans-1,3-Dichloropropene	81.9	57.5	-24.5	-29.8
29)	Bromoform	61.4	53.7	-7.8	-12.6
30)	*Chlorobenzene-d5	100.0	100.0	0.0	0.0
31)	4-Methyl-2-Pentanone	48.2	40.7	-7.6	-15.7
32)	2-Hexanone	40.1	33.2	-6.8	-17.1
33)	Tetrachloroethene	97.6	109.4	11.8	12.0
34)	1,1,2,2-Tetrachloroethane	0.0	0.2	0.2	6480.9
35)	Toluene	101.2	140.6	39.4	39.0
36)	#Toluene-d8	103.4	84.4	-19.0	-18.4
37)	Chlorobenzene	104.3	106.3	1.9	1.8
38)	Ethylbenzene	101.3	109.3	8.0	7.9
39)	Styrene	103.4	102.2	-1.1	-1.1
40)	p-Xylene	100.9	113.6	12.7	12.6
41)	o-Xylene	115.8	123.2	7.4	6.4
42)	#Bromofluorobenzene	104.8	83.0	-21.8	-20.8

* Compound is an Internal Standard

Compound is a Surrogate Spike

Table 6 shows the results obtained for a blank analyzed using 200 μ L of methanol added and blanks employed prior to analysis of tank samples that used 100 μ L of methanol added to the purge vessel.

Table 7 provides the results obtained from analysis of SST sample U110 and matrix spikes. The relative percent difference (RPD) and the matrix effect information is tabulated also.

Tables 8 and 9 provide the results obtained from analysis and duplicate analysis of SST sample B110 and matrix spikes. The RPD and the matrix effect information are tabulated also.

Table 10 provides advisory control limits to be used for analysis of tank waste samples using the extraction and cleanup procedure.

TABLE 6. Comparison of Results Obtained from a Blank Using 200 μ L of Methanol to Blanks Using 100 μ L

No.	Compound Name	High MeOH Blank (200)	B110 Blank	B110 D Blank	U110 Blank
1)	*Bromochloromethane	100.0	100.0	100.0	100.0
2)	Chloromethane	0.0	0.0	3.7	0.0
3)	Bromomethane	0.0	0.0	0.0	0.0
4)	Vinyl Chloride	0.0	0.0	0.0	0.0
5)	Chloroethane	0.0	0.0	0.0	0.0
6)	Methylene Chloride	0.0	1.3	0.0	7.2
7)	Acetone	29.9	10.5	15.7	23.5
8)	Carbon Disulfide	0.0	0.0	0.0	0.0
9)	1,1-Dichloroethene	0.0	0.0	0.0	0.0
10)	1,1-Dichloroethane	0.0	0.0	0.0	0.0
11)	trans-1,2-Dichloroethene	0.0	0.0	0.0	0.0
12)	cis-1,2-Dichloroethene	0.0	0.0	0.0	0.0
13)	Chloroform	0.0	0.0	0.3	0.0
14)	#1,2-Dichloroethane-d4	104.5	95.5	96.9	105.1
15)	1,2-Dichloroethane	0.5	0.0	0.0	0.0
16)	2-Butanone	35.8	4.6	23.2	23.5
17)	1,1,1-Trichloroethane	0.0	0.0	0.0	0.0
18)	Carbon Tetrachloride	0.0	0.0	0.0	0.0
19)	Vinyl Acetate	0.0	0.0	0.0	0.0
20)	Bromodichloromethane	0.0	0.0	0.0	0.0
21)	*1,4-Difluorobenzene	100.0	100.0	100.0	100.0
22)	1,2-Dichloropropane	0.0	0.0	0.0	0.0
23)	cis-1,3-Dichloropropene	0.0	0.0	0.0	0.0
24)	Trichloroethene	0.0	0.0	0.0	0.0
25)	Dibromochloromethane	0.0	0.0	0.0	0.0
26)	1,1,2-Trichloroethane	0.0	0.0	0.0	0.0
27)	Benzene	0.0	0.3	0.0	0.4
28)	trans-1,3-Dichloropropene	0.0	0.0	0.0	0.0
29)	Bromoform	0.9	0.0	0.0	0.0
30)	*Chlorobenzene-d5	100.0	100.0	100.0	100.0
31)	4-Methyl-2-Pentanone	9.2	0.0	0.0	0.0
32)	2-Hexanone	11.6	0.0	6.4	1.9
33)	Tetrachloroethene	0.0	0.0	0.0	0.0
34)	1,1,2,2-Tetrachloroethane	3.4	0.0	1.9	1.0
35)	Toluene	0.0	3.4	0.0	0.0
36)	#Toluene-d8	107.0	92.4	102.6	95.5
37)	Chlorobenzene	0.0	0.0	0.0	0.5
38)	Ethylbenzene	0.0	0.0	0.0	0.0
39)	Styrene	0.0	0.0	0.0	0.0
40)	p-Xylene	0.0	0.0	0.0	0.0
41)	o-Xylene	0.0	0.0	0.0	0.0
42)	#Bromofluorobenzene	103.9	92.1	103.6	101.6

* Compound is an Internal Standard

Compound is a Surrogate Spike

TABLE 7. Percent Recovery Results, RPD, and Matrix Effect for SST U110 Sample

No.	Compound Name	Sample	Spike	Spike Duplicate	RPD	Matrix Effect
1)	*Bromochloromethane	100.0	100.0	100.0	0.0	0.0
2)	Chloromethane	0.0	58.5	57.1	2.5	-42.2
3)	Bromomethane	0.0	57.4	67.0	-15.5	-37.8
4)	Vinyl Chloride	0.0	43.5	48.4	-10.6	-54.0
5)	Chloroethane	0.0	71.8	71.8	0.0	-28.2
6)	Methylene Chloride	0.0	103.6	105.0	-1.4	4.3
7)	Acetone	0.0	44.7	52.8	-16.7	-51.3
8)	Carbon Disulfide	0.0	0.0	0.4	-200.0	-99.8
9)	1,1-Dichloroethene	0.0	118.7	91.6	25.7	5.1
10)	1,1-Dichloroethane	0.0	97.9	99.6	-1.7	-1.3
11)	trans-1,2-Dichloroethene	0.0	89.5	91.5	-2.3	-9.5
12)	cis-1,2-Dichloroethene	0.0	99.8	102.6	-2.7	1.2
13)	Chloroform	0.0	102.4	105.2	-2.7	3.8
14)	#1,2-Dichloroethane-d4	89.0	90.7	95.3	-4.9	-7.0
15)	1,2-Dichloroethane	0.0	90.3	92.8	-2.8	-8.4
16)	2-Butanone	3.0	38.3	44.4	-14.9	-58.7
17)	1,1,1-Trichloroethane	0.0	98.5	98.4	0.1	-1.6
18)	Carbon Tetrachloride	0.0	97.2	98.0	-0.8	-2.4
19)	Vinyl Acetate	0.0	0.0	0.5	-200.0	-99.7
20)	Bromodichloromethane	0.0	99.6	101.6	-2.0	0.6
21)	*1,4-Difluorobenzene	100.0	100.0	100.0	0.0	0.0
22)	1,2-Dichloropropane	0.0	99.9	101.8	-1.9	0.8
23)	cis-1,3-Dichloropropene	0.0	104.2	106.7	-2.4	5.4
24)	Trichloroethene	0.0	188.6	183.8	2.6	86.2
25)	Dibromochloromethane	0.0	83.1	87.4	-5.0	-14.8
26)	1,1,2-Trichloroethane	0.0	30.7	60.2	-65.0	-54.6
27)	Benzene	0.5	97.8	98.2	-0.4	-2.0
28)	trans-1,3-Dichloropropene	0.0	80.4	83.5	-3.8	-18.1
29)	Bromoform	0.0	57.4	65.5	-13.3	-38.6
30)	*Chlorobenzene-d5	100.0	100.0	100.0	0.0	0.0
31)	4-Methyl-2-Pentanone	0.0	46.3	50.2	-8.1	-51.8
32)	2-Hexanone	0.0	37.2	43.0	-14.5	-60.0
33)	Tetrachloroethene	0.0	95.0	100.3	-5.4	-2.4
34)	1,1,2,2-Tetrachloroethane	0.0	0.0	0.0	-200.0	-100.0
35)	Toluene	15.4	99.9	102.5	-2.6	1.2
36)	#Toluene-d8	101.2	101.4	105.3	-3.8	3.4
37)	Chlorobenzene	0.0	103.9	104.7	-0.8	4.3
38)	Ethylbenzene	0.0	99.6	103.0	-3.4	1.3
39)	Styrene	0.0	102.1	104.6	-2.4	3.3
40)	p-Xylene	0.0	100.7	101.1	-0.4	0.9
41)	o-Xylene	0.8	113.9	117.7	-3.2	15.8
42)	#Bromofluorobenzene	103.0	104.9	104.7	0.2	4.8

* Compound is an Internal Standard

Compound is a Surrogate Spike

TABLE 8. Percent Recovery Results, RPD, and Matrix Effect for B110 Sample

No.	Compound Name	Sample	Spike	Spike Duplicate	RPD	Matrix Effect
1)	*Bromochloromethane	100.0	100.0	100.0	0.0	0.0
2)	Chloromethane	22.1	4.6	34.8	-153.3	-60.1
3)	Bromomethane	8.7	35.1	42.6	-19.4	-43.7
4)	Vinyl Chloride	0.0	41.2	48.2	-15.7	-22.7
5)	Chloroethane	0.0	51.4	57.9	-12.0	-18.1
6)	Methylene Chloride	0.0	85.0	94.8	-10.8	4.6
7)	Acetone	6.5	45.0	45.4	-0.9	-82.6
8)	Carbon Disulfide	0.0	38.1	44.0	-14.4	-12.6
9)	1,1-Dichloroethene	0.0	65.3	74.4	-13.0	5.4
10)	1,1-Dichloroethane	0.0	87.4	95.8	-9.2	7.6
11)	trans-1,2-Dichloroethene	0.0	76.9	87.5	-12.8	-5.7
12)	cis-1,2-Dichloroethene	0.0	92.0	102.7	-11.0	6.5
13)	Chloroform	0.0	95.1	106.3	-11.1	6.7
14)	#1,2-Dichloroethane-d4	87.4	84.9	96.0	-12.3	-5.1
15)	1,2-Dichloroethane	0.0	83.1	96.5	-14.9	-1.8
16)	2-Butanone	0.0	36.0	43.8	-19.5	-71.0
17)	1,1,1-Trichloroethane	0.0	88.6	95.3	-7.2	4.5
18)	Carbon Tetrachloride	0.0	84.6	90.1	-6.2	3.1
19)	Vinyl Acetate	0.0	0.0	0.0	NA	-81.1
20)	Bromodichloromethane	0.0	90.9	102.9	-12.3	7.4
21)	*1,4-Difluorobenzene	100.0	100.0	100.0	0.0	0.0
22)	1,2-Dichloropropane	0.0	88.1	95.4	-8.0	-4.2
23)	cis-1,3-Dichloropropene	0.0	90.1	97.0	-7.5	-9.7
24)	Trichloroethene	0.0	93.7	99.9	-6.4	1.0
25)	Dibromochloromethane	0.0	73.9	84.2	-13.0	-21.5
26)	1,1,2-Trichloroethane	0.0	70.6	80.4	-13.0	-21.9
27)	Benzene	0.3	80.0	89.7	-11.4	-8.2
28)	trans-1,3-Dichloropropene	0.0	70.6	76.9	-8.5	-13.3
29)	Bromoform	0.0	70.5	81.9	-15.0	-17.6
30)	*Chlorobenzene-d5	100.0	100.0	100.0	0.0	0.0
31)	4-Methyl-2-Pentanone	0.0	35.7	38.2	-6.8	-70.8
32)	2-Hexanone	0.0	28.9	31.0	-6.9	-70.9
33)	Tetrachloroethene	0.0	69.7	73.4	-5.2	-21.6
34)	1,1,2,2-Tetrachloroethane	0.0	38.6	40.1	-3.7	-64.8
35)	Toluene	5.1	80.8	80.9	-0.0	-51.1
36)	#Toluene-d8	75.6	76.1	78.1	-2.5	-29.0
37)	Chlorobenzene	0.0	85.5	88.0	-2.8	-14.0
38)	Ethylbenzene	0.0	79.0	82.1	-3.9	-17.8
39)	Styrene	0.0	78.3	82.4	-5.1	-19.1
40)	p-Xylene	0.0	75.9	74.4	2.1	-20.7
41)	o-Xylene	0.4	89.8	78.1	14.0	-13.5
42)	#Bromofluorobenzene	78.4	85.9	80.0	7.0	-20.6

* Compound is an Internal Standard

Compound is a Surrogate Spike

TABLE 9. Percent Recovery Results, RPD, and Matrix Effect for B110 Duplicate

No.	Compound Name	Sample	Spike	Spike Duplicate	RPD	Matrix Effect
1)	*Bromochloromethane	100.0	100.0	100.0	0.0	0.0
2)	Chloromethane	6.1	51.0	42.3	18.7	-33.1
3)	Bromomethane	0.0	73.0	60.6	18.7	-15.7
4)	Vinyl Chloride	0.0	59.7	48.8	20.0	-13.1
5)	Chloroethane	0.0	68.8	54.2	23.6	-11.2
6)	Methylene Chloride	0.0	86.6	78.2	10.2	-3.0
7)	Acetone	0.0	39.0	55.3	-34.5	-80.7
8)	Carbon Disulfide	0.0	49.9	30.9	47.1	-13.3
9)	1,1-Dichloroethene	0.0	79.3	55.6	35.0	3.0
10)	1,1-Dichloroethane	0.0	95.7	86.1	10.5	6.9
11)	trans-1,2-Dichloroethene	0.0	86.0	77.5	10.3	-6.2
12)	cis-1,2-Dichloroethene	0.0	98.0	88.2	10.6	2.3
13)	Chloroform	0.0	99.0	89.0	10.6	-0.0
14)	#1,2-Dichloroethane-d4	77.2	88.5	88.9	-0.5	-6.8
15)	1,2-Dichloroethane	0.0	85.5	87.2	-2.0	-5.3
16)	2-Butanone	0.0	36.5	62.9	-53.3	-61.2
17)	1,1,1-Trichloroethane	0.0	92.9	83.4	10.8	0.7
18)	Carbon Tetrachloride	0.0	84.7	75.8	11.1	-4.0
19)	Vinyl Acetate	0.0	0.0	0.0	NA	-81.1
20)	Bromodichloromethane	0.0	91.0	84.7	7.2	-1.7
21)	*1,4-Difluorobenzene	100.0	100.0	100.0	0.0	0.0
22)	1,2-Dichloropropane	0.0	86.5	88.4	-2.2	-8.5
23)	cis-1,3-Dichloropropene	0.0	91.7	96.4	-5.0	-9.3
24)	Trichloroethene	0.0	89.4	91.9	-2.7	-5.2
25)	Dibromochloromethane	0.0	70.6	80.6	-13.3	-24.9
26)	1,1,2-Trichloroethane	0.0	69.1	83.5	-18.9	-21.1
27)	Benzene	0.0	87.6	86.8	0.9	-5.8
28)	trans-1,3-Dichloropropene	0.0	67.5	75.7	-11.4	-15.4
29)	Bromoform	0.0	66.6	87.6	-27.2	-16.8
30)	*Chlorobenzene-d5	100.0	100.0	100.0	0.0	0.0
31)	4-Methyl-2-Pentanone	0.0	37.1	60.8	-48.3	-58.8
32)	2-Hexanone	0.0	30.2	52.8	-54.6	-59.3
33)	Tetrachloroethene	0.0	80.4	78.4	2.5	-13.8
34)	1,1,2,2-Tetrachloroethane	0.0	48.2	67.2	-32.9	-46.5
35)	Toluene	0.0	96.8	91.5	5.7	-37.8
36)	#Toluene-d8	96.1	96.3	94.0	2.4	-11.0
37)	Chlorobenzene	0.0	96.0	91.6	4.7	-6.9
38)	Ethylbenzene	0.0	91.1	91.7	-0.7	-6.9
39)	Styrene	0.0	90.6	88.4	2.4	-9.9
40)	p-Xylene	0.0	84.0	81.8	2.7	-13.0
41)	o-Xylene	0.0	101.7	85.9	16.8	-3.6
42)	#Bromofluorobenzene	95.7	100.9	87.4	14.3	-9.3

* Compound is an Internal Standard

Compound is a Surrogate Spike

TABLE 10. Calculated Advisory Control Limits for SST Samples

No.	Compound Name	Mean		-3XSTD	+3XSTD
		Recovery	STD (N-1)		
1)	*Bromochloromethane	100.0	0.0	100.0	100.0
2)	Chloromethane	56.9	25.5		
3)	Bromomethane	67.7	19.3		
4)	Vinyl Chloride	57.6	15.4		
5)	Chloroethane	69.1	12.8		
6)	Methylene Chloride	91.1	10.2		
7)	Acetone	71.3	35.1		
8)	Carbon Disulfide	40.4	25.6		
9)	1,1-Dichloroethene	77.0	19.8	17.7	136.4
10)	1,1-Dichloroethane	93.3	7.8		
11)	trans-1,2-Dichloroethene	88.7	8.3		
12)	cis-1,2-Dichloroethene	97.5	6.6		
13)	Chloroform	100.0	6.7		
14)	#1,2-Dichloroethane-d4	94.5	8.5	69.1	120.0
15)	1,2-Dichloroethane	93.4	8.4		
16)	2-Butanone	66.2	32.1		
17)	1,1,1-Trichloroethane	93.4	6.5		
18)	Carbon Tetrachloride	89.6	8.3		
19)	Vinyl Acetate	37.0	48.3		
20)	Bromodichloromethane	96.1	8.0		
21)	*1,4-Difluorobenzene	100.0	0.0	100.0	100.0
22)	1,2-Dichloropropane	95.3	5.8		
23)	cis-1,3-Dichloropropene	100.0	6.0		
24)	Trichloroethene	113.5	38.5	1.0	228.9
25)	Dibromochloromethane	87.1	10.6		
26)	1,1,2-Trichloroethane	79.7	23.0		
27)	Benzene	91.6	6.0	73.7	109.5
28)	trans-1,3-Dichloropropene	81.2	8.4		
29)	Bromoform	79.4	13.4		
30)	*Chlorobenzene-d5	100.0	0.0	100.0	100.0
31)	4-Methyl-2-Pentanone	68.2	31.4		
32)	2-Hexanone	62.7	33.9		
33)	Tetrachloroethene	90.9	15.3		
34)	1,1,2,2-Tetrachloroethane	61.5	42.7		
35)	Toluene	101.0	15.2	55.6	146.5
36)	#Toluene-d8	97.6	11.7	62.5	132.8
37)	Chlorobenzene	99.7	9.6	70.8	128.7
38)	Ethylbenzene	96.7	11.1		
39)	Styrene	96.9	11.7		
40)	p-Xylene	93.4	14.0		
41)	o-Xylene	99.6	12.5		
42)	#Bromofluorobenzene	98.0	9.7	68.9	127.2

* Compound is an Internal Standard

Compound is a Surrogate Spike

DISCUSSION

The data described in the Results section are interpreted and the development of the final procedure explained is explained in the following discussion.

INITIAL DEVELOPMENT

As illustrated in Figure 2, separation of the limited analytes, acetone and xylenes, from NPH was good. This indicated that substantial cleanup of NPH, at the solvent composition methanol/water (80:20), could be expected. Analysis of the same column eluent by GC/MS verified the removal of NPH. GC/MS analysis was performed in order to detect lower levels of NPH than the RI detector was capable of.

Table 1 shows the recoveries obtained from the partial target list standards with and without NPH. The standards used the initially developed cleanup procedure and employed a 1-mL sample introduction. They were eluted with methanol/water (80:20) to a final volume of 5 mL. The results show that generally the compounds with lower relative polarity exhibit lower recovery, indicating the need for a less polar eluting solvent composition. The presence of NPH during this experiment apparently improved recoveries. This could indicate a stationary phase loading effect from NPH. The NPH competes for the less polar adsorption sites of the stationary phase in leaving the polar mobile phase. This in turn forces more of the analyte to the mobile phase. This also introduces a differing equilibria between the new stationary phase and the mobile phase components. This effect is most conspicuous when the loading capacity of the column is approached.

The recoveries exceeding 100 percent, primarily the ketones, are presumably from contaminated methanol. Ketones are typical contaminants of methanol and the relatively large quantity of this solvent employed during this analysis magnifies the complication.

FINAL DEVELOPMENT AND VALIDATION STUDY

Initial familiarization and testing during this phase of the study revealed that a portion the analyte losses were due to the use of a nylon valve employed in the closed cleanup system. Prior to solvent optimization studies, replacement with a stainless steel valve and eventually a Hamilton Teflon® lined valve improved recoveries.

To determine the optimum column eluting solvent for greatest analyte recovery, blank spikes containing the full volatile target compound list were loaded onto the C₁₈ cartridge cleanup column and eluted using 80, 85, and 90 percent methanol in water. Table 2 shows the results obtained from GC/MS analysis when the composition of the eluting solvent was varied from 80, 85, and 90 percent methanol in water. The best recoveries were obtained at 90 percent methanol. The recoveries of the compounds with relatively lower polarity, such as the xylenes and benzenes, improved substantially.

To determine the potential loading of NPH onto the cleanup column, a solubility study was performed by saturating various methanol/water compositions and measuring the NPH present. Figure 3 shows the results of the solubility study. At up to 80 percent methanol, the solubility of the NPH is very low. The NPH removal mechanism at up to 80 percent methanol, then, is primarily one of solubility where the NPH is removed through liquid/liquid partitioning. The capacity of the cartridge column, as provided by the manufacturer, is about 10 mg and will not be exceeded until the concentration of methanol is about 96 percent.

To assess the ability of the cartridge column to remove NPH in various methanol/water solvent compositions, blanks were spiked with NPH and eluted through a cartridge column. The elution solvent composition was varied between 80 and 100 percent methanol at five different levels. The amount of NPH present in the column eluent was then measured. Figure 4 shows that no measurable quantity of NPH was present at 90 percent or less methanol in water. At 95 percent methanol, 0.43 percent recovery of the NPH was measured. The purge and trap and GC/MS system would receive about 600 ng of NPH per sample based upon the following:

- The solubility of NPH in 95 percent methanol at 21°C is about 7000 $\mu\text{g/mL}$.
- The amount of the extractant employed in the cleanup process is 1 mL.
- The amount of methanol that is currently used in the analysis is 200 μL .

Retention of this amount of NPH could be devastating to the volatiles analytical system over time. Therefore, 90 percent methanol in water was chosen as the composition of the column eluting solvent. No measurable amount of NPH should be observed using the 90 percent methanol solvent composition.

Table 3 shows the results for a blank, extracted blank spike, and the same blank spike after performing the cleanup procedure using 90 percent methanol as the eluting solvent and analyzing 200 μL of the eluent. The difference is presented as the effect of the cleanup procedure. The most volatile and polar analytes show the greatest recovery losses, as much as 38 percent, although most are quite satisfactory. The ketone recoveries show an increase due to their presence in the methanol. This contamination is best identified by inspection of the corresponding blank results.

It was also necessary to determine the maximum amount of methanol, and therefore the amount of cleaned extract that may be analyzed, that could be introduced into the purge and trap and GC/MS systems. Figure 5 shows the effect that 10, 100, 200, and 300 μL of methanol in the purge vessel had on the chromatography of the analyte chloroform. Figure 6 shows the effect that 100, 200, and 300 μL of methanol had on bromomethane. The results are quite obvious: less methanol in the purge vessel results in better chromatography. Based on these results and the chromatographic performance of other compounds, the amount of methanol extract that is purged should not exceed 100 μL . Additionally, the EPA medium-level analysis has the same limit.

Rather than allow the dilution of the method to become higher, it was proposed that 2 mL (instead of 1 mL) of the extraction mixture would be introduced into the cleanup cartridge column. The amount of methanol for analysis could then be reduced from 200 μL , as used previously, to 100 μL ; the overall dilution would remain the same, 625 for a 1 gram sample. This would also reduce the methanol ketone contamination difficulties. Table 6 shows a comparison of several blanks. The first column, "High MeOH Blank (200)," is a

blank using 200 μ L of methanol in the purge vessel during analysis. The three other columns are blanks prepared prior to analysis of SST waste samples and employ 100 μ L of methanol. The contamination is reduced as expected.

When 2 mL of the extractant is used, the small 0.5-mL rinses of the sample syringe become even more important in order to flush the compounds through the cartridge in a narrow band to yield good recovery. The 90 percent methanol worked very well at eluting the target compounds while permitting the Environmental C_{18} Sep-Pak cartridge to retain the NPH, even though the NPH was present in the original sample at high levels. The 90 percent methanol ratio permits using a 1 gram sample with 0 to 100 percent moisture. For example, if the sample was 100 percent moisture, the extraction mixture would be 75 percent methanol, which will still permit good recovery of the target compounds during the extraction process. In contrast, if a sample with zero percent moisture is extracted, the 90 percent methanol extract will still allow the NPH to remain on the Sep-Pak cartridge.

Evaluation of the above experiments yielded a final set of parameters for use in the column cleanup procedure. These empirically derived parameters were verified experimentally by analyzing spiked blanks and measuring recoveries. Validation of the procedure for use on actual samples was performed by analyzing blank spikes and spiked SST samples.

Additionally, to assess extractant storage possibilities, same-day analyses of a blank spike and SST-spiked sample U110 were compared to analyses performed 12 days later on the same residue. Table 4 shows the results of both the same-day and 12-day analyses. The difference and relative difference between the two results is also given. Methylene chloride shows the greatest loss, although all the compounds exhibit adequate comparison. Toluene shows an increase of almost 27 percent. This data is viewed as an anomaly as calibration data seems reasonable for each day of analysis and holding blanks from the refrigerated storage have not revealed toluene contamination problems. Table 5 shows the same storage time study using a residue from SST sample U110. This sample showed substantial change in the results between the two analyses. While there are considerable losses for many compounds, it seems

unreasonable to believe that an analyte native to this sample would not have undergone similar processes long before this analysis.

The results obtained from GC/MS analysis of SST sample U110 are provided in Table 7. The RPD and the matrix effect are supplied as well. The RPD is calculated by dividing the difference by the mean of the sample spike and the spike duplicate times 100. The matrix effect is calculated by subtracting the blank spike results from the mean of the sample spikes results. Therefore, a negative number represents the loss that can be expected as a result of the matrix effect of this sample relative to analysis of a blank spike.

Tables 8 and 9 provide the results tabulated in the same manner for SST sample B110 and a duplicate set of analysis. The matrix effect is substantial for both samples, although the effect of U110 was noticeably more severe. For the compounds 1,1,2,2-tetrachloroethane, carbon disulfide, and vinyl acetate, essentially complete loss is observed in sample U110. Under the alkaline conditions for this sample, approximately pH 12, it is postulated that 1,1,2,2-tetrachloroethane underwent base elimination of HCl to yield trichloroethylene (TCE). This is substantiated by the elevated recovery of TCE of about 185 percent. This was also observed in the same-day and 12-day analyses for sample U110 as shown in Table 5. Results from analysis after 12 days also indicated that 1,1,2-trichloroethane has undergone similar degradation to form 1,1-dichloroethene. The 1,1-dichloroethene yield is not as quantitative as the TCE, so it must be assumed that a competing reaction or further degradation is taking place. The reaction kinetics of the tetrachloroethane to TCE conversion are much faster than with the trichloroethane to dichloroethene conversion.

In addition to these elimination reactions, other decompositions were observed. Chloromethane and bromomethane exhibited low spike recovery, more so under the U110 sample alkaline extraction conditions. Interestingly, the recovery for bromomethane was lower than for chloromethane. This observation is consistent with the expectation that the bromide ion is a better leaving group than is the chloride ion.

Chloromethane and bromomethane were also observed in the extracts of sample B110. However, this observation may be an artifact due to the possibility

of degradation of the first internal standard, bromochloromethane, to these compounds. Responses for bromochloromethane were as much as 20 percent low during B110 analysis. This effect can not be directly associated with alkalinity because sample B110 was measured at about pH 9.5. Sample U110 was more alkaline, pH 12, but bromochloromethane response was not noticeably affected and chloromethane or bromomethane were not observed.

Other compounds that seem to exhibit alkali sensitivity are vinyl acetate, vinyl chloride, and carbon disulfide, which were decomposed by the tank sample matrix.

The effect of chelation is beyond the scope of this report, however chelators are known to exist in many of the tank wastes and are typically used to complex with metal cations and will thereby increase the effective alkalinity (and nucleophilicity) of anions.

Table 10 presents the calculated advisory control limits to be used when employing this extraction and cleanup procedure on tank waste samples. The procedure in the Appendix also contains this information.

CONCLUSIONS

Samples obtained from the Hanford SSTs are contaminated with NPH introduced as hydrostatic fluid from the sampling process or native to the tank waste. The contamination is usually high enough that a dilution of up to several orders of magnitude may be required before the sample can be analyzed by the conventional methodology of purge and trap concentration followed by gas chromatography/mass spectrometry. To eliminate this problem, use of this sample extraction and cleanup method is encouraged and is summarized as follows:

An aliquot (nominally 1 gram) of the sample is extracted with 5 mL of methanol/water (90:10) containing the EPA CLP surrogate compounds. A 2-mL aliquot is syringe-pumped through a C₁₈ Sep-Pak^R cartridge and followed by 0.5 mL rinses until a final volume of 5 mL is obtained. One hundred μ L of the eluent is transferred to a purge vessel containing 4.9 mL of water. Internal standards are added to the purge vessel and purge and trap extracted per the EPA method or the identical procedure PNL-ALO-335, "GC/MS Analysis of Volatile Organic Compounds." The final dilution incorporated into this method is 625-fold and the resulting detection limits for PNL-ALO-35 become 6250 and 3125 μ g/L.

The effectiveness of this cleanup method was demonstrated for all of the volatile target list compounds in the CLP SOW. Both blanks and SST samples spiked with all of the target compounds and NPH were used in this development and validation study. The NPH removal for samples and blank spikes is better than 99 percent. The recoveries were excellent for most of the target compounds in blank spikes and satisfactory for all compounds, although lower for the more volatile compounds. The blank spike residues were shown to store well for up to 12 days. The SST samples were observed to have a substantial matrix effect on recovery of the gases, ketones, and saturated compounds that are halogenated. Advisory quality control limits from tank waste data were established for the surrogate and matrix spike compounds.

There are several advantages to employing this procedure. The first and most obvious is the removal of NPH. The NPH cleanup has worked well to remove

large amounts of NPH from the samples. According to GC/MS analysis, tank sample extracts analyzed by this procedure had less than detectable amounts of NPH in them. These samples originally contained a level of NPH that analysis without this cleanup would require at least a 5000 to 10000 dilution factor.

A second advantage is that the methanol extracts of the samples are of much lower radioactivity than the original samples. For example, the radioactivity of some of the samples was measured at about 1000 mR per gram. In contrast, the 5-mL extracts of these samples were less than 1 mR each in radioactivity.

A third advantage is that following extraction, the residue may be stored for up to 12 days. This allows for better optimizing analytical batch sizes and resource planning.

As more data are obtained using this cleanup method, required QC limits will be put in place and a better definition of cleanup effectiveness can be established.

The NPH contamination was considered to be the overwhelming factor affecting GC/MS analysis. If other less conspicuous analytical difficulties exist, they may only now be detectable. For example, siloxanes have been detected in past analyses when large amounts of NPH were present. Following the cleanup procedure, siloxanes have still been detected during analysis of semivolatile residues. The detrimental contribution siloxanes may have, if any, has not been determined at this time.

Capillary gas chromatography as an analytical tool is well documented and offers high chromatographic resolution and sensitivity, and is specified in the current regulatory literature for semivolatile analysis. However, it lacks the ruggedness often required when characterizing complex wastes. Complementary preparation and analyses employing more robust separation techniques such as liquid chromatography should be developed and will certainly have broad application in waste tank characterization and should be explored in the near future.

REFERENCES

U.S. Environmental Protection Agency Contract Laboratory Program, Statement of Work for Organic Analysis, Multi-Media, Multi-Concentration, SOW No. 2-88.

U.S. Environmental Protection Agency Contract Laboratory Program, Statement of Work for Organic Analysis, Multi-Media, Multi-Concentration, Document Number OLM01.0, March 1990.

U.S. Environmental Protection Agency Office of Solid Waste, Test Methods for Evaluating Solid Waste, SW-846, Third Edition, 11-86.

PNL-ALO-335, GC/MS Analysis of Volatile Organic Compounds, Rev. 0, Pacific Northwest Laboratory Technical Procedure, 8-3-89.

APPENDIX A

PNL-ALO-121, CLEANUP OF HYDROCARBON CONTAINING SAMPLES
FOR THE ANALYSIS OF VOLATILE ORGANIC COMPOUNDS

APPENDIX A

PNL-ALO-121, CLEANUP OF HYDROCARBON CONTAINING SAMPLES FOR THE ANALYSIS OF VOLATILE ORGANIC COMPOUNDS

APPLICABILITY

This procedure is useful for the analysis of volatile compounds in samples containing interfering quantities of normal hydrocarbons. This procedure was developed to remove normal paraffin hydrocarbons (NPH), typically C_{12} through C_{14} , often used as a non-miscible hydrostatic fluid for obtaining radioactive tank wastes on the Hanford site. When employing conventional purge and trap sample introduction methods, large quantities of heavy hydrocarbons, such as NPH, can be introduced into the analytical system. The resulting performance of the analytical system is drastically reduced and usually requires extensive hardware cleanup and maintenance. To avoid compromising an analytical system in this manner, this procedure should be utilized. Diluting the sample to reduce the hydrocarbon can be considered, however, often the dilution factor required may be several orders of magnitude. Use of this procedure in conjunction with most GC/MS methods increases the detection limits by approximately 625 times. Any sample matrix containing excessive quantities of normal hydrocarbons, such as waters, sorbents, or soils, are applicable to this procedure.

DEFINITIONS

NPH (Normal Paraffin Hydrocarbons)

CRQL (Contract Required Quantitation Limit)

RESPONSIBLE STAFF

Cognizant Scientist

Analyst

PROCEDURE

Scope and Application

This method employs a methanolic extraction of a sample followed by cleanup of the residue using C₁₈ solid phase sorbent cartridges. A portion of the resultant residue from this preparation is analyzed using other procedures such as PNL-ALO-335, "GC/MS Analysis of Volatile Organic Compounds". Unless specifically mentioned within this procedure, there are no further deviations of the analytical methods used to analyze residues from this preparation.

This method may be used for the analysis of radioactive storage tank wastes which contain hydrocarbons in excessive quantity. This procedure does not allow for distinction between hydrocarbons introduced during the sampling process as contamination and those which are native to the tank wastes. An attempt to quantify any compound not included in the Volatiles Target Compound List (TCL) following use of this procedure without further validation is inappropriate.

This procedure may be used to remove oil from various matrices, such as soil. Before using this procedure, validation employing matrix spikes and establishing control limits for each distinct matrix should be considered.

SUMMARY OF METHOD

An aliquot (nominally 1 gram) of the sample is extracted with 5 mL of 90/10 methanol/water containing surrogate compounds. A 2 mL aliquot of the extract is syringe-pumped through a C₁₈ Sep-Pak^R Cartridge and followed with 0.5 mL rinses with 90/10 methanol/water until a final volume of 5 mL is obtained. An analysis procedure specified by the cognizant scientist is then performed on the clean residue.

INTERFERENCES

"Purge and Trap" grades of methanol are typically of sufficient quality for use in this Procedure. Continued exposure of the methanol to the ambient

air is not desirable. An approach to minimize this problem is to purchase several small quantities of methanol or aliquot larger amounts to several small containers to reduce exposure time.

Several target compounds are subject to substantial degradation under alkaline conditions. For example, under alkaline conditions, 1,1,2,2-Tetrachloroethane degrades and results for this analyte will be low. The degradation forms Trichloroethene and causes elevated results for this compound. In low moisture conditions, such as dry alkaline soils or where the presence of oil precludes water, the degradation may not be complete. Until validation of a suitable pH adjustment procedure is available, residues of low moisture alkaline soils should be analyzed immediately following the extraction procedure.

APPARATUS AND MATERIALS

Balance: Analytical, capable of accurately weighing 0.01 gram.

Vials: Pre-cleaned volatile (VOA) type 1.5 mL, 5 mL, 20 mL screw top with teflon lined silicone backed closures.

Pipets: Pasteur type or equivalent.

Centrifuge: Bench top capable of achieving 2000 rpm.

Syringes: Gas-tight, 10 uL, 100 uL, 5.0 mL and other assorted sizes as required.

Syringe Valve: Two-way with Luer ends (three each).

Sep-Pak^R Cartridges: C₁₈ Environmental (Waters) or equivalent.

Syringe Needle: Blunt tip, luer or luer-lok, 19 gauge, 10 cm long, stainless steel.

REAGENTS

Methanol: Purge and Trap grade or equivalent.

Water: Obtained from a Nanopure water purification unit equipped with an Organic-free cartridge or equivalent.

SAMPLE COLLECTION, PRESERVATION, AND HANDLING

Samples to be used for this procedure shall follow guidelines provided by the appropriate procedure and matrix, such as PNL-ALO-335.

It has been demonstrated that sample extraction residues may be stored 12 days at -10°C to -20°C, minimal or no headspace, and protected from light.

PROCEDURE

Extraction of Sample

These procedures may be performed remotely in "hot cells" if the samples are highly radioactive.

Approximately 1 gram of sample is accurately weighed into a 20 mL VOA vial. Record the weight to the nearest 0.01 grams.

To all samples, add 5.0 mL of 90/10 methanol/water (v:v) solution which has been spiked with surrogate compounds at the concentration as specified in the defined analysis method. For example, if PNL-ALO-335 is specified: 12.5 μ L of surrogate mix at a concentration of 2500 μ g/mL is added.

If this is a matrix spike sample, add the concentration specified in the defined analysis method for spike compounds. For example, if PNL-ALO-335 is specified: 12.5 μ L of the spike mix at a concentration of 2500 μ g/mL is added.

Cap the vial containing the sample and spiked solution and shake for 1 minute. Allow the contents to settle for 10 minutes or until the liquid phase appears transparent.

If the liquid phase is noticeably cloudy or is not well separated, it will be necessary to centrifuge the residue. If the centrifuge is unable to accommodate the 20 mL vial, transfer as much of the liquid as possible to a clean appropriately sized vial. Centrifuge with the vial capped for three minutes at 2000 rpm.

Transfer the extracted liquid residue to a 5 mL vial. Cap the vial tightly with the teflon side of the septa cap liner facing the sample. Remove from the hot cell if necessary. Record the date and time of extraction.

If not performing the cleanup immediately (see precautions in the "Interferences" section) store the residues for up to 12 days at -10°C to -20°C in a freezer appropriate for storage of flammable solvents. Store the residues away from other solvents as required of standards.

Residue Cleanup

Immediately prior to use, an Environmental C₁₈ Sep-Pak[®] Cartridge is pre-conditioned by passing 5 mL of methanol followed by 5 mL of 90/10 methanol/water through the cartridge from a clean 5 mL luer-lok, gas-tight syringe. Discard the eluant.

Slowly take up 2 mL of the sample extract solution in a 5 mL gas-tight syringe equipped with a teflon luer-lok and an 19 gauge, 10 cm needle. Pulling the sample too quickly into the syringe imparts excessive partial pressure to the residue and may result in lost volatile components.

Remove the needle from the syringe. Connect the sample syringe to a pre-conditioned Sep-Pak[®] cartridge. Connect the other side of the cartridge to a three-port, luer-lok, 2-way valve. Connect a clean, 5 mL gas-tight receiving syringe to another port on the valve.

Adjust the valve so that there is flow between the two ports connecting the sample and receiving syringes. Slowly syringe the sample extract solution through the Sep-Pak cartridge and into the receiving gas-tight syringe.

Fill the sample syringe repeatedly with 0.5 mL rinses of 90/10 methanol/water. Pump the liquid through the cartridge and into the receiving syringe until a 5.0 mL final volume is obtained in the receiving syringe.

Close the valve. Invert the syringe several times to insure adequate mixing of the eluent.

If not performing the analysis the same day, the receiving syringe may be removed and an aliquot or all of the cleaned residue may be transferred to a 1.5 mL or 5 mL vial. No headspace shall remain in the 1.5 mL vial and little or no headspace should remain in the 5 mL vial. Record the date of cleanup.

Store the residues at -10°C to -20°C in a freezer appropriate for storage of flammable solvents. Store the residues away from other solvents as required of standards. The time from extraction and cleanup to analysis shall not exceed 12 days.

Analysis

Transfer 100 uL of the cleaned residue to a purge vessel containing internal standards and 4.9 mL of reagent water. Attach the vessel to a purge and trap sample concentrator and purge as specified in the analytical method such as PNL-ALO-335. The attached control limits were calculated from data obtained using a heated purge. The heated purge, however, is regarded as optional for this procedure provided consistency is observed per the analytical method such as in PNL-ALO-335.

QUALITY CONTROL

Before processing samples, the analyst shall demonstrate through the analysis of a method blank that all glassware and reagents are interference free when used in the quantity as specified in this method. With each batch of samples, and whenever there is a change in reagents or the location of performing this procedure, a method blank shall be processed as a safeguard against chronic laboratory contamination. The blank samples shall be carried out through all stages of the sample preparation including matching the handling and storage times and location.

Standard quality assurance practices consistent with the analysis method shall be employed while using this method. Fortified samples shall be carried through all stages of sample preparation adhering to those parameters previously specified with the blank preparation.

Limits of uncertainty for all measurements shall be in accordance with PNL QA Plan MCS-033 Rev. 0. Briefly, the limit is ± 5 in the next digit beyond the last one stated. For example, 5.0 mL means 5.0 ± 0.05 mL.

Control Limits

Surrogate standard determinations shall be performed on all samples and blanks. All samples and blanks are fortified with surrogate spiking compounds before extraction in order to monitor the preparation and analysis of samples.

Matrix spikes shall be performed at a frequency specified by PNL-ALO-335. Briefly, matrix spike and matrix spike duplicate analyses are required for each matrix and every twenty samples, the same as specified by the EPA-CLP SOW dated 2/88.

Control limits for this method were calculated in accordance with EPA and PNL QA Plan MCS-033 Rev 0 procedures. Table 1 provides surrogate recovery control limits and Table 2 provides spiking compound recovery limits obtained from analyses using PNL-ALO-335. Due to the limited quantity of data provided from validation study samples and sample matrices, all control limits for this method shall be considered advisory only. Currently, no action will be required if these control limits can not be met; however these instances shall be noted in the case narrative.

METHOD PERFORMANCE AND SPECIFIC QUALIFICATION

Table 3 provides method accuracy and precision based on analysis of 18 surrogate spiked samples and 11 TCL spiked samples on three different matrices.

TABLE 1. Surrogate Recovery Control Limits

<u>Surrogate Compound</u>	<u>Level^(a)</u>	<u>Recovery Limit (%)</u>	<u>RPD Limit</u>
1,2-Dichloroethane-d ₄	50 ug	69-120	50
Toluene-d ₈	50 ug	63-133	50
4-Bromofluorobenzene	50 ug	69-127	50

All limits are for advisory purposes only

(a) At time of injection.

TABLE 2. Matrix Spike Recovery Limits

<u>Matrix Spike Compound</u>	<u>Level^(a)</u>	<u>Recovery Limit (%)</u>	<u>RPD Limit</u>
1,1-Dichloroethene	50 ug	18-136	50
Trichloroethene	50 ug	01-229	50
Chlorobenzene	50 ug	71-129	50
Toluene	50 ug	56-160	50
Benzene	50 ug	74-110	50

All limits are for advisory purposes only

(a) At time of injection.

The validation of this procedure was performed using spikes of the complete Volatiles Target Compound List (TLC) from the USEPA-CLP, SOW dates 2.88, on a limited number of authentic single shell tank wastes and blank samples. See Reference 1.

TABLE 3. Method Accuracy and Precision

Compound Name	Percent Recovery	
	Mean	Standard Deviation (n-1)
1) Bromochloromethane ^(a)	100.0	0.0
2) Chloromethane	62.5	18.5
3) Bromomethane	69.5	16.6
4) Vinyl Chloride	57.6	15.4
5) Chloroethane	69.1	12.8
6) Methylene Chloride	91.1	10.2
7) Acetone	81.6	45.6
8) Carbon Disulfide	40.4	25.6
9) 1,1-Dichloroethene	77.0	19.8
10) 1,1-Dichloroethane	93.3	7.8
11) trans-1,2-Dichloroethene	88.7	8.3
12) cis-1,2-Dichloroethene	97.5	6.6
13) Chloroform	100.0	6.7
14) 1,2-Dichloroethane-d ₄ ^(b)	94.5	8.5
15) 1,2-Dichloroethane	93.4	8.4
16) 2-Butanone	74.0	39.8
17) 1,1,1-Trichloroethane	93.4	6.5
18) Carbon Tetrachloride	89.6	8.3
19) Vinyl Acetate	37.0	48.3
20) Bromodichloromethane	96.1	8.0
21) 1,4-Difluorobenzene ^(a)	100.0	0.0
22) 1,2-Dichloropropane	95.3	5.8
23) cis-1,3-Dichloropropene	100.0	6.0
24) Trichloroethene	113.5	38.5
25) Dibromochloromethane	87.1	10.6
26) 1,1,2-Trichloroethane	79.7	23.0
27) Benzene	91.8	6.0
28) trans-1,3-Dichloropropene	81.2	8.4
29) Bromoform	79.6	13.6
30) Chlorobenzene-d ₅ ^(a)	100.0	0.0
31) 4-Methyl-2-Pentanone	70.0	33.6
32) 2-Hexanone	65.0	37.1
33) Tetrachloroethene	90.9	15.3
34) 1,1,2,2-Tetrachloroethane	62.1	43.5
35) Toluene	108.1	17.4
36) Toluene-d ₈ ^(b)	97.6	11.7
37) Chlorobenzene	99.7	9.6
38) Ethylbenzene	96.7	11.1
39) Styrene	96.9	11.7
40) p-Xylene	93.4	14.0
41) o-Xylene	99.9	12.6
42) Bromofluorobenzene ^(b)	98.0	9.7

(a) Compound is an Internal Standard.

(b) Compound is a Surrogate Spike.

REFERENCES

1. Hoppe, E. W.; Ross, G. A.; Lucke, R. B.; Campbell, J. A.; Development and Validation of a Preparation and Cleanup Method for Hydrocarbon Containing Samples for the Analysis of Volatile Organic Compounds, Pending Battelle-PNL Technical Report, 4/92.
2. PNL-ALO-335, GC/MS Analysis of Volatile Organic Compounds, Rev. 0, Battelle-PNL Technical Procedure, 8-3-89.
3. QA Plan for Activities Conducted by the Analytical Chemistry Laboratory (ACL), QA Plan MCS-033, Rev. 0, 9-4-91
4. USEPA Contract Laboratory Program, Statement of Work for Organic Analysis, Multi-Media, Multi-Concentration, SOW No. 2-88.
5. USEPA Contract Laboratory Program, Statement of Work for Organic Analysis, Multi-Media, Multi-Concentration, Document Number OLM01.0, March 1990.
6. USEPA Office of Solid Waste, Test Methods for Evaluating Solid Waste, SW-846, Third Edition, 11-86.

DISTRIBUTION

No. of
Copies

No. of
Copies

OFFSITE

12 DOE Office of Scientific and
Technical Information

ONSITE

2 DOE Richland Field Office

J. M. Clark, A4-02
J. R. Noble-Dial, A4-02

11 Westinghouse Hanford Company

D. A. Dodd, T6-50
S. G. McKinney, T6-50
S. G. Metcalf, T6-50
A. F. Noonan, R2-12
L. M. Sasaki (5), R2-12
K. L. Silvers, T6-08
W. I. Winters, T6-50

17 Pacific Northwest Laboratory

J. A. Campbell, P8-08
E. W. Hoppe (5), P7-22
T. E. Jones, P7-27
R. B. Lucke, P8-08
G. A. Ross, P7-22
M. J. Steele, P7-53
R. W. Stromatt, P7-22
Publishing Coordination
Technical Report Files (5)

END

**DATE
FILMED**

6 / 29 / 92