**The quality control of oceanic carbon dioxide measurements:
preparation and distribution of reference materials.**

Grant No. DE-FG03-92ER61410

Progress report for period 1/1/93–12/31/93

Andrew G. Dickson

OVERALL GOALS

The principal goal of this project — together with a related project funded through the US National Science Foundation — is to provide a mechanism for quality control that will ensure that seagoing measurement of the oceanic carbon dioxide system made as a part of the JGOFS studies are comparable and accurate, although made by different laboratories at different times.

The strategy we have chosen to achieve this has three principal thrusts:

- The preparation, certification and distribution of reference materials that investigators can use to evaluate their shipboard data quality.
- The organization of collaborative studies to test various methods and to help the community in identifying improved analytical approaches.
- The preparation and distribution of well written Standard Operating Procedures describing the analysis of the oceanic carbon dioxide system
- Active participation in various committees and meetings to proselytize the need for a systematic approach to the quality control of oceanic chemical measurements.

ACHIEVEMENTS DURING THE PAST YEAR

Production, certification, and distribution of reference materials

Since October 1992, we have bottled about 3,000 bottles of reference material: five regular batches and one special batch for a collaborative study. These batches have all been certified for total dissolved inorganic carbon in Dr. Keeling's laboratory at SIO (Table 1).

Table 1. C_T certification for recent batches of RMs mean $\pm$ std. dev. (no. of analyses)	
Batch 15	2031.66 ± 1.03 (12) $\mu\text{mol}\cdot\text{kg}^{-1}$
Batch 16	2034.28 ± 1.00 (11) $\mu\text{mol}\cdot\text{kg}^{-1}$
Batch 17	2044.54 ± 0.58 (9) $\mu\text{mol}\cdot\text{kg}^{-1}$
Batch 18	2115.59 ± 1.09 (7) $\mu\text{mol}\cdot\text{kg}^{-1}$
Batch 19	2003.31 ± 0.94 (5) $\mu\text{mol}\cdot\text{kg}^{-1}$

MASTER

Figure 1. Control chart for the NOAA C_T measurements made on their JGOFS EqPac Spring Cruise.

These reference materials were distributed to 16 different laboratories; both to investigators involved with the US JGOFS program and to investigators in Australia, Britain, Canada, Germany, Holland, Iceland, Japan, Norway, Sweden, and Taiwan. A typical use of them has been to prepare quality control charts indicating the quality of the data that are being measured on board ship (Figure 1).

Organization of collaborative studies

In late 1992, we initiated the second stage of our collaborative study to develop a standard procedure for the measurement of total dissolved inorganic carbon by extraction and coulometry. We prepared a test suite consisting of 5 different levels of spanning the range 1,980–2,200 $\mu\text{mol}\cdot\text{kg}^{-1}$ and distributed these in duplicate to the various participating groups (19 in number) for analysis as blind samples. It was clear from the results that were returned (14 groups) that there was no indication of a systematic non-linearity for the data produced. The offset of various sets of data from the certified numbers — as measured in Dr. Keeling's laboratory by extraction / manometry — are shown in figure 2.

Again, as in the first part of this collaborative study, the mean value obtained by the total group of laboratories returning results was almost identical to that obtained by the reference procedure; this further engenders our confidence in the accuracy of our certifications. However, it is apparent that

Figure 2. Results obtained from various laboratories participating in the collaborative study expressed as differences from the “true” values.

individual laboratories data are often offset by systematic amounts from the “true” values.

We are presently urging the various laboratories whose data do not match consistently the certified values for these samples (and for normal batches of RMs) to investigate and rectify the cause(s) of the discrepancies rather than relying on “calibrating with the reference materials.”

Development of a technique for alkalinity certification

In recent weeks it has become clear that there are problems with the implementation of alkalinity measurements by a variety of groups around the US. Results obtained for our RMs are sometimes in reasonable agreement (within $4 \mu\text{mol}\cdot\text{kg}^{-1}$) between laboratories and sometimes disparate (differences of about $10 \mu\text{mol}\cdot\text{kg}^{-1}$). One approach that we have taken to resolve this is to offer to check peoples calibration of the hydrochloric acid used as titrant in this analysis. During the past year we have cross-checked the acid for 5 different groups involved with JGOFS measurements — both within the US, and also from the UK. Typically our calibration number is in good agreement with theirs (within 0.1 %). This has included three laboratories that are supported by the DOE.

Mr. George Anderson (in Dr. Dickson's laboratory) is working in conjunction with Mr. Tim Lueker (a student in Dr. Keeling's laboratory) to address the problem of certifying reference materials. Although we have an accurate coulometric back titration technique available, it is not suitable for direct application to our RMs because of the presence of the mercuric chloride that is used to sterilize the sea water. We are thus investigating using a transfer standard based on "tris" in NaCl to allow us to obtain accurate analyses using a closed cell titration technique.

SOMMA workshop (Seattle, Nov 1994)

A workshop was held for SOMMA users in Seattle, WA in early November. This was funded in part by SCOR and IGBP, as well as by the DOE. I acted as a co-convenor and chairman for the workshop which was organized by Marilyn Roberts of NOAA/PMEL. The workshop was a great success, with many of the participants going away believing that they had learned a great deal from it.

Preparation and distribution of the DOE Handbook of Methods

During 1993, I substantially revised the various chapters of the DOE Handbook of Methods for the Analysis of the Various Parameters of the Carbon Dioxide System. I have also worked together with various members of the DOE science team to add additional Standard Operating Procedures that were not included in the first edition. This work is still in progress, and I expect to produce a complete second edition by April 1994.

The first edition has almost become a "best-seller" (perhaps because of the lack of a price). To date we have distributed about 90 handbooks to scientists in 18 different countries. Almost all of these handbooks were requested by the recipient, not sent as "junk mail".

Goals for 1994

- Continue the preparation, certification and distribution of RMs for the measurement of total inorganic carbon.
- Complete development of the certification procedure for total alkalinity and start certifying our RM batches routinely.
- Complete the second (revised and extended) edition of the DOE Handbook of Methods.
- Plan and conduct a collaborative study for the measurement of total alkalinity in sea water. This will initially involve the various DOE investigators.

- Plan and conduct an international comparison of continuous underway $P(\text{CO}_2)$ measurement systems. This is being run under the auspices of the SCOR / IOC CO_2 Advisory Panel, and is scheduled to occur Jun 6-10, 1994 at the Scripps Institution of Oceanography.

COMMITTEES, SCIENTIFIC MEETINGS AND PUBLICATIONS

Relevant committee meetings and presentations

3/18/93 – 3/19/93	DOE CO_2 survey science team meeting, Miami, FL
6/14/93 – 6/16/93	JGOFS/IOC CO_2 Panel, Plymouth England
9/8/93 – 9/10/93	WHP-12 meeting, Woods Hole
10/14/93 – 10/15/93	NAS Committee on Oceanic Carbon, Washington, D. C.
11/1/93 – 11/2/93	SOMMA workshop, Seattle, WA
11/3/93 – 11/4/93	DOE CO_2 survey Science Team Meeting, Seattle WA
12/7/93 – 12/8/93	GESREM Meeting on RMs, Miami, FL

Scientific Meeting Presentations

9/93	"Quality control of oceanic CO_2 measurements." paper presented at the 4th CO_2 conference, Carqueiranne
10/93	"Quality control of oceanic CO_2 measurements." paper presented at the ALEX meeting, San Francisco

Publications partially supported by this grant

- A. G. Dickson (1992) CO_2 panel proposes comparison of methods for determining pCO_2 in sea water. *U. S. JGOFS News* 4(2), 6.
- A. G. Dickson (1993) The analytical chemistry of the oceanic carbon dioxide system. In *Proceedings of the International Symposium on Global Change (IGBP), Tokyo, Japan, March 27-29, 1992*. Secretariat of International Symposium on Global Change. pp. 183-192.
- A. G. Dickson (1993) pH buffers for sea water media based on the total hydrogen ion concentration scale. *Deep-Sea Res.* 40, 107-118.
- A. G. Dickson (1993) The measurement of sea water pH. *Marine Chemistry* 44, 131-142.

DISCLAIMER

This report was prepared as an account of work sponsored by an agency of the United States Government. Neither the United States Government nor any agency thereof, nor any of their employees, makes any warranty, express or implied, or assumes any legal liability or responsibility for the accuracy, completeness, or usefulness of any information, apparatus, product, or process disclosed, or represents that its use would not infringe privately owned rights. Reference herein to any specific commercial product, process, or service by trade name, trademark, manufacturer, or otherwise does not necessarily constitute or imply its endorsement, recommendation, or favoring by the United States Government or any agency thereof. The views and opinions of authors expressed herein do not necessarily state or reflect those of the United States Government or any agency thereof.

DATE

FILMED

5/12/94

END