CEBAF PR-94-003
May 1994

Beam Energy Absolute Measurement Using
K-Edge Absorption Spectrometers

Ivan P. Karabekov† and George R. Neil‡

*†Yerevan Physics Institute
2 Alikhanian Brothers st., Yerevan, 375036, Republic of Armenia*

*‡Continuous Electron Beam Accelerator Facility
12000 Jefferson Avenue
Newport News, VA 23606*

CONTINUOUS
ELECTRON
BEAM
ACCCELERATOR
FACILITY

SURA SOUTHEASTERN UNIVERSITIES RESEARCH ASSOCIATION

CEBAF
Newport News, Virginia

MASTER

DISTRIBUTION OF THIS DOCUMENT IS UNLIMITED

Copies available from:

Library
CEBAF
12000 Jefferson Avenue
Newport News
Virginia 23606

The Southeastern Universities Research Association (SURA) operates the Continuous Electron Beam Accelerator Facility for the United States Department of Energy under contract DE-AC05-84ER40150.

DISCLAIMER

This report was prepared as an account of work sponsored by the United States government. Neither the United States nor the United States Department of Energy, nor any of their employees, makes any warranty, express or implied, or assumes any legal liability or responsibility for the accuracy, completeness, or usefulness of any information, apparatus, product, or process disclosed, or represents that its use would not infringe privately owned rights. Reference herein to any specific commercial product, process, or service by trade name, mark, manufacturer, or otherwise, does not necessarily constitute or imply its endorsement, recommendation, or favoring by the United States government or any agency thereof. The views and opinions of authors expressed herein do not necessarily state or reflect those of the United States government or any agency thereof.

Beam Energy Absolute Measurement Using K-Edge Absorption Spectrometers*

Ivan P. Karabekov† and George R. Neil‡

†Yerevan Physics Institute
2 Alikhanian Brothers st., Yerevan, 375036, Republic of Armenia

‡Continuous Electron Beam Accelerator Facility
12000 Jefferson Avenue, Newport News, Virginia 23606

Abstract

A method is presented of absolute energy measurement with an accuracy of $\Delta E \sim 10^{-4} E_0$ by direct measurement of the bend angle in a high-precision magnetic dipole using two opposite-direction short (about 2 mm long) high-field-intensity magnets ($\bar{B}_{\text{dipole}} \ll B_{\text{short mag}}$) installed at each end and two K-edge absorption spectrometers. Using these spectrometers and the hard x-ray synchrotron radiation created by the short magnets, a bend angle of 4.5 arc deg for the CEBAF energy bandwidth can be measured with an accuracy of a few units of 10^{-6} rad, and the main sources of systematic errors are the absolute measurement of the field integral and the determination of the centroid of the synchrotron beam at a wavelength equal to the K-edge absorption of the chosen substance.

Introduction

The use of magnetic spectrometers to determine absolute beam energy is well known. The best performance of the method was the SLC energy measurement facility [1]. This facility measured the bending angle using two synchrotron radiation (SR) swaths created in two short kicker magnets having magnetic field perpendicular to the field direction in the main bending magnet and installed on its sides. The field integral $\int B dl$ in the main dipole was measured with an accuracy better than 10^{-4} and monitored by a special rotating coil or NMR probes [2].

Two phosphorescent screens equipped with a system of fiducial wires and CCD video cameras measured the distance between SR beams with an accuracy of $\leq 10^{-4}$.

The distance between the center of the main bending and the phosphorescent screen equaling 15 m was measured with an absolute accuracy of 1.5 mm.

The main source of errors limiting the accuracy of energy determination at the level of 5×10^{-4} is nonorthogonality of the vectors of the magnetic field strength in the kickers and in the main bending dipole [3].

In this paper a new method of bend angle measurement using the K-edge absorption technique is discussed. The application of the method to high-precision absolute energy measurement at CEBAF is presented.

* Work supported by U.S. Department of Energy under contract DE-AC05-84ER40150.

The method

The concept of the method is the following. A beam of electrons passes through a high-precision mapped and continuously monitored dipole magnet having a field average intensity $\overline{B}_b = \frac{1}{\Delta l} \int^{\Delta l} B dl$ and two short dipoles having opposite direction of field intensity $B_s \gg \overline{B}_b$ installed on its sides. The synchrotron radiation generated in the main dipole and short magnets have different spectral distributions defined by corresponding critical wavelengths λ_c and λ_{cs} , which are chosen to satisfy the condition $\lambda_{cs} \gg \lambda_c$. So these short magnets are working as magnetic targets generating wavelength λ_0 , which is practically not represented in the spectrum radiated from electrons in the main dipole.

The principle of the method is as follows. Two K-edge absorbers (see Figure 1) having input collimators of width h and installed on the SR beam produced by the first short magnet at distance L can be tuned up at a wavelength λ_0 between the upper and lower borders of K-edge absorption of a chosen element. Scanning synchrotron light by the input collimator one can find the higher intensity at chosen wavelength which is the centroid of the SR beam generated in the first short magnet. This procedure will lower the quantum fluctuation of output (after absorption) intensity, which is one of the main sources of systematic errors. Then one of the spectrometers will be installed on the second beam axis by moving in a direction perpendicular to the axis of the first SR beam. The same wavelength λ_0 with the same output intensity can be registered in the SR beam generated in the second short magnet by scanning this beam and rotating the second spectrometer around the axis of its first crystal. When the same point of the absorption curve is reached—i.e., when the output intensities $N_{\gamma_{out}}$ of both spectrometers are equal and the gradients $\Delta N_{\gamma_{out}}/\Delta\theta$ of both spectrometers are equal—the bend angle θ_b in the main dipole will be equal to the angle of rotation of the second spectrometer.

Figure 1. The scheme of absolute energy measurement with the K-edge spectrometers.

The accuracy of the measurement depends on the angle of SR beam collimation, h/L , and the absolute value of the intensity drop gradient $\Delta N_{\gamma\text{out}}/\Delta\theta$ between the upper and lower points of the absorption curve.

To exclude from θ_b the bend angle created by the short dipoles $\delta\theta$, these magnets have opposite directions of the vectors of magnetic field strength.

For the spectrometers, ordinary EXAFS (extended x-ray absorption fine structure) spectrometers can be used (see, for example, [4,5]). Double crystal monochromators which are in use for these spectrometers have an angular resolution better than 1 arc sec. The monochromator described in [4] consists of two non-dispersive Si(422) channel-cut crystals, whose rocking curve width is $\varepsilon \sim 2.6$ arc sec, installed in a dispersive geometry (see Figure 2) having an energy resolution $\Delta E_\gamma/E_{\gamma 0}$ better than 10^{-5} and permitting measurement of the angle of about 5.6 arc deg from $\text{CuK}\alpha_{1,2}$ to $\text{CuK}\beta$ with an accuracy $\delta\phi$ better than 0.6 arc sec. This accuracy was stable in several sweeps over a period of several hours.

Figure 2 High-resolution crystal monochromator for EXAFS spectrometer.

Using this type of spectrometer the bend angle ϕ of the electron beam in the main dipole (about 4.5 arc deg) will be measured with the accuracy $\delta\phi/\phi$ better than 4×10^{-5} .

The calculations

The angular resolution of the K-edge spectrometer depends on the absolute value of a gradient of the intensity drop within the upper and lower (λ_u and λ_l) borders of K-edge absorption:

$$G = \frac{N_\gamma l - N_\gamma u}{\Delta\phi}, \quad (1)$$

and the intensity fluctuation at the lower border of absorption $N_{fl} = \sqrt{N_\gamma l}$, at a given electron beam intensity, I_e , and can be determined as follows:

$$\delta\phi \times G = \sqrt{N_\gamma l}. \quad (2)$$

The gradient of intensity drop depends on a ratio between the intensities created at λ_0 in the short and main dipoles too, because of the large divergence angle of radiation in the main dipole. This ratio can be limited to a few units of 10^{-2} .

The magnetic field intensity in the short dipole B_s can be taken as 2 T and with a bend angle of 2×10^{-5} rad.

A typical absorption curve of Cu at its K edge with an energy resolution about 2 eV [5] having the ratio $N_\gamma u/N_\gamma l$ higher than 30 has a $\Delta\phi$ equal to 4.25×10^{-4} rad (see Figure 3). For the following calculations these parameters were used.

Figure 3. A typical absorption curve of Cu.

The intensity of the photon flux generated in the short magnet at the chosen working wavelength λ_0 was calculated using following formula [6]:

$$N_\gamma = A\gamma^2(\lambda_c/\lambda_0)^2(\Delta\lambda/\lambda_0)K_{2/3}^2(\lambda_c/2\lambda_0), \quad (3)$$

where $K_{2/3}(\lambda_c/2\lambda_0)$ is the modified Bessel function, and A is a proportionality coefficient to obtain the result in units of photons per second:

$$A = 3.461 \times 10^6 I[\text{mA}]\phi[\text{mrad}]\psi[\text{mrad}], \quad (4)$$

where ϕ and ψ are horizontal and vertical angles of SR beam collimation and

$$\lambda_c = \frac{7.12 \times 10^3}{\gamma^2 B}. \quad (5)$$

Here γ is Lorentz factor of the electron.

The ratio of the intensities generated at the wavelength λ_0 in the main bending and short magnets can be calculated as follows:

$$R = \left(\frac{\lambda_{c_b}}{\lambda_{c_s}}\right)^2 \times \frac{K_{2/3}^2(\lambda_{c_b}/2\lambda_0)}{K_{2/3}^2(\lambda_{c_s}/2\lambda_0)}. \quad (6)$$

For optimization of this parameter at all CEBAF energies different substances can be used.

The calculations which affirm the high resolution of the method for CEBAF were performed using the following parameters:

- * beam current: $10 \mu\text{A}$,
- * bend angle in main dipole: $8.17 \times 10^{-2} \text{ rad}$,
- * magnetic field intensity in short magnets: 2 T,
- * SR beam horizontal divergence from short magnet limited by collimator: $2 \times 10^{-5} \text{ rad}$,
- * SR beam vertical divergence limited by collimator: $1 \times 10^{-3} \text{ rad}$,
- * the working wavelengths, λ_0 , are K-edges of the following substances: Cu ($1.38059 \text{ \AA}$), for an accelerator beam energy from 1 and 2 GeV, Ge ($1.11658 \text{ \AA}$) for 3 GeV, and Mo ($0.61978 \text{ \AA}$) for 4 GeV,
- * spectral bandwidth, $\Delta\lambda/\lambda$: 2×10^{-4} , and
- * the angular bandwidth, $\Delta\phi$, of the intensity drop at the K-edges: $4.5 \times 10^{-5} \text{ rad}$ for copper, $2 \times 10^{-5} \text{ rad}$ for germanium, and $1 \times 10^{-5} \text{ rad}$ for molybdenum.

The results of the calculations are presented in Table 1.

Table 1. The Results of the Calculations

E_e (GeV)	$\bar{B}_b$ (T)	λ_{cb} $\text{\AA}$	λ_{ce} $\text{\AA}$	N_{γ_0} photons/sec	$N_{\gamma_0}/N_{\gamma_e}$	$\Delta N_{\gamma}/\Delta\theta$ phot. sec^{-1} , rad^{-1}
1.0	0.0907	205.0	9.3	1.385×10^4	$< 10^{-5}$	3.06×10^8
2.0	0.1814	25.62	2.3	2.532×10^6	$< 10^{-5}$	5.62×10^{10}
3.0	2.721	7.59	1.03	1.86×10^6	1.37×10^{-2}	9.3×10^{10}
4.0	0.3625	3.2	0.58	1.253×10^7	6.84×10^{-2}	1.253×10^{12}

To avoid the influence of harmonics λ_0/n on the gradient of intensity drop, total external reflection (TER) will be used. The spectrum transformation by silicon TER mirror taken from [7] is presented in Figure 4.

Figure 4. The synchrotron radiation spectrum transformation by silicon TER mirror.

According to Table 1 and formula (2) the angular resolution $\delta\phi$ is better than 1 arc sec, and the limitation of absolute energy determination has mainly equal contributions from the accuracy of the magnetic field mapping in the main dipole, i.e., about 10^{-4} .

The intensity at the lower border of K-edge absorption according to [5] is taken equal to $N_{\gamma_e}/30$.

For CEBAF Halls A and C one bending magnet in each beam transport line is used. Two short (2 mm long) magnets installed at both ends of the dipole could generate SR beams extracted through a beryllium window of a modified electron beam pipe into two EXAFS spectrometers.

Acknowledgment

We would like to thank Dr. J. Bisognano for very valuable discussions.

References

1. J. Kent, M. King, C. Von Zanthier, S. Watson, "Precision measurement of the SLS beam energy", SLAC-PUB-4922, 1989.
2. M. Levi, "Precision measurement of the SLC spectrometer magnets", SLAC-PUB-4654, 1989.
3. M. Levi and F. Rouse, "Precision synchrotron radiation detectors", SLAC-PUB-4921, 1989.
4. Helio Tolento and Ricardo Rodrigues, *Rev. Sci. Instrum.* 63(1) 946-949, January 1992.
5. Yaning Xie, Hai Yang, Qiao Shan, Shaojian Xia, and D. Xian, *Rev. Sci. Instrum.* 63(1), 956-957, January 1992.
6. G. K. Green, "Spectra and optics of synchrotron radiation", BNL-50522, April 1976.
7. I. P. Karabekov, D. L. Egikian, R. A. Mikaelian, and V. G. Bagdasarian, *Nucl. Instrum. Methods* A261(1987) 318-321.

DATE

FILMED

5/23/94

END