AIM

Association for Information and Image Management

1100 Wayne Avenue, Suite 1100
Silver Spring, Maryland 20910
301/587-8202

Centimeter

Inches

MANUFACTURED TO AIM STANDARDS
BY APPLIED IMAGE, INC.

1 of 1

✓ DOE/ET/15601--47
~~FE-15601-47~~
Dist.Cat.UC-90g

DIAGNOSTIC DEVELOPMENT AND SUPPORT OF MHD TEST FACILITIES

Technical Progress Report
for the period
October, November, December 1991

W. S. Shepard, Director
R. L. Cook, Deputy Director

Diagnostic Instrumentation and Analysis Laboratory
Mississippi State University
Mississippi State, Mississippi 39762

Prepared for the United States
Department of Energy
Under Contract No. DE-AC02-80ET-15601

MASTER 87B

DISTRIBUTION OF THIS DOCUMENT IS UNLIMITED

✓

This report was prepared as an account of work sponsored by the United States Government. Neither the United States nor the United States Department of Energy, nor any of their employees, makes any warranty, expressed or implied, or assumes any legal liability or responsibility for the accuracy, completeness, or usefulness of any information, apparatus, product, or process disclosed, or represents that its use would not infringe privately owned rights. Reference herein to any specific commercial product, process, or service by trade name, mark, manufacturer, or otherwise, does not necessarily constitute or imply its endorsement, recommendation, or favoring by the United States Government or any agency thereof. The views and opinions of authors expressed herein do not necessarily state or reflect those of the United States Government or any agency thereof.

TABLE OF CONTENTS

List of Figures	iv
List of Tables	vi
Nomenclature	vii
Abstract	1
Project Objectives	2
Project Description	5
Description of Facilities	26
Technical Progress	27
Task 1. Diagnostic Instrument Development	28
A. <i>Coherent Anti-Stokes Raman Spectroscopy System (CARS)</i>	28
B. <i>Particle Size Distribution System (PSD)</i>	31
C. <i>Potassium Emission/Absorption System (PE/AS)</i>	32
D. <i>Intrusive Multi-Probe System (IMPS)</i>	33
E. <i>Faraday Rotation System (FRS)</i>	35
F. <i>Multi-Purpose Imaging System (MPIS)</i>	36
G. <i>Differential Absorption Laser Spectroscopy System (DALAS)</i>	37
H. <i>Laser Optogalvanic Spectroscopy System (LOGS)</i>	39
I. <i>Cross Correlation System (CCS)</i>	41
Task 2. Test Stand Operations	42
A. <i>Test Stand Modifications</i>	42
B. <i>Test Stand Runs</i>	42
Task 3. Technical Support for National MHD Programs	43
A. <i>Field Measurements</i>	44
B. <i>Field Tests/Instrument Modifications</i>	68
C. <i>Field Operations/Developing Systems</i>	73
Task 4. Project Management	74
Task 5. Technology Transfer	75

LIST OF FIGURES

1.A.1	Fitting of the computer-simulated spectrum with the observed N ₂ -CO CARS spectrum at the CFFF diffuser 2" from the center (east side)	29
3.A.1	Doppler burst obtained from CFFF LMF5-D test run	44
3.A.2	Histogram of particle velocity with a mean velocity value of 868 m/s and root mean square velocity of 78 m/s	45
3.A.3	PE/AS measurements for the first two days during CFFF LMF5-D test run	50
3.A.4	Plot of temperature and potassium number density vs. time for data taken 18-Nov-91, LMF5-D	51
3.A.5	Plot of temperature and electron number density vs. time for data taken 18-Nov-91, LMF5-D	52
3.A.6	Plot of temperature and potassium number density vs. time for data taken 19-Nov-91, LMF5-D	53
3.A.7	Plot of temperature and electron number density vs. time for data taken 19-Nov-91, LMF5-D	54
3.A.8	Plot of temperature and potassium number density vs. time for data taken 20-Nov-91, LMF5-D	55
3.A.9	Plot of temperature and electron number density vs. time for data taken 20-Nov-91, LMF5-D	56
3.A.10	Average particle size (Sauter mean diameter, above) and electric field meter output (below) for data collected 19-Nov-91, LMF5-D	58
3.A.11	Average particle size (Sauter mean diameter, above) and electric field meter output (below) for data collected 19-Nov-91, LMF5-D	59

3.A.12	Average Sauter diameter measured downstream of the ESP, 20-Nov-91, LMF5-D	60
3.A.13	Additional data collected 20-Nov-91, LMF5-D	61
3.A.14	Average Sauter diameter measured downstream of the ESP, 21-Nov-91, LMF5-D	62
3.A.15	Additional data collected 21-Nov-91, LMF5-D	63
3.A.16	Average Sauter diameter measured prior to and during the water injection test 23-Nov-91, LMF5-D, with increase in water injection at 1732 hours to increase the gas stream humidity to approximately 1%	64
3.A.17	Average Sauter diameter measured prior to and during the water injection test 23-Nov-91, LMF5-D, with increase in water injection at 1744 hours to increase the gas stream humidity to approximately 2%	65
3.A.18	Average Sauter diameter measured prior to and during the water injection test 23-Nov-91, LMF5-D, with increase in water injection at 1754 hours to increase the gas stream humidity to approximately 3%	66
3.A.19	Average Sauter diameter measured prior to and during the water injection test 23-Nov-91, LMF5-D, excess water injection discontinued at 1804 hours	67
3.B.1	Fitting of the computer-simulated spectrum with the observed N ₂ -CO CARS spectrum at the CFFF diffuser 2" from the center (east side)	69

LIST OF TABLES

1.A.1	Comparison of inferred HCl CARS temperature with thermocouple temperatures	29
2.B.1	Test stand runs, October through December 1991	42
3.A.1	LMF5-D hourly average two-color temperatures and emissivities	47
3.A.2	LMF5-D hourly average two-color temperatures and emissivities	48
3.A.3	LMF5-D hourly average two-color temperatures and emissivities	49
3.B.1	N ₂ CARS temperature measurements at the CFFF diffuser exit, 20-Nov-91	68
3.B.2	Simultaneous N ₂ -CO measurements at the CFFF diffuser exit, 23-Nov-91	70
3.B.3	ADP surface temperatures, CFFF LMF5-D	72

NOMENCLATURE

CFFF	Coal-Fired Flow Facility at the University of Tennessee Space Institute, Tullahoma
CDIF	Component Development and Integration Facility, Butte, Montana
LMF5-C	Low Mass Flow test sequence at the CFFF
LMF5-D	Low Mass Flow Test sequence at the CFFF
MP	melting point
Nd:YAG	neodymium: yttrium aluminum garnet

ABSTRACT

The Diagnostic Instrumentation and Analysis Laboratory (DIAL) at Mississippi State University (MSU) is developing diagnostic instruments for magnetohydrodynamic (MHD) power train data acquisition and for support of MHD component development test facilities. Microprocessor-controlled optical instruments, initially developed for Heat Recovery/Seed Recovery (HRSR) support, are being refined, and new systems to measure temperatures and gas-seed-slag stream characteristics are being developed. To further data acquisition and analysis capabilities, the diagnostic systems are being interfaced with DIAL's computers. Technical support for the diagnostic needs of the national MHD research effort is being provided. DIAL personnel also cooperate with government agencies and private industries to improve the transformation of research and development results into processes, products and services applicable to their needs.

PROJECT OBJECTIVES

Diagnostic and special instrumentation beyond that which is normally required for appropriate control and operation of an electrical power production facility is required for an MHD power system. The Department of Energy's (DOE) coal-fired MHD proof-of-concept program serves as the precursor to larger scale demonstration facilities and commercially-implemented plants. In order that the transition to commercially feasible plant designs be made quickly and effectively, it is essential to determine as much as possible about the detailed component operation of the MHD power train and the HRSR system.

Conventional diagnostic sensors are of limited use in characterizing the MHD gas stream, and special diagnostic instrumentation systems are needed to provide the means to monitor and collect engineering data to determine specific parameters, both baseline values and fluctuations, and to provide background information for continuing MHD development, solutions to technological barriers, and a basis for future plant designs. Accurate characterization of the thermal, chemical and flow properties is essential for future design optimization.

The scale-up process is a nonlinear extrapolation because many of the physical processes are nonlinear. To decrease the uncertainties in this nonlinear extrapolation to a new, larger design, it is necessary to rely not only on engineering experience with smaller systems, but also on mathematical models which describe the physical processes and the manner in which they behave as design variables are changed. Mathematical models require the input of detailed data; the greater the quantity and quality of data, the more accurate the model and the smaller the uncertainty in the extrapolation to larger designs.

The baseline data obtained by diagnostic measurements at the Component Development and Integration Facility (CDIF) and Coal-Fired Flow Facility (CFFF) are providing important engineering experience and data for developing appropriate analytical models. Tests of these models, as well as information to further develop these design models, will be provided by diagnostic measurements on these facilities. These measurements are essential to minimize the technological risks associated with the ultimate construction of a retrofit MHD facility or a commercial MHD power plant.

Moreover, measurements of selected parameters, coupled with empirical correlations or an accurate model, can contribute significantly to improved control strategies and the selection of control instrumentation. The direct measurement of a particularly useful parameter for system control may be very difficult, if not

impossible. However, another parameter which can be correlated to the parameter of interest may be much easier to measure. The application of advanced instrumentation promises to produce fundamental advances in this area. This type of information will be very important in the selection of control instrumentation for a commercial MHD power plant.

MHD flows represent possibly the most severe environment encountered by gasdynamic diagnostics. Special state-of-the-art techniques and instrumentation systems are required to monitor and collect data for MHD component systems. In addition, these systems are operating under severe environmental and magnetic field conditions. The Diagnostic Instrumentation and Analysis Laboratory at Mississippi State University is developing and applying advanced optical diagnostic techniques and instrumentation systems which provide nonintrusive, remote, real-time measurements and are designed to operate successfully in the industrial-like environment of large-scale facilities. Such diagnostic instrumentation can provide the information required to more clearly characterize the various processes involved in the ash/seed-laden gas stream of the coal-fired MHD flow. These systems will provide performance monitors giving real-time indications of

- water leaks,
- nonuniformity in the potassium seed,
- flow train discontinuities in the channel,
- flow distributions and turbulence levels,
- combustor performance, stoichiometry,
- proper operating conditions to optimize efficiency
and component lifetime,
- the extent of ash rejection by the combustor,
- conductivity, generator performance,
- hot spots, e.g., in the radiant boiler,
- baghouse and/or electrostatic precipitator performance,
- heat transfer characteristics at critical locations,
- particle loading and size distribution entering the
boiler or other system components,
- stack emissions, pollutant levels,

and will, therefore, further reduce the technical and financial risks associated with scale-up to the retrofit size integrated MHD steam plant or to a commercial size plant.

The computer-controlled, nonintrusive, advanced optical diagnostic instrumentation systems that will be used to determine these parameters are listed in the next section along with the salient features of the systems. Direct instrumentation support

and measurements will be provided to the various DOE MHD test facilities. The amount of direct diagnostic support to the MHD engineering development test facilities will be as directed by DOE.

A number of the diagnostic systems which are fully developed will be used for field measurements at the HRSR and MHD power train facilities. Additional diagnostic systems previously developed are being constructed or modified to provide field-use instruments particularly for the MHD power train. Field tests for system refinements of these instruments will be conducted on the DIAL test stand and the various DOE MHD test facilities before proceeding with field measurements.

Some additional modifications are planned for existing instrumentation systems to expand their capabilities and range. A few additional diagnostic systems requiring only limited equipment are being developed where the method used can provide the basis for a simplified instrumentation system for MHD plant control or component monitoring. In most cases, these systems will share equipment with other existing systems.

DIAL has constructed a test stand that can simulate the gas-slag-seed stream composition, temperature, and metal/fireside environmental conditions for the radiant furnace, superheater, and other MHD system components. The computer-controlled diagnostic instruments being developed are evaluated and tested on this test stand. Measurements conducted on the test stand also provide useful data for the national MHD program.

PROJECT DESCRIPTION

In order to meet the overall contract objectives, the scope of work to be performed is outlined by the following tasks.

Task 1. Diagnostic Instrument Development R. L. Cook, Section Manager

The following computer-controlled, optical diagnostic instrumentation is being developed for support of the MHD component engineering development. Objectives for the specific instrumentation systems follow.

1.A Coherent Anti-Stokes Raman Spectroscopy System (CARS) *(local gas temperature and species concentration, temperature and concentration profiles)*

The CARS system is a laser-based, nonlinear, optical technique being developed to provide temperature and species concentration measurements with spatial and temporal resolution. Of the many laser-based techniques, CARS is the best suited for thermometry in high-interference environments because of its coherent, laser-like signal character and high signal conversion efficiency. To obtain a CARS spectrum, two narrow band lasers at pump frequency ω_1 derived from the pump laser and one broadband dye laser at Stokes frequency ω_2 are phase-matched and focused in the gas stream thereby generating, through the third-order susceptibility of the medium, a spectrum at anti-Stokes frequency $\omega_3 = 2\omega_1 - \omega_2$. The CARS spectrum of a species (e.g. nitrogen) can be sensitive to both gas temperature and species concentration.

The mobile CARS instrument is being used for field measurements and additional field tests. Based on these tests, a few minor modifications may be required. Field measurements are planned throughout the contract period at both CFFF and CDIF. Point temperatures and temperature profiles from N_2 CARS spectra will be measured. Work on improving the capabilities of the present CARS system will also continue. This will include automation of the mobile CARS system. An optical fiber coupling system will be designed to couple simultaneously the CARS signal and dye laser with the spectrograph. The optical multichannel analyzer (OSMA) will be modified to record the single pulse CARS spectra, which will provide the histogram of the temperature. In addition to temperature measurements, field measurements of species concentration will also be made. For example, temperature and CO

concentration can be measured simultaneously using the multiplexed CARS spectra of N_2 and CO. The combined N_2 -CO CARS spectrum is then fit to a computer-simulated CARS model to extract the temperature and CO concentration. Work will continue on simultaneous N_2 -CO CARS to determine the detection limits of this technique for CO concentration measurements in a particle-loaded gas stream.

N_2 CARS and simultaneous N_2 -CO CARS measurements will be performed at the CFFF aerodynamic duct, which has an operating pressure above one atmosphere. To infer the temperature and CO concentration from the CARS spectra, a CARS computer model will be modified to simulate the CARS spectrum at high pressure.

1.B Particle Size Distribution System (PSD) (particle size distribution)

The PSD system uses optical, in-situ, scattering methods for measurement of particle size distribution. The three-beam, two-color particle size measurement system is designed to measure particles at slow velocity using a single-particle counting method, while an ensemble averaging technique is used for high velocity flows. The system uses the output (green and blue) from an argon ion laser. The green beam is focused on the central axis and the blue beam is divided into two equal intensity beams which are focused and crossed inside the green measurement volume producing an interference fringe pattern. The light scattered from these beams by a particle is processed to extract the particle size.

In the case of the single particle counter (SPC) instrument, a particle passing through a focused laser beam sample volume scatters a Gaussian-shaped pulse of light. The peak amplitude, detected at small angles in the near-forward scatter direction, is essentially a monotonic function of particle size and virtually independent of the particle refractive index and shape. The scattering signature is, however, dependent on the local sample volume intensity and hence, the particle trajectory through the sample volume. To eliminate the nonuniformity of response within the sample volume, three operating modes are used: the ripple-validated small-angle near-forward scatter (RVSANFS), the ripple deconvolution method (RDM), and the small-angle near-forward scatter deconvolution (ALL-SANFS). These three modes are individually optimized to measure particle size distribution under various combinations of particle size range, number density, and velocity.

To reduce the trajectory ambiguity in the RVSANFS method, the single particle scattering signal from the central green beam is used to measure large

particles (2-15 μm) with the ripple signal from the crossed blue beams used to validate the trajectory of the particles and, hence, discriminate particles not near the central region of the uniform green illumination.

On the other hand, for the RDM, the amplitude of the ripple signal from the crossed blue beams is used to size particles smaller than 5 μm . The spatial variation of the illumination intensity within the measurement volume is measured. The scattering response of the combined illuminating optics and scattered light receiver is then determined. A mathematical inversion technique (deconvolution) is used to extract the particle size distribution from the unvalidated count histogram of pulse height amplitudes. Thus, the ambiguity caused by the trajectory-dependent response is removed. This technique is applied to flow/particulate conditions with a velocity less than 100 m/s and particle number density less than $10^6 \text{ \#/cm}^3$.

For conditions of flow velocity less than 200 m/s and number densities less than $10^5 \text{ \#/cm}^3$, the ALL-SANFS method is available. The ALL-SANFS method utilizes scattering from the green beam to determine the size of a particle. The intensity response deconvolution of the unvalidated count histogram is also used with this technique to resolve the trajectory-dependent response ambiguity.

For conditions of higher velocity and large number density, a single particle counting type instrument is unable to resolve individual particles. The ensemble averaging method (EAM) is being developed for these situations. This technique does not attempt to resolve the scattering signal from individual particles thus is not limited by requiring high speed electronics with a large dynamic range. Scattered light is measured at several different angles in the forward-scattering direction. A mathematical inversion technique is then used to determine the particle distribution for the scattering signature.

Expected activities include laboratory experimentation, operation on the DIAL test stand, and operation in the field. Laboratory experimentation will be related to development of the EAM instrument and continual upgrading of the SPC instrument. It will also include development of calibration techniques and the production and evaluation of aerosols. Operation on the DIAL test stand will primarily be done for development of the EAM instrument. Operation in the field will involve the SPC instrument until the EAM instrument is refined for field operation.

The overall objective is to improve the capability for measurement of particle size distributions by providing several alternate approaches that are feasible in various situations. Realization of this objective requires continual efforts to establish credibility that instruments are accurate and measurements are meaningful. This can

only be achieved by continual attention to calibration methods and uses of alternate measurements, where possible, to supply corroborative data.

The particular objectives for the SPC are to finish construction of the system, to improve the stability of fiber optic transmission system, to prove operation of the system in the field, and to refine the system for convenience in use and performance. Receiving fibers will be added to reduce the susceptibility to noise. Development of a digital signal processor which will improve the accuracy, speed, and reliability will be started. This processor will use a DSP chip and software to discriminate ripple burst signals from the background signals due to noise and to particles below the operating size range for a particular instrument setup. Data will also be gathered for comparison of different modes of operation in the field environment; and, field measurements, to characterize the downstream components at the CFFF, will be conducted.

The specific objectives for the EAM are to prove the feasibility of the concept through experimental tests and incorporate this method into the field instrument. This will require continued theoretical and computational efforts to evaluate the parameters for the data reduction model; preliminary work on this is finished. Refinements will be needed as soon as the work is far enough along to justify them; these include the effects of lens aberrations, which will also require some experimental investigation. Experiments will be conducted as soon as feasible on the test stand and, concurrently, field tests of the method will be made to define a suitable instrument package.

1.C. Potassium Emission/Absorption System (PE/AS)

(time—resolved temperature, K—atom density, electron density)

The PE/AS system will rapidly measure (order of milliseconds) the average channel temperature by the line reversal technique, provide the average neutral potassium number density from lineshape information, and provide the average channel electron number density using the Saha equation assuming thermodynamic equilibrium.

This system has been field-tested and is presently being used to provide time-resolved temperature and potassium number density measurements in support of the national MHD effort. Some software enhancement of the instrument is planned; and, a major issue has been resolved, i.e., the over-estimation of the potassium number density. This is particularly important for wide gas streams and is primarily caused by the line profile deviating from a Voigt shape in the far wings. A new lineshape model has been developed and will be tested further in the field.

1.D. Intrusive Multi-Probe System (IMPS)
(optical temperature probes—wall and gas temperature)

An IMPS system has been developed to position various sensors at a precise location in a combustion gas stream for precise increments of time. It is completely computer-controlled and may be operated to both position the probe and acquire the data. Optical sensors, as well as conventional sensors, can be used with this probe insertion system; however, its main use is with techniques where the probe will spend a very short period of time in the gas stream. Such a system interfaced with an optical technique can provide measurement capabilities in regions where, because of path length, optical measurements would not be possible.

Development and field testing of one viable intrusive sensor has been completed for the measurement of MHD interior wall surface temperatures. This sensor is capable of measuring surface temperatures in the range of 500 to 2000 C. Field tests have been conducted at the DOE/CFFF by measuring the front surface temperatures of the superheater tubes during several test runs; the interior wall surface temperature of the furnace and the surface temperature of the ash deposition probe have also been measured.

The development of an intrusive sodium line reversal (SLR) sensor, to be used with the multi-probe systems for measuring local gas stream temperatures, is about 60 percent complete. This system uses an intrusive miniature optics system and the SLR principle to make spatially-resolved gas stream temperature measurements. The SLR sensor will provide gas stream temperature measurements from 1700 to approximately 2500 K.

Because surface temperature measurements would usually be required at distances greater than five feet in the MHD power system, an optic system is being developed to add to the existing lightpipe sensor to provide greater intrusive distances. In general, intrusive distances much greater than five feet are not possible with this sensor when accessing surfaces through the standard four-inch ports. This optics/lightpipe sensor system will provide added flexibility and a useful monitor for the larger-scale MHD development systems.

With the equipment on hand the possibility of developing an intrusive, two-sensor system for measuring gas stream temperatures and wall surface temperatures appears to be very viable. This system would provide measurements of gas stream and wall surface temperatures in the range of 500 to 2000 C and would be

especially important for surface and gas temperature measurements in the region of the superheater tubes in large-scale MHD development systems.

1.E Faraday Rotation System (FRS)
(electron density to infer conductivity)

The FRS system will provide real-time electron density measurements averaged across an MHD channel. The electrical conductivity scales very nearly linearly with the electron density. The technique is based on the Faraday effect. Linearly polarized light may be decomposed into right (+) and left (-) circularly polarized waves. Since the absorption coefficients and refractive index are different for right and left polarization, the amplitude and relative phase of the polarized waves are affected on passing through a plasma along a uniform magnetic field. Measurement of the rotation of the polarization of a beam of linearly polarized light after passing through a medium (plasma) in the direction of an applied magnetic field determines the electron number density. The system employs radiation in the far infrared derived from a CO₂-pumped far infrared laser.

The FRS system will provide, directly, the electron density and conductivity of the MHD channel which is essential to the characterization of the power train performance. The major equipment for the system has been obtained and assembled. Measurements of electron density and conductivity on a laboratory plasma are basically complete. Electron mobility modeling and measurements will be addressed and system refinements, such as waveguides and detector vibration isolation for field tests, will be carried out. Field measurements can be made when requested, however, optical access through the channel side walls is needed.

1.F Multi-Purpose Imaging System (MPIS)
(K-atom density, pressure profile)

The MPIS system is a laser-excited imaging system suitable for spatially-resolved optical diagnostics. The system is based on the phenomena of laser-induced fluorescence. Here an atomic or molecular species is raised to an excited electronic state by absorption of laser radiation. The emitted radiation when the species returns to the ground state is termed fluorescence. The instrument consists of a laser whose beam is expanded with cylindrical optics into a wide, thin sheet that is directed through the region of interest. A camera lens is then used to focus light emitted by this region onto a charge injection detector,

the signal from which is processed by a computer to produce a two dimensional image of the probed region. The versatility of the system comes from the ability to select both the excitation and observation wavelengths to determine various parameters of interest, such as the concentration mapping of selected atomic and molecular species.

The MPIS is an instrument currently in development which will allow measurements of species (K, Na, OH) concentration profiles in harsh combustion environments. The excitation intensity must be tuned to a resonance absorption with sufficient power that the fluorescence intensity is proportional to species concentration. In addition, measurements of number density and pressure profiles are possible, as well as temperature profiles using two-wavelength, excited-state fluorescence measurements. In the case of pressure determinations, the excitation intensity must be above that required for level saturation; then the fluorescence intensity is proportional to the pressure.

Previous work has involved the construction of a laboratory-scale system and subsequent measurements of Na concentration images and K laser-induced fluorescence (LIF). The latter measurements have been complicated by collisional quenching, a process whereby the excited states decay non-radiatively. For example, K LIF at 404.5 nm has been observed but at a significantly reduced quantum efficiency as compared to Na. Saturation of the 404.5-nm K line is not possible without increasing the intensity level of the incident excitation wavelength. Addition of a wavelength extension unit (WEX) to the current optical train is planned since this will drastically increase the incident laser power by mixing the 1.06-mm fundamental of the Nd:YAG laser with the low efficiency output of the dye laser in an appropriate crystal. Concentration measurements on K can then proceed. The determination of pressure profiles will begin with the construction of a variable pressure burner which will permit measurements as a function of known pressure. LIF measurements will then be performed to determine the limitations of the technique as applied to harsh environments.

Determination of the fluorescence at two wavelengths will allow measurement of the populations of the excited species which, in turn, can be used to determine spatial temperature profiles of a region. In order to accomplish this task, an additional detector (camera) will be incorporated with an extended blue wavelength response. This detector will also permit the evaluation of OH fluorescence. Measurements on the MHD power train will require optical access. Access can be expected to be a problem, especially for image capture; therefore, a flexible, image transmitting fiber bundle will be added allowing the fluorescent or scattered laser light image to be

routed to the camera which will be located away from the measurement penetration and clear of the upstream plumbing and structure.

The limitations and attributes of the system with respect to species concentration, pressure gradients, two wavelength temperature and number density measurements will be established and, concurrently, field tests with the system will be conducted. Results from these initial tests will be used to construct a stable, reliable field version of the instrument.

1.G Differential Absorption Laser Spectroscopy System (DALAS) (species concentration, SO_2 , NO)

The DALAS system is a general purpose, nonintrusive instrument for measuring time-resolved, line-of-sight average concentrations of atomic and molecular species. The system uses a two-wavelength dye laser and measures the extinction of the two wavelengths through the gas stream. One wavelength is on- and the other off-resonance of the species of interest thereby eliminating particle absorption and scattering effects.

A general purpose system for measuring average concentrations of individual species (e.g. H_2O , SO_2 , NO_2 , NO, and OH) is desirable to complement the CARS system, which does not measure the concentrations of these species. DALAS is a general purpose technique which can, in principle, be applied to any atomic or molecular species; and, can be brought on-line by additions to the existing mobile CARS instrumentation. DALAS can also provide time-resolved concentration measurements when concentration fluctuations are of interest. DALAS uses methodology and technology developed over the past decade for Differential Absorption Lidar, a mobile, laser-based remote sensing technique for measuring the concentrations of pollutants in the atmosphere.

A DALAS experiment is performed by comparing the transmission of two wavelengths through the sample. One wavelength corresponds to a strong absorption of the species of interest. The other wavelength is one that the species does not (or only weakly) absorbs. Taking the ratio of the transmitted intensities of the two wavelengths cancels out background effects, such as light scattering due to particulates, that are not associated with absorption by the species. If the absorption coefficient of the species is accurately known, or appropriate calibration experiments have been made, then the absolute concentration of the species can be determined.

The DALAS apparatus will use a single laser system to generate both wavelengths. An alternating two-wavelength dye laser pumped by the CARS Nd:YAG laser (10 Hz) will generate on alternate laser pulses the on-resonance wavelength (5 Hz) and the off-resonance wavelength (5 Hz). In particular, a rotating, rectangular optical block is used to refract the dye laser beam to independent wavelength-selection systems, thereby generating two closely spaced wavelengths. The successful operation of this dye laser is not sensitive to the alignment of the dye laser beam and the block; this ruggedness is important for field measurements. The visible dye laser beams can be frequency-doubled into the ultraviolet by sending them through two frequency doubling crystals--one crystal for each wavelength. Both wavelengths follow the same path through the plasma onto a single photodiode. The output of the photodiode for on-resonance laser pulses is fed to one boxcar averager; and, the output of the photodiode for off-resonance pulses is fed to a second boxcar averager. The ratio of the intensity outputs of the boxcar averagers enables calculation of the concentration of the species.

DALAS can, in principle, be applied to any atomic or molecular species. The focus will be on measuring the concentrations of SO₂ and NO, because the concentrations of these species, though important, are not measured by the CARS system. SO₂ and NO are of interest because they are pollutants and thus their environmental impact is important.

To date, a grazing incidence dye laser has been constructed; conversion of the dye laser to two-wavelength operation is in progress. Construction of an optical cell/vacuum system for bench-top studies of NO and SO₂ has been completed and vacuum testing has begun. Design of the photodetector/switching system is well advanced. Frequency doubling crystals, two boxcar averagers, and a data acquisition/control computer have been acquired.

The methodology for observing and recording DALAS signals will be validated. We will optimize the instrumental operating parameters for SO₂ and NO in bench-top flames and the DIAL test stand. The system will be field tested prior to field measurements.

1.H Laser Optogalvanic Spectroscopy System (LOGS) *(average gas temperature, qualitative species identification)*

The LOGS system is a general purpose instrument for measuring line-of-sight average temperatures and will provide qualitative identification of various atomic and molecular species. The system uses a tungsten electrode which is inserted into the gas stream to monitor

changes in conductivity as a dye laser is tuned into resonance with a particular species.

A general purpose system for measuring average temperature is desirable, especially one capable of working even when high particulate flow severely limits transmission of optical beams across the MHD flow. LOGS is such a general purpose technique and can be brought on-line by additions to existing instrumentation. Essentially, a frequency doubling unit with automatic scanning capability is all that is required. Because LOGS uses electrical rather than optical detection, the technique alleviates problems associated with monitoring small absorptions or weak fluorescence in the presence of a strong optical background signal or large path lengths. Hence, LOGS can identify species present and characterize their temperature under conditions where techniques employing optical detection fail. Moreover, only optical access on one side of the gas stream is required for this system.

In LOGS a pulsed dye laser is tuned into resonance with a transition of a species in the flame; the excited state concentration of that species then temporarily increases. Because the energy necessary for ionization is less for an excited electronic state than for the ground state, the rate of ionization temporarily increases. This process can be monitored as a transient (10^{-5} sec) voltage change (in lab flames, typically hundredths of a volt) if a high voltage electrode (e.g. tungsten, MP 3410 C) is inserted into the flame. Although most of the research on LOGS has concentrated on atomic species, thirty-one molecular species (such as NO_2 , HCO , N_2 , CO , CH , OH , CO_2 and H_2) have been studied to date. Efforts will concentrate on molecular species present in combustion environments, in particular, on OH and C_2 . By scanning the wavelength of a tunable dye laser and recording the optogalvanic signal as a function of laser wavelength, the optogalvanic spectrum (which is analogous to an absorption spectrum) is obtained. The temperature of the species can be ascertained by the standard technique of ratioing the intensities of different transitions and taking into account their different transition probabilities.

To date, the effects of electrode aging and fouling in clean bench-top flames has been studied. Continuation of these studies to particle-laden bench-top flames and in the DIAL test stand are now in progress. The instrumentation required for observation of molecular LOGS signals has been acquired. The instrumental operating parameters will be optimized. The precision of the temperature measurements in MHD environments will be established by comparing temperatures determined by LOGS with that obtained from CARS.

1.1 Cross Correlation System (CCS) (flow velocity)

The CCS system will measure the flow velocity nonintrusively with only limited optical access. The technique involves measuring the fluctuations in gas luminosity, e.g., at two spatially-separated locations. The cross-correlation function of these signals gives the time interval between which the signal and a time-displaced version of itself correlates.

A simple, rugged, accurate and nonintrusive system to measure the gas flow velocity would be a particularly valuable device to help monitor the MHD power train. The cross correlation technique can provide such an instrument and can be used to find the mean velocity of the gas flow. The method can find application where only limited access to the gas stream is available. The technique is based on the naturally occurring fluctuations in, e.g., species concentration, temperature, and particle concentration in a turbulent combustion environment. If a fluctuating variable is measured at two known locations--one upstream and the other downstream--then the time required for the fluctuations to travel from one location to another provides the flow velocity.

Preliminary work performed by DIAL indicates that the cross correlation method can be implemented using gas luminosity as the fluctuating variable. We also believe that the correlation of signals obtained by measuring the extinction of laser light at two or more locations can provide the flow velocity. This method would allow velocity measurements where gas luminosity is weak. Other extensions of this basic concept, employing intersecting beams, have the potential for measuring spatially-resolved velocity profiles.

A cross correlation experiment is performed by using two identical optical systems for data collection, with one located upstream and the other downstream. Each optical system consists of a focusing lens, a small monochromator, and a photomultiplier tube. The monochromators allow any wavelength of light from 200 to 800 nm to be selected. This spans the range from the ultraviolet to visible to the near infrared. The electrical signals from the photomultiplier tubes are amplified, low pass filtered, and sampled (both signals simultaneously) by an analog to a digital converter connected to a computer.

The data analysis consists of computing the cross correlation function of the two digitized signals. If the signals are correlated, i.e., only shifted in time by an amount, t , then the cross correlation function will show a peak at t . In this way, by

computing the cross-correlation function of the two photomultiplier signals, the time delay required for intensity fluctuations to propagate downstream can be found and thus the flow velocity can be determined.

Task 2. Test Stand Operations
J. A. Etheridge, Section Manager

DIAL has a computer-controlled, combustion test stand to simulate the combustion and thermal parameters present at various locations in a fossil-fueled MHD combustion system. Its versatility allows it to be used to simulate any type of combustion condition and effluent gas stream.

The existing DIAL/MHD/HRSR simulation test stand will be used for support of diagnostic instrumentation development, shake-down and evaluation. The test stand has recently been modified to better simulate the MHD gas stream environment. This will further aid in the evaluation of the instruments before they are taken to DOE's MHD engineering development test facilities. In particular, the test stand has been modified to burn a fuel oil/coal slurry which will more accurately simulate a coal fuel MHD combustor. Provisions for an increased flow rate of fuel-air mixture of approximately 700 lb/hr have been made.

The test stand control and data acquisition computer has also been replaced with a new system to enhance data collection and control as well as provide much needed memory and disk space. As the need arises the test stand may also be used to gather data systematically, under controlled conditions, to provide validation data for analytical models.

Task 3. Technical Support for the National MHD Program
R. D. Benton, Section Manager

The primary objective of this task is to provide diagnostic measurements and support to the national MHD program. The task will include attendance at contractor review meetings and contributions to technical progress reports. It will also include the preparation of a Topical Report for each diagnostic system at the time when it is available for field measurements. Objectives for this task are:

3.A *Field Measurements*

3.A.1 Sodium Line Reversal System (SLR)
(average gas temperature)

The SLR system essentially measures the temperature of the central region of the gas stream. The technique uses a calibrated light source and is based on making a series of intensity measurements on the wing of one of the sodium D-lines.

This system is field ready and has been used at various MHD facilities in support of the national program. No additional modifications are planned except to maintain the system field ready.

3.A.2 Multi-Color Pyrometer System (MCP)
(wall temperature and emissivity)

The MCP measures the wall temperature and emissivity of the wall surface. The technique is based on Planck's radiation law and requires measurement of the radiation intensity from the wall surface at two or more wavelengths, assuming the presence of a graybody radiator.

The system is field ready and has been used at various MHD facilities in support of the national program. No additional modifications are planned except to maintain the system field ready.

3.A.3 Two-Color Laser Transmissometer (TCLT)
(average particle size and particle number density)

The TCLT measures the average particle size and particle loading in the gas stream. The technique is based on measurement of the

extinction of laser light at two wavelengths (e.g., the IR and V regions) and knowledge of the refraction index of the particles.

This system is field ready and the TCLT units will continue the determination of average particle diameters and loadings at various DOE MHD facilities. Previous work has concerned the study of K_2SO_4 nucleation and real time evaluation of the performance of the CFFF electrostatic precipitator (ESP) and baghouse. Studies on the CFFF particulate collection components will continue with efforts centered on the performance characteristics of the ESP and the to-be-installed wet ESP. Measurements at other locations of the CFFF facility will be carried out as requested by facility engineers. Such experiences would be anticipated to evaluate heat transfer and the extent of ash carryover.

3.A.4 Laser Doppler Velocimeter System (LDV)

(local velocity, velocity profile and turbulence level)

The LDV measures the gas velocity and turbulence level at a given point in the gas stream. The technique is based on measurement of the modulation frequency of the scattered light from a particle traversing a measurement volume which is formed by intersection of two focused laser beams. A one-color and a two-color system are available which measure one and two components of the velocity, respectively. A velocity profile can be obtained and both forward and back scatter measurements can be made.

The LDV system is field proven and has provided a number of useful velocity flow measurements in support of the national MHD program (at CFFF, CDIF, TRW). To improve the operation and information derivable from the system, a second pair of transmitting and receiving optical fibers will be incorporated in the LDV optical train. This will allow two colors (green and blue) of incident laser light to be simultaneously transmitted to the measurement site and thus, two components (axial and radial) of particle velocity can be determined. The use of optical fibers is absolutely mandatory for field measurements, especially for difficult measurement sites such as near the combustor exits at CDIF and CFFF.

A second signal processor (Aerometrics DSA) has recently been incorporated into the LDV system. This unit has high frequency capability and will be particularly suited for measuring high speed flows (≥ 1000 m/s).

Recent experience at CDIF and CFFF proved that Doppler frequencies in the range of 50 to 70 MHz are to be expected for such flows. The only way to be

confident that a proper LDV signal is being processed is to capture the transient Doppler burst using a storage oscilloscope. It is desirable, therefore, to add a high speed digital storage oscilloscope to use in such applications. This will give a permanent record of the actual Doppler bursts and will allow post-test processing of the data to confirm the results from the signal processors.

3.A.5 Potassium Emission/Absorption System (PE/AS) (time-resolved temperature)

The PE/AS system rapidly measures (order of milliseconds) the average channel temperature by the line reversal technique.

This system has been field tested and is presently being used to provide time-resolved temperature measurements in support of the national MHD effort. Some software enhancement of the instrument is planned.

3.A.6 Gas Analysis System (GAS) (gas composition, e.g., CO, CO₂, NO, etc.)

The GAS system is capable of analyzing gases from fossil fuel combustion processes including NO_x, CO_x, SO₂, O₂ and combustibles. The system requires continuous extraction of a gas sample for analysis. The gas chromatograph (GC) can detect CO, CO₂, N₂ and O₂ using a thermal conductivity detector. Other gases of interest can also be monitored with the GC by selection of the appropriate column. Samples to be analyzed by the gas chromatograph are collected and stored in glass sample bulbs for later analysis.

This system is field ready and no additional modifications are planned except to maintain the system field ready.

3.B Field Tests/Instrument Modifications

3.B.1 Coherent Anti-Stokes Raman Spectroscopy System (CARS) (local gas temperature and species concentration, temperature and concentration profiles)

The CARS system is a laser-based nonlinear optical technique being developed to provide temperature and species concentration measurements with spatial and temporal resolution. Of the many laser-based techniques, CARS is the best suited for thermometry in high-interfer-

ence environments because of its coherent, laser-like signal character and high signal conversion efficiency. To obtain a CARS spectrum, two narrow band lasers at pump frequency derived from the pump laser and one broadband dye laser at Stokes frequency are phase-matched and focused in the gas stream thereby generating through the third-order susceptibility of the medium a spectrum at anti-Stokes frequency $\omega_3 = 2\omega_1 - \omega_2$. The CARS spectrum of a species (e.g., nitrogen) can be sensitive to both gas temperature and species concentration.

The mobile CARS instrument is ready for field measurements and additional tests. Based on these tests, a few minor modifications may be required. Field measurements are planned throughout the contract period at both CFFF and CDIF. Point temperatures and temperature profiles from N₂ CARS spectra will be measured. Work on improving the capabilities of the present CARS system will also continue. This will include automation of the mobile CARS system. In addition to temperature measurements, field measurements of species concentration will also be made.

3.B.2 Particle Size Distribution System (PSD) (particle size distribution)

The PSD system uses optical, in-situ, scattering methods for measurement of particle size distribution. The three-beam, two-color particle size measurement system is designed to measure particles at slow velocity using a single-particle counting method, while an ensemble averaging technique is used for high velocity flows. The system uses the output (green and blue) from an argon ion laser. The green beam is focused on the central axis and the blue beam is divided into two equal intensity beams which are focused and crossed inside the green measurement volume producing an interference fringe pattern. The light scattered from these beams by a particle is processed to extract the particle size.

Operation in the field will involve the single particle counter (SPC) instrument until the ensemble averaging method (EAM) instrument is refined for field operation. The particular objectives for the SPC are to finish construction of the system, prove operation of the system in the field, and refine the system for convenience in use and performance. We anticipate adding receiving fibers to reduce susceptibility to noise. Data will also be gathered for comparison of different modes of operation in the field environment, and field measurements to characterize the downstream components at UTSI will be conducted.

The specific objectives for the EAM instrument are to prove the feasibility of the concept through experimental tests and incorporate this method into the field instrument. This will require continued theoretical and computational efforts to evaluate the parameters for the data reduction model. Preliminary work on this is finished. Refinements will be needed as soon as the work is far enough along to justify them; these include the effects of lens aberrations, which will also require some experimental investigation. Experiments will be conducted on the test stand (as soon as feasible) and, concurrently, field tests of the method will be used to define a suitable instrument package.

3.B.3 Intrusive Multi-Probe System (IMPS) (wall and gas temperature)

An IMPS system has been developed to position various sensors at a precise location in a combustion gas stream for precise increments of time. It is completely computer-controlled and may be operated to both position the probe and acquire the data. Optical sensors as well as conventional sensors can be used with this probe insertion system; however, its main use is with techniques where the probe will spend a very short period of time in the gas stream. Such a system interfaced with an optical technique can provide measurement capabilities in regions where, because of path length, optical measurements would not be possible.

Development and field testing of one viable intrusive sensor has been completed for the measurement of MHD interior wall surfaces. This sensor is capable of measuring surface temperatures in the range of 500 to 2000 C. Field tests have been conducted at the DOE/CFFF by measuring the front surface temperatures of the superheater tubes during several facility test runs; the interior wall surface temperature of the furnace and the surface temperature of the ash deposition probe have also been measured.

3.B.4 Faraday Rotation System (FRS) (electron density to infer conductivity)

The FRS system will provide real-time electron density measurements averaged across an MHD channel. The electrical conductivity scales almost linearly with the electron density. The technique is based on the Faraday effect. Linearly polarized light may be decomposed into right (+) and left (-) circularly polarized waves. Since the absorption coefficients and refractive index are different for right and left

polarization, the amplitude and relative phase of the polarized waves are affected on passing through a plasma along a uniform magnetic field. Measurement of the rotation of the polarization of a beam of linearly polarized light after passing through a medium (plasma) in the direction of an applied magnetic field determines the electron number density. The system uses radiation in the far infrared derived from a CO₂-pumped far infrared laser.

The FRS system will provide directly the electron density and conductivity of the MHD channel, which is essential to the characterization of the power train performance. Most of the major equipment for the system has been obtained. The system assembly for electron density and density measurements on a laboratory plasma are basically complete. Electron mobility modeling and measurements will be addressed. System refinements for field tests (such as waveguides and vibration isolation for detectors) will follow. Field measurements could be made at CDIF if optical access through the channel sidewalls is provided.

3.C Field Operation/Developing Systems

3.C.1 Visible IR Emission Spectroscopy (VIES)

(emission spectra from ultraviolet through near infrared)

The VIES system will detect wavelength-resolved emission from the ultraviolet through the near infrared. Comparison of the resulting emission spectrum with the spectrum of an available, commercial blackbody radiation source will enable estimation of relative emissivity values. The system uses receiving optics and a computer-controlled monochromator equipped with broadband detectors.

In order to evaluate the HTAH ceramic for possible use in its combustor, CFFF would benefit from knowledge of its relative emissivity values in their gas stream. A VIES system has been built mostly from available equipment. This system records the wavelength-resolved emission spectrum of the ceramic in the CFFF gas stream. Estimates of the relative emissivity values are obtained from comparison of the emission spectrum with the spectrum of an available, commercial blackbody radiation source. In addition, this general purpose system can be utilized to obtain the relative emissivity values of other substances, or to study the emission of atomic and molecular species present in high temperature MHD gas streams.

DIAL has mobile instrument laboratories (MIL) with computer-controlled diagnostic instrumentation and trained operators for on-site measurements. Field measurements will be made at the national MHD facilities on an as-needed basis. These mobile instrument laboratories house the diagnostic equipment for transportation to, and for use at, a particular facility. The Airstream MIL has its own electric power source. Recently an 18-wheel trailer has been modified to provide an advanced mobile instrument laboratory (AMIL) to supplement the Airstream MIL which has been operational for a number of years. This field laboratory is used for the advanced diagnostic systems, such as CARS, FRS, etc., and for simultaneous operation of a large number of DIAL's instruments.

Both field laboratories have on-board computers for on-line data acquisition and control of the diagnostic equipment. The existing data analysis system, however, lacks the speed and capabilities necessary to provide quick, report quality, feedback to the facility. To this end, modifications and additions will be made to the on-line data analysis system. The particular goals are:

- speeding up analysis and display so as to provide quick feedback to the facility;
- providing on-line data display using 2D/3D techniques to help visualize and interpret the data and facility behavior;
- producing presentation quality reports for each instrument by the end of the day of measurement; and
- validating data, which is done manually at present, automatically by the analysis system.

DIAL has started evaluation of the advanced instrumentation required to control a complete MHD system. As part of the support of the national MHD program, this effort will include evaluation of diagnostic systems for monitoring/control as well as an evaluation of the dynamic characteristics of the complete MHD system when the bottoming cycle is integrated with the topping cycle. Instrumentation will hence be selected to control the MHD power system under all perturbations.

Task 4. Project Management

The objective of this task is to provide the required management for accomplishment of the Statement of Work with the proposed resources management reporting as required by DOE Order 1332.1A. Contract reporting requirements will be part of this task.

Task 5. Technology Transfer

Technology transfer is a congressionally-mandated objective of the federal government. To ensure the maximum benefits to DOE's, and subsequently DIAL's, investment in research and development, DIAL's personnel will cooperate with regional industries to demonstrate the technology for intrusive and nonintrusive diagnostic instrumentation developed for the characterization of MHD systems. Organizations which have already benefitted from this cooperative exchange are Dow Chemical, USA, Dantec Electronics, NASA and the Stennis Space Center. However important, this task shall in no way interfere with the mainstream efforts of the MHD program.

DESCRIPTION OF FACILITIES

The combustion test stand is a computer-controlled simulation test facility available for the combustion of fuel oil up to 800 lbm/hr of fuel and air. A system for using a coal slurry for fuel is presently under development. The air can be preheated to temperatures up to 1100 K (1520 F) by electrical resistance heaters. Downstream of the combustor are refractory-lined sections containing access ports on both sides and the top of each section.

The DIAL test facility has a Hewlett-Packard measurement and control system (HP-1000/A400 minicomputer-based system) with real-time software, graphics, and extensive input-output (I/O) capability.

The microprocessor development laboratory has six PC's, two IBM PS/2-80's, and two Diversified Technology development systems for developing software and firmware for special purpose microcomputers to be used in controlling diagnostic instruments, signal processing, and signal and data analysis. A variety of other PC's and microcomputers (at least twelve) are part of the field instruments and those under development.

Complete laser facilities are available, including YAG and dye lasers, spectrometers and photometers, fiber optics, and other equipment. Discharge lamps, power supplies, vacuum systems, UV optics, monochromators, detectors, and processing electronics are also available. In addition, the laboratory has several bench-top laboratory burners controlled by precision gas handling systems.

The following facilities are available in DIAL: electronics shop, machine shop, instrument shop, and gas sampling and analysis system. The Department of Electrical Engineering at MSU has the facilities and staff for the design, development, and construction of hybrid integrated circuits and printed circuit boards.

Modeling and data analysis are done on the newer, more powerful PC's (PS/2 and 386 PC's) and the VAX 11/780 and VAXstation 3200. Large scale analysis and modeling is required for various instruments including CARS, LDV, and PE/AS. Other instrument development and refinement projects also require continued modeling support in the areas of general combustion/heat transfer and gas flow modeling.

TECHNICAL PROGRESS

Task 1. Diagnostic Instrument Development

1.A Coherent Anti-Stokes Raman Spectroscopy System (CARS)

F.Y. Yueh and J.P. Singh, Principal Investigators

Work Performed

HCl CARS spectra were recorded in a quartz cell placed in a furnace at different temperatures to test the theoretical model described in a previous report.¹ Figure 1.A.1 shows the fitting of the observed HCl CARS spectra at 300, 500, 700 and 800 K with the library of the computer-simulated spectra. Temperature is the only floating parameter in those fits. The fit in the higher J regions of the spectrum is better than that in the lower J region. This might be due to the difference in the Raman linewidth at lower and higher J-values which have not been taken into account in the present model. Inferred temperature from the HCl CARS is in agreement with the furnace thermocouple temperature. Table 1.A.1 shows the comparison of the inferred HCl CARS temperatures with thermocouple temperatures. $\bar{T}_{fit}$ is the average CARS inferred temperature from five to ten measurements and σ is the standard deviation. The last column shows the error in the inferred temperature as compared with the thermocouple temperature.

Further study will continue to improve the fitting of the observed spectra with computer-simulated spectra. The better fit of the observed spectra at higher J than lower J (0-3) shows that J-dependent linewidth information is needed to improve the data analysis. More refined spectroscopic data are also needed for accurate temperature measurements with HCl CARS spectrum.

Conclusions

The theoretical HCl CARS model was tested with the HCl CARS furnace data. Though overall fit of the observed furnace spectra with the computer-simulated spectra looks reasonably good, further study will still be needed to improve the model, especially in the lower Raman wavenumber shift. HCl is not a major species in MHD, so HCl CARS work will not continue after the initial measurements in the quartz cell.

Work Forecast

Work will concentrate on high pressure N₂ and CO CARS modeling for the upcoming LMF5 test run at the CFFF aerodynamic duct, which has an operating pressure above one atmosphere.

Figure 1.A.1. Fitting of the computer-simulated spectra with the observed spectra at (a) 300 K, (b) 500 K, (c) 700 K and (d) 800 K.

Table 1.A.1. Comparison of inferred HCl CARS temperatures with thermocouple temperatures.

THERMOCOUPLE TEMPERATURE (T_{tc}), K	CARS INFERRED TEMPERATURE T_{fit} (K) (σ)	$\frac{T_{tc} - \bar{T}_{fit}}{T_{tc}}$
300	312.6 (9.6)	4.20×10^{-2}
400	421.6 (9.6)	5.40×10^{-2}
500	494.2 (2.8)	1.16×10^{-2}
700	636.1 (9.2)	9.13×10^{-2}
800	798.0 (22.2)	2.50×10^{-2}

References

1. Yueh, F.Y. and Singh, J.P. 1991. DIAL quarterly technical progress report. FE-15601-46:27.

1.B Particle Size Distribution System (PSD)

J.A. Etheridge and W. Okhuysen, Principal Investigators

Work Performed

Electronic System Development: A directly-coupled logarithmic amplifier for use with small-angle, near-forward scatter (SANFS) signals and with the pedestal part of the signal derived from crossed beam measurements has been completed. The amplifier is based on the AD640 chip manufactured by Analog Devices. At this time only a single AD640 stage is being used, which gives approximately 45 dB of signal range. Although this is adequate range for the DIAL test stand and for measurements down stream of the ESP unit at UTSI, it may not be adequate for many other situations. A second stage would provide from 70 to 80 dB but implementation of the two stage circuit is more difficult. A circuit including automatic offset correction for a two stage AD640 is under development. This circuit will replace the single stage circuit if it is successful.

The ADSP2101 digital signal processor was selected for building a prototype digital system for collection and reduction of SANFS data. The 2101 was chosen on the basis of speed, technical support available, and special university pricing of development hardware and software. The only analog circuitry in the prototype system will be a logarithmic amplifier and the circuitry associated with the PMT.

Date Reduction: A technique for reduction of SANFS data which is based on Z-transforms has been implemented on a PC. This technique requires knowledge of the receiver function, or the discrete impulse response as it is called in system theory, for the SANFS mode. Division of a polynomial representing the Z-transform of the bin count data by a polynomial representing the Z-transform of the receiver function is required to deconvolve the data. One advantage of the Z-transform technique is the ability to use digital signal processing techniques to filter or smooth the data before final data reduction. The real advantage of this technique may be in the speed of data reduction using a digital signal processor chip.

1.C Potassium Emission/Absorption System (PE/AS)
L.E. Bauman and W. Wang, Principal Investigators

Work Performed

The concurrent data acquisition/data analysis program has gone through several lab and field tests and is now a dependable program. This program is, and will continue to be, the data acquisition program during field tests. Analysis results are printed with a line printer. An on-line result display program was added to the system. This program replaces the printer and plots the results on the PC screen.

In November a hardware misconnection was discovered and corrected. This solved the double exposure problem which had been observed since May 1991.

Conclusions

The concurrent data acquisition/data analysis program has been laboratory and field tested. An on-line result display program was added.

Work Forecast

The lamp currently in use is no longer available on the market. It may be possible to replace the strip filament lamp with a coil filament lamp. This possibility should be explored both theoretically and experimentally in the near future.

1.D Intrusive Multi-Probe System (IMPS)
C.W. Bouchillon, Principal Investigator

Work Performed

Work performed during October - December, 1991, included continued efforts toward integration of the probe positioning system with the Accufiber light pipe sensor data acquisition system (LPS). The laboratory was moved in order to obtain a source of compressed air for operation of the air motor drive system. The control system program was rewritten for the new A/D board and 80386-based microcomputer system which was obtained for dedicated use for control/data acquisition operation of the IMPS.

Evaluation of various suppliers of silicon diode array-based spectrophotometers resulted in procurement of bids for a system with a 1024-element detector array. Use of this array system will allow for more information to be obtained from the IMPS system.

Experiments were conducted with a telephoto lens used to focus an image on the optical fiber at the film plane of a camera body. Preliminary results indicate that this will allow long distance observation of temperatures and emissivities with the light pipe sensor.

A modification to the LPS probe system was made to allow the sapphire/quartz sensor and the fiber optic cable to remain in the water-cooled probe system. A butt joint in the fiber cable did not produce unacceptable decay of the signal.

The SLR sensor fabrication has been delayed because of changes in priority. This task has been rescheduled to November, 1992.

Conclusions

The modification to the LPS makes the installation of the system much simpler and does not significantly degrade the operation of the system.

The procurement of a silicon diode array should make significant contributions to the information which can be collected for surface and gas measurements of high temperature situations. Having computer control of the positioning of the sensors used with the water-cooled probe will be of significant advantage, particularly while scanning across a test section.

Work Forecast

During the first quarter of calendar year 1992, the silicon diode array system for use with the water-cooled optical fiber probe system will be procured. Computer control of placement of the probe sensors will be activated. Software will be developed for the silicon diode array system. Software for improved data analysis and presentation will be developed.

1.E *Faraday Rotation System (FRS)*
L.E. Bauman and W. Wang, Principal Investigators

Work Performed

Not much progress was made during this period. Dr. Bauman is on sabbatical leave and Mr. Wang has been on assignment with PE/AS.

Work Forecast

A scheme will be configured using only the attenuation of the laser beam to measure the conductivity. The system will be made field useful while port access with magnet is unavailable.

1.F Multi-Purpose Imaging System (MPIS)

R.K. Lengel, Principal Investigator

Work Performed

The thirteen year old Nd:YAG laser was worked on this quarter to improve its performance. The major problems with the laser were beam quality and power stability. The beam quality problems were attacked by replacing the output coupling mirror, the intracavity polarizer, and the frequency doubling unit. Each of these replacements did improve the quality of the beam somewhat, but there are still irregularities that are amplified by the dye laser and subsequent non-linear frequency mixing to produce the laser beam used to excite fluorescence in the laboratory flames. Further alignment of the dye laser and WEX (wavelength extender, the frequency mixing unit) will be attempted to minimize these non-uniformities, and the final beam shape will be carefully measured so that images excited by the laser can be corrected.

The WEX unit, currently used to mix residual 1064-nm light from the Nd:YAG laser with the 652.4-nm light from the dye laser to produce the 404.4-nm light used to produce potassium fluorescence images, has a tracking system that attempts to continually maximize the efficiency of frequency mixing. The feedback circuitry of the tracker is quite sensitive to the intensity of the input beams used and has the tendency to become unstable due to power drifts caused by temperature changes in the frequency doubler of the Nd:YAG laser. The newly installed doubling unit has a regulated heating stage that maintains the non-linear crystal at a fixed temperature, which seems to have alleviated this problem.

1.G Differential Absorption Laser Spectroscopy System (DALAS)

D.L. Monts, Principal Investigator

Work Performed

Work has progressed slowly on this system for several reasons. An attempt was made to use standard catalog components rather than custom components in construction of the optical cell/vacuum system in order to minimize cost and to facilitate delivery of replacement parts should any be needed in the future. Access to the Nd:YAG laser was limited because it is shared with the CARS and LOGS projects, and access to the computer required for two-color operation was also limited because it was in use in the field.

A replacement pressure sensor for a Pirani-type pressure measurement system for the optical cell/vacuum system has been installed and verified. This pressure gauge system is capable of monitoring pressures from one atmosphere down to about 10 millitorr; and is mounted between the optical cell and the mechanical vacuum pump. A capacitance manometer-type pressure sensing system for measuring pressures in the 1 to 100-torr range has been mounted on the optical cell. This pressure gauge system will be used to measure the pressure of sample gas admitted into the optical cell after the cell has been evacuated and sealed off from the mechanical vacuum pump.

Efforts to date to achieve adequate vacuum (a few tens of millitorr) with windows trapped between two Conflat-type metal flanges have not proved successful. Standard ultrahigh vacuum viewports do not efficiently transmit the portion of the ultraviolet where the absorptions of NO and SO₂ occur. An attempt was made to avoid using a custom ultrahigh vacuum viewport with UV-grade fused silica windows because of the expense and because that ultrahigh vacuum capability is not required. Therefore, we attempted to obtain an adequate vacuum by trapping UV-grade fused silica windows (sandwiched between plastic O-rings) between metal flanges. However, this configuration has relatively large vacuum leaks that we have been so far unable to seal despite extensive effort. The most likely location of the leak is around the quartz windows. Care must be taken while tightening the bolts so as not to crack the quartz windows. During the first quarter of 1992, our extensive, but unsuccessful, efforts to seal windows between plastic O-rings continued. Ultimately, during the first quarter of 1992, we were compelled to purchase custom ultrahigh vacuum viewports with UV-grade fused silica windows. Progress was thus slowed by the duration of our extensive efforts to use less sophisticated and less expensive technology.

Efforts have begun to identify possible on- and off-resonance wavelengths for excitation of the $v' = 0 \leftarrow v'' = 0$ vibrational-electronic band of NO's γ system. These

efforts involved both a literature search and, in collaboration with Professor Chun-Fu Su of the Department of Physics and Astronomy, the calculation of theoretical spectra of this band at flame temperature using Voigt line shapes.

Some progress has been made on conversion of the grazing-incidence dye laser to two-color operation. Because of the extensive use of the Nd:YAG laser by the CARS and the LOGS projects, only limited laser time was available for DALAS. An additional Nd:YAG laser system has been requested in order to facilitate progress on all three projects.

In addition, the computer shared by the DALAS, LOGS, and VIES systems was used by the VIES system for ten days in November for calibration and participation in the LMF5-D test run at the CFFF. This prevented the DALAS and LOGS systems from using the computer during this time. An additional computer system has been requested so work can continue on DALAS and LOGS while VIES is preparing for or at CFFF.

Conclusions

Progress on the DALAS project has been satisfactory, but less than anticipated due to attempts to use standard catalog components rather than custom components in construction of the optical cell/vacuum system, in order to minimize cost and to facilitate delivery of replacement parts should any be needed in the future. Limited access to the shared Nd:YAG laser and to the computer required for two-color operation also slowed progress.

Work Forecast

During the next quarter, testing of the optical cell/vacuum system will be completed. Significant progress in the conversion of the dye laser to two-color operation will be made. Efforts to determine the concentration of NO in the optical cell will begin.

1.H Laser Optogalvanic Spectroscopy System (LOGS)

D.L. Monts, Principal Investigator

Work Performed

Slow progress was made due to the discovery that the "zero" (offset) of the detection electronics significantly drifts over periods of several hours. The sign and magnitude of the drift seemed to be arbitrary. Since both the LOGS and DALAS projects use similar equipment, and since both methods assume that the "zero" does not change with time, extensive efforts were made to characterize and diagnose this problem. The problem was due to a faulty delay generator used to trigger the computer. The synchronization pulse from the Nd:YAG laser is temporally too short to trigger the computer directly. Therefore, the synchronization pulse is fed into a pulse generator, whose output pulse's amplitude and temporal duration can be varied to conform to the requirements of the computer. After a user-selected number of laser pulses (typically 100), the computer reads the averaged output of the LOGS boxcar averager and of the photodetector boxcar averager, then resets the boxcar averagers to zero. Since the computer was being triggered at improper times by the faulty pulse generator, the computer was reading the averaged output of the boxcar averagers not after the desired number of LOGS (or photodetector) signals had been averaged together, but rather after an unknown and probably varying number of signals had been averaged together. Thus, the computer-recorded data collected at the beginning of this quarter were unreliable. The faulty delay generator has been shipped back to the manufacturer for repair.

The absorptions of molecular species of interest require pumping the dye laser with the third harmonic (355 nm) of the Nd:YAG laser rather than the second harmonic (532 nm). Since the dye laser is in a room next to the Nd:YAG laser, the Nd:YAG laser beam must be transported to the dye laser via right-angle prisms. The prisms which have been used for 532-nm Nd:YAG laser beams are made of BK-7-type borosilicate glass. Since the laser power at 355 nm is substantially less than available at 532 nm, the transmission loss of the BK-7 glass prisms is unacceptable at 355 nm. Therefore, UV-grade fused silica right-angle prisms were acquired, and the optics necessary for investigation of the LOGS spectrum of molecular species are now available.

The computer shared by the DALAS, LOGS, and VIES systems was used with the VIES system for ten days in November for calibration then participation in the LMF5-D test run at the CFFF. This prevented the LOGS and DALAS systems from using the computer during this time. An additional computer system has been

requested so DALAS and LOGS can continue working while VIES is preparing for or at CFFF.

Conclusions

Progress on the LOGS project was slowed by faulty data and the malfunctioning equipment that caused the faulty data.

Work Forecast

During the next quarter, the reliability of the detection electronics will be verified. Investigations of the effect of electrode aging and fouling, of electrode voltage, and of electrode material on the LOGS signal in a bench-top flame will continue.

1.1 Cross Correlation System (CCS)
O.P. Norton, Principal Investigator

Work Performed

Effort on the Cross Correlation System during this reporting period concentrated on software development. This software will acquire data using high-speed A/D (analog-to-digital) boards inside the PC. The data will then be processed in real time.

Task 2. Test Stand Operations

2.A Test Stand Modifications

Future work will include more intensive work toward development of a coal slurry that can be used reliably with the present test stand fuel delivery system.

2.B Test Stand Runs

Three test stand runs were completed during this quarter and are listed in Table 2.B.1. The total run time was approximately 21 hours. The first of these tests was used for evaluating a new electronic system for the LDV system. The LDV tests were extended into the second test run which was done primarily for the CARS system. The final test was a shake down test for the PE/AS system.

The test frequency is expected to continue at about the same rate as in previous months, with several tests being scheduled prior to each trip to another facility

Table 2.B.1. Test stand runs, October through December 1991.

RUN NO.	DATE	INSTRUMENTS	CONDITIONS	COMMENTS
1	11/01/91	LDV	Seed/Slag Injection	
2	11/08/91	CARS,LDV	Seed/Slag Injection	
3	11/12/91	PE/AS	Seed/Slag Injection	

Task 3. Technical Support for National MHD Programs

The LMF5-D test was conducted at the CFFF November 18-24, 1991. DIAL installed the following instruments on the CFFF to make measurements during this test:

Coherent Anti-Stokes Raman Spectroscopy System (CARS),
Intrusive Multi-Probe System (IMPS),
Laser Doppler Velocity (LDV),
Multi-Color Pyrometer (MCP),
Potassium Emission/Absorption System (PEAS),
Two Color Laser Transmissometer (TCLT),
Visible Infrared Emission Spectroscopy System (VIES), and
Particle Size Distribution System (PSD).

The IMPS measurements show a smaller change in temperature before and after sootblowing than in previous measurements. This indicates less ash deposits on the ash deposition probe than during past tests. MCP measurements at the furnace third level show typical temperature and emissivity values. This system was also used for correlation with the VIES. CARS measurements of temperature and CO concentration were made at the diffuser. TCLT results show some of the effects of water injection on ESP performance.

Development of the VIES instrument continued with successful measurements being made in spite of problems with window fouling. However, some difficulty with a chopper has delayed the data analysis. The LDV and PE/AS systems were used to make measurements for the first time at the CFFF aerodynamic duct. Technical problems were encountered with the PSD measurements. We are still working on the analysis of the measured data. Selected results of the individual measurements follow. Complete results are provided in the MSU/DIAL Diagnostic Data Report, Volume 11, Number 4.

3.A Field Measurements

Laser Doppler Velocity Measurements, LMF5-D at CFFF (A. George and W. W. Wilson): Laser Doppler velocity (LDV) measurements were made on the west side penetration of the aerodynamic duct. The experiment was done in the backscatter mode. The transmitting optics consisted of a 9-mm beam spacer, a 2.27X beam expander and a focusing lens with a focal length of 1073 mm. This combination resulted in a fringe spacing of $27.02\text{ }\mu\text{m}$, a measurement volume diameter of $206\text{ }\mu\text{m}$ and a measurement volume length of 21.68 mm . The measurements were made on November 23, 1991, between 2:00 a.m. and 7:00 a.m.

The measurement conditions presented quite a challenge for making LDV measurements. The flow was high velocity, the particle sizes were on the average small for LDV purposes, the background radiation was intense and the collection aperture was small ($\sim 1\text{-in.}$ diameter). In spite of these obstacles, we were able to obtain some indication of the flow velocity in the aerodynamic duct.

Figure 3.A.1 shows an example of a Doppler burst obtained during the test. This signal results from a single particle traversing the measurement volume. The trace of this signal was recorded using a digital oscilloscope and a number of such bursts were saved. A fast Fourier transform analysis was performed

on each of the bursts and the frequency of each burst was calculated. The velocity of the flow was inferred from these frequency values. A histogram of the particle velocity is shown in Figure 3.A.2 with a mean velocity value of 868 m/s and root mean square velocity of 78 m/s .

Figure 3.A.1. Doppler burst obtained from CFFF LMF5-D test run.

Figure 3.A.2. Histogram of particle velocity with a mean velocity value of 868 m/s and root mean square velocity of 78 m/s.

The amount of high quality LDV data obtained during the LMF5-D test is quite limited. However, we have shown that it is possible to obtain the proper LDV signal. These first time results will allow the LDV system to be configured in a way to maximize the measurement probability in future tests. We hope to have another attempt and feel confident that a complete velocity profile across the duct can be obtained.

Multi-Color Pyrometer Measurements, LMF5-D at CFFF (J. Cambre and M. McCarthy): The Multi-Color Pyrometer system (MCP) was installed at three different locations during LMF5-D. The test began at penetration 2207, located on the second level south wall of the radiant furnace, by looking at a one-inch ceramic tube. At approximately 16:00:00 on November 18, the tube shattered and was no longer visible. The heads were refocused and set up to monitor the north wall of the furnace.

At penetration 2206, located on the second level west wall of the radiant furnace, the high temperature air heater tube (HTAH) was monitored directly across from the Visible Infrared Emission Spectroscopy (VIES) system.

Another HTAH tube was monitored at penetration 2217, located on the third level north wall of the radiant furnace. When the system was brought on line the morning of November 23, it was discovered that the tube had fallen below the viewing area of the MCP. The heads were refocused and aligned on the south wall of the furnace.

Tables 3.A.1 through 3.A.3 show hourly averages of temperature and emissivity data collected during LMF5-D.

Table 3.A.1. LMF5-D hourly average two-color temperatures and emissivities.

	Port 2207		Port 2206		Port 2217	
11/18/91	Temp. (K)	Emis.	Temp. (K)	Emis.	Temp. (K)	Emis.
13:00	1679.2	0.69	1574.7	0.72	1341.1	0.72
14:00	1725.7	0.74	1625	0.76	1363.9	0.72
15:00	1736.5	0.78	1637.6	0.8	1342	0.77
16:00	1630	0.46	1639	0.86	1362	0.8
17:00	1735.5	0.54	1648.7	0.87	1380.5	0.8
18:00	1738.1	0.61	1657.2	0.86	1386.1	0.79
19:00	1762.8	0.58	1667	0.88	1392.8	0.83
20:00	1767.8	0.58	1673.3	0.88	1410.5	0.81
21:00	1768.3	0.59	1675.4	0.89	1426.7	0.81
22:00	1770.2	0.62	1683.1	0.91	1443	0.84
23:00	1770	0.62	1684.3	0.91	1444.9	0.9
11/19/91						
0:00	1768.3	0.65	1691.4	0.93	1460.6	0.9
1:00	1776.2	0.62	1691.6	0.92	1468.7	0.85
2:00	1773.9	0.64	1694.8	0.93	1467.8	0.92
3:00	1776.5	0.63	1690.9	0.93	1458	0.94
8:00	1716.7	0.81	1707	0.87	1505	0.86
9:00	1721.1	0.83	1724.8	0.84	1513.6	0.86
10:00	1727.4	0.81	1720.4	0.83	1517.8	0.85
11:00	1725.2	0.81	1709.2	0.83	1521.5	0.83
12:00	1716.8	0.8	1697.6	0.85	1510.9	0.81
13:00	1723.5	0.79	1712.9	0.83	1516	0.82
14:00	1716.5	0.83	1735.1	0.85	1537.8	0.82
15:00	1729.6	0.78	1730.3	0.81	1528.7	0.81
16:00	1733.7	0.77	1702.1	0.85	1523.7	0.8
20:00	1729.1	0.83	1736.8	0.81	1521.5	0.82
21:00	1722.2	0.83			1529.7	0.81
22:00	1732.5	0.8	1621.2	0.81	1533.5	0.81
23:00	1745.5	0.77	1727.5	0.86	1528.4	0.81

Table 3.A.2. LMF5-D hourly average two-color temperatures and emissivities.

	Port 2207		Port 2206		Port 2217	
11/20/91	Temp. (K)	Emis.	Temp. (K)	Emis.	Temp. (K)	Emis.
8:00	1687.6	0.83	1711.8	0.72	1438.7	0.79
9:00	1718	0.8	1732.7	0.7	1441	0.79
10:00	1722.1	0.79	1730.7	0.7	1440.4	0.78
11:00	1726.9	0.8	1746.6	0.67	1447.5	0.81
12:00	1702.9	0.74	1711.7	0.6	1417.8	0.79
13:00	1712.7	0.81	1736.6	0.68	1430.3	0.81
14:00	1733	0.86	1753.4	0.77	1469.6	0.89
15:00	1689.1	0.78	1682.9	0.75	1456.4	0.94
16:00	1650.7	0.78	1645.3	0.72	1414.9	0.84
17:00	1684.7	0.83	1677.5	0.77	1481.7	0.79
18:00	1721.2	0.86	1716.5	0.85	1528.2	0.86
19:00	1712.6	0.86	1701.6	0.89	1524	0.88
11/21/91						
7:00	1689.7	0.88	1705.4	0.85	1497	0.94
8:00	1604	0.85	1638.4	0.83	1439.1	0.92
10:00	1688.5	0.8	1684.4	0.84	1486.2	0.94
11:00	1695.1	0.8	1696	0.84	1503.3	0.93
12:00	1684.2	0.79	1681.6	0.81	1495.8	0.9

Table 3.A.3. LMF5-D hourly average two-color temperatures and emissivities.

11/23/91	Port 2207		Port 2206		Port 2217	
	Temp. (K)	Emis.	Temp. (K)	Emis.	Temp. (K)	Emis.
7:00			1667.6	0.74	1552.7	0.55
8:00			1665.1	0.8	1502.6	0.68
9:00			1663.7	0.82	1498.2	0.73
10:00			1673	0.78	1490.7	0.77
11:00			1547.5	0.73	1327.2	0.61
12:00			1644.1	0.85	1494.9	0.76
13:00			1641.3	0.9	1499.6	0.77
14:00	1494.7	0.75	1630.8	0.91	1509.3	0.68
15:00	1670	0.84	1653.7	0.92	1504.5	0.77
16:00	1660.1	0.83	1644.8	0.93	1501.6	0.75
17:00	1673.5	0.84	1655	0.95	1512.7	0.79
18:00	1678.7	0.83	1658.3	0.94	1510.2	0.81
19:00	1654.6	0.79	1647.1	0.92	1488.8	0.8
20:00	1599.3	0.8	1605.5	0.86	1462.2	0.58
21:00	1626.6	0.81	1604.2	0.88	1488.9	0.56
22:00	1638.8	0.8	1609.9	0.9	1504.6	0.55

Potassium Emission/Absorption System Measurements, LMF5-D at CFFF (W. Wang): The system was taken to UTSI for the LMF5-D test run, and measurements were made at the combustor section for the first time. The system was on-line continuously for more than two days. A plot of the results for the first two days is presented in Figure 3.A.3. The system continuously monitored the process for more than two days, except when computer disks were switched and during data backup interruptions.

Figure 3.A.3. PE/AS measurements for the first two days during CFFF LMF5-D test run.

PE/AS was also installed at the aerodynamic duct during the CFFF LMF5-D test run, using penetrations 1461 and 1463. Data collection began on November 18 at 1:40 p.m. The system was running continuously until 4:40 p.m. November 20. During this period, there were only a few interruptions for computer disk unit switching, data backup, lamp changing, and one facility shut down (4:10 p.m. to 8:12

Figure 3.A.4. Plot of temperature and potassium number density vs. time for data taken 18-Nov-91, LMF5-D.

p.m., November 19).

There are large uncertainties and fluctuations in the data which are understandable considering the environment where the PE/AS was running. The gas flow speed is very high (approximately 1200 m/s) and the chemical reaction did not reach an equilibrium state. The measurements provided by PE/AS do indicate the long time (1.5 minutes) average.

Figures 3.A.4 through 3.A.9 are plots of the PE/AS measurements. For each figure, the upper plot is the temperature with the scale and label to the left, while the lower plot is the potassium or electron number density with the scale and label to the right.

Analysis of data collected during the LMF5-C test run at CFFF was completed in October. Data collected from the CFFF LMF5-D test run were replotted for a better format.

Figure 3.A.5. Plot of temperature and electron number density vs. time for data taken 18-Nov-91, LMF5-D.

Figure 3.A.6. Plot of temperature and potassium number density vs. time for data taken 19-Nov-91, LMF5-D.

Figure 3.A.7. Plot of temperature and electron number density vs. time for data taken 19-Nov-91, LMF5-D.

Figure 3.A.8. Plot of temperature and potassium number density vs. time for data taken 20-Nov-91, LMF5-D.

Figure 3.A.9. Plot of temperature and electron number density vs. time for data taken 20-Nov-91, LMF5-D.

Two Color Laser Transmissometer Measurements, LMF5-D at CFFF (J. Andol and J. S. Lindner): Studies on the performance characteristics of the CFFF electrostatic precipitator (ESP) continued during facility test LMF5-D. For these experiments the TCLT unit was located downstream of the ESP (penetration 4923/4924) while two electric field meters (EFM's) were installed on the ESP. One meter was located between fields 1 and 2 (penetration 4813) while the second meter was located after field 4 in the trapezoidal duct section (penetration 4812). Prior to the start of the test CFFF engineers replaced the transformer in field 3 of the ESP (for the LMF5-C experiment fields 3 and 4 were operated from one transformer set). Based on this work, it was expected that average diameters downstream of the ESP would be larger than during the LMF5-C test and that particle collection efficiency of the device would be improved. The results which follow indicate that the ESP operated better than during the C test although considerable variations in D_{32} were observed.

Figure 3.A.10 illustrates the Sauter mean diameters and EFM signals measured on November 19th. The bottom trace of the lower panel represents the EFM unit located downstream of field 4 and results here track those of the average particle size (upper panel). The results from the EFM located between field 1 and 2 (upper trace, lower panel) do not appear to reflect any significant occurrences arising from field 1. The signal from the field 1 EFM is larger than that observed for the field 4 EFM although this difference is not as large as expected based on estimated particle loading in the various fields. The fact that the signal is low and that rapping and wire rack vibrators are not observed in field 1 may indicate that the geometry of the EFM with respect to the field needs to be examined.

Figure 3.A.11 is similar to Figure 3.A.10, i.e., little fine structure is observed in the EFM located between fields 1 and 2 and the raw voltages from this meter are somewhat less than expected. Attempts to continue the EFM experiments were halted at around 0800 on November 20th. No signal could be observed from either of the meters and they were extracted (seriously fouled) from the ESP. Additional measures must be incorporated into the EFM design or operation which prevent the relatively rapid fouling of the units. At this time, there is no other way to obtain data within or between fields 1 and 2, 2 and 3, and 3 and 4; this information could be quite valuable for future studies and some effort is, therefore, warranted.

Average Sauter diameters measured on November 20th (Figures 3.A.12 and 3.A.13) and November 21st (Figures 3.A.14 and 3.A.15) are similar to those obtained during LMF5-C. Peaks observed in, for example, Figure 3.A.12 correspond to vibration of the wire racks in various fields of the ESP. The increase in particle size during these vibrator periods should be contrasted with the decrease in D_{32} during wire

Figure 3.A.10. Average particle size (Sauter mean diameter, above) and electric field meter output (below) for data collected 19-Nov-91, LMF5-D.

Figure 3.A.11. Average particle size (Sauter mean diameter, above) and electric field meter output (below) for data collected 19-Nov-91, LMF5-D.

Figure 3.A.12. Average Sauter diameter measured downstream of the ESP, 20-Nov-91, LMF5-D.

rack vibration in Figures 3.A.10 and 3.A.11. It appears that the vibration of the wire racks produces particles of a similar particle diameter and loading regardless of whether the ESP is operating at high collection efficiency (Figures 3.A.10, 3.A.11, 3.A.15, and 3.A.16) or at a non-optimum condition (Figures 3.A.12 - 3.A.15).

Figures 3.A.16-3.A.19 correspond to the water injection test. Here water was added to increase the humidity of the gas stream. At 1732 hours in Figure 3.A.16 sufficient water was added to roughly increase the humidity in the gas stream by 1%. Only a very slight increase in D_{32} was observed and this may well lie within the error of the particle sizing measurement. The humidity was increased by 2% at 1744 hours (Figure 3.A.17) and by a total of 3% at 1754 (Figure 3.A.18). There is clearly a small increase in D_{32} which is evident in Figure 3.A.19 where the injection ceased at 1804 hours. The effect of water injection when the ESP is operating as well as in Figure 3.A.16 - 3.A.19 is quite weak. This result was observed for water injection during the LMF5-C test; in this later case, however, a larger charge was observed owing to the relatively poor (as compared to the data in Figures 3.A.10, 3.A.11 and 3.A.16 - 3.A.19) operating characteristics of the ESP during that test. It appears that the

Figure 3.A.13. Additional data collected 20-Nov-91, LMF5-D.

addition of water has a more pronounced effect when the ESP is not operating efficiently. Further studies on the performance characteristics of the ESP as well as the new wet ESP, to be installed at the CFFF this winter, will continue.

Figure 3.A.14. Average Sauter diameter measured downstream of the ESP, 21-Nov-91, LMF5-D.

Figure 3.A.15. Additional data collected 21-Nov-91, LMF5-D.

Figure 3.A.16. Average Sauter diameter measured prior to and during the water injection test 23-Nov-91, LMF5-D.

Figure 3.A.17. Average Sauter diameter measured prior to and during the water injection test 23-Nov-91, LMF5-D.

Figure 3.A.18. Average Sauter diameter measured prior to and during the water injection test 23-Nov-91, LMF5-D.

Figure 3.A.19. Average Sauter diameter measured prior to and during the water injection test 23-Nov-91, LMF5-D.

3.B Field Tests/Instrument Modifications

Coherent Anti-Stokes Raman Spectroscopy Measurements, LMF5-D at CFFF (J.P. Singh and F.Y. Yueh): The mobile CARS system was aligned at the diffuser exit (penetration 1542 and 1543) during the LMF5-D test run to record N₂ CARS data at different stoichiometries. This data was taken on November 20, 1991, at a point on the west side of the diffuser 6" from the center of the gas stream. The data were taken using the folded BOXCARS alignment with a 100-cm focal length field lens. The spatial resolution with this optical system is close to 1 cm. The inferred temperatures from fitting of this data with the library of computer-simulated spectra are summarized in Table 3.B.1. The CARS measurements show the temperature increasing slightly with stoichiometry. Since there is a 2% (± 50 K) uncertainty in the CARS temperature measurements, and the change in temperature with stoichiometry is within the uncertainty of the CARS measurement. Additional measurements are necessary to reduce the uncertainty for this type of comparison.

Table 3.B.1. N₂ CARS temperature measurements at the CFFF diffuser exit, 20-Nov-91.

STOICHIOMETRY (ϕ)	TIME	TEMPERATURE (K)
0.80	13:01	2559
0.90	13:46	2612
0.85	16:58	2589

To test the simultaneous measurement of temperature and CO concentration, N₂-CO CARS spectra were also recorded at different positions in the gas stream. There were problems with the data taken on the first two days of the run due to improper beam alignment. These problems were subsequently found and corrected. The data taken on November 23 used the same alignment for N₂ CARS measurements during this test run. The data are promising and are presently being analyzed. A typical N₂-CO CARS spectrum is shown in Figure 3.B.1. The broad nonresonant background is primarily due to the particle-induced laser breakdown and has essentially the Gaussian shape of the broadband Stokes laser. The center frequency of the Stokes laser is chosen between the CO and N₂ CARS spectra. This is necessary if both CO and N₂ CARS spectra are to be observed simultaneously. However, this gives rise to weak N₂ CARS spectra which are at the wing of the dye laser (Stokes) profile. The CO CARS spectrum is weak because of small concentration of the species and also it is on one side of the dye laser profile.

Figure 3.B.1. Fitting of the computer-simulated spectrum with the observed N_2 -CO CARS spectrum at the CFFF diffuser 2" from the center (east side).

Preliminary results from this data are listed in Table 3.B.1. Since a non-symmetrical Stokes beam profile was found in those measurements, the temperature and CO concentration had to be extracted in two steps. First, the temperature was extracted from the spectrum, and then the CO CARS region of the spectrum was fit using the temperature determined in the first step to extract CO concentration. Due to an uncertainty in the nonresonant susceptibility (χ_{nr}) of the gas stream, which is the mole-fraction weight average of the nonresonant susceptibilities of the individual molecular constituents, there is an uncertainty in the CO concentration measurement. A 10% error in the CO concentration data listed in Table 3.B.2 is expected. The CO concentration estimated from the equilibrium calculation was found to be 13.7%. The inferred temperature with N_2 -CO CARS data is about 100 K less than that of N_2 CARS data under similar test run conditions taken on different dates. It may be noted that temperature inferred from an N_2 CARS spectrum is more reliable due to a better S/N ratio in the N_2 CARS spectrum.

The alignment of the pump and Stokes laser in the N_2 -CO CARS technique is very critical for temperature and CO concentration measurements due to the nature of the Stokes beam. Special care is necessary to obtain the best N_2 -CO CARS signal. Since the N_2 -CO CARS signal is generated mainly from the wing of the Stokes profile, the S/N ratio in the N_2 -CO CARS data is generally low. For future field

measurements, efforts will be made to improve the S/N ratio of the N₂-CO CARS data.

The new feature observed at 486 nm in the N₂-CO CARS spectra recorded at the CFFF diffuser exit has been carefully studied.¹ The process to generate that signal is a nonlinear, four-wave difference mixing process (FWM) which was verified by using a sodium-seeded CH₄/air flame at DIAL. The experimental studies will continue to further understand FWM in a combustion environment. These studies will be helpful in evaluating the process for combustion diagnostics.

CARS measurements were performed during the LMF5-D test run. The temperature extracted from N₂ CARS data should be more reliable than the N₂-CO CARS data due to a better signal/noise ratio. The mobile CARS system capability needs to improve for future field measurements.

References

1. Singh, J.P. and Yueh, F.Y. 1990. DIAL quarterly technical progress report. FE-15601-43:31.

Table 3.B.2. Simultaneous N₂-CO measurements at the CFFF diffuser exit, 23-Nov-91.

POSITION (Inches)	TEMPERATURE (K)	CO CONCENTRATION (Percent)
	<i>West Side</i>	
-6	2388	11.9
-4	2464	12.8
-2	2429	13.7
0	2451 (Center)	11.9
	<i>East Side</i>	
2	2441	9.9
4	2455	10.1
6	2487	11.9

Intrusive Multi-Probe System Measurements, LMF5-D at CFFF (R.A. Green and C.W Bouchillon): The Intrusive Multi-Probe System (IMPS) with its light pipe sensor (LPS) was set up at penetration number 3015 on the superheater test module (SHTM) inlet transition section and positioned so it could monitor the surface temperature on the ash deposition probe (ADP). The position of the LPS was verified to be one inch from the surface of the ADP.

After installation of the sootblower, 17 insertions were made. Of the 17 insertions, four were made during sootblowing. The measurements made immediately before and after sootblowing revealed a very small change in ADP surface temperature.

Previous tests have indicated that an ADP surface temperature decrease of approximately 200°F is observed when sootblowing is conducted. During this test as well as during LMF5-C, smaller temperature decreases were measured. The small change in temperature indicates that either sootblowing is removing very little of the deposit or else there is very little deposit present. Vidicon observations indicate that the latter was the case. There was very little deposit on the ADP prior to sootblowing, and sootblowing again appeared to remove virtually all that was present.

All measurements taken are reported below on Table 3.B.3. The column labeled T NSB (F) is the temperature measurement between sootblowing cycles, i.e., Temperature No Sootblowing, F. The next two columns list the temperature just prior to sootblowing (T BSB (F)) and just after sootblowing (T ASB (F)). The Delta T column lists the difference between the ADP surface temperatures before sootblowing and after sootblowing.

It was also noted that a fine, white powder deposited on the LPS during this test as was also noted during LMF5-C. This ash deposition results in a small reduction in temperature recorded over time. More work needs to be done to minimize this effect.

Data analyses were performed on data collected during the LMF5-C test at the CFFF during September 23-25, 1991. Data on the surface temperature of the ash deposition probe (ADP) were collected during the LMF5-D test at the CFFF. Data collected during the LMF5-D test run during November, 1991, were analyzed and reported. The surface temperature of the ADP at the CFFF before, during, and after sootblowing did not change significantly. Observations by video indicated there was not much ash accumulation on the probe for the burning conditions of the two test runs measured during this period. Surface temperatures of the ash deposition probe will be measured at CFFF as appropriate.

Table 3.B.3. ADP surface temperatures, CFFF LMF5-D.

Insertion Number	Date	Time	Temp NSB(F)	Temp. BSB(F)	Temp. ASB(F)	Delta T (F)
1	11/18/91	14:25	1620.7			
2	11/18/91	15:29		1644.7	1595.8	48.9
3	11/18/91	18:08	1596.2			
4	11/18/91	18:44	1688.9			
5	11/19/91	09:29	1726.1			
6	11/19/91	10:35	1766.9			
7	11/19/91	10:41		1665.5	1630.7	34.8
8	11/19/91	14:02		1788.9	1781.5	7.4
9	11/19/91	14:13	1680.0			
10	11/19/91	14:33	1694.8			
11	11/19/91	15:34	1755.8			
12	11/19/91	15:52	1775.3			
13	11/20/91	10:02	1824.1			
14	11/20/91	10:24	1817.4			
15	11/20/91	10:44	1814.6			
16	11/10/91	11:59	1858.3			
17	11/20/91	15:16	1791.4			
18	11/21/91	11:45		1871.6	1812.8	

Notes: Temp. NSB(F): Temperature (f), No SootBlowing
Temp. BSB(F): Temperature (F), Before SootBlowing
Temp. ASB(F):Temperature (F), After SootBlowing

3.C *Field Operations/Developing Systems*

Nothing to report this period.

Task 4. Project Management

This is an ongoing task considered up to date with the submission of this report.

Task 5. Technology Transfer

While mindful of the mission of this task, there is nothing to report this quarter.

**DATE
FILMED**

8/11/94

END