AIM

Association for Information and Image Management

1100 Wayne Avenue, Suite 1100
Silver Spring, Maryland 20910
301/587-8202

Centimeter

Inches

MANUFACTURED TO AIM STANDARDS
BY APPLIED IMAGE, INC.

1 of 1

Issued: June 1994

*The Los Alamos Super Scrub™:
Supercritical Carbon Dioxide System
Utilities and Consumables Study*

Jerome C. Barton

Los Alamos
NATIONAL LABORATORY
Los Alamos, New Mexico 87545

als
DISTRIBUTION OF THIS DOCUMENT IS UNLIMITED

THE LOS ALAMOS SUPER SCRUB™: SUPERCRITICAL CARBON DIOXIDE SYSTEM UTILITIES AND CONSUMABLES STUDY

by

Jerome C. Barton

ABSTRACT

A study was undertaken to determine the electrical and other utilities usage of the Super Scrub™ system. The purpose of the experiments being conducted at Los Alamos is to prove or disprove the feasibility of precision cleaning for solvent and waste reduction using supercritical carbon dioxide. An actual dollars value-versus-time has been calculated for predetermined pressures, flows, and temperatures. State-of-the-art menu- and software-driven instrumentation were utilized to generate exacting results. It is hoped that these results will prove of immediate benefit to those industries interested in but concerned about the cost of changing over to carbon dioxide cleaning on both small- and large-scale operations.

INTRODUCTION

The Los Alamos Super Scrub™ is a large supercritical carbon dioxide system. It is mechanically and technically highly sophisticated with variable temperature, pressure and flow. Pressures as high as 5000 psi, temperatures from room to as high as 60°C, and flow rates as high as 500 lbs/hr are attainable utilizing forty-five separate variable numerical input commands and thirteen positive/negative machine event manual commands. Computer control is accomplished using Genesis™ software by Iconics, a Hewlett Packard (H.P.) PC, and a General Electric Fanuc model 90-30 programmable controller system.

The Super Scrub™ was developed by Autoclave Engineers of Erie, PA. At the time of this writing only two systems of this type are in use, outside of Autoclave Engineers, in the United States. The Los Alamos machine is unit number two and is an updated design of the prototype.

The four major integral components of this system are the computer control, the pumping and flow system, a 60-liter, high-pressure extraction vessel, and a contaminants separation vessel.

Figure 1 is a system overview showing the basic layout of the operation. The extraction vessel, also known as the cleaning vessel, is physically separated from the main pumping and flow operation. For clarity, it is shown closely positioned on the schematic.

Figure 2. Load curve: per cent wattage peak vs. duration. TA-35, Bldg. 125, Rm. A101, Dec. 16, 1993.

Figure 3. Instantaneous power: 5 min/div. horiz., TA-35, Bldg. 125, Rm. A101, Dec. 16, 1993.

The Super Scrub™ is served by a dedicated 480-volt ac, 4-wire delta phase 130 amp. circuit. The inductive pickups were installed directly at this switch station. Most all the machine's electrical usage is through this circuit. This includes, in addition to the high electrical usage components, such items as internal machine lighting, a small strip heater, fans, co-solvent pump, and various control devices, all small consumers of electricity. The two exceptions to this are the computers, an H.P. Vectra for machine control, and an IBM for data collection and research purposes, and the remotely located freon-to-air heat exchanger/condenser, which will be addressed in some detail later.

MAJOR POINT SOURCE CONSUMERS

There are six components within the machine that as a group consume a major portion of all electrical power used:

- a CVI positive displacement cryo/pump powered by a 10 hp, 480-volt ac, three-phase motor with a peak current usage of 13.3 amps. or 6.4 kilowatts.
- Located at four places in the carbon dioxide flow stream, 12-kilowatt/480-volt ac, three-phase screw plug immersion heaters, which collectively could potentially use up to 100 amps.
- an R22, 17-ton, refrigeration unit used by the Super Scrub™ as a heat exchanger for the gaseous carbon dioxide, returning it to the liquid state and back into the main storage tank. The refrigeration compressor is of a two-speed design capable of consuming as much as 21.2 kilowatts on high speed and 6.4 kilowatts on low speed.

THE AIR-COOLED FREON HEAT EXCHANGER

The Super Scrub™ requires cooling for the R22 refrigeration system. The original method of cooling the freon was with the facility's industrial chilled water. As a result of long-term problems with this system, it became necessary to supply another cooling method. A model 349 Libert three-fan, air-cooled heat exchanger was installed. The engineering and retrofit were accomplished by the Los Alamos facility. This equipment is not supplied or offered as a substitute with the Super Scrub™.

The new exchanger is located approximately 40 feet remote, outside, on top of the facility's roof. It has a dissipation capacity of approximately 315,000 Btus after derating for Los Alamos' 7000-foot elevation. The exchanger is served by a dedicated 25-amp, 230-volt, 3-phase circuit.

Electrical usage measurements on the heat exchanger were made as inductive pickup voltage measurements using Bell Inc. model CG-100A Guns™. Three individual pickups were attached simultaneously to each of the incoming phases. Readouts were taken first on a Tektronics 7904 oscilloscope using 7A26 plug-ins, and then on a Fluke model 8020B multimeter. In addition, visual observations were made of the fan's operation. The machine's operating conditions during data taking were pressure at 2000 psi, flow at 500 lbs/hr and cleaning vessel temperature at 40°C. The outside weather conditions were clear and sunny. The temperature was 41°F, and a light 3 mph wind was blowing. The relative humidity was 14%.

Data were recorded over a 60-minute period. During this time it was noted that one fan turned very slowly, possibly the result of heat convection. The measured electrical energy usage was less than 1 amp peak.

When comparing these results with the cost of operating an industrial chilled water system required to supply 30 gal/min at 45°F, the benefits of air cooling become quite obvious. Routine maintenance and electrical usage for pumping would make water cooling considerable more costly when compared to air cooling.

MEASUREMENT CONSIDERATIONS

It must be explained at this point that all six of the major electrical consumers in the machine could be described as pulsed power users. The emmersion heaters rarely, except at the highest operating temperatures of 50° to 60°C, supply a full 12-kilowatt heating, and then they are cycled off and on only as needed. Inductive pickups were installed on the incoming power to the 10-hp cryo-pump motor, and, as expected, current usage varied by a factor of about 80% between the power stroke and the piston return, thereby giving a peak average usage considerably lower than motor-plate specification.

Also, during this test series the pumping rate of 500 lbs/hr was never changed. Lower pumping rates would, of course, be less energy intense.

The 17-ton R22 refrigeration system used to cool the carbon dioxide gas back to liquid is in operation continuously during flow. It does, however, switch from low- to high-speed, changing its consumption from about 6 kilowatts to about 22 kilowatts.

THE COST OF IDLE TIME

When the machine is idle, the pressure slowly builds up in the liquid storage vessel and the heat exchanger/condenser. Routinely, the refrigeration compressor cycles on slow speed when vessel pressure reaches a preset level of 400 psi, thereby reducing the storage vessel pressure to 340 psi. This function is seen to occur on an average of only once per hour normally and is very inexpensive. Figures 6 and 7, instantaneous power and cash flow, are the result of one hour of idle time and show that the compressor came on only once during this hour using power at a rate of 5 kilowatts for a very short time and costing less than five cents. Also observable on these graphs is a level above zero cost per hour. This is the standard continuous use of power by controls, strip heater on the refrigeration compressor, and such, and is very low-cost.

Figure 6. Cash flow: Powerprofiler Site, 5 min/div. horiz., Jan. 27, 1994.

Figure 7. Instantaneous power: Powerprofiler Site, 5 min/div. horiz., Jan. 27, 1994.

INITIAL CARBON DIOXIDE FILLING OF THE SUPER SCRUB™

Figure 8 shows the cumulative instantaneous power usage necessary to fill the machine from scratch. The recorded time to fill was 82 minutes.

The carbon dioxide source was a gas-bottle rack containing seven 50-lb cylinders on a common manifold of one-half-inch stainless tubing with a run of about 40 feet to the machine. The cylinders were at room temperature and at a pressure of approximately 900 lbs.

At no time during the fill did the refrigeration compressor run at high speed, and at no time, as can be seen in Figure 8, did the instantaneous power usage exceed 6 kilowatts. The total electrical cost to fill the machine was 15¢ at a total kilowatt usage of roughly 1.6 as calculated at the 10¢ per kilowatt hour rate.

A total fill from scratch is not normally required during routine daily operations; however, if the carbon dioxide should somehow become contaminated or if power to the machine is interrupted for a period in excess of two hours, then venting to the atmosphere would become necessary.

One hundred liters is the total volumetric size of the carbon dioxide liquid storage vessel. However, only about 80% or 80 liters of carbon dioxide is required to activate the automatic fill shutoff. A calculation based on a cost per pound of liquid at 10¢ per pound means that a total fill cost for the carbon dioxide would be about \$17 plus electrical costs.

Figure 8. Instantaneous power: Powerprofiler Site, 5 min/div. horiz., Feb. 2, 1994.

COMPRESSED AIR USAGE

The Super Scrub™ machine uses compressed air to operate numerous control valves throughout the flow stream. A total of 10 SCFM at 90-100 psi is sufficient for this purpose. A unit such as a standard Sears 2-hp, 20-gal. home workshop air compressor would be large enough to perform this function. These typically use 120-volt ac with a peak current usage of 15 amps or less than 2 kilowatts at 10¢ per kilowatt hour, less than 20¢ per hour.

Figure 9. Electrical usage for one hour in dollars.

CONCLUSION

The Los Alamos Super Scrub™ machine is quite inexpensive to operate from an electrical as well as a consumables standpoint. Figure 9 shows a dollar-per-kilowatt-hour number for measured electrical power usage at the main machine buss. As can be seen, the higher operating temperatures are more likely than higher pressures to cause higher operating costs.

When adding up all the costs of operation previously described, it is easily seen that the electrical costs would be small when compared to operating labor, routine maintenance, and repair.

ACKNOWLEDGMENTS

Thanks to Autoclave Engineers for permission to reprint one of their computer screens. Also special thanks to Dale Spall for his guidance in the world of supercritical fluids.

This work was performed under funding from The Industrial Waste Reduction Program Office, Department of Energy, Division of Conservation and Renewable Energy.

**DATE
FILMED**

9 / 1 / 94

END