DOE/ER/60512--6

DE92 017304

**PREPARATION OF RADIOPHARMACEUTICALS LABELED
WITH METAL RADIONUCLIDES**

PROGRESS REPORT

FOR PERIOD

JULY 1, 1989 - JUNE 30, 1992

Michael J. Welch, Ph.D.

Principal Investigator

**The Edward Mallinckrodt Institute of Radiology
Washington University School of Medicine
510 South Kingshighway
St. Louis, MO 63110**

June 1992

Prepared for

U.S. DEPARTMENT OF ENERGY

UNDER GRANT DE-FG02-87ER60512

DISCLAIMER

This report was prepared as an account of work sponsored by an agency of the United States Government. Neither the United States Government nor any agency thereof, nor any of their employees, makes any warranty, express or implied, or assumes any legal liability or responsibility for the accuracy, completeness, or usefulness of any information, apparatus, product, or process disclosed, or represents that its use would not infringe privately owned rights. Reference herein to any specific commercial product, process, or service by trade name, trademark, manufacturer, or otherwise does not necessarily constitute or imply its endorsement, recommendation, or favoring by the United States Government or any agency thereof. The views and opinions of authors expressed herein do not necessarily state or reflect those of the United States Government or any agency thereof.

MASTER *MR*

DISTRIBUTION OF THIS DOCUMENT IS UNLIMITED

COMPREHENSIVE REPORT

The major goals of the proposal three years ago were:

1. The continued evaluation of gallium-68 as a radionuclide for the labeling of PET pharmaceuticals.
2. The development of new ligands for use with gallium and indium.
3. The evaluation of radiolabeled antibodies and fragments.
4. The development and evaluation of gallium-68 labeled proteins and peptides.
5. The evaluation of tracers labeled with copper-62 as blood flow imaging agents.

We have made major advances in all of these areas. The major research accomplishments performed over the prior funding period have been:

1. The demonstration of the usefulness of gallium-68 labeled transferrin in the study of lung function.
2. The labeling of a series of new ligands, proteins and peptides with gallium-68.
3. The production of a high yield, 100 mCi zinc-62/copper-62 generator and complexation of the Cu-62 with PTSM allowing high quality images of myocardial and renal blood flow utilizing PET scanning.
4. The labeling of proteins and antibodies with copper-62 and copper-64. An IND has recently been submitted for ⁶⁴Cu-labeled 1A3 antibody. The Phase I clinical trial of this antibody has begun.

Our plans for the continuation of the proposal are to proceed to develop radiopharmaceuticals as we have in the past. One major difference will be that we will investigate mechanisms of uptake for various metals attached to proteins and antibodies in a well defined model in order to determine the cellular uptake metabolism and possible release of the radionuclide. Preliminary work has already been accomplished.

During the past funding period, Susan Madsen obtained a Ph.D. working on this project and Carol Wiegers will complete her Ph.D. in the fall of 1992. Michelle Jones-Wilson and Buck Rogers are graduate students working toward their Ph.D. in the Department of Chemistry on projects directly related to this grant. Nick Franano is a M.D. student at Washington University who is working toward his M.D./M.S. on this project. James Brodack who was involved in this project is now a Senior Research Scientist at Mallinckrodt Inc. in St. Louis.

PUBLICATIONS

1. Shelton ME, Green MA, Mathias CJ, Welch MJ, Bergmann SR. Kinetics of copper-PTSM in isolated hearts: a novel tracer for measuring blood flow with positron emission tomography. *J Nucl Med* 30:1843-1847, 1989.
2. Brodack JW, Kaiser SL, Welch MJ. Laboratory robotics for the remote synthesis of generator-based positron-emitting radiopharmaceuticals. *Laboratory Robotics and Automation* 1:285-294, 1989.
3. Mathias CJ, Sun Y, Connett JM, Philpott GW, Welch MJ, Martell AE. A new bifunctional chelate, BrMe₂HBED: an effective conjugate for radiometals and antibodies. *Inorg Chem* 29:1475-1480, 1990.
4. Mathias CJ, Welch MJ, Raichle ME, Mintun MA, Lich LL, McGuire AH, Zinn KR, John EK, Green MA. Evaluation of a potential generator-produced PET tracer for cerebral perfusion imaging: single-pass cerebral extraction measurements and imaging with radiolabeled Cu-PTSM. *J Nucl Med* 31:351-359, 1990.
5. Madsen SL, Bannochie CJ, Martell AE, Mathias CJ, Welch MJ. Investigation of physicochemical and in-vivo behavior of diastereomeric iron-59, gallium-68, and indium-111-EHPG trivalent metal complexes. *J Nucl Med* 31:1662-1668, 1990.
6. Schuster DP, Haller J. A quantitative correlation of extravascular lung water accumulation with vascular permeability and hydrostatic pressure measurements: a positron emission tomography study. *J Crit Care* 5:161-168, 1990.
7. Green MA, Mathias CJ, Welch MJ, McGuire AH, Perry D, Fernandez-Rubio F, Perlmutter JS, Raichle ME, Bergmann SR. [⁶²Cu]-labeled pyruvaldehyde bis(N⁴-methylthiosemicarbazonato)copper(II): synthesis and evaluation as a positron emission tomography tracer for cerebral and myocardial perfusion. *J Nucl Med*, 31:1989-1996, 1990.
8. Shelton ME, Green MA, Mathias CJ, Welch MJ, Bergmann SR. Assessment of regional myocardial and renal blood flow using copper-PTSM and positron emission tomography. *Circulation* 82:990-997, 1990.
9. Kaplan JD, Calandrino FS, Schuster DP. A positron emission tomographic comparison of pulmonary vascular permeability during the adult respiratory distress syndrome and pneumonia. *Am Review of Respiratory Disease*, 143:150-154, 1991.

10. Sun Y, Mathias CJ, Welch MJ, Madsen SL, Martell AE. Targeting radiopharmaceuticals II: evaluation of new trivalent metal complexes with different overall charges. *Nucl Med Biol* 18:323-330, 1991 (*Int J Radiat Appl Instrum Part B*).
11. Mathias C, Margenau WH, Brodack JW, Welch MJ, Green MA. A remote system for the synthesis of copper-62 labeled Cu(PTSM). *Int J Appl Isot* 42:317-320, 1991 (*Int J Radiat Appl Instrum Part A*).
12. Moerlein SM, Daugherty A, Sobel BE, Welch MJ. Metabolic imaging with gallium-68- and indium-111-labeled low-density lipoprotein. *J Nucl Med* 32:300-307, 1991.
13. Madsen SL, Bannochie CJ, Welch MJ, Mathias CJ, Martell AE. TMPHPG: a lipophilic derivative of EHPG for iron, gallium-68 and indium-111 hepatobiliary clearance. *Nucl Med Biol* 18:289-294, 1991.
14. Mathias CJ, Welch MJ, Green MA, Diril H, Meares CF, Gropler RJ, Bergmann SR. In vivo comparison of copper blood-pool agents: potential radiopharmaceuticals for use with copper-62. *J Nucl Med* 32:475-480, 1991.
15. Velazquez M, Weibel ER, Kuhn C, Schuster DP. Positron emission tomographic evaluation of pulmonary vascular permeability: a structure-function correlation. *J Appl Physiology* 70:2206-2216, 1991.
16. Schwarz SW, Mathias CJ, Sun JY, Dilley WG, Wells Jr. SA, Martell AE, Welch MJ. Evaluation of two new bifunctional chelates for radiolabeling a parathyroid specific monoclonal antibody with In-111. *Nucl Med Biol* 18:477-481, 1991 (*Int J Radiat Appl Instrum Part B*).
17. Otsuka FL, Welch MJ, Kilbourn MR, Dence CS, Dilley WG, Wells Jr. SA. Antibody fragments labeled with fluorine-18 and gallium-68: in vivo comparison with indium-111 and iodine-125-labeled fragments. *Nucl Med Biol* 18:813-816, 1991 (*Int J Radiat Appl Instrum Part B*).
18. Mathias CJ, Welch MJ, Perry DJ, McGuire AH, Zhu X, Connett JM, Green MA. Investigation of copper-PTSM as a PET tracer for tumor blood flow. *Nucl Med Biol* 18:807-811, 1991 (*Int J Radiat Appl Instrum Part B*).
19. Kaplan JD, Calandrino FS, Schuster DP. Effect of smoking on pulmonary vascular permeability: a positron emission tomography study. *Am Rev of Respir Dis*, 145:712-715. 1992.

20. Madsen SL, Welch MJ, Motekaitis RJ, Martell AE. $^{68}\text{GaTHM}_2\text{BED}$: a potential generator-produced tracer of myocardial perfusion for positron emission tomography. *Nucl Med Biol* 19:431-444, 1992.
21. Anderson CJ, Connett JM, Schwarz SW, Rocque PA, Guo LW, Philpott GW, Zinn KR, Meares CF, Welch MJ. Copper-64 labeled antibodies for PET imaging. *J Nucl Med* (in press).
22. Wieggers CB, Welch MJ, Sharp TL, Brown JJ, Perman WH, Sun Y, Motekaitis RJ, Martell AE. Evaluation of two new gadolinium chelates as contrast agents for MRI. *Magnetic Resonance Imaging* (in press).
23. Herrero P, Markham J, Weinheimer CJ, Anderson CJ, Welch MJ, Green MA, Bergmann SR. Quantification of regional myocardial perfusion with generator-produced copper-62-PTSM and positron emission tomography. *Circulation* (submitted).

SELECTED ABSTRACTS (Not accepted as full papers)

24. Herrero P, Markham J, Weinheimer CJ, Green MA, Welch MJ, Bergmann SR. Quantification of myocardial perfusion with Cu-62-PTSM and positron emission tomography. *J Nucl Med* 32:937, 1991.
25. Moerlein SM, Daugherty A, Sobel BE, Welch MJ. Utility of Tc-99m- and In-111-labeled low-density lipoprotein as radiopharmaceuticals for metabolic imaging. *J Nucl Med* 32:926-926, 1991.
26. Herrero P, Markham J, Weinheimer CJ, Green MA, Welch MJ, Bergmann SR. Quantification of myocardial perfusion with PET: comparison of H_2^{15} and ^{62}Cu -PTSM. *Proceedings American Heart Association*, Anaheim, CA, 1991.
27. Anderson CJ, Schwarz SW, Rocque PA, Connett JM, Guo LW, Philpott GW, Zinn KR, Meares CF, Welch MJ. Copper-labeled antibodies for PET imaging. *Proceedings from the IXth International Symposium on Radiopharmaceutical Chemistry* pp 368, 1992.
28. Anderson CJ, Welch MJ, Weinheimer CJ, Bergmann SR. Preparation and in vivo comparison of ^{62}Cu -labeled blood-pool agents. *Proceedings from the IXth International Symposium on Radiopharmaceutical Chemistry* pp 558, 1992.

29. Moerlein SM, Welch MJ, Hotchkiss RM, Yahng CS, Fischmann AJ, Strauss HW, Rubin RH. A gallium-68 labeled chemotactic peptide analogue for imaging focal sites of bacterial infection by PET. *Proceedings from the IXth International Symposium on Radiopharmaceutical Chemistry* pp 426, 1992.
30. Moerlein SM, Yahng CS, Welch MJ. Radiosynthesis of technetium-99m labeled human very-low density lipoprotein for scintigraphic study of lipoprotein receptor activity in vivo. *Proceedings from the IXth International Symposium on Radiopharmaceutical Chemistry* pp 393, 1992.
31. Bergman SR, Herrero P, Anderson CJ, Welch MJ, Green MA. Measurement of regional myocardial perfusion in human subjects using copper-62-PTSM. *Journal of Nuclear Medicine* 33:837, 1992.
32. Anderson CJ, Schwarz SW, Bergmann SR, Green MA, Welch MJ. Synthesis and quality control of $^{62}\text{Cu}(\text{PTSM})$ for clinical studies. To be presented American Chemical Society Meeting, August, 1992.

All publications are included in the Progress Report as appendices. Abstracts which have subsequently been published and/or submitted as full papers are not included in this listing or in the appendices. This project has acted as a resource to all collaborators to test other new radiopharmaceuticals. A submitted publication of this type is appendix 33.

Relationship to DOE Nuclear Medicine Program Objectives

With the widespread use of sophisticated positron imaging devices, there is an increased demand for positron-emitting radiopharmaceuticals. Generator produced radiopharmaceuticals add an extra dimension to the in-house cyclotron produced radiopharmaceuticals.

Radiopharmaceuticals prepared using generator-produced isotopes compliment those produced from accelerators. This means that hospital accelerators could be of a simpler design than those currently utilized. It is interesting to note that several companies including IBA (Belgium), Science Research Laboratory (Sumerville, MA), SAIC (San Diego, CA), and AccSys Technology (Pleasanton, CA) are selling or evaluating small simple accelerators. These simple accelerators together with fluorine-18-labeled compounds from

a regional distribution center (prototypes of which are being set up in Los Angeles, Phoenix, and Florida) and generator produced radiopharmaceuticals would allow a comprehensive PET center to operate without the large staff and laboratories which are required at present.

There are only a limited number of generators that produce positron-emitting radionuclides. These generators include strontium-82/rubidium-82, zinc-62/copper-62, xenon-192/iodine-122 as well as germanium-68/gallium-68. Two of the most promising positron generators are the germanium-68/gallium-68 and the zinc-62/copper-62 generators. In this proposal major advances have been made in the development and evaluation of compounds labeled with these agents. This will allow the use of PET technology in places other than major medical centers and will be particularly significant in the field of human health and assessment letters. In providing this technology to a wider base of the population, this proposal relates directly to the DOE Nuclear Medicine program mission.

Relationship to Other DOE Work

Over the length of this project, significant collaborations have been carried out with individuals both funded by DOE and funded by NIH. The copper-67 that was initially used to evaluate copper radiopharmaceuticals was produced at Brookhaven National Lab and Los Alamos National Lab (LANL) and the germanium-68/gallium-68 generator utilizes germanium-68 produced at LANL. The work with copper labeled proteins and antibodies involves a collaboration with Claude Meares of the University of California, Davis who is a DOE funded investigator. The collaboration with the University of Missouri Research Reactor to produce Cu-64 is partially funded by a DOE reactor sharing agreement.

PROGRESS REPORT

Since the initial inception of this project, the major goal has been to develop and evaluate new radiopharmaceuticals labeled with metal radionuclides; and in particular, radionuclides that are generator produced. In the initial funding period, we evaluated germanium-68/gallium-68 generators and gallium-68 radiopharmaceuticals, developed new labeling techniques for gallium and indium radiopharmaceuticals and extended this work to labeling of MRI contrast agents with iron. We recently developed a useful zinc-62/copper-62 generator and are presently evaluating copper-62 radiopharmaceuticals for clinical

studies. While developing the copper-62 radiopharmaceuticals, we collaborated with the University of Missouri Research Reactor, Columbia in the production of copper-64. The PET images we obtained with copper-64 tracers were of such high quality that we have developed and evaluated copper-64 labeled antibodies for PET imaging. In the future we plan to continue to evaluate gallium-68, copper-62, and copper-64 radiopharmaceuticals. We will also evaluate new methods of studying the metabolism of metal-labeled compounds as well as investigating novel techniques for the labeling of proteins with metal radionuclides.

The major research activities which were planned three years ago include:

- 1) The development and assessment of gallium-68 radiopharmaceuticals.
 - a) Routine radiopharmaceuticals for the study of lung function.
 - b) The development and assessment of new ligands.
 - c) The labeling of proteins and peptides.
 - 2) The development and evaluation of a new zinc-62/copper-62 generator and the assessment of copper-62 radiopharmaceuticals.
 - a) The routine production of copper-62 PTSM.
 - b) The production of new copper-62 ligands.
 - 3) The development of new agents using the positron-emitting nuclide copper-64.
 - 4) Mechanistic studies on proteins labeled with metal radionuclides.
- These activities are described below and are addressed concurrently.

The discussion below is extended in the publications included in the appendices.

1) Routine Production of Gallium-68 Transferrin

We have collaborated with the group of Dr. Daniel Schuster for many years on the production and application of gallium-68 radiopharmaceuticals. At present, the germanium-68/gallium-68 generator purchased under this grant is being evaluated for the application of PET, particularly in the adult respiratory distress syndrome (ARDS). The current state

in this area is described in appendices 6, 9, 15, and 19.

Dr. Schuster and his co-workers have shown that the pulmonary transcapillary exchange rate (PTCR) measured by gallium transferrin is a sensitive index of lung injury. This has been achieved through a combination of animal experiments in clinical studies. Histological evidence of disease activity will be correlated with serial measurements and with PTCR in patients with the ADRS; and, the clinical usefulness of this type of study will be evaluated.

2) Evaluation of New Ligands

Over the past funding period, we have evaluated several new ligands for potential use with gallium-68. A new hexadentate ligand [N,N,N',N'-tetrakis-(2-hydroxy-3,5-dimethylbenzyl)ethylenediamine](THM₂BED) has been synthesized and complexed with gallium isotopes. This work is described in appendix 20. A conditional stability constant of $\log k = 22.3$ has been determined for the gallium complex at physiological pH. It was shown that due to the limited rotation in the molecule, two distinct isomers could be isolated by preparative HPLC and evaluated individually. The preliminary data shows that this radiotracer may be capable of giving high quality myocardial images. There is also preliminary evidence that this complex penetrates the blood-brain-barrier. We are attempting to synthesize analogs of THM₂BED with different substituents on the aromatic ring, and to prepare the analogous ligands where the benzene rings are replaced with pyridene rings. We will also evaluate THM₂BED and other complexes in isolated cell models (see later).

3) Proteins and Peptides

Over the past several years, there has been great interest in labeling of proteins and peptides as radiopharmaceuticals. Although the greatest emphasis has been placed on labeling of monoclonal antibodies, other molecules have been studied. These include labeled lipoproteins and insulin which are involved in a large number of physiological disorders. Labeled neogalactoalbumin for the study of NGA receptors in liver disease,¹ labeled chemotactic peptides for the detection of sites of infection² as well as a labeled somatostatin analogs for the quantification of somatostatin receptors in primary and

metastatic tumors, have been studied.³ We have been investigating the labeling of several of these proteins and peptides with gallium-68. During the remainder of the current funding period, we plan to label each of the proteins and peptides mentioned above with Ga-68. Gallium-68 low density lipoprotein (LDL) appears to be a useful radiopharmaceutical for the studying of the LDL receptor (appendix 12). Recent work has shown that labeled VLDL has been shown to be superior to labeled LDL. We have recently shown that both indium and technetium labeled VLDL can be utilized to distinguish lipoprotein receptor rich liver from receptor poor liver (appendices 25-29). Figure 1-2 shows the results recently obtained with technetium-99 labeled VLDL. In the case of neogalactoalbumin (NGA), we have obtained deferoximine conjugated NGA from Drs. Stadalnik and Vera of the University of California at Davis and have shown that high labeling yields can be obtained utilizing gallium-68. These agents can be utilized for the labeling of the NGA receptor. We have also obtained N-formyl-methionyl-leucyl-phenylalanine from Drs. Fishman and Strauss from the Massachusetts General Hospital. This peptide binds with high affinity to receptors on white blood cell membranes. The group at Massachusetts General Hospital has shown that a series of indium-111 labeled derivatives can be utilized for gamma scintigraphy.² Utilizing the gallium-68 from the DuPont Medical Products generator, we have labeled this peptide in relatively high yield. The DTPA derivative of the peptide was utilized. Gallium-68 is eluted from the generator using 1 molar HCl, boiled to dryness in a borosilicate glass test tube under a stream of nitrogen and re-dissolved in 0.4 M NaOAc for transchelation. After 30 minutes, labeling yields of up to 40% are obtained using 0.5 mg of labeling substrate.

We are particularly interested in labeling somatostatin analogs with gallium-68. Octreotide (or Sandostatin) is an eight amino acid peptide which binds to the somatostatin receptors. Octreotide has been conjugated with diethylenetriamine pentaacetic acid for labeling with ¹¹¹In.³ In the synthesis of DTPA-conjugated-octreotide, the ϵ -NH₂ group may be selectively protected. We have obtained several kits of DTPA-octreotide from Mallinckrodt Inc. and will utilize these for preliminary labeling studies.

To date, our work with labeled proteins and peptides have utilized amounts as low as approximately 0.5 mg of material. Due to the limited number of receptors and the pharmacological effects of octreotide, gallium-68 will have to be labeled to much less

Hepatic Localization of Tc-99m Labeled Lipoproteins in Rabbits as Determined by γ -Scintigraphy

Figure 1

Hepatic Localization of [^{99m}Tc] VLDL in Rabbits as Determined by γ -Scintigraphy

Figure 2

^{68}Ga -EDTA Complexation Effects of Multiple Elutions

Figure 3: Yield of ^{68}Ga -EDTA with addition of limiting amount of EDTA to the generator eluate.

peptide. In the clinical studies carried out in Holland, 10 micrograms of DTPA-octreotide has been labeled with indium-111.⁴ We have been studying the effects of using microgram quantities of chelates on the labeling yields with gallium-68. Figure 3 shows a model system where EDTA was labeled with gallium-68 eluted from the germanium-68/gallium-68 generator and the amount of EDTA was reduced to less than one microgram. Two different studies are presented in Figure 3, one where the generator was allowed to sit overnight and the other where the generator was eluted four hours prior to the study. It is seen in the latter case, reasonable labeling yields can be obtained with as little as 0.5 μ g of chelate. This is the amount of chelate equivalent to 2.5 μ g of DTPA octreotide. We believe that it will, therefore, be possible to label octreotide with positron-emitting ⁶⁸Ga. Labeling will be accomplished similarly to the chemotactic peptides. These labeled proteins and peptides will be utilized in imaging studies by a long term collaborator Dr. Barry A. Siegel of the Division of Nuclear Medicine at Washington University.

The Zinc/Copper Generator

1) Routine Production of Cu-PTSM

As is shown in appendices 7-11, we have developed a zinc/copper generator and prepared copper-62 PTSM which has been studied in humans. The generator has been prepared in a routine manner and a RDRC application for the administration of copper-62 PTSM for patients in the study of myocardial perfusion has been approved by the Washington University School of Medicine radioactive drug research committee (RDRC). As of June, 1992, this agent has been administered to twelve patients or normal volunteers. The generator is housed in a specially designed shield which also acts as an ionization chamber to allow measurement of the activity on the generator as well as the amount eluted.

Copper-62 ($t_{1/2}$ = 9.8 minutes) is produced via the ⁶²Zn/⁶²Cu generator. Zinc-62 ($t_{1/2}$ = 9.1 hours) in 2 N HCl is loaded onto a Dowex 1 anion exchange column. To elute ⁶²Cu with as little ⁶²Zn breakthrough as possible, the generator is eluted with a 2 N Cl⁻ solution. Initially, 2 N HCl was the generator eluate which was buffered with 2 equivalents of 3 N sodium acetate (NaAc) and reacted with μ g quantities of H₂PTSM. The yields of ⁶²Cu-PTSM were inconsistent, and experiments showed the large excess of NaAc which hindered

the reaction of ^{62}Cu with the μg quantities of H_2PTSM . The generator eluate was changed to 1.8 N $\text{NaCl}/0.2$ N HCl and buffered with 2 equivalents of 0.3 N NaAc . Under those conditions, 5 μg of H_2PTSM has been used to consistently obtain $> 90\%$ yields of ^{62}Cu -PTSM in clinical trials for myocardial imaging. ^{62}Cu -PTSM is purified by a C-18 sep-pak which is washed with distilled water and the product eluted in ethanol. The ^{62}Cu -PTSM in 0.2 ml of ethanol is diluted to 7 ml with sterile saline and sterile filtered with a 0.2 μ Millex FG filter before injection. Quality control is accomplished by silica gel thin layer chromatography in two developing solvents.

Table 1 shows the values for ^{62}Zn breakthrough. The breakthrough was measured in the generator eluant (2 N Cl^-) and in the ^{62}Cu -PTSM preparations. Although the generator eluant had a breakthrough of $> 1 \times 10^{-3}\%$, the actual breakthrough in the patient sample was significantly less ($< 1 \times 10^{-5}\%$). Because ^{62}Zn does not bind as strongly to H_2PTSM than ^{62}Cu , much of the ^{62}Zn is purified by the C-18 sep-pak.

The yields of ^{62}Cu -PTSM for the first 18 patient doses administered thus far were consistently greater than 90%, and the purities are consistently better than 95%. It was determined that ^{62}Cu -PTSM may decompose on the silica plate when ethyl acetate is the developing solvent. A TLC of ^{62}Cu -PTSM on silica in ethanol was taken as a back-up quality control.

Typically, the ^{62}Zn arrives on a Monday evening around 5 p.m. The ^{62}Zn is loaded onto the ion exchange column, and the elution profile is accomplished to determine what volume of generator eluant has the greatest amount of ^{62}Cu . Samples from the elution profile are saved to measure the ^{62}Zn breakthrough. One or two ^{62}Cu -PTSM preparations with quality control are executed. Patient doses are prepared on the following day. After the patient is positioned in the PET scanner, the patient will first be administered C^{15}O and H_2^{15}O followed by ^{62}Cu -PTSM. For a measure of how the myocardial perfusion agents perform at high blood flow rates, Dipyridamole is injected, followed by H_2^{15}O . Dipyridamole is injected a second time followed by ^{62}Cu -PTSM. Two or three patients (or normal volunteers) were studied for each generator. For the myocardial imaging studies, the generator was only used on the day after the ^{62}Zn was received (typical yields of ^{62}Cu -PTSM are shown in Table 2).

TABLE 1

Zinc-62 Breakthrough

Date	% ^{62}Zn Breakthrough Generator Eluant	% ^{62}Zn Breakthrough ^{62}Cu -PTSM Preparation
12/2/91	9.31×10^{-4}	1.15×10^{-6}
12/17/91	2.55×10^{-3}	3.89×10^{-6}
1/21/92	5.73×10^{-4}	4.45×10^{-6}
2/25/92	2.60×10^{-3}	4.89×10^{-6}
	avg = 1.66×10^{-3}	avg = 3.6×10^{-6}

TABLE 2

mCi of ^{62}Cu -PTSM Labeled for Patients

Date	mCi ^{62}Cu -PTSM Monday Evening Preps	mCi ^{62}Cu -PTSM Patient 1 Dose 1	mCi ^{62}Cu -PTSM Patient 2 Dose 2
12/3/91	85	26	18
12/17/91	75	31	24
1/21/92	80	25	15
2/25/92	65	28	15
	avg = 76	avg = 28	avg = 18

2) Labeling of Other Ligands With Copper-62

It would be desirable to have a ^{62}Cu -radiopharmaceutical for delineation of blood volume to use with ^{62}Cu -PTSM for correction of radioactivity emanating from the vascular space. One possible vascular agent is ^{62}Cu -labeled human serum albumin (HSA), utilizing the bifunctional chelate 6-p-bromoacetamidobenzyl-1,4,8,11-tetraazacyclotetradecane-N,N',N'',N'''-tetraacetic acid (Br-benzyl-TETA). ^{62}Cu -benzyl-TETA-HSA has been shown to give images that are comparable to ^{15}O -labeled carbon monoxide; however, the preparation includes a time-consuming spin column purification step to remove unlabeled copper, and the overall yield is only about 16% (without decay correction). An ideal vascular agent would be a copper-labeled protein conjugated to a bifunctional chelate which can be labeled in less than ten minutes in high yield and is stable in vivo for at least 30 minutes. Fujibayashi, et. al.⁵ have labeled ^{62}Cu to a dithiosemicarbazone-human serum albumin (DTS-HSA) conjugate for measuring blood volume in a dog model. The labeling of ^{62}Cu to DTS-HSA was accomplished by mixing the ^{62}Cu in 200 mM glycine with the DTS-HSA solution. The labeling efficiency was consistently greater than 95%, and no spin column purification was necessary to remove free copper. In our group, we have compared blood clearance of ^{67}Cu -labeled DTS-HSA and benzyl-TETA-HSA in adult female Sprague-Dawley rats (appendix 14).

Bromo-benzyl-TETA was synthesized and conjugated to HSA as described by McCall, et. al.⁶ p-Carboxyethylphenylglyoxal-N',N''-dimethyldithiosemicarbazone (CE-DTS) was prepared by a modified method of the synthesis described by McPherson, et. al.⁷ CE-DTS was conjugated to HSA via the coupling agent diphenylphosphoryl azide (DPPA) and purified using centricon membranes. Labeling of ^{67}Cu and ^{62}Cu to benzyl-TETA-HSA was accomplished by adding copper acetate to the conjugated protein, and purifying by spin column chromatography. The labeling efficiencies ranged from 60-75%. Copper-labeled DTS-HSA was prepared similarly; however, spin column chromatography showed a labeling efficiency of greater than 95%.

Blood clearance and biodistribution of ^{67}Cu -benzyl-TETA-HSA and ^{67}Cu -DTS-HSA were determined by removing a blood sample by direct cardiac puncture at the appropriate times after i.v. injection. The blood was counted in a NaI well counter, and the % injected dose/gram was determined for each sample. The blood clearance of both agents were

comparable at times up to 30 minutes, however, copper-labeled benzyl-TETA-HSA was stable for a longer time period (appendix 27).

Due to consistent labeling, efficiencies of greater than 95%, in the preparation of ^{62}Cu -DTS-HSA for dog imaging, the final spin column purification was excluded. This gave a final yield of ^{62}Cu -DTS-HSA that was over three times that of ^{62}Cu -benzyl-TETA-HSA (without decay correction). To evaluate the imaging characteristics of the ^{62}Cu -agents, 7 mCi of ^{62}Cu -benzyl-TETA-HSA and 10 mCi of ^{62}Cu -DTS-HSA were administered sequentially to an anesthetized dog. Static tomographic reconstructions were compared with those obtained after administration by inhalation of 40 mCi of C^{15}O , which binds irreversibly to erythrocytes. Images obtained with ^{62}Cu -labeled benzyl-TETA-HSA and DTS-HSA showed identical blood pool distribution compared to those obtained with C^{15}O (Figure 4). The simpler method of production and higher yield of ^{62}Cu -DTS-HSA as well as the high quality of the blood pool images suggest that this radiopharmaceutical would be of more practical use than ^{62}Cu -benzyl-TETA-HSA.

TABLE 3

Blood clearance in adult female Sprague-Dawley rats of the copper-labeled blood-pool radiotracers 1-60 minutes post i.v. injection. The data is reported as the mean $\pm$ standard deviation.

Time (minutes)	%ID/g DTS-HSA	%ID/g Benzyl-TETA-HSA
2	7.51 $\pm$ 0.24	6.67 $\pm$ 1.08
5	6.92 $\pm$ 0.39	5.59 $\pm$ 1.05
10	6.61 $\pm$ 0.57	5.30 $\pm$ 0.95
20	4.71 $\pm$ 0.48	4.93 $\pm$ 0.99
60	3.26 $\pm$ 0.43	4.62 $\pm$ 0.58

3) Labeling With Copper-64

In our work with ^{64}Cu -PTSM which was utilized prior to our studies with ^{62}Cu -PTSM, (appendix 4) high quality PET images were obtained following administration of relatively

Figure 4

Comparison of myocardial blood pool images made with $C^{15}O$ and two ^{62}Cu -labeled blood pool tracers in a dog heart. Identical images are obtained.

small amounts of this ligand. This led us to believe that copper-64 might be promising as an agent for labeling antibodies. Appendix 21 describes our work in this area. This is particularly true in light of recent advances in PET imaging techniques including whole body imaging⁸ and 3-dimensional imaging techniques⁹ that increase the sensitivity of PET when used with low levels of radioactivity. Of the long-lived PET radionuclides available for antibody labeling, we have chosen to investigate ⁶⁴Cu for reasons which include its availability from the University of Missouri Research Reactor (MURR) in Columbia, MO, and our experience in labeling proteins with copper radionuclides (appendix 14). Antibodies labeled with ⁶⁴Cu could also have use in therapy, since Apelgot and co-workers¹⁰ have shown it to be as effective in damaging DNA as ⁶⁷Cu.

We have evaluated ⁶⁴Cu-labeled 1A3, an anti-colorectal carcinoma monoclonal antibody, in the Golden Syrian Hamster Model.^{11,12} Antibodies were labeled using the bifunctional chelate 6-[p-(bromoacetamido)benzyl]-1,4,8,11-tetraazacyclo-tetradecane-N,N',N'',N'''-tetraacetic acid (Br-Benzyl-TETA) and the spacing moiety 2-iminothiolane by the techniques described by McCall, et. al.⁶ Several types of studies have been carried out utilizing this antibody and in all cases, the immunoreactivity (IR) was greater than 85% for intact antibody and approximately 70% for F(ab')₂ fragments. Both intact 1A3 and 1A3-F(ab')₂ have been shown to be virtually 100% stable in rat plasma as determined by FPLC. The absolute uptake (%ID/g) of copper-labeled intact 1A3 antibody or fragments were higher in tumor tissue than with either ¹¹¹In-Br- Φ HBED-1A3, or ¹²⁵I-labeled 1A3 antibodies which we have investigated. Tumor to nontumor ratios for the antibody and its fragments are as high or higher than those we have obtained for indium- or iodine-labeled 1A3 antibodies.

Human absorbed radiation doses for intact 1A3 and 1A3-F(ab')₂ were calculated from rat biodistribution data. The dose to the kidney from the ⁶⁴Cu-labeled fragments was extremely high, indicating that only a very small amount of this radiolabeled antibody could be administered to humans. For this reason we will carry out clinical trials utilizing ⁶⁴Cu-benzyl-TETA-1A3. If improvements in the conjugation and purification of benzyl-TETA-1A3-F(ab')₂ can be made which will reduce the kidney dose, we will utilize the fragments.

Mechanistic Studies

One of the goals of this project is to determine the metabolism of labeled proteins and peptides in the body. We have attempted to study this in intact animals, however, the results are inconclusive.

To best evaluate the metabolism of polypeptide-chelate-metal complexes, an in vitro model system is needed. Studies with tissue culture cells offer a homogeneous cell population and exacting control of the experimental conditions. For metabolism studies, one can closely correlate enzyme levels with metabolic flux and test hypotheses using specific inhibitors, many of which cannot be used in vitro. Mutant cell lines, either naturally occurring or genetically engineered, are also available. For such studies, we have developed a means of targeting polypeptide-chelate-metal complexes to the lysosome. The lysosome is a common final destination for a variety of agents such as antibodies and polypeptide hormones. Proteins such as albumin can be targeted to lysosomes by coupling mannose 6-phosphate bearing oligosaccharides to the polypeptide backbone. Cell surface mannose 6-phosphate receptors then bind these extracellular ligands and deliver them to the lysosome by receptor mediated endocytosis. Since the receptor recycles, a single receptor can deliver

more than 100 ligand molecules to the lysosome. Thus, mannose 6-phosphate-albumin can be used as a vehicle for delivering large amounts of metal-chelate complexes to the lysosome. As shown in Figure 5, when cells are incubated in media containing ^{111}In -DTPA-mannose 6-phosphate-albumin, the cells accumulate ^{111}In in a time dependent manner that is specifically inhibited by mannose 6-phosphate. The radiometal remains cell associated despite degradation of the protein backbone (results not shown). With this system, the metabolism of each component in the protein-chelate-metal complex can be studied in a wide variety of cell types.

Figure 5. Uptake of ^{111}In -DTPA-PMP-HSA by mouse fibroblasts. Cells were grown to confluence in 35 mm dishes. ^{111}In -DTPA-PMP-HSA ($3.5\mu\text{Ci}/\mu\text{g}$) was added to a final concentration of approximately 20 nM and the cells were held at 37°C for the indicated times. The media was then removed, the cell monolayer washed five times with ice cold buffer and cell associated radioactivity measured in a gamma-counter. Where indicated, the media contained 10 mM mannose 6-phosphate. The values for three wells were averaged and normalized to cellular protein. At 24 hours, cells cultured without mannose 6-phosphate contained 360,000 cpm per 35 mm dish, 14% of the added radioactivity.

We have available to us several bifunctional chelates for labeling both indium and copper and recently a series of chelates have been made commercially available (Dojindo Labs, Kumamoto, Japan). A novel acid hydrolyzable linker has been described in the literature^{13,14} and Dr. Blattner has provided us with a small quantity of this ligand. We will attach indium and copper to mannose 6-phosphate albumin allow the metal to be taken up in the tissue culture cells and analyze the cell lysate for the form of the metal. In this way, differences between the various linkers and chelators can be studied.

REFERENCES

1. Vera DR. *J Nucl Med* 33: 1160-1166, 1992.
2. Fischman AJ, Pike MC, Kroon D, et al. *J Nucl Med* 32: 483, 1991.
3. Bakker WH, Krenning EP, Reubi J-C, et al. *Life Sciences* 49: 1593-1601, 1991.
4. Krenning EP, Bakker WH, Kooij PPM, et al. *J Nucl Med* 33: 652-658, 1992.
5. Fujibayashi Y, Matsumoto K, Arano Y, et al. *Chem Pharm Bull* 38: 1946, 1990.
6. McCall MJ, Diril H, Meares CF. *Bioconj Chem* 1: 222, 1990.
7. McPherson DW, Umbricht G, Knapp Jr. FF. *J Lab Comp Radiopharm* 28: 877, 1990.
8. Guerrero TM, Hoffman EJ, Dahlbom M, et al. *IEEE Trans, Nucl Sci* 37: 676-680, 1990.
9. Cherry SR, Dahlbom M, Hoffman EJ. *J Comp Assist Tomag* 15: 655-668, 1991.
10. Apelgot S, Coppey J, Gaudemer A, et al. *Int J Radiat Biol* 55: 365-384, 1989.
11. Connett JM, Fenwick JJ, Timmcke AE, et al. *Proc. Amer. Ass. Cancer Res* 28: 352, 1990.

12. Mathias CJ, Sun Y, Welch MJ, et al. *Bioconjugate Chem* 2: 204-211, 1990.
13. Blattler WA, Kuenzi BS, Lambert JM, et al. *Biochem* 24: 1517-1524, 1985.
14. Blattner WA, Lambert JM. *Acid cleavable compound U.S. Patent* 4: 764, 1988.

All papers removed and cycled separately -

END

**DATE
FILMED**

9 / 8 / 92