DE93 001358

TASK PLAN FOR TEST OF PRBT PROTOTYPIC LIQUID SAMPLER REV. 1 (U)

by

W. J. Jenkins

Westinghouse Savannah River Company
Savannah River Site
Aiken, South Carolina 29808

DISCLAIMER

This report was prepared as an account of work sponsored by an agency of the United States Government. Neither the United States Government nor any agency thereof, nor any of their employees, makes any warranty, express or implied, or assumes any legal liability or responsibility for the accuracy, completeness, or usefulness of any information, apparatus, product, or process disclosed, or represents that its use would not infringe privately owned rights. Reference herein to any specific commercial product, process, or service by trade name, trademark, manufacturer, or otherwise does not necessarily constitute or imply its endorsement, recommendation, or favoring by the United States Government or any agency thereof. The views and opinions of authors expressed herein do not necessarily state or reflect those of the United States Government or any agency thereof.

This report was prepared in connection with work done under Contract No. DE-AC09-89SR18035 with the U.S. Department of Energy. By acceptance of this report, the publisher and/or recipient acknowledges the U.S. Government's right to retain a nonexclusive, royalty-free license in and to any copyright covering this report along with the right to reproduce and to authorize others to reproduce all or part of the copyrighted report

MASTER

DISTRIBUTION OF THIS DOCUMENT IS UNLIMITED

**WESTINGHOUSE SAVANNAH RIVER COMPANY
SAVANNAH RIVER LABORATORY**

WSRC-RP-90-1055 Rev. 1

Keywords: DWPF, DWPT, Liquid
Sampler, PHA, PRBT,
SRAT/SME, Analytical,
Process Control, Melter feed

CC: E. W. Holtzscheiter, 773-A
J. F. Ortaldo, 704-S
J. A. Gentilucci, 704-S
E. L. Graf, 704-S
H. H. Elder, 704-S
M. J. Swain, 704-28S
H. A. Helms, 704-28S
R. S. Beck, 704-28S
R. Spencer, 704-28S
D. H. McGuire, 704-27S
E. C. Guidotti, 704-27S
M. L. Mertz, 704-27S
M. J. Plodinec, 773-A
J. R. Pelfrey, 773-59A
C. E. Coffey, 773-A
K. Andringa, 773-41A
P. F. Cloessner, 773-A
J. T. Carter, 704-1T

L. F. Landon, 704-T
J. F. Sides, 679-T
C. J. Coleman, 773-A
C. P. Reeve, 773-42A
T. B. Edwards, 773-42A
M. F. Bryant, 773-A
R. W. Bean, 773-A
R. E. Eibling, 704-T
T. A. Policke, 704-1T
D. R. Best, 704-1T
L. C. Johnson, 704-T
R. E. Edwards, 704-1T
K. G. Brown, 704-1T
D. F. Bickford, 773-A
G. E. Weeks, 704-T
J. J. Gemmill, 773-A
M. A. Baich, 773-A
SRL Records File (4)

TO: C. T. Randall, 704-T

FROM: W. J. Jenkins, 704-1T

February 3, 1992

Task Plan for Test of PRBT Prototypic Liquid Sampler Rev. 1(U)

Summary of Changes

This revision updates the Task Plan requirements and responsibilities; this revision provides the complete updated Task Plan including all of the tables, figures, sample preparation sequence and analyses sequence in the APPENDIX A for ATTACHMENTS. The DWPT Analytical Lab shall do all the sample preparations and all of the sample analyses for this task. Responsibilities have been updated accordingly. A practice run shall not be required prior to the analyses of actual samples for this task.

C. T. Randall, 704-T
Authorized Derivative Classifier

This task has been included in the Integrated DWPF Schedule with a required completion Milestone of 12-Jun-92. The collection of samples was completed on 24-Jan-92. The schedule of activities for the completion of this task is shown in the table below:

DWPF Activity ID	DWPT Sub. ID	Activity Description	Start Date	Finish Date
6910063	A	TI 17.2 PRBT Sampler Task:	06-Jan-92	02-Feb-92
		<i>Collect Samples for PRBT Sample Study:</i>	06-Jan-92	02-Feb-92
6910065		PRBT Sampler Task(take Samples)	03-Feb-92	12-Jun-92
		<i>Complete & Document PRBT Sample Study:</i>	03-Feb-92	12-Jun-92
		Lab Elemental Analyses of Samples	03-Feb-92	30-Mar-92
	B	Prepare Sample Analyses Results for SCS Statistical Analyses	31-Mar-92	06-Apr-92
	C	SCS Statistical Analyses of Sample Results	07-Apr-92	06-May-92
	D	Document Results and Write Report	07-May-92	12-Jun-92
6910064		DWPF Milestone		12-Jun-92

Background

This task is part of the Defense Waste Processing Facility (DWPF) Process Control Program for Control of the Melter Feed Composition^[1]. The primary objective of the Process Control Program is to provide process measurement and calculational capability to enable DWPF to produce Melter feed within the acceptable compositional range without rework.

The ability to take representative samples of DWPF liquid slurry feed streams using the DWPF liquid sampler must be verified. The accuracy and precision of taking representative Precipitate Reactor Bottom Tank (PRBT) samples is needed.

Process cycle times for various feed tanks can lengthen if a substantial number of samples must be taken due to low precision and/or samples which are not representative of the tank contents. Non-representative samples or inaccurate analytical results can lead to out-of-specification Melter feed material which must be adjusted in the SME or can lead to the false conclusion that out-of-specification Melter feed is within specification which can then lead to out-of-specification glass. Either effect would impact the overall attainment of DWPF.

The samples taken with the Hydragard™ prototypic (peanut vial) liquid sampler in the Full-Scale PRBT tank at TNX shall be compared statistically with samples taken by a more fundamental sampling method from other locations in the tank to determine how representative the Hydragard™ prototypic samples are of the in-tank contents. More fundamental grab samples taken at two vertical depths and at a radial location different from the prototypic sampler shall be used as a reference to establish the homogeneity of the in-tank contents and the accuracy of samples taken by the Hydragard™ prototypic liquid sampler. Samples taken with the grab sampler at the lower elevation shall be used

as a fundamental reference for comparison with samples taken with the prototypic liquid sampler to establish the accuracy of the prototypic sampler.

The scatter in analytical results of the samples will be used to determine the variance and to identify the major sources of variance. The sources of variance that are excessive will identify areas where further development would yield the maximum reduction in the overall variance. The sources of variance and the overall variance are needed to predict the number of samples that must be taken and the number of replicates required for each sample to provide an acceptable uncertainty.

Objectives

The primary objective of this task is to determine just how representative PRBT samples taken from the Hydragard™ prototypic liquid sampler are of the in-tank contents and also to determine the homogeneity of the in-tank contents. This shall be accomplished by a statistical design study sampling plan for paired contrasts of the analysis results of samples taken from: 1) the Hydragard™ prototypic liquid sampler paired with the more fundamental grab sampler at the lower elevation for the Hydragard™ prototypic liquid sampler accuracy and 2) the grab sampler at the lower elevation paired with the grab sampler at the upper elevation for the in-tank homogeneity. These measurements are paired together to sharpen the contrast so that, as nearly as possible, only the conditions of the effect under study will differ between the two. This "split-plot" arrangement provides increased precision for the contrast under study by canceling out common extraneous effects, thereby enabling the detection of smaller effects.

The secondary objective of the task is to determine the level of influence of the major contributors to the overall uncertainty in the sample preparation and measurement process. These major steps include: preparation method(H_2SO_4 -HF Titanium dissolution); aliquoting and dilution within a dissolution; measurement and long-term behavior of the ICP and AA instruments, as monitored by the measurement of standards and blanks embedded within each block of samples for the measurement sequence.

This sampling task, therefore, is mostly devoted to determining the sampler characteristics and is not intended to provide a comprehensive estimate of the overall uncertainty affecting DWPF sample analysis in routine operation. It will provide us a better estimate than we now have of the sampling/measurement system margin of uncertainty around the acceptable composition control region but further efforts directed towards analysis will be required to totally specify that region.

Customer Requirements

General

The customers for this task are DWPF and Defense Waste Processing Technology (DWPT). An estimate of the uncertainty that the liquid slurry sampler provides a representative sample of the in-tank contents is needed by DWPT for development of the statistical process control to determine whether the size of the defined envelope of acceptable combinations of sludge, precipitate hydrolysis aqueous (PHA) feed, and frit, based on glass property models is sufficiently large for the accuracy and statistical variance of analytical sample results.

The statistical design sampling plan which is to be followed for this task is very similar to but not identical to the statistical study design which was used for the Test of the SME Prototypic Liquid Sampler^[2] which has previously been prescribed in detail by the Scientific Computations Division.^[3] The statistical design sampling plan for this task is similar to the statistical design which was recommended for testing the Hydragard™ Liquid Sampler during DWPF Cold Runs^[4] which has been prescribed in detail by the Scientific Computations Division.^[5] The order in which these samples shall be taken, the order in which these samples shall be prepared for analysis, and the order in which the samples shall be analyzed are important. The order which samples shall be grouped together in the same block so that they are prepared and analyzed within the same short-term time period is also important. The location where check standard solutions, check standard samples, and blanks shall be embedded into each block of samples and analyzed along with the samples is also important. The replication of samples, preparation methods, aliquots, and measurements required for this task are also important. This task plan contains the specific details of the statistical design which are shown in Appendix A. The following are shown in Appendix A of this task plan: 1) Customer Identification Numbering System; 2) Table 1 - Design Size; 3) Allocation of the Degrees of Freedom; 4) PRBT Sampler Blocking Pattern & Vial Numbering; 5) Figure 1A - Summary of Various Sampling Trees for ICP Analyses; 6) Figure 1B - Summary of Various Trees for AA Analyses; 7) Sampling Order; 8) Sample Preparation Dissolution Sequence and Blocking; and 9) Sample Analyses Sequence. The DWPT Analytical Lab is to communicate prior to and throughout the task to ensure that any deviations from these preparation sequences and analyses sequences are acceptable to W. J. Jenkins.

The check standard solutions (prepared from check standard sample as described in the following paragraph and denoted by "(K)" in this task plan and by "(*)" in the design study^[3]), check standard samples (denoted by "(S)" in both this task plan and the design study^[3]), and blanks (denoted by "(B)" in both this task plan and the design study^[3]) shall be identified with a coded alpha-numeric name similar to ones used for the samples; the specific details are shown in the preparation dissolution sequences and sample analyses sequences that are shown in Appendix A. Upon reaching the place of measurement, the samples, and these special items (K, S, B), will be renumbered in the simple sequential order of one through the number of samples plus special items (K, S, B) and cross-

referenced to their original sample number and special item origin so that the location of each within each block of samples can be readily determined; this numbering order shall be identical to the order of the measurements. All analyses results of these check standard solutions (K), check standard samples (S), and blanks (B) shall be reported along with the sample analyses results. There will be a minimum of five check standard solutions (K) measured per block of samples: one at the beginning of the block, one at the end of the block, one in the middle of the block, one between beginning of the block and the middle of the block and one between the middle of the block and the end of the block. One check standard sample (S) and one blank (B) shall be embedded into the middle of each block of samples and measured with each block of samples for elemental analyses.

These special samples (K, S, B) shall be matrix matched, that is, those prepared for the H_2SO_4 -HF Titanium dissolution shall be prepared by the H_2SO_4 -HF Titanium dissolution just as the actual samples are prepared. A typical PRBT composition shall be used as the stock for all the check standard samples (S) and for preparation of all the check standard solutions (K) used in this task. D. F. Bickford shall provide a sufficient number (32 vials) of these check standard samples (S) in the standard peanut vials to provide one check standard sample (S) for each block of the 12 blocks of this task plus 18 additional vials which are sufficient to initially prepare sufficient check standard solution (K) for all of this task; 1 vial is for use with the process control samples and the 1 vial remaining is available as a spare for unforeseen emergencies such as a broken or spilled vial. DWPT Analytical Lab shall prepare the check standard solution (K) by performing the dissolution preparation of all the 18 vials of the check standard samples initially; the 18 preparations shall then be combined and mixed to produce a single uniform source of check standard solution (K). A sufficient quantity of the check standard solution (K) shall be prepared initially so that the same identical check standard solution (K) can be used throughout this task. All the analyses results of these check standard solutions (K) shall be reported.

All analyses results are to be corrected for the weight of sample dissolved, aliquots, and dilutions and reported as elemental weight percent of the sample. The sample is a slurry of both insoluble and soluble solids in liquid; the weight percent elemental analyses results shall be based of the weight of the slurry. In the case of samples, this will enable practical reproduction of the sample composition. In the case of the check standard solutions and check standard samples (K, S), this will enable practical reproduction of the composition of the check standard sample. The elemental results will be reported as elemental weight percent of sample for all solutions (K, S, B and samples) analyzed for elemental content. The use of the same identical check standard solution (K) with all the samples in this task is most preferable.

The preparation dissolution and aliquot dilution of samples in any particular block outlined in this statistical design are to be closely coordinated with the elemental analyses by ICP and AA at DWPT Analytical Lab to minimize the time between the sample dissolution and the elemental analyses measurements. These elemental measurements per block, as assigned in this statistical design for ICP and AA, shall be completed by

both instruments within 48 hours following the preparation of the dissolved samples, which will be delivered to the ICP instrument, which will work in conjunction with the AA instrument; this time limit may be extended to 72 hours provided that sufficient tests are performed and documented to assure that these solutions are stable for this length of time. Two different dilutions will most probably be required for the ICP; if so, then an additional 12 hours (a total of 60 hours since the preparation dissolution) shall be allowed for the ICP analyses of the second dilution provided that the second dilution is analyzed within 24 of the first dilution. The analyses sequence of the second dilution shall also be the same as the analyses sequence shown in APPENDIX A for ATTACHMENTS. The aliquot dilutions shall be closely coordinated with the measurement analyses by ICP and AA to minimize the time between aliquot dilution and measurement; this time period is not to exceed 24 hours. ICP and AA elemental measurements shall be made from the same identical aliquot dilution; a second dilution of this aliquot dilution can be made if additional dilution is needed for AA analysis.

DWPF analytical methods as described in "DWPF Analytical Procedures"^[6], shall be used as a guide and a basis for the preparation and analyses of these samples. The precision shall be assessed as part of this task. Good laboratory practices shall be followed during both the sample preparation dissolutions, and the analyses.

Sample analyses for the PRBT process control shall be completed in a timely and organized manner by the DWPT Analytical Laboratory. The analyses for each process control sample are: acid; weight percent insoluble solids; weight percent soluble solids; density; elemental analyses on only the (filtered) aqueous soluble solids by: 1) ICP for Ti, B, Na, K, Cu, Fe, and Mn and 2) AA for K and Na. All of these control sample analyses shall be done by the DWPT Analytical Lab at TNX. One check standard (S) and one blank (B) shall be prepared along with the preparation of the group of process control samples. The check standard solution (K) shall also be analyzed for the desired elements by ICP and AA along with the group of process control samples as well as the check standard (S) and the blank (B). The ICP and AA elemental analyses results of the check standard (S), the blank (B) and the check standard solution (K) shall be reported along with the process control sample elemental analyses results. All measurements at TNX performed by the DWPT Analytical Laboratory will normally be analyzed on the day shift unless authorized otherwise. The analyses results shall be reported on a timely basis immediately following the sample analyses completion either directly or via the TNXLIMS. A hard copy of the process control sample analyses results shall be made available by the next work-day after completion to W. J. Jenkins either to his mailbox or via the SLTNOA All-in-One Network. These samples shall require the efforts of both a chemist, as a supervisor, and a sufficient number of analysts so that the maximum quality of results shall be completed within the minimum possible period of time.

Lab analysts performing the sample analyses required by this task, are to be trained in the specific sample analysis technique. A practice run shall not be required prior to the analyses of actual samples for this task since the DWPT Analytical Lab has already demonstrated that training is sufficient, that the system is proficient and that the

elemental analyses results produced are adequate in a prior test^[7]. The following references:

- 1) "DWPF Analytical Procedures"^{[6],[8]}
- 2) "Procedure for Measuring Total Solids, Insoluble Solids, and Soluble Solids by Microwave Drying and by Calcining"^[9]
- 3) TNX DWPT Lab Operating Procedure for Total Solids, Insoluble Solids and Soluble Solids^[10]
- 4) "Sample Preparation for the Determination of Titanium in the DWPF Precipitate Reactor Aqueous Product"^[11]
- 5) TNX DWPT Lab Operating Procedure for Sample Preparation for the Determination of Titanium^[12]
- 6) "Weight Percent Solids and Density of Slurry Material"^[13]
- 7) "Precision and Accuracy Achieved During Testing of DWPF Analytical Methods"^[14]
- 8) "Basic Data Report: Precision and Accuracy of DWPF Analytical Methods(U)"^[15]
- 9) TNX DWPT Lab Operating Procedure for Acid Determination by Autotitration^[16]
- 10) TNX DWPT Lab Operating Procedure for Direct Elemental Analysis by ICP-AES and AAS^[17]
- 11) TNX DWPT Lab Operating Procedure for Elemental Analysis by Atomic Absorption^{[18],[19]}

are to be used as a guide for the expected sample analysis accuracy, precision, and procedures which shall apply to these samples. The DWPT Analytical Lab is currently reviewing procedures on an established schedule for completion by 30-Jun-92. The procedures will be compared with the ADS procedure recommended for DWPF use and any discrepancies, changes, or deviations will be documented.

The following elements are to be analyzed by Inductively Coupled Plasma(ICP): Ti, B, Na, K, Cu, Fe and Mn. The ICP analyses of all samples at TNX will be run in sequential mode for the desired elemental lines by the DWPT Analytical Lab. Any element in the list above, which is above its reportable limit, shall be reported; the reportable limit of each of these elements shall also be provided. The elements Ti, B, Na and K are the major components of the PRBT. The elements Cu, Fe and Mn are minor components of the PRBT. No prior development work has been done of the recovery of Cu, Fe, and Mn by the H₂SO₄-HF Titanium dissolution, so the accuracy and precision of these analyses

shall be evaluated as a part of this task by the statistical analyses within this task before these results can be utilized.

The elements K and Na are to also be analyzed by Atomic Absorption (AA).

Sample analyses shall be completed in a timely and organized manner by the DWPT Analytical Laboratory. Work shall begin on the statistical samples by 18-Feb-92 and be completed by 30-Mar-92 so as to meet the current DWPF Schedule. Analyses of these samples shall require the dedicated efforts of both an analytical supervisor and a sufficient number of analysts so that the maximum quality of results shall be completed and reported in the minimum practical period of time. Each block of the statistical sample analyses results shall be reported within the following work-day after performing the measurements. The form and content of this report shall be approved by W. J. Jenkins prior to use.

Specific

No sample analysis shall be canceled nor shall any sample be dumped or discarded without the prior written approval or request from W. J. Jenkins. Non-discardable items, with respect to these samples, include the following: samples, all check standard samples, any excess check standard samples and all check standard solutions. Aliquot dilutions of samples may be discarded at the discretion of the chemist responsible for his particular analyses. Seven days after completing and reporting the results of a block of samples, all items described above for non-discard for that particular block may be returned to W. J. Jenkins unless he authorizes their disposal.

The entire contents of all pulled sample vials, which are taken for elemental analysis, are to be used for the preparation dissolution to eliminate errors inherent in taking representative aliquots of the slurry from the sample vial and thus to provide a representative homogeneous product for the analyses.

Dissolutions of the process control samples by H_2SO_4 -HF Titanium method at TNX shall be supervised by T. A. Policke and shall be done by DWPT Analytical Lab technician(s) at TNX. Dissolutions of the statistical samples by H_2SO_4 -HF Titanium method at TNX shall be supervised by T. A. Policke and shall be done by DWPT Analytical Lab technician(s) at TNX on the entire contents of each sample vial. The ICP analyses shall be run on the H_2SO_4 -HF Titanium dissolution for Ti, B, K, Na, Cu, Fe, and Mn. The elements Ti, B, Na and K are the major components of the PRBT. The elements Cu, Fe and Mn are minor components of the PRBT. No prior development work has been done of the recovery of Cu, Fe, and Mn by the H_2SO_4 -HF Titanium dissolution, so the accuracy and precision of these analyses shall be evaluated as a part of this task by the statistical analyses within this task before these results can be utilized. The H_2SO_4 -HF Titanium dissolution shall be used also for the AA analysis of K and Na.

The grab sampler shall be used as the fundamental reference for comparison with the Hydragard™ prototypic sampler for this statistical design. The grab sampler shall be

used at lower elevation equivalent to the elevation of the inlet to the prototypic sampler pump and shall also be used at a higher elevation equivalent to about 6 to 12 inches below the surface of the liquid level.

The volume of sample collected by the grab sampler shall be similar to that collected by the Hydragard™ prototypic sampler; a peanut vial holder for the grab sampler shall be utilized so that the samples are collected directly into a new peanut vial each time.

The TNX Full-Scale SRAT/SME/PRBT tank shall be flushed with water to remove the previous heel and traces of sludge and/or SME slurry until three process control samples indicate that the weight percent insoluble solids content is less than 0.1 weight percent of the total sample.

The TNX Full-Scale SRAT/SME/PRBT liquid level/density probes were calibrated by G. E. Weeks prior to the previous task^[2]; this calibration data was recorded in Laboratory Notebook E56019^[20]. The TNX Full-Scale SRAT/SME/PRBT liquid level/density probes were also calibrated by G. E. Weeks prior to this task; a summary of this calibration data was recorded in Laboratory Notebook E56019^[14] on pages 44 through 47.

The TNX Full-Scale SRAT/SME/PRBT tank volume vs. liquid level has been calibrated by S. M. Peters prior to the previous task^[2] and is described in a separate task plan "Quality Assurance Action Plan - Calibration of the TNX Full-scale SRAT/SME".^[21]

The Distributive Control System (DCS) hardware required for Full-Scale PRBT process control operations shall be verified, by G. E. Weeks, to be fully operational prior to starting this task. The DCS software shall also be verified, by G. E. Weeks, to be fully operational prior to starting this task. Historic trends of liquid level, temperature, and vessel vapor-space pressure for the operations of this task shall be saved to disk for future reference; disk storage of these trends shall be maintained until all data analyses has been completed and the report has been issued for this task.

Operators who do the Full-Scale PRBT operations required for this task shall be trained in these operations.

A log of Full-Scale PRBT operations for this task shall be maintained in Laboratory Notebook E56019^[14]. This log shall summarize all pertinent operations and observations in chronological order along with the dates, times, and names of the technical personnel.

Any calculations or other pertinent facts which are not documented elsewhere shall be documented in Laboratory Notebook E48229^[22] by W. J. Jenkins.

Responsibilities

The Process Modeling and Control Group of the Defense Waste Processing Technology Division has the overall responsibility for completing this task and producing a report summarizing the results by 12-Jun-92. This group also has the responsibility of

developing and testing prototypic sampler equipment and methods for sampling the PRBT tank contents.

Scientific Computations has the responsibility of this task with respect to its statistical foundation and statistical analyses of the sample results. The Sample analyses results shall be provided to Scientific Computations by 07-Apr-92. Scientific Computations shall communicate their results and conclusions to DWPT throughout their statistical analyses of these sample results. Scientific Computations shall complete the statistical analyses of these sample results, including a written report with conclusions, by 06-May-92.

The DWPT Analytical Laboratory is responsible for analyzing the samples, controlling the quality of the analyses results, and meeting time requirements. ADS has the responsibility for the development and technical quality of the analytical methods employed. DWPT Analytical Lab has the responsibility of: 1) Sample Receiving at 772-T, 2) developing the Sample Plans for the TNX Laboratory Information Management System (TNXLIMS), 3) logging all samples for analyses into the TNXLIMS, 4) performing all sample dissolutions, 5) performing ICP and AA measurements of samples, 6) holding the dissolution solutions, 7) sample tracking, 8) reporting the analyses results and 9) performing all process control sample analyses.

TNX Operations Section (TOS) has the responsibility of: 1) operating the Full-Scale PRBT for this task in accordance with this task plan, DWPT written daily activities sheets, and existing approved procedures; 2) transferring the PHA into the PRBT Tank; 3) pulling all samples; 4) delivering the samples to the DWPT.

M. A. Baich has the responsibility for technical support of the TNX operations for this task.

W. J. Jenkins has the overall responsibility for this task and has the following particular responsibilities of: 1) ensuring that all samples are taken in the proper order according to the design; 2) submitting the samples to the DWPT Analytical Laboratory; and 3) entering the data into a database and/or spreadsheet per his choice and/or design; 4) transmitting the data to T. B. Edwards of Scientific Computations Section by 07-Apr-92; and 5) documenting the summary and conclusions in a written report to DWPF by 12-Jun-92.

C. P. Reeve is responsible for the statistical design used for this task. C. P. Reeve is also responsible for reviewing and assisting with the statistical analyses and conclusions for this task.

T. B. Edwards is responsible for performing the statistical analyses of the sample analyses results data and reporting the statistical analyses in a written report by 06-May-92.

L. C. Johnson is responsible for general scheduling and for ensuring the quality and approval of all sample analyses, ICP, AA, density, weight percent insoluble solids and

weight percent soluble solids, performed at TNX. All sample analyses shall be completed and reported by 30-Mar-92.

D. R. Best is responsible for ensuring the quality and approval of all ICP measurements performed by the DWPT Analytical Lab at TNX. This responsibility includes supplying the procedures, all equipment, and securing the manpower necessary for these measurements. D. R. Best is also responsible for assisting T. A. Policke, as needed, on both the preparation-dissolutions of these samples and the AA Analyses measurements.

T. A. Policke is responsible for ensuring the quality and approval of all the H_2SO_4 -HF Titanium dissolutions of samples for elemental analyses by ICP and AA. This responsibility includes supplying the procedures, all equipment and securing the manpower necessary for the dissolutions. T. A. Policke is also responsible for ensuring the quality and approval of all the AA Analyses measurements of these samples. T. A. Policke is also responsible for assuring that sufficient and adequate controls are used for these dissolutions of the samples for elemental analyses at the DWPT Analytical Lab at TNX. T. A. Policke is responsible for assisting D. R. Best, as needed, on the ICP elemental analyses. T. A. Policke is also responsible for the overall sample coordination and is to assure that: 1) all samples are delivered to the appropriate location in the appropriate time frame; 2) the proper work, dissolution, and analyses are performed on the appropriate samples in the appropriate sequence and time frame; 3) the appropriate check standard solution, check standard samples and blanks are delivered to the appropriate location in the appropriate time frame; 4) the proper work, dissolution, and analyses are performed on the appropriate check standard solution, check standard samples and blanks in the appropriate sequence and time frame; and 5) that the analyses results data sheets contain all the appropriate information and are made available to W. J. Jenkins within the next work-day following completion of each block of analyses.

G. E. Weeks is responsible for the calibration of the TNX Full-Scale SRAT/SME/PRBT tank liquid level probes; this was completed prior to the previous SME task^[2] and was also completed again prior to this task. The calibration data has been recorded in Laboratory Notebook E56019^[14]. G. E. Weeks is also responsible for verifying that the DCS Hardware and Software are fully operational and performing properly throughout this task.

Activities

The activities required for the completion of this task are summarized below:

- 1) Flush the SRAT/SME/PRBT tank with water to remove the previous heel and traces of sludge and/or SME slurry until three process control samples indicate that the weight percent insoluble solids content is less than 0.1 weight percent of the total sample. Remove the heel of flush water from the tank.

- 2) Install four baffles (1" thick by 12" width by 150" height, 90 degrees apart, with 2" inside clearance to tank wall) into the Full-Scale SRAT/SME/PRBT tank as per print PV182698 Rev 47.
- 3) Transfer the PHA slurry into the PRBT tank.
- 4) Agitate the slurry at the low agitator speed (67 rpm) for a minimum of 2 hours, measure the liquid level, determine the volume, and take process control samples for analyses of weight percent insoluble solids and weight percent soluble solids, density, acid, and ICP elemental analyses for: Ti, B, K, Na, Cu, Fe, and Mn and AA elemental analysis for K and Na.
- 5) Take samples from the Hydragard™ prototypic liquid sampler and with the grab sampler at two vertical depths as per the statistical design which is shown in detail in the Appendix of this task plan.
- 6) Submit all samples to Sample Receiving at the DWPT Analytical Laboratory, 772-T, TNX.
- 7) Enter the sample analyses results into a spreadsheet and/or data base and submit them to T. B. Edwards of Scientific Computations - Computational Modeling and Applied Statistics for the statistical analyses.

Measurements

The measurement of the in-tank liquid level of the slurry in the PRBT tank shall be required when either an addition is made into the tank or any liquid or solids slurry(excluding samples) is removed from the tank. The agitator shall be turned off just prior to taking a liquid level measurement reading to provide accurate measurement; the agitator shall not remain off for more than 5 minutes to take these readings. The liquid level measurement shall also be measured and recorded both before and after any additions, removals, operations, pump priming and/or sampling. These measurements shall be taken and recorded before the first statistical sample and again after the last statistical sample but not between the individual samples of the series.

The measurement and recording of the liquid slurry temperature and the vessel vapor-space pressure of the PRBT tank shall also be required at the same frequency and same time as the liquid level measurements.

The measurement and analyses required for process control samples are listed below:

- 1) Weight percent insoluble solids and weight percent soluble solids.
- 2) Density.
- 3) Acid.

- 4) ICP elemental analyses for: Ti, B, K, Na, Cu, Fe, and Mn of the soluble solids.
- 5) AA elemental analysis for K and Na of the soluble solids.

The measurement and analyses for the samples required by the statistical design are detailed in the Appendix of this revision to the Task Plan.

Reporting Requirements

The final results and conclusions of this task shall be documented in a written report to DWPF by W. J. Jenkins with the statistical results and conclusions provided by T. B. Edwards' statistical analyses.

Sample analyses results for this task are to be reported in a timely and well organized manner. All process control sample analyses results performed at TNX by the DWPT Analytical Laboratory are to be reported in a timely and well organized manner immediately after completion of the analyses. The analyses report will include: 1) the date and time, 2) both the customer sample number and the laboratory sample number, 3) the analyses sequence number, 4) the results, 5) the analyst's name, 6) the supervisor's approval and 7) any relevant information, such as dilutions, quantity of sample, etc., per the method; items 5), 6) and 7) may instead be on a cover sheet which is attached to the analyses report. Hard copies of these analyses results shall be made available to W. J. Jenkins by the next work-day. Original copies of these results are to be maintained by the DWPT Analytical Laboratory in a well organized manner for future reference for one year.

The preparation dissolution data shall be reported within the next work-day after completion per block of samples; the form and content of this data report sheet is to be approved by W. J. Jenkins. This data sheet shall contain all data relevant to the dissolution preparation of the sample such as: 1) the weight of sample, 2) the weight of dissolved sample solution 3) the date and time that the preparation dissolution was completed, 4) the analyst's name, 5) the sequence number, 6) both the customer sample number and the laboratory sample number, 7) the total dilution factor and 8) any other relevant information per the method.

Analyses results of each day's block of the statistical design samples are to be reported by the following work-day upon completion of that block of samples. The statistical design elemental analyses results are to be reported by DWPT Analytical Lab in a format specifically approved for this task. The analyses report format shall include: 1) both the customer sample number and the laboratory sample number, 2) the analyses sequence number, 3) the analyst's name, 4) the results, 5) the date and time that the analyses was completed, 6) the supervisor's approval and 7) any relevant information, such as dilutions, quantity of sample, etc., per the method; items 5), 6) and 7) may instead be on a separate cover sheet which is attached to the analyses report. Each of the sample analyses results are to be reviewed for correctness and approved by the chemist responsible for the analyses or the laboratory supervisor responsible for the method.

Backup copies of all sample analyses result data sheets are to be maintained by DWPT Analytical Lab in a well organized manner for future reference until this task is declared complete, in a written memo, by DWPT. Copies of these sample analyses result data sheets are to be made available to W. J. Jenkins by the following work-day upon completion of the measurements for that block of samples.

The statistical design sample analyses results shall be entered into a database and/or spreadsheet and transmitted by W. J. Jenkins to T. B. Edwards for statistical analyses. T. B. Edwards is to provide a written report of the statistical analyses of these sample results along with his conclusions by 6-Apr-92 to W. J. Jenkins to incorporate into the written report for DWPF.

Computer Software

Quattro Pro Spreadsheet (Version 3.0 which currently in use or a newer version if updated before this task is completed) shall be used to enter and record sample analyses results. This program is commercially available software that has proved to be satisfactory during previous use. Any calculations done by this software shall be verified by W. J. Jenkins randomly selecting several points and redoing the calculations manually by hand; these calculations shall be documented in laboratory notebook E48229.

Statistical Analysis System(SAS) (version 6.06 which is currently in use at S.R.S. on the IBM main frame or a newer version if SAS is updated before this task is completed) software (licensed per CPU from SAS Institute, Inc., Cary, N.C. 27512) may be used by T. B. Edwards for the Statistical Analyses of the sample results. SAS software is maintained on all IBM mainframes by L. S. Carter, SAS systems programmer, of Operating Systems and Database of Computer Telecommunications Management at S.R.S..

CSS(Complete Statistical System) Software(Version 2.1 which is currently in use or a newer version if updated before this task is completed) may also be used for the Statistical Analyses of these sample results. CSS is commercially available software which is licensed by StatSoft, Inc., 2325 East 13th Street, Tulsa, Oklahoma 741004. This software is commercially available software that has proved satisfactory during previous use. Other commercially available software that has proved satisfactory during previous use may also be used, however, the software used in this task must be identified in the written report. C. P. Reeve shall be responsible for assuring that any commercial software is used properly and calculations done by this software are correct based on his technical expertise.

REFERENCES:

- [1] C. T. Randall, "Program Plan - DWPF Process Control (Control of Melter Feed Composition)", DWPTQA-89-0037, August 8, 1989.

**APPENDIX A
for
ATTACHMENTS**

03-Feb-92

Customer Identification Numbering System

Field	Field Length	Field Entry:	Field Description
#1	1 digit	1:	Preparation
#2	1 letter	R:	Hydragard™ Sampler
		G:	Lo Grab Sampler
		H:	Hi Grab Sampler
#3	2 digits	Vial #:	01 to 36
#4	1 digit	1:	First Aliquot
		2:	Second Aliquot
#5	1 digit	1:	First Measurement
		2:	Second Measurement

NOTE:

P ≡ H₂SO₄-HF Titanium Preparation-Dissolution Method

TABLE 1
DESIGN SIZE

Method	Analyses	Dissolution	X-Y Contrasts	Nx	Ny	Δ
ICP	Ti, K, Na, B	P	R-G	48	48	3.3%
ICP	Ti, K, Na, B	P	G-H	48	48	3.3%
ICP	Cu, Fe, Mn	P	R-G	48	48	3.3%
ICP	Cu, Fe, Mn	P	G-H	48	48	3.3%
AA	K, Na	P	R-G	48	48	3.3%
AA	K, Na	P	G-H	48	48	3.3%

NOTES:

P $\equiv$ H₂SO₄-HF Titanium Preparation-Dissolution Method

03-Feb-92

Allocation of the Degrees of Freedom

This test considers the following levels of the various sources of uncertainty:

<u>Source</u>	<u>Levels</u>
Samplers	R: Hydraguard™, G: Grab(Lo), H: Grab(Hi)
Vials/Sampler	1,2,3,4,5,6,7,8,9,10,11,12 (R, G & H Samplers)
Preps/Vial	1 (R, G & H Samplers)
Aliquots/Prep	1,2 (R, G & H Samplers)
Measurements/Aliquot	1,2 (R, G & H Samplers)

PRBT SAMPLER BLOCKING PATTERN & VIAL NUMBERING (ICP & AA)
Page 1 of 2

R: HYDRAGUARD™ SAMPLER
 G: LO GRAB SAMPLER
 H: HI GRAB SAMPLER

<u>Replication</u>		<u>Block</u>		↓ DIRECT CONTRAST
I	(1)	Sampler:	(R - G - H)	
		Vial:	01 02 03	
	(2)	Sampler:	(R - G - H)	
		Vial:	04 05 06	
	(3)	Sampler:	(H - G - R)	
		Vial:	07 08 09	
II	(4)	Sampler:	(R - G - H)	
		Vial:	10 11 12	
	(5)	Sampler:	(H - G - R)	
		Vial:	13 14 15	
	(6)	Sampler:	(R - G - H)	
		Vial:	16 17 18	

6

PRBT SAMPLER BLOCKING PATTERN & VIAL NUMBERING (ICP & AA)
Page 2 of 2

R: HYDRAGUARD™ SAMPLER
 G: LO GRAB SAMPLER
 H: HI GRAB SAMPLER

<u>Replication</u>	<u>Block</u>		↓ DIRECT CONTRAST
III	(7)	Sampler:	(R - G - H)
		Vial:	19 20 21
	(8)	Sampler:	(R - G - H)
		Vial:	22 23 24
	(9)	Sampler:	(H - G - R)
		Vial:	25 26 27
IV	(10)	Sampler:	(R - G - H)
		Vial:	28 29 30
	(11)	Sampler:	(H - G - R)
		Vial:	31 32 33
	(12)	Sampler:	(R - G - H)
		Vial:	34 35 36

<u>CONTRAST</u>	<u>ESTIMATES</u>	<u>DIRECT</u>	<u>INDIRECT</u>
R - G	SAMPLER ACCURACY	12	0
G - H	TANK HOMOGENEITY	12	0

FIGURE 1A

SUMMARY OF THE VARIOUS TREES

ANALYSES

ICP (Ti, K, Na, B, Cu, Fe, & Mn)

SAMPLERS

VIAL PAIRS

PREPS/DISSOLUTIONS

ALIQUOTS

MEASUREMENTS

SAMPLERS

VIAL PAIRS

PREPS/DISSOLUTIONS

ALIQUOTS

MEASUREMENTS

NOTES:

P = H₂SO₄-HF Titanium Dissolution Method

FIGURE 1B

SUMMARY OF THE VARIOUS TREES

ANALYSES

AA (K & Na)

SAMPLERS

VIAL PAIRS

PREPS/DISSOLUTIONS

ALIQUOTS

MEASUREMENTS

SAMPLERS

VIAL PAIRS

PREPS/DISSOLUTIONS

ALIQUOTS

MEASUREMENTS

NOTES:

P = H₂SO₄-HF Titanium Dissolution Method

Sampling Order for Task Plan for Test of PRBT Prototypic Liquid Sampler WSRC-RP-90-1055

Order	Sample	Block	Rep	Date Taken	Time Taken	Operator Initials
1	1 R 01	1	1	___/___/___	___:___	_____
2	1 G 02	1	1	___/___/___	___:___	_____
3	1 H 03	1	1	___/___/___	___:___	_____
4	1 R 04	2	1	___/___/___	___:___	_____
5	1 G 05	2	1	___/___/___	___:___	_____
6	1 H 06	2	1	___/___/___	___:___	_____
7	1 H 07	3	1	___/___/___	___:___	_____
8	1 G 08	3	1	___/___/___	___:___	_____
9	1 R 09	3	1	___/___/___	___:___	_____
10	1 R 10	4	2	___/___/___	___:___	_____
11	1 G 11	4	2	___/___/___	___:___	_____
12	1 H 12	4	2	___/___/___	___:___	_____
13	1 H 13	5	2	___/___/___	___:___	_____
14	1 G 14	5	2	___/___/___	___:___	_____
15	1 R 15	5	2	___/___/___	___:___	_____
16	1 R 16	6	2	___/___/___	___:___	_____
17	1 G 17	6	2	___/___/___	___:___	_____
18	1 H 18	6	2	___/___/___	___:___	_____
19	1 R 19	7	3	___/___/___	___:___	_____
20	1 G 20	7	3	___/___/___	___:___	_____
21	1 H 21	7	3	___/___/___	___:___	_____
22	1 R 22	8	3	___/___/___	___:___	_____
23	1 G 23	8	3	___/___/___	___:___	_____
24	1 H 24	8	3	___/___/___	___:___	_____
25	1 H 25	9	3	___/___/___	___:___	_____
26	1 G 26	9	3	___/___/___	___:___	_____
27	1 R 27	9	3	___/___/___	___:___	_____
28	1 R 28	10	4	___/___/___	___:___	_____
29	1 G 29	10	4	___/___/___	___:___	_____
30	1 H 30	10	4	___/___/___	___:___	_____
31	1 H 31	11	4	___/___/___	___:___	_____
32	1 G 32	11	4	___/___/___	___:___	_____
33	1 R 33	11	4	___/___/___	___:___	_____
34	1 R 34	12	4	___/___/___	___:___	_____
35	1 G 35	12	4	___/___/___	___:___	_____
36	1 H 36	12	4	___/___/___	___:___	_____

This Preparation Dissolution Sequence for use with Task Plan for
Test of PRBT Prototypic Liquid Sampler WSRC-RB-1055 by W. J. Jenkins

Customer Sample ID#s are shown inside the boxes.

1 & 2 Prefix of the Customer ID# => Replicate Vials of Sample.

P => H₂SO₄-HF Ti Dissolution

Dissolution Sequence	Prep		Blk	Rep
1	P	1 R 01 , 1 G 02 , 1 H 03	1	1
2	P	1 R 04 , 1 G 05 , 1 H 06	2	1
3	P	1 H 07 , 1 G 08 , 1 R 09	3	1
4	P	1 R 10 , 1 G 11 , 1 H 12	4	2
5	P	1 H 13 , 1 G 14 , 1 R 15	5	2
6	P	1 R 16 , 1 G 17 , 1 H 18	6	2

This Preparation Dissolution Sequence for use with Task Plan for
Test of PRBT Prototypic Liquid Sampler WSRC-RB-1055 by W. J. Jenkins

Customer Sample ID#s are shown inside the boxes.

1 & 2 Prefix of the Customer ID# => Replicate Vials of Sample.

P => H2SO4-HF Ti Dissolution

Dissolution Sequence	Prep		Blk	Rep
7	P	1 R 19 , 1 G 20 , 1 H 21	7	3
8	P	1 R 22 , 1 G 23 , 1 H 24	8	3
9	P	1 H 25 , 1 G 26 , 1 R 27	9	3
10	P	1 R 28 , 1 G 29 , 1 H 30	10	4
11	P	1 H 31 , 1 G 32 , 1 R 33	11	4
12	P	1 R 34 , 1 G 35 , 1 H 36	12	4

This Sequence of Sample Analyses for use with Task Plan for
 Test of PRBT Prototypic Liquid Sampler WSRC-RP-90-1055 Rev.1 by W. J. Jenkins.
 1 Prefix of Customer ID#=> Sample Vials for H2SO4-HF Ti Dissolution,

K=> Check Std.Soln, S=> STD, B=> Blank, PD=> Prep. Dissolution

I=> Lab Instrument Check Std., C => Lab Instrument Calibration.

The C,I can be omitted between the blocks when more than 1 block is done on the same day.

Sequence	CUSTOMER ID#	PD	ICP	AA K	DAY	BLK	REP
	C,I				1	1	
1	1 K 1 1 1	P1	ICP A1-M1	AA-A1-M1	1	1	1
2	1 R 01 1 1	P1	ICP A1-M1	AA-A1-M1	1	1	1
3	1 G 02 1 1	P1	ICP A1-M1	AA-A1-M1	1	1	1
4	1 H 03 1 1	P1	ICP A1-M1	AA-A1-M1	1	1	1
5	1 K 1 1 2	P1	ICP A1-M1	AA-A1-M1	1	1	1
6	1 R 01 2 1	P1	ICP A2-M1	AA-A2-M1	1	1	1
7	1 G 02 2 1	P1	ICP A2-M1	AA-A2-M1	1	1	1
8	1 H 03 2 1	P1	ICP A2-M1	AA-A2-M1	1	1	1
9	1 K 1 1 3	P1	ICP A1-M1	AA-A1-M1	1	1	1
10	1 S 1 1 4	P1	ICP A1-M1	AA-A1-M1	1	1	1
11	1 B 1 1 5	P1	ICP A1-M1	AA-A1-M1	1	1	1
	I				1	1	
12	1 R 01 1 2	P1	ICP A1-M2	AA-A1-M2	1	1	1
13	1 G 02 1 2	P1	ICP A1-M2	AA-A1-M2	1	1	1
14	1 H 03 1 2	P1	ICP A1-M2	AA-A1-M2	1	1	1
15	1 K 1 2 6	P1	ICP A1-M2	AA-A1-M2	1	1	1
16	1 R 01 2 2	P1	ICP A2-M2	AA-A2-M2	1	1	1
17	1 G 02 2 2	P1	ICP A2-M2	AA-A2-M2	1	1	1
18	1 H 03 2 2	P1	ICP A2-M2	AA-A2-M2	1	1	1
19	1 K 1 2 7	P1	ICP A1-M2	AA-A1-M2	1	1	1
	I				1	1	
	C,I				2	2	
20	1 K 1 1 8	P1	ICP A1-M1	AA-A1-M1	2	2	1
21	1 R 04 1 1	P1	ICP A1-M1	AA-A1-M1	2	2	1
22	1 G 05 1 1	P1	ICP A1-M1	AA-A1-M1	2	2	1
23	1 H 06 1 1	P1	ICP A1-M1	AA-A1-M1	2	2	1
24	1 K 1 1 9	P1	ICP A1-M1	AA-A1-M1	2	2	1
25	1 R 04 2 1	P1	ICP A2-M1	AA-A2-M1	2	2	1
26	1 G 05 2 1	P1	ICP A2-M1	AA-A2-M1	2	2	1
27	1 H 06 2 1	P1	ICP A2-M1	AA-A2-M1	2	2	1
28	1 K 1 1 10	P1	ICP A1-M1	AA-A1-M1	2	2	1
29	1 S 1 1 11	P1	ICP A1-M1	AA-A1-M1	2	2	1
30	1 B 1 1 12	P1	ICP A1-M1	AA-A1-M1	2	2	1
	I				2	2	
31	1 R 04 1 2	P1	ICP A1-M2	AA-A1-M2	2	2	1
32	1 G 05 1 2	P1	ICP A1-M2	AA-A1-M2	2	2	1
33	1 H 06 1 2	P1	ICP A1-M2	AA-A1-M2	2	2	1
34	1 K 1 2 13	P1	ICP A1-M2	AA-A1-M2	2	2	1
35	1 R 04 2 2	P1	ICP A2-M2	AA-A2-M2	2	2	1
36	1 G 05 2 2	P1	ICP A2-M2	AA-A2-M2	2	2	1
37	1 H 06 2 2	P1	ICP A2-M2	AA-A2-M2	2	2	1
38	1 K 1 2 14	P1	ICP A1-M2	AA-A1-M2	2	2	1
	I				2	2	

This Sequence of Sample Analyses for use with Task Plan for
 Test of PRBT Prototypic Liquid Sampler WSRC-RP-90-1055 Rev.1 by W. J. Jenkins.
 1 Prefix of Customer ID#=> Sample Vials for H2SO4-HF Ti Dissolution,

K=> Check Std.Soln, S=> STD, B=> Blank, PD => Prep. Dissolution

I=> Lab Instrument Check Std., C => Lab Instrument Calibration.

The C,I can be omitted between the blocks when more than 1 block is done on the same day.

Sequence	CUSTOMER ID#	PD	ICP	AA K	DAY	BLK	REP
	C,I				3	3	
39	1 K 1 1 15	P1	ICP A1-M1	AA-A1-M1	3	3	1
40	1 H 07 1 1	P1	ICP A1-M1	AA-A1-M1	3	3	1
41	1 G 08 1 1	P1	ICP A1-M1	AA-A1-M1	3	3	1
42	1 R 09 1 1	P1	ICP A1-M1	AA-A1-M1	3	3	1
43	1 K 1 1 16	P1	ICP A1-M1	AA-A1-M1	3	3	1
44	1 H 07 2 1	P1	ICP A2-M1	AA-A2-M1	3	3	1
45	1 G 08 2 1	P1	ICP A2-M1	AA-A2-M1	3	3	1
46	1 R 09 2 1	P1	ICP A2-M1	AA-A2-M1	3	3	1
47	1 K 1 1 17	P1	ICP A1-M1	AA-A1-M1	3	3	1
48	1 S 1 1 18	P1	ICP A1-M1	AA-A1-M1	3	3	1
49	1 B 1 1 19	P1	ICP A1-M1	AA-A1-M1	3	3	1
	I				3	3	
50	1 H 07 1 2	P1	ICP A1-M2	AA-A1-M2	3	3	1
51	1 G 08 1 2	P1	ICP A1-M2	AA-A1-M2	3	3	1
52	1 R 09 1 2	P1	ICP A1-M2	AA-A1-M2	3	3	1
53	1 K 1 2 20	P1	ICP A1-M2	AA-A1-M2	3	3	1
54	1 H 07 2 2	P1	ICP A2-M2	AA-A2-M2	3	3	1
55	1 G 08 2 2	P1	ICP A2-M2	AA-A2-M2	3	3	1
56	1 R 09 2 2	P1	ICP A2-M2	AA-A2-M2	3	3	1
57	1 K 1 2 21	P1	ICP A1-M2	AA-A1-M2	3	3	1
	I				3	3	
	C,I				4	4	
58	1 K 1 1 22	P1	ICP A1-M1	AA-A1-M1	4	4	2
59	1 R 10 1 1	P1	ICP A1-M1	AA-A1-M1	4	4	2
60	1 G 11 1 1	P1	ICP A1-M1	AA-A1-M1	4	4	2
61	1 H 12 1 1	P1	ICP A1-M1	AA-A1-M1	4	4	2
62	1 K 1 1 23	P1	ICP A1-M1	AA-A1-M1	4	4	2
63	1 R 10 2 1	P1	ICP A2-M1	AA-A2-M1	4	4	2
64	1 G 11 2 1	P1	ICP A2-M1	AA-A2-M1	4	4	2
65	1 H 12 2 1	P1	ICP A2-M1	AA-A2-M1	4	4	2
66	1 K 1 1 24	P1	ICP A1-M1	AA-A1-M1	4	4	2
67	1 S 1 1 25	P1	ICP A1-M1	AA-A1-M1	4	4	2
68	1 B 1 1 26	P1	ICP A1-M1	AA-A1-M1	4	4	2
	I				4	4	
69	1 R 10 1 2	P1	ICP A1-M2	AA-A1-M2	4	4	2
70	1 G 11 1 2	P1	ICP A1-M2	AA-A1-M2	4	4	2
71	1 H 12 1 2	P1	ICP A1-M2	AA-A1-M2	4	4	2
72	1 K 1 2 27	P1	ICP A1-M2	AA-A1-M2	4	4	2
73	1 R 10 2 2	P1	ICP A2-M2	AA-A2-M2	4	4	2
74	1 G 11 2 2	P1	ICP A2-M2	AA-A2-M2	4	4	2
75	1 H 12 2 2	P1	ICP A2-M2	AA-A2-M2	4	4	2
76	1 K 1 2 28	P1	ICP A1-M2	AA-A1-M2	4	4	2
	I				4	4	

This Sequence of Sample Analyses for use with Task Plan for
 Test of PRBT Prototypic Liquid Sampler WSRC-RP-90-1055 Rev.1 by W. J. Jenkins.
 1 Prefix of Customer ID#=> Sample Vials for H2SO4-HF Ti Dissolution,

K=> Check Std.Soln, S=> STD, B=> Blank, PD => Prep. Dissolution

I=> Lab Instrument Check Std., C => Lab Instrument Calibration.

The C,I can be omitted between the blocks when more than 1 block is done on the same day.

Sequence	CUSTOMER ID#	PD	ICP	AA K	DAY	BLK	REP
	C,I				5	5	
77	1 K 1 1 29	P1	ICP A1-M1	AA-A1-M1	5	5	2
78	1 H 13 1 1	P1	ICP A1-M1	AA-A1-M1	5	5	2
79	1 G 14 1 1	P1	ICP A1-M1	AA-A1-M1	5	5	2
80	1 R 15 1 1	P1	ICP A1-M1	AA-A1-M1	5	5	2
81	1 K 1 1 30	P1	ICP A1-M1	AA-A1-M1	5	5	2
82	1 H 13 2 1	P1	ICP A2-M1	AA-A2-M1	5	5	2
83	1 G 14 2 1	P1	ICP A2-M1	AA-A2-M1	5	5	2
84	1 R 15 2 1	P1	ICP A2-M1	AA-A2-M1	5	5	2
85	1 K 1 1 31	P1	ICP A1-M1	AA-A1-M1	5	5	2
86	1 S 1 1 32	P1	ICP A1-M1	AA-A1-M1	5	5	2
87	1 B 1 1 33	P1	ICP A1-M1	AA-A1-M1	5	5	2
	I				5	5	
88	1 H 13 1 2	P1	ICP A1-M2	AA-A1-M2	5	5	2
89	1 G 14 1 2	P1	ICP A1-M2	AA-A1-M2	5	5	2
90	1 R 15 1 2	P1	ICP A1-M2	AA-A1-M2	5	5	2
91	1 K 1 2 34	P1	ICP A1-M2	AA-A1-M2	5	5	2
92	1 H 13 2 2	P1	ICP A2-M2	AA-A2-M2	5	5	2
93	1 G 14 2 2	P1	ICP A2-M2	AA-A2-M2	5	5	2
94	1 R 15 2 2	P1	ICP A2-M2	AA-A2-M2	5	5	2
95	1 K 1 2 35	P1	ICP A1-M2	AA-A1-M2	5	5	2
	I				5	5	
	C,I				6	6	
96	1 K 1 1 36	P1	ICP A1-M1	AA-A1-M1	6	6	2
97	1 R 16 1 1	P1	ICP A1-M1	AA-A1-M1	6	6	2
98	1 G 17 1 1	P1	ICP A1-M1	AA-A1-M1	6	6	2
99	1 H 18 1 1	P1	ICP A1-M1	AA-A1-M1	6	6	2
100	1 K 1 1 37	P1	ICP A1-M1	AA-A1-M1	6	6	2
101	1 R 16 2 1	P1	ICP A2-M1	AA-A2-M1	6	6	2
102	1 G 17 2 1	P1	ICP A2-M1	AA-A2-M1	6	6	2
103	1 H 18 2 1	P1	ICP A2-M1	AA-A2-M1	6	6	2
104	1 K 1 1 38	P1	ICP A1-M1	AA-A1-M1	6	6	2
105	1 S 1 1 39	P1	ICP A1-M1	AA-A1-M1	6	6	2
106	1 B 1 1 40	P1	ICP A1-M1	AA-A1-M1	6	6	2
	I				6	6	
107	1 R 16 1 2	P1	ICP A1-M2	AA-A1-M2	6	6	2
108	1 G 17 1 2	P1	ICP A1-M2	AA-A1-M2	6	6	2
109	1 H 18 1 2	P1	ICP A1-M2	AA-A1-M2	6	6	2
110	1 K 1 2 41	P1	ICP A1-M2	AA-A1-M2	6	6	2
111	1 R 16 2 2	P1	ICP A2-M2	AA-A2-M2	6	6	2
112	1 G 17 2 2	P1	ICP A2-M2	AA-A2-M2	6	6	2
113	1 H 18 2 2	P1	ICP A2-M2	AA-A2-M2	6	6	2
114	1 K 1 2 42	P1	ICP A1-M2	AA-A1-M2	6	6	2
	I				6	6	

This Sequence of Sample Analyses for use with Task Plan for
 Test of PRBT Prototypic Liquid Sampler WSRC-RP-90-1055 Rev.1 by W. J. Jenkins.
 1 Prefix of Customer ID#=> Sample Vials for H2SO4-HF Ti Dissolution,

K=> Check Std.Soln, S=> STD, B=> Blank, PD=> Prep. Dissolution

I=> Lab Instrument Check Std., C => Lab Instrument Calibration.

The C,I can be omitted between the blocks when more than 1 block is done on the same day.

Sequence	CUSTOMER ID#	PD	ICP	AA K	DAY	BLK	REP
	C,I				7	7	
115	1 K 1 1 43	P1	ICP A1-M1	AA-A1-M1	7	7	3
116	1 R 19 1 1	P1	ICP A1-M1	AA-A1-M1	7	7	3
117	1 G 20 1 1	P1	ICP A1-M1	AA-A1-M1	7	7	3
118	1 H 21 1 1	P1	ICP A1-M1	AA-A1-M1	7	7	3
119	1 K 1 1 44	P1	ICP A1-M1	AA-A1-M1	7	7	3
120	1 R 19 2 1	P1	ICP A2-M1	AA-A2-M1	7	7	3
121	1 G 20 2 1	P1	ICP A2-M1	AA-A2-M1	7	7	3
122	1 H 21 2 1	P1	ICP A2-M1	AA-A2-M1	7	7	3
123	1 K 1 1 45	P1	ICP A1-M1	AA-A1-M1	7	7	3
124	1 S 1 1 46	P1	ICP A1-M1	AA-A1-M1	7	7	3
125	1 B 1 1 47	P1	ICP A1-M1	AA-A1-M1	7	7	3
	I				7	7	
126	1 R 19 1 2	P1	ICP A1-M2	AA-A1-M2	7	7	3
127	1 G 20 1 2	P1	ICP A1-M2	AA-A1-M2	7	7	3
128	1 H 21 1 2	P1	ICP A1-M2	AA-A1-M2	7	7	3
129	1 K 1 2 48	P1	ICP A1-M2	AA-A1-M2	7	7	3
130	1 R 19 2 2	P1	ICP A2-M2	AA-A2-M2	7	7	3
131	1 G 20 2 2	P1	ICP A2-M2	AA-A2-M2	7	7	3
132	1 H 21 2 2	P1	ICP A2-M2	AA-A2-M2	7	7	3
133	1 K 1 2 49	P1	ICP A1-M2	AA-A1-M2	7	7	3
	I				7	7	
	C,I				8	8	
134	1 K 1 1 50	P1	ICP A1-M1	AA-A1-M1	8	8	3
135	1 R 22 1 1	P1	ICP A1-M1	AA-A1-M1	8	8	3
136	1 G 23 1 1	P1	ICP A1-M1	AA-A1-M1	8	8	3
137	1 H 24 1 1	P1	ICP A1-M1	AA-A1-M1	8	8	3
138	1 K 1 1 51	P1	ICP A1-M1	AA-A1-M1	8	8	3
139	1 R 22 2 1	P1	ICP A2-M1	AA-A2-M1	8	8	3
140	1 G 23 2 1	P1	ICP A2-M1	AA-A2-M1	8	8	3
141	1 H 24 2 1	P1	ICP A2-M1	AA-A2-M1	8	8	3
142	1 K 1 1 52	P1	ICP A1-M1	AA-A1-M1	8	8	3
143	1 S 1 1 53	P1	ICP A1-M1	AA-A1-M1	8	8	3
144	1 B 1 1 54	P1	ICP A1-M1	AA-A1-M1	8	8	3
	I				8	8	
145	1 R 22 1 2	P1	ICP A1-M2	AA-A1-M2	8	8	3
146	1 G 23 1 2	P1	ICP A1-M2	AA-A1-M2	8	8	3
147	1 H 24 1 2	P1	ICP A1-M2	AA-A1-M2	8	8	3
148	1 K 1 2 55	P1	ICP A1-M2	AA-A1-M2	8	8	3
149	1 R 22 2 2	P1	ICP A2-M2	AA-A2-M2	8	8	3
150	1 G 23 2 2	P1	ICP A2-M2	AA-A2-M2	8	8	3
151	1 H 24 2 2	P1	ICP A2-M2	AA-A2-M2	8	8	3
152	1 K 1 2 56	P1	ICP A1-M2	AA-A1-M2	8	8	3
	I				8	8	

This Sequence of Sample Analyses for use with Task Plan for
 Test of PRBT Prototypic Liquid Sampler WSRC-RP-90-1055 Rev.1 by W. J. Jenkins.
 1 Prefix of Customer ID#=> Sample Vials for H2SO4-HF Ti Dissolution,

K=> Check Std.Soln, S=> STD, B=> Blank, PD=> Prep. Dissolution

I=> Lab Instrument Check Std., C=> Lab Instrument Calibration.

The C,I can be omitted between the blocks when more than 1 block is done on the same day.

Sequence	CUSTOMER ID#	PD	ICP	AA K	DAY	BLK	REP
	C,I				9	9	
153	1 K 1 1 57	P1	ICP A1-M1	AA-A1-M1	9	9	3
154	1 H 25 1 1	P1	ICP A1-M1	AA-A1-M1	9	9	3
155	1 G 26 1 1	P1	ICP A1-M1	AA-A1-M1	9	9	3
156	1 R 27 1 1	P1	ICP A1-M1	AA-A1-M1	9	9	3
157	1 K 1 1 58	P1	ICP A1-M1	AA-A1-M1	9	9	3
158	1 H 25 2 1	P1	ICP A2-M1	AA-A2-M1	9	9	3
159	1 G 26 2 1	P1	ICP A2-M1	AA-A2-M1	9	9	3
160	1 R 27 2 1	P1	ICP A2-M1	AA-A2-M1	9	9	3
161	1 K 1 1 59	P1	ICP A1-M1	AA-A1-M1	9	9	3
162	1 S 1 1 60	P1	ICP A1-M1	AA-A1-M1	9	9	3
163	1 B 1 1 61	P1	ICP A1-M1	AA-A1-M1	9	9	3
	I				9	9	
164	1 H 25 1 2	P1	ICP A1-M2	AA-A1-M2	9	9	3
165	1 G 26 1 2	P1	ICP A1-M2	AA-A1-M2	9	9	3
166	1 R 27 1 2	P1	ICP A1-M2	AA-A1-M2	9	9	3
167	1 K 1 2 62	P1	ICP A1-M2	AA-A1-M2	9	9	3
168	1 H 25 2 2	P1	ICP A2-M2	AA-A2-M2	9	9	3
169	1 G 26 2 2	P1	ICP A2-M2	AA-A2-M2	9	9	3
170	1 R 27 2 2	P1	ICP A2-M2	AA-A2-M2	9	9	3
171	1 K 1 2 63	P1	ICP A1-M2	AA-A1-M2	9	9	3
	I				9	9	
	C,I				10	10	
172	1 K 1 1 64	P1	ICP A1-M1	AA-A1-M1	10	10	4
173	1 R 28 1 1	P1	ICP A1-M1	AA-A1-M1	10	10	4
174	1 G 29 1 1	P1	ICP A1-M1	AA-A1-M1	10	10	4
175	1 H 30 1 1	P1	ICP A1-M1	AA-A1-M1	10	10	4
176	1 K 1 1 65	P1	ICP A1-M1	AA-A1-M1	10	10	4
177	1 R 28 2 1	P1	ICP A2-M1	AA-A2-M1	10	10	4
178	1 G 29 2 1	P1	ICP A2-M1	AA-A2-M1	10	10	4
179	1 H 30 2 1	P1	ICP A2-M1	AA-A2-M1	10	10	4
180	1 K 1 1 66	P1	ICP A1-M1	AA-A1-M1	10	10	4
181	1 S 1 1 67	P1	ICP A1-M1	AA-A1-M1	10	10	4
182	1 B 1 1 68	P1	ICP A1-M1	AA-A1-M1	10	10	4
	I				10	10	
183	1 R 28 1 2	P1	ICP A1-M2	AA-A1-M2	10	10	4
184	1 G 29 1 2	P1	ICP A1-M2	AA-A1-M2	10	10	4
185	1 H 30 1 2	P1	ICP A1-M2	AA-A1-M2	10	10	4
186	1 K 1 2 69	P1	ICP A1-M2	AA-A1-M2	10	10	4
187	1 R 28 2 2	P1	ICP A2-M2	AA-A2-M2	10	10	4
188	1 G 29 2 2	P1	ICP A2-M2	AA-A2-M2	10	10	4
189	1 H 30 2 2	P1	ICP A2-M2	AA-A2-M2	10	10	4
190	1 K 1 2 70	P1	ICP A1-M2	AA-A1-M2	10	10	4
	I				10	10	

This Sequence of Sample Analyses for use with Task Plan for
 Test of PRBT Prototypic Liquid Sampler WSRC-RP-90-1055 Rev.1 by W. J. Jenkins.
 1 Prefix of Customer ID#=> Sample Vials for H2SO4-HF Ti Dissolution,

K=> Check Std.Soln, S=> STD, B=> Blank, PD => Prep. Dissolution

I=> Lab Instrument Check Std., C => Lab Instrument Calibration.

The C,I can be omitted between the blocks when more than 1 block is done on the same day.

Sequence	CUSTOMER ID#	PD	ICP	AA K	DAY	BLK	REP
	C,I				11	11	
191	1 K 1 1 71	P1	ICP A1-M1	AA-A1-M1	11	11	4
192	1 H 31 1 1	P1	ICP A1-M1	AA-A1-M1	11	11	4
193	1 G 32 1 1	P1	ICP A1-M1	AA-A1-M1	11	11	4
194	1 R 33 1 1	P1	ICP A1-M1	AA-A1-M1	11	11	4
195	1 K 1 1 72	P1	ICP A1-M1	AA-A1-M1	11	11	4
196	1 H 31 2 1	P1	ICP A2-M1	AA-A2-M1	11	11	4
197	1 G 32 2 1	P1	ICP A2-M1	AA-A2-M1	11	11	4
198	1 R 33 2 1	P1	ICP A2-M1	AA-A2-M1	11	11	4
199	1 K 1 1 73	P1	ICP A1-M1	AA-A1-M1	11	11	4
200	1 S 1 1 74	P1	ICP A1-M1	AA-A1-M1	11	11	4
201	1 B 1 1 75	P1	ICP A1-M1	AA-A1-M1	11	11	4
	I				11	11	
202	1 H 31 1 2	P1	ICP A1-M2	AA-A1-M2	11	11	4
203	1 G 32 1 2	P1	ICP A1-M2	AA-A1-M2	11	11	4
204	1 R 33 1 2	P1	ICP A1-M2	AA-A1-M2	11	11	4
205	1 K 1 2 76	P1	ICP A1-M2	AA-A1-M2	11	11	4
206	1 H 31 2 2	P1	ICP A2-M2	AA-A2-M2	11	11	4
207	1 G 32 2 2	P1	ICP A2-M2	AA-A2-M2	11	11	4
208	1 R 33 2 2	P1	ICP A2-M2	AA-A2-M2	11	11	4
209	1 K 1 2 77	P1	ICP A1-M2	AA-A1-M2	11	11	4
	I				11	11	
	C,I				12	12	
210	1 K 1 1 78	P1	ICP A1-M1	AA-A1-M1	12	12	4
211	1 R 34 1 1	P1	ICP A1-M1	AA-A1-M1	12	12	4
212	1 G 35 1 1	P1	ICP A1-M1	AA-A1-M1	12	12	4
213	1 H 36 1 1	P1	ICP A1-M1	AA-A1-M1	12	12	4
214	1 K 1 1 79	P1	ICP A1-M1	AA-A1-M1	12	12	4
215	1 R 34 2 1	P1	ICP A2-M1	AA-A2-M1	12	12	4
216	1 G 35 2 1	P1	ICP A2-M1	AA-A2-M1	12	12	4
217	1 H 36 2 1	P1	ICP A2-M1	AA-A2-M1	12	12	4
218	1 K 1 1 80	P1	ICP A1-M1	AA-A1-M1	12	12	4
219	1 S 1 1 81	P1	ICP A1-M1	AA-A1-M1	12	12	4
220	1 B 1 1 82	P1	ICP A1-M1	AA-A1-M1	12	12	4
	I				12	12	
221	1 R 34 1 2	P1	ICP A1-M2	AA-A1-M2	12	12	4
222	1 G 35 1 2	P1	ICP A1-M2	AA-A1-M2	12	12	4
223	1 H 36 1 2	P1	ICP A1-M2	AA-A1-M2	12	12	4
224	1 K 1 2 83	P1	ICP A1-M2	AA-A1-M2	12	12	4
225	1 R 34 2 2	P1	ICP A2-M2	AA-A2-M2	12	12	4
226	1 G 35 2 2	P1	ICP A2-M2	AA-A2-M2	12	12	4
227	1 H 36 2 2	P1	ICP A2-M2	AA-A2-M2	12	12	4
228	1 K 1 2 84	P1	ICP A1-M2	AA-A1-M2	12	12	4
	I				12	12	

END

**DATE
FILMED**

12 / 15 / 92