CONF-880791--1

CONF-880791--1

DE88 015290

SURFACE IMAGING BY POSITRON EMISSION

L. D. Hulett, Jr.
Oak Ridge National Laboratory
Oak Ridge, Tennessee 37831

S. Pendyala
State University of New York, Fredonia
Fredonia, NY

Received by OSTI

AUG 1 1988

"The submitted manuscript has been authored by a contractor of the U.S. Government under contract No. DE-AC05-84OR21400. Accordingly, the U.S. Government retains a nonexclusive, royalty-free license to publish or reproduce the published form of this contribution, or allow others to do so, for U.S. Government purposes."

DISCLAIMER

This report was prepared as an account of work sponsored by an agency of the United States Government. Neither the United States Government nor any agency thereof, nor any of their employees, makes any warranty, express or implied, or assumes any legal liability or responsibility for the accuracy, completeness, or usefulness of any information, apparatus, product, or process disclosed, or represents that its use would not infringe privately owned rights. Reference herein to any specific commercial product, process, or service by trade name, trademark, manufacturer, or otherwise does not necessarily constitute or imply its endorsement, recommendation, or favoring by the United States Government or any agency thereof. The views and opinions of authors expressed herein do not necessarily state or reflect those of the United States Government or any agency thereof.

DISTRIBUTION OF THIS DOCUMENT IS UNLIMITED

Abstract - Monoenergetic ('slow') positrons are becoming available in the form of beams having brightness sufficiently high for use in microscope and microprobe applications. When slow positrons are injected into certain solid materials they are re-emitted without need for an externally applied extraction field. This property suggests that a positron emission microscope will have useful capabilities and be less severely impaired by problems that extraction fields impose on electron emission devices. The principles involved in a positron emission microscope are discussed, and an estimate of resolutions that might be achieved in the near future is made.

INTRODUCTION

Until recent years, positrons could not be generated in sufficient number, or with adequate definition of energy, to be focused into beams for microprobe or microscope implementation. That situation is changing rapidly, however. With the use of electron linear accelerators (1, 2, 3) and specially prepared radioisotope sources (4), it is now possible to generate currents of monoenergetic positrons of one-hundred picoamps or more. It is known that for energies above 1000 eV the total scattering cross sections of positrons and electrons are about equal. Therefore, for the conventional energies of 50-100 KeV there would appear to be no advantage for a positron microscope over the electron instrument.* For energies less than 10 eV, however, the scattering cross sections are totally different, so we are prompted to consider schemes in which this might be exploited: For certain metals injected positrons are spontaneously re-emitted, with well defined energies, without need for an external extraction field. This suggests that a positron analogue of the field emission microscope might be feasible. The ability to eliminate the extraction field is especially interesting. Perhaps some of the problems that the extraction field imposes on the electron device, such as breakdown of specimens, and image distortion, will be less severe.

MONOENERGETIC POSITRON PRODUCTION

Techniques for moderating positrons to a monoenergetic state involve the use of certain specially prepared solids which have a property known as the negative work function for positrons. The term "negative work function" means that the potential energy of positrons inside the solids is higher than that on the outside. After entering the solids the positrons become thermalized; the majority are annihilated, but a usable number are able to diffuse back to the surface of the solid before this occurs. As they come in range of the potential gradient at the surface they are ejected. At first consideration, the moderation technique would seem very wasteful, at least 99% of the positrons are annihilated inside the moderating solid. But it will be demonstrated in the next section that this is the most effective method available. Even though a great loss in the number of positrons occurs, the brightness of the beam increases.

(*) It is the opinion of the authors (5) that high energy positrons can be employed through Bohrmann measurements (channeling) to extract information about crystalline structures.

MASTER
DISTRIBUTION OF THIS DOCUMENT IS UNLIMITED

Pendyala (6) was the first to study the positron moderation characteristics of tungsten, molybdenum, and many of the other metals. The authors studied many of these metals again (7), and found that moderation efficiency tended to correlate with the magnitude of the negative work function. Tungsten, molybdenum, and iridium, having the more negative work functions, were the most efficient. It was also found that preparation of the moderators could easily be done by heating them in vacuo. It is a fortunate circumstance that the melting points of W, Mo, and Ir are much higher than those of their oxides. By heating these metals at temperatures above the sublimation points of their oxides they can easily be cleaned. The annealing of atomic vacancies and other crystalline defects which trap positrons, and prevent their escape, is also a very important effect of the vacuum heating treatment. The authors have found that annealing will increase the efficiency of tungsten by a factor of five (7).

Several institutions are now producing currents of monoenergetic positrons in excess of 10^8 per second by means of electron linear accelerators (LINAC) (1, 2, 3). Such a facility has recently been completed at Oak Ridge National Laboratory (2). The LINAC produces gamma bremsstrahlung which is intercepted by tungsten moderator plates. The high atomic mass of the tungsten promotes positron-electron pair production. The positrons are then moderated by the tungsten and ejected in a monoenergetic state. High currents of positrons can also be produced by means of specially prepared radioisotope positron sources (4).

THE BRIGHTNESS ENHANCEMENT TECHNIQUE FOR POSITRON BEAMS

It is well known in the various fields of particle beam physics that high brightness, rather than high current, is the most important factor. Brightness is quantified by the equation listed at the bottom of Figure 1, where N is the particle beam current, A is the cross sectional area, α is the solid angle of the beam, E is the energy. For a particle beam that is acted upon by only conservative forces, such as electrostatic or magnetic fields, the B-quantity is conserved. For non-conservative interactions, such as inelastic collisions of the particles, the brightness is changed. For positron beams the inelastic collisions can be orchestrated in such a way as to increase the B-quantity. The diagram on the left of Figure 1 shows a back reflection scheme for accomplishing this. A positron beam having high angular divergence, and an energy of about 3 keV, is shown being focused onto the front side of a thick moderator. Inelastic collisions cause the beam to be thermalized within the solid, and some of the positrons diffuse back to the surface and are expelled. Frieze et al. (8) report back reflection yields of about 20%. Because of the negative work function, the positrons will be expelled with predominantly forward momentum, with a reduced cross section. From the brightness formula at the bottom of Figure 1 it is seen that if the decrease in the energy, the cross sectional area and the angular divergence of the emerging beam offset the decrease in the number of particles, the B-quantity is enhanced. The diagram on the right of Figure 1 shows a brightness enhancement scheme that uses a thin film moderator. The incoming beam is injected on the backside of the film. If the film is sufficiently thin (50-100 nm), the thermalized positrons will diffuse to the front side and be emitted. In this case also, the beam emerging from the front side of the film is brighter than that of the incoming beam if the energy, the cross sectional area and angular divergence are decreased to larger degrees than the number of emerging particles.

Figure 1

Brightness enhancement is the principle that makes practical the moderation of fast positrons to a monoenergetic state. This is the only known technique by which a highly divergent particle beam, having a broad distribution of energy, can be converted to a monoenergetic, lower divergence state.

A PRELIMINARY DESIGN FOR A POSITRON EMISSION MICROSCOPE

To obtain a high resolution positron image it is necessary to effect the emission of high intensities of slow positrons from the area of interest. The brightness enhancement technique can be implemented to accomplish this. Figure 2 shows a design using the back reflection method that is presently under construction in the authors' laboratory. Point emission optics collect and focus the re-emitted positrons. The apparatus consists of a drift chamber that will be held at about 3 kV. The base of the chamber has an aperture into which a flat-tipped tungsten needle, at ground potential, protrudes. An einzel lens at the side of the drift chamber focuses the incoming slow positron beam through a grid onto the tungsten needle. The positrons penetrate the needle and are re-emitted as a brightness enhanced beam of about 2 eV energy. They are accelerated by the aperture field of the drift chamber and projected onto the face of a microchannel plate.

The Figure 2 design will be used initially for low magnification studies of the tungsten moderator surfaces. We will attempt to image grain boundaries and other defects, such as vacancy clusters, that trap positrons and prevent their escape. For the study of thin films, and possibly biological specimens, the Figure 2 design has the drawback that the incoming positron beam has to pass through the specimen. One of the advantages that we hope to obtain with the positron emission microscope is that only very low energy positrons will collide with the specimen. Also, as we attempt to increase magnification, the diameter of the needle will have to be decreased and the usual problems associated with the high field will be encountered. In future designs we will attempt to avoid these problems by using a thin film moderator, such as that of Fig. 1, which is injected with positrons from the rear. The thermalized positrons will migrate to the front side of the film and be emitted and projected to image features in or on the front surface of the film specimen.

Figure 2. Positron Re-emission microscope

L - condensing lens
W - tungsten emission needle
M - microchannel plate

ORNL-DWG 88-11612

To estimate the resolution capabilities for a positron emission microscope we will use the equation derived by Rose (9) for the electron field emission device:

$$\delta = \left(\frac{2\hbar}{mM} \right)^{1/2} \left[1 + \frac{2m\tau v^2}{\hbar M} \right]^{1/2}$$

Where δ is resolution, $\hbar$ is Plank's constant, τ is flight time from emission surface to the microchannel plate surface, M is the geometric magnification of the microscope, m is the mass of the positron, and v is the lateral velocity of the positron as it emerges from the surface. The Rose equation suggests that a positron emission microscope should have a slightly better resolution than its electron counterpart because the lateral velocity of emitted positrons will be lower. When positrons are injected into moderators they go into a band state of the same type as that for electrons. For efficient moderators, such as tungsten, the lowest band state is unbound, i. e., potential energy of positrons in this state is greater than zero. Individual positrons are captured, thermalized, and re-emitted on a time frame of about one picosecond, during which no more than one resides in the band state. Its kinetic energy can be made significantly lower than the conduction electrons if the moderator is operated at reduced temperature.

Assume that we will eventually be able to construct a positron emission microscope with a flight path of about 50 cm and 100 keV overall acceleration potential. Flight time, τ , will be about 2.6 nsec. The magnification that we can use for the microscope will be limited by the available flux of positrons:

$$\frac{F}{M^2} > 4n$$

Where F is the flux, units of $\text{cm}^{-2} \text{sec}^{-1}$, M is the geometric magnification, n is the noise level of the microchannel plate, units of $\text{cm}^{-2} \text{sec}^{-1}$. For a given magnification the intensity of re-emitted positrons reaching the microchannel plate will be decreased by the square of the magnification; this quantity must be greater than the noise level in order to record an intelligible image. For LINAC sources, positrons are delivered in 10-40 nsec bursts, each followed by a dead time of about one millisecond. Typical repetition rates are about 1000 Hz. If the microchannel plate is deactivated during the dead times, and gated on for about a microsecond during the positron pulse delivery, it should be possible to reduce the noise level to about $10^{-3} \text{cm}^{-2} \text{sec}^{-1}$. LINAC sources can deliver slow positrons at an average current of about 10^8 per second at spot sizes of one cm^2 or smaller. If the brightness enhancement is as much as 10% efficient, the re-emitted flux, F , will be about $10^7 \text{cm}^{-2} \text{sec}^{-1}$, and the magnification can be made as large as 50,000X.

If we operate the moderator at liquid nitrogen temperatures, the lateral velocity of the emitted positron will be less than $4.8 \times 10^6 \text{cm sec}^{-1}$. Substituting this value plus a magnification of 50,000X into the Rose equation, we estimate that a resolution of about 10 nm can be achieved. This is comparable to that of a typical scanning electron microscope. The authors have presented this estimate in a previous publication (10).

ACKNOWLEDGEMENT

This research is sponsored by the U.S. Department of Energy, Office of Energy Research, under contract DE-AC05-84OR21400 with Martin Marietta Energy Systems, Inc.

REFERENCES

1. R. H. Howell, I. J. Rosenberg, M. J. Fluss, Appl. Phys. A, 3, 247-255 (1987).
2. L. D. Hulett, Jr., T. A. Lewis, R. G. Alsmiller, Jr., R. Peelle, S. Pendyala, J. M. Dale, T. M. Rosseel, Nuclear Inst. Methods in Phys. Res., B 24/25, 905-908 (1987).
L. D. Hulett, Jr., T. A. Lewis, D. L. Donohue, K. B. Lim, Proc. of the Workshop on Intense Positron Beams, INEL, USA, 18-19 June, 1987, Eds. E. Ottewitte, W. Kells, World Scientific Publishing Co., 687 Hartwell St., Teaneck, NJ.
3. F. Ebel, W. Faust, C. Hahn, S. Langer, H. Schneider, Appl. Phys. A, 44, 119-121 (1987).
4. M. Weber, K. G. Lynn, L. O. Roellig, A. R. Moodenbaugh, Proc. of the Workshop on Intense Positron Beams, INEL, USA, 18-19 June, 1987, Eds. E. Ottewitte, W. Kells, World Scientific Publishing Co., 687 Hartwell St., Teaneck, NJ.
5. J. M. Dale, L. D. Hulett, Jr., S. Pendyala, Applied Spectroscopy Reviews, 19(1) 105-165 (1983).
6. S. Pendyala, P. W. Zitzewitz, J. W. McGowan, P. H. R. Orth, Phys. Lett. A, 43, 298 (1973).
7. J. M. Dale, L. D. Hulett, Jr., S. Pendyala, Surface and Interface Analysis, 2, 199 (1980).
8. W. E. Frieze, D. W. Gidley, K. G. Lynn, Phys. Rev B, 31, (1985) 5628.
9. D. J. Rose, J. Appl. Phys. 27, 215 (1956).
10. L. D. Hulett, Jr., J. M. Dale, S. Pendyala, Materials Science Forum 2, p. 133 (Trans. Tech. Publications Ltd., Ch-4711 Aedermannsdorf, Switzerland, 1984).