C00-1105-240

NOTICE
This report was prepared as an account of work sponsored by the United States Government. Neither the United States nor the United States Energy Research and Development Administration, nor any of their employees, nor any of their contractors, subcontractors, or their employees, makes any warranty, express or implied, or assumes any legal liability or responsibility for the accuracy, completeness or usefulness of any information, apparatus, product or process disclosed, or represents that its use would not infringe privately owned rights.

2

Project Summary for the Years 1973-76

D.W. Pearson and P.R. Moran

Our research of the past three years on new techniques for fast neutron dosimetry has followed three basic lines: fast neutron activation reaction, thermoluminescent dosimeter (TLD) studies with particular relevance to the detection of shallowly penetrating particles and studies of a new solid state dosimetry technique, thermocurrent dosimetry (TCD). Along the way thermoluminescent dosimetry of x-rays was extended in several interesting and useful ways.

Activation detectors have long been a major neutron dosimetry technique, the only one with complete insensitivity to gamma-rays. For fast neutrons the use of several different threshold reactions allows a calculation of the approximate neutron spectrum. A principal deterrent to extensive measurements is the long time often required to count the decay of the neutron induced activity. TL dosimeters can be used as integrators of the decay radiation eliminating the requirement for counters. The application of currently available TL dosimeters to fast neutron activation is discussed in reports C00-1105-186, 211, 213 and 227. This technique is covered in part by U.S. Patent #3,835,329 assigned to USAEC/ERDA.

Attempts to do neutron dosimetry with TL dosimeters date from "ancient times". Differences in the intrinsic response to neutrons have been exploited using the paired dosimeter technique eg., $^6\text{Li}-^7\text{Li}$, $\text{LiF}-\text{CaF}_2$. The other major technique has been to measure the proton recoil dose in a TLD from an hydrogenous cover. A major drawback to the proton recoil methods is the reduced sensitivity of most TLDs to the shallowly penetrating, high LET recoil protons. The recog-

nition of a different distribution of traps near the surface of LiF TLDs was reported in C00-1105-187. Characterization of the surface traps permits more accurate dosimetry of weakly penetrating beta rays from tritium and low energy electrons from electron microscopes. These studies, presented in reports C00-1105-234 and 239, have been carried out using electron irradiation. Electrons are a most convenient probe, being easy to control and inexpensive to produce. The knowledge obtained from these electron irradiation studies is directly extendable to proton recoil TL measurements. We are now in a position to evaluate surface modified TLDs for the dosimetry of weakly penetrating radiation, both electrons and recoil protons.

The rapid development of the field of Thermocurrent Dosimetry (TCD) has been particularly exciting. TCD has the potential to replace TLD as a general solid state dosimetry technique in many applications. Commercial development of TCD with sapphire dosimeters is being pursued by the Harshaw Chemical Co. Characterization of sapphire TC dosimeters was first reported in C00-1105-207 and 208. The advantages of TCD include the fact that the readout is a measurement of a current or a charge which makes calibration relatively easy. Therefore, TCDs have good potential as secondary radiation dosimetry standards. Additionally, TCDs do not require optically active impurity centers and can in fact be opaque or covered with opaque materials. The requirement for high purity to eliminate background ionic currents also improves the sensitivity of TCDs as indicated in C00-1105-206.

The initial characterization of sapphire TCDs was done with x-rays for convenience. We are now beginning neutron response measurements. The initial experiments with 14MeV neutrons are reported in C00-1105-236. Preliminary results

MASTER

DISTRIBUTION OF THIS DOCUMENT IS UNLIMITED

DISCLAIMER

This report was prepared as an account of work sponsored by an agency of the United States Government. Neither the United States Government nor any agency Thereof, nor any of their employees, makes any warranty, express or implied, or assumes any legal liability or responsibility for the accuracy, completeness, or usefulness of any information, apparatus, product, or process disclosed, or represents that its use would not infringe privately owned rights. Reference herein to any specific commercial product, process, or service by trade name, trademark, manufacturer, or otherwise does not necessarily constitute or imply its endorsement, recommendation, or favoring by the United States Government or any agency thereof. The views and opinions of authors expressed herein do not necessarily state or reflect those of the United States Government or any agency thereof.

DISCLAIMER

Portions of this document may be illegible in electronic image products. Images are produced from the best available original document.

indicate that the response of sapphire TCDs to recoil protons is comparable to their response to x-rays. Further investigations, particularly on thin dosimeters, seem promising.

The other class of materials with exciting potential as TCDs are high temperature polymers. Polymers are relatively inexpensive and easy to fabricate into complex shapes. The potential for obtaining near tissue equivalence (TE) exists through copolymerization or mixtures of polymers. For dosimetry of mixed neutron and gamma ray fields, paired dosimeters of fluorocarbon PTFE teflon^R and hydrogenous polymethyl pentene have been used. Measurements of the fast neutron responses of polymer TCDs are reported in C00-1105-235.

The somewhat low sensitivity of present polymer TCDs limits their use to therapy level doses. The sensitivity is strongly correlated with the microstructure of the polymers. It should be possible to increase this sensitivity considerably by proper polymer growth conditions. The non-radiation induced background currents appear to be related to impurities and we hope these can be reduced considerably. We were very fortunate to have Professor M.W. Harper or Velindre Hospital, Cardiff, Wales visit us to begin a cooperative study. We look forward to continued collaboration.

Also included in the last three years work are practical improvements to the most widely used TL dosimeters: LIF. A sensitization technique which increases the low level sensitivity of LIF dosimeters up to 8 fold is reported in C00-1105-224. A related phenomenon -- the altering of the x-ray energy dependence of the TL from LIF is reported in C00-1105-222.

One important aspect of any University research project is the education and training of students. In the past three years three students have received Ph.D. degrees supported by this contract.

The experiments on solid state ionography outlined in reports C00-1105-232 and -233 were unanticipated in the last grant proposal. However, they follow logically from studies of the trapped charge yields and carrier mobilities in sapphire TCDs. Parallel work, supported by other sources (N.C.I.), suggested the potential for x-ray imaging. The relevance of these experiments to the goals of the Division of Biomedical and Environmental Research is obvious. Report C00-1105-232 summarizes the potential for producing images by charge collection during irradiation and shows actual images produced by solid state ionography. The high resolution of this process is demonstrated. The studies summarized in Report C00-1105-233 outline the physical properties required for solid state ionographic image formation. The basic requirement is for a good insulator with a high radioconductivity in the temperature region desired.

Dr. Gene E. Fuller's thesis is presented as report C00-1105-203. This electron spin resonance study of the self trapped hole in CaF_2 provides an independent look at the dynamics of the motion of the trapped charges responsible for TL and TC phenomena. These measurements give a quantitative determination of the trapped charge yield from x-rays which cannot be obtained by TL or TC studies.

Dr. G.D. Fullerton's thesis is presented as report C00-1105-217. This thesis characterizes sapphire TC dosimeters. The commercial development of sapphire TC dosimeter is now being investigated by the Harshaw Chemical Co.

Dr. D.W. Pearson's thesis is presented as report C00-1105-227. This thesis describes the fast neutron activation of various TL dosimeters.